

# Primary Surface Ozone Standard Reference Photometers and BIPM.QM-K1

J. Viallon, BIPM




Accurate Monitoring of Surface Ozone  
Virtual Workshop  
6 October 2020


# All SRPs are equivalent and linked to the SI



The ozone **absorption cross-section** is what links the SRP to the SI




# BIPM mission since 2002 : improved accuracy



Gas Phase Titration :  $\text{NO} + \text{O}_3 \rightarrow \text{NO}_2$

- 2002 : the BIPM installs 3 SRPs and starts a Gas Phase Titration experiment
- 2005 : BIPM and NIES GPT measurements during CCQM-P28 :  $\sim 2\%$  bias
- 2007: launch of BIPM.QM-K1
- 2012 : laser-based SRP to perform relative measurements of the absorption cross-section
- 2015 : absorption cross-section measurements in pure ozone results in lower value than Hearn
- 2016 : improved GPT with traceability to NO and NO<sub>2</sub> agrees better with 2015 value


# Metrological Traceability of surface ozone measurements



# Comparisons at the BIPM since 2003

2003-2005 : Pilot Study CCQM-P28, 23 participants

Since 2007 : On-going Key Comparison BIPM.QM-K1



# Comparisons of SRPs at the BIPM

## Comparison Protocol


**One comparison** : 2 SRPs measuring different  $O_3$  concentration from the same source

**Range of  $O_3$**  : 0 to 500 nmol/mol

**One main instrument** : SRP27

**Number of points** : 10


## Triad of SRPs maintained at the BIPM



# Uncertainties (of BIPM maintained SRPs)

Standard uncertainty for the range of one ozone amount fractions measured during a comparison of SRPs

$$x(\text{O}_3) = \frac{-1}{L\sigma(\lambda)} \frac{R}{N_a} \frac{T}{P} \ln(D)$$



**Major uncertainty:** absorption cross-section, currently 1.06 % (standard)

**Light path length:** 0.3 % (increased in 2006 after BIPM work on biases<sup>1</sup>)

**Pressure:** 0.04 %


**Temperature:** 0.03 %

<sup>1</sup>Viallon J., Moussay P., Norris J.E., Guenther F.R., Wielgosz R.I., [A study of systematic biases and measurement uncertainties in ozone mole fraction measurements with the NIST Standard Reference Photometer, 2006 Metrologia 43 441](#)

# Uncertainties (of BIPM maintained SRPs)

Standard uncertainty for the range of one ozone amount fractions measured during a comparison of SRPs

$$x(\text{O}_3) = \frac{-1}{L\sigma(\lambda)} \frac{R}{N_a} \frac{T}{P} \ln(D)$$



Major uncertainty: absorption cross-section, currently 1.06 % (standard)

No uncertainty on the absorption cross-section when comparing instruments (fully correlated component)

Note for calibration : include  $u(\sigma)$  just once!



# Degrees of Equivalence

By definition:  $D = x_{NS} - x_{RS}$


RS = Reference Instrument = BIPM-SRP27 by convention

$$u(D) = \sqrt{u(x_{NS})^2 + u(x_{RS})^2}$$

10 nominal ozone amount fractions measured during a comparison :

Nominal value	Reference Standard (RS)			National standard (NS)		
	$x_{RS}$ nmol/mol	$s_{RS}$ nmol/mol	$u(x_{RS})$ nmol/mol	$x_{NS}$ nmol/mol	$s_{NS}$ nmol/mol	$u(x_{NS})$ nmol/mol
0	0.01	0.22	0.28	-0.29	0.27	0.24
220	206.50	0.22	0.66	206.74	0.27	0.53
80	76.63	0.17	0.36	76.44	0.20	0.35
420	407.25	0.44	1.22	407.58	0.33	0.81
120	119.08	0.18	0.45	119.08	0.30	0.41
320	298.71	0.18	0.92	298.84	0.31	0.66
30	29.49	0.17	0.29	29.42	0.17	0.28
370	346.30	0.24	1.05	346.77	0.31	0.73
170	163.88	0.24	0.55	163.97	0.30	0.47
500	480.25	0.42	1.43	480.74	0.36	0.91
270	251.15	0.09	0.79	251.39	0.30	0.59
0	-0.08	0.23	0.28	-0.23	0.29	0.24


2 out of 10 chosen to calculate Degrees of Equivalence



Values chosen as representative of the range  
(CCQM/GAWG Workshop 2005)

# On-going comparison since 2007

One bilateral comparison -> one report + results in KCDB  
( $D_i$  at 80 and 420 nmol/mol)



Maintained agreement  
No trend, stability maintained over years

# Results in the BIPM Key Comparison Data Base



<https://www.bipm.org/kcdb/comparison?id=1428>

All data listed in the KCDB have been reviewed and approved within the CIPM Mutual Recognition Arrangement

Home > Comparisons

INFORMATION   GRAPHS OF EQUIVALENCE   DEGREES OF EQUIVALENCE   EQUIVALENCE STATEMENTS   LABORATORY MEASUREMENTS

BIPM comparison

**BIPM.QM-K1**



→ [BIPM.QM-K1](#)

Linked comparison

INFORMATION & CONTACT   PARTICIPANTS

Metrology area, Sub-field: **Chemistry and Biology, Gases**  
 Description: **Ozone at ambient level**  
 Time of measurements: **2007 -**  
 Status: **Continuous, approved for equivalence**


Measurand: **Amount fraction of ozone in dry air  
 80 nmol/mol and 420 nmol/mol**  
 Transfer device: **Measurements made at the BIPM by comparison with the BIPM Reference Photometer SRP27**

Comparison type: **Key Comparison**  
 Consultative Committee: **CCQM (Consultative Committee for Amount of Substance)**  
 Conducted by: **BIPM (Bureau International des Poids et Mesures)**

Comments: **First results published on 09 April 2008  
 The Final Report for each bi-lateral comparison is listed in the tab "Participants"  
 Last update: 5 June 2020**


## Some rules

- Maximum period between repeats of 8 years
- Laboratories can send their national standard or a transfer standard



# Laboratory performance in Ozone comparisons since 2002

Looking at the spread amongst standards in International Comparisons versus the SRP relative uncertainty



performance maintained until now

# Acknowledgements

BIPM Staff performing the comparison measurements  
F. Idrees, P. Moussay

All participants in BIPM.QM-K1

The CCQM / Gas Analysis Working Group  
<http://www.bipm.org/en/committees/cc/wg/gawg.html>

