

# CCQM-K137, Nitrogen Monoxide (NO) in nitrogen, 30 $\mu\text{mol mol}^{-1}$ and 70 $\mu\text{mol mol}^{-1}$ Support for Calibration and Measurement Capabilities

---

## 1 Preamble

---

This guidance note is aimed at reviewers of calibration and measurement capabilities, supported by the participation in a key comparison. In principle, support to measurement capabilities is limited to those measurement results that are consistent with the key comparison reference value (KCRV). In this key comparison [1], several measurement results were not consistent with the KCRV. For those results, this guidance note provides larger expanded uncertainties, based on the GAWG strategy document [2]. The idea behind these larger uncertainties is that

- a) National Metrology Institutes (NMIs) can still use their participation in a key comparison to support their measurement service;
- b) The stated uncertainty is large enough to ensure comparability with the KCRV and the results of other NMIs;
- c) There is a harmonised way of dealing with discrepant results in relation to CMCs.

Discrepant measurement results can occur for a number of reasons. For a discussion of the measurement result in CCQM-K137, see the final report [1]. In case of incidental discrepant results, the default response would be to investigate the cause of the discrepancy and to resolve it [3]. Hence, the attached table should not be viewed as

- a) A substitute for appropriate corrective measures from the side of the NMI to resolve the discrepancy;
- b) A consent from the GAWG that the submitted measurement result is acceptable;
- c) A guarantee that a CMC submitted in accordance with this guidance note will be accepted by reviewers in the review process by the Regional Metrology Organisations;
- d) Support for the metrological traceability of the measurement result submitted;
- e) A direction or recommendation to assessors in peer reviews or accreditation visits.

## 2 Support for CMCs

---

Tables 1 and 2 show the ranges of the amount fractions and the expanded uncertainties supported by participation in CCQM-K137 [1], calculated in accordance with the GAWG Strategy document [2]. Amount fractions below 10  $\mu\text{mol mol}^{-1}$  are supported by results at 30  $\mu\text{mol mol}^{-1}$  nominal value (Table 1), while the ranges above 10  $\mu\text{mol mol}^{-1}$  are supported by results at 70  $\mu\text{mol mol}^{-1}$  nominal value (Table 2). To avoid a step change in the uncertainty at 10  $\mu\text{mol mol}^{-1}$ , the lower limit for the second range is fixed by the data from the comparison at 30  $\mu\text{mol mol}^{-1}$ .

This does not apply to VNIIM and VSL as they did not participate in the comparison at both compositions. For VNIIM, the uncertainty for the 30  $\mu\text{mol mol}^{-1}$  mixture has been used to cover the full range. For VSL, a value of 0.45% relative has been used to cover the lower range and reflects supportive evidence provided.

Table 1: Supported ranges and expanded uncertainties for amount fractions below 10  $\mu\text{mol mol}^{-1}$

|              | Amount Fraction ( $\mu\text{mol mol}^{-1}$ ) |    | Expanded Uncertainty (% Rel) |       |
|--------------|--|----|------------------------------|-------|
|              | From   | To | From                         | To    |
| <b>BFKH</b>  | 0.419  | 10 | 100                          | 4.187 |
| <b>CERI</b>  | 0.032  | 10 | 100                          | 0.316 |
| <b>GUM</b>   | 0.499  | 10 | 100                          | 4.986 |
| <b>KRISS</b> | 0.067  | 10 | 100                          | 0.667 |
| <b>LNE</b>   | 0.022  | 10 | 100                          | 0.225 |
| <b>NIM</b>   | 0.050  | 10 | 100                          | 0.501 |
| <b>NMIA</b>  | 0.087  | 10 | 100                          | 0.866 |
| <b>NMISA</b> | 0.056  | 10 | 100                          | 0.562 |
| <b>NPL</b>   | 0.020  | 10 | 100                          | 0.200 |
| <b>VNIIM</b> | 0.004  | 10 | 100                          | 0.040 |
| <b>NIST</b>  | 0.012  | 10 | 100                          | 0.121 |
| <b>VSL</b>   | 0.045  | 10 | 100                          | 0.450 |

Table 2: Supported ranges and expanded uncertainties for amount fractions above 10  $\mu\text{mol mol}^{-1}$

|              | Relative Regime                            |     |                              |       |
|--------------|--|-----|------------------------------|-------|
|              | Amount Fraction ( $\text{mmol mol}^{-1}$ ) |     | Expanded Uncertainty (% Rel) |       |
|              | From                                       | To  | From                         | To    |
| <b>BFKH</b>  | 10   | 500 | 4.187                        | 1.752 |
| <b>CERI</b>  | 10   | 500 | 0.316                        | 0.320 |
| <b>GUM</b>   | 10   | 500 | 4.986                        | 4.085 |
| <b>KRISS</b> | 10   | 500 | 0.667                        | 0.461 |
| <b>LNE</b>   | 10   | 500 | 0.225                        | 0.229 |
| <b>NIM</b>   | 10   | 500 | 0.501                        | 0.501 |
| <b>NMIA</b>  | 10   | 500 | 0.866                        | 0.443 |
| <b>NMISA</b> | 10   | 500 | 0.562                        | 0.300 |
| <b>NPL</b>   | 10   | 500 | 0.200                        | 0.143 |
| <b>VNIIM</b> | 10   | 500 | 0.040                        | 0.040 |
| <b>NIST</b>  | 10   | 500 | 0.121                        | 0.185 |
| <b>VSL</b>   | 10   | 500 | 0.450                        | 0.300 |

### 3 Bibliography

- [1]. J Viallon, E Flores, F Idrees, P Moussay, R I Wielgosz, S H Oh, S Lee, B M Kim, G Nieuwenkamp, A Van der Veen, O V Efremova, L A Konopelko, Y A Kustikov, A V Kolobova, H Shuguo, J Carney, M E Kelley, G C Rhoderick, J T Hodges, S Uehara, D Akima, P J Brewer, D Worton, S Van Aswegen, T Macé, D Smeulders, J Fükö, N Ntsasa, N Leshabane, K Ramahala, J Tshilongo, D Cieciora, T Shimosaka and N Matsumoto, Metrologia, (2020), 57 08001..
- [2]. P J Brewer, A M H van der Veen, GAWG strategy for comparisons and CMC claims. Sèvres : Gas Analysis Working Group, 2016.
- [3]. ISO. ISO/IEC 17025 General requirements for the competence of testing and calibration laboratories. Geneva : ISO, 2017.