

Report from the BIPM electricity laboratories

M. Stock

CCEM meeting 6-7 March 2025




Staff of the Physical Metrology Department





(CCEM, CCU)







José Angel **MORENO Physicist**



Benjamin ROLLAND Assistant (Engineer)

Voltage



Dr Stéphane SOLVE **Principal** Physicist (CCT)



Régis **CHAYRAMY Principal Technician**

Mass

Kibble balance



Dr Hao **FANG Principal Physicist** (CCM)



Dr Franck BIELSA Principal Physicist (CODATA TGFC)



Adrien **KISS Assistant** (Engineer)

Mass calibr. (also pressure, density)



Pedro CONCEIÇÃO **Principal** Technician

Five BIPM ongoing comparisons



Two on-site comparisons of quantum standards

BIPM.EM-K10 On-site comparison of dc and ac voltages (≤ 1 kHz) from JVSs

BIPM.EM-K12 On-site comparison of QHR standards and resistance scaling

Three bilateral comparisons using conventional BIPM transfer standards

BIPM.EM-K11 voltage at 1.018 V and 10 V

BIPM.EM-K13 resistance at 1 Ω and 10 $k\Omega$

BIPM.EM-K14 capacitance at 10 pF and 100 pF, at 1592 Hz and/or 1000 Hz

Five BIPM ongoing comparisons



Two on-site comparisons of quantum standards

BIPM.EM-K10 On-site comparison of dc and ac voltages (≤ 1 kHz) from JVSs

BIPM.EM-K12 On-site comparison of QHR standards and resistance scaling


Three bilateral comparisons using conventional BIPM transfer standards

BIPM.EM-K11 voltage at 1.018 V and 10 V

BIPM.EM-K13 resistance at 1 Ω and 10 $k\Omega$

BIPM.EM-K14 capacitance at 10 pF and 100 pF, at 1592 Hz and/or 1000 Hz

On-site comparison of ac voltages from JVSs (K10) | BIPM



Program of Pilot Studies (2015-2024)



NMIJ - 2015

CENAM - 2016

PTB - 2017, 2018, 2021, 2022, 2023, 2024 with the support of KRISS, NMIA and VNIIM

NPL - 2018

KRISS – 2018, 2022, 2023

NIST – 2024 with support of KRISS



















The BIPM Transportable PJVS



- The BIPM Transportable PJVS is a specially designed compact version of the NIST PJVS (10 V and 2 V) operated in liquid helium
- The array can be biased with three different bias sources (NIST, KRISS, Supracon)
- The system is flexible in timing can trigger or can be triggered
- The BIPM differential sampling setup can operate three different samplers (Keysight 3458A, Fluke 8588A, NI-PXI 5922)
- The BIPM sampling setup has been tested with different softwares (NIST, PTB, KRISS)

BIPM AC voltage transfer standard (sine wave generator)





Jan Kucera et al. – CPEM 2022


Improvements of CMI's digital synthesizer SWG as an AC voltage transfer standard

Pilot Study PTB - BIPM - 2024 at BIPM




Large discrepancy 2023: Assumption of a leakage effect on the BIPM side

- Repetition of the 2023 exercise using the same NI-5922 sampler and PTB software
- with a 2 V NIST array on the BIPM system
- Testing two different bias sources on the BIPM system: NIST and SupraconTM bias source




Pilot Study – NIST – August 2024 at NIST











BIPM-NIST Pilot Study - 2024



Overnight Measurements


Aperture of the windows (PJVS)

Delay Time (PJVS)

Phase shift (JAWS)

Amplitude changes (JAWS)

.... and more...



 $T_{STEP} = 1 / N \cdot f$

Typical outcome of the most recent pilot studies | | BIPM



Freq./Hz	N	difference/nV	typ. Type A unc./nV
62.5	64	5	30
125	64	7	20
250	32	44	20
625	32	100	20
1250	16	200	30

Type B uncertainty under evaluation

Type A: 2 to 5 parts in 10⁸ demonstrated for 0.75 V - 1 kHz

Parameters for the

0.7 V and 7 V rms

• f/Hz: 10, 50, 60, 62.5, 100, 312.5, 625, 976.5625, 1000, 1250







On-site comparison of ac voltages from JVSs




- For first time agreement at 10 Hz within a few 10^{-9} reached at 2 V_{rms} agreement at 1 kHz within 2 x 10^{-7}
- Technical protocol approved by CCEM support group in 2022
- Program of pilot studies now completed, KC officially launched, first two comparisons planned in 2025 with PTB and CENAM, in 2026 with NIM and ...
- Parameters:
 - 0.7 V and 7 V rms
 - o *f*/Hz : 10, 50, 60, 62.5, 100, 312.5, 625, 976.5625, 1000, 1250
- Questionnaire on interest (2023): 21 for comparison at dc, 16 for ac

On-site quantum Hall resistance key comparisons (BIPM.EM-K12)









No comparisons 2023-2024 due to liquid He shortage and other problems at NMIs


Planned comparisons were cancelled 3 times on request of the NMIs

Now planned for 2025: PTB and CEM

Questionnaire on interest (2023): 14 NMIs interested (until 2030)




On-site quantum Hall resistance key comparisons (BIPM.EM-K12)



Installation of a He liquefier






- to provide access to liquid He
- to protect He resources
- to reduce cost
- capacity:
 - ➤ 150 liter of liquid He
 - liquefying 25 liters/day

Bilateral comparisons using transfer standards






Comparison of capacitance EURAMET.EM-K4



- Coordinator: PTB
- Measurand:
 - 10 pF at 100 V and 1592 Hz
 - 100 pF at 10 V and 1592 Hz
- High number of participants (21)
- Organized in one inner loop and three outer loops
- BIPM participates as:
 - Member of the support group
 - Center of the inner loop with the three subpilots of the comparison
- Measurements of 12 capacitors simultaneously from middle April 2025 for four weeks.




Calibrations



voltage: 1.018 V, 10 V on average 6 per year

resistance: 1Ω , 100Ω , $10 k\Omega$ 38 per year

capacitance: 1 pF, 10 pF, 100 pF 30 per year



Largest number of calibrations ever (97)

4 new NMIs as customer

Trend towards more calibration requests due to more Member States


Study of the stability and performance of graphene devices



The use of a graphene-based quantized Hall resistor would simplify the BIPM.EM-K12 comparison and reduce its cost

- ⇒ Preparation to the transition from GaAs-type QHR to graphene-type QHR by characterizing devices and study their long-term stability
- ⇒ This work is carried out in collaboration with the PTB, providing epitaxial graphene-based devices to the BIPM
- ⇒ BIPM is a member of CCEM TG-Graphene







Study of the stability and performance of graphene devices



Characterization of PTB's graphene devices and study of their long term stability





Quantization of graphene-based device G33 within 2 $n\Omega/\Omega$ with low carrier density n and low s parameter

Study of long-term stability of the carrier density to be continued ...

Investigation of frequency dependence of 1 Ω resistors



Objective: find a 1 Ω standard with low or negligeable frequency variation at low frequency, between 1 Hz and dc typically, which could be used for **BIPM.EM-K12** comparisons.

- → Previous study in 2014 with most used wire wound resistance standards of type: CSIRO, L & N, Tinsley, Thomas and MI9210A.
- → This study was repeated for the recently commercialized Alpha resistors only model HRU-1R0 including prototypes having different internal wiring configuration

Collaborative work between the NMIJ, PTB and BIPM

- → Alpha HRU-1R0 resistance standards were provided by NMIJ
- → The measurements were caried at BIPM using both PTB and BIPM CCC bridges and BIPM 1 Hz bridge

Investigation of frequency dependence of 1 Ω




Comparison of the 1 Hz-dc differences of the wire wound and Alpha resistance standards

2023 study

resistors

R#	Type / SN
1	F067A
2	F068A
3	F069A
4	F112A
5	Prototype – T002
6	Prototype – T004
7	Prototype – T006
8	Prototype – T007




2014 study


R #	Туре	
1	L & N 4210	
2	L&N 4210	
3	L&N 4210	
4	CSIRO	
5	CSIRO	
6	CSIRO	
7	CSIRO	
8	CSIRO	
9	Thomas NBS	
10	Tinsley	
11	MI 9210A	
12	MI 9210A	
13	MI 9210A	

With the exception of one, all the 1 Ω Alpha resistors we tested have a 1 Hz-dc difference lower or equal to the best resistor measured in 2014

 \Rightarrow possibility of selecting resistors with a frequency dependence close to zero difference.


Calculable capacitor






- Diagonal distances within ± 1 µm on average.
- Angles adjusted at 90° within ± 0.3 mrad.

In collaboration with Dr. Almazbek Imanaliev of LNE





Diagonal displacement of one electrode

To do:

- Investigation of other important defects (on progress)
- Accuracy versus time consuming optimization

Overview of the modernization of the equipment

Recently done:

- new low frequency RT current comparator
- new CCC
- new QHR and CCC cryoprobes
- renewal of oil-bath for reference capacitors

On progress:

 final characterization of a new 1 Hz bridge for the BIPM.EM-K12 comparison

Planned:

- dual digital current source to equip a new CCC bridge
- R and C standards for quadrature bridge
- quadrature bridge including an ac-QHR






Renewed oil-bath





New 1 Hz bridge






Characterization of ceramic capacitors

The BIPM Kibble balance








- magnetic flux density 0.47 T
- bifilar coil (each 26 layers & 1400 turns)
- current 10 mA for a 1 kg mass
- standard resistor 100 Ω
- voltage drop 1 V
- velocity 1 mm/s
- induced voltage 0.5 V



Participation in CCM.M-K8.2021 – KC of kg realizations (completed 2023)






Uncertainty budget in parts in 10^9 ($k = 1$).		
Type A	10	
Alignment	(33)	
Velocity	7	
Voltage (mainly electrical grounding)	14	
Standard resistor	9	
Free-fall gravity acceleration	7	
Others	10	
Total	41	

- Standard uncertainty 4.1 x 10⁻⁸ dominated by parasitic coil motion in velocity mode → new beam mechanism
- At present preparing for participation in CCM.M-K8.2024

Other Kibble balance developments





- Determination of gravity difference between LNE and BIPM in 2024
- At present confirmation with an absolute atomic gravimeter at BIPM



Development of new beam mechanism:

- Performance improvement of existing KB
- Prototype for a compact KB





 Development of a home-made voltmeter, optimized for KB operation (one single device instead of three, voltage in sub-mV range)

Outlook into next PoW 2028-2031



Guiding principles:

- Maintain continuity with present program so that NMIs can rely on program of comparisons and calibrations
- Adapt program to new developments

Proposed activities:

- Continue the new program of comparisons of PJVSs at ac
- Launch/continue comparisons of QHR standards using graphene samples
- Development of acQHR as primary standard for impedance, for reduced uncertainty in capacitance
- Continue the provision of calibration services for voltage, resistance, capacitance for NMIs without quantum standards
- Any new ideas ?
 - Other resistance values?
 - o ...her resistance values?

