

International comparison of activity measurements
of a solution of ^{137}Cs (May 1982)

Preliminary report

by A. Rytz

November 1982

Bureau International des Poids et Mesures
F-92310 SEVRES

The present report contains information enabling the participants to see how their own results are situated with regard to the others. It was found convenient to present the data in tabular form, without any discussion or interpretation. In a few cases minor omissions were unavoidable for reasons of space.

A full report with a thorough analysis of all the data will be published at a later date.

Table 1

List of participants

| | |
|-------|---|
| AAEC | Australian Atomic Energy Commission, Sutherland, Australia |
| AECL | Atomic Energy of Canada Limited, Chalk River, Canada |
| BCMNI | Central Bureau for Nuclear Measurements, Geel, Belgium |
| BIPM | Bureau International des Poids et Mesures, Sèvres, France |
| ETL | Electrotechnical Laboratory, Ibaraki, Japan |
| IER | Institut d'électrochimie et radiochimie de l'EPFL, Lausanne, Switzerland |
| IFIN | Institut of Nuclear Physics and Engineering, Bucharest, Romania |
| LMRI | Laboratoire de Métrologie des Rayonnements Ionisants, Saclay, France |
| NAC | National Accelerator Centre, Faure, South Africa |
| NBS | National Bureau of Standards, Washington, D.C., U.S.A. |
| NIM | National Institute of Metrology, Beijing, China |
| NPL | National Physical Laboratory, Teddington, U.K. |
| NRC | National Research Council, Ottawa, Canada |
| OMH | Országos Mérésügyi Hivatal, Budapest, Hungary |
| PDS | National Atomic Energy Agency, Jakarta, Indonesia |
| PTB | Physikalisch-Technische Bundesanstalt, Braunschweig, Federal Republic of Germany |
| SCK | Studiecentrum voor Kernenergie, Mol, Belgium |
| UVVVR | Institute for research, production and application of radioisotopes, Prague, Czecho-Slovakia |

Table 2

| Laboratory | Method | Dilution | Mixing ratio $^{137}\text{Cs}/^{134}\text{Cs}$ | Sources | | β detector | | γ detector resol. at 662 keV (%) | Dead times* | | Typical count rates | |
|------------|-------------------------------|----------|---|--------------------|---|------------------|-------------------|--|---|--|---|--|
| | | | | Mass range (mg) | Backing mass ($\mu\text{g cm}^{-2}$) | Gas | Pressure (MPa) | | τ_{β} τ_{γ} τ_{c} (μs) | N_{β} N_{γ} N_{c} (s^{-1}) | | |
| AAEC | $4\pi\beta(\text{PC})-\gamma$ | yes | 0.48 | 23 - 63 | 30 | CH_4 | 0.1 | 8.8 | 9.15 (8) 18.2 (1) 1.145 (6) | | 7 000 75; 150; 0.6 65; 140; 0.45 | |
| AECL | " | no | 1.18 | 12 - 27 | 12 | " | " | 7.3 - 7.6 | 1.935 (21) 1.916 (20) 0.698 3 (19) | | 12 900 630; 450 560; 410 | |
| BCM | " | no | 0.6 - 1.4 | 7 - 15 | 50 | " | " | 6.4 | 5.97 (6) 6.09 (6) 0.785 (16) | | 14 000 - 19 000 500 - 800 400 - 600 | |
| BIRM | " | no | 1.08 | 26 - 96 | 60 | " | " | 7.5 | 4.432 (10) 4.425 (10) 1.06 (1) | | 30 000 800 670 | |
| ETL | " | no | 1.00 | 9 - 19 | 30 | " | " | 7.5 - 8.0 | 4.35 (5) 2.11 (5) 0.688 7 (28) | | 7 000 150; 250 135; 225 | |
| IER | " | yes | 1.0 | 20 - 70 | 50 | " | " | 9.6 | 3.201 (1) 3.196 (1) 1.075 (2) | | 16 800; 18 000 190; 293 168; 240 | |
| IFIN | " | no | 0.64 | 5 - 11 | 100 - 150 | " | " | 12.8 | 10.0 (5) 10.0 (5) 1.090 (5) | | 8 000 150 120 | |

* Between brackets: Uncertainty in units of the last digit

Table 2 (cont'd)

| Laboratory | Method | Dilution | Mixing ratio $^{137}\text{Cs}/^{134}\text{Cs}$ | Sources | | β detector | | γ detector resol. at 662 keV (%) | Dead times* | | Typical count rates | |
|------------|-----------------------------------|----------|---|--------------------|---|-------------------------|-------------------|--|---|--|---------------------|--|
| | | | | Mass range (mg) | Backing mass ($\mu\text{g cm}^{-2}$) | Gas | Pressure (MPa) | | τ_{β} τ_{γ} τ_{C} (μs) | N_{β} N_{γ} N_{C} (s^{-1}) | | |
| IMRI | $4\pi\beta(\text{PC})-\gamma$ | no | 1.36; 0.55 | 15 - 27 | 40 | CH_4 | 0.1 | 0.3 [Ge(Li)] | 5 ** 5 ** ≈ 1 | 14 000 40 36 | | |
| NAC | $4\pi\beta(\text{LS})-\gamma$ | " | - | - | - | - | - | 9.5 | 1.15 (2) 1.19 (3) 0.513 (1) | 32 600 197 161 | | |
| NBS | $4\pi\beta(\text{PPC})-\gamma$ AC | yes | 1.5; 0.67 | 21 - 34 | 30 | Ar/CH_4 | 1.48 | 7.1 | 4.3 10.0 (2) - | 6 000 155; 170 14 - 30 | | |
| NIM | $4\pi\beta(\text{PC})-\gamma$ 1 | " | 1.09; 1.03 | 7 - 13 | 20 | CH_4 | 0.1 | 8 | 5.98 (1); 21.901 (1) | 8 000; 11 695 | | |
| | " 2 | " | 0.9 | 10 - 15 | 20 | " | " | 7.8 | 1.52 (1); 21.869 (1) | 90; 328 | | |
| | $4\pi\beta(\text{LS})$ | no | - | 10 - 15 | - | - | - | - | 1.525(6); 1.2728(1) | 80; 226 | | |
| NPL | $4\pi\beta(\text{PC})-\gamma$ | yes | 1.0 | 14 - 15 | 30 | Ar/CH_4 | 0.1 | 8.3 | 1.54 (1); 18.70 (5) | 15 000; 4 300 | | |
| | $4\pi\beta(\text{LS})-\gamma$ | " | 1.01 | 15 - 30 | - | - | - | 9.8 | 2.80(50); 3.27 (4) 0.591(5); 0.300 (5) | 2 200; 75 1 600; 65 | | |
| NRC | $4\pi\beta(\text{PPC})-\gamma$ AC | no | 0.95 | 13 - 16 | 40 | Ar/CH_4 | 1.14 | 8.4 | 5.08 (5) 2.00 (2) - | 55; 17 5-15; 3-7 0.5 | | |

* Between brackets: Uncertainty in units of the last digit

** Cumulative dead time, common to both channels

Table 2 (cont'd)

| Laboratory | Method | Dilution | Mixing ratio $^{137}\text{Cs}/^{134}\text{Cs}$ | Sources | | β detector | | γ detector resol. at 662 keV (%) | Dead times* | | Typical count rates | |
|------------|---|----------|---|-------------------------|---|---|-------------------|--|---|--|--|--|
| | | | | Mass range (mg) | Backing mass ($\mu\text{g cm}^{-2}$) | Gas | Pressure (MPa) | | τ_{β} τ_{γ} τ_{c} (μs) | N_{β} N_{γ} N_{c} (s^{-1}) | | |
| OMH | $4\pi\beta(\text{PC})-\gamma$ | no | 0.5 1 | 3 - 16 | 30 | CH_4 | 0.1 | 8.2 | 3.067 3.021 1.022 | (5) (5) (10) | 11 500; 10 500 350; 230 290; 200 | |
| PDS | $\text{Ge}(\text{Li})\gamma$ | - | - | - | - | - | - | - | - | - | - | |
| PTB | $4\pi\beta(\text{PC})-\gamma$ $4\pi\beta(\text{PFC})-\gamma$ 1 $4\pi\beta(\text{PFC})-\gamma$ 2 | no | 0.97 " " | 5 - 60 " " | 60 " " | CH_4 Ar/CH_4 " | 0.1 1.1 1.1 | 6.6 6.6 8.9 | 5.06 4.73 1.01 | (5) (5) (2) | 7 700; 6 600; 7 000 170; 103; 650 160; 75; 550 | |
| SCK | $4\pi\beta(\text{PC})-\gamma$ $4\pi(\text{NaI})\gamma$ | yes " | 0.85 - | 4.5 - 9.5 7.5 - 11.3 | 50 - | CH_4 " | 0.1 - | 8.8 7.5 | 2.48 2.46 1.014 3 | (1) (1) (3) | 8 300 - 295; 4 000 251 - | |
| UVVVR | $4\pi\beta(\text{PC})-\gamma$ | yes | 0.55 | 9 - 18 | 45 | CH_4 | 0.1 | 7.3 | 5.724 5.710 0.992 | (3) (3) (4) | 5 840 486 387 | |

* Between brackets: Uncertainty in units of the last digit

Table 3

| Laboratory | Measured by ionization chamber | | Efficiency tracing | | | | Range of ϵ_{β} (%) | Polyn. order | $\epsilon_{\beta\gamma}$ (%) | Radioactivity concentration with combined uncertainty 1982-05-01 (Bq mg ⁻¹) |
|------------|--------------------------------|--|--------------------|--------------------------------------|-----------------------|--------------------------|------------------------------------|-----------------|---------------------------------|---|
| | Adsorption* (Bq) | Activity concentr. 1982-05-01 (Bq mg ⁻¹) | β detector | γ -channel window (keV) | Data points number | time per point (s) | | | | |
| AAEC | 780 | 603 ± 12 | PC | 795 - 900 | 50 | 1 700 | 95 - 85 | 1 | 0.1 | 608.8 ± 2.6 |
| | | | | 755 - 1 445 | 20 | 1 200 | 94 - 83 | 1 | | |
| | | | | 1 250 - 1 500 | 10 | 4 000 | 86 - 68 | 1 | | |
| AECL | 790 | 607.1 ± 1.0 | PC | 740 - 1 510 | 168 | 1 000 | 91 - 78 | 1 | 0.1 | 605.4 ± 2.4 |
| | | | | 740 - 900 | 202 | 1 000 | 93 - 81 | 1 | | |
| BCMN | 900 | - | PC | > 750 > 740 | ≈ 130 | 1 000 | 93 - 69 | 1 | 0.11 | 604.5 ± 1.8 |
| BIPM | 7 430 ** | 604.9 ± 0.4 | PC | 700 - 900 | 51 | 7 000 | 91 - 80 | 2 2 | 0.3 | 606.7 ± 2.7 |
| ETL | - | - | PC | 760 - 930 | 34 | 2 100 | 94 - 80 | 2 | 0.25 | 607.0 ± 2.3 |
| | | | | 760 - 1 500 | 28 | 2 100 | 92 - 77 | 2 | | |
| IER | 612 | - | PC | 785 - 890 | 600 | 100 | 94 - 78 | 1 | 0.1 | 595.1 ± 2.8 |
| | | | | 785 - 1 600 | 580 | 100 | 90 - 75 | 1 | | |
| IFIN | 0.08 | - | PC | 750 - 980 | 49 | 1 000 | 91 - 64 | 1 | 0.54 | 620.0 ± 2.9 |
| LMRI | - | 604.2 | PC | 760 - 820 | 22 | 3 000 | 94 - 76 | 1 | 0.11 | 604.3 ± 1.3 |
| NAC | 390 | 603 | LS | > 950 | 11 | 1 000 | 83 - 72 | 2 | 8.0 | 600.5 ± 5.2 |
| NBS | 535 | 603.3 | PPC, AC | 755 - 870 > 890 | ≈ 60 | 3 000 | 92 - 75 | 1 1 | 0.54 | 605.9 ± 1.9 |
| NIM | - | 596.8 ± 1.8 | PC 1 | 740 - 867 | 190 | 900 | 97 - 80 | 1 | 0.4 | 599.2 ± 1.6 |
| | | | PC 2 | 769 - 952 | 70 | 900 | 95 - 80 | 1 (2) | 0.29 | 597.4 ± 1.1 |
| | | | LS | - | - | - | - | 1 (2) | 5.6 | 593.1 ± 1.2 |

* Activity remaining in the "empty" ampoule after two rinsings

** Rinsings took place later, when the residue was dry

Table 3 (cont'd)

| Laboratory | Measured by ionization chamber | | Efficiency tracing | | | | Range of ϵ_β (%) | Polyn. order | $\epsilon_{\beta\gamma}$ (%) | Radioactivity concentration with combined uncertainty 1982-05-01 (Bq mg ⁻¹) |
|------------|--------------------------------|--|------------------------|--------------------------------------|-----------------------|--------------------------|----------------------------------|-----------------|---------------------------------|---|
| | Adsorption* (Bq) | Activity concentr. 1982-05-01 (Bq mg ⁻¹) | β detector | γ -channel window (keV) | Data points number | time per point (s) | | | | |
| NPL | 1 100 | 593.7 | PC | 745 - 905 | 34 | 1 000 | 83 - 70 | 2 | 0.32 | 601.5 ± 11.4 |
| | | | | > 745 | 28 | | 78 - 63 | 2 | | |
| | | | LS | 770 - 920 | 350 | 2 000 | 92 - 77 | 2 | | |
| | | | | 1 230 - 1 630 | | | 89 - 68 | 2 | | |
| | | | 1 850 - 2 200 | | | 76 - 35 | 2 | | | |
| NRC | 500 | - | PFC, AC | 780 - 810 | 240 | 700-2 000 | 91 - 73 | 1 | 0.32 | 609.30 ± 0.97 |
| | | | | 1 250 - 1 510 | | | 82 - 60 | 1 | | |
| OMH | 580 | 601.8 | PC | 750 - 1 500 | 23 | 1 000 | 94 - 74 | 2 | 0.1 | 606.4 ± 1.0 |
| | | | | | 14 | 1 500 | 93 - 74 | 2 | | |
| PDS | - | - | Ge(Li) γ | ? | - | - | - | - | - | 616.6 ± 5.9 |
| PTB | 500 | 605.09 | PC | 750 - 880 | 59 | 10 000 | 93 - 65 | 4 *** | 0.11 | 600.9 ± 2.2 |
| | | | PPC | 760 - 900 | 40 | 2 400 | 90 - 70 | 4 | 0.25 | 605.8 ± 2.2 |
| | | | PFC | 1 250 - 1 600 | ~ 60 | 1 600 | 88 - 76 | 4 | 0.25 | 605.6 ± 2.0 |
| SCK | 357 | - | PC | 745 - 930 | 54 | 5 × 1 000 | 94 - 82 | 1 | 0.25 | 604.8 ± 1.2 |
| | | | | " | 17 | " | 90 - 81 | 1 | | |
| | | | | " | 6 | " | 92 - 80 | 1 | | |
| | | | 4 π (NaI) γ | > (70 - 160) | 100 | 2 000 | - | - | | |
| UVVVR | 220 | - | PC | > 715 | 56 | 400 | 91 - 63 | 2 | 0.32 | 596.6 ± 1.6 |

* Activity remaining in the "empty" ampoule after two rinsings

*** a = a = 0

2 3

Table 4 - Variables of the efficiency functions used

| | x | y |
|-------|---|---|
| AAEC | $1 - \epsilon_\beta$ | $(1/m) \left(\frac{N_\beta}{1 - \tau_\beta N_\beta} - B_\beta \right)$ |
| AECL | $(1 - \epsilon_\beta)/\epsilon_\beta$ | $(N_\beta N_\gamma)/N_c$ |
| BCMN | $1 - N_c/N_\gamma$ | N_β |
| BIPM | $1 - \epsilon_\beta$ | N_β^{tot} |
| ETL | $\frac{1 - N_c/N_\gamma}{N_c/N_\gamma}$ | $\frac{N_\beta N_\gamma}{m N_c}$ |
| IER | " | $(1/m) \left(\frac{N_\beta N_\gamma}{N_c} - A_{134} \right)$ |
| IFIN | $1 - N_c/N_\gamma$ | N_β/m |
| LMRI | $(1 - \epsilon_\beta)/\epsilon_\beta$ | ? |
| NAC | N_γ/N_c | $N_\beta N_\gamma/N_c$ |
| NBS | $(1 - \epsilon)/\epsilon$; $\epsilon = (1 - Y)/N_\gamma$ | $N_\beta/(1 - \epsilon)$ |
| NIM | $(1 - \epsilon_\beta)/\epsilon_\beta$; $1 - \epsilon_\beta$; I_o/I_n^* | $N_\beta N_\gamma/N_c$; $(N_\beta - \epsilon_\beta A_{134})(1/m)$; N |
| NPL | $1 - \frac{N_c}{N_\gamma}$ | $N_\beta - A_{134} f(N_c/N_\gamma)$ |
| NRC | Y/N_γ | N_β^{tot} |
| OMH | $\frac{1 - N_c/N_\gamma}{N_c/N_\gamma}$ | $\frac{N_\beta N_\gamma}{m N_c}$ |
| PDS | - | - |
| PTB | $1 - \epsilon$; $1 - N_c/N_\gamma$; $\frac{1 - N_c/N_\gamma}{N_c/N_\gamma}$ | N_β/m |
| SCK | $1 - N_c/N_\gamma$ | N_β^{tot}/m |
| UVVVR | $(1 - \epsilon_\beta)/\epsilon_\beta$; $1 - \epsilon_\beta$ | $N_\beta N_\gamma/N_c$; N_β/m |

* I_n = anode current with optical filter, I_o = current without filter

