

Electricity and Magnetism Supplementary Guide for the Submission of CMCs

Version 5.2 (April 2018)

1. Scope

This Supplementary Guide reports information on how to formulate Calibration and Measurement Capability (CMC) Excel files in the field of Electricity and Magnetism (EM). The Guide is intended to be used worldwide, but makes provision for specific requirements of each Regional Metrology Organisation (RMO), particularly concerning the way the information supporting CMCs is gathered and the detailed CMC intra-regional review process.

For general information related to the CMC review process, acceptance criteria, support by key and supplementary comparisons etc. the reader is addressed to a set of CIPM documents, downloadable from the web at: <http://www.bipm.org/fr/cipm-mra/cipm-mra-documents/>.

This Guide is addressed to both the NMIs that have not yet submitted any CMCs and the NMIs that would like to add CMCs to their set or modify CMCs already published.

Suggestions about presentation of specific quantities in CMC entries are given in Appendix 1. A list of questions and answers and of pending problems is given in Appendix 2.

2. Closely related documents

The following documents are cited in this Supplementary Guide and must be used with it. They can all be found on the BIPM web site at <http://www.bipm.org/fr/cipm-mra/cipm-mra-documents/> under “Guidance on CMCs and QS”:

- 1) CIPM MRA-D-04, *Calibration and Measurement Capabilities in the context of the CIPM*
- 2) *Classification of services in electricity and magnetism*
- 3) *Additional Instructions and templates for CMC files in EM*: This zip-file includes the template for the submission of EM CMCs.
- 4) *Additional Instructions and templates for CMC file, general*: This list includes instructions for closely related CMCs and instructions for uncertainty matrices in CMC files, both developed for CMCs in electricity and magnetism.

3. Quality System requirement

Since the transition period of the Mutual Recognition Arrangement (MRA) has ended on 31 December 2004, the CMCs submitted must now be covered by a Quality System (QS) reviewed and approved by the RMO.

4. Modification of CMCs already in Appendix C

Following the document CIPM MRA-D-04 *Calibration and Measurement Capabilities in the context of the CIPM MRA*, modified CMCs fall into three categories:

- a) CMCs corrected for material or editorial errors or for improving the explanatory text; these

CMCs do not require a new review.

- b) CMCs modified to increase the uncertainty or to reduce the scope; also these CMCs do not require a new review. The proposal for change is received by the TC chairperson and transmitted to the coordinator of the BIPM database. However, in case that the change was originated by a comparison result, the TC chairperson should verify that the reduction in scope or the increase of the uncertainty is sufficient to assure the equivalence of the measurements. It is desirable in this case that the relevant RMO (or the BIPM) informs the other RMOs of the changes and their motivation.
- c) CMCs modified to change the method of measurement or to reduce the uncertainty or to increase the scope; these CMCs require a new review.

For CMCs of categories a) and b), modifications must be made on the Excel files corresponding to the published CMCs. Download these files from the restricted access JCRB CMC site from the link “Get published CMCs” (ask your TC chairperson for user name and password, or use user name “tcguest” and password “tcontact”). After downloading the file of your country, modify its name as follows:

RMO_Country_EM_date_mod_ab

where *RMO* is the acronym of the regional metrology organisation, *Country* is the ISO two- letter abbreviation for the Country¹, and *date* is the date of submission in the form yymmdd. *EM* stays for Electricity and Magnetism and *mod_ab* is to show the type of file. No other documentation is due to support the proposed changes, unless requested by the TC chairperson.

Modified CMCs of category c) will be reported together with and will follow the same rules as new CMCs (see after).

The following colour codes, suggested by the BIPM, will be used in all cases, a), b) and c), to highlight the changes in a CMC entry:

- 1) any modification of an existing entry will be written **in bold red character**;
- 2) in case of cancellation of words in a cell do **not use the strike-through character formatting** but change to **bold red** the remaining cell content;
- 3) in the rare case that the whole cell content is deleted, highlight the now empty cell by a **light pink** background;
- 4) withdrawn CMCs will be highlighted by a light **pink background extending over Columns A to P** and the words “to be deleted from the KCDB” placed in the yellow Comment Column (**Column T**).

The NMI internal service identification of the modified CMCs (**Column Q**) will not be changed. Internal service identifications of withdrawn CMCs will not be used for new CMCs. After downloading the file of your country take care to convert into black, or to delete, all previous colour coded entries, as applicable.

5. Submission of new CMCs or of modified CMCs requiring new review.

New CMCs and modified CMCs under category c) above will be reported in the same CMC Excel file, whose filename will be coded as explained in the following section 8. In this file the modified CMCs will be identified by the use of the colour codes suggested by the BIPM, as explained in the previous paragraph. Care should be taken that new CMCs do not have the same NMI internal identification (**column Q**) of CMCs already published in the Appendix C.

Each CMC Excel file with new CMCs, or modified CMCs requiring new review, shall be

¹ <https://www.iso.org/obp/ui/fr/#search>

accompanied by all information needed to substantiate claims and methods. The way this information is organised is decided by the region. For example it can be reported in a standard questionnaire approved by the region.

6. Intra-RMO review process

Even if the review process can be different in the details from region to region, it will always follow the following main steps:

- each NMI sends its CMC file together with related supporting information to the TC chairperson;
- the TC chairperson sends all the files to the technical group of his region in charge of the analysis of the CMCs, which will carry out the review following the JCRB criteria, as instructed from the TC chairperson (intra-regional review); to help the chairperson to supervise the review work, it is suggested that agreed changes are coded in colour by the reviewer, to be clearly visible: for example modifications or additions, with respect to the version submitted, will be **written in bold blue**, and deleted CMCs will be highlighted with **light blue background** and corresponding comment “to be deleted” in **Column T**;
- at the end of the review, the TC chairperson receives the files with approved CMCs from the reviewers and prepares them for the inter-RMO review, taking care that only the correct BIPM colour codes (see section 4) are maintained;
- the TC chairperson posts the CMC files, together with a report, on the JCRB web page for the inter-regional review of CMCs, which will automatically start the inter-RMO review process.

7. Inter-RMO review process

The general interregional review process is described in document CIPM MRA-D-04. In section 8 it is stated that

“CC-WG on CMCs may establish their own rules and timelines for coordinating the interregional review of CMCs. Therefore, posting, distribution and submission of comments on CMC submissions may be done without the use of the JCRB website and without following the deadlines specified for this purpose.”

The following describes the general principles of the process followed by the CCEM. The steps involved in the review process are described in the following section.

- RMOs are encouraged to cooperate in reviewing submitted CMCs. It is not necessary that all RMOs review the same CMC set. RMOs (through the chair of WGRMO and TC-chairs) may communicate with one another to divide responsibility or to decide who will participate. However, it is important that the process works in a way that ensures the confidence in the approved CMCs.
- In 2015 the CCEM agreed to further improve the efficiency of the CMC review through selective sampling of the entries to be reviewed. A proposal for the scope of the inter-RMO review is made by the chair of WGRMO, based on the magnitude of change, history of previous reviews, coverage by on-site technical review, rotation and high-level technical judgement. However, the final decision on the scope of the review lies within the RMOs.
- The review should be restricted to new and improved services, with either reduced uncertainty or wider scope.
- Technical peer review reports should be included in the CMC submission, when available, and the information be used by the reviewers.

- Since 2017 it is mandatory to submit new CMCs in matrix format.
- A new CMC set should only be submitted by an RMO if the interregional review process for the previously submitted batch is completed.
- The CCEM uses shorter deadlines as the standard JCRB procedure. The enforcement of these shorter deadlines needs to be made by the chair of WGRMO, because the JCRB website sends reminders after the - longer - deadlines specified in the document CIPM MRA-D-04.

The different steps involved in the interregional review process are the following (*in italics: specific CCEM rules*, in roman: general JCRB rules)

- RMO TCEM chair uploads new CMC set to JCRB CMC-review web page. A manual for the website is available online: http://www.bipm.org/JCRBCMCs/JCRB_webpage_manual-2.pdf.
- Automatic e-mails including the CMC files are sent to RMO representatives, RMO TCEM chairs, and the JCRB Executive Secretary.
- *A proposal for the scope of the inter-RMO review, based on selective sampling, is made by the chair of WGRMO to the RMO TCEM chairs, based on the magnitude of change, history of previous reviews, coverage by on-site technical review, rotation and high-level technical judgement. However, the final decision on the scope of the review lies within the RMOs.*
- In order to identify the TC/WGs that are interested in reviewing a particular submission, the chairs are requested to provide the date for sending comments no later than three weeks after the CMCs are posted on the website. After two weeks an automatic reminder is generated. After one further week, if the TC/WG chair has not provided a date for sending comments, the respective RMO relinquishes its right to continue with the review.
- *The review report shall be sent within four months after the announcement of the review.*
- TC/WG chairs post their first comments in the JCRB website by the date they set when they accepted to review them. Three weeks before the RMOs chosen date an automatic reminder is generated. If they have not sent their comments by their chosen date, in the absence of any other communication with the JCRB Executive Secretary, they relinquish their right to continue with the review.
- If a TC/WG chair needs to change the date specified for sending their first comments he/she should make the change through the JCRB website before the origin deadline has expired. Note, however, the JCRB Executive Secretary constantly monitors the review status and repeated or unduly long extensions may cause some questions and need discussion with the originating RMO.
- *If a reviewer requires additional information to make a decision on the submitted entry, he/she should contact the submitting NMI within three weeks after the start of the review.*
- *A maximum of three weeks is allowed for the first response of the submitting NMI to the request of the reviewer. Once the contact is established and the first response is given, further iterations of the process should take less time (typically less than two weeks per iteration).*
- The inter-regional review continues until all TC/WG chairs agree that the originating TC/WG has produced a submission that can be approved by all.
- The originating RMO shall submit a revised Excel file through the website for final approval, which will be used for the publication in the KCDB.
- The RMOs TC/WG chairs now have three weeks for considering the final vote. Two weeks after the final submission, a reminder e-mail will be sent and after one further week the RMO loses its right to vote and is considered an abstention. The final approval is done on a consensus basis. All

the RMOs should approve or abstain to vote, but a single vote against is enough for not approving the CMCs.

8. Structure of the CMC table

If an NMI declares CMCs for the first time, the empty Excel template, downloadable from the BIPM web site, as indicated in section 2 shall be used. In all other cases, the Excel table of already existing CMCs shall be used instead of the template, see the end of this section. Only one file per Country must be submitted. The submitting NMI will change the name of the file by introducing the acronym of the regional metrology organisation for *RMO* and the ISO two-letter abbreviation for *Country*. For *date* the form *yymmdd* will be used.

Header and footer are provided in the Excel template to help to identify the table when printed. In the header the acronyms of the NMI and of the Country must be introduced. In case of more than one NMI per Country, all the NMI acronyms will be reported in the header.

In the table, at the right side of the white columns, there are three blue, six yellow (divided by a white column for *Comments to be published via the database*) and three green columns. The blue columns are for supporting the CMC traceability. The yellow columns are for the administration of the table and for the names of the uncertainty matrices and of the tags for closely related CMCs. The green columns are to be used during the review processes. The white columns and the blue and yellow columns, not the green ones, are to be filled in by the submitting NMI.

Only the white columns, and the yellow columns “NMI Service Identifier”, “NMI”, and “Uncertainty Matrix” will be reported, with attached uncertainty matrices, into the Appendix C. The columns “Service category” and “Tag for closely related CMCs” are essential for a correct transfer of the table into the Key Comparison Database.

If the NMI has already CMCs in the KCDB, all changes (additions, modifications, deletions) shall be made in the Excel table of the existing CMCs. This file can be downloaded from the [JCRB CMC website](#) using the “get published CMCs” link. All changes shall be made on this file, using the color codes described in section 4. Before submission, all CMCs which are unchanged shall be hidden by using the Excel “hide” function, so that only the changes are visible to the reviewers.

9. General Criteria of compilation

- 9.1 The MRA Appendix C submission is an Excel spreadsheet where separate NMI services are listed in rows, and columns are used to organise the attributes of the services.
- 9.2 The CMCs will be reported in a single spreadsheet, following strictly the order of the EM *Classification* of services. A disordered list would give unnecessary, additional work to the TC chairperson.
- 9.3 Each entry must correspond to only one classification category and to only one measurand: for example electrical power and energies, even if in the same classification category, are different measurands and should be reported in different CMC entries.
- 9.4 Whenever possible, CMCs corresponding to different values of the same quantity shall be group together in a matrix. If the uncertainty depends on two variables (like the range of the measurand and the range of the parameter) matrices shall also be used (see 8.3 and 11.6).
- 9.5 Each CMC entry should be self consistent, without reference to other entries, because each CMC will be usually displayed alone by the database.
- 9.6 For each CMC, multiple entries in the two columns on Measurement Conditions (i.e. when CMC has more than one parameter) must be separated vertically into separate cells.

- 9.7 Do not insert headings nor blank rows in the table. Do not merge any cells vertically and do not use symbols of repetition (“”), to show that the information is the same as in the previous CMC, but repeat the information.
- 9.8 Do not use the Excel facility to append comments to the cells, but use the Comments columns on the right side of the table (see 17.2 and 17.3).

NOTE 1. For some quantities (power and energy, AC-DC transfer, voltage and current ratios, RF calibration factor, RF reflection coefficient and RF scattering parameters) recommendations for a uniform presentation are reported in Appendix 1.

10 Language and Symbols

10.1 Detailed writing rules for preparing the CMC files are given in CIPM MRA-D-04. **These Instructions must be carefully read and the rules carefully applied.** If too many violations are found, the CMC file will be sent back to the submitting NMI for correction. Compliance with the writing rules will greatly help the BIPM in the importation of the Excel files into the database and reduce the probability of errors.

10.2 The following rules complement those in CIPM MRA-D-04:

- as units for angle use the symbols °, ’, ” or rad; do not use: deg, min, sec, nor degree, minute, second, nor arc-second;
- use SI accepted prefixes for decimal multiples and sub-multiples; they must be written close to the symbol of the unit of measurement, in the appropriate columns, and not close to the numerical values;
- when reporting a list of items in the same cell use the comma (,) as the separation character [e.g.: DC voltmeter, multimeter, multifunction transfer standard];
- use a colon (:) to introduce a specification of a previous text-item of information [e.g.: reflection coefficient on coaxial: phase]

11 Electricity and Magnetism Classification scheme

11.1 The *Classification of Electricity and Magnetism CMCs*, approved by the CCEM, provides the NMIs with uniform and world-wide agreed terminology for the classification of their services. These are divided in major headings, sub-headings and categories. For each category, characterised by a 3-number classification, a list of instruments/artefacts is given.

11.2 Usually it should be possible to find classified denominations for the quantity of interest and the instrument/artefact, to be reported in **columns A and B**.

11.3 CMCs can only be submitted for services listed on the approved classification of EM services.

11.4 The most recent approved [classification of EM services](#) is available from the BIPM web site.

12 Calibration or Measurement Service (col. A, B, C)

12.1 The Quantity and the Instrument/Artefact names will usually be found in the *Classification*. If this is the case, use the given terms as explained in 12.2 and 12.3, then take note of the corresponding 3-number classification for introduction in **column R** (§ 18.3). If this is not the case, discuss the situation with the TC chairperson.

12.2 **In column A (Quantity)**, insert the appropriate description of the Quantity. This is usually

given in the *Classification* by the names of the heading, sub-heading and category. If needed for better understanding, a specification character (:) and/or separation characters (,) can be used [e.g.: capacitance: dissipation factor]. Because combining heading, sub-heading and category could lead to a cumbersome and unnecessarily long definition of the Quantity, you are very much encouraged to use one of the standardised wordings reported in the classification file *with column A definitions*, provided with this Guide.

- 12.3 **In column B (Instrument or Artefact)**, insert the appropriate name for the Instrument/Artefact. If more than one instrument applies, report a list using the separation character (,). If the measurand is not clearly identified by the Instrument/Artefact name, introduce a further specification using the specification character (:). If a symbol referring to the measurand is used in one of the subsequent columns, introduce this symbol, *in Italics*, in **column B** as the last piece of information, separated from the previous item(s) by a specification or separation character, whatever is more convenient [e.g.: multifunction calibrator: V] [e.g.: fixed capacitor: dissipation factor, *d*]. Do not use the symbol *k*, as this symbol is reserved for the uncertainty coverage factor.
- 12.4 **In column C (Method of Measurement)**, a synthetic but clear definition of the method of measurement must be reported. This information, combined with that in **column N** about the standard(s) used as the reference(s), is of fundamental importance to allow an expert to identify the measurement technique.

Example 1

| A | B | C | R |
|--|----------------------------|--|-------|
| DC resistance standards and sources: intermediate values | Fixed resistor: <i>R</i> | Comparison by means of current comparator bridge | 2.1.2 |
| AC high voltage: ratio error | Voltage transformer bridge | Generation of known reference error | 8.3.4 |

13 Measurand Level or Range (col. D, E, F)

- 13.1 **In column D (Minimum value)**, insert the minimum value of the range of the measurand as a single numerical value. If you introduce zero for this number, check that the expression for the uncertainty, reported in **column I**, is still meaningful.
- 13.2 **In column E (Maximum value)**, insert the maximum value of the range of the measurand as a single numerical value. If your measurement refers to a single value of the measurand, **column D** and **E** will report the same number.
- 13.3 **In column F (Units)**, report the unit of measurement that applies to the minimum and maximum values in the two previous columns. Use accepted SI prefixes if needed. If a unit like dB needs a reference value, include this value in **column B** (see last line of Examples 2, below). In case of quantities without dimension leave **column F** blank. In case of quantities that can be better given in relative terms it is also possible to express the unit as a ratio of different multiples of the same pertinent unit or to use the symbol %.
- 13.4 Usually it is possible to understand from the quantity if measurements can be performed on both positive and negative values. Only if this is not the case report explicitly negative range limits.

Example 2:

| A | B | C | D | E | F | R |
|---|---|---------------------------------------|--------|--------|-----|--------|
| DC Voltage ratios: up to 1100 V | Resistive divider | Comparison with reference divider | 1.E-06 | 1 | | 1.3.1 |
| AC power and energy: single phase ($f \leq 400$ Hz), active power | Power converter | Comparison with reference standard | 0.05 | 30000 | W | 7.1.1 |
| Scattering parameters: reflection coefficient (S_{ii}) in coaxial line, phase | Passive device | Dual six port | -180 | 180 | ° | 11.3.1 |
| AC resistance: argument | Fixed resistor | Calculable resistor | $-\pi$ | $+\pi$ | rad | 4.1.1 |
| Antenna properties: antenna factor | Log antenna, reference value for unit: 1 m^{-1} | Three antenna method | -30 | 80 | dB | 11.5.1 |

14 Measurement Conditions / Independent Variable (col. G, H)

- 14.1 In **column G (Parameter)**, report any condition, which imposes limitation or optional restriction to the measurement, in the form of the name of the parameter that describes this condition. If several conditions apply, the names of the parameters will be entered into separate cells in the column. If a symbol referring to the value of a parameter is used in one of the subsequent columns, report this symbol, *in Italics*, after the name of the parameter, using the specification character [e.g.: frequency: f]. Do not use the symbol k , as this symbol is reserved for the uncertainty coverage factor.
- 14.2 In **column H (Specifications)**, report the range of the parameter or other pertinent information that defines the restriction. If **column G** has several parameters, report in **column H** the corresponding ranges of variation, or pertinent information, in the corresponding cells.
- 14.3 When indicating a range in **column H** use the notation *minimum to maximum* [e.g.: 1 MHz to 500 MHz or (1 to 500) MHz]. If only one limit needs to be specified, use the notation *up to* (for maximum) or *down to* (for minimum).
- 14.4 When a parameter takes discrete values, report these values separated by commas (,) [e.g.: (1, 2, 5, 10) kV].
- 14.5 In some cases the measurand takes only discrete values. In this case the range **columns D** and **E** report the minimum and maximum values, while the parameter **columns G** and **H** report, respectively, the symbol of the measurand (previously defined in **column B**) and the list of values. This will be the first information in the parameter list. A typical case could be the decadic values of resistance within a certain range. This condensed presentation of a group of CMCs is possible if a single uncertainty declaration (by a numerical value or an expression) is given for the group or, in case of uncertainty dependence from two variables, if an uncertainty matrix is given (see § 18.3).

14.6 Usually do not report as parameters the ambient conditions of the measurement. These are intended to be standard laboratory conditions, whose limits of variation are taken into account in the reported uncertainty. Report the temperature, or temperature range, of the measurement or other ambient conditions only if it is relevant to the identification of the measurement service: **column U** is usually more appropriate for this type of information (see § 17.3).

Example 3

| A | D | E | F | G | H | I | J |
|--|-----|-------|----------|--------------|--------------|----|-------------------------|
| AC power and energy: single phase ($f \leq 400$ Hz), active power | 600 | 600 | W | Voltage | 120 V | 60 | $\mu\text{W}/\text{VA}$ |
| | | | | Current | 5 A | | |
| | | | | Power factor | 1 | | |
| | | | | Frequency | 50 Hz, 60 Hz | | |
| AC resistance: real component | 0.1 | 1E+05 | Ω | Frequency | 400 Hz | 10 | $\mu\Omega / \Omega$ |
| | | | | Max. voltage | 10 V | | |
| | | | | Max. current | 0.5 A | | |

| A | B | D | E | F | G | H |
|--|------------------------|--------|----|------------------|---------------|--|
| DC resistance standards and sources: intermediate values | Fixed resistor: R | 0.001 | 10 | $\text{k}\Omega$ | R | 1 Ω , 10 Ω , 100 Ω , 1 $\text{k}\Omega$, 10 $\text{k}\Omega$ |
| DC Voltage ratios: up to 1100 V | Resistive divider | 1.E-06 | 1 | | Input voltage | (1 to 10) V |
| Scalar RF reflection coefficient: in coaxial line | Passive device | -1 | 1 | | frequency | 0.1 GHz to 40 GHz |

15 Expanded Uncertainty (col. I, J, K, L, M)

15.1 A Calibration and Measurement Capability is the highest level of calibration or measurement normally offered to clients. This implies that the global uncertainty associated with a CMC should consider the contribution (noise, short term instability, resolution etc., but not transport) of the instrument/artefact under calibration, usually the best available. The uncertainty is usually expressed in terms of a confidence level of 95% and given by no more than 2 significant digits.

15.2 In **column I (Value)**, the expanded uncertainty value (without \pm) must be entered. This may be a fixed value, valid over the entire range of the quantity, a range of values [e.g.: (1 to

50)E-06] or a mathematical expression.

- 15.3 For AC quantities, and more in general when the uncertainty depends on two variables, usually the range of the quantity and the range of one parameter, it is recommended to use the presentation of the uncertainty in a two-dimensional matrix, following the rules given in the document listed as reference 4 on page 1 of this Guide. This technique will reduce the number of lines in the table and will improve presentation. In this case, **column I** will report the minimum and maximum uncertainty values. It is intended that the unit of measurement of the uncertainty values in the matrix is the one given in column J.
- 15.4 With the exception of presentation in matrices, if a range of values is given in **column I**, the uncertainty is considered to be varying linearly within the range, in correspondence to the variation of the measurand or of a parameter from the lower to the higher limits reported in **columns D** and **E** (for the measurand) or in **column H** (for the parameter). The functional dependence, if not clear, should be specified in the same **column I** after the uncertainty expression [e.g.: (1 to 50), depending on the frequency]. If the uncertainty variation is not linear, the user must be informed by introducing a comment in **column U**, for example specifying that the given values only define the minimum and maximum uncertainties. In this case however, the range of variation should be limited, otherwise the CMC entry would become useless.
- 15.5 Uncertainty by formula. When using a mathematical expression, defined symbols referring to the measurand or to the parameters will be entered. The units of measurement of the symbols will be specified in the same Column I after the expression (see examples). Do not introduce units of measurement within the expression. The unit for the resulting uncertainty is to be reported in **column J** (see 15.11). With the exception of the proportional uncertainty (15.6), for all other expressions (15.7 to 15.9) it is suggested to report the typical range that the uncertainty evaluates to [e.g.: $Q[0.00002, 0.02 \cdot \tan \delta]$, values from 2E-05 to 2E-02].
- 15.6 Uncertainty by formula: proportional uncertainty. For range-dependent uncertainties, the form bV can be used, where V is usually the symbol of the measurand, reported in **column B**, and b is a numerical value (proportionality factor). If the lower limit of the measurand is zero, the proportional expression alone is not acceptable, an uncertainty of zero being meaningless.
- 15.7 Uncertainty by formula: quadratic summation. Following the ISO GUM, the composition of a constant uncertainty term (for example the limit of sensitivity) and of a range dependent (proportional) uncertainty term, is done by quadratic summation. This can be written with the expression $Q[a, bV]$, where a is the numerical value of the constant term and bV is the proportional term (15.6).
- 15.8 Uncertainty by formula: linear summation. In other cases the uncertainty may be given by the linear summation of a constant and a proportional term, like $(d + cV)$.
- 15.9 Uncertainty by formula: 2nd degree polynomial. This type of uncertainty declaration will be used if it is of help with covering an extended range of the quantity or of the parameter by a single uncertainty declaration. Use of polynomials of degree higher than 2 or of other non elementary functions is strongly discouraged, as they are difficult to understand by the user and very often prone to errors, also due to the very limited capability of Excel in equation editing. For uncertainties depending on two variables the use of matrices (see 15.3) should always be preferred.
- 15.10 When an explanation is necessary in order to allow a correct interpretation of the uncertainty expression, add a separation character (,) after the uncertainty value or expression and write the comment [e.g. for the absolute uncertainty of a voltage divider: $(0.2E-06 + 0.05E-06 / U)$, U output voltage in V].

15.11 In **column J (Units)**, report the unit of measurement for the uncertainty. In case of relative uncertainties for quantities with dimensions, express the unit as a ratio of different multiples of the same pertinent unit. [e.g. $\mu\text{A}/\text{A}$] or, if the unit is an expression involving other units, use a negative power of 10 or use the symbol %. For quantities without dimensions, **column M** is essential for the correct interpretation of the uncertainty.

15.12 Do not use as unit of measurement the symbols ppt, ppm, ppb or similar.

15.13 In **column K (Coverage factor)**, report the coverage factor k .

15.14 In **column L (Level of confidence)**, report the probability corresponding to the declared uncertainty and coverage factor. A probability of 95% should in general be considered, but exceptions are allowed. The percentage sign must be reported close to the number.

15.15 In **column M (Is this exp. uncertainty a relative one ?)** answer *Y* or *N*, thus specifying if the given uncertainty is to be interpreted as a relative or an absolute one (with respect to the measurand whose units are given in **column F**). **Column M** must always be filled in, even if the nature of the uncertainty seems to be already clear.

NOTE: do not use the symbol k in the mathematical expression of the uncertainty, to avoid confusion with the uncertainty coverage factor.

Example 4:

| A | B | D | E | F | I | J | K | L | M |
|---|--|--------|-------|-----------------------|---|-----------------------|---|-----|-----|
| DC resistance meters: intermediate values | Fixed resistor | 0.001 | 10 | $k\Omega$ | 0.5 | $\mu\Omega/\Omega$ | 2 | 95% | Yes |
| DC Voltage ratios: up to 1100 V | Resistive divider | 1.E-06 | 1 | | 0.3E-06 | | 2 | 95% | No |
| Scalar RF reflection coefficient: in coaxial line | Passive device | 0.01 | 0.2 | | 0.005 to 0.010 | | 2 | 95% | No |
| DC resistance standards and sources: standards for high current | DC shunt | 1.E-04 | 100 | Ω | (15 to 2)E-06 | | 2 | 95% | Yes |
| Soft magnetic sheet and powder materials: specific apparent power | Epstein sample according to IEC 60404-2: P | 0.1 | 300 | VA/kg | $2\text{E-}03P$, P in VA/kg | VA/kg | 2 | 95% | No |
| DC voltage sources: low values | DC voltage sources | 1 | 1000 | nV | $Q[2, M]$, N is source noise in nV | nV | 2 | 95% | No |
| Electric discharge: apparent charge | partial discharge calibrator: q | 1 | 5000 | pC | (0.2 + 0.02 q), values from 0.22 to 100, q in pC | pC | 2 | 95% | No |
| Capacitance: dissipation factor for dielectric capacitors | Fixed capacitor, mica, four terminal-pair | 1E-04 | 1E-03 | | 4E-05 to 1E-04 | | 2 | 95% | No |

| | | | | | | | | | |
|-------------------------------------|--------------------------------|----|----|----|------|----|---|-----|----|
| AC voltage ratio: real component | 600-W- attenuator- boxes | 60 | 90 | dB | 0.01 | dB | 2 | 95% | No |
|-------------------------------------|--------------------------------|----|----|----|------|----|---|-----|----|

16 Reference Standard used in calibration (col. N,O)

16.1 In **column N (Standard)**, the reference standard(s) used in the measurement and making the link in the traceability chain to the SI unit, must be reported. In case several references are used for a certain measurement, these must be reported in the same cell, as a list of items separated by commas.

16.2 In **column O (Source of traceability)**, identify the laboratory(ies) which is (are) the source of traceability for the calibration of the reference standard(s) in **column N**. When the reference standard is calibrated by means of an auto-calibration process, like for some ratio standards, traceability is to the laboratory performing this process. In case of more than one laboratory, the list of laboratories should correspond to the list of the standards in **column N**.

Example 5:

| A | B | N | O | S |
|--------------------------------|------------|--|---------------|------|
| DC Voltage sources: low values | calibrator | voltage divider, solid state voltage standard | NMI1, NMI2 | NMI1 |

In this example, NMI1 is the submitting NMI (whose acronym is given in **column S**), which is able to perform the auto-calibration process of the voltage divider, while NMI2 is another NMI, which gives the traceability for DC voltage by calibrating a zener voltage standard of NMI1. This zener standard is the first line standard of NMI1, used for the calibration of the voltage standard reported in **column N**.

17 List of Comparisons Supporting the CMC, Internal Comments, Comments for the user (col. P, T, U)

17.1 In **column P (Comparisons supporting this measurement)**, enter the correct identification of one or more comparisons sufficiently related to the service being submitted, in order to demonstrate equivalence. Different types of comparisons can be considered:

- CCEM key comparisons or RMO key comparisons;
- supplementary comparisons;
- past CIPM, regional or other (e.g. pilot, BCR etc.) comparisons;
- bilateral comparisons.

In this column only comparisons for which measurements are completed and at least a draft report is available must be considered. Comparisons still running, even if close to completion, will be reported in the *Internal Comment* (yellow) **column T**.

17.2 In **column T** (yellow column) report comments related to the CMC and useful to the reviewer (this column will not be included in the Appendix C). Examples are: comparisons planned or still running, more details about traceability etc. General information supporting

the CMCs (e.g. quality system, equipment, technical publications, etc.) will be reported separately, in the form established by the RMO. More detailed information will be given to the reviewer under request.

17.3 In **column U** (white column) report comments useful to the user, as this column will be included in the Appendix C. Examples are: specific limitations to the CMC, details about the conditions of measurement, options etc.

Example 6.

| A | B | N | O | P | T | U |
|---|---------------------------------|---|------------|-------------------------------------|--|---|
| DC voltage meters: intermediate values | Multimeter | Voltage divider, solid state voltage standard | NMI1, NMI2 | EUROMET.EM-K8, BIPM.EM-K11.b | | |
| AC power and energy: single phase (f <= 400 Hz), active power | Power meter | Automated calibration system | NMI | | EUROMET.EM-K5, report in progress, Draft A | Minimum value of uncertainty with power factor equal to 1 |
| AC voltage: AC-DC transfer difference at medium voltages | Multifunction transfer standard | AC-DC transfer standard | NMI | CCEM-K6.a | | |
| DC resistance standards and sources: intermediate values | Fixed resistor | Fixed reference resistor | BIPM | PRAQ III B 5 - 97- 030 (CEN) | Report available | |
| DC resistance standards and sources: high values | Three terminal resistor | Three terminal standard resistor | NMI1 | NMI1-NMI2 intercomparison at 100 MΩ | Information on results in CPEM'98 Digest, page ... | |
| DC resistance standards and sources: intermediate values | Fixed resistor | Fixed resistor | NMI | CCEM-K1 | | Measurements only at 23 °C in oil or air bath |
| DC Voltage meters: very low value | DC voltmeter | Josephson voltage standard | NMI | BIPM.EM-K1b | | |

In the second example, in **column U**, the attention of the user is drawn to the fact that the range of the uncertainty (not shown) is not linearly related to the range of the measurand, as it should usually be.

In the second example reference to comparison is in **column T**, and not in **column P** because, at the time of submitting the given CMC, the results of that comparison were not yet available.

18 Euramet Electricity Services Administration (col. Q, R, S, T), Comments for the user (col. Q), Matrices and Tags (col. V and W) and intra/inter RMO review (col. X, Y, Z)

18.1 **Columns Q, R, S** are of particular importance for the management of the CMC table: they allow the experts to easily refer to a certain CMC, which facilitates, for example, e-mail discussions. Also, it is possible to merge rows from tables of different laboratories, for

example for comparison purposes, without losing knowledge of the service identification and originating NMI.

18.2 **Column Q (NMI service identification)** is provided for numerical identification of each service of an NMI.

If an entry is a modified CMC, aimed at substituting a CMC already published in the database, the published service identification number will not be changed.

If an entry is new, it will have a new identification number, which, in case the submitting NMI has already EM CMCs published in the database, should not be equal to numbers already used in the published set. If an NMI is submitting its CMC set for the first time, progressive numbering starting from 1 will preferably be attributed to the CMCs. If an Excel formula is used to generate these progressive numbers, the Excel command “Copy+Paste-special|Values” must be used at the end to leave only numbers in **column Q**. In case a CMC extends over more than one line, only the first line will be numbered.

18.3 In **column R (Service category)** report the service category, which can be found in the *Classification* and depends on the Quantity and the Instrument/Artefact reported in **columns A and B** (see §4). In case of lack of applicable terms see § 5.1. If more than one classification applies, one CMC entry for each classification must be provided, in order to have only one category per entry. In a few cases, different quantities have the same classification, like power and energy: in such cases different CMCs must be provided for the different quantities.

18.4 In **column S (NMI)**, report the acronym of the submitting NMI. In case of more than one NMI per Country, the acronym pertinent to each CMC will be reported. It is not possible to have in the same Country two or more NMIs submitting the same CMC.

18.5 **Columns T and U** were covered in §§ 17.2 and 17.3.

18.6 In **column V (Uncertainty matrix)** report the name of the matrix, included in the Excel file, where the uncertainties corresponding to the CMC are reported. If the uncertainty declaration has not used a matrix presentation, this column will be left blank (see document 4 listed on page 1 of this Guide for details).

18.7 In **column W (Tag for closely related CMCs)** report the tag chosen to identify the CMC as closely related to the previous or the following one having the same tag (usually for real and imaginary parts of the same quantity). This tag will alert the BIPM that these closely related CMCs must be presented close-by in the database (see document 4 listed on page 1 of this Guide for details).

18.8 **Columns X, Y, Z** are reserved for the review processes and should not be used during the preparation of the CMC tables.

Example 7

(examples for uncertainties presented in a matrix and for tagged CMCs are given in documents 5 and 6 listed on page 1 of this Guide)

| A | B | Q | R | S | T | U |
|---------------------------------------|--|-----|-------|---------|------------------------------------|---|
| AC voltage up to 1000 V: sources | Multifunction calibrator | 86a | 5.2.1 | METAS | EUROMET 464, participation planned | |
| AC voltage up to 1000 V: meters | Multimeter; multifunction transfer standard | 86b | 5.2.2 | METAS | EUROMET 464, participation planned | |
| DC voltage ratio: up to 1100 V | DC voltage divider | 9 | 1.3.1 | NMi-VSL | | |
| AC voltage ratio: imaginary component | One-decade inductive voltage divider with 11, 10 or 8 taps | 315 | 5.3.1 | PTB | | Minimum step depends on the type of divider |

Appendix 1

Presentation of some CMCs in Electricity and Magnetism

In the following some suggestions are given in order to harmonise the presentation of CMCs concerning particular quantities.

a) DC and AC Divider Ratios. Divider ratios will be presented as output/input, so that ratio values will usually be lower than 1.

b) LF Electrical Power and Energy. The following points should be considered:

- the unit of measurement must be in agreement with the type of power (W for active, var for reactive, VA for apparent), which suggests to have one separate entry for each type;
- the relative uncertainty is usually referred to the apparent power so that units like mW/VA (for active power) should be used;
- in the parameter column, the ranges for voltage, current, active or reactive power factor (with indication if inductive or capacitive) and frequency must be reported;
- for energy, the minimum and maximum measuring times must be reported;
- for three-phase power, it is usual to report the range values per phase: for clarity the information that the given range is "per phase" is to be specified in the user comment column.

c) AC-DC (voltage or current) transfer difference. In the range columns, voltage or current values must be reported and not values for the transfer difference. This approach is suggested also for HF AC-DC transfer difference. The uncertainty is generally given as relative to the voltage or current values.

d) AC-DC voltage transfer category boundaries

It is suggested to use 0.5 V and 5 V as the boundaries for the low, medium and higher voltage categories.

e) AC high voltage and current transformers

For these CMCs it is proposed to always report the ratio error for the real part (or modulus) of the ratio and the phase displacement (in rad or better in mrad) for the argument. The primary and the secondary voltage or current should be reported as parameters. This format is in accordance with the corresponding IEC standards for instrument transformers.

f) RF calibration factor

NMIs should report in the column *measurand level and range* the maximum and minimum power levels.

g) RF categories 11.2 (reflection coefficient and attenuation) and 11.3 (scattering parameters)

The first distinction between categories 11.2 and 11.3 is the type of measuring equipment used. If it only measures the magnitude of the quantity involved (and does not give any phase information) then the CMC line belongs under category 11.2, otherwise under 11.3.

For category 11.3.1 and 11.3.2, reflection coefficient is usually reported in linear terms. For categories 11.3.3 and 11.3.4, transmission/attenuation, values are often reported in dB, which is familiar to the customer. But from a physical point of view, a presentation as a complex value in linear terms should be preferred. The Euromet HF experts have agreed to use the latter format: the exact format is not yet fixed. In case of a linear presentation, information about the dB-range concerned may be given in column U (comments for the user).

h) Presentation of related quantities (complex quantities).

Real and imaginary parts of complex quantities belong to the same classification category. Such closely related CMCs will be tagged in order to alert the BIPM that they must be presented close-by in the database (see § 18.7).

Appendix 2

Answers to questions and problems

1)

Problem. If an NMI can offer the calibration of certain instruments and standards at two different levels of uncertainty, for example using a more refined (but time consuming) and a more straightforward method, can both services be submitted as two different CMCs?

Comment. In principle a CMC should correspond to the best capability of the NMI, but some CMC users, for example the secondary laboratories, would like to have all the NMI services they use listed in the database. Then the answer could be: propose as CMCs also lower level services if they employ different methods of measurement and are used to transfer traceability to secondary laboratories.

2)

Problem. Are there cases when the range columns of a CMC can be left blank?

Answer. Even if quite rare, it can happen. For example it has happened for category 11.9.1, for a CMC on the measurement of the geometrical dimensions of different types of HF waveguides. If columns D and E are left blank, the parameter columns should report sufficient information to identify the CMC.

3)

Problem. What is the meaning of “Source of traceability” (column O)?

Comment. In general it is the laboratory performing the first step of the traceability chain that ends at the corresponding standard reported in column N. But for complex traceability chains it is often the laboratory performing the last part of the chain. An example is the case of resistance scaling, where laboratories that obtain traceability from the BIPM in the range 1 Ω to 10 k Ω , report traceability to themselves for higher resistance ranges, where they use their own measurement procedures to extend the traceability given by the BIPM.

4)

Problem. Is it possible for the same standard in column N to report, as sources of traceability in column O, two different NMIs?

Comment. The only case when this can be accepted is when a CMC covers a wide range of values and traceability is different in different parts of this range. Otherwise traceability must be to only one laboratory. Of course, if this is metrologically correct, it does not take into account that, for example, commercial reasons could lead to change the NMI providing traceability. On the other side it is believed that the choice of an NMI as traceability provider is a medium or long term decision: for example, within EURAMET it is usual to register traceability agreements as EURAMET projects.

5)

Problem. It may happen that an NMI asks to reduce its uncertainty during the review process.

Answer. This situation does not give any problem, if the reviewer has enough supporting evidence for the new uncertainty. When the review process is finished, the CMC uncertainties should not be reduced any more, if not as a consequence of a new review (see CIPM MRA-D-04 on how to modify existing CMCs).

6)

Problem. Is it possible to present an uncertainty dependent only on one variable (e.g. the range of the quantity) with a matrix?

Answer. Yes, this is possible and can be used when it is difficult to express the dependence of the uncertainty on the variable by an equation.

7)

Problem. Is it possible to use an uncertainty matrix when one of the two variables on which the uncertainty depends is not a physical quantity but, for example, the type of connector, as in some HF measurements?

Answer. Yes. This has indeed been done by an NMI for RF-DC transfer difference.

8)

Problem. Is it possible to use a range declaration for the uncertainty in a cell of an uncertainty matrix?

Answer. No! The cells of an uncertainty matrix should contain only numbers. If mathematical expressions are used in the matrix to obtain the uncertainty values, after calculation these expression must be substituted by numbers having the right number of significant digits.

9)

Problem. When the uncertainty (column I) is given as a range (u_1 to u_2) sometimes it is not clear if it varies linearly with the range of the quantity or with the range of a parameter, or if the variation is not linear at all and the limits are just the extreme values of the uncertainty.

Answer. The general rule says that in case of a range declaration the uncertainty should vary linearly from the lowest to the highest value of the independent variable. Of course this is not relevant if the uncertainty declaration is supported by an uncertainty matrix, where the functional dependence is clearly visible. If there is no matrix, it is suggested to follow § 8.4: the dependence on the quantity or on a parameter or on both can be specified with a short text in the same column I; the non linear variation can be explained in column U, comment for the user. But if the dependence is not linear, as there is no way to evaluate the uncertainty in the middle of the range, the total variation of the uncertainty must be limited, or the CMC will become useless.

10)

Problem. For AC-DC transfer measurements, where the quantity is reported as a voltage or as a current (as agreed among the AC-DC experts), the uncertainty given as $\mu\text{V}/\text{V}$ or as $\mu\text{A}/\text{A}$ is relative or absolute?

Answer. It is relative, because it must be applied to the reported quantity, voltage or current.

11)

Problem. There has been the proposal to report asymmetric limits for the uncertainty in certain cases. Examples are quantities that have physical limits (like a reflection or transmission coefficient, which is always limited between 0 and 1) or quantities given in non linear units.

Answer. Reporting asymmetric uncertainties is quite difficult and not easy to understand on the part of the user. If the approximation introduced by a symmetric declaration is considered not good enough, the range in the CMC line could be adequately reduced and the CMC line split.

12)

Problem. For complex quantities, the form of presentation varies widely. For example, for AC voltage ratio, we have at least three different presentations for the imaginary component: same unit as the real component, as an angle in radian, or as an angle in degrees. For HF transmission and reflections coefficients, in some cases two lines are presented, one for real and one for imaginary, in other cases only one line for both is given. In the last case the two range-columns usually report only one value each (e.g.: -1; 1) but in other cases two values are reported in each column, with vector notation (e.g.: [-1,-1]; [1,1]).

Comment. This matter should be discussed by the LF and HF experts. When an agreement is reached, the solution suggested could be reported in Appendix 1. As it is obviously not possible to describe complex quantities in the different measurement fields in a unique way, the minimum which we should try to reach is to have only one description for one field. To give an example: for AC voltage ratios we should only use either the same description as it is used for the real component or an angle in rad or an angle in degrees. This would be much easier to understand for our customers.

13)

Problem. Different forms of presentation are still given for AC power measurements. For example the range of the power factor is given as: "1 to 0 i/c", "0 ind./cap. to 1", "0 leading to 0 lagging", "0 to 1 inductive or capacitive".

Comment. This matter should be discussed by the LF experts. When an agreement is reached, the solution suggested could be reported in Appendix 1. We should try to agree on one format. This would be much easier to understand for our customers.

14)

Problem. At present there are different ways to report the unit of measurement of the uncertainty in case of quantities without dimension or for relative uncertainty of quantities with dimension.

Comment. A way to unify this could be the following:

Absolute uncertainty for quantities without dimensions: no unit in column J (there should be no unit also in column F).

Relative uncertainty for quantities without dimension: negative power of 10 or % in column J.

Absolute uncertainty for quantities with dimensions: appropriate unit in column J (same unit as in col. F or a sub-multiple of it).

Relative uncertainty for quantities with dimension: report the pertinent ratio of submultiples of the considered unit (for example $\mu\text{V}/\text{V}$) in column J; but if the unit is a combination of other units, it is better to use a negative power of 10 or % in col. J.

History of changes

| Version number | Changes |
|---------------------|--|
| 4.2 (4 May 2009) | Note (f) on page 15: For the quantity rf calibration factor, the minimum and maximum power levels shall be stated in the column <i>measurand level and range</i> (instead of the calibration factor). |
| 4.3 (6 August 2009) | Clarification in section 1 that changes of CMCs (additions, deletions, modifications) must be made in the Excel file of already existing CMCs. |
| 4.4 (April 2011) | Insertion of chapter 0.6 on the inter-RMO review process, with deadlines and recommendations on how to accelerate the CMC review process. Modification of the answer to question 6 on page 17. It is now allowed to express the dependence of the uncertainty on one single variable by a 1 x n matrix. |
| 5.0 (March 2017) | General update of the document for review by CCEM. |
| 5.2 (April 2018) | New section 7 on interregional review process. |