

BUREAU INTERNATIONAL DES POIDS ET MESURES

Key comparison CCTF-K001.UTC - Results
 Degrees of equivalence $D_k = [UTC - UTC(k)]$ for January 2026
 Computed 2026 FEBRUARY 11, 09h UTC

Coordinated Universal Time **UTC** and its local realizations **UTC(k)** in National Metrology Institutes and Designated Institutes.
 Computed values of $[UTC - UTC(k)]$ and uncertainties valid for the period of this publication

Date 2025/26 0h UTC MJD	JAN 4	JAN 9	JAN 14	JAN 19	JAN 24	JAN 29	Uncertainty/ns		
	61044	61049	61054	61059	61064	61069	U_a	U_b	U_k
Laboratory k	$[UTC - UTC(k)]/ns$								
AzMI	507.1	522.1	538.7	557.3	570.5	584.0	0.4	14.2	14.2
BelGIM	-1.4	-0.5	-1.1	0.2	0.9	1.5	3.0	6.6	7.2
BEV	-34.5	-36.6	-25.9	-13.6	-4.6	5.6	0.4	6.2	6.2
BFKH	4634.9	-	-	-	-	-	3.0	14.6	14.8
BIM	6228.8	6277.3	6326.1	6366.9	6413.4	6458.1	0.4	5.6	5.6
BMM	4207.9	4235.5	4251.9	4276.6	4296.7	4310.4	0.4	6.0	6.0
BSJ	16.1	6.4	7.0	18.6	16.2	14.4	14.0	14.2	20.0
CENAM	-1.9	-1.9	-3.0	-3.0	-1.6	-0.0	6.0	9.0	10.8
CENAMAP AIP	-12.4	9.7	-7.6	-11.1	2.0	4.1	0.4	11.2	11.2
DEF-NAT	-	178.1	92.8	3.7	-81.5	-169.5	1.4	5.6	5.8
DFM	-26.1	-29.1	-32.1	-35.4	-39.4	-43.3	0.4	5.8	5.8
DZM	326.4	324.7	318.5	318.7	321.7	314.0	0.4	5.6	5.6
EMI	-68.9	-76.8	-81.8	-93.1	-104.2	-118.2	0.4	14.2	14.2
ESA	-0.2	-0.2	-0.0	-1.0	-1.2	-1.3	0.4	5.6	5.6
FTMC	1317.3	1326.4	1337.2	1339.6	1347.1	1340.5	0.4	5.6	5.6
GUM	-2.7	-3.2	-3.3	-3.6	-4.0	-4.2	0.4	5.6	5.6
IBMETRO	98.2	89.5	74.8	74.0	69.8	57.1	8.0	15.4	17.4
ILNAS	-7.6	-3.8	-2.9	2.9	5.3	11.3	1.0	5.6	5.6
IMBIH	-0.6	1.6	-0.8	-1.4	5.0	0.5	0.4	6.0	6.0
INACAL	23.4	38.8	21.6	12.1	0.8	17.0	10.0	8.2	13.0
INM	-	-	-	-	-	-			
INM(CO)	-42.9	-40.8	-41.7	-34.6	-24.4	-31.5	6.0	8.2	10.2
INMETRO	-3.0	1.0	-3.2	7.7	-5.9	-1.5	1.4	6.8	7.0
INPL	15.0	13.6	6.8	13.2	-2.8	-13.4	0.4	15.4	15.4
INRIM	1.8	1.2	0.9	0.8	1.2	1.9	0.4	2.4	2.4
INTI	-121.2	-109.5	-132.1	-132.1	-144.4	-137.7	0.4	6.8	6.8
IPE/ASCR	-32.5	-33.1	-27.3	-26.1	-23.7	-15.2	0.4	6.4	6.4

IPQ	-	121.5	97.0	72.8	-	13.4	1.4	6.4	6.6
JV	-1.7	-1.2	-0.6	1.2	2.5	2.6	0.4	10.0	10.0
KazStandard	-4.9	-5.4	-4.8	-5.0	-5.0	-2.7	1.4	8.8	9.0
KRISS	-0.4	0.1	0.8	1.0	1.0	1.1	0.4	6.2	6.2
LAMETRO-ICE	31.8	47.5	65.8	74.4	65.1	65.8	0.4	14.2	14.2
LNE-OP	0.2	0.2	-0.1	0.1	0.3	0.4	0.4	2.4	2.4
MASM	-	-	-	-	-	-	-	-	-
METAS	-1.9	-1.4	-0.9	-0.5	-0.2	0.2	0.4	2.4	2.4
MIKES	-7.3	-8.7	-10.5	-12.8	-11.0	-9.2	0.4	5.6	5.6
MIRS/SIQ/Metrology	1560.7	1554.6	1556.5	1552.6	1539.2	1533.0	0.4	8.4	8.4
MSL	9.7	15.3	13.0	20.7	34.2	40.5	1.4	6.2	6.4
NICT	0.0	0.2	0.6	1.1	1.5	2.2	0.4	4.2	4.2
NIM	-0.4	-0.5	-0.3	-0.3	-0.8	-0.6	0.4	4.2	4.2
NIMT	14.9	7.9	1.5	1.6	9.1	8.0	0.4	6.2	6.2
NIS	-178.7	-181.9	-188.2	-197.7	-	-197.3	1.4	14.6	14.6
NIST	1.9	0.6	-0.3	-1.7	-1.8	-1.0	0.4	4.0	4.0
NMC, A*STAR	-2.4	-6.5	-3.9	1.3	14.6	5.0	0.4	5.6	5.6
NMIA	8.4	-3.0	-9.7	-14.6	-7.8	-21.3	0.4	6.2	6.2
NMIJ AIST	-1.4	-1.2	-0.9	-0.6	-1.0	-0.9	0.4	6.0	6.0
NMIM	73.4	96.8	117.6	143.3	138.5	70.3	0.4	5.6	5.6
NMISA	10.7	3.3	10.3	-6.6	-7.3	-15.8	39.8	7.4	40.4
NPL	0.6	0.7	-0.9	-0.6	0.9	1.1	0.4	2.4	2.4
NPLI	-1.2	-0.8	-1.2	-1.3	-1.7	-1.6	0.4	5.6	5.6
NRC	-1.7	-1.9	-2.1	-1.9	-1.5	-1.2	0.4	5.6	5.6
NSAI NML	-52.9	-45.8	-39.3	-34.2	-30.0	-24.6	0.4	14.8	14.8
NSC IM	-	-10.7	-	-	-	-4.4	6.0	15.0	16.2
ON/DSHO	3.0	-0.5	-3.0	2.1	3.6	0.7	0.4	6.8	6.8
PTB	0.4	0.6	0.7	0.9	0.8	1.1	0.4	2.4	2.4
RISE	0.8	0.9	1.1	1.7	1.6	1.5	0.4	2.4	2.4
ROA	-1.9	-1.3	0.3	1.0	2.5	4.1	0.4	2.4	2.4
SASO-NMCC	900.4	902.9	899.9	893.1	892.2	896.7	0.4	7.8	7.8
SCL	-11.6	-10.8	-9.9	-1.8	5.0	7.8	0.4	7.6	7.6
SMD	0.7	1.1	1.3	1.5	1.6	1.9	0.4	9.2	9.2
SMU	417.7	429.8	406.7	370.3	364.7	280.2	3.0	12.4	12.8
SNSU-BSN	4982.1	5145.8	5297.4	5442.9	5607.3	5771.2	0.8	NC	- (*)
TL	-0.9	0.1	-0.2	-0.2	-0.1	0.2	0.4	4.0	4.0
UME	-108.8	-109.2	-109.2	-111.7	-27.8	24.1	0.4	8.4	8.4
UzNIM	112.9	119.1	111.6	105.0	107.9	109.7	0.4	5.6	5.6
VMI-STAMEQ	104.2	99.2	103.2	88.1	42.5	-28.2	1.4	6.0	6.2
VNIIFTRI	1.4	1.3	1.4	1.2	1.1	1.2	0.4	5.0	5.0
VSL	-0.1	-2.6	-4.5	-4.2	-1.2	1.6	0.4	2.6	2.6

ZMDM -10.9 -3.6 -15.3 -31.6 -28.9 -30.6 0.4 15.0 15.0

(*) U_a expanded uncertainty guarantees only the traceability in frequency