

Final Report of the Key Comparison

APMP.AUV.V-K2

Takashi Usuda¹, Akihiro Ota¹, Hideaki Nozato¹, Tamio Ishigami¹, Wataru Kokuyama¹

Virat Plangsangmas², Pairoj Rattanangkul²,

Kuang Yih Tsuei³, Yu-Chung Huang³,

Chan Chee Keong⁴, Cui Shan⁴

1: NMIJ/AIST (Japan)

2: NIMT (Thailand)

3: CMS/ITRI (Chinese Taipei)

4: NMC, A*STAR (Singapore)

Abstract

A key comparison of vibration acceleration APMP.AUV.V-K2, which is upgraded from APMP.AUV.V-K1.1, has been made within the Asia Pacific Metrology Programme (APMP) to include three national laboratories; the CMS-ITRI (Chinese Taipei), the NIMT (Thailand) and the A*Star (Singapore). Only one pilot laboratory was the NMIJ (Japan) that participated to link the RMO APMP results to the CIPM CC key comparison (CCAUV.V-K2), according to the decision of CCAUV10/D7 in 10th CCAUV meeting. The admissible acceleration amplitude ranges from 10 m/s² to 200 m/s² over the frequency range from 40 Hz to 5 kHz. The RMO APMP results demonstrate the agreement with the key comparison reference value of CCAUV.V-K2 within the expanded uncertainties considering the armature effect of vibration exciter in high-frequency range.

List of contents

| | |
|---|----|
| 1. Introduction..... | 3 |
| 2. Participants..... | 3 |
| 3. Task and Purpose of the comparison..... | 3 |
| 4. Condition and Instruction of Measurement | 4 |
| 5. Transfer Standards as Artifact | 4 |
| 6. Circulation of the Artifacts..... | 5 |
| 7. Results of the Measurements | 5 |
| 7.1 Monitoring of stability | 5 |
| 7.2 Results of the Participants..... | 7 |
| 8. Analysis of the results | 9 |
| 8.1 Unilateral degree of equivalence between the participants | 9 |
| 8.2 Bilateral degree of equivalence between the participants | 17 |
| 8.3 Linking procedure and Degree of Equivalence with the KCRV | 28 |
| 9. Conclusion | 34 |
| 10. References..... | 34 |
| Annex 1 Technical Protocol of the APMP Key Comparison | 35 |
| 1. Task and purpose of the comparison | 35 |
| 2. Pilot laboratory | 35 |
| 3. Participants..... | 36 |
| 4. Devices under Test and Measurement conditions | 38 |
| 5. Circulation and Transportation | 39 |
| 6. Measurement and Analysis Instructions | 39 |
| 7. Communication of the Results to the Pilot laboratory | 40 |
| 8. Remarks on the post processing..... | 41 |
| 9. References..... | 41 |
| Annex 2 Measurement Uncertainty Budgets reported by the participants..... | 42 |
| Annex 3 DoE to the KCRV of CCAUV.V-K2 for the BB transducer | 72 |

1. Introduction

This report presents the results of the APMP comparison in the area of ‘vibration’, which in this case means sinusoidal acceleration. It was organized to disseminate the key comparison reference values, which were established earlier in the CIPM key comparison CCAUV.V-K1 [1] and CCAUV.V-K2 [2] within the RMO APMP.

In this report the relationship between the results of the participants and the results of the CIPM comparison in the field of vibration CCAUV.V-K2 is calculated via a procedure of ‘linking’. Only one pilot laboratory NMIJ, which participated in both the CIPM comparisons, acted as the linking laboratory for that process by the decision of CCAUV10/D7 [3] according to the resolution of 10th CCAUV meeting. In order to avoid an extra bias between NMIJ result and KCRV of CCAUV.V-K2, the linking factor is defined as a ratio of two NMIJ results in CCAUV.V-K2 and APMP.AUV.V-K2. Using the linking factor, the RMO APMP results of three participants were directly compared with the results of CCAUV.V-K2.

The Technical Protocol (c.f. App. 1) specifies in detail the aim and the task of the comparison, the conditions of measurement, the transfer standards used, measurement instructions, time schedule and other items. A brief summary is given in the following sections.

2. Participants

Four National Metrology Institutes (NMIs) in APMP participated in the RMO APMP.AUV.V-K2 comparison.

Table 1 Participants of the RMO APMP.AUV.V-K2 comparison

| Participant (Laboratory name) | Acronym | Economy | Economy code |
|--|-------------|----------------|--------------|
| National Metrology Institute of Japan | NMIJ | Japan | JP |
| Center for Measurement Standards/Industrial Technology Research Institute | CMS/ITRI | Chinese Taipei | TW |
| National Institute of Metrology Thailand | NIMT | Thailand | TH |
| National Metrology Centre, Agency for Science, Technology and Research | NMC, A*Star | Singapore | SG |

3. Task and Purpose of the comparison

The main purpose of this RMO-level comparison of the APMP in the field of vibration acceleration was to disseminate the key comparison reference to three NMIs namely CMS/ITRI, NIMT,

NMC A*Star. It was also planned for CMS/ITRI to confirm their Calibration Measurement Capabilities (CMCs) of their improved primary calibration system since CMS/ITRI participated to the APMP.AUV.V-K1 [4].

In order to provide the necessary means for a linking of this RMO comparison APMP.AUV.V-K2 results to the CIPM CCAUV.V-K2 key comparison, NMIJ was requested to volunteer as the Pilot and Linking laboratory. The principal task of the comparison is to measure the charge sensitivity of two accelerometers (one of single-ended design and one of back-to-back design) at different frequencies and acceleration amplitudes specified in section 4.

The charge sensitivity is calculated as the ratio of the amplitude of the output charge of the accelerometer to the amplitude of the acceleration at its reference surface. The reference surface is the base or mounting surface of the accelerometer of single-ended design, and the top surface of the accelerometer of back-to-back design. The charge sensitivity is given in pico coulombs per metres per second squared: $\text{pC}/(\text{m}/\text{s}^2)$.

To calibrate the two accelerometers, each primary vibration calibration by laser interferometry in accordance with ISO 16063-11:1999 [5] was carried out.

4. Condition and Instruction of Measurement

All major conditions were specified in compliance with the relevant CC Key comparison CCAUV.V-K2 as follows;

- Frequencies (all values in Hz): 40, 50, 63, 80, 100, 125, 160, 200, 250, 315, 400, 500, 630, 800, 1000, 1250, 1600, 2000, 2500, 3150, 4000, 5000 (160 Hz is reference frequency)
- Acceleration amplitudes: A range of $10 \text{ m}/\text{s}^2$ to $200 \text{ m}/\text{s}^2$ is admissible.
- Ambient temperature and accelerometer temperature during the calibration:
(23 ± 2) °C (actual values to be stated within tolerances of ± 0.3 °C), relative humidity: max. 75 %
- Mounting torque of the accelerometer: (2.0 ± 0.1) N·m

The comparison was also consistent with the “Guidelines for CIPM key comparisons” [6].

5. Transfer Standards as Artifact

Two types of piezoelectric standard accelerometers were used as transfer standards.

- One transfer standard accelerometer (single-ended), type 8305-001, S/N 2644378 (manufacturer: Brüel & Kjær), subsequently named SE-transducer.
- One reference standard accelerometer (back-to-back) type 8305 S, S/N 2691032 (manufacturer: Brüel & Kjær), subsequently named BB-transducer.

The investigation of the long-term stability was continued throughout the circulation period. The results of the NMIJ stability measurements and other individual data of the transfer standards are given in Section 7.

6. Circulation of the Artifacts

The circulation Type of this comparison was a modified star type. During the comparison, an interim monitoring at the pilot laboratory was done to check the stability of the artifacts (c.f. section 7.1) and the state of the mounting surface. The artifacts were hand-carried in a closed box by a representative of the NMI.

Table 2 Circulation of the artifacts

| Participant | Economy | Planned calibration period |
|----------------------------|----------------|----------------------------|
| NMIJ | Japan | August, 2011 |
| CMS/ITRI | Chinese Taipei | September, 2011 |
| Interim monitoring at NMIJ | | |
| NMC, A*Star | Singapore | January, 2012 |
| NIMT | Thailand | March, 2012 |
| Final monitoring at NMIJ | | |

7. Results of the Measurements

7.1 Monitoring of stability

At the beginning of this comparison, it was planned to transport the artifacts by air parcel, starting in 2010. However, the pilot laboratory observed significant sensitivity drift on the back-to-back accelerometer after the comparison was completed. It was not obvious when and how the drift had occurred during the period of the comparison, but it was clear that the reported results from participants showed much larger deviation than the observed one at the pilot laboratory. We decided then to replace both the artifacts (single-ended and back-to-back) to restart the comparison. It was also decided to hand-carry the artifacts in a closed box by a representative to avoid any damage on the artifacts.

As shown in the table 2, we also adopted a modified star type circulation to monitor if any sensitivity

drift or physical damage has occurred on the artifacts.

Table 3 shows the measurements results of the pilot laboratory at the reference frequency (160 Hz) for the time of beginning, interim and final monitoring.

Table 3 Charge sensitivities of the accelerometers at 160 Hz during the comparison measurements.

| Measurement period | Back-to-back | | Single-ended | |
|--------------------|---|---------------------|---|---------------------|
| | Sq in $\mu\text{C}/(\text{m}/\text{s}^2)$ | rel. exp. unc. in % | Sq in $\mu\text{C}/(\text{m}/\text{s}^2)$ | rel. exp. unc. in % |
| Beginning | 0.12773 | 0.24 | 0.12913 | 0.39 |
| Interim | 0.12749 | 0.24 | 0.12891 | 0.39 |
| Final | 0.12744 | 0.24 | 0.12882 | 0.39 |

These monitoring measurements can be summarized in the simplest way by the statistical properties given in table 4. This analysis indicates that the stability of the artifacts was acceptable considering the uncertainty claimed although the transportation did have some negative influence.

Table 4 Mean and its relative standard deviation of the charge sensitivity of the artifacts at 160 Hz calculated from the monitoring measurements

| Back-to-back | | Single-ended | |
|---|---------------------|---|---------------------|
| Long term mean in $\mu\text{C}/(\text{m}/\text{s}^2)$ | rel. std. dev. in % | Long term mean in $\mu\text{C}/(\text{m}/\text{s}^2)$ | rel. std. dev. in % |
| 0.12755 | 0.12 | 0.12895 | 0.12 |

7.2 Results of the Participants

7.2.1 Results of the Back-to-Back Accelerometer

Table 7.2.1-1 Results of the Back-to-Back (BB) accelerometer

| Frequency in Hz | NMIJ | | CMS/ITRI | | NIMT | | A*STAR | |
|--------------------|---------------------------------|------------------------------|---------------------------------|------------------------------|---------------------------------|------------------------------|---------------------------------|------------------------------|
| | Sq in pC/(m/s ²) | rel. exp. unc. in % | Sq in pC/(m/s ²) | rel. exp. unc. in % | Sq in pC/(m/s ²) | rel. exp. unc. in % | Sq in pC/(m/s ²) | rel. exp. unc. in % |
| 40 | 0.12780 | 0.24 | 0.12794 | 0.31 | 0.12742 | 0.60 | 0.12740 | 0.46 |
| 50 | 0.12775 | 0.24 | 0.12788 | 0.28 | 0.12749 | 0.44 | | |
| 63 | 0.12788 | 0.24 | 0.12793 | 0.25 | 0.12752 | 0.43 | | |
| 80 | 0.12766 | 0.24 | 0.12791 | 0.24 | 0.12757 | 0.43 | 0.12753 | 0.48 |
| 100 | 0.12766 | 0.37 | 0.12792 | 0.27 | 0.12761 | 0.43 | | |
| 125 | 0.12773 | 0.24 | 0.12787 | 0.25 | 0.12762 | 0.38 | | |
| 160 | 0.12773 | 0.24 | 0.12793 | 0.27 | 0.12760 | 0.38 | 0.12760 | 0.49 |
| 200 | 0.12771 | 0.24 | 0.12791 | 0.27 | 0.12763 | 0.38 | | |
| 250 | 0.12776 | 0.24 | 0.12785 | 0.28 | 0.12764 | 0.38 | | |
| 315 | 0.12772 | 0.24 | 0.12794 | 0.31 | 0.12764 | 0.38 | 0.12748 | 0.48 |
| 400 | 0.12752 | 0.24 | 0.12791 | 0.29 | 0.12763 | 0.38 | | |
| 500 | 0.12763 | 0.24 | 0.12795 | 0.33 | 0.12759 | 0.40 | | |
| 630 | 0.12777 | 0.24 | 0.12795 | 0.30 | 0.12761 | 0.40 | 0.12763 | 0.48 |
| 800 | 0.12778 | 0.24 | 0.12790 | 0.35 | 0.12755 | 0.40 | | |
| 1000 | 0.12785 | 0.24 | 0.12790 | 0.37 | 0.12764 | 0.76 | | |
| 1250 | 0.12787 | 0.24 | 0.12785 | 0.39 | 0.12768 | 0.76 | 0.12755 | 0.48 |
| 1600 | 0.12803 | 0.24 | 0.12789 | 0.27 | 0.12777 | 0.76 | | |
| 2000 | 0.12798 | 0.24 | 0.12787 | 0.32 | 0.12801 | 0.76 | | |
| 2500 | 0.12836 | 0.24 | 0.12789 | 0.33 | 0.12814 | 0.76 | 0.12785 | 0.57 |
| 3150 | 0.12883 | 0.24 | 0.12819 | 0.46 | 0.12851 | 0.77 | | |
| 4000 | 0.12917 | 0.25 | 0.12921 | 0.80 | 0.12903 | 0.77 | | |
| 5000 | 0.13004 | 0.29 | 0.12948 | 0.87 | 0.12998 | 0.77 | 0.13128 | 3.1 |

7.2.2 Results of the Single-ended Accelerometer

Table 7.2.2-1 Results of the Single-ended (SE) accelerometer

| Frequency in Hz | NMIJ | | CMS/ITRI | | NIMT | | A*STAR | |
|--------------------|---------------------------------|------------------------------|---------------------------------|------------------------------|---------------------------------|------------------------------|---------------------------------|------------------------------|
| | Sq in pC/(m/s ²) | rel. exp. unc. in % | Sq in pC/(m/s ²) | rel. exp. unc. in % | Sq in pC/(m/s ²) | rel. exp. unc. in % | Sq in pC/(m/s ²) | rel. exp. unc. in % |
| 40 | 0.12914 | 0.39 | 0.12943 | 0.3 | 0.12866 | 0.61 | 0.12877 | 0.46 |
| 50 | 0.12915 | 0.47 | 0.12931 | 0.29 | 0.12875 | 0.44 | | |
| 63 | 0.12919 | 0.40 | 0.12935 | 0.25 | 0.12878 | 0.43 | | |
| 80 | 0.12911 | 0.39 | 0.12932 | 0.24 | 0.12884 | 0.43 | 0.12888 | 0.48 |
| 100 | 0.12913 | 0.42 | 0.12934 | 0.25 | 0.12888 | 0.43 | | |
| 125 | 0.12913 | 0.39 | 0.12925 | 0.25 | 0.12893 | 0.38 | | |
| 160 | 0.12913 | 0.39 | 0.12936 | 0.25 | 0.12893 | 0.38 | 0.12878 | 0.65 |
| 200 | 0.12911 | 0.40 | 0.12927 | 0.26 | 0.12892 | 0.38 | | |
| 250 | 0.12909 | 0.39 | 0.12925 | 0.27 | 0.12888 | 0.38 | | |
| 315 | 0.12918 | 0.39 | 0.12931 | 0.3 | 0.12893 | 0.38 | 0.12888 | 0.49 |
| 400 | 0.12916 | 0.39 | 0.12935 | 0.29 | 0.12895 | 0.38 | | |
| 500 | 0.12920 | 0.39 | 0.12942 | 0.3 | 0.12890 | 0.40 | | |
| 630 | 0.12921 | 0.39 | 0.12945 | 0.3 | 0.12898 | 0.40 | 0.12897 | 0.52 |
| 800 | 0.12923 | 0.39 | 0.12928 | 0.33 | 0.12893 | 0.40 | | |
| 1000 | 0.12939 | 0.39 | 0.12941 | 0.43 | 0.12903 | 0.76 | | |
| 1250 | 0.12945 | 0.39 | 0.12939 | 0.39 | 0.12902 | 0.76 | 0.12905 | 0.49 |
| 1600 | 0.12958 | 0.39 | 0.12942 | 0.29 | 0.12927 | 0.76 | | |
| 2000 | 0.12976 | 0.39 | 0.12935 | 0.34 | 0.12938 | 0.77 | | |
| 2500 | 0.13015 | 0.39 | 0.12940 | 0.32 | 0.12972 | 0.77 | 0.1301 | 0.56 |
| 3150 | 0.13085 | 0.39 | 0.13031 | 0.47 | 0.13021 | 0.77 | | |
| 4000 | 0.13170 | 0.40 | 0.13149 | 0.8 | 0.13100 | 0.76 | | |
| 5000 | 0.13274 | 0.41 | 0.13205 | 0.9 | 0.13235 | 0.77 | 0.13318 | 1.13 |

8. Analysis of the results

8.1 Unilateral degree of equivalence between the participants

In order to compare the individual results of the participating laboratories of this comparison with the linking laboratory (NMIJ), the unilateral degrees of equivalence (DoE) of pairs of results with respect to a certain frequency were calculated. For the purpose, we signify the measurand in this sequent KC as Y . The values $(y_1, u(y_1)), \dots, (y_M, u(y_M))$ denote the charge sensitivities and associated standard uncertainties of the participants. These DoE are each a pair of values of the difference D_i between the respective participants i and NMIJ including the expanded uncertainty U_i of this difference. These values are calculated for each frequency according to:

$$D_i = y_i - y_{NMIJ} \quad (1)$$

$$U_i = k \cdot \sqrt{u^2(y_i) + u^2(y_{NMIJ})} \quad (2)$$

with a coverage factor of $k = 2$.

8.1.1 Unilateral DoE of the back-to-back accelerometer

Table 8.1.1-1 Unilateral Degrees of equivalence to NMIJ for the BB at 40 Hz

| 40 Hz | CMS/ITRI | | NIMT | | A*STAR | |
|-----------------|--|-------|--|-------|--|-------|
| $i \rightarrow$ | D_i | U_i | D_i | U_i | D_{ij} | U_i |
| | in pC/(m/s ²)·10 ⁻⁴ | | in pC/(m/s ²)·10 ⁻⁴ | | in pC/(m/s ²)·10 ⁻⁴ | |
| NMIJ | 1.4 | 5.0 | -3.8 | 8.2 | -4.0 | 6.6 |

Table 8.1.1-2 Degrees of equivalence to NMIJ for the BB at 50 Hz

| 50 Hz | CMS/ITRI | | NIMT | | A*STAR | |
|-----------------|--|-------|--|-------|--|-------|
| $i \rightarrow$ | D_i | U_i | D_i | U_i | D_i | U_i |
| | in pC/(m/s ²)·10 ⁻⁴ | | in pC/(m/s ²)·10 ⁻⁴ | | in pC/(m/s ²)·10 ⁻⁴ | |
| NMIJ | 1.3 | 4.7 | -2.6 | 6.4 | | |

Table 8.1.1-3 Degrees of equivalence to NMIJ for the BB at 63 Hz

| 63 Hz | CMS/ITRI | | NIMT | | A*STAR | |
|-----------------|--|-------|--|-------|--|-------|
| $i \rightarrow$ | D_i | U_i | D_{ij} | U_i | D_{ij} | U_i |
| | in pC/(m/s ²)·10 ⁻⁴ | | in pC/(m/s ²)·10 ⁻⁴ | | in pC/(m/s ²)·10 ⁻⁴ | |
| NMIJ | 0.5 | 4.4 | -3.6 | 6.3 | | |

Table 8.1.1-4 Degrees of equivalence to NMIJ for the BB at 80 Hz

| 80 Hz | CMS/ITRI | | NIMT | | A*STAR | |
|-----------------|--|-------|--|-------|--|-------|
| $i \rightarrow$ | D_i | U_i | D_i | U_i | D_i | U_i |
| | in pC/(m/s ²)·10 ⁻⁴ | | in pC/(m/s ²)·10 ⁻⁴ | | in pC/(m/s ²)·10 ⁻⁴ | |
| NMIJ | 2.5 | 4.3 | -0.9 | 6.3 | -1.3 | 6.8 |

Table 8.1.1-5 Degrees of equivalence to NMIJ for the BB at 100 Hz

| 100 Hz | CMS/ITRI | | NIMT | | A*STAR | |
|-----------------|--|-------|--|-------|--|-------|
| $i \rightarrow$ | D_i | U_i | D_i | U_i | D_i | U_i |
| | in pC/(m/s ²)·10 ⁻⁴ | | in pC/(m/s ²)·10 ⁻⁴ | | in pC/(m/s ²)·10 ⁻⁴ | |
| NMIJ | 2.6 | 5.8 | -0.5 | 7.2 | | |

Table 8.1.1-6 Degrees of equivalence to NMIJ for the BB at 125 Hz

| 125 Hz | CMS/ITRI | | NIMT | | A*STAR | |
|-----------------|--|-------|--|-------|--|-------|
| $i \rightarrow$ | D_i | U_i | D_i | U_i | D_i | U_i |
| | in pC/(m/s ²)·10 ⁻⁴ | | in pC/(m/s ²)·10 ⁻⁴ | | in pC/(m/s ²)·10 ⁻⁴ | |
| NMIJ | 1.4 | 4.4 | -1.1 | 5.7 | | |

Table 8.1.1-7 Degrees of equivalence to NMIJ for the BB at 160 Hz

| 160 Hz | CMS/ITRI | | NIMT | | A*STAR | |
|-----------------|--|-------|--|-------|--|-------|
| $i \rightarrow$ | D_i | U_i | D_i | U_i | D_i | U_i |
| | in pC/(m/s ²)·10 ⁻⁴ | | in pC/(m/s ²)·10 ⁻⁴ | | in pC/(m/s ²)·10 ⁻⁴ | |
| NMIJ | 2.0 | 4.6 | -1.3 | 5.7 | -1.3 | 6.9 |

Table 8.1.1-8 Degrees of equivalence to NMIJ for the BB at 200 Hz

| 200 Hz | CMS/ITRI | | NIMT | | A*STAR | |
|-----------------|--|-------|--|-------|--|-------|
| $i \rightarrow$ | D_i | U_i | D_i | U_i | D_i | U_i |
| | in pC/(m/s ²)·10 ⁻⁴ | | in pC/(m/s ²)·10 ⁻⁴ | | in pC/(m/s ²)·10 ⁻⁴ | |
| NMIJ | 2.0 | 4.6 | -0.8 | 5.7 | | |

Table 8.1.1-9 Degrees of equivalence to NMIJ for the BB at 250 Hz

| 250 Hz | CMS/ITRI | | NIMT | | A*STAR | |
|-----------------|---|-------|---|-------|---|-------|
| $i \rightarrow$ | D_i | U_i | D_i | U_i | D_i | U_i |
| | in $\text{pC}/(\text{m/s}^2) \cdot 10^{-4}$ | | in $\text{pC}/(\text{m/s}^2) \cdot 10^{-4}$ | | in $\text{pC}/(\text{m/s}^2) \cdot 10^{-4}$ | |
| NMIJ | 0.9 | 4.7 | -1.2 | 5.7 | | |

Table 8.1.1-10 Degrees of equivalence to NMIJ for the BB at 315 Hz

| 315 Hz | CMS/ITRI | | NIMT | | A*STAR | |
|-----------------|---|-------|---|-------|---|-------|
| $i \rightarrow$ | D_i | U_i | D_i | U_i | D_i | U_i |
| | in $\text{pC}/(\text{m/s}^2) \cdot 10^{-4}$ | | in $\text{pC}/(\text{m/s}^2) \cdot 10^{-4}$ | | in $\text{pC}/(\text{m/s}^2) \cdot 10^{-4}$ | |
| NMIJ | 2.2 | 5.0 | -0.7 | 5.7 | -2.4 | 6.8 |

Table 8.1.1-11 Degrees of equivalence to NMIJ for the BB at 400 Hz

| 400 Hz | CMS/ITRI | | NIMT | | A*STAR | |
|-----------------|---|-------|---|-------|---|-------|
| $i \rightarrow$ | D_i | U_i | D_i | U_i | D_i | U_i |
| | in $\text{pC}/(\text{m/s}^2) \cdot 10^{-4}$ | | in $\text{pC}/(\text{m/s}^2) \cdot 10^{-4}$ | | in $\text{pC}/(\text{m/s}^2) \cdot 10^{-4}$ | |
| NMIJ | 3.9 | 4.8 | 1.0 | 5.7 | | |

Table 8.1.1-12 Degrees of equivalence to NMIJ for the BB at 500 Hz

| 500 Hz | CMS/ITRI | | NIMT | | A*STAR | |
|-----------------|---|-------|---|-------|---|-------|
| $i \rightarrow$ | D_i | U_i | D_i | U_i | D_i | U_i |
| | in $\text{pC}/(\text{m/s}^2) \cdot 10^{-4}$ | | in $\text{pC}/(\text{m/s}^2) \cdot 10^{-4}$ | | in $\text{pC}/(\text{m/s}^2) \cdot 10^{-4}$ | |
| NMIJ | 3.2 | 5.2 | -0.4 | 5.9 | | |

Table 8.1.1-13 Degrees of equivalence to NMIJ for the BB at 630 Hz

| 630 Hz | CMS/ITRI | | NIMT | | A*STAR | |
|-----------------|---|-------|---|-------|---|-------|
| $i \rightarrow$ | D_i | U_i | D_i | U_i | D_i | U_i |
| | in $\text{pC}/(\text{m/s}^2) \cdot 10^{-4}$ | | in $\text{pC}/(\text{m/s}^2) \cdot 10^{-4}$ | | in $\text{pC}/(\text{m/s}^2) \cdot 10^{-4}$ | |
| NMIJ | 1.8 | 4.9 | -1.6 | 5.9 | -1.4 | 6.8 |

Table 8.1.1-14 Degrees of equivalence to NMIJ for the BB at 800 Hz

| 800 Hz | CMS/ITRI | | NIMT | | A*STAR | |
|-----------------|---|-------|---|-------|---|-------|
| $i \rightarrow$ | D_i | U_i | D_i | U_i | D_i | U_i |
| | in $\text{pC}/(\text{m/s}^2) \cdot 10^{-4}$ | | in $\text{pC}/(\text{m/s}^2) \cdot 10^{-4}$ | | in $\text{pC}/(\text{m/s}^2) \cdot 10^{-4}$ | |
| NMIJ | 1.2 | 5.4 | -2.2 | 5.9 | | |

Table 8.1.1-15 Degrees of equivalence to NMIJ for the BB at 1000 Hz

| 1000 Hz | CMS/ITRI | | NIMT | | A*STAR | |
|-----------------|---|-------|---|-------|---|-------|
| $i \rightarrow$ | D_i | U_i | D_i | U_i | D_i | U_i |
| | in $\text{pC}/(\text{m/s}^2) \cdot 10^{-4}$ | | in $\text{pC}/(\text{m/s}^2) \cdot 10^{-4}$ | | in $\text{pC}/(\text{m/s}^2) \cdot 10^{-4}$ | |
| NMIJ | 0.5 | 5.6 | -2.1 | 10.2 | | |

Table 8.1.1-16 Degrees of equivalence to NMIJ for the BB at 1250 Hz

| 1250 Hz | CMS/ITRI | | NIMT | | A*STAR | |
|-----------------|---|-------|---|-------|---|-------|
| $i \rightarrow$ | D_i | U_i | D_i | U_i | D_i | U_i |
| | in $\text{pC}/(\text{m/s}^2) \cdot 10^{-4}$ | | in $\text{pC}/(\text{m/s}^2) \cdot 10^{-4}$ | | in $\text{pC}/(\text{m/s}^2) \cdot 10^{-4}$ | |
| NMIJ | -0.2 | 5.8 | -2.0 | 10.2 | -3.2 | 6.8 |

Table 8.1.1-17 Degrees of equivalence to NMIJ for the BB at 1600 Hz

| 1600 Hz | CMS/ITRI | | NIMT | | A*STAR | |
|-----------------|---|-------|---|-------|---|-------|
| $i \rightarrow$ | D_i | U_i | D_i | U_i | D_i | U_i |
| | in $\text{pC}/(\text{m/s}^2) \cdot 10^{-4}$ | | in $\text{pC}/(\text{m/s}^2) \cdot 10^{-4}$ | | in $\text{pC}/(\text{m/s}^2) \cdot 10^{-4}$ | |
| NMIJ | -1.4 | 4.6 | -2.6 | 10.2 | | |

Table 8.1.1-18 Degrees of equivalence to NMIJ for the BB at 2000 Hz

| 2000 Hz | CMS/ITRI | | NIMT | | A*STAR | |
|-----------------|---|-------|---|-------|---|-------|
| $i \rightarrow$ | D_i | U_i | D_i | U_i | D_i | U_i |
| | in $\text{pC}/(\text{m/s}^2) \cdot 10^{-4}$ | | in $\text{pC}/(\text{m/s}^2) \cdot 10^{-4}$ | | in $\text{pC}/(\text{m/s}^2) \cdot 10^{-4}$ | |
| NMIJ | -1.1 | 5.1 | 0.3 | 10.2 | | |

Table 8.1.1-19 Degrees of equivalence to NMIJ for the BB at 2500 Hz

| 2500 Hz | CMS/ITRI | | NIMT | | A*STAR | |
|-----------------|---|-------|---|-------|---|-------|
| $i \rightarrow$ | D_i | U_i | D_i | U_i | D_i | U_i |
| | in $\text{pC}/(\text{m/s}^2) \cdot 10^{-4}$ | | in $\text{pC}/(\text{m/s}^2) \cdot 10^{-4}$ | | in $\text{pC}/(\text{m/s}^2) \cdot 10^{-4}$ | |
| NMIJ | -4.6 | 5.2 | -2.2 | 10.2 | -5.1 | 7.9 |

Table 8.1.1-20 Degrees of equivalence to NMIJ for the BB at 3150 Hz

| 3150 Hz | CMS/ITRI | | NIMT | | A*STAR | |
|-----------------|---|-------|---|-------|---|-------|
| $i \rightarrow$ | D_i | U_i | D_i | U_i | D_i | U_i |
| | in $\text{pC}/(\text{m/s}^2) \cdot 10^{-4}$ | | in $\text{pC}/(\text{m/s}^2) \cdot 10^{-4}$ | | in $\text{pC}/(\text{m/s}^2) \cdot 10^{-4}$ | |
| NMIJ | -6.4 | 6.7 | -3.2 | 10.4 | | |

Table 8.1.1-21 Degrees of equivalence to NMIJ for the BB at 4000 Hz

| 4000 Hz | CMS/ITRI | | NIMT | | A*STAR | |
|-----------------|---|-------|---|-------|---|-------|
| $i \rightarrow$ | D_i | U_i | D_i | U_i | D_i | U_i |
| | in $\text{pC}/(\text{m/s}^2) \cdot 10^{-4}$ | | in $\text{pC}/(\text{m/s}^2) \cdot 10^{-4}$ | | in $\text{pC}/(\text{m/s}^2) \cdot 10^{-4}$ | |
| NMIJ | 0.4 | 10.8 | -1.4 | 10.5 | | |

Table 8.1.1-22 Degrees of equivalence to NMIJ for the BB at 5000 Hz

| 5000 Hz | CMS/ITRI | | NIMT | | A*STAR | |
|-----------------|---|-------|---|-------|---|-------|
| $i \rightarrow$ | D_i | U_i | D_i | U_i | D_i | U_i |
| | in $\text{pC}/(\text{m/s}^2) \cdot 10^{-4}$ | | in $\text{pC}/(\text{m/s}^2) \cdot 10^{-4}$ | | in $\text{pC}/(\text{m/s}^2) \cdot 10^{-4}$ | |
| NMIJ | -5.6 | 11.9 | -0.6 | 10.7 | 12.4 | 40.9 |

8.1.2 Unilateral DoE of the single-ended (SE) accelerometer

Table 8.1.2-1 Unilateral Degrees of equivalence to NMIJ for the SE at 40 Hz

| 40 Hz | CMS/ITRI | | NIMT | | A*STAR | |
|-----------------|---|-------|---|-------|---|-------|
| $i \rightarrow$ | D_i | U_i | D_i | U_i | D_i | U_i |
| | in $\text{pC}/(\text{m/s}^2) \cdot 10^{-4}$ | | in $\text{pC}/(\text{m/s}^2) \cdot 10^{-4}$ | | in $\text{pC}/(\text{m/s}^2) \cdot 10^{-4}$ | |
| NMIJ | 2.9 | 6.4 | -4.8 | 9.4 | -3.7 | 7.8 |

Table 8.1.2-2 Degrees of equivalence to NMIJ for the SE at 50 Hz

| 50 Hz | CMS/ITRI | | NIMT | | A*STAR | |
|-----------------|---|-------|---|-------|---|-------|
| $i \rightarrow$ | D_i | U_i | D_i | U_i | D_i | U_i |
| | in $\text{pC}/(\text{m/s}^2) \cdot 10^{-4}$ | | in $\text{pC}/(\text{m/s}^2) \cdot 10^{-4}$ | | in $\text{pC}/(\text{m/s}^2) \cdot 10^{-4}$ | |
| NMIJ | 1.6 | 7.2 | -4.0 | 8.3 | | |

Table 8.1.2-3 Degrees of equivalence to NMIJ for the SE at 63 Hz

| 63 Hz | CMS/ITRI | | NIMT | | A*STAR | |
|-----------------|---|-------|---|-------|---|-------|
| $i \rightarrow$ | D_i | U_i | D_i | U_i | D_i | U_i |
| | in $\text{pC}/(\text{m/s}^2) \cdot 10^{-4}$ | | in $\text{pC}/(\text{m/s}^2) \cdot 10^{-4}$ | | in $\text{pC}/(\text{m/s}^2) \cdot 10^{-4}$ | |
| NMIJ | 1.6 | 6.1 | -4.0 | 7.6 | | |

Table 8.1.2-4 Degrees of equivalence to NMIJ for the SE at 80 Hz

| 80 Hz | CMS/ITRI | | NIMT | | A*STAR | |
|-----------------|---|-------|---|-------|---|-------|
| $i \rightarrow$ | D_i | U_i | D_i | U_i | D_i | U_i |
| | in $\text{pC}/(\text{m/s}^2) \cdot 10^{-4}$ | | in $\text{pC}/(\text{m/s}^2) \cdot 10^{-4}$ | | in $\text{pC}/(\text{m/s}^2) \cdot 10^{-4}$ | |
| NMIJ | 2.1 | 5.9 | -2.8 | 7.5 | -3.1 | 8.0 |

Table 8.1.2-5 Degrees of equivalence to NMIJ for the SE at 100 Hz

| 100 Hz | CMS/ITRI | | NIMT | | A*STAR | |
|-----------------|---|-------|---|-------|---|-------|
| $i \rightarrow$ | D_i | U_i | D_i | U_i | D_i | U_i |
| | in $\text{pC}/(\text{m/s}^2) \cdot 10^{-4}$ | | in $\text{pC}/(\text{m/s}^2) \cdot 10^{-4}$ | | in $\text{pC}/(\text{m/s}^2) \cdot 10^{-4}$ | |
| NMIJ | 2.0 | 6.3 | -2.5 | 7.7 | | |

Table 8.1.2-6 Degrees of equivalence to NMIJ for the SE at 125 Hz

| 125 Hz | CMS/ITRI | | NIMT | | A*STAR | |
|-----------------|---|-------|---|-------|---|-------|
| $i \rightarrow$ | D_i | U_i | D_i | U_i | D_i | U_i |
| | in $\text{pC}/(\text{m/s}^2) \cdot 10^{-4}$ | | in $\text{pC}/(\text{m/s}^2) \cdot 10^{-4}$ | | in $\text{pC}/(\text{m/s}^2) \cdot 10^{-4}$ | |
| NMIJ | 1.2 | 6.0 | -2.0 | 7.1 | | |

Table 8.1.2-7 Degrees of equivalence to NMIJ for the SE at 160 Hz

| 160 Hz | CMS/ITRI | | NIMT | | A*STAR | |
|-----------------|---|-------|---|-------|---|-------|
| $i \rightarrow$ | D_i | U_i | D_i | U_i | D_i | U_i |
| | in $\text{pC}/(\text{m/s}^2) \cdot 10^{-4}$ | | in $\text{pC}/(\text{m/s}^2) \cdot 10^{-4}$ | | in $\text{pC}/(\text{m/s}^2) \cdot 10^{-4}$ | |
| NMIJ | 2.2 | 6.0 | -2.0 | 7.1 | -3.5 | 9.8 |

Table 8.1.2-8 Degrees of equivalence to NMIJ for the SE at 200 Hz

| 200 Hz | CMS/ITRI | | NIMT | | A*STAR | |
|-----------------|---|-------|---|-------|---|-------|
| $i \rightarrow$ | D_i | U_i | D_i | U_i | D_i | U_i |
| | in $\text{pC}/(\text{m/s}^2) \cdot 10^{-4}$ | | in $\text{pC}/(\text{m/s}^2) \cdot 10^{-4}$ | | in $\text{pC}/(\text{m/s}^2) \cdot 10^{-4}$ | |
| NMIJ | 1.6 | 6.2 | -1.9 | 7.1 | | |

Table 8.1.2-9 Degrees of equivalence to NMIJ for the SE at 250 Hz

| 250 Hz | CMS/ITRI | | NIMT | | A*STAR | |
|-----------------|---|-------|---|-------|---|-------|
| $i \rightarrow$ | D_i | U_i | D_i | U_i | D_i | U_i |
| | in $\text{pC}/(\text{m/s}^2) \cdot 10^{-4}$ | | in $\text{pC}/(\text{m/s}^2) \cdot 10^{-4}$ | | in $\text{pC}/(\text{m/s}^2) \cdot 10^{-4}$ | |
| NMIJ | 1.6 | 6.2 | -2.1 | 7.1 | | |

Table 8.1.2-10 Degrees of equivalence to NMIJ for the SE at 315 Hz

| 315 Hz | CMS/ITRI | | NIMT | | A*STAR | |
|-----------------|---|-------|---|-------|---|-------|
| $i \rightarrow$ | D_i | U_i | D_i | U_i | D_i | U_i |
| | in $\text{pC}/(\text{m/s}^2) \cdot 10^{-4}$ | | in $\text{pC}/(\text{m/s}^2) \cdot 10^{-4}$ | | in $\text{pC}/(\text{m/s}^2) \cdot 10^{-4}$ | |
| NMIJ | 1.3 | 6.4 | -2.5 | 7.0 | -3.8 | 8.1 |

Table 8.1.2-11 Degrees of equivalence to NMIJ for the SE at 400 Hz

| 400 Hz | CMS/ITRI | | NIMT | | A*STAR | |
|-----------------|---|-------|---|-------|---|-------|
| $i \rightarrow$ | D_i | U_i | D_i | U_i | D_i | U_i |
| | in $\text{pC}/(\text{m/s}^2) \cdot 10^{-4}$ | | in $\text{pC}/(\text{m/s}^2) \cdot 10^{-4}$ | | in $\text{pC}/(\text{m/s}^2) \cdot 10^{-4}$ | |
| NMIJ | 1.9 | 6.3 | -2.1 | 7.0 | | |

Table 8.1.2-12 Degrees of equivalence to NMIJ for the SE at 500 Hz

| 500 Hz | CMS/ITRI | | NIMT | | A*STAR | |
|-----------------|---|-------|---|-------|---|-------|
| $i \rightarrow$ | D_i | U_i | D_i | U_i | D_i | U_i |
| | in $\text{pC}/(\text{m/s}^2) \cdot 10^{-4}$ | | in $\text{pC}/(\text{m/s}^2) \cdot 10^{-4}$ | | in $\text{pC}/(\text{m/s}^2) \cdot 10^{-4}$ | |
| NMIJ | 2.2 | 6.4 | -2.9 | 7.2 | | |

Table 8.1.2-13 Degrees of equivalence to NMIJ for the SE at 630 Hz

| 630 Hz | CMS/ITRI | | NIMT | | A*STAR | |
|-----------------|---|-------|---|-------|---|-------|
| $i \rightarrow$ | D_i | U_i | D_i | U_i | D_i | U_i |
| | in $\text{pC}/(\text{m/s}^2) \cdot 10^{-4}$ | | in $\text{pC}/(\text{m/s}^2) \cdot 10^{-4}$ | | in $\text{pC}/(\text{m/s}^2) \cdot 10^{-4}$ | |
| NMIJ | 2.3 | 6.4 | -2.4 | 7.2 | -2.4 | 8.4 |

Table 8.1.2-14 Degrees of equivalence to NMIJ for the SE at 800 Hz

| 800 Hz | CMS/ITRI | | NIMT | | A*STAR | |
|-----------------|---|-------|---|-------|---|-------|
| $i \rightarrow$ | D_i | U_i | D_i | U_i | D_i | U_i |
| | in $\text{pC}/(\text{m/s}^2) \cdot 10^{-4}$ | | in $\text{pC}/(\text{m/s}^2) \cdot 10^{-4}$ | | in $\text{pC}/(\text{m/s}^2) \cdot 10^{-4}$ | |
| NMIJ | 0.5 | 6.6 | -3.0 | 7.2 | | |

Table 8.1.2-15 Degrees of equivalence to NMIJ for the SE at 1000 Hz

| 1000 Hz | CMS/ITRI | | NIMT | | A*STAR | |
|-----------------|---|-------|---|-------|---|-------|
| $i \rightarrow$ | D_i | U_i | D_i | U_i | D_i | U_i |
| | in $\text{pC}/(\text{m/s}^2) \cdot 10^{-4}$ | | in $\text{pC}/(\text{m/s}^2) \cdot 10^{-4}$ | | in $\text{pC}/(\text{m/s}^2) \cdot 10^{-4}$ | |
| NMIJ | 0.3 | 7.5 | -3.5 | 11.0 | | |

Table 8.1.2-16 Degrees of equivalence to NMIJ for the SE at 1250 Hz

| 1250 Hz | CMS/ITRI | | NIMT | | A*STAR | |
|-----------------|--|-------|--|-------|--|-------|
| $i \rightarrow$ | D_i | U_i | D_i | U_i | D_i | U_i |
| | in pC/(m/s ²)·10 ⁻⁴ | | in pC/(m/s ²)·10 ⁻⁴ | | in pC/(m/s ²)·10 ⁻⁴ | |
| NMIJ | -0.6 | 7.2 | -4.2 | 11.0 | -4.0 | 8.1 |

Table 8.1.2-17 Degrees of equivalence to NMIJ for the SE at 1600 Hz

| 1600 Hz | CMS/ITRI | | NIMT | | A*STAR | |
|-----------------|--|-------|--|-------|--|-------|
| $i \rightarrow$ | D_i | U_i | D_i | U_i | D_i | U_i |
| | in pC/(m/s ²)·10 ⁻⁴ | | in pC/(m/s ²)·10 ⁻⁴ | | in pC/(m/s ²)·10 ⁻⁴ | |
| NMIJ | -1.6 | 6.3 | -3.1 | 11.1 | | |

Table 8.1.2-18 Degrees of equivalence to NMIJ for the SE at 2000 Hz

| 2000 Hz | CMS/ITRI | | NIMT | | A*STAR | |
|-----------------|--|-------|--|-------|--|-------|
| $i \rightarrow$ | D_i | U_i | D_i | U_i | D_i | U_i |
| | in pC/(m/s ²)·10 ⁻⁴ | | in pC/(m/s ²)·10 ⁻⁴ | | in pC/(m/s ²)·10 ⁻⁴ | |
| NMIJ | -4.1 | 6.7 | -3.8 | 11.2 | | |

Table 8.1.2-19 Degrees of equivalence to NMIJ for the SE at 2500 Hz

| 2500 Hz | CMS/ITRI | | NIMT | | A*STAR | |
|-----------------|--|-------|--|-------|--|-------|
| $i \rightarrow$ | D_i | U_i | D_i | U_i | D_i | U_i |
| | in pC/(m/s ²)·10 ⁻⁴ | | in pC/(m/s ²)·10 ⁻⁴ | | in pC/(m/s ²)·10 ⁻⁴ | |
| NMIJ | -7.5 | 6.6 | -4.3 | 11.2 | -0.5 | 8.9 |

Table 8.1.2-20 Degrees of equivalence to NMIJ for the SE at 3150 Hz

| 3150 Hz | CMS/ITRI | | NIMT | | A*STAR | |
|-----------------|--|-------|--|-------|--|-------|
| $i \rightarrow$ | D_i | U_i | D_i | U_i | D_i | U_i |
| | in pC/(m/s ²)·10 ⁻⁴ | | in pC/(m/s ²)·10 ⁻⁴ | | in pC/(m/s ²)·10 ⁻⁴ | |
| NMIJ | -5.4 | 8.0 | -6.4 | 11.3 | | |

Table 8.1.2-21 Degrees of equivalence to NMIJ for the SE at 4000 Hz

| 4000 Hz | CMS/ITRI | | NIMT | | A*STAR | |
|-----------------|--|-------|--|-------|--|-------|
| $i \rightarrow$ | D_i | U_i | D_i | U_i | D_i | U_i |
| | in pC/(m/s ²)·10 ⁻⁴ | | in pC/(m/s ²)·10 ⁻⁴ | | in pC/(m/s ²)·10 ⁻⁴ | |
| NMIJ | -2.1 | 11.8 | -7.0 | 11.3 | | |

Table 8.1.2-22 Degrees of equivalence to NMIJ for the SE at 5000 Hz

| 5000 Hz | CMS/ITRI | | NIMT | | A*STAR | |
|-----------------|--|-------|--|-------|--|-------|
| $i \rightarrow$ | D_i | U_i | D_i | U_i | D_i | U_i |
| | in pC/(m/s ²)·10 ⁻⁴ | | in pC/(m/s ²)·10 ⁻⁴ | | in pC/(m/s ²)·10 ⁻⁴ | |
| NMIJ | -6.9 | 13.1 | -3.9 | 11.6 | 4.4 | 16.0 |

8.2 Bilateral degree of equivalence between the participants

In order to bilaterally compare the individual results of the participating laboratories with one another, the bilateral degrees of equivalence (DoE) with respect to a certain frequency were calculated. Again, we signify the measurand in this sequent KC as Y . The values $(y_1, u(y_1)), \dots, (y_M, u(y_M))$ denote the charge sensitivities and associated standard uncertainties of the participants. These DoE are each a pair of values of the difference D_{ij} between the respective participants i and j and the expanded uncertainty U_{ij} of this difference. These values are calculated for each frequency according to:

$$D_{ij} = y_i - y_j \quad (3)$$

$$U_{ij} = k \cdot \sqrt{u^2(y_i) + u^2(y_j)} \quad (4)$$

with a coverage factor of $k = 2$.

8.2.1 Bilateral DoE of the back-to-back accelerometer

Table 8.2.1-1 Degrees of equivalence between the participants for the BB at 40 Hz

| 40 Hz | CMS/ITRI | | NIMT | | A*STAR | |
|-----------------|--|----------|--|----------|--|----------|
| $i \rightarrow$ | D_{ij} | U_{ij} | D_{ij} | U_{ij} | D_{ij} | U_{ij} |
| $j \downarrow$ | in pC/(m/s ²)·10 ⁻⁴ | | in pC/(m/s ²)·10 ⁻⁴ | | in pC/(m/s ²)·10 ⁻⁴ | |
| NMIJ | 1.4 | 5.0 | -3.8 | 8.2 | -4.0 | 6.6 |
| CMS/ITRI | | | -5.2 | 8.6 | -5.4 | 7.1 |
| NIMT | | | | | -0.2 | 9.6 |

Table 8.2.1-2 Degrees of equivalence between the participants for the BB at 50 Hz

| 50 Hz | CMS/ITRI | | NIMT | | A*STAR | |
|-----------------|--|----------|--|----------|--|----------|
| $i \rightarrow$ | D_{ij} | U_{ij} | D_{ij} | U_{ij} | D_{ij} | U_{ij} |
| $j \downarrow$ | in $\text{pC}/(\text{m}/\text{s}^2) \cdot 10^{-4}$ | | in $\text{pC}/(\text{m}/\text{s}^2) \cdot 10^{-4}$ | | in $\text{pC}/(\text{m}/\text{s}^2) \cdot 10^{-4}$ | |
| NMIJ | 1.3 | 4.7 | -2.6 | 6.4 | | |
| CMS/ITRI | | | -3.9 | 6.7 | | |
| NIMT | | | | | | |

Table 8.2.1-3 Degrees of equivalence between the participants for the BB at 63 Hz

| 63 Hz | CMS/ITRI | | NIMT | | A*STAR | |
|-----------------|--|----------|--|----------|--|----------|
| $i \rightarrow$ | D_{ij} | U_{ij} | D_{ij} | U_{ij} | D_{ij} | U_{ij} |
| $j \downarrow$ | in $\text{pC}/(\text{m}/\text{s}^2) \cdot 10^{-4}$ | | in $\text{pC}/(\text{m}/\text{s}^2) \cdot 10^{-4}$ | | in $\text{pC}/(\text{m}/\text{s}^2) \cdot 10^{-4}$ | |
| NMIJ | 0.5 | 4.4 | -3.6 | 6.3 | | |
| CMS/ITRI | | | -4.1 | 6.3 | | |
| NIMT | | | | | | |

Table 8.2.1-4 Degrees of equivalence between the participants for the BB at 80 Hz

| 80 Hz | CMS/ITRI | | NIMT | | A*STAR | |
|-----------------|--|----------|--|----------|--|----------|
| $i \rightarrow$ | D_{ij} | U_{ij} | D_{ij} | U_{ij} | D_{ij} | U_{ij} |
| $j \downarrow$ | in $\text{pC}/(\text{m}/\text{s}^2) \cdot 10^{-4}$ | | in $\text{pC}/(\text{m}/\text{s}^2) \cdot 10^{-4}$ | | in $\text{pC}/(\text{m}/\text{s}^2) \cdot 10^{-4}$ | |
| NMIJ | 2.5 | 4.3 | -0.9 | 6.3 | -1.3 | 6.8 |
| CMS/ITRI | | | -3.5 | 6.3 | -3.8 | 6.8 |
| NIMT | | | | | -0.4 | 8.2 |

Table 8.2.1-5 Degrees of equivalence between the participants for the BB at 100 Hz

| 100 Hz | CMS/ITRI | | NIMT | | A*STAR | |
|-----------------|--|----------|--|----------|--|----------|
| $i \rightarrow$ | D_{ij} | U_{ij} | D_{ij} | U_{ij} | D_{ij} | U_{ij} |
| $j \downarrow$ | in $\text{pC}/(\text{m}/\text{s}^2) \cdot 10^{-4}$ | | in $\text{pC}/(\text{m}/\text{s}^2) \cdot 10^{-4}$ | | in $\text{pC}/(\text{m}/\text{s}^2) \cdot 10^{-4}$ | |
| NMIJ | 2.6 | 5.8 | -0.5 | 7.2 | | |
| CMS/ITRI | | | -3.1 | 6.5 | | |
| NIMT | | | | | | |

Table 8.2.1-6 Degrees of equivalence between the participants for the BB at 125 Hz

| 125 Hz | CMS/ITRI | | NIMT | | A*STAR | |
|-----------------|--|----------|--|----------|--|----------|
| $i \rightarrow$ | D_{ij} | U_{ij} | D_{ij} | U_{ij} | D_{ij} | U_{ij} |
| $j \downarrow$ | in $\text{pC}/(\text{m}/\text{s}^2) \cdot 10^{-4}$ | | in $\text{pC}/(\text{m}/\text{s}^2) \cdot 10^{-4}$ | | in $\text{pC}/(\text{m}/\text{s}^2) \cdot 10^{-4}$ | |
| NMIJ | 1.4 | 4.4 | -1.1 | 5.7 | | |
| CMS/ITRI | | | -2.5 | 5.8 | | |
| NIMT | | | | | | |

Table 8.2.1-7 Degrees of equivalence between the participants for the BB at 160 Hz

| 160 Hz | CMS/ITRI | | NIMT | | A*STAR | |
|-----------------|--|----------|--|----------|--|----------|
| $i \rightarrow$ | D_{ij} | U_{ij} | D_{ij} | U_{ij} | D_{ij} | U_{ij} |
| $j \downarrow$ | in $\text{pC}/(\text{m}/\text{s}^2) \cdot 10^{-4}$ | | in $\text{pC}/(\text{m}/\text{s}^2) \cdot 10^{-4}$ | | in $\text{pC}/(\text{m}/\text{s}^2) \cdot 10^{-4}$ | |
| NMIJ | 2.0 | 4.6 | -1.3 | 5.7 | -1.3 | 6.9 |
| CMS/ITRI | | | -3.3 | 6.0 | -3.3 | 7.1 |
| NIMT | | | | | 0.0 | 7.9 |

Table 8.2.1-8 Degrees of equivalence between the participants for the BB at 200 Hz

| 200 Hz | CMS/ITRI | | NIMT | | A*STAR | |
|-----------------|--|----------|--|----------|--|----------|
| $i \rightarrow$ | D_{ij} | U_{ij} | D_{ij} | U_{ij} | D_{ij} | U_{ij} |
| $j \downarrow$ | in $\text{pC}/(\text{m}/\text{s}^2) \cdot 10^{-4}$ | | in $\text{pC}/(\text{m}/\text{s}^2) \cdot 10^{-4}$ | | in $\text{pC}/(\text{m}/\text{s}^2) \cdot 10^{-4}$ | |
| NMIJ | 2.0 | 4.6 | -0.8 | 5.7 | | |
| CMS/ITRI | | | -2.8 | 6.0 | | |
| NIMT | | | | | | |

Table 8.2.1-9 Degrees of equivalence between the participants for the BB at 250 Hz

| 250 Hz | CMS/ITRI | | NIMT | | A*STAR | |
|-----------------|--|----------|--|----------|--|----------|
| $i \rightarrow$ | D_{ij} | U_{ij} | D_{ij} | U_{ij} | D_{ij} | U_{ij} |
| $j \downarrow$ | in $\text{pC}/(\text{m}/\text{s}^2) \cdot 10^{-4}$ | | in $\text{pC}/(\text{m}/\text{s}^2) \cdot 10^{-4}$ | | in $\text{pC}/(\text{m}/\text{s}^2) \cdot 10^{-4}$ | |
| NMIJ | 0.9 | 4.7 | -1.2 | 5.7 | | |
| CMS/ITRI | | | -2.2 | 6.0 | | |
| NIMT | | | | | | |

Table 8.2.1-10 Degrees of equivalence between the participants for the BB at 315 Hz

| 315 Hz | CMS/ITRI | | NIMT | | A*STAR | |
|-----------------|--|----------|--|----------|--|----------|
| $i \rightarrow$ | D_{ij} | U_{ij} | D_{ij} | U_{ij} | D_{ij} | U_{ij} |
| $j \downarrow$ | in $\text{pC}/(\text{m}/\text{s}^2) \cdot 10^{-4}$ | | in $\text{pC}/(\text{m}/\text{s}^2) \cdot 10^{-4}$ | | in $\text{pC}/(\text{m}/\text{s}^2) \cdot 10^{-4}$ | |
| NMIJ | 2.2 | 5.0 | -0.7 | 5.7 | -2.4 | 6.8 |
| CMS/ITRI | | | -2.9 | 6.3 | -4.6 | 7.3 |
| NIMT | | | | | -1.6 | 7.8 |

Table 8.2.1-11 Degrees of equivalence between the participants for the BB at 400 Hz

| 400 Hz | CMS/ITRI | | NIMT | | A*STAR | |
|-----------------|--|----------|--|----------|--|----------|
| $i \rightarrow$ | D_{ij} | U_{ij} | D_{ij} | U_{ij} | D_{ij} | U_{ij} |
| $j \downarrow$ | in $\text{pC}/(\text{m}/\text{s}^2) \cdot 10^{-4}$ | | in $\text{pC}/(\text{m}/\text{s}^2) \cdot 10^{-4}$ | | in $\text{pC}/(\text{m}/\text{s}^2) \cdot 10^{-4}$ | |
| NMIJ | 3.9 | 4.8 | 1.0 | 5.7 | | |
| CMS/ITRI | | | -2.9 | 6.1 | | |
| NIMT | | | | | | |

Table 8.2.1-12 Degrees of equivalence between the participants for the BB at 500 Hz

| 500 Hz | CMS/ITRI | | NIMT | | A*STAR | |
|-----------------|--|----------|--|----------|--|----------|
| $i \rightarrow$ | D_{ij} | U_{ij} | D_{ij} | U_{ij} | D_{ij} | U_{ij} |
| $j \downarrow$ | in $\text{pC}/(\text{m}/\text{s}^2) \cdot 10^{-4}$ | | in $\text{pC}/(\text{m}/\text{s}^2) \cdot 10^{-4}$ | | in $\text{pC}/(\text{m}/\text{s}^2) \cdot 10^{-4}$ | |
| NMIJ | 3.2 | 5.2 | -0.4 | 5.9 | | |
| CMS/ITRI | | | -3.6 | 6.6 | | |
| NIMT | | | | | | |

Table 8.2.1-13 Degrees of equivalence between the participants for the BB at 630 Hz

| 630 Hz | CMS/ITRI | | NIMT | | A*STAR | |
|-----------------|--|----------|--|----------|--|----------|
| $i \rightarrow$ | D_{ij} | U_{ij} | D_{ij} | U_{ij} | D_{ij} | U_{ij} |
| $j \downarrow$ | in $\text{pC}/(\text{m}/\text{s}^2) \cdot 10^{-4}$ | | in $\text{pC}/(\text{m}/\text{s}^2) \cdot 10^{-4}$ | | in $\text{pC}/(\text{m}/\text{s}^2) \cdot 10^{-4}$ | |
| NMIJ | 1.8 | 4.9 | -1.6 | 5.9 | -1.4 | 6.8 |
| CMS/ITRI | | | -3.3 | 6.4 | -3.2 | 7.2 |
| NIMT | | | | | 0.2 | 8.0 |

Table 8.2.1-14 Degrees of equivalence between the participants for the BB at 800 Hz

| 800 Hz | CMS/ITRI | | NIMT | | A*STAR | |
|-----------------|---|----------|---|----------|---|----------|
| $i \rightarrow$ | D_{ij} | U_{ij} | D_{ij} | U_{ij} | D_{ij} | U_{ij} |
| $j \downarrow$ | in $\text{pC}/(\text{m/s}^2) \cdot 10^{-4}$ | | in $\text{pC}/(\text{m/s}^2) \cdot 10^{-4}$ | | in $\text{pC}/(\text{m/s}^2) \cdot 10^{-4}$ | |
| NMIJ | 1.2 | 5.4 | -2.2 | 5.9 | | |
| CMS/ITRI | | | -3.5 | 6.8 | | |
| NIMT | | | | | | |

Table 8.2.1-15 Degrees of equivalence between the participants for the BB at 1000 Hz

| 1000 Hz | CMS/ITRI | | NIMT | | A*STAR | |
|-----------------|---|----------|---|----------|---|----------|
| $i \rightarrow$ | D_{ij} | U_{ij} | D_{ij} | U_{ij} | D_{ij} | U_{ij} |
| $j \downarrow$ | in $\text{pC}/(\text{m/s}^2) \cdot 10^{-4}$ | | in $\text{pC}/(\text{m/s}^2) \cdot 10^{-4}$ | | in $\text{pC}/(\text{m/s}^2) \cdot 10^{-4}$ | |
| NMIJ | 0.5 | 5.6 | -2.1 | 10.2 | | |
| CMS/ITRI | | | -2.5 | 10.8 | | |
| NIMT | | | | | | |

Table 8.2.1-16 Degrees of equivalence between the participants for the BB at 1250 Hz

| 1250 Hz | CMS/ITRI | | NIMT | | A*STAR | |
|-----------------|---|----------|---|----------|---|----------|
| $i \rightarrow$ | D_{ij} | U_{ij} | D_{ij} | U_{ij} | D_{ij} | U_{ij} |
| $j \downarrow$ | in $\text{pC}/(\text{m/s}^2) \cdot 10^{-4}$ | | in $\text{pC}/(\text{m/s}^2) \cdot 10^{-4}$ | | in $\text{pC}/(\text{m/s}^2) \cdot 10^{-4}$ | |
| NMIJ | -0.2 | 5.8 | -2.0 | 10.2 | -3.2 | 6.8 |
| CMS/ITRI | | | -1.8 | 10.9 | -3.0 | 7.9 |
| NIMT | | | | | -1.3 | 11.5 |

Table 8.2.1-17 Degrees of equivalence between the participants for the BB at 1600 Hz

| 1600 Hz | CMS/ITRI | | NIMT | | A*STAR | |
|-----------------|---|----------|---|----------|---|----------|
| $i \rightarrow$ | D_{ij} | U_{ij} | D_{ij} | U_{ij} | D_{ij} | U_{ij} |
| $j \downarrow$ | in $\text{pC}/(\text{m/s}^2) \cdot 10^{-4}$ | | in $\text{pC}/(\text{m/s}^2) \cdot 10^{-4}$ | | in $\text{pC}/(\text{m/s}^2) \cdot 10^{-4}$ | |
| NMIJ | -1.4 | 4.6 | -2.6 | 10.2 | | |
| CMS/ITRI | | | -1.2 | 10.3 | | |
| NIMT | | | | | | |

Table 8.2.1-18 Degrees of equivalence between the participants for the BB at 2000 Hz

| 2000 Hz | CMS/ITRI | | NIMT | | A*STAR | |
|-----------------|--|----------|--|----------|--|----------|
| $i \rightarrow$ | D_{ij} | U_{ij} | D_{ij} | U_{ij} | D_{ij} | U_{ij} |
| $j \downarrow$ | in $\text{pC}/(\text{m}/\text{s}^2) \cdot 10^{-4}$ | | in $\text{pC}/(\text{m}/\text{s}^2) \cdot 10^{-4}$ | | in $\text{pC}/(\text{m}/\text{s}^2) \cdot 10^{-4}$ | |
| NMIJ | -1.1 | 5.1 | 0.3 | 10.2 | | |
| CMS/ITRI | | | 1.4 | 10.6 | | |
| NIMT | | | | | | |

Table 8.2.1-19 Degrees of equivalence between the participants for the BB at 2500 Hz

| 2500 Hz | CMS/ITRI | | NIMT | | A*STAR | |
|-----------------|--|----------|--|----------|--|----------|
| $i \rightarrow$ | D_{ij} | U_{ij} | D_{ij} | U_{ij} | D_{ij} | U_{ij} |
| $j \downarrow$ | in $\text{pC}/(\text{m}/\text{s}^2) \cdot 10^{-4}$ | | in $\text{pC}/(\text{m}/\text{s}^2) \cdot 10^{-4}$ | | in $\text{pC}/(\text{m}/\text{s}^2) \cdot 10^{-4}$ | |
| NMIJ | -4.6 | 5.2 | -2.2 | 10.2 | -5.1 | 7.9 |
| CMS/ITRI | | | 2.5 | 10.6 | -0.4 | 8.4 |
| NIMT | | | | | -2.9 | 12.2 |

Table 8.2.1-20 Degrees of equivalence between the participants for the BB at 3150 Hz

| 3150 Hz | CMS/ITRI | | NIMT | | A*STAR | |
|-----------------|--|----------|--|----------|--|----------|
| $i \rightarrow$ | D_{ij} | U_{ij} | D_{ij} | U_{ij} | D_{ij} | U_{ij} |
| $j \downarrow$ | in $\text{pC}/(\text{m}/\text{s}^2) \cdot 10^{-4}$ | | in $\text{pC}/(\text{m}/\text{s}^2) \cdot 10^{-4}$ | | in $\text{pC}/(\text{m}/\text{s}^2) \cdot 10^{-4}$ | |
| NMIJ | -6.4 | 6.7 | -3.2 | 10.4 | | |
| CMS/ITRI | | | 3.2 | 11.5 | | |
| NIMT | | | | | | |

Table 8.2.1-21 Degrees of equivalence between the participants for the BB at 4000 Hz

| 4000 Hz | CMS/ITRI | | NIMT | | A*STAR | |
|-----------------|--|----------|--|----------|--|----------|
| $i \rightarrow$ | D_{ij} | U_{ij} | D_{ij} | U_{ij} | D_{ij} | U_{ij} |
| $j \downarrow$ | in $\text{pC}/(\text{m}/\text{s}^2) \cdot 10^{-4}$ | | in $\text{pC}/(\text{m}/\text{s}^2) \cdot 10^{-4}$ | | in $\text{pC}/(\text{m}/\text{s}^2) \cdot 10^{-4}$ | |
| NMIJ | 0.4 | 10.8 | -1.4 | 10.5 | | |
| CMS/ITRI | | | -1.8 | 14.3 | | |
| NIMT | | | | | | |

Table 8.2.1-22 Degrees of equivalence between the participants for the BB at 5000 Hz

| 5000 Hz | CMS/ITRI | | NIMT | | A*STAR | |
|-----------------|--|----------|--|----------|--|----------|
| $i \rightarrow$ | D_{ij} | U_{ij} | D_{ij} | U_{ij} | D_{ij} | U_{ij} |
| $j \downarrow$ | in $\text{pC}/(\text{m}/\text{s}^2) \cdot 10^{-4}$ | | in $\text{pC}/(\text{m}/\text{s}^2) \cdot 10^{-4}$ | | in $\text{pC}/(\text{m}/\text{s}^2) \cdot 10^{-4}$ | |
| NMIJ | -5.6 | 11.9 | -0.6 | 10.7 | 12.4 | 40.9 |
| CMS/ITRI | | | 5.0 | 15.1 | 18.0 | 42.2 |
| NIMT | | | | | 13.0 | 41.9 |

8.2.2 Bilateral DoE of the single-ended (SE) accelerometer

Table 8.2.2-1 Degrees of equivalence between the participants for the SE at 40 Hz

| 40 Hz | CMS/ITRI | | NIMT | | A*STAR | |
|-----------------|--|----------|--|----------|--|----------|
| $i \rightarrow$ | D_{ij} | U_{ij} | D_{ij} | U_{ij} | D_{ij} | U_{ij} |
| $j \downarrow$ | in $\text{pC}/(\text{m}/\text{s}^2) \cdot 10^{-4}$ | | in $\text{pC}/(\text{m}/\text{s}^2) \cdot 10^{-4}$ | | in $\text{pC}/(\text{m}/\text{s}^2) \cdot 10^{-4}$ | |
| NMIJ | 2.9 | 6.4 | -4.8 | 9.4 | -3.7 | 7.8 |
| CMS/ITRI | | | -7.7 | 8.8 | -6.6 | 7.1 |
| NIMT | | | | | 1.1 | 9.8 |

Table 8.2.2-2 Degrees of equivalence between the participants for the SE at 50 Hz

| 50 Hz | CMS/ITRI | | NIMT | | A*STAR | |
|-----------------|--|----------|--|----------|--|----------|
| $i \rightarrow$ | D_{ij} | U_{ij} | D_{ij} | U_{ij} | D_{ij} | U_{ij} |
| $j \downarrow$ | in $\text{pC}/(\text{m}/\text{s}^2) \cdot 10^{-4}$ | | in $\text{pC}/(\text{m}/\text{s}^2) \cdot 10^{-4}$ | | in $\text{pC}/(\text{m}/\text{s}^2) \cdot 10^{-4}$ | |
| NMIJ | 1.6 | 7.2 | -4.0 | 8.3 | | |
| CMS/ITRI | | | -5.6 | 6.8 | | |
| NIMT | | | | | | |

Table 8.2.2-3 Degrees of equivalence between the participants for the SE at 63 Hz

| 63 Hz | CMS/ITRI | | NIMT | | A*STAR | |
|-----------------|--|----------|--|----------|--|----------|
| $i \rightarrow$ | D_{ij} | U_{ij} | D_{ij} | U_{ij} | D_{ij} | U_{ij} |
| $j \downarrow$ | in $\text{pC}/(\text{m}/\text{s}^2) \cdot 10^{-4}$ | | in $\text{pC}/(\text{m}/\text{s}^2) \cdot 10^{-4}$ | | in $\text{pC}/(\text{m}/\text{s}^2) \cdot 10^{-4}$ | |
| NMIJ | 1.6 | 6.1 | -4.0 | 7.6 | | |
| CMS/ITRI | | | -5.6 | 6.4 | | |
| NIMT | | | | | | |

Table 8.2.2-4 Degrees of equivalence between the participants for the SE at 80 Hz

| 80 Hz | CMS/ITRI | | NIMT | | A*STAR | |
|-----------------|--|----------|--|----------|--|----------|
| $i \rightarrow$ | D_{ij} | U_{ij} | D_{ij} | U_{ij} | D_{ij} | U_{ij} |
| $j \downarrow$ | in $\text{pC}/(\text{m}/\text{s}^2) \cdot 10^{-4}$ | | in $\text{pC}/(\text{m}/\text{s}^2) \cdot 10^{-4}$ | | in $\text{pC}/(\text{m}/\text{s}^2) \cdot 10^{-4}$ | |
| NMIJ | 2.1 | 5.9 | -2.8 | 7.5 | -3.1 | 8.0 |
| CMS/ITRI | | | -4.8 | 6.4 | -5.2 | 6.9 |
| NIMT | | | | | -0.4 | 8.3 |

Table 8.2.2-5 Degrees of equivalence between the participants for the SE at 100 Hz

| 100 Hz | CMS/ITRI | | NIMT | | A*STAR | |
|-----------------|--|----------|--|----------|--|----------|
| $i \rightarrow$ | D_{ij} | U_{ij} | D_{ij} | U_{ij} | D_{ij} | U_{ij} |
| $j \downarrow$ | in $\text{pC}/(\text{m}/\text{s}^2) \cdot 10^{-4}$ | | in $\text{pC}/(\text{m}/\text{s}^2) \cdot 10^{-4}$ | | in $\text{pC}/(\text{m}/\text{s}^2) \cdot 10^{-4}$ | |
| NMIJ | 2.0 | 6.3 | -2.5 | 7.7 | | |
| CMS/ITRI | | | -4.5 | 6.4 | | |
| NIMT | | | | | | |

Table 8.2.2-6 Degrees of equivalence between the participants for the SE at 125 Hz

| 125 Hz | CMS/ITRI | | NIMT | | A*STAR | |
|-----------------|--|----------|--|----------|--|----------|
| $i \rightarrow$ | D_{ij} | U_{ij} | D_{ij} | U_{ij} | D_{ij} | U_{ij} |
| $j \downarrow$ | in $\text{pC}/(\text{m}/\text{s}^2) \cdot 10^{-4}$ | | in $\text{pC}/(\text{m}/\text{s}^2) \cdot 10^{-4}$ | | in $\text{pC}/(\text{m}/\text{s}^2) \cdot 10^{-4}$ | |
| NMIJ | 1.2 | 6.0 | -2.0 | 7.1 | | |
| CMS/ITRI | | | -3.2 | 5.9 | | |
| NIMT | | | | | | |

Table 8.2.2-7 Degrees of equivalence between the participants for the SE at 160 Hz

| 160 Hz | CMS/ITRI | | NIMT | | A*STAR | |
|-----------------|--|----------|--|----------|--|----------|
| $i \rightarrow$ | D_{ij} | U_{ij} | D_{ij} | U_{ij} | D_{ij} | U_{ij} |
| $j \downarrow$ | in $\text{pC}/(\text{m}/\text{s}^2) \cdot 10^{-4}$ | | in $\text{pC}/(\text{m}/\text{s}^2) \cdot 10^{-4}$ | | in $\text{pC}/(\text{m}/\text{s}^2) \cdot 10^{-4}$ | |
| NMIJ | 2.2 | 6.0 | -2.0 | 7.1 | -3.5 | 9.8 |
| CMS/ITRI | | | -4.2 | 5.9 | -5.8 | 9.0 |
| NIMT | | | | | -1.5 | 9.7 |

Table 8.2.2-8 Degrees of equivalence between the participants for the SE at 200 Hz

| 200 Hz | CMS/ITRI | | NIMT | | A*STAR | |
|-----------------|--|----------|--|----------|--|----------|
| $i \rightarrow$ | D_{ij} | U_{ij} | D_{ij} | U_{ij} | D_{ij} | U_{ij} |
| $j \downarrow$ | in $\text{pC}/(\text{m}/\text{s}^2) \cdot 10^{-4}$ | | in $\text{pC}/(\text{m}/\text{s}^2) \cdot 10^{-4}$ | | in $\text{pC}/(\text{m}/\text{s}^2) \cdot 10^{-4}$ | |
| NMIJ | 1.6 | 6.2 | -1.9 | 7.1 | | |
| CMS/ITRI | | | -3.5 | 5.9 | | |
| NIMT | | | | | | |

Table 8.2.2-9 Degrees of equivalence between the participants for the SE at 250 Hz

| 250 Hz | CMS/ITRI | | NIMT | | A*STAR | |
|-----------------|--|----------|--|----------|--|----------|
| $i \rightarrow$ | D_{ij} | U_{ij} | D_{ij} | U_{ij} | D_{ij} | U_{ij} |
| $j \downarrow$ | in $\text{pC}/(\text{m}/\text{s}^2) \cdot 10^{-4}$ | | in $\text{pC}/(\text{m}/\text{s}^2) \cdot 10^{-4}$ | | in $\text{pC}/(\text{m}/\text{s}^2) \cdot 10^{-4}$ | |
| NMIJ | 1.6 | 6.2 | -2.1 | 7.1 | | |
| CMS/ITRI | | | -3.7 | 6.0 | | |
| NIMT | | | | | | |

Table 8.2.2-10 Degrees of equivalence between the participants for the SE at 315 Hz

| 315 Hz | CMS/ITRI | | NIMT | | A*STAR | |
|-----------------|--|----------|--|----------|--|----------|
| $i \rightarrow$ | D_{ij} | U_{ij} | D_{ij} | U_{ij} | D_{ij} | U_{ij} |
| $j \downarrow$ | in $\text{pC}/(\text{m}/\text{s}^2) \cdot 10^{-4}$ | | in $\text{pC}/(\text{m}/\text{s}^2) \cdot 10^{-4}$ | | in $\text{pC}/(\text{m}/\text{s}^2) \cdot 10^{-4}$ | |
| NMIJ | 1.3 | 6.4 | -2.5 | 7.0 | -3.8 | 8.1 |
| CMS/ITRI | | | -3.8 | 6.2 | -5.1 | 7.4 |
| NIMT | | | | | -1.3 | 8.0 |

Table 8.2.2-11 Degrees of equivalence between the participants for the SE at 400 Hz

| 400 Hz | CMS/ITRI | | NIMT | | A*STAR | |
|-----------------|--|----------|--|----------|--|----------|
| $i \rightarrow$ | D_{ij} | U_{ij} | D_{ij} | U_{ij} | D_{ij} | U_{ij} |
| $j \downarrow$ | in $\text{pC}/(\text{m}/\text{s}^2) \cdot 10^{-4}$ | | in $\text{pC}/(\text{m}/\text{s}^2) \cdot 10^{-4}$ | | in $\text{pC}/(\text{m}/\text{s}^2) \cdot 10^{-4}$ | |
| NMIJ | 1.9 | 6.3 | -2.1 | 7.0 | | |
| CMS/ITRI | | | -4.0 | 6.2 | | |
| NIMT | | | | | | |

Table 8.2.2-12 Degrees of equivalence between the participants for the SE at 500 Hz

| 500 Hz | CMS/ITRI | | NIMT | | A*STAR | |
|-----------------|--|----------|--|----------|--|----------|
| $i \rightarrow$ | D_{ij} | U_{ij} | D_{ij} | U_{ij} | D_{ij} | U_{ij} |
| $j \downarrow$ | in $\text{pC}/(\text{m}/\text{s}^2) \cdot 10^{-4}$ | | in $\text{pC}/(\text{m}/\text{s}^2) \cdot 10^{-4}$ | | in $\text{pC}/(\text{m}/\text{s}^2) \cdot 10^{-4}$ | |
| NMIJ | 2.2 | 6.4 | -2.9 | 7.2 | | |
| CMS/ITRI | | | -5.1 | 6.5 | | |
| NIMT | | | | | | |

Table 8.2.2-13 Degrees of equivalence between the participants for the SE at 630 Hz

| 630 Hz | CMS/ITRI | | NIMT | | A*STAR | |
|-----------------|--|----------|--|----------|--|----------|
| $i \rightarrow$ | D_{ij} | U_{ij} | D_{ij} | U_{ij} | D_{ij} | U_{ij} |
| $j \downarrow$ | in $\text{pC}/(\text{m}/\text{s}^2) \cdot 10^{-4}$ | | in $\text{pC}/(\text{m}/\text{s}^2) \cdot 10^{-4}$ | | in $\text{pC}/(\text{m}/\text{s}^2) \cdot 10^{-4}$ | |
| NMIJ | 2.3 | 6.4 | -2.4 | 7.2 | -2.4 | 8.4 |
| CMS/ITRI | | | -4.7 | 6.5 | -4.8 | 7.7 |
| NIMT | | | | | -0.1 | 8.5 |

Table 8.2.2-14 Degrees of equivalence between the participants for the SE at 800 Hz

| 800 Hz | CMS/ITRI | | NIMT | | A*STAR | |
|-----------------|--|----------|--|----------|--|----------|
| $i \rightarrow$ | D_{ij} | U_{ij} | D_{ij} | U_{ij} | D_{ij} | U_{ij} |
| $j \downarrow$ | in $\text{pC}/(\text{m}/\text{s}^2) \cdot 10^{-4}$ | | in $\text{pC}/(\text{m}/\text{s}^2) \cdot 10^{-4}$ | | in $\text{pC}/(\text{m}/\text{s}^2) \cdot 10^{-4}$ | |
| NMIJ | 0.5 | 6.6 | -3.0 | 7.2 | | |
| CMS/ITRI | | | -3.5 | 6.7 | | |
| NIMT | | | | | | |

Table 8.2.2-15 Degrees of equivalence between the participants for the SE at 1000 Hz

| 1000 Hz | CMS/ITRI | | NIMT | | A*STAR | |
|-----------------|--|----------|--|----------|--|----------|
| $i \rightarrow$ | D_{ij} | U_{ij} | D_{ij} | U_{ij} | D_{ij} | U_{ij} |
| $j \downarrow$ | in $\text{pC}/(\text{m}/\text{s}^2) \cdot 10^{-4}$ | | in $\text{pC}/(\text{m}/\text{s}^2) \cdot 10^{-4}$ | | in $\text{pC}/(\text{m}/\text{s}^2) \cdot 10^{-4}$ | |
| NMIJ | 0.3 | 7.5 | -3.5 | 11.0 | | |
| CMS/ITRI | | | -3.8 | 11.3 | | |
| NIMT | | | | | | |

Table 8.2.2-16 Degrees of equivalence between the participants for the SE at 1250 Hz

| 1250 Hz | CMS/ITRI | | NIMT | | A*STAR | |
|-----------------|--|----------|--|----------|--|----------|
| $i \rightarrow$ | D_{ij} | U_{ij} | D_{ij} | U_{ij} | D_{ij} | U_{ij} |
| $j \downarrow$ | in pC/(m/s ²)·10 ⁻⁴ | | in pC/(m/s ²)·10 ⁻⁴ | | in pC/(m/s ²)·10 ⁻⁴ | |
| NMIJ | -0.6 | 7.2 | -4.2 | 11.0 | -4.0 | 8.1 |
| CMS/ITRI | | | -3.6 | 11.0 | -3.4 | 8.1 |
| NIMT | | | | | 0.3 | 11.7 |

Table 8.2.2-17 Degrees of equivalence between the participants for the SE at 1600 Hz

| 1600 Hz | CMS/ITRI | | NIMT | | A*STAR | |
|-----------------|--|----------|--|----------|--|----------|
| $i \rightarrow$ | D_{ij} | U_{ij} | D_{ij} | U_{ij} | D_{ij} | U_{ij} |
| $j \downarrow$ | in pC/(m/s ²)·10 ⁻⁴ | | in pC/(m/s ²)·10 ⁻⁴ | | in pC/(m/s ²)·10 ⁻⁴ | |
| NMIJ | -1.6 | 6.3 | -3.1 | 11.1 | | |
| CMS/ITRI | | | -1.5 | 10.5 | | |
| NIMT | | | | | | |

Table 8.2.2-18 Degrees of equivalence between the participants for the SE at 2000 Hz

| 2000 Hz | CMS/ITRI | | NIMT | | A*STAR | |
|-----------------|--|----------|--|----------|--|----------|
| $i \rightarrow$ | D_{ij} | U_{ij} | D_{ij} | U_{ij} | D_{ij} | U_{ij} |
| $j \downarrow$ | in pC/(m/s ²)·10 ⁻⁴ | | in pC/(m/s ²)·10 ⁻⁴ | | in pC/(m/s ²)·10 ⁻⁴ | |
| NMIJ | -4.1 | 6.7 | -3.8 | 11.2 | | |
| CMS/ITRI | | | 0.3 | 10.9 | | |
| NIMT | | | | | | |

Table 8.2.2-19 Degrees of equivalence between the participants for the SE at 2500 Hz

| 2500 Hz | CMS/ITRI | | NIMT | | A*STAR | |
|-----------------|--|----------|--|----------|--|----------|
| $i \rightarrow$ | D_{ij} | U_{ij} | D_{ij} | U_{ij} | D_{ij} | U_{ij} |
| $j \downarrow$ | in pC/(m/s ²)·10 ⁻⁴ | | in pC/(m/s ²)·10 ⁻⁴ | | in pC/(m/s ²)·10 ⁻⁴ | |
| NMIJ | -7.5 | 6.6 | -4.3 | 11.2 | -0.5 | 8.9 |
| CMS/ITRI | | | 3.2 | 10.8 | 7.0 | 8.4 |
| NIMT | | | | | 3.8 | 12.4 |

Table 8.2.2-20 Degrees of equivalence between the participants for the SE at 3150 Hz

| 3150 Hz | CMS/ITRI | | NIMT | | A*STAR | |
|-----------------|--|----------|--|----------|--|----------|
| $i \rightarrow$ | D_{ij} | U_{ij} | D_{ij} | U_{ij} | D_{ij} | U_{ij} |
| $j \downarrow$ | in $\text{pC}/(\text{m}/\text{s}^2) \cdot 10^{-4}$ | | in $\text{pC}/(\text{m}/\text{s}^2) \cdot 10^{-4}$ | | in $\text{pC}/(\text{m}/\text{s}^2) \cdot 10^{-4}$ | |
| NMIJ | -5.4 | 8.0 | -6.4 | 11.3 | | |
| CMS/ITRI | | | -1.0 | 11.7 | | |
| NIMT | | | | | | |

Table 8.2.2-21 Degrees of equivalence between the participants for the SE at 4000 Hz

| 4000 Hz | CMS/ITRI | | NIMT | | A*STAR | |
|-----------------|--|----------|--|----------|--|----------|
| $i \rightarrow$ | D_{ij} | U_{ij} | D_{ij} | U_{ij} | D_{ij} | U_{ij} |
| $j \downarrow$ | in $\text{pC}/(\text{m}/\text{s}^2) \cdot 10^{-4}$ | | in $\text{pC}/(\text{m}/\text{s}^2) \cdot 10^{-4}$ | | in $\text{pC}/(\text{m}/\text{s}^2) \cdot 10^{-4}$ | |
| NMIJ | -2.1 | 11.8 | -7.0 | 11.3 | | |
| CMS/ITRI | | | -4.9 | 14.5 | | |
| NIMT | | | | | | |

Table 8.2.2-22 Degrees of equivalence between the participants for the SE at 5000 Hz

| 5000 Hz | CMS/ITRI | | NIMT | | A*STAR | |
|-----------------|--|----------|--|----------|--|----------|
| $i \rightarrow$ | D_{ij} | U_{ij} | D_{ij} | U_{ij} | D_{ij} | U_{ij} |
| $j \downarrow$ | in $\text{pC}/(\text{m}/\text{s}^2) \cdot 10^{-4}$ | | in $\text{pC}/(\text{m}/\text{s}^2) \cdot 10^{-4}$ | | in $\text{pC}/(\text{m}/\text{s}^2) \cdot 10^{-4}$ | |
| NMIJ | -6.9 | 13.1 | -3.9 | 11.6 | 4.4 | 16.0 |
| CMS/ITRI | | | 3.0 | 15.7 | 11.3 | 19.2 |
| NIMT | | | | | 8.3 | 18.2 |

8.3 Linking procedure and Degree of Equivalence with the KCRV

In this section, the linking procedure to the relevant CC comparison and calculation procedure of the KCRV are described. At the CCAUV-Key Comparison Working Group (KCWG) meeting in 2015, it was recommended to consider the influence of correlation of the results of the linking laboratory(ies) based on the procedure described in the previous publication by: Clemens Elster et al [7]. Here, we apply the methodology in principle.

The linking transforms the results ($y_i, u(y_i)$) of the participants of this comparison to scaled values z_i and their respective uncertainty $u(z_i)$, which are directly comparable to the relevant CC comparison results of CCAUV.V-K2. The scaling is done with the so called linking factor r . The values $(x_1, u(x_1)), \dots, (x_N, u(x_N))$ denote the charge sensitivities and associated standard uncertainties of

the participants in the CIPM comparison. Then, r is defined by two charge sensitivities (x_{NMIJ} and y_{NMIJ}) in the CIPM and RMO comparisons of NMIJ as follows.

$$r = \frac{x_{NMIJ}}{y_{NMIJ}} \quad (5)$$

where the uncertainty associated is described as

$$u^2(r) = \left(\frac{\partial r}{\partial x_{NMIJ}} \right)^2 u^2(x_{NMIJ}) + \left(\frac{\partial r}{\partial y_{NMIJ}} \right)^2 u^2(y_{NMIJ}) + 2 \left(\frac{\partial r}{\partial x_{NMIJ}} \right) \left(\frac{\partial r}{\partial y_{NMIJ}} \right) u(x_{NMIJ}) u(y_{NMIJ}) r(x_{NMIJ}, y_{NMIJ}) \quad (6)$$

Since CCAUV.V-K2 gives KCRV only in case of SE transducer, only the scaling of SE transducer results of APMP.AUV.V-K2 is linked to CCAUV.V-K2.

Considering the correlation coefficient of uncertainty in two comparisons $r(x_{NMIJ}, y_{NMIJ})$ is 1, equation (6) can be written as equation (7).

$$u^2(r) = \left(\frac{1}{y_{NMIJ}} \right)^2 u^2(x_{NMIJ}) + \left(\frac{x_{NMIJ}}{y_{NMIJ}^2} \right)^2 u^2(y_{NMIJ}) - 2 \frac{x_{NMIJ}}{y_{NMIJ}^3} u(x_{NMIJ}) u(y_{NMIJ}) \quad (7)$$

RMO comparison participant result y_i can be transformed to the scaled value z_i of CC comparison employing r as follows

$$z_i = r y_i = \frac{x_{NMIJ}}{y_{NMIJ}} y_i \quad (8).$$

Thus, the degrees of equivalence are given as the differences between the scaled charge sensitivities in the RMO comparison and the KCRV of the CIPM comparison.

$$d_i = z_i - x_{KCRV} = \frac{x_{NMIJ}}{y_{NMIJ}} y_i - x_{KCRV}, \quad i = CMS / ITRI, NIMT, A * STAR \quad (9)$$

and the squared standard uncertainties associated with these differences are described as follows

$$u^2(d_i) = \left(\frac{y_i}{y_{NMIJ}} \right)^2 u^2(x_{NMIJ}) + \left(-\frac{x_{NMIJ} y_i}{y_{NMIJ}^2} \right)^2 u^2(y_{NMIJ}) + \left(\frac{x_{NMIJ}}{y_{NMIJ}} \right)^2 u^2(y_i) + u^2(x_{KCRV}) + 2 \left(\frac{y_i}{y_{NMIJ}} \right) \left(-\frac{x_{NMIJ} y_i}{y_{NMIJ}^2} \right) u(x_{NMIJ}, y_{NMIJ}) + 2 \left(\frac{y_i}{y_{NMIJ}} \right) (-1) u(x_{NMIJ}, x_{KCRV}) + 2 \left(-\frac{x_{NMIJ} y_i}{y_{NMIJ}^2} \right) (-1) u(y_{NMIJ}, x_{KCRV}) \quad (10).$$

Here $u(x_{NMIJ}, y_{NMIJ})$ is transformed as $u(x_{NMIJ}) u(y_{NMIJ})$ as $r(x_{NMIJ}, y_{NMIJ}) = 1$. Next, the transformation of $u(x_{NMIJ}, x_{KCRV}) = u(x_{NMIJ}) u(x_{KCRV}) r(x_{NMIJ}, x_{KCRV})$ is carried out as follows.

$$r(x_{NMIJ}, x_{KCRV}) = \frac{\frac{\partial x_{NMIJ}}{\partial x_{NMIJ}} \frac{\partial x_{KCRV}}{\partial x_{NMIJ}} u^2(x_{NMIJ})}{u(x_{NMIJ})u(x_{KCRV})} = 1 \cdot \frac{1/u^2(x_{NMIJ})u^2(x_{NMIJ})}{1/u^2(x_{KCRV})u(x_{KCRV})} = \frac{u(x_{KCRV})}{u(x_{NMIJ})} \quad (11)$$

The value of equation (11) at each frequency is shown in table 8.3.1 and lies from 0.15 to 0.47.

Table 8.3.1 Correlation coefficients of combined standard uncertainty between CCAUV.V-K2 and APMP.AUV.V-K2 international comparisons.

| Frequency | $u(x_{KCRV})$ | $u(x_{NMIJ})$ | $r(x_{NMIJ}, x_{KCRV})$ |
|-----------|------------------------|------------------------|-------------------------|
| Hz | pC/(m/s ²) | pC/(m/s ²) | |
| 40 | 0.00004 | 0.00025 | 0.16 |
| 50 | | 0.00031 | |
| 63 | 0.00004 | 0.00026 | 0.16 |
| 80 | 0.00004 | 0.00025 | 0.16 |
| 100 | 0.00004 | 0.00027 | 0.15 |
| 125 | 0.00004 | 0.00025 | 0.16 |
| 160 | 0.00004 | 0.00025 | 0.16 |
| 200 | 0.00004 | 0.00026 | 0.16 |
| 250 | 0.00004 | 0.00025 | 0.16 |
| 315 | 0.00004 | 0.00025 | 0.17 |
| 400 | 0.00004 | 0.00025 | 0.17 |
| 500 | 0.00004 | 0.00025 | 0.17 |
| 630 | 0.00004 | 0.00025 | 0.17 |
| 800 | 0.00004 | 0.00025 | 0.17 |
| 1000 | 0.00004 | 0.00025 | 0.17 |
| 1250 | 0.00005 | 0.00025 | 0.18 |
| 1600 | 0.00005 | 0.00025 | 0.18 |
| 2000 | 0.00005 | 0.00025 | 0.19 |
| 2500 | 0.00005 | 0.00026 | 0.21 |
| 3150 | 0.00007 | 0.00026 | 0.26 |
| 4000 | 0.00009 | 0.00026 | 0.33 |
| 5000 | 0.00013 | 0.00027 | 0.47 |

Also, the transformation of $u(y_{NMIJ}, x_{KCRV}) = u(y_{NMIJ})u(x_{KCRV})r(y_{NMIJ}, x_{KCRV})$ can be given as equation (12) same as equation (11) because of $r(x_{NMIJ}, y_{NMIJ}) = 1$.

$$\begin{aligned} r(y_{NMIJ}, x_{KCRV}) &= \frac{\frac{1/u^2(x_{NMIJ})}{1/u^2(x_{KCRV})} u(x_{NMIJ}, y_{NMIJ})}{u(y_{NMIJ})u(x_{KCRV})} = \frac{u(x_{KCRV})}{u(y_{NMIJ})u^2(x_{NMIJ})} u(x_{NMIJ}, y_{NMIJ}) \\ &= r(x_{NMIJ}, y_{NMIJ}) \frac{u(x_{KCRV})}{u(x_{NMIJ})} = r(x_{NMIJ}, x_{KCRV}) r(x_{NMIJ}, y_{NMIJ}) \\ &= r(x_{NMIJ}, x_{KCRV}) \end{aligned} \quad (12)$$

This also lies from 0.15 to 0.47 as shown in table 8.3.1. These DoE of the SE transducer are each a pair of values of the difference d_i between the respective corresponding participants i and KCRV with the expanded uncertainty U_i of this difference. (See table 8.3.2.) These values are calculated for each frequency according to:

$$d_i = z_i - x_{KCRV} \quad (13)$$

$$U(d_i) = k \cdot \sqrt{u^2(d_i)} \quad (14)$$

with a coverage factor of $k = 2$. Figures 8.3.1-4 show each DoE at 40 Hz, 160 Hz, 1250 Hz and 5000 Hz. Since there is no KCRV regarding the BB transducer in CCAUV.V-K2, we provide the DoE of the BB transducer as reference information in Annex 3.

Table 8.3.2 DoE to the KCRV of CCAUV.V-K2 for the SE transducer

| Frequency in Hz | CMS/ITRI | | NIMT | | A*STAR | |
|--------------------|--|----------|--|----------|--|----------|
| | d_i | $U(d_i)$ | d_i | $U(d_i)$ | d_i | $U(d_i)$ |
| | in pC/(m/s ²)-10 ⁻⁴ | | in pC/(m/s ²)-10 ⁻⁴ | | in pC/(m/s ²)-10 ⁻⁴ | |
| 40 | 2.9 | 3.9 | -4.7 | 7.8 | -3.7 | 5.9 |
| 50 | | | | | | |
| 63 | 1.9 | 3.3 | -3.6 | 5.5 | | |
| 80 | 2.0 | 3.2 | -2.8 | 5.5 | -3.2 | 6.2 |
| 100 | 2.2 | 3.3 | -2.3 | 5.6 | | |
| 125 | 1.1 | 3.3 | -2.1 | 4.9 | | |
| 160 | 1.8 | 3.3 | -2.4 | 4.9 | -3.9 | 8.3 |
| 200 | 1.4 | 3.4 | -2.1 | 4.9 | | |
| 250 | 0.8 | 3.6 | -2.8 | 4.9 | | |
| 315 | 1.7 | 3.9 | -2.1 | 4.9 | -3.4 | 6.3 |
| 400 | 0.9 | 3.8 | -3.0 | 4.9 | | |
| 500 | 2.6 | 3.9 | -2.4 | 5.2 | | |
| 630 | 2.5 | 3.9 | -2.2 | 5.2 | -2.2 | 6.7 |
| 800 | 0.8 | 4.3 | -2.7 | 5.2 | | |
| 1000 | 1.0 | 5.6 | -2.8 | 9.7 | | |
| 1250 | 1.6 | 5.1 | -2.0 | 9.8 | -1.7 | 6.3 |
| 1600 | -0.8 | 3.8 | -2.3 | 9.8 | | |
| 2000 | -0.8 | 4.5 | -0.5 | 9.9 | | |
| 2500 | -4.8 | 4.2 | -1.7 | 10.0 | 2.1 | 7.3 |
| 3150 | -1.7 | 6.2 | -2.7 | 10.0 | | |
| 4000 | 4.3 | 10.6 | -0.6 | 10.0 | | |
| 5000 | -2.8 | 12.1 | 0.2 | 10.4 | 8.4 | 15.2 |




Figure 8.3.1 Deviations and expanded uncertainties of the magnitude from the KCRV at 40 Hz.




Figure 8.3.2 Deviations and expanded uncertainties of the magnitude from the KCRV at 160 Hz.




Figure 8.3.3 Deviations and expanded uncertainties of the magnitude from the KCRV at 1250 Hz.




Figure 8.3.4 Deviations and expanded uncertainties of the magnitude from the KCRV at 5000 Hz.

9. Conclusion

The RMO key comparison in vibration and acceleration APMP.AUV.V-K2 has been successfully finished. All participating laboratories had their results linked to the KCRV of the relevant CIPM level key comparison, namely CCAUV.V-K2, for the BB and SE transducers. The degrees of equivalence of the participants to the KCRV can be used to support their calibration and measurement capabilities.

10. References

- [1] H.-J. v. Martens et al., Report on Key Comparison CCAUV.V-K1, Metrologia (2002), IOP, www.bipm.org/pdf/final_reports/AUV/V-K1/CCAUV.V-K1.pdf
- [2] T. Bruns et al., Final report of Key Comparison CCAUV.V-K2, Metrologia (2014), IOP, www.bipm.org/utis/common/pdf/final_reports/AUV/V-K2/CCAUV.V-K2.pdf
- [3] Synopsis of the actions and decisions arising from the 10th meeting of the CCAUV (2015), http://www.bipm.org/utis/common/pdf/CC/CCAUV/CCAUV10_summary.pdf
- [4] H.-J. v. Martens et al., Linking the results of the regional key comparison APMP.AUV.V-K1, Metrologia (2004), IOP, www.bipm.org/utis/common/pdf/final_reports/AUV/V-K1/CCAUV.V-K1_APMP.AUV.VK1.pdf
- [5] ISO 16063-11:1999 Methods for the calibration of vibration and shock transducers -- Part 11: Primary vibration calibration by laser interferometry
- [6] T. J. Quinn, Guidelines for CIPM key comparisons, (2003), BIPM, <http://www.bipm.org/utis/en/pdf/guidelines.pdf>
- [7] C. Elster et al., Proposal for linking the results CIPM and RMO key comparisons, Metrologia 40, 189-194 (2003).

Annex 1 Technical Protocol of the APMP Key Comparison APMP.AUV.V-K2

1. Task and purpose of the comparison

In the field of vibration and shock, this regional comparison is organized in order to compare measurements of sinusoidal linear accelerations in the frequency range from 40 Hz to 5 kHz. Moreover, the magnitude of the complex sensitivity calibration and measurement capabilities (CMCs) of the NMIs for accelerometer calibration are to be examined and compared and linked to the CIPM comparison. It is the task of the comparison to measure the magnitude of the complex sensitivity of two accelerometer standards (two piezoelectric accelerometers of back-to-back type and single ended type) at different frequencies with acceleration amplitudes as specified in section 4. The results of this APMP KC will, after approval for equivalence, be linked to CCAUV.V-K2 (only in case of the SE transducer) as the foundation for the registration of “calibration and measurement capabilities” (CMC) in the framework of the CIPM MRA.

The charge sensitivity is calculated as the ratio of the amplitude of the accelerometer output charge to the amplitude of the acceleration at its reference surface. The magnitude of the complex charge sensitivity shall be given in pico coulomb per metre per second squared ($\text{pC}/(\text{m}/\text{s}^2)$) for the different measurement conditions specified in section 4. A calibrated charge amplifier is to be used to measure the output charge of the accelerometer standards, applying appropriate electrical calibration methods.

For the calibration of the accelerometer standards, laser interferometry in compliance with any method from methods 1, 2, and 3 of the international standard ISO 16063-11:1999 has to be applied, in order to cover the entire frequency range.

The reported sensitivities and associated uncertainties will, after approval for equivalence, be used for the calculation of the “degrees of equivalence” (DoE) between the participating NMI and the key comparison reference value.

2. Pilot laboratory

Pilot laboratory for this Key Comparison, who had also participated in CCAUV.V-K2, is
Vibration and Hardness Section
Acoustics and Vibration Metrology Division

National Metrology Institute of Japan

National Institute of Advanced Industrial Science and Technology

AIST Tsukuba Central 3, 1-1-1 Umezono Tsukuba Ibaraki 305-8563 Japan

Contact Persons are

Dr. Takashi USUDA

Dr. Akihiro OTA

Tel.: +81-29-861-4356

Tel.: +81-29-861-4366

e-mail (both): vibration@m.aist.go.jp

Fax : +81-29-861-4047

3. Participants

Three National Metrology Institutes (NMIs) of APMP will participate in the comparison.

| Participant number | Participant (laboratory name) | Acronym | Country | Country Code | |
|--------------------|---|-------------|----------------|--------------|--|
| 1 | National Metrology Institute of Japan | NMIJ | Japan | JP | |
| 2 | Center for Measurement Standards/Industrial Technology Research Institute | CMS/ITRI | Chinese Taipei | TW | |
| | National Metrology Institute of Japan | NMIJ | Japan | JP | |
| 3 | National Institute of Metrology Thailand | NIMT | Thailand | TH | |
| 4 | National Metrology Centre, Agency for Science, Technology and Research | NMC, A*Star | Singapore | SG | |

Circulation and Calibration Schedule

To prevent any drift/fluctuation of transducer characteristics, all transportations among the participants must be done by hand carry.

| Date (tentative) | From | To | Carried by | Note |
|--|-------------|-------------|------------|---|
| End of August 2011 | NMIJ | CMS/ITRI | JP | Dr. Ota carries at the occasion of CMS/ITRI on-site visit |
| October 2011 | CMS/ITRI | NMIJ | TW | |
| Interim monitoring and calibration at NMIJ | | | | |
| December 2011 | NMIJ | NMC, A*Star | SG | Delegate to APMP TCAUV carries to SG on the way back. |
| January 2012 | NMC, A*Star | NIMT | SG | |
| March 2012 | NIMT | NMIJ | TH | Dr. Hirunyapruk carries to JP. |

Contact Person of each laboratory

CMS/ITRI, Mr. Yu-Chung Huang

Dynamic Engineering Measurement Laboratory –

Vibration & Acoustics Measurement

Tel; +886-3-5743758

Fax; +886-3-5726445

E-mail; JoyHuang@itri.org.tw

NIMT; Mr. Virat Plangsangmas

Head of Acoustics and Vibration Metrology Department

Tel; +66 2354 3700 ext.400

Fax; +66 2354 3692

E-mail; virat@nimt.or.th

NMC, A*Star; Ms Cui Shan

Mechanical Metrology Department

Tel; +65 6279 1963

Fax; +65 6279 1995

E-mail; cui_shan@nmc.a-star.edu.sg

4. Devices under Test and Measurement conditions

For the calibration task of this KC, a set of two piezoelectric accelerometers with connecting cable will be circulated among the participating laboratories. The individual transducers being a “single ended” (SE) type, namely a Brüel & Kjær 8305-001 and a “back to back” (BB) type, namely a Brüel & Kjær 8305. . No additional mass shall be attached to the top surface of BB type accelerometer

The accelerometers are to be calibrated for their complex charge sensitivity according to those procedures and conditions implemented by the NMI in conformance with ISO 16063-11 which provide magnitude information of the artefact. The sensitivities reported shall be for the accelerometers alone, excluding any effects from the charge amplifier.

The frequency range of the measurements was agreed to be from 40 Hz to 5 kHz. Specifically the laboratories are supposed to measure at the following frequencies (all values in Hz):

40, 50, 63, 80, 100, 125, 160, 200, 250, 315, 400, 500, 630, 800, 1000, 1250, 1600, 2000, 2500, 3150, 4000, 5000 (160 Hz is reference frequency).

The charge amplifier (CA) used for the calibration is not provided within the set of the artefacts, it must therefore be provided by the individual participant.

The measurement condition should be kept according to the laboratory's standard conditions for calibration of customer accelerometers for claiming their CMCs. This presumes that these conditions comply with those defined by the applicable ISO standards [references 1, 2 and 3], simultaneously.

Specific conditions for the measurements of this KC are:

- acceleration amplitudes: A range of 10 m/s^2 to 200 m/s^2 is admissible.
- ambient temperature and accelerometer temperature during the calibration: $(23 \pm 2) \text{ }^\circ\text{C}$ (actual values to be stated within tolerances of $\pm 0.3 \text{ }^\circ\text{C}$).
- relative humidity: max. 75 %
- mounting torque of the accelerometer: $(2.0 \pm 0.1) \text{ N}\cdot\text{m}$

5. Circulation and Transportation

The transducers will be circulated with a measurement period of four weeks for each participant. At the beginning and the end of the circulation, the transducers will be measured at the pilot laboratory in order to fix reference value and to monitor the stability of the transducers.

The cost of transportation to or from a participating laboratory shall be covered by the participating laboratory. The accelerometers have to be hand carried. Insurance of transfer devices is decided by agreement among the participants taking account of the responsibility of each participant for any damage in its country.

6. Measurement and Analysis Instructions

The participating laboratories have to observe the following instructions:

- The charge amplifier used in the laboratory is to be calibrated using a standard capacitor and standard voltmeter, both traceable to national standards. The calibration of the charge amplifier has to be carried out shortly before the calibration, using values of the electric quantities similar to those expected in the accelerometer calibration.
- In order to suppress the effect of any non-rectilinear motion, the displacement has to be measured at least at four different points. These points should be equally spaced on the top surface of the back-to-back accelerometer or the basis surface of the single-ended accelerometer.
- Primary calibration of BB accelerometer by laser interferometry: The motion is to be sensed at the top surface of the transducer without any additional reflector or dummy mass.
- Primary calibration of SE accelerometer by laser interferometry: The reference surface for acceleration measurement is by definition the base or mounting surface of the accelerometer. If this surface is covered during the calibration, the motion is to be sensed on the moving part close to the accelerometer.

Alternatively, the motion can be sensed at the mounting surface of the accelerometer via longitudinal holes in the moving part of the vibration exciter. ISO 16063-11 is to be observed.

- The mounting surface of the accelerometer and the moving part of the vibration exciter must be slightly lubricated before mounting.
- It is advised that the measurement results should be compiled from complete measurement series (normally twice) carried out at different days under nominally the same conditions, except that the accelerometer is remounted and the cable reattached. The standard deviation of the subsequent measurements should be included in the report.

7. Communication of the Results to the Pilot laboratory

Each participating laboratory will submit one printed and signed calibration report for each accelerometer to the pilot laboratory including the following:

- a description of the calibration systems used for the comparison and the mounting techniques for the accelerometer
- a description of the calibration methods used
- documented record of the ambient conditions during measurements
- the calibration results, including the **relative expanded** measurement uncertainty, and the applied coverage factor for each value
- a detailed uncertainty budget for the system covering all components of measurement uncertainty (calculated according to GUM (references 4 and 5) and ISO (reference 2)). Including among others information on the type of uncertainty (A or B), assumed distribution function and repeatability component.

In addition each participating laboratory will receive two electronic spreadsheets prepared by the pilot laboratory, where the calibration results have to be filled in following the structure given in the files. The consistency between the results in electronic form and the printed and signed calibration report is the responsibility of the participating laboratory. The data submitted in the electronic spreadsheet shall be deemed to be the official results submitted for the comparison.

The results have to be submitted to the pilot laboratory within six weeks after the measurements.

The pilot laboratory will submit its set of results to the executive secretary of CCAUV prior to the first measurement of a participating laboratory.

8. Remarks on the post processing

- Presuming consistency of the results, the degrees of equivalence will be calculated according to the established methods agreed already for CCAUV.V-K2.

9. References

- [1] ISO 16063-1:1998 “Methods for the calibration of vibration and shock transducers -- Part 1: Basic concepts
- [2] ISO 16063-11:1999 “Methods for the calibration of vibration and shock transducers -- Part 11: Primary vibration calibration by laser interferometry”
- [3] ISO/IEC 17025:2005 “General requirements for the competence of testing and calibration laboratories”
- [4] ISO/IEC Guide 98-3:2008 “Uncertainty of measurement -- Part 3: Guide to the expression of uncertainty in measurement (GUM:1995)
- [5] ISO/IEC Guide 98-3:2008/Suppl 1:2008 “Propagation of distributions using a Monte Carlo method”

Annex 2 Measurement Uncertainty Budgets reported by the participants

A2-1: NMIJ

Table A2-1-1 Uncertainty budget for the BB transducer at 40 Hz to 80 Hz

| Name of participant | NMIJ | | | | | | |
|--|--------------------------|--------|-----------------------------|-----------------------------------|-----------------------------------|-----------------------------------|-----------------------------------|
| DUT | 8305 sn2691032 | | | | | | |
| Acceleration amplitude | [m/s ²] | | | 10 | 10 | 10 | 10 |
| Vibration Frequency | [Hz] | | | 40 | 50 | 63 | 80 |
| Applied Calibration Method | | | | sine-approximation | sine-approximation | sine-approximation | sine-approximation |
| Number of measurements | | | | 240 | 240 | 240 | 240 |
| Charge Sensitivity | [pC/(m/s ²)] | | | 0.12780 | 0.12775 | 0.12788 | 0.12766 |
| Uncertainty Component | type (A or B) | factor | effective degree of freedom | Relative standard uncertainty [%] | Relative standard uncertainty [%] | Relative standard uncertainty [%] | Relative standard uncertainty [%] |
| Capacitance standard uncertainty | B | 1 | ∞ | 3.50E-03 | 3.50E-03 | 3.50E-03 | 3.50E-03 |
| Voltage standard uncertainty | B | 1 | ∞ | 5.00E-02 | 5.00E-02 | 5.00E-02 | 5.00E-02 |
| Calibration uncertainty of charge amplifier | A | 1 | 76 | 3.58E-03 | 2.19E-02 | 4.37E-03 | 4.43E-03 |
| Accuracy of voltmeter resolution | B | √3 | ∞ | 2.89E-06 | 2.89E-06 | 2.89E-06 | 2.89E-06 |
| Repeatability of sensitivity | A | 1 | 237 | 3.50E-03 | 5.06E-03 | 5.65E-03 | 2.74E-03 |
| Laser wavelength instability | B | 1 | ∞ | 1.60E-03 | 1.60E-03 | 1.60E-03 | 1.60E-03 |
| Vibration frequency instability | B | 1 | ∞ | 6.00E-03 | 6.00E-03 | 6.00E-03 | 6.00E-03 |
| Effect of phase disturbance on phase amplitude measurement | A | 1 | 237 | 2.29E-04 | 1.73E-04 | 1.68E-04 | 2.12E-04 |
| Attachment uncertainty of accelerometer | B | 1 | ∞ | 7.19E-02 | 7.19E-02 | 7.19E-02 | 7.19E-02 |
| Transverse sensitivity | B | 1 | ∞ | 7.80E-02 | 7.80E-02 | 7.80E-02 | 7.80E-02 |
| Effect of vibration distortion | B | 1 | ∞ | 6.72E-04 | 6.72E-04 | 3.23E-04 | 2.84E-04 |
| Relative combined uncertainty [%] | | | | 0.12 | 0.12 | 0.12 | 0.12 |
| Relative expanded uncertainty (<i>k</i> =2) [%] | | | | 0.24 | 0.24 | 0.24 | 0.24 |

Table A2-1-2 Uncertainty budget for the BB transducer at 100 Hz to 200 Hz

| Name of participant | NMIJ | | | | | | |
|--|--------------------------|--------|-----------------------------|-----------------------------------|-----------------------------------|-----------------------------------|-----------------------------------|
| DUT | 8305 sn2691032 | | | | | | |
| Acceleration amplitude | [m/s ²] | | | 10 | 10 | 10 | 10 |
| Vibration Frequency | [Hz] | | | 100 | 125 | 160 | 200 |
| Applied Calibration Method | | | | sine-approximation | sine-approximation | sine-approximation | sine-approximation |
| Number of measurements | | | | 240 | 240 | 240 | 240 |
| Charge Sensitivity | [pC/(m/s ²)] | | | 0.12766 | 0.12773 | 0.12773 | 0.12771 |
| Uncertainty Component | type (A or B) | factor | effective degree of freedom | Relative standard uncertainty [%] | Relative standard uncertainty [%] | Relative standard uncertainty [%] | Relative standard uncertainty [%] |
| Capacitance standard uncertainty | B | 1 | ∞ | 3.50E-03 | 3.50E-03 | 3.50E-03 | 3.50E-03 |
| Voltage standard uncertainty | B | 1 | ∞ | 5.00E-02 | 5.00E-02 | 5.00E-02 | 5.00E-02 |
| Calibration uncertainty of charge amplifier | A | 1 | 76 | 1.41E-01 | 1.34E-02 | 1.59E-02 | 3.58E-03 |
| Accuracy of voltmeter resolution | B | √3 | ∞ | 2.89E-06 | 2.89E-06 | 2.89E-06 | 2.89E-06 |
| Repeatability of sensitivity | A | 1 | 237 | 2.15E-02 | 2.02E-03 | 2.81E-03 | 6.49E-03 |
| Laser wavelength instability | B | 1 | ∞ | 1.60E-03 | 1.60E-03 | 1.60E-03 | 1.60E-03 |
| Vibration frequency instability | B | 1 | ∞ | 6.00E-03 | 6.00E-03 | 6.00E-03 | 6.00E-03 |
| Effect of phase disturbance on phase amplitude measurement | A | 1 | 237 | 2.49E-04 | 3.78E-04 | 6.55E-04 | 1.33E-03 |
| Attachment uncertainty of accelerometer | B | 1 | ∞ | 7.19E-02 | 7.19E-02 | 7.19E-02 | 7.19E-02 |
| Transverse sensitivity | B | 1 | ∞ | 7.80E-02 | 7.80E-02 | 7.80E-02 | 7.80E-02 |
| Effect of vibration distortion | B | 1 | ∞ | | | | |
| Relative combined uncertainty [%] | | | | 0.18 | 0.12 | 0.12 | 0.12 |
| Relative expanded uncertainty (<i>k</i> =2) [%] | | | | 0.37 | 0.24 | 0.24 | 0.24 |

Table A2-1-3 Uncertainty budget for the BB transducer at 250 Hz to 500 Hz

| Name of participant | NMIJ | | | | | | |
|--|--------------------------|--------|-----------------------------|-----------------------------------|-----------------------------------|-----------------------------------|-----------------------------------|
| DUT | 8305 sn2691032 | | | | | | |
| Acceleration amplitude | [m/s ²] | | | 10 | 50 | 50 | 50 |
| Vibration Frequency | [Hz] | | | 250 | 315 | 400 | 500 |
| Applied Calibration Method | | | | sine-approximation | sine-approximation | sine-approximation | sine-approximation |
| Number of measurements | | | | 240 | 240 | 240 | 240 |
| Charge Sensitivity | [pC/(m/s ²)] | | | 0.12776 | 0.12772 | 0.12752 | 0.12763 |
| Uncertainty Component | type (A or B) | factor | effective degree of freedom | Relative standard uncertainty [%] | Relative standard uncertainty [%] | Relative standard uncertainty [%] | Relative standard uncertainty [%] |
| Capacitance standard uncertainty | B | 1 | ∞ | 3.50E-03 | 3.50E-03 | 3.50E-03 | 3.50E-03 |
| Voltage standard uncertainty | B | 1 | ∞ | 5.00E-02 | 5.00E-02 | 5.00E-02 | 5.00E-02 |
| Calibration uncertainty of charge amplifier | A | 1 | 76 | 2.19E-02 | 4.37E-03 | 4.43E-03 | 1.20E-02 |
| Accuracy of voltmeter resolution | B | √3 | ∞ | 2.89E-06 | 5.77E-07 | 5.77E-07 | 5.77E-07 |
| Repeatability of sensitivity | A | 1 | 237 | 2.98E-03 | 1.46E-03 | 1.46E-03 | 1.77E-03 |
| Laser wavelength instability | B | 1 | ∞ | 1.60E-03 | 1.60E-03 | 1.60E-03 | 1.60E-03 |
| Vibration frequency instability | B | 1 | ∞ | 6.00E-03 | 6.00E-03 | 6.00E-03 | 6.00E-03 |
| Effect of phase disturbance on phase amplitude measurement | A | 1 | 237 | 1.79E-03 | 1.04E-03 | 1.19E-03 | 1.73E-03 |
| Attachment uncertainty of accelerometer | B | 1 | ∞ | 7.19E-02 | 7.19E-02 | 7.19E-02 | 7.19E-02 |
| Transverse sensitivity | B | 1 | ∞ | 7.80E-02 | 7.80E-02 | 7.80E-02 | 7.80E-02 |
| Effect of vibration distortion | B | 1 | ∞ | | | | |
| Relative combined uncertainty [%] | | | | 0.12 | 0.12 | 0.12 | 0.12 |
| Relative expanded uncertainty (<i>k</i> =2) [%] | | | | 0.24 | 0.24 | 0.24 | 0.24 |

Table A2-1-4 Uncertainty budget for the BB transducer at 630 Hz to 1250 Hz

| Name of participant | NMIJ | | | | | | |
|--|--------------------------|--------|-----------------------------|-----------------------------------|-----------------------------------|-----------------------------------|-----------------------------------|
| DUT | 8305 sn2691032 | | | | | | |
| Acceleration amplitude | [m/s ²] | | | 50 | 50 | 100 | 100 |
| Vibration Frequency | [Hz] | | | 630 | 800 | 1000 | 1250 |
| Applied Calibration Method | | | | sine-approximation | sine-approximation | sine-approximation | sine-approximation |
| Number of measurements | | | | 240 | 240 | 240 | 240 |
| Charge Sensitivity | [pC/(m/s ²)] | | | 0.12777 | 0.12778 | 0.12785 | 0.12787 |
| Uncertainty Component | type (A or B) | factor | effective degree of freedom | Relative standard uncertainty [%] | Relative standard uncertainty [%] | Relative standard uncertainty [%] | Relative standard uncertainty [%] |
| Capacitance standard uncertainty | B | 1 | ∞ | 3.50E-03 | 3.50E-03 | 3.50E-03 | 3.50E-03 |
| Voltage standard uncertainty | B | 1 | ∞ | 5.00E-02 | 5.00E-02 | 5.00E-02 | 5.00E-02 |
| Calibration uncertainty of charge amplifier | A | 1 | 76 | 4.99E-03 | 3.66E-03 | 4.54E-03 | 2.07E-03 |
| Accuracy of voltmeter resolution | B | √3 | ∞ | 5.77E-07 | 5.77E-07 | 2.89E-07 | 2.89E-07 |
| Repeatability of sensitivity | A | 1 | 237 | 3.77E-04 | 5.08E-04 | 5.45E-03 | 4.89E-03 |
| Laser wavelength instability | B | 1 | ∞ | 1.60E-03 | 1.60E-03 | 1.60E-03 | 1.60E-03 |
| Vibration frequency instability | B | 1 | ∞ | 6.00E-03 | 6.00E-03 | 6.00E-03 | 6.00E-03 |
| Effect of phase disturbance on phase amplitude measurement | A | 1 | 237 | 2.37E-03 | 3.98E-03 | 3.43E-03 | 5.61E-03 |
| Attachment uncertainty of accelerometer | B | 1 | ∞ | 7.19E-02 | 7.19E-02 | 7.19E-02 | 7.19E-02 |
| Transverse sensitivity | B | 1 | ∞ | 7.80E-02 | 7.80E-02 | 7.80E-02 | 7.80E-02 |
| Effect of vibration distortion | B | 1 | ∞ | | | | |
| Relative combined uncertainty [%] | | | | 0.12 | 0.12 | 0.12 | 0.12 |
| Relative expanded uncertainty (<i>k</i> =2) [%] | | | | 0.24 | 0.24 | 0.24 | 0.24 |

Table A2-1-5 Uncertainty budget for the BB transducer at 1600 Hz to 3150 Hz

| Name of participant | NMIJ | | | | | | |
|--|--------------------------|--------|-----------------------------|-----------------------------------|-----------------------------------|-----------------------------------|-----------------------------------|
| DUT | 8305 sn2691032 | | | | | | |
| Acceleration amplitude | [m/s ²] | | | 100 | 100 | 100 | 100 |
| Vibration Frequency | [Hz] | | | 1600 | 2000 | 2500 | 3150 |
| Applied Calibration Method | | | | sine-approximation | sine-approximation | sine-approximation | sine-approximation |
| Number of measurements | | | | 240 | 240 | 240 | 240 |
| Charge Sensitivity | [pC/(m/s ²)] | | | 0.12803 | 0.12798 | 0.12836 | 0.12883 |
| Uncertainty Component | type (A or B) | factor | effective degree of freedom | Relative standard uncertainty [%] | Relative standard uncertainty [%] | Relative standard uncertainty [%] | Relative standard uncertainty [%] |
| Capacitance standard uncertainty | B | 1 | ∞ | 3.50E-03 | 3.50E-03 | 3.50E-03 | 3.50E-03 |
| Voltage standard uncertainty | B | 1 | ∞ | 5.00E-02 | 5.00E-02 | 5.00E-02 | 5.00E-02 |
| Calibration uncertainty of charge amplifier | A | 1 | 76 | 4.37E-04 | 5.65E-04 | 3.22E-04 | 3.65E-04 |
| Accuracy of voltmeter resolution | B | √3 | ∞ | 2.89E-07 | 2.89E-07 | 2.89E-07 | 2.89E-07 |
| Repeatability of sensitivity | A | 1 | 237 | 1.56E-03 | 2.16E-03 | 2.16E-03 | 4.83E-03 |
| Laser wavelength instability | B | 1 | ∞ | 1.60E-03 | 1.60E-03 | 1.60E-03 | 1.60E-03 |
| Vibration frequency instability | B | 1 | ∞ | 6.00E-03 | 6.00E-03 | 6.00E-03 | 6.00E-03 |
| Effect of phase disturbance on phase amplitude measurement | A | 1 | 237 | 7.65E-03 | 1.24E-02 | 2.06E-02 | 3.01E-02 |
| Attachment uncertainty of accelerometer | B | 1 | ∞ | 7.19E-02 | 7.19E-02 | 7.19E-02 | 7.19E-02 |
| Transverse sensitivity | B | 1 | ∞ | 7.80E-02 | 7.80E-02 | 7.80E-02 | 7.80E-02 |
| Effect of vibration distortion | B | 1 | ∞ | | | | |
| Relative combined uncertainty [%] | | | | 0.12 | 0.12 | 0.12 | 0.12 |
| Relative expanded uncertainty (k=2) [%] | | | | 0.24 | 0.24 | 0.24 | 0.24 |

Table A2-1-6 Uncertainty budget for the BB transducer at 4000 Hz to 5000 Hz

| Name of participant | NMIJ | | | | | | |
|--|--------------------------|--------|-----------------------------|-----------------------------------|-----------------------------------|-----------------------------------|-----------------------------------|
| DUT | 8305 sn2691032 | | | | | | |
| Acceleration amplitude | [m/s ²] | | | 100 | 100 | | |
| Vibration Frequency | [Hz] | | | 4000 | 5000 | | |
| Applied Calibration Method | | | | sine-approximation | sine-approximation | | |
| Number of measurements | | | | 240 | 240 | | |
| Charge Sensitivity | [pC/(m/s ²)] | | | 0.12917 | 0.13004 | | |
| Uncertainty Component | type (A or B) | factor | effective degree of freedom | Relative standard uncertainty [%] | Relative standard uncertainty [%] | Relative standard uncertainty [%] | Relative standard uncertainty [%] |
| Capacitance standard uncertainty | B | 1 | ∞ | 3.50E-03 | 3.50E-03 | | |
| Voltage standard uncertainty | B | 1 | ∞ | 5.00E-02 | 5.00E-02 | | |
| Calibration uncertainty of charge amplifier | A | 1 | 76 | 7.56E-04 | 1.34E-04 | | |
| Accuracy of voltmeter resolution | B | √3 | ∞ | 2.89E-07 | 2.89E-07 | | |
| Repeatability of sensitivity | A | 1 | 237 | 1.29E-03 | 6.46E-04 | | |
| Laser wavelength instability | B | 1 | ∞ | 1.60E-03 | 1.60E-03 | | |
| Vibration frequency instability | B | 1 | ∞ | 6.00E-03 | 6.00E-03 | | |
| Effect of phase disturbance on phase amplitude measurement | A | 1 | 237 | 4.89E-02 | 8.12E-02 | | |
| Attachment uncertainty of accelerometer | B | 1 | ∞ | 7.19E-02 | 7.19E-02 | | |
| Transverse sensitivity | B | 1 | ∞ | 7.80E-02 | 7.80E-02 | | |
| Effect of vibration distortion | B | 1 | ∞ | | | | |
| Relative combined uncertainty [%] | | | | 0.13 | 0.14 | | |
| Relative expanded uncertainty (k=2) [%] | | | | 0.25 | 0.29 | | |

Table A2-1-7 Uncertainty budget for the SE transducer at 40 Hz to 80 Hz

| Name of participant | NMIJ | | | | | | |
|--|--------------------------|--------|-----------------------------|-----------------------------------|-----------------------------------|-----------------------------------|-----------------------------------|
| DUT | 8305-001 sn2644378 | | | | | | |
| Acceleration amplitude | [m/s ²] | | | 10 | 10 | 10 | 10 |
| Vibration Frequency | [Hz] | | | 40 | 50 | 63 | 80 |
| Applied Calibration Method | | | | sine-approximation | sine-approximation | sine-approximation | sine-approximation |
| Number of measurements | | | | 240 | 240 | 240 | 240 |
| Charge Sensitivity | [pC/(m/s ²)] | | | 0.12914 | 0.12915 | 0.12919 | 0.12911 |
| Uncertainty Component | type (A or B) | factor | effective degree of freedom | Relative standard uncertainty [%] | Relative standard uncertainty [%] | Relative standard uncertainty [%] | Relative standard uncertainty [%] |
| Capacitance standard uncertainty | B | 1 | ∞ | 3.50E-03 | 3.50E-03 | 3.50E-03 | 3.50E-03 |
| Voltage standard uncertainty | B | 1 | ∞ | 5.00E-02 | 5.00E-02 | 5.00E-02 | 5.00E-02 |
| Calibration uncertainty of charge amplifier | A | 1 | 76 | 2.60E-02 | 1.34E-01 | 4.11E-02 | 1.90E-02 |
| Accuracy of voltmeter resolution | B | √3 | ∞ | 2.89E-06 | 2.89E-06 | 2.89E-06 | 2.89E-06 |
| Repeatability of sensitivity | A | 1 | 237 | 3.19E-03 | 5.93E-03 | 6.14E-03 | 2.85E-03 |
| Laser wavelength instability | B | 1 | ∞ | 1.60E-03 | 1.60E-03 | 1.60E-03 | 1.60E-03 |
| Vibration frequency instability | B | 1 | ∞ | 6.00E-03 | 6.00E-03 | 6.00E-03 | 6.00E-03 |
| Effect of phase disturbance on phase amplitude measurement | A | 1 | 237 | 1.33E-04 | 1.36E-04 | 1.29E-04 | 1.64E-04 |
| Attachment uncertainty of accelerometer | B | 1 | ∞ | 1.72E-01 | 1.72E-01 | 1.72E-01 | 1.72E-01 |
| Transverse sensitivity | B | 1 | ∞ | 7.80E-02 | 7.80E-02 | 7.80E-02 | 7.80E-02 |
| Effect of vibration distortion | B | 1 | ∞ | 6.72E-04 | 6.72E-04 | 3.23E-04 | 2.84E-04 |
| Relative combined uncertainty [%] | | | | 0.20 | 0.24 | 0.20 | 0.20 |
| Relative expanded uncertainty (k=2) [%] | | | | 0.39 | 0.47 | 0.40 | 0.39 |

Table A2-1-8 Uncertainty budget for the SE transducer at 100 Hz to 200 Hz

| Name of participant | NMIJ | | | | | | |
|--|--------------------------|--------|-----------------------------|-----------------------------------|-----------------------------------|-----------------------------------|-----------------------------------|
| DUT | 8305-001 sn2644378 | | | | | | |
| Acceleration amplitude | [m/s ²] | | | 10 | 10 | 10 | 10 |
| Vibration Frequency | [Hz] | | | 100 | 125 | 160 | 200 |
| Applied Calibration Method | | | | sine-approximation | sine-approximation | sine-approximation | sine-approximation |
| Number of measurements | | | | 240 | 240 | 240 | 240 |
| Charge Sensitivity | [pC/(m/s ²)] | | | 0.12913 | 0.12913 | 0.12913 | 0.12911 |
| Uncertainty Component | type (A or B) | factor | effective degree of freedom | Relative standard uncertainty [%] | Relative standard uncertainty [%] | Relative standard uncertainty [%] | Relative standard uncertainty [%] |
| Capacitance standard uncertainty | B | 1 | ∞ | 3.50E-03 | 3.50E-03 | 3.50E-03 | 3.50E-03 |
| Voltage standard uncertainty | B | 1 | ∞ | 5.00E-02 | 5.00E-02 | 5.00E-02 | 5.00E-02 |
| Calibration uncertainty of charge amplifier | A | 1 | 76 | 6.82E-02 | 2.12E-02 | 1.99E-02 | 4.12E-02 |
| Accuracy of voltmeter resolution | B | √3 | ∞ | 2.89E-06 | 2.89E-06 | 2.89E-06 | 2.89E-06 |
| Repeatability of sensitivity | A | 1 | 237 | 1.91E-02 | 2.41E-03 | 3.55E-03 | 9.95E-03 |
| Laser wavelength instability | B | 1 | ∞ | 1.60E-03 | 1.60E-03 | 1.60E-03 | 1.60E-03 |
| Vibration frequency instability | B | 1 | ∞ | 6.00E-03 | 6.00E-03 | 6.00E-03 | 6.00E-03 |
| Effect of phase disturbance on phase amplitude measurement | A | 1 | 237 | 2.03E-04 | 3.08E-04 | 5.08E-04 | 8.41E-04 |
| Attachment uncertainty of accelerometer | B | 1 | ∞ | 1.72E-01 | 1.72E-01 | 1.72E-01 | 1.72E-01 |
| Transverse sensitivity | B | 1 | ∞ | 7.80E-02 | 7.80E-02 | 7.80E-02 | 7.80E-02 |
| Effect of vibration distortion | B | 1 | ∞ | | | | |
| Relative combined uncertainty [%] | | | | 0.21 | 0.20 | 0.20 | 0.20 |
| Relative expanded uncertainty ($k=2$) [%] | | | | 0.42 | 0.39 | 0.39 | 0.40 |

Table A2-1-9 Uncertainty budget for the SE transducer at 250 Hz to 500 Hz

| Name of participant | NMIJ | | | | | | |
|--|--------------------------|--------|-----------------------------|-----------------------------------|-----------------------------------|-----------------------------------|-----------------------------------|
| DUT | 8305-001 sn2644378 | | | | | | |
| Acceleration amplitude | [m/s ²] | | | 10 | 50 | 50 | 50 |
| Vibration Frequency | [Hz] | | | 250 | 315 | 400 | 500 |
| Applied Calibration Method | | | | sine-approximation | sine-approximation | sine-approximation | sine-approximation |
| Number of measurements | | | | 240 | 240 | 240 | 240 |
| Charge Sensitivity | [pC/(m/s ²)] | | | 0.12909 | 0.12918 | 0.12916 | 0.12920 |
| Uncertainty Component | type (A or B) | factor | effective degree of freedom | Relative standard uncertainty [%] | Relative standard uncertainty [%] | Relative standard uncertainty [%] | Relative standard uncertainty [%] |
| Capacitance standard uncertainty | B | 1 | ∞ | 3.50E-03 | 3.50E-03 | 3.50E-03 | 3.50E-03 |
| Voltage standard uncertainty | B | 1 | ∞ | 5.00E-02 | 5.00E-02 | 5.00E-02 | 5.00E-02 |
| Calibration uncertainty of charge amplifier | A | 1 | 76 | 2.38E-02 | 8.86E-03 | 9.51E-03 | 9.62E-03 |
| Accuracy of voltmeter resolution | B | √3 | ∞ | 2.89E-06 | 5.77E-07 | 5.77E-07 | 5.77E-07 |
| Repeatability of sensitivity | A | 1 | 237 | 4.73E-03 | 4.75E-03 | 3.52E-03 | 3.63E-03 |
| Laser wavelength instability | B | 1 | ∞ | 1.60E-03 | 1.60E-03 | 1.60E-03 | 1.60E-03 |
| Vibration frequency instability | B | 1 | ∞ | 6.00E-03 | 6.00E-03 | 6.00E-03 | 6.00E-03 |
| Effect of phase disturbance on phase amplitude measurement | A | 1 | 237 | 1.56E-03 | 6.09E-04 | 8.41E-04 | 1.35E-03 |
| Attachment uncertainty of accelerometer | B | 1 | ∞ | 1.72E-01 | 1.72E-01 | 1.72E-01 | 1.72E-01 |
| Transverse sensitivity | B | 1 | ∞ | 7.80E-02 | 7.80E-02 | 7.80E-02 | 7.80E-02 |
| Effect of vibration distortion | B | 1 | ∞ | | | | |
| Relative combined uncertainty [%] | | | | 0.20 | 0.20 | 0.20 | 0.20 |
| Relative expanded uncertainty (k=2) [%] | | | | 0.39 | 0.39 | 0.39 | 0.39 |

Table A2-1-10 Uncertainty budget for the SE transducer at 630 Hz to 1250 Hz

| Name of participant | NMIJ | | | | | | |
|--|--------------------------|--------|-----------------------------|-----------------------------------|-----------------------------------|-----------------------------------|-----------------------------------|
| DUT | 8305-001 sn2644378 | | | | | | |
| Acceleration amplitude | [m/s ²] | | | 50 | 50 | 100 | 100 |
| Vibration Frequency | [Hz] | | | 630 | 800 | 1000 | 1250 |
| Applied Calibration Method | | | | sine-approximation | sine-approximation | sine-approximation | sine-approximation |
| Number of measurements | | | | 240 | 240 | 240 | 240 |
| Charge Sensitivity | [pC/(m/s ²)] | | | 0.12921 | 0.12923 | 0.12939 | 0.12945 |
| Uncertainty Component | type (A or B) | factor | effective degree of freedom | Relative standard uncertainty [%] | Relative standard uncertainty [%] | Relative standard uncertainty [%] | Relative standard uncertainty [%] |
| Capacitance standard uncertainty | B | 1 | ∞ | 3.50E-03 | 3.50E-03 | 3.50E-03 | 3.50E-03 |
| Voltage standard uncertainty | B | 1 | ∞ | 5.00E-02 | 5.00E-02 | 5.00E-02 | 5.00E-02 |
| Calibration uncertainty of charge amplifier | A | 1 | 76 | 1.01E-02 | 1.10E-02 | 5.66E-03 | 4.25E-03 |
| Accuracy of voltmeter resolution | B | √3 | ∞ | 5.77E-07 | 5.77E-07 | 2.89E-07 | 2.89E-07 |
| Repeatability of sensitivity | A | 1 | 237 | 1.05E-03 | 1.37E-03 | 1.48E-02 | 1.03E-02 |
| Laser wavelength instability | B | 1 | ∞ | 1.60E-03 | 1.60E-03 | 1.60E-03 | 1.60E-03 |
| Vibration frequency instability | B | 1 | ∞ | 6.00E-03 | 6.00E-03 | 6.00E-03 | 6.00E-03 |
| Effect of phase disturbance on phase amplitude measurement | A | 1 | 237 | 1.83E-03 | 3.04E-03 | 2.68E-03 | 4.37E-03 |
| Attachment uncertainty of accelerometer | B | 1 | ∞ | 1.72E-01 | 1.72E-01 | 1.72E-01 | 1.72E-01 |
| Transverse sensitivity | B | 1 | ∞ | 7.80E-02 | 7.80E-02 | 7.80E-02 | 7.80E-02 |
| Effect of vibration distortion | B | 1 | ∞ | | | | |
| Relative combined uncertainty [%] | | | | 0.20 | 0.20 | 0.20 | 0.20 |
| Relative expanded uncertainty (k=2) [%] | | | | 0.39 | 0.39 | 0.39 | 0.39 |

Table A2-1-11 Uncertainty budget for the SE transducer at 1600 Hz to 3150 Hz

| Name of participant | NMIJ | | | | | | |
|--|--------------------------|--------|-----------------------------|-----------------------------------|-----------------------------------|-----------------------------------|-----------------------------------|
| DUT | 8305-001 sn2644378 | | | | | | |
| Acceleration amplitude | [m/s ²] | | | 100 | 100 | 100 | 100 |
| Vibration Frequency | [Hz] | | | 1600 | 2000 | 2500 | 3150 |
| Applied Calibration Method | | | | sine-approximation | sine-approximation | sine-approximation | sine-approximation |
| Number of measurements | | | | 240 | 240 | 240 | 240 |
| Charge Sensitivity | [pC/(m/s ²)] | | | 0.12958 | 0.12976 | 0.13015 | 0.13085 |
| Uncertainty Component | type (A or B) | factor | effective degree of freedom | Relative standard uncertainty [%] | Relative standard uncertainty [%] | Relative standard uncertainty [%] | Relative standard uncertainty [%] |
| Capacitance standard uncertainty | B | 1 | ∞ | 3.50E-03 | 3.50E-03 | 3.50E-03 | 3.50E-03 |
| Voltage standard uncertainty | B | 1 | ∞ | 5.00E-02 | 5.00E-02 | 5.00E-02 | 5.00E-02 |
| Calibration uncertainty of charge amplifier | A | 1 | 76 | 4.50E-03 | 4.99E-03 | 4.98E-03 | 4.76E-03 |
| Accuracy of voltmeter resolution | B | √3 | ∞ | 2.89E-07 | 2.89E-07 | 2.89E-07 | 2.89E-07 |
| Repeatability of sensitivity | A | 1 | 237 | 6.10E-03 | 6.84E-03 | 5.93E-03 | 1.43E-02 |
| Laser wavelength instability | B | 1 | ∞ | 1.60E-03 | 1.60E-03 | 1.60E-03 | 1.60E-03 |
| Vibration frequency instability | B | 1 | ∞ | 6.00E-03 | 6.00E-03 | 6.00E-03 | 6.00E-03 |
| Effect of phase disturbance on phase amplitude measurement | A | 1 | 237 | 5.93E-03 | 9.55E-03 | 1.57E-02 | 2.35E-02 |
| Attachment uncertainty of accelerometer | B | 1 | ∞ | 1.72E-01 | 1.72E-01 | 1.72E-01 | 1.72E-01 |
| Transverse sensitivity | B | 1 | ∞ | 7.80E-02 | 7.80E-02 | 7.80E-02 | 7.80E-02 |
| Effect of vibration distortion | B | 1 | ∞ | | | | |
| Relative combined uncertainty [%] | | | | 0.20 | 0.20 | 0.20 | 0.20 |
| Relative expanded uncertainty (k=2) [%] | | | | 0.39 | 0.39 | 0.39 | 0.39 |

Table A2-1-12 Uncertainty budget for the SE transducer at 4000 Hz to 5000 Hz

| Name of participant | NMIJ | | | | | | |
|--|--------------------------|--------|-----------------------------|-----------------------------------|-----------------------------------|-----------------------------------|-----------------------------------|
| DUT | 8305-001 sn2644378 | | | | | | |
| Acceleration amplitude | [m/s ²] | | | 100 | 100 | | |
| Vibration Frequency | [Hz] | | | 4000 | 5000 | | |
| Applied Calibration Method | | | | sine-approximation | sine-approximation | | |
| Number of measurements | | | | 240 | 240 | | |
| Charge Sensitivity | [pC/(m/s ²)] | | | 0.13170 | 0.13274 | | |
| Uncertainty Component | type (A or B) | factor | effective degree of freedom | Relative standard uncertainty [%] | Relative standard uncertainty [%] | Relative standard uncertainty [%] | Relative standard uncertainty [%] |
| Capacitance standard uncertainty | B | 1 | ∞ | 3.50E-03 | 3.50E-03 | | |
| Voltage standard uncertainty | B | 1 | ∞ | 5.00E-02 | 5.00E-02 | | |
| Calibration uncertainty of charge amplifier | A | 1 | 76 | 3.46E-03 | 3.87E-03 | | |
| Accuracy of voltmeter resolution | B | √3 | ∞ | 2.89E-07 | 2.89E-07 | | |
| Repeatability of sensitivity | A | 1 | 237 | 4.10E-03 | 2.98E-03 | | |
| Laser wavelength instability | B | 1 | ∞ | 1.60E-03 | 1.60E-03 | | |
| Vibration frequency instability | B | 1 | ∞ | 6.00E-03 | 6.00E-03 | | |
| Effect of phase disturbance on phase amplitude measurement | A | 1 | 237 | 3.72E-02 | 6.06E-02 | | |
| Attachment uncertainty of accelerometer | B | 1 | ∞ | 1.72E-01 | 1.72E-01 | | |
| Transverse sensitivity | B | 1 | ∞ | 7.80E-02 | 7.80E-02 | | |
| Effect of vibration distortion | B | 1 | ∞ | | | | |
| Relative combined uncertainty [%] | | | | 0.20 | 0.20 | | |
| Relative expanded uncertainty (<i>k</i> =2) [%] | | | | 0.40 | 0.41 | | |

A2-2: CMS-ITRI

Table A2-2-1 Uncertainty budget for the BB transducer at 40 Hz to 100 Hz

| Acceleration amplitude | [m/s ²] | | | | 10 | 15 | 15 | 15 | 20 |
|---|--|---------------|--------|-----------------------------|-----------------------------------|-----------------------------------|-----------------------------------|-----------------------------------|-----------------------------------|
| Vibration Frequency | [Hz] | | | | 40 | 50 | 63 | 80 | 100 |
| Applied Calibration Method | | | | | Sine Approximation | Sine Approximation | Sine Approximation | Sine Approximation | Sine Approximation |
| Number of measurements | | | | | 90 | 90 | 90 | 90 | 90 |
| Charge Sensitivity | [pC/(m/s ²)] | | | | 0.127941 | 0.127880 | 0.127929 | 0.127912 | 0.127920 |
| Parameter of Measurement Function | Uncertainty Component | type (A or B) | factor | effective degree of freedom | Relative standard uncertainty [%] | Relative standard uncertainty [%] | Relative standard uncertainty [%] | Relative standard uncertainty [%] | Relative standard uncertainty [%] |
| <i>V</i> (Accelerometer Output Voltage Measurement) | Voltage Measurement Repeatability | A | 1 | 87 | 0.0366% | 0.0189% | 0.0098% | 0.0055% | 0.0347% |
| | Voltage Traceability | B | 2 | ∞ | 0.08000% | 0.03500% | 0.03500% | 0.03500% | 0.03500% |
| | Voltage Measurement Resolution | B | √3 | ∞ | 0.0003% | 0.0003% | 0.0003% | 0.0003% | 0.0003% |
| | Effect of transverse, rocking and bending acceleration on accelerometer output voltage measurement | B | √3 | 200 | 0.0020% | 0.0023% | 0.0027% | 0.0029% | 0.0043% |
| | Distortion on Voltage Measurement | B | √3 | 200 | 0.0548% | 0.0751% | 0.0543% | 0.0419% | 0.0563% |
| | Effect of long-term data distribution | A | 1 | 24 | 0.0354% | 0.0370% | 0.0421% | 0.0377% | 0.0388% |
| | Effect of Charge Amplifier (with the fixed) | B | 2.57 | ∞ | 0.0973% | 0.0973% | 0.0973% | 0.0973% | 0.0973% |
| <i>f</i> (Vibration Frequency Measurement) | Frequency Measurement | A | 1 | 87 | 0.0217011% | 0.0161377% | 0.0093735% | 0.0078710% | 0.0105945% |
| | Frequency Traceability | B | 2 | ∞ | 0.000025% | 0.000025% | 0.000025% | 0.000025% | 0.000025% |
| | Frequency Measurement Resolution | B | √3 | ∞ | 0.000289% | 0.002887% | 0.002887% | 0.002887% | 0.002887% |
| <i>D</i> (Vibration Displacement Measurement) | Amplitude measurement Errors from Sine Approximation Method, Evaluated by Least-Square Method | A | 1 | 59955 | 0.0040% | 0.0045% | 0.0040% | 0.0004% | 0.0005% |
| | Amplitude Measurement Repeatability | A | 1 | 87 | 0.0427% | 0.0205% | 0.0060% | 0.0070% | 0.0369% |
| | Amplitude Traceability (Laser Tube) | B | 1.96 | ∞ | 0.00000434% | 0.00000434% | 0.00000434% | 0.00000434% | 0.00000434% |
| | Amplitude Measurement Resolution | B | √3 | ∞ | 0.0141% | 0.0141% | 0.0141% | 0.0141% | 0.0141% |
| coverage factor, <i>k</i> = | (Ref. Calculation for GUM) | | | | 1.96 | 1.96 | 1.96 | 1.96 | 1.96 |
| effective degrees of freedom, <i>v</i> _{eff} | | | | | 3419 | 1530 | 1453 | 2043 | 1852 |
| Relative combined (Standard) uncertainty [%] | | | | | 0.16% | 0.14% | 0.13% | 0.12% | 0.14% |
| Relative expanded uncertainty (<i>k</i> = 1.96) [%] | | | | | 0.31% | 0.28% | 0.25% | 0.24% | 0.27% |

Table A2-2-2 Uncertainty budget for the BB transducer at 125 Hz to 315 Hz

| Acceleration amplitude | [m/s ²] | | | | 20 | 30 | 50 | 50 | 50 |
|--|---|---------------|--------|-----------------------------|-----------------------------------|-----------------------------------|-----------------------------------|-----------------------------------|-----------------------------------|
| Vibration Frequency | [Hz] | | | | 125 | 160 | 200 | 250 | 315 |
| Applied Calibration Method | | | | | Sine Approximation | Sine Approximation | Sine Approximation | Sine Approximation | Sine Approximation |
| Number of measurements | | | | | 90 | 90 | 90 | 90 | 90 |
| Charge Sensitivity | [pC/(m/s ²)] | | | | 0.127870 | 0.127928 | 0.127909 | 0.127853 | 0.127937 |
| Parameter of Measurement Function | Uncertainty Component | type (A or B) | factor | effective degree of freedom | Relative standard uncertainty [%] | Relative standard uncertainty [%] | Relative standard uncertainty [%] | Relative standard uncertainty [%] | Relative standard uncertainty [%] |
| V (Accelerometer Output Voltage Measurement) | Voltage Measurement Repeatability | A | 1 | 87 | 0.0159% | 0.0242% | 0.0176% | 0.0096% | 0.0246% |
| | Voltage Traceability | B | 2 | ∞ | 0.03500% | 0.03500% | 0.03500% | 0.03500% | 0.03500% |
| | Voltage Measurement Resolution | B | √3 | ∞ | 0.0003% | 0.0003% | 0.0003% | 0.0003% | 0.0003% |
| | Effect of transverse, rocking and bending acceleration on accelerometer output voltage measurement (transverse sensitivity) | B | √3 | 200 | 0.0068% | 0.0136% | 0.0214% | 0.0229% | 0.0190% |
| | Distortion on Voltage Measurement | B | √3 | 200 | 0.0502% | 0.0566% | 0.0629% | 0.0635% | 0.0924% |
| | Effect of long-term data distribution | A | 1 | 24 | 0.0388% | 0.0380% | 0.0387% | 0.0507% | 0.0348% |
| | Effect of Charge Amplifier (with the fixed setting) | B | 2.57 | ∞ | 0.0973% | 0.0973% | 0.0973% | 0.0973% | 0.0973% |
| f (Vibration Frequency Measurement) | Frequency Measurement Repeatability | A | 1 | 87 | 0.0077252% | 0.0064974% | 0.0017178% | 0.0025829% | 0.0029180% |
| | Frequency Traceability | B | 2 | ∞ | 0.000025% | 0.000025% | 0.000025% | 0.000025% | 0.000025% |
| | Frequency Measurement Resolution | B | √3 | ∞ | 0.002887% | 0.002887% | 0.002887% | 0.002887% | 0.002887% |
| D (Vibration Displacement Measurement) | Amplitude measurement Errors from Sine Approximation Method, Evaluated by Least-Square Method | A | 1 | 59955 | 0.0010% | 0.0007% | 0.0233% | 0.0240% | 0.0238% |
| | Amplitude Measurement Repeatability | A | 1 | 87 | 0.0167% | 0.0415% | 0.0200% | 0.0316% | 0.0437% |
| | Amplitude Traceability (Laser Tube) | B | 1.96 | ∞ | 0.00000434% | 0.00000434% | 0.00000434% | 0.00000434% | 0.00000434% |
| | Amplitude Measurement Resolution | B | √3 | ∞ | 0.0141% | 0.0141% | 0.0141% | 0.0141% | 0.0141% |
| coverage factor, $k =$ | (Ref. Calculation for CMS) | | | | 1.96 | 1.96 | 1.96 | 1.96 | 1.96 |
| effective degrees of freedom, ν_{eff} | | | | | 1905 | 1863 | 1866 | 1056 | 1234 |
| Relative combined (Standard) uncertainty [%] | | | | | 0.13% | 0.13% | 0.13% | 0.14% | 0.16% |
| Relative expanded uncertainty ($k = 1.96$) [%] | | | | | 0.25% | 0.27% | 0.27% | 0.28% | 0.31% |

Table A2-2-3 Uncertainty budget for the BB transducer at 400 Hz to 1000 Hz

| Acceleration amplitude | [m/s ²] | | | | 50 | 50 | 50 | 50 | 50 |
|---|---|---------------|--------|-----------------------------|-----------------------------------|-----------------------------------|-----------------------------------|-----------------------------------|-----------------------------------|
| Vibration Frequency | [Hz] | | | | 400 | 500 | 630 | 800 | 1000 |
| Applied Calibration Method | | | | | Sine Approximation | Sine Approximation | Sine Approximation | Sine Approximation | Sine Approximation |
| Number of measurements | | | | | 90 | 90 | 90 | 90 | 90 |
| Charge Sensitivity | [pC/(m/s ²)] | | | | 0.127911 | 0.127946 | 0.127946 | 0.127899 | 0.127898 |
| Parameter of Measurement Function | Uncertainty Component | type (A or B) | factor | effective degree of freedom | Relative standard uncertainty [%] | Relative standard uncertainty [%] | Relative standard uncertainty [%] | Relative standard uncertainty [%] | Relative standard uncertainty [%] |
| <i>V</i> (Accelerometer Output Voltage Measurement) | Voltage Measurement Repeatability | A | 1 | 87 | 0.0034% | 0.0099% | 0.0127% | 0.0465% | 0.0221% |
| | Voltage Traceability | B | 2 | ∞ | 0.03500% | 0.03500% | 0.03500% | 0.03500% | 0.03500% |
| | Voltage Measurement Resolution | B | √3 | ∞ | 0.0003% | 0.0003% | 0.0003% | 0.0003% | 0.0003% |
| | Effect of transverse, rocking and bending acceleration on accelerometer output voltage measurement (transverse sensitivity) | B | √3 | 200 | 0.0194% | 0.0128% | 0.0237% | 0.0637% | 0.0493% |
| | Distortion on Voltage Measurement | B | √3 | 200 | 0.0924% | 0.0837% | 0.0722% | 0.0606% | 0.0456% |
| | Effect of long-term data distribution | A | 1 | 24 | 0.0401% | 0.0635% | 0.0719% | 0.0932% | 0.1306% |
| | Effect of Charge Amplifier (with the fixed setting) | B | 2.57 | ∞ | 0.0973% | 0.0973% | 0.0973% | 0.0973% | 0.0973% |
| <i>f</i> (Vibration Frequency Measurement) | Frequency Measurement Repeatability | A | 1 | 87 | 0.0024302% | 0.0041280% | 0.0029188% | 0.0025093% | 0.0244664% |
| | Frequency Traceability | B | 2 | ∞ | 0.000025% | 0.000025% | 0.000025% | 0.000025% | 0.000025% |
| | Frequency Measurement Resolution | B | √3 | ∞ | 0.002887% | 0.002887% | 0.002887% | 0.002887% | 0.002887% |
| <i>D</i> (Vibration Displacement Measurement) | Amplitude measurement Errors from Sine Approximation Method, Evaluated by Least-Square Method | A | 1 | 59955 | 0.0067% | 0.0042% | 0.0069% | 0.0109% | 0.0196% |
| | Amplitude Measurement Repeatability | A | 1 | 87 | 0.0158% | 0.0717% | 0.0117% | 0.0421% | 0.0200% |
| | Amplitude Traceability (Laser Tube) | B | 1.96 | ∞ | 0.00000434% | 0.00000434% | 0.00000434% | 0.00000434% | 0.00000434% |
| | Amplitude Measurement Resolution | B | √3 | ∞ | 0.0141% | 0.0141% | 0.0141% | 0.0141% | 0.0141% |
| coverage factor, <i>k</i> = | (Ref. Calculation for CMS) | | | | 1.96 | 1.96 | 1.96 | 1.96 | 1.96 |
| effective degrees of freedom, <i>v</i> _{eff} | | | | | 1000 | 612 | 394 | 298 | 100 |
| Relative combined (Standard) uncertainty [%] | | | | | 0.15% | 0.17% | 0.15% | 0.18% | 0.19% |
| Relative expanded uncertainty (<i>k</i> = 1.96) [%] | | | | | 0.29% | 0.33% | 0.30% | 0.35% | 0.37% |

Table A2-2-4 Uncertainty budget for the BB transducer at 1.25 kHz to 5 kHz

| Acceleration amplitude | [m/s ²] | | | | 50 | 50 | 50 | 50 | 70 | 70 | 70 |
|--|---|---------------|--------|-----------------------------|-----------------------------------|-----------------------------------|-----------------------------------|-----------------------------------|-----------------------------------|-----------------------------------|-----------------------------------|
| Vibration Frequency | [Hz] | | | | 1250 | 1600 | 2000 | 2500 | 3150 | 4000 | 5000 |
| Applied Calibration Method | | | | | Sine Approximation | Sine Approximation | Sine Approximation | Sine Approximation | Sine Approximation | Sine Approximation | Sine Approximation |
| Number of measurements | | | | | 90 | 90 | 90 | 90 | 90 | 90 | 90 |
| Charge Sensitivity | [pC/(m/s ²)] | | | | 0.127852 | 0.127888 | 0.127870 | 0.127893 | 0.128189 | 0.129207 | 0.129481 |
| Parameter of Measurement Function | Uncertainty Component | type (A or B) | factor | effective degree of freedom | Relative standard uncertainty [%] | Relative standard uncertainty [%] | Relative standard uncertainty [%] | Relative standard uncertainty [%] | Relative standard uncertainty [%] | Relative standard uncertainty [%] | Relative standard uncertainty [%] |
| V (Accelerometer Output Voltage Measurement) | Voltage Measurement Repeatability | A | 1 | 87 | 0.0192% | 0.0049% | 0.0032% | 0.0036% | 0.0038% | 0.0067% | 0.0044% |
| | Voltage Traceability | B | 2 | ∞ | 0.03500% | 0.03500% | 0.03500% | 0.03500% | 0.03500% | 0.03500% | 0.03500% |
| | Voltage Measurement Resolution | B | √3 | ∞ | 0.0003% | 0.0003% | 0.0003% | 0.0003% | 0.0003% | 0.0003% | 0.0003% |
| | Effect of transverse, rocking and bending acceleration on accelerometer output voltage measurement (transverse sensitivity) | B | √3 | 200 | 0.0609% | 0.0389% | 0.0223% | 0.0308% | 0.0899% | 0.2920% | 0.2346% |
| | Distortion on Voltage Measurement | B | √3 | 200 | 0.0404% | 0.0260% | 0.0289% | 0.0289% | 0.0346% | 0.0387% | 0.0346% |
| | Effect of long-term data distribution | A | 1 | 24 | 0.1389% | 0.0567% | 0.0766% | 0.0523% | 0.0880% | 0.0918% | 0.1430% |
| | Effect of Charge Amplifier (with the fixed setting) | B | 2.57 | ∞ | 0.0973% | 0.0973% | 0.0973% | 0.0973% | 0.0973% | 0.0973% | 0.0973% |
| f (Vibration Frequency Measurement) | Frequency Measurement Repeatability | A | 1 | 87 | 0.0105345% | 0.0060821% | 0.0060821% | 0.0060821% | 0.0060821% | 0.0105345% | 0.0060821% |
| | Frequency Traceability | B | 2 | ∞ | 0.000025% | 0.000025% | 0.000025% | 0.000025% | 0.000025% | 0.000025% | 0.000025% |
| | Frequency Measurement Resolution | B | √3 | ∞ | 0.002887% | 0.002887% | 0.002887% | 0.002887% | 0.002887% | 0.002887% | 0.002887% |
| D (Vibration Displacement Measurement) | Amplitude measurement Errors from Sine Approximation Method, Evaluated by Least-Square Method | A | 1 | 59955 | 0.03050% | 0.0452% | 0.0617% | 0.1015% | 0.1553% | 0.2356% | 0.3201% |
| | Amplitude Measurement Repeatability | A | 1 | 87 | 0.0333% | 0.0222% | 0.0550% | 0.0470% | 0.0320% | 0.0604% | 0.0702% |
| | Amplitude Traceability (Laser Tube) | B | 1.96 | ∞ | 0.00000434% | 0.00000434% | 0.00000434% | 0.00000434% | 0.00000434% | 0.00000434% | 0.00000434% |
| | Amplitude Measurement Resolution | B | √3 | ∞ | 0.0141% | 0.0141% | 0.0141% | 0.0141% | 0.0141% | 0.0141% | 0.0141% |
| coverage factor, k = | (Ref. Calculation for CMS) | | | | 1.96 | 1.96 | 1.96 | 1.96 | 1.96 | 1.96 | 1.96 |
| effective degrees of freedom, v _{eff} | | | | | 94 | 825 | 405 | 2102 | 1116 | 8619 | 2126 |
| Relative combined (Standard) uncertainty [%] | | | | | 0.20% | 0.14% | 0.16% | 0.17% | 0.23% | 0.41% | 0.44% |
| Relative expanded uncertainty (k=1.96) [%] | | | | | 0.39% | 0.27% | 0.32% | 0.33% | 0.46% | 0.80% | 0.87% |

Table A2-2-5 Uncertainty budget for the SE transducer at 40 Hz to 100 Hz

| Acceleration amplitude | [m/s ²] | | | | 10 | 15 | 15 | 15 | 20 |
|--|---|---------------|--------|-----------------------------|-----------------------------------|-----------------------------------|-----------------------------------|-----------------------------------|-----------------------------------|
| Vibration Frequency | [Hz] | | | | 40 | 50 | 63 | 80 | 100 |
| Applied Calibration Method | | | | | Sine Approximation | Sine Approximation | Sine Approximation | Sine Approximation | Sine Approximation |
| Number of measurements | | | | | 90 | 90 | 90 | 90 | 90 |
| Charge Sensitivity | [pC/(m/s ²)] | | | | 0.129433647 | 0.129311 | 0.129349 | 0.129320 | 0.129335 |
| Parameter of Measurement Function | Uncertainty Component | type (A or B) | factor | effective degree of freedom | Relative standard uncertainty [%] | Relative standard uncertainty [%] | Relative standard uncertainty [%] | Relative standard uncertainty [%] | Relative standard uncertainty [%] |
| <i>V</i> (Accelerometer Output Voltage Measurement) | Voltage Measurement Repeatability | A | 1 | 87 | 0.0202% | 0.0382% | 0.0126% | 0.0060% | 0.0070% |
| | Voltage Traceability | B | 2 | ∞ | 0.08000% | 0.03500% | 0.03500% | 0.03500% | 0.03500% |
| | Voltage Measurement Resolution | B | √3 | ∞ | 0.0003% | 0.0003% | 0.0003% | 0.0003% | 0.0003% |
| | Effect of transverse, rocking and bending acceleration on accelerometer output voltage measurement (transverse sensitivity) | B | √3 | 200 | 0.0020% | 0.0023% | 0.0027% | 0.0029% | 0.0043% |
| | Distortion on Voltage Measurement | B | √3 | 200 | 0.0548% | 0.0751% | 0.0543% | 0.0419% | 0.0563% |
| | Effect of long-term data distribution | A | 1 | 24 | 0.0354% | 0.0370% | 0.0421% | 0.0377% | 0.0388% |
| | Effect of Charge Amplifier (with the fixed setting) | B | 2.57 | ∞ | 0.0973% | 0.0973% | 0.0973% | 0.0973% | 0.0973% |
| <i>f</i> (Vibration Frequency Measurement) | Frequency Measurement Repeatability | A | 1 | 87 | 0.0217011% | 0.0161377% | 0.0093735% | 0.0078710% | 0.0105945% |
| | Frequency Traceability | B | 2 | ∞ | 0.000025% | 0.000025% | 0.000025% | 0.000025% | 0.000025% |
| | Frequency Measurement Resolution | B | √3 | ∞ | 0.000289% | 0.002887% | 0.002887% | 0.002887% | 0.002887% |
| <i>D</i> (Vibration Displacement Measurement) | Amplitude measurement Errors from Sine Approximation Method, Evaluated by Least-Square Method | A | 1 | 59955 | 0.0040% | 0.0045% | 0.0040% | 0.0004% | 0.0005% |
| | Amplitude Measurement Repeatability | A | 1 | 87 | 0.0242% | 0.0391% | 0.0099% | 0.0059% | 0.0074% |
| | Amplitude Traceability (Laser Tube) | B | 1.96 | ∞ | 0.00000434% | 0.00000434% | 0.00000434% | 0.00000434% | 0.00000434% |
| | Amplitude Measurement Resolution | B | √3 | ∞ | 0.0141% | 0.0141% | 0.0141% | 0.0141% | 0.0141% |
| coverage factor, <i>k</i> = (Ref. Calculation for CMS) | | | | | 1.96 | 1.96 | 1.96 | 1.96 | 1.96 |
| effective degrees of freedom, <i>v</i> _{eff} | | | | | 4072 | 1588 | 1474 | 2041 | 1753 |
| Relative combined (Standard) uncertainty [%] | | | | | 0.15% | 0.15% | 0.13% | 0.12% | 0.13% |
| Relative expanded uncertainty (<i>k</i> =1.96) [%] | | | | | 0.30% | 0.29% | 0.25% | 0.24% | 0.25% |

Table A2-2-6 Uncertainty budget for the SE transducer at 125 Hz to 315 Hz

| Acceleration amplitude | [m/s ²] | | | | 20 | 30 | 50 | 50 | 50 |
|---|---|---------------|--------|-----------------------------|-----------------------------------|-----------------------------------|-----------------------------------|-----------------------------------|-----------------------------------|
| Vibration Frequency | [Hz] | | | | 125 | 160 | 200 | 250 | 315 |
| Applied Calibration Method | | | | | Sine Approximation | Sine Approximation | Sine Approximation | Sine Approximation | Sine Approximation |
| Number of measurements | | | | | 90 | 90 | 90 | 90 | 90 |
| Charge Sensitivity | [pC/(m/s ²)] | | | | 0.129253 | 0.129355 | 0.129274 | 0.129251 | 0.129310 |
| Parameter of Measurement Function | Uncertainty Component | type (A or B) | factor | effective degree of freedom | Relative standard uncertainty [%] | Relative standard uncertainty [%] | Relative standard uncertainty [%] | Relative standard uncertainty [%] | Relative standard uncertainty [%] |
| <i>V</i> (Accelerometer Output Voltage Measurement) | Voltage Measurement Repeatability | A | 1 | 87 | 0.0073% | 0.0098% | 0.0115% | 0.0132% | 0.0178% |
| | Voltage Traceability | B | 2 | ∞ | 0.03500% | 0.03500% | 0.03500% | 0.03500% | 0.03500% |
| | Voltage Measurement Resolution | B | √3 | ∞ | 0.0003% | 0.0003% | 0.0003% | 0.0003% | 0.0003% |
| | Effect of transverse, rocking and bending acceleration on accelerometer output voltage measurement (transverse sensitivity) | B | √3 | 200 | 0.0068% | 0.0136% | 0.0214% | 0.0229% | 0.0190% |
| | Distortion on Voltage Measurement | B | √3 | 200 | 0.0502% | 0.0566% | 0.0629% | 0.0635% | 0.0924% |
| | Effect of long-term data distribution | A | 1 | 24 | 0.0388% | 0.0380% | 0.0387% | 0.0507% | 0.0348% |
| | Effect of Charge Amplifier (with the fixed setting) | B | 2.57 | ∞ | 0.0973% | 0.0973% | 0.0973% | 0.0973% | 0.0973% |
| <i>f</i> (Vibration Frequency Measurement) | Frequency Measurement Repeatability | A | 1 | 87 | 0.0077252% | 0.0064974% | 0.0017178% | 0.0025829% | 0.0029180% |
| | Frequency Traceability | B | 2 | ∞ | 0.000025% | 0.000025% | 0.000025% | 0.000025% | 0.000025% |
| | Frequency Measurement Resolution | B | √3 | ∞ | 0.002887% | 0.002887% | 0.002887% | 0.002887% | 0.002887% |
| <i>D</i> (Vibration Displacement Measurement) | Amplitude measurement Errors from Sine Approximation Method, Evaluated by Least-Square Method | A | 1 | 59955 | 0.0010% | 0.0007% | 0.0233% | 0.0240% | 0.0238% |
| | Amplitude Measurement Repeatability | A | 1 | 87 | 0.0085% | 0.0119% | 0.0077% | 0.0092% | 0.0249% |
| | Amplitude Traceability (Laser Tube) | B | 1.96 | ∞ | 0.00000434% | 0.00000434% | 0.00000434% | 0.00000434% | 0.00000434% |
| | Amplitude Measurement Resolution | B | √3 | ∞ | 0.0141% | 0.0141% | 0.0141% | 0.0141% | 0.0141% |
| coverage factor, <i>k</i> = | (Ref. Calculation for CMS) | | | | 1.96 | 1.96 | 1.96 | 1.96 | 1.96 |
| effective degrees of freedom, <i>v</i> _{eff} | | | | | 1829 | 1858 | 1788 | 998 | 1179 |
| Relative combined (Standard) uncertainty [%] | | | | | 0.13% | 0.13% | 0.14% | 0.14% | 0.16% |
| Relative expanded uncertainty (<i>k</i> =1.96) [%] | | | | | 0.25% | 0.25% | 0.26% | 0.27% | 0.30% |

Table A2-2-7 Uncertainty budget for the SE transducer at 400 Hz to 1000 Hz

| Acceleration amplitude | [m/s ²] | | | | 50 | 50 | 50 | 50 | 50 |
|---|---|---------------|--------|-----------------------------|-----------------------------------|-----------------------------------|-----------------------------------|-----------------------------------|-----------------------------------|
| Vibration Frequency | [Hz] | | | | 400 | 500 | 630 | 800 | 1000 |
| Applied Calibration Method | | | | | Sine Approximation | Sine Approximation | Sine Approximation | Sine Approximation | Sine Approximation |
| Number of measurements | | | | | 90 | 90 | 90 | 90 | 90 |
| Charge Sensitivity | [pC/(m/s ²)] | | | | 0.129347 | 0.129415 | 0.129446 | 0.129282 | 0.129413 |
| Parameter of Measurement Function | Uncertainty Component | type (A or B) | factor | effective degree of freedom | Relative standard uncertainty [%] | Relative standard uncertainty [%] | Relative standard uncertainty [%] | Relative standard uncertainty [%] | Relative standard uncertainty [%] |
| <i>V</i> (Accelerometer Output Voltage Measurement) | Voltage Measurement Repeatability | A | 1 | 87 | 0.0139% | 0.0073% | 0.0057% | 0.0153% | 0.0668% |
| | Voltage Traceability | B | 2 | ∞ | 0.03500% | 0.03500% | 0.03500% | 0.03500% | 0.03500% |
| | Voltage Measurement Resolution | B | √3 | ∞ | 0.0003% | 0.0003% | 0.0003% | 0.0003% | 0.0003% |
| | Effect of transverse, rocking and bending acceleration on accelerometer output voltage measurement (transverse sensitivity) | B | √3 | 200 | 0.0194% | 0.0128% | 0.0237% | 0.0637% | 0.0493% |
| | Distortion on Voltage Measurement | B | √3 | 200 | 0.0924% | 0.0837% | 0.0722% | 0.0606% | 0.0456% |
| | Effect of long-term data distribution | A | 1 | 24 | 0.0401% | 0.0635% | 0.0719% | 0.0932% | 0.1306% |
| | Effect of Charge Amplifier (with the fixed setting) | B | 2.57 | ∞ | 0.0973% | 0.0973% | 0.0973% | 0.0973% | 0.0973% |
| <i>f</i> (Vibration Frequency Measurement) | Frequency Measurement Repeatability | A | 1 | 87 | 0.0024302% | 0.0041280% | 0.0029188% | 0.0025093% | 0.0244664% |
| | Frequency Traceability | B | 2 | ∞ | 0.000025% | 0.000025% | 0.000025% | 0.000025% | 0.000025% |
| | Frequency Measurement Resolution | B | √3 | ∞ | 0.002887% | 0.002887% | 0.002887% | 0.002887% | 0.002887% |
| <i>D</i> (Vibration Displacement Measurement) | Amplitude measurement Errors from Sine Approximation Method, Evaluated by Least-Square Method | A | 1 | 59955 | 0.0067% | 0.0042% | 0.0069% | 0.0109% | 0.0196% |
| | Amplitude Measurement Repeatability | A | 1 | 87 | 0.0145% | 0.0125% | 0.0152% | 0.0248% | 0.0974% |
| | Amplitude Traceability (Laser Tube) | B | 1.96 | ∞ | 0.00000434% | 0.00000434% | 0.00000434% | 0.00000434% | 0.00000434% |
| | Amplitude Measurement Resolution | B | √3 | ∞ | 0.0141% | 0.0141% | 0.0141% | 0.0141% | 0.0141% |
| coverage factor, <i>k</i> = | (Ref. Calculation for CMS) | | | | 1.96 | 1.96 | 1.96 | 1.96 | 1.96 |
| effective degrees of freedom, <i>v</i> _{eff} | | | | | 1013 | 542 | 393 | 249 | 172 |
| Relative combined (Standard) uncertainty [%] | | | | | 0.15% | 0.15% | 0.15% | 0.17% | 0.22% |
| Relative expanded uncertainty (<i>k</i> =1.96) [%] | | | | | 0.29% | 0.30% | 0.30% | 0.33% | 0.43% |

Table A2-2-8 Uncertainty budget for the SE transducer at 1.25 kHz to 5 kHz

| Acceleration amplitude | [m/s ²] | | | | 50 | 50 | 50 | 50 | 70 | 70 | 70 |
|---|---|---------------|--------|-----------------------------|-----------------------------------|-----------------------------------|-----------------------------------|-----------------------------------|-----------------------------------|-----------------------------------|-----------------------------------|
| Vibration Frequency | [Hz] | | | | 1250 | 1600 | 2000 | 2500 | 3150 | 4000 | 5000 |
| Applied Calibration Method | | | | | Sine Approximation | Sine Approximation | Sine Approximation | Sine Approximation | Sine Approximation | Sine Approximation | Sine Approximation |
| Number of measurements | | | | | 90 | 90 | 90 | 90 | 90 | 90 | 90 |
| Charge Sensitivity | [pC/(m/s ²)] | | | | 0.129386 | 0.129415 | 0.129349 | 0.129402 | 0.130313 | 0.131491 | 0.132050 |
| Parameter of Measurement Function | Uncertainty Component | type (A or B) | factor | effective degree of freedom | Relative standard uncertainty [%] | Relative standard uncertainty [%] | Relative standard uncertainty [%] | Relative standard uncertainty [%] | Relative standard uncertainty [%] | Relative standard uncertainty [%] | Relative standard uncertainty [%] |
| V (Accelerometer Output Voltage Measurement) | Voltage Measurement Repeatability | A | 1 | 87 | 0.0231% | 0.0126% | 0.0049% | 0.0111% | 0.0068% | 0.0039% | 0.0056% |
| | Voltage Traceability | B | 2 | ∞ | 0.03500% | 0.03500% | 0.03500% | 0.03500% | 0.03500% | 0.03500% | 0.03500% |
| | Voltage Measurement Resolution | B | √3 | ∞ | 0.0003% | 0.0003% | 0.0003% | 0.0003% | 0.0003% | 0.0003% | 0.0003% |
| | Effect of transverse, rocking and bending acceleration on accelerometer output voltage measurement (transverse sensitivity) | B | √3 | 200 | 0.0609% | 0.0389% | 0.0223% | 0.0308% | 0.0899% | 0.2920% | 0.2346% |
| | Distortion on Voltage Measurement | B | √3 | 200 | 0.0404% | 0.0260% | 0.0289% | 0.0289% | 0.0346% | 0.0387% | 0.0346% |
| | Effect of long-term data distribution | A | 1 | 24 | 0.1389% | 0.0567% | 0.0766% | 0.0523% | 0.0880% | 0.0918% | 0.1430% |
| | Effect of Charge Amplifier (with the fixed setting) | B | 2.57 | ∞ | 0.0973% | 0.0973% | 0.0973% | 0.0973% | 0.0973% | 0.0973% | 0.0973% |
| f (Vibration Frequency Measurement) | Frequency Measurement Repeatability | A | 1 | 87 | 0.0105345% | 0.0060821% | 0.0060821% | 0.0060821% | 0.0060821% | 0.0105345% | 0.0060821% |
| | Frequency Traceability | B | 2 | ∞ | 0.000025% | 0.000025% | 0.000025% | 0.000025% | 0.000025% | 0.000025% | 0.000025% |
| | Frequency Measurement Resolution | B | √3 | ∞ | 0.002887% | 0.002887% | 0.002887% | 0.002887% | 0.002887% | 0.002887% | 0.002887% |
| D (Vibration Displacement Measurement) | Amplitude measurement Errors from Sine Approximation Method, Evaluated by Least-Square Method | A | 1 | 59955 | 0.0305% | 0.0452% | 0.0617% | 0.1015% | 0.1553% | 0.2356% | 0.3201% |
| | Amplitude Measurement Repeatability | A | 1 | 87 | 0.0204% | 0.0443% | 0.0800% | 0.0168% | 0.0550% | 0.0399% | 0.1375% |
| | Amplitude Traceability (Laser Tube) | B | 1.96 | ∞ | 0.00000434% | 0.00000434% | 0.00000434% | 0.00000434% | 0.00000434% | 0.00000434% | 0.00000434% |
| | Amplitude Measurement Resolution | B | √3 | ∞ | 0.0141% | 0.0141% | 0.0141% | 0.0141% | 0.0141% | 0.0141% | 0.0141% |
| coverage factor, $k =$ (Ref. Calculation for CMS) | | | | | 1.96 | 1.96 | 1.96 | 1.96 | 1.96 | 1.96 | 1.96 |
| effective degrees of freedom, ν_{eff} | | | | | 91 | 885 | 423 | 2156 | 1160 | 8743 | 2011 |
| Relative combined (Standard) uncertainty [%] | | | | | 0.20% | 0.15% | 0.17% | 0.17% | 0.24% | 0.41% | 0.46% |
| Relative expanded uncertainty ($k=1.96$) [%] | | | | | 0.39% | 0.29% | 0.34% | 0.32% | 0.47% | 0.80% | 0.90% |

A2-3: NIMT

Table A2-3-1 Uncertainty budget for the BB transducer at 40 Hz to 100 Hz

| Acceleration amplitude | [m/s ²] | | | 10 | 25 | 25 | 50 | 50 |
|---|---------------------------|--------|-----------------------------|-----------------------------------|-----------------------------------|-----------------------------------|-----------------------------------|-----------------------------------|
| Vibration Frequency | [Hz] | | | 40 | 50 | 63 | 80 | 100 |
| Applied Calibration Method | | | | Fringe-counting method | Fringe-counting method | Fringe-counting method | Fringe-counting method | Fringe-counting method |
| Number of measurements | | | | 4 | 4 | 4 | 4 | 4 |
| Charge Sensitivity | [pC/(m/s ²)] | | | 0.12742 | 0.12749 | 0.12752 | 0.12757 | 0.12761 |
| Uncertainty Component | type (A or B) | factor | effective degree of freedom | Relative standard uncertainty [%] | Relative standard uncertainty [%] | Relative standard uncertainty [%] | Relative standard uncertainty [%] | Relative standard uncertainty [%] |
| Accuracy of ratio counter | B | 0.577 | ∞ | 0.00020 | 0.00025 | 0.00032 | 0.00020 | 0.00025 |
| Repeatability of ratio counter | A | 1 | 19 | 0.019 | 0.004 | 0.002 | 0.002 | 0.004 |
| Specification of Wavelength | B | 0.577 | ∞ | 0.000046 | 0.000046 | 0.000046 | 0.000046 | 0.000046 |
| Accuracy of Sine Generator | B | 0.577 | ∞ | 0.0023 | 0.0023 | 0.0023 | 0.0023 | 0.0023 |
| Repeatability of voltage output | A | 1 | 19 | 0.019 | 0.020 | 0.003 | 0.002 | 0.004 |
| Accuracy of voltmeter | B | 0.577 | ∞ | 0.1228 | 0.053 | 0.053 | 0.044 | 0.044 |
| Uncertainty of Charge Amplifier | A | 2 | ∞ | 0.255 | 0.185 | 0.185 | 0.185 | 0.185 |
| Drift of Charge Amplifier | B | 0.577 | ∞ | 0.058 | 0.058 | 0.058 | 0.058 | 0.058 |
| Effect of Geometric Location | B | 0.577 | ∞ | 0.058 | 0.058 | 0.058 | 0.058 | 0.058 |
| Effect of distortion | B | 0.577 | ∞ | 0.0018 | 0.0018 | 0.0018 | 0.0018 | 0.0018 |
| Effect of transverse | B | 0.236 | ∞ | 0.047 | 0.047 | 0.047 | 0.047 | 0.047 |
| Repeatability of Sensitivity | A | 1 | 3 | 0.015 | 0.020 | 0.013 | 0.019 | 0.016 |
| Resolution of Sensitivity | B | 0.577 | ∞ | 0.0023 | 0.0023 | 0.0023 | 0.0023 | 0.0023 |
| Effect of quantization error | B | 0.577 | ∞ | 0.00010 | 0.00013 | 0.00016 | 0.00010 | 0.00013 |
| Relative combined uncertainty [%] | | | | 0.300 | 0.216 | 0.215 | 0.213 | 0.213 |
| Relative expanded uncertainty (k=2) [%] | | | | 0.60 | 0.44 | 0.43 | 0.43 | 0.43 |

Table A2-3-2 Uncertainty budget for the BB transducer at 125 Hz to 315 Hz

| Acceleration amplitude | [m/s ²] | | | 50 | 50 | 50 | 50 | 50 |
|---------------------------------|--------------------------|--------|-----------------------------|-----------------------------------|-----------------------------------|-----------------------------------|-----------------------------------|-----------------------------------|
| Vibration Frequency | [Hz] | | | 125 | 160 | 200 | 250 | 315 |
| Applied Calibration Method | | | | Fringe-counting method | Fringe-counting method | Fringe-counting method | Fringe-counting method | Fringe-counting method |
| Number of measurements | | | | 4 | 4 | 4 | 4 | 4 |
| Charge Sensitivity | [pC/(m/s ²)] | | | 0.12762 | 0.12760 | 0.12763 | 0.12764 | 0.12764 |
| Uncertainty Component | type (A or B) | factor | effective degree of freedom | Relative standard uncertainty [%] | Relative standard uncertainty [%] | Relative standard uncertainty [%] | Relative standard uncertainty [%] | Relative standard uncertainty [%] |
| Accuracy of ratio counter | B | 0.577 | ∞ | 0.00064 | 0.00082 | 0.00102 | 0.00127 | 0.00160 |
| Repeatability of ratio counter | A | 1 | 19 | 0.004 | 0.008 | 0.003 | 0.006 | 0.008 |
| Specification of Wavelength | B | 0.577 | ∞ | 0.000046 | 0.000046 | 0.000046 | 0.000046 | 0.000046 |
| Accuracy of Sine Generator | B | 0.577 | ∞ | 0.0023 | 0.0023 | 0.0023 | 0.0023 | 0.0023 |
| Repeatability of voltage output | A | 1 | 19 | 0.004 | 0.009 | 0.002 | 0.008 | 0.005 |
| Accuracy of voltmeter | B | 0.577 | ∞ | 0.021 | 0.021 | 0.021 | 0.021 | 0.021 |
| Uncertainty of Charge Amplifier | A | 2 | ∞ | 0.160 | 0.160 | 0.160 | 0.160 | 0.160 |
| Drift of Charge Amplifier | B | 0.577 | ∞ | 0.058 | 0.058 | 0.058 | 0.058 | 0.058 |
| Effect of Geometric Location | B | 0.577 | ∞ | 0.058 | 0.058 | 0.058 | 0.058 | 0.058 |
| Effect of distortion | B | 0.577 | ∞ | 0.0018 | 0.0018 | 0.0018 | 0.0018 | 0.0018 |
| Effect of transverse | B | 0.236 | ∞ | 0.047 | 0.047 | 0.047 | 0.047 | 0.047 |
| Repeatability of Sensitivity | A | 1 | 3 | 0.016 | 0.017 | 0.010 | 0.007 | 0.011 |
| Resolution of Sensitivity | B | 0.577 | ∞ | 0.0023 | 0.0023 | 0.0023 | 0.0023 | 0.0023 |
| Effect of quantization error | B | 0.577 | ∞ | 0.00032 | 0.00041 | 0.00051 | 0.00064 | 0.00080 |

Table A2-3-3 Uncertainty budget for the BB transducer at 400 Hz to 800 Hz

| | | | | | | | |
|---|---------------------------|--------|-----------------------------|-----------------------------------|-----------------------------------|-----------------------------------|-----------------------------------|
| Acceleration amplitude | [m/s ²] | | | 50 | 100 | 100 | 100 |
| Vibration Frequency | [Hz] | | | 400 | 500 | 630 | 800 |
| Applied Calibration Method | | | | Fringe-counting method | Fringe-counting method | Fringe-counting method | Fringe-counting method |
| Number of measurements | | | | 4 | 4 | 4 | 4 |
| Charge Sensitivity | [pC/(m/s ²)] | | | 0.12763 | 0.12759 | 0.12761 | 0.12755 |
| Uncertainty Component | type (A or B) | factor | effective degree of freedom | Relative standard uncertainty [%] | Relative standard uncertainty [%] | Relative standard uncertainty [%] | Relative standard uncertainty [%] |
| Accuracy of ratio counter | B | 0.577 | ∞ | 0.00204 | 0.00637 | 0.00803 | 0.01019 |
| Repeatability of ratio counter | A | 1 | 19 | 0.011 | 0.017 | 0.019 | 0.035 |
| Specification of Wavelength | B | 0.577 | ∞ | 0.000046 | 0.000046 | 0.000046 | 0.000046 |
| Accuracy of Sine Generator | B | 0.577 | ∞ | 0.0023 | 0.0023 | 0.0023 | 0.0023 |
| Repeatability of voltage output | A | 1 | 19 | 0.001 | 0.008 | 0.005 | 0.006 |
| Accuracy of voltmeter | B | 0.577 | ∞ | 0.021 | 0.057 | 0.057 | 0.057 |
| Uncertainty of Charge Amplifier | A | 2 | ∞ | 0.160 | 0.160 | 0.160 | 0.160 |
| Drift of Charge Amplifier | B | 0.577 | ∞ | 0.058 | 0.058 | 0.058 | 0.058 |
| Effect of Geometric Location | B | 0.577 | ∞ | 0.058 | 0.058 | 0.058 | 0.058 |
| Effect of distortion | B | 0.577 | ∞ | 0.0018 | 0.0018 | 0.0018 | 0.0018 |
| Effect of transverse | B | 0.236 | ∞ | 0.047 | 0.047 | 0.047 | 0.047 |
| Repeatability of Sensitivity | A | 1 | 3 | 0.012 | 0.012 | 0.016 | 0.014 |
| Resolution of Sensitivity | B | 0.577 | ∞ | 0.0023 | 0.0023 | 0.0023 | 0.0023 |
| Effect of quantization error | B | 0.577 | ∞ | 0.00102 | 0.00319 | 0.00401 | 0.00510 |
| Relative combined uncertainty [%] | | | | 0.188 | 0.196 | 0.196 | 0.198 |
| Relative expanded uncertainty (k=2) [%] | | | | 0.38 | 0.40 | 0.40 | 0.40 |

Table A2-3-4 Uncertainty budget for the BB transducer at 1 kHz to 2.5 kHz

| Acceleration amplitude | [m/s ²] | | | 50 | 50 | 59 | 57 | 62 |
|---|--------------------------|--------|-----------------------------|-----------------------------------|-----------------------------------|-----------------------------------|-----------------------------------|-----------------------------------|
| Vibration Frequency | [Hz] | | | 1000 | 1250 | 1600 | 2000 | 2500 |
| Applied Calibration Method | | | | minimum point method | minimum point method | minimum point method | minimum point method | minimum point method |
| Number of measurements | | | | 4 | 4 | 4 | 4 | 4 |
| Charge Sensitivity | [pC/(m/s ²)] | | | 0.12764 | 0.12768 | 0.12777 | 0.12801 | 0.12814 |
| Uncertainty Component | type (A or B) | factor | effective degree of freedom | Relative standard uncertainty [%] | Relative standard uncertainty [%] | Relative standard uncertainty [%] | Relative standard uncertainty [%] | Relative standard uncertainty [%] |
| Specification of Wavelength | B | 0.577 | ∞ | 0.00046 | 0.00046 | 0.00046 | 0.00046 | 0.00046 |
| Accuracy of Sine Generator | B | 0.577 | ∞ | 0.0023 | 0.0023 | 0.0023 | 0.0023 | 0.0023 |
| Repeatability of voltage output | A | 1 | 4 | 0.0043 | 0.0039 | 0.0043 | 0.0066 | 0.0034 |
| Accuracy of voltmeter | B | 0.577 | ∞ | 0.021 | 0.021 | 0.019 | 0.019 | 0.019 |
| Uncertainty of Charge Amplifier | A | 2 | ∞ | 0.160 | 0.160 | 0.160 | 0.160 | 0.160 |
| Drift of Charge Amplifier | B | 0.577 | ∞ | 0.058 | 0.058 | 0.058 | 0.058 | 0.058 |
| Effect of Geometric Location | B | 0.577 | ∞ | 0.058 | 0.058 | 0.058 | 0.058 | 0.058 |
| Effect of distortion | B | 0.577 | ∞ | 0.0018 | 0.0018 | 0.0018 | 0.0018 | 0.0018 |
| Effect of transverse | B | 0.236 | ∞ | 0.047 | 0.048 | 0.048 | 0.048 | 0.049 |
| Repeatability of Sensitivity | A | 1 | 3 | 0.013 | 0.014 | 0.013 | 0.012 | 0.018 |
| Resolution of FFT Analyzer | B | 0.577 | ∞ | 0.33 | 0.33 | 0.33 | 0.33 | 0.33 |
| Resolution of Sensitivity report | B | 0.577 | ∞ | 0.0023 | 0.0023 | 0.0023 | 0.0023 | 0.0023 |
| Relative combined uncertainty [%] | | | | 0.380 | 0.380 | 0.380 | 0.380 | 0.380 |
| Relative expanded uncertainty (k=2) [%] | | | | 0.76 | 0.76 | 0.76 | 0.76 | 0.76 |

Table A2-3-5 Uncertainty budget for the BB transducer at 3.15 kHz to 5 kHz

| Acceleration amplitude | [m/s ²] | | | 53 | 86 | 135 |
|---|--------------------------|--------|-----------------------------|-----------------------------------|-----------------------------------|-----------------------------------|
| Vibration Frequency | [Hz] | | | 3150 | 4000 | 5000 |
| Applied Calibration Method | | | | minimum point method | minimum point method | minimum point method |
| Number of measurements | | | | 4 | 4 | 4 |
| Charge Sensitivity | [pC/(m/s ²)] | | | 0.12851 | 0.12903 | 0.12998 |
| Uncertainty Component | type (A or B) | factor | effective degree of freedom | Relative standard uncertainty [%] | Relative standard uncertainty [%] | Relative standard uncertainty [%] |
| Specification of Wavelength | B | 0.577 | ∞ | 0.00046 | 0.00046 | 0.00046 |
| Accuracy of Sine Generator | B | 0.577 | ∞ | 0.0023 | 0.0023 | 0.0023 |
| Repeatability of voltage output | A | 1 | 4 | 0.0055 | 0.0050 | 0.0056 |
| Accuracy of voltmeter | B | 0.577 | ∞ | 0.020 | 0.017 | 0.044 |
| Uncertainty of Charge Amplifier | A | 2 | ∞ | 0.160 | 0.160 | 0.160 |
| Drift of Charge Amplifier | B | 0.577 | ∞ | 0.058 | 0.058 | 0.058 |
| Effect of Geometric Location | B | 0.577 | ∞ | 0.058 | 0.058 | 0.058 |
| Effect of distortion | B | 0.577 | ∞ | 0.0018 | 0.0018 | 0.0018 |
| Effect of transverse | B | 0.236 | ∞ | 0.050 | 0.052 | 0.056 |
| Repeatability of Sensitivity | A | 1 | 3 | 0.022 | 0.026 | 0.004 |
| Resolution of FFT Analyzer | B | 0.577 | ∞ | 0.33 | 0.33 | 0.33 |
| Resolution of Sensitivity report | B | 0.577 | ∞ | 0.0022 | 0.0022 | 0.0022 |
| Relative combined uncertainty [%] | | | | 0.380 | 0.381 | 0.382 |
| Relative expanded uncertainty (k=2) [%] | | | | 0.77 | 0.77 | 0.77 |

Table A2-3-6 Uncertainty budget for the SE transducer at 40 Hz to 100 Hz

| Acceleration amplitude | [m/s ²] | | | 25 | 25 | 25 | 50 | 50 |
|---|--------------------------|--------|-----------------------------|-----------------------------------|-----------------------------------|-----------------------------------|-----------------------------------|-----------------------------------|
| Vibration Frequency | [Hz] | | | 40 | 50 | 63 | 80 | 100 |
| Applied Calibration Method | | | | Fringe-counting method | Fringe-counting method | Fringe-counting method | Fringe-counting method | Fringe-counting method |
| Number of measurements | | | | 5 | 5 | 5 | 5 | 5 |
| Charge Sensitivity | [pC/(m/s ²)] | | | 0.12866 | 0.12875 | 0.12878 | 0.12884 | 0.12888 |
| Uncertainty Component | type (A or B) | factor | effective degree of freedom | Relative standard uncertainty [%] | Relative standard uncertainty [%] | Relative standard uncertainty [%] | Relative standard uncertainty [%] | Relative standard uncertainty [%] |
| Accuracy of ratio counter | B | 0.577 | ∞ | 0.00020 | 0.00025 | 0.00032 | 0.00020 | 0.00025 |
| Repeatability of ratio counter | A | 1 | 19 | 0.029 | 0.0078 | 0.0032 | 0.0020 | 0.0071 |
| Specification of Wavelength | B | 0.577 | ∞ | 0.000046 | 0.000046 | 0.000046 | 0.000046 | 0.000046 |
| Accuracy of Sine Generator | B | 0.577 | ∞ | 0.0023 | 0.0023 | 0.0023 | 0.0023 | 0.0023 |
| Repeatability of voltage output | A | 1 | 19 | 0.030 | 0.019 | 0.0034 | 0.0023 | 0.0074 |
| Accuracy of voltmeter | B | 0.577 | ∞ | 0.122 | 0.053 | 0.053 | 0.044 | 0.044 |
| Uncertainty of Charge Amplifier | A | 2 | ∞ | 0.255 | 0.185 | 0.185 | 0.185 | 0.185 |
| Drift of Charge Amplifier | B | 0.577 | ∞ | 0.058 | 0.058 | 0.058 | 0.058 | 0.058 |
| Effect of Geometric Location | B | 0.577 | ∞ | 0.058 | 0.058 | 0.058 | 0.058 | 0.058 |
| Effect of distortion | B | 0.577 | ∞ | 0.0018 | 0.0018 | 0.0018 | 0.0018 | 0.0018 |
| Effect of transverse | B | 0.236 | ∞ | 0.047 | 0.047 | 0.047 | 0.047 | 0.047 |
| Repeatability of Sensitivity | A | 1 | 4 | 0.0057 | 0.0092 | 0.004 | 0.0042 | 0.0046 |
| Resolution of Sensitivity | B | 0.577 | ∞ | 0.0022 | 0.0022 | 0.0022 | 0.0022 | 0.0022 |
| Effect of quantization error | B | 0.577 | ∞ | 0.00010 | 0.00013 | 0.00016 | 0.00010 | 0.00013 |
| Relative combined uncertainty [%] | | | | 0.30 | 0.22 | 0.21 | 0.21 | 0.21 |
| Relative expanded uncertainty (k=2) [%] | | | | 0.61 | 0.44 | 0.43 | 0.43 | 0.43 |

Table A2-3-7 Uncertainty budget for the SE transducer at 125 Hz to 315 Hz

| Acceleration amplitude | [m/s ²] | | | 50 | 50 | 50 | 50 | 50 |
|---|--------------------------|--------|-----------------------------|-----------------------------------|-----------------------------------|-----------------------------------|-----------------------------------|-----------------------------------|
| Vibration Frequency | [Hz] | | | 125 | 160 | 200 | 250 | 315 |
| Applied Calibration Method | | | | Fringe-counting method | Fringe-counting method | Fringe-counting method | Fringe-counting method | Fringe-counting method |
| Number of measurements | | | | 5 | 5 | 5 | 5 | 5 |
| Charge Sensitivity | [pC/(m/s ²)] | | | 0.12893 | 0.12893 | 0.12892 | 0.12888 | 0.12893 |
| Uncertainty Component | type (A or B) | factor | effective degree of freedom | Relative standard uncertainty [%] | Relative standard uncertainty [%] | Relative standard uncertainty [%] | Relative standard uncertainty [%] | Relative standard uncertainty [%] |
| Accuracy of ratio counter | B | 0.577 | ∞ | 0.00064 | 0.00082 | 0.0010 | 0.0013 | 0.0016 |
| Repeatability of ratio counter | A | 1 | 19 | 0.0087 | 0.0045 | 0.011 | 0.0071 | 0.0094 |
| Specification of Wavelength | B | 0.577 | ∞ | 0.000046 | 0.000046 | 0.000046 | 0.000046 | 0.000046 |
| Accuracy of Sine Generator | B | 0.577 | ∞ | 0.0023 | 0.0023 | 0.0023 | 0.0023 | 0.0023 |
| Repeatability of voltage output | A | 1 | 19 | 0.0085 | 0.0041 | 0.010 | 0.0080 | 0.0025 |
| Accuracy of voltmeter | B | 0.577 | ∞ | 0.020 | 0.020 | 0.020 | 0.020 | 0.020 |
| Uncertainty of Charge Amplifier | A | 2 | ∞ | 0.160 | 0.160 | 0.160 | 0.160 | 0.160 |
| Drift of Charge Amplifier | B | 0.577 | ∞ | 0.058 | 0.058 | 0.058 | 0.058 | 0.058 |
| Effect of Geometric Location | B | 0.577 | ∞ | 0.058 | 0.058 | 0.058 | 0.058 | 0.058 |
| Effect of distortion | B | 0.577 | ∞ | 0.0018 | 0.0018 | 0.0018 | 0.0018 | 0.0018 |
| Effect of transverse | B | 0.236 | ∞ | 0.047 | 0.047 | 0.047 | 0.047 | 0.047 |
| Repeatability of Sensitivity | A | 1 | 4 | 0.0055 | 0.0042 | 0.0042 | 0.006 | 0.013 |
| Resolution of Sensitivity | B | 0.577 | ∞ | 0.0022 | 0.0022 | 0.0022 | 0.0022 | 0.0022 |
| Effect of quantization error | B | 0.577 | ∞ | 0.00032 | 0.00041 | 0.00051 | 0.00064 | 0.00080 |
| Relative combined uncertainty [%] | | | | 0.19 | 0.19 | 0.19 | 0.19 | 0.19 |
| Relative expanded uncertainty (k=2) [%] | | | | 0.38 | 0.38 | 0.38 | 0.38 | 0.38 |

Table A2-3-8 Uncertainty budget for the SE transducer at 400 Hz to 800 Hz

| Acceleration amplitude | [m/s ²] | | | 50 | 100 | 100 | 100 |
|---|--------------------------|--------|-----------------------------|-----------------------------------|-----------------------------------|-----------------------------------|-----------------------------------|
| Vibration Frequency | [Hz] | | | 400 | 500 | 630 | 800 |
| Applied Calibration Method | | | | Fringe-counting method | Fringe-counting method | Fringe-counting method | Fringe-counting method |
| Number of measurements | | | | 5 | 5 | 5 | 5 |
| Charge Sensitivity | [pC/(m/s ²)] | | | 0.12895 | 0.12890 | 0.12898 | 0.12893 |
| Uncertainty Component | type (A or B) | factor | effective degree of freedom | Relative standard uncertainty [%] | Relative standard uncertainty [%] | Relative standard uncertainty [%] | Relative standard uncertainty [%] |
| Accuracy of ratio counter | B | 0.577 | ∞ | 0.0020 | 0.0064 | 0.0080 | 0.010 |
| Repeatability of ratio counter | A | 1 | 19 | 0.012 | 0.014 | 0.017 | 0.044 |
| Specification of Wavelength | B | 0.577 | ∞ | 0.000046 | 0.000046 | 0.000046 | 0.000046 |
| Accuracy of Sine Generator | B | 0.577 | ∞ | 0.0023 | 0.0023 | 0.0023 | 0.0023 |
| Repeatability of voltage output | A | 1 | 19 | 0.0043 | 0.0086 | 0.0059 | 0.0066 |
| Accuracy of voltmeter | B | 0.577 | ∞ | 0.020 | 0.056 | 0.056 | 0.056 |
| Uncertainty of Charge Amplifier | A | 2 | ∞ | 0.160 | 0.160 | 0.160 | 0.160 |
| Drift of Charge Amplifier | B | 0.577 | ∞ | 0.058 | 0.058 | 0.058 | 0.058 |
| Effect of Geometric Location | B | 0.577 | ∞ | 0.058 | 0.058 | 0.058 | 0.058 |
| Effect of distortion | B | 0.577 | ∞ | 0.0018 | 0.0018 | 0.0018 | 0.0018 |
| Effect of transverse | B | 0.236 | ∞ | 0.047 | 0.047 | 0.047 | 0.047 |
| Repeatability of Sensitivity | A | 1 | 4 | 0.012 | 0.016 | 0.019 | 0.016 |
| Resolution of Sensitivity | B | 0.577 | ∞ | 0.0022 | 0.0022 | 0.0022 | 0.0022 |
| Effect of quantization error | B | 0.577 | ∞ | 0.0010 | 0.0032 | 0.0040 | 0.0051 |
| Relative combined uncertainty [%] | | | | 0.19 | 0.20 | 0.20 | 0.20 |
| Relative expanded uncertainty ($k=2$) [%] | | | | 0.38 | 0.40 | 0.40 | 0.40 |

Table A2-3-9 Uncertainty budget for the SE transducer at 1 kHz to 2.5 kHz

| Acceleration amplitude | [m/s ²] | | | 50 | 50 | 59 | 57 | 62 |
|---|--------------------------|--------|-----------------------------|-----------------------------------|-----------------------------------|-----------------------------------|-----------------------------------|-----------------------------------|
| Vibration Frequency | [Hz] | | | 1000 | 1250 | 1600 | 2000 | 2500 |
| Applied Calibration Method | | | | minimum point method | minimum point method | minimum point method | minimum point method | minimum point method |
| Number of measurements | | | | 5 | 5 | 5 | 5 | 5 |
| Charge Sensitivity | [pC/(m/s ²)] | | | 0.12903 | 0.12902 | 0.12927 | 0.12938 | 0.12972 |
| Uncertainty Component | type (A or B) | factor | effective degree of freedom | Relative standard uncertainty [%] | Relative standard uncertainty [%] | Relative standard uncertainty [%] | Relative standard uncertainty [%] | Relative standard uncertainty [%] |
| Specification of Wavelength | B | 0.577 | ∞ | 0.000046 | 0.000046 | 0.000046 | 0.000046 | 0.000046 |
| Accuracy of Sine Generator | B | 0.577 | ∞ | 0.0023 | 0.0023 | 0.0023 | 0.0023 | 0.0023 |
| Repeatability of voltage output | A | 1 | 4 | 0.010 | 0.0034 | 0.0062 | 0.016 | 0.0068 |
| Accuracy of voltmeter | B | 0.577 | ∞ | 0.021 | 0.020 | 0.019 | 0.019 | 0.019 |
| Uncertainty of Charge Amplifier | A | 2.000 | ∞ | 0.160 | 0.160 | 0.160 | 0.160 | 0.160 |
| Drift of Charge Amplifier | B | 0.577 | ∞ | 0.058 | 0.058 | 0.058 | 0.058 | 0.058 |
| Effect of Geometric location | B | 0.577 | ∞ | 0.058 | 0.058 | 0.058 | 0.058 | 0.058 |
| Effect of distortion | B | 0.577 | ∞ | 0.0018 | 0.0018 | 0.0018 | 0.0018 | 0.0018 |
| Effect of transverse | B | 0.236 | ∞ | 0.047 | 0.048 | 0.048 | 0.048 | 0.049 |
| Repeatability of Sensitivity | A | 1 | 4 | 0.017 | 0.022 | 0.018 | 0.031 | 0.024 |
| Resolution of FFT Analyzer | B | 0.577 | ∞ | 0.33 | 0.33 | 0.33 | 0.33 | 0.33 |
| Resolution of Sensitivity report | B | 0.577 | ∞ | 0.0022 | 0.0022 | 0.0022 | 0.0022 | 0.0022 |
| Relative combined uncertainty [%] | | | | 0.38 | 0.38 | 0.38 | 0.38 | 0.38 |
| Relative expanded uncertainty (k=2) [%] | | | | 0.76 | 0.76 | 0.76 | 0.77 | 0.77 |

Table A2-3-10 Uncertainty budget for the SE transducer at 3.15 kHz to 5 kHz

| Acceleration amplitude | [m/s ²] | | | 53 | 86 | 135 |
|---|--------------------------|--------|-----------------------------|-----------------------------------|-----------------------------------|-----------------------------------|
| Vibration Frequency | [Hz] | | | 3150 | 4000 | 5000 |
| Applied Calibration Method | | | | minimum point method | minimum point method | minimum point method |
| Number of measurements | | | | 5 | 5 | 5 |
| Charge Sensitivity | [pC/(m/s ²)] | | | 0.13021 | 0.13100 | 0.13235 |
| Uncertainty Component | type (A or B) | factor | effective degree of freedom | Relative standard uncertainty [%] | Relative standard uncertainty [%] | Relative standard uncertainty [%] |
| Specification of Wavelength | B | 0.577 | ∞ | 0.000046 | 0.000046 | 0.000046 |
| Accuracy of Sine Generator | B | 0.577 | ∞ | 0.0023 | 0.0023 | 0.0023 |
| Repeatability of voltage output | A | 1 | 4 | 0.0058 | 0.0085 | 0.021 |
| Accuracy of voltmeter | B | 0.577 | ∞ | 0.020 | 0.017 | 0.043 |
| Uncertainty of Charge Amplifier | A | 2.000 | ∞ | 0.160 | 0.160 | 0.160 |
| Drift of Charge Amplifier | B | 0.577 | ∞ | 0.058 | 0.058 | 0.058 |
| Effect of Geometric location | B | 0.577 | ∞ | 0.058 | 0.058 | 0.058 |
| Effect of distortion | B | 0.577 | ∞ | 0.0018 | 0.0018 | 0.0018 |
| Effect of transverse | B | 0.236 | ∞ | 0.050 | 0.052 | 0.056 |
| Repeatability of Sensitivity | A | 1 | 4 | 0.025 | 0.008 | 0.020 |
| Resolution of FFT Analyzer | B | 0.577 | ∞ | 0.33 | 0.33 | 0.33 |
| Resolution of Sensitivity report | B | 0.577 | ∞ | 0.0022 | 0.0022 | 0.0022 |
| Relative combined uncertainty [%] | | | | 0.38 | 0.38 | 0.38 |
| Relative expanded uncertainty (k=2) [%] | | | | 0.77 | 0.76 | 0.77 |

A2-4: A@Star

Table A2-4-1 Uncertainty budget for the BB transducer

| Uncertainty Calculation: | Probability distribution model | 20 Hz | 40 Hz | 80 Hz | 160 Hz | 315 Hz | 630 Hz | 1250 Hz | 2500 Hz | 5000 Hz |
|---|--------------------------------|---------|---------|---------|---------|---------|---------|---------|---------|---------|
| <u>Output voltage measurement</u> | | | | | | | | | | |
| Output voltage | Normal | 0.062 | 0.062 | 0.062 | 0.062 | 0.062 | 0.062 | 0.062 | 0.062 | 0.062 |
| Filtering effect | Rectangular | 0.006 | 0.006 | 0.006 | 0.006 | 0.006 | 0.006 | 0.006 | 0.006 | 0.006 |
| Hum & noise | Rectangular | 0.006 | 0.006 | 0.006 | 0.006 | 0.006 | 0.006 | 0.006 | 0.006 | 0.006 |
| Transverse sensitivity | Special | 0.024 | 0.024 | 0.071 | 0.071 | 0.071 | 0.071 | 0.071 | 0.118 | 0.236 |
| <u>Displacement determination</u> | | | | | | | | | | |
| Phase amplitude measurement | Rectangular | 0.029 | 0.029 | 0.029 | 0.029 | 0.029 | 0.029 | 0.029 | 0.058 | 0.058 |
| Motion disturbance | Rectangular | 0.024 | 0.071 | 0.071 | 0.071 | 0.071 | 0.071 | 0.071 | 0.118 | 0.118 |
| <u>Vibration Frequency</u> | | | | | | | | | | |
| Frequency generator & indicator | Rectangular | 0.003 | 0.003 | 0.003 | 0.003 | 0.003 | 0.003 | 0.003 | 0.003 | 0.003 |
| <u>Reference conditioning amplifier</u> | | | | | | | | | | |
| Charge sensitivity | Normal | 0.200 | 0.200 | 0.200 | 0.200 | 0.200 | 0.200 | 0.200 | 0.200 | 0.200 |
| <u>Residual effect</u> | | | | | | | | | | |
| Gravity | Rectangular | 0.006 | 0.006 | 0.006 | 0.006 | 0.006 | 0.006 | 0.006 | 0.006 | 0.006 |
| Non-linearity | Rectangular | 0.017 | 0.017 | 0.017 | 0.017 | 0.017 | 0.017 | 0.017 | 0.017 | 0.017 |
| Exciter magnetic field | Rectangular | 0.017 | 0.017 | 0.017 | 0.017 | 0.017 | 0.017 | 0.017 | 0.017 | 0.017 |
| Temperature variation | Rectangular | 0.035 | 0.035 | 0.035 | 0.035 | 0.035 | 0.035 | 0.035 | 0.035 | 0.035 |
| Other environmental effects | Rectangular | 0.017 | 0.017 | 0.017 | 0.017 | 0.017 | 0.017 | 0.017 | 0.017 | 0.017 |
| <u>Type B total</u> | | 0.22 | 0.23 | 0.24 | 0.24 | 0.24 | 0.24 | 0.24 | 0.28 | 0.34 |
| <u>Type A</u> | | | | | | | | | | |
| Number of measurement | | 40 | 40 | 40 | 40 | 40 | 40 | 40 | 40 | 40 |
| Standard uncertainty | | 0.03 | 0.02 | 0.03 | 0.07 | 0.02 | 0.04 | 0.04 | 0.06 | 1.47 |
| <u>Combined uncertainty</u> | | 0.22 | 0.23 | 0.24 | 0.25 | 0.24 | 0.24 | 0.24 | 0.28 | 1.51 |
| Effective degree of freedom | | 7.8E+04 | 7.1E+05 | 2.0E+05 | 8.0E+03 | 4.6E+05 | 3.8E+04 | 8.7E+04 | 2.4E+04 | 4.3E+01 |
| Coverage factor | | 2.00 | 2.00 | 2.00 | 2.00 | 2.00 | 2.00 | 2.00 | 2.00 | 2.06 |
| <u>Expanded uncertainty [%]</u> | | 0.44 | 0.46 | 0.48 | 0.49 | 0.48 | 0.48 | 0.48 | 0.57 | 3.10 |

Table A2-4-2 Uncertainty budget for the SE transducer

| Uncertainty components | Probability distribution model | 20 Hz | 40 Hz | 80 Hz | 160 Hz | 315 Hz | 630 Hz | 1250 Hz | 2500 Hz | 5000 Hz |
|---|--------------------------------|-----------------------------------|---------|---------|---------|---------|---------|---------|---------|---------|
| | | Output voltage measurement | | | | | | | | |
| Output voltage | Normal | 0.062 | 0.062 | 0.062 | 0.062 | 0.062 | 0.062 | 0.062 | 0.062 | 0.062 |
| Filtering effect | Rectangular | 0.006 | 0.006 | 0.006 | 0.006 | 0.006 | 0.006 | 0.006 | 0.006 | 0.006 |
| Hum & noise | Rectangular | 0.006 | 0.006 | 0.006 | 0.006 | 0.006 | 0.006 | 0.006 | 0.006 | 0.006 |
| Transverse sensitivity | Special | 0.024 | 0.024 | 0.071 | 0.071 | 0.071 | 0.071 | 0.071 | 0.118 | 0.236 |
| Displacement determination | | | | | | | | | | |
| Phase amplitude measurement | Rectangular | 0.029 | 0.029 | 0.029 | 0.029 | 0.029 | 0.029 | 0.029 | 0.058 | 0.058 |
| Motion disturbance | Rectangular | 0.024 | 0.071 | 0.071 | 0.071 | 0.071 | 0.071 | 0.071 | 0.118 | 0.118 |
| Vibration Frequency measurement | | | | | | | | | | |
| Frequency generator & indicator | Rectangular | 0.003 | 0.003 | 0.003 | 0.003 | 0.003 | 0.003 | 0.003 | 0.003 | 0.003 |
| Reference conditioning amplifier | | | | | | | | | | |
| Charge sensitivity | Normal | 0.200 | 0.200 | 0.200 | 0.200 | 0.200 | 0.200 | 0.200 | 0.200 | 0.200 |
| Residual effect | | | | | | | | | | |
| Gravity | Rectangular | 0.006 | 0.006 | 0.006 | 0.006 | 0.006 | 0.006 | 0.006 | 0.006 | 0.006 |
| Non-linearity | Rectangular | 0.017 | 0.017 | 0.017 | 0.017 | 0.017 | 0.017 | 0.017 | 0.017 | 0.017 |
| Exciter magnetic field | Rectangular | 0.017 | 0.017 | 0.017 | 0.017 | 0.017 | 0.017 | 0.017 | 0.017 | 0.017 |
| Temperature variation | Rectangular | 0.035 | 0.035 | 0.035 | 0.035 | 0.035 | 0.035 | 0.035 | 0.035 | 0.035 |
| Other environmental effects | Rectangular | 0.017 | 0.017 | 0.017 | 0.017 | 0.017 | 0.017 | 0.017 | 0.017 | 0.017 |
| Type B Total | | 0.22 | 0.23 | 0.24 | 0.24 | 0.24 | 0.24 | 0.24 | 0.28 | 0.34 |
| Type A | | | | | | | | | | |
| Number of measurement | | 60 | 60 | 60 | 60 | 60 | 60 | 60 | 60 | 60 |
| Standard uncertainty | | 0.02 | 0.02 | 0.04 | 0.22 | 0.05 | 0.10 | 0.06 | 0.04 | 0.44 |
| Combined uncertainty | | 0.22 | 0.23 | 0.24 | 0.33 | 0.24 | 0.26 | 0.25 | 0.28 | 0.56 |
| Effective degree of freedom | | 6.0E+05 | 5.2E+05 | 1.2E+05 | 2.8E+02 | 3.6E+04 | 2.5E+03 | 1.9E+04 | 1.0E+05 | 1.5E+02 |
| Coverage factor | | 2.00 | 2.00 | 2.00 | 2.01 | 2.00 | 2.00 | 2.00 | 2.00 | 2.02 |
| Expanded uncertainty [%] | | 0.44 | 0.46 | 0.48 | 0.65 | 0.49 | 0.52 | 0.49 | 0.56 | 1.13 |

Annex 3 DoE to the KCRV of CCAUV.V-K2 for the BB transducer

| Frequency in Hz | CMS/ITRI | | NIMT | | A*STAR | |
|--------------------|--|----------|--|----------|--|----------|
| | d_i | $U(d_i)$ | d_i | $U(d_i)$ | d_i | $U(d_i)$ |
| | in $\text{pC}/(\text{m/s}^2)\cdot 10^{-4}$ | | in $\text{pC}/(\text{m/s}^2)\cdot 10^{-4}$ | | in $\text{pC}/(\text{m/s}^2)\cdot 10^{-4}$ | |
| 40 | 1.4 | 4.4 | -3.8 | 7.9 | -4.0 | 6.2 |
| 50 | | | | | | |
| 63 | 0.9 | 3.7 | -3.3 | 5.8 | | |
| 80 | 2.5 | 3.6 | -1.0 | 5.8 | -1.4 | 6.4 |
| 100 | 2.8 | 3.6 | -0.3 | 5.5 | | |
| 125 | 1.3 | 3.7 | -1.2 | 5.2 | | |
| 160 | 1.6 | 3.9 | -1.7 | 5.2 | -1.7 | 6.5 |
| 200 | 1.7 | 4.0 | -1.0 | 5.2 | | |
| 250 | 0.2 | 4.0 | -2.0 | 5.2 | | |
| 315 | 2.6 | 4.4 | -0.3 | 5.2 | -2.0 | 6.4 |
| 400 | 2.9 | 4.2 | 0.1 | 5.2 | | |
| 500 | 3.6 | 4.6 | 0.0 | 5.5 | | |
| 630 | 2.0 | 4.3 | -1.4 | 5.5 | -1.2 | 6.4 |
| 800 | 1.5 | 4.9 | -1.9 | 5.5 | | |
| 1000 | 1.2 | 5.1 | -1.3 | 9.9 | | |
| 1250 | 2.0 | 5.4 | 0.3 | 9.9 | -1.0 | 6.4 |
| 1600 | -0.6 | 4.0 | -1.8 | 9.9 | | |
| 2000 | 2.2 | 4.6 | 3.5 | 10.0 | | |
| 2500 | -2.0 | 4.7 | 0.4 | 10.0 | -2.5 | 7.6 |
| 3150 | -2.9 | 6.2 | 0.4 | 10.1 | | |
| 4000 | 6.8 | 10.6 | 4.9 | 10.2 | | |
| 5000 | -1.7 | 11.6 | 3.4 | 10.3 | 16.6 | 41.3 |