

---

# **Final Report on Key Comparison**

## **APMP.AUV.V-K3**

*This comparison was initially registered as APMP.AUV.V-S1, but was redefined as  
APMP.AUV.V-K3 in November 2013*

*Qiao SUN and Ian Veldman*

*2012-12-12*

National Institute of Metrology, P. R. China  
and  
National Metrology Institute of South Africa

---

## Table of Contents

<b>1</b>	<b>INTRODUCTION.....</b>	<b>1</b>
<b>2</b>	<b>PARTICIPANTS .....</b>	<b>1</b>
<b>3</b>	<b>TASK AND PURPOSE OF THE COMPARISON.....</b>	<b>1</b>
<b>4</b>	<b>CONDITIONS AND INSTRUCTIONS OF MEASUREMENT .....</b>	<b>2</b>
<b>5</b>	<b>TRANSFER STANDARDS AS ARTEFACTS.....</b>	<b>3</b>
<b>6</b>	<b>CIRCULATION OF THE ARTEFACTS .....</b>	<b>3</b>
<b>7</b>	<b>RESULTS OF THE MEASUREMENTS .....</b>	<b>4</b>
7.1	MONITORING OF STABILITY .....	4
7.2	RESULTS OF THE PARTICIPANTS.....	6
7.2.1	<i>Results of horizontal voltage sensitivities of the SE-1021 .....</i>	7
7.2.2	<i>Results of vertical voltage sensitivities of the SE-1021.....</i>	9
7.2.3	<i>Results of horizontal voltage sensitivities of the SE-1022 .....</i>	10
7.2.4	<i>Results of vertical voltage sensitivities of the SE-1022.....</i>	12
7.3	KEY COMPARISON REFERENCE VALUE.....	12
7.4	DEGREES OF EQUIVALENCE BETWEEN PARTICIPANTS .....	17
7.4.1	<i>Tables of DoE between participants for horizontal voltage sensitivities of the SE-1021.....</i>	18
7.4.2	<i>Tables of DoE between participants for vertical voltage sensitivities of the SE-1021.....</i>	26
7.4.3	<i>Tables of DoE between participants for horizontal voltage sensitivities of the SE-1022.....</i>	32
7.4.4	<i>Tables of DoE between participants for vertical voltage sensitivities of the SE-1022.....</i>	40
<b>8</b>	<b>CONCLUSION .....</b>	<b>45</b>
<b>9</b>	<b>BIBLIOGRAPHY .....</b>	<b>46</b>
	<b>APPENDIX A: TECHNICAL PROTOCOL .....</b>	<b>47</b>
	<b>APPENDIX B: UNCERTAINTY BUDGETS OF THE PARTICIPANTS .....</b>	<b>53</b>
	<b>APPENDIX C: DEGREES OF EQUIVALENCE RELATIVE TO THE KCRV .....</b>	<b>74</b>

---

## 1 Introduction

This report presents the results of the first Asia Pacific Metrology Programme (APMP) comparison in the area of low-frequency ‘vibration’, which in this case means sinusoidal acceleration.

The participants have reached consensus and considered the weighted mean as the most appropriate method for this particular comparison to compute the key comparison reference values (KCRVs) and the degrees of equivalence. Detailed analysis and application of the method for use in comparisons in the field of vibration, is documented in the CCAUV.V-K1 report [1]. The calculation of the KCRVs is also in accordance with the Guidelines for CIPM key comparisons [2].

The Technical Protocol of December 2010 [3] specifies in detail the aim and the task of the comparison, the conditions of measurement, the transfer standards used, measurement instructions, time schedule and other items. A brief survey is given in the following sections.

## 2 Participants

Seven metrology institutes from APMP economies have participated in the comparison APMP.AUV.V-K3 (see table 1).

Table 1: List of participating institutes

Participant	Economy	Calibration period
NIM	P. R. China	August, 2011
NIMT	Thailand	September, 2011
CMS/ITRI	Chinese Taipei	September, 2011
NMIA	Australia	October, 2011
NMISA	South Africa	November, 2011
NMIJ	Japan	December, 2011
KRISS	Korea	April, 2012

## 3 Task and Purpose of the Comparison

Recent developments in technology and improvements at the NMIs have extended the low-frequency vibration limit of calibration capabilities down to 0.5 Hz and even as low as 0.1 Hz. Therefore, during the meeting of APMP TCAUV in 2008, the decision was taken to

---

make preparations for a further comparison targeted at a low-frequency range.

In the field of vibration, this regional key comparison was organized in order to compare measurements of sinusoidal linear accelerations in the frequency range from 0.5 Hz to 20 Hz. Moreover, the magnitude of the complex sensitivity calibration and measurement capabilities (CMCs) of the participating laboratories for accelerometer calibration were to be examined and compared. The phase shift was optional. It was the task of the comparison to measure the magnitude of the complex sensitivity of two accelerometer standard sets (one set included a quartz-flexure servo accelerometer of single-ended type and a signal conditioner) at different frequencies with acceleration amplitudes as specified in section 3 of [3]. The results of this APMP comparison will, after approval by CCAUV, serve as empirical support for the registration of low-frequency vibration ‘calibration and measurement capabilities’ (CMCs) in the framework of the CIPM MRA.

The voltage sensitivity is calculated as the ratio of the amplitude of the accelerometer standard set output voltage to the amplitude of the acceleration at its reference surface. The magnitude of the complex voltage sensitivity was given in millivolts per meter per second squared ( $\text{mV}/(\text{m/s}^2)$ ) for the different measurement conditions specified in section 4 of [3].

For the calibration of the accelerometer standard sets, laser interferometry in compliance with method 1 or method 3 of the international standard ISO 16063-11:1999 [4] had to be applied, in order to cover the entire frequency range.

The reported sensitivities and associated uncertainties of measurement were used for the calculation of the key comparison reference value (KCRV).

## 4 Conditions and Instructions of Measurement

The participating laboratories observed to a large extent the conditions stated in the Technical Protocol, i.e.

- frequencies in Hz:  
0.5, 0.63, 0.8, 1, 1.25, 1.6 (reference frequency), 2, 2.5, 3.15, 4, 5, 6.3, 8, 10, 12.5, 16, 20.
- acceleration amplitudes:  
A range of  $0.1 \text{ m/s}^2$  to  $10 \text{ m/s}^2$  was admissible.
- ambient temperature and accelerometer temperature during the calibration:  
 $(23 \pm 2)^\circ\text{C}$  (actual values to be stated within tolerances of  $\pm 0.3^\circ\text{C}$ ).

- relative humidity: max. 75 % RH

The comparison was performed in compliance with the “Guidelines for CIPM key comparisons” [2].

## 5 Transfer Standards as Artefacts

For the purpose of the comparison the pilot laboratory selected two accelerometers of which monitoring data for 6 months were available and of which data were not included in any published international cooperation work. Their pictures are in Figure 1.

- One transfer standard accelerometer (single-ended), type SA704, S/N 1021 (manufacturer: NIM) referred to as SE-1021.
- One transfer standard accelerometer (single-ended), type SA704, S/N 1022 (manufacturer: NIM) referred to as SE-1022 .
- One signal conditioner, type MSA-I, S/N 02011001 (manufacturer: NIM).

The investigation of the long-term stability was continued throughout the circulation period. The results of the NIM stability measurements and other individual data of the transfer standards are given in Section 7.



Figure 1: Two transfer standard accelerometers

## 6 Circulation of the Artefacts

The transducer sets were circulated in a star type fashion with a measurement period of two weeks provided for each participating laboratory and one week for the pilot laboratory. At the beginning and the end of the circulation as well as between certain subsequent measurements of participating laboratories, the transducer sets were measured by the pilot laboratory in order to determine reference values and to monitor the stability of the transducer sets (c.f. section 7). Due to the unexpected earthquake and tsunami which occurred in March 2011, the circulation had to be modified to incorporate two loops. The first loop was from NIM to NIMT, to CMS and back to NIM. The second loop was from

---

NIM to NMIA, to NMISA, to NMIIJ, to KRISS and back to NIM.

The input range of SA-704 is  $600 \text{ m/s}^2$  and the highest environment shock  $1000 \text{ m/s}^2$ . Therefore, any violent drop could change its sensitivity or even damage it. The transducer sets had to be hand-carried with great caution, shown in Figure 2.



Figure 2: Package of transfer standard set including two accelerometers and one signal conditioner

## 7 Results of the Measurements

### 7.1 Monitoring of stability

Starting with calibration data in February 2011, the artefacts were monitored during the preparatory period and the intervals of the comparison when they were back at the pilot laboratory. As a representative of the overall change, the measurements at reference frequency (1.6 Hz) are given in the following tables, with both the horizontal and vertical installation of artefacts.

---

Table 2: Horizontal voltage sensitivities of the SE-1021 at 1.6 Hz  
during the monitoring measurements

Month rel. to 2/2011	$S_{va}$ in mV/(m/s <sup>2</sup> )	rel. exp. Uncertainty in %
0	124.09	0.2
2	124.11	0.2
4	124.12	0.2
6	124.11	0.2
8	124.09	0.2
16	124.07	0.2

Table 3: Vertical voltage sensitivities of the SE-1021 at 1.6 Hz  
during the monitoring measurements

Month rel. to 2/2011	$S_{va}$ in mV/(m/s <sup>2</sup> )	rel. exp. Uncertainty in %
0	124.07	0.2
2	124.09	0.2
4	124.10	0.2
6	124.09	0.2
8	124.08	0.2
16	124.09	0.2

Table 4: Horizontal voltage sensitivities of the SE-1022 at 1.6 Hz  
during the monitoring measurements

Month rel. to 2/2011	$S_{va}$ in mV/(m/s <sup>2</sup> )	rel. exp. Uncertainty in %
0	123.95	0.2
2	124.00	0.2
4	124.00	0.2
6	123.99	0.2
8	123.97	0.2
16	123.94	0.2

---

Table 5: Vertical voltage sensitivities of the SE-1022 at 1.6 Hz  
during the monitoring measurements

Month rel. to 2/2011	$S_{va}$ in mV/(m/s <sup>2</sup> )	rel. exp. Uncertainty in %
0	123.94	0.2
2	123.98	0.2
4	123.98	0.2
6	123.97	0.2
8	123.96	0.2
16	123.95	0.2

These monitoring measurements can in the simplest way be summarized by the following statistical properties. It indicates that the stability of the artefacts was good considering the uncertainty claimed by the pilot laboratory in Appendix B.

Table 6: Mean and standard deviation of the voltage sensitivity of the artefacts  
calculated from the monitoring measurements

Artefact	long term mean in mV/(m/s <sup>2</sup> )	rel. std. deviation in %
SE-1021 (Horizontal)	124.098	0.01
SE-1021 (Vertical)	124.098	0.01
SE-1022 (Horizontal)	123.975	0.02
SE-1022 (Vertical)	123.963	0.01

## 7.2 Results of the Participants

It should be noted that the measurement results from the pilot laboratory were sent to the CCAUV Secretary prior to the circulation of the artefacts. The results include voltage sensitivities and phase shift of the two accelerometers in both horizontal and vertical installation directions.

### 7.2.1 Results of horizontal voltage sensitivities of the SE-1021

Table 7: Reported calibration results in mV/(m/s<sup>2</sup>) of the participants for horizontal voltage sensitivities of the SE-1021 with expanded relative uncertainty ( $k = 2$ ) in %

Frequency (Hz)	NIM		CMS		NMIA	
	$S_{va}$ (mV/ (m/s <sup>2</sup> ))	$U_c$ (%)	$S_{va}$ (mV/ (m/s <sup>2</sup> ))	$U_c$ (%)	$S_{va}$ (mV/ (m/s <sup>2</sup> ))	$U_c$ (%)
0.50	124.33	0.2	124.7	2.7	124.08	0.2
0.63	124.26	0.2	125.0	2.7	124.07	0.2
0.80	124.22	0.2	124.9	2.4	124.05	0.2
1.00	124.14	0.2	125.3	2.3	124.04	0.2
1.25	124.12	0.2	125.0	2.3	124.03	0.2
1.60	124.11	0.2	125.3	2.3	124.02	0.2
2.00	124.13	0.2	124.7	2.3	124.02	0.2
2.50	124.12	0.2	124.4	2.2	124.01	0.2
3.15	124.12	0.2	124.4	0.5	124.01	0.2
4.00	124.13	0.2	124.8	0.5	124.01	0.2
5.00	124.14	0.2	124.4	0.5	124.01	0.2
6.30	124.14	0.2	124.6	0.5	124.02	0.2
8.00	124.16	0.2	124.6	0.5	124.03	0.2
10.00	124.23	0.2	124.3	0.5	124.06	0.2
12.50	124.25	0.2	124.8	0.4	124.11	0.2
16.00	124.35	0.2	125.0	0.5	124.20	0.2
20.00	124.50	0.2	124.9	0.5	124.33	0.2

Table 8: Reported calibration results in mV/(m/s<sup>2</sup>) of the participants for horizontal voltage sensitivities of the SE-1021 with expanded relative uncertainty ( $k = 2$ ) in %

Frequency (Hz)	NMISA		NMIJ		KRISS	
	$S_{va}$ (mV/ (m/s <sup>2</sup> ))	$U_c$ (%)	$S_{va}$ (mV/ (m/s <sup>2</sup> ))	$U_c$ (%)	$S_{va}$ (mV/ (m/s <sup>2</sup> ))	$U_c$ (%)
0.50	123.9	0.3	124.81	2.0	124.27	0.5
0.63	123.9	0.3	123.93	0.8	124.23	0.5
0.80	123.9	0.3	123.85	0.8	124.13	0.3
1.00	124.0	0.3	123.91	0.9	124.13	0.5
1.25	124.0	0.3	123.76	0.7	124.06	0.3
1.60	124.0	0.3	123.79	0.5	124.07	0.4
2.00	124.0	0.3	123.88	0.7	124.16	0.3
2.50	124.0	0.3	-	-	124.12	0.3
3.15	123.9	0.3	-	-	124.03	0.3
4.00	124.0	0.3	-	-	124.01	0.3
5.00	124.0	0.3	-	-	124.05	0.3
6.30	124.0	0.3	-	-	123.97	0.3
8.00	124.0	0.3	-	-	123.98	0.4
10.00	124.1	0.3	-	-	124.01	0.4
12.50	124.1	0.3	-	-	124.05	0.4
16.00	124.1	0.3	-	-	124.20	0.3
20.00	124.3	0.3	-	-	124.27	0.3

### 7.2.2 Results of vertical voltage sensitivities of the SE-1021

Table 9: Reported calibration results in mV/(m/s<sup>2</sup>) of the participants for vertical voltage sensitivities of the SE-1021 with expanded relative uncertainty ( $k = 2$ ) in %

Frequency (Hz)	NIM		NIMT		NMJ	
	$S_{va}$ (mV/ (m/s <sup>2</sup> ))	$U_c$ (%)	$S_{va}$ (mV/ (m/s <sup>2</sup> ))	$U_c$ (%)	$S_{va}$ (mV/ (m/s <sup>2</sup> ))	$U_c$ (%)
0.50	124.17	0.2	124.06	0.6	-	-
0.63	124.16	0.2	124.06	0.6	-	-
0.80	124.15	0.2	124.06	0.6	-	-
1.00	124.10	0.2	124.06	0.5	-	-
1.25	124.09	0.2	124.06	0.5	-	-
1.60	124.09	0.2	124.06	0.5	-	-
2.00	124.10	0.2	124.06	0.5	-	-
2.50	124.11	0.2	124.06	0.5	123.86	0.3
3.15	124.11	0.2	124.06	0.5	123.91	0.3
4.00	124.11	0.2	124.07	0.5	123.94	0.3
5.00	124.12	0.2	124.10	0.4	123.95	0.3
6.30	124.11	0.2	124.10	0.4	123.98	0.3
8.00	124.14	0.2	124.12	0.4	124.01	0.3
10.00	124.16	0.2	124.15	0.4	124.05	0.3
12.50	124.20	0.2	124.20	0.4	124.11	0.3
16.00	124.30	0.2	124.28	0.4	124.20	0.3
20.00	124.42	0.2	124.41	0.4	124.34	0.3

### 7.2.3 Results of horizontal voltage sensitivities of the SE-1022

Table 10: Reported calibration results in mV/(m/s<sup>2</sup>) of the participants for horizontal voltage sensitivities of the SE-1022 with expanded relative uncertainty ( $k = 2$ ) in %

Frequency (Hz)	NIM		CMS		NMIA	
	$S_{va}$ (mV/ (m/s <sup>2</sup> ))	$U_c$ (%)	$S_{va}$ (mV/ (m/s <sup>2</sup> ))	$U_c$ (%)	$S_{va}$ (mV/ (m/s <sup>2</sup> ))	$U_c$ (%)
0.50	124.20	0.2	124.9	2.6	123.94	0.2
0.63	124.13	0.2	124.9	2.7	123.93	0.2
0.80	124.09	0.2	124.8	2.5	123.91	0.2
1.00	124.01	0.2	125.1	2.3	123.90	0.2
1.25	123.97	0.2	124.8	2.3	123.89	0.2
1.60	123.99	0.2	125.0	2.3	123.88	0.2
2.00	123.99	0.2	124.7	2.3	123.88	0.2
2.50	123.99	0.2	124.3	2.2	123.87	0.2
3.15	124.00	0.2	124.9	0.6	123.87	0.2
4.00	124.00	0.2	124.7	0.6	123.87	0.2
5.00	124.01	0.2	124.1	0.5	123.87	0.2
6.30	124.01	0.2	124.1	0.5	123.87	0.2
8.00	124.04	0.2	124.2	0.5	123.89	0.2
10.00	124.11	0.2	124.6	0.5	123.92	0.2
12.50	124.11	0.2	124.4	0.5	123.96	0.2
16.00	124.20	0.2	124.8	0.5	124.05	0.2
20.00	124.34	0.2	124.6	0.5	124.16	0.2

Table 11: Reported calibration results in mV/(m/s<sup>2</sup>) of the participants for horizontal voltage sensitivities of the SE-1022 with expanded relative uncertainty ( $k = 2$ ) in %

Frequency (Hz)	NMISA		NMIJ		KRISS	
	$S_{va}$ (mV/ (m/s <sup>2</sup> ))	$U_c$ (%)	$S_{va}$ (mV/ (m/s <sup>2</sup> ))	$U_c$ (%)	$S_{va}$ (mV/ (m/s <sup>2</sup> ))	$U_c$ (%)
0.50	123.8	0.3	124.02	2.0	124.12	0.5
0.63	123.8	0.3	123.63	0.7	124.01	0.5
0.80	123.8	0.3	123.81	0.8	123.97	0.3
1.00	123.8	0.3	123.73	1.0	123.94	0.4
1.25	123.8	0.3	123.67	0.7	123.90	0.4
1.60	123.8	0.3	123.68	0.4	123.92	0.4
2.00	123.8	0.3	123.56	0.7	124.01	0.3
2.50	123.8	0.3	-	-	123.99	0.3
3.15	123.9	0.3	-	-	123.90	0.3
4.00	123.8	0.3	-	-	123.86	0.3
5.00	123.9	0.3	-	-	123.83	0.3
6.30	123.9	0.3	-	-	123.80	0.3
8.00	123.9	0.3	-	-	123.82	0.4
10.00	123.9	0.3	-	-	123.86	0.4
12.50	123.9	0.3	-	-	123.89	0.4
16.00	124.0	0.3	-	-	124.03	0.4
20.00	124.1	0.3	-	-	124.09	0.4

#### 7.2.4 Results of vertical voltage sensitivities of the SE-1022

Table 12: Reported calibration results in mV/(m/s<sup>2</sup>) of the participants for vertical voltage sensitivities of the SE-1022 with expanded relative uncertainty ( $k = 2$ ) in %

Frequency (Hz)	NIM		NIMT		NMJJ	
	$S_{va}$ (mV/ (m/s <sup>2</sup> ))	$U_c$ (%)	$S_{va}$ (mV/ (m/s <sup>2</sup> ))	$U_c$ (%)	$S_{va}$ (mV/ (m/s <sup>2</sup> ))	$U_c$ (%)
0.50	124.02	0.2	123.92	0.6	-	-
0.63	124.04	0.2	123.92	0.6	-	-
0.80	124.04	0.2	123.92	0.6	-	-
1.00	123.98	0.2	123.92	0.5	-	-
1.25	123.97	0.2	123.93	0.5	-	-
1.60	123.97	0.2	123.93	0.5	-	-
2.00	123.98	0.2	123.93	0.5	-	-
2.50	123.98	0.2	123.93	0.5	123.71	0.3
3.15	123.98	0.2	123.93	0.5	123.77	0.3
4.00	123.99	0.2	123.94	0.5	123.80	0.3
5.00	123.99	0.2	123.96	0.4	123.82	0.3
6.30	123.99	0.2	123.97	0.4	123.85	0.3
8.00	124.00	0.2	123.98	0.4	123.88	0.3
10.00	124.02	0.2	124.01	0.4	123.91	0.3
12.50	124.07	0.2	124.05	0.4	123.97	0.3
16.00	124.13	0.2	124.13	0.4	124.05	0.3
20.00	124.24	0.2	124.25	0.4	124.17	0.3

### 7.3 Key Comparison Reference Value

The weighted mean was agreed upon by all laboratories to calculate the KCRVs for the APMP.AUV.V-K3 data. KCRVs were calculated separately at each frequency point measured (17 points in total) for both accelerometers in the horizontal direction and the vertical direction.

#### Calculation of KCRVs using the weighted mean method

Tables 7 to 12 contain the data for the accelerometers reported by the participating laboratories. For each laboratory  $i$  these data are (1)  $x_{i,f}$  : best estimate of sensitivity at

---

frequency  $f$ , and (2)  $u(x_{i,f})$ : associated standard uncertainty of sensitivity reported at frequency  $f$ .

For the transfer standards and at each frequency  $f$ , a key comparison reference value  $x_{WM}$  has been determined as the weighted mean of the results of  $n$  laboratories (for this comparison,  $n = 7$ ) according to

$$x_{WM} = \frac{\sum_{i=1}^n \frac{x_{i,f}}{u^2(x_{i,f})}}{\sum_{i=1}^n \frac{1}{u^2(x_{i,f})}} \quad (1)$$

$$u^2(x_{WM}) = \frac{1}{\sum_{i=1}^n \frac{1}{u^2(x_{i,f})}} \quad (2)$$

These calculated KCRVs are reported in Table 13.

The degree of equivalence,  $D_{lab-WM}$ , and  $U_{lab-WM}$ , was determined for the magnitude measurements for the accelerometer using

$$D_{lab-WM} = X_{lab} - X_{WM}, \quad U_{lab-WM} = k \cdot \sqrt{u_{lab}^2 - u_{WM}^2} \quad (3)$$

where  $X_{lab}$  represents the measurement results obtained by the laboratory at each frequency point for the magnitude and  $X_{WM}$  represents the reference value (KCRV) calculated as the weighted mean using Eq. (1).  $U_{lab-WM}$  is the uncertainty of measurement associated with the calculated  $D_{lab-WM}$  for  $k = 2$ . The calculated degrees of equivalence are completely reported in Appendix C and partly shown in Figures 3 to 10 for SE-1021 and SE-1022 in both horizontal and vertical installation at 0.5 Hz, 1.6 Hz, 10 Hz and 16 Hz. Application of the weighted mean is justified when the data including the stated uncertainties are consistent with one another. To check this, the following criterion of chi-squared test has been applied:

$$\min_{\zeta} \sum_{i=1}^n (x_{i,f} - \zeta)^2 / u^2(x_{i,f}) \leq (n-1) + k\sqrt{2(n-1)} \quad (4)$$

where  $n$  denotes the number of laboratories concerned, cf., e.g., [1] and  $k = 3$ . All the measurement results passed the consistency test.

Table 13: Reference values and associated expanded uncertainty ( $k = 2$ ) in mV/(m/s<sup>2</sup>)  
for SE-1021 and SE-1022 in horizontal and vertical directions

Frequency (Hz)	1021H		1021V		1022H		1022V	
	$X_{WM}$	$U_{WM}$	$X_{WM}$	$U_{WM}$	$X_{WM}$	$U_{WM}$	$X_{WM}$	$U_{WM}$
	in mV/(m/s <sup>2</sup> )							
0.50	<b>124.17</b>	0.15	<b>124.16</b>	0.24	<b>124.03</b>	0.15	<b>124.01</b>	0.24
0.63	<b>124.12</b>	0.15	<b>124.15</b>	0.24	<b>123.97</b>	0.15	<b>124.03</b>	0.24
0.80	<b>124.10</b>	0.14	<b>124.14</b>	0.24	<b>123.96</b>	0.14	<b>124.03</b>	0.24
1.00	<b>124.07</b>	0.15	<b>124.09</b>	0.23	<b>123.94</b>	0.15	<b>123.97</b>	0.23
1.25	<b>124.05</b>	0.14	<b>124.09</b>	0.23	<b>123.91</b>	0.15	<b>123.96</b>	0.23
1.60	<b>124.04</b>	0.15	<b>124.09</b>	0.23	<b>123.90</b>	0.14	<b>123.96</b>	0.23
2.00	<b>124.07</b>	0.14	<b>124.09</b>	0.23	<b>123.92</b>	0.14	<b>123.97</b>	0.23
2.50	<b>124.06</b>	0.15	<b>124.04</b>	0.20	<b>123.92</b>	0.15	<b>123.90</b>	0.20
3.15	<b>124.06</b>	0.14	<b>124.05</b>	0.20	<b>123.95</b>	0.14	<b>123.92</b>	0.20
4.00	<b>124.08</b>	0.14	<b>124.06</b>	0.20	<b>123.94</b>	0.14	<b>123.93</b>	0.20
5.00	<b>124.07</b>	0.14	<b>124.07</b>	0.19	<b>123.92</b>	0.14	<b>123.94</b>	0.19
6.30	<b>124.08</b>	0.14	<b>124.07</b>	0.19	<b>123.92</b>	0.14	<b>123.95</b>	0.19
8.00	<b>124.10</b>	0.15	<b>124.10</b>	0.19	<b>123.96</b>	0.15	<b>123.96</b>	0.19
10.00	<b>124.13</b>	0.15	<b>124.13</b>	0.19	<b>124.02</b>	0.15	<b>123.99</b>	0.19
12.50	<b>124.21</b>	0.14	<b>124.17</b>	0.19	<b>124.03</b>	0.15	<b>124.04</b>	0.19
16.00	<b>124.28</b>	0.14	<b>124.27</b>	0.19	<b>124.13</b>	0.15	<b>124.11</b>	0.19
20.00	<b>124.41</b>	0.14	<b>124.40</b>	0.19	<b>124.23</b>	0.15	<b>124.22</b>	0.19




Figure 3: Degree of equivalence for horizontal voltage sensitivities of the SE-1021  
at 0.5 Hz and 1.6 Hz with error bars representing the expanded uncertainty ( $k = 2$ )




Figure 4: Degree of equivalence for horizontal voltage sensitivities of the SE-1021 at 10 Hz and 16 Hz with error bars representing the expanded uncertainty ( $k = 2$ )




Figure 5: Degree of equivalence for vertical voltage sensitivities of the SE-1021 at 0.5 Hz and 1.6 Hz with error bars representing the expanded uncertainty ( $k = 2$ )




Figure 6: Degree of equivalence for vertical voltage sensitivities of the SE-1021 at 10 Hz and 16 Hz with error bars representing the expanded uncertainty ( $k = 2$ )




Figure 7: Degree of equivalence for horizontal voltage sensitivities of the SE-1022 at 0.5 Hz and 1.6 Hz with error bars representing the expanded uncertainty ( $k = 2$ )




Figure 8: Degree of equivalence for horizontal voltage sensitivities of the SE-1022 at 10 Hz and 16 Hz with error bars representing the expanded uncertainty ( $k = 2$ )




Figure 9: Degree of equivalence for vertical voltage sensitivities of the SE-1022 at 0.5 Hz and 1.6 Hz with error bars representing the expanded uncertainty ( $k = 2$ )




Figure 10: Degree of equivalence for vertical voltage sensitivities of the SE-1022 at 10 Hz and 16 Hz with error bars representing the expanded uncertainty ( $k = 2$ )

## 7.4 Degrees of Equivalence between participants

In order to compare the individual results of the participating laboratories of this comparison with one another, the Degree of Equivalence (DoE) of pairs of results with respect to a certain frequency were calculated. These DoEs are each a pair of values of the difference  $D_{ij}$  between the respective participants i and j and the combined expanded uncertainty  $U_{ij}$  of this difference. These values are calculated for each frequency according to:

$$D_{ij} = x_i - x_j \quad (5)$$

$$U_{ij} = k \cdot \sqrt{u^2(x_i) + u^2(x_j)} \quad (6)$$

with a coverage factor of  $k = 2$ .

### 7.4.1 Tables of DoE between participants for horizontal voltage sensitivities of the SE-1021

Table 14: Degrees of equivalence between the participants for horizontal voltage sensitivities of the SE-1021 at 0.5 Hz

0.50 Hz	NIM		CMS		NMIA		NMISA		NMIJ		KRISS	
$i \rightarrow$	$D_{ij}$	$U_{ij}$	$D_{ij}$	$U_{ij}$	$D_{ij}$	$U_{ij}$	$D_{ij}$	$U_{ij}$	$D_{ij}$	$U_{ij}$	$D_{ij}$	$U_{ij}$
$j \downarrow$	(mV/(m/s <sup>2</sup> )•10 <sup>-2</sup> )		(mV/(m/s <sup>2</sup> )•10 <sup>-2</sup> )		(mV/(m/s <sup>2</sup> )•10 <sup>-2</sup> )		(mV/(m/s <sup>2</sup> )•10 <sup>-2</sup> )		(mV/(m/s <sup>2</sup> )•10 <sup>-2</sup> )		(mV/(m/s <sup>2</sup> )•10 <sup>-2</sup> )	
NIM	-	-	<b>37.0</b>	336.9	<b>-25.0</b>	27.6	<b>-43.0</b>	<b>39.1</b>	<b>48.0</b>	249.9	<b>-6.0</b>	63.3
CMS	<b>-37.0</b>	336.9	-	-	<b>-62.0</b>	336.9	<b>-80.0</b>	338.0	<b>11.0</b>	418.6	<b>-43.0</b>	341.7
NMIA	<b>25.0</b>	27.6	<b>62.0</b>	336.9	-	-	<b>-18.0</b>	39.1	<b>73.0</b>	249.9	<b>19.0</b>	63.3
NMISA	<b>43.0</b>	<b>39.1</b>	<b>80.0</b>	338.0	<b>18.0</b>	39.1	-	-	<b>91.0</b>	251.4	<b>37.0</b>	69.1
NMIJ	<b>-48.0</b>	249.9	<b>-11.0</b>	418.6	<b>-73.0</b>	249.9	<b>-91.0</b>	251.4	-	-	<b>-54.0</b>	256.3
KRISS	<b>6.0</b>	63.3	<b>43.0</b>	341.7	<b>-19.0</b>	63.3	<b>-37.0</b>	69.1	<b>54.0</b>	256.3	-	-

Table 15: Degrees of equivalence between the participants for horizontal voltage sensitivities of the SE-1021 at 0.63 Hz

0.63 Hz	NIM		CMS		NMIA		NMISA		NMIJ		KRISS	
$i \rightarrow$	$D_{ij}$	$U_{ij}$										
$j \downarrow$	(mV/(m/s <sup>2</sup> )•10 <sup>-2</sup> )											
NIM	-	-	<b>74.0</b>	337.7	<b>-19.0</b>	27.8	<b>-36.0</b>	39.2	<b>-33.0</b>	99.9	<b>-3.0</b>	63.4
CMS	<b>-74.0</b>	337.7	-	-	<b>-93.0</b>	337.7	<b>-110.0</b>	338.9	<b>-107.0</b>	351.1	<b>-77.0</b>	342.5
NMIA	<b>19.0</b>	27.8	<b>93.0</b>	337.7	-	-	<b>-17.0</b>	39.2	<b>-14.0</b>	99.9	<b>16.0</b>	63.3
NMISA	<b>36.0</b>	39.2	<b>110.0</b>	338.9	<b>17.0</b>	39.2	-	-	<b>3.0</b>	103.7	<b>33.0</b>	69.1
NMIJ	<b>33.0</b>	99.9	<b>107.0</b>	351.1	<b>14.0</b>	99.9	<b>-3.0</b>	103.7	-	-	<b>30.0</b>	115.0
KRISS	<b>3.0</b>	63.4	<b>77.0</b>	342.5	<b>-16.0</b>	63.3	<b>-33.0</b>	69.1	<b>-30.0</b>	115.0	-	-

Table 16 Degrees of equivalence between the participants for horizontal voltage sensitivities of the SE-1021 at 0.8 Hz

0.8 Hz	NIM		CMS		NMIA		NMISA		NMIJ		KRISS	
$i \rightarrow$	$D_{ij}$	$U_{ij}$										
$j \downarrow$	(mV/(m/s <sup>2</sup> )•10 <sup>-2</sup> )											
NIM	-	-	<b>68.0</b>	300.1	<b>-17.0</b>	28.6	<b>-32.0</b>	39.8	<b>-37.0</b>	100.1	<b>-9.0</b>	39.8
CMS	<b>-68.0</b>	300.1	-	-	<b>-85.0</b>	300.1	<b>-100.0</b>	301.4	<b>-105.0</b>	315.0	<b>-77.0</b>	301.4
NMIA	<b>17.0</b>	28.6	<b>85.0</b>	300.1	-	-	<b>-15.0</b>	39.8	<b>-20.0</b>	100.1	<b>8.0</b>	39.8
NMISA	<b>32.0</b>	39.8	<b>100.0</b>	301.4	<b>15.0</b>	39.8	-	-	<b>-5.0</b>	103.8	<b>23.0</b>	48.5
NMIJ	<b>37.0</b>	100.1	<b>105.0</b>	315.0	<b>20.0</b>	100.1	<b>5.0</b>	103.8	-	-	<b>28.0</b>	103.9
KRISS	<b>9.0</b>	39.8	<b>77.0</b>	301.4	<b>-8.0</b>	39.8	<b>-23.0</b>	48.5	<b>-28.0</b>	103.9	-	-

Table 17: Degrees of equivalence between the participants for horizontal voltage sensitivities of the SE-1021 at 1 Hz

1.0 Hz	NIM		CMS		NMIA		NMISA		NMIJ		KRISS	
$i \rightarrow$	$D_{ij}$	$U_{ij}$										
$j \downarrow$	(mV/(m/s <sup>2</sup> )•10 <sup>-2</sup> )											
NIM	-	-	<b>116.0</b>	288.5	<b>-10.0</b>	27.7	<b>-14.0</b>	39.2	<b>-23.0</b>	112.2	<b>-1.0</b>	63.3
CMS	<b>-116.0</b>	288.5	-	-	<b>-126.0</b>	288.5	<b>-130.0</b>	289.8	<b>-139.0</b>	308.3	<b>-117.0</b>	294.0
NMIA	<b>10.0</b>	27.7	<b>126.0</b>	288.5	-	-	<b>-4.0</b>	39.2	<b>-13.0</b>	112.2	<b>9.0</b>	63.3
NMISA	<b>14.0</b>	39.2	<b>130.0</b>	289.8	<b>4.0</b>	39.2	-	-	<b>-9.0</b>	115.6	<b>13.0</b>	69.1
NMIJ	<b>23.0</b>	112.2	<b>139.0</b>	308.3	<b>13.0</b>	112.2	<b>9.0</b>	115.6	-	-	<b>22.0</b>	125.8
KRISS	<b>1.0</b>	63.3	<b>117.0</b>	294.0	<b>-9.0</b>	63.3	<b>-13.0</b>	69.1	<b>-22.0</b>	125.8	-	-

Table 18: Degrees of equivalence between the participants for horizontal voltage sensitivities of the SE-1021 at 1.25 Hz

1.25 Hz	NIM		CMS		NMIA		NMISA		NMIJ		KRISS	
$i \rightarrow$	$D_{ij}$	$U_{ij}$										
$j \downarrow$	(mV/(m/s <sup>2</sup> )•10 <sup>-2</sup> )											
NIM	-	-	<b>88.0</b>	287.9	<b>-9.0</b>	28.6	<b>-12.0</b>	39.8	<b>-36.0</b>	87.8	<b>-6.0</b>	39.9
CMS	<b>-88.0</b>	287.9	-	-	<b>-97.0</b>	287.9	<b>-100.0</b>	289.2	<b>-124.0</b>	299.6	<b>-94.0</b>	289.2
NMIA	<b>9.0</b>	28.6	<b>97.0</b>	287.9	-	-	<b>-3.0</b>	39.8	<b>-27.0</b>	87.8	<b>3.0</b>	39.8
NMISA	<b>12.0</b>	39.8	<b>100.0</b>	289.2	<b>3.0</b>	39.8	-	-	<b>-24.0</b>	92.1	<b>6.0</b>	48.5
NMIJ	<b>36.0</b>	87.8	<b>124.0</b>	299.6	<b>27.0</b>	87.8	<b>24.0</b>	92.1	-	-	<b>30.0</b>	92.1
KRISS	<b>6.0</b>	39.9	<b>94.0</b>	289.2	<b>-3.0</b>	39.8	<b>-6.0</b>	48.5	<b>-30.0</b>	92.1	-	-

Table 19: Degrees of equivalence between the participants for horizontal voltage sensitivities of the SE-1021 at 1.6 Hz

1.6 Hz	NIM		CMS		NMIA		NMISA		NMIJ		KRISS	
$i \rightarrow$	$D_{ij}$	$U_{ij}$										
$j \downarrow$	(mV/(m/s <sup>2</sup> )•10 <sup>-2</sup> )											
NIM	-	-	<b>119.0</b>	288.5	<b>-9.0</b>	28.3	<b>-11.0</b>	39.6	<b>-32.0</b>	63.4	<b>-4.0</b>	51.5
CMS	<b>-119.0</b>	288.5	-	-	<b>-128.0</b>	288.5	<b>-130.0</b>	289.8	<b>-151.0</b>	294.0	<b>-123.0</b>	291.7
NMIA	<b>9.0</b>	28.3	<b>128.0</b>	288.5	-	-	<b>-2.0</b>	39.6	<b>-23.0</b>	63.4	<b>5.0</b>	51.5
NMISA	<b>11.0</b>	39.6	<b>130.0</b>	289.8	<b>2.0</b>	39.6	-	-	<b>-21.0</b>	69.2	<b>7.0</b>	58.5
NMIJ	<b>32.0</b>	63.4	<b>151.0</b>	294.0	<b>23.0</b>	63.4	<b>21.0</b>	69.2	-	-	<b>28.0</b>	76.6
KRISS	<b>4.0</b>	51.5	<b>123.0</b>	291.7	<b>-5.0</b>	51.5	<b>-7.0</b>	58.5	<b>-28.0</b>	76.6	-	-

Table 20: Degrees of equivalence between the participants for horizontal voltage sensitivities of the SE-1021 at 2 Hz

2.0 Hz	NIM		CMS		NMIA		NMISA		NMIJ		KRISS	
$i \rightarrow$	$D_{ij}$	$U_{ij}$										
$j \downarrow$	(mV/(m/s <sup>2</sup> )•10 <sup>-2</sup> )											
NIM	-	-	<b>57.0</b>	287.2	<b>-11.0</b>	28.6	<b>-13.0</b>	39.8	<b>-25.0</b>	87.9	<b>3.0</b>	39.9
CMS	<b>-57.0</b>	287.2	-	-	<b>-68.0</b>	287.2	<b>-70.0</b>	288.5	<b>-82.0</b>	298.9	<b>-54.0</b>	288.5
NMIA	<b>11.0</b>	28.6	<b>68.0</b>	287.2	-	-	<b>-2.0</b>	39.8	<b>-14.0</b>	87.9	<b>14.0</b>	39.9
NMISA	<b>13.0</b>	39.8	<b>70.0</b>	288.5	<b>2.0</b>	39.8	-	-	<b>-12.0</b>	92.1	<b>16.0</b>	48.6
NMIJ	<b>25.0</b>	87.9	<b>82.0</b>	298.9	<b>14.0</b>	87.9	<b>12.0</b>	92.1	-	-	<b>28.0</b>	92.2
KRISS	<b>-3.0</b>	39.9	<b>54.0</b>	288.5	<b>-14.0</b>	39.9	<b>-16.0</b>	48.6	<b>-28.0</b>	92.2	-	-

Table 21: Degrees of equivalence between the participants for horizontal voltage sensitivities of the SE-1021 at 2.5 Hz

2.5 Hz	NIM		CMS		NMIA		NMISA		NMIJ		KRISS	
$i \rightarrow$	$D_{ij}$	$U_{ij}$										
$j \downarrow$	(mV/(m/s <sup>2</sup> )•10 <sup>-2</sup> )											
NIM	-	-	<b>28.0</b>	274.0	<b>-11.0</b>	28.4	<b>-12.0</b>	39.7	-	-	<b>0.0</b>	39.7
CMS	<b>-28.0</b>	274.0	-	-	<b>-39.0</b>	274.0	<b>-40.0</b>	275.4	-	-	<b>-28.0</b>	275.4
NMIA	<b>11.0</b>	28.4	<b>39.0</b>	274.0	-	-	<b>-1.0</b>	39.7	-	-	<b>11.0</b>	39.7
NMISA	<b>12.0</b>	39.7	<b>40.0</b>	275.4	<b>1.0</b>	39.7	-	-	-	-	<b>12.0</b>	48.4
NMIJ	-	-	-	-	-	-	-	-	-	-	-	-
KRISS	<b>0.0</b>	39.7	<b>28.0</b>	275.4	<b>-11.0</b>	39.7	<b>-12.0</b>	48.4	-	-	-	-

Table 22: Degrees of equivalence between the participants for horizontal voltage sensitivities of the SE-1021 at 3.15 Hz

3.15 Hz	NIM		CMS		NMIA		NMISA		NMIJ		KRISS	
$i \rightarrow$	$D_{ij}$	$U_{ij}$										
$j \downarrow$	(mV/(m/s <sup>2</sup> )•10 <sup>-2</sup> )											
NIM	-	-	<b>28.0</b>	63.9	<b>-11.0</b>	28.8	<b>-22.0</b>	39.9	-	-	<b>-9.0</b>	40.0
CMS	<b>-28.0</b>	63.9	-	-	<b>-39.0</b>	63.9	<b>-50.0</b>	69.6	-	-	<b>-37.0</b>	69.6
NMIA	<b>11.0</b>	28.8	<b>39.0</b>	63.9	-	-	<b>-11.0</b>	39.9	-	-	<b>2.0</b>	39.9
NMISA	<b>22.0</b>	39.9	<b>50.0</b>	69.6	<b>11.0</b>	39.9	-	-	-	-	<b>13.0</b>	48.6
NMIJ	-	-	-	-	-	-	-	-	-	-	-	-
KRISS	<b>9.0</b>	40.0	<b>37.0</b>	69.6	<b>-2.0</b>	39.9	<b>-13.0</b>	48.6	-	-	-	-

Table 23: Degrees of equivalence between the participants for horizontal voltage sensitivities of the SE-1021 at 4 Hz

4.0 Hz	NIM		CMS		NMIA		NMISA		NMIJ		KRISS	
$i \rightarrow$	$D_{ij}$	$U_{ij}$										
$j \downarrow$	(mV/(m/s <sup>2</sup> )•10 <sup>-2</sup> )											
NIM	-	-	<b>67.0</b>	64.1	<b>-12.0</b>	28.8	<b>-13.0</b>	40.0	-	-	<b>-12.0</b>	40.0
CMS	<b>-67.0</b>	64.1	-	-	<b>-79.0</b>	64.1	<b>-80.0</b>	69.8	-	-	<b>-79.0</b>	69.8
NMIA	<b>12.0</b>	28.8	<b>79.0</b>	64.1	-	-	<b>-1.0</b>	39.9	-	-	<b>0.0</b>	39.9
NMISA	<b>13.0</b>	40.0	<b>80.0</b>	69.8	<b>1.0</b>	39.9	-	-	-	-	<b>1.0</b>	48.6
NMIJ	-	-	-	-	-	-	-	-	-	-	-	-
KRISS	<b>12.0</b>	40.0	<b>79.0</b>	69.8	<b>0.0</b>	39.9	<b>-1.0</b>	48.6	-	-	-	-

Table 24: Degrees of equivalence between the participants for horizontal voltage sensitivities of the SE-1021 at 5 Hz

5.0 Hz	NIM		CMS		NMIA		NMISA		NMIJ		KRISS	
$i \rightarrow$	$D_{ij}$	$U_{ij}$										
$j \downarrow$	(mV/(m/s <sup>2</sup> )•10 <sup>-2</sup> )											
NIM	-	-	<b>26.0</b>	63.9	<b>-13.0</b>	28.8	<b>-14.0</b>	40.0	-	-	<b>-9.0</b>	40.0
CMS	<b>-26.0</b>	63.9	-	-	<b>-39.0</b>	63.9	<b>-40.0</b>	69.6	-	-	<b>-35.0</b>	69.6
NMIA	<b>13.0</b>	28.8	<b>39.0</b>	63.9	-	-	<b>-1.0</b>	39.9	-	-	<b>4.0</b>	39.9
NMISA	<b>14.0</b>	40.0	<b>40.0</b>	69.6	<b>1.0</b>	39.9	-	-	-	-	<b>5.0</b>	48.6
NMIJ	-	-	-	-	-	-	-	-	-	-	-	-
KRISS	<b>9.0</b>	40.0	<b>35.0</b>	69.6	<b>-4.0</b>	39.9	<b>-5.0</b>	48.6	-	-	-	-

Table 25: Degrees of equivalence between the participants for horizontal voltage sensitivities of the SE-1021 at 6.3 Hz

6.3 Hz	NIM		CMS		NMIA		NMISA		NMIJ		KRISS	
$i \rightarrow$	$D_{ij}$	$U_{ij}$										
$j \downarrow$	(mV/(m/s <sup>2</sup> )•10 <sup>-2</sup> )											
NIM	-	-	<b>46.0</b>	64.0	<b>-12.0</b>	28.8	<b>-14.0</b>	40.0	-	-	<b>-17.0</b>	39.9
CMS	<b>-46.0</b>	64.0	-	-	<b>-58.0</b>	64.0	<b>-60.0</b>	69.7	-	-	<b>-63.0</b>	69.7
NMIA	<b>12.0</b>	28.8	<b>58.0</b>	64.0	-	-	<b>-2.0</b>	39.9	-	-	<b>-5.0</b>	39.9
NMISA	<b>14.0</b>	40.0	<b>60.0</b>	69.7	<b>2.0</b>	39.9	-	-	-	-	<b>-3.0</b>	48.6
NMIJ	-	-	-	-	-	-	-	-	-	-	-	-
KRISS	<b>17.0</b>	39.9	<b>63.0</b>	69.7	<b>5.0</b>	39.9	<b>3.0</b>	48.6	-	-	-	-

Table 26: Degrees of equivalence between the participants for horizontal voltage sensitivities of the SE-1021 at 8 Hz

8.0 Hz	NIM		CMS		NMIA		NMISA		NMIJ		KRISS	
$i \rightarrow$	$D_{ij}$	$U_{ij}$										
$j \downarrow$	(mV/(m/s <sup>2</sup> )•10 <sup>-2</sup> )											
NIM	-	-	<b>44.0</b>	63.8	<b>-13.0</b>	28.3	<b>-16.0</b>	39.6	-	-	<b>-18.0</b>	51.4
CMS	<b>-44.0</b>	63.8	-	-	<b>-57.0</b>	63.8	<b>-60.0</b>	69.5	-	-	<b>-62.0</b>	76.9
NMIA	<b>13.0</b>	28.3	<b>57.0</b>	63.8	-	-	<b>-3.0</b>	39.6	-	-	<b>-5.0</b>	51.4
NMISA	<b>16.0</b>	39.6	<b>60.0</b>	69.5	<b>3.0</b>	39.6	-	-	-	-	<b>-2.0</b>	58.4
NMIJ	-	-	-	-	-	-	-	-	-	-	-	-
KRISS	<b>18.0</b>	51.4	<b>62.0</b>	76.9	<b>5.0</b>	51.4	<b>2.0</b>	58.4	-	-	-	-

Table 27: Degrees of equivalence between the participants for horizontal voltage sensitivities of the SE-1021 at 10 Hz

10.0 Hz	NIM		CMS		NMIA		NMISA		NMIJ		KRISS	
$i \rightarrow$	$D_{ij}$	$U_{ij}$										
$j \downarrow$	(mV/(m/s <sup>2</sup> )•10 <sup>-2</sup> )											
NIM	-	-	<b>7.0</b>	63.6	<b>-17.0</b>	28.3	<b>-13.0</b>	39.6	-	-	<b>-22.0</b>	51.4
CMS	<b>-7.0</b>	63.6	-	-	<b>-24.0</b>	63.6	<b>-20.0</b>	69.4	-	-	<b>-29.0</b>	76.8
NMIA	<b>17.0</b>	28.3	<b>24.0</b>	63.6	-	-	<b>4.0</b>	39.6	-	-	<b>-5.0</b>	51.4
NMISA	<b>13.0</b>	39.6	<b>20.0</b>	69.4	<b>-4.0</b>	39.6	-	-	-	-	<b>-9.0</b>	58.4
NMIJ	-	-	-	-	-	-	-	-	-	-	-	-
KRISS	<b>22.0</b>	51.4	<b>29.0</b>	76.8	<b>5.0</b>	51.4	<b>9.0</b>	58.4	-	-	-	-

Table 28: Degrees of equivalence between the participants for horizontal voltage sensitivities of the SE-1021 at 12.5 Hz

12.5 Hz	NIM		CMS		NMIA		NMISA		NMIJ		KRISS	
$i \rightarrow$	$D_{ij}$	$U_{ij}$	$D_{ij}$	$U_{ij}$	$D_{ij}$	$U_{ij}$	$D_{ij}$	$U_{ij}$	$D_{ij}$	$U_{ij}$	$D_{ij}$	$U_{ij}$
$j \downarrow$	(mV/(m/s <sup>2</sup> )•10 <sup>-2</sup> )		(mV/(m/s <sup>2</sup> )•10 <sup>-2</sup> )		(mV/(m/s <sup>2</sup> )•10 <sup>-2</sup> )							
NIM	-	-	<b>55.0</b>	51.9	<b>-14.0</b>	28.5	<b>-15.0</b>	39.8	-	-	<b>-20.0</b>	51.6
CMS	<b>-55.0</b>	51.9	-	-	<b>-69.0</b>	51.8	<b>-70.0</b>	<b>58.8</b>	-	-	<b>-75.0</b>	67.3
NMIA	<b>14.0</b>	28.5	<b>69.0</b>	51.8	-	-	<b>-1.0</b>	39.8	-	-	<b>-6.0</b>	51.6
NMISA	<b>15.0</b>	39.8	<b>70.0</b>	58.8	<b>1.0</b>	39.8	-	-	-	-	<b>-5.0</b>	58.6
NMIJ	-	-	-	-	-	-	-	-	-	-	-	-
KRISS	<b>20.0</b>	51.6	<b>75.0</b>	67.3	<b>6.0</b>	51.6	<b>5.0</b>	58.6	-	-	-	-

Table 29: Degrees of equivalence between the participants for horizontal voltage sensitivities of the SE-1021 at 16 Hz

16.0 Hz	NIM		CMS		NMIA		NMISA		NMIJ		KRISS	
$i \rightarrow$	$D_{ij}$	$U_{ij}$	$D_{ij}$	$U_{ij}$	$D_{ij}$	$U_{ij}$	$D_{ij}$	$U_{ij}$	$D_{ij}$	$U_{ij}$	$D_{ij}$	$U_{ij}$
$j \downarrow$	(mV/(m/s <sup>2</sup> )•10 <sup>-2</sup> )		(mV/(m/s <sup>2</sup> )•10 <sup>-2</sup> )		(mV/(m/s <sup>2</sup> )•10 <sup>-2</sup> )							
NIM	-	-	<b>65.0</b>	64.2	<b>-15.0</b>	28.8	<b>-25.0</b>	40.0	-	-	<b>-15.0</b>	40.0
CMS	<b>-65.0</b>	64.2	-	-	<b>-80.0</b>	64.2	<b>-90.0</b>	<b>69.9</b>	-	-	<b>-80.0</b>	69.9
NMIA	<b>15.0</b>	28.8	<b>80.0</b>	64.2	-	-	<b>-10.0</b>	40.0	-	-	<b>0.0</b>	40.0
NMISA	<b>25.0</b>	40.0	<b>90.0</b>	69.9	<b>10.0</b>	40.0	-	-	-	-	<b>10.0</b>	48.7
NMIJ	-	-	-	-	-	-	-	-	-	-	-	-
KRISS	<b>15.0</b>	40.0	<b>80.0</b>	69.9	<b>0.0</b>	40.0	<b>-10.0</b>	48.7	-	-	-	-

Table 30: Degrees of equivalence between the participants for horizontal voltage sensitivities of the SE-1021 at 20 Hz

20.0 Hz	NIM		CMS		NMIA		NMISA		NMIJ		KRISS	
$i \rightarrow$	$D_{ij}$	$U_{ij}$										
$j \downarrow$	(mV/(m/s <sup>2</sup> )•10 <sup>-2</sup> )											
NIM	-	-	<b>40.0</b>	64.1	<b>-17.0</b>	28.8	<b>-20.0</b>	40.1	-	-	<b>-23.0</b>	40.0
CMS	<b>-40.0</b>	64.1	-	-	<b>-57.0</b>	64.1	<b>-60.0</b>	69.9	-	-	<b>-63.0</b>	69.9
NMIA	<b>17.0</b>	28.8	<b>57.0</b>	64.1	-	-	<b>-3.0</b>	40.0	-	-	<b>-6.0</b>	40.0
NMISA	<b>20.0</b>	40.1	<b>60.0</b>	69.9	<b>3.0</b>	40.0	-	-	-	-	<b>-3.0</b>	48.7
NMIJ	-	-	-	-	-	-	-	-	-	-	-	-
KRISS	<b>23.0</b>	40.0	<b>63.0</b>	69.9	<b>6.0</b>	40.0	<b>3.0</b>	48.7	-	-	-	-

#### 7.4.2 Tables of DoE between participants for vertical voltage sensitivities of the SE-1021

Table 31: Degrees of equivalence between the participants for vertical voltage sensitivities of the SE-1021 at 0.5 Hz

0.50 Hz	NIM		NIMT		NMIJ	
$i \rightarrow$	$D_{ij}$	$U_{ij}$	$D_{ij}$	$U_{ij}$	$D_{ij}$	$U_{ij}$
$j \downarrow$	(mV/(m/s <sup>2</sup> )•10 <sup>-2</sup> )		(mV/(m/s <sup>2</sup> )•10 <sup>-2</sup> )		(mV/(m/s <sup>2</sup> )•10 <sup>-2</sup> )	
NIM	-	-	<b>-11.0</b>	71.0	-	-
NIMT	<b>11.0</b>	71.0	-	-	-	-
NMIJ	-	-	-	-	-	-

Table 32: Degrees of equivalence between the participants for vertical voltage sensitivities of the SE-1021 at 0.63 Hz

0.63 Hz	NIM		NIMT		NMIJ	
$i \rightarrow$	$D_{ij}$	$U_{ij}$	$D_{ij}$	$U_{ij}$	$D_{ij}$	$U_{ij}$
$j \downarrow$	(mV/(m/s <sup>2</sup> )•10 <sup>-2</sup> )		(mV/(m/s <sup>2</sup> )•10 <sup>-2</sup> )		(mV/(m/s <sup>2</sup> )•10 <sup>-2</sup> )	
NIM	-	-	<b>-10.0</b>	71.0	-	-
NIMT	<b>10.0</b>	71.0	-	-	-	-
NMIJ	-	-	-	-	-	-

Table 33: Degrees of equivalence between the participants for vertical voltage sensitivities of the SE-1021 at 0.8 Hz

0.8 Hz	NIM		NIMT		NMIJ	
$i \rightarrow$	$D_{ij}$	$U_{ij}$	$D_{ij}$	$U_{ij}$	$D_{ij}$	$U_{ij}$
$j \downarrow$	$(\text{mV}/(\text{m/s}^2) \cdot 10^{-2})$		$(\text{mV}/(\text{m/s}^2) \cdot 10^{-2})$		$(\text{mV}/(\text{m/s}^2) \cdot 10^{-2})$	
NIM	-	-	<b>-9.0</b>	71.0	-	-
NIMT	<b>9.0</b>	71.0	-	-	-	-
NMIJ	-	-	-	-	-	-

Table 34: Degrees of equivalence between the participants for vertical voltage sensitivities of the SE-1021 at 1 Hz

1.0 Hz	NIM		NIMT		NMIJ	
$i \rightarrow$	$D_{ij}$	$U_{ij}$	$D_{ij}$	$U_{ij}$	$D_{ij}$	$U_{ij}$
$j \downarrow$	$(\text{mV}/(\text{m/s}^2) \cdot 10^{-2})$		$(\text{mV}/(\text{m/s}^2) \cdot 10^{-2})$		$(\text{mV}/(\text{m/s}^2) \cdot 10^{-2})$	
NIM	-	-	<b>-4.0</b>	58.3	-	-
NIMT	<b>4.0</b>	58.3	-	-	-	-
NMIJ	-	-	-	-	-	-

Table 35: Degrees of equivalence between the participants for vertical voltage sensitivities of the SE-1021 at 1.25 Hz

1.25 Hz	NIM		NIMT		NMIJ	
$i \rightarrow$	$D_{ij}$	$U_{ij}$	$D_{ij}$	$U_{ij}$	$D_{ij}$	$U_{ij}$
$j \downarrow$	$(\text{mV}/(\text{m/s}^2) \cdot 10^{-2})$		$(\text{mV}/(\text{m/s}^2) \cdot 10^{-2})$		$(\text{mV}/(\text{m/s}^2) \cdot 10^{-2})$	
NIM	-	-	<b>-3.0</b>	58.3	-	-
NIMT	<b>3.0</b>	58.3	-	-	-	-
NMIJ	-	-	-	-	-	-

Table 36: Degrees of equivalence between the participants for vertical voltage sensitivities of the SE-1021 at 1.6 Hz

1.6 Hz	NIM		NIMT		NMIJ	
$i \rightarrow$	$D_{ij}$	$U_{ij}$	$D_{ij}$	$U_{ij}$	$D_{ij}$	$U_{ij}$
$j \downarrow$	(mV/(m/s <sup>2</sup> )•10 <sup>-2</sup> )		(mV/(m/s <sup>2</sup> )•10 <sup>-2</sup> )		(mV/(m/s <sup>2</sup> )•10 <sup>-2</sup> )	
NIM	-	-	<b>-3.0</b>	58.3	-	-
NIMT	<b>3.0</b>	58.3	-	-	-	-
NMIJ	-	-	-	-	-	-

Table 37: Degrees of equivalence between the participants for vertical voltage sensitivities of the SE-1021 at 2 Hz

2.0 Hz	NIM		NIMT		NMIJ	
$i \rightarrow$	$D_{ij}$	$U_{ij}$	$D_{ij}$	$U_{ij}$	$D_{ij}$	$U_{ij}$
$j \downarrow$	(mV/(m/s <sup>2</sup> )•10 <sup>-2</sup> )		(mV/(m/s <sup>2</sup> )•10 <sup>-2</sup> )		(mV/(m/s <sup>2</sup> )•10 <sup>-2</sup> )	
NIM	-	-	<b>-4.0</b>	58.3	-	-
NIMT	<b>4.0</b>	58.3	-	-	-	-
NMIJ	-	-	-	-	-	-

Table 38: Degrees of equivalence between the participants for vertical voltage sensitivities of the SE-1021 at 2.5 Hz

2.5 Hz	NIM		NIMT		NMIJ	
$i \rightarrow$	$D_{ij}$	$U_{ij}$	$D_{ij}$	$U_{ij}$	$D_{ij}$	$U_{ij}$
$j \downarrow$	(mV/(m/s <sup>2</sup> )•10 <sup>-2</sup> )		(mV/(m/s <sup>2</sup> )•10 <sup>-2</sup> )		(mV/(m/s <sup>2</sup> )•10 <sup>-2</sup> )	
NIM	-	-	<b>-5.0</b>	60.8	<b>-25.0</b>	35.1
NIMT	<b>5.0</b>	60.8	-	-	<b>-20.0</b>	66.8
NMIJ	<b>25.0</b>	35.1	<b>20.0</b>	66.8	-	-

Table 39: Degrees of equivalence between the participants for vertical voltage sensitivities of the SE-1021 at 3.15 Hz

3.15 Hz	NIM		NIMT		NMIJ	
$i \rightarrow$	$D_{ij}$	$U_{ij}$	$D_{ij}$	$U_{ij}$	$D_{ij}$	$U_{ij}$
$j \downarrow$	(mV/(m/s <sup>2</sup> )•10 <sup>-2</sup> )		(mV/(m/s <sup>2</sup> )•10 <sup>-2</sup> )		(mV/(m/s <sup>2</sup> )•10 <sup>-2</sup> )	
NIM	-	-	<b>-5.0</b>	60.8	<b>-20.0</b>	35.1
NIMT	<b>5.0</b>	60.8	-	-	<b>-15.0</b>	66.8
NMIJ	<b>20.0</b>	35.1	<b>15.0</b>	66.8	-	-

Table 40: Degrees of equivalence between the participants for vertical voltage sensitivities of the SE-1021 at 4 Hz

4.0 Hz	NIM		NIMT		NMIJ	
$i \rightarrow$	$D_{ij}$	$U_{ij}$	$D_{ij}$	$U_{ij}$	$D_{ij}$	$U_{ij}$
$j \downarrow$	(mV/(m/s <sup>2</sup> )•10 <sup>-2</sup> )		(mV/(m/s <sup>2</sup> )•10 <sup>-2</sup> )		(mV/(m/s <sup>2</sup> )•10 <sup>-2</sup> )	
NIM	-	-	<b>-4.0</b>	60.8	<b>-17.0</b>	35.1
NIMT	<b>4.0</b>	60.8	-	-	<b>-13.0</b>	66.8
NMIJ	<b>17.0</b>	35.1	<b>13.0</b>	66.8	-	-

Table 41: Degrees of equivalence between the participants for vertical voltage sensitivities of the SE-1021 at 5 Hz

5.0 Hz	NIM		NIMT		NMIJ	
$i \rightarrow$	$D_{ij}$	$U_{ij}$	$D_{ij}$	$U_{ij}$	$D_{ij}$	$U_{ij}$
$j \downarrow$	(mV/(m/s <sup>2</sup> )•10 <sup>-2</sup> )		(mV/(m/s <sup>2</sup> )•10 <sup>-2</sup> )		(mV/(m/s <sup>2</sup> )•10 <sup>-2</sup> )	
NIM	-	-	<b>-2.0</b>	48.5	<b>-17.0</b>	35.7
NIMT	<b>2.0</b>	48.5	-	-	<b>-15.0</b>	55.9
NMIJ	<b>17.0</b>	35.7	<b>15.0</b>	55.9	-	-

Table 42: Degrees of equivalence between the participants for vertical voltage sensitivities of the SE-1021 at 6.3 Hz

6.3 Hz	NIM		NIMT		NMIJ	
$i \rightarrow$	$D_{ij}$	$U_{ij}$	$D_{ij}$	$U_{ij}$	$D_{ij}$	$U_{ij}$
$j \downarrow$	(mV/(m/s <sup>2</sup> )•10 <sup>-2</sup> )		(mV/(m/s <sup>2</sup> )•10 <sup>-2</sup> )		(mV/(m/s <sup>2</sup> )•10 <sup>-2</sup> )	
NIM	-	-	<b>-1.0</b>	48.5	<b>-13.0</b>	35.7
NIMT	<b>1.0</b>	48.5	-	-	<b>-12.0</b>	55.9
NMIJ	<b>13.0</b>	35.7	<b>12.0</b>	55.9	-	-

Table 43: Degrees of equivalence between the participants for vertical voltage sensitivities of the SE-1021 at 8 Hz

8.0 Hz	NIM		NIMT		NMIJ	
$i \rightarrow$	$D_{ij}$	$U_{ij}$	$D_{ij}$	$U_{ij}$	$D_{ij}$	$U_{ij}$
$j \downarrow$	(mV/(m/s <sup>2</sup> )•10 <sup>-2</sup> )		(mV/(m/s <sup>2</sup> )•10 <sup>-2</sup> )		(mV/(m/s <sup>2</sup> )•10 <sup>-2</sup> )	
NIM	-	-	<b>-2.0</b>	48.5	<b>-13.0</b>	35.7
NIMT	<b>2.0</b>	48.5	-	-	<b>-11.0</b>	55.9
NMIJ	<b>13.0</b>	35.7	<b>11.0</b>	55.9	-	-

Table 44: Degrees of equivalence between the participants for vertical voltage sensitivities of the SE-1021 at 10 Hz

10.0 Hz	NIM		NIMT		NMIJ	
$i \rightarrow$	$D_{ij}$	$U_{ij}$	$D_{ij}$	$U_{ij}$	$D_{ij}$	$U_{ij}$
$j \downarrow$	(mV/(m/s <sup>2</sup> )•10 <sup>-2</sup> )		(mV/(m/s <sup>2</sup> )•10 <sup>-2</sup> )		(mV/(m/s <sup>2</sup> )•10 <sup>-2</sup> )	
NIM	-	-	<b>-1.0</b>	48.5	<b>-11.0</b>	35.7
NIMT	<b>1.0</b>	48.5	-	-	<b>-10.0</b>	55.9
NMIJ	<b>11.0</b>	35.7	<b>10.0</b>	55.9	-	-

Table 45: Degrees of equivalence between the participants for vertical voltage sensitivities of the SE-1021 at 12.5 Hz

12.5 Hz	NIM		NIMT		NMIJ	
$i \rightarrow$	$D_{ij}$	$U_{ij}$	$D_{ij}$	$U_{ij}$	$D_{ij}$	$U_{ij}$
$j \downarrow$	(mV/(m/s <sup>2</sup> )•10 <sup>-2</sup> )		(mV/(m/s <sup>2</sup> )•10 <sup>-2</sup> )		(mV/(m/s <sup>2</sup> )•10 <sup>-2</sup> )	
NIM	-	-	<b>0.0</b>	48.6	<b>-9.0</b>	35.7
NIMT	<b>0.0</b>	48.6	-	-	<b>-9.0</b>	55.9
NMIJ	<b>9.0</b>	35.7	<b>9.0</b>	55.9	-	-

Table 46: Degrees of equivalence between the participants for vertical voltage sensitivities of the SE-1021 at 16 Hz

16.0 Hz	NIM		NIMT		NMIJ	
$i \rightarrow$	$D_{ij}$	$U_{ij}$	$D_{ij}$	$U_{ij}$	$D_{ij}$	$U_{ij}$
$j \downarrow$	(mV/(m/s <sup>2</sup> )•10 <sup>-2</sup> )		(mV/(m/s <sup>2</sup> )•10 <sup>-2</sup> )		(mV/(m/s <sup>2</sup> )•10 <sup>-2</sup> )	
NIM	-	-	<b>-2.0</b>	48.6	<b>-10.0</b>	35.7
NIMT	<b>2.0</b>	48.6	-	-	<b>-8.0</b>	56.0
NMIJ	<b>10.0</b>	35.7	<b>8.0</b>	56.0	-	-

Table 47: Degrees of equivalence between the participants for vertical voltage sensitivities of the SE-1021 at 20 Hz

20.0 Hz	NIM		NIMT		NMIJ	
$i \rightarrow$	$D_{ij}$	$U_{ij}$	$D_{ij}$	$U_{ij}$	$D_{ij}$	$U_{ij}$
$j \downarrow$	(mV/(m/s <sup>2</sup> )•10 <sup>-2</sup> )		(mV/(m/s <sup>2</sup> )•10 <sup>-2</sup> )		(mV/(m/s <sup>2</sup> )•10 <sup>-2</sup> )	
NIM	-	-	<b>-1.0</b>	48.6	<b>-8.0</b>	35.8
NIMT	<b>1.0</b>	48.6	-	-	<b>-7.0</b>	56.0
NMIJ	<b>8.0</b>	35.8	<b>7.0</b>	56.0	-	-

### 7.4.3 Tables of DoE between participants for horizontal voltage sensitivities of the SE-1022

Table 48: Degrees of equivalence between the participants for horizontal voltage sensitivities of the SE-1022 at 0.5 Hz

0.50 Hz	NIM		CMS		NMIA		NMISA		NMIJ		KRISS	
$i \rightarrow$	$D_{ij}$	$U_{ij}$	$D_{ij}$	$U_{ij}$	$D_{ij}$	$U_{ij}$	$D_{ij}$	$U_{ij}$	$D_{ij}$	$U_{ij}$	$D_{ij}$	$U_{ij}$
$j \downarrow$	(mV/(m/s <sup>2</sup> )•10 <sup>-2</sup> )		(mV/(m/s <sup>2</sup> )•10 <sup>-2</sup> )		(mV/(m/s <sup>2</sup> )•10 <sup>-2</sup> )		(mV/(m/s <sup>2</sup> )•10 <sup>-2</sup> )		(mV/(m/s <sup>2</sup> )•10 <sup>-2</sup> )		(mV/(m/s <sup>2</sup> )•10 <sup>-2</sup> )	
NIM	-	-	<b>70.0</b>	325.0	<b>-26.0</b>	27.6	<b>-40.0</b>	<b>39.1</b>	<b>-18.0</b>	248.3	<b>-8.0</b>	63.2
CMS	<b>-70.0</b>	325.0	-	-	<b>-96.0</b>	325.0	<b>-110.0</b>	326.1	<b>-88.0</b>	408.1	<b>-78.0</b>	329.9
NMIA	<b>26.0</b>	27.6	<b>96.0</b>	325.0	-	-	<b>-14.0</b>	39.0	<b>8.0</b>	248.3	<b>18.0</b>	63.2
NMISA	<b>40.0</b>	<b>39.1</b>	<b>110.0</b>	326.1	<b>14.0</b>	39.0	-	-	<b>22.0</b>	249.9	<b>32.0</b>	69.0
NMIJ	<b>18.0</b>	248.3	<b>88.0</b>	408.1	<b>-8.0</b>	248.3	<b>-22.0</b>	249.9	-	-	<b>10.0</b>	254.8
KRISS	<b>8.0</b>	63.2	<b>78.0</b>	329.9	<b>-18.0</b>	63.2	<b>-32.0</b>	69.0	<b>-10.0</b>	254.8	-	-

Table 49: Degrees of equivalence between the participants for horizontal voltage sensitivities of the SE-1022 at 0.63 Hz

0.63 Hz	NIM		CMS		NMIA		NMISA		NMIJ		KRISS	
$i \rightarrow$	$D_{ij}$	$U_{ij}$										
$j \downarrow$	(mV/(m/s <sup>2</sup> )•10 <sup>-2</sup> )											
NIM	-	-	<b>77.0</b>	337.5	<b>-20.0</b>	27.8	<b>-33.0</b>	39.2	<b>-50.0</b>	87.5	<b>-12.0</b>	63.3
CMS	<b>-77.0</b>	337.5	-	-	<b>-97.0</b>	337.5	<b>-110.0</b>	338.6	<b>-127.0</b>	347.5	<b>-89.0</b>	342.2
NMIA	<b>20.0</b>	27.8	<b>97.0</b>	337.5	-	-	<b>-13.0</b>	39.2	<b>-30.0</b>	87.4	<b>8.0</b>	63.3
NMISA	<b>33.0</b>	39.2	<b>110.0</b>	338.6	<b>13.0</b>	39.2	-	-	<b>-17.0</b>	91.7	<b>21.0</b>	69.0
NMIJ	<b>50.0</b>	87.5	<b>127.0</b>	347.5	<b>30.0</b>	87.4	<b>17.0</b>	91.7	-	-	<b>38.0</b>	104.3
KRISS	<b>12.0</b>	63.3	<b>89.0</b>	342.2	<b>-8.0</b>	63.3	<b>-21.0</b>	69.0	<b>-38.0</b>	104.3	-	-

Table 50: Degrees of equivalence between the participants for horizontal voltage sensitivities of the SE-1022 at 0.8 Hz

0.8 Hz	NIM		CMS		NMIA		NMISA		NMIJ		KRISS	
$i \rightarrow$	$D_{ij}$	$U_{ij}$										
$j \downarrow$	(mV/(m/s <sup>2</sup> )•10 <sup>-2</sup> )											
NIM	-	-	<b>71.0</b>	312.3	<b>-18.0</b>	28.5	<b>-29.0</b>	39.7	<b>-28.0</b>	100.1	<b>-12.0</b>	39.8
CMS	<b>-71.0</b>	312.3	-	-	<b>-89.0</b>	312.3	<b>-100.0</b>	313.5	<b>-99.0</b>	326.7	<b>-83.0</b>	313.5
NMIA	<b>18.0</b>	28.5	<b>89.0</b>	312.3	-	-	<b>-11.0</b>	39.7	<b>-10.0</b>	100.0	<b>6.0</b>	39.8
NMISA	<b>29.0</b>	39.7	<b>100.0</b>	313.5	<b>11.0</b>	39.7	-	-	<b>1.0</b>	103.8	<b>17.0</b>	48.4
NMIJ	<b>28.0</b>	100.1	<b>99.0</b>	326.7	<b>10.0</b>	100.0	<b>-1.0</b>	103.8	-	-	<b>16.0</b>	103.8
KRISS	<b>12.0</b>	39.8	<b>83.0</b>	313.5	<b>-6.0</b>	39.8	<b>-17.0</b>	48.4	<b>-16.0</b>	103.8	-	-

Table 51: Degrees of equivalence between the participants for horizontal voltage sensitivities of the SE-1022 at 1 Hz

1.0 Hz	NIM		CMS		NMIA		NMISA		NMIJ		KRISS	
$i \rightarrow$	$D_{ij}$	$U_{ij}$										
$j \downarrow$	(mV/(m/s <sup>2</sup> )•10 <sup>-2</sup> )											
NIM	-	-	<b>109.0</b>	288.0	<b>-11.0</b>	28.0	<b>-21.0</b>	39.3	<b>-28.0</b>	124.4	<b>-7.0</b>	51.2
CMS	<b>-109.0</b>	288.0	-	-	<b>-120.0</b>	288.0	<b>-130.0</b>	289.3	<b>-137.0</b>	312.5	<b>-116.0</b>	291.2
NMIA	<b>11.0</b>	28.0	<b>120.0</b>	288.0	-	-	<b>-10.0</b>	39.3	<b>-17.0</b>	124.4	<b>4.0</b>	51.2
NMISA	<b>21.0</b>	39.3	<b>130.0</b>	289.3	<b>10.0</b>	39.3	-	-	<b>-7.0</b>	127.4	<b>14.0</b>	58.2
NMIJ	<b>28.0</b>	124.4	<b>137.0</b>	312.5	<b>17.0</b>	124.4	<b>7.0</b>	127.4	-	-	<b>21.0</b>	131.6
KRISS	<b>7.0</b>	51.2	<b>116.0</b>	291.2	<b>-4.0</b>	51.2	<b>-14.0</b>	58.2	<b>-21.0</b>	131.6	-	-

Table 52: Degrees of equivalence between the participants for horizontal voltage sensitivities of the SE-1022 at 1.25 Hz

1.25 Hz	NIM		CMS		NMIA		NMISA		NMIJ		KRISS	
$i \rightarrow$	$D_{ij}$	$U_{ij}$										
$j \downarrow$	(mV/(m/s <sup>2</sup> )•10 <sup>-2</sup> )											
NIM	-	-	<b>83.0</b>	287.3	<b>-8.0</b>	28.1	<b>-17.0</b>	39.4	<b>-30.0</b>	87.6	<b>-7.0</b>	51.3
CMS	<b>-83.0</b>	287.3	-	-	<b>-91.0</b>	287.3	<b>-100.0</b>	288.7	<b>-113.0</b>	299.1	<b>-90.0</b>	290.5
NMIA	<b>8.0</b>	28.1	<b>91.0</b>	287.3	-	-	<b>-9.0</b>	39.4	<b>-22.0</b>	87.6	<b>1.0</b>	51.3
NMISA	<b>17.0</b>	39.4	<b>100.0</b>	288.7	<b>9.0</b>	39.4	-	-	<b>-13.0</b>	91.8	<b>10.0</b>	58.3
NMIJ	<b>30.0</b>	87.6	<b>113.0</b>	299.1	<b>22.0</b>	87.6	<b>13.0</b>	91.8	-	-	<b>23.0</b>	97.5
KRISS	<b>7.0</b>	51.3	<b>90.0</b>	290.5	<b>-1.0</b>	51.3	<b>-10.0</b>	58.3	<b>-23.0</b>	97.5	-	-

Table 53: Degrees of equivalence between the participants for horizontal voltage sensitivities of the SE-1022 at 1.6 Hz

1.6 Hz	NIM		CMS		NMIA		NMISA		NMIJ		KRISS	
$i \rightarrow$	$D_{ij}$	$U_{ij}$										
$j \downarrow$	(mV/(m/s <sup>2</sup> )•10 <sup>-2</sup> )											
NIM	-	-	<b>101.0</b>	287.8	<b>-11.0</b>	28.5	<b>-19.0</b>	39.7	<b>-31.0</b>	51.4	<b>-7.0</b>	51.5
CMS	<b>-101.0</b>	287.8	-	-	<b>-112.0</b>	287.8	<b>-120.0</b>	289.2	<b>-132.0</b>	291.0	<b>-108.0</b>	291.0
NMIA	<b>11.0</b>	28.5	<b>112.0</b>	287.8	-	-	<b>-8.0</b>	39.7	<b>-20.0</b>	51.4	<b>4.0</b>	51.5
NMISA	<b>19.0</b>	39.7	<b>120.0</b>	289.2	<b>8.0</b>	39.7	-	-	<b>-12.0</b>	58.4	<b>12.0</b>	58.5
NMIJ	<b>31.0</b>	51.4	<b>132.0</b>	291.0	<b>20.0</b>	51.4	<b>12.0</b>	58.4	-	-	<b>24.0</b>	67.0
KRISS	<b>7.0</b>	51.5	<b>108.0</b>	291.0	<b>-4.0</b>	51.5	<b>-12.0</b>	58.5	<b>-24.0</b>	67.0	-	-

Table 54: Degrees of equivalence between the participants for horizontal voltage sensitivities of the SE-1022 at 2 Hz

2.0 Hz	NIM		CMS		NMIA		NMISA		NMIJ		KRISS	
$i \rightarrow$	$D_{ij}$	$U_{ij}$										
$j \downarrow$	(mV/(m/s <sup>2</sup> )•10 <sup>-2</sup> )											
NIM	-	-	<b>71.0</b>	287.2	<b>-11.0</b>	28.6	<b>-19.0</b>	39.8	<b>-43.0</b>	87.7	<b>2.0</b>	39.8
CMS	<b>-71.0</b>	287.2	-	-	<b>-82.0</b>	287.2	<b>-90.0</b>	288.5	<b>-114.0</b>	298.9	<b>-69.0</b>	288.5
NMIA	<b>11.0</b>	28.6	<b>82.0</b>	287.2	-	-	<b>-8.0</b>	39.8	<b>-32.0</b>	87.6	<b>13.0</b>	39.8
NMISA	<b>19.0</b>	39.8	<b>90.0</b>	288.5	<b>8.0</b>	39.8	-	-	<b>-24.0</b>	91.9	<b>21.0</b>	48.5
NMIJ	<b>43.0</b>	87.7	<b>114.0</b>	298.9	<b>32.0</b>	87.6	<b>24.0</b>	91.9	-	-	<b>45.0</b>	91.9
KRISS	<b>-2.0</b>	39.8	<b>69.0</b>	288.5	<b>-13.0</b>	39.8	<b>-21.0</b>	48.5	<b>-45.0</b>	91.9	-	-

Table 55: Degrees of equivalence between the participants for horizontal voltage sensitivities of the SE-1022 at 2.5 Hz

2.5 Hz	NIM		CMS		NMIA		NMISA		NMIJ		KRISS	
$i \rightarrow$	$D_{ij}$	$U_{ij}$										
$j \downarrow$	(mV/(m/s <sup>2</sup> )•10 <sup>-2</sup> )											
NIM	-	-	<b>31.0</b>	273.8	<b>-12.0</b>	28.4	<b>-19.0</b>	39.6	-	-	<b>0.0</b>	39.7
CMS	<b>-31.0</b>	273.8	-	-	<b>-43.0</b>	273.8	<b>-50.0</b>	275.2	-	-	<b>-31.0</b>	275.2
NMIA	<b>12.0</b>	28.4	<b>43.0</b>	273.8	-	-	<b>-7.0</b>	39.6	-	-	<b>12.0</b>	39.7
NMISA	<b>19.0</b>	39.6	<b>50.0</b>	275.2	<b>7.0</b>	39.6	-	-	-	-	<b>19.0</b>	48.4
NMIJ	-	-	-	-	-	-	-	-	-	-	-	-
KRISS	<b>0.0</b>	39.7	<b>31.0</b>	275.2	<b>-12.0</b>	39.7	<b>-19.0</b>	48.4	-	-	-	-

Table 56: Degrees of equivalence between the participants for horizontal voltage sensitivities of the SE-1022 at 3.15 Hz

3.15 Hz	NIM		CMS		NMIA		NMISA		NMIJ		KRISS	
$i \rightarrow$	$D_{ij}$	$U_{ij}$										
$j \downarrow$	(mV/(m/s <sup>2</sup> )•10 <sup>-2</sup> )											
NIM	-	-	<b>90.0</b>	76.3	<b>-13.0</b>	28.6	<b>-10.0</b>	39.8	-	-	<b>-10.0</b>	39.8
CMS	<b>-90.0</b>	76.3	-	-	<b>-103.0</b>	76.3	<b>-100.0</b>	81.2	-	-	<b>-100.0</b>	81.2
NMIA	<b>13.0</b>	28.6	<b>103.0</b>	76.3	-	-	<b>3.0</b>	39.8	-	-	<b>3.0</b>	39.8
NMISA	<b>10.0</b>	39.8	<b>100.0</b>	81.2	<b>-3.0</b>	39.8	-	-	-	-	<b>0.0</b>	48.5
NMIJ	-	-	-	-	-	-	-	-	-	-	-	-
KRISS	<b>10.0</b>	39.8	<b>100.0</b>	81.2	<b>-3.0</b>	39.8	<b>0.0</b>	48.5	-	-	-	-

Table 57: Degrees of equivalence between the participants for horizontal voltage sensitivities of the SE-1022 at 4 Hz

4.0 Hz	NIM		CMS		NMIA		NMISA		NMIJ		KRISS	
$i \rightarrow$	$D_{ij}$	$U_{ij}$										
$j \downarrow$	(mV/(m/s <sup>2</sup> )•10 <sup>-2</sup> )											
NIM	-	-	<b>70.0</b>	76.2	<b>-13.0</b>	28.6	<b>-20.0</b>	39.8	-	-	<b>-14.0</b>	39.8
CMS	<b>-70.0</b>	76.2	-	-	<b>-83.0</b>	76.2	<b>-90.0</b>	81.0	-	-	<b>-84.0</b>	81.1
NMIA	<b>13.0</b>	28.6	<b>83.0</b>	76.2	-	-	<b>-7.0</b>	39.8	-	-	<b>-1.0</b>	39.8
NMISA	<b>20.0</b>	39.8	<b>90.0</b>	81.0	<b>7.0</b>	39.8	-	-	-	-	<b>6.0</b>	48.5
NMIJ	-	-	-	-	-	-	-	-	-	-	-	-
KRISS	<b>14.0</b>	39.8	<b>84.0</b>	81.1	<b>1.0</b>	39.8	<b>-6.0</b>	48.5	-	-	-	-

Table 58: Degrees of equivalence between the participants for horizontal voltage sensitivities of the SE-1022 at 5 Hz

5.0 Hz	NIM		CMS		NMIA		NMISA		NMIJ		KRISS	
$i \rightarrow$	$D_{ij}$	$U_{ij}$										
$j \downarrow$	(mV/(m/s <sup>2</sup> )•10 <sup>-2</sup> )											
NIM	-	-	<b>9.0</b>	63.7	<b>-14.0</b>	28.7	<b>-11.0</b>	39.9	-	-	<b>-18.0</b>	39.9
CMS	<b>-9.0</b>	63.7	-	-	<b>-23.0</b>	63.7	<b>-20.0</b>	69.5	-	-	<b>-27.0</b>	69.5
NMIA	<b>14.0</b>	28.7	<b>23.0</b>	63.7	-	-	<b>3.0</b>	39.9	-	-	<b>-4.0</b>	39.9
NMISA	<b>11.0</b>	39.9	<b>20.0</b>	69.5	<b>-3.0</b>	39.9	-	-	-	-	<b>-7.0</b>	48.6
NMIJ	-	-	-	-	-	-	-	-	-	-	-	-
KRISS	<b>18.0</b>	39.9	<b>27.0</b>	69.5	<b>4.0</b>	39.9	<b>7.0</b>	48.6	-	-	-	-

Table 59: Degrees of equivalence between the participants for horizontal voltage sensitivities of the SE-1022 at 6.3 Hz

6.3 Hz	NIM		CMS		NMIA		NMISA		NMIJ		KRISS	
$i \rightarrow$	$D_{ij}$	$U_{ij}$										
$j \downarrow$	(mV/(m/s <sup>2</sup> )•10 <sup>-2</sup> )											
NIM	-	-	<b>9.0</b>	63.7	<b>-14.0</b>	28.7	<b>-11.0</b>	39.9	-	-	<b>-21.0</b>	39.9
CMS	<b>-9.0</b>	63.7	-	-	<b>-23.0</b>	63.7	<b>-20.0</b>	69.5	-	-	<b>-30.0</b>	69.5
NMIA	<b>14.0</b>	28.7	<b>23.0</b>	63.7	-	-	<b>3.0</b>	39.9	-	-	<b>-7.0</b>	39.9
NMISA	<b>11.0</b>	39.9	<b>20.0</b>	69.5	<b>-3.0</b>	39.9	-	-	-	-	<b>-10.0</b>	48.6
NMIJ	-	-	-	-	-	-	-	-	-	-	-	-
KRISS	<b>21.0</b>	39.9	<b>30.0</b>	69.5	<b>7.0</b>	39.9	<b>10.0</b>	48.6	-	-	-	-

Table 60: Degrees of equivalence between the participants for horizontal voltage sensitivities of the SE-1022 at 8 Hz

8.0 Hz	NIM		CMS		NMIA		NMISA		NMIJ		KRISS	
$i \rightarrow$	$D_{ij}$	$U_{ij}$										
$j \downarrow$	(mV/(m/s <sup>2</sup> )•10 <sup>-2</sup> )											
NIM	-	-	<b>16.0</b>	63.6	<b>-15.0</b>	28.3	<b>-14.0</b>	39.6	-	-	<b>-22.0</b>	51.4
CMS	<b>-16.0</b>	63.6	-	-	<b>-31.0</b>	63.6	<b>-30.0</b>	69.3	-	-	<b>-38.0</b>	76.7
NMIA	<b>15.0</b>	28.3	<b>31.0</b>	63.6	-	-	<b>1.0</b>	39.6	-	-	<b>-7.0</b>	51.3
NMISA	<b>14.0</b>	39.6	<b>30.0</b>	69.3	<b>-1.0</b>	39.6	-	-	-	-	<b>-8.0</b>	58.3
NMIJ	-	-	-	-	-	-	-	-	-	-	-	-
KRISS	<b>22.0</b>	51.4	<b>38.0</b>	76.7	<b>7.0</b>	51.3	<b>8.0</b>	58.3	-	-	-	-

Table 61: Degrees of equivalence between the participants for horizontal voltage sensitivities of the SE-1022 at 10 Hz

10.0 Hz	NIM		CMS		NMIA		NMISA		NMIJ		KRISS	
$i \rightarrow$	$D_{ij}$	$U_{ij}$										
$j \downarrow$	(mV/(m/s <sup>2</sup> )•10 <sup>-2</sup> )											
NIM	-	-	<b>49.0</b>	63.8	<b>-19.0</b>	28.3	<b>-21.0</b>	39.6	-	-	<b>-25.0</b>	51.4
CMS	<b>-49.0</b>	63.8	-	-	<b>-68.0</b>	63.8	<b>-70.0</b>	69.5	-	-	<b>-74.0</b>	76.8
NMIA	<b>19.0</b>	28.3	<b>68.0</b>	63.8	-	-	<b>-2.0</b>	39.6	-	-	<b>-6.0</b>	51.4
NMISA	<b>21.0</b>	39.6	<b>70.0</b>	69.5	<b>2.0</b>	39.6	-	-	-	-	<b>-4.0</b>	58.4
NMIJ	-	-	-	-	-	-	-	-	-	-	-	-
KRISS	<b>25.0</b>	51.4	<b>74.0</b>	76.8	<b>6.0</b>	51.4	<b>4.0</b>	58.4	-	-	-	-

Table 62: Degrees of equivalence between the participants for horizontal voltage sensitivities of the SE-1022 at 12.5 Hz

12.5 Hz	NIM		CMS		NMIA		NMISA		NMIJ		KRISS	
$i \rightarrow$	$D_{ij}$	$U_{ij}$										
$j \downarrow$	(mV/(m/s <sup>2</sup> )•10 <sup>-2</sup> )											
NIM	-	-	<b>29.0</b>	63.7	<b>-15.0</b>	28.3	<b>-21.0</b>	39.6	-	-	<b>-22.0</b>	51.4
CMS	<b>-29.0</b>	63.7	-	-	<b>-44.0</b>	63.7	<b>-50.0</b>	69.4	-	-	<b>-51.0</b>	76.8
NMIA	<b>15.0</b>	28.3	<b>44.0</b>	63.7	-	-	<b>-6.0</b>	39.6	-	-	<b>-7.0</b>	51.4
NMISA	<b>21.0</b>	39.6	<b>50.0</b>	69.4	<b>6.0</b>	39.6	-	-	-	-	<b>-1.0</b>	58.4
NMIJ	-	-	-	-	-	-	-	-	-	-	-	-
KRISS	<b>22.0</b>	51.4	<b>51.0</b>	76.8	<b>7.0</b>	51.4	<b>1.0</b>	58.4	-	-	-	-

Table 63: Degrees of equivalence between the participants for horizontal voltage sensitivities of the SE-1022 at 16 Hz

16.0 Hz	NIM		CMS		NMIA		NMISA		NMIJ		KRISS	
$i \rightarrow$	$D_{ij}$	$U_{ij}$	$D_{ij}$	$U_{ij}$	$D_{ij}$	$U_{ij}$	$D_{ij}$	$U_{ij}$	$D_{ij}$	$U_{ij}$	$D_{ij}$	$U_{ij}$
$j \downarrow$	(mV/(m/s <sup>2</sup> )•10 <sup>-2</sup> )		(mV/(m/s <sup>2</sup> )•10 <sup>-2</sup> )		(mV/(m/s <sup>2</sup> )•10 <sup>-2</sup> )							
NIM	-	-	<b>60.0</b>	63.9	<b>-15.0</b>	28.3	<b>-20.0</b>	39.6	-	-	<b>-17.0</b>	51.4
CMS	<b>-60.0</b>	63.9	-	-	<b>-75.0</b>	63.9	<b>-80.0</b>	<b>69.6</b>	-	-	<b>-77.0</b>	77.0
NMIA	<b>15.0</b>	28.3	<b>75.0</b>	63.9	-	-	<b>-5.0</b>	39.6	-	-	<b>-2.0</b>	51.4
NMISA	<b>20.0</b>	39.6	<b>80.0</b>	69.6	<b>5.0</b>	39.6	-	-	-	-	<b>3.0</b>	58.4
NMIJ	-	-	-	-	-	-	-	-	-	-	-	-
KRISS	<b>17.0</b>	51.4	<b>77.0</b>	77.0	<b>2.0</b>	51.4	<b>-3.0</b>	58.4	-	-	-	-

Table 64: Degrees of equivalence between the participants for horizontal voltage sensitivities of the SE-1022 at 20 Hz

20.0 Hz	NIM		CMS		NMIA		NMISA		NMIJ		KRISS	
$i \rightarrow$	$D_{ij}$	$U_{ij}$										
$j \downarrow$	(mV/(m/s <sup>2</sup> )•10 <sup>-2</sup> )											
NIM	-	-	<b>26.0</b>	63.8	<b>-18.0</b>	28.3	<b>-24.0</b>	39.6	-	-	<b>-25.0</b>	51.5
CMS	<b>-26.0</b>	63.8	-	-	<b>-44.0</b>	63.8	<b>-50.0</b>	69.5	-	-	<b>-51.0</b>	76.9
NMIA	<b>18.0</b>	28.3	<b>44.0</b>	63.8	-	-	<b>-6.0</b>	39.6	-	-	<b>-7.0</b>	51.5
NMISA	<b>24.0</b>	39.6	<b>50.0</b>	69.5	<b>6.0</b>	39.6	-	-	-	-	<b>-1.0</b>	58.5
NMIJ	-	-	-	-	-	-	-	-	-	-	-	-
KRISS	<b>25.0</b>	51.5	<b>51.0</b>	76.9	<b>7.0</b>	51.5	<b>1.0</b>	58.5	-	-	-	-

#### 7.4.4 Tables of DoE between participants for vertical voltage sensitivities of the SE-1022

Table 65: Degrees of equivalence between the participants for vertical voltage sensitivities of the SE-1022 at 0.5 Hz

0.50 Hz	NIM		NIMT		NMIJ	
$i \rightarrow$	$D_{ij}$	$U_{ij}$	$D_{ij}$	$U_{ij}$	$D_{ij}$	$U_{ij}$
$j \downarrow$	(mV/(m/s <sup>2</sup> )•10 <sup>-2</sup> )		(mV/(m/s <sup>2</sup> )•10 <sup>-2</sup> )		(mV/(m/s <sup>2</sup> )•10 <sup>-2</sup> )	
NIM	-	-	<b>-10.0</b>	71.0	-	-
NIMT	<b>10.0</b>	71.0	-	-	-	-
NMIJ	-	-	-	-	-	-

Table 66: Degrees of equivalence between the participants for vertical voltage sensitivities of the SE-1022 at 0.63 Hz

0.63 Hz	NIM		NIMT		NMIJ	
$i \rightarrow$	$D_{ij}$	$U_{ij}$	$D_{ij}$	$U_{ij}$	$D_{ij}$	$U_{ij}$
$j \downarrow$	(mV/(m/s <sup>2</sup> )•10 <sup>-2</sup> )		(mV/(m/s <sup>2</sup> )•10 <sup>-2</sup> )		(mV/(m/s <sup>2</sup> )•10 <sup>-2</sup> )	
NIM	-	-	<b>-12.0</b>	71.0	-	-
NIMT	<b>12.0</b>	71.0	-	-	-	-
NMIJ	-	-	-	-	-	-

Table 67: Degrees of equivalence between the participants for vertical voltage sensitivities of the SE-1022 at 0.8 Hz

0.8 Hz	NIM		NIMT		NMIJ	
$i \rightarrow$	$D_{ij}$	$U_{ij}$	$D_{ij}$	$U_{ij}$	$D_{ij}$	$U_{ij}$
$j \downarrow$	$(\text{mV}/(\text{m/s}^2) \cdot 10^{-2})$		$(\text{mV}/(\text{m/s}^2) \cdot 10^{-2})$		$(\text{mV}/(\text{m/s}^2) \cdot 10^{-2})$	
NIM	-	-	<b>-12.0</b>	71.0	-	-
NIMT	<b>12.0</b>	71.0	-	-	-	-
NMIJ	-	-	-	-	-	-

Table 68: Degrees of equivalence between the participants for vertical voltage sensitivities of the SE-1022 at 1 Hz

1.0 Hz	NIM		NIMT		NMIJ	
$i \rightarrow$	$D_{ij}$	$U_{ij}$	$D_{ij}$	$U_{ij}$	$D_{ij}$	$U_{ij}$
$j \downarrow$	$(\text{mV}/(\text{m/s}^2) \cdot 10^{-2})$		$(\text{mV}/(\text{m/s}^2) \cdot 10^{-2})$		$(\text{mV}/(\text{m/s}^2) \cdot 10^{-2})$	
NIM	-	-	<b>-6.0</b>	58.3	-	-
NIMT	<b>6.0</b>	58.3	-	-	-	-
NMIJ	-	-	-	-	-	-

Table 69: Degrees of equivalence between the participants for vertical voltage sensitivities of the SE-1022 at 1.25 Hz

1.25 Hz	NIM		NIMT		NMIJ	
$i \rightarrow$	$D_{ij}$	$U_{ij}$	$D_{ij}$	$U_{ij}$	$D_{ij}$	$U_{ij}$
$j \downarrow$	$(\text{mV}/(\text{m/s}^2) \cdot 10^{-2})$		$(\text{mV}/(\text{m/s}^2) \cdot 10^{-2})$		$(\text{mV}/(\text{m/s}^2) \cdot 10^{-2})$	
NIM	-	-	<b>-4.0</b>	58.3	-	-
NIMT	<b>4.0</b>	58.3	-	-	-	-
NMIJ	-	-	-	-	-	-

Table 70: Degrees of equivalence between the participants for vertical voltage sensitivities of the SE-1022 at 1.6 Hz

1.6 Hz	NIM		NIMT		NMIJ	
$i \rightarrow$	$D_{ij}$	$U_{ij}$	$D_{ij}$	$U_{ij}$	$D_{ij}$	$U_{ij}$
$j \downarrow$	$(\text{mV}/(\text{m/s}^2) \cdot 10^{-2})$		$(\text{mV}/(\text{m/s}^2) \cdot 10^{-2})$		$(\text{mV}/(\text{m/s}^2) \cdot 10^{-2})$	
NIM	-	-	<b>-4.0</b>	58.3	-	-
NIMT	<b>4.0</b>	58.3	-	-	-	-
NMIJ	-	-	-	-	-	-

Table 71: Degrees of equivalence between the participants for vertical voltage sensitivities of the SE-1022 at 2 Hz

2.0 Hz	NIM		NIMT		NMIJ	
$i \rightarrow$	$D_{ij}$	$U_{ij}$	$D_{ij}$	$U_{ij}$	$D_{ij}$	$U_{ij}$
$j \downarrow$	$(\text{mV}/(\text{m/s}^2) \cdot 10^{-2})$		$(\text{mV}/(\text{m/s}^2) \cdot 10^{-2})$		$(\text{mV}/(\text{m/s}^2) \cdot 10^{-2})$	
NIM	-	-	<b>-5.0</b>	58.3	-	-
NIMT	<b>5.0</b>	58.3	-	-	-	-
NMIJ	-	-	-	-	-	-

Table 72: Degrees of equivalence between the participants for vertical voltage sensitivities of the SE-1022 at 2.5 Hz

2.5 Hz	NIM		NIMT		NMIJ	
$i \rightarrow$	$D_{ij}$	$U_{ij}$	$D_{ij}$	$U_{ij}$	$D_{ij}$	$U_{ij}$
$j \downarrow$	$(\text{mV}/(\text{m/s}^2) \cdot 10^{-2})$		$(\text{mV}/(\text{m/s}^2) \cdot 10^{-2})$		$(\text{mV}/(\text{m/s}^2) \cdot 10^{-2})$	
NIM	-	-	<b>-5.0</b>	60.7	<b>-27.0</b>	35.0
NIMT	<b>5.0</b>	60.7	-	-	<b>-22.0</b>	66.7
NMIJ	<b>27.0</b>	35.0	<b>22.0</b>	66.7	-	-

Table 73: Degrees of equivalence between the participants for vertical voltage sensitivities of the SE-1022 at 3.15 Hz

3.15 Hz	NIM		NIMT		NMIJ	
$i \rightarrow$	$D_{ij}$	$U_{ij}$	$D_{ij}$	$U_{ij}$	$D_{ij}$	$U_{ij}$
$j \downarrow$	(mV/(m/s <sup>2</sup> )•10 <sup>-2</sup> )		(mV/(m/s <sup>2</sup> )•10 <sup>-2</sup> )		(mV/(m/s <sup>2</sup> )•10 <sup>-2</sup> )	
NIM	-	-	<b>-5.0</b>	60.7	<b>-21.0</b>	35.0
NIMT	<b>5.0</b>	60.7	-	-	<b>-16.0</b>	66.7
NMIJ	<b>21.0</b>	35.0	<b>16.0</b>	66.7	-	-

Table 74: Degrees of equivalence between the participants for vertical voltage sensitivities of the SE-1022 at 4 Hz

4.0 Hz	NIM		NIMT		NMIJ	
$i \rightarrow$	$D_{ij}$	$U_{ij}$	$D_{ij}$	$U_{ij}$	$D_{ij}$	$U_{ij}$
$j \downarrow$	(mV/(m/s <sup>2</sup> )•10 <sup>-2</sup> )		(mV/(m/s <sup>2</sup> )•10 <sup>-2</sup> )		(mV/(m/s <sup>2</sup> )•10 <sup>-2</sup> )	
NIM	-	-	<b>-5.0</b>	60.7	<b>-19.0</b>	35.1
NIMT	<b>5.0</b>	60.7	-	-	<b>-14.0</b>	66.7
NMIJ	<b>19.0</b>	35.1	<b>14.0</b>	66.7	-	-

Table 75: Degrees of equivalence between the participants for vertical voltage sensitivities of the SE-1022 at 5 Hz

5.0 Hz	NIM		NIMT		NMIJ	
$i \rightarrow$	$D_{ij}$	$U_{ij}$	$D_{ij}$	$U_{ij}$	$D_{ij}$	$U_{ij}$
$j \downarrow$	(mV/(m/s <sup>2</sup> )•10 <sup>-2</sup> )		(mV/(m/s <sup>2</sup> )•10 <sup>-2</sup> )		(mV/(m/s <sup>2</sup> )•10 <sup>-2</sup> )	
NIM	-	-	<b>-3.0</b>	48.5	<b>-17.0</b>	35.6
NIMT	<b>3.0</b>	48.5	-	-	<b>-14.0</b>	55.8
NMIJ	<b>17.0</b>	35.6	<b>14.0</b>	55.8	-	-

Table 76: Degrees of equivalence between the participants for vertical voltage sensitivities of the SE-1022 at 6.3 Hz

6.3 Hz	NIM		NIMT		NMIJ	
$i \rightarrow$	$D_{ij}$	$U_{ij}$	$D_{ij}$	$U_{ij}$	$D_{ij}$	$U_{ij}$
$j \downarrow$	(mV/(m/s <sup>2</sup> )•10 <sup>-2</sup> )		(mV/(m/s <sup>2</sup> )•10 <sup>-2</sup> )		(mV/(m/s <sup>2</sup> )•10 <sup>-2</sup> )	
NIM	-	-	<b>-2.0</b>	48.5	<b>-14.0</b>	35.6
NIMT	<b>2.0</b>	48.5	-	-	<b>-12.0</b>	55.8
NMIJ	<b>14.0</b>	35.6	<b>12.0</b>	55.8	-	-

Table 77: Degrees of equivalence between the participants for vertical voltage sensitivities of the SE-1022 at 8 Hz

8.0 Hz	NIM		NIMT		NMIJ	
$i \rightarrow$	$D_{ij}$	$U_{ij}$	$D_{ij}$	$U_{ij}$	$D_{ij}$	$U_{ij}$
$j \downarrow$	(mV/(m/s <sup>2</sup> )•10 <sup>-2</sup> )		(mV/(m/s <sup>2</sup> )•10 <sup>-2</sup> )		(mV/(m/s <sup>2</sup> )•10 <sup>-2</sup> )	
NIM	-	-	<b>-2.0</b>	48.5	<b>-12.0</b>	35.6
NIMT	<b>2.0</b>	48.5	-	-	<b>-10.0</b>	55.8
NMIJ	<b>12.0</b>	35.6	<b>10.0</b>	55.8	-	-

Table 78: Degrees of equivalence between the participants for vertical voltage sensitivities of the SE-1022 at 10 Hz

10.0 Hz	NIM		NIMT		NMIJ	
$i \rightarrow$	$D_{ij}$	$U_{ij}$	$D_{ij}$	$U_{ij}$	$D_{ij}$	$U_{ij}$
$j \downarrow$	(mV/(m/s <sup>2</sup> )•10 <sup>-2</sup> )		(mV/(m/s <sup>2</sup> )•10 <sup>-2</sup> )		(mV/(m/s <sup>2</sup> )•10 <sup>-2</sup> )	
NIM	-	-	<b>-1.0</b>	48.5	<b>-11.0</b>	35.7
NIMT	<b>1.0</b>	48.5	-	-	<b>-10.0</b>	55.8
NMIJ	<b>11.0</b>	35.7	<b>10.0</b>	55.8	-	-

Table 79: Degrees of equivalence between the participants for vertical voltage sensitivities of the SE-1022 at 12.5 Hz

12.5 Hz	NIM		NIMT		NMIJ	
$i \rightarrow$	$D_{ij}$	$U_{ij}$	$D_{ij}$	$U_{ij}$	$D_{ij}$	$U_{ij}$
$j \downarrow$	$(\text{mV}/(\text{m/s}^2) \cdot 10^{-2})$		$(\text{mV}/(\text{m/s}^2) \cdot 10^{-2})$		$(\text{mV}/(\text{m/s}^2) \cdot 10^{-2})$	
NIM	-	-	<b>-2.0</b>	48.5	<b>-10.0</b>	35.7
NIMT	<b>2.0</b>	48.5	-	-	<b>-8.0</b>	55.8
NMIJ	<b>10.0</b>	35.7	<b>8.0</b>	55.8	-	-

Table 80: Degrees of equivalence between the participants for vertical voltage sensitivities of the SE-1022 at 16 Hz

16.0 Hz	NIM		NIMT		NMIJ	
$i \rightarrow$	$D_{ij}$	$U_{ij}$	$D_{ij}$	$U_{ij}$	$D_{ij}$	$U_{ij}$
$j \downarrow$	$(\text{mV}/(\text{m/s}^2) \cdot 10^{-2})$		$(\text{mV}/(\text{m/s}^2) \cdot 10^{-2})$		$(\text{mV}/(\text{m/s}^2) \cdot 10^{-2})$	
NIM	-	-	<b>0.0</b>	48.5	<b>-8.0</b>	35.7
NIMT	<b>0.0</b>	48.5	-	-	<b>-8.0</b>	55.9
NMIJ	<b>8.0</b>	35.7	<b>8.0</b>	55.9	-	-

Table 81: Degrees of equivalence between the participants for vertical voltage sensitivities of the SE-1022 at 20 Hz

20.0 Hz	NIM		NIMT		NMIJ	
$i \rightarrow$	$D_{ij}$	$U_{ij}$	$D_{ij}$	$U_{ij}$	$D_{ij}$	$U_{ij}$
$j \downarrow$	$(\text{mV}/(\text{m/s}^2) \cdot 10^{-2})$		$(\text{mV}/(\text{m/s}^2) \cdot 10^{-2})$		$(\text{mV}/(\text{m/s}^2) \cdot 10^{-2})$	
NIM	-	-	<b>1.0</b>	48.6	<b>-7.0</b>	35.7
NIMT	<b>-1.0</b>	48.6	-	-	<b>-8.0</b>	55.9
NMIJ	<b>7.0</b>	35.7	<b>8.0</b>	55.9	-	-

## 8 Conclusion

Seven NMIs measured the voltage sensitivity of two accelerometer standard sets (one set included a quartz-flexure servo accelerometer of single-ended type and a signal conditioner) at 17 frequencies from 0.5 Hz to 20 Hz, in either horizontal or vertical installation direction, except that the pilot laboratory did both directions. The results of the APMP.AUV.V-K3 are four sets of KCRVs, their uncertainties and degrees of equivalence illustrating the

---

performance of the participant laboratories with respect to one another.

The calibration results obtained for the accelerometers represent the current calibration capabilities of the participating laboratories for the voltage sensitivity of single-ended accelerometers. At the reference frequency of 1.6 Hz (specified in ISO 16063-11:1999), the participating laboratories calibrated the transfer standard sets with their claimed relative expanded uncertainty ( $k = 2$ ), the smallest of which equal to 0.2%, i.e. smaller than the limit specified by the ISO standard [4].

In conclusion, the degrees of equivalence calculated from the data submitted by the seven laboratories, support the uncertainty of measurement reported by the seven laboratories for the calibration of the magnitude of the complex sensitivities of accelerometer over the frequency range 0.5 Hz to 20 Hz, with the only exception of CMS at 3.15Hz, 4Hz, 12.5Hz and 16Hz. The completion of APMP.AUV.V-K3 can serve as part of the basis for a planned key comparison targeted at a low-frequency range at CC level.

## 9 Bibliography

- [1] von Martens, H.-J. et al., Final report on key comparison CCAUV.V-K1, 2003, Metrologia, 40, Tech. Suppl. 09001.
- [2] Guidelines for CIPM key comparisons (Appendix F to the “Mutual recognition of national measurements standards and of measurement certificates issued by national metrology institutes” (MRA)). March 1, 1999.
- [3] Technical Protocol of the APMP Key comparison APMP.AUV.V-K3 (Vibration). NIM, Qiao SUN, December 2010.
- [4] ISO 16063-11:1999 “Methods for the calibration of vibration and shock transducers - Part 11: Primary vibration calibration by laser interferometry” .
- [5] ISO/IEC Guide 98-3:2008 “Uncertainty of measurement – Part 3: Guide to the expression of uncertainty in measurement” (GUM: 1995).
- [6] ISO/IEC Guide 98-3:2008 “Suppl. 1:2008 Propagation of distributions using a Monte Carlo method”.

---

## **Appendix A: Technical Protocol**

National Institute of Metrology (NIM)

December 10th, 2010 (participants revised)

Mr. Qiao SUN

### **Technical Protocol of the APMP Key Comparison**

#### **APMP.AUV.V-K3**

### ***1 Task and Purpose of the Comparison***

According to the rules set up by the CIPM MRA the consultative committees of the CIPM have the responsibility to establish ‘degrees of equivalence’ (DoE) between the different measurement standards operated by the national NMIs. This is done by conducting key comparisons (KC) on different levels of the international metrological infrastructure. The previous top level KC in the field of Vibration metrology, CCAUV.V-K1 was completed in the year 2001 in the frequency range from 40 Hz to 5 kHz. The ongoing CCAUV.V-K2 is aimed at frequency range from 10 Hz to 10 kHz.

However, recent developments in technology and improvements at the NMIs have extended the low-frequency vibration limit of calibration capabilities down to 0.5 Hz and even 0.1 Hz. Therefore during the meeting of APMP TCAUV in 2008, the decision was taken to make preparations for a further comparison targeted at a low-frequency range.

In the field of vibration, this regional key comparison is organized in order to compare measurements of sinusoidal linear accelerations in the frequency range from 0.5 Hz to 20 Hz. Moreover, the magnitude of the complex sensitivity calibration and measurement capabilities (CMCs) of the participating laboratories for accelerometer calibration are to be examined and compared. It is the task of the comparison to measure the magnitude of the complex sensitivity of two accelerometer standard sets (one set including a quartz-flexure servo accelerometer of single-ended type and a signal conditioner) at different frequencies with acceleration amplitudes as specified in section 3. The results of this APMP Comparison will, after approval by CCAUV, serve as the foundation at low vibration frequency for the registration of ‘calibration and measurement capabilities’ (CMC) in the framework of the CIPM MRA.

The voltage sensitivity is calculated as the ratio of the amplitude of the accelerometer

---

standard set output voltage to the amplitude of the acceleration at its reference surface. The magnitude of the complex voltage sensitivity shall be given in milli volt per meter per second squared ( $\text{mV}/(\text{m/s}^2)$ ) for the different measurement conditions specified in section 4.

For the calibration of the accelerometer standard sets, laser interferometry in compliance with method 1 or method 3 of the international standard ISO 16063-11:1999 has to be applied, in order to cover the entire frequency range.

The reported sensitivities and associated uncertainties will be used for the calculation of the key comparison reference value.

## ***2 Pilot Laboratory***

Pilot laboratory for this regional key comparison is

Vibration and Shock Section  
Mechanics and Acoustics Metrology Division  
National Institute of Metrology, P.R. China  
BeiSanHuanDongLu 18, ChaoYang District, 100013 Beijing, P.R. China

This is the delivery address for the set of artefacts and the written and signed reports.

Contact Persons are

SUN Qiao	YANG Lifeng
Tel.: +86 10 64524623	Tel.: +86 10 64524606
e-mail: sunq@nim.ac.cn	e-mail: yanglf@nim.ac.cn
Fax: +86 10 64218628	

Co-Pilot laboratory for this regional key comparison is

National Metrology Institute of South Africa  
Private Bag X34, Lynnwood ridge 0040, South Africa

Contact Persons are

---

CS Veldman	ML Temba
Tel.: +27 12 841 4008	Tel.: +27 12 841 4341
e-mail: csveldman@nmisa.org	e-mail: mltemba@nmisa.org
Fax: +27 86 509 0831	

### **3 Device under Test and Measurement Conditions**

For the calibration task of this comparison two quartz-flexure accelerometer sets will be circulated between the participating laboratories. The accelerometer sets are a ‘single ended’ (SE) type, namely a SA 704 (SN: 1022) and a SA 704 (SN: 1022), with one common signal conditioner MSA-I (SN: 02011001).

The accelerometer sets are to be calibrated of their complex voltage sensitivity according to those procedures and conditions implemented by the laboratory in conformance with ISO 16063-11 which provides magnitude information of the artefact. The sensitivities reported shall be for the accelerometer sets, including all effects from the signal conditioner.

The frequency range of the measurements was agreed to be from 0.5 Hz to 20 Hz. Specifically the laboratories are supposed to measure at the following frequencies (all values in Hz).

0.5, 0.63, 0.8, 1, 1.25, 1.6, 2, 2.5, 3.15, 4, 5, 6.3, 8, 10, 12.5, 16, 20.

The measurement conditions should be kept according to the laboratory's standard conditions for calibration of customers' accelerometers for claiming their best measurement capability or CMC where applicable. This presumes that these conditions comply with those defined by the applicable ISO documentary standards [1,2,3], simultaneously.

Specific conditions for the measurements of this comparison are:

- acceleration amplitudes: a range of  $0.1 \text{ m/s}^2$  to  $10 \text{ m/s}^2$  is admissible.
- ambient temperature and accelerometer temperature during the calibration:  
 $(23 \pm 2)^\circ\text{C}$  (actual values to be stated within tolerances of  $\pm 0.3^\circ\text{C}$ ).
- relative humidity: max. 75 % RH

---

## **4 Circulation Type, Schedule and Transportation**

The transducer sets are circulated in a star type fashion with a measurement period of two weeks provided for each participating laboratory and one week for the pilot laboratory. At the beginning and the end of the circulation as well as between certain subsequent measurements of participating laboratories, the transducer sets are measured at the pilot laboratory in order to fix reference values and to monitor the stability of the transducer sets.

The schedule is planned as follows:

Participant	Measurement (calendar week)	Transportation to next Participant (calendar week)
<b>NIM</b>	23/2011	24/2011
<b>NMIJ</b>	25-26/2011	27/2011
<b>KRISS</b>	28-29/2011	30/2011
<b>NIM</b>	31/2011	32/2011
<b>NIMT</b>	33-34/2011	35/2011
<b>CMS</b>	36-37/2011	38/2011
<b>NIM</b>	39/2011	40/2011
<b>NMIA</b>	41-42/2011	43/2011
<b>NMISA</b>	44-45/2011	46/2011
<b>NIM</b>	47/2011	

The cost of transportation to the next participating laboratory shall be covered by the participating laboratory. The transducer sets have to be sent hand-carried with great caution. In case the transducer sets get damaged or lost during transportation, the participating laboratory for delivery should pay 4 000,- € to pilot laboratory for each set.

## **5 Measurement and Analysis Instructions**

The participating laboratories have to observe the following instructions:

- The motion of the quartz-flexure accelerometer should be measured on the moving part of the vertical (preferably) or horizontal vibration exciter, close to the accelerometer's mounting surface, since the mounting (reference) surface is usually not directly

---

accessible.

- The mounting surface of the accelerometer and the moving part of the exciter must be slightly lubricated before mounting.
- The cable between accelerometer and signal conditioner should be taken from the set delivered to the laboratory.
- In order to reduce the influence of non-rectilinear motion, the measurements should be performed for at least three different laser positions which are symmetrically distributed over the respective measurement surface.
- It is advised that the measurement results should be compiled from complete measurement series carried out at different days under nominally the same conditions, except that the accelerometer is remounted and the cable re-attached. The standard deviation of the subsequent measurements should be included in the report.

## ***6 Communication of the Results to Pilot Laboratory***

Each participating laboratory will submit one printed and signed calibration report for each accelerometer set to the pilot laboratory including the following:

- a description of the calibration systems used for the comparison and the mounting techniques for the accelerometer
- a description of the calibration methods used
- documented record of the ambient conditions during measurements
- the calibration results, including the relative expanded measurement uncertainty, and the applied coverage factor for each value
- a detailed uncertainty budget for the system covering all components of measurement uncertainty (calculated according to GUM [4,5]). Including, among others, information on the type of uncertainty (A or B), assumed distribution function and repeatability component.

In addition, the use of the electronic spreadsheets for reporting is mandatory. The consistency between the results in electronic form and the printed and signed calibration report is the responsibility of the participating laboratory. The data submitted in the electronic spreadsheet shall be deemed the official results submitted for the comparison.

The results have to be submitted to the pilot laboratory within six weeks after the measurements have been completed.

---

The pilot laboratory will submit its set of results to the executive secretary of CCAUV in advance to the first measurement of the participating laboratory.

## ***7 Remarks on post processing***

Presuming consistency of the results, the comparison reference value will be calculated according to the established methods agreed upon already for CCAUV.V-K1. The results of this APMP comparison will serve as the foundation of the registration of ‘calibration and measurement capabilities’ (CMC) for low-frequency vibration by the participating NMIs in the framework of the CIPM MRA.

## **References**

- [1] ISO 16063-1:1998 ‘Methods for the calibration of vibration and shock transducers -- Part 1: Basic concepts’
- [2] ISO 16063-11:1999 ‘Methods for the calibration of vibration and shock transducers-- Part 11: Primary vibration calibration by laser interferometry’
- [3] ISO/IEC 17025:2005 ‘General requirements for the competence of testing and calibration laboratories’
- [4] ISO/IEC Guide 98-3:2008 ‘Uncertainty of measurement -- Part 3: Guide to the expression of uncertainty in measurement (GUM:1995)’
- [5] ISO/IEC Guide 98-3:2008/Suppl 1:2008 ‘Propagation of distributions using a Monte Carlo method’

## Appendix B: Uncertainty Budgets of the participants

### NIM (self-developed shakers):

UNCERTAINTY BUDGET MATRIX (UBM)										Certificate No	#REF!	
										Procedure No	AUVW-0001	
Description	Sensitivity calibration (modulus) as per ISO 16063-11 method 3		Make & model: China SA-704			Range:	0.1Hz to 20 Hz			Metrologist		
			Serial number: 1021/1022							YANG Lifeng		
Mathematical Model:		$S = \hat{U}/\hat{a} = \hat{U}/(2\pi f)^2 d$										
Symbol	Input Quantity (Source of Uncertainty)	Estimated Uncertainty	Probability Distribution	k	Divisor factor	Standard Uncertainty	Sensitivity Coefficient	Standard Uncertainty Contribution $Ui(y)$	Reliabilit y	Degrees of Freedom	Remarks	
$u$	▼ Standards and Reference Equipment (Uncorrelated) ▼	( $X_i$ )	Unit	(N, R, T, U)	▼	▼	$U(X_i)$	$C_i$	Unit	%	%	
$\hat{U}_D$	Interferometer output signal disturbance on phase amplitude	0.01	%	Rectangular	2.00	1.73	5.77E-03	1	%	0.006	100	infinite
$\hat{U}_{VD}$	Effect of voltage disturbance on phase amplitude measurement	0.01	%	Rectangular	2.00	1.73	5.77E-03	0.01	%	0.000	100	infinite
$\hat{U}_{MD}$	Effect of motion disturbance on phase amplitude measurement	0.01	%	Rectangular	2.00	1.73	5.77E-03	1	%	0.006	100	infinite
$\hat{U}_{PD}$	Effect of phase disturbance on phase amplitude measurement	0.01	%	Rectangular	2.00	1.73	5.77E-03	1	%	0.006	100	infinite
$\hat{U}_{RE}$	Residual interferometric effects on phase amplitude measurement	0.02	%	Rectangular	2.00	1.73	1.15E-02	1	%	0.012	100	infinite
$f_{pg}$	Vibration frequency measurement accuracy	0.0002	%	Rectangular	2.00	1.73	1.15E-04	1	%	0.000	100	infinite
$h_u$	Uncertainty on laser wavelength measurement	2.50E-11	nm	Normal k = 2	2.00	2.00	1.25E-11	100	%	0.000	100	infinite
$\hat{U}_V$	Accelerometer output voltage measurement (ADC resolution/accuracy)	0.1	%	Rectangular	2.00	1.73	5.77E-02	1	%	0.058	100	infinite
$S_F$	Filtering effect on sensitivity measurement	0	%	Rectangular	2.00	1.73	0.00E+00	1	%	0.000	100	infinite
$G_{CA}$	Charge amplifier gain accuracy	0.00	%	Normal k = 2	2.00	2.00	0.00E+00	1	%	0.000	100	infinite
	Resolution of Standard / Equipment (If applicable)									100		
	▼ Unit Under Test / Calibration (Uncorrelated) ▼											
$\hat{U}_D$	Effect of voltage disturbance on accelerometer output voltage	0.001	%	Rectangular	1.73	5.774E-04	1	%	0.001	100	infinite	$U_{THD} = \frac{1}{2}(d/100)^2$ ; Maximum allowed by ISO
$\hat{U}_T$	Effect of transverse motion on accelerometer output voltage measurement	0.06	%	Rectangular	1.73	3.464E-02	1	%	0.035	100	infinite	Transverse error for a transverse sensitivity of 1%
$\hat{U}_{RES}$	Residual effects on accelerometer output voltage measurement	0.1	%	Normal k = 5	3.00	3.333E-02	1	%	0.033	100	infinite	Tribo-electric effect
$\hat{U}_G$	Standard deviation on accelerometer output voltage measurement	0.16	%	Normal k = 5	3.00	5.333E-02	1	%	0.053	100	infinite	ESDM for sensitivity calculation using 5 cycles min
	Resolution of UUT / Equipment (If applicable)									100		
	Data - Type "B" Evaluation Range of the results (Rectangular)									100		
About UBM	Data - Type "A" Evaluation Exp Std Dev of the Mean (Estd)			Normal k = 1						4	No of Readings	5
TOTAL COMBINED UNCERTAINTY												
Best Measurement Capability (Excluding UUT contribution)				Combined Uncertainty (Normal)	▼ Level of Confidence ▼	0.060	$V_{eff}$	infinite	Checked and Approved By:			
				Expanded Uncertainty	95.45 %	k = 2	0.120	k = 2.00				
Uncertainty of Measurement (Including UUT contribution)				Combined Uncertainty (Normal)	▼ Level of Confidence ▼	0.093	$V_{eff}$	infinite	SUN Qiao			
				Expanded Uncertainty	95.45 %	k = 2	0.2	k = 2.00				

---

## NIMT (APS 113 shaker):

*Uncertainty of Magnitude for Accelerometer SA-704 s/n1021*

Frequency (Hz)	0.5	0.63	0.8	1	1.25	1.6	2	2.5	3.15	4	5	6.3	8	10	12.5	16	20
<b>U1</b>	0.1	0.1	0.1	0.1	0.1	0.1	0.1	0.1	0.1	0.1	0.1	0.1	0.1	0.1	0.1	0.1	0.1
<b>U2</b>	0.03	0.03	0.03	0.03	0.03	0.03	0.03	0.03	0.03	0.03	0.03	0.03	0.03	0.03	0.03	0.03	0.03
<b>U3</b>	0.001	0.001	0.001	0.001	0.001	0.001	0.001	0.001	0.001	0.001	0.001	0.001	0.001	0.001	0.001	0.001	0.001
<b>U4</b>	0.02	0.02	0.02	0.02	0.02	0.02	0.02	0.02	0.02	0.02	0.02	0.02	0.02	0.02	0.02	0.02	0.02
<b>U5</b>	0.1	0.1	0.1	0.1	0.1	0.1	0.1	0.1	0.1	0.1	0.05	0.05	0.05	0.05	0.05	0.05	0.05
<b>U6</b>	0.007	0.007	0.007	0.007	0.007	0.007	0.007	0.007	0.007	0.007	0.007	0.007	0.007	0.007	0.007	0.007	0.007
<b>U7</b>	0.001	0.001	0.001	0.00	0.001	0.001	0.001	0.001	0.001	0.001	0.001	0.001	0.001	0.001	0.001	0.001	0.001
<b>U8</b>	0.01	0.01	0.01	0.01	0.01	0.01	0.01	0.01	0.01	0.01	0.01	0.01	0.01	0.01	0.01	0.01	0.01
<b>U9</b>	0.001	0.001	0.001	0.001	0.001	0.001	0.001	0.001	0.001	0.001	0.001	0.001	0.001	0.001	0.001	0.001	0.001
<b>U10</b>	0.001	0.001	0.001	0.001	0.001	0.001	0.001	0.001	0.001	0.001	0.001	0.001	0.001	0.001	0.001	0.001	0.001
<b>U11</b>	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0
<b>U12</b>	0.25	0.25	0.25	0.2	0.2	0.2	0.18	0.18	0.18	0.18	0.16	0.16	0.16	0.14	0.14	0.14	0.14
<b>U13</b>	0.05	0.05	0.05	0.02	0.02	0.02	0.01	0.01	0.01	0.01	0.01	0.01	0.01	0.001	0.001	0.001	0.001
<b>U14</b>	0.0015	0.0015	0.0015	0.0015	0.0015	0.0015	0.0015	0.0015	0.0015	0.0015	0.0015	0.0015	0.0015	0.0015	0.0015	0.0015	0.0015
<b>U15</b>	0.001	0.001	0.001	0.001	0.001	0.001	0.001	0.001	0.001	0.001	0.001	0.001	0.001	0.001	0.001	0.001	0.001
<b>U16</b>	0.001	0.001	0.001	0.001	0.001	0.001	0.001	0.001	0.001	0.001	0.001	0.001	0.001	0.001	0.001	0.001	0.001
<b>U17</b>	0.05	0.05	0.05	0.05	0.05	0.05	0.05	0.05	0.05	0.05	0.05	0.05	0.05	0.05	0.05	0.05	0.05
<b>U18</b>	0.0018	0.0038	0.0028	0.0018	0.0000	0.0017	0.0013	0.0017	0.0027	0.0000	0.0018	0.0000	0.0013	0.0013	0.0028	0.0032	0.0040
<b>Uc</b>	<b>0.298</b>	<b>0.298</b>	<b>0.298</b>	<b>0.254</b>	<b>0.254</b>	<b>0.254</b>	<b>0.238</b>	<b>0.238</b>	<b>0.238</b>	<b>0.238</b>	<b>0.205</b>	<b>0.205</b>	<b>0.205</b>	<b>0.190</b>	<b>0.190</b>	<b>0.190</b>	<b>0.190</b>
<b>Ue</b>	<b>0.597</b>	<b>0.597</b>	<b>0.597</b>	<b>0.507</b>	<b>0.507</b>	<b>0.507</b>	<b>0.475</b>	<b>0.475</b>	<b>0.475</b>	<b>0.475</b>	<b>0.411</b>	<b>0.411</b>	<b>0.411</b>	<b>0.380</b>	<b>0.380</b>	<b>0.380</b>	<b>0.380</b>

*Uncertainty of Magnitude for Accelerometer SA-704 s/n1022*

Frequency	0.5	0.63	0.8	1	1.25	1.6	2	2.5	3.15	4	5	6.3	8	10	12.5	16	20
<b>U1</b>	0.1	0.1	0.1	0.1	0.1	0.1	0.1	0.1	0.1	0.1	0.1	0.1	0.1	0.1	0.1	0.1	0.1
<b>U2</b>	0.03	0.03	0.03	0.03	0.03	0.03	0.03	0.03	0.03	0.03	0.03	0.03	0.03	0.03	0.03	0.03	0.03
<b>U3</b>	0.001	0.001	0.001	0.001	0.001	0.001	0.001	0.001	0.001	0.001	0.001	0.001	0.001	0.001	0.001	0.001	0.001
<b>U4</b>	0.02	0.02	0.02	0.02	0.02	0.02	0.02	0.02	0.02	0.02	0.02	0.02	0.02	0.02	0.02	0.02	0.02
<b>U5</b>	0.1	0.1	0.1	0.1	0.1	0.1	0.1	0.1	0.1	0.1	0.05	0.05	0.05	0.05	0.05	0.05	0.05
<b>U6</b>	0.007	0.007	0.007	0.007	0.007	0.007	0.007	0.007	0.007	0.007	0.007	0.007	0.007	0.007	0.007	0.007	0.007
<b>U7</b>	0.001	0.001	0.001	0.00	0.001	0.001	0.001	0.001	0.001	0.001	0.001	0.001	0.001	0.001	0.001	0.001	0.001
<b>U8</b>	0.01	0.01	0.01	0.01	0.01	0.01	0.01	0.01	0.01	0.01	0.01	0.01	0.01	0.01	0.01	0.01	0.01
<b>U9</b>	0.001	0.001	0.001	0.001	0.001	0.001	0.001	0.001	0.001	0.001	0.001	0.001	0.001	0.001	0.001	0.001	0.001
<b>U10</b>	0.001	0.001	0.001	0.001	0.001	0.001	0.001	0.001	0.001	0.001	0.001	0.001	0.001	0.001	0.001	0.001	0.001
<b>U11</b>	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0
<b>U12</b>	0.25	0.25	0.25	0.2	0.2	0.2	0.18	0.18	0.18	0.18	0.16	0.16	0.16	0.14	0.14	0.14	0.14
<b>U13</b>	0.05	0.05	0.05	0.02	0.02	0.02	0.01	0.01	0.01	0.01	0.01	0.01	0.01	0.001	0.001	0.001	0.001
<b>U14</b>	0.0015	0.0015	0.0015	0.0015	0.0015	0.0015	0.0015	0.0015	0.0015	0.0015	0.0015	0.0015	0.0015	0.0015	0.0015	0.0015	0.0015
<b>U15</b>	0.001	0.001	0.001	0.001	0.001	0.001	0.001	0.001	0.001	0.001	0.001	0.001	0.001	0.001	0.001	0.001	0.001
<b>U16</b>	0.001	0.001	0.001	0.001	0.001	0.001	0.001	0.001	0.001	0.001	0.001	0.001	0.001	0.001	0.001	0.001	0.001
<b>U17</b>	0.05	0.05	0.05	0.05	0.05	0.05	0.05	0.05	0.05	0.05	0.05	0.05	0.05	0.05	0.05	0.05	0.05
<b>U18</b>	0.0017	0.0038	0.0013	0.0017	0.0018	0.0013	0.0017	0.0013	0.0013	0.0018	0.0000	0.0017	0.0000	0.0013	0.0000	0.0000	0.0018
<b>Uc</b>	<b>0.298</b>	<b>0.298</b>	<b>0.298</b>	<b>0.254</b>	<b>0.254</b>	<b>0.254</b>	<b>0.238</b>	<b>0.238</b>	<b>0.238</b>	<b>0.238</b>	<b>0.205</b>	<b>0.205</b>	<b>0.205</b>	<b>0.190</b>	<b>0.190</b>	<b>0.190</b>	<b>0.190</b>
<b>Ue</b>	<b>0.597</b>	<b>0.597</b>	<b>0.597</b>	<b>0.507</b>	<b>0.507</b>	<b>0.507</b>	<b>0.475</b>	<b>0.475</b>	<b>0.475</b>	<b>0.475</b>	<b>0.411</b>	<b>0.411</b>	<b>0.411</b>	<b>0.380</b>	<b>0.380</b>	<b>0.380</b>	<b>0.380</b>

---

## DETAIL OF UNCERTAINTY BUDGET FOR MAGNITUDE CALIBRATION

Type	Source of Uncertainty	Symbol	Probability Distribution	Divisor	Ci	u(xi)	$\Upsilon_{\text{eff}}$
B	Vibration velocity (uncertainty of tracing back)	$U_1$	normal	2	1	$\frac{U_1}{2}$	$\infty$
B	Voltage $U_x$	$U_2$	normal	2	1	$\frac{U_2}{2}$	$\infty$
B	Angular frequency of $v$ signal	$U_3$	square	$\sqrt{3}$	1	$\frac{U_3}{\sqrt{3}}$	$\infty$
B	Amplifier gain	$U_4$	normal	2	1	$\frac{U_4}{2}$	$\infty$
B	Frequency response	$U_5$	normal	2	1	$\frac{U_5}{2}$	$\infty$
B	Transverse motion	$U_6$	square	$\sqrt{3}$	1	$\frac{U_6}{\sqrt{3}}$	$\infty$
B	Harmonics	$U_7$	square	$\sqrt{3}$	1	$\frac{U_7}{\sqrt{3}}$	$\infty$
B	Hum	$U_8$	normal	2	1	$\frac{U_8}{2}$	$\infty$
B	Noise	$U_9$	normal	2	1	$\frac{U_9}{2}$	$\infty$
B	Effect of geometric location	$U_{10}$	square	$\sqrt{3}$	1	$\frac{U_{10}}{\sqrt{3}}$	$\infty$
B	Sensor attachment	$U_{11}$	square	$\sqrt{3}$	1	$\frac{U_{11}}{\sqrt{3}}$	$\infty$
B	Cable routing and fixing	$U_{12}$	square	$\sqrt{3}$	1	$\frac{U_{12}}{\sqrt{3}}$	$\infty$
B	Relative motion	$U_{13}$	square	$\sqrt{3}$	1	$\frac{U_{13}}{\sqrt{3}}$	$\infty$
B	Temperature change	$U_{14}$	square	$\sqrt{3}$	1	$\frac{U_{14}}{\sqrt{3}}$	$\infty$

B	Linearity	$U_{15}$	square	$\sqrt{3}$	1	$\frac{U_{15}}{\sqrt{3}}$	$\infty$
B	Temporal instability of $v$ signal	$U_{16}$	square	$\sqrt{3}$	1	$\frac{U_{16}}{\sqrt{3}}$	$\infty$
B	Residual effects	$U_{17}$	square	$\sqrt{3}$	1	$\frac{U_{17}}{\sqrt{3}}$	$\infty$
A	Repeatability	$U_{18}$	normal	1	1	$\frac{U_{17}}{\sqrt{n}}$	n
	Combined Uncertainty	$U_c$	normal				
	Expanded Uncertainty	$U_e$	normal ( $k = 2$ )				
	Approved Uncertainty	$U$					

## CMS (APS 113 shaker):

### Uncertainty for 1021

Description	0.5	0.63	0.8	1	1.25	1.6	2	2.5	3.15	4
Uncertainty of accelerometer output voltage measurement (Voltmeter)	0.347314%	0.553592%	0.230044%	0.234321%	0.250953%	0.219515%	0.194243%	0.057703%	0.002667%	0.059548%
Quantization effect of RMS volt meter	0.800000%	0.800000%	0.750000%	0.750000%	0.750000%	0.750000%	0.750000%	0.750000%	0.020000%	0.020000%
Uncertainty of total distortion on accelerometer output voltage measurement	0.600000%	0.600000%	0.600000%	0.600000%	0.600000%	0.600000%	0.600000%	0.600000%	0.100000%	0.100000%
Uncertainty of transverse, rocking and bending acceleration on accelerometer output voltage	0.350000%	0.350000%	0.350000%	0.350000%	0.350000%	0.350000%	0.350000%	0.350000%	0.120000%	0.120000%
Uncertainty of trigger hysteresis on displacement measurement	0.030000%	0.030000%	0.030000%	0.030000%	0.030000%	0.030000%	0.030000%	0.030000%	0.030000%	0.030000%
Uncertainty of displacement measurement (interferometer signal)	0.227217%	0.227217%	0.140582%	0.067315%	0.064880%	0.084854%	0.104261%	0.058465%	0.050267%	0.126004%
Effect of ratio measurement	0.200000%	0.200000%	0.200000%	0.200000%	0.200000%	0.200000%	0.200000%	0.200000%	0.050000%	0.050000%
Uncertainty of laser wave length	0.002000%	0.002000%	0.002000%	0.002000%	0.002000%	0.002000%	0.002000%	0.002000%	0.002000%	0.002000%
Uncertainty of displacement quantization on displacement measurement	0.010200%	0.010200%	0.010200%	0.010200%	0.010200%	0.010200%	0.010200%	0.010200%	0.010200%	0.010200%
Uncertainty of vibration frequency measurement (frequency generator)	0.010000%	0.010000%	0.010000%	0.010000%	0.010000%	0.010000%	0.010000%	0.010000%	0.010000%	0.010000%
Uncertainty of standard voltage generator	0.050000%	0.050000%	0.050000%	0.050000%	0.050000%	0.050000%	0.050000%	0.050000%	0.050000%	0.050000%
Uncertainty of motion disturbance on displacement measurement (relative motion between the accelerometer reference surface and the spot sensed by the interferometer; thermal effect)	0.270000%	0.270000%	0.270000%	0.270000%	0.270000%	0.270000%	0.270000%	0.270000%	0.100000%	0.100000%
other effects on sensitivity measurement (random effect in repeat measurements; experimental standard deviation of arithmetic mean)	0.210000%	0.210000%	0.205000%	0.205000%	0.205000%	0.205000%	0.205000%	0.205000%	0.025000%	0.025000%
<i>Relative combined uncertainty <math>u_{rc}(Su)</math></i>	1.34E-02	1.35E-02	1.18E-02	1.16E-02	1.15E-02	1.14E-02	1.12E-02	1.11E-02	2.57E-03	2.49E-03
<i>Effective degrees of freedom <math>v</math></i>	420	401	318	300	292	279	268	251	514	556
<i>Coverage factor <math>k</math></i>	1.97	1.97	1.97	1.97	1.97	1.97	1.97	1.97	1.96	1.96
<i>Relative expanded uncertainty <math>U_t</math></i>	2.63%	2.65%	2.32%	2.28%	2.27%	2.24%	2.21%	2.18%	0.50%	0.49%

## Uncertainty for 1021

Description	5	6.3	8	10	12.5	16	20	Uncertainty Type
Uncertainty of accelerometer output voltage measurement (Voltmeter)	0.042694%	0.000239%	0.001782%	0.000212%	0.001555%	0.002043%	0.001848%	A
Quantization effect of RMS volt meter	0.020000%	0.020000%	0.020000%	0.020000%	0.020000%	0.020000%	0.020000%	B
Uncertainty of total distortion on accelerometer output voltage measurement	0.100000%	0.100000%	0.100000%	0.100000%	0.100000%	0.100000%	0.100000%	B
Uncertainty of transverse, rocking and bending acceleration on accelerometer output voltage	0.120000%	0.120000%	0.120000%	0.120000%	0.120000%	0.120000%	0.120000%	B
Uncertainty of trigger hysteresis on displacement measurement	0.030000%	0.030000%	0.030000%	0.030000%	0.030000%	0.030000%	0.030000%	B
Uncertainty of displacement measurement (interferometer signal)	0.108349%	0.036147%	0.042910%	0.021321%	0.021089%	0.008314%	0.039861%	A
Effect of ratio measurement	0.050000%	0.050000%	0.050000%	0.050000%	0.050000%	0.050000%	0.050000%	B
Uncertainty of laser wave length	0.002000%	0.002000%	0.002000%	0.002000%	0.002000%	0.002000%	0.002000%	B
Uncertainty of displacement quantization on displacement measurement	0.010200%	0.010200%	0.010200%	0.010200%	0.010200%	0.010200%	0.010200%	B
Uncertainty of vibration frequency measurement (frequency generator)	0.010000%	0.010000%	0.010000%	0.010000%	0.010000%	0.010000%	0.010000%	B
Uncertainty of standard voltage generator	0.050000%	0.050000%	0.050000%	0.050000%	0.050000%	0.050000%	0.050000%	B
Uncertainty of motion disturbance on displacement measurement (relative motion between the accelerometer reference surface and the spot sensed by the interferometer; thermal effect)	0.100000%	0.100000%	0.100000%	0.100000%	0.100000%	0.100000%	0.100000%	B
other effects on sensitivity measurement (random effect in repeat measurements; experimental standard deviation of arithmetic mean)	0.025000%	0.025000%	0.025000%	0.025000%	0.025000%	0.025000%	0.025000%	B
<i>Relative combined uncertainty <math>u_{tc}(Su)</math></i>	2.19E-03	2.13E-03	2.08E-03	2.07E-03	2.05E-03	2.08E-03	2.23E-03	
<i>Effective degrees of freedom <math>v</math></i>	461	418	383	375	364	384	435	
<i>Coverage factor <math>k</math></i>	1.97	1.97	1.97	1.97	1.97	1.97	1.97	
<i>Relative expanded uncertainty <math>U_t</math></i>	0.43%	0.42%	0.41%	0.41%	0.40%	0.41%	0.44%	

## Uncertainty for 1022

Description	0.5	0.63	0.8	1	1.25	1.6	2	2.5	3.15	4
Uncertainty of accelerometer output voltage measurement (Voltmeter)	0.311779%	0.558033%	0.417666%	0.257401%	0.258720%	0.172287%	0.175017%	0.035893%	0.052102%	0.079553%
Quantization effect of RMS volt meter	0.800000%	0.800000%	0.750000%	0.750000%	0.750000%	0.750000%	0.750000%	0.750000%	0.020000%	0.020000%
Uncertainty of total distortion on accelerometer output voltage measurement	0.600000%	0.600000%	0.600000%	0.600000%	0.600000%	0.600000%	0.600000%	0.600000%	0.100000%	0.100000%
Uncertainty of transverse, rocking and bending acceleration on accelerometer output voltage	0.350000%	0.350000%	0.350000%	0.350000%	0.350000%	0.350000%	0.350000%	0.350000%	0.120000%	0.120000%
Uncertainty of trigger hysteresis on displacement measurement	0.030000%	0.030000%	0.030000%	0.030000%	0.030000%	0.030000%	0.030000%	0.030000%	0.030000%	0.030000%
Uncertainty of displacement measurement (interferometer signal)	0.199727%	0.199727%	0.157201%	0.108067%	0.092004%	0.042317%	0.078550%	0.131520%	0.124935%	0.119971%
Effect of ratio measurement	0.200000%	0.200000%	0.200000%	0.200000%	0.200000%	0.200000%	0.200000%	0.200000%	0.050000%	0.050000%
Uncertainty of laser wave length	0.002000%	0.002000%	0.002000%	0.002000%	0.002000%	0.002000%	0.002000%	0.002000%	0.002000%	0.002000%
Uncertainty of displacement quantization on displacement measurement	0.010200%	0.010200%	0.010200%	0.010200%	0.010200%	0.010200%	0.010200%	0.010200%	0.010200%	0.010200%
Uncertainty of vibration frequency measurement (frequency generator)	0.010000%	0.010000%	0.010000%	0.010000%	0.010000%	0.010000%	0.010000%	0.010000%	0.010000%	0.010000%
Uncertainty of standard voltage generator	0.050000%	0.050000%	0.050000%	0.050000%	0.050000%	0.050000%	0.050000%	0.050000%	0.050000%	0.050000%
Uncertainty of motion disturbance on displacement measurement (relative motion between the accelerometer reference surface and the spot sensed by the interferometer; thermal effect)	0.270000%	0.270000%	0.270000%	0.270000%	0.270000%	0.270000%	0.270000%	0.270000%	0.100000%	0.100000%
other effects on sensitivity measurement (random effect in repeat measurements; experimental standard deviation of arithmetic mean)	0.210000%	0.210000%	0.205000%	0.205000%	0.205000%	0.205000%	0.205000%	0.205000%	0.025000%	0.025000%
<i>Relative combined uncertainty <math>u_{rc}(Su)</math></i>	1.32E-02	1.35E-02	1.23E-02	1.17E-02	1.15E-02	1.13E-02	1.13E-02	1.11E-02	2.59E-03	2.91E-03
<i>Effective degrees of freedom <math>v</math></i>	408	403	359	306	291	270	271	255	565	396
<i>Coverage factor <math>k</math></i>	1.97	1.97	1.97	1.97	1.97	1.97	1.97	1.97	1.96	1.97
<i>Relative expanded uncertainty <math>U_t</math></i>	2.60%	2.66%	2.42%	2.30%	2.27%	2.22%	2.22%	2.19%	0.51%	0.57%

## Uncertainty for 1022

Description	5	6.3	8	10	12.5	16	20	Uncertainty Type
Uncertainty of accelerometer output voltage measurement (Voltmeter)	0.058676%	0.001567%	0.000881%	0.000796%	0.001223%	0.000244%	0.001666%	A
Quantization effect of RMS volt meter	0.020000%	0.020000%	0.020000%	0.020000%	0.020000%	0.020000%	0.020000%	B
Uncertainty of total distortion on accelerometer output voltage measurement	0.100000%	0.100000%	0.100000%	0.100000%	0.100000%	0.100000%	0.100000%	B
Uncertainty of transverse, rocking and bending acceleration on accelerometer output voltage	0.120000%	0.120000%	0.120000%	0.120000%	0.120000%	0.120000%	0.120000%	B
Uncertainty of trigger hysteresis on displacement measurement	0.030000%	0.030000%	0.030000%	0.030000%	0.030000%	0.030000%	0.030000%	B
Uncertainty of displacement measurement (interferometer signal)	0.177374%	0.060510%	0.051937%	0.048200%	0.080178%	0.071004%	0.083732%	A
Effect of ratio measurement	0.050000%	0.050000%	0.050000%	0.050000%	0.050000%	0.050000%	0.050000%	B
Uncertainty of laser wave length	0.002000%	0.002000%	0.002000%	0.002000%	0.002000%	0.002000%	0.002000%	B
Uncertainty of displacement quantization on displacement measurement	0.010200%	0.010200%	0.010200%	0.010200%	0.010200%	0.010200%	0.010200%	B
Uncertainty of vibration frequency measurement (frequency generator)	0.010000%	0.010000%	0.010000%	0.010000%	0.010000%	0.010000%	0.010000%	B
Uncertainty of standard voltage generator	0.050000%	0.050000%	0.050000%	0.050000%	0.050000%	0.050000%	0.050000%	B
Uncertainty of motion disturbance on displacement measurement (relative motion between the accelerometer reference surface and the spot sensed by the interferometer; thermal effect)	0.100000%	0.100000%	0.100000%	0.100000%	0.100000%	0.100000%	0.100000%	B
other effects on sensitivity measurement (random effect in repeat measurements; experimental standard deviation of arithmetic mean)	0.025000%	0.025000%	0.025000%	0.025000%	0.025000%	0.025000%	0.025000%	B
<i>Relative combined uncertainty <math>u_{rc}(Su)</math></i>	2.28E-03	2.15E-03	2.13E-03	2.21E-03	2.17E-03	2.21E-03	2.07E-03	
<i>Effective degree of freedom <math>v</math></i>	514	430	411	440	426	434	374	
<i>Coverage factor <math>k</math></i>	1.96	1.97	1.97	1.97	1.97	1.97	1.97	
<i>Relative expanded uncertainty <math>U_t</math></i>	0.45%	0.42%	0.42%	0.43%	0.43%	0.43%	0.41%	

## NMIA (APS 113 shaker):

Primary vibration calibration by laser interferometry (ISO 16063-11:1999, Method 1, Fringe Counting). NMIA

Source of uncertainty	Symbol	Distribution	Factor	Sensitivity coefficient	Degrees of freedom	%
Accelerometer output voltage measurement (voltmeter)	u1	normal	1	1	30	0.0500
Effect of total distortion on accelerometer output voltage measurement	u2	rectangular	0.58	1	10	0.0100
Effect of spurious sensitivities on accelerometer output voltage measurement	u3	rectangular	0.58	1	10	0.0208
Effect of displacement quantization on displacement measurement	u4	rectangular	0.58	1	30	0.0015
Effect of trigger hysteresis of fringe counter on displacement measurement	u5	rectangular	0.58	1	10	0.0012
Filtering effect on displacement measurement (freq band limitation)	u6	rectangular	0.58	1	30	0.0000
Effect of voltage disturbance on displacement measurement (e.g. random noise in photoelectric measuring chain)	u7	rectangular	0.58	1	30	0.0058
Effect of motion disturbance on displacement measurements (e.g. total distortion, relative motion between accelerometer ref surface and spot sensed by interferometer)	u8	rectangular	0.58	1	10	0.0064
Effect of phase disturbance of displacement measurement (e.g. phase noise of interferometer signal)	u9	rectangular	0.58	1	30	0.0000
function)	u10	rectangular	0.58	1	30	0.0022
Vibration frequency measurement (frequency generator and indicator counter) measurements, ESDM, including any effects of the signal conditioning instrument	u11	rectangular	0.58	1	30	0.0006
Effect of motion exciter mode	u12	normal	1	1	26	0.0600
Rounding error	u13	rectangular	0.58	1	10	0.0462
	u14	rectangular	0.58	1	100	0.0023

Primary vibration calibration by laser interferometry (ISO 16063-11:1999, Method 3, Sine Approximation), NMIA

Source of uncertainty	Symbol	Distribution	Factor	Sensitivity coefficient	Degrees of freedom	%
Accelerometer output voltage measurement (voltmeter)	u1	normal	1	1	30	0.0500
Voltage filtering effect on accelerometer output voltage measurement	u2	rectangular	0.58	1	10	0.0000
Effect of voltage disturbance on accelerometer output voltage measurement	u3	rectangular	0.58	1	10	0.0000
Effect of transverse, rocking and bending acceleration on accelerometer output voltage measurement	u4	rectangular	0.58	1	30	0.0208
Effect of interferometer quadrature output signal disturbance on phase amplitude measurement	u5	rectangular	1	1	10	0.0300
Interferometer signal filtering effect on phase amplitude measurement	u6	rectangular	0.58	1	30	0.0000
Effect of voltage disturbance phase amplitude measurement	u7	rectangular	0.58	1	30	0.0058
Effect of motion disturbance on phase amplitude measurements (e.g. total distortion, relative motion between accelerometer ref surface and spot sensed by interferometer)	u8	rectangular	0.58	1	10	0.0064
Effect of phase disturbance of phase amplitude measurement (e.g. phase noise of interferometer signal function)	u9	rectangular	0.58	1	30	0.0000
Vibration frequency measurement (frequency generator and indicator counter) measurements, ESDM, including any effects of the signal conditioning instrument	u10	rectangular	0.58	1	30	0.0022
Effect of motion exciter mode	u11	rectangular	0.58	1	30	0.0006
Rounding error	u12	normal	1	1	26	0.0300
	u13	rectangular	0.58	1	10	0.0115
	u14	rectangular	0.58	1	100	0.0023
Component	Uncertainty	Degrees of freedom				
Quadrature sum of all correlated sources, Fringe Counting	0.05579%	42				
Quadrature sum of all correlated sources, Sine Approximation	0.05489%	42				
Quadrature sum of all uncorrelated sources, Fringe Counting	0.07578%	35				
Quadrature sum of all uncorrelated sources, Sine Approximation	0.04403%	33				
uc2	0.006903% <sup>2</sup>					
uc	0.08309%	32				
k uc (k=2)	0.17%					
<b>U95, rounded up</b>	<b>0.2%</b>					

## NMISA (APS 113 shaker):

UNCERTAINTY BUDGET MATRIX (UBM)										Certificate No	AVVS-2778		
										Procedure No	AVVS-0001		
Reference: Guide to the Expression of Uncertainty in Measurement, Issued by BIPM, IEC, IFCC, ISO, IUPAC, IUPAP, OIML - ISO 1995 (ISBN 92-67-10189-9)													
Description	Sensitivity calibration (modulus) as per ISO 16063-11 method 3						Make & model: Serial number:	China SA-704 1021/1022		Range:	0.4 Hz to 50 Hz		
Mathematical Model:	$S = \hat{u}/\hat{a} = \hat{u}/(2\pi f)^2 d$						Metrologist						
Metrologist	Ian Veldman												
Symbol	Input Quantity (Source of Uncertainty)		Estimated Uncertainty	Probability Distribution	k	Divisor factor	Standard Uncertainty	Sensitivity Coefficient	Standard Uncertainty Contribution $U_i(y)$	Reliability	Degrees of Freedom		
$u$	▼ Standards and Reference Equipment (Uncorrelated) ▼		(X <sub>i</sub> )	Unit	(N, R, T, U)	▼	▼	$U(X_i)$	C <sub>i</sub>	Unit	%		
$\hat{u}_D$	Interferometer output signal disturbance on phase amplitude		0.01	%	Rectangular	2.00	1.73	5.77E-03	1	%	0.006		
$\hat{u}_{VD}$	Effect of voltage disturbance on phase amplitude measurement		0.01	%	Rectangular	2.00	1.73	5.77E-03	0.01	%	0.000		
$\hat{u}_{MD}$	Effect of motion disturbance on phase amplitude measurement		0.015	%	Rectangular	2.00	1.73	8.66E-03	1	%	0.009		
$\hat{u}_{PD}$	Effect of phase disturbance on phase amplitude measurement		0.01	%	Rectangular	2.00	1.73	5.77E-03	1	%	0.006		
$\hat{u}_{RE}$	Residual interferometric effects on phase amplitude measurement		0.01	%	Rectangular	2.00	1.73	5.77E-03	1	%	0.006		
$f_{f_0}$	Vibration frequency measurement accuracy		0.05	%	Rectangular	2.00	1.73	2.89E-02	1	%	0.029		
$A_U$	Uncertainty on laser wavelength measurement		2.50E-11	nm	Normal k = 2	2.00	1.25E-11	100	%	0.000	100		
$\hat{u}_V$	Accelerometer output voltage measurement (ADC resolution/accuracy)		0.15	%	Rectangular	2.00	1.73	8.66E-02	1	%	0.087		
$S_F$	Filtering effect on sensitivity measurement		0.11	%	Rectangular	2.00	1.73	6.35E-02	1	%	0.064		
$G_{CA}$	Conditioning amplifier gain accuracy		0.00	%	Normal k = 2	2.00	0.00E+00	1	%	0.000	100		
	Resolution of Standard / Equipment (If applicable)									100			
	▼ Unit Under Test / Calibration (Uncorrelated) ▼												
$\hat{u}_D$	Effect of voltage disturbance on accelerometer output voltage		0.005	%	Rectangular	2.00	1.73	2.887E-03	1	%	0.003		
$\hat{u}_T$	Effect of transverse motion on accelerometer output voltage measurement		0.1	%	Rectangular	2.00	5.774E-02	1	%	0.058	100		
$\hat{u}_{RES}$	Residual effects on accelerometer output voltage measurement		0.1	%	Normal k = 2	2.00	5.000E-02	1	%	0.050	100		
$\hat{u}_s$	Standard deviation on accelerometer output voltage measurement		0.2	%	Normal k = 2	2.00	1.000E-01	1	%	0.100	100		
	Resolution of UUT / Equipment (If applicable)									100			
	Data - Type "B" Evaluation Range of the results (Rectangular)									100			
	Data - Type "A" Evaluation Exp Std Dev of the Mean (ESDM)				Normal k = 1					4	No of Readings 5		
About UBM	TOTAL COMBINED UNCERTAINTY										%		
Best Measurement Capability (Excluding UUT contribution)				Combined Uncertainty (Normal)		▼ Level of Confidence ▼		0.112	V <sub>eff</sub>	infinite	Checked and Approved By:		
				Expanded Uncertainty		95.46 %	k = 2	0.224	k =	2.00			
Uncertainty of Measurement (Including UUT contribution)				Combined Uncertainty (Normal)		▼ Level of Confidence ▼		0.188	V <sub>eff</sub>	infinite			
				Expanded Uncertainty		95.46 %	k = 2	0.3	k =	2.00			

## NMIJ (self-developed shakers):

DUT		Pickup : SA704 (SN: 1021) Signal Conditioner: MSA-I (SN: 02011001)																			
Frequency Hz	Calibration system	Method	Acceleration	Sensitivity	Expanded uncertainty ( $k=2$ )	Combined uncertainty	Expanded uncertainty ( $k=2$ )	Effect of internal calibration of voltmeter using voltage standard				Voltage standard uncertainty									
								%	%	Type	Distribution	DOF	%	Type	Distribution	DOF					
0.5	Ultra low	FCM	0.5	124.81	2.41	0.97	1.93	1.3E-01	A	Normal	9	5.0E-02	B	Normal	$\infty$						
0.63	Ultra low	FCM	1	123.93	0.92	0.37	0.74	8.6E-02	A	Normal	9	5.0E-02	B	Normal	$\infty$						
0.8	Ultra low	FCM	1	123.85	1.03	0.42	0.83	8.6E-02	A	Normal	9	5.0E-02	B	Normal	$\infty$						
1	Ultra low	FCM	1	123.91	1.13	0.46	0.91	8.6E-02	A	Normal	9	5.0E-02	B	Normal	$\infty$						
1.25	Ultra low	FCM	2	123.76	0.83	0.33	0.67	2.8E-02	A	Normal	9	5.0E-02	B	Normal	$\infty$						
1.6	Ultra low	FCM	2	123.79	0.53	0.22	0.43	2.8E-02	A	Normal	9	5.0E-02	B	Normal	$\infty$						
2	Ultra low	FCM	2	123.88	0.79	0.32	0.64	2.8E-02	A	Normal	9	5.0E-02	B	Normal	$\infty$						
2.5	Low	FCM	5	123.86	0.36	0.14	0.29	5.1E-03	A	Normal	9	5.0E-02	B	Normal	$\infty$						
3.15	Low	FCM	5	123.91	0.26	0.11	0.21	5.2E-03	A	Normal	9	5.0E-02	B	Normal	$\infty$						
4	Low	FCM	5	123.94	0.26	0.10	0.21	5.3E-03	A	Normal	9	5.0E-02	B	Normal	$\infty$						
5	Low	FCM	10	123.95	0.40	0.16	0.33	2.1E-03	A	Normal	9	5.0E-02	B	Normal	$\infty$						
6.3	Low	FCM	10	123.98	0.33	0.13	0.27	2.1E-03	A	Normal	9	5.0E-02	B	Normal	$\infty$						
8	Low	FCM	10	124.01	0.30	0.12	0.24	2.1E-03	A	Normal	9	5.0E-02	B	Normal	$\infty$						
10	Low	FCM	10	124.05	0.28	0.11	0.23	2.1E-03	A	Normal	9	5.0E-02	B	Normal	$\infty$						
12.5	Low	FCM	10	124.11	0.28	0.11	0.22	2.1E-03	A	Normal	9	5.0E-02	B	Normal	$\infty$						
16	Low	FCM	10	124.20	0.27	0.11	0.22	2.1E-03	A	Normal	9	5.0E-02	B	Normal	$\infty$						
20	Low	FCM	10	124.34	0.27	0.11	0.22	2.3E-03	A	Normal	9	5.0E-02	B	Normal	$\infty$						

Effect of voltage quantization on measurement by voltmeter				Effect of transverse sensitivity on accelerometer output voltage measurement				Effect of total distortion of accelerometer output voltage				Repeatability of sensitivity				Effect of laser wavelength instability			
%	Type	Distribution	DOF	%	Type	Distribution	DOF	%	Type	Distribution	DOF	%	Type	Distribution	DOF	%	Type	Distribution	DOF
2.9E-05	B	Rectangular	$\infty$	4.0E-03	B	Normal	$\infty$	1.9E-03	B	Normal	$\infty$	9.5E-01	A	Normal	20	1.6E-03	B	Normal	$\infty$
2.9E-05	B	Rectangular	$\infty$	4.0E-03	B	Normal	$\infty$	1.9E-03	B	Normal	$\infty$	3.4E-01	A	Normal	20	1.6E-03	B	Normal	$\infty$
2.9E-05	B	Rectangular	$\infty$	4.0E-03	B	Normal	$\infty$	9.1E-04	B	Normal	$\infty$	4.0E-01	A	Normal	20	1.6E-03	B	Normal	$\infty$
2.9E-05	B	Rectangular	$\infty$	4.0E-03	B	Normal	$\infty$	1.1E-03	B	Normal	$\infty$	4.4E-01	A	Normal	20	1.6E-03	B	Normal	$\infty$
2.9E-04	B	Rectangular	$\infty$	4.0E-03	B	Normal	$\infty$	1.2E-02	B	Normal	$\infty$	3.0E-01	A	Normal	20	1.6E-03	B	Normal	$\infty$
2.9E-04	B	Rectangular	$\infty$	4.0E-03	B	Normal	$\infty$	6.5E-03	B	Normal	$\infty$	1.8E-01	A	Normal	20	1.6E-03	B	Normal	$\infty$
2.9E-04	B	Rectangular	$\infty$	4.0E-03	B	Normal	$\infty$	5.5E-03	B	Normal	$\infty$	3.0E-01	A	Normal	20	1.6E-03	B	Normal	$\infty$
8.2E-05	B	Rectangular	$\infty$	1.5E-02	B	Normal	$\infty$	1.4E-02	B	Normal	$\infty$	1.5E-02	A	Normal	29	1.6E-03	B	Normal	$\infty$
8.2E-05	B	Rectangular	$\infty$	1.4E-02	B	Normal	$\infty$	5.0E-03	B	Normal	$\infty$	6.8E-03	A	Normal	29	1.6E-03	B	Normal	$\infty$
8.2E-05	B	Rectangular	$\infty$	1.3E-02	B	Normal	$\infty$	2.6E-03	B	Normal	$\infty$	1.2E-02	A	Normal	29	1.6E-03	B	Normal	$\infty$
4.1E-05	B	Rectangular	$\infty$	1.2E-02	B	Normal	$\infty$	6.6E-03	B	Normal	$\infty$	1.1E-02	A	Normal	29	1.6E-03	B	Normal	$\infty$
4.1E-05	B	Rectangular	$\infty$	1.1E-02	B	Normal	$\infty$	3.3E-03	B	Normal	$\infty$	5.0E-03	A	Normal	29	1.6E-03	B	Normal	$\infty$
4.1E-05	B	Rectangular	$\infty$	1.0E-02	B	Normal	$\infty$	1.8E-03	B	Normal	$\infty$	6.4E-03	A	Normal	29	1.6E-03	B	Normal	$\infty$
4.1E-05	B	Rectangular	$\infty$	9.0E-03	B	Normal	$\infty$	1.2E-03	B	Normal	$\infty$	5.1E-03	A	Normal	29	1.6E-03	B	Normal	$\infty$
4.1E-05	B	Rectangular	$\infty$	9.0E-03	B	Normal	$\infty$	9.8E-04	B	Normal	$\infty$	4.4E-03	A	Normal	29	1.6E-03	B	Normal	$\infty$
4.1E-05	B	Rectangular	$\infty$	1.1E-02	B	Normal	$\infty$	6.5E-04	B	Normal	$\infty$	3.6E-03	A	Normal	29	1.6E-03	B	Normal	$\infty$
4.1E-05	B	Rectangular	$\infty$	1.0E-02	B	Normal	$\infty$	5.6E-04	B	Normal	$\infty$	1.7E-02	A	Normal	29	1.6E-03	B	Normal	$\infty$

Effect of vibration frequency instability				Effect of displacement quantization on displacement measurement by universal counter				Effect of displacement quantization on measurement by laser interferometer				Effect of displacement distortion on displacement measurement				Effect of remounting accelerometer			
%	Type	Distribution	DOF	%	Type	Distribution	DOF	%	Type	Distribution	DOF	%	Type	Distribution	DOF	%	Type	Distribution	DOF
5.3E-02	B	Normal	$\infty$	6.4E-05	B	Triangular	$\infty$	6.4E-05	B	Triangular	$\infty$	4.2E-02	B	Normal	$\infty$	5.8E-02	B	Normal	$\infty$
5.7E-02	B	Normal	$\infty$	5.1E-05	B	Triangular	$\infty$	5.1E-05	B	Triangular	$\infty$	4.6E-02	B	Normal	$\infty$	5.8E-02	B	Normal	$\infty$
3.7E-02	B	Normal	$\infty$	8.2E-05	B	Triangular	$\infty$	8.2E-05	B	Triangular	$\infty$	2.7E-02	B	Normal	$\infty$	5.8E-02	B	Normal	$\infty$
1.2E-02	B	Normal	$\infty$	1.3E-04	B	Triangular	$\infty$	1.3E-04	B	Triangular	$\infty$	2.3E-02	B	Normal	$\infty$	5.8E-02	B	Normal	$\infty$
3.1E-03	B	Normal	$\infty$	1.0E-04	B	Triangular	$\infty$	1.0E-04	B	Triangular	$\infty$	1.2E-01	B	Normal	$\infty$	5.8E-02	B	Normal	$\infty$
1.6E-03	B	Normal	$\infty$	1.6E-04	B	Triangular	$\infty$	1.6E-04	B	Triangular	$\infty$	9.1E-02	B	Normal	$\infty$	5.8E-02	B	Normal	$\infty$
2.6E-03	B	Normal	$\infty$	2.5E-04	B	Triangular	$\infty$	2.5E-04	B	Triangular	$\infty$	8.4E-02	B	Normal	$\infty$	5.8E-02	B	Normal	$\infty$
3.9E-03	B	Normal	$\infty$	1.6E-04	B	Triangular	$\infty$	1.6E-04	B	Triangular	$\infty$	1.0E-01	B	Normal	$\infty$	8.6E-02	B	Normal	$\infty$
4.4E-03	B	Normal	$\infty$	2.5E-04	B	Triangular	$\infty$	2.5E-04	B	Triangular	$\infty$	3.2E-02	B	Normal	$\infty$	8.6E-02	B	Normal	$\infty$
3.9E-03	B	Normal	$\infty$	4.1E-04	B	Triangular	$\infty$	4.1E-04	B	Triangular	$\infty$	2.2E-02	B	Normal	$\infty$	8.6E-02	B	Normal	$\infty$
5.6E-03	B	Normal	$\infty$	3.2E-04	B	Triangular	$\infty$	3.2E-04	B	Triangular	$\infty$	1.3E-01	B	Normal	$\infty$	8.6E-02	B	Normal	$\infty$
5.3E-03	B	Normal	$\infty$	5.1E-04	B	Triangular	$\infty$	5.1E-04	B	Triangular	$\infty$	9.0E-02	B	Normal	$\infty$	8.6E-02	B	Normal	$\infty$
4.4E-03	B	Normal	$\infty$	8.2E-04	B	Triangular	$\infty$	8.2E-04	B	Triangular	$\infty$	6.7E-02	B	Normal	$\infty$	8.6E-02	B	Normal	$\infty$
4.9E-03	B	Normal	$\infty$	1.3E-03	B	Triangular	$\infty$	1.3E-03	B	Triangular	$\infty$	5.4E-02	B	Normal	$\infty$	8.6E-02	B	Normal	$\infty$
3.7E-04	B	Normal	$\infty$	2.0E-03	B	Triangular	$\infty$	2.0E-03	B	Triangular	$\infty$	4.9E-02	B	Normal	$\infty$	8.6E-02	B	Normal	$\infty$
6.2E-04	B	Normal	$\infty$	3.3E-03	B	Triangular	$\infty$	3.3E-03	B	Triangular	$\infty$	4.0E-02	B	Normal	$\infty$	8.6E-02	B	Normal	$\infty$
4.7E-04	B	Normal	$\infty$	5.1E-03	B	Triangular	$\infty$	5.1E-03	B	Triangular	$\infty$	3.7E-02	B	Normal	$\infty$	8.6E-02	B	Normal	$\infty$

DUT		Pickup : SA704 (SN: 1022)		Signal Conditioner: MSA-I (SN: 02011001)													
Frequency Hz	Calibration system	Method	Acceleration m/s <sup>2</sup>	Sensitivity mV/(m/s <sup>2</sup> )	Expanded uncertainty (k=2) mV/(m/s <sup>2</sup> )	Combined uncertainty %	Expanded uncertainty (k=2) %	Effect of internal calibration of voltmeter using voltage standard				Voltage standard uncertainty					
								%	Type	Distribution	DOF	%	Type	Distribution	DOF		
0.5	Ultra low	FCM	0.5	124.02	2.19	0.88	1.77	1.3E-01	A	Normal	9	5.0E-02	B	Normal	$\infty$		
0.63	Ultra low	FCM	1	123.63	0.81	0.33	0.66	8.6E-02	A	Normal	9	5.0E-02	B	Normal	$\infty$		
0.8	Ultra low	FCM	1	123.81	0.92	0.37	0.74	8.6E-02	A	Normal	9	5.0E-02	B	Normal	$\infty$		
1	Ultra low	FCM	1	123.73	1.34	0.54	1.08	8.6E-02	A	Normal	9	5.0E-02	B	Normal	$\infty$		
1.25	Ultra low	FCM	2	123.67	0.87	0.35	0.70	2.8E-02	A	Normal	9	5.0E-02	B	Normal	$\infty$		
1.6	Ultra low	FCM	2	123.68	0.47	0.19	0.38	2.8E-02	A	Normal	9	5.0E-02	B	Normal	$\infty$		
2	Ultra low	FCM	2	123.56	0.84	0.34	0.68	2.8E-02	A	Normal	9	5.0E-02	B	Normal	$\infty$		
2.5	Low	FCM	5	123.71	0.35	0.14	0.29	5.1E-03	A	Normal	9	5.0E-02	B	Normal	$\infty$		
3.15	Low	FCM	5	123.77	0.26	0.11	0.21	5.2E-03	A	Normal	9	5.0E-02	B	Normal	$\infty$		
4	Low	FCM	5	123.80	0.26	0.10	0.21	5.3E-03	A	Normal	9	5.0E-02	B	Normal	$\infty$		
5	Low	FCM	10	123.82	0.40	0.16	0.33	2.1E-03	A	Normal	9	5.0E-02	B	Normal	$\infty$		
6.3	Low	FCM	10	123.85	0.33	0.13	0.27	2.1E-03	A	Normal	9	5.0E-02	B	Normal	$\infty$		
8	Low	FCM	10	123.88	0.30	0.12	0.24	2.1E-03	A	Normal	9	5.0E-02	B	Normal	$\infty$		
10	Low	FCM	10	123.91	0.28	0.11	0.23	2.1E-03	A	Normal	9	5.0E-02	B	Normal	$\infty$		
12.5	Low	FCM	10	123.97	0.28	0.11	0.22	2.1E-03	A	Normal	9	5.0E-02	B	Normal	$\infty$		
16	Low	FCM	10	124.05	0.27	0.11	0.22	2.1E-03	A	Normal	9	5.0E-02	B	Normal	$\infty$		
20	Low	FCM	10	124.17	0.27	0.11	0.21	2.3E-03	A	Normal	9	5.0E-02	B	Normal	$\infty$		

Effect of voltage quantization on measurement by voltmeter				Effect of transverse sensitivity on accelerometer output voltage measurement				Effect of total distortion of accelerometer output voltage				Repeatability of sensitivity				Effect of laser wavelength instability			
%	Type	Distribution	DOF	%	Type	Distribution	DOF	%	Type	Distribution	DOF	%	Type	Distribution	DOF	%	Type	Distribution	DOF
2.9E-05	B	Rectangular	$\infty$	4.0E-03	B	Normal	$\infty$	1.9E-03	B	Normal	$\infty$	8.7E-01	A	Normal	20	1.6E-03	B	Normal	$\infty$
2.9E-05	B	Rectangular	$\infty$	4.0E-03	B	Normal	$\infty$	1.9E-03	B	Normal	$\infty$	3.0E-01	A	Normal	20	1.6E-03	B	Normal	$\infty$
2.9E-05	B	Rectangular	$\infty$	4.0E-03	B	Normal	$\infty$	9.1E-04	B	Normal	$\infty$	3.5E-01	A	Normal	20	1.6E-03	B	Normal	$\infty$
2.9E-05	B	Rectangular	$\infty$	4.0E-03	B	Normal	$\infty$	1.1E-03	B	Normal	$\infty$	5.3E-01	A	Normal	20	1.6E-03	B	Normal	$\infty$
2.9E-04	B	Rectangular	$\infty$	4.0E-03	B	Normal	$\infty$	1.2E-02	B	Normal	$\infty$	3.2E-01	A	Normal	20	1.6E-03	B	Normal	$\infty$
2.9E-04	B	Rectangular	$\infty$	4.0E-03	B	Normal	$\infty$	6.5E-03	B	Normal	$\infty$	1.4E-01	A	Normal	20	1.6E-03	B	Normal	$\infty$
2.9E-04	B	Rectangular	$\infty$	4.0E-03	B	Normal	$\infty$	5.5E-03	B	Normal	$\infty$	3.2E-01	A	Normal	20	1.6E-03	B	Normal	$\infty$
8.2E-05	B	Rectangular	$\infty$	1.5E-02	B	Normal	$\infty$	1.4E-02	B	Normal	$\infty$	1.1E-02	A	Normal	29	1.6E-03	B	Normal	$\infty$
8.2E-05	B	Rectangular	$\infty$	1.4E-02	B	Normal	$\infty$	5.0E-03	B	Normal	$\infty$	1.2E-02	A	Normal	29	1.6E-03	B	Normal	$\infty$
8.2E-05	B	Rectangular	$\infty$	1.3E-02	B	Normal	$\infty$	2.6E-03	B	Normal	$\infty$	1.3E-02	A	Normal	29	1.6E-03	B	Normal	$\infty$
4.1E-05	B	Rectangular	$\infty$	1.2E-02	B	Normal	$\infty$	6.6E-03	B	Normal	$\infty$	1.1E-02	A	Normal	29	1.6E-03	B	Normal	$\infty$
4.1E-05	B	Rectangular	$\infty$	1.1E-02	B	Normal	$\infty$	3.3E-03	B	Normal	$\infty$	1.2E-02	A	Normal	29	1.6E-03	B	Normal	$\infty$
4.1E-05	B	Rectangular	$\infty$	1.0E-02	B	Normal	$\infty$	1.8E-03	B	Normal	$\infty$	1.5E-02	A	Normal	29	1.6E-03	B	Normal	$\infty$
4.1E-05	B	Rectangular	$\infty$	9.0E-03	B	Normal	$\infty$	1.2E-03	B	Normal	$\infty$	6.6E-03	A	Normal	29	1.6E-03	B	Normal	$\infty$
4.1E-05	B	Rectangular	$\infty$	9.0E-03	B	Normal	$\infty$	9.8E-04	B	Normal	$\infty$	3.6E-03	A	Normal	29	1.6E-03	B	Normal	$\infty$
4.1E-05	B	Rectangular	$\infty$	1.1E-02	B	Normal	$\infty$	6.5E-04	B	Normal	$\infty$	4.3E-03	A	Normal	29	1.6E-03	B	Normal	$\infty$
4.1E-05	B	Rectangular	$\infty$	1.0E-02	B	Normal	$\infty$	5.6E-04	B	Normal	$\infty$	1.3E-02	A	Normal	29	1.6E-03	B	Normal	$\infty$

Effect of vibration frequency instability				Effect of displacement quantization on displacement measurement by universal counter				Effect of displacement quantization on measurement by laser interferometer				Effect of displacement distortion on displacement measurement				Effect of remounting accelerometer			
%	Type	Distribution	DOF	%	Type	Distribution	DOF	%	Type	Distribution	DOF	%	Type	Distribution	DOF	%	Type	Distribution	DOF
5.3E-02	B	Normal	$\infty$	6.4E-05	B	Triangular	$\infty$	6.4E-05	B	Triangular	$\infty$	4.2E-02	B	Normal	$\infty$	5.8E-02	B	Normal	$\infty$
5.7E-02	B	Normal	$\infty$	5.1E-05	B	Triangular	$\infty$	5.1E-05	B	Triangular	$\infty$	4.6E-02	B	Normal	$\infty$	5.8E-02	B	Normal	$\infty$
3.7E-02	B	Normal	$\infty$	8.2E-05	B	Triangular	$\infty$	8.2E-05	B	Triangular	$\infty$	2.7E-02	B	Normal	$\infty$	5.8E-02	B	Normal	$\infty$
1.2E-02	B	Normal	$\infty$	1.3E-04	B	Triangular	$\infty$	1.3E-04	B	Triangular	$\infty$	2.3E-02	B	Normal	$\infty$	5.8E-02	B	Normal	$\infty$
3.1E-03	B	Normal	$\infty$	1.0E-04	B	Triangular	$\infty$	1.0E-04	B	Triangular	$\infty$	1.2E-01	B	Normal	$\infty$	5.8E-02	B	Normal	$\infty$
1.6E-03	B	Normal	$\infty$	1.6E-04	B	Triangular	$\infty$	1.6E-04	B	Triangular	$\infty$	9.1E-02	B	Normal	$\infty$	5.8E-02	B	Normal	$\infty$
2.6E-03	B	Normal	$\infty$	2.5E-04	B	Triangular	$\infty$	2.5E-04	B	Triangular	$\infty$	8.4E-02	B	Normal	$\infty$	5.8E-02	B	Normal	$\infty$
3.9E-03	B	Normal	$\infty$	1.6E-04	B	Triangular	$\infty$	1.6E-04	B	Triangular	$\infty$	1.0E-01	B	Normal	$\infty$	8.6E-02	B	Normal	$\infty$
4.4E-03	B	Normal	$\infty$	2.5E-04	B	Triangular	$\infty$	2.5E-04	B	Triangular	$\infty$	3.2E-02	B	Normal	$\infty$	8.6E-02	B	Normal	$\infty$
3.9E-03	B	Normal	$\infty$	4.1E-04	B	Triangular	$\infty$	4.1E-04	B	Triangular	$\infty$	2.2E-02	B	Normal	$\infty$	8.6E-02	B	Normal	$\infty$
5.6E-03	B	Normal	$\infty$	3.2E-04	B	Triangular	$\infty$	3.2E-04	B	Triangular	$\infty$	1.3E-01	B	Normal	$\infty$	8.6E-02	B	Normal	$\infty$
5.3E-03	B	Normal	$\infty$	5.1E-04	B	Triangular	$\infty$	5.1E-04	B	Triangular	$\infty$	9.0E-02	B	Normal	$\infty$	8.6E-02	B	Normal	$\infty$
4.4E-03	B	Normal	$\infty$	8.2E-04	B	Triangular	$\infty$	8.2E-04	B	Triangular	$\infty$	6.7E-02	B	Normal	$\infty$	8.6E-02	B	Normal	$\infty$
4.9E-03	B	Normal	$\infty$	1.3E-03	B	Triangular	$\infty$	1.3E-03	B	Triangular	$\infty$	5.4E-02	B	Normal	$\infty$	8.6E-02	B	Normal	$\infty$
3.7E-04	B	Normal	$\infty$	2.0E-03	B	Triangular	$\infty$	2.0E-03	B	Triangular	$\infty$	4.9E-02	B	Normal	$\infty$	8.6E-02	B	Normal	$\infty$
6.2E-04	B	Normal	$\infty$	3.3E-03	B	Triangular	$\infty$	3.3E-03	B	Triangular	$\infty$	4.0E-02	B	Normal	$\infty$	8.6E-02	B	Normal	$\infty$
4.7E-04	B	Normal	$\infty$	5.1E-03	B	Triangular	$\infty$	5.1E-03	B	Triangular	$\infty$	3.7E-02	B	Normal	$\infty$	8.6E-02	B	Normal	$\infty$

**KRISS (APS 129 shaker):** Uncertainty budget of the modulus of the voltage sensitivity of standard accelerometer (SA-704 S/N1021).

Freq. Hz	Accel. m/s <sup>2</sup>	Relative standard uncertainty %												$u_c(S)$ %	$U(S)$ % ( $k=2$ )
		$u_1$	$u_2$	$u_3$	$u_4$	$u_5$	$u_6$	$u_7$	$u_8$	$u_9$	$u_{10}$	$u_{11}$	$u_{12}$		
0.5	0.4	0.12	0.06	0.12	0.00	0.00	0.00	0.01	0.12	0.00	0.00	0.00	0.10	0.24	0.47
0.63	0.6	0.09	0.03	0.12	0.00	0.00	0.00	0.01	0.10	0.00	0.00	0.00	0.09	0.20	0.40
0.8	0.8	0.06	0.02	0.08	0.00	0.00	0.00	0.01	0.05	0.00	0.00	0.00	0.09	0.14	0.28
1	1.6	0.02	0.01	0.07	0.00	0.00	0.00	0.01	0.07	0.00	0.00	0.00	0.18	0.21	0.42
1.25	2.0	0.02	0.00	0.06	0.00	0.00	0.00	0.01	0.04	0.00	0.00	0.00	0.12	0.14	0.28
1.6	2.5	0.01	0.00	0.05	0.00	0.00	0.00	0.01	0.04	0.00	0.00	0.00	0.19	0.20	0.40
2	6.0	0.03	0.01	0.06	0.00	0.00	0.00	0.00	0.10	0.00	0.00	0.00	0.08	0.15	0.29
2.5	6.0	0.03	0.00	0.06	0.00	0.00	0.00	0.00	0.07	0.00	0.00	0.00	0.08	0.13	0.26
3.15	7.1	0.02	0.00	0.05	0.00	0.00	0.00	0.00	0.09	0.00	0.00	0.00	0.08	0.14	0.27
4	6.1	0.02	0.00	0.05	0.00	0.00	0.00	0.01	0.09	0.00	0.00	0.00	0.09	0.14	0.28
5	5.4	0.01	0.01	0.05	0.00	0.00	0.00	0.01	0.09	0.00	0.00	0.00	0.10	0.15	0.29
6.3	5.0	0.01	0.01	0.05	0.00	0.00	0.00	0.01	0.09	0.00	0.00	0.00	0.09	0.14	0.29
8	5.0	0.01	0.01	0.06	0.00	0.00	0.00	0.02	0.10	0.00	0.00	0.00	0.09	0.15	0.30
10	5.0	0.01	0.01	0.05	0.00	0.00	0.00	0.02	0.11	0.00	0.00	0.00	0.09	0.16	0.31
12.5	5.0	0.01	0.01	0.05	0.01	0.00	0.00	0.03	0.11	0.00	0.00	0.00	0.08	0.15	0.31
16	4.0	0.01	0.01	0.05	0.01	0.00	0.00	0.04	0.09	0.00	0.00	0.00	0.08	0.14	0.29
20	5.0	0.01	0.01	0.05	0.01	0.00	0.00	0.04	0.12	0.00	0.00	0.00	0.06	0.15	0.30

Uncertainty budget of the modulus of the voltage sensitivity of standard accelerometer (SA-704 S/N1022).

Freq. Hz	Accel. m/s <sup>2</sup>	Relative standard uncertainty %												$u_c(S)$ %	$U(S)$ % ( $k=2$ )
		$u_1$	$u_2$	$u_3$	$u_4$	$u_5$	$u_6$	$u_7$	$u_8$	$u_9$	$u_{10}$	$u_{11}$	$u_{12}$		
0.5	0.4	0.12	0.07	0.12	0.00	0.00	0.00	0.01	0.14	0.00	0.00	0.00	0.11	0.25	0.50
0.63	0.6	0.10	0.03	0.12	0.00	0.00	0.00	0.01	0.10	0.00	0.00	0.00	0.09	0.20	0.40
0.8	0.8	0.06	0.02	0.08	0.00	0.00	0.00	0.01	0.05	0.00	0.00	0.00	0.08	0.14	0.28
1	1.6	0.02	0.01	0.07	0.00	0.00	0.00	0.01	0.07	0.00	0.00	0.00	0.11	0.15	0.30
1.25	2.0	0.02	0.00	0.06	0.00	0.00	0.00	0.01	0.04	0.00	0.00	0.00	0.15	0.17	0.33
1.6	2.5	0.01	0.00	0.05	0.00	0.00	0.00	0.01	0.04	0.00	0.00	0.00	0.14	0.15	0.30
2	6.0	0.03	0.01	0.06	0.00	0.00	0.00	0.00	0.11	0.00	0.00	0.00	0.08	0.15	0.30
2.5	6.0	0.03	0.00	0.05	0.00	0.00	0.00	0.00	0.08	0.00	0.00	0.00	0.08	0.13	0.26
3.15	7.1	0.02	0.00	0.05	0.00	0.00	0.00	0.00	0.10	0.00	0.00	0.00	0.08	0.14	0.28
4	6.1	0.02	0.00	0.05	0.00	0.00	0.00	0.01	0.09	0.00	0.00	0.00	0.09	0.15	0.29
5	5.4	0.01	0.01	0.05	0.00	0.00	0.00	0.01	0.10	0.00	0.00	0.00	0.10	0.15	0.30
6.3	5.0	0.01	0.01	0.05	0.00	0.00	0.00	0.01	0.10	0.00	0.00	0.00	0.09	0.15	0.29
8	5.0	0.01	0.01	0.06	0.00	0.00	0.00	0.02	0.11	0.00	0.00	0.00	0.09	0.15	0.31
10	5.0	0.01	0.01	0.05	0.00	0.00	0.00	0.02	0.11	0.00	0.00	0.00	0.09	0.16	0.31
12.5	5.0	0.01	0.01	0.05	0.00	0.00	0.00	0.03	0.12	0.00	0.00	0.00	0.10	0.16	0.33
16	4.0	0.01	0.01	0.05	0.00	0.00	0.00	0.04	0.10	0.00	0.00	0.00	0.10	0.16	0.31
20	5.0	0.01	0.01	0.05	0.00	0.00	0.00	0.04	0.12	0.00	0.00	0.00	0.08	0.16	0.32

Uncertainty components and descriptions for the calibration of standard accelerometers (SA-704).

Symbol	Uncertainty source and description	Type	Distribution function
$u_1$	Voltage: accelerometer output voltage measurement(voltmeter)	B	rectangular
$u_2$	Voltage: effect of total distortion and noise on accelerometer output voltage measurement	B	rectangular
$u_3$	Voltage: transverse motion measurement, etc.	B	rectangular
$u_4$	Displacement: quantization error of laser interferometer, quantization error of ratio counter	B	triangular
	Displacement: counter calibration	B	normal
$u_5$	Displacement: trigger hysteresis	B	rectangular
$u_6$	Displacement: filtering effect	B	rectangular
$u_7$	Displacement: voltage disturbance	B	rectangular
$u_8$	Displacement: vibration distortion	B	rectangular
$u_9$	Displacement: phase disturbance	B	rectangular
$u_{10}$	Displacement: He-Ne laser stability, misalignment	B	triangular
$u_{11}$	Frequency: function generator stability	B	rectangular
	Frequency: function generator calibration	B	normal
$u_{12}$	Sensitivity: repeatability	A	normal
	Sensitivity: deviation of the subsequent measurements, excitation direction etc.	B	rectangular

---

## Appendix C: Degrees of equivalence relative to the KCRV

### Results of horizontal voltage sensitivities of the SE-1021:

Frequency →	0.5 Hz		0.63 Hz		0.8 Hz		1 Hz		1.25 Hz		1.6 Hz	
	$D_i$	$U_i$										
Lab $i \downarrow$	in mV/(m/s <sup>2</sup> )•10 <sup>-2</sup>											
NIM	<b>16.9</b>	19.6	<b>13.9</b>	19.7	<b>12.5</b>	20.2	<b>6.3</b>	19.6	<b>6.5</b>	20.2	<b>6.7</b>	20.0
CMS	<b>53.9</b>	336.3	<b>87.9</b>	337.2	<b>80.5</b>	299.4	<b>122.3</b>	287.8	<b>94.5</b>	287.1	<b>125.7</b>	287.8
NMIA	<b>-8.1</b>	19.5	<b>-5.1</b>	19.6	<b>-4.5</b>	20.2	<b>-3.7</b>	19.6	<b>-2.5</b>	20.2	<b>-2.3</b>	20.0
NMISA	<b>-26.1</b>	33.9	<b>-22.1</b>	33.9	<b>-19.5</b>	34.3	<b>-7.7</b>	33.9	<b>-5.5</b>	34.3	<b>-4.3</b>	34.2
NMIJ	<b>64.9</b>	249.1	<b>-19.1</b>	98.0	<b>-24.5</b>	98.0	<b>-16.7</b>	110.5	<b>-29.5</b>	85.4	<b>-25.3</b>	60.1
KRISS	<b>10.9</b>	60.2	<b>10.9</b>	60.2	<b>3.5</b>	34.3	<b>5.3</b>	60.2	<b>0.5</b>	34.3	<b>2.7</b>	47.4

**Results of horizontal voltage sensitivities of the SE-1021 (continued):**

Frequency →	2 Hz		2.5 Hz		3.15 Hz		4 Hz		5 Hz		6.3 Hz	
	$D_i$	$U_i$	$D_i$	$U_i$	$D_i$	$U_i$	$D_i$	$U_i$	$D_i$	$U_i$	$D_i$	$U_i$
Lab $i \downarrow$	in mV/(m/s <sup>2</sup> )•10 <sup>-2</sup>		in mV/(m/s <sup>2</sup> )•10 <sup>-2</sup>		in mV/(m/s <sup>2</sup> )•10 <sup>-2</sup>							
	NIM	<b>5.7</b>	20.2	<b>5.6</b>	20.1	<b>6.7</b>	20.4	<b>4.1</b>	20.4	<b>6.3</b>	20.4	<b>6.0</b>
CMS	<b>62.7</b>	286.4	<b>33.6</b>	273.3	<b>34.7</b>	60.6	<b>71.1</b>	<b>60.8</b>	<b>32.3</b>	60.6	<b>52.0</b>	60.7
NMIA	<b>-5.3</b>	20.2	<b>-5.4</b>	20.1	<b>-4.3</b>	20.3	<b>-7.9</b>	20.3	<b>-6.7</b>	20.3	<b>-6.0</b>	20.3
NMISA	<b>-7.3</b>	34.3	<b>-6.4</b>	34.2	<b>-15.3</b>	34.3	<b>-8.9</b>	34.4	<b>-7.7</b>	34.4	<b>-8.0</b>	34.4
NMIJ	<b>-19.3</b>	85.5										
KRISS	<b>8.7</b>	34.4	<b>5.6</b>	34.3	<b>-2.3</b>	34.4	<b>-7.9</b>	34.4	<b>-2.7</b>	34.4	<b>-11.0</b>	34.4

Frequency →	8 Hz		10 Hz		12.5 Hz		16 Hz		20 Hz	
	$D_i$	$U_i$								
Lab $i \downarrow$	in mV/(m/s <sup>2</sup> )•10 <sup>-2</sup>									
NIM	<b>6.2</b>	20.0	<b>9.5</b>	20.0	<b>4.0</b>	20.2	<b>7.3</b>	20.4	<b>9.7</b>	20.4
CMS	<b>50.2</b>	60.5	<b>16.5</b>	60.4	<b>59.0</b>	47.8	<b>72.3</b>	60.9	<b>49.7</b>	60.8
NMIA	<b>-6.8</b>	20.0	<b>-7.5</b>	20.0	<b>-10.0</b>	20.2	<b>-7.7</b>	20.4	<b>-7.3</b>	20.4
NMISA	<b>-9.8</b>	34.2	<b>-3.5</b>	34.2	<b>-11.0</b>	34.3	<b>-17.7</b>	34.4	<b>-10.3</b>	34.5
NMIJ										
KRISS	<b>-11.8</b>	47.4	<b>-12.5</b>	47.4	<b>-16.0</b>	47.5	<b>-7.7</b>	34.4	<b>-13.3</b>	34.5

**Results of vertical voltage sensitivities of the SE-1021:**

Frequency→		0.5 Hz		0.63 Hz		0.8 Hz		1 Hz		1.25 Hz		1.6 Hz	
		$D_i$	$U_i$										
Lab $i \downarrow$		in mV/(m/s <sup>2</sup> )•10 <sup>-2</sup>		in mV/(m/s <sup>2</sup> )•10 <sup>-2</sup>		in mV/(m/s <sup>2</sup> )•10 <sup>-2</sup>		in mV/(m/s <sup>2</sup> )•10 <sup>-2</sup>		in mV/(m/s <sup>2</sup> )•10 <sup>-2</sup>		in mV/(m/s <sup>2</sup> )•10 <sup>-2</sup>	
NIM		<b>1.1</b>	7.9	<b>1.0</b>	7.9	<b>0.9</b>	7.9	<b>0.6</b>	9.2	<b>0.4</b>	9.2	<b>0.4</b>	9.2
NIMT		<b>-9.9</b>	70.6	<b>-9.0</b>	70.6	<b>-8.1</b>	70.6	<b>-3.4</b>	57.6	<b>-2.6</b>	57.6	<b>-2.6</b>	57.6

Frequency→		2 Hz		2.5 Hz		3.15 Hz		4 Hz		5 Hz		6.3 Hz	
		$D_i$	$U_i$										
Lab $i \downarrow$		in mV/(m/s <sup>2</sup> )•10 <sup>-2</sup>		in mV/(m/s <sup>2</sup> )•10 <sup>-2</sup>		in mV/(m/s <sup>2</sup> )•10 <sup>-2</sup>		in mV/(m/s <sup>2</sup> )•10 <sup>-2</sup>		in mV/(m/s <sup>2</sup> )•10 <sup>-2</sup>		in mV/(m/s <sup>2</sup> )•10 <sup>-2</sup>	
NIM		<b>0.6</b>	9.2	<b>7.4</b>	15.3	<b>6.0</b>	15.2	<b>5.1</b>	15.2	<b>4.8</b>	15.9	<b>3.6</b>	15.9
NIMT		<b>-3.4</b>	57.6	<b>2.4</b>	58.9	<b>1.0</b>	58.9	<b>1.1</b>	58.9	<b>2.8</b>	45.8	<b>2.6</b>	45.8
NMIJ				<b>-17.6</b>	31.6	<b>-14.0</b>	31.6	<b>-11.9</b>	31.6	<b>-12.2</b>	31.9	<b>-9.4</b>	31.9

Frequency→		8 Hz		10 Hz		12.5 Hz		16 Hz		20 Hz	
		$D_i$	$U_i$								
Lab $i \downarrow$		in mV/(m/s <sup>2</sup> )•10 <sup>-2</sup>		in mV/(m/s <sup>2</sup> )•10 <sup>-2</sup>		in mV/(m/s <sup>2</sup> )•10 <sup>-2</sup>		in mV/(m/s <sup>2</sup> )•10 <sup>-2</sup>		in mV/(m/s <sup>2</sup> )•10 <sup>-2</sup>	
NIM		<b>3.7</b>	15.9	<b>3.0</b>	15.9	<b>2.4</b>	15.9	<b>2.9</b>	15.9	<b>2.2</b>	15.9
NIMT		<b>1.7</b>	45.8	<b>2.0</b>	45.9	<b>2.4</b>	45.9	<b>0.9</b>	45.9	<b>1.2</b>	45.9
NMIJ		<b>-9.3</b>	31.9	<b>-8.0</b>	32.0	<b>-6.6</b>	32.0	<b>-7.1</b>	32.0	<b>-5.8</b>	32.0

---

## Results of horizontal voltage sensitivities of the SE-1022:

Frequency →	0.5 Hz		0.63 Hz		0.8 Hz		1 Hz		1.25 Hz		1.6 Hz	
	$D_i$	$U_i$										
Lab $i \downarrow$	in mV/(m/s <sup>2</sup> )•10 <sup>-2</sup>											
NIM	<b>17.1</b>	19.5	<b>15.0</b>	19.7	<b>12.7</b>	20.2	<b>8.2</b>	19.8	<b>6.9</b>	19.9	<b>9.5</b>	20.2
CMS	<b>87.1</b>	324.4	<b>92.0</b>	336.9	<b>83.7</b>	311.7	<b>117.2</b>	287.3	<b>89.9</b>	286.7	<b>110.5</b>	287.1
NMIA	<b>-8.9</b>	19.5	<b>-5.0</b>	19.6	<b>-5.3</b>	20.2	<b>-2.8</b>	19.7	<b>-1.1</b>	19.8	<b>-1.5</b>	20.1
NMISA	<b>-22.9</b>	33.8	<b>-18.0</b>	33.9	<b>-16.3</b>	34.2	<b>-12.8</b>	34.0	<b>-10.1</b>	34.0	<b>-9.5</b>	34.2
NMIJ	<b>-0.9</b>	247.6	<b>-35.0</b>	85.2	<b>-15.3</b>	98.0	<b>-19.8</b>	122.8	<b>-23.1</b>	85.3	<b>-21.5</b>	47.3
KRISS	<b>9.1</b>	60.1	<b>3.0</b>	60.1	<b>0.7</b>	34.3	<b>1.2</b>	47.3	<b>-0.1</b>	47.3	<b>2.5</b>	47.4

**Results of horizontal voltage sensitivities of the SE-1022 (continued):**

Frequency→	2 Hz		2.5 Hz		3.15 Hz		4 Hz		5 Hz		6.3 Hz	
	$D_i$	$U_i$										
Lab $i \downarrow$	in mV/(m/s <sup>2</sup> )•10 <sup>-2</sup>											
NIM	<b>7.2</b>	20.2	<b>7.0</b>	20.1	<b>4.0</b>	20.3	<b>6.8</b>	20.3	<b>8.3</b>	20.3	<b>8.8</b>	20.3
CMS	<b>78.2</b>	286.5	<b>38.0</b>	273.1	<b>94.0</b>	73.6	<b>76.8</b>	73.4	<b>17.3</b>	60.4	<b>17.8</b>	60.4
NMIA	<b>-3.8</b>	20.2	<b>-5.0</b>	20.0	<b>-9.0</b>	20.2	<b>-6.2</b>	20.2	<b>-5.7</b>	20.3	<b>-5.2</b>	20.3
NMISA	<b>-11.8</b>	34.3	<b>-12.0</b>	34.2	<b>-6.0</b>	34.3	<b>-13.2</b>	34.3	<b>-2.7</b>	34.4	<b>-2.2</b>	34.4
NMIJ	<b>-35.8</b>	85.3										
KRISS	<b>9.2</b>	34.3	<b>7.0</b>	34.2	<b>-6.0</b>	34.3	<b>-7.2</b>	34.3	<b>-9.7</b>	34.3	<b>-12.2</b>	34.3

Frequency→	8 Hz		10 Hz		12.5 Hz		16 Hz		20 Hz	
	$D_i$	$U_i$								
Lab $i \downarrow$	in mV/(m/s <sup>2</sup> )•10 <sup>-2</sup>									
NIM	<b>8.5</b>	20.0	<b>9.4</b>	20.0	<b>8.8</b>	20.0	<b>6.5</b>	20.0	<b>10.8</b>	20.1
CMS	<b>24.5</b>	60.3	<b>58.4</b>	60.5	<b>37.8</b>	60.4	<b>66.5</b>	60.6	<b>36.8</b>	60.5
NMIA	<b>-6.5</b>	20.0	<b>-9.6</b>	20.0	<b>-6.2</b>	20.0	<b>-8.5</b>	20.0	<b>-7.2</b>	20.0
NMISA	<b>-5.5</b>	34.2	<b>-11.6</b>	34.1	<b>-12.2</b>	34.1	<b>-13.5</b>	34.2	<b>-13.2</b>	34.2
NMIJ										
KRISS	<b>-13.5</b>	47.3	<b>-15.6</b>	47.3	<b>-13.2</b>	47.3	<b>-10.5</b>	47.4	<b>-14.2</b>	47.4

**Results of vertical voltage sensitivities of the SE-1022:**

Frequency→		0.5 Hz		0.63 Hz		0.8 Hz		1 Hz		1.25 Hz		1.6 Hz	
		$D_i$	$U_i$										
Lab $i \downarrow$		in mV/(m/s <sup>2</sup> )•10 <sup>-2</sup>		in mV/(m/s <sup>2</sup> )•10 <sup>-2</sup>		in mV/(m/s <sup>2</sup> )•10 <sup>-2</sup>		in mV/(m/s <sup>2</sup> )•10 <sup>-2</sup>		in mV/(m/s <sup>2</sup> )•10 <sup>-2</sup>		in mV/(m/s <sup>2</sup> )•10 <sup>-2</sup>	
NIM		<b>1.0</b>	7.8	<b>1.2</b>	7.9	<b>1.2</b>	7.9	<b>0.8</b>	9.2	<b>0.6</b>	9.2	<b>0.6</b>	9.2
NIMT		<b>-9.0</b>	70.5	<b>-10.8</b>	70.5	<b>-10.8</b>	70.5	<b>-5.2</b>	57.5	<b>-3.4</b>	57.5	<b>-3.4</b>	57.5

Frequency→		2 Hz		2.5 Hz		3.15 Hz		4 Hz		5 Hz		6.3 Hz	
		$D_i$	$U_i$										
Lab $i \downarrow$		in mV/(m/s <sup>2</sup> )•10 <sup>-2</sup>		in mV/(m/s <sup>2</sup> )•10 <sup>-2</sup>		in mV/(m/s <sup>2</sup> )•10 <sup>-2</sup>		in mV/(m/s <sup>2</sup> )•10 <sup>-2</sup>		in mV/(m/s <sup>2</sup> )•10 <sup>-2</sup>		in mV/(m/s <sup>2</sup> )•10 <sup>-2</sup>	
NIM		<b>0.7</b>	9.2	<b>8.0</b>	15.2	<b>6.3</b>	15.2	<b>5.8</b>	15.2	<b>4.9</b>	15.9	<b>4.0</b>	15.9
NIMT		<b>-4.3</b>	57.5	<b>3.0</b>	58.8	<b>1.3</b>	58.8	<b>0.8</b>	58.8	<b>1.9</b>	45.8	<b>2.0</b>	45.8
NMIJ				<b>-19.0</b>	31.5	<b>-14.7</b>	31.6	<b>-13.2</b>	31.6	<b>-12.1</b>	31.9	<b>-10.0</b>	31.9

Frequency→		8 Hz		10 Hz		12.5 Hz		16 Hz		20 Hz	
		$D_i$	$U_i$								
Lab $i \downarrow$		in mV/(m/s <sup>2</sup> )•10 <sup>-2</sup>		in mV/(m/s <sup>2</sup> )•10 <sup>-2</sup>		in mV/(m/s <sup>2</sup> )•10 <sup>-2</sup>		in mV/(m/s <sup>2</sup> )•10 <sup>-2</sup>		in mV/(m/s <sup>2</sup> )•10 <sup>-2</sup>	
NIM		<b>3.4</b>	15.9	<b>3.0</b>	15.9	<b>2.9</b>	15.9	<b>2.1</b>	15.9	<b>1.7</b>	15.9
NIMT		<b>1.4</b>	45.8	<b>2.0</b>	45.8	<b>0.9</b>	45.8	<b>2.1</b>	45.8	<b>2.7</b>	45.9
NMIJ		<b>-8.6</b>	31.9	<b>-8.0</b>	31.9	<b>-7.1</b>	31.9	<b>-5.9</b>	32.0	<b>-5.3</b>	32.0