



CIPM MRA
Comparison reports

CCQM-K3.2019.1

Automotive emission gases

KEY COMPARISON

© 2026, M.H. Adriaan *et al*

This report is published by the BIPM.

Original content from this Report may be used under the terms of the [Creative Commons Attribution 4.0 International \(CC BY 4.0\) Licence](https://creativecommons.org/licenses/by/4.0/).

Any further distribution of this Report must be cited as:
M.H. Adriaan *et al* 2026 CIPM MRA Comparison reports 08006

<https://doi.org/10.59161/ZYIU7759>

The CIPM MRA Comparison reports are made available under the Creative Commons Attribution International licence:

Attribution 4.0 International (CC BY 4.0)



By using this Report, you accept to be bound by the terms of this licence

(<https://creativecommons.org/licenses/by/4.0/>).

Distribution – you may distribute the Report according to the stipulations below.

Attribution – you must cite the Report.

Adaptations – you must cite the original Report, identify changes to the original and add the following text: This is an adaptation of an original Report by the Author(s). The opinions expressed and arguments employed in this adaptation should not be reported as representing the views of the Authors.

Translations – you must cite the original Report, identify changes to the original and add the following text: In the event of any discrepancy between the original work and the translation, only the text of the original Report should be considered valid.

Third-party material – the licence does not apply to third-party material in the Report. If using such material, you are responsible for obtaining permission from the third-party and of any claims of

International key comparison CCQM-K3.2019.1

Adriaan M.H. van der Veen¹, Noor Abdulhussain¹, Serhiy Kisel², Ovsiy Levbarg², Yaroslav Nikolaiev², Serhiy Shpilnyi², and Stanislav Yakubov²

¹Van Swinden Laboratorium, Thijsseweg 11, 2629 JA Delft, The Netherlands

²Pan-Ukrainian State Research and Production Center for Standardization, Metrology, Certification and Consumers Rights Protection, Metrolohichna St, 4, Kyiv, Ukraine, 03143

5 December 2025

Table of Contents

| | |
|--|----------|
| Field | 1 |
| Subject | 1 |
| 1 Introduction | 1 |
| 2 Design and organisation of the key comparison | 1 |
| 2.1 Participants | 1 |
| 2.2 Measurement standards | 2 |
| 2.3 Measurement protocol | 2 |
| 2.4 Schedule | 2 |
| 2.5 Measurement equation | 3 |
| 2.6 Link to CCQM-K3.2019 | 4 |
| 2.7 Measurement methods | 6 |
| 2.8 Degrees of equivalence | 6 |
| 3 Results | 7 |
| 4 Support to CMC claims | 7 |
| 5 Conclusions | 7 |
| Coordinator | 9 |
| Project reference | 9 |
| Completion date | 9 |
| A Measurement report of the participating institute | 9 |

Field

Amount of substance

Subject

Comparison of amount fractions of carbon monoxide, carbon dioxide, oxygen and propane in nitrogen (track A – core competences)

1 Introduction

This key comparison is one of a series of key comparisons in the gas analysis area assessing core competences (track A key comparisons). Such competences include, among others, the capabilities to prepare primary standard gas mixtures (PSMs) using static gravimetry [1, 2], perform the necessary purity analysis on the materials used in the gas mixture preparation, the verification of the composition of newly prepared PSMs against existing ones, and the capability of calibrating a gas mixture.

For this key comparison, a mixture containing carbon monoxide, carbon dioxide, oxygen, and propane in nitrogen has been chosen which is used for calibration of exhaust automotive gas analysers. This key comparison is a follow-up of CCQM-K3.2019 [3]. In the characterisation of the travelling standard, a link was created with the results of said key comparison based on the verification measurements [4]. This key comparison was organised as a ‘model 1’ key comparison [5].

This key comparison aims to support Calibration and Measurement Capabilities (CMCs) for the amount fractions of four components (carbon monoxide, carbon dioxide, propane and oxygen) in car exhaust gas mixtures. It is organised on behalf of the CCQM Gas Analysis Working Group (CCQM-GAWG) of the Consultative Committee for Amount of Substance: Metrology in Chemistry and Biology (CCQM) in accordance with the provisions of the CIPM Mutual Recognition Arrangement (CIPM MRA) [6, 7] of the International Committee for Weights and Measures (CIPM).

2 Design and organisation of the key comparison

2.1 Participants

Table 1 lists the participants in this key comparison.

Table 1: Participating national metrology institutes (NMIs) in CCQM-K3.2019.1

| Acronym | Country | Institute |
|---------|---------|---|
| UMTS | UA | SE “Pan-Ukrainian State Research and Production Center for Standardization, Metrology, Certification and Consumers Rights Protection” (SE “Ukrmetrteststandart”), Kyiv, Ukraine |
| VSL | NL | Van Swinden Laboratorium, Delft, The Netherlands |

2.2 Measurement standards

A mixture was prepared gravimetrically by VSL, using carbon dioxide grade 5.5, oxygen grade 6.0, carbon monoxide grade 4.7, propane grade 3.5, and nitrogen grade 6.0 BIP+. Carbon dioxide and oxygen were transferred from the pure gas. Carbon monoxide was transferred from the pre-mixture obtained with one dilution step from the pure gas. Propane was transferred from the pre-mixture obtained with two dilution steps. The mixture was verified against a set of VSL PSMs. All pure gases were subjected to a purity analysis in accordance with ISO 19229 [8] prior to use for preparation of the gas mixture.

The filling pressure in the cylinder was approximately 11.5 MPa. An aluminium cylinder having a 5 L volume from Luxfer UK with an Aculife IV treatment was used. The mixture composition and its associated uncertainty was calculated in accordance with ISO 6142-1 [1,2]. The amount fractions as obtained from gravimetry and purity analysis of the parent gases were used as key comparison reference values (KCRVs). The expanded uncertainties include uncertainty contributions from weighing, purity analysis of the parent gases, the standard atomic weights and the verification of the gas mixture.

The nominal ranges of amount fractions of the targeted components in the mixtures are given in table 2.

Table 2: Nominal composition of mixtures, given in amount fractions (cmol mol^{-1}) [3]

| Component | Amount fraction range |
|-----------------|-----------------------|
| Carbon monoxide | 0.5 – 2 |
| Carbon dioxide | 2 – 5 |
| Oxygen | 1 – 4 |
| Propane | 0.01 – 0.03 |
| Nitrogen | Balance |

2.3 Measurement protocol

The measurement protocol requested UMTS to perform at least three (3) measurements, with independent calibrations. The replicates, leading to a measurement, were to be carried out under repeatability conditions. The protocol informed the participant about the nominal amount fraction ranges (see table 2). The participant was requested to submit a description of their method and a full description of their uncertainty evaluation used for evaluating the uncertainty of their result.

2.4 Schedule

The schedule of this key comparison was as follows (see table 3).

Table 3: Schedule

| Date | Event |
|----------|--|
| Apr 2023 | Agreement of protocol |
| Feb 2024 | Preparation of the mixture |
| Mar 2024 | Characterisation of mixture compositions |
| Jun 2024 | Dispatch of the mixture |
| Oct 2024 | Measurements by the participant |
| Nov 2024 | Measurement report due |
| Feb 2025 | Return of cylinder |
| Mar 2025 | Re-characterisation of the mixture |
| Mar 2025 | Draft A report available |

2.5 Measurement equation

The KCRVs are based on the weighing data, the molar masses of the components and the purity verification of the parent gases. The mixture underwent verification prior to shipping them to UMTS. In the preparation, the following four groups of uncertainty components have been considered:

- gravimetric preparation (weighing process) ($x_{i,\text{grav}}$)
- purity of the parent gases ($\Delta x_{i,\text{purity}}$)
- stability of the gas mixture ($\Delta x_{i,\text{stab}}$)
- correction due to partial recovery of a component ($\Delta x_{i,\text{nr}}$)

Previous experience has indicated that there are no stability issues and no correction is needed for the partial recovery of a component [3, 9–12]. These terms are set to zero, and so are their associated standard uncertainties.

The amount fraction $x_{i,\text{prep}}$ of a particular component in mixture i , as it appears during use of the cylinder, can now be expressed as

$$x_{i,\text{prep}} = x_{i,\text{grav}} + \Delta x_{i,\text{purity}} \quad (1)$$

Using the law of propagation of uncertainty (LPU) from the Guide to the Expression of Uncertainty in Measurement (GUM) [13] for mutually independent quantities provides the equation for calculating the associated standard uncertainty which takes the form

$$u^2(x_{i,\text{prep}}) = u^2(x_{i,\text{grav}}) + u^2(\Delta x_{i,\text{purity}}) \quad (2)$$

The results of the amount fractions of the travelling standard are shown in table 4. For this key comparison, a new travelling standard was prepared with different target amount fractions than those of CCQM-K3.2019 [3]. The latter travelling standards had been used in CCQM-K3.2019 and EURAMET.QM-K3.2019 [12], so that their KCRVs were known, which makes them unfit for this follow-up key comparison.

Table 4: Composition of the travelling standard, expressed in amount fractions (cmol mol^{-1}). The standard uncertainties cover the preparation only.

| Component | x_{prep} | $u(x_{\text{prep}})$ |
|-----------------|-------------------|----------------------|
| Carbon monoxide | 1.117 718 | 0.000 055 |
| Carbon dioxide | 2.035 843 | 0.000 090 |
| Propane | 0.019 576 1 | 0.000 001 3 |
| Oxygen | 2.938 81 | 0.000 13 |

The validity of the composition of the mixture has been demonstrated by verifying the composition as calculated from the preparation data with that obtained from (analytical chemical) measurement. In order to have a positive demonstration of the preparation data (including uncertainty), the following condition should be met [1]:

$$|x_{i,\text{prep}} - x_{i,\text{ver}}| \leq 2\sqrt{u^2(x_{i,\text{prep}}) + u^2(x_{i,\text{ver}})} \quad (3)$$

The factor 2 is a coverage factor (assuming a normal distribution, 95 % level of confidence). The assumption was made that both preparation and verification are unbiased. Such bias has never been observed. The uncertainty associated with the verification highly depends on the experimental design followed. In this particular key comparison, an approach has been chosen which is consistent with CCQM-K3 [10] and takes advantage of the work done in the gravimetry study CCQM-P41 [14].

The verification experiments have demonstrated that within the uncertainty of these measurements, the gravimetric values of the key comparison mixtures agreed with older measurement standards.

The expression for the standard uncertainty of the key comparison reference value is

$$u^2(x_{i,\text{KCRV}}) = u^2(x_{i,\text{prep}}) + u^2(x_{i,\text{ver}}) \quad (4)$$

2.6 Link to CCQM-K3.2019

A link with CCQM-K3.2019 [3] was established through the verification measurements of the travelling standard. In the same sequences, one of the travelling standards of CCQM-K3.2019 was included and the analytical composition was compared with the composition from preparation. The background of this way of linking key comparisons is described elsewhere [4].

The verification measurements were run on a gas chromatograph (GC) equipped with four different columns and thermal conductivity detectors (TCDs) and a flame ionisation detector (FID). The GC primarily operates with hydrogen for FID-B, argon for TCD-D, and helium for TCD-B and TCD-C. The method includes four simultaneous analyses, each with its own detector, ensuring a complete measurement in approximately 14 minutes.

Each component is analyzed using specific columns. Propane is measured on FID-B using a GS alumina plot KCl column. Oxygen is detected on TCD-D using a combination of micropacked molsieve 5A and silica gel 80/100 mesh 22" for separation. Carbon dioxide is analyzed on TCD-A using a micropacked precolumn with Hayesep A 80/100 mesh 6. Carbon monoxide is measured on TCD-C using a micropacked precolumn (Hayesep A 80/100 mesh 6) combined with a CP molsieve 5A column. A backflush mechanism optimizes separation and prevents contamination.

The new GC performs better than the GCs used in CCQM-K3.2019 and EURAMET.QM-K3 [3, 12]. The relative expanded uncertainty for the verification of the travelling standard was set at 0.10 %, which is the same as in CCQM-K3.2019 and EURAMET.QM-K3 [3, 12].

The results of the verification measurements are shown in tables 5–8. All verifications were performed against the PSMs of the coordinating laboratory. VSL428449 is one of the travelling standards from the previous key comparisons and was verified again to establish a link with CCQM-K3.2019 [4].

For carbon dioxide (table 5), the relative differences from verification are 0.02 % for the travelling standard from CCQM-K3.2019 and 0.06 % for the travelling standard used in this key comparison. The column labelled y_i shows the obtained instrument responses and the next column the associated standard uncertainty $u(y_i)$. Δx_i denotes the difference between the amount fractions from preparation and analysis. These results confirm the link between this key comparison and CCQM-K3.2019 and equivalence between the travelling standards.

For carbon monoxide (table 6), the relative differences from verification are -0.08 % for the travelling standard from CCQM-K3.2019 and -0.14 % for the travelling standard used in this key comparison. These larger differences are due to the shape of the calibration function in the neighbourhood of 1 cmol mol^{-1} . A comparison using a single point calibration [15] confirmed the equivalence.

For oxygen (table 7), the relative differences from verification are -0.09 % for the travelling standard from CCQM-K3.2019 and -0.01 % for the travelling standard used in this key comparison. The equivalence was confirmed.

For propane (table 8), the relative differences from verification are 0.01 % for the travelling standard from CCQM-K3.2019 and 0.01 % for the travelling standard used in this key comparison. Equivalence was confirmed.

Table 5: Verification results for the amount fraction carbon dioxide using a multipoint calibration in accordance with ISO 6143

| Mixture | y_i a.u. | $u(y_i)$ a.u. | $x_{i,\text{prep}}$ cmol mol^{-1} | $x_{i,\text{ver}}$ cmol mol^{-1} | $u(x_{i,\text{ver}})$ cmol mol^{-1} | $\Delta x_i/x_{i,\text{prep}}$ % |
|-----------|---------------|------------------|---|--|---|-------------------------------------|
| VSL502704 | 509.66 | 0.04 | 2.035 84 | 2.037 09 | 0.000 22 | 0.06 |
| VSL428449 | 501.09 | 0.08 | 2.002 46 | 2.002 78 | 0.000 35 | 0.02 |

Table 6: Verification results for the amount fraction carbon monoxide using a multipoint calibration in accordance with ISO 6143

| Mixture | y_i a.u. | $u(y_i)$ a.u. | $x_{i,\text{prep}}$ cmol mol^{-1} | $x_{i,\text{ver}}$ cmol mol^{-1} | $u(x_{i,\text{ver}})$ cmol mol^{-1} | $\Delta x_i/x_{i,\text{prep}}$ % |
|-----------|---------------|------------------|---|--|---|-------------------------------------|
| VSL502704 | 1115.62 | 0.07 | 1.117 72 | 1.116 19 | 0.000 09 | -0.14 |
| VSL428449 | 999.21 | 0.19 | 1.000 71 | 0.999 91 | 0.000 20 | -0.08 |

Upon return, the travelling standard was checked again. This re-verification was performed on 16 June 2025. The GC was calibrated using a multipoint calibration and one of the travelling standards from CCQM-K3.2019 [3, 12] was used as reference. For the amount fraction carbon dioxide, the results differed 0.04 % (relative to the amount fraction), for propane 0.01 %, for oxygen 0.01 %

Table 7: Verification results for the amount fraction oxygen using a multipoint calibration in accordance with ISO 6143

| Mixture | y a.u. | $u(y)$ a.u. | x_{prep} cmol mol^{-1} | x_{ver} cmol mol^{-1} | $u(x_{\text{ver}})$ cmol mol^{-1} | $\Delta x/x_{\text{prep}}$ % |
|-----------|-------------|----------------|---|--|---|---------------------------------|
| VSL502704 | 2930.25 | 0.60 | 2.938 81 | 2.936 28 | 0.000 65 | -0.09 |
| VSL428449 | 3005.22 | 0.43 | 3.011 64 | 3.011 48 | 0.000 49 | -0.01 |

Table 8: Verification results for the amount fraction propane using a multipoint calibration in accordance with ISO 6143

| Mixture | y a.u. | $u(y)$ a.u. | x_{prep} cmol mol^{-1} | x_{ver} cmol mol^{-1} | $u(x_{\text{ver}})$ cmol mol^{-1} | $\Delta x/x_{\text{prep}}$ % |
|-----------|-------------|----------------|---|--|---|---------------------------------|
| VSL502704 | 490.93 | 0.03 | 0.019 576 1 | 0.019 577 8 | 0.000 002 0 | 0.01 |
| VSL428449 | 498.26 | 0.02 | 0.019 868 4 | 0.019 870 3 | 0.000 001 8 | 0.01 |

and for carbon monoxide 0.03 %. Based on these verification data, the KCRVs and their associated uncertainties were confirmed.

2.7 Measurement methods

A summary of the methods, the dates of measurement, and the way in which metrological traceability has been established is given in table 9.

Table 9: Summary of the analytical techniques, calibration approaches and measurement periods

| NMI | Date | Calibration | Standards | Matrix | Technique |
|------|----------|-------------|----------------------------|----------|---------------|
| VSL | Mar 2024 | Multipoint | Own standards (ISO 6142-1) | Nitrogen | GC/FID+GC/TCD |
| UMTS | Oct 2024 | Multipoint | Own standards (ISO 6142-1) | Nitrogen | GC/FID+GC/TCD |

2.8 Degrees of equivalence

A unilateral degree of equivalence in key comparisons is defined in the CIPM MRA as [6]

$$\Delta x_i = d_i = x_i - x_{i,\text{KCRV}}, \quad (5)$$

and the uncertainty of the difference d_i at 95 % level of confidence. Here $x_{i,\text{KCRV}}$ denotes the KCRV (the amount fraction from preparation, $x_{i,\text{prep}}$), and x_i the result of laboratory i . The standard uncertainty of d_i can be expressed as (see also equation (4))

$$u^2(d_i) = u^2(x_i) + u^2(x_{i,\text{prep}}) + u^2(x_{i,\text{ver}}) \quad (6)$$

assuming that the aggregated error terms are uncorrelated. As discussed, the combined standard uncertainty of the reference value comprises that from preparation and that from verification for the mixture involved.

3 Results

In table 10, the KCRVs, results and degrees of equivalence are shown. The agreement is good for the amount fractions carbon dioxide, propane and oxygen. The difference for the amount fraction carbon monoxide slightly exceeds the associated expanded uncertainty.

Table 10: Key comparison reference values, results and degrees of equivalence (cmol mol^{-1}). The coverage factor for all expanded uncertainties is $k = 2$.

| Component | x_{KCRV} | $u(x_{\text{KCRV}})$ | $U(x_{\text{KCRV}})$ | x_{lab} | $U(x_{\text{lab}})$ | d | $U(d)$ |
|-----------------|-------------------|----------------------|----------------------|------------------|---------------------|----------|----------|
| Carbon monoxide | 1.1177 | 0.0011 | 0.0022 | 1.1232 | 0.0033 | 0.0055 | 0.0040 |
| Carbon dioxide | 2.0358 | 0.0020 | 0.0041 | 2.0359 | 0.0056 | 0.0001 | 0.0069 |
| Propane | 0.019 576 | 0.000 020 | 0.000 039 | 0.019 63 | 0.000 10 | 0.000 05 | 0.000 11 |
| Oxygen | 2.9388 | 0.0029 | 0.0059 | 2.9379 | 0.0084 | -0.001 | 0.010 |

4 Support to CMC claims

Participation in this key comparison can be used to support CMCs for core competences under the flexible regime [5] and for CMCs for the amount fractions carbon monoxide, carbon dioxide, propane and oxygen in automotive and binary mixtures in nitrogen. The full support for CMCs is described in the report of CCQM-K3.2019 [3].

5 Conclusions

Equivalence was shown for the amount fractions carbon dioxide, propane and oxygen in automotive mixtures.

This key comparison was successfully linked to CCQM-K3.2019. No additional uncertainty had to be taken into account.

References

- [1] ISO 6142-1 Gas analysis — Preparation of calibration gas mixtures — Gravimetric method for Class I mixtures. ISO, International Organization for Standardization, Geneva, Switzerland, 2015. First edition.
- [2] ISO 6142-1:2015/Amd 1:2020 Amendment 1: Corrections to formulae in Annex E and Annex G. ISO, International Organization for Standardization, Geneva, Switzerland, 2020.
- [3] Adriaan M H van der Veen, Ewelina T Zalewska, Janneke I T van Wijk, Midori Kobayashi, Dai Akima, Shinji Uehara, Andreia L Fioravante, Cristiane R Augusto, Claudia C Ribeiro, Viviane Silva, Florbela Dias, Alda Botas, Carlos Costa, Joengsoon Lee, Jinbok Lee, Jeongsik Lim, Hyun-Kil Bae, Namgoo Kang, Christina E Cecelski, Kimberly J Harris, Walter R Miller Jr, Jennifer Carney, James Tshilongo, Napo G Ntsasa, Mudalo I Jozela, Nompumelelo Leshabane,

- Prelylly Mohweledi Marebane, David R Worton, Eric B Mussell Webber, Sergi Moreno, Paul J Brewer, Leonid A Konopelko, Anna V Kolobova, V V Pankratov, and Olga V Efremova. International comparison CCQM-K3.2019 automotive exhaust gases. *Metrologia*, 60(1A):08008, jan 2023.
- [4] Adriaan M H van der Veen and Maurice G Cox. Degrees of equivalence across key comparisons in gas analysis. *Metrologia*, 40(2):18, 2003.
- [5] P Brewer and A. M. H. van der Veen. GAWG strategy for comparisons and CMC claims. GAWG 19/41, Gas Analysis Working Group, Sèvres, France, October 2019.
- [6] CIPM. Mutual recognition of national measurement standards and of calibration and measurement certificates issued by national metrology institutes, October 1999.
- [7] CIPM MRA-G-11 Measurement comparisons in the CIPM MRA. guidelines for organizing, participating and reporting. CIPM MRA, Bureau International des Poids et Mesures, BIPM, Pavillon de Breteuil, F-92312 Sèvres Cedex, 2021. Version 1.1.
- [8] ISO 19229 Gas analysis — Purity analysis and the treatment of purity data. ISO, International Organization for Standardization, Geneva, Switzerland, 2019. Second edition.
- [9] A Alink. The first key comparison of primary standard gas mixtures. *Metrologia*, 37(1):35–49, feb 2000.
- [10] A M H van der Veen. CCQM key comparison CCQM-K3 of measurements of CO, CO₂, and C₃H₈ in N₂. *Metrologia*, 39(1):121–122, feb 2002.
- [11] A M H van der Veen, J I T van Wijk, R P van Otterloo, R M Wessel, E W B de Leer, J Novak, M Sega, A Rakowska, I Castanheira, A Botha, and S Musil. EUROMET.QM-K3: automotive emission gas measurements. *Metrologia*, 39(1A):08005–08005, January 2002.
- [12] Ruben P Ziel, E Zalewska, Adriaan M H van der Veen, Heinrich Kipphardt, Dirk Tuma, Andres Rojo Esteban, Teresa E Fernández Vicente, Tatiana Macé, Christophe Sutour, Fabrice Marioni, Judit Fükó, László Szűcs, Tamás Büki, Zsófia Nagyné Szilágyi, Dariusz Cieciora, Grzegorz Ochman, Przemysław Romanowski, Anna Lis, Magdalena Garnuszek, Miroslava Valkova, Andreas Ackermann, Celine Pascale, Bernhard Niederhauser, Tanil Tarhan, Aylin Boztepe, Jan Beránek, and Miroslav Bárta. International comparison EURAMET.QM-K3.2019 – automotive exhaust gases. *Metrologia*, 60(1A):08030, January 2023.
- [13] BIPM, IEC, IFCC, ILAC, ISO, IUPAC, IUPAP, and OIML. Evaluation of measurement data — Guide to the expression of uncertainty in measurement, JCGM 100:2008, GUM:1995 with minor corrections, 2008.
- [14] Adriaan M H van der Veen, Freek N C Brinkmann, Mile Arnautovic, Laurie Besley, Hans-Joachim Heine, Teresa Lopez Esteban, Michela Sega, Kenji Kato, Jin Seog Kim, Alejandro Perez Castorena, Agata Rakowska, Martin J T Milton, Frank R Guenther, R Francey, and E Dlugokencky. International comparison CCQM-P41 Greenhouse gases. 1. Measurement capability. *Metrologia*, 44(1A):08002, 2007.
- [15] ISO 12963 Gas analysis — Comparison methods for the determination of the composition of gas mixtures based on one- and two-point calibration. ISO, International Organization for Standardization, Geneva, Switzerland, 2017. First edition.

Coordinator

VSL
Adriaan M.H. van der Veen
Thijsseweg 11
2629 JA Delft
the Netherlands
Phone +31 6 1202 1712
E-mail avdveen@vsl.nl

Project reference

CCQM-K3.2019.1

Completion date

March 2025

A Measurement report of the participating institute

Measurement report of SE “Ukrmetrteststandart”

1. Results

Cylinder number: 2704 E.

Results are given in Tables 1 to 5.

Table 1. Measurement 1

| Component | Date | Fraction (cmol/mol) | Rel. std. deviation (%) | Number of replicates | Standard uncertainty (cmol/mol) |
|-----------------|------------|---------------------|-------------------------|----------------------|---------------------------------|
| Oxygen | 18.09.2024 | 2.9391 | 0.06 | 4 | 0.0023 |
| Carbon dioxide | 18.09.2024 | 2.0358 | 0.02 | 4 | 0.0016 |
| Propane | 24.09.2024 | 0.019689 | 0.06 | 4 | 0.000018 |
| Carbon monoxide | 18.09.2024 | 1.1236 | 0.08 | 4 | 0.0014 |

Table 2. Measurement 2

| Component | Date | Fraction (cmol/mol) | Rel. std. deviation (%) | Number of replicates | Standard uncertainty (cmol/mol) |
|-----------------|------------|---------------------|-------------------------|----------------------|---------------------------------|
| Oxygen | 19.09.2024 | 2.9416 | 0.11 | 3 | 0.0035 |
| Carbon dioxide | 19.09.2024 | 2.0393 | 0.10 | 3 | 0.002 |
| Propane | 25.09.2024 | 0.019574 | 0.13 | 4 | 0.00003 |
| Carbon monoxide | 19.09.2024 | 1.124 | 0.10 | 3 | 0.0014 |

Table 3. Measurement 3

| Component | Date | Fraction (cmol/mol) | Rel. std. deviation (%) | Number of replicates | Standard uncertainty (cmol/mol) |
|-----------------|------------|---------------------|-------------------------|----------------------|---------------------------------|
| Oxygen | 09.10.2024 | 2.9321 | 0.03 | 3 | 0.0016 |
| Carbon dioxide | 09.10.2024 | 2.0328 | 0.02 | 3 | 0.0016 |
| Propane | 26.09.2024 | 0.019614 | 0.12 | 4 | 0.000027 |
| Carbon monoxide | 09.10.2024 | 1.121 | 0.04 | 3 | 0.0011 |

Table 4. Measurement 4

| Component | Date | Fraction (cmol/mol) | Rel. std. deviation (%) | Number of replicates | Standard uncertainty (cmol/mol) |
|-----------------|------------|---------------------|-------------------------|----------------------|---------------------------------|
| Oxygen | 10.10.2024 | 2.9383 | 0.03 | 3 | 0.0016 |
| Carbon dioxide | 10.10.2024 | 2.0357 | 0.04 | 3 | 0.0016 |
| Propane | 10.10.2024 | 0.019631 | 0.17 | 4 | 0.000036 |
| Carbon monoxide | 10.10.2024 | 1.1243 | 0.03 | 3 | 0.0011 |

Table 5. Final results

| Component | Fraction (cmol/mol) | Expanded uncertainty (cmol/mol) | Coverage factor |
|-----------------|---------------------|---------------------------------|-----------------|
| Oxygen | 2.9379 | 0.0084 | 2 |
| Carbon dioxide | 2.0359 | 0.0056 | 2 |
| Propane | 0.01963 | 0.00010 | 2 |
| Carbon monoxide | 1.1232 | 0.0033 | 2 |

2. Calibration standards

Three Primary Standard Gas Mixtures (PSGM) were used as calibration standards. PSGMs were prepared gravimetrically according to ISO 6142-1 [1]. Parent gases were pure nitrogen, pure oxygen, pure carbon dioxide, pure carbon monoxide and previously prepared pre-mixture with propane amount fraction of 1 % in nitrogen.

Composition of the PSGMs is given in Tables 6 to 8.

Table 6. Composition of PSGM No 1, cylinder number 81405010

| Component | Amount fraction, $\mu\text{mol/mol}$ | Std. uncertainty (gravimetric), $\mu\text{mol/mol}$ |
|-------------------------------|--------------------------------------|---|
| N ₂ | 942272 | 22 |
| O ₂ | 27424 | 14 |
| CO ₂ | 19288 | 10 |
| CO | 10817 | 16 |
| C ₃ H ₈ | 185.88 | 0.16 |
| Ar | 7.07 | 0.24 |
| H ₂ O | 4.06 | 0.19 |
| H ₂ | 0.328 | 0.009 |
| CH ₄ | 0.057 | 0.005 |
| C ₂ H ₆ | 0.0483 | 0.0012 |

Table 7. Composition of PSGM No 2, cylinder number M083222

| Component | Amount fraction, $\mu\text{mol/mol}$ | Std. uncertainty (gravimetric), $\mu\text{mol/mol}$ |
|-------------------------------|--------------------------------------|---|
| N ₂ | 939262 | 19 |
| O ₂ | 28795 | 11 |
| CO ₂ | 20232 | 8 |
| CO | 11501 | 13 |
| C ₃ H ₈ | 195.63 | 0.13 |
| Ar | 7.18 | 0.24 |
| H ₂ O | 4.07 | 0.18 |

| Component | Amount fraction, $\mu\text{mol/mol}$ | Std. uncertainty (gravimetric), $\mu\text{mol/mol}$ |
|-------------------------------|--------------------------------------|---|
| H ₂ | 0.349 | 0.009 |
| CH ₄ | 0.058 | 0.005 |
| C ₂ H ₆ | 0.0508 | 0.0013 |

Table 8. Composition of PSGM No 3, cylinder number M083217

| Component | Amount fraction, $\mu\text{mol/mol}$ | Std. uncertainty (gravimetric), $\mu\text{mol/mol}$ |
|-------------------------------|--------------------------------------|---|
| N ₂ | 935979 | 20 |
| O ₂ | 30455 | 13 |
| CO ₂ | 21362 | 9 |
| CO | 11986 | 14 |
| C ₃ H ₈ | 205.29 | 0.14 |
| Ar | 7.28 | 0.24 |
| H ₂ O | 4.08 | 0.18 |
| H ₂ | 0.36 | 0.01 |
| CH ₄ | 0.059 | 0.005 |
| C ₂ H ₆ | 0.0533 | 0.0013 |

Composition of the premixture C₃H₈ – N₂ is given in Table 9.

Table 9. Composition of the premixture C₃H₈ – N₂, cylinder number D621473

| Component | Amount fraction, $\mu\text{mol/mol}$ | Std. uncertainty, $\mu\text{mol/mol}$ |
|-------------------------------|--------------------------------------|---------------------------------------|
| N ₂ | 89846 | 3 |
| C ₃ H ₈ | 10015.9 | 0.9 |
| Ar | 131 | 3 |
| O ₂ | 3.06 | 0.15 |
| C ₂ H ₆ | 2.60 | 0.06 |
| H ₂ O | 1.29 | 0.05 |
| CO ₂ | 0.099 | 0.025 |
| CO | 0.049 | 0.029 |
| CH ₄ | 0.049 | 0.029 |
| H ₂ | 0.049 | 0.029 |

Purity data are given in Tables 10 to 15.

Table 10. Purity data for nitrogen used to prepare premixture C₃H₈ – N₂

| Component | Amount fraction, mol/mol | Std. uncertainty, mol/mol |
|------------------|--------------------------|---------------------------|
| N ₂ | 0.999863 | 0.000003 |
| O ₂ | 0.00000300 | 0.00000015 |
| H ₂ O | 0.00000130 | 0.00000005 |
| CO ₂ | 0.000000100 | 0.000000025 |
| Ar | 0.000132 | 0.000003 |
| CO | 0.000000050 | 0.000000029 |
| CH ₄ | 0.000000050 | 0.000000029 |
| H ₂ | 0.000000050 | 0.000000029 |

Table 11. Purity data for propane used to prepare premixture C₃H₈ – N₂

| Component | Amount fraction, mol/mol | Std. uncertainty, mol/mol |
|-------------------------------|--------------------------|---------------------------|
| N ₂ | 0.000040 | 0.000001 |
| O ₂ | 0.00000900 | 0.00000025 |
| C ₂ H ₆ | 0.000260 | 0.000006 |
| C ₃ H ₈ | 0.999691 | 0.000007 |

Table 12. Purity data for nitrogen used to prepare PSGMs

| Component | Amount fraction, mol/mol | Std. uncertainty, mol/mol |
|------------------|--------------------------|---------------------------|
| N ₂ | 0.9999909 | 0.0000003 |
| O ₂ | 0.000000060 | 0.000000005 |
| H ₂ O | 0.00000039 | 0.00000002 |
| CO ₂ | 0.000000040 | 0.000000005 |
| Ar | 0.00000500 | 0.00000025 |
| CO | 0.000000005 | 0.000000003 |
| CH ₄ | 0.000000040 | 0.000000005 |
| H ₂ | 0.000000005 | 0.000000003 |

Table 13. Purity data for oxygen used to prepare PSGMs

| Component | Amount fraction, mol/mol | Std. uncertainty, mol/mol |
|------------------|--------------------------|---------------------------|
| N ₂ | 0.00000050 | 0.00000029 |
| O ₂ | 0.9999975 | 0.0000007 |
| H ₂ O | 0.00000050 | 0.00000029 |
| CO ₂ | 0.000000170 | 0.000000017 |
| Ar | 0.0000010 | 0.0000006 |
| CO | 0.00000010 | 0.00000006 |
| CH ₄ | 0.00000010 | 0.00000006 |
| H ₂ | 0.00000010 | 0.00000006 |

Table 14. Purity data for carbon monoxide used to prepare PSGMs

| Component | Amount fraction, mol/mol | Std. uncertainty, mol/mol |
|-----------------|--------------------------|---------------------------|
| N ₂ | 0.0000466 | 0.0000011 |
| O ₂ | 0.00000290 | 0.00000015 |
| CO ₂ | 0.0000018 | 0.0000001 |
| CO | 0.9999191 | 0.0000014 |
| H ₂ | 0.0000296 | 0.0000007 |

Table 15. Purity data for carbon dioxide used to prepare PSGMs

| Component | Amount fraction, mol/mol | Std. uncertainty, mol/mol |
|------------------|--------------------------|---------------------------|
| N ₂ | 0.00000980 | 0.00000025 |
| O ₂ | 0.0000024 | 0.0000001 |
| H ₂ O | 0.000022 | 0.000001 |
| CO ₂ | 0.999965 | 0.000001 |
| CH ₄ | 0.00000084 | 0.00000004 |

PSGMs were prepared gravimetrically according to ISO 6142-1 [1] using an automatic weighing device. PSGM composition was calculated by formula (3) from ISO 6142-1 [1].

Amount fraction of impurities was measured/calculated according to ISO 19229 [2].

Example weighing data for PSGM No 1 are given in Table 16.

Table 16. Weighing data for PSGM No 1

| Parent gas | Transferred mass, g | Std. uncertainty, g |
|---|---------------------|---------------------|
| Premixture C ₃ H ₈ - N ₂ | 11.12800 | 0.00949 |
| CO | 6.44800 | 0.00941 |
| CO ₂ | 18.06400 | 0.00960 |

| Parent gas | Transferred mass, g | Std. uncertainty, g |
|----------------|---------------------|---------------------|
| O ₂ | 18.67400 | 0.00961 |
| N ₂ | 550.7660 | 0.0185 |

The PSGMs were verified by examining their mutual consistency according to ISO 6143 [3]. Some details are given in the *Uncertainty* section below.

3. Instrumentation

Balance used for PSGM preparation by gravimetric method: Mettler Toledo electronic balance XP26003L.

Oxygen, carbon dioxide, and carbon monoxide amount fractions were measured using gas chromatograph Agilent 6890N with a thermal conductivity detector (GC-TCD). Two columns were used: Porapak Q 9ft 80/100 mesh and Mole sieve 5A 6ft 60/80 mesh. Temperature mode – isothermal at 60 °C.

Propane amount fraction was measured using gas chromatograph Agilent 6890N with a flame ionisation detector (GC-FID) and column GS-Alumina 50 m 0,53 mm. Temperature mode – isothermal at 100 °C.

In both cases carrier gas was helium, and sample volume was 0,5 mL.

Purity analysis of parent gases was performed using Agilent 6890N gas chromatographs with helium ionization detector, flame ionization detector, thermal conductivity detector and mass spectrometry detector.

4. Calibration method and value assignment

Calibration was performed according to ISO 6143 [3] using linear model. Four measurements were made, each for one day, except for propane that was measured separately (see Tables 1 to 4). Three or four runs, each including three PSGMs and the mixture to be analysed, were performed for each measurement. For each measurement, peak areas obtained during the runs were averaged, and the averaged values were used to obtain linear function coefficients and, consequently, measured values of the component amount fraction. Calculations were performed using B_LEAST software. The values from the four measurements were pooled to give the final values (see Table 5).

5. Uncertainty evaluation

5.1 Uncertainty of PSGMs

PSGM gravimetric uncertainty was calculated by formula (4) from ISO 6142-1 [1], with no correlations taken into account. Stability term was not included into the uncertainty budget as no significant short-term drift was revealed. Also, from the long-time experience it is known that for the kind of gas mixtures concerned this term can be regarded negligible compared to the claimed uncertainty level. Thus, preparation uncertainty to be calculated by formula (7) from ISO 6142-1 [1] becomes equal to the gravimetric uncertainty.

Uncertainty term from verification, which was performed by examining PSGMs mutual consistency, was estimated for each component in each PSGM via the differences, Δ , between gravimetric values of amount fractions, y_{grav} , and those determined using linear analysis function for each component in each PSGM, y_{ver} (formula 1):

$$\Delta = y_{\text{grav}} - y_{\text{ver}} \quad (1)$$

Value of Δ for each component was determined as an average of five measurements. In each case the compatibility criterion established by formula (8) from ISO 6142-1 [1] was met. Then the combined uncertainties in component amount fractions, $u_c(y)$, were calculated by formula (2) in accordance with formula (9) from ISO 6142-1 [1]:

$$u_c(y) = \frac{1}{2} \sqrt{u^2(y_{\text{grav}}) + u^2(\Delta) + \Delta^2} \quad (2)$$

where $u(y_{\text{grav}})$ and $u(\Delta)$ are standard uncertainties of y_{grav} and Δ , respectively.

For all components except for CO₂ the uncertainty values calculated by formula (2) appeared to be less than gravimetric uncertainty. To be conservative, the gravimetric uncertainties for these components stated in Tables 6 to 8 were used for the subsequent uncertainty evaluation for analysis. For CO₂, increased standard uncertainty values, calculated by formula (2), were used: 20 $\mu\text{mol/mol}$ for PSGM No 1, 22 $\mu\text{mol/mol}$ for PSGM No 2, and 13 $\mu\text{mol/mol}$ for PSGM No 3.

5.2 Uncertainty of analysis results

Uncertainty of analysis results was evaluated according to ISO 6143 [3], taking into account uncertainty of PSGMs and variability of output signal (peak areas). Calculations were performed using B_LEAST software. Standards uncertainties for each component for each measurement are given in Tables 1 to 4. Then these uncertainties were combined for each component j to obtain the standard uncertainty of a final averaged amount fraction, $u(y_j)$, by formula (3):

$$u(y_j) = \sqrt{\frac{\sum_{i=1}^n u_i^2(y_{j,i})}{n}} \quad (3),$$

where $u_i(y_{j,i})$ is a standard uncertainty of an amount fraction of component j in the i -th measurement (given in Tables 1 to 4) and n is a number of measurements. In our case $n = 4$.

For each measurement, the results were obtained in repeatability conditions. To take into account the reproducibility conditions, the between-measurement (between-day) variability has to be included into the uncertainty budget. Between-day variability was evaluated as a between-day standard deviation, $s_{\text{b-d}}$, and a combined standard uncertainty, u_c , for each component was evaluated by formula (4):

$$u_c = \sqrt{u^2(y_j) + s_{\text{b-d}}^2} \quad (4)$$

Values of $u(y_j)$, $s_{\text{b-d}}$ and u_c are given in Table 17.

Table 17. Combined standard uncertainties of final analysis results

| Component | $u(y_j)$ (by formula (3)), cmol/mol | Between-day std deviation, $s_{\text{b-d}}$, cmol/mol | Combined std unc-ty, u_c , cmol/mol |
|-------------------------------|--|---|--|
| O ₂ | 0.0012 | 0.0040 | 0.0042 |
| CO ₂ | 0.0008 | 0.0027 | 0.0028 |
| C ₃ H ₈ | 0.000014 | 0.000048 | 0.000050 |
| CO | 0.0006 | 0.0015 | 0.0016 |

It can be seen from Table 17 that between-day variability was a dominating contribution to the combined uncertainty of all the components.

Combined standard uncertainties were multiplied by the coverage factor of 2 to obtain expanded uncertainties that are given in Table 5.

References

1. ISO 6142-1:2015 Gas analysis – Preparation of calibration gas mixtures – Part 1: Gravimetric method for Class I mixtures
2. ISO 19229:2019 Gas analysis – Purity analysis and the treatment of purity data
3. ISO 6143:2001 Gas analysis – Comparison methods for determining and checking the composition of calibration gas mixtures