Comparisons of the radiation protection standards for air kerma of the NIST and the BIPM for ⁶⁰Co and ¹³⁷Cs gamma radiation

C. Kessler¹, P.J. Allisy-Roberts¹, R. Minniti²

¹Bureau International des Poids et Mesures, F-92312 Sèvres Cedex

² National Institute of Standards and Technology, Gaithersburg, MD, USA

Abstract

An indirect comparison of the standards for air kerma of the National Institute of Standards and Technology (NIST), USA, and of the Bureau International des Poids et Mesures (BIPM) was carried out in the ⁶⁰Co and ¹³⁷Cs radiation protection-level beams of the BIPM in September 2011. The comparison results, based on the calibration coefficient for two transfer standards and expressed as a ratio of the NIST and the BIPM standards for air kerma, are 1.0023 with a combined standard uncertainty of 3.2×10^{-3} in ⁶⁰Co, and 0.9990 with a combined standard uncertainty of 3.5×10^{-3} in ¹³⁷Cs.

The result in the ⁶⁰Co beam for radiation protection agrees within the uncertainties with the result of the comparison carried out at the same time in the CIS Bio ⁶⁰Co beam for radiotherapy-level air kerma.

1. Introduction

The previous indirect comparison of the air kerma standards for radiation protection of the National Institute of Standards and Technology (NIST), USA, and the Bureau International des Poids et Mesures (BIPM) was made in 1994 for ¹³⁷Cs gamma radiation (Leitner *et al* 1997). The comparison result was 1.0017 (42) but when updated for the changes made to the standards at both labs, this became 1.0077 (42); the changes made to the standards (Seltzer *et al* 2003, Kessler *et al* 2009) are explained in Section 2. No comparison was made previously in the ⁶⁰Co radiation protection beam. A new indirect comparison between both laboratories was carried out in September 2011 in both the ⁶⁰Co and ¹³⁷Cs radiation protection-level beams at the BIPM, although only results for ¹³⁷Cs appear in the online BIPM key comparison database (KCDB 2014) under the reference BIPM.RI(I)-K5.

The NIST air kerma standard for ⁶⁰Co and ¹³⁷Cs is determined using a suite of six standard graphite spherical cavity ionization chambers with nominal volumes of 1 cm³, 10 cm³, 30 cm³ and 50 cm³. Three of these six standards have the same nominal volume of 50 cm³ but different wall thickness (Loftus *et al* 1974). The main characteristics of the standards are listed in Table 1. The BIPM standards are parallel-plate graphite cavity ionization chambers with volume of about 6.8 cm³ as described by Boutillon *et al* (1973) and Boutillon *et al* (1996) and the results of recent evaluations of calculated correction factors and new volume estimations are described by Burns *et al* (2007) and Kessler *et al* (2009).

The comparison was undertaken using two ionization chambers of the NIST as transfer standards. The results of the comparison are given in terms of the mean ratio of the calibration coefficients of the transfer standards determined at the two laboratories under the same reference conditions. Two ionization chambers of type Exradin^a A5, serial numbers XY051604 and XY051605, were used for the present comparison.

		NIST	standards			
Chamber						
Chamber name	Std 1	Std 10	Std 30	Std 50-1	Std 50-2	Std 50-3
Outer diameter / mm	20.65	34.28	46.07	53.40	55.80	58.00
Inner diameter / mm	12.70	26.77	38.57	46.10	45.63	45.74
Wall thickness / mm	3.980	3.76	3.75	3.65	5.09	6.13
Electrode						
Shape	cylindrical	cylindrical	cylindrical	cylindrical	cylindrical	cylindrical
Diameter / mm	1.016	1.0	1.0	3.0	3.0	3.0
Net volume ⁽¹⁾						
Air cavity / cm ³	1.1309	10.069	30.24	51.634	50.089	50.155
Wall						
Material	graphite	graphite	graphite	graphite	graphite	graphite
Density / g cm ⁻³	1.73	1.72	1.74	1.73	1.73	1.73
Applied voltage / V						
Both polarities	300	300	300	500	500	500

Table 1. Characteristics of the NIST standard for air k	erma
---	------

(1) Difference between cavity volume and electrode volume

2. Determination of the air kerma

The air kerma rate is determined by

$$\dot{K} = \frac{I}{m} \frac{W}{e} \frac{1}{1 - \overline{g}} \left(\frac{\mu_{\text{en}}}{\rho}\right)_{\text{a,c}} \overline{s}_{\text{c,a}} \prod k_i \quad , \tag{1}$$

where

I/m is the ionization current per unit mass of air measured by the standard,

- W is the average energy spent by an electron of charge e to produce an ion pair in dry air,
- \overline{g} is the fraction of electron energy lost by bremsstrahlung production in air,

^a Certain commercial equipment, instruments, and materials are identified in this work in order to specify adequately the experimental procedure. Such identification does not imply recommendation nor endorsement by the NIST nor does it imply that the material or equipment identified is the best available for the purposes described in this work.

- $(\mu_{en}/\rho)_{a,c}$ is the ratio of the mean mass energy-absorption coefficients of air and graphite,
- $\bar{s}_{c,a}$ is the ratio of the mean stopping powers of graphite and air,
- $\prod k_i$ is the product of the correction factors to be applied to the standard.

Physical data and correction factors

The values for the physical data used in (1) are consistent with the Consultative Committee for Ionizing Radiation recommendations (CCEMRI(I) 1985) and the correction factors needed for ⁶⁰Co and ¹³⁷Cs radiation are given in Table 2 for the BIPM standard and in Tables 3a and 3b for the NIST standards, together with their associated uncertainties.

Table 2.Physical constants and correction factors with their estimated relativeuncertainties of the BIPM standards for the ⁶⁰Co and ¹³⁷Cs radiation beams at the BIPM

	BIPM standard	values	uncert	ainty (1)	values	uncerta	ainty ⁽¹⁾
		⁶⁰ Co	100 s_i	$100 \ u_{\rm i}$	¹³⁷ Cs	100 s_i	$100 u_{\rm i}$
Physical	Constants						
$ ho_{\mathrm{a}}$	dry air density $^{(2)}$ / kg m ⁻³	1.2930		0.01	1.2930	_	0.01
$(\mu_{ m en}/ ho)_{ m a,c}$	ratio of mass energy-absorption coefficients	0.9989	0.01	0.04	0.9990		0.05
S _{c,a}	ratio of mass stopping powers	1.0003		$0.11^{(3)}$	1.0104	_	0.11 ⁽³⁾
W/e	mean energy per charge / J C^{-1}	33.97			33.97	_	
g a	fraction of energy lost in radiative processes	0.0031		0.02	0.0012	_	0.02
Correctio	on factors:						
$k_{ m g}$	re-absorption	0.9996	_	0.01	_	_	_
ks	recombination losses	1.0018	0.01	0.02	1.0018	0.01	0.02
$k_{ m h}$	humidity	0.9970	_	0.03	0.9970	_	0.03
$k_{ m st}$	stem scattering	1.0000	0.01	_	0.9998	0.01	_
$k_{ m wall}$	wall attenuation and scattering	1.0011	_	(4)	1.0002	0.01	(4)
$k_{ m an}$	axial non-uniformity	1.0020	_	(4)	1.0018	_	0.04
k _{rn}	radial non-uniformity	1.0003	_	0.02	1.0011	0.01	0.10
Measure	ment of I / υ						
υ	chamber volume / cm ³	$6.8570^{(5)}$	_	0.08	6.7967 ⁽⁶⁾	_	0.08
Ι	ionization current / pA		0.01	0.02		0.02	0.02
Relative s	standard uncertainty						
quadratic	summation		0.02	0.15		0.03	0.19
combined	l uncertainty		0.	.15		0.	19

⁽¹⁾ Expressed as one standard deviation

 s_i represents the type A relative standard uncertainty estimated by statistical methods, u_i represents the type B relative standard uncertainty estimated by other means

⁽²⁾ At 101 325 Pa and 273.15 K

⁽³⁾ Combined uncertainty for the product of $\overline{s}_{c,a}$ and W/e

⁽⁴⁾ Uncertainties included in the uncertainty of the chamber effective volume

⁽⁵⁾ For standard CH2, the original measured volume 6.8116 cm³ increased by the factor 1.0076 determined ionometrically following a chamber repair and reduced by the factor 1.0009 (Burns *et al* 2007)

⁽⁶⁾ For standard CH5-1, the measured volume 6.8028 cm³ reduced by the factor 1.0009 (Burns *et al* 2007).

	NIST	Std 1	uncer	tainty	Std 10	uncert	ainty	Std 30	uncert	uinty	Std 50-1	unceri	ainty	Std 50-2	uncert	ainty	Std 50-3	uncert	uinty
	Physical Constants	value	Type A	Type B	value	Type A	Type B	value	Type A 7	Type B	value	Type A	Type B	value	Type A	Type B	value	Type A	Type B
$ ho_0$	Density of dry air at T=0C and P=1atm (kg/m3)	1.2930		0.02	1.2930		0.02	1.2930		0.02	1.2930		0.02	1.2930		0.02	1.2930		0.02
$(\mu_{_{\mathrm{m}}}/ ho)_{_{\mathrm{m}}}$	ratio energy-absorption coefficients, air/graphite	0666.0		0.06	0666.0		0.06	0.9990		0.06	0666.0		0.06	0666.0		0.06	0666.0		0.06
$S_{c,a}$	ratio of stopping powers, graphite/air	1.0009		0.11	1.0004		0.11	1.0002		0.11	1.0001		0.11	1.0001		0.11	1.0001		0.11
W/e	Wair/e (J/C)	33.97			33.97			33.97			33.97			33.97			33.97		
100	(1-g), radiative loss correction.	0.9967		0.02	0.9967		0.02	0.9967		0.02	0.9967		0.02	0.9967		0.02	0.9967		0.02
	Correction Factors																		
$k_{ m h}$	humidity correction (corrects to dry air)	0.9971		0.06	0.9971		0.06	0.9971		0.06	0.9971		0.06	0.9971		0.06	0.9971		0.06
$k_{\rm s}$	ksat, loss of ionization due to recombination	1.000	0.05	0.10	1.000	0.05	0.10	1.001	0.05	0.10	1.001	0.05	0.10	1.001	0.05	0.10	1.001	0.05	0.10
$k_{\rm st}$	stem scatter	0.9982		0.05	0.9992		0.05	0.9992		0.05	0666.0		0.05	0666.0		0.05	0666.0		0.05
$k_{ m wall}$	kwall, zero wall thickness	1.0207		0.17	1.0236		0.17	1.0260		0.17	1.0263		0.17	1.0363		0.17	1.0429		0.17
$k_{ m an}$	axial nonuniformity	1.000		0.02	1.000		0.05	1.000		0.05	1.000		0.05	1.000		0.05	1.000		0.05
$k_{ m m}$	radial nonuniformity	1.000		0.01	1.000		0.01	1.000		0.01	1.000		0.01	1.000		0.01	1.000		0.01
k_{TP}	air density			0.03			0.03			0.03			0.03			0.03			0.03
V	chamber net volume/ cm3	1.1309	0.10	0.10	10.0690	0.16	0.10	30.2400	0.06	0.05	51.6340	0.06	0.05	50.0890	0.06	0.05	50.1550	0.06	0.05
Ι	ionization current/ pA		0.10	0.10		0.10	0.10		0.10	0.10		0.10	0.10		0.10	0.10		0.10	0.10
t	time			0.05			0.05			0.05			0.05			0.05			0.05
d	distance (axial)			0.02			0.02			0.02			0.02			0.02			0.02
	Relative standard uncertainty																		
	quadratic summation		0.15	0.29		0.20	0.30		0.13	0.28		0.13	0.28		0.13	0.28		0.13	0.28
	combined uncertainty (for each chamber)		0.	33		0.3	Ş		0.3	_		0.3	1		0.3	1		0.3	-
	combined uncertainty of K										0.29								
	expanded uncertainty of K (k=2)										0.59								

Table 3a. Physical constants and correction factors with their estimated relative uncertainties of the NIST standards for the ⁶⁰Co radiation beams at the NIST

	NIST	Std 1	uncer	tainty	Std 10	uncert	ainty	Std 30	uncert	ainty	Std 50-1	uncert	ainty	Std 50-2	uncert	ainty	Std 50-3	uncert	ainty
	Physical Constants	value	Type A	Type B	value	Type A	Lype B	value	Type A	Type B	value	Type A	Type B	value	Type A	Type B	value	Lype A	lype B
ρ_0	Density of dry air at T=0C and P=1atm (kg/m3)	1.2930		0.02	1.2930		0.02	1.2930		0.02	1.2930		0.02	1.2930		0.02	1.2930		0.02
$(\mu_{_{\mathrm{m}}}/ ho)_{_{\mathrm{m}}}$	ratio energy-absorption coefficients, air/graphite	0.9993		0.06	0.9993		0.06	0.9993		0.06	0.9993		0.06	0.9993		0.06	0.9993		0.06
$S_{\rm c,a}$	ratio of stopping powers, graphite/air	1.0096		0.11	1.0090		0.11	1.0087		0.11	1.0086		0.11	1.0086		0.11	1.0086		0.11
W/e	Wair/e (J/C)	33.97			33.97			33.97			33.97			33.97			33.97		
100	(1-g), radiative loss correction.	0.9982		0.02	0.9982		0.02	0.9982		0.02	0.9982		0.02	0.9982		0.02	0.9982		0.02
	Correction Factors																		
$k_{ m h}$	humidity correction (corrects to dry air)	0.9971		0.06	0.9971		0.06	0.9971		0.06	0.9971		0.06	0.9971		0.06	0.9971		0.06
k_{s}	ksat, loss of ionization due to recombination	1.000	0.05	0.10	1.000	0.05	0.10	1.001	0.05	0.10	1.001	0.05	0.10	1.001	0.05	0.10	1.001	0.05	0.10
$k_{ m st}$	stem scatter	0.9964		0.05	0.9979		0.05	0.9920		0.05	0.9987		0.05	0.9987		0.05	0.9987		0.05
$k_{ m wall}$	kwall, zero wall thickness	1.0285		0.17	1.0314		0.17	1.0347		0.17	1.0348		0.17	1.0468		0.17	1.0537		0.17
$k_{ m an}$	axial nonuniformity	1.000		0.02	1.000		0.05	1.000		0.05	1.000		0.05	1.000		0.05	1.000		0.05
$k_{ m m}$	radial nonuniformity	1.000		0.01	1.000		0.01	1.000		0.01	1.000		0.01	1.000		0.01	1.000		0.01
k_{TP}	air density			0.03			0.03			0.03			0.03			0.03			0.03
Λ	chamber volume/ cm3	1.1309	0.10	0.10	10.0690	0.16	0.10	30.2400	0.06	0.05	51.6340	0.06	0.05	50.0890	0.06	0.05	50.1550	0.06	0.05
I	ionization current/ pA		0.10	0.10		0.10	0.10		0.10	0.10		0.10	0.10		0.10	0.10		0.10	0.10
t	time			0.05			0.05			0.05			0.05			0.05			0.05
d	distance (axial)			0.02			0.02			0.02			0.02			0.02			0.02
	Relative standard uncertainty																		
	quadratic summation		0.15	0.29		0.20	0.30		0.13	0.28		0.13	0.28		0.13	0.28		0.13	0.28
	combined uncertainty (for each chamber)		0.	33		0.3	5		0.3	-		0.3			0.3	1		0.3	-
	combined uncertainty of K										0.29								
	expanded uncertainty of K ($k=2$)										0.59								

Table 3b.Physical constants and correction factors with their estimated relative
uncertainties of the NIST standards for the ¹³⁷Cs radiation beams at the NIST

The correction factors for the BIPM standards were re-evaluated in 2009 and the changes to the air-kerma rate determination arise from the results of Monte Carlo calculations of correction factors for the standards, a re-evaluation of the correction factor for saturation and a new evaluation of the air volume of the standards using an experimental chamber of variable volume. The combined effect of these changes is an increase in the BIPM determination of air kerma by the factor 1.0038 and 1.0030 for the ⁶⁰Co and ¹³⁷Cs radiation protection beams, respectively, and a reduction of the relative standard uncertainty of this determination to 1.5 parts in 10³ for ⁶⁰Co and 1.9 parts in 10³ for ¹³⁷Cs (Kessler *et al* 2009 and Burns *et al* 2007). The NIST standards were similarly re-evaluated resulting in an increase in the air kerma values by a factor of 1.0105 and 1.0090 for the ⁶⁰Co and ¹³⁷Cs radiation protection beams, respectively, as reported previously (Minniti *et al* 2011 and Seltzer *et al* 2003).

Reference conditions

Air kerma at the BIPM is determined under the following conditions (Allisy-Roberts *et al* 2011):

- the distance from source to reference plane is 1.12 m and 1 m for the ⁶⁰Co and ¹³⁷Cs beams, respectively;
- the field size in air at the reference plane is 26 cm diameter and 20 cm diameter for the ⁶⁰Co and ¹³⁷Cs beams, respectively.

The reference conditions for the measurements made at the NIST are the following:

- the distance from the source to the reference plane is 1.5 m and 1.95 m for the 60 Co and 137 Cs beams, respectively.
- the field size in air at the reference plane is 59 cm diameter and 81 cm diameter for the ⁶⁰Co and ¹³⁷Cs beams, respectively.

Reference values

The $\dot{K}_{\rm BIPM}$ value is the mean of four measurements made over a period of seven months before the comparison and is about 5 µGy s⁻¹ and 15 µGy s⁻¹ for the ⁶⁰Co and ¹³⁷Cs beams, respectively. By convention they are given at the reference date of 2011-01-01 T 00:00:00 UTC using the half-life value of 1925.21 d, $\sigma = 0.29$ d for ⁶⁰Co and 10 976 d, $\sigma = 29$ d for ¹³⁷Cs (Bé *et al* 2006).

The values of $\dot{K}_{\rm NIST}$ used to calibrate the transfer chambers are 0.7 µGy s⁻¹ and 25 µGy s⁻¹ for the ⁶⁰Co and ¹³⁷Cs beams, respectively; they are based on the air kerma rate determined with the primary standards at the reference date of 1984-12-31 (Lamperti 1988) and correcting for decay using the half-life of 1925.21 d, $\sigma = 0.29$ d for ⁶⁰Co and 30.05 y, $\sigma = 0.08$ y for ¹³⁷Cs (Bé *et al* 2006).

Beam characteristics

The characteristics of the BIPM and the NIST beams are given in Table 4.

Table 4.	Parameters of the radiation protection beams at the NIST and the BIPM
----------	---

Sou	ırce	Nominal \dot{K}	Scatter contribution / energy fluence	Field size
¹³⁷ Ca	NIST	25 μ Gy s ⁻¹ (2011-01-01)	Not available	81 cm \u00f6 at 1.95 m
Cs	BIPM	15 μ Gy s ⁻¹ (2011-01-01)	30 %	20 cm \u00f6 at 1 m
⁶⁰ Ca	NIST	$0.7 \ \mu Gy \ s^{-1} (2011-01-01)$	Not available	59 cm \u00f6 at 1.5 m
Co	BIPM	5 μ Gy s ⁻¹ (2011-01-01)	8 %	26 cm \u00f6 at 1.12 m

3. The transfer chambers and their calibration

The comparison of the NIST and BIPM standards was made indirectly using the calibration coefficients N_{κ} for two transfer chambers given by

$$N_{K,\text{lab}} = \dot{K}_{\text{lab}} / I_{\text{lab}} , \qquad (2)$$

where \dot{K}_{lab} is the air kerma rate at each lab and I_{lab} is the ionization current of a transfer chamber measured at the NIST or the BIPM. The current is corrected for the effects and influences described in this section.

The ionization chambers Exradin A5 serial numbers XY051604 and XY051605, belonging to the NIST, were the transfer chambers used for this comparison. Their main characteristics are listed in Table 5.

Characteristic	c/Nominal values of the Exradin	A5 (spherical chamber)
		Nominal value
Dimensions	Outer diameter / mm	63
	Wall thickness / mm	3
Electrode	Electrode shape	cylindrical
	Electrode diameter / mm	6.5
	Electrode length / mm	37.3
Volume	Air cavity / cm ³	100
Wall	Material	Air equivalent plastic C552
	Density / g cm ⁻³	1.76
Applied voltage	Negative polarity / V $^{(1)}$	800

Table 5.Characteristics of the NIST transfer chambers

negative polarity to the outer electrode

The experimental method for calibrations at the BIPM is described by Allisy-Roberts *et al* (2011); the essential details for calibration at both laboratories are reproduced here.

Positioning

At each laboratory the chamber was positioned with the stem perpendicular to the beam direction and with the appropriate marking on the stem (white dot) facing the source.

Applied voltage and polarity

At the BIPM and at the NIST, a collecting voltage of 800 V (negative polarity) was applied to the outer electrode of the chambers at least 30 min before any measurements were made. No corrections were applied at either laboratory for polarity.

Volume recombination

Volume recombination is negligible at the dose rate of the radiation protection beams for this type of chamber, and the initial recombination loss will be the same in the two laboratories. Consequently, no correction for recombination was applied at either laboratory.

Charge and leakage measurements

The charge Q collected by each transfer chamber was measured using a Keithley electrometer, model 642 at the BIPM. The source is operational during the entire exposure series and the charge is collected for the appropriate, electronically controlled, time interval. At the NIST, the shutter that exposes the source is opened and closed for a fixed period of time that can range between 1 and 2 minutes. During this time interval, the ionization charge produced in the chamber is collected using an electrometer Keithley model 616. This process of opening and closing the shutter is repeated multiple times. The chambers were exposed to a pre-irradiation period of at least 40 min at the BIPM and at least 20 min at the NIST before any measurements were made. The ionization current measured for the transfer standards was corrected for the leakage current both at the BIPM and the NIST. This correction was less than 1×10^{-4} in relative value.

Ambient conditions

During a series of measurements, the air temperature is measured for each current measurement and it was stable to better than 0.01 °C at the BIPM. At the NIST the temperature range during the comparison was between 21.5 °C and 24.0 °C and during a series of measurements the air temperature was stable to better than 0.05 °C. The measurements are normalized to 293.15 K and 101.325 kPa at the BIPM and the NIST. Relative humidity is controlled at (50 ± 5) % at the BIPM; at the NIST, humidity varies between 22% and 50%. No correction for humidity is applied to the ionization current measured.

Radial non-uniformity

The effect of the radial non-uniformity of the BIPM ¹³⁷Cs and ⁶⁰Co beams over the section of the transfer chambers is 1.0017 and 1.0010, respectively; at the NIST, this effect is assumed to be negligible in both beams. No corrections are made at either laboratory and an uncertainty of 0.0010 is included in Tables 7 and 8 to account for this effect.

4. **Result of the comparison**

The result of the comparison, R_{κ} , is expressed in the form

$$R_K = N_{K,\text{NIST}} / N_{K,\text{BIPM}} , \qquad (3)$$

in which, for each chamber, the mean value of measurements made at the NIST prior to those made at the BIPM (pre-BIPM) and those made afterwards (post-BIPM) is compared with the mean of the measurements made at the BIPM. The BIPM calibration coefficient is the mean of two calibrations with repositioning, in series of 30 measurements for each chamber. The NIST calibration coefficient is the mean of three series of ten measurements on each occasion (before and after the measurements at the BIPM).

Table 6 lists the relevant values of N_K at the stated reference conditions.

Table	6.		Results of	the compar	ison		
	Beam	Transfer Chambers Exradin A5	$N_{K,\rm NIST}$ / Gy μC^{-1} pre-BIPM	$N_{K, \text{BIPM}}$ / Gy μC^{-1}	$N_{K,\rm NIST}$ / Gy μC^{-1} post-BIPM	$N_{K,\rm NIST}$ / Gy $\mu \rm C^{-1}$ overall mean	R _K
		XY051604	0.29798	0.29718	0.29773	0.29785	1.0023
	⁶⁰ Co	XY051605	0.29807	0.29743	0.29815	0.29811	1.0023
						mean	1.0023
		XY051604	0.30209	0.30224	0.30183	0.30196	0.9991
	¹³⁷ Cs	XY051605	0.30227	0.30254	0.30217	0.30222	0.9989
						mean	0.9990

The comparison result is $R_{K} = 1.0023$ for ⁶⁰Co and $R_{K} = 0.9990$ for ¹³⁷Cs. The uncertainties of the comparison results are evaluated in the following section.

5. Uncertainties

Contributions to the relative standard uncertainty of $N_{K,lab}$ are listed in Tables 7 and 8. The relative standard uncertainty of the mean ionization current measured with the transfer chamber over the short period of calibration was estimated to be 1×10^{-4} and 2×10^{-4} at the BIPM and at the NIST, respectively.

Table 7.	Estimated relative standard uncertainties of the calibration coefficient, $N_{K,lab}$, of
	the transfer chambers and of the comparison result, R_K for ⁶⁰ Co beams

Deletive standard uncertainty	NIST	⁶⁰ Co	BIPM	I ⁶⁰ Co
Relative standard uncertainty	$100 \ s_{i}$	$100 u_{\rm i}$	$100 \ s_{i}$	$100 \ u_{\rm i}$
Air kerma	0.06	0.29	0.02	0.15
Ionization current of the transfer chamber	0.02	0.10	0.01	0.02
Repeatability of measurement	0.02	—	0.01	_
Distance and orientation	—	0.02	0.01	_
Air density correction	—	0.03	—	_
Decay correction	—	0.08	—	_
Radial non-uniformity	—	_	—	0.10
Relative standard uncertainty of $N_{\rm K,lab}^{}$				
quadratic summation	0.07	0.32	0.03	0.18
combined uncertainty	0.	32	0.	18
Relative standard uncertainties of R_{K}^{-1}				
combined uncertainty		0	32	

¹ Taking correlation into account.

Delative standard uncertainty	NIST	¹³⁷ Cs	BIPM	I ¹³⁷ Cs
	$100 \ s_{i}$	$100 \ s_{i}$	$100 \ s_{i}$	$100 \ u_{\rm i}$
Air kerma	0.06	0.29	0.03	0.19
Ionization current of the transfer chamber	0.02	0.10	0.01	0.02
Repeatability of measurement	0.02	_	0.01	_
Distance and orientation	-	0.02	0.01	_
Air density correction	_	0.03	_	_
Decay correction	_	0.10	_	_
Radial non-uniformity	_	_	_	0.10
Relative standard uncertainty of $N_{_{K,{ m lab}}}$				
quadratic summation	0.07	0.32	0.03	0.22
combined uncertainty	0.	33	0.	.22
Relative standard uncertainties of R_K^{-1}				
combined uncertainty		0.	35	

Table 8. Estimated relative standard uncertainties of the calibration coefficient, $N_{K,\text{lab}}$, of the transfer chambers and of the comparison result, R_K for ¹³⁷Cs beams

¹ Taking correlation into account

Some of the uncertainties in \dot{K} that appear in both the BIPM and the NIST determinations (namely air density, W/e, $\mu_{en}/\rho, \bar{g}$, $\bar{s}_{c,a}$ and k_h) are correlated and therefore cancel when evaluating the uncertainty of the comparison result R_K .

6. Degrees of equivalence

Comparison of a given NMI with the key comparison reference value

Following a decision of the CCRI, the BIPM determination of the dosimetric quantity for each of the CCRI radiation qualities is taken as the key comparison reference value (KCRV) (Allisy-Roberts *et al* 2009). It follows that for each NMI *i* having a BIPM comparison result $R_{K,i}$ (denoted x_i in the KCDB) with combined standard uncertainty, u_i , the degree of equivalence with respect to the reference value for the NMI comparison, here denoted $K_{R,i}$, is given by a pair of terms:

the relative difference
$$D_i = (K_i - K_{R,i})/K_{R,i} = R_{Ki} - 1$$
 (4)
and the expanded uncertainty (k = 2) of this difference,
 $U_i = 2 u_i$. (5)

The results for D_i and U_i are expressed in mGy/Gy. Table 9 gives the values for D_i and U_i for each NMI, *i*, taken from the KCDB of the CIPM MRA (1999) and this report, using (4) and (5). These data are presented graphically in Figure 1.

Table 9. Table of degrees of equivalence and introductory text for air kerma in 137 Cs

Key comparison BIPM.RI(I)-K5

MEASURAND : Air kerma

The key comparison reference value is the BIPM evaluation of air kerma

The degree of equivalence of each laboratory *i* with respect to the key comparison reference value is given by a pair of terms:

the relative difference D_i and $U_i = 2u_i$, its expanded uncertainty (k = 2), both expressed in mGy/Gy. In evaluating u_i , account is taken of correlation between laboratory *i* and the BIPM.

When required, the degree of equivalence between two laboratories *i* and *j* can be evaluated by a pair of terms:

 $D_{ij} = D_i - D_j$ and $U_{ij} = 2u_{ij}$, its expanded uncertainty (k = 2), both expressed in mGy/Gy.

In evaluating u_{ii} , account should be taken of correlation between u_i and u_i .

Lab i	Di	Ui
	/ (mGy/Gy)	
РТВ	3.4	5.6
GUM	-0.5	5.8
IST/ITN	1.3	4.2
LNE-LNHB	-1.6	5.2
BEV	4.1	6.2
МКЕН	5.3	5.0
VNIIM	1.3	5.4
KRISS	-1.4	4.4
NIST	-1.0	7.0
NMIJ	-2.3	5.2

Figure 1. Graph of degrees of equivalence with the KCRV



Comparison of any two NMIs with each other

The degree of equivalence between any pair of national measurement standards, when required, is expressed in terms of the difference between the two comparison results and the expanded uncertainty of this difference; consequently, it is independent of the choice of key comparison reference value.

The degree of equivalence, D_{ij} , between any pair of NMIs, *i* and *j*, is thus expressed as the difference

$$D_{ij} = D_i - D_j = R_i - R_j \tag{6}$$

and the expanded uncertainty (k = 2) of this difference, $U_{ij} = 2 u_{ij}$, where

$$u_{ij}^{2} = u_{c,i}^{2} + u_{c,j}^{2} - \sum_{k} (f_{k} u_{k,\text{corr}})_{i}^{2} - \sum_{k} (f_{k} u_{k,\text{corr}})_{j}^{2}$$
(7)

and the final two terms are used to take into account correlation between the primary standards, notably that arising from the physical constants and correction factors for similar types of standard.

Note that the data presented in the table, while correct at the time of publication of the present report, become out-of-date as NMIs make new comparisons. The formal results under the CIPM MRA are those available in the BIPM key comparison database KCDB.

6. Conclusion

Two comparisons have been carried out between the NIST and the BIPM for air kerma in ⁶⁰Co and ¹³⁷Cs radiation protection-level beams using two transfer chambers.

The ⁶⁰Co comparison result, 1.0023(32) is in agreement within the uncertainties with the result of a recent comparison carried out at the BIPM in the ⁶⁰Co CISBio beam for radiotherapy-level air kerma (Kessler *et al* 2013).

The NIST standard for air kerma in ¹³⁷Cs gamma radiation compared with the present BIPM air kerma standard gives a comparison result of 0.9990(35) and so is in agreement with the KCRV within the uncertainties.

References

Allisy-Roberts P.J., Burns D.T., Kessler C., 2011, Measuring conditions and uncertainties for the comparison and calibration of national dosimetric standards at the BIPM, <u>*Rapport BIPM-11/04*</u>, 20 pp.

Allisy-Roberts P.J., Burns D.R., Andreo P., International framework of traceability for radiation dosimetry quantities, *Metrologia*, 2009, **46(2)**, S1-S8.

Bé M.-M., Chisté V, Dulieu C., Browne E., Baglin C., Chechev V., Kuzmenco N., Helmer R., Kondev F., MacMahon D., Lee K.B., Table of Radionuclides (Vol. 3 - A = 3 to 244) *Monographie BIPM-5*.

BIPM Key Comparison Database KCDB, ¹³⁷Cs air kerma comparisons, <u>BIPM.RI(I)-K5</u>

Boutillon M., Niatel M.-TA., Study of a graphite cavity chamber for absolute measurements of ⁶⁰Co gamma rays, 1973, *Metrologia*, **9**, 139-146.

Boutillon M., Allisy-Roberts P.J., Measurement of air kerma and ambient dose equivalent in a ¹³⁷Cs beam, 1996, <u>*Rapport BIPM*-96/7</u>, 12 pp.

Burns, D.T., Allisy-Roberts, P.J., Kessler, C., 2007, Re-evaluation of the BIPM international standard for air kerma in ⁶⁰Co gamma radiation, *Metrologia*, 44(6), L53-L56.

CIPM MRA: *Mutual recognition of national measurement standards and of calibration and measurement certificates issued by national metrology institutes*, International Committee for Weights and Measures, 1999, 45 pp. <u>http://www.bipm.org/pdf/mra.pdf</u>

Comité Consultatif pour les Étalons de Mesures des Rayonnements Ionisants, Constantes physiques pour les étalons de mesure de rayonnement, 1985, *CCEMRI Section (I)*, **11**, R45.

Kessler C., Burns D.T., Allisy-Roberts P.J., Re-evaluation of the BIPM international standard for air kerma in ¹³⁷Cs gamma radiation, 2009, *Metrologia*, 46, L24-L25.

Kessler C., Allisy-Roberts P.J., Minitti R., Comparison of the standards for air kerma of the NIST and the BIPM for ⁶⁰Co gamma radiation, 2013, *Metrologia*, 50, 06002

Lamperti P., Calibration of X-Ray and Gamma-Ray Measuring Instruments, NBS Special Publication 250-16 (1988).

Leitner A., Witzani J., Boutillon M., Allisy-Roberts P., Delaunay F., Leroy E., Lamperti P., Strachotinsky C., Csete I., International comparisons of air kerma standards in ¹³⁷Cs gamma radiation, *Metrologia*, 1997, 34(2), 169-175.

Loftus, T.P., Weaver, J.T., 1974. Standardization of ⁶⁰Co and ¹³⁷Cs gamma-ray beams in terms of exposure. *J. Res. Nat. Bur. Stand.* **78A**, 465-476.

Minniti R., Czap L., Comparison of the air kerma standards for ¹³⁷Cs and ⁶⁰Co gamma-ray beams between the IAEA and the NIST, 2011, *Metrologia* **48** *Tech. Supp.* 06005. doi:10.1088/0026-1394/48/1A/06005

Seltzer, S.M., Bergstrom, P.M., 2003. Changes in the U.S. primary standards for the air kerma for gamma-ray beams. *J. Res. Natl. Inst. Stand. Technol.* **108**, 359-381.

Unterweger M., Fitzgerald R., Update of NIST half-life results corrected for ionization chamber source holder instability, Applied Radiation and Isotopes (2013), http://dx.doi.org/10.1016/j.apradiso.2013.11.017