

## REPORT TO THE EURAMET TC-F



# EURAMET project 1325 (KCDB identifier EURAMET.M.FF-S10): Comparison for gas flow range 5 ml/min to 30 l/min

## FINAL REPORT

Author:

Richard Högström

VTT Technical Research Centre of Finland Ltd, Centre for Metrology MIKES

P.O. Box 1000, 02044 VTT, Tekniikantie 1, 02150 Espoo, Finland

Email: [richard.hogstrom@vtt.fi](mailto:richard.hogstrom@vtt.fi)



## Contents

---

|  |    |
|--|----|
| Contents.....  | 2  |
| 1. Introduction.....                                   | 3  |
| 2. Organization .....                                  | 4  |
| 2.1 Comparison scheme and participants.....            | 4  |
| 2.2 Comparison schedule .....                          | 5  |
| 3. Transfer standards.....                             | 7  |
| 3.1 Description of the transfer standards .....        | 7  |
| 3.2 Operating principle.....                           | 8  |
| 3.3 Handling .....                                     | 8  |
| 4. Comparison method.....                              | 9  |
| 4.1 Scope of the comparison .....                      | 9  |
| 4.2 Calibration points .....                           | 9  |
| 4.3 Calibration procedure.....                         | 9  |
| 5. Stability of the transfer standards.....            | 11 |
| 6. Measurement results .....                           | 14 |
| 6.1 Summary of the results .....                       | 14 |
| 6.2 Notes on the results.....                          | 18 |
| 7. Bilateral equivalence.....                          | 20 |
| 7.1 Analysis method .....                              | 20 |
| 7.2 Results of analysis.....                           | 21 |
| 8. EURAMET comparison reference values (ERV) .....     | 29 |
| 8.1 ERV calculation method.....                        | 29 |
| 8.2 Consistency check.....                             | 29 |
| 8.3 Final results of the ERV analysis .....            | 30 |
| 9. Conclusion and discussion.....                      | 35 |
| 9.1 General.....                                       | 35 |
| 9.2 Identified discrepancies in the results .....      | 35 |
| 9.3 Discussion .....                                   | 37 |
| References.....  | 38 |
| Appendix 1: Detailed information of participants ..... | 40 |
| Appendix 2: Technical protocol.....                    | 44 |
| Appendix 3: Results reported by the participants.....  | 56 |

## 1. Introduction

At the EURAMET TC-F meeting at VSL in Delft on 12<sup>th</sup> to 14<sup>th</sup> of March 2013 it was proposed by MIKES to initiate an inter-comparison for low gas flows in the range 5 ml/min to 30 l/min. The comparison was motivated by the fact that the last wider comparison in the low flow range was performed in years 2005 to 2007, almost ten years ago. Based on discussions during and after the meeting, it was decided to initiate such a comparison. Twelve participants expressed their interest in participating and MIKES volunteered to coordinate the comparison and act as a pilot. The comparison is registered as EURAMET project number 1325 and as supplementary comparison EURAMET.M.FF-S10 in the BIPM database.

The aim of the project is to compare measurement capabilities of the participating laboratories in the gas flow range 5 ml/min to 30 l/min. Due to the high amount of participants the comparison was carried out in two parallel loops, with one circulating transfer standard in each loop. In order to link the results between the loops, PTB and INRIM performed measurement in both loops. The stability of the transfer standards was investigated at the end of the comparison by PTB and INRIM.

Mesa Laboratories Inc. loaned two sets of piston provers to be used as transfer standards in the comparison. The instruments were delivered to the pilot laboratory in August 2014. Based on initial tests on the transfer standards, the technical protocol (draft issued in June 2014) was revised. After this only minor revisions, which were mainly related to the comparison schedule and participants were made.

In an early phase of the comparison, instrument failure was observed, and as a result the comparison was terminated. The transfer standards were sent back to Mesa Labs for repair. After receiving the transfer standards back, the comparison was restarted in May 2015. This time, the comparison was successfully completed as planned. A detailed timeline of the comparison is given below (table 1).



Figure 1. Timeline of the EURAMET1325 comparison.

## 2. Organization

### 2.1 Comparison scheme and participants

The comparison was carried out in two parallel loops (figure 2), such that one transfer standard circulated in each loop. Before the actual comparison measurements, MIKES conducted performance tests on the transfer standards. The aim of these tests was to find out the optimal measurement procedure, in which the influence of the transfer standard itself is as small as possible. In order to establish a link between the loops, INRIM and PTB performed measurements on both transfer standards. After the actual comparison measurements, INRIM and PTB performed additional measurements on the transfer standards of loops 1 and 2, respectively, in order to assess the long-term stability of the standards.




Figure 2. Scheme of the comparison. INRIM and PTB (in red) acted as linking laboratories.

A list of participating laboratories is presented in table 1. Detailed information of laboratories is given in Appendix 1. After the EURAMET project was registered, one participant decided not to participate and one new partner (UL) joined the comparison. The Norwegian Metrology Service (Justervesenet) withdrew themselves from the comparison because they do not provide calibration services in the range applicable to the comparison anymore.

Table 1. List of participants.

| Participating laboratory |  | Country        |
|--------------------------|--|----------------|
| MIKES                    | Centre for Metrology MIKES                     | Finland        |
| CMI                      | Czech Metrology Institute                      | Czech Republic |
| UL                       | University of Ljubljana                        | Slovenia       |
| PTB                      | Physikalisch-Technische Bundesanstalt          | Germany        |
| FORCE                    | FORCE Technology                               | Denmark        |
| INRIM                    | Istituto Nazionale di Ricerca Metrologica      | Italy          |
| METAS                    | Federal Institute of Metrology METAS           | Switzerland    |
| VSL                      | VSL Dutch Metrology Institute                  | Netherlands    |
| LNE                      | Laboratoire national de métrologie et d'essais | France         |
| LEI                      | Lithuanian Energy Institute                    | Lithuania      |
| EIM                      | Hellenic Institute of Metrology                | Greece         |
| GUM                      | Central Office of Measures / Główny Urząd Miar | Poland         |

## 2.2 Comparison schedule

In the very beginning of the comparison, failure of both transfer standards was observed. Most probably, the transfer standards were damaged during transportation. The small flow cell of the transfer standard of loop 2 exhibited uneven movement of the piston and occasionally piston sticking. Clearly visible cracks were found on the flow cells glass tubes, indicating that the flow cells had experienced an impact. The pressure sensor of the other transfer standard was damaged during transportation and as a result, no pressure readings nor flow readings were displayed. Both transfer standards were returned to Mesa Labs for inspection and repair.

As a result of the extensive instrument failure, it was decided to start over the comparison. This was agreed with all the participants. In order to avoid similar problems from recurring, the packaging of the transfer standards was improved. A new timetable was drawn up and the protocol was updated accordingly. Table 4 in the Protocol (see Appendix 2) show the provisional timetables for both loops.

For each laboratory, four weeks was allowed for measurements and shipping to the next laboratory. The timetable was modified several times due to a number of delays and changes in the order of the participants in the loops. All changes were made in agreement with the participants. The delays were mainly caused by non-technical reasons, such as delays in customs clearance, other delays in transportation, summer and Christmas holidays, and other personal absences.

After completing the comparison measurements, stability measurements were carried out by INRIM and PTB to investigate the long-term stability and reproducibility of the transfer standards.

*Table 2. Actual dates of measurements at the laboratories*

| <b>Loop 1<br/>Laboratory</b> | <b>Date</b>       | <b>Loop 2<br/>Laboratory</b> | <b>Date</b>                |
|------------------------------|-------------------|------------------------------|----------------------------|
| INRIM1                       | 05/2016           | INRIM                        | 05/2016                    |
| PTB                          | 06/2016 - 07/2016 | PTB1                         | 06/2016 - 07/2016          |
| MIKES                        | 09/2016           | GUM                          | 10/2016, results withdrawn |
| FORCE                        | 10/2016           | UL                           | 11/2016                    |
| METAS                        | 01/2017           | LEI                          | No results reported        |
| VSL                          | 02/2017 - 03/2017 | CMI                          | 01/2017                    |
| EIM                          | 05/2017           | LNE                          | 02/2017 - 03/2017          |
| INRIM2                       | 06/2017           | PTB2                         | 05/2017 - 07/2017          |

### 3. Transfer standards

#### 3.1 Description of the transfer standards

The transfer standards used in the comparison are commercially available Mesa Labs DryCal 800 equipped with three measuring cells covering the whole gas flow range of the comparison. The transfer standards serial numbers are shown in table 3. After the failure of the transfer standards, both bases were changed to the newest model (figure 3). The same flow cells were used throughout the project, only repair and verification at Mesa Labs was performed after instrument failure. After these service and repair measures, the stability of the instrument response cannot be guaranteed, and thus the comparison was decided to start over and the previously measured results were discarded.

Table 3. Transfer standards used for comparison measurements

| <b>LOOP 1 instrument</b> |                  |               |                   |
|--------------------------|------------------|---------------|-------------------|
| Model                    | Unit             | Serial number | Flow range        |
| ML-800-B                 | Base             | 147457        |                   |
| ML-800-10                | Measurement cell | 135207        | 5-500 ml/min      |
| ML-800-24                | Measurement cell | 134909        | 50-5,000 ml/min   |
| ML-800-44                | Measurement cell | 135198        | 500-50,000 ml/min |
| <b>LOOP 2 instrument</b> |                  |               |                   |
| Model                    | Unit             | Serial number |                   |
| ML-800-B                 | Base             | 147461        |                   |
| ML-800-10                | Measurement cell | 135208        | 5-500 ml/min      |
| ML-800-24                | Measurement cell | 134910        | 50-5,000 ml/min   |
| ML-800-44                | Measurement cell | 135199        | 500-50,000 ml/min |



Figure 3. Mesa Labs DryCal 800 system including base and flow cell. A set of three flow cells were used to cover the flow range of the comparison.

### 3.2 Operating principle

The operation of the transfer standards is based on the positive displacement principle, in which the volumetric flow is determined by measuring the time it takes for the gas flow to “displace” a graphite piston through a glass cylinder with known dimensions. By means of internal temperature and pressure measurements, actual flow readings are converted to standardized flow readings (0 °C and 101325 Pa). There is a small clearance between the piston and the cylinder to allow frictionless motion of the piston. This causes a small internal leak, which is corrected by the transfer standard. The leak rate depends on the viscosity of the calibration gas and therefore participants were advised to use nitrogen (at least grade 5.0 and 99.999% purity) in order to minimize uncertainties caused by the gas composition. All participants used nitrogen (at least grade 5.0 purity) as the calibration gas as advised.

### 3.3 Handling

Special care was taken in packaging and handling the transfer standards in order to avoid damage and/or contamination. The transfer standards were packed in custom-made tough pelican cases to ensure safe transportation. The packaging was further improved after the instrument failure, by including additional bubble wrap around the delicate flow cells. In this way a tighter fit of the flow cells inside the case was achieved to prevent movement of the parts during transport.

Additional guidelines on handling and operation of the instruments were provided to all participants. Special care needs to be taken to avoid particulate contamination from entering the flow cells. Any micrometer size particle will build up on the walls of the flow tube. The clearance between the piston and the cylinder is only a few micrometer, and thus even a small amount of particulate debris will influence the leakage between the piston and the cylinder. This will in turn have an influence on the response, i.e. stability, of the instrument and may eventually cause piston sticking. To minimize this effect, participants were advised to use a 5 micron filter at the inlet of the transfer standard.



## 4. Comparison method

---

### 4.1 Scope of the comparison

The purpose of the comparison was to compare the volumetric gas flow measurements of the participating laboratories in the range from 5 ml/min to 30 l/min. The transfer standard is able to measure both actual volumetric flow and standard flow. In this comparison, calibrations were performed in terms of standard readings at 0 °C and 101325 Pa. The instruments were pre-set to display standard readings at the aforementioned conditions.

### 4.2 Calibration points

The calibration flow rates and the associated flow cells are shown in table 4. The flow rates for each flow cell were chosen based on results of the initial performance tests at MIKES. In the overlapping flow range, the flow cell with the best performance in terms of reproducibility and repeatability was selected. In addition, measurement were performed with both flow cells at one calibration point in the overlapping range in order to check the consistency of the measurements results.

Participants were only presumed to measure at points included in their CMCs. However, measurements at other points was strongly encouraged to support future CMC entries. The participating laboratories were asked to match the flow reading indicated by the transfer standard as close as possible (at least within  $\pm 3\%$ ) to the nominal flow in table 4 to allow good comparability between participants results. This is especially important for the low flow cell (-10), which experienced a non-linear response in the low end of the flow range.

*Table 4. Calibration points and corresponding transfer standard flow cells.*

| Nominal flow rate (ml/min) | Transfer standard flow cell |
|----------------------------|-----------------------------|
| 5                          | ML-800-10                   |
| 10                         | ML-800-10                   |
| 20                         | ML-800-10                   |
| 80                         | ML-800-10, ML-800-24        |
| 300                        | ML-800-24                   |
| 600                        | ML-800-24                   |
| 1250                       | ML-800-24, ML-800-44        |
| 5000                       | ML-800-44                   |
| 10000                      | ML-800-44                   |
| 20000                      | ML-800-44                   |
| 30000                      | ML-800-44                   |

### 4.3 Calibration procedure

The transfer standard was allowed to stabilize to laboratory conditions for at least 24 hours before starting measurements. After this, pressure and temperature readings of the transfer standard were compared against a participant's calibrated laboratory reference. This allowed to check for damage or drift of the sensors.

Before starting actual comparison measurements, it is necessary to allow the flow cell temperature to stabilize by passing gas flow through the flow cell and continuously launching the piston for 180 min. Performance tests of the transfer standards showed that an initial

“heating” is necessary for achieving consistent and comparable results, especially in the lower flow range of the measurement cells. When changing the flow rate, three minutes of continuous operation is sufficient for reaching a stable temperature.

At each nominal calibration point, four separate calibration measurements were performed. For a single calibration measurement, the flow rate of the transfer standard was obtained as the mean of at least ten (10) readings. As supporting measurements, temperature and pressure readings of the transfer standard were recorded for each measurement. Details of the recorded data can be seen from the results reported by the participants (Appendix 3).

## 5. Stability of the transfer standards

---

Measurements performed after the comparison (INRIM2 and PTB2) indicate that the response of the low flow cells (ML-800-10) have changed during the comparison (figures 4 and 5). The response of both flow cells have shifted about -0.02 ml/min (tables 5 and 6). The most obvious explanation would be a leakage in the instrument. However, according to the manufacturer, this rarely happens and no evidence of leakage was found when inspecting the instruments after the comparison. A possible reason for the shift is contamination, i.e. build up of particulate debris on the inner wall of the flow cell. Any contamination will affect the flow around the piston during measurements, and thus influence the response of the instrument. Also piston wear would have an influence on the leakage flow around the piston, but this is not likely considering the limited amount of measurements. According to the manufacturer, a typical value for the bypass flow, i.e. the piston tare value (PTV), is 0.1 ml/min, 0.2 ml/min and 1.4 ml/min for the low (ML-800-10), medium (ML-800-24) and high flow cell (ML-800-44), respectively. Thus, the low flow cell is most sensitive to contamination and even a small change in the PTV will affect the response.

A shift in the response was also observed for the medium flow cell of loop 1 (figure 4). Similar to the low flow cell, the relative change is larger in the lower end of the measurement range. In this case, however, the shift in the response is not constant and it is much larger than the typical PTV. The temperature and pressure sensors of the transfer standard were found stable during the comparison. Thus, the reasons for the observed instability remain unclear. The rest of the flow cells indicated a good stability, i.e. the change in the response was typically less than 0.05%, and thus much smaller than the measurement uncertainty ( $k=2$ ) of about 0.2 %.

Although a shift was observed for some of the flow cells, no corrections to the participants' results were made. Instead, the change in the response was considered as an additional uncertainty when evaluating results (see section 6 for calculations). It was not possible to reliably determine a correction for the shift, because the stability measurements were only performed at the end of the comparison, and thus the trend of the response could not be determined. Also, based on the participants' results (e.g. figure 6 and 7) the trend cannot be reliably deduced.

Assuming that the shift in the response of the low flow cells is caused by contamination, it will depend on the measurement times (vary among participants) and the cleanliness of the calibration gas (filtration and gas purity vary among participants). Condensation of humid air inside the cell, caused by temperature variations during transportation, is also a possible source of contamination. In any case, a linear trend cannot be assumed. However, apart from flow rates below 20 ml/min, the uncertainties of stability (tables 5 and 6) are in most cases much smaller than the uncertainties reported by the participants.

### Long-term stability of loop 1 TS




Figure 4. Measurements of loop 1 transfer standard (TS) relative error in the beginning (INRIM1) and at the end (INRIM2) of the comparison. Measurement uncertainties ( $k=2$ ) are shown as error bars. Note that at some points, the flow rates are shifted for clarity.

### Long-term stability of loop 2 TS




Figure 5. Measurements of loop 2 transfer standard (TS) relative error in the beginning (PTB1) and at the end (PTB2) of the comparison. Measurement uncertainties ( $k=2$ ) are shown as error bars. Note that at some points, the flow rates are shifted for clarity.

Table 5. Stability and associated uncertainty of loop 1 transfer standard.

| Nominal Flow Rate<br>(std ml/min)   | Transfer standard error (INRIM1)<br>(%) | Transfer standard error (INRIM2)<br>(%) | INRIM2 - INRIM1<br>(%) | INRIM2 - INRIM1<br>(ml/min) | Uncertainty due to stability<br>(%) | Uncertainty due to stability<br>(ml/min) |
|-------------------------------------|---|---|------------------------|-----------------------------|-------------------------------------|--|
| <b>Flow cell: -10 (S/N: 135207)</b> |   |   |                        |                             |                                     |  |
| 5                                   | -0.26                                   | -0.67                                   | -0.41                  | -0.02                       | 0.24                                | 0.01                                     |
| 10                                  | -0.19                                   | -0.44                                   | -0.25                  | -0.02                       | 0.14                                | 0.01                                     |
| 20                                  | -0.15                                   | -0.31                                   | -0.16                  | -0.03                       | 0.09                                | 0.02                                     |
| 80                                  | -0.15                                   | -0.15                                   | 0.01                   | 0.01                        | 0.004                               | 0.003                                    |
| <b>Flow cell: -24 (S/N: 134909)</b> |   |   |                        |                             |                                     |  |
| 80                                  | -0.44                                   | -0.70                                   | -0.26                  | -0.20                       | 0.15                                | 0.12                                     |
| 300                                 | -0.43                                   | -0.55                                   | -0.12                  | -0.37                       | 0.07                                | 0.21                                     |
| 600                                 | -0.43                                   | -0.53                                   | -0.10                  | -0.65                       | 0.06                                | 0.36                                     |
| 1250                                | -0.45                                   | -0.52                                   | -0.07                  | -0.92                       | 0.04                                | 0.51                                     |
| <b>Flow cell: -44 (S/N: 135198)</b> |   |   |                        |                             |                                     |  |
| 1250                                | 0.01                                    | 0.00                                    | 0.00                   | -0.06                       | 0.003                               | 0.03                                     |
| 5000                                | -0.03                                   | -0.04                                   | 0.00                   | -0.23                       | 0.003                               | 0.13                                     |
| 10000                               | -0.07                                   | -0.06                                   | 0.01                   | 1.26                        | 0.007                               | 0.73                                     |
| 20000                               | -0.05                                   | -0.04                                   | 0.00                   | 0.78                        | 0.002                               | 0.45                                     |
| 30000                               | -0.04                                   | 0.00                                    | 0.04                   | 11.3                        | 0.022                               | 6.51                                     |

Table 6. Stability and associated uncertainty for of loop 2 transfer standard.

| Nominal Flow Rate<br>(std ml/min)   | Transfer standard error (PTB1)<br>(%) | Transfer standard error (PTB2)<br>(%) | PTB2 - PTB1<br>(%) | PTB2 - PTB1<br>(ml/min) | Uncertainty due to stability<br>(%) | Uncertainty due to stability<br>(ml/min) |
|-------------------------------------|---------------------------------------|---------------------------------------|--------------------|-------------------------|-------------------------------------|--|
| <b>Flow cell: -10 (S/N: 135208)</b> |                                       |                                       |                    |                         |                                     |  |
| 5                                   | -0.17                                 | -0.48                                 | -0.30              | -0.02                   | 0.18                                | 0.01                                     |
| 10                                  | -0.08                                 | -0.27                                 | -0.19              | -0.02                   | 0.11                                | 0.01                                     |
| 20                                  | 0.15                                  | -0.12                                 | -0.27              | -0.05                   | 0.16                                | 0.03                                     |
| 80                                  | 0.15                                  | 0.11                                  | -0.03              | -0.03                   | 0.02                                | 0.01                                     |
| <b>Flow cell: -24 (S/N: 134910)</b> |                                       |                                       |                    |                         |                                     |  |
| 80                                  | -0.19                                 | -0.20                                 | -0.01              | -0.01                   | 0.01                                | 0.01                                     |
| 300                                 | 0.01                                  | -0.08                                 | -0.09              | -0.27                   | 0.05                                | 0.15                                     |
| 600                                 | 0.00                                  | -0.01                                 | -0.02              | -0.10                   | 0.01                                | 0.06                                     |
| 1250                                | -0.05                                 | -0.09                                 | -0.04              | -0.51                   | 0.02                                | 0.29                                     |
| <b>Flow cell: -44 (S/N: 135199)</b> |                                       |                                       |                    |                         |                                     |  |
| 1250                                | 0.01                                  | 0.05                                  | 0.04               | 0.51                    | 0.02                                | 0.29                                     |
| 5000                                | -0.04                                 | 0.06                                  | 0.10               | 5.00                    | 0.06                                | 2.89                                     |
| 10000                               | -0.04                                 | 0.01                                  | 0.05               | 5.26                    | 0.03                                | 3.04                                     |
| 20000                               | -0.07                                 | 0.00                                  | 0.06               | 12.7                    | 0.04                                | 7.32                                     |
| 30000                               | -0.04                                 | 0.00                                  | 0.04               | 10.6                    | 0.02                                | 6.13                                     |

## 6. Measurement results

---

### 6.1 Summary of the results

Measurement results reported by the participants are given in Appendix 3. For each measurement, the laboratories calculated the relative error ( $E$ ) of the transfer standard as:

$$E_i = \frac{\dot{V}_{ts,i} - \dot{V}_{ref,i}}{\dot{V}_{ref,i}} \times 100\%, \quad (1)$$

where  $\dot{V}_{ts}$  and  $\dot{V}_{ref}$  is the volumetric flow of the transfer standard and the laboratory flow standard, respectively. The result at each flow rate was calculated as the mean of four repeated measurements:

$$E_i = \frac{1}{4} (\sum_{j=1}^4 E_j) + \delta_{stab, loop(i)}, \quad (2)$$

where  $\delta_{stab, loop(i)}$  is the error due to the instability of the transfer standard.

The standard uncertainty of the error was calculated as follows:

$$u^2(E_i) = u^2(\dot{V}_{ref,i}) + u^2(\dot{V}_{ts,i}) + u^2(\delta_{stab, loop(i)}). \quad (3)$$

The uncertainty of the reference  $u(\dot{V}_{ref,i})$  was calculated as the average uncertainty of the four repeated measurements. The uncertainty of the transfer standard  $u(\dot{V}_{ts,i})$  was calculated from the spread of the error readings:

$$u(\dot{V}_{ts,i}) = \frac{1}{2\sqrt{3}} [\max(E_j) - \min(E_j)]. \quad (4)$$

The stability of the transfer standards was calculated as:

$$u(\delta_{stab, loop(i)}) = \frac{1}{\sqrt{3}} |E_{loop(i),2} - E_{loop(i),1}|, \quad (5)$$

where  $E_{loop(i),1}$  and  $E_{loop(i),2}$  is the transfer standard error measured at the beginning and at the end of the comparison, respectively.

The stability of the transfer standards was included in the uncertainty of the results for all laboratories as the full-width of the change in the response assuming a uniform distribution, i.e. type B uncertainty. This approach is justified due to the limited understanding of the reasons for the instability. Although the response of some of the flow cells appeared to have permanently shifted, a definite conclusion on the reasons for the change and the behaviour of the shift (linear or step-like) cannot be made based on the participants' results and the limited amount of stability data. In addition, initial tests on the transfer standards indicated that the reproducibility is inferior in the lower end of the measurement range, i.e. in the same range where a shift was noticed. Thus, the uncertainty derived from the stability measurements includes contributions from both long-term stability and short-term stability (reproducibility).

The results obtained with equations (2) to (5) are shown in figure 6 to 11. Error bars show the expanded uncertainty of the results. The expanded uncertainty was obtained by multiplying the standard uncertainty (equation 3) with 2. A summary of the results can be found in appendix 3. Note that the results are shifted in sake of clarity of presentation.

### Loop 1, ML-800-10




Figure 6. Measurements results of the participating laboratories in loop 1 for the low flow cell ML-800-10. Error bars show expanded uncertainties ( $k=2$ ) including stability of the transfer standard. Note that the x-axis values are shifted for clarity of presentation.

### Loop 2, ML-800-10




Figure 7. Measurements results of the participating laboratories in loop 2 for the low flow cell ML-800-10. Error bars show expanded uncertainties ( $k=2$ ) including stability of the transfer standard. Note that the x-axis values are shifted for clarity of presentation.




Figure 8. Measurements results of the participating laboratories in loop 1 for the medium flow cell ML-800-24. Error bars show expanded uncertainties ( $k=2$ ) including stability of the transfer standard. Note that the x-axis values are shifted for clarity of presentation.




Figure 9. Measurements results of the participating laboratories in loop 2 for the medium flow cell ML-800-24. Error bars show expanded uncertainties ( $k=2$ ) including stability of the transfer standard. Note that the x-axis values are shifted for clarity of presentation.






Figure 10. Measurements results of the participating laboratories in loop 1 for the medium flow cell ML-800-44. Error bars show expanded uncertainties ( $k=2$ ) including stability of the transfer standard. Note that the x-axis values are shifted for clarity of presentation.




Figure 11. Measurements results of the participating laboratories in loop 2 for the medium flow cell ML-800-44. Error bars show expanded uncertainties ( $k=2$ ) including stability of the transfer standard. Note that the x-axis values are shifted for clarity of presentation.

## 6.2 Notes on the results

The calibration points measured by the participating laboratories are summarized in table 7. Most of the laboratories performed measurements at all flow rates, or at the flow rates that were included in their calibration and measurement capabilities (CMCs). Exceptions to the measurement protocol and calibration scheme are listed below:

*FORCE*: At calibration points 600 ml/min and 1250 ml/min with the medium flow cell (ML-800-24) and at flow rate 30 000 ml/min with the high flow cell (ML-800-44), less than 10 flow indications from the transfer standard were recorded. The reason for this is that the reference standard of FORCE is a primary standard based on the displacement principle with fixed volume flow tubes. Thus, the number of flow indications during a calibration cycle depends on the flow rate and the volume of the selected flow tube. Anyhow, a minimum of 7 flow indications was recorded at these points, which is sufficient for obtaining a reliable result for the transfer standard (insignificant uncertainty due to repeatability of transfer standard reading). In addition, FORCE did not perform measurement with the small flow cell, because the pulsation induced by the transfer standard piston was found to influence the results of the primary standard. Therefore, FORCE decided not to participate in the comparison of the low flow cell (ML-800-10).

*VSL*: At calibration points 5 ml/min, 10 ml/min and 20 ml/min less repeated measurements were performed than advised in the protocol. Instead of four separate measurement, VSL performed three measurement at 5 ml/min and only two measurement at 10 ml/min and 20 ml/min. However, these measurements indicate good repeatability, and thus the results are considered to provide a reliable estimate of the error of the transfer standard.

*EIM*: Calibration points with flow rates 5 ml/min and 10 ml/min were not measured by EIM. The current CMCs of the laboratory start from 150 ml/min. In addition, at 10000 ml/min no measurements were made, because this flow rate corresponds to the lower limit of the laboratory bell prover reference standard.

*GUM*: Results were withdrawn from the comparison, because the measurements were not performed according to protocol. Actual volumetric flow readings were recorded instead of standardized readings as advised in the protocol. An attempt to convert the results to standardized readings based on pressure and temperature readings was unsuccessful.

*LEI*: Owing to technical reasons related to the laboratory flow standard, LEI was not able to provide measurement results according to the protocol.

*LNE*: LNE decided to participate the comparison only in the lower flow range.



## 7. Bilateral equivalence

### 7.1 Analysis method

The applied method of linking results between laboratories in different loops is adapted from the method presented by Heinonen *et al.* [1]. In this case, however, there are only two loops and two linking laboratories, which simplifies the analysis.

The bilateral equivalence can be calculated as:

$$D_{ij} = E_i - E_j = E_{i,loop(i)} + B_{ij} + E_{j,loop(j)} \quad (6)$$

where subscripts  $i$  and  $j$  identify the laboratories and  $loop(i)$  gives the loop number of the laboratory no.  $i$ .

The linking function  $B$  is determined based on the results of both linking laboratories as:

$$\begin{cases} B_{ij} = \frac{\frac{b_{INRIM}}{u^2[b_{INRIM}] + u^2[b_{PTB}]} + \frac{b_{PTB}}{u^2[b_{INRIM}] + u^2[b_{PTB}]} + \delta_B; & loop(i) \neq loop(j) \\ B_{ij} = 0; & loop(i) = loop(j) \end{cases} \quad (7)$$

where the linking value  $b$  for respective linking laboratory is calculated as:

$$b_{INRIM} = -E_{INRIM,loop1} + E_{INRIM,loop2} \quad (8)$$

$$b_{PTB} = -E_{PTB,loop1} + E_{PTB,loop2} \quad (9)$$

The error in the linking function  $\delta_B$  is assumed to be zero. However, due to the deviation of the linking values  $b$ , an uncertainty of the error needs to be included:

$$u(\delta_B) = \frac{1}{2\sqrt{3}} |b_{INRIM} - b_{PTB}| \quad (10)$$

By following well-known principles of uncertainty estimation, the uncertainty of the linking function  $B$  can be given as:

$$\begin{cases} u^2[B_{ij}] = \left( \frac{1}{u^2[b_{INRIM}]} + \frac{1}{u^2[b_{PTB}]} \right)^{-1} + u^2(\delta_B); & loop(i) \neq loop(j) \\ u^2[B_{ij}] = 0; & loop(i) = loop(j) \end{cases} \quad (11)$$

where:

$$u^2[b_{INRIM}] = u^2(E_{INRIM,loop1}) + u^2(E_{INRIM,loop2}) \quad (12)$$

$$-2 \cdot u(\dot{V}_{ref,INRIM,loop1})u(\dot{V}_{ref,INRIM,loop2}) \cdot r_{INRIM}$$

$$u^2[b_{PTB}] = u^2(E_{PTB, loop1}) + u^2(E_{PTB, loop2}) \quad (13)$$

$$-2 \cdot u(\dot{V}_{ref,PTB,loop1})u(\dot{V}_{ref,PTB,loop2}) \cdot r_{PTB}$$

and  $r$  is the correlation coefficient. The results obtained by the linking laboratories in both loops are correlated, because the same reference equipment was used. Based on the uncertainty budgets provided by the linking laboratories, a correlation coefficient of  $r=1$  was used. This is justified, since the uncertainty of the laboratory references is dominated by non-random

uncertainty sources. Anyhow, the influence of the correlation coefficient on the uncertainty of the bilateral equivalence was found insignificant.

As mentioned in section 6.1, the stability of the transfer standard is included in the uncertainties of each laboratory, and therefore it does not show up in the equations above. Table 8 shows the calculated estimates and uncertainties of  $B$ .

Table 8. Calculated estimates of the linking function  $B_{ij}$  when  $loop(i) \neq loop(j)$ .

| Nominal Flow Rate<br>(std ml/min) | $B_{ij}$<br>(%) | $u(B_{ij})$<br>(%) |
|-----------------------------------|-----------------|--------------------|
| <b>Flow cell: -10</b>             |                 |                    |
| 5                                 | 0.10            | 0.21               |
| 10                                | -0.13           | 0.16               |
| 20                                | 0.11            | 0.13               |
| 80                                | 0.18            | 0.02               |
| <b>Flow cell: -24</b>             |                 |                    |
| 80                                | 0.25            | 0.11               |
| 300                               | 0.35            | 0.06               |
| 600                               | 0.35            | 0.04               |
| 1250                              | 0.32            | 0.03               |
| <b>Flow cell: -44</b>             |                 |                    |
| 1250                              | -0.05           | 0.03               |
| 5000                              | -0.04           | 0.04               |
| 10000                             | -0.03           | 0.03               |
| 20000                             | -0.04           | 0.04               |
| 30000                             | 0.06            | 0.05               |

The uncertainty of the bilateral equivalence is calculated as:

$$u^2(D_{ij}) = u^2(E_{i, loop(i)}) + u^2(B_{ij}) + u^2(E_{j, loop(j)}) \quad (14)$$

In the case of linking laboratories the uncertainty is calculated as follows:

$$D_{ij} = E_i - E_j \quad (15)$$

where:

$$E_i = \frac{1}{2}(E_{i, loop1} + E_{i, loop2}) \quad (16)$$

The uncertainty can be expressed as:

$$u^2(E_i) = \left( \frac{1}{2}u(E_{i, loop1}) + \frac{1}{2}u(E_{i, loop2}) \right)^2 \quad (17)$$

## 7.2 Results of analysis

The bilateral degrees of equivalence (DoE) is determined as  $(D_{ij}, U_{ij}) = (D_{ij}, 2 \cdot u_{ij})$  [2]. The DoE was calculated for each pair of participants at each nominal measurement point. The results are summarized in tables 9 to 21. The DoE of equivalence is expressed as the difference of the relative error of the transfer standard determined by the participants.

Table 9. Degree of equivalence ( $D_{ij}$ ) between the participants of EURAMET1325 at the flow rate of 5 ml/min with the low flow cell (-10). DoE are given as relative values (%).

| ↓i \ j→ | <i>DoE ± U(DoE) @ 5 ml/min (flow cell: -10)</i> |              |              |       |             |              |     |              |              |              |
|---------|---|--------------|--------------|-------|-------------|--------------|-----|--------------|--------------|--------------|
|         | INRIM   | PTB          | MIKES        | FORCE | METAS       | VSL          | EIM | UL           | CMI          | LNE          |
| INRIM   |   | 0.01 ± 0.66  | -0.12 ± 0.70 |       | 0.30 ± 0.78 | -0.07 ± 0.79 |     | -0.13 ± 0.61 | -0.31 ± 0.54 | -0.28 ± 0.60 |
| PTB     | -0.01 ± 0.66                                    |              | -0.19 ± 0.77 |       | 0.23 ± 0.83 | -0.14 ± 0.84 |     | -0.09 ± 0.68 | -0.26 ± 0.62 | -0.23 ± 0.67 |
| MIKES   | 0.12 ± 0.70                                     | 0.19 ± 0.77  |              |       | 0.41 ± 0.81 | 0.05 ± 0.81  |     | 0.04 ± 0.72  | -0.13 ± 0.66 | -0.11 ± 0.71 |
| FORCE   |   |              |              |       |             |              |     |              |              |              |
| METAS   | -0.30 ± 0.78                                    | -0.23 ± 0.83 | -0.41 ± 0.81 |       |             | -0.37 ± 0.88 |     | -0.37 ± 0.79 | -0.55 ± 0.74 | -0.52 ± 0.79 |
| VSL     | 0.07 ± 0.79                                     | 0.14 ± 0.84  | -0.05 ± 0.81 |       | 0.37 ± 0.88 |              |     | 0.00 ± 0.80  | -0.18 ± 0.75 | -0.15 ± 0.79 |
| EIM     |   |              |              |       |             |              |     |              |              |              |
| UL      | 0.13 ± 0.61                                     | 0.09 ± 0.68  | -0.04 ± 0.72 |       | 0.37 ± 0.79 | 0.00 ± 0.80  |     |              | -0.18 ± 0.64 | -0.15 ± 0.69 |
| CMI     | 0.31 ± 0.54                                     | 0.26 ± 0.62  | 0.13 ± 0.66  |       | 0.55 ± 0.74 | 0.18 ± 0.75  |     | 0.18 ± 0.64  |              | 0.03 ± 0.63  |
| LNE     | 0.28 ± 0.60                                     | 0.23 ± 0.67  | 0.11 ± 0.71  |       | 0.52 ± 0.79 | 0.15 ± 0.79  |     | 0.15 ± 0.69  | -0.03 ± 0.63 |              |

Table 10. Degree of equivalence ( $D_{ij}$ ) between the participants of EURAMET1325 at the flow rate of 10 ml/min with the low flow cell (-10). DoE are given as relative values (%).

| ↓i \ j→ | <i>DoE ± U(DoE) @ 10 ml/min (flow cell: -10)</i> |              |              |       |             |              |     |              |              |              |
|---------|--|--------------|--------------|-------|-------------|--------------|-----|--------------|--------------|--------------|
|         | INRIM  | PTB          | MIKES        | FORCE | METAS       | VSL          | EIM | UL           | CMI          | LNE          |
| INRIM   |  | -0.24 ± 0.48 | -0.16 ± 0.44 |       | 0.41 ± 0.53 | -0.04 ± 0.57 |     | -0.20 ± 0.39 | -0.20 ± 0.34 | -0.26 ± 0.39 |
| PTB     | 0.24 ± 0.48                                      |              | 0.25 ± 0.54  |       | 0.82 ± 0.61 | 0.38 ± 0.65  |     | -0.13 ± 0.48 | -0.13 ± 0.45 | -0.19 ± 0.48 |
| MIKES   | 0.16 ± 0.44                                      | -0.25 ± 0.54 |              |       | 0.57 ± 0.55 | 0.12 ± 0.59  |     | -0.21 ± 0.45 | -0.21 ± 0.41 | -0.27 ± 0.45 |
| FORCE   |  |              |              |       |             |              |     |              |              |              |
| METAS   | -0.41 ± 0.53                                     | -0.82 ± 0.61 | -0.57 ± 0.55 |       |             | -0.45 ± 0.66 |     | -0.78 ± 0.54 | -0.78 ± 0.50 | -0.84 ± 0.54 |
| VSL     | 0.04 ± 0.57                                      | -0.38 ± 0.65 | -0.12 ± 0.59 |       | 0.45 ± 0.66 |              |     | -0.33 ± 0.58 | -0.33 ± 0.55 | -0.39 ± 0.58 |
| EIM     |  |              |              |       |             |              |     |              |              |              |
| UL      | 0.20 ± 0.39                                      | 0.13 ± 0.48  | 0.21 ± 0.45  |       | 0.78 ± 0.54 | 0.33 ± 0.58  |     |              | 0.00 ± 0.39  | -0.06 ± 0.43 |
| CMI     | 0.20 ± 0.34                                      | 0.13 ± 0.45  | 0.21 ± 0.41  |       | 0.78 ± 0.50 | 0.33 ± 0.55  |     | 0.00 ± 0.39  |              | -0.06 ± 0.39 |
| LNE     | 0.26 ± 0.39                                      | 0.19 ± 0.48  | 0.27 ± 0.45  |       | 0.84 ± 0.54 | 0.39 ± 0.58  |     | 0.06 ± 0.43  | 0.06 ± 0.39  |              |

Table 11. Degree of equivalence ( $D_{ij}$ ) between the participants of EURAMET1325 at the flow rate of 20 ml/min with the low flow cell (-10). DoE are given as relative values (%).

| ↓i \ j→ | DoE ± U(DoE) @ 20 ml/min (flow cell: -10) |              |              |       |             |              |              |              |              |              |
|---------|---|--------------|--------------|-------|-------------|--------------|--------------|--------------|--------------|--------------|
|         | INRIM                                     | PTB          | MIKES        | FORCE | METAS       | VSL          | EIM          | UL           | CMI          | LNE          |
| INRIM   |   | -0.19 ± 0.47 | -0.04 ± 0.33 |       | 0.24 ± 0.41 | 0.07 ± 0.48  | -0.28 ± 0.37 | -0.10 ± 0.48 | -0.12 ± 0.46 | -0.11 ± 0.47 |
| PTB     | 0.19 ± 0.47                               |              | 0.10 ± 0.44  |       | 0.39 ± 0.51 | 0.22 ± 0.57  | -0.13 ± 0.48 | 0.13 ± 0.56  | 0.11 ± 0.55  | 0.12 ± 0.56  |
| MIKES   | 0.04 ± 0.33                               | -0.10 ± 0.44 |              |       | 0.29 ± 0.45 | 0.11 ± 0.51  | -0.24 ± 0.41 | -0.02 ± 0.44 | -0.04 ± 0.42 | -0.03 ± 0.44 |
| FORCE   |   |              |              |       |             |              |              |              |              |              |
| METAS   | -0.24 ± 0.41                              | -0.39 ± 0.51 | -0.29 ± 0.45 |       |             | -0.17 ± 0.57 | -0.52 ± 0.48 | -0.30 ± 0.51 | -0.32 ± 0.49 | -0.31 ± 0.51 |
| VSL     | -0.07 ± 0.48                              | -0.22 ± 0.57 | -0.11 ± 0.51 |       | 0.17 ± 0.57 |              | -0.35 ± 0.54 | -0.13 ± 0.57 | -0.15 ± 0.55 | -0.14 ± 0.56 |
| EIM     | 0.28 ± 0.37                               | 0.13 ± 0.48  | 0.24 ± 0.41  |       | 0.52 ± 0.48 | 0.35 ± 0.54  |              | 0.22 ± 0.48  | 0.20 ± 0.46  | 0.21 ± 0.47  |
| UL      | 0.10 ± 0.48                               | -0.13 ± 0.56 | 0.02 ± 0.44  |       | 0.30 ± 0.51 | 0.13 ± 0.57  | -0.22 ± 0.48 |              | -0.02 ± 0.48 | -0.01 ± 0.50 |
| CMI     | 0.12 ± 0.46                               | -0.11 ± 0.55 | 0.04 ± 0.42  |       | 0.32 ± 0.49 | 0.15 ± 0.55  | -0.20 ± 0.46 | 0.02 ± 0.48  |              | 0.01 ± 0.48  |
| LNE     | 0.11 ± 0.47                               | -0.12 ± 0.56 | 0.03 ± 0.44  |       | 0.31 ± 0.51 | 0.14 ± 0.56  | -0.21 ± 0.47 | 0.01 ± 0.50  | -0.01 ± 0.48 |              |

Table 12. Degree of equivalence ( $D_{ij}$ ) between the participants of EURAMET1325 at the flow rate of 80 ml/min with the low flow cell (-10). DoE are given as relative values (%).

| ↓i \ j→ | DoE ± U(DoE) @ 80 ml/min (flow cell: -10) |              |              |       |              |             |              |              |              |              |
|---------|---|--------------|--------------|-------|--------------|-------------|--------------|--------------|--------------|--------------|
|         | INRIM                                     | PTB          | MIKES        | FORCE | METAS        | VSL         | EIM          | UL           | CMI          | LNE          |
| INRIM   |   | -0.14 ± 0.31 | -0.12 ± 0.17 |       | -0.07 ± 0.33 | 0.14 ± 0.41 | -0.11 ± 0.23 | -0.21 ± 0.18 | -0.22 ± 0.13 | -0.28 ± 0.14 |
| PTB     | 0.14 ± 0.31                               |              | 0.00 ± 0.34  |       | 0.04 ± 0.44  | 0.26 ± 0.51 | 0.01 ± 0.37  | -0.04 ± 0.34 | -0.04 ± 0.32 | -0.11 ± 0.32 |
| MIKES   | 0.12 ± 0.17                               | 0.00 ± 0.34  |              |       | 0.05 ± 0.36  | 0.26 ± 0.44 | 0.01 ± 0.27  | -0.04 ± 0.22 | -0.05 ± 0.19 | -0.11 ± 0.20 |
| FORCE   |   |              |              |       |              |             |              |              |              |              |
| METAS   | 0.07 ± 0.33                               | -0.04 ± 0.44 | -0.05 ± 0.36 |       |              | 0.21 ± 0.52 | -0.04 ± 0.39 | -0.09 ± 0.36 | -0.09 ± 0.34 | -0.16 ± 0.35 |
| VSL     | -0.14 ± 0.41                              | -0.26 ± 0.51 | -0.26 ± 0.44 |       | -0.21 ± 0.52 |             | -0.25 ± 0.46 | -0.30 ± 0.44 | -0.31 ± 0.42 | -0.38 ± 0.42 |
| EIM     | 0.11 ± 0.23                               | -0.01 ± 0.37 | -0.01 ± 0.27 |       | 0.04 ± 0.39  | 0.25 ± 0.46 |              | -0.05 ± 0.27 | -0.06 ± 0.24 | -0.12 ± 0.25 |
| UL      | 0.21 ± 0.18                               | 0.04 ± 0.34  | 0.04 ± 0.22  |       | 0.09 ± 0.36  | 0.30 ± 0.44 | 0.05 ± 0.27  |              | 0.00 ± 0.19  | -0.07 ± 0.19 |
| CMI     | 0.22 ± 0.13                               | 0.04 ± 0.32  | 0.05 ± 0.19  |       | 0.09 ± 0.34  | 0.31 ± 0.42 | 0.06 ± 0.24  | 0.00 ± 0.19  |              | -0.07 ± 0.16 |
| LNE     | 0.28 ± 0.14                               | 0.11 ± 0.32  | 0.11 ± 0.20  |       | 0.16 ± 0.35  | 0.38 ± 0.42 | 0.12 ± 0.25  | 0.07 ± 0.19  | 0.07 ± 0.16  |              |













## 8. EURAMET comparison reference values (ERV)

### 8.1 ERV calculation method

In this project, the calculation of the EURAMET comparison reference value (ERV) is carried out according to Heinonen [1], which is based on the principles presented by M. Cox [2, 3]. Because there was no common transfer standard for all participants, absolute ERV values were not determined. Only the difference between ERVs and the results of each laboratory were calculated.

The calculations were carried out using the weighed mean of the results normalised to the loop of the laboratory under study ( $i$ ). The normalisation was realised using the linking function  $B$  defined by equation (7):

$$\Delta E_i = E_i - E_{ERV} = E_i - \frac{\sum_{j=1}^N \frac{E_j}{u^2(E_j)}}{\sum_{j=1}^N \frac{1}{u^2(E_j)}} = E_i - \frac{\sum_{j=1}^N \frac{E_j + B_{ij}}{u^2(E_j) + u^2(B_{ij})}}{\sum_{j=1}^N \frac{1}{u^2(E_j) + u^2(B_{ij})}} \quad (18)$$

$$u^2(\Delta E_i) = \left( \sum_{j=1}^N [u^2(E_j) + u^2(B_{ij})] \right)^{-1} + u^2(E_i) \quad (19)$$

where  $N$  is the number of participants contributing to the ERV and  $E_j$  is the result of  $j$ th laboratory included in the ERV calculation.

Due to participation in both loops, the link laboratories form a special case, where results are combined in the following way:

$$\Delta E_i = \frac{1}{2} (\Delta E_{i, loop1} + \Delta E_{i, loop2}) + \delta \Delta E_i \quad (20)$$

$$u^2(\Delta E_i) = \left[ \frac{1}{2} u(\Delta E_{i, loop1}) + \frac{1}{2} u(\Delta E_{i, loop2}) \right]^2 + u^2(\delta \Delta E_i) \quad (21)$$

where  $\delta \Delta E_i = 0$  and

$$u(\delta \Delta E_i) = \frac{1}{2\sqrt{3}} |\Delta E_{i, loop1} - \Delta E_{i, loop2}| \quad (22)$$

Note that the results of the stability measurements were not included in the calculations, because the linking laboratories performed stability measurements only for one transfer standard.

### 8.2 Consistency check

The consistency of the participating laboratories results was evaluated by performing a chi-square test applying principles presented by M. Cox [2, 3]. As a first step, the chi-squared value was calculated as:

$$\chi_{obs}^2 = \sum_{i=1}^n \frac{(E_i - E_{ERV})^2}{u^2(\Delta E_i)} = \sum_{i=1}^n \frac{(\Delta E_i)^2}{u^2(\Delta E_i)} \quad (23)$$

where  $n$  is the number of laboratories contributing the reference value. The results were evaluated against the 95 % confidence interval. The consistency test was performed for each flow rate separately. In cases where the consistency check failed, outliers were identified as results deviating more than two times the standard uncertainty:

$$|\Delta E_i| > 2 \cdot u(\Delta E_i) \tag{24}$$

As a next step, the outlier deviating the most is removed from the calculation of the ERV at that specific flow rate and the chi-square test described above is repeated. This iterative process is repeated until a consistent subset is achieved. The outcome of the consistency check including chi-square values is summarized in table 22.

Table 22. Outcome of consistency check including chi-square values.

| Chi-square value for each participant $(\Delta E_i)^2 / u^2(\Delta E_i)$ |      |       |       |       |       |      |      |      |      | Chi-square value (sum) | Degrees of freedom | Chi-square criteria | Laboratories excluded from ERV calculations |
|--|------|-------|-------|-------|-------|------|------|------|------|------------------------|--------------------|---------------------|---|
| INRIM  | PTB  | MIKES | FORCE | METAS | VSL   | EIM  | UL   | CMI  | LNE  |                        |                    |                     |   |
| (%)  | (%)  | (%)   | (%)   | (%)   | (%)   | (%)  | (%)  | (%)  | (%)  |                        |                    |                     |   |
| <b>Flow cell: -10</b>  |      |       |       |       |       |      |      |      |      |                        |                    |                     |   |
| 0.13   | 0.12 | 0.07  |       | 1.10  | 0.01  |      | 0.00 | 0.70 | 0.38 | 2.5                    | 7                  | 14.1                |   |
| 0.66   | 0.30 | 0.00  |       | 6.00  | 0.21  |      | 0.47 | 0.73 | 1.10 | 9.5                    | 7                  | 14.1                |   |
| 0.16   | 0.40 | 0.01  |       | 2.01  | 0.19  | 2.31 | 0.00 | 0.04 | 0.02 | 5.1                    | 8                  | 15.6                |   |
| 5.43   | 0.05 | 0.22  |       | 0.00  | 1.19  | 0.06 | 0.76 | 1.65 | 5.23 | 14.6                   | 8                  | 15.6                |   |
| <b>Flow cell: -24</b>  |      |       |       |       |       |      |      |      |      |                        |                    |                     |   |
| 0.06   | 0.01 | 1.04  | 0.28  | 3.19  | 10.50 | 4.74 | 0.36 | 0.32 |      | 10.0                   | 7                  | 14.1                | VSL   |
| 0.75   | 0.03 | 0.80  | 1.64  | 3.42  | 0.69  | 4.91 | 0.05 | 0.40 |      | 12.7                   | 8                  | 15.6                |   |
| 0.84   | 0.05 | 1.61  | 2.46  | 2.59  | 1.39  | 4.69 | 0.04 | 1.65 |      | 15.3                   | 8                  | 15.6                |   |
| 0.32   | 0.14 | 1.48  | 9.22  | 1.19  | 1.45  | 2.52 | 0.69 | 1.00 |      | 8.8                    | 7                  | 14.1                | FORCE                                       |
| <b>Flow cell: -44</b>  |      |       |       |       |       |      |      |      |      |                        |                    |                     |   |
| 0.00   | 0.04 | 1.75  | 1.58  | 0.52  | 0.76  | 0.38 | 0.41 | 0.09 |      | 5.5                    | 8                  | 15.6                |   |
| 0.08   | 0.00 | 0.00  | 0.00  | 0.02  | 0.05  | 0.06 | 0.39 | 0.00 |      | 0.6                    | 8                  | 15.6                |   |
| 0.18   | 0.01 | 0.01  | 0.54  | 0.15  | 0.50  |      | 0.51 | 0.53 |      | 2.4                    | 7                  | 14.1                |   |
| 0.22   | 0.03 | 0.04  | 0.01  | 0.01  | 0.17  | 0.27 | 0.61 | 0.31 |      | 1.7                    | 8                  | 15.6                |   |
| 0.08   | 0.10 | 0.08  | 18.84 | 0.02  | 0.01  | 0.05 | 0.10 | 0.09 |      | 0.5                    | 7                  | 14.1                | FORCE                                       |

### 8.3 Final results of the ERV analysis

The results of all laboratories are compared to the ERV according to the analysis described in the previous sections. The results are summarized below in figures 12 to 14 and table 23. All uncertainties are at the 95 % confidence level ( $k=2$ ). Table 24 summarizes the ERVs. The ERVs are shown for each laboratory because they depend on the loop. Also, the ERVs for the link laboratories differ from the others (as seen from eq. 21).




Figure 12. Difference between the ERV and the results of the laboratories for the low flow cell ML-800-10. Error bars show expanded uncertainties ( $k=2$ ). Note that the x-axis values are shifted for clarity of presentation.




Figure 13. Difference between the ERV and the results of the laboratories for the medium flow cell ML-800-24. Error bars show expanded uncertainties ( $k=2$ ). Note that the x-axis values are shifted for clarity of presentation.




Figure 14. Difference between the ERV and the results of the laboratories for the high flow cell ML-800-44. Error bars show expanded uncertainties ( $k=2$ ). Note that the x-axis values are shifted for clarity of presentation.



Table 23. Estimates of the difference between the ERVs and the results of the laboratories expressed as relative values (%). Uncertainties are given as expanded uncertainties ( $k=2$ ).

| Nominal Flow Rate     | INRIM        | PTB          | MIKES        | FORCE        | METAS        | VSL          | EIM          | UL           | CMI         | LNE         |
|-----------------------|--------------|--------------|--------------|--------------|--------------|--------------|--------------|--------------|-------------|-------------|
| (std ml/min)          | (%)          | (%)          | (%)          | (%)          | (%)          | (%)          | (%)          | (%)          | (%)         | (%)         |
| <b>Flow cell: -10</b> |              |              |              |              |              |              |              |              |             |             |
| 5                     | -0.08 ± 0.46 | -0.10 ± 0.55 | 0.07 ± 0.56  |              | -0.34 ± 0.65 | 0.03 ± 0.66  |              | 0.01 ± 0.52  | 0.18 ± 0.44 | 0.16 ± 0.51 |
| 10                    | -0.12 ± 0.30 | 0.12 ± 0.45  | 0.01 ± 0.36  |              | -0.56 ± 0.46 | -0.12 ± 0.51 |              | 0.11 ± 0.33  | 0.12 ± 0.27 | 0.17 ± 0.33 |
| 20                    | -0.06 ± 0.28 | 0.13 ± 0.41  | 0.02 ± 0.28  |              | -0.27 ± 0.38 | -0.10 ± 0.45 | 0.25 ± 0.33  | 0.01 ± 0.38  | 0.04 ± 0.35 | 0.02 ± 0.37 |
| 80                    | -0.11 ± 0.09 | 0.03 ± 0.30  | 0.04 ± 0.17  |              | -0.01 ± 0.33 | -0.22 ± 0.41 | 0.03 ± 0.22  | 0.07 ± 0.16  | 0.08 ± 0.12 | 0.14 ± 0.13 |
| <b>Flow cell: -24</b> |              |              |              |              |              |              |              |              |             |             |
| 80                    | -0.02 ± 0.19 | -0.02 ± 0.37 | -0.18 ± 0.35 | 0.10 ± 0.37  | -0.32 ± 0.36 | -0.83 ± 0.51 | 0.42 ± 0.38  | 0.05 ± 0.16  | 0.03 ± 0.11 |             |
| 300                   | -0.06 ± 0.14 | 0.03 ± 0.33  | -0.10 ± 0.21 | 0.16 ± 0.24  | -0.21 ± 0.23 | 0.11 ± 0.26  | 0.29 ± 0.26  | -0.02 ± 0.19 | 0.05 ± 0.16 |             |
| 600                   | -0.05 ± 0.10 | 0.03 ± 0.32  | -0.13 ± 0.20 | 0.18 ± 0.23  | -0.17 ± 0.21 | 0.14 ± 0.24  | 0.34 ± 0.32  | 0.02 ± 0.16  | 0.07 ± 0.11 |             |
| 1250                  | -0.03 ± 0.09 | 0.06 ± 0.31  | -0.11 ± 0.18 | 0.36 ± 0.24  | -0.11 ± 0.20 | 0.13 ± 0.22  | 0.43 ± 0.54  | 0.07 ± 0.16  | 0.06 ± 0.12 |             |
| <b>Flow cell: -44</b> |              |              |              |              |              |              |              |              |             |             |
| 1250                  | 0.00 ± 0.21  | 0.03 ± 0.31  | -0.11 ± 0.16 | 0.12 ± 0.19  | -0.07 ± 0.19 | -0.09 ± 0.21 | 0.06 ± 0.19  | 0.05 ± 0.17  | 0.02 ± 0.13 |             |
| 5000                  | -0.03 ± 0.21 | 0.00 ± 0.32  | 0.00 ± 0.17  | 0.00 ± 0.19  | 0.01 ± 0.18  | -0.02 ± 0.21 | -0.02 ± 0.17 | 0.06 ± 0.20  | 0.00 ± 0.17 |             |
| 10000                 | -0.04 ± 0.21 | 0.02 ± 0.31  | 0.01 ± 0.17  | -0.07 ± 0.20 | -0.04 ± 0.20 | -0.07 ± 0.21 |              | 0.06 ± 0.17  | 0.05 ± 0.13 |             |
| 20000                 | -0.05 ± 0.20 | -0.03 ± 0.31 | -0.02 ± 0.20 | -0.01 ± 0.19 | 0.01 ± 0.19  | -0.04 ± 0.21 | -0.07 ± 0.25 | 0.07 ± 0.18  | 0.04 ± 0.16 |             |
| 30000                 | -0.03 ± 0.18 | -0.05 ± 0.31 | 0.03 ± 0.20  | -0.44 ± 0.20 | -0.01 ± 0.20 | -0.01 ± 0.22 | -0.03 ± 0.29 | 0.03 ± 0.17  | 0.05 ± 0.31 |             |

Table 24. ERVs and associated expanded uncertainties ( $k=2$ ) for each participant expressed as relative values (%).

| Nominal Flow Rate<br>(std ml/min) | INRIM<br>(%) | PTB<br>(%)   | MIKES<br>(%) | FORCE<br>(%) | METAS<br>(%) | VSL<br>(%)   | EIM<br>(%)   | UL<br>(%)    | CMI<br>(%)   | LNE<br>(%)   |
|-----------------------------------|--------------|--------------|--------------|--------------|--------------|--------------|--------------|--------------|--------------|--------------|
| <b>Flow cell: -10</b>             |              |              |              |              |              |              |              |              |              |              |
| 5                                 | -0.16 ± 0.19 | -0.16 ± 0.19 | -0.22 ± 0.20 |              | -0.22 ± 0.20 | -0.22 ± 0.20 |              | -0.10 ± 0.17 | -0.10 ± 0.17 | -0.10 ± 0.17 |
| 10                                | -0.05 ± 0.13 | -0.05 ± 0.13 | -0.04 ± 0.14 |              | -0.04 ± 0.14 | -0.04 ± 0.14 |              | -0.06 ± 0.12 | -0.06 ± 0.12 | -0.06 ± 0.12 |
| 20                                | -0.06 ± 0.12 | -0.06 ± 0.12 | -0.12 ± 0.12 |              | -0.12 ± 0.12 | -0.12 ± 0.12 | -0.12 ± 0.12 | 0.01 ± 0.13  | 0.01 ± 0.13  | 0.01 ± 0.13  |
| 80                                | 0.02 ± 0.05  | 0.02 ± 0.05  | -0.07 ± 0.05 |              | -0.07 ± 0.05 | -0.07 ± 0.05 | -0.07 ± 0.05 | 0.11 ± 0.05  | 0.11 ± 0.05  | 0.11 ± 0.05  |
| <b>Flow cell: -24</b>             |              |              |              |              |              |              |              |              |              |              |
| 80                                | -0.29 ± 0.08 | -0.29 ± 0.08 | -0.41 ± 0.11 | -0.41 ± 0.11 | -0.41 ± 0.11 | -0.41 ± 0.11 | -0.41 ± 0.11 | -0.17 ± 0.04 | -0.17 ± 0.04 |              |
| 300                               | -0.21 ± 0.07 | -0.21 ± 0.07 | -0.37 ± 0.07 | -0.37 ± 0.07 | -0.37 ± 0.07 | -0.37 ± 0.07 | -0.37 ± 0.07 | -0.04 ± 0.06 | -0.04 ± 0.06 |              |
| 600                               | -0.20 ± 0.05 | -0.20 ± 0.05 | -0.36 ± 0.06 | -0.36 ± 0.06 | -0.36 ± 0.06 | -0.36 ± 0.06 | -0.36 ± 0.06 | -0.04 ± 0.04 | -0.04 ± 0.04 |              |
| 1250                              | -0.26 ± 0.05 | -0.26 ± 0.05 | -0.42 ± 0.06 | -0.42 ± 0.06 | -0.42 ± 0.06 | -0.42 ± 0.06 | -0.42 ± 0.06 | -0.11 ± 0.04 | -0.11 ± 0.04 |              |
| <b>Flow cell: -44</b>             |              |              |              |              |              |              |              |              |              |              |
| 1250                              | 0.02 ± 0.06  | 0.02 ± 0.04  | 0.04 ± 0.06  | 0.04 ± 0.06  | 0.04 ± 0.06  | 0.04 ± 0.06  | 0.04 ± 0.06  | 0.00 ± 0.06  | 0.00 ± 0.06  |              |
| 5000                              | -0.03 ± 0.06 | -0.03 ± 0.05 | -0.01 ± 0.06 | -0.01 ± 0.06 | -0.01 ± 0.06 | -0.01 ± 0.06 | -0.01 ± 0.06 | -0.05 ± 0.07 | -0.05 ± 0.07 |              |
| 10000                             | -0.04 ± 0.06 | -0.04 ± 0.05 | -0.02 ± 0.06 | -0.02 ± 0.06 | -0.02 ± 0.06 | -0.02 ± 0.06 |              | -0.05 ± 0.06 | -0.05 ± 0.06 |              |
| 20000                             | -0.03 ± 0.06 | -0.03 ± 0.05 | -0.02 ± 0.06 | -0.02 ± 0.06 | -0.02 ± 0.06 | -0.02 ± 0.06 | -0.02 ± 0.06 | -0.05 ± 0.07 | -0.05 ± 0.07 |              |
| 30000                             | -0.03 ± 0.07 | -0.03 ± 0.06 | -0.05 ± 0.07 | -0.05 ± 0.07 | -0.05 ± 0.07 | -0.05 ± 0.07 | -0.05 ± 0.07 | 0.00 ± 0.07  | -0.01 ± 0.07 |              |

## 9. Conclusion and discussion

---

### 9.1 General

The comparison had a rough start as the transfer standards broke down in the very beginning. The instruments were sent for repair, and as a result, the comparison had to be re-started. This time, the comparison was successfully completed, without any significant problems reported by the participants.

Past EURAMET comparisons [4, 5, 6], Molbloc flow standards (sonic nozzles and laminar flow elements) have been used as high-precision transfer standards. This time it was decided to use high-end piston provers, instead. The nominal accuracy of such gas flow meters is similar to the Molblobs. The main reasons for this choice was to gain experience of using such flow meters for inter-comparisons, and to investigate the laboratories competence in calibrating this type of flow meters.

Based on the experience and the results of this comparison, the transfer standards were found suitable for high-level inter-comparisons with some reservations. Special care should be taken when handling and transporting these delicate instruments to avoid damage and to minimize possible shifts in the response. Especially, the low flow cell indicated a change in the response at flow rates below 20 ml/min. To reduce the risks of such an event, the amount of participants in each loop should be reduced, or alternatively intermediate stability measurements should be performed to better characterize and compensate for the shift. Anyway, the uncertainty caused by the transfer standards was in many cases smaller than the participating laboratories calibration uncertainties, and thus it had only a small effect on the overall uncertainty.

The comparison method applied in this project was found successful, and the results can be used for reviewing the CMCs of the participants. Uncertainty estimations carried out by the participants seem to be realistic.

### 9.2 Identified discrepancies in the results

Generally a good agreement between laboratories was found, with only a few exceptions. It was concluded that the observed discrepancies are in most cases not caused by stability problems of the transfer standards. Discrepant results were identified as results deviating more than the estimated expanded uncertainty (see tables 23 and 25, and figures 12 - 14). Results with normalized error values  $|E_n| > 1$  are considered discrepant and highlighted in red in table 25. Results where the standard uncertainty of the transfer standard is larger than the reported measurement uncertainty ( $k=2$ ) is highlighted in orange. In these cases, the results are inconclusive for evaluating the validity of reported measurement uncertainties. Such results were mainly found for the low flow cell.

Table 25. Normalized error values  $E_n$  for each participant. Discrepant results ( $|E_n| > 1$ ) are shown in red and inconclusive results in orange.

| Normalized error values $E_n$ $E_n = \Delta E_i / U(\Delta E_i)$ |           |         |           |           |           |         |         |        |         |         |
|--|-----------|---------|-----------|-----------|-----------|---------|---------|--------|---------|---------|
| Nominal Flow Rate (std ml/min)                                   | INRIM (%) | PTB (%) | MIKES (%) | FORCE (%) | METAS (%) | VSL (%) | EIM (%) | UL (%) | CMI (%) | LNE (%) |
| <b>Flow cell: -10</b>  |           |         |           |           |           |         |         |        |         |         |
| 5  | -0.2      | -0.2    | 0.1       |           | -0.5      | 0.0     |         | 0.0    | 0.4     | 0.3     |
| 10   | -0.4      | 0.3     | 0.0       |           | -1.2      | -0.2    |         | 0.3    | 0.4     | 0.5     |
| 20   | -0.2      | 0.3     | 0.1       |           | -0.7      | -0.2    | 0.8     | 0.0    | 0.1     | 0.1     |
| 80   | -1.2      | 0.1     | 0.2       |           | 0.0       | -0.5    | 0.1     | 0.4    | 0.6     | 1.1     |
| <b>Flow cell: -24</b>  |           |         |           |           |           |         |         |        |         |         |
| 80   | -0.1      | 0.0     | -0.5      | 0.3       | -0.9      | -1.6    | 1.1     | 0.3    | 0.3     |         |
| 300  | -0.4      | 0.1     | -0.4      | 0.6       | -0.9      | 0.4     | 1.1     | -0.1   | 0.3     |         |
| 600  | -0.5      | 0.1     | -0.6      | 0.8       | -0.8      | 0.6     | 1.1     | 0.1    | 0.6     |         |
| 1250   | -0.3      | 0.2     | -0.6      | 1.5       | -0.5      | 0.6     | 0.8     | 0.4    | 0.5     |         |
| <b>Flow cell: -44</b>  |           |         |           |           |           |         |         |        |         |         |
| 1250   | 0.0       | 0.1     | -0.7      | 0.6       | -0.4      | -0.4    | 0.3     | 0.3    | 0.2     |         |
| 5000   | -0.1      | 0.0     | 0.0       | 0.0       | 0.1       | -0.1    | -0.1    | 0.3    | 0.0     |         |
| 10000  | -0.2      | 0.1     | 0.1       | -0.4      | -0.2      | -0.4    |         | 0.4    | 0.4     |         |
| 20000  | -0.2      | -0.1    | -0.1      | 0.0       | 0.1       | -0.2    | -0.3    | 0.4    | 0.3     |         |
| 30000  | -0.1      | -0.2    | 0.1       | -2.2      | -0.1      | 0.0     | -0.1    | 0.2    | 0.1     |         |

A detailed description of identified discrepancies is given below:

- INRIM - Discrepant results were found at flow rate 80 ml/min with the low flow cell (-10). The result was, however, well in-line with the rest of the INRIM results with the low flow cell (-10), but due to the very low uncertainty it was classified as discrepant. On the other hand, a good agreement was found at 80 ml/min with the medium flow cell (-24).
- FORCE - Discrepant results were found at flow rate 1250 ml/min with the medium flow cell (-24) and at flow rate 30 000 ml/min with the high flow cell (-44). Interestingly, the results at 1250 ml/min with the high flow cell (-44) was found consistent. The laboratory suspects that the pressure pulsation of the transfer standard might have influenced the results of the primary standard, especially at the higher flow rates, similar to what was observed with the low flow cell (-10). In addition, FORCE did not perform measurement with the low flow cell, because the pulsation induced by the transfer standard piston was found to influence the results of the primary standard. Therefore, FORCE decided not to participate in the comparison of the low flow cell (ML-800-10).
- METAS - A discrepant results was found at flow rate 10 ml/min with the low flow cell (-10).
- VSL - The results at flow rate 80 ml/min with the medium flow cell (-24) was found discrepant. However, a good agreement was found with the low flow cell (-10) at the same flow rate.
- EIM - Slightly discrepant results were found at flow rates 80 ml/min, 300 ml/min and 600 ml/min with the medium flow cell (-24). Again, at flow rate 80 ml/min the results with the low flow cell (-10) show a good agreement. Discussion on redundant measurements performed with different flow cells is given in the next section.
- LNE - The results at flow rate 80 ml/min with the low flow cell (-10) was found slightly discrepant. As with INRIM, the result was well in-line with the rest of the LNE results, but due to the very low uncertainty it was classified as discrepant. LNE performed

measurement only with the low flow cell (-10), and therefore a comparison at redundant measurement point with the medium flow cell (-24) could not be made.

### 9.3 Discussion

Redundant measurement were performed at 80 ml/min and 1250 ml/min in order to investigate the influence of the transfer standard on the measurement results. Clear differences in the results for the different flow cells were found for many laboratories in loop 1, whereas the results were consistent for laboratories in loop 2 (see table 23 and figures 12, 13 and 14). At flow rate 80 ml/min MIKES, METAS, VSL and EIM got clearly different results with the low (-10) and medium flow cells (-24). Similar findings were made at flow rate 1250 ml/min with the medium (-24) and high flow cells (-44), but this time only FORCE, VSL and EIM got inconsistent results.

These findings are discussed in more detail below for each laboratory:

- EIM - At 80 ml/min and 1250 ml/min the result with the medium flow cell (-24) are inconsistent with the other laboratories, whereas results with the low flow cell (-10) at 80 ml/min and with the high flow cell (-44) at 1250 ml/min show good agreement. Comparing the results of EIM for all flow cells, it seems that there is an offset of +0.3% to +0.4 % in the results of the medium flow cell (-24).
- VSL - The result at flow rate 80 ml/min with the medium flow cell (-24) was found discrepant and inconsistent with the rest of the VSL results. At 1250 ml/min there is a clear difference in the VSL results of the medium (-24) and high flow cell (-44). Similar to the EIM results, it seems that, apart from the 80 ml/min measurement point, there is an offset of +0.2 % to +0.3 % in the results of the medium flow cell (-24) compared to the other flow cells. In this case, however, the results with the medium flow cell (-24) were found to agree within estimated expanded uncertainties, except for the 80 ml/min measurement point.
- METAS - The result at 80 ml/min with the medium flow cell (-24) was lower than the results of other laboratories but still within estimated expanded uncertainties, while the result at 80 ml/min with the low flow cell (-10) shows good agreement. However, all METAS results, also the discrepant result at 10 ml/min with the low flow cell (-10), are perfectly concordant with the INRIM2 results. This could lead to the interpretation, that METAS was already subject to the equal shift of medium and low flow cells of the transfer standard as measured by INRIM at the end of the loop (see figures 4, 6 and 8). Therefore, the discrepancies found may not be real but caused by the shift of the transfer standard.
- FORCE - The results at 1250 ml/min with the medium flow cell was found discrepant, while the results at the same flow rate with the high flow cell (-44) agree with the other laboratories. As with EIM and VSL, there seems to be an offset of about +0.2 % in the results of the medium flow cell.
- MIKES - At flow rate 80 ml/min there is a deviation in the results of the low (-10) and medium flow cells (-24). Also in the MIKES results, a small offset of about -0.1 % can be seen in the medium flow cell (-24) results when comparing to the other flow cells. This time, however, the offset is opposite to what was observed in the results of EIM, VSL and FORCE. Anyhow, the deviation is rather small and all results were found to agree within estimated expanded uncertainties.

The spread of the laboratories results was larger for the low (-10) and medium flow cells (-24) than for the high flow cell (-44). Also the instability of these flow cells was larger, especially in the lower end of the measurement range. It can be clearly seen from the results of the low flow

cells (-10) (figures 6 and 7) that the spread of the results is larger at low flow rates where the instability was found larger. However, a clear trend in response cannot be seen from the results (laboratories are placed in measurement order in the figures 6 to 11). The measured shift in the response of the transfer standards is probably also to some extent influenced by random effects caused by unideal reproducibility of the flow cell indications.

Initial tests on the transfer standards and previous studies [5] indicate that piston prover measurements are sensitive to the heat exchange effect that results from differences between the piston-cylinder temperature and the gas temperature. This effect is larger for the low (-10) and medium (-24) flow cells that require long time for temperature stabilization, because of small gas flow rates. Due to these effects, a long stabilization time of 180 min was included in the measurement protocol. However, if the surrounding temperature conditions are not stable enough, such temperature differences could occur during measurement even with long stabilization period; e.g., the cylinder with a high thermal capacity is stabilized at different temperature than the inlet gas flow, which temperature mainly depends on the temperature of the connection lines. Such temperature related effects might explain to some extent the observed spread of result for the low (-10) and medium (-24) flow cells.

The reasons for the constant offset in the results of some laboratories for the medium flow cell (-24) remains unclear. As this observation was only made for laboratories in loop 1, the possibility of an instrument malfunction cannot be completely ruled out. However, this seems unlikely based on the measurements performed after the comparison. An alternative explanation for the deviations could be an interference between the calibration system and the transfer standard. Launching the piston introduces a pressure pulse that might influence the flow stability and thus the operation of the calibration system depending on the design. Anyhow, further investigations are necessary to understand the reasons for the deviations.

This comparison provided valuable experience of using piston provers as high precision transfer standards. Such an extensive inter-comparison using piston provers has not been done before. As such, it gives a good overview on the capabilities of European laboratories of calibrating such instruments that are commonly used in industrial calibration laboratories worldwide.

## References

---

- [1] Heinonen, Martti, et al. Investigation of the equivalence of national dew-point temperature realizations in the  $-50^{\circ}\text{C}$  to  $+20^{\circ}\text{C}$  range. *International journal of thermophysics*, 2012, 33.8-9: 1422-1437.
- [2] Cox, M. G. The evaluation of key comparison data. *Metrologia*, 2002, 39.6: 589.
- [3] Cox, M. G. The evaluation of key comparison data: determining the largest consistent subset. *Metrologia*, 2007, 44.3: 187.
- [4] Probert, C., Johns, A. I., Metaxiotou, Z. An intercomparison of low flow gas facilities at eleven European laboratories using a Molbloc transfer package: EURAMET. MM. FF-S3 (EURAMET Project no 806). *Metrologia*, 2008, 45.1A: 07012.
- [5] Kutin, J., G. Bobovnik, and I. Bajsić. Heat exchange effects on the performance of a clearance-sealed piston prover for gas flow measurements. *Metrologia* 52.6 (2015): 857.
- [6] B. Niederhauser and J. Barbe. Bilateral comparison of primary low-gas-flow standards between the BNM-LNE and METAS. *Metrologia*, 2002, 39.6: 573.

- [7] S. Sillanpää, B. Niederhauser and M. Heinonen, Comparison of the primary low gasflow standards between MIKES and METAS, Measurement, 2006, 39, 26 - 33.

## Appendix 1: Detailed information of participants

---

Table A1. Updated contact information

| Name of the laboratory   | Country     | Address  | Contact               | E-mail                         |
|--|-------------|--|-----------------------|--------------------------------|
| Istituto Nazionale di Ricerca Metrologica (INRIM)                                | Italy       | Strada delle Cacce, 91, I-10135 Torino                                     | Pier Giorgio Spazzini | piergiorgio.spazzini@polito.it |
| Physikalisch-Technische Bundesanstalt (PTB)                                      | Germany     | Bundesallee 100, 38116 Braunschweig  | Rainer Kramer         | rainer.kramer@ptb.de           |
| VTT Technical Research Centre of Finland Ltd, Centre for Metrology MIKES (MIKES) | Finland     | Tekniikantie 1, 02150 Espoo  | Richard Högström      | richard.hogstrom@vtt.fi        |
| FORCE Technology (FORCE)   | Denmark     | Navervej 1, 6600 Vejen   | Jesper Busk           | jrb@force.dk                   |
| Federal Institute of Metrology (METAS)   | Switzerland | Lindenweg 50, CH-3003 Bern-Wabern  | Bernhard Niederhauser | bernhard.niederhauser@metas.ch |
| Hellenic Institute of Metrology (EIM)  | Greece      | Industrial Area of Thessaloniki, Block 45, GR 57 022, Sindos, Thessaloniki | Zoe Metaxiotou        | zoe@eim.gr                     |
| VSL (VSL)  | Netherlands | Thijsseweg 11, 2629 JA Delft   | Gerard Blom           | gblom@vsl.nl                   |
| Central Office of Measures (GUM)   | Poland      | ul. Elektoralna 2, 00-139 Warszawa   | Arkadiusz Zadworny    | a.zadworny@gum.gov.pl          |



|  |                |  |                     |                            |
|--|----------------|--|---------------------|----------------------------|
| University of Ljubljana (UL)                         | Slovenia       | Askerceva 6, SI-1000 Ljubljana               | Joze Kutin          | joze.kutin@fs.uni-lj.si    |
| Lithuanian Energy Institute (LEI)                    | Lithuania      | Breslaujos str. 3, 44403 Kaunas              | Arūnas Stankevičius | arunas.stankevicius@lei.lt |
| Czech Metrology Institute (CMI)                      | Czech Republic | Okruzni 31, Brno 638 00                      | Zdenek Krajicek     | zkrajicek@cmi.cz           |
| Laboratoire national de métrologie et d'essais (LNE) | France         | 1, rue Gaston Boissier, 75724 Paris Cedex 15 | Jean Barbe          | jean.barbe@lne.fr          |

Table A2. Detailed information on the measurement standards used by the participants in EURAMET1325

| Name of the laboratory                            | Country | Type of standards  | Identification of the standard   |
|---|---------|--|--|
| Istituto Nazionale di Ricerca Metrologica (INRIM) | Italy   | Piston prover (MICROGas)<br>Bell prover (BELLGas)                                  | MICROGas test rig: used for calibrating the low (ML-800-10) and medium flow cells (ML-800-24).<br>BELLGas test rig: used for calibrating the high flow cells (ML-800-44).  |
| Physikalisch-Technische Bundesanstalt (PTB)       | Germany | 3 mercury sealed piston provers with interferometric distance/velocity measurement | The flow cells were calibrated by critical nozzles installed upstream to the flow cells. The critical nozzles were recalibrated with the mercury sealed primary standard meter immediately before the calibration. By constant pressure and temperature upstream to the critical nozzles the flow rate was stabilised during the calibration of the flow cells and the comparison with the mercury sealed piston prover. |

|  |             |   |  |
|--|-------------|---|--|
| VTT Technical Research Centre of Finland Ltd, Centre for Metrology MIKES (MIKES) | Finland     | Dynamic gravimetric weighing (DWS1 and DWS2)                      | DWS1: used for calibrating flow rates up to 300 ml/min.<br>DWS2: used for calibrating flow rates above 300 ml/min  |
| FORCE Technology (FORCE)   | Denmark     | Piston prover (Cal-bench no C02-006)                              | Cal-bench with flow tubes GT001, GT002, GT003.   |
| Federal Institute of Metrology (METAS)   | Switzerland | Piston prover   | Three flow tubes with following ranges:<br>3 to 180 ml/min No. 3 (small) used for cell -10<br>20 to 2000 ml/min No. 2 (medium) used for cell -24<br>200 to 30000 ml/min No. 1 (large) used for cell -44                                |
| Hellenic Institute of Metrology (EIM)  | Greece      | Piston prover (Brooks VOL-U-METER, 2 units)<br>Sierra Bell Prover | Brooks 1064: used for flow rates 20 - 600 ml/min<br>Brooks 1066: used for flow rates 1250 - 5000 ml/min<br>Bell Prover: used for flow rates 20 000 - 30 000 ml/min   |
| VSL (VSL)  | Netherlands | Piston prover (mercury sealed)                                    | Tube 1 (range 3.7 to 45 ml/min): used for cell 10 and cell -24<br>Tube 2 (range 45 to 385 ml/min): used for cell -24<br>Tube 3 (range 440 to 3500 ml/min): used for cell -44<br>Tube 4 (range 5000 to 13000 ml/min): used for cell -44 |
| University of Ljubljana (UL)   | Slovenia    | Piston prover (3 flow cells)                                      | Cell A: used for calibrating the low flow cell (ML-800-10).<br>Cell B: used for calibrating the medium flow cell (ML-800-24).  |

|  |                |  |  |
|--|----------------|--|--|
|  |                |  | Cell C: used for calibrating the high flow cell (ML-800-44).   |
| Czech Metrology Institute (CMI)                      | Czech Republic | Dynamic gravimetric flow standard GFS DHI Fluke upgraded for operation under vacuum and hermetic modes | GFS used for calibrating all flow rates from 5 ml/min to 30 000 ml/min. Except for the 30 000 ml/min calibration point which was measured using a molbloc-S sonic nozzle traceable to Gravimetric Flow Standard in vacuum. |
| Laboratoire national de métrologie et d'essais (LNE) | France         | Set of 2 Molblocs calibrated with the dynamic gravimetric method                                       | 1E1 : 5 and 10 ml/min<br>1E2 : 20 and 80 ml/min  |

## Appendix 2: Technical protocol

---

### EURAMET PROJECT 1325

#### Comparison for gas flow range 5 ml/min to 30 l/min

#### Technical Protocol

Version 5.0 (17 May 2016)

#### Contents

|   |    |
|---|----|
| 1. INTRODUCTION .....   | 45 |
| 2. PARTICIPANTS .....   | 45 |
| 3. METHOD OF COMPARISON.....  | 46 |
| 4. HANDLING AND TRANSPORT .....   | 46 |
| 4.1 In the case of failure of the transfer standard .....               | 47 |
| 4.2 In the case of an unexpected delay at a participant laboratory..... | 47 |
| 5. DESCRIPTION OF THE TRANSFER STANDARD .....                           | 47 |
| 6. MEASUREMENT PROCESS .....  | 49 |
| 7. REPORTING MEASUREMENT RESULTS.....                                   | 50 |
| 8. ANALYSIS OF THE COMPARISON RESULTS .....                             | 51 |
| 9. TENTATIVE SCHEDULE OF COMPARISON.....                                | 51 |
| APPENDIX 1. DETAILS OF PARTICIPATING NMIS .....                         | 53 |
| APPENDIX 2. PACKING LIST OF TRANSFER STANDARD .....                     | 55 |

## 1. INTRODUCTION

At the EURAMET TC-F meeting at VSL in Delft on 12<sup>th</sup> to 14<sup>th</sup> of March 2013 it was proposed by MIKES to initiate an inter-comparison for low gas flows in the range 5 ml/min to 30 l/min. The comparison was motivated by the fact that the last wider comparison in the low flow range was performed in years 2005 to 2007, almost ten years ago.

Based on discussions during and after the meeting, it was decided to initiate such a comparison. Twelve participants expressed their interest in participating and MIKES volunteered to coordinate the comparison and act as a pilot. The comparison is registered as EURAMET project number 1325.

The aim of the project is to compare measurement capabilities of the participating laboratories in the gas flow range 5 ml/min to 30 l/min. The comparison will be carried out using commercially available piston provers as high precision transfer standards. Due to the high amount of participants the comparison will be carried out in two parallel loops, with one circulating transfer standard in each loop. MIKES will perform initial tests on the transfer standards and based on this experience the measurement protocol (this document) will be refined before the actual comparison measurements.

This document serves as the technical protocol for the comparison and includes e.g. specific instructions for calibrating the transfer standards and timetables for a successful and timely completion of the project.

## 2. PARTICIPANTS

There are twelve participants in the comparison, as showed in table 1. The details of all participants are reported in Appendix 1.

Table 1 - NMIs participating in the comparison.

| Participating NMI |  | Country        |
|-------------------|--|----------------|
| MIKES             | Centre for Metrology and Accreditation         | Finland        |
| CMI               | Czech Metrology Institute                      | Czech Republic |
| UL                | University of Ljubljana                        | Slovenia       |
| PTB               | Physikalisch-Technische Bundesanstalt          | Germany        |
| FORCE             | FORCE Technology                               | Denmark        |
| INRIM             | Istituto Nazionale di Ricerca Metrologica      | Italy          |
| METAS             | Federal Institute of Metrology METAS           | Switzerland    |
| VSL               | VSL  | Netherlands    |
| LNE               | Laboratoire national de métrologie et d'essais | France         |
| JV                | Justervesenet - Norwegian Metrology Service    | Norway         |
| LEI               | Lithuanian Energy Institute                    | Lithuania      |
| EIM               | Hellenic Institute of Metrology                | Greece         |
| GUM               | Central Office of Measures / Główny Urząd Miar | Poland         |

MIKES is the coordinator and pilot of the comparison and supplies the transfer standards.

### 3. METHOD OF COMPARISON

The purpose of the comparison is to compare the volumetric gas flow measurements of the participating NMIs in the range from 5 ml/min to 30 l/min using two transfer standards. The instruments loaned by the manufacturer will be thoroughly tested at the pilot laboratory before actual comparison measurements. Nitrogen (grade 5.0 with purity at least 99.999 %) should be used as the calibration gas. The comparison is carried out in two parallel loops (see Table 4), such that one transfer standard circulates in each loop. Before measurements in any other laboratory, MIKES as the pilot will test both transfer. INRIM and PTB will perform calibrations with both transfer standards and thus establish a link between the loops. After the actual comparison measurements, INRIM and PTB will perform additional measurements on the transfer standards of loops 1 and 2, respectively, in order to assess the long-term stability of the standards including stability of pressure and temperature sensors.

### 4. HANDLING AND TRANSPORT

The transfer standard is supplied with its shipping box, which is sufficiently robust to ensure safe transportation. The transfer standard should be opened by authorised persons upon receipt at the laboratory.

#### **On receipt:**

- Inform the sending laboratory, next laboratory in the chain and the pilot laboratory to confirm receipt. Inform the pilot laboratory and the next laboratory of your expected measurement schedule to confirm planned dispatch time
- check the package casing for damage and on opening the package check the contents against the packing list. Check the transfer standard for damage. Report any damage or missing items to the pilot laboratory and the previous laboratory.

#### **On completion:**

- Pack all items (including manuals) into the shipping box in a similar way as received
- check the contents against the packing list
- check that all paperwork is available to the carrier
- inform the next laboratory and the pilot laboratory on the day of dispatch.

**For shipment:**

- Co-ordinate the shipping with the recipient: Obtain the recipient's exact shipping address. The sending laboratory should provide the recipient with the carrier, the exact travel mode, and the estimated time of arrival
- the recipient should be aware of any customs issues in their country that may cause problems at the customs
- the shipping laboratory must be aware of any national regulations covering the travelling standard to be exported
- mark the shipping box with "FRAGILE SCIENTIFIC INSTRUMENTS", "TO BE OPENED ONLY BY LABORATORY STAFF"
- each laboratory is responsible for the cost of shipping to the next participant including any customs charges and insurance. The insurance should be sufficient to cover the costs of the travelling standards (**about 15 000 Euro per transfer standard**) and any damages that could occur
- **for shipment to METAS and JV only:** The travelling standard will be accompanied by a customs ATA Carnet for temporary import/export and uniquely identifying the instrument.

**Report:**

- Complete within 3 weeks of measurements and send the results to the pilot laboratory.

**Timing:**

- Allow 3 weeks for measurements and 1 week for transport to the next laboratory.

**4.1 In the case of failure of the transfer standard**

If a participant suspects failure of the transfer standard it shall be reported immediately to the pilot laboratory. The pilot laboratory shall decide if repair is required and make arrangements for any repairs. The total costs for repairing (including shipping) will be shared equally among all participants.

**4.2 In the case of an unexpected delay at a participant laboratory**

With their agreement the laboratories will be allotted a time slot (see section 9), within the schedule, to conduct the comparison measurements. If a laboratory is unable to keep to the schedule, then it will be allowed to perform the measurements at the end of the comparison schedule, but it will be responsible for any additional transport costs.

**Inform the pilot laboratory immediately if a problem or delay occurs.**

**5. DESCRIPTION OF THE TRANSFER STANDARD**

The travelling standards are commercially available BIOS Met Lab ML-800 equipped with three measuring cells covering the whole gas flow range of the comparison. The travelling standards serial numbers are shown in table 2. The operating manuals will accompany the transfer standard. A complete packing list of all items included with the transfer standard is given in appendix 2.

Table 2 – Transfer standards used for comparison measurements

| <b>LOOP 1 instrument</b> |                  |               |
|--------------------------|------------------|---------------|
| Model                    | Unit             | Serial number |
| ML-800-B                 | Base             | 147457        |
| ML-800-10                | Measurement cell | 135207        |
| ML-800-24                | Measurement cell | 134909        |
| ML-800-44                | Measurement cell | 135198        |
| <b>LOOP 2 instrument</b> |                  |               |
| Model                    | Unit             | Serial number |
| ML-800-B                 | Base             | 147461        |
| ML-800-10                | Measurement cell | 135208        |
| ML-800-24                | Measurement cell | 134910        |
| ML-800-44                | Measurement cell | 135199        |

The instrument type is based on the positive displacement principle, in which the volumetric flow is determined by measuring the time it takes for the gas flow to “displace” a graphite piston through a glass cylinder with known dimensions. By means of internal temperature and pressure measurements, actual flow readings are converted to standardized flow readings (0 °C and 101325 Pa). There is a small clearance between the piston and the cylinder to allow frictionless motion of the piston. This causes a small internal leak, which is corrected by the transfer standard. The leak rate depends on the viscosity of the calibration gas and therefore comparison measurements should preferably be performed using the same gas, in this case nitrogen (at least grade 5.0 and 99.999% purity).



Figure 2 – Met Lab ML-800 travelling standard



## 6. MEASUREMENT PROCESS

All participants should refer to the operating manuals for instructions of use of the travelling standard. Participant laboratories may perform any initial checks of the operation of the transfer standard that would be performed for a normal calibration. In case of an unexpected instrument failure, the coordinator should be informed as soon as possible.

The gas flow is connected to the transfer standard inlet using Swagelok ¼” ID compression fittings. Flow direction is depicted on top of the measurement cell with an arrow pointing in the flow direction. **The flow cell should not be pressurized since it might result in damaging the instrument.** This can be easily avoided by leaving the outlet open to atmosphere. The instrument is able to measure both actual volumetric flow and standard flow. In this comparison, **calibrations are performed in terms of standard readings at 0 °C and 101325 Pa.** The instruments are pre-set to display standard readings at the aforementioned conditions and therefore no configuration of the settings is necessary.

The calibration points of the comparison and the associated transfer standard flow cells to be used are shown in table 3. Participants are only presumed to measure at points included in their CMCs. However, measurements at other points are strongly encouraged to support future CMCs entries. The flow reading indicated by the transfer standard should be matched as exactly as possible (at least within  $\pm 3\%$ ) to the nominal flows in table 3 to allow good comparability between participants results. The flow ranges of the transfer standard flow cells are partially overlapping, which allows checking the consistency of the measurements results obtained with different flow cells at the same nominal flow. Any drift in a flow cell can therefore be detected and taken into account when evaluating the final comparison results.

Table 3 – Calibration points and corresponding flow cells to be used as transfer standard.

| Nominal flow rate (ml/min) | Transfer standard flow cell       |
|----------------------------|-----------------------------------|
| 5                          | ML-800-10                         |
| 10                         | ML-800-10                         |
| 20                         | ML-800-10                         |
| 80                         | ML-800-10, ML-800-24 <sup>1</sup> |
| 300                        | ML-800-24                         |
| 600                        | ML-800-24                         |
| 1250                       | ML-800-24, ML-800-44 <sup>1</sup> |
| 5000                       | ML-800-44                         |
| 10000                      | ML-800-44                         |
| 20000                      | ML-800-44                         |
| 30000                      | ML-800-44                         |

<sup>1</sup> Measurements are performed with both cells for consistency checking

The transfer standard should be allowed to stabilize to laboratory conditions for at least 24 hours before starting measurements. After this, the transfer standard pressure and temperature sensor readings are checked against participant’s temperature measurement at the transfer standard and laboratory ambient pressure (no flow through transfer standard). This allows considering any drifts in temperature and pressure sensors when evaluating the comparison results.

**Before starting actual comparison measurements, it is necessary to allow the flow cell temperature to settle by passing gas flow through the flow cell and continuously launching the piston for 180 min (referred to in the manual as continuous auto-read mode).** This initial “heating” is necessary for achieving consistent and comparable results, especially in the

lower flow range of the measurement cells. When changing the flow rate, three minutes of continuous measurements is sufficient for reaching a stable temperature. For each nominal calibration point, four separate calibrations are performed. For a single calibration at a nominal point, the flow rate of the transfer standard is obtained as the mean of at least ten (10) consecutive single readings. As supporting measurements, temperature and pressure readings of the transfer standard are recorded for each measurement.

## **7. REPORTING MEASUREMENT RESULTS**

The following data shall be recorded and reported:

1. A description and identification of the laboratory flow standard and traceability including an uncertainty budget of the flow standard. An estimate of the connection tubing (between reference and transfer standard) inner volume as this might influence the results.
2. A short description of the applied calibration method including a schematic drawing indicating the position of the transfer standard.
3. Prior to any calibration measurements for each flow cell: Temperature and pressure indicated by transfer standard and temperature at the transfer standard and laboratory pressure.
4. Required data for each measurement (four per calibration point):
  - a. Mean volume flow of transfer standard at standard conditions.
  - b. Mean volume flow at standard conditions realized by laboratory flow standard and the associated standard uncertainty.
  - c. Standard deviation of transfer standard flow readings.
  - d. Number of transfer standard flow readings contributing to the mean value.
  - e. Temperature and pressure readings indicated by the transfer standard.
  - f. Laboratory temperature, pressure and relative humidity.

An excel spreadsheet will be provided to the participants for recording and reporting results.

Participants shall report their measurement results to the coordinator within three weeks of completion of measurements. If not reported on time as requested, the laboratory results may be excluded from the comparison.

## 8. ANALYSIS OF THE COMPARISON RESULTS

The coordinator will be responsible for analysing the comparison results.

The parameter to be compared is the difference found at each laboratory between the transfer standard and the laboratory flow standard and it is calculated as:

$$E = \frac{\dot{V}_t - \dot{V}_r}{\dot{V}_r} \times 100, \quad (1)$$

where  $\dot{V}_t$  and  $\dot{V}_r$  is the volumetric flow of the transfer standard and laboratory flow standard, respectively.

The uncertainty of the comparison results will be derived from:

- the quoted uncertainty of the flow rate realisation
- the estimated uncertainty from the standard deviation of four flow measurements with the transfer standard
- the estimated uncertainty from the short-term stability of the transfer standard at the time of measurements derived from the standard deviation of transfer standard flow readings
- the estimated uncertainty due to any drift of the travelling standard in the comparison period
- any other components of uncertainty that might be significant, such as stability of temperature and pressure sensors of the transfer standard

The outputs of the comparison are expected to be:

- differences between the realisations of volume flow rate of the participants at each calibration point evaluated with reference to a comparison reference value calculated as the uncertainty weighted mean
- estimates of bilateral equivalence between every pair of participants at each calibration point.

## 9. TENTATIVE SCHEDULE OF COMPARISON

Each participant will be allowed three weeks for performing the comparison measurements, after which the traveling standard is shipped to the next institute (see table 4). One week is reserved for shipping of the transfer standard to the next institute. After all institutes in both loops have performed their measurements, the travelling standards will return to INRIM and PTB for evaluation of possible drifts during the comparison. Based on the data reported by the participants, the comparison results will be evaluated and a comparison report will be prepared by the co-ordinator and circulated to the participants by the end of May 2017.

Table 4 – Tentative schedule for comparison measurements.

| <b>Date</b>                | <b>Loop 1</b> | <b>Loop 2</b> | <b>TASK</b>   |
|----------------------------|---------------|---------------|---|
| April 2016 – May/June 2016 | INRIM         |               | Comparison measurements                             |
| June 2016 – July 2016      | PTB           |               | Comparison measurements                             |
| September 2016             | EIM           | VSL           | Comparison measurements                             |
| October 2016               | FORCE         | GUM           | Comparison measurements                             |
| November 2016              | MIKES         | UL            | Comparison measurements                             |
| December 2016              |               | LEI           | Comparison measurements                             |
| January 2017               | METAS         | CMI           | Comparison measurements                             |
| February 2017              | JV            | LNE           | Comparison measurements                             |
| March 2017                 | INRIM         | PTB           | Evaluation of possible drifts in transfer standards |
| May 2017                   | MIKES         |               | Draft A of comparison report                        |

**APPENDIX 1. DETAILS OF PARTICIPATING NMIs**
**MIKES Metrology, VTT Ltd (MIKES)**

Finland

Address: Tekniikantie 1  
 P.O. Box 9  
 02150 Espoo  
 Finland  
 Contact: Dr. Martti Heinonen,  
 Dr. Richard Högström  
 Phone: +358 400 686 553 (Martti Heinonen),  
 +358 50 303 9341 (Richard Högström)  
 E-mail: [martti.heinonen@vtt.fi](mailto:martti.heinonen@vtt.fi)  
[richard.hogstrom@vtt.fi](mailto:richard.hogstrom@vtt.fi)

**Czech Metrology Institute (CMI)**

Czech Republic

Address: Okružni 31, Brno 638 00, Czech Republic  
 Contact: Zdenek Krajicek  
 Phone: +420 724121129  
 E-mail: [zkrajicek@cmi.cz](mailto:zkrajicek@cmi.cz)

**Physikalisch-Technische Bundesanstalt (PTB)**

Germany

Address: Bundesallee 100, 38116 Braunschweig  
 Contact: Rainer Kramer  
 Phone: +49 531 592 1330  
 E-mail: [rainer.kramer@ptb.de](mailto:rainer.kramer@ptb.de)

**FORCE Technology (FORCE)**

Denmark

Address: Navervej 1, 6600 Vejen  
 Contact: Kurt Rasmussen  
 Phone: +45 76961600 (+45 22697632 Kurt Rasmussen direct)  
 E-mail: [ktr@force.dk](mailto:ktr@force.dk)

**Istituto Nazionale di Ricerca Metrologica (INRIM)**

Italy

Address: Strada delle Cacce, 91  
 I-10135 TORINO  
 Contact: Pier Giorgio Spazzini  
 Phone: +39 11 3919941  
 E-mail: [piergioorgio.spazzini@polito.it](mailto:piergioorgio.spazzini@polito.it)

**University of Ljubljana (UL)**

Slovenia

Address: Askerceva 6,  
 SI-1000 Ljubljana,  
 Slovenia  
 Contact: Joze Kutin  
 Phone: +386 1 4771 307, +386 1 4771 131  
 E-mail: [joze.kutin@fs.uni-lj.si](mailto:joze.kutin@fs.uni-lj.si)

**Federal Institute of Metrology (METAS)**

Switzerland

Address: Lindenweg 50  
CH-3003 Bern-Wabern  
Switzerland  
Contact: Bernhard Niederhauser  
Phone: +41 58 387 02 62  
E-mail: bernhard.niederhauser@metas.ch

**VSL (VSL)**

Netherlands

Address:  
Contact: Gerard Blom  
Phone:  
E-mail: gblom@vsl.nl

**Laboratoire national de métrologie et d'essais (LNE)**

France

Address: 1, rue Gaston Boissier, 75724 Paris Cedex 15  
Contact: Jean Barbe  
Phone:  
E-mail: jean.barbe@lne.fr

**Justervesenet - Norwegian Metrology Service (JV)**

Norway

Address:  
Contact: Tore Mortensen  
Phone:  
E-mail: tmo@justervesenet.no

**Lithuanian Energy Institute (LEI)**

Lithuania

Address: Breslaujos str. 3, 44403 Kaunas  
Contact: Gediminas Zygmantas  
Arūnas Stankevičius  
Phone:  
E-mail: zygmanta@mail.lei.lt

**Hellenic Institute of Metrology (EIM)**

Greece

Address:  
Contact: Zoe Metaxiotou  
Phone:  
E-mail: zoe@eim.gr

**Central Office of Measures / Główny Urząd Miar (GUM)**

Poland

Address:  
Contact: Arkadiusz Zadworny  
Phone:  
E-mail: a.zadworny@gum.gov.pl

**APPENDIX 2. PACKING LIST OF TRANSFER STANDARD**

The contents of the transfer standard packages for loops 1 and 2 are listed in tables A1 and A2, respectively

Table A1 – Loop 1 transfer standard package content

| Item number | Description                                   | Model     | Serial number |
|-------------|---|-----------|---------------|
| 1           | Base  | ML-800-B  | 147457        |
| 2           | Low flow cell                                 | ML-800-10 | 135207        |
| 3           | Medium flow cell                              | ML-800-24 | 134909        |
| 4           | High flow cell                                | ML-800-44 | 135198        |
| 5           | RS-232 Serial cable                           |           |               |
| 6           | USB-cable                                     |           |               |
| 7           | Power supply                                  |           |               |
| 8           | Power supply adapter                          |           |               |
| 9           | ML-800 user manual                            |           |               |
| 10          | Serial communication manuals                  |           |               |
| 11          | Additional instruction for debug and handling |           |               |
| 12          | Certificate of calibration                    |           |               |

Table A2 – Loop 2 transfer standard package content

| Item number | Description                                   | Model     | Serial number |
|-------------|---|-----------|---------------|
| 1           | Base  | ML-800-B  | 147461        |
| 2           | Low flow cell                                 | ML-800-10 | 135208        |
| 3           | Medium flow cell                              | ML-800-24 | 134910        |
| 4           | High flow cell                                | ML-800-44 | 135199        |
| 5           | RS-232 Serial cable                           |           |               |
| 6           | USB-cable                                     |           |               |
| 7           | Power supply                                  |           |               |
| 8           | Power supply adapter                          |           |               |
| 9           | ML-800 user manual                            |           |               |
| 10          | Serial communication manuals                  |           |               |
| 11          | Additional instruction for debug and handling |           |               |
| 12          | Certificate of calibration                    |           |               |

## Appendix 3: Results reported by the participants

### INRIM loop1 results

#### INITIAL TESTS

Compare BIOS temperature and pressure readings with calibrated temperature and ambient pressure instrument readings.

| Laboratory temperature<br>(°C)                 | Laboratory pressure<br>(kPa) | BIOS temperature<br>(°C) | BIOS pressure<br>(kPa) |
|--|------------------------------|--------------------------|------------------------|
| Stabilization 24 hours in laboratory (no flow) |                              |                          |                        |
| 19.84  | 97.987                       | 19.96                    | 97.978                 |

#### COMPARISON MEASUREMENTS

Flow cell: -10 (S/N: 135207)

|                                   |     |                     | ENVIRONMENTAL PARAMETERS         |                             |                                       | TRANSFER STANDARD PARAMETERS  |                             |  |  |                                    | REFERENCE PARAMETERS |                                     |   |
|-----------------------------------|-----|---------------------|----------------------------------|-----------------------------|---------------------------------------|-------------------------------|-----------------------------|--|--|------------------------------------|----------------------|-------------------------------------|---|
| Nominal Flow Rate<br>(std ml/min) | RUN | Date<br>(dd.mm.yyy) | Laboratory temperature<br>(°C)   | Laboratory pressure<br>(Pa) | Laboratory relative humidity<br>(%rh) | BIOS mean temperature<br>(°C) | BIOS mean pressure<br>(kPa) | Mean Indicated Flow Rate<br>(std ml/min) | Standard deviation of indication<br>(std ml/min) | Number of flow indications in mean | RUN                  | Reference Flow Rate<br>(std ml/min) | Standard uncertainty of reference<br>(std ml/min) |
| <b>Set Flow Rate</b>              |     |                     | <b>Stabilization 180 minutes</b> |                             |                                       |                               |                             |  |  |                                    |                      |                                     |   |
| 5                                 | 1   | 19.05.2016          | 19.903                           | 98034                       | 54                                    | 20.462                        | 98.1312                     | 5.01                                     | 0.00   | 10                                 | 1                    | 5.0296                              | 0.0011  |
| 5                                 | 2   | 19.05.2016          | 19.907                           | 98020                       | 54                                    | 20.558                        | 98.1188                     | 5.01                                     | 0.00   | 10                                 | 2                    | 5.0207                              | 0.0011  |
| 5                                 | 3   | 19.05.2016          | 19.902                           | 98016                       | 54                                    | 20.554                        | 98.1191                     | 5.01                                     | 0.00   | 10                                 | 3                    | 5.0241                              | 0.0012  |
| 5                                 | 4   | 19.05.2016          | 19.917                           | 98003                       | 54                                    | 20.596                        | 98.1073                     | 5.00                                     | 0.00   | 10                                 | 4                    | 5.0182                              | 0.0014  |
|                                   |     |                     |                                  |                             |                                       |                               |                             | Average=                                 | 5.01   |                                    |                      | Average=                            | 5.02  |
| <b>Change Flow Rate</b>           |     |                     | <b>Stabilization 3 minutes</b>   |                             |                                       |                               |                             |  |  |                                    |                      |                                     |   |
| 10                                | 1   | 19.05.2016          | 19.908                           | 98028                       | 54                                    | 20.49                         | 98.1272                     | 10.01                                    | 0.01   | 10                                 | 1                    | 10.023                              | 0.002   |
| 10                                | 2   | 19.05.2016          | 19.906                           | 98023                       | 54                                    | 20.538                        | 98.1227                     | 9.99                                     | 0.00   | 10                                 | 2                    | 10.018                              | 0.002   |
| 10                                | 3   | 19.05.2016          | 19.915                           | 98018                       | 54                                    | 20.55                         | 98.1184                     | 9.99                                     | 0.00   | 10                                 | 3                    | 10.010                              | 0.002   |
| 10                                | 4   | 19.05.2016          | 19.912                           | 97998                       | 54                                    | 20.596                        | 98.0985                     | 9.99                                     | 0.00   | 10                                 | 4                    | 10.007                              | 0.002   |
|                                   |     |                     |                                  |                             |                                       |                               |                             | Average=                                 | 10.00  |                                    |                      | Average=                            | 10.01   |
| <b>Change Flow Rate</b>           |     |                     | <b>Stabilization 3 minutes</b>   |                             |                                       |                               |                             |  |  |                                    |                      |                                     |   |
| 20                                | 1   | 19.05.2016          | 19.912                           | 98021                       | 54                                    | 20.504                        | 98.1246                     | 19.99                                    | 0.00   | 10                                 | 1                    | 20.017                              | 0.004   |
| 20                                | 2   | 19.05.2016          | 19.910                           | 98025                       | 54                                    | 20.522                        | 98.1263                     | 19.96                                    | 0.00   | 10                                 | 2                    | 20.001                              | 0.004   |
| 20                                | 3   | 19.05.2016          | 19.910                           | 98018                       | 54                                    | 20.548                        | 98.1153                     | 19.97                                    | 0.00   | 10                                 | 3                    | 19.998                              | 0.005   |
| 20                                | 4   | 19.05.2016          | 19.914                           | 98012                       | 54                                    | 20.588                        | 98.1122                     | 19.95                                    | 0.01   | 10                                 | 4                    | 19.980                              | 0.004   |
|                                   |     |                     |                                  |                             |                                       |                               |                             | Average=                                 | 19.97  |                                    |                      | Average=                            | 20.00   |
| <b>Change Flow Rate</b>           |     |                     | <b>Stabilization 3 minutes</b>   |                             |                                       |                               |                             |  |  |                                    |                      |                                     |   |
| 80                                | 1   | 19.05.2016          | 19.905                           | 98023                       | 54                                    | 20.514                        | 98.1325                     | 79.84                                    | 0.02   | 10                                 | 1                    | 80.011                              | 0.017   |
| 80                                | 2   | 19.05.2016          | 19.911                           | 98020                       | 54                                    | 20.516                        | 98.1261                     | 79.84                                    | 0.03   | 10                                 | 2                    | 79.945                              | 0.017   |
| 80                                | 3   | 19.05.2016          | 19.911                           | 98016                       | 54                                    | 20.56                         | 98.1271                     | 79.80                                    | 0.03   | 10                                 | 3                    | 79.908                              | 0.017   |
| 80                                | 4   | 19.05.2016          | 19.913                           | 98012                       | 54                                    | 20.578                        | 98.1183                     | 79.79                                    | 0.01   | 10                                 | 4                    | 79.893                              | 0.017   |
|                                   |     |                     |                                  |                             |                                       |                               |                             | Average=                                 | 79.82  |                                    |                      | Average=                            | 79.94   |



**INITIAL TESTS**

Compare BIOS temperature and pressure readings with calibrated temperature and ambient pressure instrument readings.

| Laboratory temperature<br>(°C)                 | Laboratory pressure<br>(kPa) | BIOS temperature<br>(°C) | BIOS pressure<br>(kPa) |
|--|------------------------------|--------------------------|------------------------|
| Stabilization 24 hours in laboratory (no flow) |                              |                          |                        |
| 19.54  | 97.049                       | 19.98                    | 97.045                 |

**COMPARISON MEASUREMENTS**

**Flow cell: -24 (S/N: 134909)**

| ENVIRONMENTAL PARAMETERS          |     |                     | TRANSFER STANDARD PARAMETERS     |                             |                                       |                               |                             | REFERENCE PARAMETERS                     |  |                                    |          |                                     |   |
|-----------------------------------|-----|---------------------|----------------------------------|-----------------------------|---------------------------------------|-------------------------------|-----------------------------|--|--|------------------------------------|----------|-------------------------------------|---|
| Nominal Flow Rate<br>(std ml/min) | RUN | Date<br>(dd.mm.yyy) | Laboratory temperature<br>(°C)   | Laboratory pressure<br>(Pa) | Laboratory relative humidity<br>(%rh) | BIOS mean temperature<br>(°C) | BIOS mean pressure<br>(kPa) | Mean Indicated Flow Rate<br>(std ml/min) | Standard deviation of indication<br>(std ml/min) | Number of flow indications in mean | RUN      | Reference Flow Rate<br>(std ml/min) | Standard uncertainty of reference<br>(std ml/min) |
| <b>Set Flow Rate</b>              |     |                     | <b>Stabilization 180 minutes</b> |                             |                                       |                               |                             |  |  |                                    |          |                                     |   |
| 80                                | 1   | 12.05.2016          | 19.919                           | 97167                       | 59.5                                  | 20.658                        | 97.2741                     | 79.54                                    | 0.02   | 10                                 | 1        | 79.893                              | 0.013   |
| 80                                | 2   | 12.05.2016          | 19.921                           | 97148                       | 59.5                                  | 20.78                         | 97.2549                     | 79.40                                    | 0.02   | 10                                 | 2        | 79.755                              | 0.013   |
| 80                                | 3   | 12.05.2016          | 19.918                           | 97135                       | 59.5                                  | 20.776                        | 97.2439                     | 79.41                                    | 0.02   | 10                                 | 3        | 79.745                              | 0.014   |
| 80                                | 4   | 12.05.2016          | 19.926                           | 97099                       | 59.5                                  | 20.878                        | 97.206                      | 79.30                                    | 0.01   | 10                                 | 4        | 79.664                              | 0.013   |
|                                   |     |                     |                                  |                             |                                       |                               |                             | Average=                                 | <b>79.41</b>                                     |                                    | Average= | <b>79.76</b>                        |   |
| <b>Change Flow Rate</b>           |     |                     | <b>Stabilization 3 minutes</b>   |                             |                                       |                               |                             |  |  |                                    |          |                                     |   |
| 300                               | 1   | 12.05.2016          | 19.920                           | 97166                       | 59.5                                  | 20.684                        | 97.2968                     | 298.17                                   | 0.04   | 10                                 | 1        | 299.45                              | 0.05  |
| 300                               | 2   | 12.05.2016          | 19.927                           | 97155                       | 59.5                                  | 20.726                        | 97.2842                     | 298.63                                   | 0.03   | 10                                 | 2        | 299.97                              | 0.05  |
| 300                               | 3   | 12.05.2016          | 19.920                           | 97122                       | 59.5                                  | 20.782                        | 97.2519                     | 298.79                                   | 0.02   | 10                                 | 3        | 300.12                              | 0.05  |
| 300                               | 4   | 12.05.2016          | 19.927                           | 97107                       | 59.5                                  | 20.86                         | 97.2355                     | 298.58                                   | 0.03   | 10                                 | 4        | 299.80                              | 0.05  |
|                                   |     |                     |                                  |                             |                                       |                               |                             | Average=                                 | <b>298.54</b>                                    |                                    | Average= | <b>299.84</b>                       |   |
| <b>Change Flow Rate</b>           |     |                     | <b>Stabilization 3 minutes</b>   |                             |                                       |                               |                             |  |  |                                    |          |                                     |   |
| 600                               | 1   | 12.05.2016          | 19.917                           | 97166                       | 59.5                                  | 20.68                         | 97.3246                     | 596.42                                   | 0.04   | 10                                 | 1        | 598.92                              | 0.10  |
| 600                               | 2   | 12.05.2016          | 19.918                           | 97160                       | 59.5                                  | 20.706                        | 97.3179                     | 595.99                                   | 0.05   | 10                                 | 2        | 598.64                              | 0.10  |
| 600                               | 3   | 12.05.2016          | 19.918                           | 97121                       | 59.5                                  | 20.782                        | 97.2787                     | 596.39                                   | 0.05   | 10                                 | 3        | 598.90                              | 0.10  |
| 600                               | 4   | 12.05.2016          | 19.923                           | 97106                       | 59.5                                  | 20.832                        | 97.2656                     | 596.22                                   | 0.04   | 10                                 | 4        | 598.81                              | 0.10  |
|                                   |     |                     |                                  |                             |                                       |                               |                             | Average=                                 | <b>596.25</b>                                    |                                    | Average= | <b>598.82</b>                       |   |
| <b>Change Flow Rate</b>           |     |                     | <b>Stabilization 3 minutes</b>   |                             |                                       |                               |                             |  |  |                                    |          |                                     |   |
| 1250                              | 1   | 12.05.2016          | 19.918                           | 97163                       | 59.5                                  | 20.668                        | 97.4172                     | 1245.09                                  | 0.14   | 10                                 | 1        | 1250.73                             | 0.20  |
| 1250                              | 2   | 12.05.2016          | 19.923                           | 97162                       | 59.5                                  | 20.682                        | 97.4153                     | 1244.60                                  | 0.13   | 10                                 | 2        | 1250.38                             | 0.20  |
| 1250                              | 3   | 12.05.2016          | 19.917                           | 97118                       | 59.5                                  | 20.778                        | 97.3742                     | 1245.11                                  | 0.13   | 10                                 | 3        | 1250.65                             | 0.20  |
| 1250                              | 4   | 12.05.2016          | 19.922                           | 97113                       | 59.5                                  | 20.802                        | 97.3702                     | 1245.18                                  | 0.04   | 10                                 | 4        | 1250.64                             | 0.20  |
|                                   |     |                     |                                  |                             |                                       |                               |                             | Average=                                 | <b>1245.00</b>                                   |                                    | Average= | <b>1250.60</b>                      |   |

**INITIAL TESTS**

Compare BIOS temperature and pressure readings with calibrated temperature and ambient pressure instrument readings.

| Laboratory temperature (°C)                    | Laboratory pressure (kPa) | BIOS temperature (°C) | BIOS pressure (kPa) |
|--|---------------------------|-----------------------|---------------------|
| Stabilization 24 hours in laboratory (no flow) |                           |                       |                     |
| 19.9   | 98.422                    | 19.96                 | 98.411              |

**COMPARISON MEASUREMENTS**

Flow cell: -44 (S/N: 135198)

|                                |     |                  | ENVIRONMENTAL PARAMETERS         |                          |                                    | TRANSFER STANDARD PARAMETERS |                          |                                       |   |                                    | REFERENCE PARAMETERS     |                                  |  |
|--------------------------------|-----|------------------|----------------------------------|--------------------------|------------------------------------|------------------------------|--------------------------|---------------------------------------|---|------------------------------------|--------------------------|----------------------------------|--|
| Nominal Flow Rate (std ml/min) | RUN | Date (dd.mm.yyy) | Laboratory temperature (°C)      | Laboratory pressure (Pa) | Laboratory relative humidity (%rh) | BIOS mean temperature (°C)   | BIOS mean pressure (kPa) | Mean Indicated Flow Rate (std ml/min) | Standard deviation of indication (std ml/min) | Number of flow indications in mean | RUN                      | Reference Flow Rate (std ml/min) | Standard uncertainty of reference (std ml/min) |
| <b>Set Flow Rate</b>           |     |                  | <b>Stabilization 180 minutes</b> |                          |                                    |                              |                          |                                       |   |                                    |                          |                                  |  |
| 1250                           | 1   | 25.05.2016       | 19.685                           | 98563                    | 50.5                               | 20.477                       | 98.7117                  | 1247.4                                | 0.5   | 80.000                             | 1                        | 1246.8                           | 1.1  |
| 1250                           | 2   | 25.05.2016       | 19.720                           | 98562                    | 50.5                               | 20.784                       | 98.7119                  | 1237.5                                | 2.2   | 70.000                             | 2                        | 1237.8                           | 1.2  |
| 1250                           | 3   | 25.05.2016       | 19.745                           | 98567                    | 50.5                               | 20.738                       | 98.7189                  | 1242.8                                | 1.1   | 70.000                             | 3                        | 1243.6                           | 1.2  |
| 1250                           | 4   | 25.05.2016       | 19.750                           | 98555                    | 50.5                               | 20.843                       | 98.7103                  | 1236.2                                | 2.6   | 70.000                             | 4                        | 1235.3                           | 1.2  |
| Average=                       |     |                  |                                  |                          |                                    |                              |                          | <b>1240.97</b>                        |   |                                    | Average= <b>1240.88</b>  |                                  |  |
| <b>Change Flow Rate</b>        |     |                  | <b>Stabilization 3 minutes</b>   |                          |                                    |                              |                          |                                       |   |                                    |                          |                                  |  |
| 5000                           | 1   | 25.05.2016       | 19.685                           | 98563                    | 50.5                               | 20.816                       | 98.7363                  | 5005.3                                | 3.3   | 60                                 | 1                        | 5009.4                           | 4.6  |
| 5000                           | 2   | 25.05.2016       | 19.700                           | 98563                    | 50.5                               | 20.944                       | 98.7411                  | 4972.3                                | 1.6   | 50                                 | 2                        | 4974.1                           | 4.6  |
| 5000                           | 3   | 25.05.2016       | 19.715                           | 98561                    | 50.5                               | 20.824                       | 98.7398                  | 5008.2                                | 2.5   | 60                                 | 3                        | 5008.1                           | 4.5  |
| 5000                           | 4   | 25.05.2016       | 19.740                           | 98555                    | 50.5                               | 21.040                       | 98.7370                  | 4969.9                                | 3.0   | 60                                 | 4                        | 4970.7                           | 4.6  |
| Average=                       |     |                  |                                  |                          |                                    |                              |                          | <b>4988.91</b>                        |   |                                    | Average= <b>4990.57</b>  |                                  |  |
| <b>Change Flow Rate</b>        |     |                  | <b>Stabilization 3 minutes</b>   |                          |                                    |                              |                          |                                       |   |                                    |                          |                                  |  |
| 10000                          | 1   | 25.05.2016       | 19.695                           | 98557                    | 50.5                               | 21.156                       | 98.7945                  | 10009.7                               | 3.4   | 50                                 | 1                        | 10022.3                          | 8.8  |
| 10000                          | 2   | 25.05.2016       | 19.695                           | 98562                    | 50.5                               | 21.050                       | 98.7958                  | 9950.2                                | 5.5   | 40                                 | 2                        | 9956.1                           | 9.2  |
| 10000                          | 3   | 25.05.2016       | 19.695                           | 98557                    | 50.5                               | 21.179                       | 98.8058                  | 10001.1                               | 5.0   | 40                                 | 3                        | 10009.0                          | 9.4  |
| 10000                          | 4   | 25.05.2016       | 19.715                           | 98550                    | 50.5                               | 21.101                       | 98.7762                  | 9951.0                                | 4.9   | 40                                 | 4                        | 9951.7                           | 9.3  |
| Average=                       |     |                  |                                  |                          |                                    |                              |                          | <b>9978.01</b>                        |   |                                    | Average= <b>9984.79</b>  |                                  |  |
| <b>Change Flow Rate</b>        |     |                  | <b>Stabilization 3 minutes</b>   |                          |                                    |                              |                          |                                       |   |                                    |                          |                                  |  |
| 20000                          | 1   | 25.05.2016       | 19.680                           | 98561                    | 50.5                               | 21.255                       | 98.9757                  | 19999.9                               | 9.7   | 30                                 | 1                        | 20013.0                          | 17.9   |
| 20000                          | 2   | 25.05.2016       | 19.695                           | 98558                    | 50.5                               | 21.105                       | 98.9910                  | 19926.4                               | 9.5   | 30                                 | 2                        | 19942.7                          | 17.6   |
| 20000                          | 3   | 25.05.2016       | 19.695                           | 98557                    | 50.5                               | 21.215                       | 99.0073                  | 20006.8                               | 11.1  | 30                                 | 3                        | 20014.7                          | 17.7   |
| 20000                          | 4   | 25.05.2016       | 19.710                           | 98551                    | 50.5                               | 21.116                       | 99.0150                  | 19917.4                               | 11.2  | 30                                 | 4                        | 19916.9                          | 17.4   |
| Average=                       |     |                  |                                  |                          |                                    |                              |                          | <b>19962.62</b>                       |   |                                    | Average= <b>19971.85</b> |                                  |  |
| <b>Change Flow Rate</b>        |     |                  | <b>Stabilization 3 minutes</b>   |                          |                                    |                              |                          |                                       |   |                                    |                          |                                  |  |
| 30000                          | 1   | 25.05.2016       | 19.685                           | 98559                    | 50.5                               | 21.162                       | 99.1954                  | 29989.9                               | 17.5  | 30                                 | 1                        | 30016.2                          | 22.0   |
| 30000                          | 2   | 25.05.2016       | 19.695                           | 98556                    | 50.5                               | 21.121                       | 99.2060                  | 29947.1                               | 14.6  | 20                                 | 2                        | 29963.6                          | 20.9   |
| 30000                          | 3   | 25.05.2016       | 19.700                           | 98558                    | 50.5                               | 21.161                       | 99.2282                  | 30008.6                               | 19.5  | 30                                 | 3                        | 29998.2                          | 21.3   |
| 30000                          | 4   | 25.05.2016       | 19.705                           | 98551                    | 50.5                               | 21.103                       | 99.2151                  | 29945.9                               | 18.6  | 30                                 | 4                        | 29960.5                          | 21.6   |
| Average=                       |     |                  |                                  |                          |                                    |                              |                          | <b>29972.86</b>                       |   |                                    | Average= <b>29984.63</b> |                                  |  |
| <b>End of measurements</b>     |     |                  |                                  |                          |                                    |                              |                          |                                       |   |                                    |                          |                                  |  |

**INRIM loop2 results**

**INITIAL TESTS**

Compare BIOS temperature and pressure readings with calibrated temperature and ambient pressure instrument readings.

| Laboratory temperature                         | Laboratory pressure | BIOS temperature | BIOS pressure |
|--|---------------------|------------------|---------------|
| (°C)   | (kPa)               | (°C)             | (kPa)         |
| Stabilization 24 hours in laboratory (no flow) |                     |                  |               |
| 19.84  | 98.768              | 20.14            | 98.79         |

**COMPARISON MEASUREMENTS**

Flow cell: -10 (S/N: 135208)

|                         |     |             | ENVIRONMENTAL PARAMETERS         |                     |                              | TRANSFER STANDARD PARAMETERS |                    |                          |                                  |                                    | REFERENCE PARAMETERS |                     |                                   |  |              |  |
|-------------------------|-----|-------------|----------------------------------|---------------------|------------------------------|------------------------------|--------------------|--------------------------|----------------------------------|------------------------------------|----------------------|---------------------|-----------------------------------|--|--------------|--|
| Nominal Flow Rate       | RUN | Date        | Laboratory temperature           | Laboratory pressure | Laboratory relative humidity | BIOS mean temperature        | BIOS mean pressure | Mean Indicated Flow Rate | Standard deviation of indication | Number of flow indications in mean | RUN                  | Reference Flow Rate | Standard uncertainty of reference |  |              |  |
| (std ml/min)            |     | (dd.mm.yyy) | (°C)                             | (Pa)                | (%rh)                        | (°C)                         | (kPa)              | (std ml/min)             | (std ml/min)                     |                                    |                      | (std ml/min)        | (std ml/min)                      |  |              |  |
| <b>Set Flow Rate</b>    |     |             | <b>Stabilization 180 minutes</b> |                     |                              |                              |                    |                          |                                  |                                    |                      |                     |                                   |  |              |  |
| 5                       | 1   | 18.05.2016  | 19.897                           | 98743               | 50.5                         | 20.814                       | 98.8859            | 5.04                     | 0.00                             | 10                                 | 1                    | 5.0489              | 0.0011                            |  |              |  |
| 5                       | 2   | 18.05.2016  | 19.908                           | 98704               | 50.5                         | 20.836                       | 98.8446            | 5.02                     | 0.00                             | 10                                 | 2                    | 5.0279              | 0.0012                            |  |              |  |
| 5                       | 3   | 18.05.2016  | 19.906                           | 98683               | 50.5                         | 20.79                        | 98.8239            | 5.01                     | 0.00                             | 10                                 | 3                    | 5.0221              | 0.0014                            |  |              |  |
| 5                       | 4   | 18.05.2016  | 19.909                           | 98613               | 50.5                         | 20.738                       | 98.7553            | 5.01                     | 0.00                             | 10                                 | 4                    | 5.0199              | 0.0012                            |  |              |  |
| Average=                |     |             |                                  |                     |                              |                              |                    | <b>5.02</b>              |                                  | Average=                           |                      |                     |                                   |  | <b>5.03</b>  |  |
| <b>Change Flow Rate</b> |     |             | <b>Stabilization 3 minutes</b>   |                     |                              |                              |                    |                          |                                  |                                    |                      |                     |                                   |  |              |  |
| 10                      | 1   | 18.05.2016  | 19.902                           | 98742               | 50.5                         | 20.84                        | 98.884             | 10.03                    | 0.01                             | 10                                 | 1                    | 10.040              | 0.002                             |  |              |  |
| 10                      | 2   | 18.05.2016  | 19.901                           | 98715               | 50.5                         | 20.884                       | 98.8581            | 10.01                    | 0.00                             | 10                                 | 2                    | 10.028              | 0.002                             |  |              |  |
| 10                      | 3   | 18.05.2016  | 19.908                           | 98661               | 50.5                         | 20.76                        | 98.8052            | 10.00                    | 0.01                             | 10                                 | 3                    | 10.012              | 0.002                             |  |              |  |
| 10                      | 4   | 18.05.2016  | 19.911                           | 98630               | 50.5                         | 20.754                       | 98.7747            | 9.99                     | 0.01                             | 10                                 | 4                    | 10.008              | 0.002                             |  |              |  |
| Average=                |     |             |                                  |                     |                              |                              |                    | <b>10.01</b>             |                                  | Average=                           |                      |                     |                                   |  | <b>10.02</b> |  |
| <b>Change Flow Rate</b> |     |             | <b>Stabilization 3 minutes</b>   |                     |                              |                              |                    |                          |                                  |                                    |                      |                     |                                   |  |              |  |
| 20                      | 1   | 18.05.2016  | 19.902                           | 98733               | 50.5                         | 20.844                       | 98.8752            | 20.02                    | 0.01                             | 10                                 | 1                    | 20.038              | 0.005                             |  |              |  |
| 20                      | 2   | 18.05.2016  | 19.908                           | 98720               | 50.5                         | 20.89                        | 98.8626            | 19.98                    | 0.01                             | 10                                 | 2                    | 19.995              | 0.004                             |  |              |  |
| 20                      | 3   | 18.05.2016  | 19.915                           | 98654               | 50.5                         | 20.76                        | 98.8017            | 19.99                    | 0.00                             | 10                                 | 3                    | 19.999              | 0.004                             |  |              |  |
| 20                      | 4   | 18.05.2016  | 19.906                           | 98640               | 50.5                         | 20.76                        | 98.7883            | 19.96                    | 0.01                             | 10                                 | 4                    | 19.984              | 0.005                             |  |              |  |
| Average=                |     |             |                                  |                     |                              |                              |                    | <b>19.99</b>             |                                  | Average=                           |                      |                     |                                   |  | <b>20.00</b> |  |
| <b>Change Flow Rate</b> |     |             | <b>Stabilization 3 minutes</b>   |                     |                              |                              |                    |                          |                                  |                                    |                      |                     |                                   |  |              |  |
| 80                      | 1   | 18.05.2016  | 19.901                           | 98724               | 50.5                         | 20.86                        | 98.8755            | 79.90                    | 0.01                             | 10                                 | 1                    | 79.961              | 0.016                             |  |              |  |
| 80                      | 2   | 18.05.2016  | 19.901                           | 98722               | 50.5                         | 20.874                       | 98.8725            | 79.88                    | 0.01                             | 10                                 | 2                    | 79.941              | 0.017                             |  |              |  |
| 80                      | 3   | 18.05.2016  | 19.909                           | 98652               | 50.5                         | 20.764                       | 98.8076            | 79.92                    | 0.02                             | 10                                 | 3                    | 79.899              | 0.017                             |  |              |  |
| 80                      | 4   | 18.05.2016  | 19.909                           | 98646               | 50.5                         | 20.76                        | 98.8023            | 79.87                    | 0.01                             | 10                                 | 4                    | 79.870              | 0.017                             |  |              |  |
| Average=                |     |             |                                  |                     |                              |                              |                    | <b>79.90</b>             |                                  | Average=                           |                      |                     |                                   |  | <b>79.92</b> |  |

**INITIAL TESTS**

Compare BIOS temperature and pressure readings with calibrated temperature and ambient pressure instrument readings.

| Laboratory temperature<br>(°C)                 | Laboratory pressure<br>(kPa) | BIOS temperature<br>(°C) | BIOS pressure<br>(kPa) |
|--|------------------------------|--------------------------|------------------------|
| Stabilization 24 hours in laboratory (no flow) |                              |                          |                        |
| 19.67  | 98.795                       | 20.42                    | 98.818                 |

**COMPARISON MEASUREMENTS**

**Flow cell: -24 (S/N: 134910)**

|                                   |     |                     | ENVIRONMENTAL PARAMETERS         |                             |                                       | TRANSFER STANDARD PARAMETERS  |                             |  |  |                                    | REFERENCE PARAMETERS |                                     |   |
|-----------------------------------|-----|---------------------|----------------------------------|-----------------------------|---------------------------------------|-------------------------------|-----------------------------|--|--|------------------------------------|----------------------|-------------------------------------|---|
| Nominal Flow Rate<br>(std ml/min) | RUN | Date<br>(dd.mm.yyy) | Laboratory temperature<br>(°C)   | Laboratory pressure<br>(Pa) | Laboratory relative humidity<br>(%rh) | BIOS mean temperature<br>(°C) | BIOS mean pressure<br>(kPa) | Mean Indicated Flow Rate<br>(std ml/min) | Standard deviation of indication<br>(std ml/min) | Number of flow indications in mean | RUN                  | Reference Flow Rate<br>(std ml/min) | Standard uncertainty of reference<br>(std ml/min) |
| <b>Set Flow Rate</b>              |     |                     | <b>Stabilization 180 minutes</b> |                             |                                       |                               |                             |  |  |                                    |                      |                                     |   |
| 80                                | 1   | 17.05.2016          | 19.904                           | 98942                       | 45.8                                  | 20.468                        | 99.0871                     | 80.433                                   | 0.02   | 10                                 | 1                    | 80.600                              | 0.013   |
| 80                                | 2   | 17.05.2016          | 19.899                           | 98931                       | 45.8                                  | 20.616                        | 99.0752                     | 80.451                                   | 0.03   | 10                                 | 2                    | 80.603                              | 0.013   |
| 80                                | 3   | 17.05.2016          | 19.899                           | 98928                       | 45.8                                  | 20.634                        | 99.0723                     | 80.481                                   | 0.02   | 10                                 | 3                    | 80.595                              | 0.013   |
| 80                                | 4   | 17.05.2016          | 19.904                           | 98894                       | 45.8                                  | 20.748                        | 99.0397                     | 80.027                                   | 0.02   | 10                                 | 4                    | 80.184                              | 0.013   |
|                                   |     |                     |                                  |                             |                                       |                               |                             | Average=                                 | <b>80.35</b>                                     |                                    |                      | Average=                            | <b>80.50</b>                                      |
| <b>Change Flow Rate</b>           |     |                     | <b>Stabilization 3 minutes</b>   |                             |                                       |                               |                             |  |  |                                    |                      |                                     |   |
| 300                               | 1   | 17.05.2016          | 19.896                           | 98939                       | 45.8                                  | 20.48                         | 99.1029                     | 299.33                                   | 0.04   | 10                                 | 1                    | 299.75                              | 0.05  |
| 300                               | 2   | 17.05.2016          | 19.897                           | 98935                       | 45.8                                  | 20.58                         | 99.0998                     | 301.36                                   | 0.03   | 10                                 | 2                    | 301.62                              | 0.05  |
| 300                               | 3   | 17.05.2016          | 19.901                           | 98919                       | 45.8                                  | 20.66                         | 99.0846                     | 299.11                                   | 0.02   | 10                                 | 3                    | 299.37                              | 0.05  |
| 300                               | 4   | 17.05.2016          | 19.900                           | 98899                       | 45.8                                  | 20.72                         | 99.0647                     | 298.85                                   | 0.03   | 10                                 | 4                    | 299.16                              | 0.05  |
|                                   |     |                     |                                  |                             |                                       |                               |                             | Average=                                 | <b>299.66</b>                                    |                                    |                      | Average=                            | <b>299.97</b>                                     |
| <b>Change Flow Rate</b>           |     |                     | <b>Stabilization 3 minutes</b>   |                             |                                       |                               |                             |  |  |                                    |                      |                                     |   |
| 600                               | 1   | 17.05.2016          | 19.898                           | 98942                       | 45.8                                  | 20.488                        | 99.1346                     | 600.05                                   | 0.02   | 10                                 | 1                    | 600.70                              | 0.10  |
| 600                               | 2   | 17.05.2016          | 19.899                           | 98934                       | 45.8                                  | 20.552                        | 99.1297                     | 599.83                                   | 0.03   | 10                                 | 2                    | 600.31                              | 0.10  |
| 600                               | 3   | 17.05.2016          | 19.901                           | 98916                       | 45.8                                  | 20.66                         | 99.1109                     | 599.72                                   | 0.03   | 10                                 | 3                    | 600.06                              | 0.10  |
| 600                               | 4   | 17.05.2016          | 19.900                           | 98903                       | 45.8                                  | 20.674                        | 99.0985                     | 599.55                                   | 0.04   | 10                                 | 4                    | 599.89                              | 0.10  |
|                                   |     |                     |                                  |                             |                                       |                               |                             | Average=                                 | <b>599.79</b>                                    |                                    |                      | Average=                            | <b>600.24</b>                                     |
| <b>Change Flow Rate</b>           |     |                     | <b>Stabilization 3 minutes</b>   |                             |                                       |                               |                             |  |  |                                    |                      |                                     |   |
| 1250                              | 1   | 17.05.2016          | 19.894                           | 98938                       | 45.8                                  | 20.496                        | 99.2262                     | 1250.48                                  | 0.12   | 10                                 | 1                    | 1252.10                             | 0.20  |
| 1250                              | 2   | 17.05.2016          | 19.898                           | 98934                       | 45.8                                  | 20.51                         | 99.223                      | 1250.22                                  | 0.09   | 10                                 | 2                    | 1251.73                             | 0.20  |
| 1250                              | 3   | 17.05.2016          | 19.902                           | 98913                       | 45.8                                  | 20.664                        | 99.2058                     | 1249.44                                  | 0.09   | 10                                 | 3                    | 1251.28                             | 0.20  |
| 1250                              | 4   | 17.05.2016          | 19.900                           | 98905                       | 45.8                                  | 20.668                        | 99.1974                     | 1249.51                                  | 0.07   | 10                                 | 4                    | 1251.15                             | 0.20  |
|                                   |     |                     |                                  |                             |                                       |                               |                             | Average=                                 | <b>1249.91</b>                                   |                                    |                      | Average=                            | <b>1251.56</b>                                    |

**INITIAL TESTS**

Compare BIOS temperature and pressure readings with calibrated temperature and ambient pressure instrument readings.

| Laboratory temperature<br>(°C)                 | Laboratory pressure<br>(kPa) | BIOS temperature<br>(°C) | BIOS pressure<br>(kPa) |
|--|------------------------------|--------------------------|------------------------|
| Stabilization 24 hours in laboratory (no flow) |                              |                          |                        |
| 19.62  | 98.751                       | 19.6                     | 98.775                 |

**COMPARISON MEASUREMENTS**

Flow cell: -44 (S/N: 135199)

| Nominal Flow Rate<br>(std ml/min) | RUN | Date<br>(dd.mm.yyy) | ENVIRONMENTAL PARAMETERS         |                             |                                       | TRANSFER STANDARD PARAMETERS  |                             |  |  |                                    | REFERENCE PARAMETERS |                                     |   |
|-----------------------------------|-----|---------------------|----------------------------------|-----------------------------|---------------------------------------|-------------------------------|-----------------------------|--|--|------------------------------------|----------------------|-------------------------------------|---|
|                                   |     |                     | Laboratory temperature<br>(°C)   | Laboratory pressure<br>(Pa) | Laboratory relative humidity<br>(%rh) | BIOS mean temperature<br>(°C) | BIOS mean pressure<br>(kPa) | Mean Indicated Flow Rate<br>(std ml/min) | Standard deviation of indication<br>(std ml/min) | Number of flow indications in mean | RUN                  | Reference Flow Rate<br>(std ml/min) | Standard uncertainty of reference<br>(std ml/min) |
| <b>Set Flow Rate</b>              |     |                     | <b>Stabilization 180 minutes</b> |                             |                                       |                               |                             |  |  |                                    |                      |                                     |   |
| 1250                              | 1   | 26.05.2016          | 19.710                           | 98805                       | 62.3                                  | 20.475                        | 98.9865                     | 1245.0                                   | 0.6  | 70.000                             | 1                    | 1244.5                              | 1.2   |
| 1250                              | 2   | 26.05.2016          | 19.740                           | 98787                       | 62.3                                  | 20.799                        | 98.9707                     | 1234.8                                   | 2.8  | 70.000                             | 2                    | 1234.9                              | 1.2   |
| 1250                              | 3   | 30.05.2016          | 19.655                           | 98109                       | 66.0                                  | 20.433                        | 98.2879                     | 1249.8                                   | 0.5  | 70.000                             | 3                    | 1249.5                              | 1.2   |
| 1250                              | 4   | 26.05.2016          | 19.750                           | 98776                       | 62.3                                  | 20.874                        | 98.9625                     | 1254.3                                   | 2.9  | 70.000                             | 4                    | 1253.9                              | 1.2   |
|                                   |     |                     |                                  |                             |                                       |                               |                             | Average=                                 | <b>1245.99</b>                                   |                                    |                      | Average=                            | <b>1245.70</b>                                    |
| <b>Change Flow Rate</b>           |     |                     | <b>Stabilization 3 minutes</b>   |                             |                                       |                               |                             |  |  |                                    |                      |                                     |   |
| 5000                              | 1   | 26.05.2016          | 19.715                           | 98820                       | 62.3                                  | 20.761                        | 99.0235                     | 5002.9                                   | 2.5  | 60                                 | 1                    | 5007.8                              | 4.6   |
| 5000                              | 2   | 26.05.2016          | 19.715                           | 98802                       | 62.3                                  | 21.005                        | 99.0097                     | 4968.0                                   | 2.1  | 60                                 | 2                    | 4970.9                              | 4.6   |
| 5000                              | 3   | 30.05.2016          | 19.675                           | 98107                       | 66.0                                  | 20.741                        | 98.3001                     | 5010.7                                   | 2.3  | 50                                 | 3                    | 5015.3                              | 4.8   |
| 5000                              | 4   | 26.05.2016          | 19.735                           | 98786                       | 62.3                                  | 21.053                        | 98.9912                     | 4965.0                                   | 2.3  | 60                                 | 4                    | 4970.0                              | 4.5   |
|                                   |     |                     |                                  |                             |                                       |                               |                             | Average=                                 | <b>4986.63</b>                                   |                                    |                      | Average=                            | <b>4991.00</b>                                    |
| <b>Change Flow Rate</b>           |     |                     | <b>Stabilization 3 minutes</b>   |                             |                                       |                               |                             |  |  |                                    |                      |                                     |   |
| 10000                             | 1   | 26.05.2016          | 19.695                           | 98825                       | 62.3                                  | 21.234                        | 99.0770                     | 9971.4                                   | 5.7  | 50                                 | 1                    | 9978.9                              | 9.0   |
| 10000                             | 2   | 26.05.2016          | 19.720                           | 98813                       | 62.3                                  | 21.127                        | 99.0862                     | 9931.4                                   | 2.5  | 40                                 | 2                    | 9934.4                              | 9.3   |
| 10000                             | 3   | 30.05.2016          | 19.670                           | 98105                       | 66.0                                  | 21.189                        | 98.3550                     | 9999.6                                   | 4.3  | 40                                 | 3                    | 10013.2                             | 9.4   |
| 10000                             | 4   | 26.05.2016          | 19.715                           | 98791                       | 62.3                                  | 21.132                        | 99.0536                     | 9919.1                                   | 3.9  | 40                                 | 4                    | 9934.0                              | 9.4   |
|                                   |     |                     |                                  |                             |                                       |                               |                             | Average=                                 | <b>9955.38</b>                                   |                                    |                      | Average=                            | <b>9965.12</b>                                    |
| <b>Change Flow Rate</b>           |     |                     | <b>Stabilization 3 minutes</b>   |                             |                                       |                               |                             |  |  |                                    |                      |                                     |   |
| 20000                             | 1   | 26.05.2016          | 19.695                           | 98822                       | 62.3                                  | 21.368                        | 99.2069                     | 20011.2                                  | 13.2   | 40                                 | 1                    | 20041.7                             | 15.9  |
| 20000                             | 2   | 26.05.2016          | 19.715                           | 98820                       | 62.3                                  | 21.175                        | 99.1936                     | 19879.6                                  | 10.8   | 30                                 | 2                    | 19899.8                             | 17.2  |
| 20000                             | 3   | 30.05.2016          | 19.680                           | 98106                       | 66.0                                  | 21.280                        | 98.4873                     | 19976.0                                  | 13.1   | 30                                 | 3                    | 19989.9                             | 17.4  |
| 20000                             | 4   | 26.05.2016          | 19.720                           | 98787                       | 62.3                                  | 21.157                        | 99.1757                     | 19855.5                                  | 14.0   | 30                                 | 4                    | 19882.2                             | 17.7  |
|                                   |     |                     |                                  |                             |                                       |                               |                             | Average=                                 | <b>19930.56</b>                                  |                                    |                      | Average=                            | <b>19953.40</b>                                   |
| <b>Change Flow Rate</b>           |     |                     | <b>Stabilization 3 minutes</b>   |                             |                                       |                               |                             |  |  |                                    |                      |                                     |   |
| 30000                             | 1   | 26.05.2016          | 19.695                           | 98828                       | 62.3                                  | 21.255                        | 99.3876                     | 29974.6                                  | 19.7   | 30                                 | 1                    | 29999.7                             | 21.9  |
| 30000                             | 2   | 26.05.2016          | 19.705                           | 98827                       | 62.3                                  | 21.163                        | 99.3970                     | 29944.3                                  | 16.6   | 30                                 | 2                    | 29962.4                             | 21.9  |
| 30000                             | 3   | 30.05.2016          | 19.675                           | 98102                       | 66.0                                  | 21.149                        | 98.6915                     | 29987.2                                  | 15.0   | 30                                 | 3                    | 30002.3                             | 21.6  |
| 30000                             | 4   | 26.05.2016          | 19.715                           | 98786                       | 62.3                                  | 21.177                        | 99.3715                     | 29912.1                                  | 20.5   | 30                                 | 4                    | 29952.4                             | 22.0  |
|                                   |     |                     |                                  |                             |                                       |                               |                             | Average=                                 | <b>29954.57</b>                                  |                                    |                      | Average=                            | <b>29979.18</b>                                   |
| <b>End of measurements</b>        |     |                     |                                  |                             |                                       |                               |                             |  |  |                                    |                      |                                     |   |

**INRIM loop1 stability results (INRIM2)**

**INITIAL TESTS**

Compare BIOS temperature and pressure readings with calibrated temperature and ambient pressure instrument readings.

|  |                     |                  |               |
|--|---------------------|------------------|---------------|
| Laboratory temperature                         | Laboratory pressure | BIOS temperature | BIOS pressure |
| (°C)   | (kPa)               | (°C)             | (kPa)         |
| Stabilization 24 hours in laboratory (no flow) |                     |                  |               |
| 20.1   | 98.716              | 20.9             | 98.709        |

Initial

|  |                     |                  |               |
|--|---------------------|------------------|---------------|
| Laboratory temperature                         | Laboratory pressure | BIOS temperature | BIOS pressure |
| (°C)   | (kPa)               | (°C)             | (kPa)         |
| Stabilization 24 hours in laboratory (no flow) |                     |                  |               |
| 20.5   | 98.61               | 21.09            | 98.73         |

Final

**COMPARISON MEASUREMENTS**

Flow cell: -10 (S/N: 135207)

| ENVIRONMENTAL PARAMETERS          |     |                     | TRANSFER STANDARD PARAMETERS     |                             |                                       |                               |                             |  |  |                                    | REFERENCE PARAMETERS |                                     |   |
|-----------------------------------|-----|---------------------|----------------------------------|-----------------------------|---------------------------------------|-------------------------------|-----------------------------|--|--|------------------------------------|----------------------|-------------------------------------|---|
| Nominal Flow Rate<br>(std ml/min) | RUN | Date<br>(dd.mm.yyy) | Laboratory temperature<br>(°C)   | Laboratory pressure<br>(Pa) | Laboratory relative humidity<br>(%rh) | BIOS mean temperature<br>(°C) | BIOS mean pressure<br>(kPa) | Mean Indicated Flow Rate<br>(std ml/min) | Standard deviation of indication<br>(std ml/min) | Number of flow indications in mean | RUN                  | Reference Flow Rate<br>(std ml/min) | Standard uncertainty of reference<br>(std ml/min) |
| <b>Set Flow Rate</b>              |     |                     | <b>Stabilization 180 minutes</b> |                             |                                       |                               |                             |  |  |                                    |                      |                                     |   |
| 5                                 | 1   | 13.06.2017          | 19.938                           | 98724                       | 71                                    | 20.924                        | 98.83                       | 5.01                                     | 0.00   | 10                                 | 1                    | 5.0525                              | 0.0022  |
| 5                                 | 2   | 13.06.2017          | 19.975                           | 98687                       | 71                                    | 21.048                        | 98.7945                     | 5.01                                     | 0.00   | 10                                 | 2                    | 5.0456                              | 0.0022  |
| 5                                 | 3   | 13.06.2017          | 19.978                           | 98676                       | 71                                    | 21.054                        | 98.7856                     | 5.01                                     | 0.00   | 10                                 | 3                    | 5.0421                              | 0.0022  |
| 5                                 | 4   | 13.06.2017          | 19.957                           | 98631                       | 71                                    | 21.1                          | 98.7383                     | 5.01                                     | 0.00   | 10                                 | 4                    | 5.0429                              | 0.0022  |
|                                   |     |                     |                                  |                             |                                       |                               |                             | Average=                                 | <b>5.01</b>                                      |                                    |                      | Average=                            | <b>5.05</b>                                       |
| <b>Change Flow Rate</b>           |     |                     | <b>Stabilization 3 minutes</b>   |                             |                                       |                               |                             |  |  |                                    |                      |                                     |   |
| 10                                | 1   | 13.06.2017          | 19.958                           | 98719                       | 71                                    | 20.968                        | 98.8224                     | 10.00                                    | 0.00   | 10                                 | 1                    | 10.055                              | 0.003   |
| 10                                | 2   | 13.06.2017          | 19.969                           | 98697                       | 71                                    | 21.04                         | 98.8076                     | 10.00                                    | 0.01   | 10                                 | 2                    | 10.044                              | 0.003   |
| 10                                | 3   | 13.06.2017          | 19.962                           | 98676                       | 71                                    | 21.064                        | 98.7816                     | 10.00                                    | 0.00   | 10                                 | 3                    | 10.046                              | 0.003   |
| 10                                | 4   | 13.06.2017          | 19.968                           | 98646                       | 71                                    | 21.108                        | 98.7522                     | 10.00                                    | 0.01   | 10                                 | 4                    | 10.039                              | 0.003   |
|                                   |     |                     |                                  |                             |                                       |                               |                             | Average=                                 | <b>10.00</b>                                     |                                    |                      | Average=                            | <b>10.05</b>                                      |
| <b>Change Flow Rate</b>           |     |                     | <b>Stabilization 3 minutes</b>   |                             |                                       |                               |                             |  |  |                                    |                      |                                     |   |
| 20                                | 1   | 13.06.2017          | 19.962                           | 98714                       | 71                                    | 20.988                        | 98.8217                     | 20.00                                    | 0.00   | 10                                 | 1                    | 20.072                              | 0.005   |
| 20                                | 2   | 13.06.2017          | 19.964                           | 98702                       | 71                                    | 21.027                        | 98.8112                     | 19.99                                    | 0.00   | 10                                 | 2                    | 20.052                              | 0.005   |
| 20                                | 3   | 13.06.2017          | 19.962                           | 98670                       | 71                                    | 21.07                         | 98.7799                     | 20.00                                    | 0.00   | 10                                 | 3                    | 20.055                              | 0.005   |
| 20                                | 4   | 13.06.2017          | 19.961                           | 98651                       | 71                                    | 21.096                        | 98.763                      | 19.99                                    | 0.00   | 10                                 | 4                    | 20.051                              | 0.005   |
|                                   |     |                     |                                  |                             |                                       |                               |                             | Average=                                 | <b>19.99</b>                                     |                                    |                      | Average=                            | <b>20.06</b>                                      |
| <b>Change Flow Rate</b>           |     |                     | <b>Stabilization 3 minutes</b>   |                             |                                       |                               |                             |  |  |                                    |                      |                                     |   |
| 80                                | 1   | 13.06.2017          | 19.962                           | 98709                       | 71                                    | 21                            | 98.8251                     | 79.99                                    | 0.01   | 10                                 | 1                    | 80.117                              | 0.02  |
| 80                                | 2   | 13.06.2017          | 19.956                           | 98705                       | 71                                    | 21.018                        | 98.8205                     | 80.00                                    | 0.01   | 10                                 | 2                    | 80.097                              | 0.02  |
| 80                                | 3   | 13.06.2017          | 19.966                           | 98662                       | 71                                    | 21.074                        | 98.7788                     | 79.98                                    | 0.01   | 10                                 | 3                    | 80.089                              | 0.02  |
| 80                                | 4   | 13.06.2017          | 19.959                           | 98654                       | 71                                    | 21.088                        | 98.7693                     | 79.99                                    | 0.02   | 10                                 | 4                    | 80.132                              | 0.02  |
|                                   |     |                     |                                  |                             |                                       |                               |                             | Average=                                 | <b>79.99</b>                                     |                                    |                      | Average=                            | <b>80.11</b>                                      |

**INITIAL TESTS**

Compare BIOS temperature and pressure readings with calibrated temperature and ambient pressure instrument readings.

| Laboratory temperature<br>(°C)                 | Laboratory pressure<br>(kPa) | BIOS temperature<br>(°C) | BIOS pressure<br>(kPa) |
|--|------------------------------|--------------------------|------------------------|
| Stabilization 24 hours in laboratory (no flow) |                              |                          |                        |
| 20.5   | 98.822                       | 20.8                     | 99.091                 |

Initial

| Laboratory temperature<br>(°C)                 | Laboratory pressure<br>(kPa) | BIOS temperature<br>(°C) | BIOS pressure<br>(kPa) |
|--|------------------------------|--------------------------|------------------------|
| Stabilization 24 hours in laboratory (no flow) |                              |                          |                        |
| 21.31  | 98.792                       | 21.51                    | 98.913                 |

Final

**COMPARISON MEASUREMENTS**

Flow cell: -24 (S/N: 134909)

| Flow cell: -24 (S/N: 134909)      |     |                     | ENVIRONMENTAL PARAMETERS         |                             |                                       | TRANSFER STANDARD PARAMETERS  |                             |  |  |                                    | REFERENCE PARAMETERS |                                     |   |
|-----------------------------------|-----|---------------------|----------------------------------|-----------------------------|---------------------------------------|-------------------------------|-----------------------------|--|--|------------------------------------|----------------------|-------------------------------------|---|
| Nominal Flow Rate<br>(std ml/min) | RUN | Date<br>(dd.mm.yyy) | Laboratory temperature<br>(°C)   | Laboratory pressure<br>(Pa) | Laboratory relative humidity<br>(%rh) | BIOS mean temperature<br>(°C) | BIOS mean pressure<br>(kPa) | Mean Indicated Flow Rate<br>(std ml/min) | Standard deviation of indication<br>(std ml/min) | Number of flow indications in mean | RUN                  | Reference Flow Rate<br>(std ml/min) | Standard uncertainty of reference<br>(std ml/min) |
| <b>Set Flow Rate</b>              |     |                     | <b>Stabilization 180 minutes</b> |                             |                                       |                               |                             |  |  |                                    |                      |                                     |   |
| 80                                | 1   | 16.06.2017          | 19.962                           | 98841                       | 71.1                                  | 21.29                         | 98.9526                     | 79.43                                    | 0.01   | 10                                 | 1                    | 80.011                              | 0.020   |
| 80                                | 2   | 16.06.2017          | 19.965                           | 98805                       | 71.1                                  | 21.404                        | 98.9153                     | 79.17                                    | 0.03   | 10                                 | 2                    | 79.742                              | 0.020   |
| 80                                | 3   | 16.06.2017          | 19.963                           | 98813                       | 71.1                                  | 21.428                        | 98.9237                     | 79.36                                    | 0.01   | 10                                 | 3                    | 79.866                              | 0.020   |
| 80                                | 4   | 16.06.2017          | 19.965                           | 98804                       | 71.1                                  | 21.508                        | 98.9155                     | 78.88                                    | 0.07   | 10                                 | 4                    | 79.445                              | 0.020   |
|                                   |     |                     |                                  |                             |                                       |                               |                             | Average=                                 | <b>79.21</b>                                     |                                    | Average=             | <b>79.77</b>                        |   |
| <b>Change Flow Rate</b>           |     |                     | <b>Stabilization 3 minutes</b>   |                             |                                       |                               |                             |  |  |                                    |                      |                                     |   |
| 300                               | 1   | 16.06.2017          | 19.954                           | 98839                       | 71.1                                  | 21.312                        | 98.97                       | 300.00                                   | 0.02   | 10                                 | 1                    | 301.81                              | 0.075   |
| 300                               | 2   | 16.06.2017          | 19.963                           | 98819                       | 71.1                                  | 21.374                        | 98.95                       | 299.75                                   | 0.02   | 10                                 | 2                    | 301.28                              | 0.075   |
| 300                               | 3   | 16.06.2017          | 19.963                           | 98816                       | 71.1                                  | 21.43                         | 98.9444                     | 299.92                                   | 0.03   | 10                                 | 3                    | 301.67                              | 0.075   |
| 300                               | 4   | 16.06.2017          | 19.971                           | 98804                       | 71.1                                  | 21.47                         | 98.9345                     | 299.73                                   | 0.01   | 10                                 | 4                    | 301.28                              | 0.075   |
|                                   |     |                     |                                  |                             |                                       |                               |                             | Average=                                 | <b>299.85</b>                                    |                                    | Average=             | <b>301.51</b>                       |   |
| <b>Change Flow Rate</b>           |     |                     | <b>Stabilization 3 minutes</b>   |                             |                                       |                               |                             |  |  |                                    |                      |                                     |   |
| 600                               | 1   | 16.06.2017          | 19.958                           | 98839                       | 71.1                                  | 21.3                          | 98.999                      | 599.75                                   | 0.05   | 10                                 | 1                    | 603.01                              | 0.15  |
| 600                               | 2   | 16.06.2017          | 19.960                           | 98825                       | 71.1                                  | 21.352                        | 98.985                      | 599.38                                   | 0.04   | 10                                 | 2                    | 602.50                              | 0.15  |
| 600                               | 3   | 16.06.2017          | 19.963                           | 98811                       | 71.1                                  | 21.426                        | 98.9737                     | 599.70                                   | 0.07   | 10                                 | 3                    | 602.99                              | 0.15  |
| 600                               | 4   | 16.06.2017          | 19.968                           | 98807                       | 71.1                                  | 21.45                         | 98.9681                     | 599.51                                   | 0.04   | 10                                 | 4                    | 602.69                              | 0.15  |
|                                   |     |                     |                                  |                             |                                       |                               |                             | Average=                                 | <b>599.59</b>                                    |                                    | Average=             | <b>602.80</b>                       |   |
| <b>Change Flow Rate</b>           |     |                     | <b>Stabilization 3 minutes</b>   |                             |                                       |                               |                             |  |  |                                    |                      |                                     |   |
| 1250                              | 1   | 16.06.2017          | 19.957                           | 98833                       | 71.1                                  | 21.304                        | 99.0952                     | 1249.97                                  | 0.13   | 10                                 | 1                    | 1256.63                             | 0.31  |
| 1250                              | 2   | 16.06.2017          | 19.959                           | 98827                       | 71.1                                  | 21.32                         | 99.0951                     | 1249.42                                  | 0.12   | 10                                 | 2                    | 1255.92                             | 0.31  |
| 1250                              | 3   | 16.06.2017          | 19.967                           | 98810                       | 71.1                                  | 21.41                         | 99.0772                     | 1249.83                                  | 0.10   | 10                                 | 3                    | 1256.50                             | 0.31  |
| 1250                              | 4   | 16.06.2017          | 19.967                           | 98810                       | 71.1                                  | 21.428                        | 99.0808                     | 1249.70                                  | 0.09   | 10                                 | 4                    | 1255.97                             | 0.31  |
|                                   |     |                     |                                  |                             |                                       |                               |                             | Average=                                 | <b>1249.73</b>                                   |                                    | Average=             | <b>1256.25</b>                      |   |

**INITIAL TESTS**

Compare BIOS temperature and pressure readings with calibrated temperature and ambient pressure instrument readings.

| Laboratory temperature (°C)                    | Laboratory pressure (kPa) | BIOS temperature (°C) | BIOS pressure (kPa) |
|--|---------------------------|-----------------------|---------------------|
| Stabilization 24 hours in laboratory (no flow) |                           |                       |                     |
| 19.45  | 98.978                    | 20.84                 | 99.555              |

Initial

| Laboratory temperature (°C)                    | Laboratory pressure (kPa) | BIOS temperature (°C) | BIOS pressure (kPa) |
|--|---------------------------|-----------------------|---------------------|
| Stabilization 24 hours in laboratory (no flow) |                           |                       |                     |
| 19.56  | 98.916                    | 20.92                 | 99.494              |

Final

**COMPARISON MEASUREMENTS**

Flow cell: -44 (S/N: 135198)

|                                |     |                  | ENVIRONMENTAL PARAMETERS         |                          |                                    | TRANSFER STANDARD PARAMETERS |                          |                                       |   |                                    | REFERENCE PARAMETERS |                                  |  |
|--------------------------------|-----|------------------|----------------------------------|--------------------------|------------------------------------|------------------------------|--------------------------|---------------------------------------|---|------------------------------------|----------------------|----------------------------------|--|
| Nominal Flow Rate (std ml/min) | RUN | Date (dd.mm.yyy) | Laboratory temperature (°C)      | Laboratory pressure (Pa) | Laboratory relative humidity (%rh) | BIOS mean temperature (°C)   | BIOS mean pressure (kPa) | Mean Indicated Flow Rate (std ml/min) | Standard deviation of indication (std ml/min) | Number of flow indications in mean | RUN                  | Reference Flow Rate (std ml/min) | Standard uncertainty of reference (std ml/min) |
| <b>Set Flow Rate</b>           |     |                  | <b>Stabilization 180 minutes</b> |                          |                                    |                              |                          |                                       |   |                                    |                      |                                  |  |
| 1250                           | 1   | 09.06.2017       | 19.420                           | 98979                    | 62.2                               | 20.855                       | 99.5568                  | 1246.6                                | 0.4   | 70                                 | 1                    | 1247.3                           | 1.2  |
| 1250                           | 2   | 09.06.2017       | 19.475                           | 98952                    | 62.2                               | 21.109                       | 99.5310                  | 1246.6                                | 2.1   | 80                                 | 2                    | 1246.4                           | 1.2  |
| 1250                           | 3   | 09.06.2017       | 19.515                           | 98940                    | 62.2                               | 21.002                       | 99.5167                  | 1250.9                                | 0.5   | 80                                 | 3                    | 1251.1                           | 1.2  |
| 1250                           | 4   | 09.06.2017       | 19.590                           | 98913                    | 62.2                               | 20.954                       | 99.4932                  | 1248.3                                | 2.4   | 80                                 | 4                    | 1247.5                           | 1.2  |
|                                |     |                  |                                  |                          |                                    |                              |                          | Average=                              | <b>1248.09</b>                                |                                    |                      | Average=                         | <b>1248.06</b>                                 |
| <b>Change Flow Rate</b>        |     |                  | <b>Stabilization 3 minutes</b>   |                          |                                    |                              |                          |                                       |   |                                    |                      |                                  |  |
| 5000                           | 1   | 09.06.2017       | 19.435                           | 98987                    | 62.2                               | 21.176                       | 99.5843                  | 5014.8                                | 2.0   | 60                                 | 1                    | 5020.2                           | 4.9  |
| 5000                           | 2   | 09.06.2017       | 19.455                           | 98954                    | 62.2                               | 21.323                       | 99.5472                  | 4981.6                                | 2.1   | 60                                 | 2                    | 4981.7                           | 4.8  |
| 5000                           | 3   | 09.06.2017       | 19.440                           | 98931                    | 62.2                               | 21.167                       | 99.5269                  | 5007.0                                | 3.3   | 60                                 | 3                    | 5006.7                           | 4.9  |
| 5000                           | 4   | 09.06.2017       | 19.585                           | 98913                    | 62.2                               | 21.156                       | 99.5083                  | 4983.6                                | 2.3   | 60                                 | 4                    | 4985.8                           | 4.8  |
|                                |     |                  |                                  |                          |                                    |                              |                          | Average=                              | <b>4996.73</b>                                |                                    |                      | Average=                         | <b>4998.61</b>                                 |
| <b>Change Flow Rate</b>        |     |                  | <b>Stabilization 3 minutes</b>   |                          |                                    |                              |                          |                                       |   |                                    |                      |                                  |  |
| 10000                          | 1   | 09.06.2017       | 19.420                           | 98983                    | 62.2                               | 21.562                       | 99.6370                  | 10016.2                               | 3.2   | 50                                 | 1                    | 10025.4                          | 9.2  |
| 10000                          | 2   | 09.06.2017       | 19.485                           | 98958                    | 62.2                               | 21.388                       | 99.6057                  | 9950.4                                | 4.2   | 50                                 | 2                    | 9953.3                           | 9.3  |
| 10000                          | 3   | 09.06.2017       | 19.445                           | 98931                    | 62.2                               | 21.493                       | 99.5716                  | 10007.3                               | 3.0   | 50                                 | 3                    | 10013.1                          | 9.2  |
| 10000                          | 4   | 09.06.2017       | 19.565                           | 98916                    | 62.2                               | 21.266                       | 99.5586                  | 9967.0                                | 5.7   | 50                                 | 4                    | 9971.1                           | 9.2  |
|                                |     |                  |                                  |                          |                                    |                              |                          | Average=                              | <b>9985.21</b>                                |                                    |                      | Average=                         | <b>9990.73</b>                                 |
| <b>Change Flow Rate</b>        |     |                  | <b>Stabilization 3 minutes</b>   |                          |                                    |                              |                          |                                       |   |                                    |                      |                                  |  |
| 20000                          | 1   | 09.06.2017       | 19.450                           | 98977                    | 62.2                               | 21.636                       | 99.7594                  | 20007.4                               | 11.2  | 40                                 | 1                    | 20014.2                          | 16.6   |
| 20000                          | 2   | 09.06.2017       | 19.495                           | 98964                    | 62.2                               | 21.420                       | 99.7468                  | 19923.0                               | 15.1  | 40                                 | 2                    | 19929.7                          | 17.1   |
| 20000                          | 3   | 09.06.2017       | 19.465                           | 98929                    | 62.2                               | 21.614                       | 99.7028                  | 20001.0                               | 12.0  | 40                                 | 3                    | 20014.5                          | 17.1   |
| 20000                          | 4   | 09.06.2017       | 19.575                           | 98912                    | 62.2                               | 21.311                       | 99.6795                  | 19974.2                               | 13.9  | 40                                 | 4                    | 19981.1                          | 17.0   |
|                                |     |                  |                                  |                          |                                    |                              |                          | Average=                              | <b>19976.42</b>                               |                                    |                      | Average=                         | <b>19984.88</b>                                |
| <b>Change Flow Rate</b>        |     |                  | <b>Stabilization 3 minutes</b>   |                          |                                    |                              |                          |                                       |   |                                    |                      |                                  |  |
| 30000                          | 1   | 09.06.2017       | 19.475                           | 98977                    | 62.2                               | 21.487                       | 99.9346                  | 29965.3                               | 21.3  | 40                                 | 1                    | 29962.3                          | 21.7   |
| 30000                          | 2   | 09.06.2017       | 19.485                           | 98970                    | 62.2                               | 21.424                       | 99.9261                  | 29976.2                               | 16.3  | 40                                 | 2                    | 29977.6                          | 21.8   |
| 30000                          | 3   | 09.06.2017       | 19.500                           | 98928                    | 62.2                               | 21.507                       | 99.8922                  | 29972.4                               | 22.5  | 40                                 | 3                    | 29977.2                          | 21.7   |
| 30000                          | 4   | 09.06.2017       | 19.585                           | 98911                    | 62.2                               | 21.232                       | 99.8608                  | 30051.6                               | 20.1  | 40                                 | 4                    | 30050.4                          | 21.1   |
|                                |     |                  |                                  |                          |                                    |                              |                          | Average=                              | <b>29991.36</b>                               |                                    |                      | Average=                         | <b>29991.86</b>                                |
| <b>End of measurements</b>     |     |                  |                                  |                          |                                    |                              |                          |                                       |   |                                    |                      |                                  |  |



**PTB loop1 results**

**INITIAL TESTS**

Compare BIOS temperature and pressure readings with calibrated temperature and ambient pressure instrument readings.

| Laboratory temperature<br>(°C)                 | Laboratory pressure<br>(kPa) | BIOS temperature<br>(°C) | BIOS pressure<br>(kPa) |
|--|------------------------------|--------------------------|------------------------|
| Stabilization 24 hours in laboratory (no flow) |                              |                          |                        |
| 21.85  | 1010.3                       | 21.75                    | 1010.2                 |

**COMPARISON MEASUREMENTS**

Flow cell: -10 (S/N: 135207)

| Flow cell: -10 (S/N: 135207)      |     |                     | ENVIRONMENTAL PARAMETERS         |                              |                                       | TRANSFER STANDARD PARAMETERS  |                             |  |  |                                    | REFERENCE PARAMETERS  |                                     |   |
|-----------------------------------|-----|---------------------|----------------------------------|------------------------------|---------------------------------------|-------------------------------|-----------------------------|--|--|------------------------------------|-----------------------|-------------------------------------|---|
| Nominal Flow Rate<br>(std ml/min) | RUN | Date<br>(dd.mm.yyy) | Laboratory temperature<br>(°C)   | Laboratory pressure<br>(kPa) | Laboratory relative humidity<br>(%rh) | BIOS mean temperature<br>(°C) | BIOS mean pressure<br>(kPa) | Mean Indicated Flow Rate<br>(std ml/min) | Standard deviation of indication<br>(std ml/min) | Number of flow indications in mean | RUN                   | Reference Flow Rate<br>(std ml/min) | Standard uncertainty of reference<br>(std ml/min) |
| <b>Set Flow Rate</b>              |     |                     | <b>Stabilization 180 minutes</b> |                              |                                       |                               |                             |  |  |                                    |                       |                                     |   |
| 5                                 | 1   | 09-08-2016          | 21.76                            | 100.99                       |                                       | 22.72                         | 101.12                      | 4.96                                     | 0.00   | 10                                 | 1                     | 4.978                               | 0.0075  |
| 5                                 | 2   | 09-08-2017          | 21.77                            | 100.99                       |                                       | 22.72                         | 101.12                      | 4.96                                     | 0.00   | 10                                 | 2                     | 4.978                               | 0.0075  |
| 5                                 | 3   | 09-08-2018          | 21.67                            | 100.98                       |                                       | 22.70                         | 101.11                      | 4.96                                     | 0.00   | 10                                 | 3                     | 4.979                               | 0.0075  |
| 5                                 | 4   | 09-08-2019          | 21.64                            | 101.00                       |                                       | 22.66                         | 101.11                      | 4.96                                     | 0.00   | 10                                 | 4                     | 4.980                               | 0.0075  |
|                                   |     |                     |                                  |                              |                                       |                               |                             | Average=                                 | <b>4.96</b>                                      |                                    | Average= <b>4.98</b>  |                                     |   |
| <b>Change Flow Rate</b>           |     |                     | <b>Stabilization 3 minutes</b>   |                              |                                       |                               |                             |  |  |                                    |                       |                                     |   |
| 10                                | 1   | 09-08-2019          | 21.60                            | 100.98                       |                                       | 22.66                         | 101.11                      | 10.01                                    | 0.00   | 10                                 | 1                     | 9.981                               | 0.015   |
| 10                                | 2   | 09-08-2020          | 21.60                            | 100.99                       |                                       | 22.66                         | 101.12                      | 10.01                                    | 0.00   | 10                                 | 2                     | 9.982                               | 0.015   |
| 10                                | 3   | 09-08-2021          | 21.63                            | 100.99                       |                                       | 22.66                         | 101.12                      | 10.00                                    | 0.00   | 10                                 | 3                     | 9.981                               | 0.015   |
| 10                                | 4   | 09-08-2022          | 21.66                            | 100.99                       |                                       | 22.68                         | 101.12                      | 9.99                                     | 0.00   | 10                                 | 4                     | 9.980                               | 0.015   |
|                                   |     |                     |                                  |                              |                                       |                               |                             | Average=                                 | <b>10.00</b>                                     |                                    | Average= <b>9.98</b>  |                                     |   |
| <b>Change Flow Rate</b>           |     |                     | <b>Stabilization 3 minutes</b>   |                              |                                       |                               |                             |  |  |                                    |                       |                                     |   |
| 20                                | 1   | 19.7.2016           | 22.02                            | 101.06                       |                                       | 22.62                         | 101.17                      | 19.85                                    |  | 10                                 | 1                     | 19.848                              | 0.03  |
| 20                                | 2   | 19.7.2016           | 22.02                            | 101.06                       |                                       | 22.62                         | 101.17                      | 19.85                                    |  | 10                                 | 2                     | 19.847                              | 0.03  |
| 20                                | 3   | 19.7.2016           | 22.02                            | 101.06                       |                                       | 22.62                         | 101.17                      | 19.85                                    |  | 10                                 | 3                     | 19.848                              | 0.03  |
| 20                                | 4   | 19.7.2016           | 22.02                            | 101.06                       |                                       | 22.64                         | 101.18                      | 19.85                                    |  | 10                                 | 4                     | 19.850                              | 0.03  |
|                                   |     |                     |                                  |                              |                                       |                               |                             | Average=                                 | <b>19.85</b>                                     |                                    | Average= <b>19.85</b> |                                     |   |
| <b>Change Flow Rate</b>           |     |                     | <b>Stabilization 3 minutes</b>   |                              |                                       |                               |                             |  |  |                                    |                       |                                     |   |
| 80                                | 1   | 28.7.2016           | 22.37                            | 100.64                       |                                       | 23.13                         | 100.76                      | 80.19                                    |  | 10                                 | 1                     | 80.222                              | 0.12  |
| 80                                | 2   | 28.7.2016           | 22.40                            | 100.64                       |                                       | 23.13                         | 100.76                      | 80.19                                    |  | 10                                 | 2                     | 80.219                              | 0.12  |
| 80                                | 3   | 28.7.2016           | 22.40                            | 100.64                       |                                       | 23.13                         | 100.76                      | 80.19                                    |  | 10                                 | 3                     | 80.220                              | 0.12  |
| 80                                | 4   | 28.7.2016           | 22.40                            | 100.64                       |                                       | 23.15                         | 100.76                      | 80.19                                    |  | 10                                 | 4                     | 80.222                              | 0.12  |
|                                   |     |                     |                                  |                              |                                       |                               |                             | Average=                                 | <b>80.19</b>                                     |                                    | Average= <b>80.22</b> |                                     |   |

**INITIAL TESTS**

Compare BIOS temperature and pressure readings with calibrated temperature and ambient pressure instrument readings.

| Laboratory temperature<br>(°C)                 | Laboratory pressure<br>(kPa) | BIOS temperature<br>(°C) | BIOS pressure<br>(kPa) |
|--|------------------------------|--------------------------|------------------------|
| Stabilization 24 hours in laboratory (no flow) |                              |                          |                        |
| 22.02  | 1010.3                       | 21.91                    | 1010.2                 |

**COMPARISON MEASUREMENTS**

Flow cell: -24 (S/N: 134909)

|                                   |     |                     | ENVIRONMENTAL PARAMETERS         |                              |                                       | TRANSFER STANDARD PARAMETERS  |                             |  |  |                                    | REFERENCE PARAMETERS    |                                     |   |
|-----------------------------------|-----|---------------------|----------------------------------|------------------------------|---------------------------------------|-------------------------------|-----------------------------|--|--|------------------------------------|-------------------------|-------------------------------------|---|
| Nominal Flow Rate<br>(std ml/min) | RUN | Date<br>(dd.mm.yyy) | Laboratory temperature<br>(°C)   | Laboratory pressure<br>(kPa) | Laboratory relative humidity<br>(%rh) | BIOS mean temperature<br>(°C) | BIOS mean pressure<br>(kPa) | Mean Indicated Flow Rate<br>(std ml/min) | Standard deviation of indication<br>(std ml/min) | Number of flow indications in mean | RUN                     | Reference Flow Rate<br>(std ml/min) | Standard uncertainty of reference<br>(std ml/min) |
| <b>Set Flow Rate</b>              |     |                     | <b>Stabilization 180 minutes</b> |                              |                                       |                               |                             |  |  |                                    |                         |                                     |   |
| 80                                | 1   | 28.7.2016           | 22.37                            | 100.39                       |                                       | 23.03                         | 100.52                      | 79.91                                    |  | 10                                 | 1                       | 80.248                              | 0.12  |
| 80                                | 2   | 29.7.2016           | 22.39                            | 100.38                       |                                       | 23.09                         | 100.51                      | 79.90                                    |  | 10                                 | 2                       | 80.232                              | 0.12  |
| 80                                | 3   | 30.7.2016           | 22.39                            | 100.38                       |                                       | 23.11                         | 100.51                      | 79.88                                    |  | 10                                 | 3                       | 80.218                              | 0.12  |
| 80                                | 4   | 31.7.2016           | 22.40                            | 100.38                       |                                       | 23.11                         | 100.50                      | 79.87                                    |  | 10                                 | 4                       | 80.211                              | 0.12  |
|                                   |     |                     |                                  |                              |                                       |                               |                             | Average=                                 | <b>79.89</b>                                     |                                    | Average= <b>80.23</b>   |                                     |   |
| <b>Change Flow Rate</b>           |     |                     | <b>Stabilization 3 minutes</b>   |                              |                                       |                               |                             |  |  |                                    |                         |                                     |   |
| 300                               | 1   | 13.7.2016           | 22.03                            | 100.66                       |                                       | 22.86                         | 100.78                      | 303.07                                   |  | 10                                 | 1                       | 304.2118                            | 0.46  |
| 300                               | 2   | 13.7.2016           | 22.03                            | 100.66                       |                                       | 22.88                         | 100.79                      | 303.09                                   |  | 10                                 | 2                       | 304.1972                            | 0.46  |
| 300                               | 3   | 13.7.2016           | 22.03                            | 100.66                       |                                       | 22.88                         | 100.79                      | 303.10                                   |  | 10                                 | 3                       | 304.1842                            | 0.46  |
| 300                               | 4   | 13.7.2016           | 22.03                            | 100.66                       |                                       | 22.90                         | 100.79                      | 303.09                                   |  | 10                                 | 4                       | 304.1644                            | 0.46  |
|                                   |     |                     |                                  |                              |                                       |                               |                             | Average=                                 | <b>303.09</b>                                    |                                    | Average= <b>304.19</b>  |                                     |   |
| <b>Change Flow Rate</b>           |     |                     | <b>Stabilization 3 minutes</b>   |                              |                                       |                               |                             |  |  |                                    |                         |                                     |   |
| 600                               | 1   | 14.7.2016           | 22.03                            | 100.67                       |                                       | 22.80                         | 100.80                      | 599.76                                   |  | 10                                 | 1                       | 601.780                             | 0.90  |
| 600                               | 2   | 14.7.2016           | 22.03                            | 100.67                       |                                       | 22.80                         | 100.80                      | 599.72                                   |  | 10                                 | 2                       | 601.823                             | 0.90  |
| 600                               | 3   | 14.7.2016           | 22.02                            | 100.67                       |                                       | 22.82                         | 100.80                      | 599.74                                   |  | 10                                 | 3                       | 601.827                             | 0.90  |
| 600                               | 4   | 14.7.2016           | 22.02                            | 100.67                       |                                       | 22.82                         | 100.80                      | 599.71                                   |  | 10                                 | 4                       | 601.829                             | 0.90  |
|                                   |     |                     |                                  |                              |                                       |                               |                             | Average=                                 | <b>599.73</b>                                    |                                    | Average= <b>601.81</b>  |                                     |   |
| <b>Change Flow Rate</b>           |     |                     | <b>Stabilization 3 minutes</b>   |                              |                                       |                               |                             |  |  |                                    |                         |                                     |   |
| 1250                              | 1   | 14.7.2016           | 22.14                            | 100.61                       |                                       | 22.63                         | 100.75                      | 1250.70                                  |  | 10                                 | 1                       | 1255.406                            | 1.88  |
| 1250                              | 2   | 14.7.2016           | 22.14                            | 100.61                       |                                       | 22.64                         | 100.75                      | 1249.50                                  |  | 10                                 | 2                       | 1254.173                            | 1.88  |
| 1250                              | 3   | 14.7.2016           | 22.14                            | 100.61                       |                                       | 22.64                         | 100.75                      | 1249.50                                  |  | 10                                 | 3                       | 1253.782                            | 1.88  |
| 1250                              | 4   | 14.7.2016           | 22.14                            | 100.61                       |                                       | 22.66                         | 100.77                      | 1248.80                                  |  | 10                                 | 4                       | 1253.577                            | 1.88  |
|                                   |     |                     |                                  |                              |                                       |                               |                             | Average=                                 | <b>1249.63</b>                                   |                                    | Average= <b>1254.23</b> |                                     |   |
| <b>Change Flow Rate</b>           |     |                     | <b>Stabilization 3 minutes</b>   |                              |                                       |                               |                             |  |  |                                    |                         |                                     |   |
| 1000                              | 1   | 14.7.2016           | 22.07                            | 100.67                       |                                       | 22.72                         | 100.82                      | 1003.70                                  |  | 10                                 | 1                       | 1007.111                            | 1.51  |
| 1000                              | 2   | 14.7.2016           | 22.07                            | 100.67                       |                                       | 22.72                         | 100.81                      | 1003.60                                  |  | 10                                 | 2                       | 1007.117                            | 1.51  |
| 1000                              | 3   | 14.7.2016           | 22.07                            | 100.67                       |                                       | 22.74                         | 100.82                      | 1003.60                                  |  | 10                                 | 3                       | 1007.123                            | 1.51  |
| 1000                              | 4   | 14.7.2016           | 22.07                            | 100.67                       |                                       | 22.74                         | 100.81                      | 1003.40                                  |  | 10                                 | 4                       | 1007.146                            | 1.51  |
|                                   |     |                     |                                  |                              |                                       |                               |                             | Average=                                 | <b>1003.58</b>                                   |                                    | Average= <b>1007.12</b> |                                     |   |

**INITIAL TESTS**

Compare BIOS temperature and pressure readings with calibrated temperature and ambient pressure instrument readings.

| Laboratory temperature (°C)                    | Laboratory pressure (kPa) | BIOS temperature (°C) | BIOS pressure (kPa) |
|--|---------------------------|-----------------------|---------------------|
| Stabilization 24 hours in laboratory (no flow) |                           |                       |                     |
| 23.74  | 101.01                    | 23.45                 | 100.85              |

**COMPARISON MEASUREMENTS**

Flow cell: -44 (S/N: 135198)

|                                |     |                  | ENVIRONMENTAL PARAMETERS         |                           |                                    | TRANSFER STANDARD PARAMETERS |                          |                                       |   |                                    | REFERENCE PARAMETERS |                                  |  |
|--------------------------------|-----|------------------|----------------------------------|---------------------------|------------------------------------|------------------------------|--------------------------|---------------------------------------|---|------------------------------------|----------------------|----------------------------------|--|
| Nominal Flow Rate (std ml/min) | RUN | Date (dd.mm.yyy) | Laboratory temperature (°C)      | Laboratory pressure (kPa) | Laboratory relative humidity (%rh) | BIOS mean temperature (°C)   | BIOS mean pressure (kPa) | Mean Indicated Flow Rate (std ml/min) | Standard deviation of indication (std ml/min) | Number of flow indications in mean | RUN                  | Reference Flow Rate (std ml/min) | Standard uncertainty of reference (std ml/min) |
| <b>Set Flow Rate</b>           |     |                  | <b>Stabilization 180 minutes</b> |                           |                                    |                              |                          |                                       |   |                                    |                      |                                  |  |
| 1250                           | 1   | 14.7.2016        | 22.08                            | 100.63                    |                                    | 22.24                        | 100.8                    | 1254.90                               |   | 10                                 | 1                    | 1253.866                         | 1.88   |
| 1250                           | 2   | 14.7.2016        | 22.08                            | 100.64                    |                                    | 22.26                        | 100.8                    | 1255.00                               |   | 10                                 | 2                    | 1253.908                         | 1.88   |
| 1250                           | 3   | 14.7.2016        | 22.07                            | 100.64                    |                                    | 22.28                        | 100.8                    | 1255.00                               |   | 10                                 | 3                    | 1253.914                         | 1.88   |
| 1250                           | 4   | 14.7.2016        | 22.07                            | 100.64                    |                                    | 22.28                        | 100.81                   | 1255.00                               |   | 10                                 | 4                    | 1253.875                         | 1.88   |
|                                |     |                  |                                  |                           |                                    |                              |                          | Average=                              |   |                                    | Average=             | 1253.89                          |  |
| <b>Change Flow Rate</b>        |     |                  | <b>Stabilization 3 minutes</b>   |                           |                                    |                              |                          |                                       |   |                                    |                      |                                  |  |
| 5000                           | 1   | 22.6.2016        | 22.00                            | 101.44                    | 49.14                              | 22.66                        | 101.58                   | 5004.19                               | 0.62  | 10                                 | 1                    | 5005.397                         | 7.51   |
| 5000                           | 2   | 22.6.2016        | 22.00                            | 101.44                    | 49.16                              | 22.63                        | 101.58                   | 5004.86                               | 0.49  | 10                                 | 2                    | 5005.628                         | 7.51   |
| 5000                           | 3   | 22.6.2016        | 22.00                            | 101.44                    | 49.10                              | 22.62                        | 101.59                   | 5005.22                               | 0.34  | 10                                 | 3                    | 5005.805                         | 7.51   |
| 5000                           | 4   | 22.6.2016        | 22.00                            | 101.44                    | 49.01                              | 22.62                        | 101.59                   | 5005.30                               | 0.43  | 10                                 | 4                    | 5006.075                         | 7.51   |
|                                |     |                  |                                  |                           |                                    |                              |                          | Average=                              |   |                                    | Average=             | 5005.73                          |  |
| <b>Change Flow Rate</b>        |     |                  | <b>Stabilization 3 minutes</b>   |                           |                                    |                              |                          |                                       |   |                                    |                      |                                  |  |
| 10000                          | 1   | 22.6.2016        | 22.04                            | 101.43                    | 48.60                              | 22.49                        | 101.62                   | 9992.29                               | 1.70  | 10                                 | 1                    | 9992.807                         | 14.99  |
| 10000                          | 2   | 22.6.2016        | 22.00                            | 101.43                    | 48.60                              | 22.46                        | 101.62                   | 9991.86                               | 1.27  | 10                                 | 2                    | 9992.122                         | 14.99  |
| 10000                          | 3   | 22.6.2016        | 22.00                            | 101.43                    | 48.60                              | 22.45                        | 101.62                   | 9991.15                               | 1.32  | 10                                 | 3                    | 9991.389                         | 14.99  |
| 10000                          | 4   | 22.6.2016        | 22.00                            | 101.43                    | 48.60                              | 22.45                        | 101.62                   | 9990.22                               | 0.96  | 10                                 | 4                    | 9990.681                         | 14.99  |
|                                |     |                  |                                  |                           |                                    |                              |                          | Average=                              |   |                                    | Average=             | 9991.75                          |  |
| <b>Change Flow Rate</b>        |     |                  | <b>Stabilization 3 minutes</b>   |                           |                                    |                              |                          |                                       |   |                                    |                      |                                  |  |
| 20000                          | 1   | 22.6.2016        | 22.10                            | 101.43                    | 48.64                              | 22.47                        | 101.74                   | 20058.50                              | 8.73  | 10                                 | 1                    | 20077.09                         | 30.12  |
| 20000                          | 2   | 22.6.2016        | 22.10                            | 101.43                    | 48.54                              | 22.41                        | 101.73                   | 20064.10                              | 7.00  | 10                                 | 2                    | 20074.03                         | 30.11  |
| 20000                          | 3   | 22.6.2016        | 22.10                            | 101.43                    | 48.47                              | 22.40                        | 101.73                   | 20069.00                              | 4.40  | 10                                 | 3                    | 20078.00                         | 30.12  |
| 20000                          | 4   | 22.6.2016        | 22.10                            | 101.43                    | 48.40                              | 22.40                        | 101.73                   | 20073.20                              | 4.87  | 10                                 | 4                    | 20078.01                         | 30.12  |
|                                |     |                  |                                  |                           |                                    |                              |                          | Average=                              |   |                                    | Average=             | 20076.78                         |  |
| <b>Change Flow Rate</b>        |     |                  | <b>Stabilization 3 minutes</b>   |                           |                                    |                              |                          |                                       |   |                                    |                      |                                  |  |
| 30000                          | 1   | 23.6.2016        | 22.50                            | 100.94                    | 49.52                              | 22.64                        | 101.34                   | 29942.78                              | 9.74  | 10                                 | 1                    | 29983.77                         | 44.98  |
| 30000                          | 2   | 23.6.2016        | 22.50                            | 100.94                    | 49.32                              | 22.57                        | 101.33                   | 29949.90                              | 9.59  | 10                                 | 2                    | 29991.51                         | 44.99  |
| 30000                          | 3   | 23.6.2016        | 22.50                            | 100.94                    | 49.26                              | 22.56                        | 101.33                   | 29953.00                              | 10.71   | 10                                 | 3                    | 29993.86                         | 44.99  |
| 30000                          | 4   | 23.6.2016        | 22.50                            | 100.94                    | 49.14                              | 22.55                        | 101.33                   | 29961.20                              | 8.64  | 10                                 | 4                    | 29998.27                         | 45.00  |
|                                |     |                  |                                  |                           |                                    |                              |                          | Average=                              |   |                                    | Average=             | 29991.85                         |  |
| <b>End of measurements</b>     |     |                  |                                  |                           |                                    |                              |                          | Average=                              |   |                                    | Average=             | 29951.72                         |  |

**PTB loop2 results**

**INITIAL TESTS**

Compare BIOS temperature and pressure readings with calibrated temperature and ambient pressure instrument readings.

| Laboratory temperature<br>(°C)                 | Laboratory pressure<br>(kPa) | BIOS temperature<br>(°C) | BIOS pressure<br>(kPa) |
|--|------------------------------|--------------------------|------------------------|
| Stabilization 24 hours in laboratory (no flow) |                              |                          |                        |
| 21.85  | 1010.3                       | 21.81                    | 1010.5                 |

**COMPARISON MEASUREMENTS**

Flow cell: -10 (S/N: 135208)

|                                   |     |                     | ENVIRONMENTAL PARAMETERS         |                              |                                       | TRANSFER STANDARD PARAMETERS  |                             |  |  |                                    | REFERENCE PARAMETERS |                                     |   |
|-----------------------------------|-----|---------------------|----------------------------------|------------------------------|---------------------------------------|-------------------------------|-----------------------------|--|--|------------------------------------|----------------------|-------------------------------------|---|
| Nominal Flow Rate<br>(std ml/min) | RUN | Date<br>(dd.mm.yyy) | Laboratory temperature<br>(°C)   | Laboratory pressure<br>(kPa) | Laboratory relative humidity<br>(%rh) | BIOS mean temperature<br>(°C) | BIOS mean pressure<br>(kPa) | Mean Indicated Flow Rate<br>(std ml/min) | Standard deviation of indication<br>(std ml/min) | Number of flow indications in mean | RUN                  | Reference Flow Rate<br>(std ml/min) | Standard uncertainty of reference<br>(std ml/min) |
| <b>Set Flow Rate</b>              |     |                     | <b>Stabilization 180 minutes</b> |                              |                                       |                               |                             |  |  |                                    |                      |                                     |   |
| 5                                 | 1   | 08.08.2016          | 21.69                            | 100.88                       |                                       | 22.62                         | 101.05                      | 4.9739                                   | 0.0013   | 10                                 | 1                    | 4.982                               | 0.0075  |
| 5                                 | 2   | 08.08.2017          | 21.78                            | 100.87                       |                                       | 22.68                         | 101.04                      | 4.9703                                   | 0.0023   | 10                                 | 2                    | 4.979                               | 0.0075  |
| 5                                 | 3   | 08.08.2018          | 21.76                            | 100.87                       |                                       | 22.66                         | 101.03                      | 4.9699                                   | 0.0010   | 10                                 | 3                    | 4.980                               | 0.0075  |
| 5                                 | 4   | 08.08.2019          | 21.77                            | 100.86                       |                                       | 22.68                         | 101.03                      | 4.9712                                   | 0.0010   | 10                                 | 4                    | 4.980                               | 0.0075  |
|                                   |     |                     |                                  |                              |                                       |                               |                             | Average=                                 | <b>4.97</b>                                      |                                    |                      | Average=                            | <b>4.98</b>                                       |
| <b>Change Flow Rate</b>           |     |                     | <b>Stabilization 3 minutes</b>   |                              |                                       |                               |                             |  |  |                                    |                      |                                     |   |
| 10                                | 1   | 08.08.2019          | 21.77                            | 100.83                       |                                       | 22.64                         | 100.99                      | 9.9972                                   | 0.0035   | 10                                 | 1                    | 10.005                              | 0.015   |
| 10                                | 2   | 08.08.2019          | 21.76                            | 100.83                       |                                       | 22.64                         | 101.00                      | 9.9978                                   | 0.0030   | 10                                 | 2                    | 10.006                              | 0.015   |
| 10                                | 3   | 08.08.2019          | 21.76                            | 100.83                       |                                       | 22.64                         | 101.00                      | 9.9978                                   | 0.0025   | 10                                 | 3                    | 10.006                              | 0.015   |
| 10                                | 4   | 08.08.2019          | 21.75                            | 100.83                       |                                       | 22.64                         | 101.00                      | 10                                       | 0.0015   | 10                                 | 4                    | 10.006                              | 0.015   |
|                                   |     |                     |                                  |                              |                                       |                               |                             | Average=                                 | <b>10.00</b>                                     |                                    |                      | Average=                            | <b>10.01</b>                                      |
| <b>Change Flow Rate</b>           |     |                     | <b>Stabilization 3 minutes</b>   |                              |                                       |                               |                             |  |  |                                    |                      |                                     |   |
| 20                                | 1   | 19.7.2016           | 21.64                            | 101.244                      |                                       | 22.4                          | 101.41                      | 19.945                                   |  | 10                                 | 1                    | 19.916                              | 0.03  |
| 20                                | 2   | 19.7.2016           | 21.64                            | 101.25                       |                                       | 22.36                         | 101.41                      | 19.952                                   |  | 10                                 | 2                    | 19.922                              | 0.03  |
| 20                                | 3   | 19.7.2016           | 21.64                            | 101.248                      |                                       | 22.38                         | 101.40                      | 19.946                                   |  | 10                                 | 3                    | 19.918                              | 0.03  |
| 20                                | 4   | 19.7.2016           | 21.64                            | 101.248                      |                                       | 22.38                         | 101.41                      | 19.95                                    |  | 10                                 | 4                    | 19.918                              | 0.03  |
|                                   |     |                     |                                  |                              |                                       |                               |                             | Average=                                 | <b>19.95</b>                                     |                                    |                      | Average=                            | <b>19.92</b>                                      |
| <b>Change Flow Rate</b>           |     |                     | <b>Stabilization 3 minutes</b>   |                              |                                       |                               |                             |  |  |                                    |                      |                                     |   |
| 80                                | 1   | 28.7.2016           | 22.41                            | 100.549                      |                                       | 23.19                         | 100.71                      | 80.318                                   |  | 10                                 | 1                    | 80.198                              | 0.12  |
| 80                                | 2   | 28.7.2016           | 22.41                            | 100.549                      |                                       | 23.19                         | 100.71                      | 80.318                                   |  | 10                                 | 2                    | 80.198                              | 0.12  |
| 80                                | 3   | 28.7.2016           | 22.4                             | 100.547                      |                                       | 23.21                         | 100.71                      | 80.317                                   |  | 10                                 | 3                    | 80.200                              | 0.12  |
| 80                                | 4   | 28.7.2016           | 22.41                            | 100.549                      |                                       | 23.21                         | 100.71                      | 80.314                                   |  | 10                                 | 4                    | 80.200                              | 0.12  |
|                                   |     |                     |                                  |                              |                                       |                               |                             | Average=                                 | <b>80.32</b>                                     |                                    |                      | Average=                            | <b>80.20</b>                                      |

**INITIAL TESTS**

Compare BIOS temperature and pressure readings with calibrated temperature and ambient pressure instrument readings.

| Laboratory temperature<br>(°C)                 | Laboratory pressure<br>(kPa) | BIOS temperature<br>(°C) | BIOS pressure<br>(kPa) |
|--|------------------------------|--------------------------|------------------------|
| Stabilization 24 hours in laboratory (no flow) |                              |                          |                        |
| 22.02  | 1010.3                       | 21.97                    | 1010.5                 |

**COMPARISON MEASUREMENTS**

Flow cell: -24 (S/N: 134910)

|                                   |     |                     | ENVIRONMENTAL PARAMETERS         |                              |                                       | TRANSFER STANDARD PARAMETERS  |                             |  |  |                                    | REFERENCE PARAMETERS |                                     |   |
|-----------------------------------|-----|---------------------|----------------------------------|------------------------------|---------------------------------------|-------------------------------|-----------------------------|--|--|------------------------------------|----------------------|-------------------------------------|---|
| Nominal Flow Rate<br>(std ml/min) | RUN | Date<br>(dd.mm.yyy) | Laboratory temperature<br>(°C)   | Laboratory pressure<br>(kPa) | Laboratory relative humidity<br>(%rh) | BIOS mean temperature<br>(°C) | BIOS mean pressure<br>(kPa) | Mean Indicated Flow Rate<br>(std ml/min) | Standard deviation of indication<br>(std ml/min) | Number of flow indications in mean | RUN                  | Reference Flow Rate<br>(std ml/min) | Standard uncertainty of reference<br>(std ml/min) |
| <b>Set Flow Rate</b>              |     |                     | <b>Stabilization 180 minutes</b> |                              |                                       |                               |                             |  |  |                                    |                      |                                     |   |
| 80                                | 1   | 28.7.2016           | 22.34                            | 100.43                       |                                       | 22.99                         | 100.58                      | 80.10                                    |  | 10                                 | 1                    | 80.232                              | 0.12  |
| 80                                | 2   | 28.7.2016           | 22.34                            | 100.42                       |                                       | 23.01                         | 100.58                      | 80.06                                    |  | 10                                 | 2                    | 80.224                              | 0.12  |
| 80                                | 3   | 28.7.2016           | 22.38                            | 100.414                      |                                       | 23.03                         | 100.58                      | 80.04                                    |  | 10                                 | 3                    | 80.211                              | 0.12  |
| 80                                | 4   | 28.7.2016           | 22.38                            | 100.414                      |                                       | 23.05                         | 100.58                      | 80.08                                    |  | 10                                 | 4                    | 80.200                              | 0.12  |
|                                   |     |                     |                                  |                              |                                       |                               |                             | Average=                                 | <b>80.07</b>                                     |                                    |                      | Average=                            | <b>80.22</b>                                      |
| <b>Change Flow Rate</b>           |     |                     | <b>Stabilization 3 minutes</b>   |                              |                                       |                               |                             |  |  |                                    |                      |                                     |   |
| 300                               | 1   | 13.7.2016           | 22.09                            | 100.282                      |                                       | 22.5                          | 100.44                      | 303.21                                   |  | 10                                 | 1                    | 303.122                             | 0.45  |
| 300                               | 2   | 13.7.2016           | 22.13                            | 100.284                      |                                       | 22.54                         | 100.44                      | 303.12                                   |  | 10                                 | 2                    | 303.081                             | 0.45  |
| 300                               | 3   | 13.7.2016           | 22.13                            | 100.284                      |                                       | 22.54                         | 100.44                      | 303.11                                   |  | 10                                 | 3                    | 303.076                             | 0.45  |
| 300                               | 4   | 13.7.2016           | 22.13                            | 100.29                       |                                       | 22.54                         | 100.44                      | 303.10                                   |  | 10                                 | 4                    | 303.100                             | 0.45  |
|                                   |     |                     |                                  |                              |                                       |                               |                             | Average=                                 | <b>303.14</b>                                    |                                    |                      | Average=                            | <b>303.09</b>                                     |
| <b>Change Flow Rate</b>           |     |                     | <b>Stabilization 3 minutes</b>   |                              |                                       |                               |                             |  |  |                                    |                      |                                     |   |
| 600                               | 1   | 13.7.2016           | 22.05                            | 100.281                      |                                       | 22.56                         | 100.43                      | 600.90                                   |  | 10                                 | 1                    | 600.889                             | 0.90  |
| 600                               | 2   | 13.7.2016           | 22.05                            | 100.281                      |                                       | 22.58                         | 100.44                      | 600.93                                   |  | 10                                 | 2                    | 600.927                             | 0.90  |
| 600                               | 3   | 13.7.2016           | 22.05                            | 100.284                      |                                       | 22.58                         | 100.44                      | 600.82                                   |  | 10                                 | 3                    | 600.791                             | 0.90  |
| 600                               | 4   | 13.7.2016           | 22.05                            | 100.292                      |                                       | 22.6                          | 100.45                      | 600.88                                   |  | 10                                 | 4                    | 600.831                             | 0.90  |
|                                   |     |                     |                                  |                              |                                       |                               |                             | Average=                                 | <b>600.88</b>                                    |                                    |                      | Average=                            | <b>600.86</b>                                     |
| <b>Change Flow Rate</b>           |     |                     | <b>Stabilization 3 minutes</b>   |                              |                                       |                               |                             |  |  |                                    |                      |                                     |   |
| 1250                              | 1   | 14.7.2016           | 22.05                            | 100.711                      |                                       | 22.62                         | 100.88                      | 1255.40                                  |  | 10                                 | 1                    | 1256.060                            | 1.88  |
| 1250                              | 2   | 14.7.2016           | 22.05                            | 100.714                      |                                       | 22.62                         | 100.89                      | 1255.30                                  |  | 10                                 | 2                    | 1255.942                            | 1.88  |
| 1250                              | 3   | 14.7.2016           | 22.07                            | 100.714                      |                                       | 22.62                         | 100.89                      | 1255.30                                  |  | 10                                 | 3                    | 1255.840                            | 1.88  |
| 1250                              | 4   | 14.7.2016           | 22.07                            | 100.714                      |                                       | 22.64                         | 100.9                       | 1255.30                                  |  | 10                                 | 4                    | 1255.757                            | 1.88  |
|                                   |     |                     |                                  |                              |                                       |                               |                             | Average=                                 | <b>1255.33</b>                                   |                                    |                      | Average=                            | <b>1255.90</b>                                    |
| <b>Change Flow Rate</b>           |     |                     | <b>Stabilization 3 minutes</b>   |                              |                                       |                               |                             |  |  |                                    |                      |                                     |   |
| 1000                              | 1   | 13.7.2016           | 22.04                            | 100.342                      |                                       | 22.62                         | 100.53                      | 1002.60                                  |  | 10                                 | 1                    | 1003.222                            | 1.50  |
| 1000                              | 2   | 13.7.2016           | 22.04                            | 100.342                      |                                       | 22.64                         | 100.51                      | 1002.60                                  |  | 10                                 | 2                    | 1003.185                            | 1.50  |
| 1000                              | 3   | 13.7.2016           | 22.04                            | 100.345                      |                                       | 22.66                         | 100.51                      | 1002.60                                  |  | 10                                 | 3                    | 1003.140                            | 1.50  |
| 1000                              | 4   | 13.7.2016           | 22.05                            | 100.348                      |                                       | 22.66                         | 100.53                      | 1002.70                                  |  | 10                                 | 4                    | 1003.100                            | 1.50  |
|                                   |     |                     |                                  |                              |                                       |                               |                             | Average=                                 | <b>1002.63</b>                                   |                                    |                      | Average=                            | <b>1003.16</b>                                    |

**INITIAL TESTS**

Compare BIOS temperature and pressure readings with calibrated temperature and ambient pressure instrument readings.

| Laboratory temperature (°C)                    | Laboratory pressure (kPa) | BIOS temperature (°C) | BIOS pressure (kPa) |
|--|---------------------------|-----------------------|---------------------|
| Stabilization 24 hours in laboratory (no flow) |                           |                       |                     |
| 22.66  | 101.17                    | 22.74                 | 101.04              |

**COMPARISON MEASUREMENTS**

Flow cell: -44 (S/N: 135199)

| ENVIRONMENTAL PARAMETERS       |     |                  | TRANSFER STANDARD PARAMETERS     |                           |                                    |                            |                          | REFERENCE PARAMETERS                  |   |                                    |     |                                  |  |
|--------------------------------|-----|------------------|----------------------------------|---------------------------|------------------------------------|----------------------------|--------------------------|---------------------------------------|---|------------------------------------|-----|----------------------------------|--|
| Nominal Flow Rate (std ml/min) | RUN | Date (dd.mm.yyy) | Laboratory temperature (°C)      | Laboratory pressure (kPa) | Laboratory relative humidity (%rh) | BIOS mean temperature (°C) | BIOS mean pressure (kPa) | Mean Indicated Flow Rate (std ml/min) | Standard deviation of indication (std ml/min) | Number of flow indications in mean | RUN | Reference Flow Rate (std ml/min) | Standard uncertainty of reference (std ml/min) |
| <b>Set Flow Rate</b>           |     |                  | <b>Stabilization 180 minutes</b> |                           |                                    |                            |                          |                                       |   |                                    |     |                                  |  |
| 1250                           | 1   | 18.7.2016        | 21.34                            | 1010.58                   |                                    | 22.07                      | 1012.5                   | 1255.10                               |   | 10                                 | 1   | 1255.045                         | 1.88   |
| 1250                           | 2   | 18.7.2016        | 21.34                            | 1010.6                    |                                    | 22.07                      | 1012.5                   | 1255.00                               |   | 10                                 | 2   | 1254.762                         | 1.88   |
| 1250                           | 3   | 18.7.2016        | 21.35                            | 1010.63                   |                                    | 22.07                      | 1012.7                   | 1254.70                               |   | 10                                 | 3   | 1254.61                          | 1.88   |
| 1250                           | 4   | 18.7.2016        | 21.37                            | 1010.65                   |                                    | 22.09                      | 1012.5                   | 1254.70                               |   | 10                                 | 4   | 1254.468                         | 1.88   |
|                                |     |                  |                                  |                           |                                    |                            |                          | Average=                              | <b>1254.88</b>                                |                                    |     | Average=                         | <b>1254.72</b>                                 |
| <b>Change Flow Rate</b>        |     |                  | <b>Stabilization 3 minutes</b>   |                           |                                    |                            |                          |                                       |   |                                    |     |                                  |  |
| 5000                           | 1   | 28.6.2016        | 21.63                            | 101.04                    | 38.48                              | 22.7                       | 101.22                   | 5003.85                               | 0.72  | 10                                 | 1   | 5006.223                         | 7.51   |
| 5000                           | 2   | 28.6.2016        | 21.61                            | 101.05                    | 38.40                              | 22.6                       | 101.22                   | 5005.06                               | 0.72  | 10                                 | 2   | 5006.636                         | 7.51   |
| 5000                           | 3   | 28.6.2016        | 21.69                            | 101.05                    | 38.37                              | 22.6                       | 101.22                   | 5005.34                               | 0.47  | 10                                 | 3   | 5006.874                         | 7.51   |
| 5000                           | 4   | 28.6.2016        | 21.70                            | 101.05                    | 38.34                              | 22.6                       | 101.22                   | 5004.78                               | 0.47  | 10                                 | 4   | 5006.9                           | 7.51   |
|                                |     |                  |                                  |                           |                                    |                            |                          | Average=                              | <b>5004.76</b>                                |                                    |     | Average=                         | <b>5006.66</b>                                 |
| <b>Change Flow Rate</b>        |     |                  | <b>Stabilization 3 minutes</b>   |                           |                                    |                            |                          |                                       |   |                                    |     |                                  |  |
| 10000                          | 1   | 28.6.2016        | 21.70                            | 101.03                    | 37.93                              | 22.4                       | 101.27                   | 9996.81                               | 3.25  | 10                                 | 1   | 10002.78                         | 15.00  |
| 10000                          | 2   | 28.6.2016        | 21.70                            | 101.03                    | 37.93                              | 22.3                       | 101.26                   | 9999.92                               | 2.67  | 10                                 | 2   | 10002.74                         | 15.00  |
| 10000                          | 3   | 28.6.2016        | 21.70                            | 101.03                    | 37.91                              | 22.3                       | 101.25                   | 9998.82                               | 1.59  | 10                                 | 3   | 10002.97                         | 15.00  |
| 10000                          | 4   | 28.6.2016        | 21.70                            | 101.03                    | 37.99                              | 22.3                       | 101.26                   | 10000.50                              | 3.74  | 10                                 | 4   | 10003.26                         | 15.00  |
|                                |     |                  |                                  |                           |                                    |                            |                          | Average=                              | <b>9999.01</b>                                |                                    |     | Average=                         | <b>10002.94</b>                                |
| <b>Change Flow Rate</b>        |     |                  | <b>Stabilization 3 minutes</b>   |                           |                                    |                            |                          |                                       |   |                                    |     |                                  |  |
| 20000                          | 1   | 28.6.2016        | 21.70                            | 100.89                    | 37.50                              | 22.1                       | 101.20                   | 20071.70                              | 6.13  | 10                                 | 1   | 20080.65                         | 30.12  |
| 20000                          | 2   | 28.6.2016        | 21.70                            | 100.89                    | 37.50                              | 22.0                       | 101.18                   | 20064.10                              | 4.18  | 10                                 | 2   | 20080.4                          | 30.12  |
| 20000                          | 3   | 28.6.2016        | 21.70                            | 100.89                    | 37.44                              | 22.0                       | 101.18                   | 20066.70                              | 4.69  | 10                                 | 3   | 20080.89                         | 30.12  |
| 20000                          | 4   | 28.6.2016        | 21.70                            | 100.89                    | 37.40                              | 22.0                       | 101.18                   | 20064.90                              | 10.30   | 10                                 | 4   | 20080.24                         | 30.12  |
|                                |     |                  |                                  |                           |                                    |                            |                          | Average=                              | <b>20066.85</b>                               |                                    |     | Average=                         | <b>20080.54</b>                                |
| <b>Change Flow Rate</b>        |     |                  | <b>Stabilization 3 minutes</b>   |                           |                                    |                            |                          |                                       |   |                                    |     |                                  |  |
| 30000                          | 1   | 28.6.2016        | 21.70                            | 100.88                    | 36.86                              | 22.0                       | 101.27                   | 29944.00                              | 11.34   | 10                                 | 1   | 29955.96                         | 44.93  |
| 30000                          | 2   | 28.6.2016        | 21.70                            | 100.88                    | 36.90                              | 21.9                       | 101.27                   | 29950.60                              | 8.41  | 10                                 | 2   | 29956.11                         | 44.93  |
| 30000                          | 3   | 28.6.2016        | 21.70                            | 100.88                    | 36.90                              | 21.9                       | 101.27                   | 29939.70                              | 12.35   | 10                                 | 3   | 29955.81                         | 44.93  |
| 30000                          | 4   | 28.6.2016        | 21.70                            | 100.88                    | 36.94                              | 21.9                       | 101.28                   | 29946.30                              | 8.04  | 10                                 | 4   | 29955.11                         | 44.93  |
|                                |     |                  |                                  |                           |                                    |                            |                          | Average=                              | <b>29945.15</b>                               |                                    |     | Average=                         | <b>29955.75</b>                                |
| <b>End of measurements</b>     |     |                  |                                  |                           |                                    |                            |                          |                                       |   |                                    |     |                                  |  |

**PTB loop2 stability results (PTB2)**

**INITIAL TESTS**

Compare BIOS temperature and pressure readings with calibrated temperature and ambient pressure instrument readings.

| Laboratory temperature (°C)                    | Laboratory pressure (kPa) | BIOS temperature (°C) | BIOS pressure (kPa) |
|--|---------------------------|-----------------------|---------------------|
| Stabilization 24 hours in laboratory (no flow) |                           |                       |                     |
| 21.85  | 101.03                    | 21.8                  | 101.05              |

**COMPARISON MEASUREMENTS (LOOP2)**

Base: S/N: 135516

Flow cell: -10 (S/N: 135208)

|                                |     |                  | ENVIRONMENTAL PARAMETERS         |                           |                                    | TRANSFER STANDARD PARAMETERS |                          |                                       |   |                                    | REFERENCE PARAMETERS |                                  |  |
|--------------------------------|-----|------------------|----------------------------------|---------------------------|------------------------------------|------------------------------|--------------------------|---------------------------------------|---|------------------------------------|----------------------|----------------------------------|--|
| Nominal Flow Rate (std ml/min) | RUN | Date (dd.mm.yyy) | Laboratory temperature (°C)      | Laboratory pressure (kPa) | Laboratory relative humidity (%rh) | BIOS mean temperature (°C)   | BIOS mean pressure (kPa) | Mean Indicated Flow Rate (std ml/min) | Standard deviation of indication (std ml/min) | Number of flow indications in mean | RUN                  | Reference Flow Rate (std ml/min) | Standard uncertainty of reference (std ml/min) |
| <b>Set Flow Rate</b>           |     |                  | <b>Stabilization 180 minutes</b> |                           |                                    |                              |                          |                                       |   |                                    |                      |                                  |  |
| 5                              | 1   | 21.7.2017        | 21.97                            | 100.45                    | 45                                 | 22.62                        | 100.58                   | 5.029                                 | 0.0010  | 10                                 | 1                    | 5.050                            | 0.008  |
| 5                              | 2   | 21.7.2017        | 21.92                            | 100.42                    | 45                                 | 22.66                        | 100.58                   | 5.029                                 | 0.0011  | 10                                 | 2                    | 5.051                            | 0.008  |
| 5                              | 3   | 21.7.2017        | 21.92                            | 100.40                    | 45                                 | 22.74                        | 100.55                   | 5.025                                 | 0.0017  | 10                                 | 3                    | 5.050                            | 0.008  |
| 5                              | 4   | 21.7.2017        | 21.97                            | 100.39                    | 45                                 | 22.74                        | 100.55                   | 5.021                                 | 0.0007  | 10                                 | 4                    | 5.050                            | 0.008  |
|                                |     |                  |                                  |                           |                                    |                              |                          | Average=                              | <b>5.03</b>                                   |                                    |                      | Average=                         | <b>5.05</b>                                    |
| <b>Change Flow Rate</b>        |     |                  | <b>Stabilization 3 minutes</b>   |                           |                                    |                              |                          |                                       |   |                                    |                      |                                  |  |
| 10                             | 1   | 24.7.2017        | 21.94                            | 99.79                     | 45                                 | 22.46                        | 99.925                   | 10.342                                | 0.0021  | 10                                 | 1                    | 10.365                           | 0.016  |
| 10                             | 2   | 24.7.2017        | 21.96                            | 99.76                     | 45                                 | 22.52                        | 99.918                   | 10.339                                | 0.0037  | 10                                 | 2                    | 10.363                           | 0.016  |
| 10                             | 3   | 24.7.2017        | 21.96                            | 99.75                     | 45                                 | 22.56                        | 99.91                    | 10.333                                | 0.0021  | 10                                 | 3                    | 10.362                           | 0.016  |
| 10                             | 4   | 24.7.2017        | 21.96                            | 99.75                     | 45                                 | 22.6                         | 99.908                   | 10.326                                | 0.0030  | 10                                 | 4                    | 10.362                           | 0.016  |
|                                |     |                  |                                  |                           |                                    |                              |                          | Average=                              | <b>10.34</b>                                  |                                    |                      | Average=                         | <b>10.36</b>                                   |
| <b>Change Flow Rate</b>        |     |                  | <b>Stabilization 3 minutes</b>   |                           |                                    |                              |                          |                                       |   |                                    |                      |                                  |  |
| 20                             | 1   | 3.7.2017         | 21.85                            | 100.81                    | 48                                 | 22.76                        | 101                      | 19.915                                | 0.0050  | 10                                 | 1                    | 19.952                           | 0.030  |
| 20                             | 2   | 3.7.2017         | 21.85                            | 100.81                    | 48                                 | 22.76                        | 101                      | 19.927                                | 0.0031  | 10                                 | 2                    | 19.952                           | 0.030  |
| 20                             | 3   | 3.7.2017         | 21.85                            | 100.80                    | 48                                 | 22.76                        | 100.97                   | 19.932                                | 0.0049  | 10                                 | 3                    | 19.952                           | 0.030  |
| 20                             | 4   | 3.7.2017         | 21.85                            | 100.80                    | 48                                 | 22.76                        | 100.97                   | 19.938                                | 0.0032  | 10                                 | 4                    | 19.952                           | 0.030  |
|                                |     |                  |                                  |                           |                                    |                              |                          | Average=                              | <b>19.93</b>                                  |                                    |                      | Average=                         | <b>19.95</b>                                   |
| <b>Change Flow Rate</b>        |     |                  | <b>Stabilization 3 minutes</b>   |                           |                                    |                              |                          |                                       |   |                                    |                      |                                  |  |
| 80                             | 1   | 30.6.2017        | 21.88                            | 99.33                     | 43                                 | 22.56                        | 99.5                     | 80.442                                | 0.0064  | 10                                 | 1                    | 80.350                           | 0.121  |
| 80                             | 2   | 30.6.2017        | 21.88                            | 99.33                     | 43                                 | 22.58                        | 99.49                    | 80.442                                | 0.0075  | 10                                 | 2                    | 80.350                           | 0.121  |
| 80                             | 3   | 30.6.2017        | 21.88                            | 99.33                     | 43                                 | 22.58                        | 99.49                    | 80.440                                | 0.0092  | 10                                 | 3                    | 80.350                           | 0.121  |
| 80                             | 4   | 30.6.2017        | 21.88                            | 99.33                     | 43                                 | 22.58                        | 99.49                    | 80.445                                | 0.0068  | 10                                 | 4                    | 80.349                           | 0.121  |
|                                |     |                  |                                  |                           |                                    |                              |                          | Average=                              | <b>80.44</b>                                  |                                    |                      | Average=                         | <b>80.35</b>                                   |
| <b>Change Flow Rate</b>        |     |                  | <b>Stabilization 3 minutes</b>   |                           |                                    |                              |                          |                                       |   |                                    |                      |                                  |  |
| 50                             | 1   | 3.7.2017         | 21.85                            | 100.80                    | 48                                 | 22.66                        | 100.99                   | 49.719                                | 0.0066  | 10                                 | 1                    | 49.753                           | 0.075  |
| 50                             | 2   | 3.7.2017         | 21.85                            | 100.80                    | 48                                 | 22.66                        | 100.99                   | 49.717                                | 0.0037  | 10                                 | 2                    | 49.752                           | 0.075  |
| 50                             | 3   | 3.7.2017         | 21.85                            | 100.80                    | 48                                 | 22.68                        | 101                      | 49.715                                | 0.0045  | 10                                 | 3                    | 49.752                           | 0.075  |
| 50                             | 4   | 3.7.2017         | 21.85                            | 100.80                    | 48                                 | 22.7                         | 101                      | 49.711                                | 0.0039  | 10                                 | 4                    | 49.752                           | 0.075  |
|                                |     |                  |                                  |                           |                                    |                              |                          | Average=                              | <b>49.72</b>                                  |                                    |                      | Average=                         | <b>49.75</b>                                   |

**INITIAL TESTS**

Compare BIOS temperature and pressure readings with calibrated temperature and ambient pressure instrument readings.

| Laboratory temperature<br>(°C)                 | Laboratory pressure<br>(kPa) | BIOS temperature<br>(°C) | BIOS pressure<br>(kPa) |
|--|------------------------------|--------------------------|------------------------|
| Stabilization 24 hours in laboratory (no flow) |                              |                          |                        |
| 22.02  | 1010.3                       | 21.97                    | 1010.5                 |

**COMPARISON MEASUREMENTS (LOOP2)**

Base: (S/N: 135516)

Flow cell: -24 (S/N: 134910)

|                                   |     |                     | ENVIRONMENTAL PARAMETERS         |                              |                                       | TRANSFER STANDARD PARAMETERS  |                             |  |  |                                    | REFERENCE PARAMETERS |                                     |   |
|-----------------------------------|-----|---------------------|----------------------------------|------------------------------|---------------------------------------|-------------------------------|-----------------------------|--|--|------------------------------------|----------------------|-------------------------------------|---|
| Nominal Flow Rate<br>(std ml/min) | RUN | Date<br>(dd.mm.yyy) | Laboratory temperature<br>(°C)   | Laboratory pressure<br>(kPa) | Laboratory relative humidity<br>(%rh) | BIOS mean temperature<br>(°C) | BIOS mean pressure<br>(kPa) | Mean Indicated Flow Rate<br>(std ml/min) | Standard deviation of indication<br>(std ml/min) | Number of flow indications in mean | RUN                  | Reference Flow Rate<br>(std ml/min) | Standard uncertainty of reference<br>(std ml/min) |
| <b>Set Flow Rate</b>              |     |                     | <b>Stabilization 180 minutes</b> |                              |                                       |                               |                             |  |  |                                    |                      |                                     |   |
| 80                                | 1   | 30.6.2017           | 21.50                            | 99.31                        | 43                                    | 22.32                         | 99.47                       | 80.231                                   | 0.0130   | 10                                 | 1                    | 80.380                              | 0.12  |
| 80                                | 2   | 30.6.2017           | 21.56                            | 99.30                        | 43                                    | 22.34                         | 99.46                       | 80.225                                   | 0.0100   | 10                                 | 2                    | 80.377                              | 0.12  |
| 80                                | 3   | 30.6.2017           | 21.60                            | 99.30                        | 43                                    | 22.38                         | 99.47                       | 80.205                                   | 0.0180   | 10                                 | 3                    | 80.370                              | 0.12  |
| 80                                | 4   | 30.6.2017           | 21.62                            | 99.30                        | 43                                    | 22.44                         | 99.47                       | 80.187                                   | 0.0190   | 10                                 | 4                    | 80.363                              | 0.12  |
|                                   |     |                     |                                  |                              |                                       |                               |                             | Average=                                 | <b>80.21</b>                                     |                                    |                      | Average=                            | <b>80.37</b>                                      |
| <b>Change Flow Rate</b>           |     |                     | <b>Stabilization 3 minutes</b>   |                              |                                       |                               |                             |  |  |                                    |                      |                                     |   |
| 300                               | 1   | 22.5.2017           | 20.85                            | 100.94                       | 45                                    | 21.81                         | 101.09                      | 302.880                                  |  | 10                                 | 1                    | 303.130                             | 0.45  |
| 300                               | 2   | 22.5.2017           | 20.85                            | 100.94                       | 45                                    | 21.83                         | 101.09                      | 302.880                                  |  | 10                                 | 2                    | 303.120                             | 0.45  |
| 300                               | 3   | 22.5.2017           | 20.87                            | 100.94                       | 45                                    | 21.85                         | 101.09                      | 302.880                                  |  | 10                                 | 3                    | 303.099                             | 0.45  |
| 300                               | 4   | 22.5.2017           | 20.88                            | 100.94                       | 45                                    | 21.85                         | 101.09                      | 302.890                                  |  | 10                                 | 4                    | 303.099                             | 0.45  |
|                                   |     |                     |                                  |                              |                                       |                               |                             | Average=                                 | <b>302.88</b>                                    |                                    |                      | Average=                            | <b>303.11</b>                                     |
| <b>Change Flow Rate</b>           |     |                     | <b>Stabilization 3 minutes</b>   |                              |                                       |                               |                             |  |  |                                    |                      |                                     |   |
| 600                               | 1   | 22.5.2017           | 20.88                            | 100.94                       | 45                                    | 21.95                         | 101.08                      | 601.810                                  |  | 10                                 | 1                    | 601.838                             | 0.90  |
| 600                               | 2   | 22.5.2017           | 20.88                            | 100.94                       | 45                                    | 21.93                         | 101.08                      | 601.780                                  |  | 10                                 | 2                    | 601.859                             | 0.90  |
| 600                               | 3   | 22.5.2017           | 20.86                            | 100.94                       | 45                                    | 21.93                         | 101.08                      | 601.790                                  |  | 10                                 | 3                    | 601.880                             | 0.90  |
| 600                               | 4   | 22.5.2017           | 20.88                            | 100.94                       | 45                                    | 21.93                         | 101.08                      | 601.760                                  |  | 10                                 | 4                    | 601.880                             | 0.90  |
|                                   |     |                     |                                  |                              |                                       |                               |                             | Average=                                 | <b>601.79</b>                                    |                                    |                      | Average=                            | <b>601.86</b>                                     |
| <b>Change Flow Rate</b>           |     |                     | <b>Stabilization 3 minutes</b>   |                              |                                       |                               |                             |  |  |                                    |                      |                                     |   |
| 1250                              | 1   | 22.5.2017           | 21.08                            | 100.89                       | 42                                    | 21.97                         | 101.07                      | 1253.500                                 |  | 10                                 | 1                    | 1255.218                            | 1.88  |
| 1250                              | 2   | 22.5.2017           | 21.08                            | 100.89                       | 42                                    | 21.99                         | 101.07                      | 1253.400                                 |  | 10                                 | 2                    | 1254.272                            | 1.88  |
| 1250                              | 3   | 22.5.2017           | 21.11                            | 100.89                       | 42                                    | 21.99                         | 101.07                      | 1253.500                                 |  | 10                                 | 3                    | 1254.315                            | 1.88  |
| 1250                              | 4   | 22.5.2017           | 21.11                            | 100.89                       | 42                                    | 21.99                         | 101.07                      | 1253.400                                 |  | 10                                 | 4                    | 1254.337                            | 1.88  |
|                                   |     |                     |                                  |                              |                                       |                               |                             | Average=                                 | <b>1253.45</b>                                   |                                    |                      | Average=                            | <b>1254.54</b>                                    |



**INITIAL TESTS**

Compare BIOS temperature and pressure readings with calibrated temperature and ambient pressure instrument readings.

|  |                           |                       |                     |
|--|---------------------------|-----------------------|---------------------|
| Laboratory temperature (°C)                    | Laboratory pressure (kPa) | BIOS temperature (°C) | BIOS pressure (kPa) |
| Stabilization 24 hours in laboratory (no flow) |                           |                       |                     |
|  |                           |                       |                     |

**COMPARISON MEASUREMENTS (LOOP2)**

Base: (S/N: 135516)

Flow cell: -44 (S/N: 135199)

|                                |     |                  | ENVIRONMENTAL PARAMETERS         |                           |                                    | TRANSFER STANDARD PARAMETERS |                          |                                       |   |                                    | REFERENCE PARAMETERS |                                  |  |
|--------------------------------|-----|------------------|----------------------------------|---------------------------|------------------------------------|------------------------------|--------------------------|---------------------------------------|---|------------------------------------|----------------------|----------------------------------|--|
| Nominal Flow Rate (std ml/min) | RUN | Date (dd.mm.yyy) | Laboratory temperature (°C)      | Laboratory pressure (kPa) | Laboratory relative humidity (%rh) | BIOS mean temperature (°C)   | BIOS mean pressure (kPa) | Mean Indicated Flow Rate (std ml/min) | Standard deviation of indication (std ml/min) | Number of flow indications in mean | RUN                  | Reference Flow Rate (std ml/min) | Standard uncertainty of reference (std ml/min) |
| <b>Set Flow Rate</b>           |     |                  | <b>Stabilization 180 minutes</b> |                           |                                    |                              |                          |                                       |   |                                    |                      |                                  |  |
| 1250                           | 1   | 11.5.2017        | 21.6                             | 99.19                     | 37.36                              | 22.57                        | 99.44                    | 1250.39                               | 0.12  | 10                                 | 1                    | 1249.747                         | 1.87   |
| 1250                           | 2   | 11.5.2017        | 21.6                             | 99.18                     | 37.31                              | 22.56                        | 99.43                    | 1250.42                               | 0.13  | 10                                 | 2                    | 1249.689                         | 1.87   |
| 1250                           | 3   | 11.5.2017        | 21.6                             | 99.19                     | 37.30                              | 22.55                        | 99.43                    | 1250.34                               | 0.15  | 10                                 | 3                    | 1249.689                         | 1.87   |
| 1250                           | 4   | 11.5.2017        | 21.6                             | 99.20                     | 37.42                              | 22.56                        | 99.43                    | 1250.28                               | 0.21  | 10                                 | 4                    | 1249.655                         | 1.87   |
|                                |     |                  |                                  |                           |                                    |                              |                          | Average=                              | <b>1250.36</b>                                |                                    | Average=             | <b>1249.70</b>                   |  |
| <b>Change Flow Rate</b>        |     |                  | <b>Stabilization 3 minutes</b>   |                           |                                    |                              |                          |                                       |   |                                    |                      |                                  |  |
| 5000                           | 1   | 12.5.2017        | 21.7                             | 99.33                     | 40.95                              | 22.46                        | 99.63                    | 4999.70                               | 0.62  | 10                                 | 1                    | 4996.92                          | 7.50   |
| 5000                           | 2   | 12.5.2017        | 21.7                             | 99.33                     | 41.03                              | 22.42                        | 99.63                    | 5000.40                               | 0.11  | 10                                 | 2                    | 4997.15                          | 7.50   |
| 5000                           | 3   | 12.5.2017        | 21.7                             | 99.33                     | 41.11                              | 22.42                        | 99.63                    | 5000.37                               | 0.07  | 10                                 | 3                    | 4997.20                          | 7.50   |
| 5000                           | 4   | 12.5.2017        | 21.7                             | 99.33                     | 41.09                              | 22.42                        | 99.63                    | 5000.26                               | 0.45  | 10                                 | 4                    | 4997.08                          | 7.50   |
|                                |     |                  |                                  |                           |                                    |                              |                          | Average=                              | <b>5000.18</b>                                |                                    | Average=             | <b>4997.08</b>                   |  |
| <b>Change Flow Rate</b>        |     |                  | <b>Stabilization 3 minutes</b>   |                           |                                    |                              |                          |                                       |   |                                    |                      |                                  |  |
| 10000                          | 1   | 12.5.2017        | 21.5                             | 99.32                     | 41.63                              | 22.39                        | 99.71                    | 9990.61                               | 2.43  | 10                                 | 1                    | 9990.07                          | 14.99  |
| 10000                          | 2   | 12.5.2017        | 21.6                             | 99.33                     | 41.60                              | 22.33                        | 99.71                    | 9991.54                               | 1.31  | 10                                 | 2                    | 9990.28                          | 14.99  |
| 10000                          | 3   | 12.5.2017        | 21.6                             | 99.33                     | 41.60                              | 22.33                        | 99.70                    | 9991.96                               | 1.13  | 10                                 | 3                    | 9989.94                          | 14.98  |
| 10000                          | 4   | 12.5.2017        | 21.6                             | 99.33                     | 41.60                              | 22.32                        | 99.71                    | 9990.96                               | 0.45  | 10                                 | 4                    | 9989.42                          | 14.98  |
|                                |     |                  |                                  |                           |                                    |                              |                          | Average=                              | <b>9991.27</b>                                |                                    | Average=             | <b>9989.93</b>                   |  |
| <b>Change Flow Rate</b>        |     |                  | <b>Stabilization 3 minutes</b>   |                           |                                    |                              |                          |                                       |   |                                    |                      |                                  |  |
| 20000                          | 1   | 15.5.2017        | 21.5                             | 101.95                    | 44.10                              | 21.95                        | 102.56                   | 19984.90                              | 3.41  | 10                                 | 1                    | 19988.02                         | 29.98  |
| 20000                          | 2   | 15.5.2017        | 21.5                             | 101.95                    | 44.42                              | 21.93                        | 102.56                   | 19988.30                              | 4.08  | 10                                 | 2                    | 19987.17                         | 29.98  |
| 20000                          | 3   | 15.5.2017        | 21.5                             | 101.95                    | 44.57                              | 21.93                        | 102.55                   | 19985.20                              | 6.53  | 10                                 | 3                    | 19985.68                         | 29.98  |
| 20000                          | 4   | 15.5.2017        | 21.5                             | 101.95                    | 44.55                              | 21.93                        | 102.56                   | 19983.00                              | 4.64  | 10                                 | 4                    | 19984.34                         | 29.98  |
|                                |     |                  |                                  |                           |                                    |                              |                          | Average=                              | <b>19985.35</b>                               |                                    | Average=             | <b>19986.30</b>                  |  |
| <b>Change Flow Rate</b>        |     |                  | <b>Stabilization 3 minutes</b>   |                           |                                    |                              |                          |                                       |   |                                    |                      |                                  |  |
| 30000                          | 1   | 24.5.2017        | 21.7                             | 101.92                    | 41.52                              | 22.07                        | 102.62                   | 29993.20                              | 3.26  | 10                                 | 1                    | 29990.43                         | 44.99  |
| 30000                          | 2   | 24.5.2017        | 21.7                             | 101.93                    | 41.38                              | 22.04                        | 102.62                   | 29986.70                              | 12.06   | 10                                 | 2                    | 29989.86                         | 44.98  |
| 30000                          | 3   | 24.5.2017        | 21.7                             | 101.92                    | 41.34                              | 22.03                        | 102.62                   | 29991.00                              | 11.77   | 10                                 | 3                    | 29988.85                         | 44.98  |
| 30000                          | 4   | 24.5.2017        | 21.7                             | 101.93                    | 41.26                              | 22.03                        | 102.62                   | 29984.10                              | 10.55   | 10                                 | 4                    | 29985.88                         | 44.98  |
|                                |     |                  |                                  |                           |                                    |                              |                          | Average=                              | <b>29988.75</b>                               |                                    | Average=             | <b>29988.75</b>                  |  |
| <b>End of measurements</b>     |     |                  |                                  |                           |                                    |                              |                          |                                       |   |                                    |                      |                                  |  |

**MIKES results (loop1)**

**INITIAL TESTS**

Compare BIOS temperature and pressure readings with calibrated temperature and ambient pressure instrument readings.

| Laboratory temperature<br>(°C)                 | Laboratory pressure<br>(kPa) | BIOS temperature<br>(°C) | BIOS pressure<br>(kPa) |
|--|------------------------------|--------------------------|------------------------|
| Stabilization 24 hours in laboratory (no flow) |                              |                          |                        |
| 23.7   | 102.00                       | 24.3                     | 101.98                 |

**COMPARISON MEASUREMENTS**

**Flow cell: -10 (S/N: 135207)**

|                                   |     |                     | ENVIRONMENTAL PARAMETERS         |                              |                                       | TRANSFER STANDARD PARAMETERS  |                             |  |  |                                    | REFERENCE PARAMETERS |                                     |   |
|-----------------------------------|-----|---------------------|----------------------------------|------------------------------|---------------------------------------|-------------------------------|-----------------------------|--|--|------------------------------------|----------------------|-------------------------------------|---|
| Nominal Flow Rate<br>(std ml/min) | RUN | Date<br>(dd.mm.yyy) | Laboratory temperature<br>(°C)   | Laboratory pressure<br>(kPa) | Laboratory relative humidity<br>(%rh) | BIOS mean temperature<br>(°C) | BIOS mean pressure<br>(kPa) | Mean Indicated Flow Rate<br>(std ml/min) | Standard deviation of indication<br>(std ml/min) | Number of flow indications in mean | RUN                  | Reference Flow Rate<br>(std ml/min) | Standard uncertainty of reference<br>(std ml/min) |
| <b>Set Flow Rate</b>              |     |                     | <b>Stabilization 180 minutes</b> |                              |                                       |                               |                             |  |  |                                    |                      |                                     |   |
| 5                                 | 1   | 28.9.2016           | 23.6                             | 101.1                        | 35.9                                  | 24.40                         | 101.1                       | 4.981                                    | 0.002  | 37                                 | 1                    | 4.982                               | 0.004   |
| 5                                 | 2   | 28.9.2016           | 23.6                             | 100.9                        | 35.6                                  | 24.41                         | 100.9                       | 4.978                                    | 0.002  | 37                                 | 2                    | 4.989                               | 0.004   |
| 5                                 | 3   | 28.9.2016           | 23.6                             | 100.7                        | 36.0                                  | 24.39                         | 100.7                       | 4.975                                    | 0.001  | 37                                 | 3                    | 4.981                               | 0.004   |
| 5                                 | 4   | 29.9.2016           | 23.6                             | 99.8                         | 35.4                                  | 24.47                         | 99.8                        | 4.971                                    | 0.001  | 38                                 | 4                    | 4.982                               | 0.004   |
|                                   |     |                     |                                  |                              |                                       |                               |                             | Average=                                 | <b>4.976</b>                                     |                                    | Average=             | <b>4.984</b>                        |   |
| <b>Change Flow Rate</b>           |     |                     | <b>Stabilization 3 minutes</b>   |                              |                                       |                               |                             |  |  |                                    |                      |                                     |   |
| 10                                | 1   | 28.9.2016           | 23.6                             | 100.0                        | 36.3                                  | 24.36                         | 100.1                       | 9.984                                    | 0.001  | 37                                 | 1                    | 9.982                               | 0.007   |
| 10                                | 2   | 28.9.2016           | 23.6                             | 100.0                        | 35.4                                  | 24.36                         | 100.1                       | 9.984                                    | 0.001  | 38                                 | 2                    | 9.994                               | 0.007   |
| 10                                | 3   | 28.9.2016           | 23.6                             | 99.9                         | 35.5                                  | 24.36                         | 100.0                       | 9.984                                    | 0.001  | 38                                 | 3                    | 9.980                               | 0.007   |
| 10                                | 4   | 30.9.2016           | 23.6                             | 98.7                         | 35.4                                  | 24.47                         | 98.8                        | 9.986                                    | 0.002  | 38                                 | 4                    | 9.994                               | 0.007   |
|                                   |     |                     |                                  |                              |                                       |                               |                             | Average=                                 | <b>9.984</b>                                     |                                    | Average=             | <b>9.987</b>                        |   |
| <b>Change Flow Rate</b>           |     |                     | <b>Stabilization 3 minutes</b>   |                              |                                       |                               |                             |  |  |                                    |                      |                                     |   |
| 20                                | 1   | 28.9.2016           | 23.6                             | 99.9                         | 35.4                                  | 24.37                         | 100.0                       | 19.954                                   | 0.002  | 36                                 | 1                    | 19.984                              | 0.015   |
| 20                                | 2   | 28.9.2016           | 23.6                             | 99.9                         | 35.3                                  | 24.38                         | 100.0                       | 19.954                                   | 0.001  | 36                                 | 2                    | 19.987                              | 0.015   |
| 20                                | 3   | 28.9.2016           | 23.6                             | 99.9                         | 35.5                                  | 24.38                         | 100.0                       | 19.954                                   | 0.001  | 36                                 | 3                    | 19.949                              | 0.015   |
| 20                                | 4   | 28.9.2016           | 23.6                             | 99.9                         | 35.3                                  | 24.39                         | 100.0                       | 19.954                                   | 0.002  | 36                                 | 4                    | 19.980                              | 0.015   |
|                                   |     |                     |                                  |                              |                                       |                               |                             | Average=                                 | <b>19.954</b>                                    |                                    | Average=             | <b>19.975</b>                       |   |
| <b>Change Flow Rate</b>           |     |                     | <b>Stabilization 3 minutes</b>   |                              |                                       |                               |                             |  |  |                                    |                      |                                     |   |
| 80                                | 1   | 30.9.2016           | 23.6                             | 98.9                         | 34.8                                  | 24.59                         | 99.0                        | 78.693                                   | 0.046  | 81                                 | 1                    | 78.678                              | 0.057   |
| 80                                | 2   | 30.9.2016           | 23.6                             | 99.0                         | 34.6                                  | 24.59                         | 99.1                        | 78.615                                   | 0.031  | 81                                 | 2                    | 78.663                              | 0.057   |
| 80                                | 3   | 30.9.2016           | 23.6                             | 99.0                         | 34.7                                  | 24.59                         | 99.1                        | 79.995                                   | 0.027  | 81                                 | 3                    | 80.004                              | 0.058   |
| 80                                | 4   | 30.9.2016           | 23.6                             | 99.1                         | 34.8                                  | 24.59                         | 99.2                        | 80.002                                   | 0.045  | 81                                 | 4                    | 80.074                              | 0.058   |
|                                   |     |                     |                                  |                              |                                       |                               |                             | Average=                                 | <b>79.326</b>                                    |                                    | Average=             | <b>79.355</b>                       |   |

**INITIAL TESTS**

Compare BIOS temperature and pressure readings with calibrated temperature and ambient pressure instrument readings.

| Laboratory temperature<br>(°C)                 | Laboratory pressure<br>(kPa) | BIOS temperature<br>(°C) | BIOS pressure<br>(kPa) |
|--|------------------------------|--------------------------|------------------------|
| Stabilization 24 hours in laboratory (no flow) |                              |                          |                        |
| 23.6   | 101.48                       | 23.7                     | 101.46                 |

**COMPARISON MEASUREMENTS**

Flow cell: -24 (S/N: 134909)

| ENVIRONMENTAL PARAMETERS          |     |                     | TRANSFER STANDARD PARAMETERS     |                              |                                       |                               |                             |  |  |                                    | REFERENCE PARAMETERS    |                                     |   |
|-----------------------------------|-----|---------------------|----------------------------------|------------------------------|---------------------------------------|-------------------------------|-----------------------------|--|--|------------------------------------|-------------------------|-------------------------------------|---|
| Nominal Flow Rate<br>(std ml/min) | RUN | Date<br>(dd.mm.yyy) | Laboratory temperature<br>(°C)   | Laboratory pressure<br>(kPa) | Laboratory relative humidity<br>(%rh) | BIOS mean temperature<br>(°C) | BIOS mean pressure<br>(kPa) | Mean Indicated Flow Rate<br>(std ml/min) | Standard deviation of indication<br>(std ml/min) | Number of flow indications in mean | RUN                     | Reference Flow Rate<br>(std ml/min) | Standard uncertainty of reference<br>(std ml/min) |
| <b>Set Flow Rate</b>              |     |                     | <b>Stabilization 180 minutes</b> |                              |                                       |                               |                             |  |  |                                    |                         |                                     |   |
| 80                                | 1   | 20.9.2016           | 23.6                             | 101.8                        | 34.7                                  | 24.51                         | 101.9                       | 79.49                                    | 0.04   | 13                                 | 1                       | 79.95                               | 0.06  |
| 80                                | 2   | 20.9.2016           | 23.6                             | 101.9                        | 34.6                                  | 24.53                         | 101.9                       | 79.42                                    | 0.01   | 13                                 | 2                       | 79.89                               | 0.06  |
| 80                                | 3   | 20.9.2016           | 23.6                             | 101.9                        | 34.6                                  | 24.53                         | 101.9                       | 79.40                                    | 0.03   | 13                                 | 3                       | 79.86                               | 0.06  |
| 80                                | 4   | 20.9.2016           | 23.6                             | 101.9                        | 34.7                                  | 24.55                         | 101.9                       | 79.43                                    | 0.03   | 13                                 | 4                       | 79.92                               | 0.06  |
| Average=                          |     |                     |                                  |                              |                                       |                               |                             | <b>79.44</b>                             |  |                                    | Average= <b>79.90</b>   |                                     |   |
| <b>Change Flow Rate</b>           |     |                     | <b>Stabilization 3 minutes</b>   |                              |                                       |                               |                             |  |  |                                    |                         |                                     |   |
| 300                               | 1   | 20.9.2016           | 23.6                             | 101.8                        | 34.7                                  | 24.57                         | 101.9                       | 299.97                                   | 0.02   | 33                                 | 1                       | 301.30                              | 0.22  |
| 300                               | 2   | 20.9.2016           | 23.6                             | 101.8                        | 34.5                                  | 24.58                         | 101.9                       | 299.97                                   | 0.02   | 33                                 | 2                       | 301.47                              | 0.22  |
| 300                               | 3   | 20.9.2016           | 23.6                             | 101.8                        | 34.4                                  | 24.58                         | 101.9                       | 299.97                                   | 0.02   | 33                                 | 3                       | 301.49                              | 0.22  |
| 300                               | 4   | 20.9.2016           | 23.6                             | 101.8                        | 34.5                                  | 24.59                         | 101.9                       | 299.97                                   | 0.02   | 33                                 | 4                       | 301.31                              | 0.22  |
| Average=                          |     |                     |                                  |                              |                                       |                               |                             | <b>299.97</b>                            |  |                                    | Average= <b>301.39</b>  |                                     |   |
| <b>Change Flow Rate</b>           |     |                     | <b>Stabilization 3 minutes</b>   |                              |                                       |                               |                             |  |  |                                    |                         |                                     |   |
| 600                               | 1   | 15.9.2016           | 23.6                             | 101.3                        | 35.2                                  | 24.56                         | 101.4                       | 599.85                                   | 0.06   | 50                                 | 1                       | 602.78                              | 0.44  |
| 600                               | 2   | 15.9.2016           | 23.6                             | 101.3                        | 35.3                                  | 24.57                         | 101.4                       | 599.81                                   | 0.04   | 50                                 | 2                       | 602.81                              | 0.44  |
| 600                               | 3   | 15.9.2016           | 23.6                             | 101.3                        | 35.5                                  | 24.57                         | 101.4                       | 599.78                                   | 0.05   | 50                                 | 3                       | 602.82                              | 0.44  |
| 600                               | 4   | 15.9.2016           | 23.6                             | 101.3                        | 35.4                                  | 24.57                         | 101.4                       | 599.77                                   | 0.05   | 50                                 | 4                       | 602.65                              | 0.44  |
| Average=                          |     |                     |                                  |                              |                                       |                               |                             | <b>599.80</b>                            |  |                                    | Average= <b>602.77</b>  |                                     |   |
| <b>Change Flow Rate</b>           |     |                     | <b>Stabilization 3 minutes</b>   |                              |                                       |                               |                             |  |  |                                    |                         |                                     |   |
| 1250                              | 1   | 15.9.2016           | 23.6                             | 101.3                        | 35.6                                  | 24.57                         | 101.4                       | 1249.88                                  | 0.19   | 72                                 | 1                       | 1256.63                             | 0.92  |
| 1250                              | 2   | 15.9.2016           | 23.6                             | 101.3                        | 35.3                                  | 24.57                         | 101.4                       | 1249.77                                  | 0.17   | 72                                 | 2                       | 1256.42                             | 0.92  |
| 1250                              | 3   | 15.9.2016           | 23.6                             | 101.3                        | 35.3                                  | 24.57                         | 101.4                       | 1249.66                                  | 0.24   | 72                                 | 3                       | 1256.35                             | 0.92  |
| 1250                              | 4   | 15.9.2016           | 23.6                             | 101.7                        | 34.8                                  | 24.53                         | 101.8                       | 1249.46                                  | 0.22   | 72                                 | 4                       | 1255.95                             | 0.92  |
| Average=                          |     |                     |                                  |                              |                                       |                               |                             | <b>1249.69</b>                           |  |                                    | Average= <b>1256.34</b> |                                     |   |

**INITIAL TESTS**

Compare BIOS temperature and pressure readings with calibrated temperature and ambient pressure instrument readings.

| Laboratory temperature (°C)                    | Laboratory pressure (kPa) | BIOS temperature (°C) | BIOS pressure (kPa) |
|--|---------------------------|-----------------------|---------------------|
| Stabilization 24 hours in laboratory (no flow) |                           |                       |                     |
| 23.6   | 100.79                    | 23.6                  | 100.76              |

**COMPARISON MEASUREMENTS**

Flow cell: -44 (S/N: 135198)

|                                |     |                  | ENVIRONMENTAL PARAMETERS         |                           |                                    | TRANSFER STANDARD PARAMETERS |                          |                                       |   |                                    | REFERENCE PARAMETERS |                                  |  |
|--------------------------------|-----|------------------|----------------------------------|---------------------------|------------------------------------|------------------------------|--------------------------|---------------------------------------|---|------------------------------------|----------------------|----------------------------------|--|
| Nominal Flow Rate (std ml/min) | RUN | Date (dd.mm.yyy) | Laboratory temperature (°C)      | Laboratory pressure (kPa) | Laboratory relative humidity (%rh) | BIOS mean temperature (°C)   | BIOS mean pressure (kPa) | Mean Indicated Flow Rate (std ml/min) | Standard deviation of indication (std ml/min) | Number of flow indications in mean | RUN                  | Reference Flow Rate (std ml/min) | Standard uncertainty of reference (std ml/min) |
| <b>Set Flow Rate</b>           |     |                  | <b>Stabilization 180 minutes</b> |                           |                                    |                              |                          |                                       |   |                                    |                      |                                  |  |
| 1250                           | 1   | 9.9.2016         | 24.0                             | 101.5                     | 35.2                               | 24.49                        | 101.7                    | 1250.56                               | 0.18  | 15                                 | 1                    | 1251.35                          | 0.91   |
| 1250                           | 2   | 9.9.2016         | 24.0                             | 101.5                     | 35.0                               | 24.50                        | 101.7                    | 1250.45                               | 0.10  | 15                                 | 2                    | 1250.94                          | 0.91   |
| 1250                           | 3   | 9.9.2016         | 24.0                             | 101.5                     | 35.0                               | 24.51                        | 101.7                    | 1250.54                               | 0.11  | 15                                 | 3                    | 1251.52                          | 0.91   |
| 1250                           | 4   | 9.9.2016         | 24.0                             | 101.5                     | 34.8                               | 24.51                        | 101.7                    | 1250.55                               | 0.10  | 15                                 | 4                    | 1251.55                          | 0.91   |
|                                |     |                  |                                  |                           |                                    |                              |                          | Average=                              | <b>1250.53</b>                                |                                    |                      | Average=                         | <b>1251.34</b>                                 |
| <b>Change Flow Rate</b>        |     |                  | <b>Stabilization 3 minutes</b>   |                           |                                    |                              |                          |                                       |   |                                    |                      |                                  |  |
| 5000                           | 1   | 9.9.2016         | 24.0                             | 101.5                     | 34.8                               | 24.41                        | 101.7                    | 4983.95                               | 3.20  | 24                                 | 1                    | 4984.00                          | 3.65   |
| 5000                           | 2   | 9.9.2016         | 24.0                             | 101.5                     | 34.7                               | 24.41                        | 101.7                    | 4984.10                               | 4.79  | 24                                 | 2                    | 4987.45                          | 3.65   |
| 5000                           | 3   | 9.9.2016         | 24.0                             | 101.5                     | 34.7                               | 24.42                        | 101.7                    | 4990.03                               | 1.52  | 24                                 | 3                    | 4990.20                          | 3.66   |
| 5000                           | 4   | 9.9.2016         | 24.0                             | 101.5                     | 34.5                               | 24.42                        | 101.7                    | 4993.18                               | 1.91  | 24                                 | 4                    | 4992.08                          | 3.66   |
|                                |     |                  |                                  |                           |                                    |                              |                          | Average=                              | <b>4987.81</b>                                |                                    |                      | Average=                         | <b>4988.43</b>                                 |
| <b>Change Flow Rate</b>        |     |                  | <b>Stabilization 3 minutes</b>   |                           |                                    |                              |                          |                                       |   |                                    |                      |                                  |  |
| 10000                          | 1   | 9.9.2016         | 24.0                             | 101.5                     | 34.5                               | 24.37                        | 101.7                    | 9984.00                               | 2.17  | 33                                 | 1                    | 9989.53                          | 7.31   |
| 10000                          | 2   | 9.9.2016         | 24.0                             | 101.5                     | 34.5                               | 24.37                        | 101.7                    | 9987.45                               | 2.87  | 33                                 | 2                    | 9990.62                          | 7.31   |
| 10000                          | 3   | 9.9.2016         | 24.0                             | 101.5                     | 34.4                               | 24.37                        | 101.7                    | 9987.12                               | 2.29  | 33                                 | 3                    | 9987.98                          | 7.31   |
| 10000                          | 4   | 9.9.2016         | 24.0                             | 101.5                     | 34.4                               | 24.37                        | 101.7                    | 9987.96                               | 2.70  | 33                                 | 4                    | 9984.55                          | 7.31   |
|                                |     |                  |                                  |                           |                                    |                              |                          | Average=                              | <b>9986.63</b>                                |                                    |                      | Average=                         | <b>9988.17</b>                                 |
| <b>Change Flow Rate</b>        |     |                  | <b>Stabilization 3 minutes</b>   |                           |                                    |                              |                          |                                       |   |                                    |                      |                                  |  |
| 20000                          | 1   | 9.9.2016         | 24.0                             | 101.5                     | 34.3                               | 24.44                        | 101.8                    | 20032.04                              | 4.95  | 24                                 | 1                    | 20014.44                         | 14.77  |
| 20000                          | 2   | 9.9.2016         | 24.0                             | 101.5                     | 34.3                               | 24.45                        | 101.8                    | 20022.63                              | 6.03  | 24                                 | 2                    | 20045.23                         | 14.74  |
| 20000                          | 3   | 9.9.2016         | 24.0                             | 101.5                     | 34.2                               | 24.45                        | 101.8                    | 20021.04                              | 5.73  | 24                                 | 3                    | 20035.48                         | 14.73  |
| 20000                          | 4   | 9.9.2016         | 24.0                             | 101.5                     | 34.3                               | 24.44                        | 101.8                    | 20019.46                              | 6.34  | 24                                 | 4                    | 20029.16                         | 14.73  |
|                                |     |                  |                                  |                           |                                    |                              |                          | Average=                              | <b>20023.79</b>                               |                                    |                      | Average=                         | <b>20031.08</b>                                |
| <b>Change Flow Rate</b>        |     |                  | <b>Stabilization 3 minutes</b>   |                           |                                    |                              |                          |                                       |   |                                    |                      |                                  |  |
| 30000                          | 1   | 9.9.2016         | 23.7                             | 101.5                     | 34.8                               | 24.34                        | 101.9                    | 30135.88                              | 28.40   | 24                                 | 1                    | 30147.04                         | 22.17  |
| 30000                          | 2   | 9.9.2016         | 23.7                             | 101.5                     | 34.7                               | 24.36                        | 101.9                    | 30079.38                              | 53.25   | 24                                 | 2                    | 30059.59                         | 22.10  |
| 30000                          | 3   | 9.9.2016         | 23.7                             | 101.5                     | 34.9                               | 24.35                        | 102.0                    | 30009.46                              | 17.74   | 24                                 | 3                    | 30042.36                         | 22.07  |
| 30000                          | 4   | 9.9.2016         | 23.7                             | 101.5                     | 34.8                               | 24.36                        | 102.0                    | 30059.00                              | 14.96   | 24                                 | 4                    | 30065.32                         | 22.09  |
|                                |     |                  |                                  |                           |                                    |                              |                          | Average=                              | <b>30070.93</b>                               |                                    |                      | Average=                         | <b>30078.58</b>                                |
| <b>End of measurements</b>     |     |                  |                                  |                           |                                    |                              |                          | Average=                              | <b>30070.93</b>                               |                                    |                      | Average=                         | <b>30078.58</b>                                |

## FORCE results (loop1)

### INITIAL TESTS

Compare BIOS temperature and pressure readings with calibrated temperature and ambient pressure instrument readings.

| Laboratory temperature<br>(°C)                 | Laboratory pressure<br>(kPa) | BIOS temperature<br>(°C) | BIOS pressure<br>(kPa) |
|--|------------------------------|--------------------------|------------------------|
| Stabilization 24 hours in laboratory (no flow) |                              |                          |                        |
| 20.29  | 101.143                      | 20.6                     | 102.15                 |

### COMPARISON MEASUREMENTS

Flow cell: -24 (S/N: 134909)

| ENVIRONMENTAL PARAMETERS          |     |                     | TRANSFER STANDARD PARAMETERS     |                              |                                       |                               |                             |  |  |                                    | REFERENCE PARAMETERS |                                     |   |
|-----------------------------------|-----|---------------------|----------------------------------|------------------------------|---------------------------------------|-------------------------------|-----------------------------|--|--|------------------------------------|----------------------|-------------------------------------|---|
| Nominal Flow Rate<br>(std ml/min) | RUN | Date<br>(dd.mm.yyy) | Laboratory temperature<br>(°C)   | Laboratory pressure<br>(kPa) | Laboratory relative humidity<br>(%rh) | BIOS mean temperature<br>(°C) | BIOS mean pressure<br>(kPa) | Mean Indicated Flow Rate<br>(std ml/min) | Standard deviation of indication<br>(std ml/min) | Number of flow indications in mean | RUN                  | Reference Flow Rate<br>(std ml/min) | Standard uncertainty of reference<br>(std ml/min) |
| <b>Set Flow Rate</b>              |     |                     | <b>Stabilization 180 minutes</b> |                              |                                       |                               |                             |  |  |                                    |                      |                                     |   |
| 80                                | 1   | 21.10.2016          | 20.35                            | 101.10                       | 42                                    | 21.45                         | 101.60                      | 79.823                                   | 0.05   | 10                                 | 1                    | 80.061                              | 0.076   |
| 80                                | 2   | 21.10.2016          | 20.38                            | 101.10                       | 42                                    | 21.49                         | 101.60                      | 79.858                                   | 0.02   | 10                                 | 2                    | 80.078                              | 0.076   |
| 80                                | 3   | 21.10.2016          | 20.43                            | 101.10                       | 42                                    | 21.55                         | 101.60                      | 79.881                                   | 0.04   | 10                                 | 3                    | 80.151                              | 0.076   |
| 80                                | 4   | 21.10.2016          | 20.44                            | 101.09                       | 42                                    | 21.58                         | 101.59                      | 79.853                                   | 0.02   | 10                                 | 4                    | 80.118                              | 0.076   |
|                                   |     |                     |                                  |                              |                                       |                               |                             | Average=                                 | <b>79.85</b>                                     |                                    |                      | Average=                            | <b>80.10</b>                                      |
| <b>Change Flow Rate</b>           |     |                     | <b>Stabilization 3 minutes</b>   |                              |                                       |                               |                             |  |  |                                    |                      |                                     |   |
| 300                               | 1   | 21.10.2016          | 20.48                            | 101.07                       | 42                                    | 21.75                         | 101.57                      | 301.018                                  | 0.12   | 10                                 | 1                    | 301.652                             | 0.287   |
| 300                               | 2   | 21.10.2016          | 20.47                            | 101.06                       | 41                                    | 21.75                         | 101.57                      | 300.801                                  | 0.11   | 10                                 | 2                    | 301.467                             | 0.286   |
| 300                               | 3   | 21.10.2016          | 20.45                            | 101.06                       | 42                                    | 21.75                         | 101.57                      | 301.126                                  | 0.12   | 10                                 | 3                    | 301.757                             | 0.287   |
| 300                               | 4   | 21.10.2016          | 20.43                            | 101.06                       | 42                                    | 21.75                         | 101.57                      | 301.108                                  | 0.10   | 10                                 | 4                    | 301.811                             | 0.287   |
|                                   |     |                     |                                  |                              |                                       |                               |                             | Average=                                 | <b>301.01</b>                                    |                                    |                      | Average=                            | <b>301.67</b>                                     |
| <b>Change Flow Rate</b>           |     |                     | <b>Stabilization 3 minutes</b>   |                              |                                       |                               |                             |  |  |                                    |                      |                                     |   |
| 600                               | 1   | 21.10.2016          | 20.47                            | 101.06                       | 42                                    | 21.8                          | 101.58                      | 607.661                                  | 0.13   | 9                                  | 1                    | 608.769                             | 0.578   |
| 600                               | 2   | 21.10.2016          | 20.46                            | 101.07                       | 42                                    | 21.8                          | 101.58                      | 608.584                                  | 0.15   | 9                                  | 2                    | 609.734                             | 0.579   |
| 600                               | 3   | 21.10.2016          | 20.47                            | 101.06                       | 42                                    | 21.8                          | 101.58                      | 609.003                                  | 0.19   | 9                                  | 3                    | 610.054                             | 0.580   |
| 600                               | 4   | 21.10.2016          | 20.47                            | 101.06                       | 42                                    | 21.8                          | 101.58                      | 609.104                                  | 0.20   | 9                                  | 4                    | 610.243                             | 0.580   |
|                                   |     |                     |                                  |                              |                                       |                               |                             | Average=                                 | <b>608.59</b>                                    |                                    |                      | Average=                            | <b>609.70</b>                                     |
| <b>Change Flow Rate</b>           |     |                     | <b>Stabilization 3 minutes</b>   |                              |                                       |                               |                             |  |  |                                    |                      |                                     |   |
| 1250                              | 1   | 21.10.2016          | 20.45                            | 101.05                       | 42                                    | 21.8                          | 101.57                      | 1248.586                                 | 8.63   | 7                                  | 1                    | 1248.516                            | 1.186   |
| 1250                              | 2   | 21.10.2016          | 20.45                            | 101.05                       | 42                                    | 21.8                          | 101.57                      | 1251.600                                 | 12.02  | 7                                  | 2                    | 1252.227                            | 1.190   |
| 1250                              | 3   | 21.10.2016          | 20.46                            | 101.05                       | 42                                    | 21.8                          | 101.57                      | 1252.614                                 | 9.81   | 7                                  | 3                    | 1254.639                            | 1.192   |
| 1250                              | 4   | 21.10.2016          | 20.46                            | 101.05                       | 42                                    | 21.8                          | 101.57                      | 1247.171                                 | 5.78   | 7                                  | 4                    | 1247.784                            | 1.185   |
|                                   |     |                     |                                  |                              |                                       |                               |                             | Average=                                 | <b>1249.99</b>                                   |                                    |                      | Average=                            | <b>1250.79</b>                                    |

**INITIAL TESTS**

Compare BIOS temperature and pressure readings with calibrated temperature and ambient pressure instrument readings.

| Laboratory temperature<br>(°C)                 | Laboratory pressure<br>(kPa) | BIOS temperature<br>(°C) | BIOS pressure<br>(kPa) |
|--|------------------------------|--------------------------|------------------------|
| Stabilization 24 hours in laboratory (no flow) |                              |                          |                        |
| 20.29  | 101.369                      | 20.36                    | 101.37                 |

**COMPARISON MEASUREMENTS**

Flow cell: -44 (S/N: 135198)

|                                   |     |                     | ENVIRONMENTAL PARAMETERS         |                              |                                       | TRANSFER STANDARD PARAMETERS  |                             |  |  |                                    | REFERENCE PARAMETERS |                                     |   |
|-----------------------------------|-----|---------------------|----------------------------------|------------------------------|---------------------------------------|-------------------------------|-----------------------------|--|--|------------------------------------|----------------------|-------------------------------------|---|
| Nominal Flow Rate<br>(std ml/min) | RUN | Date<br>(dd.mm.yyy) | Laboratory temperature<br>(°C)   | Laboratory pressure<br>(kPa) | Laboratory relative humidity<br>(%rh) | BIOS mean temperature<br>(°C) | BIOS mean pressure<br>(kPa) | Mean Indicated Flow Rate<br>(std ml/min) | Standard deviation of indication<br>(std ml/min) | Number of flow indications in mean | RUN                  | Reference Flow Rate<br>(std ml/min) | Standard uncertainty of reference<br>(std ml/min) |
| <b>Set Flow Rate</b>              |     |                     | <b>Stabilization 180 minutes</b> |                              |                                       |                               |                             |  |  |                                    |                      |                                     |   |
| 1250                              | 1   | 25.10.2016          | 20.35                            | 101.45                       | 38                                    | 21.51                         | 101.96                      | 1261.73                                  | 3.48   | 24                                 | 1                    | 1259.88                             | 1.13  |
| 1250                              | 2   | 25.10.2016          | 20.39                            | 101.45                       | 38                                    | 21.53                         | 101.96                      | 1222.16                                  | 3.54   | 25                                 | 2                    | 1220.20                             | 1.10  |
| 1250                              | 3   | 25.10.2016          | 20.41                            | 101.45                       | 38                                    | 21.56                         | 101.96                      | 1235.81                                  | 3.56   | 25                                 | 3                    | 1233.61                             | 1.11  |
| 1250                              | 4   | 25.10.2016          | 20.39                            | 101.46                       | 38                                    | 21.57                         | 101.97                      | 1251.13                                  | 3.28   | 24                                 | 4                    | 1249.23                             | 1.12  |
|                                   |     |                     |                                  |                              |                                       |                               |                             | Average=                                 | <b>1242.71</b>                                   |                                    | Average=             | <b>1240.73</b>                      |   |
| <b>Change Flow Rate</b>           |     |                     | <b>Stabilization 3 minutes</b>   |                              |                                       |                               |                             |  |  |                                    |                      |                                     |   |
| 5000                              | 1   | 24.10.2016          | 20.39                            | 101.50                       | 38                                    | 21.49                         | 102.08                      | 4998.36                                  | 1.85   | 21                                 | 1                    | 4999.53                             | 4.50  |
| 5000                              | 2   | 24.10.2016          | 20.41                            | 101.50                       | 38                                    | 21.50                         | 102.09                      | 5003.98                                  | 1.29   | 20                                 | 2                    | 5004.31                             | 4.50  |
| 5000                              | 3   | 24.10.2016          | 20.45                            | 101.50                       | 38                                    | 21.48                         | 102.08                      | 5013.42                                  | 1.60   | 19                                 | 3                    | 5014.33                             | 4.51  |
| 5000                              | 4   | 24.10.2016          | 20.45                            | 101.49                       | 39                                    | 21.50                         | 102.07                      | 5033.79                                  | 1.59   | 19                                 | 4                    | 5033.92                             | 4.53  |
|                                   |     |                     |                                  |                              |                                       |                               |                             | Average=                                 | <b>5012.39</b>                                   |                                    | Average=             | <b>5013.03</b>                      |   |
| <b>Change Flow Rate</b>           |     |                     | <b>Stabilization 3 minutes</b>   |                              |                                       |                               |                             |  |  |                                    |                      |                                     |   |
| 10000                             | 1   | 24.10.2016          | 20.44                            | 101.49                       | 39                                    | 21.41                         | 102.24                      | 9936.65                                  | 1.56   | 14                                 | 1                    | 9945.92                             | 8.95  |
| 10000                             | 2   | 24.10.2016          | 20.44                            | 101.50                       | 39                                    | 21.41                         | 102.25                      | 9942.17                                  | 2.01   | 15                                 | 2                    | 9946.33                             | 8.95  |
| 10000                             | 3   | 24.10.2016          | 20.45                            | 101.50                       | 39                                    | 23.42                         | 102.25                      | 9946.49                                  | 2.67   | 15                                 | 3                    | 9959.26                             | 8.96  |
| 10000                             | 4   | 24.10.2016          | 20.46                            | 101.50                       | 39                                    | 21.42                         | 102.26                      | 9963.58                                  | 1.72   | 13                                 | 4                    | 9976.10                             | 8.98  |
|                                   |     |                     |                                  |                              |                                       |                               |                             | Average=                                 | <b>9947.22</b>                                   |                                    | Average=             | <b>9956.90</b>                      |   |
| <b>Change Flow Rate</b>           |     |                     | <b>Stabilization 3 minutes</b>   |                              |                                       |                               |                             |  |  |                                    |                      |                                     |   |
| 20000                             | 1   | 25.10.2016          | 20.34                            | 101.83                       | 37                                    | 20.69                         | 103.12                      | 20144.83                                 | 6.37   | 12                                 | 1                    | 20147.12                            | 18.13   |
| 20000                             | 2   | 25.10.2016          | 20.34                            | 101.83                       | 37                                    | 20.72                         | 103.10                      | 20157.58                                 | 9.13   | 12                                 | 2                    | 20163.33                            | 18.15   |
| 20000                             | 3   | 25.10.2016          | 20.33                            | 101.83                       | 37                                    | 20.73                         | 103.10                      | 20176.67                                 | 7.29   | 12                                 | 3                    | 20178.12                            | 18.16   |
| 20000                             | 4   | 25.10.2016          | 20.31                            | 101.83                       | 37                                    | 20.73                         | 103.09                      | 20190.25                                 | 9.27   | 12                                 | 4                    | 20200.32                            | 18.18   |
|                                   |     |                     |                                  |                              |                                       |                               |                             | Average=                                 | <b>20167.33</b>                                  |                                    | Average=             | <b>20172.22</b>                     |   |
| <b>Change Flow Rate</b>           |     |                     | <b>Stabilization 3 minutes</b>   |                              |                                       |                               |                             |  |  |                                    |                      |                                     |   |
| 30000                             | 1   | 24.10.2016          | 20.53                            | 101.48                       | 39                                    | 21.15                         | 103.50                      | 30217.11                                 | 20.36  | 9                                  | 1                    | 30380.72                            | 27.34   |
| 30000                             | 2   | 24.10.2016          | 20.55                            | 101.47                       | 39                                    | 21.17                         | 103.52                      | 30425.67                                 | 17.31  | 9                                  | 2                    | 30574.36                            | 27.52   |
| 30000                             | 3   | 24.10.2016          | 20.56                            | 101.48                       | 39                                    | 21.13                         | 103.51                      | 30340.33                                 | 14.21  | 9                                  | 3                    | 30487.88                            | 27.44   |
| 30000                             | 4   | 24.10.2016          | 20.56                            | 101.47                       | 39                                    | 21.11                         | 103.47                      | 29913.56                                 | 11.47  | 9                                  | 4                    | 30055.90                            | 27.05   |
|                                   |     |                     |                                  |                              |                                       |                               |                             | Average=                                 | <b>30224.17</b>                                  |                                    | Average=             | <b>30374.72</b>                     |   |
| <b>End of measurements</b>        |     |                     |                                  |                              |                                       |                               |                             |  |  |                                    |                      |                                     |   |

**METAS results (loop1)**

**INITIAL TESTS**

Compare BIOS temperature and pressure readings with calibrated temperature and ambient pressure instrument readings.

| Laboratory temperature                         | Laboratory pressure | BIOS temperature | BIOS pressure |
|--|---------------------|------------------|---------------|
| (°C)   | (kPa)               | (°C)             | (kPa)         |
| Stabilization 24 hours in laboratory (no flow) |                     |                  |               |
| 20.118   | 95.995              | 20.56            | 95.987        |

**COMPARISON MEASUREMENTS**

*Flow cell: -10 (S/N: 135207)*

| ENVIRONMENTAL PARAMETERS |     |             | TRANSFER STANDARD PARAMETERS     |                     |                              |                       |                    |                          |                                  |                                    | REFERENCE PARAMETERS |                     |                                   |                            |                         |
|--------------------------|-----|-------------|----------------------------------|---------------------|------------------------------|-----------------------|--------------------|--------------------------|----------------------------------|------------------------------------|----------------------|---------------------|-----------------------------------|----------------------------|-------------------------|
| Nominal Flow Rate        | RUN | Date        | Laboratory temperature           | Laboratory pressure | Laboratory relative humidity | BIOS mean temperature | BIOS mean pressure | Mean Indicated Flow Rate | Standard deviation of indication | Number of flow indications in mean | RUN                  | Reference Flow Rate | Standard uncertainty of reference | Reference mean temperature | Reference mean pressure |
| (std ml/min)             |     | (dd.mm.yyy) | (°C)                             | (kPa)               | (%rh)                        | (°C)                  | (kPa)              | (std ml/min)             | (std ml/min)                     |                                    |                      | (std ml/min)        | (std ml/min)                      | (°C)                       | (kPa)                   |
| <b>Set Flow Rate</b>     |     |             | <b>Stabilization 180 minutes</b> |                     |                              |                       |                    |                          |                                  |                                    |                      |                     |                                   |                            |                         |
| 5                        | 1   | 18.1.2017   | 20.095                           | 96.017              | 46.5                         | 20.61                 | 96.162             | 4.9944                   | 0.0012                           | 10                                 | 1                    | 5.008               | 0.008                             | 20.113                     | 96.047                  |
| 5                        | 2   | 18.1.2017   | 20.111                           | 96.044              | 46.4                         | 20.65                 | 96.184             | 4.9984                   | 0.0013                           | 10                                 | 2                    | 5.028               | 0.008                             | 20.114                     | 96.104                  |
| 5                        | 3   | 18.1.2017   | 20.098                           | 96.111              | 46.7                         | 20.66                 | 96.216             | 4.9998                   | 0.0015                           | 10                                 | 3                    | 5.035               | 0.008                             | 20.103                     | 96.109                  |
| 5                        | 4   | 18.1.2017   | 20.099                           | 96.125              | 46.7                         | 20.69                 | 96.220             | 4.9963                   | 0.0013                           | 10                                 | 4                    | 5.031               | 0.008                             | 20.092                     | 96.128                  |
|                          |     |             |                                  |                     |                              |                       | Average=           | <b>4.9972</b>            |                                  |                                    | Average=             | <b>5.025</b>        |                                   |                            |                         |
| <b>Change Flow Rate</b>  |     |             | <b>Stabilization 3 minutes</b>   |                     |                              |                       |                    |                          |                                  |                                    |                      |                     |                                   |                            |                         |
| 10                       | 1   | 18.1.2017   | 20.093                           | 96.160              | 46.8                         | 20.64                 | 96.262             | 10.114                   | 0.002                            | 10                                 | 1                    | 10.185              | 0.016                             | 20.072                     | 96.161                  |
| 10                       | 2   | 18.1.2017   | 20.063                           | 96.161              | 47.2                         | 20.65                 | 96.257             | 10.104                   | 0.003                            | 10                                 | 2                    | 10.153              | 0.016                             | 20.128                     | 96.162                  |
| 10                       | 3   | 18.1.2017   | 20.125                           | 96.157              | 46.6                         | 20.65                 | 96.259             | 10.089                   | 0.002                            | 10                                 | 3                    | 10.151              | 0.016                             | 20.128                     | 96.156                  |
| 10                       | 4   | 18.1.2017   | 20.073                           | 96.152              | 46.8                         | 20.66                 | 96.251             | 10.090                   | 0.002                            | 10                                 | 4                    | 10.152              | 0.016                             | 20.064                     | 96.151                  |
|                          |     |             |                                  |                     |                              |                       | Average=           | <b>10.099</b>            |                                  |                                    | Average=             | <b>10.160</b>       |                                   |                            |                         |
| <b>Change Flow Rate</b>  |     |             | <b>Stabilization 3 minutes</b>   |                     |                              |                       |                    |                          |                                  |                                    |                      |                     |                                   |                            |                         |
| 20                       | 1   | 18.1.2017   | 20.063                           | 96.142              | 46.6                         | 20.63                 | 96.227             | 20.069                   | 0.002                            | 10                                 | 1                    | 20.162              | 0.031                             | 20.062                     | 96.140                  |
| 20                       | 2   | 18.1.2017   | 20.072                           | 96.130              | 46.7                         | 20.63                 | 96.225             | 20.074                   | 0.002                            | 10                                 | 2                    | 20.149              | 0.031                             | 20.114                     | 96.130                  |
| 20                       | 3   | 18.1.2017   | 20.036                           | 96.124              | 46.7                         | 20.64                 | 96.220             | 20.065                   | 0.008                            | 10                                 | 3                    | 20.138              | 0.031                             | 20.121                     | 96.126                  |
| 20                       | 4   | 18.1.2017   | 20.111                           | 96.113              | 46.7                         | 20.64                 | 96.208             | 20.064                   | 0.003                            | 10                                 | 4                    | 20.139              | 0.031                             | 20.092                     | 96.111                  |
|                          |     |             |                                  |                     |                              |                       | Average=           | <b>20.068</b>            |                                  |                                    | Average=             | <b>20.147</b>       |                                   |                            |                         |
| <b>Change Flow Rate</b>  |     |             | <b>Stabilization 3 minutes</b>   |                     |                              |                       |                    |                          |                                  |                                    |                      |                     |                                   |                            |                         |
| 80                       | 1   | 18.1.2017   | 20.085                           | 96.114              | 47.1                         | 20.60                 | 96.203             | 80.006                   | 0.004                            | 10                                 | 1                    | 80.09               | 0.13                              | 20.087                     | 96.114                  |
| 80                       | 2   | 18.1.2017   | 20.114                           | 96.104              | 47.2                         | 20.60                 | 96.203             | 79.989                   | 0.015                            | 10                                 | 2                    | 80.05               | 0.13                              | 20.117                     | 96.106                  |
| 80                       | 3   | 18.1.2017   | 20.120                           | 96.102              | 47.0                         | 20.60                 | 96.200             | 79.978                   | 0.006                            | 10                                 | 3                    | 80.04               | 0.13                              | 20.119                     | 96.105                  |
| 80                       | 4   | 18.1.2017   | 20.105                           | 96.108              | 47.1                         | 20.61                 | 96.202             | 79.958                   | 0.014                            | 10                                 | 4                    | 80.02               | 0.13                              | 20.102                     | 96.104                  |
|                          |     |             |                                  |                     |                              |                       | Average=           | <b>79.983</b>            |                                  |                                    | Average=             | <b>80.05</b>        |                                   |                            |                         |

**INITIAL TESTS**

Compare BIOS temperature and pressure readings with calibrated temperature and ambient pressure instrument readings.

| Laboratory temperature (°C)                    | Laboratory pressure (kPa) | BIOS temperature (°C) | BIOS pressure (kPa) |
|--|---------------------------|-----------------------|---------------------|
| Stabilization 24 hours in laboratory (no flow) |                           |                       |                     |
| 20.109   | 95.138                    | 20.60                 | 95.134              |
| 20.077   | 96.346                    | 20.74                 | 96.341              |

Date  
11.1.2017  
19.1.2017

**COMPARISON MEASUREMENTS**

Flow cell: -24 (S/N: 134909)

|                         |     |             | ENVIRONMENTAL PARAMETERS         |                     |                              | TRANSFER STANDARD PARAMETERS |                    |                          |                                  |                                    | REFERENCE PARAMETERS |                     |                                   |                            |                         |  |
|-------------------------|-----|-------------|----------------------------------|---------------------|------------------------------|------------------------------|--------------------|--------------------------|----------------------------------|------------------------------------|----------------------|---------------------|-----------------------------------|----------------------------|-------------------------|--|
| Nominal Flow Rate       | RUN | Date        | Laboratory temperature           | Laboratory pressure | Laboratory relative humidity | BIOS mean temperature        | BIOS mean pressure | Mean Indicated Flow Rate | Standard deviation of indication | Number of flow indications in mean | RUN                  | Reference Flow Rate | Standard uncertainty of reference | Reference mean temperature | Reference mean pressure |  |
| (std ml/min)            |     | (dd.mm.yyy) | (°C)                             | (kPa)               | (%rh)                        | (°C)                         | (kPa)              | (std ml/min)             | (std ml/min)                     |                                    |                      | (std ml/min)        | (std ml/min)                      | (°C)                       | (kPa)                   |  |
| <b>Set Flow Rate</b>    |     |             | <b>Stabilization 180 minutes</b> |                     |                              |                              |                    |                          |                                  |                                    |                      |                     |                                   |                            |                         |  |
| 80                      | 1   | 19.1.2017   | 20.105                           | 96.384              | 46.5                         | 20.68                        | 96.475             | 80.035                   | 0.008                            | 10                                 | 1                    | 80.65               | 0.07                              | 20.105                     | 96.381                  |  |
| 80                      | 2   | 19.1.2017   | 20.093                           | 96.391              | 46.8                         | 20.69                        | 96.473             | 80.055                   | 0.003                            | 10                                 | 2                    | 80.60               | 0.07                              | 20.094                     | 96.396                  |  |
| 80                      | 3   | 19.1.2017   | 20.093                           | 96.402              | 46.5                         | 20.70                        | 96.490             | 80.008                   | 0.005                            | 10                                 | 3                    | 80.60               | 0.07                              | 20.093                     | 96.401                  |  |
| 80                      | 4   | 19.1.2017   | 20.102                           | 96.394              | 46.6                         | 20.71                        | 96.485             | 80.013                   | 0.006                            | 10                                 | 4                    | 80.61               | 0.07                              | 20.103                     | 96.395                  |  |
|                         |     |             |                                  |                     |                              |                              |                    | Average=                 | <b>80.028</b>                    |                                    | Average=             |                     | <b>80.62</b>                      |                            |                         |  |
| <b>Change Flow Rate</b> |     |             | <b>Stabilization 3 minutes</b>   |                     |                              |                              |                    |                          |                                  |                                    |                      |                     |                                   |                            |                         |  |
| 300                     | 1   | 19.1.2017   | 20.100                           | 96.401              | 46.6                         | 20.70                        | 96.493             | 300.07                   | 0.01                             | 10                                 | 1                    | 301.92              | 0.25                              | 20.100                     | 96.400                  |  |
| 300                     | 2   | 19.1.2017   | 20.097                           | 96.410              | 47.0                         | 20.70                        | 96.495             | 300.10                   | 0.01                             | 10                                 | 2                    | 301.86              | 0.25                              | 20.096                     | 96.404                  |  |
| 300                     | 3   | 19.1.2017   | 20.105                           | 96.415              | 46.6                         | 20.69                        | 96.496             | 300.07                   | 0.01                             | 10                                 | 3                    | 301.80              | 0.25                              | 20.104                     | 96.407                  |  |
| 300                     | 4   | 19.1.2017   | 20.125                           | 96.397              | 46.6                         | 20.70                        | 96.488             | 300.07                   | 0.01                             | 10                                 | 4                    | 301.81              | 0.25                              | 20.128                     | 96.398                  |  |
|                         |     |             |                                  |                     |                              |                              |                    | Average=                 | <b>300.08</b>                    |                                    | Average=             |                     | <b>301.84</b>                     |                            |                         |  |
| <b>Change Flow Rate</b> |     |             | <b>Stabilization 3 minutes</b>   |                     |                              |                              |                    |                          |                                  |                                    |                      |                     |                                   |                            |                         |  |
| 600                     | 1   | 11.1.2017   | 20.026                           | 95.184              | 46.9                         | 20.59                        | 95.297             | 599.94                   | 0.02                             | 10                                 | 1                    | 603.32              | 0.49                              | 20.027                     | 95.183                  |  |
| 600                     | 2   | 11.1.2017   | 20.043                           | 95.145              | 46.8                         | 20.59                        | 95.284             | 600.06                   | 0.07                             | 10                                 | 2                    | 603.32              | 0.49                              | 20.044                     | 95.141                  |  |
| 600                     | 3   | 11.1.2017   | 20.054                           | 95.123              | 46.7                         | 20.59                        | 95.237             | 599.90                   | 0.10                             | 10                                 | 3                    | 603.18              | 0.49                              | 20.047                     | 95.130                  |  |
| 600                     | 4   | 11.1.2017   | 20.055                           | 95.114              | 46.9                         | 20.59                        | 95.234             | 600.09                   | 0.04                             | 10                                 | 4                    | 603.13              | 0.49                              | 20.056                     | 95.112                  |  |
|                         |     |             |                                  |                     |                              |                              |                    | Average=                 | <b>600.00</b>                    |                                    | Average=             |                     | <b>603.24</b>                     |                            |                         |  |
| <b>Change Flow Rate</b> |     |             | <b>Stabilization 3 minutes</b>   |                     |                              |                              |                    |                          |                                  |                                    |                      |                     |                                   |                            |                         |  |
| 1250                    | 1   | 11.1.2017   | 20.054                           | 95.104              | 46.9                         | 20.56                        | 95.224             | 1250.8                   | 0.1                              | 10                                 | 1                    | 1257.4              | 1.1                               | 20.054                     | 95.107                  |  |
| 1250                    | 2   | 11.1.2017   | 20.044                           | 95.084              | 47.0                         | 20.56                        | 95.224             | 1251.0                   | 0.1                              | 10                                 | 2                    | 1257.6              | 1.1                               | 20.044                     | 95.085                  |  |
| 1250                    | 3   | 11.1.2017   | 20.041                           | 95.065              | 46.9                         | 20.59                        | 95.201             | 1250.6                   | 0.1                              | 10                                 | 3                    | 1257.4              | 1.1                               | 20.043                     | 95.072                  |  |
| 1250                    | 4   | 11.1.2017   | 20.037                           | 95.073              | 46.9                         | 20.58                        | 95.202             | 1250.6                   | 0.1                              | 10                                 | 4                    | 1257.4              | 1.1                               | 20.037                     | 95.071                  |  |
|                         |     |             |                                  |                     |                              |                              |                    | Average=                 | <b>1250.8</b>                    |                                    | Average=             |                     | <b>1257.4</b>                     |                            |                         |  |



**INITIAL TESTS**

Compare BIOS temperature and pressure readings with calibrated temperature and ambient pressure instrument readings.

| Laboratory temperature (°C)                    | Laboratory pressure (kPa) | BIOS temperature (°C) | BIOS pressure (kPa) |
|--|---------------------------|-----------------------|---------------------|
| Stabilization 24 hours in laboratory (no flow) |                           |                       |                     |
| 20.008   | 95.301                    | 20.44                 | 95.300              |

**COMPARISON MEASUREMENTS**

Flow cell: -44 (S/N: 135198)

|                            |     |             | ENVIRONMENTAL PARAMETERS         |                     |                              | TRANSFER STANDARD PARAMETERS |                    |                          |                                  |                                    | REFERENCE PARAMETERS |                     |                                   |                            |                         |
|----------------------------|-----|-------------|----------------------------------|---------------------|------------------------------|------------------------------|--------------------|--------------------------|----------------------------------|------------------------------------|----------------------|---------------------|-----------------------------------|----------------------------|-------------------------|
| Nominal Flow Rate          | RUN | Date        | Laboratory temperature           | Laboratory pressure | Laboratory relative humidity | BIOS mean temperature        | BIOS mean pressure | Mean Indicated Flow Rate | Standard deviation of indication | Number of flow indications in mean | RUN                  | Reference Flow Rate | Standard uncertainty of reference | Reference mean temperature | Reference mean pressure |
| (std ml/min)               |     | (dd.mm.yyy) | (°C)                             | (kPa)               | (%rh)                        | (°C)                         | (kPa)              | (std ml/min)             | (std ml/min)                     |                                    |                      | (std ml/min)        | (std ml/min)                      | (°C)                       | (kPa)                   |
| <b>Set Flow Rate</b>       |     |             | <b>Stabilization 180 minutes</b> |                     |                              |                              |                    |                          |                                  |                                    |                      |                     |                                   |                            |                         |
| 1250                       | 1   | 12.1.2017   | 20.068                           | 95.257              | 46.7                         | 20.63                        | 95.402             | 1244.3                   | 0.1                              | 10                                 | 1                    | 1245.5              | 1.0                               | 20.071                     | 95.254                  |
| 1250                       | 2   | 12.1.2017   | 20.071                           | 95.254              | 46.8                         | 20.63                        | 95.391             | 1244.8                   | 0.1                              | 10                                 | 2                    | 1245.3              | 1.0                               | 20.068                     | 95.231                  |
| 1250                       | 3   | 12.1.2017   | 20.071                           | 95.234              | 46.9                         | 20.63                        | 95.374             | 1245.9                   | 0.1                              | 10                                 | 3                    | 1245.3              | 1.0                               | 20.068                     | 95.229                  |
| 1250                       | 4   | 12.1.2017   | 20.066                           | 95.223              | 47.0                         | 20.62                        | 95.377             | 1245.1                   | 0.2                              | 10                                 | 4                    | 1245.5              | 1.0                               | 20.065                     | 95.218                  |
|                            |     |             |                                  |                     |                              |                              |                    | Average=                 | <b>1245.0</b>                    |                                    | Average=             |                     | <b>1245.4</b>                     |                            |                         |
| <b>Change Flow Rate</b>    |     |             | <b>Stabilization 3 minutes</b>   |                     |                              |                              |                    |                          |                                  |                                    |                      |                     |                                   |                            |                         |
| 5000                       | 1   | 12.1.2017   | 20.046                           | 95.224              | 46.9                         | 20.59                        | 95.405             | 4988.7                   | 2.2                              | 10                                 | 1                    | 4989.8              | 4.0                               | 20.046                     | 95.229                  |
| 5000                       | 2   | 12.1.2017   | 20.044                           | 95.229              | 47.0                         | 20.58                        | 95.401             | 4987.0                   | 2.5                              | 10                                 | 2                    | 4989.0              | 4.0                               | 20.043                     | 95.200                  |
| 5000                       | 3   | 12.1.2017   | 20.045                           | 95.204              | 46.9                         | 20.59                        | 95.388             | 4989.1                   | 1.2                              | 10                                 | 3                    | 4988.4              | 4.0                               | 20.044                     | 95.198                  |
| 5000                       | 4   | 12.1.2017   | 20.042                           | 95.216              | 47.0                         | 20.57                        | 95.386             | 4990.8                   | 2.0                              | 10                                 | 4                    | 4989.0              | 4.0                               | 20.040                     | 95.206                  |
|                            |     |             |                                  |                     |                              |                              |                    | Average=                 | <b>4988.9</b>                    |                                    | Average=             |                     | <b>4989.0</b>                     |                            |                         |
| <b>Change Flow Rate</b>    |     |             | <b>Stabilization 3 minutes</b>   |                     |                              |                              |                    |                          |                                  |                                    |                      |                     |                                   |                            |                         |
| 10000                      | 1   | 12.1.2017   | 20.038                           | 95.199              | 46.8                         | 20.56                        | 95.342             | 9985.7                   | 6.6                              | 10                                 | 1                    | 9986                | 8                                 | 20.037                     | 95.140                  |
| 10000                      | 2   | 12.1.2017   | 20.042                           | 95.128              | 47.0                         | 20.58                        | 95.339             | 9982.7                   | 5.1                              | 10                                 | 2                    | 9986                | 8                                 | 20.045                     | 95.123                  |
| 10000                      | 3   | 12.1.2017   | 20.058                           | 95.118              | 47.0                         | 20.58                        | 95.317             | 9968.4                   | 3.8                              | 10                                 | 3                    | 9986                | 8                                 | 20.054                     | 95.120                  |
| 10000                      | 4   | 12.1.2017   | 20.062                           | 95.116              | 47.0                         | 20.60                        | 95.314             | 9980.2                   | 3.4                              | 10                                 | 4                    | 9985                | 8                                 | 20.064                     | 95.117                  |
|                            |     |             |                                  |                     |                              |                              |                    | Average=                 | <b>9979.3</b>                    |                                    | Average=             |                     | <b>9986</b>                       |                            |                         |
| <b>Change Flow Rate</b>    |     |             | <b>Stabilization 3 minutes</b>   |                     |                              |                              |                    |                          |                                  |                                    |                      |                     |                                   |                            |                         |
| 20000                      | 1   | 12.1.2017   | 20.095                           | 94.897              | 46.5                         | 20.54                        | 95.188             | 20154                    | 4                                | 10                                 | 1                    | 20142               | 17                                | 20.095                     | 94.898                  |
| 20000                      | 2   | 12.1.2017   | 20.091                           | 94.891              | 46.7                         | 20.56                        | 95.186             | 20156                    | 5                                | 10                                 | 2                    | 20166               | 17                                | 20.092                     | 94.886                  |
| 20000                      | 3   | 12.1.2017   | 20.091                           | 94.882              | 46.5                         | 20.58                        | 95.174             | 20173                    | 5                                | 10                                 | 3                    | 20169               | 17                                | 20.092                     | 94.881                  |
| 20000                      | 4   | 12.1.2017   | 20.081                           | 94.865              | 46.5                         | 20.60                        | 95.175             | 20170                    | 3                                | 10                                 | 4                    | 20181               | 17                                | 20.081                     | 94.865                  |
|                            |     |             |                                  |                     |                              |                              |                    | Average=                 | <b>20163</b>                     |                                    | Average=             |                     | <b>20165</b>                      |                            |                         |
| <b>Change Flow Rate</b>    |     |             | <b>Stabilization 3 minutes</b>   |                     |                              |                              |                    |                          |                                  |                                    |                      |                     |                                   |                            |                         |
| 30000                      | 1   | 12.1.2017   | 20.078                           | 94.844              | 46.3                         | 20.59                        | 95.274             | 30105                    | 12                               | 10                                 | 1                    | 30149               | 25                                | 20.078                     | 94.846                  |
| 30000                      | 2   | 12.1.2017   | 20.081                           | 94.830              | 46.4                         | 20.60                        | 95.278             | 30113                    | 13                               | 10                                 | 2                    | 30132               | 25                                | 20.081                     | 94.829                  |
| 30000                      | 3   | 12.1.2017   | 20.090                           | 94.805              | 46.3                         | 20.62                        | 95.243             | 30096                    | 7                                | 10                                 | 3                    | 30112               | 25                                | 20.084                     | 94.824                  |
| 30000                      | 4   | 12.1.2017   | 20.096                           | 94.787              | 46.4                         | 20.62                        | 95.242             | 30100                    | 3                                | 10                                 | 4                    | 30103               | 25                                | 20.096                     | 94.787                  |
|                            |     |             |                                  |                     |                              |                              |                    | Average=                 | <b>30104</b>                     |                                    | Average=             |                     | <b>30124</b>                      |                            |                         |
| <b>End of measurements</b> |     |             |                                  |                     |                              |                              |                    |                          |                                  |                                    |                      |                     |                                   |                            |                         |

**VSL results (loop1)**

**INITIAL TESTS**

Compare BIOS temperature and pressure readings with calibrated temperature and ambient pressure instrument readings.

| Laboratory temperature                         | Laboratory pressure | BIOS temperature | BIOS pressure |
|--|---------------------|------------------|---------------|
| (°C)   | (kPa)               | (°C)             | (kPa)         |
| Stabilization 24 hours in laboratory (no flow) |                     |                  |               |
|  |                     |                  |               |

**COMPARISON MEASUREMENTS**

**Flow cell: -10 (S/N: 135207)**

| Flow cell: -10 (S/N: 135207) |     |             | ENVIRONMENTAL PARAMETERS         |                     |                              | TRANSFER STANDARD PARAMETERS |                    |                          |                                  |                                    | REFERENCE PARAMETERS |                     |                                   |
|------------------------------|-----|-------------|----------------------------------|---------------------|------------------------------|------------------------------|--------------------|--------------------------|----------------------------------|------------------------------------|----------------------|---------------------|-----------------------------------|
| Nominal Flow Rate            | RUN | Date        | Laboratory temperature           | Laboratory pressure | Laboratory relative humidity | BIOS mean temperature        | BIOS mean pressure | Mean Indicated Flow Rate | Standard deviation of indication | Number of flow indications in mean | RUN                  | Reference Flow Rate | Standard uncertainty of reference |
| (std ml/min)                 |     | (dd.mm.yyy) | (°C)                             | (kPa)               | (%rh)                        | (°C)                         | (kPa)              | (std ml/min)             | (std ml/min)                     |                                    |                      | (std ml/min)        | (std ml/min)                      |
| <b>Set Flow Rate</b>         |     |             | <b>Stabilization 180 minutes</b> |                     |                              |                              |                    |                          |                                  |                                    |                      |                     |                                   |
| 5                            | 1   | 14.03.2017  | 20.12                            | 103.109             | 45                           | 20.40                        | 103.19             | 4.968                    | 0.003                            | 10                                 | 1                    | 4.981               | 0.010                             |
| 5                            | 2   | 14.03.2017  | 20.11                            | 103.125             | 45                           | 20.48                        | 103.20             | 4.972                    | 0.002                            | 10                                 | 2                    | 4.979               | 0.010                             |
| 5                            | 3   | 14.03.2017  | 20.14                            | 103.128             | 45                           | 20.52                        | 103.21             | 4.971                    | 0.003                            | 10                                 | 3                    | 4.980               | 0.010                             |
| 5                            | 4   |             |                                  |                     |                              |                              |                    |                          |                                  |                                    | 4                    |                     |                                   |
|                              |     |             |                                  |                     |                              |                              |                    | Average=                 | <b>4.97</b>                      |                                    |                      | Average=            | <b>4.98</b>                       |
| <b>Change Flow Rate</b>      |     |             | <b>Stabilization 3 minutes</b>   |                     |                              |                              |                    |                          |                                  |                                    |                      |                     |                                   |
| 10                           | 1   | 14.03.2017  | 20.14                            | 103.136             | 45                           | 20.66                        | 103.21             | 10.141                   | 0.005                            | 10                                 | 1                    | 10.155              | 0.020                             |
| 10                           | 2   | 14.03.2017  | 20.14                            | 103.136             | 45                           | 20.68                        | 103.22             | 10.141                   | 0.008                            | 10                                 | 2                    | 10.159              | 0.020                             |
| 10                           | 3   |             |                                  |                     |                              |                              |                    |                          |                                  |                                    | 3                    |                     |                                   |
| 10                           | 4   |             |                                  |                     |                              |                              |                    |                          |                                  |                                    | 4                    |                     |                                   |
|                              |     |             |                                  |                     |                              |                              |                    | Average=                 | <b>10.14</b>                     |                                    |                      | Average=            | <b>10.16</b>                      |
| <b>Change Flow Rate</b>      |     |             | <b>Stabilization 3 minutes</b>   |                     |                              |                              |                    |                          |                                  |                                    |                      |                     |                                   |
| 20                           | 1   | 14.03.2017  | 20.14                            | 103.126             | 45                           | 20.72                        | 103.21             | 20.252                   | 0.004                            | 10                                 | 1                    | 20.29               | 0.04                              |
| 20                           | 2   | 14.03.2017  | 20.15                            | 103.126             | 45                           | 20.7                         | 103.21             | 20.248                   | 0.009                            | 10                                 | 2                    | 20.30               | 0.04                              |
| 20                           | 3   |             |                                  |                     |                              |                              |                    |                          |                                  |                                    | 3                    |                     |                                   |
| 20                           | 4   |             |                                  |                     |                              |                              |                    |                          |                                  |                                    | 4                    |                     |                                   |
|                              |     |             |                                  |                     |                              |                              |                    | Average=                 | <b>20.25</b>                     |                                    |                      | Average=            | <b>20.30</b>                      |
| <b>Change Flow Rate</b>      |     |             | <b>Stabilization 3 minutes</b>   |                     |                              |                              |                    |                          |                                  |                                    |                      |                     |                                   |
| 80                           | 1   | 13.03.2017  | 20.19                            | 102.894             | 45                           | 20.60                        | 102.98             | 79.336                   | 0.016                            | 10                                 | 1                    | 79.50               | 0.16                              |
| 80                           | 2   | 13.03.2017  | 20.19                            | 102.894             | 45                           | 20.62                        | 102.98             | 79.333                   | 0.023                            | 10                                 | 2                    | 79.60               | 0.16                              |
| 80                           | 3   | 13.03.2017  | 20.17                            | 102.902             | 45                           | 20.64                        | 102.98             | 79.319                   | 0.039                            | 10                                 | 3                    | 79.60               | 0.16                              |
| 80                           | 4   | 13.03.2017  | 20.17                            | 102.902             | 45                           | 20.64                        | 103.00             | 79.331                   | 0.020                            | 10                                 | 4                    | 79.57               | 0.16                              |
|                              |     |             |                                  |                     |                              |                              |                    | Average=                 | <b>79.33</b>                     |                                    |                      | Average=            | <b>79.57</b>                      |

**INITIAL TESTS**

Compare BIOS temperature and pressure readings with calibrated temperature and ambient pressure instrument readings.

| Laboratory temperature<br>(°C)                 | Laboratory pressure<br>(kPa) | BIOS temperature<br>(°C) | BIOS pressure<br>(kPa) |
|--|------------------------------|--------------------------|------------------------|
| Stabilization 24 hours in laboratory (no flow) |                              |                          |                        |
|  |                              |                          |                        |

**COMPARISON MEASUREMENTS**

**Flow cell: -24 (S/N: 134909)**

|                                   |     |                     | ENVIRONMENTAL PARAMETERS         |                              |                                       | TRANSFER STANDARD PARAMETERS  |                             |  |  |                                    | REFERENCE PARAMETERS |                                     |   |
|-----------------------------------|-----|---------------------|----------------------------------|------------------------------|---------------------------------------|-------------------------------|-----------------------------|--|--|------------------------------------|----------------------|-------------------------------------|---|
| Nominal Flow Rate<br>(std ml/min) | RUN | Date<br>(dd.mm.yyy) | Laboratory temperature<br>(°C)   | Laboratory pressure<br>(kPa) | Laboratory relative humidity<br>(%rh) | BIOS mean temperature<br>(°C) | BIOS mean pressure<br>(kPa) | Mean Indicated Flow Rate<br>(std ml/min) | Standard deviation of indication<br>(std ml/min) | Number of flow indications in mean | RUN                  | Reference Flow Rate<br>(std ml/min) | Standard uncertainty of reference<br>(std ml/min) |
| <b>Set Flow Rate</b>              |     |                     | <b>Stabilization 180 minutes</b> |                              |                                       |                               |                             |  |  |                                    |                      |                                     |   |
| 80                                | 1   | 21.2.2017           | 20.71                            | 101.607                      | 45                                    | 20.92                         | 101.69                      | 79.95                                    | 0.04   | 10                                 | 1                    | 80.89                               | 0.16  |
| 80                                | 2   | 21.2.2017           | 20.66                            | 101.607                      | 45                                    | 20.96                         | 101.65                      | 79.92                                    | 0.03   | 10                                 | 2                    | 80.94                               | 0.16  |
| 80                                | 3   | 21.2.2017           | 20.59                            | 101.607                      | 45                                    | 20.96                         | 101.65                      | 79.91                                    | 0.02   | 10                                 | 3                    | 80.95                               | 0.16  |
| 80                                | 4   | 21.2.2017           | 20.55                            | 101.607                      | 45                                    | 20.98                         | 101.65                      | 79.88                                    | 0.03   | 10                                 | 4                    | 80.88                               | 0.16  |
|                                   |     |                     |                                  |                              |                                       |                               |                             | Average=                                 | <b>79.91</b>                                     |                                    | Average=             | <b>80.92</b>                        |   |
| <b>Change Flow Rate</b>           |     |                     | <b>Stabilization 3 minutes</b>   |                              |                                       |                               |                             |  |  |                                    |                      |                                     |   |
| 300                               | 1   | 21.2.2017           | 20.48                            | 101.553                      | 45                                    | 20.86                         | 101.68                      | 300.32                                   | 0.14   | 10                                 | 1                    | 301.2                               | 0.3   |
| 300                               | 2   | 21.2.2017           | 20.48                            | 101.553                      | 45                                    | 20.88                         | 101.69                      | 300.41                                   | 0.09   | 10                                 | 2                    | 301.2                               | 0.3   |
| 300                               | 3   | 21.2.2017           | 20.46                            | 101.553                      | 45                                    | 20.9                          | 101.69                      | 300.44                                   | 0.12   | 10                                 | 3                    | 301.1                               | 0.3   |
| 300                               | 4   | 21.2.2017           | 20.43                            | 101.598                      | 45                                    | 20.88                         | 101.7                       | 300.44                                   | 0.15   | 10                                 | 4                    | 301.4                               | 0.3   |
|                                   |     |                     |                                  |                              |                                       |                               |                             | Average=                                 | <b>300.40</b>                                    |                                    | Average=             | <b>301.21</b>                       |   |
| <b>Change Flow Rate</b>           |     |                     | <b>Stabilization 3 minutes</b>   |                              |                                       |                               |                             |  |  |                                    |                      |                                     |   |
| 600                               | 1   | 21.2.2017           | 20.47                            | 101.532                      | 45                                    | 20.74                         | 101.63                      | 600.36                                   | 0.3  | 10                                 | 1                    | 601.7                               | 0.6   |
| 600                               | 2   | 21.2.2017           | 20.47                            | 101.532                      | 45                                    | 20.74                         | 101.63                      | 600.36                                   | 0.26   | 10                                 | 2                    | 601.7                               | 0.6   |
| 600                               | 3   | 21.2.2017           | 20.48                            | 101.532                      | 45                                    | 20.78                         | 101.63                      | 600.36                                   | 0.24   | 10                                 | 3                    | 601.8                               | 0.6   |
| 600                               | 4   | 21.2.2017           | 20.49                            | 101.532                      | 45                                    | 20.78                         | 101.63                      | 600.41                                   | 0.25   | 10                                 | 4                    | 601.6                               | 0.6   |
|                                   |     |                     |                                  |                              |                                       |                               |                             | Average=                                 | <b>600.37</b>                                    |                                    | Average=             | <b>601.71</b>                       |   |
| <b>Change Flow Rate</b>           |     |                     | <b>Stabilization 3 minutes</b>   |                              |                                       |                               |                             |  |  |                                    |                      |                                     |   |
| 1250                              | 1   | 20.2.2017           | 20.52                            | 101.43                       | 45                                    | 20.88                         | 101.54                      | 1248.2                                   | 0.75   | 10                                 | 1                    | 1251.7                              | 1.3   |
| 1250                              | 2   | 20.2.2017           | 20.46                            | 101.43                       | 45                                    | 20.86                         | 101.54                      | 1247.7                                   | 0.49   | 10                                 | 2                    | 1251.2                              | 1.3   |
| 1250                              | 3   | 20.2.2017           | 20.45                            | 101.43                       | 45                                    | 20.86                         | 101.55                      | 1247.4                                   | 0.37   | 10                                 | 3                    | 1251.1                              | 1.3   |
| 1250                              | 4   | 20.2.2017           | 20.47                            | 101.43                       | 45                                    | 20.86                         | 101.54                      | 1247.3                                   | 0.38   | 10                                 | 4                    | 1250.9                              | 1.3   |
|                                   |     |                     |                                  |                              |                                       |                               |                             | Average=                                 | <b>1247.65</b>                                   |                                    | Average=             | <b>1251.24</b>                      |   |

**INITIAL TESTS**

Compare BIOS temperature and pressure readings with calibrated temperature and ambient pressure instrument readings.

| Laboratory temperature<br>(°C)                 | Laboratory pressure<br>(kPa) | BIOS temperature<br>(°C) | BIOS pressure<br>(kPa) |
|--|------------------------------|--------------------------|------------------------|
| Stabilization 24 hours in laboratory (no flow) |                              |                          |                        |
|  |                              |                          |                        |

**COMPARISON MEASUREMENTS**

Flow cell: -44 (S/N: 135198)

|                                   |     |                     | ENVIRONMENTAL PARAMETERS         |                              |                                       | TRANSFER STANDARD PARAMETERS  |                             |  |  |                                    | REFERENCE PARAMETERS |                                     |   |
|-----------------------------------|-----|---------------------|----------------------------------|------------------------------|---------------------------------------|-------------------------------|-----------------------------|--|--|------------------------------------|----------------------|-------------------------------------|---|
| Nominal Flow Rate<br>(std ml/min) | RUN | Date<br>(dd.mm.yyy) | Laboratory temperature<br>(°C)   | Laboratory pressure<br>(kPa) | Laboratory relative humidity<br>(%rh) | BIOS mean temperature<br>(°C) | BIOS mean pressure<br>(kPa) | Mean Indicated Flow Rate<br>(std ml/min) | Standard deviation of indication<br>(std ml/min) | Number of flow indications in mean | RUN                  | Reference Flow Rate<br>(std ml/min) | Standard uncertainty of reference<br>(std ml/min) |
| <b>Set Flow Rate</b>              |     |                     | <b>Stabilization 180 minutes</b> |                              |                                       |                               |                             |  |  |                                    |                      |                                     |   |
| 1250                              | 1   | 16.2.2017           | 20.38                            | 102.78                       | 45                                    | 20.80                         | 102.93                      | 1249.4                                   |  | 10                                 | 1                    | 1249.8                              | 1.2   |
| 1250                              | 2   | 16.2.2017           | 20.38                            | 102.78                       | 45                                    | 20.80                         | 102.92                      | 1248.9                                   |  | 10                                 | 2                    | 1249.9                              | 1.2   |
| 1250                              | 3   | 16.2.2017           | 20.38                            | 102.78                       | 45                                    | 20.80                         | 102.92                      | 1249.3                                   |  | 10                                 | 3                    | 1250.1                              | 1.3   |
| 1250                              | 4   | 16.2.2017           | 20.39                            | 102.78                       | 45                                    | 20.80                         | 102.92                      | 1249.2                                   |  | 10                                 | 4                    | 1249.6                              | 1.2   |
|                                   |     |                     |                                  |                              |                                       |                               |                             | Average=                                 |  |                                    | Average=             | 1249.84                             |   |
| <b>Change Flow Rate</b>           |     |                     | <b>Stabilization 3 minutes</b>   |                              |                                       |                               |                             |  |  |                                    |                      |                                     |   |
| 5000                              | 1   | 16.2.2017           | 20.42                            | 102.83                       | 45                                    | 20.72                         | 102.98                      | 4992.8                                   |  | 10                                 | 1                    | 4994                                | 5   |
| 5000                              | 2   | 16.2.2017           | 20.46                            | 102.83                       | 45                                    | 20.68                         | 102.97                      | 4992.9                                   |  | 10                                 | 2                    | 4994                                | 5   |
| 5000                              | 3   | 16.2.2017           | 20.36                            | 102.822                      | 45                                    | 20.66                         | 102.97                      | 4993.1                                   |  | 10                                 | 3                    | 4996                                | 5   |
| 5000                              | 4   | 16.2.2017           | 20.39                            | 102.822                      | 45                                    | 20.68                         | 102.96                      | 4992.9                                   |  | 10                                 | 4                    | 4995                                | 5   |
|                                   |     |                     |                                  |                              |                                       |                               |                             | Average=                                 |  |                                    | Average=             | 4994.81                             |   |
| <b>Change Flow Rate</b>           |     |                     | <b>Stabilization 3 minutes</b>   |                              |                                       |                               |                             |  |  |                                    |                      |                                     |   |
| 10000                             | 1   | 16.2.2017           | 20.47                            | 102.87                       | 45                                    | 20.72                         | 103.03                      | 10014                                    |  | 10                                 | 1                    | 10024                               | 10  |
| 10000                             | 2   | 16.2.2017           | 20.47                            | 102.87                       | 45                                    | 20.7                          | 103.03                      | 10016                                    |  | 10                                 | 2                    | 10025                               | 10  |
| 10000                             | 3   | 16.2.2017           | 20.45                            | 102.87                       | 45                                    | 20.7                          | 103.02                      | 10015                                    |  | 10                                 | 3                    | 10027                               | 10  |
| 10000                             | 4   | 16.2.2017           | 20.48                            | 102.87                       | 45                                    | 20.7                          | 103.02                      | 10017                                    |  | 10                                 | 4                    | 10027                               | 10  |
|                                   |     |                     |                                  |                              |                                       |                               |                             | Average=                                 |  |                                    | Average=             | 10025.43                            |   |
| <b>Change Flow Rate</b>           |     |                     | <b>Stabilization 3 minutes</b>   |                              |                                       |                               |                             |  |  |                                    |                      |                                     |   |
| 20000                             | 1   | 16.2.2017           | 20.56                            | 102.884                      | 45                                    | 20.76                         | 103.14                      | 20083                                    |  | 10                                 | 1                    | 20092                               | 20  |
| 20000                             | 2   | 16.2.2017           | 20.52                            | 102.884                      | 45                                    | 20.78                         | 103.14                      | 20083                                    |  | 10                                 | 2                    | 20097                               | 20  |
| 20000                             | 3   | 16.2.2017           | 20.53                            | 102.884                      | 45                                    | 20.76                         | 103.14                      | 20083                                    |  | 10                                 | 3                    | 20097                               | 20  |
| 20000                             | 4   | 16.2.2017           | 20.55                            | 102.884                      | 45                                    | 20.74                         | 103.14                      | 20084                                    |  | 10                                 | 4                    | 20095                               | 20  |
|                                   |     |                     |                                  |                              |                                       |                               |                             | Average=                                 |  |                                    | Average=             | 20095.40                            |   |
| <b>Change Flow Rate</b>           |     |                     | <b>Stabilization 3 minutes</b>   |                              |                                       |                               |                             |  |  |                                    |                      |                                     |   |
| 30000                             | 1   | 16.2.2017           | 20.66                            | 102.899                      | 45                                    | 20.76                         | 103.25                      | 29935                                    |  | 10                                 | 1                    | 29958                               | 30  |
| 30000                             | 2   | 16.2.2017           | 20.55                            | 102.899                      | 45                                    | 20.82                         | 102.85                      | 29947                                    |  | 10                                 | 2                    | 29967                               | 30  |
| 30000                             | 3   | 16.2.2017           | 20.61                            | 102.899                      | 45                                    | 20.78                         | 103.26                      | 29934                                    |  | 10                                 | 3                    | 29959                               | 30  |
| 30000                             | 4   | 16.2.2017           | 20.65                            | 102.899                      | 45                                    | 20.78                         | 103.26                      | 29946                                    |  | 10                                 | 4                    | 29954                               | 30  |
|                                   |     |                     |                                  |                              |                                       |                               |                             | Average=                                 |  |                                    | Average=             | 29959.44                            |   |
| <b>End of measurements</b>        |     |                     |                                  |                              |                                       |                               |                             | Average=                                 |  |                                    | Average=             | 29940.50                            |   |

**EIM results (loop1)**

**INITIAL TESTS**

Compare BIOS temperature and pressure readings with calibrated temperature and ambient pressure instrument readings.

| Laboratory temperature<br>(°C)                 | Laboratory pressure<br>(kPa) | BIOS temperature<br>(°C) | BIOS pressure<br>(kPa) |
|--|------------------------------|--------------------------|------------------------|
| Stabilization 24 hours in laboratory (no flow) |                              |                          |                        |
| 21.9   | 1012.5                       | 21.99                    | 1013.5                 |

**COMPARISON MEASUREMENTS**

Flow cell: -10 (S/N: 135207)

|                                   |     |                     | ENVIRONMENTAL PARAMETERS         |                              |                                       | TRANSFER STANDARD PARAMETERS  |                             |  |  |                                    | REFERENCE PARAMETERS |                                     |   |
|-----------------------------------|-----|---------------------|----------------------------------|------------------------------|---------------------------------------|-------------------------------|-----------------------------|--|--|------------------------------------|----------------------|-------------------------------------|---|
| Nominal Flow Rate<br>(std ml/min) | RUN | Date<br>(dd.mm.yyy) | Laboratory temperature<br>(°C)   | Laboratory pressure<br>(kPa) | Laboratory relative humidity<br>(%rh) | BIOS mean temperature<br>(°C) | BIOS mean pressure<br>(kPa) | Mean Indicated Flow Rate<br>(std ml/min) | Standard deviation of indication<br>(std ml/min) | Number of flow indications in mean | RUN                  | Reference Flow Rate<br>(std ml/min) | Standard uncertainty of reference<br>(std ml/min) |
| <b>Set Flow Rate</b>              |     |                     | <b>Stabilization 180 minutes</b> |                              |                                       |                               |                             |  |  |                                    |                      |                                     |   |
| 5                                 | 1   |                     |                                  |                              |                                       |                               |                             |  |  |                                    | 1                    |                                     |   |
| 5                                 | 2   |                     |                                  |                              |                                       |                               |                             |  |  |                                    | 2                    |                                     |   |
| 5                                 | 3   |                     |                                  |                              |                                       |                               |                             |  |  |                                    | 3                    |                                     |   |
| 5                                 | 4   |                     |                                  |                              |                                       |                               |                             |  |  |                                    | 4                    |                                     |   |
|                                   |     |                     |                                  |                              |                                       |                               |                             | Average=                                 | #DIV/0!  |                                    | Average=             | #DIV/0!                             |   |
| <b>Change Flow Rate</b>           |     |                     | <b>Stabilization 3 minutes</b>   |                              |                                       |                               |                             |  |  |                                    |                      |                                     |   |
| 10                                | 1   |                     |                                  |                              |                                       |                               |                             |  |  |                                    | 1                    |                                     |   |
| 10                                | 2   |                     |                                  |                              |                                       |                               |                             |  |  |                                    | 2                    |                                     |   |
| 10                                | 3   |                     |                                  |                              |                                       |                               |                             |  |  |                                    | 3                    |                                     |   |
| 10                                | 4   |                     |                                  |                              |                                       |                               |                             |  |  |                                    | 4                    |                                     |   |
|                                   |     |                     |                                  |                              |                                       |                               |                             | Average=                                 | #DIV/0!  |                                    | Average=             | #DIV/0!                             |   |
| <b>Change Flow Rate</b>           |     |                     | <b>Stabilization 3 minutes</b>   |                              |                                       |                               |                             |  |  |                                    |                      |                                     |   |
| 20                                | 1   | 07.05.2017          | 23.2                             | 1005                         | 42.8                                  | 23.74                         | 1007.2                      | 19.960                                   | 0.017  | 10                                 | 1                    | 19.963                              | 0.020   |
| 20                                | 2   |                     |                                  |                              |                                       | 23.80                         | 1006.5                      | 19.755                                   | 0.012  | 10                                 | 2                    | 19.743                              | 0.020   |
| 20                                | 3   |                     |                                  |                              |                                       | 23.83                         | 1006.3                      | 19.618                                   | 0.008  | 10                                 | 3                    | 19.570                              | 0.020   |
| 20                                | 4   |                     |                                  |                              |                                       | 23.90                         | 1005.6                      | 19.497                                   | 0.014  | 10                                 | 4                    | 19.451                              | 0.019   |
|                                   |     |                     |                                  |                              |                                       |                               |                             | Average=                                 | 19.707   |                                    | Average=             | 19.682                              |   |
| <b>Change Flow Rate</b>           |     |                     | <b>Stabilization 3 minutes</b>   |                              |                                       |                               |                             |  |  |                                    |                      |                                     |   |
| 80                                | 1   | 18.05.2017          | 23.9                             | 1015                         | 52                                    | 24.02                         | 1016.6                      | 82.756                                   | 0.058  | 10                                 | 1                    | 82.879                              | 0.083   |
| 80                                | 2   |                     | 23.9                             | 1015                         | 52                                    | 24.03                         | 1016.4                      | 82.621                                   | 0.133  | 10                                 | 2                    | 82.619                              | 0.083   |
| 80                                | 3   |                     | 23.9                             | 1015                         | 52                                    | 24.06                         | 1016.4                      | 82.474                                   | 0.047  | 10                                 | 3                    | 82.483                              | 0.082   |
| 80                                | 4   |                     | 23.9                             | 1015                         | 52                                    | 24.06                         | 1016.4                      | 82.347                                   | 0.052  | 10                                 | 4                    | 82.372                              | 0.082   |
|                                   |     |                     |                                  |                              |                                       |                               |                             | Average=                                 | 82.550   |                                    | Average=             | 82.588                              |   |

**INITIAL TESTS**

Compare BIOS temperature and pressure readings with calibrated temperature and ambient pressure instrument readings.

| Laboratory temperature<br>(°C)                 | Laboratory pressure<br>(kPa) | BIOS temperature<br>(°C) | BIOS pressure<br>(kPa) |
|--|------------------------------|--------------------------|------------------------|
| Stabilization 24 hours in laboratory (no flow) |                              |                          |                        |
| 24   | 1021.1                       | 23.78                    | 1021.4                 |

**COMPARISON MEASUREMENTS**

**Flow cell: -24 (S/N: 134909)**

| Flow cell: -24 (S/N: 134909)      |     |                     | ENVIRONMENTAL PARAMETERS         |                              |                                       | TRANSFER STANDARD PARAMETERS  |                             |  |  |                                    | REFERENCE PARAMETERS    |                                     |   |
|-----------------------------------|-----|---------------------|----------------------------------|------------------------------|---------------------------------------|-------------------------------|-----------------------------|--|--|------------------------------------|-------------------------|-------------------------------------|---|
| Nominal Flow Rate<br>(std ml/min) | RUN | Date<br>(dd.mm.yyy) | Laboratory temperature<br>(°C)   | Laboratory pressure<br>(kPa) | Laboratory relative humidity<br>(%rh) | BIOS mean temperature<br>(°C) | BIOS mean pressure<br>(kPa) | Mean Indicated Flow Rate<br>(std ml/min) | Standard deviation of indication<br>(std ml/min) | Number of flow indications in mean | RUN                     | Reference Flow Rate<br>(std ml/min) | Standard uncertainty of reference<br>(std ml/min) |
| <b>Set Flow Rate</b>              |     |                     | <b>Stabilization 180 minutes</b> |                              |                                       |                               |                             |  |  |                                    |                         |                                     |   |
| 80                                | 1   | 16.05.2017          | 23.9                             | 1021.2                       | 58.9                                  | 23.96                         | 1022.9                      | 78.967                                   | 0.105  | 10                                 | 1                       | 79.008                              | 0.079   |
| 80                                | 2   |                     |                                  |                              |                                       | 24.02                         | 1022.7                      | 78.226                                   | 0.057  | 10                                 | 2                       | 78.208                              | 0.078   |
| 80                                | 3   |                     |                                  |                              |                                       | 24.10                         | 1022.5                      | 79.143                                   | 0.900  | 10                                 | 3                       | 79.147                              | 0.079   |
| 80                                | 4   |                     |                                  |                              |                                       | 24.18                         | 1022.3                      | 78.910                                   | 0.143  | 10                                 | 4                       | 78.846                              | 0.079   |
|                                   |     |                     |                                  |                              |                                       |                               |                             | Average=                                 | <b>78.812</b>                                    |                                    | Average= <b>78.802</b>  |                                     |   |
| <b>Change Flow Rate</b>           |     |                     | <b>Stabilization 3 minutes</b>   |                              |                                       |                               |                             |  |  |                                    |                         |                                     |   |
| 300                               | 1   | 17.05.2017          | 23.2                             | 1014.7                       | 55.3                                  | 23.86                         | 1015.9                      | 300.58                                   | 0.27   | 10                                 | 1                       | 300.89                              | 0.30  |
| 300                               | 2   |                     |                                  |                              |                                       | 23.90                         | 1015.7                      | 302.07                                   | 0.55   | 10                                 | 2                       | 302.40                              | 0.30  |
| 300                               | 3   |                     |                                  |                              |                                       | 24.00                         | 1015.3                      | 299.74                                   | 0.21   | 10                                 | 3                       | 299.75                              | 0.30  |
| 300                               | 4   |                     |                                  |                              |                                       | 24.05                         | 1015.1                      | 302.98                                   | 0.49   | 10                                 | 4                       | 303.35                              | 0.30  |
|                                   |     |                     |                                  |                              |                                       |                               |                             | Average=                                 | <b>301.34</b>                                    |                                    | Average= <b>301.60</b>  |                                     |   |
| <b>Change Flow Rate</b>           |     |                     | <b>Stabilization 3 minutes</b>   |                              |                                       |                               |                             |  |  |                                    |                         |                                     |   |
| 600                               | 1   | 17.05.2017          | 23.2                             | 1014.7                       | 55.3                                  | 24.32                         | 1014.5                      | 609.17                                   | 1.12   | 10                                 | 1                       | 609.67                              | 0.61  |
| 600                               | 2   |                     |                                  |                              |                                       | 24.32                         | 1014.4                      | 610.27                                   | 0.41   | 10                                 | 2                       | 609.00                              | 0.61  |
| 600                               | 3   |                     |                                  |                              |                                       | 24.24                         | 1014.0                      | 609.82                                   | 0.73   | 10                                 | 3                       | 610.73                              | 0.61  |
| 600                               | 4   |                     |                                  |                              |                                       | 24.28                         | 1014.5                      | 609.21                                   | 1.70   | 10                                 | 4                       | 609.60                              | 0.61  |
|                                   |     |                     |                                  |                              |                                       |                               |                             | Average=                                 | <b>609.62</b>                                    |                                    | Average= <b>609.75</b>  |                                     |   |
| <b>Change Flow Rate</b>           |     |                     | <b>Stabilization 3 minutes</b>   |                              |                                       |                               |                             |  |  |                                    |                         |                                     |   |
| 1250                              | 1   | 08.05.2017          | 23.6                             | 1001.8                       | 39.5                                  | 24.13                         | 1004.0                      | 1242.11                                  | 0.28   | 10                                 | 1                       | 1239.93                             | 0.93  |
| 1250                              | 2   |                     |                                  |                              |                                       | 24.19                         | 1004.0                      | 1234.56                                  | 0.13   | 10                                 | 2                       | 1237.59                             | 0.93  |
| 1250                              | 3   |                     |                                  |                              |                                       | 24.22                         | 1003.9                      | 1254.99                                  | 0.53   | 10                                 | 3                       | 1248.75                             | 0.94  |
| 1250                              | 4   |                     |                                  |                              |                                       | 24.26                         | 1004.0                      | 1242.49                                  | 0.10   | 10                                 | 4                       | 1247.31                             | 0.94  |
|                                   |     |                     |                                  |                              |                                       |                               |                             | Average=                                 | <b>1243.54</b>                                   |                                    | Average= <b>1243.40</b> |                                     |   |

**INITIAL TESTS**

Compare BIOS temperature and pressure readings with calibrated temperature and ambient pressure instrument readings.

| Laboratory temperature (°C)                    | Laboratory pressure (kPa) | BIOS temperature (°C) | BIOS pressure (kPa) |
|--|---------------------------|-----------------------|---------------------|
| Stabilization 24 hours in laboratory (no flow) |                           |                       |                     |
| 23   | 1013                      | 22.9                  | 1013.5              |
| 12.05.2017                                     |                           |                       |                     |

**COMPARISON MEASUREMENTS**

Flow cell: -44 (S/N: 135198)

|                                |     |                  | ENVIRONMENTAL PARAMETERS         |                           |                                    | TRANSFER STANDARD PARAMETERS |                          |                                       |   |                                    | REFERENCE PARAMETERS |                                  |  |
|--------------------------------|-----|------------------|----------------------------------|---------------------------|------------------------------------|------------------------------|--------------------------|---------------------------------------|---|------------------------------------|----------------------|----------------------------------|--|
| Nominal Flow Rate (std ml/min) | RUN | Date (dd.mm.yyy) | Laboratory temperature (°C)      | Laboratory pressure (kPa) | Laboratory relative humidity (%rh) | BIOS mean temperature (°C)   | BIOS mean pressure (kPa) | Mean Indicated Flow Rate (std ml/min) | Standard deviation of indication (std ml/min) | Number of flow indications in mean | RUN                  | Reference Flow Rate (std ml/min) | Standard uncertainty of reference (std ml/min) |
| <b>Set Flow Rate</b>           |     |                  | <b>Stabilization 180 minutes</b> |                           |                                    |                              |                          |                                       |   |                                    |                      |                                  |  |
| 1250                           | 1   | 12.05.2017       | 23.4                             | 1012.8                    | 54                                 | 24.04                        | 1015.4                   | 1255.54                               | 1.08  | 10                                 | 1                    | 1255.27                          | 0.94   |
| 1250                           | 2   |                  |                                  |                           |                                    | 24.06                        | 1015.2                   | 1247.81                               | 1.01  | 10                                 | 2                    | 1245.48                          | 0.93   |
| 1250                           | 3   |                  |                                  |                           |                                    | 24.04                        | 1015.4                   | 1250.08                               | 0.78  | 10                                 | 3                    | 1248.15                          | 0.94   |
| 1250                           | 4   |                  |                                  |                           |                                    | 24.04                        | 1015.5                   | 1244.80                               | 1.55  | 10                                 | 4                    | 1244.41                          | 0.93   |
| Average=                       |     |                  |                                  |                           |                                    |                              |                          | <b>1249.56</b>                        |   |                                    | Average=             |                                  | <b>1248.33</b>                                 |
| <b>Change Flow Rate</b>        |     |                  | <b>Stabilization 3 minutes</b>   |                           |                                    |                              |                          |                                       |   |                                    |                      |                                  |  |
| 5000                           | 1   | 12.05.2017       | 23.4                             | 1012.8                    | 54                                 | 23.07                        | 1015.4                   | 5071.6                                | 1.1   | 10                                 | 1                    | 5071.7                           | 3.8  |
| 5000                           | 2   |                  |                                  |                           |                                    | 23.05                        | 1015.5                   | 5066.5                                | 1.1   | 10                                 | 2                    | 5066.1                           | 3.8  |
| 5000                           | 3   |                  |                                  |                           |                                    | 23.1                         | 1015.2                   | 5069.4                                | 1.1   | 10                                 | 3                    | 5073.2                           | 3.8  |
| 5000                           | 4   |                  |                                  |                           |                                    | 23.12                        | 1015                     | 5043.3                                | 3.5   | 10                                 | 4                    | 5046.6                           | 3.8  |
| Average=                       |     |                  |                                  |                           |                                    |                              |                          | <b>5062.7</b>                         |   |                                    | Average=             |                                  | <b>5064.4</b>                                  |
| <b>Change Flow Rate</b>        |     |                  | <b>Stabilization 3 minutes</b>   |                           |                                    |                              |                          |                                       |   |                                    |                      |                                  |  |
| <b>Change Flow Rate</b>        |     |                  | <b>Stabilization 3 minutes</b>   |                           |                                    |                              |                          |                                       |   |                                    |                      |                                  |  |
| 20000                          | 1   | 12.05.2017       | 23.4                             | 1012.8                    | 54                                 | 22.65                        | 1017.3                   | 19976                                 | 5   | 10                                 | 1                    | 19959                            | 20   |
| 20000                          | 2   |                  |                                  |                           |                                    | 22.65                        | 1017.2                   | 19858                                 | 7   | 10                                 | 2                    | 19888                            | 20   |
| 20000                          | 3   |                  |                                  |                           |                                    | 22.71                        | 1016.6                   | 20041                                 | 7   | 10                                 | 3                    | 20061                            | 20   |
| 20000                          | 4   |                  |                                  |                           |                                    | 22.67                        | 1016.7                   | 20225                                 | 7   | 10                                 | 4                    | 20238                            | 20   |
| 20000                          | 5   |                  |                                  |                           |                                    | 22.72                        | 1016.8                   | 20354                                 | 9   | 10                                 | 5                    | 20382                            | 20   |
| 20000                          | 6   |                  |                                  |                           |                                    | 22.72                        | 1017                     | 20588                                 | 9   | 10                                 | 6                    | 20614                            | 21   |
| Average=                       |     |                  |                                  |                           |                                    |                              |                          | <b>20174</b>                          |   |                                    | Average=             |                                  | <b>20190</b>                                   |
| <b>Change Flow Rate</b>        |     |                  | <b>Stabilization 3 minutes</b>   |                           |                                    |                              |                          |                                       |   |                                    |                      |                                  |  |
| 30000                          | 1   | 11.05.2017       | 23.4                             | 1012.8                    | 54                                 | 23.01                        | 1020                     | 29943                                 | 17  | 10                                 | 1                    | 29992                            | 30   |
| 30000                          | 2   |                  |                                  |                           |                                    | 23.02                        | 1020.2                   | 30139                                 | 15  | 10                                 | 2                    | 30141                            | 30   |
| 30000                          | 3   |                  |                                  |                           |                                    | 23.02                        | 1020.2                   | 29602                                 | 20  | 10                                 | 3                    | 29696                            | 30   |
| 30000                          | 4   |                  |                                  |                           |                                    | 22.91                        | 1020.3                   | 30391                                 | 12  | 10                                 | 4                    | 30401                            | 30   |
| 30000                          | 5   |                  |                                  |                           |                                    | 23.03                        | 1020.4                   | 30046                                 | 11  | 10                                 | 5                    | 30039                            | 30   |
| 30000                          | 6   |                  |                                  |                           |                                    | 23.06                        | 1020.4                   | 30234                                 | 12  | 10                                 | 6                    | 30244                            | 30   |
| Average=                       |     |                  |                                  |                           |                                    |                              |                          | <b>30059</b>                          |   |                                    | Average=             |                                  | <b>30085</b>                                   |
| <b>End of measurements</b>     |     |                  |                                  |                           |                                    |                              |                          |                                       |   |                                    |                      |                                  |  |

**UL results (loop2)**

**INITIAL TESTS**

Compare BIOS temperature and pressure readings with calibrated temperature and ambient pressure instrument readings.

| Laboratory temperature (°C)                    | Laboratory pressure (kPa) | BIOS temperature (°C) | BIOS pressure (kPa) |
|--|---------------------------|-----------------------|---------------------|
| Stabilization 24 hours in laboratory (no flow) |                           |                       |                     |
| 22.50  | 98.895                    | 22.72                 | 98.925              |

**COMPARISON MEASUREMENTS**

**Flow cell: -10 (S/N: 135208)**

| ENVIRONMENTAL PARAMETERS       |     |                  | TRANSFER STANDARD PARAMETERS     |                           |                                    |                            |                          | REFERENCE PARAMETERS                  |   |                                    |          |                                  |  |  |
|--------------------------------|-----|------------------|----------------------------------|---------------------------|------------------------------------|----------------------------|--------------------------|---------------------------------------|---|------------------------------------|----------|----------------------------------|--|--|
| Nominal Flow Rate (std ml/min) | RUN | Date (dd.mm.yyy) | Laboratory temperature (°C)      | Laboratory pressure (kPa) | Laboratory relative humidity (%rh) | BIOS mean temperature (°C) | BIOS mean pressure (kPa) | Mean Indicated Flow Rate (std ml/min) | Standard deviation of indication (std ml/min) | Number of flow indications in mean | RUN      | Reference Flow Rate (std ml/min) | Standard uncertainty of reference (std ml/min) |  |
| <b>Set Flow Rate</b>           |     |                  | <b>Stabilization 180 minutes</b> |                           |                                    |                            |                          |                                       |   |                                    |          |                                  |  |  |
| 5                              | 1   | 18.11.2016       | 22.8                             | 98.051                    | 38                                 | 23.49                      | 98.193                   | 5.0417                                | 0.0019  | 10                                 | 1        | 5.0403                           | 0.0083   |  |
| 5                              | 2   | 18.11.2016       | 22.8                             | 98.041                    | 38                                 | 23.44                      | 98.174                   | 5.0431                                | 0.0008  | 10                                 | 2        | 5.0473                           | 0.0083   |  |
| 5                              | 3   | 18.11.2016       | 22.8                             | 97.998                    | 38                                 | 23.64                      | 98.152                   | 5.0400                                | 0.0016  | 10                                 | 3        | 5.0484                           | 0.0083   |  |
| 5                              | 4   | 18.11.2016       | 22.8                             | 97.996                    | 38                                 | 23.90                      | 98.142                   | 5.0354                                | 0.0015  | 10                                 | 4        | 5.0423                           | 0.0083   |  |
| Average=                       |     |                  |                                  |                           |                                    |                            |                          | <b>5.0401</b>                         |   |                                    | Average= |                                  | <b>5.0446</b>                                  |  |
| <b>Change Flow Rate</b>        |     |                  | <b>Stabilization 3 minutes</b>   |                           |                                    |                            |                          |                                       |   |                                    |          |                                  |  |  |
| 10                             | 1   | 21.11.2016       | 22.6                             | 98.248                    | 37                                 | 23.32                      | 98.400                   | 9.983                                 | 0.002   | 10                                 | 1        | 9.975                            | 0.010  |  |
| 10                             | 2   | 21.11.2016       | 22.6                             | 98.248                    | 37                                 | 23.33                      | 98.400                   | 9.985                                 | 0.003   | 10                                 | 2        | 9.978                            | 0.010  |  |
| 10                             | 3   | 21.11.2016       | 22.6                             | 98.266                    | 37                                 | 23.33                      | 98.436                   | 9.986                                 | 0.003   | 10                                 | 3        | 9.980                            | 0.010  |  |
| 10                             | 4   | 21.11.2016       | 22.6                             | 98.290                    | 37                                 | 23.34                      | 98.443                   | 9.984                                 | 0.003   | 10                                 | 4        | 9.985                            | 0.010  |  |
| Average=                       |     |                  |                                  |                           |                                    |                            |                          | <b>9.984</b>                          |   |                                    | Average= |                                  | <b>9.979</b>                                   |  |
| <b>Change Flow Rate</b>        |     |                  | <b>Stabilization 3 minutes</b>   |                           |                                    |                            |                          |                                       |   |                                    |          |                                  |  |  |
| 20                             | 1   | 21.11.2016       | 22.3                             | 98.192                    | 40                                 | 23.13                      | 98.337                   | 20.005                                | 0.005   | 10                                 | 1        | 19.993                           | 0.016  |  |
| 20                             | 2   | 21.11.2016       | 22.3                             | 98.200                    | 40                                 | 23.13                      | 98.335                   | 19.999                                | 0.006   | 10                                 | 2        | 19.998                           | 0.016  |  |
| 20                             | 3   | 21.11.2016       | 22.3                             | 98.179                    | 40                                 | 23.12                      | 98.326                   | 19.998                                | 0.005   | 10                                 | 3        | 20.001                           | 0.016  |  |
| 20                             | 4   | 21.11.2016       | 22.3                             | 98.186                    | 40                                 | 23.13                      | 98.334                   | 20.005                                | 0.004   | 10                                 | 4        | 19.999                           | 0.016  |  |
| Average=                       |     |                  |                                  |                           |                                    |                            |                          | <b>20.002</b>                         |   |                                    | Average= |                                  | <b>19.998</b>                                  |  |
| <b>Change Flow Rate</b>        |     |                  | <b>Stabilization 3 minutes</b>   |                           |                                    |                            |                          |                                       |   |                                    |          |                                  |  |  |
| 80                             | 1   | 21.11.2016       | 22.5                             | 98.247                    | 40                                 | 23.32                      | 98.388                   | 80.024                                | 0.009   | 10                                 | 1        | 79.877                           | 0.060  |  |
| 80                             | 2   | 21.11.2016       | 22.5                             | 98.236                    | 40                                 | 23.33                      | 98.380                   | 80.026                                | 0.012   | 10                                 | 2        | 79.877                           | 0.060  |  |
| 80                             | 3   | 21.11.2016       | 22.5                             | 98.229                    | 40                                 | 23.33                      | 98.374                   | 80.017                                | 0.011   | 10                                 | 3        | 79.871                           | 0.060  |  |
| 80                             | 4   | 21.11.2016       | 22.5                             | 98.227                    | 40                                 | 23.32                      | 98.378                   | 80.001                                | 0.015   | 10                                 | 4        | 79.857                           | 0.060  |  |
| Average=                       |     |                  |                                  |                           |                                    |                            |                          | <b>80.017</b>                         |   |                                    | Average= |                                  | <b>79.871</b>                                  |  |



**INITIAL TESTS**

Compare BIOS temperature and pressure readings with calibrated temperature and ambient pressure instrument readings.

| Laboratory temperature<br>(°C)                 | Laboratory pressure<br>(kPa) | BIOS temperature<br>(°C) | BIOS pressure<br>(kPa) |
|--|------------------------------|--------------------------|------------------------|
| Stabilization 24 hours in laboratory (no flow) |                              |                          |                        |
| 22.51  | 98.895                       | 22.64                    | 98.925                 |

**COMPARISON MEASUREMENTS**

*Flow cell: -24 (S/N: 134910)*

|                                   |     |                     | ENVIRONMENTAL PARAMETERS         |                              |                                       | TRANSFER STANDARD PARAMETERS  |                             |  |  |                                    | REFERENCE PARAMETERS |                                     |   |
|-----------------------------------|-----|---------------------|----------------------------------|------------------------------|---------------------------------------|-------------------------------|-----------------------------|--|--|------------------------------------|----------------------|-------------------------------------|---|
| Nominal Flow Rate<br>(std ml/min) | RUN | Date<br>(dd.mm.yyy) | Laboratory temperature<br>(°C)   | Laboratory pressure<br>(kPa) | Laboratory relative humidity<br>(%rh) | BIOS mean temperature<br>(°C) | BIOS mean pressure<br>(kPa) | Mean Indicated Flow Rate<br>(std ml/min) | Standard deviation of indication<br>(std ml/min) | Number of flow indications in mean | RUN                  | Reference Flow Rate<br>(std ml/min) | Standard uncertainty of reference<br>(std ml/min) |
| <b>Set Flow Rate</b>              |     |                     | <b>Stabilization 180 minutes</b> |                              |                                       |                               |                             |  |  |                                    |                      |                                     |   |
| 80                                | 1   | 25.11.2016          | 22.9                             | 98.172                       | 45                                    | 23.79                         | 98.310                      | 79.864                                   | 0.009  | 10                                 | 1                    | 79.963                              | 0.060   |
| 80                                | 2   | 25.11.2016          | 22.9                             | 98.166                       | 45                                    | 23.82                         | 98.291                      | 79.868                                   | 0.011  | 10                                 | 2                    | 79.964                              | 0.060   |
| 80                                | 3   | 25.11.2016          | 22.9                             | 98.143                       | 45                                    | 23.84                         | 98.279                      | 79.857                                   | 0.008  | 10                                 | 3                    | 79.963                              | 0.060   |
| 80                                | 4   | 25.11.2016          | 22.9                             | 98.129                       | 45                                    | 23.84                         | 98.271                      | 79.858                                   | 0.010  | 10                                 | 4                    | 79.948                              | 0.060   |
|                                   |     |                     |                                  |                              |                                       |                               |                             | Average=                                 | <b>79.862</b>                                    |                                    | Average=             | <b>79.959</b>                       |   |
| <b>Change Flow Rate</b>           |     |                     | <b>Stabilization 3 minutes</b>   |                              |                                       |                               |                             |  |  |                                    |                      |                                     |   |
| 300                               | 1   | 24.11.2016          | 22.7                             | 98.350                       | 50                                    | 23.51                         | 98.498                      | 299.86                                   | 0.04   | 10                                 | 1                    | 300.03                              | 0.23  |
| 300                               | 2   | 24.11.2016          | 22.7                             | 98.354                       | 50                                    | 23.53                         | 98.492                      | 299.82                                   | 0.02   | 10                                 | 2                    | 300.01                              | 0.23  |
| 300                               | 3   | 24.11.2016          | 22.7                             | 98.344                       | 50                                    | 23.53                         | 98.488                      | 299.84                                   | 0.02   | 10                                 | 3                    | 300.02                              | 0.23  |
| 300                               | 4   | 24.11.2016          | 22.7                             | 98.341                       | 50                                    | 23.54                         | 98.486                      | 299.85                                   | 0.02   | 10                                 | 4                    | 299.99                              | 0.22  |
|                                   |     |                     |                                  |                              |                                       |                               |                             | Average=                                 | <b>299.84</b>                                    |                                    | Average=             | <b>300.01</b>                       |   |
| <b>Change Flow Rate</b>           |     |                     | <b>Stabilization 3 minutes</b>   |                              |                                       |                               |                             |  |  |                                    |                      |                                     |   |
| 600                               | 1   | 24.11.2016          | 22.8                             | 98.453                       | 46                                    | 23.59                         | 98.591                      | 599.66                                   | 0.05   | 10                                 | 1                    | 599.83                              | 0.45  |
| 600                               | 2   | 24.11.2016          | 22.8                             | 98.447                       | 46                                    | 23.60                         | 98.587                      | 599.65                                   | 0.05   | 10                                 | 2                    | 599.83                              | 0.45  |
| 600                               | 3   | 24.11.2016          | 22.8                             | 98.440                       | 46                                    | 23.59                         | 98.580                      | 599.71                                   | 0.03   | 10                                 | 3                    | 599.85                              | 0.45  |
| 600                               | 4   | 24.11.2016          | 22.8                             | 98.438                       | 46                                    | 23.59                         | 98.579                      | 599.71                                   | 0.04   | 10                                 | 4                    | 599.89                              | 0.45  |
|                                   |     |                     |                                  |                              |                                       |                               |                             | Average=                                 | <b>599.68</b>                                    |                                    | Average=             | <b>599.85</b>                       |   |
| <b>Change Flow Rate</b>           |     |                     | <b>Stabilization 3 minutes</b>   |                              |                                       |                               |                             |  |  |                                    |                      |                                     |   |
| 1250                              | 1   | 24.11.2016          | 23.0                             | 98.487                       | 44                                    | 23.53                         | 98.646                      | 1250.03                                  | 0.14   | 10                                 | 1                    | 1250.56                             | 0.94  |
| 1250                              | 2   | 24.11.2016          | 23.0                             | 98.481                       | 44                                    | 23.53                         | 98.644                      | 1250.06                                  | 0.14   | 10                                 | 2                    | 1250.54                             | 0.94  |
| 1250                              | 3   | 24.11.2016          | 23.0                             | 98.482                       | 44                                    | 23.52                         | 98.647                      | 1250.06                                  | 0.12   | 10                                 | 3                    | 1250.59                             | 0.94  |
| 1250                              | 4   | 24.11.2016          | 23.0                             | 98.482                       | 44                                    | 23.52                         | 98.643                      | 1250.15                                  | 0.10   | 10                                 | 4                    | 1250.62                             | 0.94  |
|                                   |     |                     |                                  |                              |                                       |                               |                             | Average=                                 | <b>1250.08</b>                                   |                                    | Average=             | <b>1250.58</b>                      |   |

**INITIAL TESTS**

Compare BIOS temperature and pressure readings with calibrated temperature and ambient pressure instrument readings.

| Laboratory temperature<br>(°C)                 | Laboratory pressure<br>(kPa) | BIOS temperature<br>(°C) | BIOS pressure<br>(kPa) |
|--|------------------------------|--------------------------|------------------------|
| Stabilization 24 hours in laboratory (no flow) |                              |                          |                        |
| 22.46  | 98.895                       | 22.62                    | 98.925                 |

**COMPARISON MEASUREMENTS**

Flow cell: -44 (S/N: 135199)

|                                   |     |                     | ENVIRONMENTAL PARAMETERS         |                              |                                       | TRANSFER STANDARD PARAMETERS  |                             |  |  |                                    | REFERENCE PARAMETERS    |                                     |   |
|-----------------------------------|-----|---------------------|----------------------------------|------------------------------|---------------------------------------|-------------------------------|-----------------------------|--|--|------------------------------------|-------------------------|-------------------------------------|---|
| Nominal Flow Rate<br>(std ml/min) | RUN | Date<br>(dd.mm.yyy) | Laboratory temperature<br>(°C)   | Laboratory pressure<br>(kPa) | Laboratory relative humidity<br>(%rh) | BIOS mean temperature<br>(°C) | BIOS mean pressure<br>(kPa) | Mean Indicated Flow Rate<br>(std ml/min) | Standard deviation of indication<br>(std ml/min) | Number of flow indications in mean | RUN                     | Reference Flow Rate<br>(std ml/min) | Standard uncertainty of reference<br>(std ml/min) |
| <b>Set Flow Rate</b>              |     |                     | <b>Stabilization 180 minutes</b> |                              |                                       |                               |                             |  |  |                                    |                         |                                     |   |
| 1250                              | 1   | 23.11.2016          | 22.6                             | 98.572                       | 40                                    | 23.59                         | 98.772                      | 1250.26                                  | 0.15   | 10                                 | 1                       | 1249.60                             | 0.94  |
| 1250                              | 2   | 23.11.2016          | 22.6                             | 98.576                       | 40                                    | 23.62                         | 98.772                      | 1250.07                                  | 0.11   | 10                                 | 2                       | 1249.55                             | 0.94  |
| 1250                              | 3   | 23.11.2016          | 22.6                             | 98.581                       | 40                                    | 23.62                         | 98.778                      | 1250.22                                  | 0.17   | 10                                 | 3                       | 1249.59                             | 0.94  |
| 1250                              | 4   | 23.11.2016          | 22.6                             | 98.583                       | 40                                    | 23.63                         | 98.775                      | 1250.25                                  | 0.22   | 10                                 | 4                       | 1249.59                             | 0.94  |
|                                   |     |                     |                                  |                              |                                       |                               |                             | Average=                                 | <b>1250.20</b>                                   |                                    | Average= <b>1249.58</b> |                                     |   |
| <b>Change Flow Rate</b>           |     |                     | <b>Stabilization 3 minutes</b>   |                              |                                       |                               |                             |  |  |                                    |                         |                                     |   |
| 5000                              | 1   | 23.11.2016          | 22.6                             | 98.607                       | 40                                    | 22.86                         | 98.821                      | 4996.1                                   | 1.0  | 10                                 | 1                       | 4995.8                              | 3.7   |
| 5000                              | 2   | 23.11.2016          | 22.6                             | 98.599                       | 40                                    | 22.84                         | 98.820                      | 4996.9                                   | 0.8  | 10                                 | 2                       | 4996.1                              | 3.7   |
| 5000                              | 3   | 23.11.2016          | 22.6                             | 98.597                       | 40                                    | 22.84                         | 98.824                      | 4997.1                                   | 1.3  | 10                                 | 3                       | 4996.3                              | 3.7   |
| 5000                              | 4   | 23.11.2016          | 22.6                             | 98.595                       | 40                                    | 22.82                         | 98.814                      | 4997.4                                   | 0.4  | 10                                 | 4                       | 4996.5                              | 3.7   |
|                                   |     |                     |                                  |                              |                                       |                               |                             | Average=                                 | <b>4996.9</b>                                    |                                    | Average= <b>4996.2</b>  |                                     |   |
| <b>Change Flow Rate</b>           |     |                     | <b>Stabilization 3 minutes</b>   |                              |                                       |                               |                             |  |  |                                    |                         |                                     |   |
| 10000                             | 1   | 23.11.2016          | 22.6                             | 98.665                       | 40                                    | 23.01                         | 98.950                      | 10005.4                                  | 1.1  | 10                                 | 1                       | 10004.3                             | 7.5   |
| 10000                             | 2   | 23.11.2016          | 22.6                             | 98.666                       | 40                                    | 23.02                         | 98.949                      | 10006.6                                  | 1.7  | 10                                 | 2                       | 10005.4                             | 7.5   |
| 10000                             | 3   | 23.11.2016          | 22.6                             | 98.666                       | 40                                    | 23.03                         | 98.947                      | 10005.9                                  | 1.0  | 10                                 | 3                       | 10005.0                             | 7.5   |
| 10000                             | 4   | 23.11.2016          | 22.6                             | 98.663                       | 40                                    | 23.03                         | 98.945                      | 10005.2                                  | 1.7  | 10                                 | 4                       | 10004.5                             | 7.5   |
|                                   |     |                     |                                  |                              |                                       |                               |                             | Average=                                 | <b>10005.8</b>                                   |                                    | Average= <b>10004.8</b> |                                     |   |
| <b>Change Flow Rate</b>           |     |                     | <b>Stabilization 3 minutes</b>   |                              |                                       |                               |                             |  |  |                                    |                         |                                     |   |
| 20000                             | 1   | 23.11.2016          | 22.4                             | 98.692                       | 39                                    | 23.10                         | 99.055                      | 20001                                    | 6  | 10                                 | 1                       | 20000                               | 15  |
| 20000                             | 2   | 23.11.2016          | 22.4                             | 98.689                       | 39                                    | 23.09                         | 99.055                      | 20005                                    | 5  | 10                                 | 2                       | 20000                               | 15  |
| 20000                             | 3   | 23.11.2016          | 22.4                             | 98.687                       | 39                                    | 23.09                         | 99.062                      | 20005                                    | 5  | 10                                 | 3                       | 20000                               | 15  |
| 20000                             | 4   | 23.11.2016          | 22.4                             | 98.685                       | 39                                    | 23.07                         | 99.057                      | 20006                                    | 4  | 10                                 | 4                       | 20001                               | 15  |
|                                   |     |                     |                                  |                              |                                       |                               |                             | Average=                                 | <b>20004</b>                                     |                                    | Average= <b>20000</b>   |                                     |   |
| <b>Change Flow Rate</b>           |     |                     | <b>Stabilization 3 minutes</b>   |                              |                                       |                               |                             |  |  |                                    |                         |                                     |   |
| 30000                             | 1   | 23.11.2016          | 22.6                             | 98.695                       | 40                                    | 23.05                         | 99.193                      | 30024                                    | 7  | 10                                 | 1                       | 30023                               | 23  |
| 30000                             | 2   | 23.11.2016          | 22.6                             | 98.679                       | 40                                    | 23.02                         | 99.186                      | 30028                                    | 14   | 10                                 | 2                       | 30023                               | 23  |
| 30000                             | 3   | 23.11.2016          | 22.6                             | 98.676                       | 40                                    | 23.01                         | 99.190                      | 30032                                    | 15   | 10                                 | 3                       | 30024                               | 23  |
| 30000                             | 4   | 23.11.2016          | 22.6                             | 98.673                       | 40                                    | 23.00                         | 99.189                      | 30031                                    | 14   | 10                                 | 4                       | 30028                               | 23  |
|                                   |     |                     |                                  |                              |                                       |                               |                             | Average=                                 | <b>30029</b>                                     |                                    | Average= <b>30025</b>   |                                     |   |
| <b>End of measurements</b>        |     |                     |                                  |                              |                                       |                               |                             |  |  |                                    |                         |                                     |   |

**CMI results (loop2)**

**INITIAL TESTS**

Compare BIOS temperature and pressure readings with calibrated temperature and ambient pressure instrument readings.

| Laboratory temperature<br>(°C)                 | Laboratory pressure<br>(kPa) | BIOS temperature<br>(°C) | BIOS pressure<br>(kPa) |
|--|------------------------------|--------------------------|------------------------|
| Stabilization 24 hours in laboratory (no flow) |                              |                          |                        |
| 20.76  | 97.42                        | 22.07                    | 97.45                  |

**COMPARISON MEASUREMENTS**

**Flow cell: -10 (S/N: 135208)**

| Flow cell: -10 (S/N: 135208)      |     |                     | ENVIRONMENTAL PARAMETERS         |                              |                                       | TRANSFER STANDARD PARAMETERS  |                             |  |  |                                    | REFERENCE PARAMETERS |                                     |   |
|-----------------------------------|-----|---------------------|----------------------------------|------------------------------|---------------------------------------|-------------------------------|-----------------------------|--|--|------------------------------------|----------------------|-------------------------------------|---|
| Nominal Flow Rate<br>(std ml/min) | RUN | Date<br>(dd.mm.yyy) | Laboratory temperature<br>(°C)   | Laboratory pressure<br>(kPa) | Laboratory relative humidity<br>(%rh) | BIOS mean temperature<br>(°C) | BIOS mean pressure<br>(kPa) | Mean Indicated Flow Rate<br>(std ml/min) | Standard deviation of indication<br>(std ml/min) | Number of flow indications in mean | RUN                  | Reference Flow Rate<br>(std ml/min) | Standard uncertainty of reference<br>(std ml/min) |
| <b>Set Flow Rate</b>              |     |                     | <b>Stabilization 180 minutes</b> |                              |                                       |                               |                             |  |  |                                    |                      |                                     |   |
| 5                                 | 1   | 17.01.2017          | 20.46                            | 99.016                       | 53                                    | 21.59                         | 99.155                      | 5.0045                                   | 0.0002   | 10                                 | 1                    | 5.0000                              | 0.0050  |
| 5                                 | 2   | 17.01.2017          | 20.45                            | 99.018                       | 53                                    | 21.61                         | 99.159                      | 5.0043                                   | 0.0002   | 10                                 | 2                    | 5.0000                              | 0.0050  |
| 5                                 | 3   | 17.01.2017          | 20.48                            | 98.024                       | 53                                    | 21.61                         | 99.164                      | 5.0040                                   | 0.0002   | 10                                 | 3                    | 5.0000                              | 0.0050  |
| 5                                 | 4   | 17.01.2017          | 20.51                            | 99.023                       | 53                                    | 21.63                         | 99.167                      | 5.0045                                   | 0.0002   | 10                                 | 4                    | 4.9999                              | 0.0050  |
|                                   |     |                     |                                  |                              |                                       |                               |                             | Average=                                 | <b>5.0043</b>                                    |                                    | Average=             | <b>5.0000</b>                       |   |
| <b>Change Flow Rate</b>           |     |                     | <b>Stabilization 3 minutes</b>   |                              |                                       |                               |                             |  |  |                                    |                      |                                     |   |
| 10                                | 1   | 09.01.2017          | 20.76                            | 98.536                       | 21                                    | 22.11                         | 98.699                      | 10.0055                                  | 0.0006   | 10                                 | 1                    | 10.0007                             | 0.0050  |
| 10                                | 2   | 09.01.2017          | 20.82                            | 98.532                       | 21                                    | 22.09                         | 98.697                      | 10.0062                                  | 0.0009   | 10                                 | 2                    | 10.0007                             | 0.0050  |
| 10                                | 3   | 09.01.2017          | 20.83                            | 98.533                       | 21                                    | 22.09                         | 98.7                        | 10.0056                                  | 0.0018   | 10                                 | 3                    | 10.0007                             | 0.0050  |
| 10                                | 4   | 09.01.2017          | 20.84                            | 98.532                       | 21                                    | 22.09                         | 98.701                      | 10.0068                                  | 0.0012   | 10                                 | 4                    | 10.0007                             | 0.0050  |
|                                   |     |                     |                                  |                              |                                       |                               |                             | Average=                                 | <b>10.0060</b>                                   |                                    | Average=             | <b>10.0007</b>                      |   |
| <b>Change Flow Rate</b>           |     |                     | <b>Stabilization 3 minutes</b>   |                              |                                       |                               |                             |  |  |                                    |                      |                                     |   |
| 20                                | 1   | 13.1.2017           | 20.71                            | 95.833                       | 24                                    | 22.07                         | 95.986                      | 20.004                                   | 0.001  | 10                                 | 1                    | 19.999                              | 0.010   |
| 20                                | 2   | 13.1.2017           | 20.72                            | 95.839                       | 24                                    | 22.07                         | 95.985                      | 20.008                                   | 0.002  | 10                                 | 2                    | 19.999                              | 0.010   |
| 20                                | 3   | 13.1.2017           | 20.72                            | 95.841                       | 24                                    | 22.09                         | 95.987                      | 20.009                                   | 0.001  | 10                                 | 3                    | 19.999                              | 0.010   |
| 20                                | 4   | 13.1.2017           | 20.66                            | 95.828                       | 25                                    | 22.09                         | 95.975                      | 20.009                                   | 0.001  | 10                                 | 4                    | 19.999                              | 0.010   |
|                                   |     |                     |                                  |                              |                                       |                               |                             | Average=                                 | <b>20.008</b>                                    |                                    | Average=             | <b>19.999</b>                       |   |
| <b>Change Flow Rate</b>           |     |                     | <b>Stabilization 3 minutes</b>   |                              |                                       |                               |                             |  |  |                                    |                      |                                     |   |
| 80                                | 1   | 18.01.2017          | 20.52                            | 99.636                       | 57.7                                  | 21.87                         | 99.79                       | 80.164                                   | 0.003  | 10                                 | 1                    | 80.012                              | 0.040   |
| 80                                | 2   | 18.01.2017          | 20.54                            | 99.632                       | 58.7                                  | 21.87                         | 99.787                      | 80.163                                   | 0.003  | 10                                 | 2                    | 80.012                              | 0.040   |
| 80                                | 3   | 18.01.2017          | 20.52                            | 99.633                       | 58.7                                  | 21.87                         | 99.787                      | 80.160                                   | 0.002  | 10                                 | 3                    | 80.012                              | 0.040   |
| 80                                | 4   | 18.01.2017          | 20.54                            | 99.622                       | 58.9                                  | 21.89                         | 99.776                      | 80.160                                   | 0.002  | 10                                 | 4                    | 80.012                              | 0.040   |
|                                   |     |                     |                                  |                              |                                       |                               |                             | Average=                                 | <b>80.162</b>                                    |                                    | Average=             | <b>80.012</b>                       |   |

**INITIAL TESTS**

Compare BIOS temperature and pressure readings with calibrated temperature and ambient pressure instrument readings.

| Laboratory temperature<br>(°C)                 | Laboratory pressure<br>(kPa) | BIOS temperature<br>(°C) | BIOS pressure<br>(kPa) |
|--|------------------------------|--------------------------|------------------------|
| Stabilization 24 hours in laboratory (no flow) |                              |                          |                        |
| 20.7   | 99.809                       | 21.89                    | 99.887                 |

**COMPARISON MEASUREMENTS**

Flow cell: -24 (S/N: 134910)

| Flow cell: -24 (S/N: 134910)      |     |                     | ENVIRONMENTAL PARAMETERS         |                              |                                       | TRANSFER STANDARD PARAMETERS  |                             |  |  |                                    | REFERENCE PARAMETERS |                                     |   |
|-----------------------------------|-----|---------------------|----------------------------------|------------------------------|---------------------------------------|-------------------------------|-----------------------------|--|--|------------------------------------|----------------------|-------------------------------------|---|
| Nominal Flow Rate<br>(std ml/min) | RUN | Date<br>(dd.mm.yyy) | Laboratory temperature<br>(°C)   | Laboratory pressure<br>(kPa) | Laboratory relative humidity<br>(%rh) | BIOS mean temperature<br>(°C) | BIOS mean pressure<br>(kPa) | Mean Indicated Flow Rate<br>(std ml/min) | Standard deviation of indication<br>(std ml/min) | Number of flow indications in mean | RUN                  | Reference Flow Rate<br>(std ml/min) | Standard uncertainty of reference<br>(std ml/min) |
| <b>Set Flow Rate</b>              |     |                     | <b>Stabilization 180 minutes</b> |                              |                                       |                               |                             |  |  |                                    |                      |                                     |   |
| 80                                | 1   | 19.01.2017          | 20.57                            | 99.78                        | 60                                    | 21.87                         | 99.926                      | 79.893                                   | 0.005  | 10                                 | 1                    | 80.002                              | 0.040   |
| 80                                | 2   | 19.01.2018          | 20.6                             | 99.76                        | 60                                    | 21.88                         | 99.904                      | 79.896                                   | 0.004  | 10                                 | 2                    | 80.002                              | 0.040   |
| 80                                | 3   | 19.01.2019          | 20.63                            | 99.758                       | 60                                    | 21.89                         | 99.904                      | 79.889                                   | 0.004  | 10                                 | 3                    | 80.004                              | 0.040   |
| 80                                | 4   | 19.01.2020          | 20.63                            | 99.76                        | 60                                    | 21.89                         | 99.904                      | 79.893                                   | 0.002  | 10                                 | 4                    | 80.004                              | 0.040   |
|                                   |     |                     |                                  |                              |                                       |                               |                             | Average=                                 | <b>79.893</b>                                    |                                    | Average=             | <b>80.003</b>                       |   |
| <b>Change Flow Rate</b>           |     |                     | <b>Stabilization 3 minutes</b>   |                              |                                       |                               |                             |  |  |                                    |                      |                                     |   |
| 300                               | 1   | 20.01.2017          | 20.75                            | 99.675                       | 62                                    | 21.63                         | 99.824                      | 300.229                                  | 0.008  | 10                                 | 1                    | 300.171                             | 0.150   |
| 300                               | 2   | 20.01.2017          | 20.7                             | 99.679                       | 62                                    | 21.65                         | 99.826                      | 300.220                                  | 0.009  | 10                                 | 2                    | 300.171                             | 0.150   |
| 300                               | 3   | 20.01.2017          | 20.74                            | 99.679                       | 62                                    | 21.65                         | 99.826                      | 300.198                                  | 0.010  | 10                                 | 3                    | 300.171                             | 0.150   |
| 300                               | 4   | 20.01.2017          | 20.74                            | 99.679                       | 62                                    | 21.65                         | 99.826                      | 300.204                                  | 0.011  | 10                                 | 4                    | 300.171                             | 0.150   |
|                                   |     |                     |                                  |                              |                                       |                               |                             | Average=                                 | <b>300.213</b>                                   |                                    | Average=             | <b>300.171</b>                      |   |
| <b>Change Flow Rate</b>           |     |                     | <b>Stabilization 3 minutes</b>   |                              |                                       |                               |                             |  |  |                                    |                      |                                     |   |
| 600                               | 1   | 23.01.2017          | 20.95                            | 98.794                       | 62                                    | 21.87                         | 98.93                       | 600.71                                   | 0.03   | 10                                 | 1                    | 600.60                              | 0.30  |
| 600                               | 2   | 23.01.2017          | 20.95                            | 98.793                       | 62                                    | 21.89                         | 98.93                       | 600.78                                   | 0.06   | 10                                 | 2                    | 600.60                              | 0.30  |
| 600                               | 3   | 23.01.2017          | 20.95                            | 98.795                       | 62                                    | 21.89                         | 98.932                      | 600.77                                   | 0.10   | 10                                 | 3                    | 600.57                              | 0.30  |
| 600                               | 4   | 23.01.2017          | 20.93                            | 98.796                       | 62                                    | 21.89                         | 98.932                      | 600.68                                   | 0.07   | 10                                 | 4                    | 600.57                              | 0.30  |
|                                   |     |                     |                                  |                              |                                       |                               |                             | Average=                                 | <b>600.73</b>                                    |                                    | Average=             | <b>600.58</b>                       |   |
| <b>Change Flow Rate</b>           |     |                     | <b>Stabilization 3 minutes</b>   |                              |                                       |                               |                             |  |  |                                    |                      |                                     |   |
| 1250                              | 1   | 23.01.2017          | 20.75                            | 98.804                       | 63                                    | 21.99                         | 98.959                      | 1250.59                                  | 0.06   | 10                                 | 1                    | 1251.10                             | 0.63  |
| 1250                              | 2   | 23.01.2017          | 20.76                            | 98.803                       | 63                                    | 21.99                         | 98.954                      | 1250.75                                  | 0.07   | 10                                 | 2                    | 1251.10                             | 0.63  |
| 1250                              | 3   | 23.01.2017          | 20.76                            | 98.803                       | 63                                    | 21.99                         | 98.94                       | 1250.41                                  | 0.10   | 10                                 | 3                    | 1251.10                             | 0.63  |
| 1250                              | 4   | 23.01.2017          | 20.79                            | 98.801                       | 63                                    | 21.99                         | 98.952                      | 1250.28                                  | 0.15   | 10                                 | 4                    | 1251.10                             | 0.63  |
|                                   |     |                     |                                  |                              |                                       |                               |                             | Average=                                 | <b>1250.51</b>                                   |                                    | Average=             | <b>1251.10</b>                      |   |

**INITIAL TESTS**

Compare BIOS temperature and pressure readings with calibrated temperature and ambient pressure instrument readings.

| Laboratory temperature (°C)                    | Laboratory pressure (kPa) | BIOS temperature (°C) | BIOS pressure (kPa) |
|--|---------------------------|-----------------------|---------------------|
| Stabilization 24 hours in laboratory (no flow) |                           |                       |                     |
| 20.81  | 99.09                     | 22.05                 | 99.11               |

**COMPARISON MEASUREMENTS**

Flow cell: -44 (S/N: 135199)

|                                |     |                  | ENVIRONMENTAL PARAMETERS         |                           |                                    | TRANSFER STANDARD PARAMETERS |                          |                                       |   |                                    | REFERENCE PARAMETERS    |                                  |  |
|--------------------------------|-----|------------------|----------------------------------|---------------------------|------------------------------------|------------------------------|--------------------------|---------------------------------------|---|------------------------------------|-------------------------|----------------------------------|--|
| Nominal Flow Rate (std ml/min) | RUN | Date (dd.mm.yyy) | Laboratory temperature (°C)      | Laboratory pressure (kPa) | Laboratory relative humidity (%rh) | BIOS mean temperature (°C)   | BIOS mean pressure (kPa) | Mean Indicated Flow Rate (std ml/min) | Standard deviation of indication (std ml/min) | Number of flow indications in mean | RUN                     | Reference Flow Rate (std ml/min) | Standard uncertainty of reference (std ml/min) |
| <b>Set Flow Rate</b>           |     |                  | <b>Stabilization 180 minutes</b> |                           |                                    |                              |                          |                                       |   |                                    |                         |                                  |  |
| 1250                           | 1   | 23.1.2017        | 20.64                            | 98.636                    | 63                                 | 21.79                        | 98.8                     | 1250.93                               | 0.10  | 10                                 | 1                       | 1250.85                          | 0.63   |
| 1250                           | 2   | 23.1.2018        | 20.63                            | 98.636                    | 63                                 | 21.81                        | 98.81                    | 1251.18                               | 0.14  | 10                                 | 2                       | 1250.85                          | 0.63   |
| 1250                           | 3   | 23.1.2019        | 20.63                            | 98.637                    | 63                                 | 21.81                        | 98.796                   | 1251.02                               | 0.17  | 10                                 | 3                       | 1250.79                          | 0.63   |
| 1250                           | 4   | 23.1.2020        | 20.63                            | 98.636                    | 63                                 | 21.81                        | 98.815                   | 1250.87                               | 0.13  | 10                                 | 4                       | 1250.79                          | 0.63   |
|                                |     |                  |                                  |                           |                                    |                              |                          | Average=                              | <b>1251.00</b>                                |                                    | Average= <b>1250.82</b> |                                  |  |
| <b>Change Flow Rate</b>        |     |                  | <b>Stabilization 3 minutes</b>   |                           |                                    |                              |                          |                                       |   |                                    |                         |                                  |  |
| 5000                           | 1   | 24.1.2017        | 20.68                            | 98.667                    | 63                                 | 22.2                         | 98.87                    | 5002.4                                | 0.5   | 10                                 | 1                       | 5004.7                           | 2.5  |
| 5000                           | 2   | 24.1.2017        | 20.68                            | 98.667                    | 63                                 | 22.21                        | 98.81                    | 5003.4                                | 0.5   | 10                                 | 2                       | 5004.9                           | 2.5  |
| 5000                           | 3   | 24.1.2017        | 20.68                            | 98.669                    | 63                                 | 22.22                        | 98.87                    | 5002.2                                | 0.3   | 10                                 | 3                       | 5005.3                           | 2.5  |
| 5000                           | 4   | 24.1.2017        | 20.68                            | 98.67                     | 63                                 | 22.24                        | 98.87                    | 5002.6                                | 0.2   | 10                                 | 4                       | 5005.3                           | 2.5  |
|                                |     |                  |                                  |                           |                                    |                              |                          | Average=                              | <b>5002.7</b>                                 |                                    | Average= <b>5005.0</b>  |                                  |  |
| <b>Change Flow Rate</b>        |     |                  | <b>Stabilization 3 minutes</b>   |                           |                                    |                              |                          |                                       |   |                                    |                         |                                  |  |
| 10000                          | 1   | 24.1.2017        | 20.62                            | 98.817                    | 62                                 | 21.69                        | 99.058                   | 10001.1                               | 1.2   | 10                                 | 1                       | 10002.0                          | 5.0  |
| 10000                          | 2   | 24.1.2017        | 20.62                            | 98.817                    | 62                                 | 21.69                        | 99.052                   | 9998.4                                | 1.0   | 10                                 | 2                       | 10000.0                          | 5.0  |
| 10000                          | 3   | 24.1.2017        | 20.62                            | 98.819                    | 62                                 | 21.69                        | 99.058                   | 9999.7                                | 0.9   | 10                                 | 3                       | 9999.8                           | 5.0  |
| 10000                          | 4   | 24.1.2017        | 20.62                            | 98.819                    | 62                                 | 21.71                        | 99.054                   | 10001.1                               | 1.1   | 10                                 | 4                       | 9999.8                           | 5.0  |
|                                |     |                  |                                  |                           |                                    |                              |                          | Average=                              | <b>10000.0</b>                                |                                    | Average= <b>10000.4</b> |                                  |  |
| <b>Change Flow Rate</b>        |     |                  | <b>Stabilization 3 minutes</b>   |                           |                                    |                              |                          |                                       |   |                                    |                         |                                  |  |
| 20000                          | 1   | 26.1.2016        | 20.78                            | 99.528                    | 62                                 | 21.77                        | 99.531                   | 20043.4                               | 1.8   | 10                                 | 3                       | 20046.6                          | 12.5   |
| 20000                          | 2   | 26.1.2016        | 20.78                            | 99.528                    | 62                                 | 21.77                        | 99.531                   | 20034.3                               | 2.0   | 10                                 | 4                       | 20037.0                          | 12.5   |
| 20000                          | 3   | 25.1.2017        | 20.81                            | 99.09                     | 63                                 | 22.03                        | 99.436                   | 20005.1                               | 2.5   | 10                                 | 1                       | 20004.6                          | 12.4   |
| 20000                          | 4   | 25.1.2017        | 20.8                             | 99.092                    | 63                                 | 22.01                        | 99.435                   | 19997.8                               | 0.9   | 10                                 | 2                       | 19998.6                          | 12.4   |
|                                |     |                  |                                  |                           |                                    |                              |                          | Average=                              | <b>20020.2</b>                                |                                    | Average= <b>20021.7</b> |                                  |  |
| <b>Change Flow Rate</b>        |     |                  | <b>Stabilization 3 minutes</b>   |                           |                                    |                              |                          |                                       |   |                                    |                         |                                  |  |
| 30000                          | 1   | 25.1.2017        | 20.76                            | 98.057                    | 62                                 | 21.75                        | 99.53                    | 30020                                 | 2   | 10                                 | 1                       | 30009                            | 45   |
| 30000                          | 2   | 25.1.2017        | 20.76                            | 98.057                    | 62                                 | 21.73                        | 99.524                   | 30016                                 | 3   | 10                                 | 2                       | 30006                            | 45   |
| 30000                          | 3   | 25.1.2017        | 20.77                            | 98.057                    | 62                                 | 21.71                        | 99.534                   | 30009                                 | 2   | 10                                 | 3                       | 30000                            | 45   |
| 30000                          | 4   | 25.1.2017        | 20.77                            | 98.057                    | 62                                 | 21.73                        | 99.544                   | 30005                                 | 3   | 10                                 | 4                       | 29996                            | 45   |
|                                |     |                  |                                  |                           |                                    |                              |                          | Average=                              | <b>30013</b>                                  |                                    | Average= <b>30003</b>   |                                  |  |
| <b>End of measurements</b>     |     |                  |                                  |                           |                                    |                              |                          |                                       |   |                                    |                         |                                  |  |

**LNE results (loop2)**

**INITIAL TESTS**

Compare BIOS temperature and pressure readings with calibrated temperature and ambient pressure instrument readings.

| Laboratory temperature<br>(°C)                 | Laboratory pressure<br>(kPa) | BIOS temperature<br>(°C) | BIOS pressure<br>(kPa) |
|--|------------------------------|--------------------------|------------------------|
| Stabilization 24 hours in laboratory (no flow) |                              |                          |                        |
| 21.8   | 100.97                       | 22.05                    | 101.09                 |

**COMPARISON MEASUREMENTS (LOOP2)**

Base: **S/N: 147461**

Flow cell: **-10 (S/N: 135208)**

|                                   |     |                     | ENVIRONMENTAL PARAMETERS         |                              |                                       | TRANSFER STANDARD PARAMETERS  |                             |  |  |                                    | REFERENCE PARAMETERS |                                     |   |
|-----------------------------------|-----|---------------------|----------------------------------|------------------------------|---------------------------------------|-------------------------------|-----------------------------|--|--|------------------------------------|----------------------|-------------------------------------|---|
| Nominal Flow Rate<br>(std ml/min) | RUN | Date<br>(dd.mm.yyy) | Laboratory temperature<br>(°C)   | Laboratory pressure<br>(kPa) | Laboratory relative humidity<br>(%rh) | BIOS mean temperature<br>(°C) | BIOS mean pressure<br>(kPa) | Mean Indicated Flow Rate<br>(std ml/min) | Standard deviation of indication<br>(std ml/min) | Number of flow indications in mean | RUN                  | Reference Flow Rate<br>(std ml/min) | Standard uncertainty of reference<br>(std ml/min) |
| <b>Set Flow Rate</b>              |     |                     | <b>Stabilization 180 minutes</b> |                              |                                       |                               |                             |  |  |                                    |                      |                                     |   |
| 5                                 | 1   | 24.2.2017           | 21.4                             | 100.91                       | 43                                    | 22.34                         | 101.04                      | 5.0180                                   | 0.0015   | 10                                 | 1                    | 5.0138                              | 0.0076  |
| 5                                 | 2   | 27.2.2017           | 21.7                             | 99.08                        | 46                                    | 22.54                         | 99.21                       | 5.0128                                   | 0.0017   | 10                                 | 2                    | 5.0078                              | 0.0079  |
| 5                                 | 3   | 1.3.2017            | 21.0                             | 99.27                        | 45                                    | 22.68                         | 99.39                       | 5.0141                                   | 0.0018   | 10                                 | 3                    | 5.0173                              | 0.0081  |
| 5                                 | 4   | 2.3.2017            | 21.2                             | 100.85                       | 43                                    | 22.48                         | 100.97                      | 5.0180                                   | 0.0014   | 10                                 | 4                    | 5.0121                              | 0.0078  |
|                                   |     |                     |                                  |                              |                                       |                               |                             | Average=                                 | <b>5.0157</b>                                    |                                    | Average=             | <b>5.0128</b>                       |   |
| <b>Change Flow Rate</b>           |     |                     | <b>Stabilization 3 minutes</b>   |                              |                                       |                               |                             |  |  |                                    |                      |                                     |   |
| 10                                | 1   | 24.2.2017           | 21.4                             | 100.91                       | 46                                    | 22.46                         | 101.09                      | 10.022                                   | 0.001  | 10                                 | 1                    | 10.017                              | 0.010   |
| 10                                | 2   | 27.2.2017           | 21.2                             | 99.14                        | 42                                    | 22.44                         | 99.26                       | 10.012                                   | 0.003  | 10                                 | 2                    | 9.997                               | 0.010   |
| 10                                | 3   | 1.3.2017            | 21.5                             | 99.31                        | 43                                    | 22.60                         | 99.41                       | 10.034                                   | 0.001  | 10                                 | 3                    | 10.022                              | 0.010   |
| 10                                | 4   | 2.3.2017            | 20.8                             | 100.85                       | 44                                    | 22.64                         | 100.97                      | 10.016                                   | 0.003  | 10                                 | 4                    | 10.004                              | 0.010   |
|                                   |     |                     |                                  |                              |                                       |                               |                             | Average=                                 | <b>10.021</b>                                    |                                    | Average=             | <b>10.010</b>                       |   |
| <b>Change Flow Rate</b>           |     |                     | <b>Stabilization 3 minutes</b>   |                              |                                       |                               |                             |  |  |                                    |                      |                                     |   |
| 20                                | 1   | 24.2.2017           | 21.4                             | 100.91                       | 43                                    | 22.34                         | 101.04                      | 20.053                                   | 0.003  | 10                                 | 1                    | 20.044                              | 0.014   |
| 20                                | 2   | 27.2.2017           | 21.8                             | 98.79                        | 42                                    | 22.46                         | 98.91                       | 20.048                                   | 0.007  | 10                                 | 2                    | 20.039                              | 0.019   |
| 20                                | 3   | 1.3.2017            | 20.8                             | 99.33                        | 45                                    | 22.52                         | 99.46                       | 20.019                                   | 0.005  | 10                                 | 3                    | 20.017                              | 0.014   |
| 20                                | 4   | 2.3.2017            | 21.4                             | 100.82                       | 43                                    | 22.58                         | 100.94                      | 20.016                                   | 0.003  | 10                                 | 4                    | 20.010                              | 0.014   |
|                                   |     |                     |                                  |                              |                                       |                               |                             | Average=                                 | <b>20.034</b>                                    |                                    | Average=             | <b>20.028</b>                       |   |
| <b>Change Flow Rate</b>           |     |                     | <b>Stabilization 3 minutes</b>   |                              |                                       |                               |                             |  |  |                                    |                      |                                     |   |
| 80                                | 1   | 24.2.2017           | 21.5                             | 100.91                       | 45                                    | 22.48                         | 101.03                      | 80.097                                   | 0.010  | 10                                 | 1                    | 79.885                              | 0.040   |
| 80                                | 2   | 27.2.2017           | 21.5                             | 98.81                        | 41                                    | 22.64                         | 98.92                       | 80.064                                   | 0.017  | 10                                 | 2                    | 79.834                              | 0.042   |
| 80                                | 3   | 1.3.2017            | 21.2                             | 99.33                        | 44                                    | 22.52                         | 99.46                       | 80.078                                   | 0.008  | 10                                 | 3                    | 79.902                              | 0.041   |
| 80                                | 4   | 2.3.2017            | 21.7                             | 100.82                       | 45                                    | 22.68                         | 100.95                      | 80.093                                   | 0.006  | 10                                 | 4                    | 79.895                              | 0.040   |
|                                   |     |                     |                                  |                              |                                       |                               |                             | Average=                                 | <b>80.083</b>                                    |                                    | Average=             | <b>79.879</b>                       |   |