

Revision: Original  
Date: October 16, 2009

**Final Report**  
**Bilateral Comparison Report on Hardness Measurement**  
**Rockwell scale A and B Between PTB and NIMT**

Tassanai Sanponput<sup>1</sup>, Apichaya Meesaplak<sup>2</sup>, Konrad Herrmann<sup>3</sup>, Febo Menelao<sup>4</sup>

<sup>1,2</sup> NIMT(Pilot Institute): National Institute of Metrology(Thailand), 3/5 Moo 3, Klong 5, Klong Luang, Pathumthani Province, 12120, Thailand

<sup>3,4</sup> PTB : Physikalisch-Technische Bundesanstalt, Bundesallee 100, 38116 Braunschweig, Germany

Revision: Original  
Date: October 16, 2009

## Abstract

This bilateral comparison hardness measurement Rockwell scale A and B was arranged by National Institute of Metrology (Thailand), NIMT as a pilot laboratory comparing with Physikalisch-Technische Bundesanstalt (PTB). The objective of the comparison is to confirm the calibration and measurement capabilities of NIMT in hardness measurement. The period was between March to August 2009. There are two sets of artifacts: scale A artifact set and scale B artifact set. The scale A artifact consists of 7 hardness blocks; 35 HRA, 40 HRA, 55 HRA, 60 HRA, 70 HRA, 80 HRA, 85 HRA. The artifact for scale B consists of 9 hardness blocks; 25 HRB, 30 HRB, 40 HRB, 50 HRB, 60 HRB, 70 HRB, 80 HRB, 90 HRB, 100 HRB. Laboratory must ensure that the primary Rockwell hardness machine pass the verification according to ISO 6508-3. Then, participants measure hardness value by making 10 indentations in the designated area for each artifact block. Hardness measurement result and uncertainty budget are then reported to pilot laboratory and are used to compute the degree of equivalence in terms of Comparison Reference Value (CRV) and En ratio.

Revision: Original  
Date: October 16, 2009

## Content

1. Introduction	4
2. Participating Institutes	4
3. Artifacts	4
4. Measurement of Artifacts	5
5. Measurement Data	6
6. Analyzing Method of Comparison Results	10
7. Comparison Results	12
8. Discussion and Conclusion	16
9. References	16

## 1. Introduction

The National Institute of Metrology (Thailand), NIMT established the hardness national measurement standard on hardness scale HRA and HRB since 2006. NIMT would like to participate in an inter-laboratory comparison with a well-known national metrology institute to confirm the calibration and measurement capabilities (CMCs) of NIMT in hardness. Physikalisch-Technische Bundesanstalt (PTB) is one of the NMI's that have well supported NIMT. Also this time, the bilateral comparison on HRA and HRB, piloted by NIMT, has been arranged with Physikalisch-Technische Bundesanstalt (PTB).

## 2. Participating Institutes

- National Institute of Metrology (Thailand), NIMT
- Physikalisch-Technische Bundesanstalt, PTB

## 3. Artifacts


The artifacts are Rockwell hardness scale A and scale B reference block manufactured by Asahi Giken Co.,Ltd. Japan. Each scale consists of hardness block with serial number as in the table 1. The dimensions of the blocks are 65 mm in diameter and thickness of each block as in the table 1.

**Table 1**

HRA Set	S/N	Thickness	HRB Set	S/N	Thickness
35 HRA	A32208	11.50 mm	25 HRB	A30431	11.50 mm
40 HRA	A34095	11.50 mm	30 HRB	A30591	11.50 mm
55 HRA	24282	15 mm	40 HRB	A31146	11.50 mm
60 HRA	A12603	15 mm	50 HRB	A32205	11.50 mm
70 HRA	B15027	15 mm	60 HRB	A33301	11.50 mm
80 HRA	44010	15 mm	70 HRB	A34092	11.50 mm
85 HRA	A22039	15 mm	80 HRB	A35202	11.50 mm
			90 HRB	24280	15 mm
			100 HRB	30734	15 mm

The surface of each artifact block was engraved with divided segments for indicating the indentation positions (Figure 1). Each segment is used for single indentation in the comparison. There are 10 areas of measurement on each block. The rest of the segments on the block surface are reserved as alternative places for trial and error measurements

Revision: Original  
Date: October 16, 2009



**Figure 1**


The plastic sheet indicating the indentation position of each participant will be enclosed with the artifact sets. It shall be placed on the artifact block in order to ensure the indentation position before making the indentation.

The two set of artifacts, a set for scale A and a set for scale B will be transferred between participants according to a timetable in which the date of measurement has been agreed between participants.

Institute/Country	Time of measurement
NIMT, Thailand	March, 2009
PTB, Germany	May 2009

#### 4. Measurement of Artifacts

The primary Rockwell hardness machines used for the comparison must be passed the verification according to ISO 6508-3 [1]. Before making the indentation on each block, the testing cycle must be adjusted to comply with ISO 6508-3. Then start to measure the hardness value of artifact blocks by making 10 indentations each block on the designated area on the indentation-positioning sheet.



**Figure 2**

Revision: Original  
 Date: October 16, 2009

## 5. Measurement Data

**Table 2 NIMT HRA Measurement Result**

Hardness Block Number	Measured value						
	A32208	A34095	24282	A12603	B15027	44010	A22039
Nominal Value (HRA)	35	40	55	60	70	80	85
Section 1	34.09	44.13	54.99	60.58	70.65	81.53	86.74
Section 2	34.08	44.30	54.95	60.57	70.67	81.55	86.74
Section 3	34.15	44.29	54.93	60.60	70.66	81.52	86.73
Section 4	33.97	44.17	55.01	60.70	70.70	81.49	86.71
Section 5	34.10	44.23	55.07	60.52	70.66	81.50	86.72
Section 6	34.02	44.27	55.07	60.47	70.66	81.53	86.58
Section 7	34.12	44.19	54.94	60.51	70.67	81.53	86.65
Section 8	34.05	44.22	54.99	60.50	70.70	81.57	86.62
Section 9	33.96	44.30	54.98	60.63	70.72	81.55	86.62
Section 10	34.11	44.29	55.01	60.57	70.76	81.53	86.63
Average Value	34.07	44.24	54.99	60.57	70.69	81.53	86.67
Correction Value	0.00	0.00	0.00	0.00	0.00	0.00	0.00
Corrected Value	34.07	44.24	54.99	60.57	70.69	81.53	86.67

Revision: Original  
 Date: October 16, 2009

**Table 3 PTB HRA Measurement Result**

Hardness Block Number	Measured value						
	A32208	A34095	24282	A12603	B15027	44010	A22039
Nominal Value (HRA)	35	40	55	60	70	80	85
Section 1	33.78	44.02	55.23	60.77	70.97	81.64	86.80
Section 2	33.90	44.20	55.22	60.80	70.98	81.66	86.78
Section 3	33.76	44.16	55.32	60.91	70.98	81.63	86.71
Section 4	33.80	44.16	55.29	60.85	70.93	81.60	86.67
Section 5	34.07	43.99	55.30	60.80	70.95	81.59	86.74
Section 6	33.75	43.95	55.25	60.73	70.93	81.58	86.60
Section 7	33.91	44.20	55.26	60.78	70.95	81.60	86.57
Section 8	33.89	43.40	55.15	60.87	70.97	81.71	86.61
Section 9	33.79	44.06	55.22	60.89	71.01	81.62	86.56
Section 10	33.91	44.12	55.28	60.78	70.97	81.64	86.58
Average Value	33.86	44.03	55.25	60.82	70.96	81.63	86.66
Correction Value	0.10	0.10	0.10	0.10	0.10	0.10	0.10
Corrected Value	33.96	44.13	55.35	60.92	71.06	81.73	86.76

Revision: Original  
 Date: October 16, 2009

**Table 4 NIMT HRB Measurement Result**

	Measured value								
	A30431	A30591	A33146	A32205	A33301	A34092	A35202	24280	30734
Nominal Value (HRB)	25	30	40	50	60	70	80	90	100
Section 1	26.01	31.11	41.16	49.28	60.46	70.23	78.45	90.59	100.38
Section 2	26.10	31.19	40.83	49.28	60.60	70.09	78.58	90.48	100.41
Section 3	25.95	31.17	40.78	49.35	60.53	70.20	78.62	90.56	100.39
Section 4	25.92	31.24	40.61	49.38	60.46	70.16	78.49	90.41	100.42
Section 5	25.92	31.26	40.91	49.38	60.49	70.15	78.48	90.50	100.46
Section 6	25.77	30.88	40.47	49.19	60.46	70.13	78.41	90.51	100.48
Section 7	26.12	30.83	40.50	49.06	60.48	70.12	78.53	90.54	100.39
Section 8	25.99	30.75	40.69	49.16	60.37	70.16	78.32	90.48	100.28
Section 9	25.20	31.30	40.46	49.36	60.21	70.21	78.53	90.51	100.38
Section 10	25.16	31.32	40.22	49.24	60.27	70.04	78.52	90.57	100.46
Average Value	25.81	31.11	40.66	49.27	60.43	70.15	78.49	90.52	100.41
Correction Value	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00
Corrected Value	25.81	31.11	40.66	49.27	60.43	70.15	78.49	90.52	100.41

Revision: Original  
 Date: October 16, 2009

**Table 5 PTB HRB Measurement Result**

Hardness Block Number	Measured value								
	A30431	A30591	A33146	A32205	A33301	A34092	A35202	24280	30734
Nominal Value (HRB)	25	30	40	50	60	70	80	90	100
Section 1	25.90	30.56	40.49	48.69	60.49	70.04	78.56	90.24	99.99
Section 2	25.43	30.64	40.61	48.77	60.51	70.16	78.49	90.27	100.10
Section 3	24.91	30.84	40.18	48.77	60.51	70.20	78.48	90.24	100.16
Section 4	25.01	31.30	40.22	48.78	60.50	70.17	78.31	90.27	100.16
Section 5	25.31	30.84	40.30	48.81	60.63	70.09	78.27	90.27	100.05
Section 6	25.41	30.28	40.36	48.67	60.55	70.15	78.39	90.17	100.06
Section 7	25.82	30.71	40.14	48.76	60.43	70.32	78.47	90.14	99.96
Section 8	25.22	30.92	40.39	48.83	60.38	70.19	78.57	90.13	99.99
Section 9	24.91	31.09	40.18	48.72	60.35	70.16	78.53	90.13	100.05
Section 10	24.84	30.91	40.15	48.84	60.58	70.03	78.41	90.21	100.10
Average Value	25.28	30.81	40.30	48.76	60.49	70.15	78.45	90.21	100.06
Correction Value	0.15	0.15	0.15	0.15	0.15	0.15	0.15	0.15	0.15
Corrected Value	25.43	30.96	40.45	48.91	60.64	70.30	78.60	90.36	100.21

## 6. Analyzing Method of Comparison Results

The measurement results are used to compute the degree of equivalence in Comparison Reference Value (CRV) and En ratio. The calculation is shown in following formulas:

-Calculation of Comparison Reference Value (CRV)

Pilot laboratory determined CRV by calculating the weighted mean of measurements of all participants ( $x_{ref}$ ).

$$x_{ref} = \frac{x_1/u^2(x_1) + x_2/u^2(x_2) + \dots + x_n/u^2(x_n)}{1/u^2(x_1) + 1/u^2(x_2) + \dots + 1/u^2(x_n)} \quad \text{Eq. 1}$$

-The uncertainty of the CRV was calculated by following expression:

$$\frac{1}{u^2(x_{ref})} = \frac{1}{u^2(x_1)} + \frac{1}{u^2(x_2)} + \dots + \frac{1}{u^2(x_n)} \quad \text{Eq. 2}$$

Where:

$x_i$  = The measured value of participating institute  $i$  ( $i=1,2,\dots,n$ )

$u(x_i)$  = The standard uncertainty of  $x_i$

-Its deviation from CRV

$$d_i = x_i - x_{ref} \quad \text{Eq. 3}$$

-The uncertainty of this deviation at a 95% level of confidence

$$U(d_i) = k \cdot u(d_i) \quad \text{Eq. 4}$$

Where  $u(d_i)$  was given by

$$u^2(d_i) = u^2(x_i) + u^2(x_{ref}) \quad \text{Eq. 5}$$

And  $k = 2$

-Evaluation of Coefficient  $E_n$

The equivalence between the measurements of participating institutes was expressed by coefficient  $E_n$  as well.

$$E_n = \frac{x_i - x_{ref}}{\sqrt{U^2(x_i) + U^2(x_{ref})}} \quad \text{Eq. 6}$$

Where:

Revision: Original  
Date: October 16, 2009

$$U(x_i) = k \cdot u(x_i)$$

$$U(x_{ref}) = k \cdot u(x_{ref})$$

**Eq. 7**

The  $x_i$  was equivalent with CRV  $x_{ref}$  at 95% confidence level, if  $|E_n| \leq 1$ .

Where

CRV	is comparison Reference Value
$U(CRV)$	is uncertainty of CRV
$D_{PTB}$	is deviation value of PTB from CRV
$U(D_{PTB})$	is uncertainty of deviation value of PTB from CRV
En ratio PTB	is degree of equivalence of PTB's result expressed in En ratio
$D_{NIMT}$	is deviation value of NIMT from CRV
$U(D_{NIMT})$	is uncertainty of deviation value of NIMT from CRV
En ratio NIMT	is degree of equivalence of NIMT's result expressed in En ratio
$U_e$	is expanded uncertainty that evaluated by each NMIs

Revision: Original

Date: October 16, 2009

## 7. Comparison Results

The HRA comparison results, comparison reference value (CRV), the deviation value of PTB and NIMT from CRV and their uncertainty ( $D_{PTB}$ ,  $D_{NIMT}$ ,  $U(D_{PTB})$ ,  $U(D_{NIMT})$ ) and En ratio, were calculated and shown in Table 6. Figure 3 shows

**Table 6 Conclusion of comparison evaluation on Rockwell hardness scale A**

<b>PTB</b>	<b>PTB Comparison Result (HRA)</b>							
	No.	A32208	A34095	24282	A12603	B15027	44010	A22039
Nominal value	35	40	55	60	70	80	85	
Measured value	33.96	44.13	55.35	60.92	71.06	81.73	86.76	
$U_e$	0.25	0.49	0.16	0.17	0.13	0.14	0.22	

<b>NIMT</b>	<b>NIMT Comparison Result (HRA)</b>							
	No.	A32208	A34095	24282	A12603	B15027	44010	A22039
Nominal value	35	40	55	60	70	80	85	
Measured value	34.07	44.24	54.99	60.57	70.69	81.53	86.67	
$U_e$	0.40	0.40	0.39	0.40	0.39	0.38	0.40	

<b>Comparison evaluation</b>							
Norminal (HRA)	35	40	55	60	70	80	85
CRV	33.99	44.19	55.30	60.87	71.03	81.71	86.74
$U(CRV)$	0.21	0.31	0.15	0.16	0.12	0.13	0.19
$D_{PTB}$	-0.03	-0.07	0.05	0.05	0.04	0.02	0.02
$U(D_{PTB})$	0.33	0.58	0.22	0.23	0.17	0.19	0.29
En ratio PTB	-0.09	-0.12	0.23	0.23	0.21	0.12	0.07
$D_{NIMT}$	0.08	0.04	-0.31	-0.30	-0.34	-0.18	-0.07
$U(D_{NIMT})$	0.46	0.51	0.42	0.43	0.41	0.41	0.44
En ratio NIMT	0.17	0.09	-0.73	-0.69	-0.85	-0.43	-0.16

Revision: Original  
Date: October 16, 2009




Figure 3




Figure 4

Revision: Original  
 Date: October 16, 2009

**Table 7 Conclusion of comparison evaluation on Rockwell hardness scale B**

<b>PTB</b>	Measure value (HRB)								
No.	A30431	A30591	A31146	A32205	A33301	A34092	A35202	24280	30734
Nominal value	25	30	40	50	60	70	80	90	100
Measured value	25.43	30.96	40.45	48.91	60.64	70.30	78.60	90.36	100.21
Ue	0.76	0.58	0.34	0.17	0.21	0.21	0.23	0.15	0.16

<b>NIMT</b>	Measure value (HRB)								
No.	A30431	A30591	A31146	A32205	A33301	A34092	A35202	24280	30734
Nominal value	25	30	40	50	60	70	80	90	100
Measured value	25.81	31.11	40.66	49.27	60.43	70.15	78.49	90.52	100.41
Ue	0.78	0.54	0.64	0.41	0.42	0.37	0.39	0.37	0.37

Comparison evaluation									
Norminal (HRA)	25	30	40	50	60	70	80	90	100
CRV	25.62	31.04	40.50	48.96	60.60	70.27	78.57	90.38	100.24
U(CRV)	0.55	0.40	0.30	0.16	0.19	0.18	0.20	0.14	0.15
D_PTB	-0.19	-0.08	-0.05	-0.05	0.04	0.04	0.03	-0.02	-0.03
U(D_PTB)	0.94	0.71	0.46	0.23	0.29	0.27	0.31	0.21	0.22
En ratio PTB	-0.20	-0.11	-0.10	-0.22	0.15	0.13	0.09	-0.11	-0.14
D_NIMT	0.20	0.07	0.16	0.30	-0.17	-0.12	-0.08	0.13	0.16
U(D_NIMT)	0.95	0.67	0.71	0.44	0.46	0.41	0.44	0.39	0.40
En ratio NIMT	0.21	0.10	0.23	0.69	-0.36	-0.28	-0.18	0.34	0.41

Revision: Original  
Date: October 16, 2009




Figure 5




Figure 6

Revision: Original

Date: October 16, 2009

## 8. Discussion and Conclusion

This comparison used the deviation from CRV and the uncertainty of this deviation of a 95% level of confidence as the degree of equivalence. The  $E_n$  numbers were also calculated to express the equivalence between measurements of the participants. Comparison result of Rockwell hardness scale A measurement is shown in figure 3 and 4. The range of the deviation from CRV is between -0.34 HRA to 0.08 HRA. The  $E_n$  numbers ranges from -0.85 to +0.23. Figure 5 and 6 shows the comparison result of Rockwell hardness scale B measurement. The range of the deviation from CRV is between -0.17 HRB to 0.30 HRB. The  $E_n$  numbers ranges from -0.36 to +0.69. It has been shown that the NIMT and the PTB hardness machines compare well. Both laboratories show the ability of measurement in hardness scale of 35 HRA to 85 HRA and 25 HRB to 100 HRB.

## 9. References

- [1] ISO 6508-3: Metallic materials-Rockwell hardness test-Part 3: Calibration of reference blocks (scales A, B, C, D, E, F, G, H, K, N, T)