

**Final Report**  
**on**  
**EURAMET comparison of 500 kg mass standard**  
**EURAMET.M.M-S7**

**EURAMET project No. 1300**

Pilot

Matej Grum – MIRS, SI

Participants

Jaroslav Zůda - CMI, CZ  
Goran Grgić - MIRS, SI  
Michał Nawotka - GUM, PL  
Tatjana Žandarova - LATMB, LV  
Allar Pärn - Metrosert, EE  
Sauli Kilponen - MIKES, FI  
Bengt Gutfelt - RISE, SE  
Pekka T. Neuvonen - JV, NO  
Zoltan Zelenka - BEV, AT  
Christian Wüthrich - METAS, CH  
Mannie Panesar - NMO, UK  
Pedro Conceição - IPQ, PT  
Hugo Pirée - SMD-ENS, BE  
Csilla Vámosy - BFKH, HU  
Šejla Ališić - IMBIH, BA  
Goran Vukoslavović - MBM, MN  
Dragan Pantić - DMDM, RS  
Bianka Mangutova-Stoilkovska - BoM, MK  
Mariana Miteva - BIM, BG  
George Florian Popa - INM, RO

**Abstract**

In order to demonstrate the equivalence in calibration of 500 kg mass standard among National Metrology Institutes a supplementary comparison has been carried out by 20 members of EURAMET. The overall result shows a good consistency among the participants.

July 2021

## 1. Objectives

The objectives of the comparison were to facilitate the demonstration of metrological equivalence between the participating national laboratories and to check, confirm or improve the capabilities of quoted calibration measurement capabilities (CMC) at 500 kg. Details relevant for the comparison were specified in the technical protocol.

## 2. Participants

The pilot laboratory for the comparison was Metrology Institute of the Republic of Slovenia (MIRS) represented by Matej Grum. The list of participants with their contact details is given in the Table 1.

Table 1: Participants in the comparison.

Laboratory	Address (Measurements)	Contact Person
Czech Metrology Institute ( <b>CMI</b> )	Romana Havelky 17, CZ-58601	Jaroslav Žůda +420 602 551 921 jzuda@cmi.cz
Metrology Institute of the Republic of Slovenia ( <b>MIRS</b> )	Tkalska 15, SI-3000 Celje, Slovenia	Goran Grgić + 386 1 24 42 732 goran.grgic@gov.si
Central Office of Measures ( <b>GUM</b> )	2 Elektoralna Str., 00-139 Warsaw, Poland	Michał Nawotka (+48) 22 581 9335 m.nawotka@gum.gov.pl
Latvian National Metrology Centre, Metrology Bureau ( <b>LATMB</b> )	K. Valdemara, 157, Riga, LV-1013, Latvia	Tatjana Žandarova +371 67 517 728; +371 67 517 726 tatjana.zandarova@latmb.lv
AS <b>Metrosert</b> - Estonian Central Office of Metrology	Teaduspargi 8, 12618 Tallinn, Estonia	Allar Pärn +37253889335 allar.parn@metrosert.ee
VTT, <b>MIKES</b> Metrology	Tehdaskatu 15, FI-87100 Kajaani, Finland	Sauli Kilponen +358 50 4434178 Sauli.kilponen@vtt.fi
Research Institutes of Sweden ( <b>RISE</b> )	Brinellgatan 4, SE-502 64 Borås, Sweden	Bengt Gutfelt +46 10 516 54 76 bengt.gutfelt@ri.se
Justervesenet – The Norwegian Metrology Service ( <b>JV</b> )	Fetveien 99, NO-2007 Kjeller, Norway	Pekka T. Neuvonen + 47 64 84 84 70 ptn@justervesenet.no
Bundesamt für Eich- und Vermessungswesen ( <b>BEV</b> )	Arltgasse 35, 1160 Vienna	Zoltan Zelenka + 43 1 211 10 6607 zoltan.zelenka@bev.gv.at
Eidgenössisches Institut für Metrologie ( <b>METAS</b> )	Lindenweg 50, 3003 Bern-Wabern, Switzerland	Christian Wüthrich +41 58 387 04 23 christian.wuethrich@metas.ch
Physikalisch-Technische Bundesanstalt ( <b>PTB</b> )	Bundesallee 100, 38116 Braunschweig, Germany	Michael Borys + 49 531 592 1110 michael.borys@ptb.de
National Measurement and Regulation Office ( <b>NMO</b> )	Stanton Avenue, Teddington, Middlesex, United Kingdom, TW11 0JZ	Mannie Panesar +44 (0)20 8943 7246 mannie.panesar@nmo.beis.gov.uk

Instituto Português da Qualidade ( <b>IPQ</b> )	Rua António Gião, 2; Almada, 2829-513 Caparica, PORTUGAL	Pedro Conceição +351212948170 pconceicao@ipq.pt
FPS Economy, Metrology Division ( <b>SMD-ENS</b> )	Haachtsesteenweg 1795, 1130 Brussel, Belgium	Hugo Pirée +32 2 277 76 10 hugo.piree@economie.fgov.be
Budapest Főváros Kormányhivatala ( <b>BFKH</b> )	Nagytétényi út 15, 1222 Budapest, Hungary	Csilla Vámosy +36 1 4585 947 vamosscy@bfkh.gov.hu
Institute for Metrology of Bosnia and Herzegovina ( <b>IMBIH</b> )	Čatići bb, 72240 Kakanj, Bosna i Hercegovina	Šejla Ališić +387 33 568 920; +387 33 568 948 sejla.alisic@met.gov.ba;
Bureau of Metrology ( <b>MBM</b> )	Arsenija Boljeviča bb, 81000 Podgorica, Montenegro	Goran Vukoslavović +382 20 601 360; +382 67 596 125 goran.vukoslavovic@metrologija.gov.me
Directorate for measures and precious metals ( <b>DMDM</b> )	Mike Alasa 14, Belgrade, Serbia	Dragan Pantić + 381 11 2024 417 pantic@dmdm.rs
Ministry of economy - Bureau of metrology ( <b>BoM</b> )	Blvd.Jane Sandanski 109a, 1000 Skopje, R. Macedonia	Bianka Mangutova-Stoilkovska +389 2 24 03 676, ext.029 bianka.stoilkovska@bom.gov.mk
Bulgarian Institute of Metrology (DG NCM) ( <b>BIM</b> )	52 B »G. M. Dimitrov« Blvd., Sofia 1040, Bulgaria	Mariana Miteva + 3592 9702 759, +359 885 747 411 m.miteva@bim.government.bg
National Institute of Metrology ( <b>INM</b> )	Sos. Vitan-Barzesti no. 11, sector 4, 042122 Bucharest, Romania	George Florian Popa +40754824217 george.popa@inm.ro

### 3. Transfer standard

The transfer standard for the comparison was a stainless steel cylindrical weight with the nominal mass 500 kg provided by CMI. The density of the standard was 7888,2 kg/m<sup>3</sup>, associated uncertainty ( $k=2$ ) was 1,5 kg/m<sup>3</sup>. Producer: ZDAS. Serial number: 1. Dimensions: diameter 60 cm, height 25 cm.



Figure 1: Transfer standard - side view

In the technical protocol values of the magnetic properties, i.e. the magnetic polarisation and magnetic susceptibility of the standard were not reported. However, during the comparison PTB reported that the susceptibility seems to be increased in some areas at the bottom of the weight. They observed values up to the limit (and at one position slightly above) of the OIML class F1. The polarisation of the weight was found suitable. Later the susceptibility was measured also by MIRS ( $0,1 < \mu \leq 1,5$ ; measured by the attracting method, with a limited choice of permanent magnets) and at the end by CMI ( $\mu = 0,06$ ; measured by the attracting method).

Stability check of the transfer standard was performed by CMI at the beginning, three times during the comparison and at the end of the circulation schedule (see Section 5 for details).

The transfer standard was housed in a wooden container on a wooden pallet.

#### 4. Circulation schedule (actual)

The transfer standard circulated between participants according to the list in Table 2. Due to difficulties related to the custom clearance and Covid-19 pandemic the actual circulation time exceeded the planned one.

Table 2: Actual circulation schedule of the transfer standard.

<b>NMI</b>	<b>Country</b>	<b>Date</b>	<b>Remarks</b>
<b>CMI</b>	Czech Republic	June 2017	also used for stability evaluation
<b>GUM</b>	Poland	July 2017	
<b>LATMB</b>	Latvia	September 2017	
<b>Metrosert</b>	Estonia	September 2017	
<b>MIKES</b>	Finland	October 2017	
<b>RISE</b>	Sweden	November 2017	
<b>JV</b>	Norway	December 2017	
<b>CMI</b>	Czech Republic	February 2018	stability measurement only
<b>BEV</b>	Austria	April 2018	
<b>METAS</b>	Switzerland	June 2018	
<b>IPQ</b>	Portugal	August 2018	
<b>CMI</b>	Czech Republic	January 2019	stability measurement only
<b>PTB</b>	Germany	February 2019	
<b>NMO</b>	United Kingdom	March 2019	
<b>SMD</b>	Belgium	May 2019	
<b>MIRS</b>	Slovenia	July 2019	
<b>CMI</b>	Czech Republic	August 2019	stability measurement only
<b>BFKH</b>	Hungary	October 2019	
<b>IMBIH</b>	Bosnia and Herzegovina	December 2019	
<b>MBM</b>	Montenegro	January 2020	
<b>DMDM</b>	Serbia	March 2020	
<b>BOM</b>	Macedonia	June 2020	
<b>BIM</b>	Bulgaria	July 2020	
<b>INM</b>	Romania	September 2020	
<b>CMI</b>	Czech Republic	December 2020	stability measurements only
<b>BEV</b>	Austria	February 2021	repeated measurement

It was the responsibility of the participating laboratories to organize the transport to the next participant. When necessary, the standard was accompanied by an ATA Carnet, which was provided

by CMI. The transfer standard was transported between the participating laboratories mostly by courier companies.

When the standard arrived at the participating laboratory, the transportation container and its contents were checked for damage. A visual inspection of the surfaces of the standard was made and the results noted on the measurement report. No significant accidents occurred during the circulation.

## 5. Calculation of reference value and degrees of equivalence

The participating laboratories determined the conventional mass value of the transfer standard according their normal calibration procedure. Each participating laboratory reported its measured conventional mass difference from the nominal value of the transfer standard together with the expanded uncertainty. Reporting the conventional mass instead of the mass for calibration of 500 kg standards is most common practice of the participants.

The stability of the transfer standards was monitored by CMI. The transfer standards were measured five times, measurements at the beginning were taken also as a part of the comparison and the other four measurements only as additional stability measurements. The results of the stability measurements are given in the Table 3.  $m_s$  represents measured deviation of the conventional mass from the nominal mass of the transfer standard during the stability measurement and  $U$  the corresponding expanded uncertainty.

Table 3: Stability of the transfer standard

Date	$m_s$	$U$
06/2017	1,31 g	0,19 g
02/2018	1,29 g	0,19 g
01/2019	1,24 g	0,21 g
09/2019	1,20 g	0,23 g
12/2020	1,39 g	0,23 g

The standard uncertainty of the mass stability measurements by CMI  $u(m_s)$ , which only includes the contributions relevant for the relative mass measurements was estimated to be 0,07 g.

According to Clause 9.3 OIML R 111 [1], the limit values of permanent magnetization and magnetic susceptibilities are defined in such a way that, at magnetic fields and magnetic field gradients possibly present on balance pans, they produce a change of the conventional mass of less than 1/10 of the maximum permissible error of the test weight. However, it is not easy (and probably not even possible) to quantify the effect of a magnetised/susceptible weight on the mass determination results since it depends so much on the equipment used. For the purpose of this report the influence of magnetic properties to instability of the transfer standard, which was determined to be approximately on the level of requirements for class F1, was taken into account as

$$u(m_{magnetic}) = \frac{1}{\sqrt{3}} \frac{MPE(F1 \text{ for } 500 \text{ kg})}{10}$$

The analysis of the stability shows that it is not necessary to consider the drift of the transfer standard. However, the uncertainty of the instability  $u(m_{inst})$  of the standard due to:

- the maximum mass change  $\Delta m_s$  of all five stability measurements of the transfer standard obtained from Table 2 (rectangular distribution),
  - the uncertainty of the stability measurements and
  - the uncertainty due to influence of magnetic properties of the transfer standard
- was taken into account

$$u(m_{inst}) = \sqrt{\frac{\Delta m_s^2}{12} + u^2(m_s) + u^2(m_{magnetic})} = 0,17 \text{ g}$$

The reference value for the transfer standards was calculated by weighted least squares analysis of the measurement results taking into account the uncertainties of the measured values [2,3]. It was assumed that there was no correlation between the results of the participants.

The uncertainty component due to instability of the transfer standard was included in the calculation of the weighted average and the chi-square test by combining it with the uncertainty given by the laboratory  $u(m_i)$

$$u'^2(m_i) = u^2(m_i) + u^2(m_{inst})$$

A chi-squared test (with 19 degrees of freedom and at the 0,05 level of significance) was applied [2] to carry out an overall consistency check of the results obtained.

The degree of equivalence  $d_i$  was calculated as a difference between the participant's measured value  $m_i$  and the reference value  $m_{ref}$ , both given as the difference from nominal mass 500 kg of the transfer standard

$$d_i = m_i - m_{ref}$$

The standard uncertainty of the degree of equivalence  $u(d_i)$  for measurement results which were used in the calculation of the reference value was calculated by

$$u(d_i) = \sqrt{u'^2(m_i) - u^2(m_{ref})}$$

where  $u'(m_i)$  is the uncertainty given by the laboratory combined with the uncertainty component due to instability of the transfer standard and  $u(m_{ref})$  uncertainty of the reference value.

In a case where the consistency check had failed, the criterion  $|d_i| > 2u(d_i)$  was used to identify results which were discrepant compared with the reference value. If the absolute value of degree of equivalence was larger than the expanded uncertainty ( $k=2$ ) of the degree of equivalence, then the result was considered as discrepant at a 5 % level of significance.

The result with the highest discrepancy was excluded from the next round of calculation of the new reference value and reference standard uncertainty. When none of the results which were used in the calculation of the reference value were discrepant then the last calculated reference value is considered as final reference value.

For the measurements which were considered discrepant and were not used in the calculation of the reference value the standard uncertainty  $u(d_i)$  was calculated by

$$u(d_i) = \sqrt{u'^2(m_i) + u^2(m_{ref})}$$

After the consistency check was applied in the first calculation of the comparison reference value, discrepant results were found for 3 laboratories. As per CIPM MRA protocol [4] these participants were contacted to check that there is no arithmetic, typographical or transcription errors without providing them with initial comparison reference value.

IMBiH and BoM had an arithmetic error and sent the corrected results.

Also BEV submitted an obviously discrepant measurement result. The pilot let the laboratory know that the result was discrepant without any further information. After a careful investigation BEV identified the cause and asked to repeat the measurements. BEV stated that: "It looks like the balance had more serious problems after the EURAMET measurements, and it was repaired several times. It

could have already affected the EURAMET measurements.” BEV repeated measurements and reported new values. This is the reason that both BEV initial and repeated measurement results are included in the comparison report.

After taking the corrected results mentioned above into account, the consistency check didn't fail any more.

## 6. Summary of the results

Each participating laboratory was requested to report the following information:

- The conventional mass value of the transfer standard and associated expanded measurement uncertainty.
- Uncertainty budget.
- Details on the used reference mass standards including their traceability.
- Details of the balance used in the comparison.
- Description of the measurement procedure.
- Instruments used for measurement of ambient conditions.
- Laboratory environmental conditions during the measurements.

Table 4 provides information on how 500 kg reference mass was composed and which balances were used by the participants. In order to indirectly define possible magnetic effect of the standard on the balances the participating laboratories were subsequently asked to provide an estimate of the distance of the closest surface of the transfer standard from the balance force transducer.

Table 4: Equipment used

Laboratory	Composition	Balance	Resolution	Estimated distance
CMI	10 x 50 kg	KC 500-1	0,1 g	150 mm
GUM	10 x 50 kg	HRP.500.4Y.KO	0,1 g	220-250 mm
LATMB	500 kg	MC 0,5 t	1 g	(35 ÷ 38) mm
Metrosert	9 x 50 kg + 50 kg platform	KC 500-1	0,1 g	50 mm
MIKES	10 x 50 kg	Equal arm beam	0,02 g	Not applicable. Mechanical balance.
RISE	8 x 50 + 100 kg	KC500	0,1 g	130 0m
JV	500 kg	XP604KM	0,1 g	~ 128 0m
BEV	24 x 20 kg + 20 kg basket	XP604KMC	0,1 g	(150 ± 25) mm
METAS	stack of 25 x 20 kg or 500 kg monolithic (2 pieces)	XP604	0,1 g	100 mm
IPQ	500 kg (3 pieces)	XP1003K	0,5 g	55 mm
NMO	500 kg (2 pieces)	KC500-1	0,01 g	No estimate provided.
SMD	500 kg	XPE604KMC	0,1 g	No estimate provided.
MIRS	24 x 20 kg + 20 kg basket	XP604KM	0,1 g	120 mm
BFKH	500 kg	KA-500	0,1 g	100-120 0m
IMBIH	500 kg	PTF 600	1 g	No estimate provided.
MBM	22 x 20 kg + 60 kg basket	XP604KM	0,1 g	30 mm
DMDM	22 x 20 kg + 60 kg basket	KC500-1	0,1 g	(100 ± 10) mm
BOM	500 kg	CCS600K	1 g	(70 ± 4) mm
BIM	500 kg	KCC1000	1 g	180 mm
INM	23 x 20 kg + 40 kg rack	XP604KM, KC1000	0,1 g	at least 200 mm

The uncertainties claimed by each participant were supported by the relevant uncertainty budgets, which followed the templates provided in the technical protocol.

The participants' results for comparison of the 500 kg transfer standard (the measured conventional mass difference from the nominal value of the transfer standard  $m_i$  with the corresponding standard uncertainty  $u(m_i)$ ), the degrees of equivalence  $d_i$  with the corresponding standard uncertainty  $u(d_i)$  and the ratio  $d_i/u(d_i)$  are reported in Table 5 together with the calculated reference value  $m_{ref}$  with its corresponding standard uncertainty  $u(m_{ref})$ . The results are plotted on Figures 2 and 3.

Based on its observations concerning the magnetic properties of the transfer standard as mentioned in Section 3, PTB didn't report its results. A confirmation of magnetic properties of the transfer standard on the level of OIML R 111 [1] class F1 would be a limiting factor for their measurements and would not allow them to support their CMCs.

BEV results are reported twice. The initial discrepant result is reported under BEV (1) and the repeated result is reported under BEV (2). The result BEV (1) was not included in calculation of the reference value.

Table 5: Participants' results, reference value and degrees of equivalence for the 500 kg standard comparison.

Laboratory	$m_i$ g	$u(m_i)$ g	$d_i$ g	$u(d_i)$ g	$d_i/u(d_i)$
CMI	1,31	0,10	0,13	0,17	0,76
GUM	1,90	1,00	0,72	1,01	0,71
LATMB	0,73	2,20	-0,45	2,20	-0,21
Metrosert	3,10	1,00	1,92	1,01	1,90
MIKES	1,37	0,38	0,19	0,40	0,47
RISE	1,40	0,75	0,22	0,76	0,29
JV	-0,40	0,75	-1,58	0,76	-2,07
BEV (1)	-0,30	0,40	-1,48	0,41	-3,58
METAS	0,61	0,25	-0,57	0,29	-1,99
IPQ	-0,25	0,80	-1,43	0,81	-1,76
NRO	1,12	0,13	-0,06	0,19	-0,34
SMD	2,22	0,62	1,03	0,63	1,63
MIRS	1,13	0,37	-0,05	0,39	-0,13
BFKH	1,99	0,52	0,81	0,53	1,52
IMBIH	-2,56	4,00	-3,74	4,00	-0,93
MBM	0,23	1,16	-0,95	1,17	-0,81
DMDM	1,50	1,30	0,32	1,31	0,24
BOM	1,47	1,20	0,29	1,21	0,24
BIM	1,40	1,45	0,22	1,46	0,15
INM	0,70	0,75	-0,48	0,76	-0,63
BEV (2)	1,40	0,35	0,22	0,38	0,58

$$m_{ref} = \mathbf{1,18\ g} \quad u(m_{ref}) = \mathbf{0,10\ g}$$

Due to the magnetic properties of the transfer standard (see Chapter 3 for details), in general the results of the comparison cannot support the calibration capabilities of the laboratories with the expanded measurement uncertainty of less than 0,80 g. However, it needs to be taken into account that the magnetic impact depends on the equipment used, where in particular the distance between the centre of gravity of the transfer standard and the load cell is important in the case of the load cell with electromagnetic compensation. No such influence is relevant if a mechanical balance is used. The "limit" value of the expanded measurement uncertainty of 0,80 g was determined based on the provisions of OIML R 111 for a 500 kg weight of accuracy class F1, because values of susceptibility up to the limit (and at one position slightly above) of the OIML class F1 were observed during the comparison.




Figure 2: Results for the 500 kg standard with expanded uncertainties ( $k=2$ ) as reported by participants. Red data point represents the discrepant result, which was not included in the calculation of the reference value. Black data points represent the stability measurements.




Figure 3: Degrees of equivalence for the 500 kg standard with expanded uncertainties ( $k=2$ )

## 7. Conclusion

The comparison EURAMET.M.M-S7 was piloted by MIRS, twenty laboratories calibrated the conventional mass of the 500 kg transfer standard. The standard and its stability measurements were provided by CMI.

The comparison was conducted between June 2017 and February 2021.

The uncertainty component due to instability of the transfer standard was considered in the calculation of the weighted average and in the chi-square test.

BEV results were reported twice. The initial result was discrepant and therefore not included in calculation of the reference value.

With the exception mentioned above, the chi-square test consistency check did not fail and all results were included in the calculation of the reference value.

The degrees of equivalence of the results mostly show a good agreement of participants' results with the reference value, however one laboratory shows the absolute value of the ratio  $d_i / u(d_i)$  slightly above 2.

Due to the magnetic properties of the transfer standard, in general the results of the comparison cannot support the calibration capabilities of the laboratories with the expanded measurement uncertainty of less than 0,80 g.

EURAMET Guide on Comparisons [5] suggests that the participants should give a written statement indicating if their results are consistent with the CMC claims or not. The participants' statements regarding the CMCs are listed below:

Laboratory	Written statement indicating if the result is consistent with the CMC claims or not. If not, corrective actions should be described.
CMI	The result is consistent with the CMC claims.
GUM	The result is consistent with the CMC claims.
LATMB	No CMC published for 500 kg yet.
Metrosert	No CMC published for 500 kg yet.
MIKES	The result is consistent with the CMC claims.
RISE	The result is consistent with the CMC claims.
JV	<p>We discovered a technical problem (corner load) with the balance during the measurements for this comparison. This has been communicated to the pilot laboratory throughout the whole comparison.</p> <p>According to the manufacturer of the balance, this balance type is known to have this type of technical issues. The balance was repaired and we have implemented better routines, and shortened the testing interval, to discover technical issues earlier in the future.</p> <p>An extra uncertainty contribution is added due to the corner load in accordance with EURAMET G18, however this does not account for a possible systematic error in the measurement due to faulty balance. Since the balance is repaired now (2018), uncertainty contributions due to corner load and repeatability can be reduced. Thus our uncertainty falls close to 1,1 g, which is our CMC.</p> <p>Due to the possibility of systematic error, it is difficult to state from these results if our CMC is confirmed. However, after the technical problem was solved, calibration results for our internal reference weight yields results consistent with historical data. This was not the case just before the comparison.</p>

BEV	The result is consistent with the CMC claims.
METAS	The offset respective to the reference value (0,57 g) as well as the claimed uncertainty (0,5 g, k=2) are within our CMC for 500 kg which is 1 g.
IPQ	No CMC published for 500 kg yet.
NMRO	No CMC published for 500 kg yet.
SMD	The result is consistent with the CMC claims.
MIRS	No CMC published for 500 kg yet.
BFKH	BFKH do not want to change the measurement uncertainty in the CMC table because a new balance and standards weights are under procurement. (BFKH reported higher uncertainty (U = 1,04 g) than published CMC at 500 kg (U = 0,80 g).)
IMBIH	No CMC published for 500 kg yet.
MBM	No CMC published for 500 kg yet.
DMDM	The result is consistent with the CMC claims.
BOM	No CMC published for 500 kg yet.
BIM	No CMC published for 500 kg yet.
INM	The result is consistent with the CMC claims.

## 8. References

- [1] Weights of classes E<sub>1</sub>, E<sub>2</sub>, F<sub>1</sub>, F<sub>2</sub>, M<sub>1</sub>, M<sub>1-2</sub>, M<sub>2</sub>, M<sub>2-3</sub>, M<sub>3</sub>, International recommendation OIML R111-1, OIML, 2004.
- [2] M.G. Cox, The evaluation of key comparison data, Metrologia, 2002, Vol. 39, 589-595.
- [3] Nielsen, L. "Evaluation of measurement intercomparisons by the method of least squares", DFM-99-R39 (2000).
- [4] Measurement comparisons in the CIPM MRA, CIPM MRA-D-05 Version 1.6
- [5] EURAMET Guide on Comparisons, EURAMET Guide No. 4, Version 1.0 (05/2016)