

Analysis of “problem” molecules: Ammonia, and Hydrogen Chloride

Lyn Gameson

NIST Gas Sensing Metrology Group

Gaithersburg, MD

301-975-6558

LGameson@nist.gov


CCQM-GWAG, October 13th 2016

Lisbon, Portugal

Outline

- ❖ Gas Auto Sampling - Standard Format
- ❖ Gas Auto Sampling - Refined Format
- ❖ Results - Ammonia
- ❖ Results - Hydrogen Chloride
- ❖ Gas Auto Sampling - Major Change

Computer Operated Gas Auto Sampler (COGAS) Standard Format



Hardware:

PSMs Available

Gas Regulator - 30 psig

PFA Tubing

Other Samples Dead Ended

Acquisition:

2 Samples per Control

Up to 6 Minute Purge

Up to 2 Minute Data Averaging

Components:


CO₂; CO;

CH₄; C₃H₈; O₂

SO₂; NO; NO₂; H₂S

Precision ≤ 0.2% Rel.

Computer Operated Gas Auto Sampler (COGAS) Small Refinement



Hardware:

PSMs Available

Gas Regulator - 30 psig

SS Tubing – Silconert 2000 coated

Other Samples Dead Ended

Acquisition:

1 Sample per Control

Up to **10** Minute Purge


Up to 2 Minute Data Averaging

Permeation Tube Vs. Compressed Gas Cylinder

Production of Gas Mixture Standard:

- Single Component ($\leq 50 \mu\text{mol/mol}$)
- Nitrogen or Air Balance

Category	Permeation Tube	Compressed Gas Cylinder
Standard Type	Dynamic	Static
Standard is Portable	No	Yes
Number of Dilution Steps	1	≥ 4
Possible Immediate Loss to Walls	N/A	Yes
Possible Unstable Gas Mixture	N/A	Yes
Direct Handling of Hazardous Material	No	Yes



Fixed Gas Flow


Perm. Tube

Outlet


Vary Gas Flow

Permeation Device System

Continuous-weighing mode compliant with ISO 6145 Part 10



Computer Operated Gas Auto Sampler (COGAS) Refined Format



Hardware:

Gas Regulator - 30 psig

No PSMs Available

Other Samples Dead Ended


Acquisition:

1 Sample per Control

Up to 10 Minute Purge

Up to 2 Minute Data Averaging

Computer Operated Gas Auto Sampler (COGAS) Further Refined Format



Hardware:

Gas Regulator - 30 psig

No PSMs Available

Switching Valve


Acquisition:

Sample used as Control


Up to 10 Minute Purge

Up to 2 Minute Data Averaging

NH₃ Analysis - Preparation






NH₃ Analysis - Preparation





9 Nominal $\mu\text{mol/mol}$ NH_3 Response by Chemiluminescence

Raw Instrument Response Vrs Time





	Seq#: 1	Seq#: 2	Seq#: 3	Seq#: 4	Seq#: 5	Seq#: 6
m:	1.27200E-01	1.28054E-01	1.27173E-01	1.26179E-01	1.27238E-01	1.27238E-01
c:	-1.30739E-01	-1.39998E-01	-1.31043E-01	-1.24344E-01	-1.34317E-01	-1.34317E-01
Control (ppm):	8.889	8.902	8.894	8.911	8.915	8.915
Previous (ppm):	8.88	8.88	8.88	8.88	8.88	8.88
%Diff:	0.11	0.25	0.15	0.35	0.39	0.39



Cylinder Number	2011 Analysis NH ₃ (μmol/mol)	2013 Analysis NH ₃ (μmol/mol)	%Difference
KAL002649	27.74 ± 0.34	27.67 ± 0.28	-0.25
KAL002651	29.44 ± 0.36	29.41 ± 0.29	-0.09
AAL071147	9.51 ± 0.12	9.62 ± 0.10	1.15
AAL071153	9.38 ± 0.12	9.38 ± 0.10	-0.01
AAL071013	9.50 ± 0.12	9.51 ± 0.10	0.10
AAL071027	9.25 ± 0.11	9.37 ± 0.10	1.30
AAL072344	8.88 ± 0.11	8.89 ± 0.10	0.11

↑
No Flow of
Sample when
Perm Tube is
Analyzed

↑
Continuous
Sample Flow

NH₃ Certification - Current Capabilities and Methodology


- Certify from 5 to 50 µmol/mol at **1%** Expanded Uncertainty
- Condition Analyzer using Perm Tube Process Flow
- Confirm Equivalence of Perm Tube Process Flow through COGAS and directly to Analyzer
- Sample used as Control
- Initial **One** Hour Sample Purge, 5 minutes thereafter
- Sample flow at 300 cc/min to Analyzer
- Sample flow at 100 cc/min to Vent, while Analyzing Perm Tube Flow to Analyzer

Original Roadmap of Certification of HCl

Objective: Expanded Uncertainty of $\leq 1\%$ Relative

Range: 1 to 1000 $\mu\text{mol/mol}$ HCl in Balance N_2

Certify Working Standards



HCl - Preparation




Wash with 1 M HCl


Wash with Water

Wash with Acetone

Dry with N₂




1.8 Nominal $\mu\text{mol}/\text{mol}$ HCl Response by CRDS





HCl Halo Response vrs Time



	Mean	Control	Ratio
Perm	2199.7		
Cyl	1453.7	2201.4	0.6604
Perm	2203.0		
Cyl	1524.7	2208.7	0.6903
Perm	2214.3		
Cyl	1538.2	2208.3	0.6966
Perm	2202.2		
Perm	2207.3		
Cyl	1542.2	2205.2	0.6993
Perm	2203.1		

6.6% Increase

5 Hrs

3 Hrs

1 Hr

1 Hr

Low HCl Certification - Current Capabilities and Methodology

- Certify from 1 to **5** $\mu\text{mol/mol}$ at **2%** Expanded Uncertainty
- Condition Analyzer using Perm Tube Process Flow
- Confirm Equivalence of Perm Tube Process Flow through COGAS and directly to Analyzer
- Sample used as Control
- Initial **Eight** Hour Sample Purge, 10 minutes thereafter
- Sample flow at 300 cc/min to Analyzer
- Sample flow at 100 cc/min to Vent, while Analyzing Perm Tube Flow to Analyzer

Low HCl Certification - Next Steps

Reduce the **EIGHT** hour Purge Time !!

Replace Two Stage Regulator with:


- Silconert 2000 Coated, Low Volume, Single Stage Regulator
- Restriction Valve – gas must ALWAYS be flowing

Coatings:


- Sample Selection Valve
- Use the highly Hydrophobic Dursan coating


Develop a High Pressure COGAS with One Regulator

High Pressure COGAS - Purging Mode



High Pressure COGAS - Analyzing Mode







Questions ?