

Traceability in External Quality Assessment:

How Weqas ensures traceability in EQA and stresses its importance to users

David Ducroq

Weqas


Unit 6, Parc Tŷ Glas

Llanishen

Cardiff

UK

www.weqas.com



Programme Design: Lab and PoCT

Serum Chemistry

Lipids / Bilirubin

Common Report format
(quantitative)

ED Toxicology

Urine Chemistry / Oxalate & Citrate

Blood Gases / Co-oximetry

Endocrine / Haematinics / Cardiac Marker / BNP

Homocysteine / Bile Acids / Serum ACE / Serum HCG
/ Porphyrin / Ammonia / CRP / TDM / IS

HbA1c

POCT Creatinine

Urine Drugs of Abuse

Multiple samples are important

- Identifies components of both Imprecision and Inaccuracy (*traceability across the measurement range*)
- Identifies systematic errors
- Assesses method linearity – required for ISO 15189
- Powerful error detection tool



Value of Reference Targets

- Traceable to higher order
- Establishes method traceability for the lab – requirement of ISO 15189
- Highlights the pitfalls of using the trimmed overall mean as an accuracy target in EQA Schemes
- Useful in the post market vigilance of the IVD - Directive
- Required for UK MAPS



***Detailed in
Participants Manual
and on website***

Reference Methods

Flame Atomic Absorption/ Emission Spectrometry

- Sodium, Potassium, Calcium
- Magnesium, Lithium

IFCC Enzymes

- AST, ALT, LDH, GGT

- HbA1c *

* Provided by IFCC Ref lab, Netherlands

ID-MS

- Progesterone
- Testosterone
- Cortisol
- Bile Acids
- Creatinine
- Cholesterol**
- Glucose
- Urate
- Triglyceride
- HDL ***

** Currently provided by CDC lab Rotterdam & Weqas Ref Lab

*** Currently provided by CDC lab Rotterdam

Drugs Of Abuse / Therapeutic Drug Monitoring: Gravimetric values

Measurand (DOA; Urine)	Range Covered	Measurand (TDM; Serum)	Range Covered
Amphetamine	0 – 3000 µg/L	Amikacin	0 – 35 mg/L
Benzodiazepine	0 – 1000 µg/L	Carbamazepine	0 – 20 mg/L
Barbiturate	0 - 1000 µg/L	Digoxin	0 – 6 ug/L
Buprenorphine	0 - 50 µg/L	Gentamicin	0 - 20 mg/L
Cocaine	0 - 1000 µg/L	*Lamotrigine	0- 30 mg/L
Cannabis	0 - 150 µg/L	Lithium	0 – 2.5 mmol/L
6-Acetylmorphine (heroin)	0 - 50 µg/L	Methotrexate	0 -1.5 umol/L
Ketamine	0 - 3000 µg/L	Phenobarbital	0 - 65 mg/L
Methadone	0 - 1000 µg/L	Phenytoin	0 - 30 mg/L
EDDP	0 - 1000 µg/L	*Teicoplanin	0 - 70 mg/L
Methamphetamine	0 - 3000 µg/L	Theophylline	0 - 30 mg/L
Opiates	0 – 3000 µg/L	Tobramycin	0 - 15 mg/L
Phencyclidine (PCP)	0 - 100 µg/L	Valproic acid	0 - 175 mg/L
Tricyclic antidepressants	0 - 3000 µg/L	Vancomycin	0 - 50 mg/L
MDMA	0 – 3000 µg/L		
Amphetamines Group Screen	Qualitative only		

- High order drug/metabolite gravimetrically added to base material
- Pools mixed with the negative base material to produce a panel of intermediate pools.
- The “weighed-in” value incorporating purity of the spike used as target value.
- All microbalances are calibrated by ISO17025 accredited organisation

The WEQAS Report

Target values used in Statistical Analysis

Hierarchy



Reference values – used for bias plot /SDI calculation and σ score

Method mean – used for SDI calc if no ref and $n > 8$

Overall mean – used for SDI calc and bias plot if no ref and $n < 8$

Analyser mean – on report for information only

Analytical Specification Requirements

The National QA Advisory Panel in the UK has devised a **Minimum Analytical Performance Specification (MAPS)** which has been adopted since 2010

	Concentration	Allowable Bias vs Reference Value	Allowable variability	Allowable Total Error
Total Cholesterol	5.0 mmol/L [Desirable ¹]	4.00%	2.70%	8.50%
HDL-Cholesterol	1.0 mmol/L [Desirable ¹]	5.20%	3.60%	11.10%
	1.0 mmol/L [Achievable]	10.00%	3.60%	15.90%
Glucose	7.0 mmol/L [Desirable ¹]	2.20%	2.90%	6.90%
	2.0 mmol/L [Achievable]	+/- 10% absolute		
HbA1c	50 mmol/mol [Desirable ¹]	2.2%*	2.5%*	6.3%*
	50 mmol/mol [Achievable]	3.60%	2.50%	7.70%
Creatinine	75 umol/L [Desirable ¹]	3.80%	2.70%	8.20%
	75 umol/L [Achievable]	5.00%	2.70%	9.50%

Scheme: Mainline Chemistry. Distribution Code: PQ.
Distribution Date: 2/06/14. Final Report Issued: 1/07/14

Creatinine (µmol/L)		1	2	3	4	Analyte SDI
Reported Result		245	40	452	314	
Method Corrected Result		245.0	40.0	452.0	314.0	
Kinetic Jaffe	Mean	247.1	43.2	448.4	313.1	
	SD	6.9	4.6	18.2	10.1	
	Number	53	52	54	53	
	Uncert.	0.94	0.64	2.48	1.38	
Architect	Mean	249.6	44.1	464.3	320.9	
	SD	5.1	0.9	7.1	4.2	
	Number	9	8	9	9	
	Uncert.	1.71	0.33	2.37	1.39	
Overall	Mean	245.5	40.9	448.3	313.0	
	SD	7.6	4.7	17.0	10.4	
	Number	244	239	250	247	
	Uncert.	0.49	0.30	1.08	0.66	
Reference Values						
ID-GCMS		245.9	41.6	443.9	310.5	
Ref. Value Uncertainty		1.52	0.26	2.75	1.92	
Non-scoring Reference Values						
WeQas SD		9.7	5.1	21.1	12.7	
SDI		-0.09	-0.31	0.38	0.28	0.27
Sigma Metrics						
Critical Level 1: 75 µmol/L						
Minimum Acceptable score		1.67	Critical Level 1 Sigma score		2.1	
MAPS Allowable TE		9.5%	Lab bias %		2.9%	
MAPS Allowable bias %		5.0%	Lab CV %		3.2%	
MAPS Allowable CV %		2.7%				

Please note: Linear regression uses CF corrected data.

“True” Value

Performance Criteria


Target ± 9.5%

Total Error


SDI is a measurement of your total error and will include both in

This Distribution PQ
Your average analyte SDI for the 4 samples is 0.27


Previous SDI



This Distribution PQ



Previous Distributions



Assesses how far the lab's results are away from the “true” value and general performance over time.

Traceability Communication to Users


Reports sent to participants, published at conferences, journals, website etc.

Annual Participant Meeting often features talks on traceability


Example:

Creatinine (presented at Euromedlab, Athens)


EQA Assigned Method Groups



Enzymatic methods users: the number of users risen from 20% in 2012 to 38% in 2017




Major group is the Roche Cobas, which will influence the Jaffe overall mean




Major group again is the Roche Cobas, which will influence the enzymatic overall mean


Overall Method Groups Bias Plot



Jaffe ID-MS Traceable Bias Plot



Enzymatic: Bias Plot



In Summary

- Laboratories are scored against Reference Measurement Values where available
- Gravimetric spiking of material with higher order material is used as an alternative traceable target
- Both Reference Measurement and gravimetric target values aid in identification of methods where traceable calibration may be an issue
- Traceability of methods is therefore assured aiding ISO 15189 accreditation for laboratories
- The use of reference targets as opposed to comparison with mean data eloquently highlights the variability of results and reduces the pitfalls of laboratory trimmed data comparisons

Thank you for your attention

