

## Final Report for Supplementary Comparison

### APMP.QM-S15: Carbon Dioxide in Nitrogen at 1000 $\mu\text{mol/mol}$

Jeongsoon Lee<sup>1</sup>, JinBok Lee<sup>1</sup>, Jeongsik Lim<sup>1</sup>, Dongmin Moon<sup>1</sup>, Shankar G. Aggarwal<sup>2</sup>, Prabha Johri<sup>2</sup>, and Daya Soni<sup>2</sup>, Liu Hui<sup>3</sup> and Kai Fu Ming<sup>3</sup>, Ratirat Sinweeruthai<sup>4</sup> and Soponrat Rattanasombat<sup>4</sup>, Oman Zuas<sup>5</sup>, Harry Budiman<sup>5</sup>, and Muhammad Rizky Mulyana<sup>5</sup>, Vladimir Alexandrov<sup>6</sup>

<sup>1</sup>Korea Research Institute of Standards and Science, Daejeon, 34113, Republic of Korea

<sup>2</sup>CSIR-National Physical Laboratory India (NPLI), New Delhi, 110012, India

<sup>3</sup>National Metrology Centre, A\*STAR 118221, Singapore

<sup>4</sup>National Institute of Metrology, 12120 Thailand

<sup>5</sup>National Measurement Standards, National Standardization Agency of Indonesia (SNSU-BSN), 15314, Banten – Indonesia

<sup>6</sup>Kazakhstan Institute of Metrology (KazInMetr), 100009, Kazakhstan

#### **Field**

Amount of substance

#### **Subject**

Carbon dioxide 1000  $\mu\text{mol/mol}$  in Nitrogen

#### **Participants**

A total of six laboratories participated in this supplementary comparison. Table 1 lists the participants in this supplementary comparison

**Table 1: List of participants**

| Acronym  | Country    | Institute  |
|----------|------------|--|
| NPLI     | India      | CSIR-National Physical Laboratory India                                      |
| KAZ      | Kazakhstan | Kazakhstan Institute of Metrology  |
| NIMT     | Thailand   | National Institute of Metrology (Thailand)                                   |
| NMC      | Singapore  | National Metrology Centre  |
| SNSU-BSN | Indonesia  | National Measurement Standards, National Standardization Agency of Indonesia |
| KRISS    | Korea      | Korea Research Institute of Standards and Science                            |

**Organizing Body**

APMP TCQM

**Background**

Carbon dioxide (CO<sub>2</sub>) in nitrogen was one of the first types of gas mixtures performed at an international key comparison. The comparison dates back to 1998 (CCQMK1a) [1]. Since then, many National Metrology Institutes (NMIs) have developed Calibration and Measurement Capabilities (CMCs) for these mixtures. The international comparison of CO<sub>2</sub> at ambient level through CCQM-K52 was compared in 2007 [2]. Recently, NMIs in the APMP region have focused on developing emission standards for regulating CO<sub>2</sub> released by various powered vehicles such as automobiles (motor cars). At the 2017 APMP meeting, several NMIs requested a CO<sub>2</sub>/N<sub>2</sub> comparison to establish their own standards related with automotive regulation, which was to be coordinated by KRISS. Consequently, this comparison provides a CMC for APMP regional NMIs to develop CO<sub>2</sub>/N<sub>2</sub> CMC claim. The nominal amount-of-substance fraction for the CO<sub>2</sub> in nitrogen supplementary comparison is presented in table 2.

**Table 2: Nominal Amount-of-substance fraction**

| Component | Nominal amount |
|-----------|----------------|
|-----------|----------------|

|                |                          |
|----------------|--------------------------|
| Carbon dioxide | 1000 $\mu\text{mol/mol}$ |
| Nitrogen       | Balance                  |

## Schedule

The schedule for this part of the comparison is presented in Table 3.


**Table 3: Schedule**

| Time     | Event   |
|----------|---|
| Nov 2016 | Proposal for the supplementary comparison     |
| Apr 2017 | Protocol preparation by KRISS                 |
| Jun 2017 | Registration and protocol circulation         |
| Nov 2017 | Preparation/Distribution of mixtures by KRISS |
| Aug 2018 | Returning cylinder to KRISS                   |
| Dec 2018 | Reanalysis                                    |
| Sep 2019 | Draft report A                                |
| Oct 2019 | Draft report B –undergoing                    |

## Preparation of measurement standards

A total of eight gas mixtures were prepared gravimetrically in August 2017 [4] by diluting the first step cylinders of CCQM-K120 [3] and verified with a GC (Gas Chromatograph)/TCD (Thermal Conductivity Detector) analyzer in October 2017. The amount-of-substance fraction was determined based on the gravimetric method after purity analysis, which was assigned as a reference value. This implies that each cylinder has a unique reference value. CO<sub>2</sub> in the raw N<sub>2</sub> gas with high purity was around 0.011  $\mu\text{mol/mol}$ , which was considered negligible. Purity results are shown in the report of the former CCQM-K120 comparison. Accordingly, the final amount of each cylinder was assigned after applying the CO<sub>2</sub> purity result to the gravimetric one.

After weighing, all prepared mixtures were analyzed to verify their reliability [5]. As shown in figure 1, they agree within 0.1 %. Therefore expanded uncertainties of verification were evaluated as 0.050 % ( $k = 2$ ), as shown in Table 4.



**Figure 1. Consistency between gravimetrically prepared mixtures for this comparison (October 2017)**

**Table 4: Preparation of measurement standards**

| Cylinder number | Gravimetric value<br>[ $\mu\text{mol/mol}$ ] | $U$ from gravimetry<br>( $k=2$ ) [ $\mu\text{mol/mol}$ ] | $U$ from verification<br>( $k=2$ ) [ $\mu\text{mol/mol}$ ] |
|-----------------|--|--|--|
| D581078         | 999.49                                       | 0.15   | 0.50   |
| D581092         | 999.87                                       | 0.16   | 0.50   |
| D581146         | 999.83                                       | 0.14   | 0.50   |
| D581054         | 999.70                                       | 0.15   | 0.50   |
| D581240         | 999.29                                       | 0.15   | 0.50   |
| D581075         | 999.54                                       | 0.14   | 0.50   |
| D581070         | 999.94                                       | 0.16   | 0.50   |
| D581103         | 1001.00                                      | 0.14   | 0.50   |

A reference mixture (Rm) was analyzed between every sample mixture (Sm) to measure ratios of samples to reference and to monitor analyzer drift, for example, in a sequential set of (Rm<sub>*j-1*</sub>-Sm<sub>*i,j*</sub>- Rm<sub>*j+1*</sub>) for *i*-th sample cylinder of *j*-th analysis. The D015343 cylinder was used as the reference (Rm). In

equation (1),  $R_i$  is the ratio ( $S_i/S_{i^{th}\text{-drift corrected}}$ ) where sensitivity ( $S_i$ ) was defined as the analyzer response ( $A_i$ ) of  $i^{th}$  cylinder divided by its reference value ( $C_i$ ). Ratio in figure 1 denotes  $R_i$  given by equation (1).


$$R_{i,j} = \frac{S_{Sm_{i,j}}}{S_{j^{th}\text{-drift corrected}}} \quad (\text{eq.1})$$

where  $S_{Sm_{i,j}} = \frac{A_{Sm_{i,j}}}{C_{Sm_i}}$  for sample  $i$ ,  $S_{j^{th}\text{-drift corrected}} = \frac{S_{Rm_{j-1}} + S_{Rm_{j+1}}}{2}$  for analytical sequence  $j$

and  $S_{Rm_{j-1}}$  is a sensitivity of a reference for  $(j-1)$ -th analysis, defined as  $S_{Rm_{j-1}} = \frac{A_{Rm_{j-1}}}{C_{Rm}}$ .

All cylinders showed agreement with the gravimetric reference value within  $\pm 0.05\%$  uncertainty. The prepared mixtures are summarized in Table 4, where uncertainty includes uncertainty components generated from verification analysis and gravimetric weighing. Among the eight cylinders, six mixtures were used for this comparison.

All cylinders were returned with sufficient pressure and re-analyzed in October 2018. The results indicated that the mixtures remained stable during transport.



**Figure 2. Reanalysis results of the returned cylinders (October 2018)**

### Results and Discussion

Some important items reported by the participants are summarized in Table 5. They all prepared their own standards for calibration. SNSU-BSN, NIMT and KRISS used GC-TCD calibrated with single point, while others used GC-FID, where two of them calibrated their analyzers with a single point, only

KAZ used multiple points for calibration. The details of the analytical methods used by the participants are described in the individual participant reports.

**Table 5: Summary of the analysis methods of the participants**

| Laboratory | Cylinder | Measurement period | Calibration standards | Instrument calibration | Measurement technique |
|------------|----------|--------------------|-----------------------|------------------------|-----------------------|
| KRISS      | D581070  | Oct. 2018          | Own standards         | Single point           | GC/TCD                |
| SNSU-BSN   | D581092  | Feb. 2018          | Own standards         | Single point           | GC/TCD                |
| NIMT       | D581146  | Jan. 2018          | Own standards         | Single point           | GC/TCD                |
| NMC        | D581054  | Feb. 2018          | Own standards         | Single point           | GC/FID/Methaniser     |
| NPLI       | D581240  | Mar. 2018          | Own standards         | Single point           | GC/FID/Methaniser     |
| KAZ        | D581075  | Apr. 2018          | Own standards         | Multiple point         | GC/FID/Methaniser     |

The results of the comparison are summarized in Table 6.

**Table 6: Summary of the comparison of APMP.QM-S15**

| Lab.     | Cylinder | $X_{prep}$ | $u_{prep}$ | $x_{lab}$               | $U_{lab}$ | $k_{lab}$ | $\Delta x$              | $U(\Delta x)$ | $k$ |
|----------|----------|------------|------------|-------------------------|-----------|-----------|-------------------------|---------------|-----|
|          |          |            |            | [ $\mu\text{mol/mol}$ ] |           |           | [ $\mu\text{mol/mol}$ ] |               |     |
| KRISS    | D581070  | 999.94     | 0.25       | 999.86                  | 1.08      | 2         | -0.08                   | 1.19          | 2   |
| SNSU-BSN | D581092  | 999.87     | 0.25       | 1000.13<br>44           | 6.538     | 2         | 0.26                    | 6.56          | 2   |
| NIMT     | D581146  | 999.83     | 0.25       | 999.66                  | 1.60      | 2         | -0.17                   | 1.68          | 2   |
| NMC      | D581054  | 999.70     | 0.25       | 999.23                  | 2.16      | 2         | -0.47                   | 2.22          | 2   |
| NPLI     | D581240  | 999.29     | 0.25       | 999.58                  | 2.52      | 2         | 0.29                    | 2.57          | 2   |
| KAZ      | D581075  | 999.54     | 0.25       | 1003.0                  | 14.7      | 2         | 3.5                     | 14.7          | 2   |

As shown in table 6, all participants agreed with their SCR<sub>V</sub> within their associated uncertainties.

### Degrees of equivalence

The degree of equivalence ( $D_i$ ) of the comparisons is defined as


$$D_i(= \Delta x_i) = x_{i,lab} - x_{i,ref},$$

where  $x_{i,ref}$  denotes the supplementary comparison reference value and  $x_i$  the result of laboratory  $i$ .

The standard uncertainty of  $D_i$  can be expressed as

$$u^2(D_i) = u_{i,lab}^2 + u_{i,prep}^2$$

The degrees of equivalence (DoE) for the APMP.QM-S15 is presented in figure 3. As shown in figure 3, all results were consistent within the uncertainties  $u(D_i)$ .



**Figure 3: Degrees of equivalence for the APMP.QM-S15 ( $K=2$ )**

## Conclusions

In the comparison, all the results of the participants were consistent with their SCRv within the associated uncertainties.

## How Far Does the Light Shine?

The goal of this supplementary comparison is to support CMC claim for carbon dioxide in  $N_2$  at the range of 50  $\mu\text{mol/mol}$  - 500  $\text{mmol/mol}$ . An extended range may be supported as described in the GAWG strategy for comparisons and CMC claims [6].

Table 7: HFTLS list of each participant for CMC claims

| Participant |      | Amount fraction         | Uncertainty (%) | Amount fraction         | Uncertainty (%) |
|-------------|------|-------------------------|-----------------|-------------------------|-----------------|
|             |      | ( $\mu\text{mol/mol}$ ) |                 | ( $\mu\text{mol/mol}$ ) |                 |
| KRISS       | from | 1.08                    | 1.00            | 10                      | 0.11            |
|             | to   | 10                      | 0.11            | 500 000                 | 0.11            |
| SNSU-BSN    | from | 6.538                   | 0.99            | 10                      | 0.65            |
|             | to   | 10                      | 0.65            | 500 000                 | 0.65            |
| NIMT        | from | 1.6                     | 1.00            | 10                      | 0.16            |
|             | to   | 10                      | 0.16            | 500 000                 | 0.16            |
| NMC         | from | 2.16                    | 1.02            | 10                      | 0.22            |
|             | to   | 10                      | 0.22            | 500 000                 | 0.22            |
| NPLI        | from | 2.52                    | 0.99            | 10                      | 0.25            |
|             | to   | 10                      | 0.25            | 500 000                 | 0.25            |
| KAZ         | from | -                       |                 | 14.7                    | 1.47            |
|             | to   | -                       |                 | 500 000                 | 1.47            |

## References

- [1] A. Alink: The first key comparison of primary standard gas mixtures, *Metrologia* 37 (1). 2000
- [2] Rob M Wessel et al. International comparison CCQM-K52: Carbon dioxide in synthetic air, *Metrologia* 45 08011
- [3] Edgar Flores et al. International comparison CCQM-K120: (Carbon dioxide at background and urban level), *Metrologia* 56 1A
- [4] International organization for standardization, ISO 6142-1:2015. “Gas analysis — Preparation of calibration gas mixtures — Part 1: Gravimetric method for Class I mixtures, ISO, 2015 (E)
- [5] International organization for standardization, ISO 6143:2015. “Gas analysis – Comparison methods for determining and checking the composition of calibration gas mixtures”, ISO, 2015(E).
- [6] CCQM-GAWG strategy for comparisons and CMC claims, GAWG/19-41, [https://www.bipm.org/wg/CCQM/GAWG/Restricted/October\\_2019/GAWG19-41-CCQM-GAWG\\_strategy\\_for\\_comparisons\\_and\\_CMC\\_claims.pdf](https://www.bipm.org/wg/CCQM/GAWG/Restricted/October_2019/GAWG19-41-CCQM-GAWG_strategy_for_comparisons_and_CMC_claims.pdf)



## APMP-QM-S15: Carbon Dioxide in Nitrogen (1000 $\mu\text{mol/mol}$ )

---

Laboratory name: Gas Metrology Laboratory, National Metrology Centre, A\*STAR

Author: Liu Hui, Fang Jie, Thomas Wu, Kai Fuu Ming, Mou Jianqiang

Cylinder number: KRISS D581054

### Measurement #1

| Component       | Date<br>(dd/mm/yy) | Result<br>( $\mu\text{mol/mol}$ ) | Standard<br>Deviation<br>(% relative) | Number of<br>Replicates |
|-----------------|--------------------|-----------------------------------|---------------------------------------|-------------------------|
| CO <sub>2</sub> | 09/02/2018         | 999.022                           | 0.036                                 | 5                       |

### Measurement #2

| Component       | Date<br>(dd/mm/yy) | Result<br>( $\mu\text{mol/mol}$ ) | Standard<br>Deviation<br>(% relative) | Number of<br>Replicates |
|-----------------|--------------------|-----------------------------------|---------------------------------------|-------------------------|
| CO <sub>2</sub> | 12/02/2018         | 999.307                           | 0.080                                 | 5                       |

### Measurement #3

| Component       | Date<br>(dd/mm/yy) | Result<br>( $\mu\text{mol/mol}$ ) | Standard<br>Deviation<br>(% relative) | Number of<br>Replicates |
|-----------------|--------------------|-----------------------------------|---------------------------------------|-------------------------|
| CO <sub>2</sub> | 13/02/2018         | 999.353                           | 0.051                                 | 5                       |

### Results

| Component       | Result<br>( $\mu\text{mol/mol}$ ) | Expanded Uncertainty<br>( $\mu\text{mol/mol}$ ) | Coverage factor <sup>1</sup> |
|-----------------|-----------------------------------|---|------------------------------|
| CO <sub>2</sub> | 999.23                            | 2.16  | 2                            |

---

<sup>1</sup> The coverage factor shall be based on approximately 95% confidence.

### **Details of the Measurement Method Used:**

A customized FID-GC (Flame Ionised Detector - Gas Chromatography) with methanizer was used to conduct the comparison. One Reference Standard which is close to transfer standard's concentration were chosen to as the one-point calibration standard. The sample cylinder was analysed with the Reference Standard in the model of Reference Standard – Sample Cylinder – Reference Standard. The number of injections from each cylinder was 8, and only the last 5 injections were used for the calculation of the mole fraction of the sample cylinder. The Reference Standards and the sample cylinder were injected directly into the FID-GC through the sampling tube. Average results obtained in each individual analysis were combined and averaged to produce a single measurement result on that day.

The purity of balance gas, nitrogen and the sample gas, carbon dioxide was analysed using PDHID/FID-GC (Pulsed Discharge Ionization Detector/ Flame Ionised Detector - Gas Chromatograph). The regulator used was SS Verifo single stage without gauges, which was purged at least 10 times based on the standard operation procedure.

### **Details of Sample Handling:**

The sample cylinder and reference standards were stored at a room temperature ( $21 \pm 2$ ) °C for 3 days before an analysis. The gas mixture in the sample cylinder KRISS D581054 was analysed over 5 days against Reference Standard I maintained at NMC using FID-GC and a sampling system consisting of valves, pressure regulator and flow meter. Modified Teflon was used in the sampling line. The measurements were carried out under ambient temperature of ( $21 \pm 2$ ) °C and (60

± 15) % relative humidity.

**Details of the Reference Standards Used:**

Reference Standards used for the comparison were maintained in 2 separate cylinders with the following details:

| <b>Reference Standard</b> | <b>Cylinder No.</b> | <b>Concentration (µmol/mol)</b> | <b>Gravimetric Preparation Uncertainty (µmol/mol)</b> |
|---------------------------|---------------------|---------------------------------|---|
| Reference Standard        | PSM218695           | 1000.93                         | 0.41  |

Reference Standard, which was selected as the Reference Standard in Measurement #1, Measurement #2 and Measurement #3 to measure the concentration of the sample cylinder, were prepared in NMC using gravimetric method following ISO6142 Standard.

Cylinders for Reference Standard was 5-litre aluminium cylinder with Aculife-3 treatment supplied by Scott Specialty Gases. The Reference Standards had been verified against Consistency Check with internal reference materials. The concentration of Reference Standard was the concentration after verification. The preparation of gas mixtures and measurements were carried out under ambient

temperature of  $(21 \pm 2)$  °C and  $(60 \pm 15)$  % relative humidity based on the standard operation procedure.

### Details on Uncertainty Budget:

The purity analysis of carbon dioxide and nitrogen were measured by PDHID/FID-GC. The core impurities e.g. O<sub>2</sub>, CO, CO<sub>2</sub>, N<sub>2</sub>, Ar, H<sub>2</sub>, CH<sub>4</sub>, etc., were analysed.

#### 1. Uncertainty Evaluation of Reference Standard

Two type of uncertainty were evaluated for the Combined Uncertainty of Reference Standard as below.

- Uncertainty of pure CO<sub>2</sub> and the matrix gas, pure nitrogen
- Uncertainty of reference gas mixtures by gravimetric method

#### 1.1. Uncertainty Budget of Pure CO<sub>2</sub>

| Components |                               | Concentration<br>(mol/mol) | Distribution | Standard Uncertainty<br>(mol/mol) |
|------------|-------------------------------|----------------------------|--------------|-----------------------------------|
| Impurity   | N <sub>2</sub>                | 2.000E-07                  | Normal       | 3.325E-09                         |
| Impurity   | O <sub>2</sub>                | 6.000E-08                  | Normal       | 3.139E-09                         |
| Impurity   | CH <sub>4</sub>               | 5.000E-08                  | Rectangular  | 2.887E-08                         |
| Impurity   | C <sub>2</sub> H <sub>4</sub> | 5.000E-08                  | Rectangular  | 2.887E-08                         |
| Impurity   | H <sub>2</sub> O              | 2.500E-06                  | Rectangular  | 1.443E-06                         |
| Impurity   | CO                            | 2.500E-08                  | Rectangular  | 1.443E-08                         |

|             |                 |            |        |           |
|-------------|-----------------|------------|--------|-----------|
| Impurity    | H <sub>2</sub>  | 1.200E-07  | Normal | 6.928E-08 |
| Impurity    | Ar              | 9.000E-08  | Normal | 6.213E-09 |
| Balance gas | CO <sub>2</sub> | 0.99999691 |        | 1.446E-06 |

### 1.2. Uncertainty Budget of Pure N<sub>2</sub>

| Components  |                  | Concentration<br>(mol/mol) | Distribution | Standard Uncertainty<br>(mol/mol) |
|-------------|------------------|----------------------------|--------------|-----------------------------------|
| Impurity    | O <sub>2</sub>   | 2.500E-08                  | Rectangular  | 1.44342E-08                       |
| Impurity    | CO               | 2.500E-08                  | Rectangular  | 1.44342E-08                       |
| Impurity    | H <sub>2</sub>   | 2.500E-08                  | Rectangular  | 1.44342E-08                       |
| Impurity    | CO <sub>2</sub>  | 5.000E-08                  | Rectangular  | 2.88684E-08                       |
| Impurity    | CH <sub>4</sub>  | 5.000E-08                  | Rectangular  | 2.88684E-08                       |
| Impurity    | H <sub>2</sub> O | 1.000E-08                  | Rectangular  | 5.77367E-09                       |
| Balance gas | N <sub>2</sub>   | 0.999999815                |              | 4.82197E-08                       |

### 1.3. Gravimetric Uncertainty Budget for Reference Standard

Gravimetric uncertainty of the Reference Standard at 3% CO<sub>2</sub> in nitrogen mixture was evaluated in the below tables.

| <b>Uncertainty source</b>  | <b>Estimated Value</b> | <b>Standard Uncertainty</b> | <b>Distribution</b> | <b>Contribution to Standard Uncertainty (<math>\mu</math> mol/mol)</b> |
|--|------------------------|-----------------------------|---------------------|--|
| Mass of CO <sub>2</sub> (g)  | 26.696                 | 0.0065                      | normal              | 7.17   |
| Mass of N <sub>2</sub> (g)   | 545.398                | 0.0082                      | normal              | 0.44   |
| Concentration of CO <sub>2</sub> in pure CO <sub>2</sub> gas (mol/mol) | 9.9999696E-01          | 2.04E-06                    | normal              | 0.062  |
| Concentration of CO <sub>2</sub> in N <sub>2</sub> gas (mol/mol)       | 5.00E-08               | 2.88E-08                    | normal              | 0.028  |
| Molar mass of CO <sub>2</sub> (g/mol)                                  | 44.00900               | 0.00072                     | normal              | 0.48   |
| Molar mass of N <sub>2</sub> (g/mol)                                   | 28.01400               | 0.00049                     | normal              | 0.51   |
| <b>Combined Uncertainty (k = 1)</b>                                    |                        |                             |                     | <b>7.2</b>   |

Gravimetric uncertainty budget for Reference Standard at 1000.93  $\mu\text{mol/mol}$ , which was diluted from the 3%  $\text{CO}_2$  premix gas.

| Uncertainty source   | Estimated Value | Standard Uncertainty | Distribution | Contribution to Standard Uncertainty ( $\mu\text{mol/mol}$ ) |
|--|-----------------|----------------------|--------------|--|
| Mass of premix (g)   | 19.270          | 0.0065               | normal       | 0.33   |
| Mass of $\text{N}_2$ (g)                                     | 552.970         | 0.0082               | normal       | 0.014  |
| Concentration of $\text{CO}_2$ in premix gas (mol/mol)       | 3.02E-02        | 7.22E-06             | normal       | 0.24   |
| Concentration of $\text{CO}_2$ in $\text{N}_2$ gas (mol/mol) | 5.00E-08        | 2.88E-08             | normal       | 0.028  |
| Molar mass of premix (g/mol)                                 | 28.49704        | 0.00048              | normal       | 0.016  |
| Molar mass of $\text{N}_2$ (g/mol)                           | 28.01400        | 0.00049              | normal       | 0.017  |
| <b>Combined Uncertainty (k = 1)</b>                          |                 |                      |              | <b>0.41</b>  |

## 2. Uncertainty Evaluation for the Measurement

The GC analyser was calibrated with the calibration standard gas prepared by the gravimetric method. The A-B-A method and the one the point calibration model was used.

The concentration of sample gas was determined by the following equation:

$$X_{sample} = \frac{Y_{sample}}{Y_{std}} X_{std}$$

*Where,  $X_{sample}$ : Concentration of sample*

*$Y_{sample}$ : GC analysis results of the sample cylinder*

*$Y_{std}$  : GC analysis results of Reference Standard*

*$X_{std}$  : Concentration of Reference Standard*

The uncertainties of  $Y_{sample}$  and  $Y_{std}$  have been estimated using the pooled standard deviation of analysis. The uncertainties of the  $X_{std}$  has been estimated by the uncertainty of standard concentration in preparation including the uncertainties of the standard concentration in gravimetric process and purity analysis, verification and stability check. The reproducibility of the measurements was estimated by the standard deviation of the pooled mean value of the  $X_{sample}$ .



The uncertainty of  $X_{std}$  was the combined by the uncertainties of the standard concentration in gravimetric process and purity analysis, verification and stability check. As the CO<sub>2</sub> gas mixtures which NMC used for this comparison were new prepared. We estimated the uncertainty of stability was negligible.

$$u^2(X_{std}) = u^2(X_{std,prep}) + u^2(X_{std,veri})$$

**and:**  $u^2(X_{std,prep}) = u^2(X_{std,gravi}) + u^2(X_{std,pur}) + u^2(X_{std,stab})$

**If:**  $|X_{std,prep} - X_{std,veri}| \leq 2\sqrt{u_{std,prep}^2 + u_{std,veri}^2}$

**Then:**  $X_{std} = X_{std,prep}$

Where,  $u(X_{std})$  : Uncertainty of Standard concentration

$u(X_{std,prep})$  : Uncertainty of Standard concentration in preparation

$u(X_{std,veri})$  : Uncertainty of Standard concentration in analytical verification

$u(X_{std,gravi})$  : Uncertainty of Standard concentration in gravimetric gas mixing process

$u(X_{std,pur})$  : Uncertainty of purity analysis

$u(X_{std,stab})$  : Uncertainty of stability check

| Uncertainty Evaluation for the Measurement                    |         |                      |              |                         |  |
|---|---------|----------------------|--------------|-------------------------|--|
| Uncertainty Source  | Value   | Standard Uncertainty | Distribution | Sensitivity Coefficient | Uncertainty Contribution ( $\mu\text{mol/mol}$ ) |
| $Y_{\text{sample}}$   | 1097.35 | 0.59                 | Normal       | 0.911                   | 0.54   |
| $Y_{\text{std}}$  | 1099.22 | 0.29                 | Normal       | -0.909                  | 0.26   |
| $X_{\text{std}}$  | 1000.93 | 0.91                 | Normal       | 0.998                   | 0.90   |
| Reproducibility   |         | 0.103                | Normal       | 1                       | 0.103  |
| Combined Uncertainty ( $\mu\text{mol/mol}$ )                  |         |                      |              |                         | 1.08   |
| Expanded Combined Uncertainty ( $\mu\text{mol/mol}$ ) ; K = 2 |         |                      |              |                         | 2.16   |
| Expanded Combined Uncertainty (Relative %) ; K = 2            |         |                      |              |                         | 0.22%  |

Based on the above uncertainty model, the final uncertainty evaluation for the measurement was evaluated as shown in the below table.

Measurand: 999.23  $\mu\text{mol/mol}$

Coverage factor: k = 2

Expanded Uncertainty: 2.16  $\mu\text{mol/mol}$ , relative 0.22%

# Report Form

## Carbon dioxide in nitrogen

---

Laboratory name: CSIR-NPLI

Cylinder number: D 581240

### Measurement #1

| Component | Date (dd/mm/yy) | Result ( $\mu\text{mol/mol}$ ) | Standard deviation (% relative) | number of replicates |
|-----------|-----------------|--------------------------------|---------------------------------|----------------------|
| CO2       | 05/03/18        | 998.57                         | 0.09                            | 8                    |

### Measurement #2

| Component | Date (dd/mm/yy) | Result ( $\mu\text{mol/mol}$ ) | Standard deviation (% relative) | number of replicates |
|-----------|-----------------|--------------------------------|---------------------------------|----------------------|
| CO2       | 06/03/18        | 999.43                         | 0.17                            | 8                    |

### Measurement #3<sup>2</sup>

| Component | Date (dd/mm/yy) | Result ( $\mu\text{mol/mol}$ ) | Standard deviation (% relative) | number of replicates |
|-----------|-----------------|--------------------------------|---------------------------------|----------------------|
| CO2       | 06/03/18        | 1000.73                        | 0.16                            | 8                    |

## Results

| Component | Result ( $\mu\text{mol/mol}$ ) | Expanded Uncertainty ( $\mu\text{mol/mol}$ ) | Coverage factor <sup>3</sup> |
|-----------|--------------------------------|--|------------------------------|
| CO2       | 999.58                         | 2.52   | 2                            |

---

<sup>2</sup> If more than three measurements are taken, please copy and insert a table of the appropriate format as necessary

<sup>3</sup> The coverage factor shall be based on approximately 95% confidence.

### **Details of the measurement method used:**

GC FID (Agilent 6890N) with Methanizer

Column used: Haysep D; length 12 ft, Dia 1/8" and mesh range 100/120

Oven temp: 80 °C

Carrier gas: He (20 ml/min)

Methanizer temp: 350 °C

Detector Temp: 250 °C

GSV loop: 0.25 ml

Hydrogen and air flow rate were 20 ml/min and 300 ml/min respectively

The APMP.QM-S15 gas cylinder was maintained inside a laboratory at a nominal temperature for  $22 \pm 5^\circ\text{C}$  for all the period of its storage at NPL India. A dual stage regulator is fitted on the cylinder to inject the gas sample through GSV into the GC-FID system for its analysis. The cylinders were rolled for two hours on homogenization system before measurement.

### **Details of the calibration method used:**

Single point calibration method was used for the analysis of the inter-comparison cylinder. Calibration standard of concentration  $1095.64 \pm 2.59 \mu\text{mol/mol}$  is used for

the calibration of GC-FID system during the analysis of APMP QM-S15 cylinder and value evaluation.

### **Details of the standards used:**

The preparation of Primary Reference Gas Mixtures (PRGM) was done in accordance to ISO 6142: Gas Analysis -Preparation of calibration gas mixtures - Gravimetric Method.

The preconditioning of 10 litre aluminium cylinder was done by evacuation (filling of N<sub>2</sub> gas + evacuation + heating at 60-70 °C & evacuation) of cylinders. This process has been repeated three times for each cylinder before preparation of gas mixture. The evacuation of cylinders is carried out using PFEIFFER HiCube 80 Eco vacuum System. The theoretical calculations for the calculation of mole fraction were carried out for the desired concentrations using model equation from ISO 6142-1:2015.

Gas mixtures of CO<sub>2</sub> in nitrogen gas from pure gas were prepared in two series in the concentrations around  $20544 \pm 33.58$  and  $20392.67 \pm 34.73$   $\mu\text{mol/mol}$ . The pre-mixture of  $20392.67 \pm 34.73$   $\mu\text{mol/mol}$  was used for further dilution in the concentration  $1095.64 \pm 2.59$   $\mu\text{mol/mol}$  for APMP.QM S15 cylinder measurement. The initial weighing of components transferred was done using a top pan balance. And the final weighing was done using an equal arm double pan balance Raymor HCE 25G max capacity 25kg with 1mg sensitivity. These cylinders were validated in accordance to ISO 6143:2001 "Gas analysis - Comparison method for determining and checking the composition of calibration gas mixtures". Thus the prepared gas mixtures were certified as CO<sub>2</sub> in Nitrogen gas (Primary Reference Gas Mixtures (PRGMs)).

## **Purity Analysis**

The purity of N<sub>2</sub> parent gases was determined using tiger optics CRDS analyzers model for the following H<sub>2</sub>O, CH<sub>4</sub> and CO gas components. The moisture of the gases was determined using Tiger Optics moisture analyzer model Laser Trace. CH<sub>4</sub> was determined using Tiger Optics methane analyzer model MTO-1000-CH4 and CO gas was determined using Tiger Optics CO analyzer model HALO 3-CO.

### **Details on uncertainty budget:**

The Uncertainty for the prepared gas mixtures has been evaluated according to guideline prescribed in ISO 6142-1:2015 and EURACHEM Guide taking account of following gravimetric and analytical components:

- I. Uncertainty Components in Gravimetric Preparation of calibration gas mixture (Calibration standard)
  1. Raymor Balance
  2. Mass Pieces
  3. Buoyancy effect
  4. Handling of cylinder

5. Residual gas
6. Expansion of the cylinder due to filling of gas at High pressure

## II. Uncertainty Components in Analytical method

- Repeatability
- Reproducibility
- GC Response

Pooled standard deviation is taken as standard uncertainty of assigned value and GC response taking account of repeatability and reproducibility.

| Date of Analysis | n <sub>i</sub> | X <sub>i</sub><br>(μmol/mol) | (SD) <sub>i</sub><br>(μmol/mol) | RSD (%) | SD <sub>pooled</sub> |
|------------------|----------------|------------------------------|---------------------------------|---------|----------------------|
| 05-03-2018       | 8              | 998.57                       | 0.94                            | 0.09    |                      |
| 06-03-2018       | 8              | 999.43                       | 1.67                            | 0.17    |                      |
| 06-03-2018       | 8              | 1000.73                      | 1.63                            | 0.16    |                      |
| <b>RESULT</b>    | <b>24</b>      | <b>999.58</b>                |                                 |         | <b>0.32</b>          |

$$SD_{pooled} = \sqrt{\frac{s_1^2(n_1 - 1) + s_2^2(n_2 - 2) + \dots}{(n_1 + n_2 + n_3) - 3}}$$

### Measurement Uncertainty Budget:

| Sources of Uncertainty               | Estimates $x_i$             | Distribution/ Type A & B | Standard uncertainty $u(x_i)$ | Sensitivity coefficient $c_i$ | Contribution to standard uncertainty $u_i(y)$ |
|--------------------------------------|-----------------------------|--------------------------|-------------------------------|-------------------------------|---|
| Assigned value                       | 999.58 $\mu\text{mol/mol}$  | Normal, Type A           | 0.32 $\mu\text{mol/mol}$      | 1                             | 0.00032                                       |
| Conc. of Cal Std                     | 1095.64 $\mu\text{mol/mol}$ | Normal, Type B, 2        | 1.29 $\mu\text{mol/mol}$      | 1                             | 0.00118                                       |
| GC Response                          | 2064.50 mV                  | Normal, Type A           | 0.67 mV                       | 1                             | 0.00033                                       |
| Combined standard Uncertainty, $u_c$ | 1.26 $\mu\text{mol/mol}$    |                          |                               |                               |   |
| Expanded Uncertainty, U              | 2.52 $\mu\text{mol/mol}$    | $k = 2$                  |                               |                               |   |
| U                                    | 0.25 %                      |                          |                               |                               |   |

**Cylinder Pressure after Analysis ~ 90 bar**

**Team Members: Dr Daya Soni, Dr Khem Singh, Ms Sulakshina Bhat, Dr Shankar G Aggarwal and Dr Prabha Johri.**



**REPORT ON  
APMP- QM-S15**

**APMP Regional Comparison  
Carbon Dioxide in Nitrogen (1000  $\mu\text{mol/mol}$ )**

Oman Zuas,  
Harry Budiman,  
Muhammad Rizky Mulyana

Research Centre for Metrology-Indonesian Institute of Sciences (SNSU-BSN). Building 420,  
Kawasan PUSPIPTEK Serpong 15314, Tangerang Selatan, Banten, Indonesia

02 March 2018

# Report Form

## Carbon dioxide in nitrogen

---

Laboratory name: **Research Centre for Metrology-Indonesian Institutes of Sciences (SNSU-BSN)**

Cylinder number: **D581092 (APMP QM S-15)**

### Measurement:

#### Measurement #1

| Component       | Date (dd/mm/yy) | Result ( $\mu\text{mol/mol}$ ) | Standard deviation (% relative) | number of replicates |
|-----------------|-----------------|--------------------------------|---------------------------------|----------------------|
| CO <sub>2</sub> | 07/02/2018      | 999.959                        | 0.645                           | 3                    |

#### Measurement #2

| Component       | Date (dd/mm/yy) | Result ( $\mu\text{mol/mol}$ ) | Standard deviation (% relative) | number of replicates |
|-----------------|-----------------|--------------------------------|---------------------------------|----------------------|
| CO <sub>2</sub> | 08/02/2018      | 999.936                        | 0.575                           | 3                    |

#### Measurement #3

| Component       | Date (dd/mm/yy) | Result ( $\mu\text{mol/mol}$ ) | Standard deviation (% relative) | number of replicates |
|-----------------|-----------------|--------------------------------|---------------------------------|----------------------|
| CO <sub>2</sub> | 16/02/2018      | 1000.307                       | 0.572                           | 5                    |

#### Measurement #4

| Component       | Date (dd/mm/yy) | Result ( $\mu\text{mol/mol}$ ) | Standard deviation (% relative) | number of replicates |
|-----------------|-----------------|--------------------------------|---------------------------------|----------------------|
| CO <sub>2</sub> | 17/02/2018      | 1000.528                       | 0.718                           | 6                    |

#### Measurement #5

| Component       | Date (dd/mm/yy) | Result ( $\mu\text{mol/mol}$ ) | Standard deviation (% relative) | number of replicates |
|-----------------|-----------------|--------------------------------|---------------------------------|----------------------|
| CO <sub>2</sub> | 18/02/2018      | 999.834                        | 0.360                           | 7                    |

## Results

| Component       | Date<br>(dd/mm/yy) | Result<br>( $\mu\text{mol/mol}$ ) | Expanded Uncertainty<br>( $\mu\text{mol/mol}$ ) | Coverage<br>factor <sup>4</sup> |
|-----------------|--------------------|-----------------------------------|---|---------------------------------|
| CO <sub>2</sub> | 01/03/2018         | 1000.1344                         | 6.538   | <i>k</i> = 2                    |

## Details of the measurement method used:

### Reference Method

Gas chromatography equipped with thermal conductivity detector (GC-TCD).

### Instruments

Gas chromatography equipped with thermal conductivity detector (GC-TCD) was used to determine the concentration of CO<sub>2</sub> in gas mixtures. Separation of CO<sub>2</sub> from the gas mixture was conducted on a stainless steel packed column (Porapak Q, 6 feet, 1/8" outer diameter). The oven temperature was isothermal at 40°C. The TCD gas used was He at 20 mL/min and 7 mL/min as reference and make-up gases, respectively. The TCD temperature was kept at 250°C with negative polarity. Ultra high purity of helium (99.999%) was used as a carrier gas at a flow rate of 28 mL/min. The valve box temperature was maintained at 100°C. The flow rate of gas mixture was set at 30 mL/min (checked at GC gas outlet by using a digital flow meter) and the gas mixture was passed through on a 500  $\mu\text{L}$  sample loop. A mass flow controller was used to keep the gas mixture flow at constant rate.

## Details of the calibration method used:

The calibration standard gas mixtures (hereinafter called as CSGMs) of CO<sub>2</sub> in N<sub>2</sub> were prepared by SNSU-BSN using gravimetric method in accordance to ISO 6142:2001<sup>(1)</sup>. The pre-mixtures were prepared from CO<sub>2</sub> (ultra-high purity grade, Air Liquid Indonesia) and N<sub>2</sub> (ultra-high purity grade, SII-Indonesia). The purity (compositions) assessment of CO<sub>2</sub> and N<sub>2</sub> were conducted prior to use and the result are presented in Table 1 and Table2, respectively. Two-step dilution processes (Figure 1) were adopted to prepare each of six cylinders of CSGMs

---

<sup>4</sup> The coverage factor shall be based on approximately 95% confidence.


containing CO<sub>2</sub> in N<sub>2</sub> with a nominal concentration ~1000 μmol/mol. After that, the gravimetric concentrations of CO<sub>2</sub> in all prepared CSGM cylinders were verified using the method described in ISO 6143:2001<sup>(2)</sup>, and the results are presented in Table 3.

**Table 1.** Purity table of pure CO<sub>2</sub> (parent)

| Component        | Mole fraction (μmol/mol) | Standard uncertainty (μmol/mol) | Method of evaluation (type A or type B) | Analysis method            |
|------------------|--------------------------|---------------------------------|---|----------------------------|
| H <sub>2</sub> O | 1.000                    | 0.577                           | B                                       | Manufacturer specification |
| CO               | 0.846                    | 0.073                           | A                                       | PDHID                      |
| O <sub>2</sub>   | 0.846                    | 0.063                           | A                                       | PDHID                      |
| CH <sub>4</sub>  | 0.306                    | 0.021                           | A                                       | PDHID                      |
| Ar               | 0.495                    | 0.038                           | A                                       | PDHID                      |
| N <sub>2</sub>   | 2.254                    | 0.167                           | A                                       | PDHID                      |
| CO <sub>2</sub>  | 999994.253               | 0.610                           | A                                       | Mass balance               |

**Table 2.** Purity table of pure N<sub>2</sub> (parent)

| Component        | Mole fraction (μmol/mol) | Standard uncertainty (μmol/mol) | Method of evaluation (type A or type B) | Analysis method            |
|------------------|--------------------------|---------------------------------|---|----------------------------|
| H <sub>2</sub> O | 1.500                    | 0.866                           | B                                       | Manufacturer specification |
| CO <sub>2</sub>  | 0.022                    | 0.012                           | A                                       | PDHID                      |
| O <sub>2</sub>   | 5.846                    | 0.348                           | A                                       | PDHID                      |
| CH <sub>4</sub>  | 2.685                    | 0.231                           | A                                       | PDHID                      |
| Ar               | 0.915                    | 0.064                           | A                                       | PDHID                      |
| CO               | 1.320                    | 0.114                           | A                                       | PDHID                      |
| N <sub>2</sub>   | 999987.712               | 0.970                           | A                                       | Mass balance               |



|                            |                            |                            |                            |                             |                             |
|----------------------------|----------------------------|----------------------------|----------------------------|-----------------------------|-----------------------------|
| <b>975.231</b><br>μmol/mol | <b>985.929</b><br>μmol/mol | <b>995.441</b><br>μmol/mol | <b>999.208</b><br>μmol/mol | <b>1005.107</b><br>μmol/mol | <b>1010.438</b><br>μmol/mol |
|----------------------------|----------------------------|----------------------------|----------------------------|-----------------------------|-----------------------------|

**Figure 1.** Two step dilution process of CSGMs at ~1000 μmol/mol CO<sub>2</sub> in N<sub>2</sub>

**Table 3.** Calibration standard gas mixture (CSGMs)

| # Cylinder        | Concentration<br>(μmol/mol) | U <sub>combined</sub><br>(μmol/mol) | U <sub>expanded</sub> *<br>(μmol/mol) | U <sub>expanded</sub> relative<br>(%) |
|-------------------|-----------------------------|-------------------------------------|---------------------------------------|---------------------------------------|
| L150721016        | 975.231                     | 3.056                               | 6.111                                 | 0.627                                 |
| L150721015        | 985.929                     | 3.091                               | 6.182                                 | 0.627                                 |
| L150721006        | 995.441                     | 3.119                               | 6.237                                 | 0.627                                 |
| <b>L150721012</b> | <b>999.208</b>              | <b>3.212</b>                        | <b>6.423</b>                          | <b>0.643</b>                          |
| L150721005        | 1005.107                    | 3.170                               | 6.341                                 | 0.631                                 |
| L150721003        | 1010.438                    | 3.170                               | 6.340                                 | 0.627                                 |

\* The coverage factor ( $k=2$ ) was based on approximately 95% confidence level.

### Weighing Data

Weighing data for cylinder #L150721012 are summarized as follows:

1<sup>st</sup> dilution mixture:

1. Evacuated cylinder #L150721004 - tare cylinder = 70.210 g (cylinder #L150721004)

2. Cylinder #L150721004 filled with parent CO<sub>2</sub> - tare cylinder = 80.185 g (amount of parent CO<sub>2</sub> transferred into #L150721004 = 9.974 g)
3. Cylinder #L150721004 filled with parent N<sub>2</sub> - tare cylinder = 299.669 g (amount of parent N<sub>2</sub> transferred into #L150721004 = 219.462 g)
4. Cylinder #L150721004 filled with parent CO<sub>2</sub> (9.974 g) + parent N<sub>2</sub> (219.462 g) = 1<sup>st</sup> dilution mixture (cylinder #L150721004).

2<sup>nd</sup> dilution mixture:

1. Evacuated cylinder #L150721012 - tare cylinder = 3.260 g (cylinder #L150721012)
2. Cylinder #L150721012 filled with 1<sup>st</sup> dilution mixture #L150721004 - tare cylinder = 13.445 g (amount of 1<sup>st</sup> dilution mixture #L150721004 transferred into #L150721012 = 10.183 g)
3. Cylinder #L150721012 filled with parent N<sub>2</sub> - tare cylinder = 285.450 g (amount of N<sub>2</sub> transferred into cylinder #L150721012 = 271.975 g)
4. Cylinder #L150721012 filled with 1<sup>st</sup> dilution mixture #L150721004 (10.183 g) + parent N<sub>2</sub> (271.975 g) = 2<sup>nd</sup> dilution mixture (cylinder #L150721012)

## **Details of the standards used for instrument calibration:**

Preliminary evaluation of CO<sub>2</sub> concentration in the sample cylinder #D581092 was performed by constructing a calibration curve using the six prepared CSGMs. Such evaluation was conducted under identical conditions and for a CSGM having a GC signal response close to that of sample cylinder #D581092 was selected for single-point calibration to determine the concentration of CO<sub>2</sub> in sample cylinder. Our evaluation showed that the CO<sub>2</sub> in cylinder #L150721012 was found to be the closest GC signal response relative to that of GC signal response of CO<sub>2</sub> in sample cylinder #D581092. Therefore, the CSGM cylinder #L150721012 was chosen as a reference standard for the single-point calibration process to determine the CO<sub>2</sub> concentration in the sample #D581092.

The single-point calibration consists of several sets of measurement (at least three set of measurements) in different days. The order of measurement was A-B-A (where A is the cylinder #L150721012 as reference standard, and B is the sample cylinder #D581092). Each set of measurement comprised of at least seven replications of analysis and the first injection was excluded from measurement repeatability evaluation. The mathematical model (Eq. 5) was used to calculate the concentration of CO<sub>2</sub> in sample cylinder #D581092.

### Sample handling

The sample cylinder #D581092 was conditioned in the laboratory environmental by keeping the sample cylinder in the laboratory for 48 h. Each cylinder (CSGMs #L150721012 .and sample #D581092) was equipped with an Alphagaz double stage pressure regulator that was adequately purged.

### Details of uncertainty budgets:

#### Uncertainty evaluation for the prepared CSGM #L150721012 .

- **Model equation:** a model formula (Eq. 1) below was used to calculate CO<sub>2</sub> concentration in the prepared CSGMs (measurand). The concentrations of CO<sub>2</sub> in in CSGMs ( $C_{CO_2}$ ) were calculated as the gravimetric concentration based on ISO 6142 using equation 1 (Eq. 1).

$$C_{CO_2} = \frac{\sum_{A=1}^P \left[ \frac{x_{CO_2,A} \cdot m_A}{\sum_{i=1}^n x_{i,A} \cdot M_i} \right]}{\sum_{A=1}^P \left[ \frac{m_A}{\sum_{i=1}^n x_{i,A} \cdot M_i} \right]}$$

(1)



The  $C_{CO_2}$  was calculated as the mole of the total  $CO_2$  transferred from each parent gas (  $\sum_{A=1}^P \left[ \frac{x_{CO_2,A} \cdot m_A}{\sum_{i=1}^n x_{i,A} \cdot M_i} \right]$  ) divided by the total mole of gas components in the CSGM cylinder (  $\sum_{A=1}^P \left[ \frac{m_A}{\sum_{i=1}^n x_{i,A} \cdot M_i} \right]$  ). Notation  $A$  corresponds to the parent gases in the amount of  $P$ , while  $i$  is corresponding to each gas components in the mixture with fraction of  $x_i$ , including the impurities, in a total of  $n$  components.  $M_i$  is the molecular mass of each component and  $m_A$  is the mass of transferred parent gas.

- **Uncertainty budgets.** For the uncertainty estimation of the CSGM #L150721012, the uncertainty contributors are including gravimetric uncertainty ( $u_{grav}$ ), uncertainty from verification ( $u_{ver}$ ), and uncertainty form stability ( $u_{stab}$ ). For that, the combined uncertainty of the CSGMs #L150721012 was calculated by means of equation 2 (Eq. 2).

$$u_{CO_2} = \sqrt{u_{grav}^2 + u_{ver}^2 + u_{stab}^2}$$

(2)

For the uncertainty from gravimetric preparation ( $u_{grav}$ ) the estimation was done by modifying Eq. 1 based on the propagation rules”, resulting in an equation 3 (Eq. 3) below.

$$u^2(C_{CO_2}) = \sum_{A=1}^P \left[ \frac{\partial x_{CO_2}}{\partial m_A} \right]^2 \cdot u^2(m_A) + \sum_{A=1}^P \left[ \frac{\partial x_{CO_2}}{\partial M_i} \right]^2 \cdot u^2(M_{CO_2}) + \sum_{A=1}^P \sum_{i=1}^n \left[ \frac{\partial x_{CO_2}}{\partial x_{i,A}} \right]^2 \cdot u^2(x_{CO_2,A})$$

(3)

where  $u^2(m_A)$  is the uncertainty from the weighing of the transferred parent gas  $A$ . The  $u^2(M_i)$  is the uncertainty of molecular mass for all gas components  $i$  in the mixture. The  $u^2(x_{i,A})$  is the uncertainty of the mole fraction for all of gas components  $i$ , including the impurities of the parent gas  $A$ .

Moreover, the uncertainty from the verification ( $u_{ver}$ ) was estimated from the standard deviations of the CSGM verification. The uncertainty from the stability of the CSGM ( $u_{stab}$ ) was estimated from the concentration difference between some days of measurement.


For the uncertainty of weighing process of the transferred parent gas ( $u^2(m_A)$ ), it was estimated by using following equation 4 (Eq. 4).

$$u^2_{m_A} = (\Delta w_A - \Delta w_{A-1})^2 u^2(e) + (-e)^2 u^2(\Delta w_A) + (-e)^2 u^2(\Delta w_{A-1}) + (\Delta P \rho_{air})^2 u^2(K) + (K \rho_{air})^2 u^2(\Delta P) + u^2(\Delta L) + (\rho_{air})^2 u^2(\delta V) + (K \Delta P + \delta V)^2 u^2(\rho_{air}) \quad (4)$$

where:

- $u^2(e)$  is the uncertainty of the linearity of the balance.
- $u^2(\Delta w_A)$  is the repeatability of mass difference between the tare cylinder and CSGM cylinder after the transfer of parent gas by repeated weighing.
- $u^2(\Delta w_{A-1})$  is the repeatability of mass difference between the tare cylinder and CSGM cylinder before the transfer of parent gas by repeated weighing.
- $u^2(K)$  is the uncertainty caused by expansion of cylinder volume due to pressure change.
- $u^2(\Delta P)$  is the uncertainty from the pressure change due to gas transferring.
- $u^2(\Delta L)$  is the uncertainty caused by random loss of mass or sticking dirt in cylinder's wall.
- $u^2(\delta V)$  is the uncertainty from the volume change due to temperature rise during transfer of parent gas.
- $u^2(\rho_{air})$  is the uncertainty from the buoyancy correction caused by air density change in the weighing chamber.

By combining those three aforementioned equations represented by the Eq. 2, Eq. 3 and Eq. 4 and applying them to evaluate the sources of uncertainty of the CSGM concentration, the below fishbone diagram was obtained.



**Figure 1.** Fishbone diagram of uncertainty sources affecting the final concentration of CSGMs

**Table 4.** Uncertainty budgets for the CSGM #L150721012

| Uncertainty sources   | Value      | Standard Uncertainty | Type |
|---|------------|----------------------|------|
| Gravimetric preparation of CSGM;<br>Combined from :   |            |                      |      |
| 1. Weighing of the transferred parent gas (a combination of the uncertainty sources stated in equation 4) : |            |                      |      |
| a. Weighing of transferred parent gas #L150721004   | 10.1832 g  | 0.0064 g             | A    |
| b. Weighing of transferred parent gas N <sub>2</sub>  |            |                      |      |
| 2. Mole fraction of components in parent gas #L150721004 and parent N <sub>2</sub> :                        | 271.9754 g | 0.0075 g             | A    |

|   |                    |                    |   |
|---|--------------------|--------------------|---|
| a. H <sub>2</sub> O in #L150721004                                    |                    |                    |   |
| b. CO in #L150721004  |                    |                    |   |
| c. CO <sub>2</sub> in #L150721004                                     |                    |                    |   |
| d. O <sub>2</sub> in #L150721004                                      |                    |                    |   |
| e. Ar in #L150721004  |                    |                    |   |
| f. CH <sub>4</sub> in #L150721004                                     |                    |                    |   |
| g. N <sub>2</sub> in #L150721004                                      | 0.00000149 mol/mol | 0.00000084 mol/mol | B |
| h. H <sub>2</sub> O in parent N <sub>2</sub>                          |                    |                    |   |
| i. CO in parent N <sub>2</sub>  | 0.00000131 mol/mol | 0.00000010 mol/mol | B |
| j. CO <sub>2</sub> in parent N <sub>2</sub>                           | 0.02811440 mol/mol | 0.00001024 mol/mol | B |
| k. O <sub>2</sub> in parent N <sub>2</sub>                            |                    |                    |   |
| l. Ar in parent N <sub>2</sub>  | 0.00000571 mol/mol | 0.00000034 mol/mol | B |
| m. CH <sub>4</sub> in parent N <sub>2</sub>                           |                    |                    |   |
| n. N <sub>2</sub> in parent N <sub>2</sub>                            | 0.00000091 mol/mol | 0.00000006 mol/mol | B |
| 3. Molecular mass of all components in the mixture (based on IUPAC) : | 0.00000262 mol/mol | 0.00000021 mol/mol | B |
| a. H <sub>2</sub> O   | 0.97187357 mol/mol | 0.00001028 mol/mol | B |
| b. CO   | 0.00000150 mol/mol | 0.00000087 mol/mol | A |
| c. CO <sub>2</sub>  | 0.00000132 mol/mol | 0.00000011 mol/mol | A |
| d. O <sub>2</sub>   | 0.00000002 mol/mol | 0.00000001 mol/mol | A |
| e. Ar   | 0.00000585 mol/mol | 0.00000035 mol/mol | A |
| f. CH <sub>4</sub>  | 0.00000092 mol/mol | 0.00000006 mol/mol | A |
| g. N <sub>2</sub>   | 0.00000269 mol/mol | 0.00000023 mol/mol | A |
|   | 0.99998771 mol/mol | 0.00000097 mol/mol | A |
|   |                    |                    |   |
|   | 18.01528000 g/mol  | 0.00000087 g/mol   | B |
|   | 28.01040000 g/mol  | 0.00000011 g/mol   | B |
|   | 44.00950000 g/mol  | 0.00000001 g/mol   | B |
|   | 31.99880000 g/mol  | 0.00000035 g/mol   | B |
|   | 39.94800000 g/mol  | 0.00000006 g/mol   | B |
|   | 16.04246000 g/mol  | 0.00000023 g/mol   | B |
|   | 28.01348000 g/mol  | 0.00000097 g/mol   | B |
| Combined uncertainty from the gravimetric preparation (using Eq. 3)   |                    | 0.00000070 mol/mol | B |

|   |  |   |
|---|--|---|
| Verification of the CSGM concentration                            | 0.00000261 mol/mol   | A |
| Stability testing of the CSGM                                     | 0.00000173 mol/mol   | A |
| <b>Combined uncertainty</b> of the CSGM (using Eq. 2)             | 0.00000321 mol/mol   |   |
| <b>Expanded uncertainty</b> for confidence level of 95% ( $k=2$ ) | 0.00000642 mol/mol<br><b>(6.423 <math>\mu</math>mol/mol)</b> |   |

- **Measurand and expanded uncertainty** . Measurand and expanded uncertainty of prepared CSGM #L150721012 are listed in Table 5.

**Table 5.** Measurand ( $C_{\text{CSGM}\#\text{L150721012}}$ ) and expanded uncertainty ( $U_{\text{CSGMs}\#\text{L150721012}}$ )

| CSGM               | Assigned value<br>( $\mu$ mol/mol)     | Expanded uncertainty<br>( $\mu$ mol/mol) | Coverage factor* |
|--------------------|--|--|------------------|
| <b>#L150721012</b> | <b>999.208 <math>\mu</math>mol/mol</b> | <b>6.423 <math>\mu</math>mol/mol</b>     | <b>k = 2</b>     |

\* The coverage factor ( $k=2$ ) was based on approximately 95% confidence level.

### Uncertainty evaluation for the Sample #D581092

- **Model equation:** a model equation (Eq. 5) below was used to calculate  $\text{CO}_2$  concentration in the sample #D581092 (measurand).

$$C_{\text{sample}\#\text{D581092}} = \left( \frac{\text{Response area}_{\text{sample}\#\text{D581092}}}{\text{Response area}_{\text{standard}\#\text{L150721012}}} \right) \times C_{\text{standard}\#\text{L150721012}}$$

(5)

- **Uncertainty budget:** For the uncertainty of sample #D581092, the estimation was performed by modifying Eq. 5 based on the propagation rules<sup>(3)</sup>, resulting in an equation 6 (Eq. 6) below.

$$\left(\frac{u_{C_{sample \#D581092}}}{C_{sample \#D581092}}\right)^2 = \left(\frac{u_{(A_{sample \#D581092}/A_{standard \#L150721012})}}{A_{sample \#D581092} / A_{standard \#L150721012}}\right)^2 + \left(\frac{u_{C_{standard \#L150721012}}}{C_{standard \#L150721012}}\right)^2$$

(6)

Based on Eq. 6, there are two sources of uncertainty of the sample #D581092 concentration, i.e., 1). Repeatability of the ratio between detector's response of sample #D581092 and detector's response of standard #L150721012. ( $u_{(A_{sample \#D581092}/A_{standard \#L150721012})}$ ), and 2). Uncertainty of standard concentration #L150721012. ( $u_{C_{standard \#L150721012}}$ ). The details of uncertainty budgets for the sample #D581092 are listed in Table 6.

**Table 6.** Uncertainty budgets for the sample #D581092

| <b>Uncertainty source</b><br><b>X<sub>i</sub></b>                                       | <b>Estimated value</b><br><b>x<sub>i</sub></b> | <b>Assumed distribution</b> | <b>Standard uncertainty</b><br><b>u (x<sub>i</sub>)</b> | <b>Sensitivity coefficient</b><br><b>c<sub>i</sub></b> | <b>Contribution to standard uncertainty</b><br><b>u<sub>i</sub> (%)</b> |
|---|--|-----------------------------|---|--|---|
| Ratio of detector's response to sample and standard,<br>$u_{(A_{sample}/A_{standard})}$ | 1.001  | normal                      | 0.001   | 999.208  | 3.285   |

|  |                             |        |                                  |       |        |
|--|-----------------------------|--------|----------------------------------|-------|--------|
| Uncertainty of CSGM #L150721012                              | 999.208 $\mu\text{mol/mol}$ | normal | 3.212 $\mu\text{mol/mol}$        | 1.001 | 96.716 |
| <b>Combined Uncertainty</b> of sample #D581092               |                             |        | <b>3.269</b> $\mu\text{mol/mol}$ |       |        |
| <b>Expanded Uncertainty</b> , confidence level 95% ( $k=2$ ) |                             |        | <b>6.538</b> $\mu\text{mol/mol}$ |       |        |

- **Measurand and expanded uncertainty** : Measurand and expanded uncertainty of sample #D581092 are listed in Table 7.

**Table 7.** Measurand and expanded uncertainty of sample #D581092.

| Sample          | Concentration ( $\mu\text{mol/mol}$ ) | Expanded uncertainty ( $\mu\text{mol/mol}$ ) | Coverage factor*          |
|-----------------|---------------------------------------|--|---------------------------|
| <b>#D581092</b> | <b>1000.134</b>                       | <b>6.538</b>                                 | <b><math>k = 2</math></b> |

## References

- [1]. International Organization for Standardization, ISO 6142:2001 “Gas analysis - Preparation of calibration gas mixtures – Gravimetric method”, 2<sup>nd</sup> Edition.
- [2]. International Organization for Standardization, ISO 6143:2001 “Gas analysis - Comparison methods for determining and checking the composition of calibration gas mixtures”, 2<sup>nd</sup> Edition.
- [3]. Joint Committee for Guides in Metrology, JCGM 2008. Evaluation of measurement data - Guide to the expression of uncertainty in measurement (GUM).

# Report Form

## Carbon dioxide in nitrogen

---

Laboratory name: National Institute of Metrology (Thailand)

Cylinder number: D581146

### Measurement #1

| Component       | Date (dd/mm/yy) | Result ( $\mu\text{mol/mol}$ ) | Standard deviation (% relative) | number of replicates |
|-----------------|-----------------|--------------------------------|---------------------------------|----------------------|
| CO <sub>2</sub> | 24/01/2018      | 999.80                         | 0.01                            | 3                    |

### Measurement #2

| Component       | Date (dd/mm/yy) | Result ( $\mu\text{mol/mol}$ ) | Standard deviation (% relative) | number of replicates |
|-----------------|-----------------|--------------------------------|---------------------------------|----------------------|
| CO <sub>2</sub> | 24/01/2018      | 1000.28                        | 0.03                            | 3                    |

### Measurement #3<sup>5</sup>

| Component       | Date (dd/mm/yy) | Result ( $\mu\text{mol/mol}$ ) | Standard deviation (% relative) | number of replicates |
|-----------------|-----------------|--------------------------------|---------------------------------|----------------------|
| CO <sub>2</sub> | 24/01/2018      | 998.89                         | 0.02                            | 3                    |

### Results

| Component       | Result ( $\mu\text{mol/mol}$ ) | Expanded Uncertainty (% relative) | Coverage factor <sup>6</sup> |
|-----------------|--------------------------------|-----------------------------------|------------------------------|
| CO <sub>2</sub> | 999.66                         | 0.16                              | 2                            |

---

<sup>5</sup> If more than three measurements are taken, please copy and insert a table of the appropriate format as necessary

<sup>6</sup> The coverage factor shall be based on approximately 95% confidence.



### **Details of the measurement method used:**

The measurements were performed using a 6890 Gas Chromatograph with Thermal conductivity detector (TCD). The measurement procedure is shown as follow; "PGRM (Calibration) – Sample – PGRM (Calibration) – PGRM (Assurance) – PGRM (Calibration)– Sample – PGRM (Calibration) – PGRM (Assurance) – PGRM (Calibration) – Sample – PGRM (Calibration) – PGRM (Assurance) – PGRM (Calibration)". The single point calibration was used for determining the mole fraction of carbon dioxide in sample. In the addition, one of the primary gas reference material (PGRM) was used in the measure to assure consistency with the standard gas mixtures used and measurement system. The average response was calculated by using the last three of six times for each cylinder.

### **Details of the calibration method used:**

The GC-TCD was performed following the single point calibration by primary gas reference material (PGRM). The mole fraction of carbon dioxide in the PGRM used was closed to the target mole fraction of carbon dioxide in the sample cylinder. The GC column used was Hayesep Q, 8 ft and mesh 80/100, and with helium as carrier gas. The measured condition was sample loop 1 ml, temperature of oven 45°C, temperature of detector 250°C, reference gas flow 30 ml/min. The flow rate of gas mixtures was controlled by using a mass flow controller at 40 ml/min.

### **Details of the standards used:**

The PGRMs used in the measurements are binary mixtures of the carbon dioxide in nitrogen. They are traceable to the National Institute of Metrology (Thailand). The mole fraction of PGRMs used was determined in compliance with ISO 6142-1 by using gravimetric method and verified by using GC-TCD calibrated by using one of the PGRMs. The purity of Nitrogen is more than 99.9995% and the purity of carbon dioxide is more than 99.995%. These standard gas mixtures used were prepared by 2-step of dilution. Uncertainty values of PGRM s are evaluated from the gravimetry, verification, stability and measurement bias. The characteristics of the standard gas mixtures used are listed in Table as below.

Table 1. Concentration of PGRMs.

| Cylinder number | Assigned value              | Expanded uncertainty<br>(Relative value, $k = 2$ ) |
|-----------------|-----------------------------|--|
| PRM 112694      | 1000.68 $\mu\text{mol/mol}$ | 0.14%  |
| PRM 112684      | 1001.33 $\mu\text{mol/mol}$ | 0.18%  |

### Details on uncertainty budget:

The certified value applies to only this cylinder, and the uncertainty is expressed as an expanded uncertainty obtained by multiplying the standard uncertainty at 95% confidence interval by the coverage factor  $k=2$ . The standard uncertainty  $u(x_s)$  of the sample gas mixture is calculated from the following equations;

$$u(x_s) = x_s \sqrt{\frac{u^2(X_{crm})}{X_{crm}^2} + \frac{u^2(Y_{crm})}{Y_{crm}^2} + \frac{u^2(Y_s)}{Y_s^2}}$$

Where

- $u(X_{crm})$  is the standard uncertainty of the standard gas mixture
- $u(Y_{crm})$  is the standard uncertainties of measurement response of standard gas mixture
- $u(Y_s)$  is the standard uncertainties of measurement response of sample
- $X_{crm}$  is the standard gas mixture contents
- $Y_{crm}$  is average measurement response of standard gas mixture
- $Y_s$  is average measurement response of sample

## Uncertainty Budget for of CO<sub>2</sub> measurement

| Quantity<br>(Uncertainty source), $X_i$ | Estimate<br>$x_i$<br>( $\mu\text{mol/mol}$ ) | Evaluation type<br>(A or B)                  | Distribution | Standard uncertainty<br>(%relative) | Sensitivity coefficient<br>$c_i$ | Contribution<br>(%relative) |
|---|--|--|--------------|-------------------------------------|----------------------------------|-----------------------------|
| The standard gas mixture                | 1000.68                                      | B  | Normal       | 0.070                               | 1.0                              | 0.070                       |
| Response of standard gas mixture        | 330.44                                       | A  | Normal       | 0.014                               | 1.0                              | 0.014                       |
| Response of sample gas mixture          | 330.10                                       | A  | Normal       | 0.013                               | 1.0                              | 0.013                       |
| Analytical content of sample            | 999.66                                       | Combined Uncertainty, (%relative)            |              |                                     |                                  | 0.08                        |
|   |  | Expanded Uncertainty, ( $k=2$ ), (%relative) |              |                                     |                                  | 0.16                        |

### Authorship

Mr.Arnuttachai Wongjuk , Ms.Ratirat Sinweeruthai

**APMP Regional Comparison****Carbon dioxide in Nitrogen (1000  $\mu\text{mol/mol}$ )**

**Laboratory: Karaganda branch of RSE "Kazakhstan Institute of Metrology"  
Republic of Kazakhstan**

**Cylinder number: D581075**

**I. Measurement**

| <b>Measurement #1</b>      | <b>Date</b> | <b>Result, <math>\mu\text{mol/mol}</math></b> | <b>Standart deviation<br/>(% relative)</b> | <b>Number of<br/>replicates</b> |
|----------------------------|-------------|---|--|---------------------------------|
| Carbon dioxide in Nitrogen | 03.04.2018  | 1004,7  | 3,2  | 5                               |

| <b>Measurement #2</b>      | <b>Date</b> | <b>Result, <math>\mu\text{mol/mol}</math></b> | <b>Standart deviation<br/>(% relative)</b> | <b>Number of<br/>replicates</b> |
|----------------------------|-------------|---|--|---------------------------------|
| Carbon dioxide in Nitrogen | 04.04.2018  | 1001,3  | 2,9  | 5                               |

| <b>Measurement #3</b>      | <b>Date</b> | <b>Result, <math>\mu\text{mol/mol}</math></b> | <b>Standart deviation<br/>(% relative)</b> | <b>Number of<br/>replicates</b> |
|----------------------------|-------------|---|--|---------------------------------|
| Carbon dioxide in Nitrogen | 06.04.2018  | 999,3   | 3,0  | 5                               |

| <b>Measurement #4</b>      | <b>Date</b> | <b>Result, <math>\mu\text{mol/mol}</math></b> | <b>Standart deviation<br/>(% relative)</b> | <b>Number of<br/>replicates</b> |
|----------------------------|-------------|---|--|---------------------------------|
| Carbon dioxide in Nitrogen | 10.04.2018  | 1007,6  | 3,1  | 5                               |

| <b>Measurement #4</b>      | <b>Date</b> | <b>Result, <math>\mu\text{mol/mol}</math></b> | <b>Standart deviation<br/>(% relative)</b> | <b>Number of<br/>replicates</b> |
|----------------------------|-------------|---|--|---------------------------------|
| Carbon dioxide in Nitrogen | 10.04.2018  | 1002,0  | 2,9  | 5                               |

**Result**

| <b>Component</b>              | <b>Result, <math>\mu\text{mol/mol}</math></b> | <b>Coverage factor*)</b> | <b>Expanded<br/>Uncertainty,<br/><math>\mu\text{mol/mol}</math></b> |
|-------------------------------|---|--------------------------|---|
| Carbon dioxide in<br>Nitrogen | 1003,0  | 2                        | 14,7  |

\*) The coverage factor based on 95% confidence.

## II. Measurement Details for APMP-QM-S15.2017

### Instruments

Measurements were carried out using gas chromatograph "Crystal 5000" combined with flame-ionization detector and methanizer for conversion carbon dioxide.

Carrier gas: argon.

Volume size: 1 ml.

Chromatographic column: Hayesep N 80/100 mesh, 2m x 2mm.

Computers and software "Chromatech Analytic" were used to control chromatograph and collect and process chromatographic data.

### Calibration standards

1. The calibration gas standards were prepared by gravimetric method multiple dilutions, according to ISO 6142. An electronic mass-comparator (Mettler Toledo model XP10003S, capacity 10.1 kg, readability 1 mg) was used for preparation of all calibration gas standards. Manufacturer, type and metrological characteristics of the equipment used for the preparation of the gravimetric gas mixtures are given in Table 1.

Table 1.

| Type                    | Manufacturer                          | Metrological characteristics  |
|-------------------------|---------------------------------------|---|
| Model XP10003S          | «Mettler-Toledo», Switzerland         | The maximum limit weighing 10100 g<br>Resolution 1 mg<br>The standard deviation of 10 mg<br>The maximum change in temperature for 1 h. $\pm 0.5$ °C |
| Gas mixing plant, GSU-3 | OOO «PGS-Servise», Russian Federation | Pressure measuring range: from 0.001 to 16.0 MPa<br>Residual pressure cylinders before filling 10 Pa.   |

For the production of calibration gas mixtures were used aluminum cylinders with a capacity of 4 dm<sup>3</sup> complete with brass diaphragm valve type VBM-1. The internal surface of the cylinders was coated by paraffin grade P2.

2. Analysis of the purity of the clean gases.

Analysis of the purity of the original pure gases was based on information provided by the suppliers of pure gases (passports, certificates), as well as on the results of the measurement of impurities in pure gases using measurement techniques developed and approved by the RSE "KazInMetr".

In cases where the analytical method can not determine the content of the alleged impurities molar fraction of the expected impurity was assumed to be half the detection limit of the analytical method. The content of impurities unmeasured assumes a rectangular probability distribution, whereby the standard uncertainty is calculated as half the detection limit.

Determination of impurities in the starting pure gases (carbon dioxide, nitrogen) used to prepare calibration samples was conducted by gas chromatography using a flame ionization, thermal chemical and thermal conductivity detector.

The content of impurities in the pure gas used for preparing the calibration gas

mixtures shown in Table 2.

Table 2.

The metrological characteristics of pure gases.

| Clean gas  | Component       | Content mole fraction, % | The standard uncertainty, mole fraction, % |
|--|-----------------|--------------------------|--|
| Carbon dioxide<br>The cylinder capacity of 40 dm <sup>3</sup><br>Manufacturer: -<br>Kazakhstan         | CO <sub>2</sub> | 99,99                    | 0,01                                       |
| Nitrogen<br>The cylinder capacity of 40 dm <sup>3</sup><br>Manufacturer: The Republic of<br>Kazakhstan | N <sub>2</sub>  | 99,9999                  | 0,0001                                     |

In measuring the mass of gas filled and comparative cylinder identical volume being weighed according to the scheme and the method of substitution RMMR.

Based on previous studies RMS measurement result is taken to be 30 mg (standard uncertainty evaluated by type A).

3. After making the balloon with the calibration gas mixture was placed in a laboratory, where the at least 72 hours. Before the measurement tanks rolled the calibration gas mixtures for 10 minutes.

### Calibration of instrument

1. Calibration was performed using GC calibration gas mixtures are identical in composition to sample comparisons. The content of each component and its expanded uncertainty ( $k = 2$ ) is shown in Table 3.

Table 3 - Calibration gas mixtures

| Cylinder number, passport number, size, material, date of manufacture | Component       | Content, x (μmol/mol) | The standard uncertainty of the calibration samples (rel.), u (x), % |
|---|-----------------|-----------------------|--|
| PV-214, 4 dm <sup>3</sup>   | CO <sub>2</sub> | 1055,4                | 0,5  |
|   | N <sub>2</sub>  | -                     |  |
| PV-216, 4 dm <sup>3</sup>   | CO <sub>2</sub> | 962,1                 | 0,5  |
|   | N <sub>2</sub>  | -                     |  |
| PV-218, 4 dm <sup>3</sup>   | CO <sub>2</sub> | 1032,7                | 0,5  |
|   | N <sub>2</sub>  | -                     |  |
| PV-219, 4 dm <sup>3</sup>   | CO <sub>2</sub> | 1022,7                | 0,5  |
|   | N <sub>2</sub>  | -                     |  |

The total content of components standard uncertainty in calibration gas mixtures are calculated according to the formula:

$$u_{total} = \sqrt{u_m + u_p}$$

$u_m$  – standard uncertainty weighing, %;

$u_p$  – standard uncertainty of frequency source gases, %.

The standard uncertainty of the molar mass of gases, as well as uncertainty due to air buoyancy, with the pressure and volume of a cylinder is filled not taken into account in connection with a minor contribution.

2. The measurements were carried out under repeatability conditions. Before each measurement was conducted by the chromatograph calibration. Each measurement includes 5 observations.

3. Analytical function used to determine the components in the sample is as follows:

$$x(y) = b_1y + b_0$$

were,

$x$  – measured content,  $\mu\text{mol/mol}$ ;

$y$  – chromatographic response of the analyte;

$b_1$  – coefficient of linear dependence;

$b_0$  – offset coefficient.

### **Sample preparation**

The sample with the sample and the calibration sample was stored prior to measurement for 24 hours in the laboratory. The change in temperature in the laboratory at the time of measurement is  $\pm 2$  ° C, the change in pressure within  $\pm 0,5$  kPa.

### **Calculation of measurement uncertainty**

Uncertainty value  $u(x)$  was calculated in accordance with ISO 6143 taking into account the uncertainties of the calibration standards and instrument response variability during calibration and measurements under reproducibility conditions:

$$u(x) = \sqrt{u^2(x, x_{cs}) + u^2(x, y)}$$

were

$u(x, x_{cs})$  – the standard uncertainty associated with the amount-of-substance fractions of the calibration standards;

$$u(x, x_{cs}) = \sqrt{\sum_{i=1}^n \left(\frac{u(x_{csi})}{n}\right)^2},$$

where

$u(x_{csi})$  – uncertainty of the calibration standards;

$n$  - number of the calibration standards;

$u(x, y)$  – uncertainty associated with the instrument response,

$$u(x, y) = \sqrt{\frac{u^2(y) + x^2 \cdot u^2(a_1) + 2x \cdot \text{cov}(a_0, a_1) + u^2(a_0)}{a_1^2}},$$

where

$u(y)$  – uncertainty of instrument response during measurements;



$a_0, a_1, u(a_0), u(a_1), cov(a_0, a_1)$  – calibration function parameters obtained from B\_Least for linear function.

The calibration function parameters of all measurements are given in Table 4.

Table 4 - Calibration function parameters.

| Parameter                                   | 1 <sup>st</sup><br>measurement | 2 <sup>nd</sup><br>measurement | 3 <sup>rd</sup><br>measurement | 4 <sup>rd</sup><br>measurement | 5 <sup>rd</sup><br>measurement |
|---|--------------------------------|--------------------------------|--------------------------------|--------------------------------|--------------------------------|
| Regression coefficient $a_1$                | 4,097824E+01                   | 4,142879E+01                   | 3,839535E+01                   | 4,017443E+01                   | 4,931432E+01                   |
| Regression coefficient uncertainty $u(a_1)$ | 3,071293E+00                   | 3,087130E+00                   | 2,802749E+00                   | 3,162043E+00                   | 3,765568E+00                   |
| Regression coefficient $a_0$                | -2,816330E+03                  | -2,813569E+03                  | 5,315462E+02                   | -8,128793E+02                  | 1,576403E+04                   |
| Regression coefficient uncertainty $u(a_0)$ | 3,128914E+03                   | 3,148256E+03                   | 2,854800E+03                   | 3,213045E+03                   | 3,836918E+03                   |
| Covariance $cov(a_0, a_1)$                  | -9,602676E+03                  | -9,713637E+03                  | 7,996633E+03                   | -1,015268E+04                  | -1,444015E+04                  |

Table 5

Uncertainty table

| Uncertainty source $X_i$   | Estimate $x_i$ | Assumed distribution | Standart uncertainty $u(x_i)$ | Sensitivity coefficient $c_i$ | Contribution to standard uncertainty $u_i(y)$ |
|--|----------------|----------------------|-------------------------------|-------------------------------|---|
| Uncertainty associated with the amount-of substance fractions of the calibration standards | -              | Normal               | 2,9 $\mu\text{mol/mol}$       | 1                             | 2,9 $\mu\text{mol/mol}$                       |
| Uncertainty associated with the instrument response  | -              | Normal               | 7,3 $\mu\text{mol/mol}$       | 1                             | 7,3 $\mu\text{mol/mol}$                       |

Coverage factor:  $k = 2$ .

Expanded uncertainty: 14,7  $\mu\text{mol/mol}$ .

## APMP.QM-S15 Carbon dioxide in nitrogen

---

Laboratory name: KRISS

Cylinder number: D581070

### Measurement #1

| Component       | Date (dd/mm/yy) | Result ( $\mu\text{mol/mol}$ ) | Standard deviation ( $\mu\text{mol/mol}$ ) | number of replicates |
|-----------------|-----------------|--------------------------------|--|----------------------|
| CO <sub>2</sub> | 11.8.2017       | 1000.13                        | 0.55                                       | 4                    |

### Measurement #2

| Component       | Date (dd/mm/yy) | Result ( $\mu\text{mol/mol}$ ) | Standard deviation ( $\mu\text{mol/mol}$ ) | number of replicates |
|-----------------|-----------------|--------------------------------|--|----------------------|
| CO <sub>2</sub> | 18.9.2018       | 999.78                         | 0.50                                       | 4                    |

### Measurement #3

| Component       | Date (dd/mm/yy) | Result ( $\mu\text{mol/mol}$ ) | Standard deviation ( $\mu\text{mol/mol}$ ) | number of replicates |
|-----------------|-----------------|--------------------------------|--|----------------------|
| CO <sub>2</sub> | 1.10.2018       | 999.70                         | 0.50                                       | 5                    |

### Measurement #4

| Component       | Date (dd/mm/yy) | Result ( $\mu\text{mol/mol}$ ) | Standard deviation ( $\mu\text{mol/mol}$ ) | number of replicates |
|-----------------|-----------------|--------------------------------|--|----------------------|
| CO <sub>2</sub> | 2.10.2018       | 999.84                         | 0.50                                       | 5                    |

### Results

| Component       | Result ( $\mu\text{mol/mol}$ ) | Expanded Uncertainty ( $\mu\text{mol/mol}$ ) | Coverage factor <sup>7</sup> |
|-----------------|--------------------------------|--|------------------------------|
| CO <sub>2</sub> | 999.86                         | 1.08   | 2                            |

---

<sup>7</sup> The coverage factor shall be based on approximately 95% confidence.


## **Details of the measurement method used:**

### **Analysis method:**

Carbon dioxide concentration in nitrogen has been quantified using gas chromatograph thermal conductivity detector (GC-TCD). Figure 1 shows an analytical condition of the analyzer and its chromatogram.

Configuration of analysis system: gas cylinder >> regulator >> MFC >> sample injection valve >> column >> detector >> integrator >> area comparison >> results

To achieve analytical interval of  $\pm 0.1\%$  (standard deviation) the instrument drift and standard deviation of the response were controlled carefully. Two cylinders D581078 and D581103 were analyzed as the reference mixture against the distributed sample cylinder (D581070).



**Figure 1. Analytical condition and chromatogram of CO<sub>2</sub>**

### **Details of the calibration method used:**

Instrument calibration is performed using KRIS primary standard mixtures. One point calibration was done with a cylinder of nominal value  $\sim 1000 \mu\text{mol/mol}$  which was very close to the target cylinder.

### **Sample handling:**

The sample cylinder had put in the laboratory with room temperature for several days after preparation. Each cylinder was equipped with a stainless steel pressure regulator that was purged more than 5 times after connection to the analysis line. Samples were transferred to sample loop at flow rate of 75 mL/min using the mass-flow controller.

### Calibration standards:

#### Preparation method

Total 8 cylinders were prepared for this comparison (figure 1) and 2 primary standard mixtures among them were used for determining amount of carbon dioxide in Nitrogen. The standards were prepared from pure carbon dioxide and pure nitrogen in accordance with ISO6142:2001 (Gas analysis-preparation of calibration gases-gravimetric method). Pure carbon dioxide was diluted by 2 step and purity analysis for every pure gases were done (pure CO<sub>2</sub> used at CCQM-K120). Table 1 shows gravimetric value and expanded uncertainty of the calibration standards.




Figure 2. Preparation cylinder tree of CO<sub>2</sub> for this comparison

Table 1. Gravimetric value and expanded uncertainty in calibration standards

| Cylinder number | Gravimetric value<br>(μmol/mol) | Expanded uncertainty<br>[k=2] (μmol/mol) |
|-----------------|---------------------------------|--|
| D581078         | 999.49                          | 0.15                                     |
| D581103         | 1001.00                         | 0.14                                     |

#### Purity analysis

The impurities of carbon dioxide and nitrogen were determined by analytical methods and the amount of the major component is conventionally determined from the following equation,

$$x_{pure} = 1 - \sum_{i=1}^N x_i$$

Where

$x_i$  : the mole fraction of impurity  $i$ , determined by analysis;

N: the number of impurities likely to be present in the final mixture;

$x_{pure}$ : the mole fraction “purity” of the “pure” parent gas.

Table 2 and 3 show the results of purity analysis of CO<sub>2</sub> and N<sub>2</sub>.

**Table 2. Results of purity analysis of Carbon dioxide (NB16027)**

| component        | Analytical conc.<br>( $\mu\text{mol/mol}$ ) | Detector        | distribution | Applied conc.<br>( $\mu\text{mol/mol}$ ) | Standard<br>uncertainty<br>( $\mu\text{mol/mol}$ ) |
|------------------|---|-----------------|--------------|--|--|
| H <sub>2</sub>   | 6.4   | GC/TCD          | Normal       | 6.4                                      | 0.6  |
| O <sub>2</sub>   | 79.5  | GC/TCD          | Normal       | 79.5                                     | 7.9  |
| Ar               | 3.5   | GC/TCD          | normal       | 3.5                                      | 0.4  |
| N <sub>2</sub>   | 214.8                                       | GC/TCD          | Normal       | 214.8                                    | 21.5   |
| CO               | 39.6  | GC/TCD          | Normal       | 39.6                                     | 4.0  |
| CH <sub>4</sub>  | 15.2  | GC/TCD          | Normal       | 15.2                                     | 1.5  |
| H <sub>2</sub> O | 0.54  | dew point meter | Normal       | 0.54                                     | 0.05   |
| THC              | <0.2  | GC/AED          | Rectangular  | 0.1                                      | 0.06   |
| Total Sulfur     | <0.01                                       | GC/AED          | Rectangular  | 0.01                                     | 0.003  |
|                  |   |                 |              | impurities                               | 23.30  |
|                  |   |                 |              | <b>CO<sub>2</sub> purity</b>             | <b>46.60 (<math>k=2</math>)</b>                    |

**Table 3. Results of purity analysis of Nitrogen (NK02608)**

| component        | Analytical conc.<br>( $\mu\text{mol/mol}$ ) | Detector                           | distribution | Applied conc.<br>( $\mu\text{mol/mol}$ ) | Standard<br>uncertainty<br>( $\mu\text{mol/mol}$ ) |
|------------------|---|------------------------------------|--------------|--|--|
| H <sub>2</sub>   | < 0.1                                       | GC/PDD                             | Rectangular  | 0.050                                    | 0.029  |
| O <sub>2</sub>   | 0.11  | Galvanic Sensor<br>oxygen analyzer | Normal       | 0.110                                    | 0.011  |
| Ar               | 4.48  | GC/TCD                             | Normal       | 4.480                                    | 0.448  |
| CO               | < 0.003                                     | GC/FID                             | Rectangular  | 0.002                                    | 0.001  |
| CO <sub>2</sub>  | 0.011                                       | GC/FID                             | Normal       | 0.011                                    | 0.001  |
| CH <sub>4</sub>  | < 0.002                                     | GC/FID                             | Rectangular  | 0.001                                    | 0.001  |
| H <sub>2</sub> O | 0.55  | dew point meter                    | Normal       | 0.550                                    | 0.055  |
| N <sub>2</sub> O | 0.00014                                     | GC/ $\mu$ ECD                      | Normal       | 0.00014                                  | 0.00001  |
| THC              | < 0.5                                       | GC/FID                             | Rectangular  | 0.250                                    | 0.144  |
|                  |   |                                    |              | impurities                               | 0.47   |
|                  |   |                                    |              | <b>N<sub>2</sub> purity</b>              | <b>0.95 (<math>k = 2</math>)</b>                   |

### Uncertainty:

The uncertainty used for the calibration mixtures contains all source of gravimetric preparation. Uncertainty for stability is not included because no instability has been detected. An analysis uncertainty is calculated based on repeatability and drift of analyzer of the acquired area.

### Detailed uncertainty budget:

Please include a list of the uncertainty contribution, the estimate of the standard uncertainty, probability distribution, sensitivity coefficients, etc.

$$C_{\text{final}} = \frac{A_{\text{sample}}}{A_{\text{ref}}} \times C_{\text{crm\_ref}}$$

Typical evaluation of the of each measurement uncertainty for CO<sub>2</sub>:

| Quantity $X_i$  | Estimate $x_i$<br>Area[arb.]<br>[ $\mu\text{mol/mol}$ ] | Evaluation<br>Type<br>(A or B) | Distribution | Standard<br>uncertainty<br>$u(x_i)$<br>Area[arb.]<br>[ $\mu\text{mol/mol}$ ] | Sensitivity<br>coefficient | Contribution<br>$u_i(y)$ |
|---|---|--------------------------------|--------------|--|----------------------------|--------------------------|
| Response_reference<br>D581078(before)                 | 3319.512  | A                              | Gaussian     | 0.322  | -0.15                      | -0.049                   |
| D581078(after)  | 3316.000  |                                |              | 0.370  | -0.15                      | -0.056                   |
| Area[arb.]  |   |                                |              |  |                            |                          |
| Response_Sample<br>D581070                            | 3318.690  | A                              | Gaussian     | 0.205  | 0.30                       | 0.062                    |
| Area[arb.]  |   |                                |              |  |                            |                          |
| Reference<br>prepared<br>grav.                        | D581078<br>1001.00                                      | B                              | Gaussian     | 0.5  | 1.0                        | 0.50                     |
|   | D081103<br>1001.00                                      |                                |              | -  | -                          |                          |
| [ $\mu\text{mol/mol}$ ]                               |   |                                |              |  |                            |                          |
| Amount_sample   | 999.78  |                                |              |  |                            |                          |
| [ $\mu\text{mol/mol}$ ]                               |   |                                |              |  |                            |                          |
| Combined standard uncertainty [ $\mu\text{mol/mol}$ ] |   |                                |              | 0.5  |                            |                          |