



CCM.M-K8.2019

First CCM key comparison of kg realizations

M. Stock, BIPM

CCM

20 May 2021



Bureau
International des
Poids et
Mesures

Key comparison of realizations of the kilogram

Objectives	<ul style="list-style-type: none">- test the consistency of realizations based on different realization experiments (Kibble balances, joule balance, XRCD method)- contribute to the first consensus value for coordinated dissemination												
Pilot laboratory	BIPM												
Conditions for participation	<ul style="list-style-type: none">- $u(m) < 200 \mu\text{g}$ at 1 kg (2×10^{-7})- peer reviewed publication incl. detailed uncertainty budget												
Participants	<table><tr><td>Kibble balances:</td><td>BIPM, KRIS, NIST, NRC</td></tr><tr><td>Joule balance:</td><td>NIM</td></tr><tr><td>XRCD method:</td><td>NMIJ, PTB</td></tr></table>	Kibble balances:	BIPM, KRIS, NIST, NRC	Joule balance:	NIM	XRCD method:	NMIJ, PTB						
Kibble balances:	BIPM, KRIS, NIST, NRC												
Joule balance:	NIM												
XRCD method:	NMIJ, PTB												
Timeline	<table><tr><td>Technical protocol</td><td>21 October 2019</td></tr><tr><td>BIPM measurements</td><td>Nov 2019 – Feb 2020</td></tr><tr><td>last NMI results</td><td>29 May 2020</td></tr><tr><td>Draft A-1</td><td>20 July 2020</td></tr><tr><td>Draft B</td><td>12 October 2020</td></tr><tr><td>Final Report</td><td>16 October 2020 (delay of 4 months due to covid-19)</td></tr></table>	Technical protocol	21 October 2019	BIPM measurements	Nov 2019 – Feb 2020	last NMI results	29 May 2020	Draft A-1	20 July 2020	Draft B	12 October 2020	Final Report	16 October 2020 (delay of 4 months due to covid-19)
Technical protocol	21 October 2019												
BIPM measurements	Nov 2019 – Feb 2020												
last NMI results	29 May 2020												
Draft A-1	20 July 2020												
Draft B	12 October 2020												
Final Report	16 October 2020 (delay of 4 months due to covid-19)												

CCM.M-K8.2019: organization of comparison

- Each participants selected 1 or 2 travelling 1 kg standards:


- 1 Pt-Ir standard
- 1 optional standard (Pt-Ir, stainless steel, Si-sphere, none)

for calibration under vacuum using realization experiment (Kibble balance, joule balance, Avo sphere),

based on : $h = 6.626\ 070\ 15 \times 10^{-34}$ Js (revised SI)

- NMIs sent travelling standards to BIPM
- At BIPM, all travelling standards were compared under vacuum with a BIPM reference standard
- The mass of the reference standard under vacuum is known in terms of BIPM as-maintained mass unit (traceable to h via the IPK)


CCM.M-K8.2019: principle of the comparison




CCM.M-K8.2019: travelling standards

Institute	Identification of standard	Manufacturer	Type	Estimated air-vacuum surface sorption / μg	Magnetic susceptibility	Magnetic polarization / μT
BIPM	691s	BIPM	Pt-Ir standard with small Pt wire	4	24×10^{-5}	<0.02
KRISS	111	BIPM	Pt-Ir prototype		24×10^{-5}	< 0.02
	SN17	Mettler Toledo	stainless steel cylinder	6.3	<0.002	<0.2
NIM	110	BIPM	Pt-Ir prototype		24×10^{-5}	0
	6601	Changzhou Accurate Weight Co., China	stainless steel cylindrical	5.5	5.16×10^{-4}	0.01
NIST	85	BIPM	Pt-Ir prototype	7.2	24×10^{-5}	<0.02
	104	BIPM	Pt-Ir prototype	3.5	24×10^{-5}	<0.02
NMIJ	94	BIPM	Pt-Ir prototype	5.7(3.3)	3×10^{-4}	< 0.02
	E59	Stanton Instruments	Pt-Ir standard	11.6(3.3)	2×10^{-4}	< 0.02
NRC	106	BIPM	Pt-Ir prototype	3.4	24×10^{-5}	<0.02
	F18	BIPM	Stack of 8 Pt-Ir disks	12.2	24×10^{-5}	<0.02
PTB	109	BIPM	Pt-Ir prototype	2	< 0.001	< 0.1
	Si14-02	PTB	Si sphere	20	-2.6×10^{-7}	0

CCM.M-K8.2019: Results




CCM.M-K8.2019: Results




standard uncertainty of weighted mean: 7.5 μg

CCM.R-kg-P1 (2016) compared to CCM.M-K.2019



kg realizations have changed with respect to BIPM mass unit from -14 μg to -60 μg .

Pilot Study CCM.R-kg-P1 (2016) compared to CCM.M-K.2019



blue points: recalculated from Pilot Study 2016 and change of NMI's value for h or N_A from 2016 (Pilot Study) to 2017 (CODATA fundamental constants adjustment)

Summary and Conclusions

- All except two results agree within their standard uncertainties.
- The two results with the smallest uncertainties do not agree.
- The weighted mean of the seven results has an uncertainty of 7.5 μg .
- The weighted mean of the independent realizations agrees within the expanded uncertainty with the BIPM as-maintained mass unit, traceable to the IPK.
- Some of the independent kg realizations have changed significantly since 2016.
- Not all travelling standards could be returned due to the travel restrictions. Those who were returned showed mass changes between +2 μg and -8 μg (for a standard with bad surface): mass stability in general not an issue.

Recommendations for future key comparisons

- Follow same approach as for CCM.M-K8.2019
- Slightly less tight schedule
- Conditions for participation: $u < 200 \mu\text{g}$, peer-reviewed publication
- Travel restriction might make it difficult to use Pt-Ir prototypes, which are in general hand-carried. Use only stainless steel and Si-spheres?
- Keep the travelling standards at the BIPM until the NMI measurement report is available so that possible inconsistencies can be investigated
- The time for sending the NMI measurement report was in some cases very long (up to 6 months), it would be desirable that the reports are sent within 2 months.

Proposed timescale for CCM.M-K8.2021

June 2021	Call for participation
June/July 2021	Agreement on technical protocol (similar to K8.2019)
September – December 2021	Measurement of travelling standards at NMIs
January – March 2022	Comparison measurements at BIPM
April-May 2022	Stability checks at NMIs after return of standards
June 2022	Draft A
September 2022	Final report

Thank you for your attention !



Bureau
International des
Poids et
Mesures