

# **Comparison of primary absorbed dose to water standards in the medium-energy x-ray range**

## **Technical Protocol**

Pilot laboratory:

Physikalisch-Technische Bundesanstalt (PTB)

Contact person: Ludwig Büermann

## **Contents**

|  |           |
|--|-----------|
| <b>1. Object and participants</b>  | <b>2</b>  |
| <b>2 Transfer chambers</b>   | <b>2</b>  |
| <b>3 Radiation qualities</b>   | <b>3</b>  |
| <b>4 Measurement procedure and reference conditions</b>                                | <b>3</b>  |
| <b>4.1 Measurement procedure</b>   | <b>3</b>  |
| <b>4.2 Reference conditions for the calibration in terms of absorbed dose to water</b> | <b>4</b>  |
| <b>4.3 Reference conditions for the calibration in terms of air kerma</b>              | <b>4</b>  |
| <b>4.4 Additional measurements in the depth of 5 g/cm<sup>2</sup></b>                  | <b>4</b>  |
| <b>5 Course of comparison</b>  | <b>5</b>  |
| <b>5.1 Prospective time schedule</b>   | <b>5</b>  |
| <b>5.2 Report of results</b>   | <b>5</b>  |
| <b>5.3 Procedure for handling the results of the pilot laboratory</b>                  | <b>6</b>  |
| <b>5.4 Evaluation of the results</b>   | <b>6</b>  |
| <b>5.4.1 Calibration coefficients in terms of air kerma</b>                            | <b>6</b>  |
| <b>5.4.2 Calibration coefficients in terms of absorbed dose to water</b>               | <b>6</b>  |
| <b>5.5 Publication of the results</b>  | <b>8</b>  |
| <b>6 References</b>  | <b>9</b>  |
| <b>APPENDIX B: Complete addresses of the participants (used for shipment!)</b>         | <b>10</b> |

## 1. Object and participants

The objective of this project is to carry out a comparison of primary measurement standards of absorbed dose to water for the medium-energy x-ray range. The prospective participants are listed in Table 1. Three of the participants (VSL B.V., PTB, LNE-LNHB) have operating primary standards based on water calorimeters as published in references [1 – 3]. The remaining two participants are developing new primary standards based on a water-graphite calorimeter (ENEA) and a graphite extrapolation chamber (MKEH). According to the schedule of deliverables the newly developed primary standards should be available by the end of May 2014.

**Table 1:** Participants and contact persons

| Participant | Institute | Country     | Contact person  | E-mail of contact person   |
|-------------|-----------|-------------|-----------------|--|
| 1           | PTB       | Germany     | Ludwig Büermann | <a href="mailto:ludwig.bueermann@ptb.de">ludwig.bueermann@ptb.de</a> |
| 2           | LNE-LNHB  | France      | Benjamin Rapp   | <a href="mailto:benjamin.rapp@cea">benjamin.rapp@cea</a>             |
| 3           | VSL B.V.  | Netherlands | Leon de Prez    | <a href="mailto:ldprez@vsl.nl">ldprez@vsl.nl</a>                     |
| 4           | MKEH      | Hungary     | Gábor Machula   | <a href="mailto:machulag@mkeh.hu">machulag@mkeh.hu</a>               |
| 5           | ENEA      | Italy       | Massimo Pinto   | <a href="mailto:Massimo.pinto@enea.it">Massimo.pinto@enea.it</a>     |

Note: Complete addresses are given in Appendix B.

Three waterproof farmer type chambers (see clause 2) will be used as transfer chambers. Each participant calibrates the transfer chambers under reference conditions as defined below (clause 4) in terms of air kerma free in air and absorbed dose to water in the depth of 2 g/cm<sup>2</sup>. Additional charge measurements shall be made with the transfer chambers in a depth of 5 g/cm<sup>2</sup> (clause 4.3).

## 2 Transfer chambers

Three ionization chambers of the same type PTW TM30013 with a nominal volume of 0,6 cm<sup>3</sup> will be provided by PTB and used as transfer chambers. The chambers are manufactured by PTW. The chambers have a PTW-M type connector. An appropriate adapter to other connections will be provided by PTB. Participants are pleased to inform the pilot laboratory about their needs.

**Table 2:** Main technical data of the transfer chambers

|                           |  |
|---------------------------|--|
| Type:                     | TM30013, serial numbers 7463, 7484, 7485 |
| Nominal sensitive volume: | 0,6 cm <sup>3</sup>                      |
| Chamber high voltage:     | 300 V                                    |
| Leakage current:          | ≤ +4fA                                   |
| Nominal response:         | 20 nC/Gy                                 |
| Minimum air kerma rate:   | 0,2 mGy/s (yields about 4 pA)            |

### 3 Radiation qualities

The CCRI radiation qualities in the range from 100 kV to 250 kV (CCEMRI 1972, Ref. [8]) shall be used for the comparison. The PTB realisation of these qualities are shown in Table 3.

**Table 3:** Characteristics of the PTB reference radiation qualities

| Radiation quality                                     | 100 kV | 135 kV | 180 kV | 250 kV |
|---|--------|--------|--------|--------|
| Generating potential / kV                             | 100    | 135    | 180    | 250    |
| Additional Al filtration / mm                         | 3.506  | 2.302  | 2.302  | 2.302  |
| Additional Cu filtration / mm                         | -      | 0.222  | 0.512  | 1.590  |
| Al HVL / mm   | 4.142  | -      | -      | -      |
| Cu HVL / mm   | 0.157  | 0.489  | 1.013  | 2.482  |
| $(\mu/\rho)_{\text{air}} / \text{cm}^2 \text{g}^{-1}$ | 0.287  | 0.196  | 0.168  | 0.143  |
| $\dot{K}_{\text{PTB}} / \text{mGy s}^{-1}$            | 0.50   | 0.50   | 0.50   | 0.50   |

### 4 Measurement procedure and reference conditions

#### 4.1 Measurement procedure

Measurements free-in-air in the water phantom should not start before the chamber has equalized to the ambient temperature. After connection of the high voltage it is advisable to wait at least 1 hour until the measurements begin. The currents of the transfer chambers at the place of measurement should always be measured with and without the radiation beam. The signal to background ratio of the currents should not be less than 1000. The background current shall be subtracted from the signal current. A complete measurement should consist of at least 10 repeated single measurements and the mean value should be taken as the result. The relative percentage Type A standard uncertainty of the repeated measurements shall not exceed 0,1%.

The calibration coefficients of the transfer chambers shall be measured at both polarities and given in terms of air kerma or absorbed dose to water per unit charge in units of Gy/C referring to standard conditions of air temperature, pressure and relative humidity of  $T = 293,15 \text{ K}$ ,  $P = 101,325 \text{ kPa}$  and  $h = 50 \%$ . All measurements shall be performed in the same x-ray beam. The relative air humidity shall be between 20 % and 80 % during the calibrations otherwise a correction to  $h = 50 \%$  should be applied. Participants do not need to apply any correction for the incomplete charge collection.

The transfer chambers shall be calibrated at the reference conditions listed in 4.2 and 4.3. Unfortunately, it was not possible to agree on exactly the same reference conditions at all participants' sites. However, it was confirmed that these differences should not cause significant differences in the calibration coefficients of the transfer chambers.

#### 4.2 Reference conditions for the calibration in terms of absorbed dose to water

|   |   |
|---|---|
| Dimensions of the water phantom                                       | 30 cm x 30 cm x 30 cm<br>20 cm x 20 cm x 20 cm (ENEA) |
| Material  | Water   |
| Radiation qualities   | According to Table 3                                  |
| Air temperature:  | 20 °C   |
| Air pressure  | 1013,25 hPa   |
| Rel. humidity   | 50 %  |
| Absorbed dose rate to water   | > 0,2 mGy/s   |
| Size of the radiation field at the point of measurement (50% Isodose) | 10 cm x 10 cm<br>or circular with diameter 10 cm      |
| Measurement depth   | 2 g/cm <sup>2</sup>                                   |
| Distance between source and point of measurement                      | 100 cm<br>50 cm (LNE-LNHB)<br>61,5 cm (VSL b.V.)      |
| Polarizing voltage  | 300 V (measure at both polarities)                    |

#### 4.3 Reference conditions for the calibration in terms of air kerma

|  |  |
|--|--|
| Radiation qualities:   | According to Table 3   |
| Air temperature:   | 20 °C  |
| Air pressure:  | 1013,25 hPa  |
| Rel. humidity:   | 50 %   |
| Air kerma rate:  | > 0,2 mGy/s  |
| Size of the radiation field at the point of measurement (50% Isodose): | About 10 cm x 10 cm<br>or circular with diameter of about 10 cm<br>Circular, diameter 7 cm (at LNE-LNHB) |
| Distance between source and point of measurement:                      | 100 cm<br>120 cm (LNE-LNHB)  |
| Polarizing voltage   | 300 V (measure at both polarities)   |

#### 4.4 Additional measurements in the depth of 5 g/cm<sup>2</sup>

The reference depth in the water phantom was agreed to be 2 g/cm<sup>2</sup> (clause 4.2). It was agreed that all participants shall perform additional charge measurements under otherwise equal conditions with the transfer chambers positioned in the depth of 5 g/cm<sup>2</sup> at all comparison radiation qualities (clause 3). Note that the distance between the source and point of measurement shall remain constant (i.e. move the phantom, not the chamber). If possible, these measurements shall also be done at both polarities. The charge ratio obtained for the two depths will be used to evaluate a suitable beam qualifier.

## 5 Course of comparison

There will be a star-shaped circulation of the chambers between PTB and the participants. PTB pays for the transport of the chambers to the participants. The participants pay for the transport of the chambers back to PTB. After every participants calibration PTB will perform chamber constancy checks. The chambers should stay at the participant's site for no longer than 2 weeks. The report of results (clause 5.2) should be sent to the coordinator within 2 weeks after the calibration.

### 5.1 Prospective time schedule

The comparison is scheduled to commence in June 2014 (starting with PTB's measurements) and expected to be completed in November 2014. The proposed schedule is shown in Table 2.

**Table 2:** Proposed schedule of the comparison (June 2014 until November 2014)

| Participant               | Date of Chambers leaving PTB for participant | Measurement duration at laboratory | Date of chamber leaving participant for PTB |
|---------------------------|--|------------------------------------|---|
| PTB, Germany              |  | 16-27 June 2014                    |   |
| CEA, LIST, France         | 07 July 2014                                 | 14-25 July 2014                    | 28 July 2014                                |
| ENEA, Italy               | 25 Aug. 2014                                 | 01-12 Sep. 2014                    | 15 Sept. 2014                               |
| MKEH, Hungary             | 29 Sept. 2014                                | 06-17 Oct. 2014                    | 20 Oct. 2014                                |
| VSL B.V., The Netherlands | 03 Nov 2014                                  | 10-21 Nov. 2014                    | 24 Nov. 2014                                |

Notes:

1. Duration of measurement for one laboratory is two weeks.
2. Transportation time for the chambers from the PTB to a participant is about one week, vice versa.
3. Duration of constancy measurements at PTB laboratory is about one week.

### 5.2 Report of results

The **report of the results** shall at least contain the following information:

- Description of the radiation field (type of x-ray source, field size (50% Isodose), approximate air kerma rate at 1 m distance and 10 mA tube current, photon fluence spectra if measured, HVL and mean energies of the radiation qualities used for the comparison)
- Description of the set-up (electrometer type, connector types used, traceability of the electrometer calibration)
- Climatic conditions prevailing in the calibration laboratory during the calibration (Temperature, Pressure, Humidity)
- Complete uncertainty budget of the air kerma rate measured with the primary standard
- Complete uncertainty budget of the absorbed dose to water rate measured with the primary standard
- Leakage and signal currents of the transfer chambers during the measurements
- Calibration coefficients of the three transfer chambers for both polarities in terms of air kerma free in air and absorbed dose to water
- Results of the additional charge measurements in the depth of 5 g/cm<sup>2</sup>

- Complete uncertainty budget of the calibration coefficients

The uncertainties shall be given in accordance with the ISO Guide to the expression of uncertainties in measurements (GUM) [4].

### 5.3 Procedure for handling the results of the pilot laboratory

The pilot laboratory will participate in the comparison. It shall finish its measurements and report of results prior to all other participants. The report on these measurements will be sent to the EURAMET TC-IR Chair Lena Johansson before the first participant following the pilot's measurements has finished its measurements (i.e. before July 25 according to Table 2). This procedure should be a measure of confidence. For the purpose of constancy checks, the pilot laboratory will repeat its determination of the calibration coefficients at selected radiation qualities after every participant's measurements.

### 5.4 Evaluation of the results

The pilot laboratory will evaluate the comparison on the basis of the reports of results (see 5.2) given by the participants. In general, the mean value obtained from measurements at both polarities will be regarded as the result for each transfer chamber at one radiation quality. The final result will be evaluated as the mean value of the single results obtained for each of the three transfer chambers.

#### 5.4.1 Calibration coefficients in terms of air kerma

This comparison is not aimed at a comparison of air kerma standards but the calibration coefficients in terms of air kerma,  $N_K$ , give valuable additional information about the consistency of the absorbed dose calibration coefficients,  $N_{D_w}$ , for the different radiation qualities used at the participant's sites (see next clause). If one wants to use the  $N_K$  results in such a way, it is necessary to know the differences of  $N_K$  between the participants. Therefore it is suggested, to evaluate the ratios and uncertainties of those of the  $N_K$  values obtained by the participants for identical radiation qualities. These results will mainly be used for the evaluation procedure according to equation (1) described in following clause. In addition, the consistencies of the results for the CCRI-qualities obtained by the participants in the current exercise with those obtained in the bilateral key comparisons with the BIPM (BIPM.RI(I)-K3) will be checked. But it is not intended to publish these results as an additional indirect K3-key comparison result.


#### 5.4.2 Calibration coefficients in terms of absorbed dose to water

The consistency of the calibration coefficients in terms of absorbed dose to water and air kerma free in air obtained by the participants will be evaluated by the equation:

$$N_{D_w} = N_K * \left( \frac{\bar{\mu}_{en}}{\rho} \right)_{w,a}^d * k_{ch} * k_{sheath}, \quad (1)$$

where  $(\bar{\mu}_{\text{en}}/\rho)_{\text{w,a}}^d$  is the mean value of the water-to-air ratio of the mass-energy-absorption coefficients at the depth  $d$  in water and  $k_{\text{ch}}$  is the overall chamber correction factor that accounts for the change in the chamber response due to the displacement of water by the ionization chamber (air cavity plus wall) and the presence of the chamber stem, the change in the energy, and the angular distribution of the photon beam in the phantom compared to that used for the calibration in air, and  $k_{\text{sheath}}$  is a correction factor for the effects of a waterproofing sheath. The transfer chamber of type TM30013 is waterproof, thus  $k_{\text{sheath}}$  must not be applied. The correction factor  $k_{\text{ch}}$  was shown [2] to vary by less than 0,5 % for qualities generated with tube voltages between 70 and 280 kV (TH 70 to TH 280 in Ref. [2]).

$(\bar{\mu}_{\text{en}}/\rho)_{\text{w,a}}^d$  is obtained with relative uncertainties of about 0,3%. Mean values for the radiation qualities of Table 3 were calculated based on the mono-energetic ratios shown in Figure 1 and averaged with photon fluence spectra measured at PTB. Photon fluence spectra in 2 cm and 5 cm depth of a water phantom were obtained by the EGSnrc [5] user code FLURZnrc [6] using a parallel beam of photons of circular diameter 10 cm and the measured free-in-air spectra incident on a cylindrical water phantom of diameter 30 cm. Results for the qualities TH 70 – TH 280 are shown in Table 4 together with the chamber correction factor  $k_{\text{ch}}$  as published in [2]. Similar values can be expected for the transfer ionisation chambers of the same type.



**Figure 1:** Ratio of water-to-air mass energy-absorption coefficients as a function of photon energy. Mass energy-absorption coefficients of air and water were taken from [7].

From Table 4 it is obvious that values of  $(\bar{\mu}_{\text{en}}/\rho)_{\text{w,a}}^{d=2\text{cm}}$  and  $(\bar{\mu}_{\text{en}}/\rho)_{\text{w,a}}^{d=5\text{cm}}$  differ by no more than about 0.2 % so that calibrations in different depths will not result in significantly

different calibration coefficients. The quality correction factor  $k_Q$  with respect to the air kerma calibration coefficient normalized at TH280 of the chamber TM30013 was shown to be almost equal to one for TH70 to TH280 [2]. Thus it can be expected that this chamber type will indicate correctly the air kerma in the different depths of a water phantom because the air kerma response is almost independent of the photon fluence spectrum in this energy range. In summary, according to the results described above it is expected that the ratios  $N_{Dw} / N_K$  of the transfer chambers are close to  $(\bar{\mu}_{en}/\rho)_{w,a}^d$  within less than 1 %. Therefore, this chamber type is well suited for this type of comparison and the conclusions which can be made.

**Table 4:** Ratio for water-to-air of the mean mass energy-absorption coefficients averaged over the incident photon spectrum and at 2 cm and 5 cm depth in a reference water phantom, the chamber correction factor  $k_{ch}$  and the radiation quality correction factor with respect to the air kerma calibration coefficient at TH280 of the chamber TM30013-425 as published in [2]. The values for the CCRI qualities were obtained from additional calculations and measurements according to the methods described in [2].

| Radiation quality | Tube voltage<br>kV | Mean energy<br>keV | $(\bar{\mu}_{en}/\rho)_{w,a}^{air}$ | $(\bar{\mu}_{en}/\rho)_{w,a}^{d=2\text{ cm}}$ | $(\bar{\mu}_{en}/\rho)_{w,a}^{d=5\text{ cm}}$ | $k_{ch}$ | $k_Q$ |
|-------------------|--------------------|--------------------|-------------------------------------|---|---|----------|-------|
| TH 70             | 70                 | 42.0               | 1.020                               | 1.020   | 1.021   | 1.004    | 1.000 |
| CCRI 100          | 100                | 50.9               | 1.030                               | 1.029   | 1.031   | 1.005    | 0.999 |
| TH 100            | 100                | 52.8               | 1.033                               | 1.031   | 1.032   | 1.005    | 0.999 |
| TH 120            | 120                | 60.0               | 1.043                               | 1.038   | 1.040   | 1.007    | 0.997 |
| CCRI 135          | 135                | 67.2               | 1.055                               | 1.047   | 1.047   | 1.007    | 0.998 |
| TH 140            | 140                | 67.6               | 1.055                               | 1.047   | 1.048   | 1.007    | 0.998 |
| TH 150            | 150                | 78.5               | 1.071                               | 1.060   | 1.059   | 1.006    | 0.999 |
| CCRI 180          | 180                | 84.8               | 1.076                               | 1.065   | 1.064   | 1.006    | 0.999 |
| TH 200            | 200                | 99.4               | 1.089                               | 1.078   | 1.076   | 1.005    | 0.998 |
| TH 250            | 250                | 122.2              | 1.099                               | 1.090   | 1.089   | 1.003    | 0.999 |
| CCRI 250          | 250                | 122.6              | 1.098                               | 1.089   | 1.087   | 1.003    | 0.999 |
| TH 280            | 280                | 146.3              | 1.105                               | 1.098   | 1.097   | 1.002    | 1.000 |

The final result will be the comparison of the primary absorbed dose to water standards of the participants reflected by the mean values and uncertainties of the final calibration coefficients of the three transfer chambers obtained at the same or at least at sufficiently similar radiation qualities. In addition, information is gained about the consistency of the ratio  $N_{Dw} / N_K$  for the chamber type TM30013 as obtained by the participants.

## 5.5 Publication of the results

After finishing the whole program with all the participants the pilot laboratory will prepare a Draft A of the report of the results of this comparison. The Draft A will be circulated for comments and discussion of the results and once agreement is reached, the revised report Draft B, will be produced as the official final report of the comparison. This will be the first



comparison between recently developed primary standards of absorbed dose to water in the medium-energy x-ray range. In addition, results of  $N_{Dw} / N_K$  for this chamber type will be obtained. Therefore it is justified and suggested to publish the results in *Phys. Med. Biol.* and/or *Metrologia*. The results may also be useful for the support of CMC entrances of the participant's institutes. For this purpose it is suggested to publish the results on the BIPM Key Comparison Data Bank (KCDB).

## 6 References

- [1] De Prez L A and de Pooter J A 2008 The new NMi orthovolt X-rays absorbed dose to water primary standard based on water calorimetry *Phys. Med. Biol.* 53 3531-42
- [2] Krauss A, Büermann L, Kramer H-M and Selbach H-J 2012 Calorimetric determination of the absorbed dose to water for medium-energy X-rays with generating voltages from 70 kV to 280 kV *Phys. Med. Biol.* 57 6245-68
- [3] Rapp B, Perichon N, Denoziere M, Daures J, Ostrowski A and Bordy J-M 2013 The LNE-LNHB water calorimeter for primary measurement of absorbed dose at low depth in water: application to medium-energy x-rays *Phys. Med. Biol.* 58 2769-86
- [4] Guide to the Expression of Uncertainty in Measurement 2<sup>nd</sup> edn (Geneva: International Organization for Standardization) 1995 ISBN 92-67-10188-9
- [5] Kawrakow I and Rogers D W O 2006 The EGSnrc code system: Monte Carlo simulation of electron and photon transport Report PIRS-701 (Ottawa: National Research Council of Canada (NRCC)) <http://www.irs.inms.nrc.ca/inms/irs/EGSnrc>
- [6] Rogers D W O, Kawrakow I, Seuntjens J P, Walters B R B and Mainegra-Hing E 2005 NRC user codes for EGSnrc NRCC Report PIRS-702 (revB) <http://www.irs.inms.nrc.ca/EGSnrc/pirs702/>
- [7] Seltzer S M 1993 Calculation of photon mass energy-transfer and mass energy-absorption coefficients *Radiat. Res.* 136 147-70
- [8] CCEMRI 1972 Qualités de rayonnement *Comité Consultatif pour les Étalons de Mesures des Rayonnements Ionisants (Section I)* [2<sup>nd</sup> meeting R15-16](#)

**APPENDIX B: Complete addresses of the participants (used for shipment!)****PTB / Germany**postal address:

Physikalisch-Technische Bundesanstalt (PTB)  
Department 6.2  
Bundesallee 100  
38116 Braunschweig  
Germany

contact person:

Ludwig Büermann  
Tel.: +49 531 592 6250  
Fax: +49 531 592 6205  
e-mail: [ludwig.bueermann@ptb.de](mailto:ludwig.bueermann@ptb.de)  
Internet: <http://www.ptb.de>

**Magyar Kereskedelmi Engedélyezési Hivatal (MKEH)**

*Hungarian Trade Licensing Office*  
Metrológiai Hatóság - *Metrology Authority*  
*Radiation Physics and Chemistry Section*  
H-1124 Budapest, Németvölgyi út 37-39.  
Hungary

## Contact person:

Gábor Machula  
Tel: +36 1 4585909  
Fax: +36 1 4585937  
E-mail: [machulag@mkeh.gov.hu](mailto:machulag@mkeh.gov.hu)  
web: <http://www.mkeh.gov.hu>

**ENEA/Italy**

## Postal address:

ENEA-INMRI Istituto Nazionale di Metrologia delle Radiazioni Ionizzanti  
Centro Ricerche ENEA Casaccia  
Via Anguillarese, 301  
I-00123 Santa Maria di Galeria (RM)

## Contact person:

Massimo Pinto  
Tel: +39.06.3048.4662  
Fax: +39.06.3048.6074  
email: [massimo.pinto@enea.it](mailto:massimo.pinto@enea.it)  
web: <http://inmri.enea.it>

**CEA SACLAY / France**

Postal address:

DRT / LIST / LNHB-LMD

Bât. 534 – PC 104

Route de Chateaufort (Départementale 36)

F-91191 GIF SUR YVETTE Cedex

FRANCE

Contact person:

Benjamin Rapp

Tel:

Fax:

email: [benjamin.rapp@cea](mailto:benjamin.rapp@cea)

web:

**VSL B.V./ Netherlands**

Postal address:

Contact person:

Leon de Prez

Tel:

Fax:

email: [ldprez@vsl.nl](mailto:ldprez@vsl.nl)

web: