

Technical Protocol for the CIPM key comparison

CCM.D_K4 “Hydrometer”

S. Lorefice - Istituto Nazionale di Ricerca Metrologica (INRIM) - Italy

1 Outline

A CIPM key comparison concerning the calibration of hydrometers was proposed during the meeting of the Working Group on Density (WG-Density) of the Consultative Committee for Mass and Related Quantities (CCM) held on April 22nd, 2008 at the BIPM with the purpose of linking the regional comparisons previously performed under the auspices of the Regional Metrology Organizations APMP, EURAMET and SIM.

This CIPM key comparison, designated as CCM.D-K4, will be coordinated by the Istituto Nazionale di Ricerca Metrologica (INRIM, Italy) supported by all participants (Appendix A).

For the purpose of this project and to speed up the key comparison, the participating laboratories will be divided into two groups (petals) connected with each other by three independent laboratories INRIM (IT), CENAM (MX) and PTB (DE).

Each laboratory, belonging to an individual petal, will determine the corrections to be applied to three stated scale readings at 20 °C of different transfer standards in the density range between 600 kg/m³ to 2 000 kg/m³.

The linking laboratories INRIM, CENAM and PTB should calibrate all transfer standards involved in the comparison.

The INRIM will carry out the measurement at the beginning and at the end of the comparison.

The measurement are scheduled to end in June 2012. Draft A of the report should be available within December 2012 (Appendix B).

2 Purpose of this document

The purpose of this document is to provide the participating laboratories with instructions for handling the transfer standards (hydrometers) and to report on the measurement results, the measuring procedure and the apparatus.

It is important that all instructions given in this document are followed. This will ensure that the measurement data are obtained under comparable conditions and presented in the same format. Any deviation from the instructions has to be reported to the Pilot Laboratory.

3 Transfer Standards as Artefacts (Hydrometer samples)

For the comparison INRIM supplies two similar sets of four artefacts to be assigned to each petal and used as transfer standards at 20 °C. Each set will be concerned with two hydrometers with a scale division of 0.1 kg/m³ working near 600 kg/m³, and 1 300 kg/m³, an alcoholometer (alcohol hydrometers) cl.1 and an hydrometer with a scale division of 0.2 kg/m³ working near 2 000 kg/m³.

4 Circulation of the Artefacts

Each participating laboratory will send to the subsequent one the assigned set of four hydrometers as soon as calibrated, after that the laboratory, in accordance with the circulation scheme (Appendix C), and the pilot laboratory have been informed. The completion of the measurements and the date of dispatch, giving details of the transportation will be given by the form "Appendix E"..

Each set will be packed into a suitable container. The package will be transported by courier or directly by the personnel of the last laboratory. The package will be provided with a warning: To be opened only by laboratory personnel.

After arrival of the package, the participating laboratories will inform the Pilot Laboratory without delay giving details of the arrival date, the state of the package and its contents (Appendix D).

Each participating laboratory bears its own expenses for the transportation of the travelling set(s) to the next laboratory and any customs charges. The participants are responsible for completing the local customs formalities.

Note: An ATA carnet will be issued by INRIM for each package. If it is used, please verify that the carnet is presented to the customs authorities and correctly stamped. Each laboratory must provide all customs formalities in the shortest time to avoid delay in the circulation. A month has been estimated for the transportation, the export and the import formalities.

At the end of October 2011 both packages have to come back to INRIM to arrange, if it is necessary, for two new carnets.

5 Measurements

At least 5 weighing-in-air sequences have to be carried out for the weight determination of each hydrometer and to evaluate the experimental standard deviation. The mean of the parameters contributing to the air density evaluation are to be recorded, i.e. pressure, temperature, relative humidity (or dew point), and CO₂ content (whether measured or assumed). For the calculation of the air density, the revised CIPM formula (2007) is to be used [1]. The mean of this value is to be reported.

5.1 Mass measurement

Mass least 5 weighing-in-air sequences have to be carried out for the weight determination of each hydrometer and to evaluate the experimental standard deviation.

The mean of the parameters contributing to the air density evaluation are to be recorded, i.e. pressure measurements should be made under ambient laboratory conditions close to 20 °C. At, temperature, relative humidity (or dew point), and CO₂ content (whether measured or assumed). For the calculation of the air density, the revised CIPM formula (2007) is to be used [1]. The mean of this value is to be reported.

5.2 Hydrostatic weighing

After the mass measurement, the hydrostatic weighing is to be performed without cleaning. The measurements should be made close to 20 °C. The cubic expansion coefficient for all hydrometers is assumed to be $25 \cdot 10^{-6} \text{ } ^\circ\text{C}^{-1}$ with an uncertainty of $2 \cdot 10^{-6} \text{ } ^\circ\text{C}^{-1}$, rectangular




Figure 1. Alignment of the stated scale reading with the horizontal plane of liquid.

distribution.

At least 5 weighing sequences have to be carried out for the weight determination at each of the three scale readings stated and to evaluate the experimental standard deviation. Adjustment of the scale readings to the liquid level should be made when the middle of the line is aligned with the horizontal plane of liquid (Figure 1).

The mean of the parameters contributing to the air density evaluation are to be recorded, i.e. pressure, temperature, relative humidity (or dew point), and CO₂ content (measured or assumed). The mean value of the air density is to be reported.

6 Reports

As soon as the measurements are over, each laboratory will send the Pilot Laboratory a summary of the procedure used describing the apparatus, including, if possible, any reference, giving the mathematical model equations for calculating the corrections and how the standard uncertainties of the individual influence quantities are estimated. Besides, information and results should be made up using the enclosed MS Excel Report Form.

The MS Excel Report Form consists of two parts:

6.1 Report Form 1

It concerns information about the instrumentation used in the project. Please add any additional information obtained in your measurements.

For the balance/s used to determine the weighing value in air/liquid, the following information is to be given:

- a) Manufacturer and type of balance,
- b) Maximum capacity, electronic range, resolution,
- c) Standard deviation, maximum non-linearity, out-of-centre error,
- d) Calibration uncertainty, method and frequency.

If mass standards are used as "substitutional weight" for the comparison of mass of hydrometer to the mass of weights, the following information is to be given:

- a) Manufacturer, name and material of weights,
- b) Identification of weights, their masses and standard uncertainties ($k=1$),
- c) Date of last calibration and traceability.

For the thermostat system used to stabilize the temperature of the buoyant liquid, the following information is to be given:

- a) Manufacturer and type of thermostat,

- b) Capacity,
- c) Temperature stability and uniformity at 20 °C.

For the alignment system used for alignment the scale readings, the following information is to be given:

- a) Type of magnifier,
- b) Method of alignment,
- c) Uncertainty of alignment.

For the instruments used for the determination of the density of air (air pressure, temperature, humidity and CO₂ content), the following information is to be given:

- a) Manufacturer, type,
- b) Resolution,
- c) Frequency of measurement,
- d) Calibration uncertainty, date and traceability.

For the instruments used for the measurement of the liquid temperature, the following information is to be given:

- a) Manufacturer and type of sensor,
- b) Manufacturer and type of resistance bridge and standard resistor (if applicable),
- c) Resolution of temperature measurement,
- d) Calibration uncertainty, date and traceability of thermometer (whole temperature range).

For the apparatus used for determining the surface tension of the buoyant liquid, the following information is to be given:

- a) Manufacturer and measuring method,
- b) Calibration uncertainty and traceability.

For the reference buoyant liquid the following information is to be given:

- a) Manufacturer,
- b) product name,
- c) molecular formula and assay.

6.2 *Report Form 2(ID Hydrometer)*

For each hydrometer, general information and results of measurements are to be given.

General information is to be given in:

Table 1-1: Movement of the travelling standard

- a) Date of arrival of the hydrometer at the laboratory,

- b) Condition of package and of the individual hydrometer,
- c) Date of shipment of hydrometers,
- d) Company responsible for transportation.

Table 1-2: Measurement period

- a) Date for hydrometer mass measurement,
- b) Date for hydrostatic weighing.

Measurement information is to be given in:

Table 1-3: Mass determination

- a) Substitutional weights, if they are used, and mean value of the hydrometer mass for five weights in air and its standard deviation of the mean.
- b) Ambient conditions during mass measurements including data about air density, air temperature, air pressure, humidity and CO₂ content.

Table 1-4: Hydrostatic weighing

- a) Mean value of the additional weight (ballast) if a stainless steel weight is added to the hydrometer to cause it to sink, and the standard deviation for the mean.
- b) Mean values of the hydrometer weighing value at the indicated scale reading. The standard deviation for the mean and the property of buoyant liquid during the hydrostatic weighing.
- c) Ambient conditions during hydrostatic weighing, including data about air density, air temperature, air pressure, humidity and CO₂ content.

Table 1-5: Results

- a) Corrections at each scale reading as determined at 20 °C. In the last column the reference surface tension is given.

Table 1-6: Uncertainty budget for the hydrometer corrections

A list of main components of the uncertainty budget is given. Please add any additional component occurring in your measurements.

The uncertainty evaluation should include a list of all influence quantities, values, their degrees of freedom and their combined standard uncertainty. This is obtained by combining the individual standard uncertainties obtained from Type A and Type B evaluations, according to ISO "Guide to the Expression of Uncertainty in Measurement" [2].

A pooled experimental standard deviation characterizing the hydrometer weighing value in buoyant liquid is to be given [2 (H.3.6)].

Table 1-7: Uncertainty of the hydrometer corrections

The uncertainty of measurement for the corrections is calculated here with the Excel sheet from the above given data, as well as the effective degrees of freedom ν_{eff} of the combined standard uncertainty u_c , the t-factor $t_{95}(\nu_{\text{eff}})$ taken from the t-distribution for a 95% confidence level and the expanded uncertainty for the corrections as $U_{95} = t_{95}(\nu_{\text{eff}}) \cdot u_c$.

7 Deadline

The reports are to be sent to the Pilot Laboratory as soon as possible but six weeks after the measurements are completed at the latest. A result will not be considered complete unless an associated uncertainty supported by a complete uncertainty budget is given. The results are confidential until all the participants have completed their measurements and all the results have been received (or until the deadline for receipt of results is over).

8 Special problem

Please, do not hesitate to contact the Pilot Laboratory for any questions.

8.1 Breach of travelling standard

In case a hydrometer is broken, its substitution will be decided at that time.

8.3 Late entry of a participant

Due to the tight timetable, it is not possible for any additional participant to join after the circulation has started. Only if an additional comparison will be carried out due to breach of the hydrometer/s or due to unexpected delays might a late entry be possible.

9 References

- [1] A. Picard, R.S. Davies, M. Glaser, and K. Fujii : "Revised formula for the density of moist air (CIPM 2007)," *Metrologia*, 2008, 45, pp. 149 - 155.
- [2] "Guide to the Expression of Uncertainty in Measurement," International Organization for Standardization (ISO), 1995.

Appendices

A. Participants

Laboratory	Country code	Contact Person
Istituto Nazionale di Ricerca Metrologica (INRiM) Strada delle Cacce, 91 -73 10135 Torino - ITALY	IT	Salvatore Lorefice S.Lorefice@inrim.it Tel: +39 011 3977 929 Fax: +39 011 3977 937
Centro Nacional de Metrologia (CENAM) Km 4.5 Carretera a los Cués, Mpio El Marqués, Querétaro - MEXICO	MX	Luis Omar Becerra lbecerra@cenam.mx Tel.: +52 442 2 110573 +52 442 2 110500 to 04 ext. 3602 Fax: +52 442 2 110568
Korea Research Institute of Standards and Science (KRISS) Div. Physical Metrology Center for Mass and Related Quantities 267 Gajeong-Ro Yuseong-Gu, Daejeon 305-340 – REPUBLIC of KOREA	KR	Yong Jae Lee yjlee@kriss.re.kr Tel.: +82 42 868 5244 Fax: +82 42 868 5012
Physikalisch-Technische Bundesanstalt (PTB) Bundesallee 100 38116 Braunschweig - GERMANY	DE	Hans Toth Hans.Toth@PTB.DE Tel.: +49 531 592 3114 Fax: +49 531 592 3015
Magyar Kereskedelmi Engedélyezési Hivatal (MKEH) 1124 Budapest., Németvölgyi út 37-39 Budapest - HUNGARY	HU	Csilla Vámosy vamosyacs@mkeh.hu Tel.: +36 1 4585 947 Fax: +36 1 4585 927
National Institute of Standards and Technology (NIST) 100 Bureau Dr., Mail Stop 8361, Gaithersburg, MD 20899 – United States of America	US	John Wright John.wright@nist.gov Tel. +301 975 5937 Fax: +301 258 9201
National Metrology Institute of Japan (NMIJ) 1-8-31, Midirigaoka, Ikeda, Osaka, 563-8577 - JAPAN	JP	Kunihiko Toda k.toda@aist.go.jp Tel.: +81 72 751 8644 Fax: +81 72 751 8696
Central Office of Measures - Główny Urząd Miar (GUM) Physical Chemistry Department, Density, Viscosity and Spectral Analysis Laboratory, ul. Elektoralna 2 00-139 Warszawa - POLAND	PL	Elżbieta Lenard, density@gum.gov.pl Tel :+ 48 22 581 9410 Fax :+ 48 22 581 9395
Laboratoire National d'Essais (LNE) 1 rue Gaston Boissier 75724 Paris Cedex 15 - FRANCE	FR	Tanguy Madec, Paul-Andre Meury tanguy.madec@lne.fr Paul-Andre.Meury@lne.fr Tel: +33 1 40 43 39 34 Fax: +33 1 40 43 37 37
National Measurement Institute (NMIA) Bradfield Rd, Lindfield, NSW 2070 - AUSTRALIA	AU	John Man, Kitty Fen John.man@measurement.gov.au Kitty.fen@measurement.gov.au Tel.: +612 8467 3513 +612 8467 3519 Fax: +612 84673754
Laboratorio Tecnológico del Uruguay (LATU) Av. Italia 6201 Montevideo - URUGUAY	UY	Joselaine Cáceres jcaceres@latu.org.uy Tel.: +598 2601372 office: 1298 Fax: +598 26018554

B. Timetable for the comparison

September 2010	Answers to questionnaire
December 2010	Agreement on Technical Protocol
December 2010 – January 2011	Registration of CCM.DK4 key comparison
January 2011	Start of measurements (Pilot Laboratory)
March 2011 to June 2012	Measurements by all participants and Reports from all participants
December 2012	Draft A of comparison report
February 2013	Draft B of comparison report: end of comparison

C. Circulation scheme for the comparison

Group A		Group B	
Laboratory	Data	Laboratory	Data
INRIM	January 2011	INRIM	January 2011
CENAM	March – April 2011	CENAM	March – April 2011
LATU	June 2011	NMIA	June 2011
NMIJ	August 2011	NIST	August 2011
LNE	October 2011	KRISS	October 2011
to INRIM for ATA Carnet formality			
GUM	December 2011	MKEH	December 2011
PTB	February – March 2012	PTB	February – March 2012
INRIM	May - June 2012	INRIM	May - June 2012

D. Fax/e-mail: Receipt of a comparison package

To monitor the progress of the comparison, we ask on receipt of the package to kindly send a report by telefax to

Salvatore Lorefice
INRIM
Strada delle Cacce, 91 - 73
10135 Torino
ITALY
Fax: +39 011 3977 937
e-mail: S.Lorefice@inrim.it

This report should contain the following information:

Participating laboratory
Contact person
Telephone
Fax

and a text like this:

The package of CMC.D-K4 "Hydrometer" was received on (date).

The package seems, after short inspection, (not) to be damaged.

If damaged: (not) seriously.

The contents is probably (not) suitable for use.

Remarks:

Date, signed

E. Fax/e-mail: Progress report

To monitor the progress of the comparison, we ask to kindly send a report by fax/e-mail to

Salvatore Lorefice
INRIM
Strada delle Cacce, 91 - 73
10135 Torino
ITALY
Fax: +39 011 3977 937
e-mail: S.Lorefice@inrim.it

This report should contain the following information:

Participating laboratory
Contact person
Telephone
Fax

and a text like this:

Object: CCM.D-K4 "Hydrometer"

The measurements were completed on (date).

[The comparison package was shipped on (date) through the forwarding agency xx.]

[The results were sent to the Pilot Laboratory on (date).]

Remarks:

Date, signed