

# **REPORT ON KEY COMPARISON APMP.AUV.A–K3**

**1 December, 2008**

**Dr Hyu-sang Kwon**

**Korea Research Institute of Standards and Science**

**(KRISS)**

## **Abstract**

A regional key-comparison APMP.AUV.A-K3 has been carried out for the pressure sensitivity of laboratory standard microphones. The Korea Research Institute of Standards and Science (KRISS) was the pilot laboratory for this project. Two LS2P microphones were circulated through ten national metrology institutes and calibrated in the frequency range from 31.5 Hz to 25 kHz (31.5 kHz was optional). Deviations from the mean value for all participants are within their declared expanded uncertainties for all frequencies except the optional 31.5 kHz. The results will be linked to the CIPM key comparison CCAUV.A-K3 in a separate report.

## **1. Introduction**

This is the final report for the regional key-comparison APMP.AUV.A-K3. It has taken into account the comments from the participants and has been approved by them. It has also been approved by the CCAUV.

This report includes calibration results from the participants and an outline of proposed linking to the CIPM key-comparison CCAUV.A-K3 <sup>[1]</sup>. The Korea Research Institute of Standards and Science (KRISS) prepared this report as the pilot laboratory.

## **2. Protocol**

The basis of this key-comparison was pressure calibration of laboratory standard microphones. Two LS2P microphones, a Bruel & Kjaer type 4180 with serial number 1763688 and a Bruel & Kjaer type 4180 with serial number 2341431 were supplied by KRISS.

A technical protocol instructed participants to submit their uncertainty budget prior to the key comparison and then to report the pressure sensitivity in the frequency range from 31.5 Hz to 25 kHz (31.5 kHz was optional) and at the reference environmental conditions in their usual certificate forms. Additional information was requested on any deviation from the requirements of IEC 61094-2, together with the estimated uncertainty.

Ten national metrology institutes took part in the project, as listed in Table 1.

Table 1. Participants in the key-comparison APMP.AUV.A-K3.

Participant	Acronym	Economy
National Institute of Metrology (Thailand)	NIMT	Thailand
Center for Measurement Standards Industrial Technology Research Institute	CMS/ITRI	Taiwan
National Measurement Institute (Australia)	NMIA	Australia
National Physical Laboratory of India	NPLI	India
Standards and Calibration Laboratory	SCL	Hong Kong
National Metrology Laboratory SIRIM Berhad	NML/SIRIM	Malaysia
National Metrology Institute of Japan	NMIJ	Japan
National Institute of Metrology	NIM	China
National Metrology Centre Agency for Science, Technology and Research, Singapore	NMC	Singapore
Korea Research Institute of Science and Standards	KRISS	Korea

### 2.1. Circulation of the microphones

This key comparison permitted international delivery services for the transportation of the microphones. Therefore KRISS paid close attention to their travelling container. The container was supplied by NMIJ and previously used in key comparison APMP.AUV.A-K1. The microphones were packaged in an aluminum box padded with cushioning material and containing small holes in the outside casing to avoid sudden shocks and to minimize extreme changes in temperature or pressure, which could cause an irreversible change in the sensitivity or degrade the stability of the microphones. The KRISS also recommended a reliable shipping agency that had handled other kinds of travelling standards previously.

Microphones were circulated from July 2006 to June 2007 and returned to the KRISS for a check calibration each time two participants completed their calibrations. The circulation proceeded exactly as planned due to the participants' cooperation.

### 2.2. Measurement frequencies

The protocol specified the calibration of the microphones at the nominal frequencies as in the CCAUV.A-K3; the nominal preferred octave frequencies from 31.5 Hz to 4 kHz (i.e. 31.5, 63.0, 125.0, 250.0, 500.0, 1000.0, 2000.0 and

4000.0 Hz) and the nominal preferred 1/3 octave frequencies from 6.3 kHz to 25 kHz (i.e. 6.3, 8.0, 10.0, 12.5, 16.0, 20.0 and 25.0 kHz). The 31.5 kHz frequency was optional. Experience in the CIPM key comparison CCAUV.A-K3 showed that the set of frequencies to be used should be specified clearly <sup>[1]</sup> and this improvement proved to be largely effective in this key comparison.

### **3. Travelling standards**

The KRISS regularly monitored the stability of the microphones on their return to the pilot laboratory. Fig. 1 shows the deviation of each calibration from the average value at the KRISS. Observed changes over the period are well below the uncertainty quoted by the KRISS, thus confirming that the microphones had an acceptable level of stability during the circulation.

Therefore the very first measurement from the KRISS was used as the reported KRISS result for this key comparison. This practice follows previous CCAUV key comparisons.

Fig. 2 gives the temporal variation of the sensitivities at 250 Hz. For reference, results declared by the participants are included. There appears to be little significant correlation between the stability of the microphones and the results from participants.





Fig. 1. KRISS measurements during APMP.AUV.A-K3.





Fig. 2. Sensitivity variation during APMP.AUV.A-K3 at 250 Hz.

## 4. Methodologies

The protocol required that the calibration method should be based on IEC 61094-2, but this Standard does not mention any specific equipment to be used. The following descriptions give the methodologies and facilities used in this key comparison by each participant. Any variation from the requirements of IEC 61094-2 was also included, if declared.

**NIMT** – The NIMT used the reciprocity calibration system. In this system, the voltage transfer function was measured using the insert voltage technique. The transmitter current was determined by measuring the voltage across a calibrated capacitance connected in series with the transmitter.

Two plain wave couplers (cavity volume 3.098 cm<sup>3</sup> and 5.135 cm<sup>3</sup> respectively) were used, filled with air at all measurement frequencies. The microphones were set into the coupler without using grease on the assumption that leakage of air and sound was negligible. Capillary tube correction was not applied because a needle bung was fitted with each coupler instead of the capillary tube. The needle bung was attached to the coupler to equalize the static pressure and to act as an acoustic seal. No corrections were applied.

**CMS/ITRI** – The CMS/ITRI used two reference microphones and one microphone to be calibrated by selecting two microphones a group to carry out the calibration, one being as the transmitting microphone and the other as the receiving microphone during calibration. They placed the microphones separately into the cavity coupler and measured the ratio of voltage attenuation, thus obtaining three individual sets of sensitivity equations to solve the sound pressure sensitivity of microphone under calibration by the reciprocity method.

**NMIA** - Pressure sensitivity by the reciprocity technique was determined using 3 type LS2P laboratory standard microphones according to IEC 61094-2. The measurement data was analysed using the computer software MP.EXE microphone pressure sensitivity calibration program. The radial wave motion corrections were applied. Three different plane wave couplers of nominal lengths of 3 mm, 4 mm and 6 mm with maximum frequency limits of

31.5 kHz, 25 kHz and 16 kHz respectively were utilized and the results averaged in their overlapping frequency range.

The front cavity volume, equivalent volume and resonance frequency were determined by data fitting. The front cavity depth was measured using a microscope fitted with an electronic depth gauge and the damping factor was determined from the ratio of the sensitivities where they are at maxima and in the vicinity of 1000 Hz where they level out.

**NPLI** - The capacitor microphones were calibrated by absolute method in the frequency range 31.5 Hz to 25 kHz using plane wave couplers. Using three standard microphones in successive pairs the open-circuit voltage sensitivities of the microphones were obtained by using reciprocity calibration. The microphones were acoustically coupled in pairs by the air enclosed in a coupler. For each pair, one microphone was used as a sound source (transmitter) and the other as a receiver. The task was to measure the electrical transfer impedance  $U_R/i_T$  where  $U_R$  is the open circuit voltage of the receiver microphone and  $i_T$  is the current through the transmitter microphone. The current through the transmitter microphone was found by measuring the voltage across a reference capacitor connected in series with the microphone. The measurements were controlled by PC software running on the Windows platform. For each frequency five sets of measurements were made and the measurement data was stored in a text file along with microphone identification and ambient parameters. Using this information the calculation program calculates the sensitivities of the three microphones in accordance with IEC-61094-2. Also the sensitivities valid at reference ambient conditions (Room Temperature  $T = 23\text{ }^\circ\text{C}$ , Static Pressure  $P_s = 101.325\text{ kPa}$  and Relative Humidity  $\text{RH} = 50\%$ ) were calculated through the use of built-in microphone correction data. To increase the accuracy, repeated calibrations were made using different couplers thus allowing the determination of the total volume along with other microphone parameters by an iterative process giving convergent results.

**SCL** - The open circuit pressure sensitivity of the microphone was determined by reciprocity technique in accordance with the international standard IEC 61094-2:1992. A long plane wave and a short plane wave coupler were used in the measurement. For the frequencies 31.5 Hz to



2000 Hz, both the long and short couplers were used. The measurement results reported in the certificate for the frequency range from 31.5 Hz to 2000 Hz were the average of the two results. For frequencies above 2000 Hz, only the short plane wave coupler was used.

**NML/SIRIM** - Absolute calibration method using Reciprocity Calibration System according to IEC 61094-2:1992 was used. The open circuit sensitivities of the microphones are obtained by using a combination of reciprocity calibration and insert voltage techniques. The microphones are acoustically coupled in pairs by the air enclosed in a coupler. There are two couplers (short and long) used in this measurement with precisely determined dimensions. For each pair of microphones, one microphone is used as a sound source (transmitter), and the other as a receiver. Included with the system are two PC programs, one to control measurements and the other to handle the sensitivity calculations according to the IEC Standard.

**NMIJ** - The pressure sensitivity was determined in compliance with IEC 61094-2, using a reciprocity calibration system developed by NMIJ. In the system, both the signal generation and the signal processing were executed by a dual-channel FFT analyzer, model CF-5220 of ONO SOKKI Co. Signal to noise ratio was improved by the synchronous waveform averaging method. The insert voltage technique was used to cancel the effect of the gain and impedance of an electrical circuit. The calibration was performed by using software of our own making.

A Brüel & Kjær UA1430 plane-wave (short) coupler was used for the reciprocity calibration and a long coupler type UA1414 was also used for determining the equivalent volume of the microphones under test. Both couplers were filled with air and no grease was used to the contacting surfaces between the microphones and the coupler. Capillary tube correction was considered to be unnecessary because a capillary tube was blocked by a needle bung DA5563 so that both equalization of static pressure and an acoustic seal could be ensured. Correction for radial wave-motion was not applied either.

All the measurements were conducted within a room whose temperature and relative humidity were controlled ( $23.0 \pm 0.5$  °C and  $50 \pm 5$  %RH, respectively). The sensitivity was corrected to the reference environmental

conditions by using K. Rasmussen's method [2]. At 31.5 kHz, only pressure dependency has been corrected because of the lack of a reliable temperature coefficient.

Microphone parameters were determined as follows: The resonance frequency, loss factor and the cavity volume of the couplers were taken from Brüel & Kjær's nominal values. Front depth was measured using a microscope calibrated by a block gauge. Equivalent volume was calculated as an averaged value from 250 Hz to 4 kHz.

**NIM** - Conforming to the Reference Standard IEC 61094-2 Primary method for pressure calibration of laboratory standard microphones by the reciprocity technique, NIM carried out six complete measurements. The Reciprocity calibration system and the computer software MP.EXE determined the open-circuit pressure sensitivity of the microphones. The front cavity depth of the microphones was determined by using an optical method, and then the front volume calculated. The loss factor, pressure coefficient and temperature coefficient that we used were the typical values of the microphone. Equivalent volume was determined by analysis of the measurement curve using the software MP.EXE. Two plane wave couplers were used during measurements and coupler parameters that we used were also the typical values.

**NMC** - Before the measurements, the microphones were conditioned in the laboratory for 24 hours. The 200 volt polarisation voltage was measured and adjusted to within  $\pm 0.01$  V during test set-up, and was verified again after all the measurements were completed. A precision pressure sensor was used to monitor static pressure inside the measurement chamber during measurement.

The measurements were made according to IEC 61094-2 using Brüel & Kjær Reciprocity Calibration System Type 9699. Two reference microphones were paired with the two microphones under comparison. A short coupler (4.7 mm) was used for measuring the pressure sensitivities in the frequency range of 31 Hz to 25 kHz, and a long coupler (9.4 mm) was used for the frequency range of 31 Hz to 2 kHz. The measurement result for each frequency point was obtained by averaging the pressure sensitivities measured using the short and long couplers at the same frequency.

In the calculation of pressure sensitivities for the standard with serial

number 1763688, nominal values for microphone parameters were used. For the standard with serial number 2341431, measured values provided by Brüel & Kjær (B&K) were used in the calculation.

**KRISS** - The calibration is performed by reciprocity calibration according to IEC 61094-2 using three reciprocal microphones. The microphones are coupled in pairs with two plane-wave couplers of different length (nominal length: 4.7 mm and 9.4 mm).

The receiver microphone is connected to a preamplifier B&K type 2673/WH 3291 with insert voltage facilities and the transmitter microphone is connected to a similar housing but with grounded shield.

The electrical current through the transmitter is measured as the voltage across the 4.7 nF capacitor connected in series with the transmitter. The measuring instruments are: Sine Generator B&K 1051, Digital Multimeter Wavetek 1281 and 1/3 Octave band Pass Filter B&K 1617. The measurements are made at discrete frequencies controlled via computer.

During the calibrations the coupler and the microphones are located under a cylindrical bulb of volume of about 20 litre and the cylindrical bulb is again enclosed by the acryl box with dimensions of 650 mm(W) × 859 mm(H) × 700 mm(D). The static pressure is measured using a Multifunction Pressure Indicator, Druck DPI 145, the temperature and the relative humidity are measured using a Testo 650.

The front cavity depths of the microphones are measured using the Video Measuring Scope, Nikon, VMH-300N. The equivalent volume is determined by fitting the final results for the two couplers in the frequency range up to about 2 kHz. Nominal values are assumed for the resonance frequency and the loss factor of the microphone diaphragm.

## 5. Results

The pressure sensitivities of the two microphones determined by each participant are shown in Table 2. They are presented with two or three decimals, depending on the way they were reported. Table 3 gives the associated measurement uncertainties. For the optional 31.5 kHz frequency, only three participants reported their results. The SCL did not measure the pressure sensitivity at 6300 Hz.

Table 2 (a). Pressure sensitivity (dB re 1 V/Pa) for B&K 4180 No 1763688.

Freq. (Hz)	NIMT	CMS/ ITRI	NMIA	NPLI	SCL	SIRIM	NMIJ	NIM	NMC	KRISS
31.5	-39.53	-39.55	-39.571	-39.58	-39.559	-39.52	-39.58	-39.558	-39.520	-39.558
63	-39.54	-39.56	-39.583	-39.58	-39.573	-39.54	-39.58	-39.568	-39.534	-39.553
125	-39.55	-39.57	-39.589	-39.59	-39.583	-39.54	-39.59	-39.574	-39.543	-39.559
250	-39.56	-39.58	-39.596	-39.60	-39.590	-39.55	-39.59	-39.584	-39.551	-39.564
500	-39.56	-39.58	-39.601	-39.60	-39.594	-39.55	-39.60	-39.588	-39.556	-39.567
1000	-39.56	-39.58	-39.600	-39.60	-39.594	-39.55	-39.59	-39.588	-39.555	-39.566
2000	-39.55	-39.56	-39.582	-39.58	-39.575	-39.53	-39.58	-39.569	-39.539	-39.547
4000	-39.46	-39.48	-39.501	-39.50	-39.495	-39.46	-39.50	-39.488	-39.450	-39.461
6300	-39.31	-39.33	-39.352	-39.34	–	-39.32	-39.35	-39.337	-39.305	-39.310
8000	-39.18	-39.18	-39.216	-39.19	-39.214	-39.19	-39.22	-39.201	-39.179	-39.172
10000	-39.02	-39.01	-39.049	-39.04	-39.049	-39.04	-39.06	-39.032	-39.019	-39.010
12500	-38.88	-38.83	-38.886	-38.85	-38.891	-38.89	-38.91	-38.866	-38.874	-38.858
16000	-38.94	-38.92	-38.907	-38.93	-38.946	-38.97	-38.95	-38.933	-38.918	-38.932
20000	-39.68	-39.67	-39.658	-39.67	-39.703	-39.75	-39.64	-39.675	-39.660	-39.695
25000	-41.52	-41.63	-41.620	-41.66	-41.684	-41.69	-41.51	-41.540	-41.681	-41.770
31500	–	–	-44.793	–	–	–	-46.46	–	–	-45.835

Table 2 (b). Pressure sensitivity (dB re 1 V/Pa) for B&K 4180 No 2341431.

Freq. (Hz)	NIMT	CMS/ ITRI	NMIA	NPLI	SCL	SIRIM	NMIJ	NIM	NMC	KRISS
31.5	-38.70	-38.72	-38.748	-38.76	-38.728	-38.71	-38.73	-38.744	-38.730	-38.727
63	-38.71	-38.73	-38.752	-38.76	-38.743	-38.72	-38.75	-38.755	-38.744	-38.725
125	-38.73	-38.74	-38.756	-38.76	-38.751	-38.73	-38.76	-38.759	-38.752	-38.728
250	-38.73	-38.74	-38.760	-38.77	-38.757	-38.73	-38.76	-38.769	-38.758	-38.732
500	-38.74	-38.75	-38.762	-38.77	-38.759	-38.73	-38.76	-38.771	-38.760	-38.733
1000	-38.73	-38.74	-38.758	-38.76	-38.757	-38.73	-38.76	-38.769	-38.757	-38.729
2000	-38.71	-38.72	-38.734	-38.74	-38.732	-38.71	-38.73	-38.744	-38.733	-38.703
4000	-38.59	-38.61	-38.629	-38.63	-38.628	-38.60	-38.63	-38.639	-38.632	-38.593
6300	-38.40	-38.41	-38.432	-38.44	–	-38.41	-38.43	-38.440	-38.432	-38.395
8000	-38.22	-38.22	-38.251	-38.24	-38.253	-38.24	-38.25	-38.255	-38.258	-38.211
10000	-37.99	-37.99	-38.025	-38.02	-38.028	-38.02	-38.03	-38.026	-38.024	-37.992
12500	-37.78	-37.76	-37.806	-37.78	-37.811	-37.82	-37.82	-37.797	-37.799	-37.778
16000	-37.86	-37.88	-37.867	-37.87	-37.893	-37.92	-37.89	-37.879	-37.850	-37.873
20000	-38.93	-38.95	-38.918	-38.90	-38.981	-39.00	-38.92	-38.934	-38.902	-38.906
25000	-41.44	-41.49	-41.443	-41.33	-41.534	-41.53	-41.44	-41.414	-41.511	-41.357
31500	–	–	-45.023	–	–	–	-46.16	–	–	-46.118

Table 3. Declared expanded uncertainties at  $k = 2$  (dB).

Freq. (Hz)	NIMT	CMS/ ITRI	NMIA	NPLI	SCL	SIRIM	NMIJ	NIM	NMC	KRISS
31.5	0.06	0.04	0.08	0.12	0.08	0.04	0.06	0.08	0.06	0.08
63	0.04	0.04	0.06	0.11	0.04	0.04	0.06	0.05	0.04	0.05
125	0.04	0.04	0.05	0.08	0.04	0.04	0.06	0.05	0.04	0.05
250	0.04	0.04	0.05	0.05	0.04	0.04	0.05	0.05	0.04	0.04
500	0.04	0.04	0.04	0.05	0.04	0.04	0.05	0.05	0.04	0.04
1000	0.04	0.04	0.04	0.05	0.04	0.04	0.05	0.05	0.04	0.04
2000	0.04	0.04	0.04	0.08	0.04	0.04	0.05	0.05	0.04	0.04
4000	0.04	0.04	0.04	0.08	0.04	0.04	0.05	0.05	0.04	0.04
6300	0.04	0.04	0.04	0.10	–	0.04	0.05	0.05	0.04	0.04
8000	0.04	0.04	0.04	0.10	0.04	0.04	0.05	0.05	0.04	0.04
10000	0.04	0.06	0.05	0.11	0.04	0.05	0.05	0.05	0.04	0.05
12500	0.04	0.08	0.06	0.11	0.04	0.06	0.06	0.10	0.05	0.06
16000	0.05	0.08	0.08	0.12	0.05	0.07	0.07	0.10	0.06	0.07
20000	0.08	0.08	0.10	0.13	0.08	0.10	0.13	0.10	0.09	0.10
25000	0.14	0.12	0.15	0.17	0.14	0.17	0.18	0.12	0.15	0.20
31500	–	–	0.25	–	–	–	0.66	–	–	1.00

Every participant declared the same uncertainties for each microphone. Detailed uncertainty budgets for the participants are reproduced in Appendix A.

Microphone parameters are listed in Table 4, and temperature and pressure coefficients of the sensitivity in Table 5. Every participant submitted the same coefficients for each microphone. These coefficients are frequency dependent, but most of the participants reported just a single value. In this case, the pilot laboratory regarded the reported coefficients as the values at 250 Hz, if not specified. Furthermore, information on the couplers used in the key comparison is described in Table 6.

Table 4 (a). Microphone parameters for B&K4180 No 1763688.

	NIMT	CMS/ ITRI	NMIA	NPLI	SCL	SIRIM	NMIJ	NIM	NMC	KRISS
Front cavity volume /mm <sup>3</sup>	32.759	33.1	32.0	34	33.3	34	32.6	32.7	34	35
Front cavity depth /mm	0.488	0.48	0.482	0.463	0.5	0.5	0.48	0.479	0.5	0.505
Equivalent volume /mm <sup>3</sup>	9.2	6.6	8.1	7.56	7.1	9.2	8.1	7.9	9.2	6.7
Resonance freq. /kHz	22	22	23.0	22.4	23	22	22	22	22	22
Loss factor	1.05	1.05	1.05	1.05	1.05	1.05	1.05	1.05	1.05	1.05

Table 4 (b). Microphone parameters for B&K4180 No 2341431.

	NIMT	CMS/ ITRI	NMIA	NPLI	SCL	SIRIM	NMIJ	NIM	NMC	KRISS
Front cavity volume /mm <sup>3</sup>	33.818	33.4	33.0	32.8	33.6	34	33.2	32.9	33.3	33.1
Front cavity depth /mm	0.507	0.48	0.498	0.504	0.5	0.5	0.49	0.479	0.496	0.478
Equivalent volume /mm <sup>3</sup>	9.2	7.5	8.5	9.64	8	9.2	9.1	8.2	8.1	10.1
Resonance freq. /kHz	22	22	22.0	22.4	23	22	22	22	22	22
Loss factor	1.05	1.05	1.05	1.07	1.05	1.05	1.05	1.05	1.05	1.05

Table 5 (a). Temperature coefficients for pressure sensitivity (dB/K).

Freq. /Hz	NIMT	CMS/ ITRI	NMIA	NPLI	SCL	SIRIM	NMIJ	NIM	NMC	KRISS
31.5	-0.001	–	–	–	–	–	-0.002	–	–	-0.001
63	-0.001	–	–	–	–	–	-0.002	–	–	-0.001
125	-0.001	–	–	–	–	–	-0.002	–	–	-0.001
250	-0.001	-0.002	-0.002	-0.0012	-0.002	-0.002	-0.002	-0.002	-0.0012	-0.001
500	-0.001	–	–	–	–	–	-0.002	–	–	-0.001
1000	-0.001	–	–	–	–	–	-0.002	–	–	-0.001
2000	-0.002	–	–	–	–	–	-0.003	–	–	-0.002
4000	-0.003	–	–	–	–	–	-0.003	–	–	-0.003
6300	-0.005	–	–	–	–	–	-0.006	–	–	-0.005
8000	-0.007	–	–	–	–	–	-0.008	–	–	-0.007
10000	-0.010	–	–	–	–	–	-0.011	–	–	-0.011
12500	-0.014	–	–	–	–	–	-0.014	–	–	-0.015
16000	-0.015	–	–	–	–	–	-0.017	–	–	-0.016
20000	-0.009	–	–	–	–	–	-0.008	–	–	-0.008
25000	0.006	–	–	–	–	–	0.009	–	–	0.006
31500	–	–	–	–	–	–	–	–	–	0.012



Table 5 (b). Pressure coefficients for pressure sensitivity (dB/kPa).

Freq. /Hz	NIMT	CMS/ ITRI	NMIA	NPLI	SCL	SIRIM	NMIJ	NIM	NMC	KRISS
31.5	-0.005	–	–	–	–	–	-0.006	–	–	-0.005
63	-0.005	–	–	–	–	–	-0.006	–	–	-0.005
125	-0.005	–	–	–	–	–	-0.006	–	–	-0.005
250	-0.005	-0.0055	-0.0055	-0.0055	-0.007	-0.007	-0.006	-0.0055	-0.0064	-0.006
500	-0.005	–	–	–	–	–	-0.006	–	–	-0.006
1000	-0.005	–	–	–	–	–	-0.006	–	–	-0.006
2000	-0.005	–	–	–	–	–	-0.005	–	–	-0.005
4000	-0.003	–	–	–	–	–	-0.004	–	–	-0.004
6300	-0.001	–	–	–	–	–	-0.001	–	–	-0.001
8000	0.002	–	–	–	–	–	0.001	–	–	0.002
10000	0.006	–	–	–	–	–	0.005	–	–	0.005
12500	0.008	–	–	–	–	–	0.008	–	–	0.008
16000	-0.001	–	–	–	–	–	0.004	–	–	0.001
20000	-0.019	–	–	–	–	–	-0.017	–	–	-0.021
25000	-0.046	–	–	–	–	–	-0.043	–	–	-0.046
31500	–	–	–	–	–	–	-0.048	–	–	-0.049

Table 6. Information on the couplers.

(a) short plane-wave coupler

	NIMT	CMS/ ITRI	NMIA	NPLI	SCL	SIRIM	NMIJ	NIM	NMC	KRISS
Diameter /mm	9.3	9.3	9.3	–	9.3	9.3057	9.3	–	9.30729	9.3
Length /mm	4.7	4.7	3.0, 4.0	–	4.7	4.7019	4.7	–	4.69958	4.7
Freq. Range /Hz	20 to 25k	20 to 25k	20 to -31.5k, -25k	–	31.5 to 25k	20 to 25k	31.5 to 31.5k	–	31 to 25k	20 to 25k

(b) long plane-wave coupler

	NIMT	CMS/ ITRI	NMIA	NPLI	SCL	SIRIM	NMIJ	NIM	NMC	KRISS
Diameter /mm	9.3	9.3	9.3	–	9.3	9.3	9.3	–	9.30488	9.3
Length /mm	9.4	9.4	6.0	–	9.4	9.4	9.4	–	9.40285	9.4
Freq. Range /Hz	20 to 13k	20 to 13k	20 to 16k	–	31.5 to 2k	20 to 13k	250 to 4k	–	31 to 2k	20 to 13k

(c) capillary tube

	NIMT	CMS/ ITRI	NMIA	NPLI	SCL	SIRIM	NMIJ	NIM	NMC	KRISS
Number	–	0	none	–	none	none	none	–	–	–
Inner diameter /mm	–	0.335	–	–	–	–	–	–	–	–
Length /mm	–	50	–	–	–	–	–	–	–	–

For each of the two microphones and at each of the frequencies, the mean value of the pressure sensitivity level was determined from all the submitted data. Fig. 3 shows the results of individual participants expressed as the difference from this mean value.

Fig. 3 indicates that the trend for each participant is mostly common to both microphones, as observed in the CCAUV.A-K3. Therefore the average difference for the two microphones was taken to specify the performance of each participant and presented in Fig. 4.




Fig. 3 (a). Sensitivity deviations from the mean value for B&K4180 No 1763688.




Fig. 3 (b). Sensitivity deviations from the mean value for B&K4180 No 2341431.

Table 7 (a). Sensitivity deviations from the mean value for B&K4180 No 1763688 (dB).

Freq. /Hz	NIMT	CMS/ ITRI	NMIA	NPLI	SCL	SIRIM	NMIJ	NIM	NMC	KRISS
31.5	0.02	0.00	-0.02	-0.03	-0.01	0.03	-0.03	-0.01	0.03	-0.01
63	0.02	0.00	-0.02	-0.02	-0.01	0.02	-0.02	-0.01	0.03	0.01
125	0.02	0.00	-0.02	-0.02	-0.01	0.03	-0.02	-0.01	0.03	0.01
250	0.02	0.00	-0.02	-0.02	-0.01	0.03	-0.01	-0.01	0.03	0.01
500	0.02	0.00	-0.02	-0.02	-0.01	0.03	-0.02	-0.01	0.02	0.01
1000	0.02	0.00	-0.02	-0.02	-0.02	0.03	-0.01	-0.01	0.02	0.01
2000	0.01	0.00	-0.02	-0.02	-0.01	0.03	-0.02	-0.01	0.02	0.01
4000	0.02	0.00	-0.02	-0.02	-0.02	0.02	-0.02	-0.01	0.03	0.02
6300	0.02	0.00	-0.02	-0.01	–	0.01	-0.02	-0.01	0.02	0.02
8000	0.01	0.01	-0.02	0.00	-0.02	0.00	-0.03	-0.01	0.02	0.02
10000	0.01	0.02	-0.02	-0.01	-0.02	-0.01	-0.03	0.00	0.01	0.02
12500	-0.01	0.04	-0.01	0.02	-0.02	-0.02	-0.04	0.01	0.00	0.02
16000	-0.01	0.01	0.03	0.00	-0.01	-0.04	-0.02	0.00	0.02	0.00
20000	0.00	0.01	0.02	0.01	-0.02	-0.07	0.04	0.01	0.02	-0.01
25000	0.11	0.00	0.01	-0.03	-0.05	-0.06	0.12	0.09	-0.05	-0.14
31500	–	–	0.90	–	–	–	-0.76	–	–	-0.14

Table 7 (b). Sensitivity deviations from the mean value for B&K4180 No 2341431 (dB).

Freq. /Hz	NIMT	CMS/ ITRI	NMIA	NPLI	SCL	SIRIM	NMIJ	NIM	NMC	KRISS
31.5	0.03	0.01	-0.02	-0.03	0.00	0.02	0.00	-0.01	0.00	0.00
63	0.03	0.01	-0.01	-0.02	0.00	0.02	-0.01	-0.02	-0.01	0.01
125	0.02	0.01	-0.01	-0.01	0.00	0.02	-0.01	-0.01	-0.01	0.02
250	0.02	0.01	-0.01	-0.02	-0.01	0.02	-0.01	-0.02	-0.01	0.02
500	0.01	0.00	-0.01	-0.02	-0.01	0.02	-0.01	-0.02	-0.01	0.02
1000	0.02	0.01	-0.01	-0.01	-0.01	0.02	-0.01	-0.02	-0.01	0.02
2000	0.02	0.01	-0.01	-0.01	-0.01	0.02	0.00	-0.02	-0.01	0.02
4000	0.03	0.01	-0.01	-0.01	-0.01	0.02	-0.01	-0.02	-0.01	0.03
6300	0.02	0.01	-0.01	-0.02	–	0.01	-0.01	-0.02	-0.01	0.03
8000	0.02	0.02	-0.01	0.00	-0.01	0.00	-0.01	-0.02	-0.02	0.03
10000	0.02	0.02	-0.01	-0.01	-0.01	-0.01	-0.02	-0.01	-0.01	0.02
12500	0.02	0.04	-0.01	0.02	-0.02	-0.02	-0.02	0.00	0.00	0.02
16000	0.02	0.00	0.01	0.01	-0.01	-0.04	-0.01	0.00	0.03	0.01
20000	0.00	-0.02	0.02	0.03	-0.05	-0.07	0.01	0.00	0.03	0.03
25000	0.01	-0.04	0.01	0.12	-0.09	-0.08	0.01	0.03	-0.06	0.09
31500	–	–	0.74	–	–	–	-0.39	–	–	-0.35




Fig. 4. Deviations in amalgamated results from the mean value.

Table 8. Deviations in amalgamated results from the mean value /dB.

Freq. /Hz	NIMT	CMS/ ITRI	NMIA	NPLI	SCL	SIRIM	NMIJ	NIM	NMC	KRISS
31.5	0.03	0.01	-0.02	-0.03	0.00	0.03	-0.01	-0.01	0.02	0.00
63	0.02	0.00	-0.02	-0.02	-0.01	0.02	-0.02	-0.01	0.01	0.01
125	0.02	0.00	-0.01	-0.02	-0.01	0.02	-0.02	-0.01	0.01	0.01
250	0.02	0.00	-0.01	-0.02	-0.01	0.02	-0.01	-0.01	0.01	0.02
500	0.02	0.00	-0.01	-0.02	-0.01	0.03	-0.01	-0.01	0.01	0.02
1000	0.02	0.00	-0.02	-0.02	-0.01	0.02	-0.01	-0.01	0.01	0.02
2000	0.01	0.00	-0.01	-0.02	-0.01	0.02	-0.01	-0.01	0.01	0.02
4000	0.02	0.00	-0.02	-0.02	-0.01	0.02	-0.02	-0.01	0.01	0.02
6300	0.02	0.00	-0.02	-0.02	–	0.01	-0.02	-0.01	0.01	0.02
8000	0.02	0.02	-0.02	0.00	-0.02	0.00	-0.02	-0.01	0.00	0.03
10000	0.02	0.02	-0.01	-0.01	-0.01	-0.01	-0.02	-0.01	0.00	0.02
12500	0.00	0.04	-0.01	0.02	-0.02	-0.02	-0.03	0.00	0.00	0.02
16000	0.01	0.01	0.02	0.01	-0.01	-0.04	-0.01	0.00	0.02	0.00
20000	0.00	0.00	0.02	0.02	-0.03	-0.07	0.03	0.00	0.03	0.01
25000	0.06	-0.02	0.01	0.04	-0.07	-0.07	0.06	0.06	-0.06	-0.02
31500	–	–	0.82	–	–	–	-0.58	–	–	-0.24



## 6. Linking model for the CCAUV.A-K3

The goal is to establish a linkage between the results of the APMP.AUV.A-K3 and those of the CCAUV.A-K3. Four laboratories (NMIJ, KRISS, NIM and NMIA) listed in table 9 participated in both key-comparisons and play the important role as the linking laboratories.

In December 2002, the protocol for the CCAUV.A-K3 was issued containing the procedures to circulate the standards and a general framework for the measurements. The comparison consisted of two circulations (A and B) with two different standards per circulation.

Participant	Acronym	Country	Circulation
National Metrology Institute of Japan	NMIJ	Japan	A
Korea Research Institute of Science and Standards	KRISS	Korea	A
National Institute of Metrology	NIM	China	B
National Measurement Institute (Australia)	NMIA	Australia	B

Table 9. Participants in both key comparisons for linkage.

It is planned to link the APMP.AUV.A-K3 results with the CCAUV.A-K3 results using the results from the four laboratories that also participated in the CCAUV.A-K3 comparison. The methodology described in the CCAUV.A-K3 comparison will be used for the linkage and to determine the degrees of equivalence. The results of the linking will be included in a further report.

## 7. Conclusion

The KRISS has piloted the regional key comparison APMP.AUV.A-K3 and reported the results. Deviations from the mean value for all participants are within their declared expanded uncertainties for all frequencies except the optional 31.5 kHz. The results from the participants will be linked to the key comparison reference values of the CIPM key comparison CCAUV.A-K3.

## **Acknowledgement**

The authors gratefully acknowledge all the participating institutes for their thorough cooperation and fruitful discussion.

## **References**

- [1] Cutanda Henríquez V., Rasmussen K., Final report on the key comparison CCAUV.A-K3, *Metrologia* **43** (2006) *Tech. Suppl.* 09001.
- [2] Rasmussen K., The static pressure and temperature coefficients of laboratory standard microphones, *Metrologia* **36** (1999) 265-273.

## Appendix A. Uncertainty budgets

The uncertainty budgets submitted by the participants are reproduced here.

NIMT

National Institute of Metrology (Thailand)

Uncertainty Components of the Pressure Sensitivity Level for LS2P Microphone (unit quote in dB)

Frequency /Hz	31.5	63	125	250	500	1 k	2k	4k	6.3k	8k	10k	12.5k	16k	20k	25k
$U_{\text{Pol.V}}$	0.0004	0.0004	0.0004	0.0004	0.0004	0.0004	0.0004	0.0004	0.0004	0.0004	0.0004	0.0004	0.0004	0.0004	0.0004
$U_{\text{PCorr.}}$	0.0045	0.0045	0.0045	0.0045	0.0045	0.0045	0.0045	0.0045	0.0045	0.0045	0.0045	0.0046	0.0073	0.0187	0.0200
$U_{\text{Tcorr.}}$	0.0006	0.0006	0.0006	0.0006	0.0006	0.0006	0.0006	0.0006	0.0006	0.0006	0.0006	0.0006	0.0013	0.0032	0.0100
$U_{\text{RH.Corr.}}$	0.0010	0.0010	0.0010	0.0010	0.0010	0.0010	0.0010	0.0010	0.0010	0.0010	0.0010	0.0010	0.0010	0.0010	0.0010
$U_{\text{HWcf.}}$	0.0100	0.0100	0.0100	0.0100	0.0100	0.0100	0.0100	0.0100	0.0100	0.0100	0.0100	0.0100	0.0100	0.0100	0.0100
$U_{\text{P}}$	0.0008	0.0008	0.0008	0.0008	0.0008	0.0008	0.0008	0.0008	0.0008	0.0008	0.0008	0.0008	0.0008	0.0008	0.0008
$U_{\text{Spec.H}}$	0.0020	0.0020	0.0020	0.0020	0.0020	0.0020	0.0020	0.0020	0.0020	0.0020	0.0020	0.0020	0.0020	0.0020	0.0020
$U_{\text{Cap.}}$	0.0060	0.0060	0.0060	0.0060	0.0060	0.0060	0.0060	0.0060	0.0060	0.0060	0.0060	0.0060	0.0060	0.0060	0.0060
$U_{\text{Elect.Para.}}$	0.0073	0.0073	0.0073	0.0073	0.0073	0.0073	0.0073	0.0073	0.0073	0.0073	0.0073	0.0073	0.0073	0.0073	0.0073
$U_{\text{Coupler}}$	0.0176	0.0053	0.0032	0.0031	0.0031	0.0030	0.0030	0.0030	0.0030	0.0030	0.0030	0.0030	0.0030	0.0030	0.0030
$U_{\text{mic.}}$	0.0070	0.0070	0.0070	0.0070	0.0070	0.0070	0.0070	0.0070	0.0070	0.0070	0.0070	0.0070	0.0100	0.0300	0.0500
$U_{\text{A Repeat}}$	0.0120	0.0040	0.0030	0.0030	0.0030	0.0040	0.0030	0.0020	0.0020	0.0030	0.0040	0.0060	0.0100	0.0110	0.0400
Combined Standard Uncertainty	0.0268	0.0176	0.0169	0.0169	0.0169	0.0170	0.0168	0.0167	0.0167	0.0168	0.0170	0.0176	0.0214	0.0398	0.0693
Expanded Uncertainty ( $k = 2$ )	0.0536	0.0352	0.0337	0.0337	0.0337	0.0341	0.0337	0.0334	0.0334	0.0337	0.0341	0.0353	0.0429	0.0796	0.1386
Reported uncertainty /dB	0.06	0.04	0.04	0.04	0.04	0.04	0.04	0.04	0.04	0.04	0.04	0.04	0.05	0.08	0.14

Uncertainty Budget For LS2P Microphone (dB)

Symbol	Freq.	31.5	63	125	250	500	1000	2000	4000	6300	8000	10000	12500	16000	20000	25000
		Source of uncertainty, $x_i$														
<i>Cor<sub>R,n</sub></i>																
<i>U<sub>meas,deviation</sub></i>	Voltage ratio correction	0.009	0.009	0.009	0.009	0.009	0.009	0.009	0.009	0.009	0.009	0.009	0.009	0.009	0.009	0.009
<i>U<sub>meas,specification</sub></i>	Voltage Ratio, accuracy	0.003	0.003	0.003	0.003	0.003	0.003	0.003	0.003	0.003	0.003	0.003	0.003	0.003	0.003	0.003
<i>Cor<sub>CV</sub></i>																
<i>V<sub>coupl,specification</sub></i>	Coupler volume correction	0.011	0.011	0.011	0.011	0.011	0.011	0.011	0.011	0.011	0.011	0.011	0.011	0.011	0.011	0.011
<i>Cor<sub>P<sub>s</sub></sub></i>																
<i>P<sub>s,deviation</sub></i>	Static pressure correction	0.002	0.002	0.002	0.002	0.002	0.002	0.002	0.002	0.002	0.002	0.002	0.002	0.002	0.002	0.002
<i>Cor<sub>C</sub></i>																
<i>C<sub>traceability</sub></i>	Capacitance correction	0.002	0.002	0.002	0.002	0.002	0.002	0.002	0.002	0.002	0.002	0.002	0.002	0.002	0.002	0.002
<i>S<sub>ref</sub></i>																
<i>Cor<sub>HW</sub></i>	Heat conduction correction	0.006	0.006	0.006	0.006	0.006	0.006	0.006	0.006	0.006	0.006	0.006	0.006	0.006	0.006	0.006
<i>k</i>	Ratio of specific heats	0.004	0.004	0.004	0.004	0.004	0.004	0.004	0.004	0.004	0.004	0.004	0.004	0.004	0.004	0.004
<i>P<sub>s,error</sub></i>	Static pressure	0.000	0.000	0.000	0.000	0.000	0.000	0.000	0.000	0.000	0.000	0.000	0.000	0.000	0.000	0.000
<i>T<sub>error</sub></i>	Ambient temperature	0.000	0.000	0.000	0.000	0.000	0.000	0.000	0.001	0.002	0.004	0.006	0.014	0.004	0.008	0.018
<i>H<sub>errot</sub></i>	Ambient humidity	0.000	0.000	0.000	0.000	0.000	0.000	0.000	0.000	0.000	0.001	0.001	0.002	0.001	0.001	0.003
<i>L<sub>F,error</sub></i>	Microphone Cavity depth	0.000	0.000	0.000	0.000	0.000	0.000	0.000	0.001	0.003	0.005	0.008	0.018	0.011	0.019	0.043
<i>M<sub>P,repeat</sub></i>	Repeatability	0.003	0.005	0.005	0.005	0.006	0.006	0.006	0.005	0.005	0.005	0.008	0.013	0.019	0.016	0.026
<i>u<sub>c</sub></i>	Combined standard uncertainty	0.02	0.02	0.02	0.02	0.02	0.02	0.02	0.02	0.02	0.02	0.03	0.04	0.03	0.04	0.06
<i>U</i>	Expanded uncertainty ( $k = 2$ )	0.04	0.04	0.04	0.04	0.04	0.04	0.04	0.04	0.04	0.04	0.06	0.08	0.06	0.08	0.12
Stated Uncertainty		0.04	0.04	0.04	0.04	0.04	0.04	0.04	0.04	0.04	0.04	0.06	0.08	0.08	0.08	0.12

		Uncertainty Estimates		APMP-K3 Key Comparison on LS2P Microphones														NMI Australia		Jun-06		Meldrum/Bell/Narang			R	Dof	Distrib'n
Component	Symbol	Unit	Source	31.5	63	125	250	500	1000	2000	4000	6300	8000	10000	12500	16000	20000	25000	31500								
				31.5	63	125	250	500	1k	2k	4k	6.3k	8k	10k	12.5k	16k	20k	25k	31.5k								
<b>Electrical measurements</b>				0.004	0.00426	0.00426	0.00426	0.00426	0.00426	0.00426	0.00426	0.00426	0.00426	0.00426	0.00426	0.00426	0.00426	0.00426	0.00426	0.00426							
1.1	Series Z	C	0.001 nF	NMI	3.3E-10	3.3E-10	3.3E-10	3.3E-10	3.3E-10	3.3E-10	3.3E-10	3.3E-10	3.3E-10	3.3E-10	3.3E-10	3.3E-10	3.3E-10	3.3E-10	3.3E-10	3.3E-10	3.3E-10	3.3E-10	93				
1.2	Voltage ratio	Vr	V	NMI	0.0019	0.0019	0.0019	0.0019	0.0019	0.0019	0.0019	0.0019	0.0019	0.0019	0.0019	0.0019	0.0019	0.0019	0.0019	0.0019	0.0019	0.0019	30	Rect			
1.3	Frequency	f	Hz	NMI	0.0027	0.0027	0.0027	0.0027	0.0027	0.0027	0.0027	0.0027	0.0027	0.0027	0.0027	0.0027	0.0027	0.0027	0.0027	0.0027	0.0027	0.0027	30	Rect			
1.4	Inherent Noise	V	NMI	0.0001	0.0001	0.0001	0.0001	0.0001	0.0001	0.0001	0.0001	0.0001	0.0001	0.0001	0.0001	0.0001	0.0001	0.0001	0.0001	0.0001	0.0001	0.0001	9	Normal			
1.5	Distortion	D	ratio	B&K	0.0001	0.0001	0.0001	0.0001	0.0001	0.0001	0.0001	0.0001	0.0001	0.0001	0.0001	0.0001	0.0001	0.0001	0.0001	0.0001	0.0001	0.0001	30	Rect			
1.6	Cross talk	ratio	B&K	0.0025	0.0025	0.0025	0.0025	0.0025	0.0025	0.0025	0.0025	0.0025	0.0025	0.0025	0.0025	0.0025	0.0025	0.0025	0.0025	0.0025	0.0025	0.0025	30	Rect			
1.7	Polarising V	Fv	V	NMI	0.0010	0.0010	0.0010	0.0010	0.0010	0.0010	0.0010	0.0010	0.0010	0.0010	0.0010	0.0010	0.0010	0.0010	0.0010	0.0010	0.0010	0.0010	30	Rect			
<b>Coupler properties</b>				0.010	0.00711	0.00613	0.00515	0.00125	0.00125	0.00125	0.00128	0.0012	0.00113	0.00106	0.00103	0.00117	0.0017	0.00298	0.00775								
2.1	Length	Cl	0.007 mm	NMI	1.03E-08	2.6E-09	1.4E-09	7.1E-10	2.4E-12	2.4E-12	2.4E-12	2.7E-12	2.1E-12	1.6E-12	1.3E-12	1.1E-12	1.8E-12	8.4E-12	7.9E-11	3.6E-09			13				
2.2	Diameter	Cd	0.004 mm	NMI	0.0010	0.0010	0.0010	0.0010	0.0010	0.0010	0.0010	0.0009	0.0008	0.0007	0.0005	0.0006	0.0011	0.0023	0.0067				30	Normal			
2.3	Volume	Cv	1.E-08 mm^3	NMI	0.0007	0.0007	0.0007	0.0007	0.0007	0.0007	0.0007	0.0008	0.0008	0.0008	0.0009	0.0010	0.0013	0.0019	0.0039				30	Normal			
2.4	Surface Area	CA	1.E-08 mm^2	NMI	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0				30	Normal			
2.5	Air Leakage			NMI	0.010	0.007	0.008	0.005	0.000	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0				20	12.5			
<b>Microphone parameters</b>				0.014	0.01423	0.01419	0.01416	0.01415	0.01415	0.01414	0.01415	0.01349	0.0123	0.01137	0.01069	0.01018	0.01128	0.02316	0.06613								
3.1	Front depth	Fd	0.01 mm	NMI	4.2E-08	4.1E-08	4E-08	4E-08	4E-08	4E-08	4E-08	3.3E-08	2.3E-08	1.7E-08	1.3E-08	1.1E-08	1.6E-08	2.9E-07	1.9E-05				63				
3.2	Front volume	Fv	0.5 mm^3	NMI	0.0022	0.0018	0.0011	0.0008	0.0008	0.0004	0.0002	0.0001	0.0002	0.0004	0.0008	0.0012	0.0018	0.0025	0.0042	0.0198			30	Normal			
3.3	Equiv volume	Ve	0.5 mm^3	NMI	0.0100	0.0100	0.0100	0.0100	0.0100	0.0100	0.0100	0.0100	0.0100	0.0100	0.0100	0.0100	0.0100	0.0100	0.0100	0.0100			30	Normal			
3.4	Resonance freq	Fr	200 Hz	NMI	0.0100	0.0100	0.0100	0.0100	0.0100	0.0100	0.0100	0.0090	0.0070	0.0050	0.0030	0.0020	0.0004	0.0130	0.0580			30	Normal				
3.5	Loss factor	Loss	0.02 ratio	NMI	0.0	0.0	0.0	0.0	0.0	0.0	0.0001	0.0001	0.0001	0.0001	0.0000	0.0005	0.0022	0.0050	0.0148			30	Normal				
				NMI	0.0	0.0	0.0	0.0	0.0	0.0	0.0001	0.0004	0.0010	0.0015	0.0019	0.0019	0.0001	0.0040	0.0150	0.0230			30	Rect			
<b>Ambient Conditions</b>				0.011	0.0022	0.00219	0.00219	0.00214	0.00213	0.00219	0.00219	0.00226	0.00238	0.00245	0.00259	0.00259	0.00613	0.00305	0.00733								
4.1	Static pressure	Pamb	0.03 kPa	NMI	1.66E-08	2.3E-11	2.3E-11	2.3E-11	2.1E-11	2.1E-11	2.3E-11	2.3E-11	2.6E-11	3.2E-11	3.6E-11	4.5E-11	4.5E-11	1.4E-09	8.7E-11	2.9E-09			59.8				
4.2	Drift in pressure	Pdrift	0.01 kPa	NMI	0.0013	0.0013	0.0013	0.0013	0.0012	0.0012	0.0013	0.0013	0.0014	0.0015	0.0016	0.0018	0.0014	0.0014	0.0010	0.0010			30	Triangular			
4.3	Pcoeff unknown	Pcoeff	dB/kPa	NMI	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0			30	Rect				
4.4	Mic temperature	Tmic	0.3 °C	NMI	0.0088	0.0001	0.0001	0.0001	0.0001	0.0001	0.0001	0.0001	0.0001	0.0001	0.0001	0.0001	0.0001	0.0001	0.0001	0.0001			30	Rect			
4.5	Tcoeff unknown	Tcoeff	dB/°C	NMI	0.0004	0.0004	0.0004	0.0004	0.0004	0.0004	0.0004	0.0004	0.0004	0.0004	0.0005	0.0005	0.0005	0.0003	0.0001	0.0001			30	Normal			
4.6	Relative Humidity	%RH	3 %	NMI	0.0072	0.0017	0.0017	0.0017	0.0017	0.0017	0.0017	0.0017	0.0017	0.0017	0.0017	0.0017	0.0017	0.0051	0.0008	0.0019			30	Rect			
				NMI	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0005	0.0005	0.0005	0.0010	0.0020	0.0070			30	Normal				
<b>Physical Corrections</b>				0.003	0.003	0.003	0.003	0.003	0.003	0.003	0.004	0.005	0.007	0.011	0.016	0.023	0.031	0.043	0.062								
5.1	Radial wave motion Corr.		dB	KR	1.34E-10	1.3E-10	1.3E-10	1.3E-10	1.3E-10	1.3E-10	1.3E-10	3.4E-10	8E-10	3.1E-09	1.3E-08	6.8E-08	2.8E-07	9.1E-07	3.4E-06	1.4E-05			19.1				
5.2	Viscosity of air		dB	KR	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0010	0.0020	0.0030	0.0050	0.0100	0.0200	0.0300	20	12.5	Normal					
5.3	Air properties		dB	KR	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0002	0.0005	0.0010	0.0030	0.0050	0.0100	0.0150	0.0200	20	12.5	Normal				
5.4	Correction to ref conditions		dB	KR	0.0018	0.0016	0.0016	0.0016	0.0016	0.0016	0.0016	0.0016	0.0016	0.0015	0.0015	0.0014	0.0014	0.0013	0.0013	0.0012	20	12.5	Normal				
				KR	0.0030	0.0030	0.0030	0.0030	0.0030	0.0030	0.0030	0.0040	0.0050	0.0070	0.0100	0.0150	0.0200	0.0250	0.0350	0.0500	20	12.5	Normal				
<b>Rounding and Type A</b>				0.030	0.025	0.020	0.015	0.012	0.010	0.010	0.010	0.010	0.010	0.015	0.020	0.030	0.040	0.050	0.064								
6.1	Rounding		0.0005	NMI	8.1E-07	4E-07	1.7E-07	5.4E-08	2.3E-08	1.2E-08	1.2E-08	1.2E-08	1.2E-08	1.2E-08	5.4E-08	1.7E-07	8.3E-07	2.6E-06	6.3E-06	1.7E-05			12.5				
6.2	Type A		dB	Data	0.0003	0.0029	0.0029	0.0029	0.0029	0.0029	0.0029	0.0029	0.0029	0.0029	0.0029	0.0029	0.0029	0.0029	0.0029	0.0029			30	Rect			
				Data	0.0300	0.0250	0.0200	0.0150	0.0120	0.0100	0.0100	0.0100	0.0100	0.0100	0.0100	0.0150	0.0200	0.0300	0.0400	0.0500	0.0640	20	12.5	Normal			
<b>Combined Uncertainty, uc</b>				0.037	0.030	0.026	0.022	0.020	0.019	0.019	0.019	0.019	0.018	0.022	0.028	0.040	0.052	0.070	0.111								
<b>Effective dof</b>				28.0	25.7	32.9	48.4	60.2	74.7	74.8	77.4	78.0	78.8	47.3	38.3	30.3	29.6	35.4	63.9								
<b>k factor</b>				2.05	2.06	2.04	2.01	2.00	1.99	1.99	1.99	1.99	1.99	2.01	2.03	2.04	2.05	2.03	2.00								
<b>Expanded Uncertainty, U95 for k=2</b>				0.074	0.061	0.052	0.045	0.039	0.037	0.037	0.037	0.037	0.037	0.045	0.057	0.079	0.105	0.140	0.223								
<b>U95 rounded to 2 dP</b>				0.07	0.06	0.05	0.04	0.04	0.04	0.04	0.04	0.04	0.04	0.04	0.04	0.06	0.08	0.10	0.14	0.22							
<b>Stated Uncertainty</b>				0.08	0.06	0.05	0.05	0.04	0.04	0.04	0.04	0.04	0.04	0.04	0.05	0.06	0.08	0.1	0.15	0.25							

Uncertainty of 4180 Microphones																	
S.No.	Sources of Uncertainty	Probability Distribution Type - A or B	Sensitivity Coefficient	Uncertainty Contribution													
				20 Hz	31.5 Hz	63 Hz	125 Hz	250 Hz	1KHz	4 KHz	8 KHz	10 KHz	12.5 kHz	16 kHz	20 kHz	25 kHz	
1	Static Pressure	Type-B, Rectangular	1	0.0016	0.0016	0.0016	0.0016	0.0016	0.0016	0.0016	0.0016	0.0016	0.0017	0.0017	0.0017	0.0017	
2	Temperature	Type-B, Rectangular	1	0.0004	0.0002	0.0001	0.0001	0.0001	0.0001	0.0001	0.0002	0.0013	0.0022	0.004	0.0072	0.0112	0.024
3	Relative Humidity	Type-B, Rectangular	1	0.0003	0.0003	0.0003	0.0003	0.0003	0.0003	0.0003	0.0003	0.0004	0.0006	0.0008	0.0012	0.002	0.004
4	Coupler Volume	Type-B, Normal	1	0.02	0.02	0.02	0.02	0.02	0.02	0.02	0.02	0.02	0.02	0.02	0.02	0.02	0.02
5	Equivalent & Front Volume	Type-B, Normal	1	0.0067	0.0068	0.0069	0.007	0.007	0.007	0.007	0.0068	0.0058	0.0049	0.0035	0.0014	0.0013	0.0075
6	Microphone Front Length	Type-B, Normal	1	0.0002	0.0002	0.0001	0.0001	0.0001	0.0001	0.0001	0.0004	0.0016	0.0026	0.0045	0.0075	0.012	0.028
7	Series capacitance	Type-B, Rectangular	1	0.0065	0.0048	0.0028	0.0024	0.0018	0.0018	0.0024	0.0027	0.0028	0.0032	0.0035	0.0038	0.005	
8	Voltage Ratio (DVM)	Type-B, Rectangular	1	0.015	0.015	0.01	0.01	0.01	0.01	0.01	0.01	0.01	0.01	0.015	0.015	0.015	
9	Voltage Ratio(Cr.talk)	Type-B, Rectangular	1	0.0025	0.0025	0.0025	0.0025	0.0025	0.0025	0.0025	0.0025	0.0025	0.0025	0.0025	0.0025	0.0025	0.0025
10	Voltage Ratio (Noise)	Type-B, Rectangular	1	0.0001	0.0001	0.0001	0.0001	0.0001	0.0001	0.0001	0.0001	0.0001	0.0001	0.0001	0.0001	0.0001	0.0001
11	Voltage Ratio(Distortion)	Type-B, Rectangular	1	0.0001	0.0001	0.0001	0.0001	0.0001	0.0001	0.0001	0.0001	0.0001	0.0001	0.0001	0.0001	0.0001	0.0001
12	Polarization Voltage	Type-B, Normal	1	0.0036	0.0036	0.0036	0.0036	0.0036	0.0036	0.0036	0.0036	0.0036	0.0036	0.0036	0.0036	0.0036	0.0036
13	Electrical Parameters incl. Transfer impedance,acc.of voltmeter	Type-B, Rectangular	1	0.05	0.05	0.05	0.02	0.02	0.02	0.03	0.04	0.05	0.05	0.05	0.05	0.05	0.05
14	Specific Heat Ratio	Type-B, Normal	1	0.002	0.002	0.002	0.002	0.002	0.002	0.002	0.002	0.002	0.002	0.002	0.002	0.002	0.002
15	Wave motion correction	Type-B, Rectangular	1	0.001	0.001	0.001	0.001	0.001	0.001	0.001	0.001	0.001	0.001	0.001	0.001	0.001	0.001
16	Heat Conduction correction	Type-B, Rectangular	1	0.01	0.01	0.01	0.01	0.01	0.01	0.01	0.01	0.01	0.01	0.01	0.01	0.01	0.01
17	Reproducibility	Type-A, Normal		0.005	0.005	0.01	0.01	0.01	0.01	0.01	0.01	0.01	0.005	0.005	0.005	0.015	0.045
<b>Uncertainty of Sensitivity at Meas. conditions</b>																	
	Sensitivity Correction for Static pressure	Type-B, Rectangular	1	0.0005	0.0005	0.0005	0.0005	0.0005	0.0005	0.0005	0.0008	0.0022	0.0024	0.0014	0.008	0.0175	0.014
	Sensitivity Correction for Temperature	Type-B, Rectangular	1	0.0027	0.0027	0.0027	0.0027	0.0027	0.0027	0.0027	0.0027	0.003	0.0033	0.0032	0.0031	0.0065	0.0068
<b>Uncertainty of Sensitivity at ref. Cond.</b>				0.058067	0.057912	0.054	0.04	0.025	0.025	0.039	0.05	0.054	0.053	0.058945	0.06415	0.083526	
<b>Expanded Uncertainty dB</b>				<b>0.12</b>	<b>0.12</b>	<b>0.11</b>	<b>0.08</b>	<b>0.05</b>	<b>0.05</b>	<b>0.08</b>	<b>0.1</b>	<b>0.11</b>	<b>0.11</b>	<b>0.12</b>	<b>0.13</b>	<b>0.17</b>	<b>dB</b>

## Uncertainty Budget for LS2P Microphones

Uncertainty Components	Standard uncertainty in 0.001 dB													
	31.5 Hz	63.0 Hz	125.0 Hz	250.0 Hz	500.0 Hz	1.0 kHz	2.0 kHz	4.0 kHz	8.0 kHz	10.0 kHz	12.5 kHz	16.0 kHz	20.0 kHz	25.0 kHz
<b>Electrical Measurements</b>														
1. Voltage ratios	4	4	4	4	4	4	4	4	4	4	4	4	4	4
2. Polarizing voltage	2	2	2	2	2	2	2	2	2	2	2	2	2	2
3. Frequency	1	1	1	1	1	1	1	1	1	1	1	1	1	1
4. Capacitance	2	2	2	2	2	2	2	2	2	2	2	2	2	2
<b>Coupler Dimensions</b>														
5. Coupler length	5	5	5	5	5	5	5	5	5	5	5	10	15	50
6. Coupler diameter	2	2	2	2	2	2	2	2	2	2	2	5	10	20
7. Coupler surface area	1	1	1	1	1	1	1	1	1	1	1	2	10	20
<b>Microphone Parameters</b>														
8. Font cavity depth	2	2	2	2	2	2	2	2	2	2	2	2	2	5
9. Front cavity volume	5	5	5	5	5	5	5	5	5	5	5	5	5	10
10. Equivalent volume	5	5	5	5	5	5	5	5	5	5	5	5	5	10
11. Resonant frequency	3	3	3	3	3	3	3	3	3	3	3	10	10	15
12. Loss factor	3	3	3	3	3	3	3	3	3	3	3	10	10	15
<b>Ambient Conditions</b>														
13. Static pressure	3	3	3	3	3	3	3	3	3	3	3	3	3	3
14. Temperature	4	4	4	4	4	4	4	4	4	4	4	4	4	4
15. Pressure coeff.	10	2	2	2	2	2	2	2	2	2	2	5	10	10
16. Temperature coeff.	10	2	2	2	2	2	2	2	2	2	2	5	10	10
<b>Others</b>														
17. Theory imperfection	25	5	2	2	2	2	2	2	2	2	2	5	10	10
18. Repeatability	20	10	10	10	10	10	10	10	10	10	10	10	20	20
Combined standard uncert.	37.1	16.9	16.2	16.2	16.2	16.2	16.2	16.2	16.2	16.2	16.2	24.7	37.8	68.7
Coverage factor	2	2	2	2	2	2	2	2	2	2	2	2	2	2
Expanded combined uncert.	74.2	33.8	32.5	32.5	32.5	32.5	32.5	32.5	32.5	32.5	32.5	49.3	75.6	137.5
Stated Uncertainty	80	40	40	40	40	40	40	40	40	40	40	50	80	140

Uncertainty for LS2P Microphones - NML, SIRIM Berhad, MALAYSIA for Regional Key Comparison APMP.AUV.A-K3

No	Input Parameter	Std U	Unit	31.5	63	125	250	500	1000	2000	4000	6300	8000	10000	12500	16000	20000	25000
2.1.1	Static Pressure	0.039	kPa	0.0007	0.0005	0.0004	0.0003	0.0002	0.0002	0.0003	0.0008	0.0020	0.0033	0.0053	0.0086	0.0151	0.0268	0.0594
2.1.2	Temperature	3.057	K	0.0015	0.0009	0.0007	0.0005	0.0004	0.0003	0.0005	0.0016	0.0037	0.0060	0.0098	0.0157	0.0277	0.0492	0.1093
2.1.3	Relative Humidity	6.066	%	0.0003	0.0003	0.0003	0.0003	0.0003	0.0003	0.0004	0.0004	0.0005	0.0007	0.0008	0.0012	0.0018	0.0029	0.0060
	<b>Amb. Conditions</b>			<b>0.0017</b>	<b>0.0011</b>	<b>0.0009</b>	<b>0.0008</b>	<b>0.0005</b>	<b>0.0005</b>	<b>0.0007</b>	<b>0.0018</b>	<b>0.0043</b>	<b>0.0069</b>	<b>0.0110</b>	<b>0.0179</b>	<b>0.0318</b>	<b>0.0561</b>	<b>0.1248</b>
2.2.1	Coupler Length	0.00101	mm	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.0001	0.0001	0.0002	0.0003	0.0006	0.0010	0.0017	0.0038
2.2.2	Coupler Diameter	0.00473	mm	0.0001	0.0001	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.0001
2.2.3	Coupler Volume	0.35063	mm <sup>3</sup>	0.0038	0.0038	0.0038	0.0038	0.0038	0.0038	0.0038	0.0038	0.0038	0.0038	0.0038	0.0039	0.0039	0.0039	0.0038
2.2.4	Coup. Surf. Area	0.2796	mm <sup>2</sup>	0.0003	0.0003	0.0003	0.0003	0.0003	0.0003	0.0003	0.0003	0.0003	0.0003	0.0003	0.0003	0.0003	0.0003	0.0003
	<b>Coupler</b>			<b>0.0038</b>	<b>0.0038</b>	<b>0.0038</b>	<b>0.0038</b>	<b>0.0038</b>	<b>0.0038</b>	<b>0.0038</b>	<b>0.0038</b>	<b>0.0038</b>	<b>0.0038</b>	<b>0.0038</b>	<b>0.0039</b>	<b>0.0040</b>	<b>0.0043</b>	<b>0.0052</b>
2.3.1	Mic. Front Length	0.01	mm	0.0004	0.0003	0.0002	0.0001	0.0001	0.0001	0.0003	0.0010	0.0025	0.0040	0.0065	0.0106	0.0187	0.0331	0.0734
2.3.2	Front Vol	0.6	mm <sup>3</sup>	0.0129	0.0129	0.0129	0.0129	0.0129	0.0129	0.0129	0.0129	0.0130	0.0130	0.0131	0.0133	0.0135	0.0135	0.0125
2.3.3	Eqv. Vol	1	mm <sup>3</sup>	0.0202	0.0206	0.0206	0.0210	0.0211	0.0211	0.0210	0.0200	0.0180	0.0157	0.0122	0.0087	0.0010	0.0011	0.0292
2.3.4	Diaphragm Res.	1300	Hz	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.0001	0.0005	0.0010	0.0019	0.0025	0.0040	0.0271	0.0490
2.3.5	Diaph. Damp. Fac.	0.1	-	0.0000	0.0000	0.0000	0.0000	0.0000	0.0001	0.0003	0.0013	0.0030	0.0044	0.0054	0.0048	0.0027	0.0058	0.0425
	<b>Microphones</b>			<b>0.0240</b>	<b>0.0243</b>	<b>0.0245</b>	<b>0.0248</b>	<b>0.0247</b>	<b>0.0248</b>	<b>0.0246</b>	<b>0.0239</b>	<b>0.0225</b>	<b>0.0213</b>	<b>0.0199</b>	<b>0.0190</b>	<b>0.0236</b>	<b>0.0453</b>	<b>0.1030</b>
2.4.1	Series Cap.	see table	pF	0.0014	0.0009	0.0009	0.0009	0.0009	0.0009	0.0009	0.0009	0.0009	0.0009	0.0009	0.0009	0.0009	0.0010	0.0010
2.4.2	Volt. Ratio, DVM	see table	-	0.0094	0.0094	0.0093	0.0093	0.0093	0.0093	0.0094	0.0094	0.0094	0.0094	0.0105	0.0105	0.0105	0.0105	0.0105
2.4.3	Volt. Ratio, Cr-talk	< -86 dB	-	0.0025	0.0025	0.0025	0.0025	0.0025	0.0025	0.0025	0.0025	0.0025	0.0025	0.0025	0.0025	0.0025	0.0025	0.0025
2.4.4	Volt. Ratio, Noise	< -46 dB	-	0.0001	0.0001	0.0001	0.0001	0.0001	0.0001	0.0001	0.0001	0.0001	0.0001	0.0001	0.0001	0.0001	0.0001	0.0001
2.4.5	Volt. Ratio, Distort.	< -46 dB	-	0.0001	0.0001	0.0001	0.0001	0.0001	0.0001	0.0001	0.0001	0.0001	0.0001	0.0001	0.0001	0.0001	0.0001	0.0001
2.4.6	Frequency	12	ppm	0.0001	0.0001	0.0001	0.0001	0.0001	0.0001	0.0001	0.0001	0.0002	0.0003	0.0004	0.0004	0.0004	0.0004	0.0004
2.4.7	Pol. Voltage	0.023	V	0.0014	0.0014	0.0014	0.0014	0.0014	0.0014	0.0014	0.0014	0.0014	0.0014	0.0014	0.0014	0.0014	0.0014	0.0014
	<b>Electrical Parameters</b>			<b>0.0099</b>	<b>0.0098</b>	<b>0.0098</b>	<b>0.0098</b>	<b>0.0098</b>	<b>0.0098</b>	<b>0.0098</b>	<b>0.0098</b>	<b>0.0098</b>	<b>0.0098</b>	<b>0.0109</b>	<b>0.0109</b>	<b>0.0109</b>	<b>0.0109</b>	<b>0.0109</b>
2.5	Repeatability		dB	0.0097	0.0100	0.0101	0.0101	0.0099	0.0097	0.0092	0.0078	0.0055	0.0062	0.0103	0.0156	0.0155	0.0223	0.0162
2.6	Result Rounding	< 0.005	dB	0.0058	0.0058	0.0058	0.0058	0.0058	0.0058	0.0058	0.0058	0.0058	0.0058	0.0058	0.0058	0.0058	0.0058	0.0058
	<b>Uncertainty at Measurement Conditions</b>		<b>dB</b>	<b>0.0286</b>	<b>0.0289</b>	<b>0.0291</b>	<b>0.0292</b>	<b>0.0292</b>	<b>0.0292</b>	<b>0.0289</b>	<b>0.0279</b>	<b>0.0264</b>	<b>0.0261</b>	<b>0.0281</b>	<b>0.0331</b>	<b>0.0443</b>	<b>0.0766</b>	<b>0.1630</b>
2.7	Sensitivity Correction for Static Pressure		dB	0.0011	0.0011	0.0011	0.0012	0.0013	0.0013	0.0012	0.0005	0.0010	0.0028	0.0059	0.0086	0.0033	0.0187	0.0150
2.7	Sensitivity Correction for Temperature		dB	0.0010	0.0010	0.0010	0.0010	0.0010	0.0011	0.0018	0.0035	0.0066	0.0101	0.0166	0.0250	0.0289	0.0269	0.0281
	<b>Sens Corr to Reference Environmental Cond.</b>		<b>dB</b>	<b>0.0015</b>	<b>0.0015</b>	<b>0.0015</b>	<b>0.0018</b>	<b>0.0016</b>	<b>0.0017</b>	<b>0.0022</b>	<b>0.0035</b>	<b>0.0067</b>	<b>0.0105</b>	<b>0.0176</b>	<b>0.0264</b>	<b>0.0291</b>	<b>0.0328</b>	<b>0.0319</b>
	<b>Expanded Uncertainty (k=2)</b>		<b>dB</b>	<b>0.0286</b>	<b>0.0289</b>	<b>0.0291</b>	<b>0.0292</b>	<b>0.0293</b>	<b>0.0292</b>	<b>0.0290</b>	<b>0.0281</b>	<b>0.0273</b>	<b>0.0282</b>	<b>0.0332</b>	<b>0.0423</b>	<b>0.0530</b>	<b>0.0833</b>	<b>0.1661</b>
	<b>Reported Uncertainty</b>		<b>dB</b>	<b>0.04</b>	<b>0.04</b>	<b>0.04</b>	<b>0.04</b>	<b>0.04</b>	<b>0.04</b>	<b>0.04</b>	<b>0.04</b>	<b>0.04</b>	<b>0.04</b>	<b>0.05</b>	<b>0.06</b>	<b>0.07</b>	<b>0.10</b>	<b>0.17</b>



## Uncertainty Budget for APMP.AUV.A-K3 (NMIJ/AIST)

Uncertainty components		sensitivity coefficient	31.5	63	125	250	500	1000	2000	4000	6300	8000	10000	12500	16000	20000	25000	31500
static pressure	A normal	0.5	0.001	0.001	0.001	0.001	0.001	0.001	0.001	0.001	0.001	0.001	0.001	0.001	0.001	0.001	0.001	0.001
specific heat ratio	A normal	0.5	0.002	0.002	0.002	0.002	0.002	0.002	0.002	0.002	0.002	0.002	0.002	0.002	0.002	0.002	0.002	0.002
coupler volume (including front cavity volumes)	A normal	0.5	0.016	0.016	0.016	0.016	0.016	0.016	0.016	0.016	0.016	0.016	0.016	0.016	0.016	0.016	0.016	0.016
equivalent volumes	A normal	0.5	0.016	0.016	0.016	0.016	0.016	0.016	0.016	0.016	0.016	0.016	0.016	0.016	0.016	0.016	0.016	0.016
voltage transfer function between input terminal of transmitter and output terminal of receiver	combined	0.866	0.010	0.010	0.010	0.007	0.007	0.007	0.007	0.006	0.006	0.006	0.005	0.004	0.004	0.060	0.060	0.060
cross-talk	B rectangular		0.001	0.001	0.001	0.001	0.001	0.001	0.001	0.001	0.001	0.001	0.001	0.001	0.001	0.001	0.001	0.001
distortion (linearity of FFT analyzer)	B rectangular		0.001	0.001	0.001	0.001	0.001	0.001	0.001	0.001	0.001	0.001	0.001	0.001	0.001	0.001	0.001	0.001
attenuator	B rectangular		0.001	0.001	0.001	0.001	0.001	0.001	0.001	0.001	0.001	0.001	0.001	0.001	0.001	0.001	0.001	0.001
transmitter and receiver ground shield	A normal		0.003	0.003	0.003	0.003	0.003	0.003	0.003	0.003	0.003	0.003	0.003	0.003	0.003	0.003	0.003	0.003
repeatability	A normal		0.009	0.009	0.009	0.006	0.006	0.006	0.006	0.005	0.005	0.005	0.004	0.003	0.003	0.060	0.060	0.060
leakage of sound	A normal		0.000	0.000	0.000	0.000	0.000	0.000	0.000	0.000	0.000	0.000	0.000	0.000	0.000	0.000	0.000	0.000
electrical impedance of transmitter	combined	0.5	0.002	0.002	0.002	0.002	0.002	0.002	0.002	0.002	0.002	0.002	0.002	0.002	0.002	0.002	0.002	0.002
distortion (linearity of FFT analyzer)	B rectangular		0.001	0.001	0.001	0.001	0.001	0.001	0.001	0.001	0.001	0.001	0.001	0.001	0.001	0.001	0.001	0.001
repeatability	A normal		0.002	0.002	0.002	0.002	0.002	0.002	0.002	0.002	0.002	0.002	0.002	0.002	0.002	0.002	0.002	0.002
polarizing voltage	A normal	1	0.001	0.001	0.001	0.001	0.001	0.001	0.001	0.001	0.001	0.001	0.001	0.001	0.001	0.001	0.001	0.001
rounding error	B rectangular	1	0.001	0.001	0.001	0.001	0.001	0.001	0.001	0.001	0.001	0.001	0.001	0.001	0.001	0.001	0.001	0.001
static pressure coefficient of pressure sensitivity	A normal	1	0.000	0.000	0.000	0.000	0.000	0.000	0.000	0.000	0.000	0.000	0.000	0.000	0.002	0.003	0.003	0.001
temperature coefficient of pressure sensitivity	A normal	1	0.000	0.000	0.000	0.000	0.000	0.000	0.000	0.000	0.001	0.001	0.001	0.002	0.002	0.001	0.002	0.000
coupler correction(heat conduction, capillary tube and wave motion)	combined	1	0.008	0.006	0.005	0.006	0.006	0.006	0.006	0.006	0.006	0.006	0.007	0.010	0.017	0.031	0.068	0.324
static pressure, temperature and relative humidity	A normal		0.000	0.000	0.000	0.000	0.000	0.000	0.000	0.000	0.000	0.000	0.000	0.001	0.001	0.003	0.006	0.026
coupler length and diameter	A normal		0.006	0.002	0.001	0.001	0.001	0.001	0.001	0.002	0.003	0.004	0.006	0.010	0.017	0.030	0.067	0.303
capillary tube length and diameter	A normal		0.000	0.000	0.000	0.000	0.000	0.000	0.000	0.000	0.000	0.000	0.000	0.000	0.000	0.000	0.000	0.000
resonance frequency, quality factor and tension of microphone diaphragm	A normal		0.005	0.005	0.005	0.006	0.006	0.006	0.006	0.005	0.005	0.004	0.004	0.002	0.001	0.008	0.012	0.111
Combined standard uncertainty of pressure sensitivity			0.016	0.015	0.015	0.014	0.014	0.014	0.014	0.014	0.014	0.014	0.014	0.016	0.021	0.062	0.087	0.328
Expanded uncertainty of pressure sensitivity (coverage factor $k=2$ )			0.04	0.04	0.04	0.03	0.03	0.03	0.03	0.03	0.03	0.03	0.03	0.04	0.05	0.13	0.18	0.66
Stated uncertainty			0.06	0.06	0.06	0.05	0.05	0.05	0.05	0.05	0.05	0.05	0.05	0.06	0.07	0.13	0.18	0.66

Table: uncertainty budget of LS2aP

No.	resource	symbol	20Hz	31.5Hz	50Hz	63Hz	125Hz	250Hz	1kHz	2kHz	4kHz	8kHz	10kHz	12.5kHz	16kHz	20kHz	25kHz
1	Sensitive calibration	$v_1=S_1$	0.0118	0.0101	0.0100	0.0077	0.0093	0.0086	0.0085	0.0085	0.0089	0.0113	0.0133	0.0159	0.0194	0.0219	0.0271
2	Polarization volt.	$v_2$	0.00004	0.00004	0.00004	0.00004	0.00004	0.00004	0.00004	0.00004	0.00004	0.00004	0.00004	0.00004	0.00004	0.00004	0.00004
3	Volt. Ratio	$v_3$	0.0009	0.0009	0.0007	0.0007	0.0004	0.0004	0.0004	0.0004	0.0007	0.0007	0.0007	0.0024	0.0024	0.0024	0.0024
4	Cross talk	$v_3$	0.0025	0.0025	0.0025	0.0025	0.0025	0.0025	0.0025	0.0025	0.0025	0.0025	0.0025	0.0025	0.0025	0.0025	0.0025
5	Noise	$v_5$	0.0001	0.0001	0.0001	0.0001	0.0001	0.0001	0.0001	0.0001	0.0001	0.0001	0.0001	0.0001	0.0001	0.0001	0.0001
6	Distortion	$v_6$	0.0001	0.0001	0.0001	0.0001	0.0001	0.0001	0.0001	0.0001	0.0001	0.0001	0.0001	0.0001	0.0001	0.0001	0.0001
7	Frequency	$v_7$	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000
8	Length of coupler	$v_8$	0.0001	0.0001	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.0002	0.0006	0.0010	0.0016	0.0026	0.0046	0.0100
9	Diameter of coupler	$v_9$	0.0001	0.0001	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000
10	Surface area of coupler	$v_{10}$	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000
11	Volume of coupler	$v_{11}$	0.0037	0.0037	0.0037	0.0037	0.0037	0.0037	0.0037	0.0037	0.0037	0.0037	0.0037	0.0037	0.0037	0.0037	0.0037
12	Leak of coupler	$v_{12}$	0.0217	0.0087	0.0050	0.0022	0.0011	0.0003	0.0001	0.0001	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000
13	Length of front cavity of microphone	$v_{13}$	0.0002	0.0002	0.0002	0.0001	0.0001	0.0000	0.0000	0.0000	0.0004	0.0016	0.0026	0.0045	0.0075	0.0100	0.0140
14	Affix front cavity	$v_{14}$	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--

## NMC uncertainty budget for L92P microphone - Key Comparison A PMP.AU.V.A-K3

	32 Hz	63 Hz	125 Hz	250 Hz	500 Hz	1 kHz	4 kHz	8 kHz	10 kHz	12.5 kHz	16 kHz	20 kHz	25 kHz
<b>Ambient</b>													
Static Pressure Correction	0.0010	0.0010	0.0010	0.0010	0.0010	0.0008	0.0005	0.0018	0.0027	0.0011	0.0081	0.0184	
Temperature Correction	0.0015	0.0015	0.0015	0.0015	0.0015	0.0026	0.0062	0.0087	0.0124	0.0160	0.0108	0.0104	
	0.0018	0.0018	0.0018	0.0018	0.0018	0.0027	0.0062	0.0089	0.0127	0.0160	0.0135	0.0211	
<b>Plane Wave Couplers</b>													
Length	0.0001	0.0000	0.0000	0.0000	0.0000	0.0002	0.0006	0.0010	0.0016	0.0026	0.0046	0.0100	
Diameter	0.0001	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	
Volume	0.0037	0.0037	0.0037	0.0037	0.0037	0.0037	0.0037	0.0037	0.0037	0.0038	0.0038	0.0037	
Surface Area	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	
Coupler Leakage	0.0173	0.0044	0.0011	0.0003	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	
	0.0177	0.0057	0.0039	0.0037	0.0037	0.0037	0.0037	0.0038	0.0040	0.0046	0.0060	0.0107	
<b>1/2" Microphone Type 4180</b>													
Mic Front Length	0.0003	0.0002	0.0002	0.0000	0.0000	0.0006	0.0024	0.0039	0.0068	0.0113	0.0180	0.0420	
Eq. S Front Vol.	0.0068	0.0069	0.0070	0.0070	0.0070	0.0068	0.0058	0.0049	0.0035	0.0014	0.0013	0.0075	
Diaphragm Res.	0.0000	0.0000	0.0000	0.0000	0.0000	0.0002	0.0010	0.0022	0.0026	0.0060	0.0280	0.0580	
Diaph. Damp. Fac.	0.0000	0.0000	0.0000	0.0000	0.0001	0.0001	0.0040	0.0052	0.0044	0.0040	0.0070	0.0300	
	0.0068	0.0069	0.0070	0.0070	0.0070	0.0068	0.0075	0.0084	0.0092	0.0135	0.0340	0.0780	
<b>Series Coupling</b>													
Voltage Ratio	0.0015	0.0012	0.0010	0.0010	0.0010	0.0010	0.0010	0.0010	0.0010	0.0010	0.0010	0.0010	
DVM	0.0015	0.0012	0.0008	0.0008	0.0008	0.0012	0.0012	0.0012	0.0042	0.0042	0.0042	0.0042	
Cross-talk	0.0025	0.0025	0.0025	0.0025	0.0025	0.0025	0.0025	0.0025	0.0025	0.0025	0.0025	0.0025	
Noise	0.0001	0.0001	0.0001	0.0001	0.0001	0.0001	0.0001	0.0001	0.0001	0.0001	0.0001	0.0001	
Distortion	0.0001	0.0001	0.0001	0.0001	0.0001	0.0001	0.0001	0.0001	0.0001	0.0001	0.0001	0.0001	
Frequency	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.0001	0.0003	
Polarization Voltage	0.0036	0.0036	0.0036	0.0036	0.0036	0.0036	0.0036	0.0036	0.0036	0.0036	0.0036	0.0036	
	0.0046	0.0045	0.0044	0.0044	0.0044	0.0045	0.0045	0.0045	0.0061	0.0061	0.0061	0.0061	
Reproducibility	0.0080	0.0080	0.0080	0.0080	0.0080	0.0080	0.0080	0.0080	0.0080	0.0147	0.0200	0.0347	
Result Rounding	0.00058	0.00058	0.00058	0.00058	0.00058	0.00058	0.00058	0.00058	0.00058	0.00058	0.00058	0.00058	
Combine (Type B)	0.021	0.013	0.012	0.012	0.012	0.012	0.014	0.016	0.019	0.027	0.043	0.089	
RefMic uncertainty 95%	0.05	0.03	0.03	0.03	0.03	0.03	0.03	0.03	0.04	0.05	0.07	0.12	
Combine with Type B (1s)	0.03	0.02	0.02	0.02	0.02	0.02	0.02	0.02	0.02	0.03	0.04	0.07	
Type A (STD EV 200ppm)	0.01225	0.01225	0.01225	0.01225	0.01225	0.01225	0.01225	0.01225	0.01225	0.01225	0.01225	0.01225	
Combine Type A + B (1s)	0.02960	0.02043	0.02032	0.02032	0.02032	0.02033	0.02059	0.02092	0.02532	0.03088	0.04276	0.07564	
DOF (n=55)	1.59E+03	4.26E+02	4.17E+02	4.17E+02	4.16E+02	4.18E+02	4.39E+02	4.69E+02	1.00E+03	2.22E+03	8.17E+03	8.00E+04	
k at 95%	2.01	2.02	2.02	2.02	2.02	2.02	2.02	2.01	2.01	2.01	2.01	2.01	
Expanded Uncertainty	0.06	0.04	0.04	0.04	0.04	0.04	0.04	0.04	0.06	0.06	0.09	0.16	

KRISS: Uncertainty Budget for LS2P

MEASURED QUANTITY	Symbol	Unc.	Unit	31,5	63	126	250	500	1000	2000	3150	4000	5000	6300	8000	10000	12600	16000	20000	25000	31500	
<b>Electrical Transfer Impedance</b>				<b>0,0112</b>	<b>0,0109</b>	<b>0,0107</b>	<b>0,0067</b>	<b>0,0054</b>	<b>0,0060</b>	<b>0,0055</b>	<b>0,0067</b>	<b>0,0067</b>	<b>0,0067</b>	<b>0,0067</b>	<b>0,0067</b>	<b>0,0083</b>	<b>0,0117</b>	<b>0,0117</b>	<b>0,0117</b>	<b>0,0117</b>	<b>0,0159</b>	
Series Capacitance	C	Table	nF	0,0102	0,0098	0,0097	0,0048	0,0029	0,0019	0,0029	0,0048	0,0048	0,0048	0,0048	0,0048	0,0048	0,0096	0,0096	0,0096	0,0096	0,0096	
Voltage Ratio	VR	Table		0,0031	0,0029	0,0028	0,0028	0,0028	0,0028	0,0028	0,0029	0,0029	0,0029	0,0029	0,0029	0,0037	0,0057	0,0057	0,0057	0,0057	0,0057	
Cross-talk				0,0035	0,0035	0,0035	0,0035	0,0035	0,0035	0,0035	0,0035	0,0035	0,0035	0,0035	0,0035	0,0035	0,0035	0,0035	0,0035	0,0035	0,0035	
Inherent Noise				0,0009	0,0009	0,0009	0,0009	0,0009	0,0009	0,0009	0,0009	0,0009	0,0009	0,0009	0,0009	0,0009	0,0009	0,0009	0,0009	0,0009	0,0009	
Distortion				0,0004	0,0004	0,0004	0,0004	0,0004	0,0004	0,0004	0,0004	0,0005	0,0005	0,0005	0,0005	0,0004	0,0003	0,0003	0,0003	0,0003	0,0003	
Frequency	f		Hz	0,0000	0,0000	0,0000	0,0000	0,0000	0,0000	0,0000	0,0000	0,0000	0,0000	0,0000	0,0000	0,0000	0,0000	0,0000	0,0000	0,0000	0,0000	
Receiver Ground Shield	B&K			0,0003	0,0003	0,0003	0,0003	0,0003	0,0003	0,0003	0,0003	0,0003	0,0003	0,0003	0,0003	0,0003	0,0003	0,0003	0,0003	0,0003	0,0003	
Transmitter Ground Shield	B&K			0,0003	0,0003	0,0003	0,0003	0,0003	0,0003	0,0003	0,0003	0,0003	0,0003	0,0003	0,0003	0,0003	0,0003	0,0003	0,0003	0,0003	0,0003	
<b>Coupler Properties</b>				<b>0,0039</b>	<b>0,0039</b>	<b>0,0039</b>	<b>0,0039</b>	<b>0,0040</b>	<b>0,0039</b>	<b>0,0039</b>	<b>0,0039</b>	<b>0,0039</b>	<b>0,0039</b>	<b>0,0039</b>	<b>0,0039</b>	<b>0,0040</b>	<b>0,0042</b>	<b>0,0050</b>	<b>0,0059</b>	<b>0,0141</b>	<b>0,0563</b>	
Coupler Length	lcoup	0,0020	mm	0,0014	0,0014	0,0014	0,0014	0,0015	0,0014	0,0014	0,0014	0,0014	0,0013	0,0012	0,0011	0,0008	0,0004	0,0003	0,0017	0,0058	0,0345	
Coupler Diameter	dcoup	0,0030	mm	0,0022	0,0022	0,0022	0,0022	0,0022	0,0022	0,0022	0,0022	0,0022	0,0022	0,0022	0,0022	0,0023	0,0023	0,0023	0,0023	0,0023	0,0022	0,0017
Coupler Volume	Vcoup	0,2470	mm <sup>3</sup>	0,0027	0,0027	0,0027	0,0027	0,0027	0,0027	0,0027	0,0027	0,0027	0,0027	0,0027	0,0027	0,0027	0,0027	0,0027	0,0028	0,0028	0,0026	0,0020
Coupler Surface Area	Scoup	0,1440	mm <sup>2</sup>	0,0000	0,0000	0,0000	0,0000	0,0000	0,0000	0,0000	0,0000	0,0000	0,0000	0,0000	0,0000	0,0000	0,0000	0,0000	0,0000	0,0000	0,0000	
Static Pressure	Ps	0,0275	kPa	0,0012	0,0012	0,0011	0,0011	0,0011	0,0011	0,0011	0,0011	0,0011	0,0011	0,0011	0,0011	0,0012	0,0012	0,0012	0,0012	0,0012	0,0011	0,0008
Temperature	T	0,3464	K	0,0001	0,0001	0,0001	0,0001	0,0000	0,0000	0,0001	0,0001	0,0002	0,0003	0,0004	0,0007	0,0011	0,0018	0,0030	0,0053	0,0119	0,0548	
Relative Humidity	RH	3,3164	%	0,0002	0,0002	0,0002	0,0002	0,0002	0,0002	0,0002	0,0002	0,0002	0,0002	0,0003	0,0004	0,0005	0,0006	0,0010	0,0016	0,0033	0,0141	
<b>Microphone Parameters</b>				<b>0,0096</b>	<b>0,0097</b>	<b>0,0099</b>	<b>0,0099</b>	<b>0,0100</b>	<b>0,0100</b>	<b>0,0099</b>	<b>0,0096</b>	<b>0,0099</b>	<b>0,0100</b>	<b>0,0106</b>	<b>0,0117</b>	<b>0,0128</b>	<b>0,0121</b>	<b>0,0120</b>	<b>0,0519</b>	<b>0,1363</b>	<b>0,8243</b>	
Front Cavity Depth	Lf	0,0030	mm	0,0001	0,0001	0,0001	0,0000	0,0000	0,0000	0,0001	0,0002	0,0003	0,0005	0,0007	0,0012	0,0019	0,0032	0,0054	0,0097	0,0216	0,1006	
Front Cavity Volume	Vf	0,2100	mm <sup>3</sup>	0,0045	0,0045	0,0045	0,0045	0,0045	0,0045	0,0045	0,0045	0,0045	0,0045	0,0045	0,0045	0,0046	0,0047	0,0047	0,0047	0,0044	0,0035	
Equivalent Volume	Veq	0,4200	mm <sup>3</sup>	0,0085	0,0085	0,0088	0,0088	0,0089	0,0089	0,0088	0,0086	0,0085	0,0081	0,0076	0,0068	0,0053	0,0029	0,0002	0,0004	0,0130	0,0564	
Resonance Frequency	fo	2440	Hz	0,0000	0,0000	0,0000	0,0000	0,0000	0,0000	0,0000	0,0001	0,0002	0,0005	0,0010	0,0022	0,0042	0,0065	0,0078	0,0494	0,1018	0,7684	
Loss Factor	D	0,1800		0,0000	0,0000	0,0000	0,0000	0,0000	0,0001	0,0006	0,0015	0,0024	0,0037	0,0056	0,0079	0,0097	0,0087	0,0057	0,0118	0,0654	0,2751	
<b>Additional Heat Conduction Caused by Front Cavity Thread</b>																						
Polarizing Voltage	Uo	0,0048	V	0,0003	0,0003	0,0003	0,0003	0,0003	0,0003	0,0003	0,0003	0,0003	0,0003	0,0003	0,0003	0,0003	0,0003	0,0003	0,0003	0,0003	0,0003	
<b>Imperfection of Theory</b>				<b>0,0000</b>	<b>0,0000</b>	<b>0,0000</b>	<b>0,0000</b>	<b>0,0000</b>	<b>0,0000</b>	<b>0,0000</b>	<b>0,0000</b>	<b>0,0000</b>	<b>0,0000</b>	<b>0,0000</b>	<b>0,0000</b>	<b>0,0000</b>	<b>0,0000</b>	<b>0,0000</b>	<b>0,0000</b>	<b>0,0000</b>	<b>0,0000</b>	
Heat Conduction Theory				0,0000	0,0000	0,0000	0,0000	0,0000	0,0000	0,0000	0,0000	0,0000	0,0000	0,0000	0,0000	0,0000	0,0000	0,0000	0,0000	0,0000	0,0000	
Adding of Excess Volume				0,0000	0,0000	0,0000	0,0000	0,0000	0,0000	0,0000	0,0000	0,0000	0,0000	0,0000	0,0000	0,0000	0,0000	0,0000	0,0000	0,0000	0,0000	
Radial Wave Motion				0,0000	0,0000	0,0000	0,0000	0,0000	0,0000	0,0000	0,0000	0,0000	0,0000	0,0000	0,0000	0,0000	0,0000	0,0000	0,0000	0,0000	0,0000	
<b>Processing of Results</b>				<b>0,0182</b>	<b>0,0125</b>	<b>0,0117</b>	<b>0,0117</b>	<b>0,0117</b>	<b>0,0117</b>	<b>0,0117</b>	<b>0,0116</b>	<b>0,0118</b>	<b>0,0119</b>	<b>0,0122</b>	<b>0,0129</b>	<b>0,0154</b>	<b>0,0190</b>	<b>0,0221</b>	<b>0,0262</b>	<b>0,0348</b>	<b>0,0417</b>	
Rounding Error		0,0050	dB	0,0058	0,0058	0,0058	0,0058	0,0058	0,0058	0,0058	0,0058	0,0058	0,0058	0,0058	0,0058	0,0058	0,0058	0,0058	0,0058	0,0058	0,0058	
Repeatability of Measurements			dB	0,0140	0,0110	0,0100	0,0100	0,0100	0,0100	0,0100	0,0100	0,0100	0,0100	0,0100	0,0100	0,0110	0,0120	0,0150	0,0180	0,0210	0,0240	
Static Pressure Corrections				0,0011	0,0011	0,0011	0,0011	0,0012	0,0012	0,0011	0,0010	0,0009	0,0007	0,0003	0,0006	0,0020	0,0032	0,0004	0,0107	0,0261	0,0308	
Temperature Corrections				0,0009	0,0009	0,0009	0,0009	0,0009	0,0010	0,0014	0,0019	0,0023	0,0028	0,0008	0,0066	0,0089	0,0132	0,0152	0,0146	0,0061	0,0132	
<b>Sum</b>				<b>0,0216</b>	<b>0,0196</b>	<b>0,0191</b>	<b>0,0171</b>	<b>0,0166</b>	<b>0,0167</b>	<b>0,0172</b>	<b>0,0173</b>	<b>0,0174</b>	<b>0,0178</b>	<b>0,0190</b>	<b>0,0221</b>	<b>0,0267</b>	<b>0,0282</b>	<b>0,0597</b>	<b>0,1408</b>	<b>0,8282</b>		
<b>Total Uncertainty with Residual Effects</b>				<b>0,0270</b>	<b>0,0245</b>	<b>0,0238</b>	<b>0,0214</b>	<b>0,0209</b>	<b>0,0208</b>	<b>0,0209</b>	<b>0,0215</b>	<b>0,0216</b>	<b>0,0217</b>	<b>0,0223</b>	<b>0,0238</b>	<b>0,0276</b>	<b>0,0322</b>	<b>0,0353</b>	<b>0,0746</b>	<b>0,1760</b>	<b>1,0352</b>	
<b>Stated Uncertainty</b>				<b>0,06</b>	<b>0,04</b>	<b>0,03</b>	<b>0,03</b>	<b>0,03</b>	<b>0,03</b>	<b>0,03</b>	<b>0,03</b>	<b>0,03</b>	<b>0,03</b>	<b>0,03</b>	<b>0,03</b>	<b>0,04</b>	<b>0,06</b>	<b>0,06</b>	<b>0,08</b>	<b>0,20</b>	<b>1,20</b>	