EURAMET EM-K5

Key comparison of 50/60 Hz power

TECHNICAL PROTOCOL

Organizing laboratories: VSL – PTB – NPL – LNE

TABLE OF CONTENTS

1.	Intro	roduction							
2.	Trave	elling Re	ference standards	3					
	2.1	Descri	otion of the reference standards	3					
	2.2	Measu	rement Quantity	6					
	2.3	Releva	nt parameters	6					
	2.4	Softwa	re for data logging the power readings from the reference standards	7					
	2.5	The ke	y comparison reference value and the degrees of equivalence	7					
	2.6	Prepar	ation of Drafts A and B of this key comparison.	7					
3.	Orga	nization		8					
	3.1	Coordi	nator and members of the support group	8					
	3.2	Partici	pants	8					
	3.3	Time S	chedule	9					
	3.4	EURAN	/IET Toolbox	9					
	3.5	"Door-	to-door delivery scheme"	9					
	3.6	Unpac	king, handling and packing	9					
	3.7	Failure	of the travelling standard	10					
	3.8	Financ	ial aspects, insurance, customs	10					
4.	Meas	suremer	nt instructions	11					
	4.1	Genera	al instructions	11					
	4.2	Particu	lar requirements for energizing and connecting the travelling standards	11					
		4.2.1	Energizing the reference standards at 24 V DC.	11					
		4.2.2	Connecting the reference standards	11					
	4.3	Metho	d of measurement of active power	12					
5.	Unce	rtainty	of measurement	12					
6.	Meas	suremer	nt Report	13					
7.	Refer	rences		13					
A1.	Detai	iled list (of participants	14					
A2.	Schedule of the measurements1								
A3.	Туріс	al scher	ne of an uncertainty budget	18					
A4.	Layou	ut of the	e measurement report	19					
A5.	Confi	rmatior	note of receipt	20					
A6.	Confirmation note of shipment								

1. INTRODUCTION

Committed to underpin the Mutual Recognition Arrangement as proposed by the Committee International of Weights and Measures, CIPM, a key comparison of 50/60 Hz of power standards is currently being organized under the auspices of the Consultative Committee of Electromagnetism, CCEM.

In order to link the National Metrology Institutes organised in EURAMET to the key comparison CCEM-K5, the EURAMET-EM-K5 is proposed.

This protocol has been prepared based on the protocol of the CCEM-K5 and following the EURAMET guide on Comparison N0. 4 Version 1.0 (05/2016).

VSL, PTB, NPL and LNE jointly coordinate the EURAMET comparison

- 4 VSL and NPL are the pilot laboratories and in charge of the analysis and reporting task
- **FIB** and VSL will each supply one travelling standard
- **PTB** is in charge of the reference measurements
- LNE is in charge of the production and the monitoring of the schedule

2. TRAVELLING REFERENCE STANDARDS

2.1 Description of the reference standards

Two traveling reference standards, of the type RADIAN RD-22-332S, will be used in this key comparison. These standards have been adapted to measure active power at 120 V and 240 V and 5 A with outstanding stability in time.

VSL and PTB will provide the reference standards for this EURAMET key comparison.

- ✓ PTB provides the reference standard S/N 207172 that will be use for the loop 1
- ✓ VSL provides the reference standard S/N 208014 that will be use for the loop 2



Figure 1: reference standard: RD-22-3328, serial number: 208 014



Figure 2: reference standard: RD-22-3328, serial number: 207 172

SPECIAL CONNECTIONS ON THE REFERENCE STANDARDS

24 V DC Power Supply:

The travelling standards should be powered using a 24 V DC power supply delivered with the standards. See Figure 4 for the location of the connections.

MAKE SURE that the voltage input selector of the 24 V DC power supply is set to the correct value – 120 V or 240 V – before connecting the power supply to the AC mains.

AC current measurements:

As shown in Figure 3, the reference standards have been fitted to provide high accuracy measurements of 5 A current. As shown in the picture below, a couple of posts out of the reference standard allow for the measurement of 5 A current. This allows for the connection of cables with either banana plugs or spades.

To ensure that the standard reference correctly measures the phase angle between the applied voltage and current, take post in red as the HIGH side of the 5 A current measurement circuit.



Figure 3. Special connection for current measurements on the reference standard

AC voltage measurements:

As shown in Figure 4, the V_{CA} input voltage of the reference standards are on the left-hand side of the panel of the references. Be aware of the connection of the auxiliary power input of 24 V DC, which is located below the V_{CA} posts.



Figure 4: The V_{CA} input posts on the reference standards. The 24 V DC auxiliary power input is below the V_{CA} input posts

2.2 Measurement Quantity

The quantity to be reported is the calibration error of the travelling standard when measuring active power; the calibration error will be expressed in μ W/VA.

The participating laboratory should report a single measurement result and its uncertainty for each of the testing points as given in Table 2.

Parameter	Value
RMS voltages	120 V 240 V
RMS current	5 A
power factor	1.0 0.5 lead 0.5 lag 0 lead 0 lag
testing frequency	53 Hz
Total amount of testing points	10

Table 2. Testing points for the measurement of Active Power.

Notes:

- 1. The "lead" is defined as the current phase leading the voltage phase, and "lag" as the current phase lagging the voltage phase.
- **2.** The measurement result to be reported is the calibration error of the travelling standard, defined as the difference between the measured quantity indicated by the travelling standard and the quantity applied to it, and divided by the nominal apparent power VA.
- 3. The error and uncertainty of the calibration should be expressed in terms of μ W/VA.
- **4.** The travelling standard should be de-energized between each set of measurements for 1 minute, followed by at least 15-minute warm-up period.
- **5.** The total estimated expanded uncertainty quoted in the report shall encompass the Type A uncertainty and the Type B uncertainty of the corresponding NMI calibration system. The expanded uncertainty should be estimated for a level of confidence of 95.45 %.

2.3 Relevant parameters

The values of RMS voltage, RMS current, PF (power factor/phase angle) and frequency are relevant parameters to the 50/60 Hz key comparison power. As indicated in Annex A4 *Layout of the Measurement Report*, participants must record the mean value of voltage, current, PF (power factor/phase angle) and frequency for each of the testing points in Table 1. As shown in Annex A3 *Typical scheme of an uncertainty budget*, the participant laboratory must report all uncertainty sources pertinent to the measurement of power.

2.4 Software for data logging the power readings from the reference standards

EURAMET is grateful to Radian Research Inc. for supporting the CCEM-K5 key comparison of power for the provision of a software package that the participating laboratories may download.

Laboratories may download the following two ftp sites where versions of PC Suite software can be found.

Note that the first file 02.00.13 is for Microsoft XP, whereas version 04.00.04.zip is for Microsoft WIN 7 platform.

PC suit software	platform
www.radianresearch.com/upgrade/PC_Suite_4_00_20.zip	Microsoft WIN 7 or higher
www.radianresearch.com/upgrade/PCSuite_02_00_13.zip	Microsoft XP

2.5 The key comparison reference value and the degrees of equivalence

This protocol has been prepared based on the protocol of the CCEM-K5/2017 following the EURAMET guide on Comparison N0. 4 Version 1.0 (05/2016). The principles of the method of computation of the reference values are as follows.

The comparison will follow the data analysis method used during the last Key Comparison CCEM-K5 and EURAMET.EM-K5.1 and described in the document "Link Between CCEM-K5 and EURAMET.EM-K5" from March 2011.

2.6 Preparation of Drafts A and B of this key comparison.

As stated in the EURAMET guide on Comparison No. 4 Version 1.0 (05/2016), for this key comparison, the following information shall be considered in preparing Drafts A and B with the results of the comparison:

- individual values for each institute together with their declared uncertainties;
- the key comparison reference value with its associated uncertainty;
- for each institute, the deviation from the key comparison reference value and the uncertainty in that deviation (at a 95.45 % level of confidence), i.e. its degree of equivalence.

3. ORGANIZATION

3.1 Coordinator and members of the support group

The organization of the comparison is shared between four institutes: VSL, NPL, PTB, and LNE. VSL with NPL are the pilot laboratories and in charge with the final report. PTB is performing multiple measurements during the comparison in order to determine the stability of the standards, and LNE is in charge with the schedule of the comparison, including the proposition of the technical protocol for EURAMET loop.

Name	NMI	Address	e-mail address	Phone	
Helko van den Brom	VSL	Thijsseweg 11, 2629 JA Delft The Netherlands	hvdbrom@vsl.nl	+31 15 269 1726	
Kristian Dauke	РТВ	Bundesallee 100, 38116 Braunschweig Germany	kristian.dauke@ptb.de	+49 531 592 2336	
Adrian Wheaton	NPL	National Physical Laboratory Hampton Road Teddington, Middlesex UK TW11 0LW	adrian.wheaton@npl.co.uk	+44 20 8943 6235	
Daniela Istrate	LNE	29 avenue Roger Hennequin	daniela.istrate@lne.fr	+33 1 30 69 32 05	
Pierre-Jean Janin	LITE	78197 Trappes Cedex France	pierre-jean.janin@lne.fr	+33 1 30 69 11 26	

3.2 Participants

(see also Annex A1 Detailed list of participants)

Participating NMI	Contact person	e-mail address
BEV	Wolfgang Waldmann	wolfgang.waldmann@bev.gv.at
BIM	Antoaneta Yovcheva	a.yovcheva@bim.government.bg
CEM	Marisa Romero	mlromero@cem.es
CMI	Jiri Streit	jstreit@cmi.cz
EIM	Myrto Holiastou	<u>holiastou@eim.gr</u>
GUM	Grzegorz Sadkowski	j.szutkowski@gum.gov.pl
INM	Doru Flamanzeanu	dflamanzeanu@inm.ro
INRIM	Gabriella Crotti	<u>g.crotti@inrim.it</u>
JV	Kristian Ellingsberg	kbe@justervesenet.no
LNE	Pierre-Jean Janin	pierre-jean.janin@lne.fr
METAS	Jean-Pierre Braun	jean-pierre.braun@metas.ch
METROSERT	Andrei Pokatilov	andrei.pokatilov@metrosert.ee
MIKES VTT	Tapio Lehtonen	<u>tapio.lehtonen@vtt.fi</u>
MIRS/SIQ	Matjaž Lindič	<u>matjaz.lindic@siq.si</u>
NPL	Adrian Wheaton	adrian.wheaton@npl.co.uk
PTB	Kristian Dauke	kristian.dauke@ptb.de
RISE	Stefan Svensson	<u>stefan.svensson@ri.se</u>
SMU	Juraj Slučiak	<u>sluciak@smu.gov.sk</u>
TRESCAL	Torsten Lippert	torsten.lippert@trescal.com
UME	Hüseyin Çayci	huseyin.cayci@tubitak.gov.tr
UMTS	Oleh Velychko	velychko@ukrcsm.kiev.ua
VSL	Helko van den Brom	hvdbrom@vsl.nl

3.3 Time Schedule

The comparison will be organized in two loops. Each loop has one travelling reference standard. See Annex 2 for details of schedule of the measurements

3.4 EURAMET Toolbox

The EURAMET Toolbox will be used to follow and manage this comparison. The adress is <u>http://tssl.de/euramet</u> Information are available at <u>https://tssl.de/euramet/keynote</u> Each participant will be registered by the Support group which is composed of VSL, PBT, NPL, LNE

3.5 "Door-to-door delivery scheme"

In order to reduce delay times at customs, the travelling standards shall be sent directly to the address of the following laboratory according to the schedule program given. Make sure that the address of the following laboratory is as given in Annex 1.

The standards will be provided with an individual rugged plastic container, suitable for shipping the standards on airplane. The standards and accompanying accessories, like connectors and 24 V DC power supply are to be transported inside the rugged plastic container.

The standards are packaged along with a temperature/humidity miniature logger. During measurements at the participant's laboratory, make sure that the logger remains on the top surface of the travelling standard, mainly close to the backlit LCD of the travelling standard, in order to log measurements of ambient temperature and humidity. There is no need for the participating laboratory trying to gain access to the logger; its contents will be downloaded at PTB in order to keep track of the changes of temperature or humidity which may have occurred during transportation or while staying at the participating laboratory.

IMPORTANT: In order to have the final time schedule of the comparison, all the participants should inform LNE whether they agree to send the travelling standard by a customs agency or may want to be responsible of a different way to transport the travelling standard.

The contact person of the participant laboratory must inform LNE of all the pertinent information when sending the travelling standards to the destine customs agency.

Costs involved with sending the travelling standard either to PTB or to the next participant, should be paid by the participant laboratory.

3.6 Unpacking, handling and packing

Unpacking:

- 1. Upon arrival, the participant laboratory must carefully inspect the travelling standard before it is removed from the plastic rugged container. In case of damage to the travelling standard, please take a photograph and send it to the pilot laboratory.
- 2. For removing the travelling standards from their containers, please refer below to the handling instructions.
- 3. Fill the Confirmation note of receipt (Annex A5) and send it to the pilot laboratory.

Handling:

Extreme care should be taken to remove the travelling standards out the rugged plastic container.

For removing the RD-22 from the container, the metal handlers located at both sides of its platform must be used.

Packing:

- 1. For packing the travelling standards, make sure that the travelling standard is comfortably put inside the container. For maneuvering the travelling standards, please refer to the above handling instructions.
- 2. All accompanying accessories should be put inside the container. Make sure that the miniature data logger is put along the standards in the container.
- 3. Inform the pilot laboratory of pertinent details regarding the clearance of customs, shipping and flights. If the travelling standards should be sent to another participant, make sure that both the recipient and the pilot laboratories are informed about travelling details.
- 4. Fill the Confirmation note of shipment (Annex A6) and send it to the pilot laboratory.

3.7 Failure of the travelling standard

In case of failure:

- 1. Unplug the travelling standard.
- 2. Write immediately to the pilot laboratory describing the behavior of the travelling standard. Send along any message displayed on the backlit LCD. Take photographs of the measuring system and of the connecting cables to the travelling standard.
- 3. Wait for instructions from the pilot laboratory on the way to proceed.

3.8 Financial aspects, insurance, customs

In every case, the participant laboratory should ensure that its organization covers insurance for any defect to the travelling standards while they are in its premises.

All the participants will be responsible for expenses related with sending the travelling standards either to the next participant or to the pilot laboratory.

4. MEASUREMENT INSTRUCTIONS

4.1 General instructions

There are no performance tests on the reference standard to be performed before measurements at the participant's laboratory.

4.2 Particular requirements for energizing and connecting the travelling standards

4.2.1 Energizing the reference standards at 24 V DC.

The reference standards are provided with a 24 V DC power supply, which shall be connected to the mains at either 120 V or 240 V, 45 Hz or 65 Hz.

Before connecting the 24 V DC power supply to the mains:

- 1. MAKE SURE that the voltage input selector of the 24 V DC power supply is set to the correct value 120 V or 240 V.
- 2. Make sure that this supply is properly connected to the Auxiliary Power input port of the reference standard as shown in Figure 5.
- 3. Then, plug the 24 V DC power supply to the mains power.

The auxiliary power to the travelling standard should be applied at least 4 hours before starting the tests.

IMPORTANT: In order to achieve the best measurement capability of the travelling standard, allow it to warm up at least 4 hours at room temperature. Then, switch the reference standard off and on again. This action will set the internal DC voltage reference of the standard being corrected at its current internal temperature.

4.2.2 Connecting the reference standards

Make sure that the testing voltage and current sources are properly connected to the input terminals of the reference standard. Refer to Figure 5 for details.



V_{CA} input for 120 and 240 V DC auxiliary power = 24 V DC



I_{CA} input for 5 A

Figure 5. Connecting the reference standard for $V_{CA} = 120$ V and 240 V, $I_{CA} = 5$ A, and auxiliary power of 24 V DC.

4.3 Method of measurement of active power

The measurement method is that used by the participating laboratory for the provision of a calibration. Make sure that testing voltage and current are within at least 0.2 % of the values shown in Table 1.

- 1. Upon arrival, allow the reference standard to warm up at least 4 hours after powering it with the 24 V DC power supply. Then switch the unit off and on in order for its internal DC voltage reference to correct according to its internal temperature.
- 2. For the voltage and current sources, make sure that their frequency is set at 53 Hz, according to the testing points shown in Table 1.
- 3. At every power factor/phase angle shown in Table 1, make as many independent measurements as stated on the calibration procedures of your laboratory.
- 4. Record the readings of active power, voltage, current, power factor/phase angle and frequency displayed on the backlit LCD of the travelling standard.
- 5. Calculate the calibration error of the travelling standard at the testing points shown in Table 1. The calibration error is defined as the difference between the measured quantity indicated by the reference standard and the quantity applied to it, and divided by the applied VA. The calibration error should be expressed in μ W/VA for active power. The error is positive if the reference standard's indication is more positive than the applied quantity.
- 6. The average of at least five sets of measurements should be reported. The reference standard should be de-energized between each set of measurements for 1 minute, followed by at least 15 minute warm-up period.
- 7. The total estimated expanded uncertainty quoted in the laboratory's report should encompass the Type A and Type B uncertainties of the corresponding NMI calibration service. The expanded uncertainty should be estimated for a level of confidence of 95.45 %.
- 8. Report the mean value and spread of the ambient temperature and relative humidity of the laboratory.
- 9. The measurement report of the participant may be completed according to Annex A4 *Layout of the measurement report*.

5. UNCERTAINTY OF MEASUREMENT

(See Annex A3 Typical scheme of uncertainty budget)

The uncertainty of measurement must be determined following the Guide for the Expression of Uncertainty in Measurement [4].

Participant laboratories are requested to report the main uncertainty components of their measurement systems, identifying all the pertinent uncertainty sources and quantifying their contribution to the expanded uncertainty.

In order to have a comparable uncertainty evaluation, each laboratory is asked to report the following information in the form of an uncertainty budget:

- i. The result of the type A method of uncertainty estimation which yields the standard deviation of the mean values of data sets recorded by the participant in order to calculate its final report value.
- ii. The result of the type B method of uncertainty estimation.
- iii. The expanded uncertainty estimated at a 95.45 % level of confidence.
- iv. The degrees of freedom for the estimation of the expanded uncertainty at a 95.45 % level of confidence.

6. MEASUREMENT REPORT

(See Annex A4 *Layout of the measurement report* and the template Excel reporting file) The participant laboratory should report the following information within 2 (two) months from the end of measurements:

- 1. The mean of the calibration error of the travelling standard. The calibration error should be expressed in terms of $\mu W/VA$.
- 2. The expanded uncertainty of the calibration error of the travelling standard should be estimated at a 95.45 % level of confidence. The expanded uncertainty should be expressed in terms of μ W/VA.
- 3. The uncertainty components coming out from the Type A and B methods of uncertainty estimation of the calibration error of the travelling standard.
- 4. The mean value and standard uncertainty of the relevant parameters of the active power value, such as: RMS voltage, RMS current, power factor / phase angle, frequency. The standard uncertainty of the relevant parameters should be determined for a 95.45 % level of confidence (k = 2 for a normal distribution).
- 5. The final results for all 10 test points must also be reported using the Excel template file associated with the comparison.
- 6. Measurement uncertainty estimation shall comply with the Guide to the Expression of Uncertainty in Measurement [4].
- 7. A complete uncertainty budget and any additional information must be reported as shown in Annex A3.

All participants must report the results of the comparison to VSL as soon as possible, in every case, not later than 2 (two) months after the measurements are completed.

7. **References**

- N. Oldham, T. Nelson. R. Bergeest. G. Ramm, R. Carranza. A. C. Corney, M. Gibbes, G. Kyriazis, H. Laiz, L. Liu, Z. Lu, U. Pogliano, K. Rydler, E. Shapiro, E. So, M. Temba, and P. Wright, "An International Comparison of 50/60 Hz Power (1996-1999), IEEE Trans. Instr. and Meas., Vol. 50, Num. 2, pp. 356-360, Apr. 2001.
- N. Oldham, T. Nelson, T. N. F. Zhang and H. K. Liu, "CCEM-K5 Comparison of 50/60 Hz Power. Final Report", Metrologia 40, Technical Supplement 01003.
- 3. EURAMET guide on Comparison N0. 4 Version 1.0 (05/2016)
- 4. Link Between CCEM-K5 and EURAMET.EM-K5 March 2011
- 5. Guide to the Expression of Uncertainty in Measurement, 1999. International Organization of Standards, Geneva, Switzerland. ISO ENV 13005:1999.

A1. DETAILED LIST OF PARTICIPANTS

NMI	Country	Contact person	Email Contact person	Adresse
BEV	Austria	Wolfgang Waldmann	wolfgang.waldmann@bev .gv.at	BEV - Bundesamt für Eich- und Vermessungswesen Section Electrical Quantities A-1160 Wien, Arltgasse 35, Austria
BIM	Bulgaria	Antoaneta Yovcheva Gergana Georgieva	a.yovcheva@bim.govern ment.bg g.georgieva@bim.govern ment.bg	Bulgarian Institute of Metrology (BIM) 52B, G.M. Dimitrov blvd. 1040 Sofia Bulgaria
CEM	Spain	Marisa Romero Javier Díaz de Aguilar	<u>mlromero@cem.es</u> jdiaz@cem.es	Centro Español de Metrología (CEM) C/del Alfar 2 28760 Tres Cantos (Madrid) Spain
СМІ	Czech Republic	Jiri Streit	jstreit@cmi.cz	Czech Metrology Institute (CMI) Okružní 31 638 00 Brno Czech Republic
EIM	Greece	Myrto Holiastou	holiastou@eim.gr	HELLENIC INSTITUTE OF METROLOGY (EIM) Industrial Area of Thessaloniki Block 45, Sindos, GR 57022 Thessaloniki ,Greece
GUM	Poland	Grzegorz Sadkowski	j.szutkowski@gum.gov.pl	Laboratory of Electricity and Magnetism Central Office of Measures ul. Elektoralna 2, 00-139 Warszawa - Poland
INM	Romania	Doru Flamanzeanu	dflamanzeanu@inm.ro	National Institute of Metrology (INM) Sos. Vitan-Bârzesti 11 042122 Bucuresti Romania
INRIM	Italy	Gabriella Crotti	<u>g.crotti@inrim.it</u>	INRIM Istituto Nazionale di Ricerca Metrologica Division of Metrology for the Quality of Life Strada delle Cacce, 91 10135 Torino - ITALY
JV	Norway	Kristian Ellingsberg	kbe@justervesenet.no	Justervesenet - Norwegian Metrology Service (JV) Fetveien 99 2007 Kjeller Norway
LNE	France	Pierre-Jean Janin Daniela Istrate	pierre-jean.janin@lne.fr daniela.istrate@lne.fr	29 avenue Roger Hennequin 78197 Trappes Cedex France

The indicated NMI addresses are the shipping addresses.

METAS	Switzerland	Jean-Pierre Braun Christian Mester	jean- pierre.braun@metas.ch christian.Mester@metas.c h	Federal Institute of Metrology METAS (METAS) Lindenweg 50 3003 Bern-Wabern Switzerland
METROSERT	Estonia	Andrei Pokatilov	andrei.pokatilov@metrose rt.ee	Research Scientist National Standards Laboratory for Electrical Quantities AS Metrosert R&D division Teaduspargi 8, 12618 Tallinn Estonia
MIKES VTT	Finland	Tapio Lehtonen	tapio.lehtonen@vtt.fi	VTT Technical Research Centre of Finland Ltd, Centre for Metrology MIKES Tekniikantie 1 02150 Espoo Finland
MIRS/SIQ	Slovenia	Matjaž Lindič	matjaz.lindic@siq.si	Metrology Department SIQ LJUBLJANA, Tržaška c. 2 SI-1000 Ljubljana, Slovenia
NPL	United Kingdom	Adrian Wheaton	adrian.wheaton@npl.co.u k	National Physical Laboratory Hampton Road Teddington, Middlesex UK TW11 0LW
РТВ	Germany	Kristian Dauke	kristian.dauke@ptb.de	Bundesallee 100, 38116 Braunschweig Germany
RISE	Sweden	Stefan Svensson	stefan.svensson@ri.se	RISE Research Institutes of Sweden AB 50115 Borås Sweden
SMU	Slovakia	Juraj Slučiak	<u>sluciak@smu.gov.sk</u>	Slovak Insitute of Metrology Department of Electricity and Time Karloveská 63, 842 55 Bratislava Slovakia
TRESCAL	Denmark	Torsten Lippert	torsten.lippert@trescal.co m	Trescal A/S Mads Clausens Vej 12 8600 Silkeborg, Denmark
UME	Turkey	Hüseyin Çayci	huseyin.cayci@tubitak.go v.tr	Ulusal Metroloji Enstitüsü (UME) Barış Mah. Dr. Zeki Acar Cad. No:1 41470 Gebze, Kocaeli Turkey
SE "Ukrmetr- teststandard" (UMTS)	Ukraine	Oleh Velychko	velychko@ukrcsm.kiev.ua	SE "Ukrmetrteststandard" 4, Metrologichna Str., 03143, Kyiv, Ukraine
VSL	Netherlands	Helko van den Brom	hvdbrom@vsl.nl	Thijsseweg 11, 2629 JA Delft The Netherlands

A2. SCHEDULE OF THE MEASUREMENTS

-				Planning			Accomplish		
	NMI	Country	Start date Initial schedule	End date Initial schedule	Duration (calendar days)	Start date Updated	End date Updated	Duration (calendar days) Updated	
	PTB	Germany		04/02/2019					
	Transport from PTB to GUM		04/02/2019	11/02/2019	7	04/02/2019	04/02/2019	0	
	GUM	Poland	11/02/2019	11/03/2019	28	04/02/2019	28/02/2019	24	
	Transpor	t from GUM to PTB	11/03/2019	18/03/2019	7	28/02/2019	04/03/2019	4	
	PTB	Germany	18/03/2019	18/03/2019	0	04/03/2019	08/03/2019	4	
	Transpo	rt from PTB to CMI	18/03/2019	18/03/2019	0	08/03/2019	15/03/2019	7	
	СМІ	Czech Republic	18/03/2019	15/04/2019	28	15/03/2019	16/04/2019	32	
	Transpor	t from CMI to SMU	15/04/2019	22/04/2019	7	16/04/2019	16/04/2019	0	
	SMU	Slovakia	22/04/2019	20/05/2019	28	16/04/2019	22/05/2019	36	
	Transpor	t from SMU to BEV	20/05/2019	27/05/2019	7	22/05/2019	22/05/2019	0	
	BEV	Austria	27/05/2019	24/06/2019	28	22/05/2019	24/06/2019	33	
	Transpor	t from BEV to INM	24/06/2019	01/07/2019	7	24/06/2019	26/06/2019	2	
	INM	Romania	01/07/2019	26/08/2019	56	26/06/2019	26/08/2019	61	
	Transport from INM to PTB		26/08/2019	02/09/2019	7	26/08/2019	29/08/2019	3	
	PTB	Germany	02/09/2019	30/09/2019	28	29/08/2019	30/09/2019	32	
Loop No. 1	Transport from PTB to UME		30/09/2019	14/10/2019	14	30/09/2019	04/11/2019	35	
Standard S/N 207172	UME	Turkey	14/10/2019	13/11/2019	28	04/11/2019	15/11/2019	11	
Provided by PTB	Transport from UME to PTB		14/11/2019	28/11/2019	14	15/11/2019	02/12/2019	17	
	РТВ	Germany	28/11/2019	28/11/2019	0	02/12/2019	03/12/2019	1	
	Transport from PTB to MIRS/SIQ		28/11/2019	05/12/2019	7	03/12/2019	06/12/2019	3	
	MIRS/SIQ	Slovenia	05/12/2019	17/01/2020	43	06/12/2019	16/01/2020	41	
	Transport from MIRS/SIQ to INRIM		17/01/2020	21/01/2020	4	16/01/2020	17/01/2020	1	
	INRIM	Italy	21/01/2020	18/02/2020	28	17/01/2020	24/02/2020	38	
	Transport	from INRIM to BIM	18/02/2020	25/02/2020	7	24/02/2020	27/02/2020	3	
	BIM	Bulgaria	25/02/2020	10/06/2020	106	27/02/2020	11/06/2020	105	
	Transpo	rt from BIM to EIM	10/06/2020	17/06/2020	7	11/06/2020	12/06/2020	1	
	EIM	Greece	17/06/2020	15/07/2020	28	12/06/2020	13/07/2020	31	
	Transpo	rt from EIM to PTB	15/07/2020	22/07/2020	7	06/08/2020	07/08/2020	1	
	PTB	Germany	22/07/2020	19/08/2020	28	10/08/2020	14/08/2020	4	
	Transport from PTB to UMTS		19/08/2020	13/10/2020	55	14/08/2020			
	UMTS	Ukraine	13/10/2020	10/11/2020	28				
	Transport	from UMTS to PTB	10/11/2020	24/11/2020	14				
	PTB	Germany	24/11/2020	22/12/2020	28				

 Table A.2.1. Circulation schedule of the reference standard

 Loop 1 - Standard S/N 207172 - Provided by PTB

			Planning			Accomplish		
	NMI	Country	Start date Initial schedule	End date Initial schedule	Duration (calendar days)	Start date Updated	End date Updated	Duration (calendar days) Updated
	РТВ	Germany		04/02/2019				
	Transport from PTB to TRESCAL		04/02/2019	11/02/2019	7	04/02/2019	05/02/2019	1
	TRESCAL	Denmark	11/02/2019	11/03/2019	28	05/02/2019	05/03/2019	28
	Transport f	from TRESCAL to PTB	11/03/2019	18/03/2019	7	05/03/2019	11/03/2019	6
	РТВ	Germany	18/03/2019	18/03/2019	0	11/03/2019	14/03/2019	3
	Transpo	rt from PTB to RISE	18/03/2019	18/03/2019	0	14/03/2019	18/03/2019	4
	RISE	Sweden	18/03/2019	15/04/2019	28	18/03/2019	23/04/2019	36
	Transport fr	om RISE to MIKES VTT	15/04/2019	22/04/2019	7	23/04/2019	24/04/2019	1
	MIKES VTT	Finland	22/04/2019	20/05/2019	28	24/04/2019	24/05/2019	30
	Transport from I	MIKES VTT to METROSERT	20/05/2019	27/05/2019	7	24/05/2019	27/05/2019	3
	METROSERT	Estonia	27/05/2019	24/06/2019	28	27/05/2019	21/06/2019	25
	Transport fro	om METROSERT to VSL	24/06/2019	01/07/2019	7	21/06/2019	26/06/2019	5
	VSL	Netherlands	01/07/2019	26/08/2019	56	26/06/2019	26/08/2019	61
	Transport from VSL to PTB		26/08/2019	02/09/2019	7	26/08/2019	27/08/2019	1
	РТВ	Germany	02/09/2019	30/09/2019	28	27/08/2019	30/09/2019	34
Loop No. 2	Transport from PTB to VSL		30/09/2019	14/10/2019	14	30/09/2019	14/10/2019	14
Standard	Transport from VSL to JV					16/10/2019	01/11/2019	16
S/N 208014 Provided by	VL	Norway	01/11/2019	29/11/2019	28	01/11/2019	02/12/2019	31
VSL	Transport from JV to VSL		29/11/2019	13/12/2019	14	04/12/2019	16/12/2019	12
	Transport from VSL to METAS		06/01/2020	13/01/2020	7	13/01/2020	13/01/2020	0
	METAS	Switzerland	13/01/2020	13/03/2020	28	13/01/2020	17/03/2020	64
	Transport	from METAS to VSL	13/03/2020	27/03/2020	14	17/03/2020	18/03/2020	1
	VSL	Netherlands	27/03/2020	27/03/2020	0	18/03/2020	18/03/2020	0
	Transpo	rt from VSL to CEM	13/05/2020	18/05/2020	5	13/05/2020	18/05/2020	5
	СЕМ	Spain	18/05/2020	15/06/2020	28	18/05/2020	18/06/2020	31
	Transpor	t from CEM to LNE	15/06/2020	22/06/2020	7	23/06/2020	26/06/2020	3
	LNE	France	22/06/2020	20/07/2020	28	26/06/2020	04/08/2020	39
	Transpo	rt from LNE to PTB	20/07/2020	03/08/2020	14	10/08/2020	17/08/2020	7
	РТВ	Germany	03/08/2020	31/08/2020	28	18/08/2020	21/08/2020	3
	Transpo	rt from PTB to VSL	31/08/2020	07/09/2020	7	31/08/2020	03/09/2020	3
	Transpo	rt from VSL to NPL	07/09/2020	21/09/2020	14	03/09/2020	08/09/2020	5
	NPL	United Kingdom	21/09/2020	19/10/2020	28	08/09/2020		
	Transpo	rt from NPL to VSL	19/10/2020	02/11/2020	14			
	Transpo	rt from VSL to PTB	02/11/2020	09/11/2020	7			
	РТВ	Germany	09/11/2020	07/12/2020	28			

Loop 2 - Standard S/N 208014 - Provided by VSL

A3. TYPICAL SCHEME OF AN UNCERTAINTY BUDGET

The uncertainty of measurement must be determined following the Guide for the Expression of Uncertainty in Measurement [4]. Information of the uncertainty of measurements must be provided in this form.

Participant laboratories are requested to report the main uncertainty components of their measurement systems, identifying all the pertinent uncertainty sources and quantifying their contribution to the expanded uncertainty.

In order to have a comparable uncertainty evaluation, each laboratory is asked to report the following information in the form of an uncertainty budget:

- i. The result of the type A method of uncertainty estimation which yields the standard deviation of the mean values of data sets recorded by the participant in order to calculate its final report value.
- ii. The result of the type B method of uncertainty estimation.
- iii. The expanded uncertainty estimated at a 95.45 % level of confidence.
- iv. The degrees of freedom for the estimation of the expanded uncertainty at a 95.45 % level of confidence.

Main uncertainty	Standard	Type method A or B of	Sensitivity	Uncertainty	Degrees of				
components	uncertainty	evaluation/probability	coefficient	contribution	freedom				
yi yi	u(yi)	distribution function	Ci	$u(R_i)$	n _i				
1) Standard deviation of		Type A							
the calibration error of									
the travelling standard		pdf: Normal							
2) uncertainty									
components of the									
reference standard of the									
participant									
3) Ambient conditions									
3.1) temperature									
3.2) humidity									
Root square sum of Type A	Root square sum of Type A standard uncertainties and effective degrees of freedom								
Root square sum of Type B standard uncertainties and effective degrees of freedom									
Combined standard uncert	Combined standard uncertainty and effective degrees of freedom								
Expanded uncertainty (95.	45 % coverage fa	ctor)							

A4. LAYOUT OF THE MEASUREMENT REPORT

- 1. Identification of the travelling standard: RD-22
- 2. Identification of the participant laboratory and its representative
- 3. Measurement set-up and traceability scheme
- 4. Measurement procedure
- 5. Results:
 - a. Mean value of the calibration error of reference standard at the active power values shown in Table 1, expressed in μ W/VA.
 - b. Expanded uncertainty estimated at a 95.45 % confidence level and the degrees of freedom of the calibration error of the travelling standard at active power.
 - c. Mean value of the relevant parameters measured by the reference standard: voltage, current, power factor, frequency.
 - d. Mean date of measurement.
 - e. Ambient conditions: mean value and spread of temperature and humidity measurements.

These measurement results must (also) be reported using the template xls file that comes with this protocol

- 6. Detailed uncertainty budget as in Annex 3.
- 7. Report the date and time when the reference standard is de-energized and energized.
- 8. Signature and title of the laboratory representative.

A5. CONFIRMATION NOTE OF RECEIPT

To:

Daniela Istrate and Pierre-Jean Janin, <u>daniela.istrate@lne.fr</u>; <u>pierre-jean.janin@lne.fr</u> LNE, France

From: (participating laboratory)

Re: EURAMET-K5 Receipt of travelling standard (identification of the travelling standard: RD-22

We confirm having received the travelling standard of the EURAMET-K5 key comparison 50/60 Hz power on (date)....

After visual inspection:

No damage of the transport package and the travelling standard has been noticed.

Or

The following damage(s) are reported. In such a case, please add pictures of damages.

Date:

A6. CONFIRMATION NOTE OF SHIPMENT

To:

Daniela Istrate and Pierre-Jean Janin, <u>daniela.istrate@lne.fr</u> ; <u>pierre-jean.janin@lne.fr</u> LNE, France

Re: EURAMET -K5 Dispatch of travelling standard (identification of the travelling standard: RD-22)

We confirm having shipped the travelling standard of the EURAMET -K5 key comparison 50/60 Hz power on (date).... to the following address:

a) if the travelling standard will be sent to the pilot laboratory:

PTB Braunschweig Kristian Dauke Working Group 2.33 Bundesallee 100 38116 Braunschweig Germany

b) if the travelling standard will be sent to another participant laboratory: Shipping address of the next participant (see Annex 1)

We confirm that we have included in both shipping containers, the accessories:

- \Box 24 DC power supply
- \Box Cable for PC read out of the RD22
- □ Datalogger for temperature / humidity

In any case, add the following information:

- flight number,
- airline,
- estimated time of arrival,
- air bill number and
- any other pertinent information about the shipment

Date: