

September 13, 2023

Technical Protocol for Regional Key Comparison on LS2P Microphone Calibration in Pressure Field - SIM.AUV.A-K6

Amendment 1

Page 2, Section “**Microphone to be circulated**”

Replace by

Microphones to be circulated

For this comparison, the microphones to be circulated will be two LS2P microphones, manufactured by Brüel and Kjær, type 4180 [1], serial numbers **2787481** and 3257821. These microphones require a polarizing voltage of 200 V. The stability of the microphones will be monitored by the pilot laboratory prior to and during the period of their circulation for calibration by the participants.
