# Protocol APMP.QM-K3.2019

# Automotive emission gases

Kiryong Hong KRISS, 267 Gajeon-ro, Daejeon 34113, Republic of Korea March 13, 2023

# **Table of contents**

Table of contents	II
Background	1
Measurands	1
Mixtures	1
Supported claims	2
Participants	2
Schedule	2
Coordinator	2
Measurement report	3
Report form	3
Calibration standards	4
Instrumentation	4
Calibration method and value assignment	4
Uncertainty evaluation	4
References	5
Authorship	5

## Background

To support calibration and measurement capabilities (CMCs) for gas mixtures presenting little analytical challenge, a key comparison on a 'core' gas mixture is organised.

## Measurands

Amount fractions carbon monoxide, carbon dioxide, oxygen and propane (in a matrix of high-purity nitrogen)

### Mixtures

A set of mixtures will be prepared gravimetrically and subsequently verified against other PSMs. The mixtures will be prepared by the coordinating laboratory KRISS. The nominal composition of the mixtures is within the following ranges:

Component	Amount fraction		
	x		
Carbon monoxide	(0.5 - 2) cmol/mol		
Carbon dioxide	(2-5) cmol/mol		
Oxygen	(1-4) cmol/mol		
Propane	(100 – 300) µmol/mol		
Nitrogen	Balance		

Table 1: Nominal ranges of amount of substance fractions

The pressure in the cylinders will be approximately 10 MPa; cylinders of 10 L nominal volume will be used. The amount fractions as calculated from gravimetry and purity data will be used as key comparison reference value. Each cylinder will have its own reference value and associated expanded uncertainty.

Please note that:

- 1. Participating laboratories are responsible for the calibration of their own equipment. For a proper evaluation of the data, it is necessary that the calibration method, as well as the way in which the calibration mixtures have been prepared is reported to the coordinator in sufficient detail to check the data for obvious omissions.
- 2. After the measurements, the participants have to return the cylinders with a sufficient amount of gas (pressure shall be at least 5 MPa) to KRISS for second verification analysis.
- 3. Transport of the cylinders from coordinating laboratory (KRISS) to the participating laboratories will be paid for by the KRISS and back to the coordinating laboratory (KRISS) will be paid for by the participants. The participants are also responsible for returning the cylinders promptly after completing the measurements.
- 4. The measurement report requires per cylinder at least three independent measurements, obtained under repeatability conditions. This is a strict requirement to come to proper statistical analysis of the reported data.
- 5. Participating laboratories are requested to specify in detail which analytical method(s) have been used and how the evaluation of measurement uncertainty was performed.
- 6. Additional measurement reports and additional information can be submitted jointly with the report forms to KRISS and will be taken into consideration during the evaluation.

### **Supported claims**

This key comparison can be used to support CMC claims for mixtures of propane in nitrogen, propane in air, and mixtures falling into the category of "core gas mixtures".

### **Participants**

Participants: KRISS (coordinating laboratory), NIM, NMIM, NMC/A\*STAR, NIMT, NMIJ, NMIA, CMS/ITRI, NPLI, CENAM, and CERI

#### Schedule

The schedule for this key comparison reads as follows

Table 2: Key comparison schedule

Date	Event		
March 2023	Agreement of protocol		
April 2023	Registration of participants		
October 2023	Preparation and verification of mixtures		
November 2023	Dispatch of mixtures		
March 2024	Reports and cylinders arrived back at		
	KRISS		
May 2024	Re-verification of the mixtures		
September 2024	Draft A report available		
March 2025	Draft B report available		

#### Coordinator

Korea Research Institute of Standards and Science (KRISS) Gas Metrology Group Kiryong Hong 267 Gajeong-ro, Yuseong-gu Daejeon 34113 Republic of Korea Phone: +82-42-868-5236 E-mail: khong@kriss.re.kr

# Measurement report

## **Report form**

Laboratory name:

Cylinder number:

## Measurement 1<sup>#</sup>

Component	Date (dd/mm/yy)	Result (cmol/mol)	Standard deviation (% relative)	Number of replicates
СО				
CO <sub>2</sub>				
O <sub>2</sub>				
C <sub>3</sub> H <sub>8</sub>				

#### Measurement 2<sup>#</sup>

Component	Date (dd/mm/yy)	Result (cmol/mol)	Standard deviation (% relative)	Number of replicates
СО				
CO <sub>2</sub>				
O <sub>2</sub>				
C <sub>3</sub> H <sub>8</sub>				

# Measurement 3<sup>#</sup>

Component	Date (dd/mm/yy)	Result (cmol/mol)	Standard deviation (% relative)	Number of replicates
СО				
CO <sub>2</sub>				
O <sub>2</sub>				
C <sub>3</sub> H <sub>8</sub>				

Measurement X<sup>#</sup>

Component	Date (dd/mm/yy)	Result (cmol/mol)	Standard deviation (% relative)	Number of replicates
СО				
CO <sub>2</sub>				
O <sub>2</sub>				
C <sub>3</sub> H <sub>8</sub>				

#### Results

Component	Result (cmol/mol)	Expanded uncertainty (cmol/mol)	Coverage factor
СО			
CO <sub>2</sub>			
O <sub>2</sub>			
$C_3H_8$			

#### **Calibration standards**

Please provide a brief description of the calibration standards used

#### Instrumentation

Please provide a brief description of the particulars of the instrument(s) used in this key comparison:

#### Calibration method and value assignment

Please provide a brief description how the equipment was calibrated and how the assigned value was calculated (including the necessary formulae):

#### **Uncertainty evaluation**

Please provide a brief description of the evaluation of measurement uncertainty.

#### References

Any literature references you may wish to refer to come here ...

## Authorship

Please provide the authorship of the measurement report (2-3 persons typically)