

LEr 29/1/2013

## **TECHNICAL PROTOCOL**

### **1. INTRODUCTION**

The comparison is organised within the EU-Indonesia Trade Support Programme II, Sub-project Number APE12-06b, "Improvement of traceability of Metrology and Calibration measurements of Puslit KIM".

Two National Metrology Institutes take part in this comparison: LNE (France) and KIM-LIPI (Indonesia).

LNE is acting as the pilot laboratory and in this function is responsible for providing the travelling standard, the evaluation of the measurement results and the final report.

The comparison will be accomplished in accordance with the EURAMET Guidelines on Conducting Comparisons and CCEM Guidelines for Planning, Organising, Conducting and Reporting Key, Supplementary and Pilot Comparisons.

### **2. TRAVELLING STANDARDS**

2.1. The travelling standard is a resistance (YEW, type 2792) having the nominal value of 1mΩ.

2.2. Specifications

Nominal value of the resistance:	1mΩ
Dimensions of the case:	40 mm x 30 mm x 20 mm
Mass Approx:	3 kg

### **3. Quantities to be measured**

- **R**: resistance of the standard (four terminals)
- **I**: DC current through the resistor
- **T<sub>ext</sub>**: the temperature (°C) of the environment where the standard is measured.

### **4. Measurement instructions**

The measurements should be performed under the following conditions:

- DC current: 1 A, 10 A and 20A
- Temperature of the environment: 23°C ± 2°C
- Relative humidity of air: between 30 % and 70 %.

### **5. Reporting of results**

A report should be sent to the pilot laboratory within one month after the measurements are completed. The report should include:

- Description of the measurement method
- The reference standard
- The traceability to the SI
- The results of the quantities to be measured (list of section 3)
- The associated standard uncertainties, the effective degrees of freedom and the expanded uncertainties.

# BILATERAL COMPARAISON of DC Resistance 1 mΩ

The measurement of the DC current and the temperature of the environment must also be recorded and reported.

## 6. Uncertainty of measurement

The uncertainty must be calculated following the ISO "Guide to the expression of uncertainty in measurement" (GUM) and the complete uncertainty budget must be reported.

## 7. Transportation

The travelling standard must be transported in the original case and protected from mechanical loads, vibration etc. for transport by plane.

The travel box contains the following items:

- Resistance standard
- Operating instructions of the travelling standard (this document).

## 8. CONTACT

**Pilot Laboratory :** Laboratoire national de métrologie et d'essais (LNE)  
ZA de Trappes-Élancourt  
29, avenue Roger Hennequin  
78197 TRAPPES Cedex  
France

**Contact :** Mrs. BLANC Isabelle  
Tél : 01 30 69 21 08  
Fax : 01 30 16 24 52  
Mail : [isabelle.blanc.@lne.fr](mailto:isabelle.blanc.@lne.fr)

**KIM LIPI :** Pusat Penelitian Kalibrasi, Instrumentasi, dan Metrologi  
Lembaga Ilmu Pengetahuan Indonesia (Puslit KIM-LIPI)  
Kompleks PUSPIPTEK Gedung 420  
Tangerang Selatan, Banten Indonesia

**Contacts:** Mr. R. Hadi Sardjono  
Tel : (+6221) 7560562  
Fax : (+6221) 7560064  
Email : [hadisarjono@kim.lipi.go.id](mailto:hadisarjono@kim.lipi.go.id)

Mrs. Lukluk Khairiyati  
Tel : (+6221) 7560562  
Fax : (+6221) 7560064  
Email : [luluk@kim.lipi.go.id](mailto:luluk@kim.lipi.go.id)

\*\*\*\*\*

### Laboratoire national de métrologie et d'essais

Établissement public à caractère industriel et commercial • Siège social : 1, rue Gaston Boissier - 75724 Paris Cedex 15 • Tél. : 01 40 43 37 00  
Fax : 01 40 43 37 37 • E-mail : [info@lne.fr](mailto:info@lne.fr) • Internet : [www.lne.fr](http://www.lne.fr) • Siret : 313 320 244 00012 • NAF : 743 B • TVA : FR 92 313 320 244  
Barclays Paris Centrale IBAN : FR76 3058 8600 0149 7267 4010 170 BIC : BARCFRPP