

Key comparison CCAUV.V-K2

DEVICE : Single-Ended Accelerometer

MEASURAND : Phase

FREQUENCIES : 10 Hz, 40 Hz, 80 Hz, 160 Hz, 800 Hz, 2 kHz, 5 kHz and 10 kHz

The laboratories' individual measurement may be found in Section 7.2.1 of the Final Report (starting page 18).

The computation of the key comparison reference value, x_R , and of its expanded uncertainty ($k = 2$), U_R , is explained in Section 8 of the Final Report.

10 Hz	
x_R	U_R
/ °	/ °
0.012	0.107

40 Hz	
x_R	U_R
/ °	/ °
-0.011	0.103

80 Hz	
x_R	U_R
/ °	/ °
-0.018	0.104

160 Hz	
x_R	U_R
/ °	/ °
-0.020	0.104

800 Hz	
x_R	U_R
/ °	/ °
-0.028	0.100

2 kHz	
x_R	U_R
/ °	/ °
-0.103	0.120

5 kHz	
x_R	U_R
/ °	/ °
-0.108	0.152

10 kHz	
x_R	U_R
/ °	/ °
0.038	0.206

The degree of equivalence of laboratory i with respect to the key comparison reference value is given by a pair of terms: D_i and its expanded uncertainty ($k = 2$), U_i , computed as explained in Section 8 of the Final Report.

The pair-wise degrees of equivalence are available from the Final Report starting on page 58.

Key comparison CCAUV.V-K2


DEVICE : Single-Ended Accelerometer

MEASURAND : Phase


FREQUENCIES : 10 Hz, 40 Hz, 80 Hz, 160 Hz, 800 Hz, 2 kHz, 5 kHz and 10 kHz

Lab <i>i</i>	10 Hz		40 Hz		80 Hz		160 Hz		800 Hz		2 kHz		5 kHz		10 kHz	
	<i>D_i</i>	<i>U_i</i>														
	<i>I</i> °		<i>I</i> °		<i>I</i> °		<i>I</i> °		<i>I</i> °		<i>I</i> °		<i>I</i> °		<i>I</i> °	
PTB	-0.04	0.17	-0.03	0.17	-0.01	0.17	0.00	0.17	0.02	0.17	0.06	0.49	-0.02	0.48	-0.29	0.46
CMI	0.02	0.49	0.00	0.49	-0.02	0.49	-0.05	0.49	-0.28	0.49	-0.59	0.49	-1.43	0.52	-2.81	0.54
NMISA	-0.07	0.39	-0.06	0.39	-0.03	0.39	-0.08	0.39	0.07	0.39	0.22	0.49	0.42	0.79	0.57	0.77
DPLA	-0.07	0.28	-0.01	0.28	0.00	0.28	-0.01	0.28	-0.06	0.28	-0.12	0.27	-0.32	0.26	-0.68	0.98
CEM	-0.06	0.49	0.02	0.49	0.04	0.49	0.08	0.49	0.08	0.49	0.18	0.99	0.25	0.99	0.04	0.98
GUM	-0.04	0.69	-0.04	0.59	-0.03	0.59	0.00	0.59	0.05	0.59	0.19	0.59	0.27	0.79	0.13	0.98
METAS	-0.01	0.39	0.02	0.37	-0.03	0.37	0.04	0.37	0.06	0.37	0.13	0.46	0.19	0.46	-0.25	0.83
NMIJ	-0.05	0.81	-0.01	0.41	0.03	0.47	-0.01	0.51	0.00	0.28	-0.03	0.25	0.15	0.24	0.19	0.22
KRISS																
LNE	-0.35	2.00	-0.07	2.00	0.17	2.00	0.12	2.00	0.17	2.00	0.48	5.00	1.08	5.00	1.98	5.00
NIM	0.15	0.49	0.06	0.49	0.06	0.49	0.01	0.49	-0.06	0.49	-0.20	0.49	-0.64	0.52	-1.53	1.02
CENAM	0.01	0.99	-0.03	0.99	-0.02	0.99	0.01	0.99	0.00	0.99	0.00	0.99	-0.13	0.99	-0.33	0.98
INMETRO	-0.01	0.21	0.01	0.22	0.03	0.22	0.04	0.22	0.08	0.22	0.20	0.21	0.38	0.48	0.05	0.98
UME	0.04	0.49	0.03	0.49	0.02	0.49	-0.03	0.49	-0.02	0.49	-0.16	0.99	-0.86	0.99	-1.91	1.49
VNIIM	2.30	0.99	0.34	0.75	-0.11	0.75	-0.21	0.75	-0.55	0.75	-0.62	0.75	-0.93	0.99	-3.74	1.51


CCAUV.V-K2 Single-Ended Accelerometer (phase) 10 Hz




CCAUV.V-K2 Single-Ended Accelerometer (phase) 40 Hz




CCAUV.V-K2 Single-Ended Accelerometer (phase) 80 Hz




CCAU.V-K2 Single-Ended Accelerometer (phase) 160 Hz




CCAUV.V-K2 Single-Ended Accelerometer (phase) 800 Hz



CCAU.V-K2 Single-Ended Accelerometer (phase) 2 kHz



CCAUV.V-K2 Single-Ended Accelerometer (phase) 5 kHz



CCAUV.V-K2 Single-Ended Accelerometer (phase) 10 kHz

