### Addendum to final report

CCT-K5: Comparison of local realizations of the ITS-90 between the silver point and 1700 °C using vacuum tungsten strip lamps as transfer standards

## Combining the final results to the KCDB

based on the Final report

Prepared by
E.W.M. van der Ham
NMi-Van Swinden Laboratorium B.V.
Delft
the Netherlands

May 2008

### One page summary

Four GEC high-stability vacuum Tungsten-strip lamps were used as transfer standards for radiance temperature measurements at specific currents corresponding to nominal temperatures  $T_{nom}$ . To shorten the measurement time significantly the set of transfer standards was split in two sets of two lamps for simultaneous comparisons in two loops. The pilot of each loop measured on both lamp sets in order to establish a linkage mechanism between both loops. The established link  $\Delta t$  ( $T_{nom}$ ) is described in detail in the CCT-K5 final report. Using this link the four key comparison reference values,  $T_R$  (K,  $T_{nom}$ ), were calculated at each nominal temperature  $T_{nom}$ , that is, one value for each individual lamp K, with  $K \in \{C864, C681\}$  or  $K \in \{C860, C864\}$  for loop 1 and loop 2, respectively.


The key comparison reference value  $T_R(k, T_{nom})$  is calculated for each nominal temperature  $T_{nom}$  on the basis of the median of measured radiance temperatures  $T_i(k, T_{nom})$ . For each participant i, either from loop 1 or loop 2, the difference from the key comparison reference value is calculated for each lamp k in the loop. As a result two differences are calculated,  $D_i(C564, T_{nom}) = T_i(C564, T_{nom}) - T_R(C564, T_{nom})$  and  $D_i(C681, T_{nom}) = T_i(C681, T_{nom}) - T_R(C681, T_{nom})$  for loop 1 and similarly for loop 2. For each difference the associated uncertainty  $U_i \equiv U_i(D_i(k, T_{nom}))$  is calculated based on the sum of squares of  $U_i(T_i(k, T_{nom}))$  and  $U_i(T_R(k, T_{nom}))$ .

Consequently four different combinations determine the difference between two participants. For participants in the same loop this condenses to the average of two differences. Again, for loop 1  $D_{ij}$  (loop 1,  $T_{nom}$ ) =  $\frac{1}{4}$   $D_{ij}$  (C564,  $T_{nom}$ ) +  $\frac{1}{4}$   $D_{ji}$  (C564,  $T_{nom}$ ) +  $\frac{1}{4}$   $D_{ij}$  (C681,  $T_{nom}$ ). =  $\frac{1}{4}$   $D_{ij}$  (C564,  $T_{nom}$ ) +  $\frac{1}{4}$   $D_{ij}$  (C681,  $T_{nom}$ ),

where  $D_{ij}(\mathbf{k}, T_{nom}) = D_i(\mathbf{k}, T_{nom}) - D_i(\mathbf{k}, T_{nom})$ .

The degree of equivalence of each temperature  $T_i$  with respect to the key comparison reference value is given by a pair of terms: the average difference  $D_i = \Sigma_k D_i(k, T_{nom}) / \Sigma_k 1$  and associated uncertainty  $U_i$ , its expanded uncertainty at 95 % confidence, both expressed in K. The uncertainty  $U_i$  includes the uncertainties in the original laboratory calibrations, the standard deviation of the average difference, the key comparison reference values and, when applicable, the link between the loops.

For the difference between two inter-loop participants the four different combinations cannot be condensed and even an additional term arises describing the difference  $\Delta t$  ( $T_{nom}$ ) between the two loops, that is  $D_{ij}$  (inter-loop,  $T_{nom}$ ) =  $D_i$  (loop i,  $T_{nom}$ ) -  $D_i$  (loop j,  $T_{nom}$ )  $\pm \Delta t$  ( $T_{nom}$ ), where the  $\pm$  sign relates to either adding or subtracting the differences between the loops depending on whether participant j is in loop 1 or loop 2. In this latter case the uncertainties will be larger as they include the uncertainty of established link U ( $\Delta t$  ( $T_{nom}$ )).




### **Contents**

One page summary	2
Resumé	
From 4-to-1 or 2-to-1?	4
Single entry solution for KCDB	5
Final results	6
Nominal temperature 961°C	6
Nominal temperature 1000°C	10
	14
Nominal temperature 1084°C	18
	22
Nominal temperature 1200°C	26
•	30
<u> •</u>	34
<u>.</u>	38
1	42
Nominal temperature 1700°C	

### Finalize CCT-K5 from two-to-one entry

#### Resumé

Two sets of two GEC high-stability vacuum tungsten-strip lamps were used as transfer standards during the comparison. The each set circulated in a loop of participants. The pilot and co-pilot of the comparison measured on both lamp sets. The main issue to solve in this comparison was to identify the linkage mechanism between two sets of measurements performed on four lamps.



The link was realized through the pilot institutes of each loop, as they are the only participants that measured all four lamps. Using the link the median as representative value for the KCRV was calculated on the basis of a single contribution from each institute for each nominal temperature.

Based on the calculated KCRV, one then can determine the differences from the participant entries with respect to the KCRV and, together with the KCRV, the difference between two participants. The KCRV is global, that is, one set of values as a function of the nominal temperature, without a lamp- or loop-dependence. The set of differences per participant  $\delta t_i(k,t_{nom})$ , however, is still specified either from loop 1 or from loop 2. The latter is important when calculating the uncertainty associated with the difference between two participants, as one has to take the uncertainty of the loop difference into account as well. Based on this argument, one can distinguish interloop and intraloop intercomparisons.

In the agreed draft B each participant has two entries per nominal temperature. From these entries the difference between two participants, either interloop or intraloop, was calculated. The result is in total 4 combinations of differences since both participants have measured 2 lamps.

#### From 4-to-1 or 2-to-1?

The issue raised is how to condense either the participant result to a single one or the difference between two participants to a single value. Very straight forward is to simply average out the four differences that can be calculated between two participants. This was brought forward at the CCT-WG5 meeting May 2007 and was unanimously accepted.

#### Single entry solution for KCDB

The differences between the participants through the KCRVs are already determined and reported in the agreed draft B. The results are presented in the tables. Now the average values are calculated accordingly to fill the KCDB. Based on the formulas from page 26 the average differences and associated uncertainties are calculated along the following route.

First, the average difference is calculated,

$$\overline{\Delta t_{nm}(t_{nom})} \equiv \sum_{\mathbf{k}} \Delta t_{nm}(\mathbf{k}, t_{nom}) / \sum_{\mathbf{k}} 1$$
, based on 2 and 4 differences for the intra- and the inter-loop, respectively.

The uncertainty associated with this average difference is composed of two items. One is based on the standard deviation associated with the average, stdev( $\Delta t_{nm}$  (k,  $t_{nom}$ ))/ $\sqrt{\sum_k 1}$ , and the other is based on the average of already calculated uncertainties,  $u(\Delta t_{nm}$  (k,  $t_{nom}$ )). The combination is calculated as:

$$u(\overline{\Delta t_{nm}(t_{nom})})^2 \equiv stdev^2(\Delta t_{nm}(k, t_{nom}))/\sum_k 1 + u(\overline{\Delta t_{nm}(k, t_{nom})})^2$$

More explicitly the average difference and uncertainty for each loop combination are calculated according to the following expressions:

+  $[\frac{1}{2}u(\Delta t_{nm} (C564, t_{nom}) + \frac{1}{2}u(\Delta t_{nm} (C681, t_{nom}))]^2$ .

#### Intra-loop in loop 1

$$\Delta t_{nm} (\text{loop 1}, t_{nom}) \equiv \frac{1}{2} \Delta t_{nm} (\text{C564}, t_{nom}) + \frac{1}{2} \Delta t_{nm} (\text{C681}, t_{nom}) \text{ and}$$
  
 $u^2 (\Delta t_{nm} (\text{loop 1}, t_{nom})) \equiv \frac{1}{2} \left[ \text{stdev}(\Delta t_{nm} (\text{C564}, t_{nom}), \Delta t_{nm} (\text{C681}, t_{nom})) \right]^2$ 

#### Intra-loop in loop 2

$$\Delta t_{nm} (\text{loop 2}, t_{nom}) \equiv \frac{1}{2} \Delta t_{nm} (\text{C860}, t_{nom}) + \frac{1}{2} \Delta t_{nm} (\text{C864}, t_{nom})$$
 and

$$\begin{aligned} \mathbf{u}^2(\Delta t_{nm}(\text{loop2},\,\mathbf{t}_{\text{nom}})) &\equiv \frac{1}{2}[\text{stdev}(\Delta t_{nm}\,(\text{C860},\,t_{nom}),\,\Delta t_{nm}\,(\text{C864},\,t_{nom}))]^2 \\ &\quad + \left[\frac{1}{2}\mathbf{u}(\Delta t_{nm}\,(\text{C860},\,t_{nom})) + \frac{1}{2}\mathbf{u}(\Delta t_{nm}\,(\text{C864},\,t_{nom}))\right]^2. \end{aligned}$$

#### Inter-loop between loops 1 and 2

$$\Delta t_{nm} (\text{loop } 1/2, t_{nom}) \equiv {}^{1}\!\!/4 \Delta t_{nm} (\text{C}564/\text{C}860, t_{nom}) + {}^{1}\!\!/4 \Delta t_{nm} (\text{C}564/\text{C}864, t_{nom}) + {}^{1}\!\!/4 \Delta t_{nm} (\text{C}681/\text{C}864, t_{nom}) + {}^{1}\!\!/4 \Delta t_{nm} (\text{C}681/\text{C}864, t_{nom})$$
and

$$\begin{aligned} \mathbf{u}^2(\Delta t_{nm}(\text{loop1/2},\,\mathbf{t}_{\text{nom}})) &\equiv \frac{1}{4}[\text{stdev}(\Delta t_{nm}\,(\text{C564/C860},\,t_{nom}),\,\Delta t_{nm}\,(\text{C564/C864},\,t_{nom}),\\ &\quad \Delta t_{nm}\,(\text{C681/C860},\,t_{nom}),\,\Delta t_{nm}\,(\text{C681/C864},\,t_{nom}))\,\big]^2\\ &\quad + \big[\frac{1}{4}\mathbf{u}(\Delta t_{nm}\,(\text{C564/C860},\,t_{nom})) + \frac{1}{4}\mathbf{u}(\Delta t_{nm}\,(\text{C564/C864},\,t_{nom}))\\ &\quad + \frac{1}{4}\mathbf{u}(\Delta t_{nm}\,(\text{C681/C860},\,t_{nom})) + \frac{1}{4}\mathbf{u}(\Delta t_{nm}\,(\text{C681/C864},\,t_{nom})\big]^2. \end{aligned}$$

Although the pilot laboratories NPL and VSL have measured the lamp sets several times during the intercomparison only their first contribution to the set of measurement data is used for the KCDB database, NPL1 and VSL1 respectively. The results are presented below.

## Final results


## Nominal temperature 961°C

Lab, S/N j ∏			VSL		N	PL	CSI	RO	KR	ISS	NI	М	NMC	
	D; L	i	<b>D</b> <sub>ij</sub> /	U <sub>ij</sub> K	<b>D</b> <sub>ij</sub> /	υ <sub>ij</sub> Κ	D <sub>ij</sub>	U <sub>ij</sub> K	<b>D</b> <sub>ij</sub> /	υ <sub>ij</sub> Κ	<b>D</b> <sub>ij</sub> /	<i>υ<sub>ij</sub></i> κ	<b>D</b> <sub>ij</sub> /	U <sub>ij</sub> K
VSL	<b>-0.010</b> 0.2	27			0.113	0.408	0.085	0.205	0.040	0.287	-0.065	0.300	0.470	0.377
NPL	<b>-0.124</b> 0.3	10	-0.113	0.408			-0.028	0.359	-0.073	0.407	-0.178	0.421	0.357	0.465
CSIRO	<b>-0.095</b> 0.1	23	-0.085	0.205	0.028	0.359			-0.045	0.203	-0.150	0.226	0.385	0.328
KRISS	<b>-0.050</b> 0.2	55	-0.040	0.287	0.073	0.407	0.045	0.203			-0.105	0.289	0.430	0.408
NIM	<b>0.055</b> 0.2	59	0.065	0.300	0.178	0.421	0.150	0.226	0.105	0.289			0.535	0.403
NMC	<b>-0.480</b> 0.3	19	-0.470	0.377	-0.357	0.465	-0.385	0.328	-0.430	0.427	-0.535	0.403		
NRLM	<b>0.010</b> 0.2	63	0.020	0.312	0.133	0.429	0.105	0.243	0.060	0.329	-0.045	0.329	0.490	0.399
VNIIM	<b>0.555</b> 0.3	33	0.565	0.380	0.678	0.477	0.650	0.328	0.605	0.398	0.500	0.398	1.035	0.444
NIST														
NRC	<b>0.311</b> 0.2	62	0.322	0.378	0.435	0.372	0.407	0.325	0.362	0.378	0.257	0.392	0.792	0.440
CENAM	<b>0.091</b> 0.4	33	0.102	0.512	0.215	0.507	0.187	0.474	0.142	0.512	0.037	0.522	0.572	0.559
INM	<b>-0.294</b> 0.3	50	-0.283	0.410	-0.170	0.400	-0.198	0.360	-0.243	0.409	-0.348	0.423	0.187	0.467
IMGC	<b>0.111</b> 0.1	71	0.122	0.324	0.235	0.337	0.207	0.259	0.162	0.323	0.057	0.339	0.592	0.394
РТВ	<b>0.051</b> 0.1	69	0.062	0.420	0.175	0.442	0.147	0.372	0.102	0.420	-0.003	0.433	0.532	0.476

Lab, S/N j ∏			NRLM		VNIIM		NIST		NR	C	CENAM		INM	
Å.	<b>D</b> <sub>i</sub>	U <sub>i</sub> K	D <sub>ij</sub>	U <sub>ij</sub> K	<b>D</b> <sub>ij</sub> /	U <sub>ij</sub> K								
VSL	-0.010	0.227	-0.020	0.312	-0.565	0.380			-0.322	0.378	-0.102	0.512	0.283	0.410
NPL	-0.124	0.310	-0.133	0.429	-0.678	0.477			-0.435	0.372	-0.215	0.091	0.170	0.400
CSIRO	-0.095	0.123	-0.105	0.243	-0.650	0.328			-0.407	0.325	-0.187	0.474	0.198	0.360
KRISS	-0.050	0.255	-0.060	0.316	-0.605	0.395			-0.362	0.378	-0.142	0.512	0.243	0.409
NIM	0.055	0.259	0.045	0.329	-0.500	0.398			-0.257	0.392	-0.037	0.522	0.348	0.423
NMC	-0.480	0.319	-0.490	0.399	-1.035	0.444			-0.792	0.440	-0.572	0.559	-0.187	0.467
NRLM	0.010	0.263			-0.545	0.403			-0.302	0.401	-0.082	0.529	0.303	0.431
VNIIM	0.555	0.333	0.545	0.403					0.243	0.452	0.463	0.569	0.848	0.478
NIST														
NRC	0.311	0.262	0.302	0.401	-0.243	0.452					0.220	0.484	0.605	0.391
CENAM	0.091	0.433	0.082	0.529	-0.463	0.569			-0.220	0.484			0.385	0.521
INM	-0.294	0.350	-0.303	0.431	-0.848	0.478			-0.605	0.391	-0.385	0.521		
IMGC	0.111	0.171	0.102	0.350	-0.443	0.407			-0.200	0.289	0.020	0.450	0.405	0.383
РТВ	0.051	0.169	0.042	0.441	-0.503	0.488			-0.260	0.402	-0.040	0.530	0.345	0.489

Lab, S/N j ∏			IMC	GC	PΊ	TB
JĮ	$D_i$	<b>U</b> <sub>i</sub>	$D_{ii}$	U <sub>ij</sub>	D <sub>ii</sub>	U <sub>ii</sub>
V	1	K	1	K	1	K
VSL	-0.010	0.227	-0.122	0.324	-0.062	0.420
NPL	-0.124	0.310	-0.235	0.337	-0.175	0.442
CSIRO	-0.095	0.123	-0.207	0.259	-0.147	0.372
KRISS	-0.050	0.255	-0.162	0.323	-0.102	0.420
NIM	0.055	0.259	-0.057	0.339	0.003	0.433
NMC	-0.480	0.319	-0.592	0.394	-0.532	0.476
NRLM	0.010	0.263	-0.102	0.350	-0.042	0.441
VNIIM	0.555	0.333	0.443	0.407	0.503	0.488
NIST						
NRC	0.311	0.262	0.200	0.289	0.260	0.402
CENAM	0.091	0.433	-0.020	0.450	0.040	0.530
INM	-0.294	0.350	-0.405	0.383	-0.345	0.489
IMGC	0.111	0.171			0.060	0.333
РТВ	0.051	0.169	-0.060	0.333		

CCT-K5 : Nominal temperature,  $T_{90}$  = 1234 K Degrees of equivalence,  $D_i$ , and expanded uncertainties (k = 2),  $U_i$ , expressed in K




# Nominal temperature 1000°C

Lab, S/N j ∏		\ \ \ \ \ \ \ \ \ \ \ \ \ \ \ \ \ \ \	VSL		PL	CSI	RO	KRI	ISS	NIM		NN	<b>IC</b>
	<i>D</i> ; <i>U</i> ; /K	D <sub>ij</sub>	U <sub>ij</sub> / K	D <sub>ij</sub>	υ <sub>ij</sub> Κ	<b>D</b> <sub>ij</sub> /	U <sub>ij</sub> K	<b>D</b> <sub>ij</sub> /	U <sub>ij</sub> K	<b>D</b> <sub>ij</sub> /	<i>υ<sub>ij</sub></i> Κ	<b>D</b> <sub>ij</sub> /	U <sub>ij</sub> K
VSL	<b>-0.002</b> 0.22			0.113	0.386	0.055	0.206	0.020	0.281	-0.090	0.320	0.500	0.375
NPL	<b>-0.114</b> 0.29	-0.113	0.386			-0.058	0.334	-0.093	0.382	-0.203	0.413	0.388	0.446
CSIRO	<b>-0.057</b> 0.11	-0.055	0.206	0.058	0.334			-0.035	0.192	-0.145	0.247	0.445	0.328
KRISS	<b>-0.022</b> 0.23	-0.020	0.281	0.093	0.382	0.035	0.192			-0.110	0.300	0.480	0.394
NIM	<b>0.088</b> 0.27	0.090	0.320	0.203	0.413	0.145	0.247	0.110	0.300			0.590	0.420
NMC	<b>-0.502</b> 0.31	-0.500	0.375	-0.388	0.446	-0.445	0.328	-0.480	0.424	-0.590	0.420		
NRLM	<b>0.053</b> 0.25	0.055	0.312	0.168	0.408	0.110	0.244	0.075	0.345	-0.035	0.345	0.555	0.398
VNIIM	<b>0.333</b> 0.34	0.335	0.408	0.448	0.463	0.390	0.369	0.355	0.459	0.245	0.459	0.835	0.442
NIST	<b>0.151</b> 0.48	0.153	0.548	0.265	0.561	0.208	0.513	0.173	0.546	0.063	0.568	0.653	0.592
NRC	<b>0.251</b> 0.25	0.253	0.358	0.365	0.369	0.308	0.301	0.273	0.354	0.163	0.387	0.753	0.422
CENAM	<b>0.051</b> 0.44	0.053	0.514	0.165	0.522	0.108	0.476	0.073	0.511	-0.038	0.534	0.553	0.560
INM	<b>-0.304</b> 0.36	-0.303	0.406	-0.190	0.435	-0.248	0.357	-0.283	0.402	-0.393	0.431	0.198	0.463
IMGC	<b>0.111</b> 0.17	0.113	0.310	0.225	0.328	0.168	0.241	0.133	0.305	0.023	0.342	0.613	0.381
PTB	<b>-0.054</b> 0.18	-0.053	0.416	0.060	0.437	0.003	0.369	-0.033	0.413	-0.143	0.442	0.448	0.473

Lab, S/N j ∏			NRLM		VN	IIM	NIS	ST T	NF	RC .	CENAM		INM	
ĮĮ,	$D_i$	U <sub>i</sub>	$D_{ij}$	U <sub>ij</sub>	D <sub>ij</sub>	U <sub>ij</sub>								
	1	K	1	K	1	K	1	K	1	K	1	K	1	K
VSL	-0.002	0.221	-0.055	0.312	-0.335	0.408	-0.153	0.548	-0.253	0.358	-0.053	0.514	0.303	0.406
NPL	-0.114	0.294	-0.168	0.408	-0.448	0.463	-0.265	0.561	-0.365	0.369	-0.165	0.077	0.190	0.435
CSIRO	-0.057	0.116	-0.110	0.244	-0.390	0.369	-0.208	0.513	-0.308	0.301	-0.108	0.476	0.248	0.357
KRISS	-0.022	0.235	-0.075	0.309	-0.355	0.436	-0.173	0.546	-0.273	0.354	-0.073	0.511	0.283	0.402
NIM	0.088	0.279	0.035	0.345	-0.245	0.459	-0.063	0.568	-0.163	0.387	0.038	0.534	0.393	0.431
NMC	-0.502	0.314	-0.555	0.398	-0.835	0.442	-0.653	0.592	-0.753	0.422	-0.553	0.560	-0.198	0.463
NRLM	0.053	0.258			-0.280	0.431	-0.098	0.564	-0.198	0.382	0.003	0.530	0.358	0.427
VNIIM	0.333	0.345	0.280	0.431			0.183	0.605	0.083	0.440	0.283	0.574	0.638	0.479
NIST	0.151	0.487	0.098	0.564	-0.183	0.605			-0.100	0.542	0.100	0.655	0.455	0.612
NRC	0.251	0.255	0.198	0.382	-0.083	0.440	0.100	0.542			0.200	0.501	0.555	0.413
CENAM	0.051	0.449	-0.003	0.530	-0.283	0.574	-0.100	0.655	-0.200	0.501			0.355	0.554
INM	-0.304	0.366	-0.358	0.427	-0.638	0.479	-0.455	0.612	-0.555	0.413	-0.355	0.554		
IMGC	0.111	0.179	0.058	0.337	-0.223	0.401	-0.040	0.506	-0.140	0.294	0.060	0.472	0.415	0.405
PTB	-0.054	0.183	-0.108	0.437	-0.388	0.489	-0.205	0.578	-0.305	0.411	-0.105	0.552	0.250	0.509

Lab, S/N j ∏			IM	GC	PI	В
ĮĮ,	$D_i$	U <sub>i</sub>	$D_{ij}$	U <sub>ij</sub>	$D_{ij}$	U <sub>ij</sub>
V	1	K	1	K	1	K
VSL	-0.002	0.221	-0.113	0.310	0.053	0.416
NPL	-0.114	0.294	-0.225	0.328	-0.060	0.437
CSIRO	-0.057	0.116	-0.168	0.241	-0.003	0.369
KRISS	-0.022	0.235	-0.133	0.305	0.033	0.413
NIM	0.088	0.279	-0.023	0.342	0.143	0.442
NMC	-0.502	0.314	-0.613	0.381	-0.448	0.473
NRLM	0.053	0.258	-0.058	0.337	0.108	0.437
VNIIM	0.333	0.345	0.223	0.401	0.388	0.489
NIST	0.151	0.487	0.040	0.506	0.205	0.578
NRC	0.251	0.255	0.140	0.294	0.305	0.411
CENAM	0.051	0.449	-0.060	0.472	0.105	0.552
INM	-0.304	0.366	-0.415	0.405	-0.250	0.509
IMGC	0.111	0.179			0.165	0.360
PTB	-0.054	0.183	-0.165	0.360		

CCT-K5 : Nominal temperature,  $T_{90}$  = 1273 K Degrees of equivalence,  $D_i$ , and expanded uncertainties (k = 2),  $U_i$ , expressed in K



# Nominal temperature 1064°C


Lab, S/N i <del>□</del>

Lab, S/N j ∏			VSL		N	PL	CSI	RO	KRI	ISS	NIM		NMC	
	D <sub>i</sub>	U <sub>i</sub> K	D <sub>ij</sub>	U <sub>ij</sub> K										
VSL	-0.033	0.257			0.149	0.410	0.005	0.227	-0.040	0.313			0.505	0.408
NPL	-0.181	0.329	-0.149	0.410			-0.144	0.351	-0.189	0.409			0.357	0.473
CSIRO	-0.038	0.155	-0.005	0.227	0.144	0.351			-0.045	0.214			0.500	0.360
KRISS	0.007	0.279	0.040	0.313	0.189	0.409	0.045	0.214					0.545	0.439
NIM														
NMC	-0.538	0.350	-0.505	0.408	-0.357	0.473	-0.500	0.360	-0.545	0.225				
NRLM	0.062	0.291	0.095	0.341	0.244	0.433	0.100	0.266	0.055	0.000			0.600	0.429
VNIIM	-0.093	0.365	-0.060	0.428	0.089	0.477	-0.055	0.387	-0.100	0.000			0.445	0.456
NIST	0.134	0.536	0.167	0.590	0.315	0.600	0.172	0.551	0.127	0.589			0.672	0.635
NRC	0.149	0.287	0.182	0.381	0.330	0.399	0.187	0.317	0.142	0.379			0.687	0.447
CENAM	0.104	0.515	0.137	0.573	0.285	0.585	0.142	0.531	0.097	0.571			0.642	0.618
INM	-0.316	0.401	-0.284	0.441	-0.135	0.466	-0.279	0.387	-0.324	0.439			0.222	0.500
IMGC	0.129	0.206	0.162	0.322	0.310	0.354	0.167	0.243	0.122	0.320			0.667	0.399
РТВ	-0.046	0.205	-0.014	0.441	0.135	0.465	-0.009	0.386	-0.054	0.439			0.492	0.499

Lab, S/N j ∏			NRLM		VN	IIM	NIS	ST T	NF	RC .	CENAM		INM	
Įļ,	$D_i$	<b>U</b> <sub>i</sub>	D ij	U <sub>ij</sub>	$D_{ij}$	U <sub>ij</sub>	D <sub>ij</sub>	U <sub>ij</sub>						
	1	K	1	K	1	K	1	K	1	K	1	K	1	K
VSL	-0.033	0.257	-0.095	0.341	-0.060	0.428	-0.167	0.590	-0.182	0.381	-0.137	0.573	0.284	0.441
NPL	-0.181	0.329	-0.244	0.433	0.089	0.477	-0.315	0.600	-0.330	0.399	-0.285	0.121	0.135	0.466
CSIRO	-0.038	0.155	-0.100	0.266	-0.055	0.387	-0.172	0.551	-0.187	0.317	-0.142	0.531	0.279	0.387
KRISS	0.007	0.279	-0.055	0.344	-0.100	0.470	-0.127	0.589	-0.142	0.379	-0.097	0.571	0.324	0.439
NIM														
NMC	-0.538	0.350	-0.600	0.429	0.445	0.456	-0.672	0.635	-0.687	0.447	-0.642	0.618	-0.222	0.500
NRLM	0.062	0.291			-0.155	0.448	-0.072	0.606	-0.087	0.406	-0.042	0.589	0.379	0.462
VNIIM	-0.093	0.365	-0.155	0.448			-0.227	0.638	-0.242	0.451	-0.197	0.621	0.224	0.503
NIST	0.134	0.536	0.072	0.606	0.227	0.638			-0.015	0.582	0.030	0.722	0.450	0.632
NRC	0.149	0.287	0.087	0.406	0.242	0.451	0.015	0.582			0.045	0.564	0.465	0.456
CENAM	0.104	0.515	0.042	0.589	0.197	0.621	-0.030	0.722	-0.045	0.564			0.420	0.627
INM	-0.316	0.401	-0.379	0.462	-0.224	0.503	-0.450	0.632	-0.465	0.456	-0.420	0.627		
IMGC	0.129	0.206	0.067	0.351	0.222	0.404	-0.005	0.552	-0.020	0.311	0.025	0.528	0.445	0.440
PTB	-0.046	0.205	-0.109	0.462	0.047	0.502	-0.180	0.629	-0.195	0.433	-0.150	0.608	0.270	0.534

Lab, S/N j ∏			IM	GC	PT	В
][	D,	U,	D <sub>ij</sub>	U <sub>ii</sub>	D <sub>ij</sub>	U <sub>ii</sub>
V		K		K ຶ	_	K
VSL	-0.033	0.257	-0.162	0.322	0.014	0.441
NPL	-0.181	0.329	-0.310	0.354	-0.135	0.465
CSIRO	-0.038	0.155	-0.167	0.243	0.009	0.386
KRISS	0.007	0.279	-0.122	0.320	0.054	0.439
NIM	0.000	0.000	0.000	0.000	0.000	0.000
NMC	-0.538	0.350	-0.667	0.399	-0.492	0.499
NRLM	0.062	0.291	-0.067	0.351	0.109	0.462
VNIIM	-0.093	0.365	-0.222	0.404	-0.047	0.502
NIST	0.134	0.536	0.005	0.552	0.180	0.629
NRC	0.149	0.287	0.020	0.311	0.195	0.433
CENAM	0.104	0.515	-0.025	0.528	0.150	0.608
INM	-0.316	0.401	-0.445	0.440	-0.270	0.534
IMGC	0.129	0.206			0.175	0.376
РТВ	-0.046	0.205	-0.175	0.376		

CCT-K5 : Nominal temperature,  $T_{90}$  = 1337 K Degrees of equivalence,  $D_i$ , and expanded uncertainties (k = 2),  $U_i$ , expressed in K




# Nominal temperature 1084°C

Lab, S/N j ∏			VSL		N	PL	CSI	RO	KRI	ISS	NIM		NN	<b>IC</b>
	<b>D</b> ;	U <sub>i</sub> K	D <sub>ij</sub>	U <sub>ij</sub> K	D <sub>ij</sub>	U <sub>ij</sub> K	D <sub>ij</sub>	U <sub>ij</sub> K	D <sub>ij</sub> /	<i>υ<sub>ij</sub></i> Κ	D <sub>ij</sub>	<i>U<sub>ij</sub></i> К	D <sub>ij</sub> /	U <sub>ij</sub> K
VSL	-0.013	0.254			0.188	0.405	0.000	0.226	-0.040	0.284			0.550	0.409
NPL	-0.201	0.323	-0.188	0.405			-0.188	0.344	-0.228	0.381			0.363	0.467
CSIRO	-0.013	0.148	0.000	0.226	0.188	0.344			-0.040	0.172			0.550	0.360
KRISS	0.027	0.238	0.040	0.284	0.228	0.381	0.040	0.172					0.590	0.415
NIM														
NMC	-0.563	0.348	-0.550	0.409	-0.363	0.467	-0.550	0.360	-0.590	0.210				
NRLM	0.097	0.288	0.110	0.341	0.298	0.427	0.110	0.266	0.070	0.000			0.660	0.429
VNIIM	-0.103	0.408	-0.090	0.464	0.098	0.511	-0.090	0.422	-0.130	0.000			0.460	0.497
NIST														
NRC	0.124	0.302	0.138	0.389	0.325	0.410	0.138	0.326	0.098	0.366			0.688	0.454
CENAM	0.099	0.532	0.113	0.586	0.300	0.600	0.113	0.546	0.073	0.570			0.663	0.632
INM	-0.306	0.403	-0.293	0.437	-0.105	0.476	-0.293	0.381	-0.333	0.416			0.258	0.496
IMGC	0.124	0.217	0.138	0.326	0.325	0.358	0.138	0.246	0.098	0.297			0.688	0.401
РТВ	-0.001	0.213	0.013	0.435	0.200	0.462	0.013	0.380	-0.028	0.415			0.563	0.495

Lab, S/N j ∏			NRLM		VN	IIM	NI	ST	NR	C	CEN	AM	IN	М
ĮĮ,	$D_i$	U <sub>i</sub>	D ij	U <sub>ij</sub>	D <sub>ij</sub>	U <sub>ij</sub>								
v	1	K	1	K	1	K	1	K	1	K	1	K	1	K
VSL	-0.013	0.254	-0.110	0.341	0.090	0.464			-0.138	0.389	-0.113	0.586	0.293	0.437
NPL	-0.201	0.323	-0.298	0.427	-0.098	0.511			-0.325	0.410	-0.300	0.101	0.105	0.476
CSIRO	-0.013	0.148	-0.110	0.266	0.090	0.422			-0.138	0.326	-0.113	0.546	0.293	0.381
KRISS	0.027	0.238	-0.070	0.318	0.130	0.472			-0.098	0.366	-0.073	0.570	0.333	0.416
NIM														
NMC	-0.563	0.348	-0.660	0.429	-0.460	0.497			-0.688	0.454	-0.663	0.632	-0.258	0.496
NRLM	0.097	0.288			0.200	0.481			-0.028	0.413	-0.003	0.602	0.403	0.458
VNIIM	-0.103	0.408	-0.200	0.481					-0.228	0.499	-0.203	0.664	0.203	0.538
NIST														
NRC	0.124	0.302	0.028	0.413	0.228	0.499					0.025	0.591	0.430	0.470
CENAM	0.099	0.532	0.003	0.602	0.203	0.664			-0.025	0.591			0.405	0.641
INM	-0.306	0.403	-0.403	0.458	-0.203	0.538			-0.430	0.470	-0.405	0.641		
IMGC	0.124	0.217	0.028	0.354	0.228	0.452			0.000	0.338	0.025	0.554	0.430	0.450
РТВ	-0.001	0.213	-0.098	0.457	0.103	0.537			-0.125	0.446	-0.100	0.626	0.305	0.542

Lab, S/N j	Π.			IMO	GC	PΊ	В
	][	$D_i$	U <sub>i</sub>	D <sub>ij</sub>	U <sub>ij</sub>	$D_{ij}$	U <sub>ij</sub>
	<b>V</b>	1	K	1	K	1	K
VSL		-0.013	0.254	-0.138	0.326	-0.013	0.435
NPL		-0.201	0.323	-0.325	0.358	-0.200	0.462
CSIRO		-0.013	0.148	-0.138	0.246	-0.013	0.380
KRISS		0.027	0.238	-0.098	0.297	0.028	0.415
NIM							
NMC		-0.563	0.348	-0.688	0.401	-0.563	0.495
NRLM		0.097	0.288	-0.028	0.354	0.098	0.457
VNIIM		-0.103	0.408	-0.228	0.452	-0.103	0.537
NIST							
NRC		0.124	0.302	0.000	0.338	0.125	0.446
CENAM		0.099	0.532	-0.025	0.554	0.100	0.626
INM		-0.306	0.403	-0.430	0.450	-0.305	0.542
IMGC		0.124	0.217			0.125	0.385
РТВ		-0.001	0.213	-0.125	0.385		

CCT-K5 : Nominal temperature,  $T_{90}$  = 1357 K Degrees of equivalence,  $D_i$ , and expanded uncertainties (k = 2),  $U_i$ , expressed in K



# Nominal temperature 1100°C


Lab, S/N i <del>□</del>

Lab, S/N j ∏			VSL		N	PL	CSI	RO	KR	ISS	NI	М	NN	<b>MC</b>
	<b>D</b> ;	U <sub>i</sub> K	<b>D</b> <sub>ij</sub> /	D <sub>ij</sub> U <sub>ij</sub> / K		U <sub>ij</sub> K	D <sub>ij</sub>	U <sub>ij</sub> K	<b>D</b> <sub>ij</sub> /	U <sub>ij</sub> K	<b>D</b> <sub>ij</sub> /	U <sub>ij</sub> K	<b>D</b> <sub>ij</sub> /	U <sub>ij</sub> K
VSL	-0.017	0.250			0.278	0.412	-0.005	0.227	-0.025	0.304	-0.175	0.371	0.560	0.406
NPL	-0.295	0.342	-0.278	0.412			-0.283	0.353	-0.303	0.406	-0.453	0.455	0.282	0.474
CSIRO	-0.012	0.145	0.005	0.227	0.283	0.353			-0.020	0.206	-0.170	0.292	0.565	0.357
KRISS	0.008	0.255	0.025	0.304	0.303	0.406	0.020	0.206			0.150	0.346	0.585	0.420
NIM	0.158	0.339	0.175	0.371	0.453	0.455	0.170	0.292	0.150	0.346			0.735	0.481
NMC	-0.577	0.345	-0.560	0.406	-0.282	0.474	-0.565	0.357	-0.585	0.464	-0.735	0.481		
NRLM	0.103	0.287	0.120	0.341	0.398	0.435	0.115	0.265	0.095	0.396	-0.055	0.396	0.680	0.429
VNIIM	-0.197	0.437	-0.180	0.484	0.098	0.546	-0.185	0.443	-0.205	0.545	-0.355	0.545	0.380	0.528
NIST	0.235	0.556	0.252	0.600	0.530	0.641	0.247	0.561	0.227	0.595	0.077	0.630	0.812	0.644
NRC	0.075	0.300	0.092	0.380	0.370	0.427	0.087	0.316	0.067	0.374	-0.083	0.427	0.652	0.447
CENAM	0.070	0.531	0.087	0.581	0.365	0.612	0.082	0.540	0.062	0.576	-0.088	0.612	0.647	0.625
INM	-0.320	0.424	-0.303	0.447	-0.025	0.499	-0.308	0.392	-0.328	0.442	-0.478	0.486	0.257	0.504
IMGC	0.100	0.216	0.117	0.316	0.395	0.382	0.112	0.233	0.092	0.308	-0.058	0.370	0.677	0.393
РТВ	-0.035	0.211	-0.018	0.444	0.260	0.492	-0.023	0.389	-0.043	0.438	-0.193	0.483	0.542	0.501

Lab, S/N j ∏			NRLM		VN	IIM	NIS	ST TE	NR	RC .	CEN	IAM	IN	М
Å.	<b>D</b> ;	U <sub>i</sub> K	D <sub>ij</sub>	U <sub>ij</sub> K	D <sub>ij</sub>	U <sub>ij</sub> K	<b>D</b> <sub>ij</sub> /	U <sub>ij</sub> K	<b>D</b> <sub>ij</sub> /	υ <sub>ij</sub> Κ	<b>D</b> <sub>ij</sub> /	U <sub>ij</sub> K	<b>D</b> <sub>ij</sub> /	U <sub>ij</sub> K
VSL	-0.017	0.250	-0.120	0.341	0.180	0.484	-0.252	0.600	-0.092	0.380	-0.087	0.581	0.303	0.447
NPL	-0.295	0.342	-0.398	0.435	-0.098	0.546	-0.530	0.641	-0.370	0.427	-0.365	0.103	0.025	0.499
CSIRO	-0.012	0.145	-0.115	0.265	0.185	0.443	-0.247	0.561	-0.087	0.316	-0.082	0.540	0.308	0.392
KRISS	0.008	0.255	-0.095	0.334	0.205	0.493	-0.227	0.595	-0.067	0.374	-0.062	0.576	0.328	0.442
NIM	0.158	0.339	0.055	0.396	0.355	0.545	-0.077	0.630	0.083	0.427	0.088	0.612	0.478	0.486
NMC	-0.577	0.345	-0.680	0.429	-0.380	0.528	-0.812	0.644	-0.652	0.447	-0.647	0.625	-0.257	0.504
NRLM	0.103	0.287			0.300	0.504	-0.132	0.615	0.028	0.405	0.033	0.597	0.423	0.468
VNIIM	-0.197	0.437	-0.300	0.504			-0.432	0.698	-0.272	0.522	-0.267	0.682	0.123	0.572
NIST	0.235	0.556	0.132	0.615	0.432	0.698			0.160	0.615	0.165	0.756	0.555	0.707
NRC	0.075	0.300	-0.028	0.405	0.272	0.522	-0.160	0.615			0.005	0.591	0.395	0.490
CENAM	0.070	0.531	-0.033	0.597	0.267	0.682	-0.165	0.756	-0.005	0.591			0.390	0.656
INM	-0.320	0.424	-0.423	0.468	-0.123	0.572	-0.555	0.707	-0.395	0.490	-0.390	0.656		
IMGC	0.100	0.216	-0.003	0.345	0.297	0.477	-0.135	0.570	0.025	0.339	0.030	0.554	0.420	0.476
PTB	-0.035	0.211	-0.138	0.465	0.162	0.569	-0.270	0.650	-0.110	0.460	-0.105	0.635	0.285	0.566

Lab, S/N j ∏			IM	GC	P1	В
][	D,	U <sub>i</sub>	D ij	U <sub>ii</sub>	$D_{ij}$	U <sub>ii</sub>
V	1	K		K	_	K
VSL	-0.017	0.250	-0.117	0.316	0.018	0.444
NPL	-0.295	0.342	-0.395	0.382	-0.260	0.492
CSIRO	-0.012	0.145	-0.112	0.233	0.023	0.389
KRISS	0.008	0.255	-0.092	0.308	0.043	0.438
NIM	0.158	0.339	0.058	0.370	0.193	0.483
NMC	-0.577	0.345	-0.677	0.393	-0.542	0.501
NRLM	0.103	0.287	0.003	0.345	0.138	0.465
VNIIM	-0.197	0.437	-0.297	0.477	-0.162	0.569
NIST	0.235	0.556	0.135	0.570	0.270	0.650
NRC	0.075	0.300	-0.025	0.339	0.110	0.460
CENAM	0.070	0.531	-0.030	0.554	0.105	0.635
INM	-0.320	0.424	-0.420	0.476	-0.285	0.566
IMGC	0.100	0.216			0.135	0.402
PTB	-0.035	0.211	-0.135	0.402		

CCT-K5 : Nominal temperature,  $T_{90}$  = 1373 K Degrees of equivalence,  $D_i$ , and expanded uncertainties (k = 2),  $U_i$ , expressed in K




# Nominal temperature 1200°C

Lab, S/N j ⊓			V	SL	N	DI .	CSI	PΩ	KRI	22	l NI	M	NM	MC .
Lub, Girt j	<i>D</i> ; / K	<i>U<sub>i</sub></i>	D ij	U <sub>ij</sub> K	D <sub>ij</sub>	U <sub>ij</sub> K								
VSL	-0.055	0.262			0.322	0.478	-0.030	0.244	-0.055	0.343	-0.140	0.421	0.650	0.458
NPL	-0.377	0.411	-0.322	0.478			-0.352	0.416	-0.377	0.481	-0.462	0.532	0.328	0.550
CSIRO	-0.025	0.122	0.030	0.244	0.352	0.416			-0.025	0.246	-0.110	0.340	0.680	0.400
KRISS	0.000	0.278	0.055	0.343	0.377	0.481	0.025	0.246			-0.085	0.407	0.705	0.479
NIM	0.085	0.379	0.140	0.421	0.462	0.532	0.110	0.340	0.085	0.407		,		0.557
NMC	-0.705	0.384	-0.650	0.458	-0.328	0.550	-0.680	0.400	-0.705	0.524	-0.790	0.557		
NRLM	0.145	0.320	0.200	0.384	0.522	0.511	0.170	0.303	0.145	0.455	0.060	0.455	0.850	0.496
VNIIM	-0.445	0.686	-0.390	0.724	-0.068	0.795	-0.420	0.687	-0.445	0.778	-0.530	0.778	0.260	0.773
NIST	0.183	0.632	0.238	0.674	0.560	0.745	0.208	0.632	0.183	0.676	0.098	0.714	0.888	0.728
NRC	-0.117	0.333	-0.062	0.413	0.260	0.512	-0.092	0.340	-0.117	0.416	-0.202	0.476	0.588	0.495
CENAM	0.003	0.628	0.058	0.673	0.380	0.738	0.028	0.631	0.003	0.675	-0.082	0.713	0.708	0.727
INM	-0.427	0.445	-0.372	0.477	-0.050	0.584	-0.402	0.415	-0.427	0.480	-0.512	0.532	0.278	0.550
IMGC	0.048	0.222	0.103	0.330	0.425	0.450	0.073	0.231	0.048	0.334	-0.037	0.404	0.753	0.428
РТВ	0.033	0.220	0.088	0.461	0.410	0.552	0.058	0.397	0.033	0.464	-0.052	0.518	0.738	0.536

Lab, S/N j ∏		NE	NRLM		IIM	NI	ST	NF	RC .	CEN	IAM	IN	М
ŢĻ	$D_i$ $U_i$	D ij	U <sub>ij</sub>	D <sub>ij</sub>	U <sub>ij</sub>								
	/ <b>K</b>		/ <b>K</b>		K	1	K	1	K	1	K	1	K
VSL	<b>-0.055</b> 0.26	-0.200	0.384	0.390	0.724	-0.238	0.674	0.062	0.413	-0.058	0.673	0.372	0.477
NPL	<b>-0.377</b> 0.41	-0.522	0.511	0.068	0.795	-0.560	0.745	-0.260	0.512	-0.380	0.174	0.050	0.584
CSIRO	<b>-0.025</b> 0.12	-0.170	0.303	0.420	0.687	-0.208	0.632	0.092	0.340	-0.028	0.631	0.402	0.415
KRISS	<b>0.000</b> 0.27	-0.145	0.387	0.445	0.732	-0.183	0.676	0.117	0.416	-0.003	0.675	0.427	0.480
NIM	<b>0.085</b> 0.37	-0.060	0.455	0.530	0.778	-0.098	0.714	0.202	0.476	0.082	0.713	0.512	0.532
NMC	<b>-0.705</b> 0.38	-0.850	0.496	-0.260	0.773	-0.888	0.728	-0.588	0.495	-0.708	0.727	-0.278	0.550
NRLM	<b>0.145</b> 0.32			0.590	0.747	-0.038	0.698	0.262	0.451	0.142	0.697	0.572	0.510
VNIIM	<b>-0.445</b> 0.68	-0.590	0.747			-0.628	0.927	-0.328	0.758	-0.448	0.926	-0.018	0.795
NIST	<b>0.183</b> 0.63	0.038	0.698	0.628	0.927			0.300	0.704	0.180	0.884	0.610	0.783
NRC	<b>-0.117</b> 0.33	-0.262	0.451	0.328	0.758	-0.300	0.704			-0.120	0.698	0.310	0.536
CENAM	<b>0.003</b> 0.62	-0.142	0.697	0.448	0.926	-0.180	0.884	0.120	0.698			0.430	0.749
INM	<b>-0.427</b> 0.44	-0.572	0.510	0.018	0.795	-0.610	0.783	-0.310	0.536	-0.430	0.749		
IMGC	<b>0.048</b> 0.22		0.376	0.493	0.716	-0.135	0.653	0.165	0.380	0.045	0.654	0.475	0.494
PTB	<b>0.033</b> 0.22	-0.112	0.495	0.478	0.785	-0.150	0.731	0.150	0.497	0.030	0.728	0.460	0.577

Lab, S/N j ∏			IMO	GC .	P1	В
, <u>,</u>	D <sub>i</sub>	<b>U</b> <sub>i</sub>	D <sub>ij</sub>	U <sub>ij</sub>	D ij	<b>U</b> <sub>ij</sub>
	1	K	1	K	- 1	K
VSL	-0.055	0.262	-0.103	0.330	-0.088	0.461
NPL	-0.377	0.411	-0.425	0.450	-0.410	0.552
CSIRO	-0.025	0.122	-0.073	0.231	-0.058	0.397
KRISS	0.000	0.278	-0.048	0.334	-0.033	0.464
NIM	0.085	0.379	0.037	0.404	0.052	0.518
NMC	-0.705	0.384	-0.753	0.428	-0.738	0.536
NRLM	0.145	0.320	0.097	0.376	0.112	0.495
VNIIM	-0.445	0.686	-0.493	0.716	-0.478	0.785
NIST	0.183	0.632	0.135	0.653	0.150	0.731
NRC	-0.117	0.333	-0.165	0.380	-0.150	0.497
CENAM	0.003	0.628	-0.045	0.654	-0.030	0.728
INM	-0.427	0.445	-0.475	0.494	-0.460	0.577
IMGC	0.048	0.222			0.015	0.430
РТВ	0.033	0.220	-0.015	0.430		

CCT-K5 : Nominal temperature,  $T_{90}$  = 1473 K Degrees of equivalence,  $D_i$ , and expanded uncertainties (k = 2),  $U_i$ , expressed in K




# Nominal temperature 1300°C

Lab, S/N j ∏			V:	SL	N	PL	CSI	RO	KR	ISS	NI	М	NN	ИC
ĮĮ.	$D_i$	U <sub>i</sub>	D <sub>ij</sub>	U <sub>ij</sub>	$D_{ij}$	U <sub>ij</sub>								
	1	K	1	/ <b>K</b>		K	1	K	1	K	1	K	1	K
VSL	-0.045	0.322			0.452	0.545	-0.045	0.283	-0.065	0.403	-0.185	0.478	0.750	0.530
NPL	-0.497	0.486	-0.452	0.545			-0.497	0.471	-0.517	0.550	-0.637	0.601	0.298	0.631
CSIRO	0.000	0.165	0.045	0.283	0.497	0.471			-0.020	0.292	-0.140	0.388	0.795	0.453
KRISS	0.020	0.343	0.065	0.403	0.517	0.550	0.020	0.292			-0.120	0.461	0.815	0.557
NIM	0.140	0.437	0.185	0.478	0.637	0.601	0.140	0.388	0.120	0.461			0.935	0.632
NMC	-0.795	0.459	-0.750	0.530	-0.298	0.631	-0.795	0.453	-0.815	0.601	-0.935	0.632		
NRLM	0.210	0.393	0.255	0.456	0.707	0.591	0.210	0.362	0.190	0.531	0.070	0.531	1.005	0.574
VNIIM	-0.660	0.956	-0.615	0.990	-0.163	1.051	-0.660	0.951	-0.680	1.045	-0.800	1.045	0.135	1.030
NIST	0.164	0.719	0.208	0.760	0.660	0.845	0.163	0.708	0.143	0.763	0.023	0.800	0.958	0.823
NRC	-0.187	0.415	-0.142	0.482	0.310	0.598	-0.187	0.394	-0.207	0.486	-0.327	0.544	0.608	0.577
CENAM	-0.047	0.755	-0.002	0.796	0.450	0.872	-0.047	0.746	-0.067	0.798	-0.187	0.835	0.748	0.857
INM	-0.492	0.580	-0.447	0.566	0.005	0.706	-0.492	0.493	-0.512	0.570	-0.632	0.619	0.303	0.649
IMGC	0.064	0.284	0.108	0.379	0.560	0.524	0.063	0.259	0.043	0.385	-0.077	0.455	0.858	0.494
РТВ	0.024	0.283	0.068	0.512	0.520	0.625	0.023	0.430	0.003	0.516	-0.117	0.571	0.818	0.602

Lab, S/N j ∏			NRLM		VN	IIM	NIS	ST	NR	C	CEN	AM	IN	М
Įļ	$D_i$	U <sub>i</sub>	D ij	U <sub>ij</sub>	D <sub>ij</sub>	U <sub>ij</sub>	D <sub>ij</sub>	U <sub>ij</sub>	D <sub>ij</sub>	U <sub>ij</sub>	D ij	U <sub>ij</sub>	D ij	U <sub>ij</sub>
	/ K		/ <b>K</b>		- 1	K	1	K	- 1	K	- 1	K	1	K
VSL	-0.045	0.322	-0.255	0.456	0.615	0.990	-0.208	0.760	0.142	0.482	0.002	0.796	0.447	0.566
NPL	-0.497	0.486	-0.707	0.591	0.163	1.051	-0.660	0.845	-0.310	0.598	-0.450	0.169	-0.005	0.706
CSIRO	0.000	0.165	-0.210	0.362	0.660	0.951	-0.163	0.708	0.187	0.394	0.047	0.746	0.492	0.493
KRISS	0.020	0.343	-0.190	0.463	0.680	1.003	-0.143	0.763	0.207	0.486	0.067	0.798	0.512	0.570
NIM	0.140	0.437	-0.070	0.531	0.800	1.045	-0.023	0.800	0.327	0.544	0.187	0.835	0.632	0.619
NMC	-0.795	0.459	-1.005	0.574	-0.135	1.030	-0.958	0.823	-0.608	0.577	-0.748	0.857	-0.303	0.649
NRLM	0.210	0.393			0.870	1.015	0.047	0.792	0.397	0.532	0.257	0.827	0.702	0.609
VNIIM	-0.660	0.956	-0.870	1.015			-0.823	1.177	-0.473	1.019	-0.613	1.200	-0.168	1.061
NIST	0.164	0.719	-0.047	0.792	0.823	1.177			0.350	0.808	0.210	1.021	0.655	0.925
NRC	-0.187	0.415	-0.397	0.532	0.473	1.019	-0.350	0.808			-0.140	0.835	0.305	0.643
CENAM	-0.047	0.755	-0.257	0.827	0.613	1.200	-0.210	1.021	0.140	0.835			0.445	0.928
INM	-0.492	0.580	-0.702	0.609	0.168	1.061	-0.655	0.925	-0.305	0.643	-0.445	0.928		
IMGC	0.064	0.284	-0.147	0.441	0.723	0.974	-0.100	0.741	0.250	0.460	0.110	0.779	0.555	0.626
РТВ	0.024	0.283	-0.187	0.560	0.683	1.034	-0.140	0.818	0.210	0.572	0.070	0.851	0.515	0.704

Lab, S/N j ∏			IM	GC	PΊ	TB
ŢĹ	$D_i$	U,	D <sub>ij</sub>	U <sub>ij</sub>	D <sub>ij</sub>	U <sub>ij</sub>
V	1	K		K	_	K
VSL	-0.045	0.322	-0.108	0.379	-0.068	0.512
NPL	-0.497	0.486	-0.560	0.524	-0.520	0.625
CSIRO	0.000	0.165	-0.063	0.259	-0.023	0.430
KRISS	0.020	0.343	-0.043	0.385	-0.003	0.516
NIM	0.140	0.437	0.077	0.455	0.117	0.571
NMC	-0.795	0.459	-0.858	0.494	-0.818	0.602
NRLM	0.210	0.393	0.147	0.441	0.187	0.560
VNIIM	-0.660	0.956	-0.723	0.974	-0.683	1.034
NIST	0.164	0.719	0.100	0.741	0.140	0.818
NRC	-0.187	0.415	-0.250	0.460	-0.210	0.572
CENAM	-0.047	0.755	-0.110	0.779	-0.070	0.851
INM	-0.492	0.580	-0.555	0.626	-0.515	0.704
IMGC	0.064	0.284			0.040	0.484
РТВ	0.024	0.283	-0.040	0.484		

CCT-K5 : Nominal temperature,  $T_{90} = 1573 \text{ K}$ Degrees of equivalence,  $D_i$ , and expanded uncertainties (k = 2),  $U_i$ , expressed in K



## Nominal temperature 1400°C


Lab, S/N i ===>

Lab, S/N j ∏			VSL		N	PL	CSI	RO	KR	ISS	NI	М	NM	<b>IC</b>
	<b>D</b> ;	<i>U<sub>i</sub></i> K	D <sub>ij</sub> U <sub>ij</sub> / K		D <sub>ij</sub>	υ <sub>ij</sub> Κ	D <sub>ij</sub>	U <sub>ij</sub> K	D <sub>ij</sub> /	U <sub>ij</sub> K	<b>D</b> <sub>ij</sub> /	U <sub>ij</sub> K	<b>D</b> <sub>ij</sub> /	U <sub>ij</sub> K
VSL	-0.045	0.385			0.465	0.656	-0.090	0.322	-0.100	0.464	-0.240	0.558	0.840	0.585
NPL	-0.510	0.600	-0.465	0.656			-0.555	0.575	-0.565	0.661	-0.705	0.722	0.375	0.738
CSIRO	0.045	0.221	0.090	0.322	0.555	0.575			-0.010	0.334	-0.150	0.455	0.930	0.496
KRISS	0.055	0.410	0.100	0.464	0.565	0.661	0.010	0.334			-0.140	0.534	0.940	0.622
NIM	0.195	0.522	0.240	0.558	0.705	0.722	0.150	0.455	0.140	0.534			1.080	0.715
NMC	-0.885	0.519	-0.840	0.585	-0.375	0.738	-0.930	0.496	-0.940	0.679	-1.080	0.715		
NRLM	0.310	0.473	0.355	0.528	0.820	0.710	0.265	0.422	0.255	0.616	0.115	0.616	1.195	0.650
VNIIM	-0.880	1.228	-0.835	1.259	-0.370	1.332	-0.925	1.220	-0.935	1.328	-1.075	1.328	0.005	1.285
NIST	0.260	0.815	0.305	0.854	0.770	0.972	0.215	0.792	0.205	0.858	0.065	0.905	1.145	0.918
NRC	-0.730	0.496	-0.685	0.559	-0.220	0.714	-0.775	0.462	-0.785	0.565	-0.925	0.635	0.155	0.653
CENAM	-0.050	0.906	-0.005	0.944	0.460	1.043	-0.095	0.889	-0.105	0.947	-0.245	0.991	0.835	1.003
INM	-0.510	0.629	-0.465	0.643	0.000	0.801	-0.555	0.560	-0.565	0.648	-0.705	0.710	0.375	0.726
IMGC	0.115	0.352	0.160	0.443	0.625	0.632	0.070	0.308	0.060	0.450	-0.080	0.535	1.000	0.556
РТВ	0.000	0.352	0.045	0.589	0.510	0.739	-0.045	0.496	-0.055	0.594	-0.195	0.661	0.885	0.679

Lab, S/N j ∏			NRLM		VN	IIM	NIS	ST .	NF	RC .	CEN	IAM	IN	М
Å.	<b>D</b> ;	U <sub>i</sub> K	D <sub>ij</sub>	U <sub>ij</sub> K	<b>D</b> <sub>ij</sub> /	U <sub>ij</sub> K								
VSL	-0.045	0.385	-0.355	0.528	0.835	1.259	-0.305	0.854	0.685	0.559	0.005	0.944	0.465	0.643
NPL	-0.510	0.600	-0.820	0.710	0.370	1.332	-0.770	0.972	0.220	0.714	-0.460	0.252	0.000	0.801
CSIRO	0.045	0.221	-0.265	0.422	0.925	1.220	-0.215	0.792	0.775	0.462	0.095	0.889	0.555	0.560
KRISS	0.055	0.410	-0.255	0.535	0.935	1.279	-0.205	0.858	0.785	0.565	0.105	0.947	0.565	0.648
NIM	0.195	0.522	-0.115	0.616	1.075	1.328	-0.065	0.905	0.925	0.635	0.245	0.991	0.705	0.710
NMC	-0.885	0.519	-1.195	0.650	-0.005	1.285	-1.145	0.918	-0.155	0.653	-0.835	1.003	-0.375	0.726
NRLM	0.310	0.473			1.190	1.290	0.050	0.896	1.040	0.623	0.360	0.983	0.820	0.699
VNIIM	-0.880	1.228	-1.190	1.290			-1.140	1.440	-0.150	1.287	-0.830	1.496	-0.370	1.326
NIST	0.260	0.815	-0.050	0.896	1.140	1.440			0.990	0.918	0.310	1.184	0.770	1.015
NRC	-0.730	0.496	-1.040	0.623	0.150	1.287	-0.990	0.918			-0.680	0.986	-0.220	0.707
CENAM	-0.050	0.906	-0.360	0.983	0.830	1.496	-0.310	1.184	0.680	0.986			0.460	1.052
INM	-0.510	0.629	-0.820	0.699	0.370	1.326	-0.770	1.015	0.220	0.707	-0.460	1.052		
IMGC	0.115	0.352	-0.195	0.520	0.995	1.241	-0.145	0.832	0.845	0.539	0.165	0.926	0.625	0.674
PTB	0.000	0.352	-0.310	0.649	0.880	1.301	-0.260	0.923	0.730	0.657	0.050	1.002	0.510	0.761

Lab, S/N j ∏			IMGC		PTB		
][	D;	<b>U</b> i	D ij	U ij	$D_{ij}$	U <sub>ii</sub>	
V	/ <b>K</b>			* / <b>K</b>		*/ <b>K</b>	
VSL	-0.045	0.385	-0.160	0.443	-0.045	0.589	
NPL	-0.510	0.600	-0.625	0.632	-0.510	0.739	
CSIRO	0.045	0.221	-0.070	0.308	0.045	0.496	
KRISS	0.055	0.410	-0.060	0.450	0.055	0.594	
NIM	0.195	0.522	0.080	0.535	0.195	0.661	
NMC	-0.885	0.519	-1.000	0.556	-0.885	0.679	
NRLM	0.310	0.473	0.195	0.520	0.310	0.649	
VNIIM	-0.880	1.228	-0.995	1.241	-0.880	1.301	
NIST	0.260	0.815	0.145	0.832	0.260	0.923	
NRC	-0.730	0.496	-0.845	0.539	-0.730	0.657	
CENAM	-0.050	0.906	-0.165	0.926	-0.050	1.002	
INM	-0.510	0.629	-0.625	0.674	-0.510	0.761	
IMGC	0.115	0.352			0.115	0.558	
РТВ	0.000	0.352	-0.115	0.558			

CCT-K5 : Nominal temperature,  $T_{90}$  = 1673 K Degrees of equivalence,  $D_i$ , and expanded uncertainties (k = 2),  $U_i$ , expressed in K



# Nominal temperature 1500°C


Lab, S/N i ==⇒

Lab, S/N j ⊓			V	SL SL	N	PL	CSI	RO	KRI	SS	NI	М	NN	<b>MC</b>
	<b>D</b> ;	<i>U<sub>i</sub></i> K	D <sub>ij</sub>	U <sub>ij</sub> K	<b>D</b> <sub>ij</sub> /	<i>υ<sub>ij</sub></i> Κ	<b>D</b> <sub>ij</sub> /	U <sub>ij</sub> K	<b>D</b> <sub>ij</sub> /	U <sub>ij</sub> K	<b>D</b> <sub>ij</sub> /	<i>υ<sub>ij</sub></i> Κ	<b>D</b> <sub>ij</sub> /	U <sub>ij</sub> K
VSL	-0.082	0.414			0.467	0.747	-0.140	0.349	-0.110	0.518	-0.295	0.623	0.970	0.656
NPL	-0.550	0.683	-0.467	0.747			-0.607	0.671	-0.577	0.770	-0.762	0.834	0.503	0.849
CSIRO	0.058	0.259	0.140	0.349	0.607	0.671			0.030	0.387	-0.155	0.508	1.110	0.584
KRISS	0.028	0.467	0.110	0.518	0.577	0.770	-0.030	0.387			-0.185	0.623	1.080	0.710
NIM	0.213	0.593	0.295	0.623	0.762	0.834	0.155	0.508	0.185	0.623			1.265	0.821
NMC	-1.052	0.588	-0.970	0.656	-0.503	0.849	-1.110	0.584	-1.080	0.769	-1.265	0.821		
NRLM	0.353	0.538	0.435	0.589	0.902	0.822	0.295	0.485	0.325	0.702	0.140	0.702	1.405	0.746
VNIIM	-0.912	1.520	-0.830	1.554	-0.363	1.617	-0.970	1.532	-0.940	1.655	-1.125	1.655	0.140	1.563
NIST	0.336	0.918	0.418	0.963	0.885	1.104	0.278	0.905	0.308	0.981	0.123	1.032	1.388	1.044
NRC	-0.850	0.565	-0.767	0.636	-0.300	0.815	-0.907	0.543	-0.877	0.663	-1.062	0.736	0.203	0.752
CENAM	0.000	1.026	0.083	1.072	0.550	1.190	-0.057	1.020	-0.027	1.088	-0.212	1.134	1.053	1.144
INM	-0.580	0.730	-0.497	0.741	-0.030	0.924	-0.637	0.663	-0.607	0.764	-0.792	0.828	0.473	0.843
IMGC	0.146	0.396	0.228	0.502	0.695	0.724	0.088	0.377	0.118	0.535	-0.067	0.623	1.198	0.643
РТВ	-0.025	0.394	0.058	0.648	0.525	0.833	-0.082	0.557	-0.052	0.674	-0.237	0.746	1.028	0.763

Lab, S/N j ∏			NR	NRLM		IIM	NIS	ST T	NF	RC .	CEN	IAM	INM	
Įļ	$D_i$	U <sub>i</sub>	D <sub>ij</sub>	U <sub>ij</sub>										
	1	K	1	/ K		/ <b>K</b>		/ <b>K</b>		K	/ <b>K</b>		1	K
VSL	-0.082	0.414	-0.435	0.589	0.830	1.554	-0.418	0.963	0.767	0.636	-0.083	1.072	0.497	0.741
NPL	-0.550	0.683	-0.902	0.822	0.363	1.617	-0.885	1.104	0.300	0.815	-0.550	0.272	0.030	0.924
CSIRO	0.058	0.259	-0.295	0.485	0.970	1.532	-0.278	0.905	0.907	0.543	0.057	1.020	0.637	0.663
KRISS	0.028	0.467	-0.325	0.615	0.940	1.590	-0.308	0.981	0.877	0.663	0.027	1.088	0.607	0.764
NIM	0.213	0.593	-0.140	0.702	1.125	1.655	-0.123	1.032	1.062	0.736	0.212	1.134	0.792	0.828
NMC	-1.052	0.588	-1.405	0.746	-0.140	1.563	-1.388	1.044	-0.203	0.752	-1.053	1.144	-0.473	0.843
NRLM	0.353	0.538			1.265	1.597	0.017	1.022	1.202	0.721	0.352	1.125	0.932	0.816
VNIIM	-0.912	1.520	-1.265	1.597			-1.248	1.728	-0.063	1.569	-0.913	1.791	-0.333	1.615
NIST	0.336	0.918	-0.017	1.022	1.248	1.728			1.185	1.044	0.335	1.336	0.915	1.166
NRC	-0.850	0.565	-1.202	0.721	0.063	1.569	-1.185	1.044			-0.850	1.128	-0.270	0.811
CENAM	0.000	1.026	-0.352	1.125	0.913	1.791	-0.335	1.336	0.850	1.128			0.580	1.223
INM	-0.580	0.730	-0.932	0.816	0.333	1.615	-0.915	1.166	0.270	0.811	-0.580	1.223		
IMGC	0.146	0.396	-0.207	0.606	1.058	1.519	-0.190	0.939	0.995	0.619	0.145	1.050	0.725	0.786
PTB	-0.025	0.394	-0.377	0.732	0.888	1.574	-0.360	1.024	0.825	0.744	-0.025	1.128	0.555	0.891

Lab, S/N j ∏			IM	GC	PT	В	
][	D,	U,	D ij	U <sub>ii</sub>	$D_{ij}$	U <sub>ii</sub>	
V		K		K ້	" / <b>K</b>		
VSL	-0.082	0.414	-0.228	0.502	-0.058	0.648	
NPL	-0.550	0.683	-0.695	0.724	-0.525	0.833	
CSIRO	0.058	0.259	-0.088	0.377	0.082	0.557	
KRISS	0.028	0.467	-0.118	0.535	0.052	0.674	
NIM	0.213	0.593	0.067	0.623	0.237	0.746	
NMC	-1.052	0.588	-1.198	0.643	-1.028	0.763	
NRLM	0.353	0.538	0.207	0.606	0.377	0.732	
VNIIM	-0.912	1.520	-1.058	1.519	-0.888	1.574	
NIST	0.336	0.918	0.190	0.939	0.360	1.024	
NRC	-0.850	0.565	-0.995	0.619	-0.825	0.744	
CENAM	0.000	1.026	-0.145	1.050	0.025	1.128	
INM	-0.580	0.730	-0.725	0.786	-0.555	0.891	
IMGC	0.146	0.396			0.170	0.611	
РТВ	-0.025	0.394	-0.170	0.611			

CCT-K5 : Nominal temperature,  $T_{90} = 1773 \text{ K}$ Degrees of equivalence,  $D_i$ , and expanded uncertainties (k = 2),  $U_i$ , expressed in K



# Nominal temperature 1600°C


Lab, S/N i <del>□</del>

Lab, S/N j ∏			V:	VSL		PL	CSI	RO	KR	ISS	NI	М	NMC	
Ų Į	<b>D</b> ;	U <sub>i</sub> K	D <sub>ij</sub>	U <sub>ij</sub> K	<b>D</b> <sub>ij</sub> /	U <sub>ij</sub> K								
VSL	-0.113	0.452			0.427	0.853	-0.165	0.387	-0.140	0.572	-0.350	0.702	1.020	0.724
NPL	-0.540	0.781	-0.427	0.853			-0.592	0.768	-0.567	0.875	-0.777	0.951	0.593	0.972
CSIRO	0.053	0.262	0.165	0.387	0.592	0.768			0.025	0.426	-0.185	0.580	1.185	0.628
KRISS	0.028	0.502	0.140	0.572	0.567	0.875	-0.025	0.426			-0.210	0.702	1.160	0.764
NIM	0.238	0.660	0.350	0.702	0.777	0.951	0.185	0.580	0.210	0.702			1.370	0.892
NMC	-1.133	0.652	-1.020	0.724	-0.593	0.972	-1.185	0.628	-1.160	0.836	-1.370	0.892		
NRLM	0.443	0.599	0.555	0.662	0.982	0.938	0.390	0.543	0.415	0.788	0.205	0.788	1.575	0.828
VNIIM	-0.578	1.797	-0.465	1.835	-0.038	1.901	-0.630	1.807	-0.605	1.955	-0.815	1.955	0.555	1.860
NIST	0.591	1.014	0.703	1.069	1.130	1.240	0.538	1.002	0.563	1.086	0.353	1.148	1.723	1.167
NRC	-1.020	0.643	-0.907	0.713	-0.480	0.939	-1.072	0.607	-1.047	0.738	-1.257	0.826	0.113	0.851
CENAM	-0.005	1.224	0.108	1.272	0.535	1.411	-0.057	1.217	-0.032	1.287	-0.242	1.340	1.128	1.355
INM	-0.615	0.820	-0.502	0.833	-0.075	1.056	-0.667	0.744	-0.642	0.854	-0.852	0.932	0.518	0.955
IMGC	0.126	0.436	0.238	0.557	0.665	0.831	0.073	0.413	0.098	0.589	-0.112	0.696	1.258	0.726
РТВ	0.001	0.433	0.113	0.720	0.540	0.950	-0.052	0.616	-0.027	0.745	-0.237	0.832	1.133	0.858

Lab, S/N j ∏			NR	NRLM		IIM	NIS	ST T	NR	RC .	CEN	AM	INM	
Ţ	<b>D</b> ;	U <sub>i</sub> K	<b>D</b> <sub>ij</sub> /	U <sub>ij</sub> K	<b>D</b> <sub>ij</sub> /	U <sub>ij</sub> K	D <sub>ij</sub> /	U <sub>ij</sub> K	D <sub>ij</sub> /	U <sub>ij</sub> K	<b>D</b> <sub>ij</sub> /	<i>υ<sub>ij</sub></i> Κ	<b>D</b> <sub>ij</sub> /	U <sub>ij</sub> K
VSL	-0.113	0.452	-0.555	0.662	0.465	1.835	-0.703	1.069	0.907	0.713	-0.108	1.272	0.502	0.833
NPL	-0.540	0.781	-0.982	0.938	0.038	1.901	-1.130	1.240	0.480	0.939	-0.535	0.297	0.075	1.056
CSIRO	0.053	0.262	-0.390	0.543	0.630	1.807	-0.538	1.002	1.072	0.607	0.057	1.217	0.667	0.744
KRISS	0.028	0.502	-0.415	0.685	0.605	1.868	-0.563	1.086	1.047	0.738	0.032	1.287	0.642	0.854
NIM	0.238	0.660	-0.205	0.788	0.815	1.955	-0.353	1.148	1.257	0.826	0.242	1.340	0.852	0.932
NMC	-1.133	0.652	-1.575	0.828	-0.555	1.860	-1.723	1.167	-0.113	0.851	-1.128	1.355	-0.518	0.955
NRLM	0.443	0.599			1.020	1.888	-0.148	1.137	1.462	0.811	0.447	1.330	1.057	0.918
VNIIM	-0.578	1.797	-1.020	1.888			-1.168	2.007	0.442	1.841	-0.573	2.122	0.037	1.892
NIST	0.591	1.014	0.148	1.137	1.168	2.007			1.610	1.171	0.595	1.554	1.205	1.292
NRC	-1.020	0.643	-1.462	0.811	-0.442	1.841	-1.610	1.171			-1.015	1.337	-0.405	0.900
CENAM	-0.005	1.224	-0.447	1.330	0.573	2.122	-0.595	1.554	1.015	1.337			0.610	1.428
INM	-0.615	0.820	-1.057	0.918	-0.037	1.892	-1.205	1.292	0.405	0.900	-0.610	1.428		
IMGC	0.126	0.436	-0.317	0.678	0.703	1.787	-0.465	1.045	1.145	0.713	0.130	1.255	0.740	0.886
PTB	0.001	0.433	-0.442	0.817	0.578	1.845	-0.590	1.139	1.020	0.853	0.005	1.336	0.615	1.008

Lab, S/N j ∏			IM	GC	PΊ	В	
][	D,	U,	D <sub>ij</sub>	U <sub>ii</sub>	$D_{ij}$	U <sub>ii</sub>	
V	. 1	K		K	" / <b>K</b>		
VSL	-0.113	0.452	-0.238	0.557	-0.113	0.720	
NPL	-0.540	0.781	-0.665	0.831	-0.540	0.950	
CSIRO	0.053	0.262	-0.073	0.413	0.052	0.616	
KRISS	0.028	0.502	-0.098	0.589	0.027	0.745	
NIM	0.238	0.660	0.112	0.696	0.237	0.832	
NMC	-1.133	0.652	-1.258	0.726	-1.133	0.858	
NRLM	0.443	0.599	0.317	0.678	0.442	0.817	
VNIIM	-0.578	1.797	-0.703	1.787	-0.578	1.845	
NIST	0.591	1.014	0.465	1.045	0.590	1.139	
NRC	-1.020	0.643	-1.145	0.713	-1.020	0.853	
CENAM	-0.005	1.224	-0.130	1.255	-0.005	1.336	
INM	-0.615	0.820	-0.740	0.886	-0.615	1.008	
IMGC	0.126	0.436			0.125	0.683	
РТВ	0.001	0.433	-0.125	0.683			

CCT-K5 : Nominal temperature,  $T_{90}$  = 1873 K Degrees of equivalence,  $D_i$ , and expanded uncertainties (k = 2),  $U_i$ , expressed in K



# Nominal temperature 1700°C

Lab, S/N i □

Lab, S/N j ∏			V	SL .	N	PL	CSI	RO	KRI	SS	NI	М	NN	<b>IC</b>
ŢĮ.	<b>D</b> ;	U <sub>i</sub> K	D <sub>ij</sub>	U <sub>ij</sub> K	D <sub>ij</sub>	U <sub>ij</sub> K	<b>D</b> <sub>ij</sub> /	U <sub>ij</sub> K	D <sub>ij</sub>	U <sub>ij</sub> K	<b>D</b> <sub>ij</sub> /	U <sub>ij</sub> K	<b>D</b> <sub>ij</sub> /	U <sub>ij</sub> K
VSL	-0.235	0.480			0.330	0.929	-0.265	0.433	-0.235	0.667	-0.485	0.813	1.155	0.816
NPL	-0.565	0.854	-0.330	0.929			-0.595	0.834	-0.565	0.972	-0.815	1.052	0.825	1.075
CSIRO	0.030	0.259	0.265	0.433	0.595	0.834			0.030	0.511	-0.220	0.673	1.420	0.723
KRISS	0.000	0.571	0.235	0.667	0.565	0.972	-0.030	0.511			-0.250	0.816	1.390	0.896
NIM	0.250	0.740	0.485	0.813	0.815	1.052	0.220	0.673	0.250	0.816			1.640	1.038
NMC	-1.390	0.734	-1.155	0.816	-0.825	1.075	-1.420	0.723	-1.390	0.968	-1.640	1.038		
NRLM	0.455	0.667	0.690	0.752	1.020	1.036	0.425	0.625	0.455	0.911	0.205	0.911	1.845	0.950
VNIIM	0.035	2.022	0.270	2.054	0.600	2.155	0.005	2.025	0.035	2.179	-0.215	2.179	1.425	2.102
NIST	0.500	1.129	0.735	1.185	1.065	1.383	0.470	1.112	0.500	1.219	0.250	1.284	1.890	1.302
NRC	-1.340	0.740	-1.105	0.792	-0.775	1.064	-1.370	0.678	-1.340	0.841	-1.590	0.933	0.050	0.960
CENAM	-0.170	1.321	0.065	1.371	0.395	1.538	-0.200	1.309	-0.170	1.401	-0.420	1.458	1.220	1.475
INM	-0.805	0.892	-0.570	0.907	-0.240	1.169	-0.835	0.809	-0.805	0.950	-1.055	1.033	0.585	1.056
IMGC	0.005	0.482	0.240	0.606	0.570	0.928	-0.025	0.446	0.005	0.669	-0.245	0.782	1.395	0.812
РТВ	-0.100	0.480	0.135	0.775	0.465	1.046	-0.130	0.657	-0.100	0.825	-0.350	0.918	1.290	0.945

Lab, S/N j ∏			NR	LM	VN	IIM	NIS	ST	NF	C	CEN	IAM	IN	М
Įļ	D <sub>i</sub>	U <sub>i</sub>	D <sub>ij</sub>	U ij	D <sub>ij</sub>	U ij	D <sub>ij</sub>	U ij	D <sub>ij</sub>	U <sub>ij</sub>	D <sub>ij</sub>	U <sub>ij</sub>	D <sub>ij</sub>	U <sub>ij</sub>
•	1	K	/ <b>K</b>		1	/ <b>K</b>		/ <b>K</b>		K	/ <b>K</b>		I	K
VSL	-0.235	0.480	-0.690	0.752	-0.270	2.054	-0.735	1.185	1.105	0.792	-0.065	1.371	0.570	0.907
NPL	-0.565	0.854	-1.020	1.036	-0.600	2.155	-1.065	1.383	0.775	1.064	-0.395	0.351	0.240	1.169
CSIRO	0.030	0.259	-0.425	0.625	-0.005	2.025	-0.470	1.112	1.370	0.678	0.200	1.309	0.835	0.809
KRISS	0.000	0.571	-0.455	0.799	-0.035	2.099	-0.500	1.219	1.340	0.841	0.170	1.401	0.805	0.950
NIM	0.250	0.740	-0.205	0.911	0.215	2.179	-0.250	1.284	1.590	0.933	0.420	1.458	1.055	1.033
NMC	-1.390	0.734	-1.845	0.950	-1.425	2.102	-1.890	1.302	-0.050	0.960	-1.220	1.475	-0.585	1.056
NRLM	0.455	0.667			0.420	2.117	-0.045	1.271	1.795	0.915	0.625	1.446	1.260	1.016
VNIIM	0.035	2.022	-0.420	2.117			-0.465	2.277	1.375	2.100	0.205	2.379	0.840	2.146
NIST	0.500	1.129	0.045	1.271	0.465	2.277			1.840	1.334	0.670	1.708	1.305	1.433
NRC	-1.340	0.740	-1.795	0.915	-1.375	2.100	-1.840	1.334			-1.170	1.474	-0.535	1.001
CENAM	-0.170	1.321	-0.625	1.446	-0.205	2.379	-0.670	1.708	1.170	1.474			0.635	1.554
INM	-0.805	0.892	-1.260	1.016	-0.840	2.146	-1.305	1.433	0.535	1.001	-0.635	1.554		
IMGC	0.005	0.482	-0.450	0.760	-0.030	2.036	-0.495	1.179	1.345	0.836	0.175	1.368	0.810	0.978
РТВ	-0.100	0.480	-0.555	0.900	-0.135	2.093	-0.600	1.274	1.240	0.967	0.070	1.451	0.705	1.093

Lab, S/N j ∏			IMC	GC	PΊ	В	
][	$D_i$	<b>U</b> <sub>i</sub>	D <sub>ij</sub>	U <sub>ii</sub>	$D_{ij}$	U <sub>ij</sub>	
V	1	K	_	K	" / <b>K</b>		
VSL	-0.235	0.480	-0.240	0.606	-0.135	0.775	
NPL	-0.565	0.854	-0.570	0.928	-0.465	1.046	
CSIRO	0.030	0.259	0.025	0.446	0.130	0.657	
KRISS	0.000	0.571	-0.005	0.669	0.100	0.825	
NIM	0.250	0.740	0.245	0.782	0.350	0.918	
NMC	-1.390	0.734	-1.395	0.812	-1.290	0.945	
NRLM	0.455	0.667	0.450	0.760	0.555	0.900	
VNIIM	0.035	2.022	0.030	2.036	0.135	2.093	
NIST	0.500	1.129	0.495	1.179	0.600	1.274	
NRC	-1.340	0.740	-1.345	0.836	-1.240	0.967	
CENAM	-0.170	1.321	-0.175	1.368	-0.070	1.451	
INM	-0.805	0.892	-0.810	0.978	-0.705	1.093	
IMGC	0.005	0.482			0.105	0.766	
РТВ	-0.100	0.480	-0.105	0.766			

CCT-K5 : Nominal temperature,  $T_{90}$  = 1973 K Degrees of equivalence,  $D_i$ , and expanded uncertainties (k = 2),  $U_i$ , expressed in K

