

Key comparisons CCT-K7, EUROMET.T-K7, COOMET.T-K7, EURAMET.T-K7.2, EURAMET.T-K7.3, EURAMET.T-K7.1 and APMP.T-K7.1

Key comparison CCT-K7

MEASURAND : Temperature of triple point of water

NOMINAL VALUE : 273.16 K

x_i : temperature difference between the national reference of laboratory i and the BIPM reference,

$T(\text{Lab } i) - T(\text{BIPM})$, where T is the temperature

u_i : combined standard uncertainty of x_i

Lab i	x_i / μK	u_i / μK
BIPM	0	44
LNE-INM	-54	66
CEM	-14	41
CENAM	-5	27
NMISA	105	74
NMIA	-29	34
INRIM	-15	27
IPQ	40	160
KRISS	69	56
MSL	117	16
NIM	33	61
NIST	-40	33
NMIJ	54	151
NMi-VSL	16	55
NPL	45	39
NRC	85	23
PTB	-14	56
SMU	69	53
A*STAR	34	71
UME	-53	91
VNIIM	22	46

Key comparison EUROMET.T-K7

MEASURAND : Temperature of triple point of water

NOMINAL VALUE : 273.16 K

$x_{i,\text{EUR}}$: temperature difference between the national reference of laboratory i and the NMi-VSL reference cell (VSL-094), $T(\text{Lab } i) - T(\text{VSL-094})$, where T is the temperature

$u_{i,\text{EUR}}$: combined standard uncertainty of $x_{i,\text{EUR}}$

Lab i	$x_{i,\text{EUR}}$ / μK	$u_{i,\text{EUR}}$ / μK
INM(RO)	-3	64
MKEH	-77	69
EIM	45	113
MIKES	-17	65
PTB	39	37
DTI	-51	68
JV	-146	194
VNIIM	63	52
DZM/LPM	10	50
UME	-86	77
DMDM	-151	119
BEV	-20	174
IPQ	30	80
NML(IE)	40	106
CMI	-310	61
GUM	-309	75
VMT/PFI	-62	119
CEM	51	44
INRIM	2	29
LNE-INM	16	56
MIRS/FE-LMK	33	25
SMD	44	34
SMU	-4	58
NMi-VSL	14	28

Key comparison COOMET.T-K7**MEASURAND : Temperature of triple point of water****NOMINAL VALUE : 273.16 K**

The comparison was performed in three stages: 1) each participating laboratory carried out comparisons of the transfer cell with its national standard cell in local conditions; 2) the transfer cells were forwarded together with measurement results to BelGIM, where they were compared with the standard BelGIM cell using its laboratory equipment; and 3) the transfer cells were sent back to laboratories, where they were compared again to the national standard cells in order to determine the stability of the transfer cells.

The measurement results of the participating laboratories are given in section 2, 3, and 4 of the COOMET.T-K7 Final Report.

Key comparison EURAMET.T-K7.2**MEASURAND : Temperature of triple point of water****NOMINAL VALUE : 273.16 K**

The measurement results of this bilateral key comparison between VSL (The Netherlands) and INTiBS (Poland), carried out in 2010, are summarized in Sections 3 and 4, on pages 5 and 6 of the EURAMET.T-K7.2 Final Report.

They lead to the following numbers:

$T^{\text{INTiBS}}_{\text{Nat. Ref.}} - T^{\text{VSL}}_{\text{Nat. Ref.}} = 2.9 \mu\text{K}$ and associated combined standard uncertainty ($k = 1$) $u = 46 \mu\text{K}$, where "Nat. Ref." stands for "National Reference".

Key comparison EURAMET.T-K7.3**MEASURAND : Temperature of triple point of water****NOMINAL VALUE : 273.16 K**

The measurement results of this bilateral key comparison between VSL (The Netherlands) and GUM (Poland), carried out in 2011, are summarized in Sections 3 and 4, on pages 5 and 6 of the EURAMET.T-K7.3 Final Report.

They lead to the following numbers:

$T^{\text{GUM}}_{\text{Nat. Ref.}} - T^{\text{VSL}}_{\text{Nat. Ref.}} = 11.5 \mu\text{K}$ and associated combined standard uncertainty ($k = 1$) $u = 60 \mu\text{K}$, where "Nat. Ref." stands for "National Reference".

Key comparison EURAMET.T-K7.1**MEASURAND : Temperature of triple point of water****NOMINAL VALUE : 273.16 K**

The measurement results of this bilateral key comparison between CMI (Czech Republic) and SMU (Slovakia), carried out in 2009, are summarized in Sections 2 and 3 of the EURAMET.T-K7.1 Final Report.

It should be noted that SMU replaced its national reference cell with a cell with known isotopic composition after completion of key comparison EUROMET.T-K7.

Key comparison APMP.T-K7.1**MEASURAND : Temperature of triple point of water****NOMINAL VALUE : 273.16 K**

The measurement results of this bilateral key comparison between VMI-STAMEQ (Viet Nam) and NMIJ (Japan), carried out in 2014 - 2015, are summarized in Sections 3 to 6 of the APMP.T-K7.1 Final Report.

They lead to the following numbers:

$T^{\text{VMI-STAMEQ}}_{\text{Nat. Ref.}} - T^{\text{NMIJ}}_{\text{Nat. Ref.}} = -19.9 \mu\text{K}$ and associated combined standard uncertainty ($k = 1$) $u = 46.3 \mu\text{K}$, where "Nat. Ref." stands for "National Reference".

Key comparisons CCT-K7, EUROMET.T-K7, COOMET.T-K7, EURAMET.T-K7.2, EURAMET.T-K7.3, EURAMET.T-K7.1 and APMP.T-K7.1

MEASURAND : Temperature of triple point of water

NOMINAL VALUE : 273.16 K

Key comparison CCT-K7

The key comparison reference value, x_R , is calculated as the arithmetic mean of the results, x_i , from all participants, including some laboratories who made corrections for the influence of chemical impurities and isotopic composition, and some who did not. $x_R = 22 \mu\text{K}$ (expressed as the difference from the BIPM value).

The uncertainty of x_R , u_R , is calculated as the standard deviation of the mean of the data set. $u_R = 11 \mu\text{K}$. Because the distribution of the pooled data is multimodal, care should be taken when using this quantity for calculating confidence intervals.

The degree of equivalence of laboratory i with respect to the key comparison reference value is given by a pair of terms:

$D_i = x_i - x_R$, and its expanded uncertainty U_i ($k = 2$), both expressed in μK . The uncertainty includes the participant's uncertainty, the comparison uncertainty and the uncertainty of the key comparison reference value.

The degree of equivalence between two laboratories i and j is given by a pair of terms: $D_{ij} = (D_i - D_j) = (x_i - x_j)$ and its expanded uncertainty U_{ij} ($k = 2$), both expressed in μK . U_{ij} includes both participants' uncertainties and twice the comparison uncertainty.

Key comparison EUROMET.T-K7

The EUROMET reference value referred to the NMi-VSL reference cell (VSL-094), $x_{R\text{ EUR}}$, and its standard uncertainty, $u_{R\text{ EUR}}$, are computed from the weighted mean of the individual results, with the exclusion of the CMI and GUM results, as explained in Section 7.1 on page 29 of the EUROMET.T-K7 Final Report: $x_{R\text{ EUR}} = 15.6 \mu\text{K}$, and $u_{R\text{ EUR}} = 10.5 \mu\text{K}$.

The degree of equivalence of laboratory i , participant in EUROMET.T-K7, with respect to the EUROMET reference value is given by a pair of terms: $D_{i\text{ EUR}} = x_{i\text{ EUR}} - x_{R\text{ EUR}}$, and its expanded uncertainty $U_{i\text{ EUR}}$ ($k = 2$), both expressed in μK .

Linking EUROMET.T-K7 to CCT-K7

The linkage process of EUROMET.T-K7 results to those of CCT-K7 is explained in Section 7.3 on page 36 of the EUROMET.T-K7 Final Report. It is based on the common participation of CEM, IPQ, LNE-INM, NMi-VSL, and SMU in both comparisons.

The degree of equivalence of laboratory i , participant in EUROMET.T-K7, with respect to the CCT-K7 reference value is given by a pair of terms: $D_i = D_{i\text{ EUR}} + 64.7 \mu\text{K}$, and its expanded uncertainty ($k = 2$) $U_i = 2[(U_{i\text{ EUR}}/2)^2 + (19.7)^2]^{1/2}$, both expressed in μK .

Pair-wise degrees of equivalence are computed inside EUROMET (see Section 7.2 on page 30 of the EUROMET.T-K7 Final Report).

Linking COOMET.T-K7 to CCT-K7

The linkage process of COOMET.T-K7 results to those of CCT-K7 is explained in Section 6 of the COOMET.T-K7 Final Report. It is based on the common participation of VNIIM in both comparisons.

The degree of equivalence of laboratory i , participant in COOMET.T-K7, with respect to the CCT-K7 reference value is given by a pair of terms: D_i , and its expanded uncertainty ($k = 2$) U_i , both expressed in μK .

No pair-wise degrees of equivalence are computed inside COOMET.T-K7.

Linking EURAMET.T-K7.2 to CCT-K7

The degree of equivalence of INTiBS relative to the CCT-K7 key comparison reference value is computed through the common participation of VSL in EURAMET.T-K7.2 and EUROMET.T-K7 (see Section 6 of the EURAMET.T-K7.2 Final Report).

Linking EURAMET.T-K7.3 to CCT-K7

The degree of equivalence of GUM relative to the CCT-K7 key comparison reference value is computed through the common participation of VSL in EURAMET.T-K7.3 and EUROMET.T-K7 (see Sections 6 and 7 of the EURAMET.T-K7.3 Final Report).

Linking EURAMET.T-K7.1 to CCT-K7

New degrees of equivalence relative to the CCT-K7 key comparison reference value are obtained in EURAMET.T-K7.1 for SMU and CMI, through the common participation of SMU in EURAMET.T-K7.1 and EUROMET.T-K7, taking into account the change of national reference cell at SMU since the completion of EUROMET.T-K7.

Linking APMP.T-K7.1 to CCT-K7

The degree of equivalence of VMI-STAMEQ relative to the CCT-K7 key comparison reference value is computed through the common participation of NMIJ in APMP.T-K7 and CCT-K7 (see Section 8 of the APMP.T-K7.1 Final Report).

Key comparisons CCT-K7, EUROMET.T-K7, COOMET.T-K7, EURAMET.T-K7.2, EURAMET.T-K7.3, EURAMET.T-K7.1 and APMP.T-K7.1

MEASURAND : Temperature of triple point of water
 NOMINAL VALUE : 273.16 K

Degrees of equivalence relative to the CCT-K7 key comparison reference value

Lab i	D_i	U_i / μK
BIPM	-22	90
LNE-INM	-76	134
CEM	-36	85
CENAM	-27	58
NMISA	83	150
NMIA	-51	72
INRIM	-37	58
IPQ	18	322
KRISS	47	115
MSL	95	39
NIM	11	124
NIST	-62	69
NMIJ	32	304
NMi-VSL	-6	113
NPL	23	81
NRC	62	52
PTB	-37	113
SMU	47	109
A*STAR	11	144
UME	-75	183
VNIIM	0	94

	D_i	U_i / μK
INM(RO)	46	136
MKEH	-28	145
EIM	95	230
MIKES	33	137
PTB	88	86
DTI	-2	145
JV	-97	390
VNIIM	112	113
DZM/LPM	59	110
UME	-37	161
DMDM	-101	242
BEV	29	352
IPQ	79	166
NML(IE)	89	212
CMI	-261	130
GUM	-260	155
VMT/PFI	-13	243
CEM	100	98
INRIM	51	73
LNE-INM	65	120
MIRS/FE-LMK	82	68
SMD	94	82
SMU	45	124
NMi-VSL	63	72
INTiBS	66	118
GUM	75	140

	D_i	U_i / μK
BelGIM	50	172
GEOSTM	-95	345
INSM	47	242
INIMET	-43	191
VNIIM	0	175
NSC IM	-19	249
KazInMetr	-102	215


SMU	52	149
CMI	73	206
VMI-STAMEQ	43	286

Black: participants in CCT-K7
 Green: participants in EUROMET.T-K7
 Blue: participants in COOMET.T-K7
 Orange: participant in EURAMET.T-K7.2
 Braun: participant in EURAMET.T-K7.3
 Grey: participants in EURAMET.T-K7.1
 Pink: participants in APMP.T-K7.1

CCT-K7, EUROMET.T-K7, EURAMET.T-K7.2, 7.3 & 7.1, COOMET.T-K7 and APMPT-K7.1

Water triple point cells

Degrees of equivalence [D_i and its expanded uncertainty ($k = 2$), U_i]



Red diamonds: CCT-K7 participants Green triangles: EUROMET.T-K7 participants Blue circles: COOMET.T-K7 participants Orange, brown and grey squares: EURAMET.T-K7.2,

EURAMET.T-K7.3 and EURAMET.T-K7.1 participants Pink circle: APMPT-K7.1 participant

In CCT-K7, NMISA, MSL and NRC realized systematically higher temperatures, because they are the only laboratories which based their realizations on the recommendation of the Supplementary Information for the ITS-90 to use water with the isotopic composition of the standard mean ocean water. Since the publication of CCT-K7 results, the CIPM approved Recommendation 2 (CI-2005) "Clarification of the definition of the kelvin, unit of thermodynamic temperature": see on pages 23 and 24 of the [EUROMET.T-K7 Final Report](#) for detailed

Key comparison CCT-K7

MEASURAND : Temperature of triple point of water

NOMINAL VALUE : 273.16 K

Matrix of equivalence

Lab j \longrightarrow

Lab i			BIPM		LNE-INM		CEM		CENAM		NMISA		NMIA	
	D_i	U_i	D_{ij}	U_{ij}										
	$/ \mu\text{K}$		$/ \mu\text{K}$		$/ \mu\text{K}$		$/ \mu\text{K}$		$/ \mu\text{K}$		$/ \mu\text{K}$		$/ \mu\text{K}$	
BIPM	-22	90			54	158	14	120	5	103	-105	172	29	111
LNE-INM	-76	134	-54	158			-40	156	-49	143	-159	199	-25	149
CEM	-36	85	-14	120	40	156			-9	98	-119	170	15	107
CENAM	-27	58	-5	103	49	143	9	98			-110	158	24	87
NMISA	83	150	105	172	159	199	119	170	110	158			134	163
NMIA	-51	72	-29	111	25	149	-15	107	-24	87	-134	163		
INRIM	-37	58	-15	103	39	143	-1	98	-10	76	-120	158	14	87
IPQ	18	322	40	333	94	347	54	331	45	325	-65	354	69	328
KRISS	47	115	69	143	122	174	83	139	73	125	-36	186	98	132
MSL	95	39	117	93	171	136	131	88	122	63	12	152	146	76
NIM	11	124	33	150	87	180	47	147	38	134	-72	192	62	140
NIST	-62	69	-40	109	13	147	-26	105	-36	85	-145	162	-11	95
NMIJ	32	304	54	315	108	331	68	314	59	308	-51	337	83	311
NMi-VSL	-6	113	16	141	69	172	30	138	20	123	-90	185	45	130
NPL	23	81	45	117	99	154	59	114	50	95	-60	168	74	104
NRC	62	52	84	99	138	140	98	95	89	71	-21	155	114	83
PTB	-37	113	-14	141	39	173	-1	138	-10	123	-120	185	15	130
SMU	47	109	69	138	123	170	83	135	74	119	-36	183	98	127
A*STAR	11	144	34	167	87	194	47	164	38	152	-72	205	63	158
UME	-75	183	-53	202	1	225	-39	199	-48	189	-158	234	-24	194
VNIIM	0	94	22	126	76	161	36	123	27	106	-83	174	51	114

Key comparison CCT-K7

MEASURAND : Temperature of triple point of water

NOMINAL VALUE : 273.16 K

Matrix of equivalence (Cont.)

Lab *j* \longrightarrow

Lab <i>i</i>			INRIM		IPQ		KRISS		MSL		NIM		NIST	
	<i>D_i</i>	<i>U_i</i>	<i>D_{ij}</i>	<i>U_{ij}</i>										
BIPM	-22	90	15	103	-40	333	-69	143	-117	93	-33	150	40	109
LNE-INM	-76	134	-39	143	-94	347	-122	174	-171	136	-87	180	-13	147
CEM	-36	85	1	98	-54	331	-83	139	-131	88	-47	147	26	105
CENAM	-27	58	10	76	-45	325	-73	125	-122	63	-38	134	36	85
NMISA	83	150	120	158	65	354	36	186	-12	152	72	192	145	162
NMIA	-51	72	-14	87	-69	328	-98	132	-146	76	-62	140	11	95
INRIM	-37	58			-55	325	-84	125	-132	63	-48	134	25	85
IPQ	18	322	55	325			-29	340	-77	323	7	343	80	327
KRISS	47	115	84	125	29	340			-48	117	36	166	109	130
MSL	95	39	132	63	77	323	48	117			84	127	157	73
NIM	11	124	48	134	-7	343	-36	166	-84	127			73	139
NIST	-62	69	-25	85	-80	327	-109	130	-157	73	-73	139		
NMIJ	32	304	69	308	14	441	-14	323	-63	305	21	327	95	310
NMi-VSL	-6	113	31	123	-24	339	-53	158	-101	115	-17	165	56	129
NPL	23	81	60	95	5	330	-23	137	-72	85	12	145	86	102
NRC	62	52	99	71	44	324	16	122	-33	57	51	131	125	80
PTB	-37	113	1	123	-55	340	-83	158	-132	116	-48	165	26	129
SMU	47	109	84	119	29	338	0	155	-48	112	36	162	109	125
A*STAR	11	144	48	152	-7	351	-35	181	-84	146	0	187	74	156
UME	-75	183	-38	189	-93	369	-121	214	-170	184	-86	219	-12	193
VNIIM	0	94	37	106	-18	334	-47	145	-95	97	-11	153	62	112

Key comparison CCT-K7

MEASURAND : Temperature of triple point of water

NOMINAL VALUE : 273.16 K

Matrix of equivalence (Cont.)

Lab *j* \longrightarrow

Lab <i>i</i>			NMIIJ		NMi-VSL		NPL		NRC		PTB		SMU	
	<i>D_i</i>	<i>U_i</i>	<i>D_{ij}</i>	<i>U_{ij}</i>										
BIPM	-22	90	-54	315	-16	141	-45	117	-84	99	14	141	-69	138
LNE-INM	-76	134	-108	331	-69	172	-99	154	-138	140	-39	173	-123	170
CEM	-36	85	-68	314	-30	138	-59	114	-98	95	1	138	-83	135
CENAM	-27	58	-59	308	-20	123	-50	95	-89	71	10	123	-74	119
NMISA	83	150	51	337	90	185	60	168	21	155	120	185	36	183
NMIA	-51	72	-83	311	-45	130	-74	104	-114	83	-15	130	-98	127
INRIM	-37	58	-69	308	-31	123	-60	95	-99	71	-1	123	-84	119
IPQ	18	322	-14	441	24	339	-5	330	-44	324	55	340	-29	338
KRISS	47	115	14	323	53	158	23	137	-16	122	83	158	0	155
MSL	95	39	63	305	101	115	72	85	33	57	132	116	48	112
NIM	11	124	-21	327	17	165	-12	145	-51	131	48	165	-36	162
NIST	-62	69	-95	310	-56	129	-86	102	-125	80	-26	129	-109	125
NMIIJ	32	304			39	323	9	313	-30	307	69	323	-15	321
NMi-VSL	-6	113	-39	323			-30	136	-69	120	30	157	-53	154
NPL	23	81	-9	313	30	136			-39	91	60	136	-24	132
NRC	62	52	30	307	69	120	39	91			99	120	16	116
PTB	-37	113	-69	323	-30	157	-60	136	-99	120			-83	154
SMU	47	109	15	321	53	154	24	132	-16	116	83	154		
A*STAR	11	144	-21	335	18	180	-12	162	-51	150	48	180	-35	178
UME	-75	183	-107	353	-68	213	-98	198	-137	187	-38	213	-122	211
VNIIM	0	94	-32	316	7	143	-23	120	-62	102	37	144	-47	140

Key comparison CCT-K7

MEASURAND : Temperature of triple point of water

NOMINAL VALUE : 273.16 K

Matrix of equivalence (Cont.)

Lab *j* →

Lab <i>i</i>			A*STAR		UME		VNIIM	
	<i>D_i</i>	<i>U_i</i>	<i>D_{ij}</i>	<i>U_{ij}</i>	<i>D_{ij}</i>	<i>U_{ij}</i>	<i>D_{ij}</i>	<i>U_{ij}</i>
	/ μK		/ μK		/ μK		/ μK	
BIPM	-22	90	-34	167	53	202	-22	126
LNE-INM	-76	134	-87	194	-1	225	-76	161
CEM	-36	85	-47	164	39	199	-36	123
CENAM	-27	58	-38	152	48	189	-27	106
NMISA	83	150	72	205	158	234	83	174
NMIA	-51	72	-63	158	24	194	-51	114
INRIM	-37	58	-48	152	38	189	-37	106
IPQ	18	322	7	351	93	369	18	334
KRISS	47	115	35	181	121	214	47	145
MSL	95	39	84	146	170	184	95	97
NIM	11	124	0	187	86	219	11	153
NIST	-62	69	-74	156	12	193	-62	112
NMIJ	32	304	21	335	107	353	32	316
NMi-VSL	-6	113	-18	180	68	213	-7	143
NPL	23	81	12	162	98	198	23	120
NRC	62	52	51	150	137	187	62	102
PTB	-37	113	-48	180	38	213	-37	144
SMU	47	109	35	178	122	211	47	140
A*STAR	11	144			86	231	11	169
UME	-75	183	-86	231			-75	203
VNIIM	0	94	-11	169	75	203		

Key comparison EURAMET.T-K7

MEASURAND : Temperature of triple point of water

NOMINAL VALUE : 273.16 K

Matrix of equivalence

Lab j \longrightarrow

Lab i			CEM		INM(RO)		MKEH		INRIM		EIM		MIKES	
	D_i	U_i	D_{ij}	U_{ij}										
	$/ \mu\text{K}$		$/ \mu\text{K}$		$/ \mu\text{K}$		$/ \mu\text{K}$		$/ \mu\text{K}$		$/ \mu\text{K}$		$/ \mu\text{K}$	
CEM	100	98			54	146	129	155	49	98	6	240	68	152
INM(RO)	46	136	-54	146			74	176	-5	136	-48	257	14	178
MKEH	-28	145	-129	155	-74	176			-79	145	-123	262	-61	186
INRIM	51	73	-49	98	5	136	79	145			-44	228	18	132
EIM	95	230	-6	240	48	257	123	262	44	228			62	254
MIKES	33	137	-68	152	-14	178	61	186	-18	132	-62	254		
PTB	88	86	-12	108	42	143	117	152	37	78	-6	231	56	137
LNE-INM	65	120	-35	137	19	166	93	174	14	120	-30	250	32	167
DTI	-2	145	-102	159	-48	184	26	191	-53	145	-97	239	-35	185
JV	-97	390	-197	405	-143	416	-69	419	-148	400	-192	456	-130	416
VNIIM	112	113	12	130	66	161	141	169	61	113	18	246	80	148
MIRS/FE-LMK	82	68	-18	94	36	133	110	142	31	68	-12	229	50	134
DZM/LPM	59	110	-42	128	13	159	87	167	8	110	-36	245	26	160
UME	-37	161	-137	173	-83	197	-8	204	-88	161	-131	271	-69	198
DMDM	-101	242	-201	251	-147	268	-73	273	-152	242	-196	326	-134	269
SMD	94	82	-7	105	47	141	122	150	43	82	-1	233	61	142
BEV	29	352	-72	357	-17	369	57	372	-22	351	-66	413	-4	369
IPQ	79	166	-22	178	33	202	107	208	28	166	-16	275	46	202
NML(IE)	89	212	-11	226	43	248	117	250	38	217	-6	308	56	246
SMU	45	124	-55	140	-1	169	74	176	-6	124	-49	251	13	170
CMI	-261	130	-361	146	-307	173	-233	181	-312	130	-356	255	-294	174
GUM	-260	155	-360	170	-306	194	-231	201	-311	157	-354	269	-292	195
VMT/PFI	-13	243	-114	251	-59	268	15	273	-64	243	-108	327	-46	269
NMi-VSL	63	72	-37	96	17	135	91	144	12	71	-32	230	30	136

Key comparison EURAMET.T-K7

MEASURAND : Temperature of triple point of water

NOMINAL VALUE : 273.16 K

Matrix of equivalence (Cont.)

Lab *j* →

Lab <i>i</i>	PTB		LNE-INM		DTI		JV		VNIIM		MIRS/FE-LMK			
	<i>D_i</i>	<i>U_i</i>	<i>D_{ij}</i>	<i>U_{ij}</i>										
CEM	100	98	12	108	35	137	102	159	197	405	-12	130	18	94
INM(RO)	46	136	-42	143	-19	166	48	184	143	416	-66	161	-36	133
MKEH	-28	145	-117	152	-93	174	-26	191	69	419	-141	169	-110	142
INRIM	51	73	-37	78	-14	120	53	145	148	400	-61	113	-31	68
EIM	95	230	6	231	30	250	97	239	192	456	-18	246	12	229
MIKES	33	137	-56	137	-33	167	35	185	129	416	-80	148	-50	134
PTB	88	86			23	128	90	151	185	402	-24	116	6	81
LNE-INM	65	120	-23	128			67	170	162	409	-47	144	-17	117
DTI	-2	145	-90	151	-67	170			-228	412	-19	151	-49	142
JV	-97	390	-185	402	-162	409	228	412		-209	402	-179	399	
VNIIM	112	113	24	116	47	144	19	151	209	402			30	110
MIRS/FE-LMK	82	68	-6	81	17	117	49	142	179	399	-30	110		
DZM/LPM	59	110	-30	119	-6	146	73	166	156	406	-54	140	-23	100
UME	-37	161	-125	167	-102	187	168	203	60	424	-149	182	-119	154
DMDM	-101	242	-190	247	-167	261	233	273	-5	462	-214	257	-184	238
SMD	94	82	5	93	29	126	38	149	191	401	-19	119	11	77
BEV	29	352	-60	353	-36	116	103	372	126	527	-84	361	-53	350
IPQ	79	166	-10	172	14	192	53	208	176	427	-34	187	-3	164
NML(IE)	89	212	1	221	24	237	42	250	186	449	-23	233	7	215
SMU	45	124	-43	132	-20	157	86	176	142	412	-67	151	-37	121
CMI	-261	130	-349	138	-326	162	393	181	-164	414	-373	156	-343	128
GUM	-260	155	-348	163	-325	184	391	201	-163	423	-372	179	-342	155
VMT/PFI	-13	243	-102	247	-78	261	145	273	84	462	-126	258	-95	241
NMi-VSL	63	72	-25	84	-2	119	68	144	160	397	-49	112	-19	66

Key comparison EURAMET.T-K7

MEASURAND : Temperature of triple point of water

NOMINAL VALUE : 273.16 K

Matrix of equivalence (Cont.)

Lab *j* \longrightarrow

Lab <i>i</i>	DZM/LPM				UME		DMDM		SMD		BEV		IPQ	
	<i>D_i</i>	<i>U_i</i>	<i>D_{ij}</i>	<i>U_{ij}</i>										
CEM	100	98	42	128	137	173	202	251	7	105	72	357	22	178
INM(RO)	46	136	-13	159	83	197	148	268	-47	141	17	369	-33	202
MKEH	-28	145	-87	167	8	204	73	273	-122	150	-57	372	-107	208
INRIM	51	73	-8	110	88	161	152	242	-43	82	22	351	-28	166
EIM	95	230	36	245	131	271	196	326	1	233	66	413	16	275
MIKES	33	137	-26	160	69	198	134	269	-61	142	4	369	-46	202
PTB	88	86	30	119	125	167	190	247	-5	93	60	353	10	172
LNE-INM	65	120	6	146	102	187	167	261	-29	126	36	116	-14	192
DTI	-2	145	-73	166	-168	203	-233	273	-38	149	-103	372	-53	208
JV	-97	390	-156	406	-60	424	5	462	-191	401	-126	527	-176	427
VNIIM	112	113	54	140	149	182	214	257	19	119	84	361	34	187
MIRS/FE-LMK	82	68	23	100	119	154	184	238	-11	77	53	350	3	164
DZM/LPM	59	110			95	175	160	252	-35	116	30	360	-20	185
UME	-37	161	-95	175			65	278	-130	165	-65	379	-115	219
DMDM	-101	242	-160	252	-65	278			-195	245	-130	420	-180	285
SMD	94	82	35	116	130	165	195	245			65	351	15	166
BEV	29	352	-30	360	65	379	130	420	-65	351			-50	379
IPQ	79	166	20	185	115	219	180	285	-15	166	50	379		
NML(IE)	89	212	30	232	126	260	191	317	-4	217	60	403	10	260
SMU	45	124	-13	149	82	189	147	262	-48	130	17	365	-33	194
CMI	-261	130	-320	154	-224	194	-160	265	-355	136	-290	365	-340	198
GUM	-260	155	-318	177	-223	212	-158	279	-353	161	-288	377	-338	217
VMT/PFI	-13	243	-72	256	23	282	88	335	-107	246	-42	420	-92	285
NMi-VSL	63	72	4	109	100	150	165	242	-31	80	34	350	-16	165

Key comparison EURAMET.T-K7

MEASURAND : Temperature of triple point of water

NOMINAL VALUE : 273.16 K

Matrix of equivalence (Cont.)

Lab *j* \longrightarrow

Lab <i>i</i>	NML(IE)		SMU		CMI		GUM		VMT/PFI		NMi-VSL			
	<i>D_i</i>	<i>U_i</i>	<i>D_{ij}</i>	<i>U_{ij}</i>										
CEM	100	98	11	226	55	140	361	146	360	170	114	251	37	96
INM(RO)	46	136	-43	248	1	169	307	173	306	194	59	268	-17	135
MKEH	-28	145	-117	250	-74	176	233	181	231	201	-15	273	-91	144
INRIM	51	73	-38	217	6	124	312	130	311	157	64	243	-12	71
EIM	95	230	6	308	49	251	356	255	354	269	108	327	32	230
MIKES	33	137	-57	246	-13	170	294	174	292	195	46	269	-31	136
PTB	88	86	-1	221	43	132	349	138	348	163	102	247	25	84
LNE-INM	65	120	-24	237	20	157	326	162	325	184	78	261	2	119
DTI	-2	145	-42	250	-86	176	-393	181	-391	201	-145	273	-68	144
JV	-97	390	-186	449	-142	412	164	414	163	423	-84	462	-160	397
VNIIM	112	113	23	233	67	151	373	156	372	179	126	258	49	112
MIRS/FE-LMK	82	68	-7	215	37	121	343	128	342	155	95	241	19	66
DZM/LPM	59	110	-30	232	13	149	320	154	318	177	72	256	-4	109
UME	-37	161	-126	260	-82	189	224	194	223	212	-23	282	-100	150
DMDM	-101	242	-191	317	-147	262	160	265	158	279	-88	335	-165	242
SMD	94	82	4	217	48	130	355	136	353	161	107	246	31	80
BEV	29	352	-60	403	-17	365	290	365	288	377	42	420	-34	350
IPQ	79	166	-10	260	33	194	340	198	338	217	92	285	16	165
NML(IE)	89	212			44	239	350	242	349	258	102	317	26	216
SMU	45	124	-44	239			306	161	305	183	59	260	-18	123
CMI	-261	130	-350	242	-306	161			-1	182	-248	260	-324	129
GUM	-260	155	-349	258	-305	183	1	182			-246	274	-323	156
VMT/PFI	-13	243	-102	317	-59	260	248	260	246	274			-76	242
NMi-VSL	63	72	-26	216	18	123	324	129	323	156	76	242		