

CCM.P-K4 .2012

MEASURAND: Pressure

NOMINAL VALUES: 1 Pa, 3 Pa, 10 Pa, 30 Pa, 100 Pa, 300 Pa, 1 000 Pa, 3 000 Pa, 7 000 Pa and 10 000 Pa

$p_i$ : corrected mean gauge reading obtained by laboratory  $i$  (see CCM.P-K4.2012 Final Report, page 15)

$u_i$ : combined standard uncertainty of  $p_i$  (see CCM.P-K4 Final Report, page 22)

Lab $i$	Nominal pressure / Pa	$p_i$ / Pa	$u_i$ / Pa	Measurement method	Date of measurement
PTB	1	1.002	0.006	Static expansion system and FRS5	June 2012
	3	3.004	0.006		
	10	10.007	0.010		
	30	30.00	0.02		
	100	99.99	0.02		
	300	299.95	0.02		
	1 000	999.83	0.03		
	3 000	2999.56	0.08		
	7 000	6998.94	0.17		
10 000	9998.6	0.2			
CMI	1	1.005	0.019	Force balanced piston gauge	September 2012
	3	3.017	0.019		
	10	10.031	0.019		
	30	30.039	0.019		
	100	100.015	0.019		
	300	299.97	0.02		
	1 000	999.85	0.03		
	3 000	2999.54	0.06		
	7 000	6998.95	0.11		
10 000	9998.73	0.15			
NMIJ	1	1.002	0.007	Static expansion system and calibrated transducer	November 2012
	3	3.006	0.007		
	10	10.01	0.03		
	30	30.00	0.08		
	100	99.96	0.09		
	300	299.92	0.08		
	1 000	999.80	0.09		
	3 000	2999.47	0.16		
	7 000	6998.9	0.3		
10 000	9998.6	0.4			

CCM.P-K4 .2012

MEASURAND: Pressure

NOMINAL VALUES: 1 Pa, 3 Pa, 10 Pa, 30 Pa, 100 Pa, 300 Pa, 1 000 Pa, 3 000 Pa, 7 000 Pa and 10 000 Pa

$p_i$ : corrected mean gauge reading obtained by laboratory  $i$  (see CCM.P-K4.2012 Final Report, page 15)

$u_i$ : combined standard uncertainty of  $p_i$  (see CCM.P-K4 Final Report, page 22)

Lab $i$	Nominal pressure / Pa	$p_i$ / Pa	$u_i$ / Pa	Measurement method	Date of measurement
NIST	1	1.008	0.004	Mercury and oil ultrasonic interferometer manometers	April 2012, January 2013 and September 2013
	3	3.010	0.004		
	10	10.012	0.004		
	30	30.012	0.004		
	100	99.993	0.006		
	300	299.944	0.005		
	1 000	999.799	0.013		
	3 000	2999.43	0.02		
	7 000	6998.73	0.03		
10 000	9998.41	0.04			
CENAM	1	0.992	0.014	Static expansion system and piston gauge	March 2013
	3	2.989	0.015		
	10	9.96	0.02		
	30	29.87	0.05		
	100	99.82	0.11		
	300	299.8	0.3		
	1 000	999.50	1.07		
	3 000	2999.51	0.06		
	7 000	6998.74	0.08		
10 000	9998.39	0.09			
VNIIM	1	1.001	0.014	Mercury and oil laser interferometer manometers	April 2013
	3	2.999	0.014		
	10	10.000	0.015		
	30	29.998	0.015		
	100	99.951	0.014		
	300	299.865	0.019		
	1 000	999.57	0.06		
	3 000	2999.8	0.5		
	7 000	6999.2	0.5		
10 000	9998.9	0.5			

CCM.P-K4 .2012

MEASURAND: Pressure

NOMINAL VALUES: 1 Pa, 3 Pa, 10 Pa, 30 Pa, 100 Pa, 300 Pa, 1 000 Pa, 3 000 Pa, 7 000 Pa and 10 000 Pa

For each nominal pressure value, the key comparison reference value,  $p_R$ , is obtained from an unweighted mean of the measurement method means, excluding CENAM for the determination of  $p_R$  at 10 Pa and 30 Pa.

The combined standard uncertainty of  $p_R$  is  $u_R$ .

Nominal pressure / Pa	1	3	10	30	100
$p_R$ / Pa	1.002	3.004	10.003	29.987	99.955
$u_R$ / Pa	1.01E-05	3.00E-05	4.63E-05	1.12E-05	1.77E-03

Nominal pressure / Pa	300	1000	3000	7000	10000
$p_R$ / Pa	299.904	999.726	2999.553	6998.902	9998.604
$u_R$ / Pa	1.60E-03	7.56E-03	5.56E-03	9.38E-03	1.29E-02

The degree of equivalence of each laboratory with respect to the reference value is given by a pair of terms:  $D_i = (p_i - p_R)$  and  $U_i$ , its expanded uncertainty at a 95% level of confidence, both expressed in Pa.

The degree of equivalence between two laboratories is given at p. 29 of the Final Report.

Lab <i>i</i>	1 Pa	
	$D_i$ / Pa	$U_i$ / Pa
PTB	0.000	0.014
CMI	0.003	0.032
NMIJ	0.001	0.014
NIST	0.006	0.012
CENAM	-0.010	0.025
VNIIM	-0.001	0.025

	3 Pa	
	$D_i$ / Pa	$U_i$ / Pa
	0.000	0.014
	0.013	0.032
	0.001	0.015
	0.006	0.012
	-0.015	0.026
	-0.005	0.025

	10 Pa	
	$D_i$ / Pa	$U_i$ / Pa
	0.004	0.021
	0.028	0.034
	0.005	0.048
	0.010	0.016
	-0.044	0.041
	-0.003	0.028

	30 Pa	
	$D_i$ / Pa	$U_i$ / Pa
	0.018	0.050
	0.052	0.046
	0.013	0.135
	0.024	0.035
	-0.118	0.095
	0.011	0.042

	100 Pa	
	$D_i$ / Pa	$U_i$ / Pa
	0.039	0.053
	0.061	0.049
	0.004	0.059
	0.038	0.039
	-0.138	0.177
	-0.004	0.045

Lab <i>i</i>	300 Pa	
	$D_i$ / Pa	$U_i$ / Pa
PTB	0.042	0.117
CMI	0.063	0.115
NMIJ	0.017	0.171
NIST	0.039	0.110
CENAM	-0.122	0.528
VNIIM	-0.039	0.114

	1 000 Pa	
	$D_i$ / Pa	$U_i$ / Pa
	0.11	0.36
	0.12	0.36
	0.08	0.39
	0.07	0.36
	-0.23	1.78
	-0.15	0.37

	3 000 Pa	
	$D_i$ / Pa	$U_i$ / Pa
	0.01	0.21
	-0.01	0.19
	-0.08	0.31
	-0.12	0.17
	-0.04	0.20
	0.25	0.79





	7 000 Pa	
	$D_i$ / Pa	$U_i$ / Pa
	0.04	0.34
	0.05	0.27
	-0.03	0.49
	-0.18	0.21
	-0.17	0.24
	0.28	0.82

	10 000 Pa	
	$D_i$ / Pa	$U_i$ / Pa
	-0.03	0.46
	0.12	0.34
	0.00	0.62
	-0.19	0.24
	-0.21	0.27
	0.31	0.85

CCM.P-K4 .2012

MEASURAND: Pressure





NOMINAL VALUES: 1 Pa, 3 Pa, 10 Pa, 30 Pa, 100 Pa, 300 Pa, 1 000 Pa, 3 000 Pa, 7 000 Pa and 10 000 Pa



CCM.P-K4 .2012

MEASURAND: Pressure

NOMINAL VALUES: 1 Pa, 3 Pa, 10 Pa, 30 Pa, 100 Pa, 300 Pa, 1 000 Pa, 3 000 Pa, 7 000 Pa and 10 000 Pa



CCM.P-K4 .2012

MEASURAND: Pressure

NOMINAL VALUES: 1 Pa, 3 Pa, 10 Pa, 30 Pa, 100 Pa, 300 Pa, 1 000 Pa, 3 000 Pa, 7 000 Pa and 10 000 Pa

