

Key comparison CCT-K1

MEASURAND : Temperature T_{90} / K

Equivalence statements

For each temperature in the comparison the key comparison reference value, T_R , is the weighted mean of T_i , the individual temperature values of thermometers.

The weighted mean is calculated using the laboratory uncertainty combined with the comparison uncertainty, to set the weights. T_R is used as the baseline for the comparison, but has no special significance with respect to the ITS-90, and is used without uncertainty.

The degree of equivalence of each temperature T_i with respect to the key comparison reference value, T_R , is given by a pair of terms: $D_i = (T_i - T_R)$ and U_i , its expanded uncertainty at 95 % confidence, both expressed in mK.
 U_i includes the uncertainties in the original laboratory calibrations and in the comparison measurements, but not in T_R .

The degrees of equivalence between each pair of thermometers is given by:

$D_{ij} = (D_i - D_j) = (T_i - T_R) - (T_j - T_R)$ and $U_{ij} = (U_i^2 + U_j^2)^{1/2}$, its expanded uncertainty (at 95 % confidence), both expressed in mK.

T_R / K	T_R / K	T_R / K	T_R / K
0.649875	1.996554	4.477522	16.999335
0.676928	2.248485	5.000458	18.597377
0.704354	2.600776	5.948165	20.298899
0.761580	2.699911	7.201544	21.575444
0.858421	2.896733	8.296372	22.676998
0.991223	2.996648	8.399612	23.496448
1.031584	3.099398	9.508020	24.101970
1.224991	3.400235	10.803390	24.340317
1.249542	3.429250	12.297309	24.446403
1.503370	3.800903	13.798183	24.551354
1.754822	4.224794	15.499566	

Key comparison CCT-K1

MEASURAND : Temperature T_{90} / K

NOMINAL TEMPERATURE : $T_{90} = 0.650$ K

Laboratory individual measurements

Lab: Laboratory

S/N i : serial number of Thermometer i

T_i : temperature value for Thermometer i

u_i : standard uncertainty in the calibration of Thermometer i

u_{comp} : standard uncertainty in the comparison measurements

$$u_{\text{comp}} = 0.205 \text{ mK}$$

Lab, S/N i	T_i / mK	u_i / mK
NIST A128	0.650152	0.139
NIST A129	0.649998	0.139
NMi-VSL 226246	0.650370	0.182
NPL 221481	0.649615	0.254
NPL 221483	0.649625	0.254
NPL 221485	0.649681	0.254
PTB 229074	0.649546	0.180
PTB 229075	0.649551	0.180
VNIIFTRI 79	0.650145	0.540
VNIIFTRI 89	0.650376	0.550

Key comparison CCT-K1

MEASURAND : Temperature T_{90} / K

KEY COMPARISON REFERENCE VALUE: $T_R = 0.649875$ K


Matrix of equivalence

Lab, S/N <i>i</i>	Lab, S/N <i>j</i> \longrightarrow													
	NIST A128		NIST A129		NMi-VSL 226246		NPL 221481		NPL 221483		NPL 221485			
	D_i	U_i		D_{ij}	U_{ij}		D_{ij}	U_{ij}		D_{ij}	U_{ij}			
NIST A128	0.277	0.495		0.154	0.700	-0.218	0.739	0.536	0.819	0.526	0.819	0.470	0.819	
NIST A129	0.123	0.495	-0.154	0.700		-0.372	0.739	0.383	0.819	0.373	0.819	0.316	0.819	
NMi-VSL 226246	0.495	0.548	0.218	0.739	0.372	0.739		0.755	0.852	0.745	0.852	0.689	0.852	
NPL 221481	-0.260	0.653	-0.536	0.819	-0.383	0.819	-0.755	0.852		-0.010	0.923	-0.066	0.923	
NPL 221483	-0.250	0.653	-0.526	0.819	-0.373	0.819	-0.745	0.852	0.010	0.923		-0.056	0.923	
NPL 221485	-0.194	0.653	-0.470	0.819	-0.316	0.819	-0.689	0.852	0.066	0.923	0.056	0.923		
PTB 229074	-0.329	0.546	-0.606	0.737	-0.452	0.737	-0.824	0.773	-0.069	0.851	-0.079	0.851	-0.135	0.851
PTB 229075	-0.324	0.546	-0.601	0.737	-0.447	0.737	-0.819	0.773	-0.064	0.851	-0.074	0.851	-0.130	0.851
VNIIFTRI 79	0.270	1.155	-0.007	1.257	0.147	1.257	-0.225	1.279	0.529	1.327	0.519	1.327	0.463	1.327
VNIIFTRI 89	0.501	1.174	0.224	1.274	0.378	1.274	0.005	1.295	0.760	1.343	0.750	1.343	0.694	1.343

Matrix of equivalence (Continued)

Lab, S/N <i>i</i>	Lab, S/N <i>j</i> \longrightarrow										
	PTB 229074		PTB 229075		VNIIIFTRI 79		VNIIIFTRI 89				
	D_i	U_i		D_{ij}	U_{ij}		D_{ij}	U_{ij}			
	/ mK	/ mK		/ mK	/ mK		/ mK	/ mK			
NIST A128	0.277	0.495		0.606	0.737	0.601	0.737	0.007	1.257	-0.224	1.274
NIST A129	0.123	0.495		0.452	0.737	0.447	0.737	-0.147	1.257	-0.378	1.274
NMi-VSL 226246	0.495	0.548		0.824	0.773	0.819	0.773	0.225	1.279	-0.005	1.295
NPL 221481	-0.260	0.653		0.069	0.851	0.064	0.851	-0.529	1.327	-0.760	1.343
NPL 221483	-0.250	0.653		0.079	0.851	0.074	0.851	-0.519	1.327	-0.750	1.343
NPL 221485	-0.194	0.653		0.135	0.851	0.130	0.851	-0.463	1.327	-0.694	1.343
PTB 229074	-0.329	0.546				-0.005	0.772	-0.599	1.278	-0.830	1.294
PTB 229075	-0.324	0.546		0.005	0.772			-0.594	1.278	-0.825	1.294
VNIIIFTRI 79	0.270	1.155		0.599	1.278	0.594	1.278			-0.231	1.647
VNIIIFTRI 89	0.501	1.174		0.8296	1.2945	0.825	1.294	0.231	1.647		

CCT-K1 : Nominal temperature, $T_{90} = 0.650$ K
Degrees of equivalence, D_i , and expanded uncertainties ($k = 2$), U_i , expressed in mK



Key comparison CCT-K1

MEASURAND : Temperature T_{90} / K

NOMINAL TEMPERATURE : $T_{90} = 0.677$ K

Laboratory individual measurements

Lab: Laboratory

S/N i : serial number of Thermometer i

T_i : temperature value for Thermometer i

u_i : standard uncertainty in the calibration of Thermometer i

u_{comp} : standard uncertainty in the comparison measurements

$$u_{\text{comp}} = 0.201 \text{ mK}$$

Lab, S/N i	T_i / mK	u_i / mK
NIST A128	0.677079	0.121
NIST A129	0.676957	0.121
NMi-VSL 226246	0.677410	0.182
NPL 221481	0.677030	0.245
NPL 221483	0.676970	0.245
NPL 221485	0.676438	0.245
PTB 229074	0.676573	0.180
PTB 229075	0.676440	0.180
VNIIFTRI 79	0.677630	0.533
VNIIFTRI 89	0.677956	0.548

Key comparison CCT-K1

MEASURAND : Temperature T_{90} / K

KEY COMPARISON REFERENCE VALUE: $T_R = 0.676928$ K


Matrix of equivalence

Lab, S/N <i>i</i>	Lab, S/N <i>j</i> \longrightarrow														
	NIST A128		NIST A129		NMi-VSL 226246		NPL 221481		NPL 221483		NPL 221485				
	D_i	U_i		D_{ij}	U_{ij}		D_{ij}	U_{ij}		D_{ij}	U_{ij}				
	/ mK	/ mK		/ mK	/ mK		/ mK	/ mK		/ mK	/ mK				
NIST A128	0.151	0.470			0.122	0.664	-0.330	0.718	0.049	0.790	0.109	0.790	0.641	0.790	
NIST A129	0.029	0.470		-0.122	0.664		-0.452	0.718	-0.073	0.790	-0.013	0.790	0.519	0.790	
NMi-VSL 226246	0.482	0.543		0.330	0.718	0.452	0.718		0.380	0.835	0.440	0.835	0.971	0.835	
NPL 221481	0.102	0.635		-0.049	0.790	0.073	0.790	-0.380	0.835		0.060	0.898	0.592	0.898	
NPL 221483	0.042	0.635		-0.109	0.790	0.013	0.790	-0.440	0.835	-0.060	0.898		0.532	0.898	
NPL 221485	-0.490	0.635		-0.641	0.790	-0.519	0.790	-0.971	0.835	-0.592	0.898	-0.532	0.898		
PTB 229074	-0.355	0.540		-0.506	0.716	-0.384	0.716	-0.836	0.766	-0.457	0.833	-0.397	0.833	0.135	0.833
PTB 229075	-0.488	0.540		-0.639	0.716	-0.517	0.716	-0.970	0.766	-0.590	0.833	-0.530	0.833	0.002	0.833
VNIIFTRI 79	0.702	1.139		0.551	1.232	0.673	1.232	0.220	1.262	0.600	1.304	0.660	1.304	1.192	1.304
VNIIFTRI 89	1.028	1.167		0.877	1.258	0.999	1.258	0.547	1.287	0.926	1.328	0.986	1.328	1.518	1.328

Matrix of equivalence (Continued)

Lab, S/N <i>i</i>	Lab, S/N <i>j</i> \longrightarrow										
	PTB 229074		PTB 229075		VNIIIFTRI 79		VNIIIFTRI 89				
	D_i	U_i		D_{ij}	U_{ij}		D_{ij}	U_{ij}			
	/ mK	/ mK		/ mK	/ mK		/ mK	/ mK			
NIST A128	0.151	0.470		0.506	0.716	0.639	0.716	-0.551	1.232	-0.877	1.258
NIST A129	0.029	0.470		0.384	0.716	0.517	0.716	-0.673	1.232	-0.999	1.258
NMi-VSL 226246	0.482	0.543		0.836	0.766	0.970	0.766	-0.220	1.262	-0.547	1.287
NPL 221481	0.102	0.635		0.457	0.833	0.590	0.833	-0.600	1.304	-0.926	1.328
NPL 221483	0.042	0.635		0.397	0.833	0.530	0.833	-0.660	1.304	-0.986	1.328
NPL 221485	-0.490	0.635		-0.135	0.833	-0.002	0.833	-1.192	1.304	-1.518	1.328
PTB 229074	-0.355	0.540				0.133	0.764	-1.057	1.261	-1.383	1.286
PTB 229075	-0.488	0.540				-0.133	0.764	-1.190	1.261	-1.516	1.286
VNIIIFTRI 79	0.702	1.139				1.057	1.261	1.190	1.261	-0.326	1.631
VNIIIFTRI 89	1.028	1.167				1.383	1.2858	1.516	1.286	0.326	1.631

CCT-K1 : Nominal temperature, $T_{90} = 0.677$ K
Degrees of equivalence, D_i , and expanded uncertainties ($k = 2$), U_i , expressed in mK



Key comparison CCT-K1

MEASURAND : Temperature T_{90} / K

NOMINAL TEMPERATURE : $T_{90} = 0.704$ K

Laboratory individual measurements

Lab: Laboratory

S/N i : serial number of Thermometer i

T_i : temperature value for Thermometer i

u_i : standard uncertainty in the calibration of Thermometer i

u_{comp} : standard uncertainty in the comparison measurements

$$u_{\text{comp}} = 0.198 \text{ mK}$$

Lab, S/N i	T_i / mK	u_i / mK
NIST A128	0.704853	0.110
NIST A129	0.704179	0.110
NMi-VSL 226246	0.704626	0.182
NPL 221481	0.704350	0.236
NPL 221483	0.704305	0.236
NPL 221485	0.704029	0.236
PTB 229074	0.704040	0.180
PTB 229075	0.703875	0.180
VNIIFTRI 79	0.705043	0.526
VNIIFTRI 89	0.705293	0.545

Key comparison CCT-K1

MEASURAND : Temperature T_{90} / K

KEY COMPARISON REFERENCE VALUE: $T_R = 0.704354$ K


Matrix of equivalence

Lab, S/N <i>i</i>	Lab, S/N <i>j</i> \longrightarrow														
	NIST A128		NIST A129		NMi-VSL 226246		NPL 221481		NPL 221483		NPL 221485				
	D_i	U_i		D_{ij}	U_{ij}		D_{ij}	U_{ij}		D_{ij}	U_{ij}				
	/ mK	/ mK		/ mK	/ mK		/ mK	/ mK		/ mK	/ mK				
NIST A128	0.499	0.453			0.674	0.640	0.227	0.702	0.503	0.765	0.548	0.765	0.824	0.765	
NIST A129	-0.175	0.453		-0.674	0.640		-0.447	0.702	-0.171	0.765	-0.126	0.765	0.150	0.765	
NMi-VSL 226246	0.272	0.537		-0.227	0.702	0.447	0.702		0.276	0.817	0.321	0.817	0.598	0.817	
NPL 221481	-0.004	0.616		-0.503	0.765	0.171	0.765	-0.276	0.817		0.045	0.872	0.322	0.872	
NPL 221483	-0.049	0.616		-0.548	0.765	0.126	0.765	-0.321	0.817	-0.045	0.872		0.277	0.872	
NPL 221485	-0.325	0.616		-0.824	0.765	-0.150	0.765	-0.598	0.817	-0.322	0.872	-0.277	0.872		
PTB 229074	-0.314	0.535		-0.813	0.701	-0.139	0.701	-0.586	0.758	-0.310	0.816	-0.265	0.816	0.012	0.816
PTB 229075	-0.479	0.535		-0.978	0.701	-0.304	0.701	-0.752	0.758	-0.476	0.816	-0.431	0.816	-0.154	0.816
VNIIFTRI 79	0.689	1.124		0.190	1.211	0.864	1.211	0.416	1.245	0.693	1.281	0.738	1.281	1.014	1.281
VNIIFTRI 89	0.939	1.160		0.440	1.245	1.114	1.245	0.667	1.278	0.943	1.314	0.988	1.314	1.264	1.314

Matrix of equivalence (Continued)

Lab, S/N <i>i</i>	Lab, S/N <i>j</i> \longrightarrow											
	PTB 229074		PTB 229075		VNIIIFTRI 79		VNIIIFTRI 89					
	D_i	U_i		D_{ij}	U_{ij}		D_{ij}	U_{ij}				
	/ mK	/ mK		/ mK	/ mK		/ mK	/ mK				
NIST A128	0.499	0.453		0.813	0.701	0.978	0.701	-0.190	1.211	-0.440	1.245	
NIST A129	-0.175	0.453		0.139	0.701	0.304	0.701	-0.864	1.211	-1.114	1.245	
NMi-VSL 226246	0.272	0.537		0.586	0.758	0.752	0.758	-0.416	1.245	-0.667	1.278	
NPL 221481	-0.004	0.616		0.310	0.816	0.476	0.816	-0.693	1.281	-0.943	1.314	
NPL 221483	-0.049	0.616		0.265	0.816	0.431	0.816	-0.738	1.281	-0.988	1.314	
NPL 221485	-0.325	0.616		-0.012	0.816	0.154	0.816	-1.014	1.281	-1.264	1.314	
PTB 229074	-0.314	0.535				0.166	0.756	-1.003	1.244	-1.253	1.277	
PTB 229075	-0.479	0.535				-0.166	0.756		-1.168	1.244	-1.418	1.277
VNIIIFTRI 79	0.689	1.124				1.003	1.244	1.168			-0.250	1.615
VNIIIFTRI 89	0.939	1.160				1.2527	1.2772	1.418	1.277	0.250	1.615	

CCT-K1 : Nominal temperature, $T_{90} = 0.704$ K
Degrees of equivalence, D_i , and expanded uncertainties ($k = 2$), U_i , expressed in mK



Key comparison CCT-K1

MEASURAND : Temperature T_{90} / K

NOMINAL TEMPERATURE : $T_{90} = 0.762$ K

Laboratory individual measurements

Lab: Laboratory

S/N i : serial number of Thermometer i

T_i : temperature value for Thermometer i

u_i : standard uncertainty in the calibration of Thermometer i

u_{comp} : standard uncertainty in the comparison measurements

$$u_{\text{comp}} = 0.190 \text{ mK}$$

Lab, S/N i	T_i / mK	u_i / mK
NIST A128	0.761654	0.082
NIST A129	0.761491	0.082
NMi-VSL 226246	0.761866	0.182
NPL 221481	0.761820	0.218
NPL 221483	0.761770	0.218
NPL 221485	0.761112	0.218
PTB 229074	0.761082	0.180
PTB 229075	0.761437	0.180
VNIIFTRI 79	0.762375	0.511
VNIIFTRI 89	0.762682	0.540

Key comparison CCT-K1

MEASURAND : Temperature T_{90} / K

KEY COMPARISON REFERENCE VALUE: $T_R = 0.761580$ K


Matrix of equivalence

Lab, S/N <i>i</i>	Lab, S/N <i>j</i> \longrightarrow														
	NIST A128		NIST A129		NMi-VSL 226246		NPL 221481		NPL 221483		NPL 221485				
	D_i	U_i		D_{ij}	U_{ij}		D_{ij}	U_{ij}		D_{ij}	U_{ij}				
	/ mK	/ mK		/ mK	/ mK		/ mK	/ mK		/ mK	/ mK				
NIST A128	0.074	0.413			0.163	0.585	-0.213	0.669	-0.166	0.711	-0.116	0.711	0.541	0.711	
NIST A129	-0.089	0.413		-0.163	0.585		-0.376	0.669	-0.329	0.711	-0.279	0.711	0.379	0.711	
NMi-VSL 226246	0.286	0.526		0.213	0.669	0.376	0.669		0.046	0.781	0.096	0.781	0.754	0.781	
NPL 221481	0.240	0.578		0.166	0.711	0.329	0.711	-0.046	0.781		0.050	0.817	0.708	0.817	
NPL 221483	0.190	0.578		0.116	0.711	0.279	0.711	-0.096	0.781	-0.050	0.817		0.658	0.817	
NPL 221485	-0.468	0.578		-0.541	0.711	-0.379	0.711	-0.754	0.781	-0.708	0.817	-0.658	0.817		
PTB 229074	-0.498	0.523		-0.572	0.667	-0.409	0.667	-0.785	0.742	-0.738	0.780	-0.688	0.780	-0.031	0.780
PTB 229075	-0.143	0.523		-0.217	0.667	-0.054	0.667	-0.430	0.742	-0.383	0.780	-0.333	0.780	0.324	0.780
VNIIFTRI 79	0.795	1.090		0.721	1.166	0.884	1.166	0.508	1.211	0.555	1.234	0.605	1.234	1.262	1.234
VNIIFTRI 89	1.102	1.145		1.028	1.218	1.191	1.218	0.815	1.260	0.862	1.283	0.912	1.283	1.569	1.283

Matrix of equivalence (Continued)

Lab, S/N <i>i</i>	Lab, S/N <i>j</i> \longrightarrow										
	PTB 229074		PTB 229075		VNIIIFTRI 79		VNIIIFTRI 89				
	D_i	U_i		D_{ij}	U_{ij}		D_{ij}	U_{ij}			
	/ mK	/ mK		/ mK	/ mK		/ mK	/ mK			
NIST A128	0.074	0.413		0.572	0.667	0.217	0.667	-0.721	1.166	-1.028	1.218
NIST A129	-0.089	0.413		0.409	0.667	0.054	0.667	-0.884	1.166	-1.191	1.218
NMi-VSL 226246	0.286	0.526		0.785	0.742	0.430	0.742	-0.508	1.211	-0.815	1.260
NPL 221481	0.240	0.578		0.738	0.780	0.383	0.780	-0.555	1.234	-0.862	1.283
NPL 221483	0.190	0.578		0.688	0.780	0.333	0.780	-0.605	1.234	-0.912	1.283
NPL 221485	-0.468	0.578		0.031	0.780	-0.324	0.780	-1.262	1.234	-1.569	1.283
PTB 229074	-0.498	0.523				-0.355	0.740	-1.293	1.210	-1.600	1.259
PTB 229075	-0.143	0.523		0.355	0.740			-0.938	1.210	-1.245	1.259
VNIIIFTRI 79	0.795	1.090		1.293	1.210	0.938	1.210			-0.307	1.582
VNIIIFTRI 89	1.102	1.145		1.6001	1.2593	1.245	1.259	0.307	1.582		

CCT-K1 : Nominal temperature, $T_{90} = 0.762$ K
Degrees of equivalence, D_i , and expanded uncertainties ($k = 2$), U_i , expressed in mK



Key comparison CCT-K1

MEASURAND : Temperature T_{90} / K

NOMINAL TEMPERATURE : $T_{90} = 0.858$ K

Laboratory individual measurements

Lab: Laboratory

S/N i : serial number of Thermometer i

T_i : temperature value for Thermometer i

u_i : standard uncertainty in the calibration of Thermometer i

u_{comp} : standard uncertainty in the comparison measurements

$$u_{\text{comp}} = 0.177 \text{ mK}$$

Lab, S/N i	T_i / mK	u_i / mK
NIST A128	0.858267	0.076
NIST A129	0.858267	0.076
NMi-VSL 226246	0.858603	0.182
NPL 221481	0.858740	0.186
NPL 221483	0.858680	0.186
NPL 221485	0.858267	0.186
PTB 229074	0.858122	0.180
PTB 229075	0.858288	0.180
VNIIFTRI 79	0.859183	0.486
VNIIFTRI 89	0.859376	0.532

Key comparison CCT-K1

MEASURAND : Temperature T_{90} / K

KEY COMPARISON REFERENCE VALUE: $T_R = 0.858421$ K


Matrix of equivalence

Lab, S/N <i>i</i>	Lab, S/N <i>j</i> \longrightarrow														
	NIST A128		NIST A129		NMi-VSL 226246		NPL 221481		NPL 221483		NPL 221485				
	D_i	U_i		D_{ij}	U_{ij}		D_{ij}	U_{ij}		D_{ij}	U_{ij}				
	/ mK	/ mK		/ mK	/ mK		/ mK	/ mK		/ mK	/ mK				
NIST A128	-0.154	0.384			0.000	0.544	-0.336	0.636	-0.473	0.642	-0.413	0.642	0.000	0.642	
NIST A129	-0.154	0.384		0.000	0.544		-0.336	0.636	-0.473	0.642	-0.413	0.642	0.000	0.642	
NMi-VSL 226246	0.182	0.507		0.336	0.636	0.336	0.636		-0.137	0.722	-0.077	0.722	0.336	0.722	
NPL 221481	0.319	0.514		0.473	0.642	0.473	0.642	0.137	0.722		0.060	0.727	0.473	0.727	
NPL 221483	0.259	0.514		0.413	0.642	0.413	0.642	0.077	0.722	-0.060	0.727		0.413	0.727	
NPL 221485	-0.154	0.514		0.000	0.642	0.000	0.642	-0.336	0.722	-0.473	0.727	-0.413	0.727		
PTB 229074	-0.299	0.505		-0.145	0.634	-0.145	0.634	-0.481	0.715	-0.619	0.720	-0.559	0.720	-0.145	0.720
PTB 229075	-0.133	0.505		0.021	0.634	0.021	0.634	-0.315	0.715	-0.452	0.720	-0.392	0.720	0.021	0.720
VNIIFTRI 79	0.762	1.035		0.917	1.104	0.917	1.104	0.581	1.152	0.443	1.155	0.503	1.155	0.916	1.155
VNIIFTRI 89	0.955	1.121		1.109	1.185	1.109	1.185	0.773	1.231	0.636	1.233	0.696	1.233	1.109	1.233

Matrix of equivalence (Continued)

Lab, S/N <i>i</i>	Lab, S/N <i>j</i> \longrightarrow										
	PTB 229074		PTB 229075		VNIIIFTRI 79		VNIIIFTRI 89				
	D_i	U_i		D_{ij}	U_{ij}		D_{ij}	U_{ij}			
	/ mK	/ mK		/ mK	/ mK		/ mK	/ mK			
NIST A128	-0.154	0.384		0.145	0.634	-0.021	0.634	-0.917	1.104	-1.109	1.185
NIST A129	-0.154	0.384		0.145	0.634	-0.021	0.634	-0.917	1.104	-1.109	1.185
NMi-VSL 226246	0.182	0.507		0.481	0.715	0.315	0.715	-0.581	1.152	-0.773	1.231
NPL 221481	0.319	0.514		0.619	0.720	0.452	0.720	-0.443	1.155	-0.636	1.233
NPL 221483	0.259	0.514		0.559	0.720	0.392	0.720	-0.503	1.155	-0.696	1.233
NPL 221485	-0.154	0.514		0.145	0.720	-0.021	0.720	-0.916	1.155	-1.109	1.233
PTB 229074	-0.299	0.505				-0.166	0.714	-1.062	1.151	-1.255	1.230
PTB 229075	-0.133	0.505		0.166	0.714			-0.895	1.151	-1.088	1.230
VNIIIFTRI 79	0.762	1.035		1.062	1.151	0.895	1.151			-0.193	1.526
VNIIIFTRI 89	0.955	1.121		1.2545	1.2296	1.088	1.230	0.193	1.526		

CCT-K1 : Nominal temperature, $T_{90} = 0.858$ K
Degrees of equivalence, D_i , and expanded uncertainties ($k = 2$), U_i , expressed in mK



Key comparison CCT-K1

MEASURAND : Temperature T_{90} / K

NOMINAL TEMPERATURE : $T_{90} = 0.991$ K

Laboratory individual measurements

Lab: Laboratory

S/N i : serial number of Thermometer i

T_i : temperature value for Thermometer i

u_i : standard uncertainty in the calibration of Thermometer i

u_{comp} : standard uncertainty in the comparison measurements

$$u_{\text{comp}} = 0.159 \text{ mK}$$

Lab, S/N i	T_i / mK	u_i / mK
NIST A128	0.991179	0.078
NIST A129	0.991012	0.078
NMi-VSL 226246	0.991361	0.182
NPL 221481	0.991590	0.144
NPL 221485	0.991375	0.144
PTB 229074	0.991133	0.180
PTB 229075	0.990995	0.180

Key comparison CCT-K1


MEASURAND : Temperature T_{90} / K

KEY COMPARISON REFERENCE VALUE: $T_R = 0.991223$ K

Matrix of equivalence

Lab, S/N i	Lab, S/N j \longrightarrow													
	NIST A128		NIST A129		NMi-VSL 226246		NPL 221481		NPL 221485		PTB 229074		PTB 229075	
	D_i	U_i	D_{ij}	U_{ij}	D_{ij}	U_{ij}	D_{ij}	U_{ij}	D_{ij}	U_{ij}	D_{ij}	U_{ij}	D_{ij}	U_{ij}
NIST A128	-0.044	0.354			0.168	0.500	-0.182	0.599	-0.411	0.556	-0.196	0.556	0.046	0.597
NIST A129	-0.211	0.354	-0.168	0.500			-0.349	0.599	-0.578	0.556	-0.363	0.556	-0.121	0.597
NMi-VSL 226246	0.138	0.483	0.182	0.599	0.349	0.599			-0.229	0.646	-0.014	0.646	0.228	0.681
NPL 221481	0.367	0.429	0.411	0.556	0.578	0.556	0.229	0.646			0.215	0.607	0.457	0.644
NPL 221485	0.152	0.429	0.196	0.556	0.363	0.556	0.014	0.646	-0.215	0.607			0.242	0.644
PTB 229074	-0.090	0.480	-0.046	0.597	0.121	0.597	-0.228	0.681	-0.457	0.644	-0.242	0.644		0.138
PTB 229075	-0.228	0.480	-0.184	0.597	-0.016	0.597	-0.366	0.681	-0.595	0.644	-0.380	0.644	-0.138	0.679

CCT-K1 : Nominal temperature, $T_{90} = 0.991$ K
Degrees of equivalence, D_i , and expanded uncertainties ($k = 2$), U_i , expressed in mK



Key comparison CCT-K1

MEASURAND : Temperature T_{90} / K

NOMINAL TEMPERATURE : $T_{90} = 1.032$ K

Laboratory individual measurements

Lab: Laboratory

S/N i : serial number of Thermometer i

T_i : temperature value for Thermometer i

u_i : standard uncertainty in the calibration of Thermometer i

u_{comp} : standard uncertainty in the comparison measurements

$$u_{\text{comp}} = 0.157 \text{ mK}$$

Lab, S/N i	T_i / mK	u_i / mK
NIST A128	1.031395	0.078
NIST A129	1.031369	0.078
NMi-VSL 226246	1.031678	0.182
NPL 221481	1.031835	0.145
NPL 221483	1.031765	0.145
NPL 221485	1.031717	0.145
PTB 229074	1.031415	0.180
PTB 229075	1.031397	0.180
VNIIFTRI 79	1.032281	0.454
VNIIFTRI 89	1.032306	0.526

Key comparison CCT-K1

MEASURAND : Temperature T_{90} / K

KEY COMPARISON REFERENCE VALUE: $T_R = 1.031584$ K


Matrix of equivalence

Lab, S/N <i>i</i>	Lab, S/N <i>j</i> \longrightarrow												
	NIST A128		NIST A129		NMi-VSL 226246		NPL 221481		NPL 221483		NPL 221485		
	D_i	U_i		D_{ij}	U_{ij}		D_{ij}	U_{ij}		D_{ij}	U_{ij}		
	/ mK	/ mK		/ mK	/ mK		/ mK	/ mK		/ mK	/ mK		
NIST A128	-0.189	0.351											
NIST A129	-0.215	0.351		0.026	0.496	-0.283	0.595	-0.440	0.553	-0.370	0.553	-0.322	0.553
NMi-VSL 226246	0.094	0.481		-0.026	0.496	-0.310	0.595	-0.467	0.553	-0.397	0.553	-0.348	0.553
NPL 221481	0.251	0.428		0.283	0.595	0.310	0.595	-0.157	0.644	-0.087	0.644	-0.039	0.644
NPL 221483	0.181	0.428		0.440	0.553	0.467	0.553	0.157	0.644	0.070	0.605	0.118	0.605
NPL 221485	0.133	0.428		0.370	0.553	0.397	0.553	0.087	0.644	-0.070	0.605	0.048	0.605
PTB 229074	-0.169	0.478		0.322	0.553	0.348	0.553	0.039	0.644	-0.118	0.605	-0.048	0.605
PTB 229075	-0.187	0.478		0.021	0.593	0.047	0.593	-0.263	0.678	-0.420	0.642	-0.350	0.642
VNIIFTRI 79	0.697	0.961		0.002	0.593	0.028	0.593	-0.281	0.678	-0.438	0.642	-0.368	0.642
VNIIFTRI 89	0.722	1.099		0.886	1.023	0.913	1.023	0.603	1.074	0.446	1.052	0.516	1.052
				0.911	1.153	0.937	1.153	0.628	1.199	0.471	1.179	0.541	1.179
												0.589	1.179

Matrix of equivalence (Continued)

Lab, S/N <i>i</i>	Lab, S/N <i>j</i> \longrightarrow										
	PTB 229074		PTB 229075		VNIIIFTRI 79		VNIIIFTRI 89				
	D_i	U_i		D_{ij}	U_{ij}		D_{ij}	U_{ij}			
	/ mK	/ mK		/ mK	/ mK		/ mK	/ mK			
NIST A128	-0.189	0.351		-0.021	0.593	-0.002	0.593	-0.886	1.023	-0.911	1.153
NIST A129	-0.215	0.351		-0.047	0.593	-0.028	0.593	-0.913	1.023	-0.937	1.153
NMi-VSL 226246	0.094	0.481		0.263	0.678	0.281	0.678	-0.603	1.074	-0.628	1.199
NPL 221481	0.251	0.428		0.420	0.642	0.438	0.642	-0.446	1.052	-0.471	1.179
NPL 221483	0.181	0.428		0.350	0.642	0.368	0.642	-0.516	1.052	-0.541	1.179
NPL 221485	0.133	0.428		0.302	0.642	0.320	0.642	-0.564	1.052	-0.589	1.179
PTB 229074	-0.169	0.478				0.018	0.676	-0.866	1.073	-0.890	1.198
PTB 229075	-0.187	0.478		-0.018	0.676			-0.884	1.073	-0.909	1.198
VNIIIFTRI 79	0.697	0.961		0.866	1.073	0.884	1.073			-0.025	1.460
VNIIIFTRI 89	0.722	1.099		0.8904	1.1983	0.909	1.198	0.025	1.460		

CCT-K1 : Nominal temperature, $T_{90} = 1.032$ K
Degrees of equivalence, D_i , and expanded uncertainties ($k = 2$), U_i , expressed in mK



Key comparison CCT-K1

MEASURAND : Temperature T_{90} / K

NOMINAL TEMPERATURE : $T_{90} = 1.225$ K

Laboratory individual measurements

Lab: Laboratory

S/N i : serial number of Thermometer i

T_i : temperature value for Thermometer i

u_i : standard uncertainty in the calibration of Thermometer i

u_{comp} : standard uncertainty in the comparison measurements

$$u_{\text{comp}} = 0.149 \text{ mK}$$

Lab, S/N i	T_i / mK	u_i / mK
NIST A128	1.224814	0.065
NIST A129	1.224828	0.065
NMi-VSL 226246	1.225123	0.256
NPL 221481	1.225340	0.151
NPL 221485	1.225201	0.151
PTB 229074	1.224998	0.180
PTB 229075	1.224931	0.180

Key comparison CCT-K1


MEASURAND : Temperature T_{90} / K

KEY COMPARISON REFERENCE VALUE: $T_R = 1.224991$ K

Matrix of equivalence

Lab, S/N i	Lab, S/N j \longrightarrow											
	D_i	U_i	D_{ij}	U_{ij}								
	/ mK	/ mK	/ mK	/ mK	/ mK	/ mK	/ mK	/ mK	/ mK	/ mK	/ mK	/ mK
NIST A128	-0.177	0.325			-0.014	0.459	-0.309	0.675	-0.526	0.534	-0.388	0.534
NIST A129	-0.163	0.325	0.014	0.459			-0.295	0.675	-0.512	0.534	-0.373	0.534
NMi-VSL 226246	0.132	0.592	0.309	0.675	0.295	0.675			-0.217	0.729	-0.078	0.729
NPL 221481	0.349	0.425	0.526	0.534	0.512	0.534	0.217	0.729			0.124	0.754
NPL 221485	0.210	0.425	0.388	0.534	0.373	0.534	0.078	0.729	-0.139	0.600	0.341	0.631
PTB 229074	0.007	0.467	0.185	0.569	0.171	0.569	-0.124	0.754	-0.341	0.631	-0.203	0.631
PTB 229075	-0.060	0.467	0.118	0.569	0.103	0.569	-0.191	0.754	-0.408	0.631	-0.270	0.631

CCT-K1 : Nominal temperature, $T_{90} = 1.225$ K
Degrees of equivalence, D_i , and expanded uncertainties ($k = 2$), U_i , expressed in mK



Key comparison CCT-K1

MEASURAND : Temperature T_{90} / K

NOMINAL TEMPERATURE : $T_{90} = 1.250$ K

Laboratory individual measurements

Lab: Laboratory

S/N i : serial number of Thermometer i

T_i : temperature value for Thermometer i

u_i : standard uncertainty in the calibration of Thermometer i

u_{comp} : standard uncertainty in the comparison measurements

$$u_{\text{comp}} = 0.148 \text{ mK}$$

Lab, S/N i	T_i / mK	u_i / mK
NIST A128	1.249254	0.065
NIST A129	1.249283	0.065
NMi-VSL 226246	1.249645	0.256
NPL 221481	1.250010	0.152
NPL 221483	1.249780	0.152
NPL 221485	1.249814	0.152
PTB 229074	1.249406	0.180
PTB 229075	1.249394	0.180
VNIIFTRI 79	1.249940	0.480
VNIIFTRI 89	1.250150	0.570

Key comparison CCT-K1

MEASURAND : Temperature T_{90} / K

KEY COMPARISON REFERENCE VALUE: $T_R = 1.249542$ K


Matrix of equivalence

Lab, S/N <i>i</i>	Lab, S/N <i>j</i> \longrightarrow													
	D _{<i>i</i>} U _{<i>i</i>} / mK		NIST A128		NIST A129		NMi-VSL 226246		NPL 221481		NPL 221483		NPL 221485	
	D _{<i>ij</i>}	U _{<i>ij</i>}	D _{<i>ij</i>}	U _{<i>ij</i>}	D _{<i>ij</i>}	U _{<i>ij</i>}	D _{<i>ij</i>}	U _{<i>ij</i>}	D _{<i>ij</i>}	U _{<i>ij</i>}	D _{<i>ij</i>}	U _{<i>ij</i>}	D _{<i>ij</i>}	U _{<i>ij</i>}
NIST A128	-0.288	0.323			-0.029	0.456	-0.390	0.673	-0.755	0.533	-0.525	0.533	-0.559	0.533
NIST A129	-0.259	0.323	0.029	0.456			-0.361	0.673	-0.726	0.533	-0.496	0.533	-0.530	0.533
NMi-VSL 226246	0.103	0.591	0.390	0.673	0.361	0.673			-0.365	0.728	-0.135	0.728	-0.169	0.728
NPL 221481	0.468	0.424	0.755	0.533	0.726	0.533	0.365	0.728			0.230	0.600	0.196	0.600
NPL 221483	0.238	0.424	0.525	0.533	0.496	0.533	0.135	0.728	-0.230	0.600			-0.034	0.600
NPL 221485	0.272	0.424	0.559	0.533	0.530	0.533	0.169	0.728	-0.196	0.600	0.034	0.600		
PTB 229074	-0.136	0.466	0.152	0.567	0.123	0.567	-0.239	0.753	-0.604	0.630	-0.374	0.630	-0.408	0.630
PTB 229075	-0.148	0.466	0.140	0.567	0.111	0.567	-0.250	0.753	-0.615	0.630	-0.385	0.630	-0.419	0.630
VNIIFTRI 79	0.398	1.004	0.686	1.055	0.657	1.055	0.296	1.166	-0.069	1.090	0.161	1.090	0.127	1.090
VNIIFTRI 89	0.608	1.178	0.896	1.221	0.867	1.221	0.505	1.318	0.141	1.252	0.371	1.252	0.336	1.252

Matrix of equivalence (Continued)

Lab, S/N <i>i</i>	Lab, S/N <i>j</i> \longrightarrow										
	PTB 229074		PTB 229075		VNIIIFTRI 79		VNIIIFTRI 89				
	D_i	U_i		D_{ij}	U_{ij}		D_{ij}	U_{ij}			
	/ mK	/ mK		/ mK	/ mK		/ mK	/ mK			
NIST A128	-0.288	0.323		-0.152	0.567	-0.140	0.567	-0.686	1.055	-0.896	1.221
NIST A129	-0.259	0.323		-0.123	0.567	-0.111	0.567	-0.657	1.055	-0.867	1.221
NMi-VSL 226246	0.103	0.591		0.239	0.753	0.250	0.753	-0.296	1.166	-0.505	1.318
NPL 221481	0.468	0.424		0.604	0.630	0.615	0.630	0.069	1.090	-0.141	1.252
NPL 221483	0.238	0.424		0.374	0.630	0.385	0.630	-0.161	1.090	-0.371	1.252
NPL 221485	0.272	0.424		0.408	0.630	0.419	0.630	-0.127	1.090	-0.336	1.252
PTB 229074	-0.136	0.466				0.012	0.659	-0.534	1.107	-0.744	1.267
PTB 229075	-0.148	0.466		-0.012	0.659			-0.546	1.107	-0.756	1.267
VNIIIFTRI 79	0.398	1.004		0.534	1.107	0.546	1.107			-0.210	1.548
VNIIIFTRI 89	0.608	1.178		0.744	1.2665	0.756	1.267	0.210	1.548		

CCT-K1 : Nominal temperature, $T_{90} = 1.250$ K
Degrees of equivalence, D_i , and expanded uncertainties ($k = 2$), U_i , expressed in mK



Key comparison CCT-K1

MEASURAND : Temperature T_{90} / K

NOMINAL TEMPERATURE : $T_{90} = 1.503$ K

Laboratory individual measurements

Lab: Laboratory

S/N i : serial number of Thermometer i

T_i : temperature value for Thermometer i

u_i : standard uncertainty in the calibration of Thermometer i

u_{comp} : standard uncertainty in the comparison measurements

$$u_{\text{comp}} = 0.137 \text{ mK}$$

Lab, S/N i	T_i / mK	u_i / mK
NIST A128	1.503103	0.062
NIST A129	1.503008	0.062
NMi-VSL 226246	1.503751	0.256
NPL 221481	1.503610	0.160
NPL 221483	1.503650	0.160
NPL 221485	1.503561	0.160
NRC A138	1.503715	0.225
NRC A140	1.503918	0.225
PTB 229074	1.503155	0.180
PTB 229075	1.503100	0.180
VNIIFTRI 79	1.503876	0.510
VNIIFTRI 89	1.504156	0.620

Key comparison CCT-K1

MEASURAND : Temperature T_{90} / K

KEY COMPARISON REFERENCE VALUE: $T_R = 1.503370$ K


Matrix of equivalence

Lab, S/N i	Lab, S/N j \longrightarrow											
	NIST A128		NIST A129		NMi-VSL 226246		NPL 221481		NPL 221483		NPL 221485	
	D_i	U_i		D_{ij}	U_{ij}		D_{ij}	U_{ij}		D_{ij}	U_{ij}	
	/ mK	/ mK		/ mK	/ mK		/ mK	/ mK		/ mK	/ mK	
NIST A128	-0.267	0.301										
NIST A129	-0.362	0.301		0.095	0.426		-0.648	0.654		-0.507	0.518	
NMi-VSL 226246	0.381	0.581		-0.095	0.426		-0.743	0.654		-0.602	0.518	
NPL 221481	0.240	0.421		0.648	0.654	0.743	0.654		0.141	0.717	0.101	0.717
NPL 221483	0.280	0.421		0.507	0.518	0.602	0.518	-0.141	0.717		-0.040	0.596
NPL 221485	0.191	0.421		0.547	0.518	0.642	0.518	-0.101	0.717	0.040	0.596	0.090
NRC A138	0.345	0.527		0.457	0.518	0.552	0.518	-0.190	0.717	-0.050	0.596	-0.090
NRC A140	0.548	0.527		0.611	0.607	0.706	0.607	-0.036	0.784	0.104	0.675	0.064
PTB 229074	-0.215	0.452		0.815	0.607	0.910	0.607	0.167	0.784	0.308	0.675	0.268
PTB 229075	-0.270	0.452		0.052	0.543	0.147	0.543	-0.596	0.736	-0.455	0.618	-0.495
VNIIFTRI 79	0.506	1.055		-0.004	0.543	0.092	0.543	-0.651	0.736	-0.511	0.618	-0.551
VNIIFTRI 89	0.786	1.269		0.773	1.098	0.868	1.098	0.125	1.205	0.266	1.136	0.226
	1.052	1.304	1.147	1.304	0.405	1.396	0.545	1.337	0.505	1.337	0.595	1.337

Matrix of equivalence (Continued)

Lab, S/N <i>i</i>	Lab, S/N <i>j</i> \longrightarrow															
	NRC A138		NRC A140		PTB 229074		PTB 229075		VNIIIFTRI 79		VNIIIFTRI 89					
	D_i	U_i		D_{ij}	U_{ij}		D_{ij}	U_{ij}		D_{ij}	U_{ij}					
	$/ \text{mK}$			$/ \text{mK}$			$/ \text{mK}$			$/ \text{mK}$						
NIST A128	-0.267	0.301		-0.611	0.607	-0.815	0.607	-0.052	0.543	0.004	0.543	-0.773	1.098	-1.052	1.304	
NIST A129	-0.362	0.301		-0.706	0.607	-0.910	0.607	-0.147	0.543	-0.092	0.543	-0.868	1.098	-1.147	1.304	
NMi-VSL 226246	0.381	0.581		0.036	0.784	-0.167	0.784	0.596	0.736	0.651	0.736	-0.125	1.205	-0.405	1.396	
NPL 221481	0.240	0.421		-0.104	0.675	-0.308	0.675	0.455	0.618	0.511	0.618	-0.266	1.136	-0.545	1.337	
NPL 221483	0.280	0.421		-0.064	0.675	-0.268	0.675	0.495	0.618	0.551	0.618	-0.226	1.136	-0.505	1.337	
NPL 221485	0.191	0.421		-0.154	0.675	-0.357	0.675	0.405	0.618	0.461	0.618	-0.316	1.136	-0.595	1.337	
NRC A138	0.345	0.527				-0.203	0.745	0.559	0.694	0.615	0.694	-0.161	1.180	-0.441	1.374	
NRC A140	0.548	0.527				0.203	0.745		0.763	0.694	0.818	0.694	0.042	1.180	-0.238	1.374
PTB 229074	-0.215	0.452				-0.559	0.694	-0.763	0.694		0.056	0.640	-0.721	1.148	-1.000	1.347
PTB 229075	-0.270	0.452				-0.615	0.694	-0.818	0.694	-0.056	0.640		-0.776	1.148	-1.056	1.347
VNIIIFTRI 79	0.506	1.055				0.161	1.180	-0.042	1.180	0.721	1.148	0.776	1.148		-0.279	1.651
VNIIIFTRI 89	0.786	1.269				0.441	1.374	0.238	1.374	1.000	1.347	1.056	1.347	0.279	1.651	

CCT-K1 : Nominal temperature, $T_{90} = 1.503$ K
Degrees of equivalence, D_i , and expanded uncertainties ($k = 2$), U_i , expressed in mK



Key comparison CCT-K1

MEASURAND : Temperature T_{90} / K

NOMINAL TEMPERATURE : $T_{90} = 1.755$ K

Laboratory individual measurements

Lab: Laboratory

S/N i : serial number of Thermometer i

T_i : temperature value for Thermometer i

u_i : standard uncertainty in the calibration of Thermometer i

u_{comp} : standard uncertainty in the comparison measurements

$$u_{\text{comp}} = 0.132 \text{ mK}$$

Lab, S/N i	T_i / mK	u_i / mK
NIST A128	1.754643	0.038
NIST A129	1.754505	0.038
NMi-VSL 226246	1.755116	0.256
NPL 221481	1.755060	0.157
NPL 221483	1.755070	0.157
NPL 221485	1.754965	0.157
NRC A138	1.755172	0.225
NRC A140	1.755291	0.225
PTB 229074	1.754594	0.180
PTB 229075	1.754562	0.180
VNIIFTRI 79	1.755577	0.484
VNIIFTRI 89	1.755761	0.592

Key comparison CCT-K1

MEASURAND : Temperature T_{90} / K

KEY COMPARISON REFERENCE VALUE: $T_R = 1.754822$ K


Matrix of equivalence

Lab, S/N <i>i</i>	Lab, S/N <i>j</i> \longrightarrow												
	NIST A128		NIST A129		NMi-VSL 226246		NPL 221481		NPL 221483		NPL 221485		
	D_i	U_i		D_{ij}	U_{ij}		D_{ij}	U_{ij}		D_{ij}	U_{ij}		
	/ mK	/ mK		/ mK	/ mK		/ mK	/ mK		/ mK	/ mK		
NIST A128	-0.179	0.275			0.138	0.388	-0.473	0.638	-0.418	0.493	-0.428	0.493	-0.322 0.493
NIST A129	-0.317	0.275		-0.138	0.388		-0.611	0.638	-0.556	0.493	-0.566	0.493	-0.460 0.493
NMi-VSL 226246	0.294	0.576		0.473	0.638	0.611	0.638		0.055	0.707	0.045	0.707	0.151 0.707
NPL 221481	0.238	0.410		0.418	0.493	0.556	0.493	-0.055	0.707		-0.010	0.579	0.095 0.579
NPL 221483	0.248	0.410		0.428	0.493	0.566	0.493	-0.045	0.707	0.010	0.579		0.105 0.579
NPL 221485	0.143	0.410		0.322	0.493	0.460	0.493	-0.151	0.707	-0.095	0.579	-0.105	0.579
NRC A138	0.350	0.522		0.529	0.590	0.668	0.590	0.056	0.777	0.112	0.663	0.102	0.663 0.207 0.663
NRC A140	0.469	0.522		0.649	0.590	0.787	0.590	0.176	0.777	0.231	0.663	0.221	0.663 0.326 0.663
PTB 229074	-0.228	0.446		-0.049	0.524	0.090	0.524	-0.521	0.729	-0.466	0.606	-0.476	0.606 -0.371 0.606
PTB 229075	-0.260	0.446		-0.081	0.524	0.057	0.524	-0.554	0.729	-0.499	0.606	-0.509	0.606 -0.403 0.606
VNIIFTRI 79	0.755	1.004		0.934	1.041	1.073	1.041	0.461	1.158	0.517	1.085	0.507	1.085 0.612 1.085
VNIIFTRI 89	0.939	1.213		1.118	1.243	1.256	1.243	0.645	1.343	0.701	1.280	0.691	1.280 0.796 1.280

Matrix of equivalence (Continued)

Lab, S/N <i>i</i>	Lab, S/N <i>j</i> \longrightarrow															
	NRC A138		NRC A140		PTB 229074		PTB 229075		VNIIIFTRI 79		VNIIIFTRI 89					
	D_i	U_i		D_{ij}	U_{ij}		D_{ij}	U_{ij}		D_{ij}	U_{ij}					
	$/ \text{mK}$			$/ \text{mK}$			$/ \text{mK}$			$/ \text{mK}$						
NIST A128	-0.179	0.275		-0.529	0.590	-0.649	0.590	0.049	0.524	0.081	0.524	-0.934	1.041	-1.118	1.243	
NIST A129	-0.317	0.275		-0.668	0.590	-0.787	0.590	-0.090	0.524	-0.057	0.524	-1.073	1.041	-1.256	1.243	
NMi-VSL 226246	0.294	0.576		-0.056	0.777	-0.176	0.777	0.521	0.729	0.554	0.729	-0.461	1.158	-0.645	1.343	
NPL 221481	0.238	0.410		-0.112	0.663	-0.231	0.663	0.466	0.606	0.499	0.606	-0.517	1.085	-0.701	1.280	
NPL 221483	0.248	0.410		-0.102	0.663	-0.221	0.663	0.476	0.606	0.509	0.606	-0.507	1.085	-0.691	1.280	
NPL 221485	0.143	0.410		-0.207	0.663	-0.326	0.663	0.371	0.606	0.403	0.606	-0.612	1.085	-0.796	1.280	
NRC A138	0.350	0.522				-0.119	0.738	0.578	0.686	0.610	0.686	-0.405	1.132	-0.589	1.320	
NRC A140	0.469	0.522				0.119	0.738	0.697	0.686	0.730	0.686	-0.286	1.132	-0.470	1.320	
PTB 229074	-0.228	0.446				-0.578	0.686	-0.697	0.686		0.033	0.631	-0.983	1.099	-1.167	1.292
PTB 229075	-0.260	0.446				-0.610	0.686	-0.730	0.686	-0.033	0.631		-1.015	1.099	-1.199	1.292
VNIIIFTRI 79	0.755	1.004				0.405	1.132	0.286	1.132	0.983	1.099	1.015	1.099		-0.184	1.575
VNIIIFTRI 89	0.939	1.213				0.589	1.320	0.470	1.320	1.167	1.292	1.199	1.292	0.184	1.575	

CCT-K1 : Nominal temperature, $T_{90} = 1.755$ K
Degrees of equivalence, D_i , and expanded uncertainties ($k = 2$), U_i , expressed in mK



Key comparison CCT-K1

MEASURAND : Temperature T_{90} / K

NOMINAL TEMPERATURE : $T_{90} = 1.997$ K

Laboratory individual measurements

Lab: Laboratory

S/N i : serial number of Thermometer i

T_i : temperature value for Thermometer i

u_i : standard uncertainty in the calibration of Thermometer i

u_{comp} : standard uncertainty in the comparison measurements

$$u_{\text{comp}} = 0.127 \text{ mK}$$

Lab, S/N i	T_i / mK	u_i / mK
NIST A128	1.996327	0.039
NIST A129	1.996288	0.039
NMi-VSL 226246	1.996820	0.256
NPL 221481	1.996780	0.154
NPL 221483	1.996840	0.154
NPL 221485	1.996757	0.154
NRC A138	1.996922	0.225
NRC A140	1.996943	0.225
PTB 229074	1.996278	0.180
PTB 229075	1.996315	0.180
VNIIFTRI 79	1.997300	0.460
VNIIFTRI 89	1.997462	0.565

Key comparison CCT-K1

MEASURAND : Temperature T_{90} / K

KEY COMPARISON REFERENCE VALUE: $T_R = 1.996554$ K


Matrix of equivalence

Lab, S/N <i>i</i>	Lab, S/N <i>j</i> \longrightarrow												
	NIST A128		NIST A129		NMi-VSL 226246		NPL 221481		NPL 221483		NPL 221485		
	D_i	U_i		D_{ij}	U_{ij}		D_{ij}	U_{ij}		D_{ij}	U_{ij}		
	/ mK	/ mK		/ mK	/ mK		/ mK	/ mK		/ mK	/ mK		
NIST A128	-0.227	0.266			0.040	0.376	-0.493	0.630	-0.453	0.479	-0.513	0.479	-0.430 0.479
NIST A129	-0.266	0.266		-0.040	0.376		-0.532	0.630	-0.492	0.479	-0.552	0.479	-0.470 0.479
NMi-VSL 226246	0.266	0.572		0.493	0.630	0.532 0.630			0.040	0.697	-0.020	0.697	0.063 0.697
NPL 221481	0.226	0.399		0.453	0.479	0.492 0.479	-0.040	0.697			-0.060	0.564	0.023 0.564
NPL 221483	0.286	0.399		0.513	0.479	0.552 0.479	0.020	0.697	0.060	0.564		0.083	0.564
NPL 221485	0.203	0.399		0.430	0.479	0.470 0.479	-0.063	0.697	-0.023	0.564	-0.083	0.564	
NRC A138	0.368	0.517		0.594	0.581	0.634 0.581	0.102	0.771	0.142	0.653	0.082	0.653	0.165 0.653
NRC A140	0.389	0.517		0.615	0.581	0.655 0.581	0.123	0.771	0.163	0.653	0.103	0.653	0.185 0.653
PTB 229074	-0.276	0.441		-0.049	0.515	-0.010 0.515	-0.542	0.722	-0.502	0.594	-0.562	0.594	-0.479 0.594
PTB 229075	-0.239	0.441		-0.012	0.515	0.027 0.515	-0.505	0.722	-0.465	0.594	-0.525	0.594	-0.442 0.594
VNIIFTRI 79	0.746	0.955		0.973	0.991	1.012 0.991	0.480	1.113	0.520	1.035	0.460	1.035	0.543 1.035
VNIIFTRI 89	0.908	1.159		1.135	1.189	1.174 1.189	0.642	1.292	0.682	1.225	0.622	1.225	0.705 1.225

Matrix of equivalence (Continued)

Lab, S/N <i>i</i>	Lab, S/N <i>j</i> \longrightarrow														
	NRC A138		NRC A140		PTB 229074		PTB 229075		VNIIIFTRI 79		VNIIIFTRI 89				
	D_i	U_i		D_{ij}	U_{ij}		D_{ij}	U_{ij}		D_{ij}	U_{ij}				
	$/ \text{mK}$			$/ \text{mK}$			$/ \text{mK}$			$/ \text{mK}$					
NIST A128	-0.227	0.266		-0.594	0.581	-0.615	0.581	0.049	0.515	0.012	0.515	-0.973	0.991	-1.135	1.189
NIST A129	-0.266	0.266		-0.634	0.581	-0.655	0.581	0.010	0.515	-0.027	0.515	-1.012	0.991	-1.174	1.189
NMi-VSL 226246	0.266	0.572		-0.102	0.771	-0.123	0.771	0.542	0.722	0.505	0.722	-0.480	1.113	-0.642	1.292
NPL 221481	0.226	0.399		-0.142	0.653	-0.163	0.653	0.502	0.594	0.465	0.594	-0.520	1.035	-0.682	1.225
NPL 221483	0.286	0.399		-0.082	0.653	-0.103	0.653	0.562	0.594	0.525	0.594	-0.460	1.035	-0.622	1.225
NPL 221485	0.203	0.399		-0.165	0.653	-0.185	0.653	0.479	0.594	0.442	0.594	-0.543	1.035	-0.705	1.225
NRC A138	0.368	0.517				-0.021	0.731	0.644	0.679	0.607	0.679	-0.378	1.086	-0.540	1.269
NRC A140	0.389	0.517		0.021	0.731		0.665	0.679	0.628	0.679	-0.357	1.086	-0.519	1.269	
PTB 229074	-0.276	0.441		-0.644	0.679	-0.665	0.679		-0.037	0.623	-1.022	1.052	-1.184	1.240	
PTB 229075	-0.239	0.441		-0.607	0.679	-0.628	0.679	0.037	0.623		-0.985	1.052	-1.147	1.240	
VNIIIFTRI 79	0.746	0.955		0.378	1.086	0.357	1.086	1.022	1.052	0.985	1.052		-0.162	1.502	
VNIIIFTRI 89	0.908	1.159		0.540	1.269	0.519	1.269	1.184	1.240	1.147	1.240	0.162	1.502		

CCT-K1 : Nominal temperature, $T_{90} = 1.997$ K
Degrees of equivalence, D_i , and expanded uncertainties ($k = 2$), U_i , expressed in mK



Key comparison CCT-K1

MEASURAND : Temperature T_{90} / K

NOMINAL TEMPERATURE : $T_{90} = 2.248$ K

Laboratory individual measurements

Lab: Laboratory

S/N i : serial number of Thermometer i

T_i : temperature value for Thermometer i

u_i : standard uncertainty in the calibration of Thermometer i

u_{comp} : standard uncertainty in the comparison measurements

$$u_{\text{comp}} = 0.122 \text{ mK}$$

Lab, S/N i	T_i / mK	u_i / mK
NIST A128	2.248374	0.039
NIST A129	2.248385	0.039
NMi-VSL 226246	2.248743	0.225
NPL 221481	2.248710	0.150
NPL 221483	2.248750	0.150
NPL 221485	2.248685	0.150
NRC A138	2.248713	0.225
NRC A140	2.248734	0.225
PTB 229074	2.248321	0.180
PTB 229075	2.248306	0.180
VNIIFTRI 79	2.249248	0.435
VNIIFTRI 89	2.249363	0.538
NMIJ B271	2.248277	0.095
NMIJ B310	2.248308	0.094

Key comparison CCT-K1

MEASURAND : Temperature T_{90} / K

KEY COMPARISON REFERENCE VALUE: $T_R = 2.248485$ K

Matrix of equivalence

Lab, S/N i	Lab, S/N j \longrightarrow													
	NIST A128		NIST A129		NMi-VSL 226246		NPL 221481		NPL 221483		NPL 221485		NRC A138	
	D_i	U_i	D_{ij}	U_{ij}	D_{ij}	U_{ij}	D_{ij}	U_{ij}	D_{ij}	U_{ij}	D_{ij}	U_{ij}	D_{ij}	U_{ij}
NIST A128	-0.111	0.256			-0.011	0.363	-0.368	0.573	-0.336	0.465	-0.376	0.465	-0.311	0.465
NIST A129	-0.100	0.256	0.011	0.363			-0.358	0.573	-0.325	0.465	-0.365	0.465	-0.300	0.465
NMi-VSL 226246	0.258	0.512	0.368	0.573	0.358	0.573			0.033	0.642	-0.007	0.642	0.057	0.642
NPL 221481	0.225	0.387	0.336	0.465	0.325	0.465	-0.033	0.642			-0.040	0.548	0.024	0.548
NPL 221483	0.265	0.387	0.376	0.465	0.365	0.465	0.007	0.642	0.040	0.548			0.064	0.548
NPL 221485	0.200	0.387	0.311	0.465	0.300	0.465	-0.057	0.642	-0.024	0.548	-0.064	0.548		
NRC A138	0.228	0.512	0.339	0.573	0.328	0.573	-0.029	0.724	0.003	0.642	-0.037	0.642	0.028	0.642
NRC A140	0.249	0.512	0.360	0.573	0.349	0.573	-0.009	0.724	0.024	0.642	-0.016	0.642	0.049	0.642
PTB 229074	-0.164	0.435	-0.053	0.505	-0.064	0.505	-0.421	0.672	-0.388	0.582	-0.428	0.582	-0.364	0.582
PTB 229075	-0.179	0.435	-0.068	0.505	-0.079	0.505	-0.437	0.672	-0.404	0.582	-0.444	0.582	-0.380	0.582
VNIIFTRI 79	0.763	0.904	0.873	0.939	0.863	0.939	0.505	1.039	0.538	0.983	0.498	0.983	0.562	0.983
VNIIFTRI 89	0.878	1.102	0.989	1.132	0.978	1.132	0.621	1.216	0.654	1.169	0.614	1.169	0.678	1.169
NMIJ B271	-0.208	0.309	-0.097	0.402	-0.108	0.402	-0.465	0.598	-0.432	0.496	-0.472	0.496	-0.408	0.496
NMIJ B310	-0.177	0.308	-0.066	0.401	-0.077	0.401	-0.435	0.597	-0.402	0.495	-0.442	0.495	-0.378	0.495


MEASURAND : Temperature T_{90} / K

KEY COMPARISON REFERENCE VALUE: $T_R = 2.248485$ K

Matrix of equivalence (Continued)

Lab, S/N <i>i</i>	Lab, S/N <i>j</i> \longrightarrow																
	D _{<i>i</i>}		U _{<i>i</i>}		NRC A140		PTB 229074		PTB 229075		VNIIIFTRI 79		VNIIIFTRI 89		NMIJ B271		NMIJ B310
	D _{<i>ij</i>}	U _{<i>ij</i>}	D _{<i>ij</i>}	U _{<i>ij</i>}	D _{<i>ij</i>}	U _{<i>ij</i>}	D _{<i>ij</i>}	U _{<i>ij</i>}	D _{<i>ij</i>}	U _{<i>ij</i>}	D _{<i>ij</i>}	U _{<i>ij</i>}	D _{<i>ij</i>}	U _{<i>ij</i>}	D _{<i>ij</i>}	U _{<i>ij</i>}	
NIST A128	-0.111	0.256	-0.360	0.573	0.053	0.505	0.068	0.505	-0.873	0.939	-0.989	1.132	0.097	0.402	0.066	0.401	
NIST A129	-0.100	0.256	-0.349	0.573	0.064	0.505	0.079	0.505	-0.863	0.939	-0.978	1.132	0.108	0.402	0.077	0.401	
NMi-VSL 226246	0.258	0.512	0.009	0.724	0.421	0.672	0.437	0.672	-0.505	1.039	-0.621	1.216	0.465	0.598	0.435	0.597	
NPL 221481	0.225	0.387	-0.024	0.642	0.388	0.582	0.404	0.582	-0.538	0.983	-0.654	1.169	0.432	0.496	0.402	0.495	
NPL 221483	0.265	0.387	0.016	0.642	0.428	0.582	0.444	0.582	-0.498	0.983	-0.614	1.169	0.472	0.496	0.442	0.495	
NPL 221485	0.200	0.387	-0.049	0.642	0.364	0.582	0.380	0.582	-0.562	0.983	-0.678	1.169	0.408	0.496	0.378	0.495	
NRC A138	0.228	0.512	-0.021	0.724	0.392	0.672	0.407	0.672	-0.534	1.039	-0.650	1.216	0.436	0.598	0.405	0.597	
NRC A140	0.249	0.512			0.413	0.672	0.428	0.672	-0.514	1.039	-0.629	1.216	0.457	0.598	0.426	0.597	
PTB 229074	-0.164	0.435	-0.413	0.672			0.016	0.615	-0.926	1.003	-1.042	1.185	0.044	0.534	0.014	0.533	
PTB 229075	-0.179	0.435	-0.428	0.672	-0.016	0.615			-0.942	1.003	-1.058	1.185	0.028	0.534	-0.002	0.533	
VNIIIFTRI 79	0.763	0.904	0.514	1.039	0.926	1.003	0.942	1.003			-0.116	1.426	0.970	0.955	0.940	0.955	
VNIIIFTRI 89	0.878	1.102	0.629	1.216	1.042	1.185	1.058	1.185	0.116	1.426			1.086	1.145	1.056	1.145	
NMIJ B271	-0.208	0.309	-0.457	0.598	-0.044	0.534	-0.028	0.534	-0.970	0.955	-1.086	1.145			-0.030	0.436	
NMIJ B310	-0.177	0.308	-0.426	0.597	-0.014	0.533	0.002	0.533	-0.940	0.955	-1.056	1.145	0.030	0.436			

CCT-K1 : Nominal temperature, $T_{90} = 2.248$ K
Degrees of equivalence, D_i , and expanded uncertainties ($k = 2$), U_i , expressed in mK



Key comparison CCT-K1

MEASURAND : Temperature T_{90} / K

NOMINAL TEMPERATURE : $T_{90} = 2.601$ K

Laboratory individual measurements

Lab: Laboratory

S/N i : serial number of Thermometer i

T_i : temperature value for Thermometer i

u_i : standard uncertainty in the calibration of Thermometer i

u_{comp} : standard uncertainty in the comparison measurements

$$u_{\text{comp}} = 0.115 \text{ mK}$$

Lab, S/N i	T_i / mK	u_i / mK
NIST A128	2.600642	0.040
NIST A129	2.600648	0.040
NMi-VSL 226246	2.600862	0.181
NPL 221481	2.600827	0.146
NPL 221483	2.600842	0.146
NPL 221485	2.600863	0.146
NRC A138	2.600854	0.225
NRC A140	2.600896	0.225
PTB 229074	2.600646	0.180
PTB 229075	2.600798	0.180
VNIIFTRI 79	2.601411	0.406
VNIIFTRI 89	2.601495	0.498
NMIJ B271	2.600793	0.084
NMIJ B310	2.600816	0.082

Key comparison CCT-K1

MEASURAND : Temperature T_{90} / K

KEY COMPARISON REFERENCE VALUE: $T_R = 2.600776$ K


Matrix of equivalence

Lab, S/N i	Lab, S/N j \longrightarrow													
	NIST A128		NIST A129		NMi-VSL 226246		NPL 221481		NPL 221483		NPL 221485		NRC A138	
	D_i	U_i	D_{ij}	U_{ij}	D_{ij}	U_{ij}	D_{ij}	U_{ij}	D_{ij}	U_{ij}	D_{ij}	U_{ij}	D_{ij}	U_{ij}
NIST A128	-0.134	0.243			-0.006	0.344	-0.219	0.493	-0.184	0.444	-0.199	0.444	-0.220	0.444
NIST A129	-0.128	0.243	0.006	0.344			-0.214	0.493	-0.179	0.444	-0.194	0.444	-0.215	0.444
NMi-VSL 226246	0.086	0.429	0.219	0.493	0.214	0.493			0.035	0.567	0.020	0.567	-0.001	0.567
NPL 221481	0.051	0.372	0.184	0.444	0.179	0.444	-0.035	0.567			-0.015	0.526	-0.036	0.526
NPL 221483	0.066	0.372	0.199	0.444	0.194	0.444	-0.020	0.567	0.015	0.526			-0.021	0.526
NPL 221485	0.087	0.372	0.220	0.444	0.215	0.444	0.001	0.567	0.036	0.526	0.021	0.526		0.009
NRC A138	0.078	0.505	0.212	0.561	0.206	0.561	-0.008	0.663	0.028	0.627	0.013	0.627	-0.009	0.627
NRC A140	0.120	0.505	0.253	0.561	0.248	0.561	0.034	0.663	0.069	0.627	0.054	0.627	0.033	0.627
PTB 229074	-0.130	0.427	0.004	0.492	-0.002	0.492	-0.215	0.605	-0.180	0.566	-0.195	0.566	-0.216	0.566
PTB 229075	0.022	0.427	0.155	0.492	0.150	0.492	-0.064	0.605	-0.029	0.566	-0.044	0.566	-0.065	0.566
VNIIFTRI 79	0.635	0.844	0.769	0.878	0.763	0.878	0.549	0.947	0.584	0.922	0.569	0.922	0.548	0.922
VNIIFTRI 89	0.719	1.022	0.852	1.050	0.846	1.050	0.633	1.108	0.668	1.087	0.653	1.087	0.632	1.087
NMIJ B271	0.017	0.285	0.151	0.375	0.145	0.375	-0.068	0.515	-0.033	0.468	-0.048	0.468	-0.069	0.468
NMIJ B310	0.040	0.282	0.174	0.373	0.168	0.373	-0.046	0.513	-0.011	0.467	-0.026	0.467	-0.047	0.467

Matrix of equivalence (Continued)

Lab, S/N <i>i</i>	Lab, S/N <i>j</i> →																
	D _{<i>i</i>}		U _{<i>i</i>}		NRC A140		PTB 229074		PTB 229075		VNIIIFTRI 79		VNIIIFTRI 89		NMJJ B271		NMJJ B310
	D _{<i>ij</i>}	U _{<i>ij</i>}															
NIST A128	-0.134	0.243	-0.253	0.561	-0.004	0.492	-0.155	0.492	-0.769	0.878	-0.852	1.050	-0.151	0.375	-0.174	0.373	
NIST A129	-0.128	0.243	-0.248	0.561	0.002	0.492	-0.150	0.492	-0.763	0.878	-0.846	1.050	-0.145	0.375	-0.168	0.373	
NMi-VSL 226246	0.086	0.429	-0.034	0.663	0.215	0.605	0.064	0.605	-0.549	0.947	-0.633	1.108	0.068	0.515	0.046	0.513	
NPL 221481	0.051	0.372	-0.069	0.627	0.180	0.566	0.029	0.566	-0.584	0.922	-0.668	1.087	0.033	0.468	0.011	0.467	
NPL 221483	0.066	0.372	-0.054	0.627	0.195	0.566	0.044	0.566	-0.569	0.922	-0.653	1.087	0.048	0.468	0.026	0.467	
NPL 221485	0.087	0.372	-0.033	0.627	0.216	0.566	0.065	0.566	-0.548	0.922	-0.632	1.087	0.069	0.468	0.047	0.467	
NRC A138	0.078	0.505	-0.042	0.715	0.208	0.662	0.056	0.662	-0.557	0.984	-0.640	1.140	0.061	0.580	0.038	0.579	
NRC A140	0.120	0.505			0.250	0.662	0.098	0.662	-0.515	0.984	-0.599	1.140	0.103	0.580	0.080	0.579	
PTB 229074	-0.130	0.427	-0.250	0.662			-0.152	0.604	-0.765	0.946	-0.848	1.108	-0.147	0.513	-0.170	0.512	
PTB 229075	0.022	0.427	-0.098	0.662	0.152	0.604			-0.613	0.946	-0.697	1.108	0.005	0.513	-0.018	0.512	
VNIIIFTRI 79	0.635	0.844	0.515	0.984	0.765	0.946	0.613	0.946			-0.084	1.325	0.618	0.891	0.595	0.890	
VNIIIFTRI 89	0.719	1.022	0.599	1.140	0.848	1.108	0.697	1.108	0.084	1.325			0.701	1.061	0.679	1.060	
NMJJ B271	0.017	0.285	-0.103	0.580	0.147	0.513	-0.005	0.513	-0.618	0.891	-0.701	1.061			-0.023	0.401	
NMJJ B310	0.040	0.282	-0.080	0.579	0.170	0.512	0.018	0.512	-0.595	0.890	-0.679	1.060	0.023	0.401			

CCT-K1 : Nominal temperature, $T_{90} = 2.601$ K
Degrees of equivalence, D_i , and expanded uncertainties ($k = 2$), U_i , expressed in mK



Key comparison CCT-K1

MEASURAND : Temperature T_{90} / K

NOMINAL TEMPERATURE : $T_{90} = 2.700$ K

Laboratory individual measurements

Lab: Laboratory

S/N i : serial number of Thermometer i

T_i : temperature value for Thermometer i

u_i : standard uncertainty in the calibration of Thermometer i

u_{comp} : standard uncertainty in the comparison measurements

$$u_{\text{comp}} = 0.115 \text{ mK}$$

Lab, S/N i	T_i / mK	u_i / mK
NIST A128	2.699899	0.040
NIST A129	2.699752	0.040
NMi-VSL 226246	2.700027	0.169
NPL 221481	2.699890	0.147
NPL 221483	2.699940	0.147
NPL 221485	2.699849	0.147
NRC A138	2.699987	0.225
NRC A140	2.700099	0.225
PTB 229074	2.699794	0.180
PTB 229075	2.699936	0.180
VNIIFTRI 79	2.700463	0.402
VNIIFTRI 89	2.700602	0.486
NMIJ B271	2.699928	0.085
NMIJ B310	2.699939	0.082

Key comparison CCT-K1

MEASURAND : Temperature T_{90} / K

KEY COMPARISON REFERENCE VALUE: $T_R = 2.699911$ K

Matrix of equivalence


Lab, S/N j \longrightarrow

Lab, S/N i	D_i U_i / mK		NIST A128		NIST A129		NMi-VSL 226246		NPL 221481		NPL 221483		NPL 221485		NRC A138	
	D_{ij} / mK	U_{ij} / mK	D_{ij} / mK	U_{ij} / mK	D_{ij} / mK	U_{ij} / mK	D_{ij} / mK	U_{ij} / mK	D_{ij} / mK	U_{ij} / mK	D_{ij} / mK	U_{ij} / mK	D_{ij} / mK	U_{ij} / mK	D_{ij} / mK	U_{ij} / mK
NIST A128	-0.012	0.243			0.147	0.344	-0.129	0.475	0.008	0.446	-0.042	0.446	0.049	0.446	-0.089	0.561
NIST A129	-0.159	0.243	-0.147	0.344			-0.275	0.475	-0.138	0.446	-0.188	0.446	-0.097	0.446	-0.236	0.561
NMi-VSL 226246	0.116	0.408	0.129	0.475	0.275	0.475			0.137	0.553	0.087	0.553	0.178	0.553	0.040	0.649
NPL 221481	-0.021	0.373	-0.008	0.446	0.138	0.446	-0.137	0.553			-0.050	0.528	0.041	0.528	-0.097	0.628
NPL 221483	0.029	0.373	0.042	0.446	0.188	0.446	-0.087	0.553	0.050	0.528			0.091	0.528	-0.047	0.628
NPL 221485	-0.062	0.373	-0.049	0.446	0.097	0.446	-0.178	0.553	-0.041	0.528	-0.091	0.528			-0.138	0.628
NRC A138	0.076	0.505	0.089	0.561	0.236	0.561	-0.040	0.649	0.097	0.628	0.047	0.628	0.138	0.628		
NRC A140	0.188	0.505	0.200	0.561	0.347	0.561	0.072	0.649	0.209	0.628	0.159	0.628	0.250	0.628	0.111	0.714
PTB 229074	-0.117	0.427	-0.104	0.491	0.042	0.491	-0.233	0.590	-0.096	0.567	-0.146	0.567	-0.055	0.567	-0.193	0.661
PTB 229075	0.025	0.427	0.038	0.491	0.184	0.491	-0.091	0.590	0.046	0.567	-0.004	0.567	0.087	0.567	-0.051	0.661
VNIIFTRI 79	0.552	0.836	0.564	0.871	0.711	0.871	0.436	0.930	0.573	0.916	0.523	0.916	0.614	0.916	0.476	0.977
VNIIFTRI 89	0.691	0.999	0.704	1.028	0.851	1.028	0.575	1.079	0.712	1.066	0.662	1.066	0.753	1.066	0.615	1.119
NMIJ B271	0.017	0.286	0.030	0.375	0.176	0.375	-0.099	0.498	0.038	0.470	-0.012	0.470	0.079	0.470	-0.059	0.580
NMIJ B310	0.028	0.282	0.041	0.372	0.188	0.372	-0.088	0.496	0.049	0.468	-0.001	0.468	0.090	0.468	-0.048	0.579

Matrix of equivalence (Continued)

Lab, S/N <i>i</i>	Lab, S/N <i>j</i> →																
	NRC A140		PTB 229074		PTB 229075		VNIIIFTRI 79		VNIIIFTRI 89		NMJJ B271		NMJJ B310				
	D_i	U_i		D_{ij}	U_{ij}		D_{ij}	U_{ij}		D_{ij}	U_{ij}		D_{ij}	U_{ij}			
NIST A128	-0.012	0.243		-0.200	0.561	0.104	0.491	-0.038	0.491	-0.564	0.871	-0.704	1.028	-0.030	0.375	-0.041	0.372
NIST A129	-0.159	0.243		-0.347	0.561	-0.042	0.491	-0.184	0.491	-0.711	0.871	-0.851	1.028	-0.176	0.375	-0.188	0.372
NMi-VSL 226246	0.116	0.408		-0.072	0.649	0.233	0.590	0.091	0.590	-0.436	0.930	-0.575	1.079	0.099	0.498	0.088	0.496
NPL 221481	-0.021	0.373		-0.209	0.628	0.096	0.567	-0.046	0.567	-0.573	0.916	-0.712	1.066	-0.038	0.470	-0.049	0.468
NPL 221483	0.029	0.373		-0.159	0.628	0.146	0.567	0.004	0.567	-0.523	0.916	-0.662	1.066	0.012	0.470	0.001	0.468
NPL 221485	-0.062	0.373		-0.250	0.628	0.055	0.567	-0.087	0.567	-0.614	0.916	-0.753	1.066	-0.079	0.470	-0.090	0.468
NRC A138	0.076	0.505		-0.111	0.714	0.193	0.661	0.051	0.661	-0.476	0.977	-0.615	1.119	0.059	0.580	0.048	0.579
NRC A140	0.188	0.505				0.305	0.661	0.162	0.661	-0.364	0.977	-0.504	1.119	0.171	0.580	0.159	0.579
PTB 229074	-0.117	0.427		-0.305	0.661			-0.142	0.604	-0.669	0.939	-0.808	1.086	-0.134	0.514	-0.145	0.512
PTB 229075	0.025	0.427		-0.162	0.661	0.142	0.604			-0.527	0.939	-0.666	1.086	0.008	0.514	-0.003	0.512
VNIIIFTRI 79	0.552	0.836		0.364	0.977	0.669	0.939	0.527	0.939			-0.139	1.302	0.535	0.884	0.524	0.882
VNIIIFTRI 89	0.691	0.999		0.504	1.119	0.808	1.086	0.666	1.086	0.139	1.302			0.674	1.039	0.663	1.038
NMJJ B271	0.017	0.286		-0.171	0.580	0.134	0.514	-0.008	0.514	-0.535	0.884	-0.674	1.039			-0.011	0.401
NMJJ B310	0.028	0.282		-0.159	0.579	0.145	0.512	0.003	0.512	-0.524	0.882	-0.663	1.038	0.011	0.401		

CCT-K1 : Nominal temperature, $T_{90} = 2.700$ K
Degrees of equivalence, D_i , and expanded uncertainties ($k = 2$), U_i , expressed in mK



Key comparison CCT-K1

MEASURAND : Temperature T_{90} / K

NOMINAL TEMPERATURE : $T_{90} = 2.897$ K

Laboratory individual measurements

Lab: Laboratory

S/N i : serial number of Thermometer i

T_i : temperature value for Thermometer i

u_i : standard uncertainty in the calibration of Thermometer i

u_{comp} : standard uncertainty in the comparison measurements

$$u_{\text{comp}} = 0.114 \text{ mK}$$

Lab, S/N i	T_i / mK	u_i / mK
NIST A128	2.896679	0.040
NIST A129	2.896544	0.040
NMi-VSL 226246	2.896816	0.144
NPL 221481	2.896790	0.150
NPL 221483	2.896800	0.150
NPL 221485	2.896663	0.150
NRC A138	2.896787	0.225
NRC A140	2.896897	0.225
PTB 229074	2.896512	0.180
PTB 229075	2.896793	0.180
VNIIFTRI 79	2.897272	0.394
VNIIFTRI 89	2.897367	0.462
NMIJ B271	2.896750	0.081
NMIJ B310	2.896850	0.084

Key comparison CCT-K1

MEASURAND : Temperature T_{90} / K

KEY COMPARISON REFERENCE VALUE: $T_R = 2.896733$ K


Matrix of equivalence

Lab, S/N i	Lab, S/N j \longrightarrow															
	NIST A128		NIST A129		NMi-VSL 226246		NPL 221481		NPL 221483		NPL 221485		NRC A138			
	D_i	U_i	D_{ij}	U_{ij}	D_{ij}	U_{ij}	D_{ij}	U_{ij}	D_{ij}	U_{ij}	D_{ij}	U_{ij}	D_{ij}	U_{ij}		
NIST A128	-0.054	0.242			0.135	0.343	-0.136	0.440	-0.111	0.448	-0.121	0.448	0.017	0.448	-0.108	0.560
NIST A129	-0.189	0.242	-0.135	0.343			-0.272	0.440	-0.246	0.448	-0.256	0.448	-0.118	0.448	-0.243	0.560
NMi-VSL 226246	0.083	0.368	0.136	0.440	0.272	0.440			0.026	0.526	0.016	0.526	0.153	0.526	0.029	0.624
NPL 221481	0.057	0.377	0.111	0.448	0.246	0.448	-0.026	0.526			-0.010	0.533	0.128	0.533	0.003	0.630
NPL 221483	0.067	0.377	0.121	0.448	0.256	0.448	-0.016	0.526	0.010	0.533			0.138	0.533	0.013	0.630
NPL 221485	-0.070	0.377	-0.017	0.448	0.118	0.448	-0.153	0.526	-0.128	0.533	-0.138	0.533			-0.125	0.630
NRC A138	0.054	0.505	0.108	0.560	0.243	0.560	-0.029	0.624	-0.003	0.630	-0.013	0.630	0.125	0.630		
NRC A140	0.164	0.505	0.218	0.560	0.353	0.560	0.081	0.624	0.107	0.630	0.097	0.630	0.235	0.630	0.110	0.714
PTB 229074	-0.221	0.426	-0.168	0.490	-0.033	0.490	-0.304	0.563	-0.279	0.569	-0.289	0.569	-0.151	0.569	-0.275	0.661
PTB 229075	0.060	0.426	0.114	0.490	0.249	0.490	-0.023	0.563	0.003	0.569	-0.007	0.569	0.131	0.569	0.006	0.661
VNIIFTRI 79	0.539	0.821	0.593	0.856	0.728	0.856	0.457	0.899	0.482	0.903	0.472	0.903	0.610	0.903	0.485	0.963
VNIIFTRI 89	0.634	0.952	0.687	0.983	0.823	0.983	0.551	1.021	0.577	1.024	0.567	1.024	0.704	1.024	0.580	1.078
NMIJ B271	0.017	0.280	0.070	0.370	0.205	0.370	-0.066	0.462	-0.041	0.470	-0.051	0.470	0.087	0.470	-0.038	0.577
NMIJ B310	0.117	0.284	0.170	0.373	0.305	0.373	0.034	0.464	0.059	0.472	0.049	0.472	0.187	0.472	0.062	0.579

Matrix of equivalence (Continued)

Lab, S/N <i>i</i>	Lab, S/N <i>j</i> →																
	NRC A140		PTB 229074		PTB 229075		VNIIIFTRI 79		VNIIIFTRI 89		NMJJ B271		NMJJ B310				
	D_i	U_i		D_{ij}	U_{ij}		D_{ij}	U_{ij}		D_{ij}	U_{ij}		D_{ij}	U_{ij}			
NIST A128	-0.054	0.242		-0.218	0.560	0.168	0.490	-0.114	0.490	-0.593	0.856	-0.687	0.983	-0.070	0.370	-0.170	0.373
NIST A129	-0.189	0.242		-0.353	0.560	0.033	0.490	-0.249	0.490	-0.728	0.856	-0.823	0.983	-0.205	0.370	-0.305	0.373
NMi-VSL 226246	0.083	0.368		-0.081	0.624	0.304	0.563	0.023	0.563	-0.457	0.899	-0.551	1.021	0.066	0.462	-0.034	0.464
NPL 221481	0.057	0.377		-0.107	0.630	0.279	0.569	-0.003	0.569	-0.482	0.903	-0.577	1.024	0.041	0.470	-0.059	0.472
NPL 221483	0.067	0.377		-0.097	0.630	0.289	0.569	0.007	0.569	-0.472	0.903	-0.567	1.024	0.051	0.470	-0.049	0.472
NPL 221485	-0.070	0.377		-0.235	0.630	0.151	0.569	-0.131	0.569	-0.610	0.903	-0.704	1.024	-0.087	0.470	-0.187	0.472
NRC A138	0.054	0.505		-0.110	0.714	0.275	0.661	-0.006	0.661	-0.485	0.963	-0.580	1.078	0.038	0.577	-0.062	0.579
NRC A140	0.164	0.505				0.385	0.661	0.104	0.661	-0.375	0.963	-0.470	1.078	0.148	0.577	0.047	0.579
PTB 229074	-0.221	0.426		-0.385	0.661			-0.281	0.603	-0.761	0.925	-0.855	1.044	-0.238	0.510	-0.338	0.512
PTB 229075	0.060	0.426		-0.104	0.661	0.281	0.603			-0.479	0.925	-0.574	1.044	0.043	0.510	-0.057	0.512
VNIIIFTRI 79	0.539	0.821		0.375	0.963	0.761	0.925	0.479	0.925			-0.094	1.257	0.523	0.867	0.423	0.868
VNIIIFTRI 89	0.634	0.952		0.470	1.078	0.855	1.044	0.574	1.044	0.094	1.257			0.617	0.993	0.517	0.994
NMJJ B271	0.017	0.280		-0.148	0.577	0.238	0.510	-0.043	0.510	-0.523	0.867	-0.617	0.993			-0.100	0.399
NMJJ B310	0.117	0.284		-0.047	0.579	0.338	0.512	0.057	0.512	-0.423	0.868	-0.517	0.994	0.100	0.399		

CCT-K1 : Nominal temperature, $T_{90} = 2.897$ K
Degrees of equivalence, D_i , and expanded uncertainties ($k = 2$), U_i , expressed in mK



Key comparison CCT-K1

MEASURAND : Temperature T_{90} / K

NOMINAL TEMPERATURE : $T_{90} = 2.997$ K

Laboratory individual measurements

Lab: Laboratory

S/N i : serial number of Thermometer i

T_i : temperature value for Thermometer i

u_i : standard uncertainty in the calibration of Thermometer i

u_{comp} : standard uncertainty in the comparison measurements

$$u_{\text{comp}} = 0.114 \text{ mK}$$

Lab, S/N i	T_i / mK	u_i / mK
NIST A128	2.996572	0.041
NIST A129	2.996576	0.041
NMi-VSL 226246	2.996750	0.131
NPL 221481	2.996710	0.151
NPL 221485	2.996632	0.151
NRC A138	2.996624	0.225
NRC A140	2.996781	0.225
PTB 229074	2.996582	0.180
PTB 229075	2.996720	0.180
NMIJ B271	2.996679	0.082
NMIJ B310	2.996688	0.080

Key comparison CCT-K1

MEASURAND : Temperature T_{90} / K

KEY COMPARISON REFERENCE VALUE: $T_R = 2.996648$ K


Matrix of equivalence

Lab, S/N <i>i</i>	Lab, S/N <i>j</i> \longrightarrow											
	NIST A128		NIST A129		NMi-VSL 226246		NPL 221481		NPL 221485		NRC A138	
	D_i	U_i		D_{ij}	U_{ij}		D_{ij}	U_{ij}		D_{ij}	U_{ij}	
NIST A128	-0.076	0.242			-0.004	0.343	-0.178	0.423	-0.138	0.450	-0.061	0.450
NIST A129	-0.072	0.242		0.004	0.343		-0.174	0.423	-0.134	0.450	-0.057	0.450
NMi-VSL 226246	0.102	0.347		0.178	0.423	0.174	0.423		0.040	0.514	0.118	0.514
NPL 221481	0.062	0.379		0.138	0.450	0.134	0.450	-0.040	0.514		0.126	0.612
NPL 221485	-0.016	0.379		0.061	0.450	0.057	0.450	-0.118	0.514	-0.077	0.536	0.086
NRC A138	-0.024	0.504		0.052	0.560	0.048	0.560	-0.126	0.612	-0.086	0.631	-0.009
NRC A140	0.133	0.504		0.209	0.560	0.205	0.560	0.031	0.612	0.071	0.631	0.148
PTB 229074	-0.066	0.426		0.010	0.490	0.007	0.490	-0.168	0.550	-0.127	0.570	-0.050
PTB 229075	0.072	0.426		0.149	0.490	0.145	0.490	-0.030	0.550	0.011	0.570	0.088
NMIJ B271	0.031	0.281		0.107	0.371	0.103	0.371	-0.071	0.447	-0.031	0.472	0.046
NMIJ B310	0.040	0.279		0.116	0.369	0.112	0.369	-0.062	0.445	-0.022	0.470	0.055

Matrix of equivalence (Continued)

Lab, S/N <i>i</i>	Lab, S/N <i>j</i> \longrightarrow												
	NRC A140		PTB 229074		PTB 229075		NMIJ B271		NMIJ B310				
	D_i	U_i		D_{ij}	U_{ij}		D_{ij}	U_{ij}		D_{ij}	U_{ij}		
	$/ \text{mK}$			$/ \text{mK}$			$/ \text{mK}$			$/ \text{mK}$			
NIST A128	-0.076	0.242		-0.209	0.560	-0.010	0.490	-0.149	0.490	-0.107	0.371	-0.116	0.369
NIST A129	-0.072	0.242		-0.205	0.560	-0.007	0.490	-0.145	0.490	-0.103	0.371	-0.112	0.369
NMi-VSL 226246	0.102	0.347		-0.031	0.612	0.168	0.550	0.030	0.550	0.071	0.447	0.062	0.445
NPL 221481	0.062	0.379		-0.071	0.631	0.127	0.570	-0.011	0.570	0.031	0.472	0.022	0.470
NPL 221485	-0.016	0.379		-0.148	0.631	0.050	0.570	-0.088	0.570	-0.046	0.472	-0.055	0.470
NRC A138	-0.024	0.504		-0.157	0.713	0.042	0.660	-0.097	0.660	-0.055	0.577	-0.064	0.576
NRC A140	0.133	0.504				0.198	0.660	0.060	0.660	0.102	0.577	0.093	0.576
PTB 229074	-0.066	0.426		-0.198	0.660			-0.138	0.603	-0.097	0.510	-0.106	0.509
PTB 229075	0.072	0.426		-0.060	0.660	0.138	0.603			0.042	0.510	0.033	0.509
NMIJ B271	0.031	0.281		-0.102	0.577	0.097	0.510	-0.042	0.510			-0.009	0.396
NMIJ B310	0.040	0.279		-0.093	0.576	0.106	0.509	-0.033	0.509	0.009	0.396		

CCT-K1 : Nominal temperature, $T_{90} = 2.997$ K
Degrees of equivalence, D_i , and expanded uncertainties ($k = 2$), U_i , expressed in mK



Key comparison CCT-K1

MEASURAND : Temperature T_{90} / K

NOMINAL TEMPERATURE : $T_{90} = 3.099$ K

Laboratory individual measurements

Lab: Laboratory

S/N i : serial number of Thermometer i

T_i : temperature value for Thermometer i

u_i : standard uncertainty in the calibration of Thermometer i

u_{comp} : standard uncertainty in the comparison measurements

$$u_{\text{comp}} = 0.114 \text{ mK}$$

Lab, S/N i	T_i / mK	u_i / mK
NIST A128	3.099314	0.041
NIST A129	3.099213	0.041
NMi-VSL 226246	3.099533	0.130
NPL 221481	3.099520	0.153
NPL 221483	3.099480	0.153
NPL 221485	3.099376	0.153
NRC A138	3.099336	0.225
NRC A140	3.099494	0.225
PTB 229074	3.099343	0.180
PTB 229075	3.099502	0.180
VNIIFTRI 79	3.099793	0.392
VNIIFTRI 89	3.099869	0.451
NMIJ B271	3.099409	0.082
NMIJ B310	3.099440	0.080

Key comparison CCT-K1

MEASURAND : Temperature T_{90} / K

KEY COMPARISON REFERENCE VALUE: $T_R = 3.099398$ K


Matrix of equivalence

Lab, S/N <i>i</i>	Lab, S/N <i>j</i> \longrightarrow													
	NIST A128		NIST A129		NMi-VSL 226246		NPL 221481		NPL 221483		NPL 221485		NRC A138	
	D_i	U_i	D_{ij}	U_{ij}	D_{ij}	U_{ij}	D_{ij}	U_{ij}	D_{ij}	U_{ij}	D_{ij}	U_{ij}	D_{ij}	U_{ij}
NIST A128	-0.084	0.242			0.101	0.342	-0.218	0.422	-0.205	0.451	-0.165	0.451	-0.062	0.451
NIST A129	-0.185	0.242	-0.101	0.342			-0.320	0.422	-0.307	0.451	-0.267	0.451	-0.163	0.451
NMi-VSL 226246	0.135	0.346	0.218	0.422	0.320	0.422			0.013	0.514	0.053	0.514	0.157	0.514
NPL 221481	0.122	0.381	0.205	0.451	0.307	0.451	-0.013	0.514			0.040	0.538	0.144	0.538
NPL 221483	0.082	0.381	0.165	0.451	0.267	0.451	-0.053	0.514	-0.040	0.538			0.104	0.538
NPL 221485	-0.022	0.381	0.062	0.451	0.163	0.451	-0.157	0.514	-0.144	0.538	-0.104	0.538		
NRC A138	-0.062	0.504	0.021	0.559	0.123	0.559	-0.197	0.611	-0.184	0.632	-0.144	0.632	-0.040	0.632
NRC A140	0.096	0.504	0.180	0.559	0.281	0.559	-0.038	0.611	-0.026	0.632	0.014	0.632	0.118	0.632
PTB 229074	-0.055	0.426	0.029	0.490	0.130	0.490	-0.189	0.548	-0.176	0.571	-0.136	0.571	-0.032	0.571
PTB 229075	0.104	0.426	0.188	0.490	0.289	0.490	-0.030	0.548	-0.017	0.571	0.023	0.571	0.127	0.571
VNIIFTRI 79	0.395	0.816	0.479	0.851	0.580	0.851	0.261	0.886	0.274	0.900	0.314	0.900	0.417	0.900
VNIIFTRI 89	0.471	0.931	0.555	0.962	0.656	0.962	0.336	0.993	0.349	1.005	0.389	1.005	0.493	1.005
NMIJ B271	0.011	0.280	0.095	0.370	0.196	0.370	-0.124	0.445	-0.111	0.473	-0.071	0.473	0.033	0.473
NMIJ B310	0.042	0.278	0.125	0.369	0.227	0.369	-0.093	0.444	-0.080	0.471	-0.040	0.471	0.064	0.471

Matrix of equivalence (Continued)

Lab, S/N <i>i</i>	Lab, S/N <i>j</i> →																
	NRC A140		PTB 229074		PTB 229075		VNIIIFTRI 79		VNIIIFTRI 89		NMJJ B271		NMJJ B310				
	D_i	U_i		D_{ij}	U_{ij}		D_{ij}	U_{ij}		D_{ij}	U_{ij}		D_{ij}	U_{ij}			
NIST A128	-0.084	0.242		-0.180	0.559	-0.029	0.490	-0.188	0.490	-0.479	0.851	-0.555	0.962	-0.095	0.370	-0.125	0.369
NIST A129	-0.185	0.242		-0.281	0.559	-0.130	0.490	-0.289	0.490	-0.580	0.851	-0.656	0.962	-0.196	0.370	-0.227	0.369
NMi-VSL 226246	0.135	0.346		0.038	0.611	0.189	0.548	0.030	0.548	-0.261	0.886	-0.336	0.993	0.124	0.445	0.093	0.444
NPL 221481	0.122	0.381		0.026	0.632	0.176	0.571	0.017	0.571	-0.274	0.900	-0.349	1.005	0.111	0.473	0.080	0.471
NPL 221483	0.082	0.381		-0.014	0.632	0.136	0.571	-0.023	0.571	-0.314	0.900	-0.389	1.005	0.071	0.473	0.040	0.471
NPL 221485	-0.022	0.381		-0.118	0.632	0.032	0.571	-0.127	0.571	-0.417	0.900	-0.493	1.005	-0.033	0.473	-0.064	0.471
NRC A138	-0.062	0.504		-0.158	0.713	-0.008	0.660	-0.167	0.660	-0.458	0.959	-0.533	1.058	-0.073	0.577	-0.104	0.576
NRC A140	0.096	0.504				0.151	0.660	-0.008	0.660	-0.299	0.959	-0.375	1.058	0.085	0.577	0.055	0.576
PTB 229074	-0.055	0.426		-0.151	0.660			-0.159	0.602	-0.450	0.920	-0.525	1.023	-0.065	0.510	-0.096	0.509
PTB 229075	0.104	0.426		0.008	0.660	0.159	0.602			-0.291	0.920	-0.366	1.023	0.094	0.510	0.063	0.509
VNIIIFTRI 79	0.395	0.816		0.299	0.959	0.450	0.920	0.291	0.920			-0.076	1.237	0.384	0.862	0.354	0.862
VNIIIFTRI 89	0.471	0.931		0.375	1.058	0.525	1.023	0.366	1.023	0.076	1.237			0.460	0.972	0.429	0.971
NMJJ B271	0.011	0.280		-0.085	0.577	0.065	0.510	-0.094	0.510	-0.384	0.862	-0.460	0.972			-0.031	0.395
NMJJ B310	0.042	0.278		-0.055	0.576	0.096	0.509	-0.063	0.509	-0.354	0.862	-0.429	0.971	0.031	0.395		

CCT-K1 : Nominal temperature, $T_{90} = 3.099$ K
Degrees of equivalence, D_i , and expanded uncertainties ($k = 2$), U_i , expressed in mK



Key comparison CCT-K1

MEASURAND : Temperature T_{90} / K

NOMINAL TEMPERATURE : $T_{90} = 3.400$ K

Laboratory individual measurements

Lab: Laboratory

S/N i : serial number of Thermometer i

T_i : temperature value for Thermometer i

u_i : standard uncertainty in the calibration of Thermometer i

u_{comp} : standard uncertainty in the comparison measurements

$$u_{\text{comp}} = 0.113 \text{ mK}$$

Lab, S/N i	T_i / mK	u_i / mK
NIST A128	3.400154	0.042
NIST A129	3.400049	0.042
NMi-VSL 226246	3.400396	0.127
NPL 221481	3.400390	0.157
NPL 221483	3.400290	0.157
NPL 221485	3.400242	0.157
NRC A138	3.400163	0.225
NRC A140	3.400390	0.225
PTB 229074	3.400225	0.180
PTB 229075	3.400352	0.180
VNIIFTRI 79	3.400410	0.396
VNIIFTRI 89	3.400480	0.455
NMIJ B271	3.400259	0.082
NMIJ B310	3.400234	0.081

Key comparison CCT-K1

MEASURAND : Temperature T_{90} / K

KEY COMPARISON REFERENCE VALUE: $T_R = 3.400235$ K


Matrix of equivalence

Lab, S/N i	Lab, S/N j \longrightarrow															
	NIST A128		NIST A129		NMi-VSL 226246		NPL 221481		NPL 221483		NPL 221485		NRC A138			
	D_i	U_i	D_{ij}	U_{ij}	D_{ij}	U_{ij}	D_{ij}	U_{ij}	D_{ij}	U_{ij}	D_{ij}	U_{ij}	D_{ij}	U_{ij}		
NIST A128	-0.081	0.241			0.106	0.341	-0.242	0.417	-0.235	0.455	-0.135	0.455	-0.088	0.455	-0.009	0.558
NIST A129	-0.186	0.241	-0.106	0.341			-0.348	0.417	-0.341	0.455	-0.241	0.455	-0.193	0.455	-0.114	0.558
NMi-VSL 226246	0.161	0.341	0.242	0.417	0.348	0.417			0.007	0.515	0.107	0.515	0.154	0.515	0.233	0.608
NPL 221481	0.155	0.386	0.235	0.455	0.341	0.455	-0.007	0.515			0.100	0.546	0.148	0.546	0.227	0.635
NPL 221483	0.055	0.386	0.135	0.455	0.241	0.455	-0.107	0.515	-0.100	0.546			0.048	0.546	0.127	0.635
NPL 221485	0.007	0.386	0.088	0.455	0.193	0.455	-0.154	0.515	-0.148	0.546	-0.048	0.546			0.079	0.635
NRC A138	-0.072	0.504	0.009	0.558	0.114	0.558	-0.233	0.608	-0.227	0.635	-0.127	0.635	-0.079	0.635		
NRC A140	0.155	0.504	0.236	0.558	0.341	0.558	-0.006	0.608	0.000	0.635	0.100	0.635	0.148	0.635	0.227	0.712
PTB 229074	-0.010	0.425	0.070	0.489	0.176	0.489	-0.172	0.545	-0.165	0.574	-0.065	0.574	-0.017	0.574	0.062	0.659
PTB 229075	0.117	0.425	0.198	0.489	0.304	0.489	-0.044	0.545	-0.037	0.574	0.063	0.574	0.110	0.574	0.189	0.659
VNIIFTRI 79	0.175	0.824	0.255	0.859	0.361	0.859	0.013	0.892	0.020	0.910	0.120	0.910	0.168	0.910	0.247	0.966
VNIIFTRI 89	0.245	0.937	0.325	0.968	0.431	0.968	0.083	0.997	0.090	1.014	0.190	1.014	0.237	1.014	0.316	1.064
NMIJ B271	0.024	0.279	0.105	0.369	0.211	0.369	-0.137	0.440	-0.130	0.476	-0.030	0.476	0.017	0.476	0.096	0.576
NMIJ B310	-0.001	0.278	0.080	0.368	0.186	0.368	-0.162	0.440	-0.156	0.476	-0.056	0.476	-0.008	0.476	0.071	0.575

Matrix of equivalence (Continued)

Lab, S/N <i>i</i>	Lab, S/N <i>j</i> →																
	NRC A140		PTB 229074		PTB 229075		VNIIIFTRI 79		VNIIIFTRI 89		NMJJ B271		NMJJ B310				
	D_i	U_i		D_{ij}	U_{ij}		D_{ij}	U_{ij}		D_{ij}	U_{ij}		D_{ij}	U_{ij}			
NIST A128	-0.081	0.241		-0.236	0.558	-0.070	0.489	-0.198	0.489	-0.255	0.859	-0.325	0.968	-0.105	0.369	-0.080	0.368
NIST A129	-0.186	0.241		-0.341	0.558	-0.176	0.489	-0.304	0.489	-0.361	0.859	-0.431	0.968	-0.211	0.369	-0.186	0.368
NMi-VSL 226246	0.161	0.341		0.006	0.608	0.172	0.545	0.044	0.545	-0.013	0.892	-0.083	0.997	0.137	0.440	0.162	0.440
NPL 221481	0.155	0.386		0.000	0.635	0.165	0.574	0.037	0.574	-0.020	0.910	-0.090	1.014	0.130	0.476	0.156	0.476
NPL 221483	0.055	0.386		-0.100	0.635	0.065	0.574	-0.063	0.574	-0.120	0.910	-0.190	1.014	0.030	0.476	0.056	0.476
NPL 221485	0.007	0.386		-0.148	0.635	0.017	0.574	-0.110	0.574	-0.168	0.910	-0.237	1.014	-0.017	0.476	0.008	0.476
NRC A138	-0.072	0.504		-0.227	0.712	-0.062	0.659	-0.189	0.659	-0.247	0.966	-0.316	1.064	-0.096	0.576	-0.071	0.575
NRC A140	0.155	0.504				0.165	0.659	0.038	0.659	-0.020	0.966	-0.089	1.064	0.131	0.576	0.156	0.575
PTB 229074	-0.010	0.425		-0.165	0.659			-0.128	0.601	-0.185	0.927	-0.255	1.029	-0.035	0.509	-0.009	0.508
PTB 229075	0.117	0.425		-0.038	0.659	0.128	0.601			-0.058	0.927	-0.127	1.029	0.093	0.509	0.118	0.508
VNIIIFTRI 79	0.175	0.824		0.020	0.966	0.185	0.927	0.058	0.927			-0.070	1.248	0.150	0.870	0.176	0.870
VNIIIFTRI 89	0.245	0.937		0.089	1.064	0.255	1.029	0.127	1.029	0.070	1.248			0.220	0.978	0.245	0.978
NMJJ B271	0.024	0.279		-0.131	0.576	0.035	0.509	-0.093	0.509	-0.150	0.870	-0.220	0.978			0.025	0.394
NMJJ B310	-0.001	0.278		-0.156	0.575	0.009	0.508	-0.118	0.508	-0.176	0.870	-0.245	0.978	-0.025	0.394		

CCT-K1 : Nominal temperature, $T_{90} = 3.400$ K
Degrees of equivalence, D_i , and expanded uncertainties ($k = 2$), U_i , expressed in mK



Key comparison CCT-K1

MEASURAND : Temperature T_{90} / K

NOMINAL TEMPERATURE : $T_{90} = 3.429$ K

Laboratory individual measurements

Lab: Laboratory

S/N i : serial number of Thermometer i

T_i : temperature value for Thermometer i

u_i : standard uncertainty in the calibration of Thermometer i

u_{comp} : standard uncertainty in the comparison measurements

$$u_{\text{comp}} = 0.113 \text{ mK}$$

Lab, S/N i	T_i / mK	u_i / mK
NIST A128	3.429195	0.042
NIST A129	3.429129	0.042
NMi-VSL 226246	3.429411	0.127
NPL 221481	3.429460	0.157
NPL 221485	3.429250	0.157
NRC A138	3.429182	0.225
NRC A140	3.429382	0.225
PTB 229074	3.429166	0.180
PTB 229075	3.429333	0.180
NMIJ B271	3.429255	0.082
NMIJ B310	3.429244	0.081

Key comparison CCT-K1

MEASURAND : Temperature T_{90} / K

KEY COMPARISON REFERENCE VALUE: $T_R = 3.429250$ K


Matrix of equivalence

Lab, S/N <i>i</i>	Lab, S/N <i>j</i> \longrightarrow											
	NIST A128		NIST A129		NMi-VSL 226246		NPL 221481		NPL 221485		NRC A138	
	D_i	U_i		D_{ij}	U_{ij}		D_{ij}	U_{ij}		D_{ij}	U_{ij}	
NIST A128	-0.055	0.241			0.066	0.340	-0.217	0.417	-0.266	0.455	-0.055	0.455
NIST A129	-0.121	0.241		-0.066	0.340		-0.282	0.417	-0.331	0.455	-0.121	0.455
NMi-VSL 226246	0.161	0.340		0.217	0.417	0.282	0.417		-0.049	0.515	0.161	0.515
NPL 221481	0.210	0.387		0.266	0.455	0.331	0.455	0.049	0.515		0.210	0.547
NPL 221485	0.000	0.387		0.055	0.455	0.121	0.455	-0.161	0.515	-0.210	0.547	0.279
NRC A138	-0.068	0.503		-0.013	0.558	0.053	0.558	-0.229	0.608	-0.279	0.635	
NRC A140	0.132	0.503		0.187	0.558	0.253	0.558	-0.029	0.608	-0.078	0.635	0.132
PTB 229074	-0.084	0.425		-0.029	0.488	0.037	0.488	-0.245	0.544	-0.294	0.575	-0.084
PTB 229075	0.083	0.425		0.138	0.488	0.204	0.488	-0.078	0.544	-0.127	0.575	0.083
NMIJ B271	0.005	0.279		0.060	0.369	0.126	0.369	-0.156	0.440	-0.205	0.477	0.005
NMIJ B310	-0.006	0.278		0.050	0.368	0.115	0.368	-0.167	0.439	-0.216	0.476	-0.006

Matrix of equivalence (Continued)

Lab, S/N <i>i</i>	Lab, S/N <i>j</i> \longrightarrow											
	NRC A140		PTB 229074		PTB 229075		NMIJ B271		NMIJ B310			
	D_i	U_i		D_{ij}	U_{ij}		D_{ij}	U_{ij}		D_{ij}	U_{ij}	
	$/ \text{mK}$			$/ \text{mK}$			$/ \text{mK}$			$/ \text{mK}$		
NIST A128	-0.055	0.241		-0.187	0.558	0.029	0.488	-0.138	0.488	-0.060	0.369	
NIST A129	-0.121	0.241		-0.253	0.558	-0.037	0.488	-0.204	0.488	-0.126	0.369	
NMi-VSL 226246	0.161	0.340		0.029	0.608	0.245	0.544	0.078	0.544	0.156	0.440	
NPL 221481	0.210	0.387		0.078	0.635	0.294	0.575	0.127	0.575	0.205	0.477	
NPL 221485	0.000	0.387		-0.132	0.635	0.084	0.575	-0.083	0.575	-0.005	0.477	
NRC A138	-0.068	0.503		-0.200	0.712	0.016	0.659	-0.151	0.659	-0.073	0.576	
NRC A140	0.132	0.503				0.216	0.659	0.049	0.659	0.127	0.576	
PTB 229074	-0.084	0.425		-0.216	0.659		-0.167	0.601	-0.089	0.508	-0.078	0.508
PTB 229075	0.083	0.425		-0.049	0.659	0.167	0.601		0.078	0.508	0.089	0.508
NMIJ B271	0.005	0.279		-0.127	0.576	0.089	0.508	-0.078	0.508		0.011	0.394
NMIJ B310	-0.006	0.278		-0.138	0.575	0.078	0.508	-0.089	0.508	-0.011	0.394	

CCT-K1 : Nominal temperature, $T_{90} = 3.429$ K
Degrees of equivalence, D_i , and expanded uncertainties ($k = 2$), U_i , expressed in mK



Key comparison CCT-K1

MEASURAND : Temperature T_{90} / K

NOMINAL TEMPERATURE : $T_{90} = 3.801$ K

Laboratory individual measurements

Lab: Laboratory

S/N i : serial number of Thermometer i

T_i : temperature value for Thermometer i

u_i : standard uncertainty in the calibration of Thermometer i

u_{comp} : standard uncertainty in the comparison measurements

$$u_{\text{comp}} = 0.112 \text{ mK}$$

Lab, S/N i	T_i / mK	u_i / mK
NIST A128	3.800754	0.042
NIST A129	3.800731	0.042
NMi-VSL 226246	3.801097	0.124
NPL 221481	3.801025	0.162
NPL 221483	3.800935	0.162
NPL 221485	3.800972	0.162
NRC A138	3.800894	0.225
NRC A140	3.800958	0.225
PTB 229074	3.800798	0.180
PTB 229075	3.801003	0.180
VNIIFTRI 79	3.801008	0.403
VNIIFTRI 89	3.801137	0.460
NMIJ B271	3.800962	0.085
NMIJ B310	3.800991	0.084

Key comparison CCT-K1

MEASURAND : Temperature T_{90} / K

KEY COMPARISON REFERENCE VALUE: $T_R = 3.800903$ K


Matrix of equivalence

Lab, S/N i	Lab, S/N j \longrightarrow														
	NIST A128		NIST A129		NMi-VSL 226246		NPL 221481		NPL 221483		NPL 221485		NRC A138		
	D_i	U_i	D_{ij}	U_{ij}	D_{ij}	U_{ij}	D_{ij}	U_{ij}	D_{ij}	U_{ij}	D_{ij}	U_{ij}	D_{ij}	U_{ij}	
NIST A128	-0.149	0.239			0.023	0.338	-0.343	0.411	-0.270	0.460	-0.180	0.460	-0.218	0.460	-0.140 0.557
NIST A129	-0.172	0.239	-0.023	0.338			-0.366	0.411	-0.294	0.460	-0.204	0.460	-0.241	0.460	-0.163 0.557
NMi-VSL 226246	0.194	0.334	0.343	0.411	0.366	0.411			0.073	0.516	0.163	0.516	0.125	0.516	0.203 0.603
NPL 221481	0.122	0.394	0.270	0.460	0.294	0.460	-0.073	0.516			0.090	0.556	0.053	0.556	0.130 0.638
NPL 221483	0.032	0.394	0.180	0.460	0.204	0.460	-0.163	0.516	-0.090	0.556			-0.037	0.556	0.040 0.638
NPL 221485	0.069	0.394	0.218	0.460	0.241	0.460	-0.125	0.516	-0.053	0.556	0.037	0.556			0.078 0.638
NRC A138	-0.009	0.503	0.140	0.557	0.163	0.557	-0.203	0.603	-0.130	0.638	-0.040	0.638	-0.078	0.638	
NRC A140	0.055	0.503	0.204	0.557	0.227	0.557	-0.139	0.603	-0.067	0.638	0.023	0.638	-0.014	0.638	0.064 0.711
PTB 229074	-0.105	0.424	0.043	0.487	0.067	0.487	-0.299	0.540	-0.227	0.578	-0.137	0.578	-0.174	0.578	-0.096 0.658
PTB 229075	0.100	0.424	0.249	0.487	0.272	0.487	-0.094	0.540	-0.021	0.578	0.069	0.578	0.032	0.578	0.109 0.658
VNIIFTRI 79	0.105	0.836	0.253	0.869	0.276	0.869	-0.090	0.900	-0.017	0.924	0.073	0.924	0.036	0.924	0.113 0.975
VNIIFTRI 89	0.234	0.946	0.383	0.976	0.406	0.976	0.040	1.003	0.112	1.025	0.202	1.025	0.165	1.025	0.243 1.071
NMIJ B271	0.059	0.281	0.207	0.369	0.231	0.369	-0.136	0.437	-0.063	0.484	0.027	0.484	-0.010	0.484	0.068 0.576
NMIJ B310	0.088	0.280	0.237	0.368	0.260	0.368	-0.106	0.436	-0.033	0.483	0.057	0.483	0.019	0.483	0.097 0.575

Matrix of equivalence (Continued)

Lab, S/N <i>i</i>	Lab, S/N <i>j</i> →																
	NRC A140		PTB 229074		PTB 229075		VNIIIFTRI 79		VNIIIFTRI 89		NMJJ B271		NMJJ B310				
	D_i	U_i		D_{ij}	U_{ij}		D_{ij}	U_{ij}		D_{ij}	U_{ij}		D_{ij}	U_{ij}			
NIST A128	-0.149	0.239		-0.204	0.557	-0.043	0.487	-0.249	0.487	-0.253	0.869	-0.383	0.976	-0.207	0.369	-0.237	0.368
NIST A129	-0.172	0.239		-0.227	0.557	-0.067	0.487	-0.272	0.487	-0.276	0.869	-0.406	0.976	-0.231	0.369	-0.260	0.368
NMi-VSL 226246	0.194	0.334		0.139	0.603	0.299	0.540	0.094	0.540	0.090	0.900	-0.040	1.003	0.136	0.437	0.106	0.436
NPL 221481	0.122	0.394		0.067	0.638	0.227	0.578	0.021	0.578	0.017	0.924	-0.112	1.025	0.063	0.484	0.033	0.483
NPL 221483	0.032	0.394		-0.023	0.638	0.137	0.578	-0.069	0.578	-0.073	0.924	-0.202	1.025	-0.027	0.484	-0.057	0.483
NPL 221485	0.069	0.394		0.014	0.638	0.174	0.578	-0.032	0.578	-0.036	0.924	-0.165	1.025	0.010	0.484	-0.019	0.483
NRC A138	-0.009	0.503		-0.064	0.711	0.096	0.658	-0.109	0.658	-0.113	0.975	-0.243	1.071	-0.068	0.576	-0.097	0.575
NRC A140	0.055	0.503				0.160	0.658	-0.046	0.658	-0.050	0.975	-0.179	1.071	-0.004	0.576	-0.033	0.575
PTB 229074	-0.105	0.424		-0.160	0.658			-0.206	0.600	-0.210	0.937	-0.339	1.037	-0.164	0.509	-0.193	0.508
PTB 229075	0.100	0.424		0.046	0.658	0.206	0.600			-0.004	0.937	-0.133	1.037	0.042	0.509	0.012	0.508
VNIIIFTRI 79	0.105	0.836		0.050	0.975	0.210	0.937	0.004	0.937			-0.129	1.262	0.046	0.882	0.016	0.881
VNIIIFTRI 89	0.234	0.946		0.179	1.071	0.339	1.037	0.133	1.037	0.129	1.262			0.175	0.987	0.146	0.987
NMJJ B271	0.059	0.281		0.004	0.576	0.164	0.509	-0.042	0.509	-0.046	0.882	-0.175	0.987			-0.029	0.397
NMJJ B310	0.088	0.280		0.033	0.575	0.193	0.508	-0.012	0.508	-0.016	0.881	-0.146	0.987	0.029	0.397		

CCT-K1 : Nominal temperature, $T_{90} = 3.801$ K
Degrees of equivalence, D_i , and expanded uncertainties ($k = 2$), U_i , expressed in mK



Key comparison CCT-K1

MEASURAND : Temperature T_{90} / K

NOMINAL TEMPERATURE : $T_{90} = 4.225$ K

Laboratory individual measurements

Lab: Laboratory

S/N i : serial number of Thermometer i

T_i : temperature value for Thermometer i

u_i : standard uncertainty in the calibration of Thermometer i

u_{comp} : standard uncertainty in the comparison measurements

$$u_{\text{comp}} = 0.111 \text{ mK}$$

Lab, S/N i	T_i / mK	u_i / mK
NIST A128	4.224713	0.042
NIST A129	4.224680	0.042
NMi-VSL 226246	4.224827	0.122
NPL 221481	4.224760	0.167
NPL 221483	4.224731	0.167
NPL 221485	4.224823	0.167
NRC A138	4.224794	0.225
NRC A140	4.224867	0.225
PTB 229074	4.224704	0.180
PTB 229075	4.224800	0.180
VNIIFTRI 79	4.224908	0.409
VNIIFTRI 89	4.225064	0.465
NMIJ B271	4.224928	0.088
NMIJ B310	4.224923	0.087

Key comparison CCT-K1

MEASURAND : Temperature T_{90} / K

KEY COMPARISON REFERENCE VALUE: $T_R = 4.224794$ K


Matrix of equivalence

Lab, S/N <i>i</i>	Lab, S/N <i>j</i> \longrightarrow													
	NIST A128		NIST A129		NMi-VSL 226246		NPL 221481		NPL 221483		NPL 221485		NRC A138	
	D_i	U_i	D_{ij}	U_{ij}	D_{ij}	U_{ij}	D_{ij}	U_{ij}	D_{ij}	U_{ij}	D_{ij}	U_{ij}	D_{ij}	U_{ij}
NIST A128	-0.081	0.238			0.033	0.336	-0.114	0.407	-0.046	0.466	-0.018	0.466	-0.110	0.466
NIST A129	-0.114	0.238	-0.033	0.336			-0.146	0.407	-0.079	0.466	-0.051	0.466	-0.143	0.466
NMi-VSL 226246	0.033	0.330	0.114	0.407	0.146	0.407			0.067	0.519	0.095	0.519	0.003	0.519
NPL 221481	-0.034	0.401	0.046	0.466	0.079	0.466	-0.067	0.519			0.028	0.567	-0.064	0.567
NPL 221483	-0.063	0.401	0.018	0.466	0.051	0.466	-0.095	0.519	-0.028	0.567			-0.092	0.567
NPL 221485	0.029	0.401	0.110	0.466	0.143	0.466	-0.003	0.519	0.064	0.567	0.092	0.567		0.029
NRC A138	0.000	0.502	0.081	0.555	0.114	0.555	-0.033	0.601	0.035	0.642	0.063	0.642	-0.029	0.642
NRC A140	0.073	0.502	0.154	0.555	0.186	0.555	0.040	0.601	0.107	0.642	0.135	0.642	0.043	0.642
PTB 229074	-0.090	0.423	-0.009	0.485	0.023	0.485	-0.123	0.537	-0.056	0.583	-0.028	0.583	-0.120	0.583
PTB 229075	0.006	0.423	0.087	0.485	0.120	0.485	-0.026	0.537	0.041	0.583	0.069	0.583	-0.023	0.583
VNIIFTRI 79	0.114	0.848	0.195	0.881	0.228	0.881	0.082	0.910	0.149	0.938	0.177	0.938	0.085	0.938
VNIIFTRI 89	0.270	0.956	0.351	0.985	0.383	0.985	0.237	1.011	0.304	1.036	0.332	1.036	0.240	1.036
NMIJ B271	0.134	0.283	0.214	0.370	0.247	0.370	0.101	0.435	0.168	0.491	0.196	0.491	0.104	0.491
NMIJ B310	0.129	0.282	0.210	0.369	0.243	0.369	0.097	0.434	0.164	0.490	0.192	0.490	0.100	0.490

Matrix of equivalence (Continued)

Lab, S/N <i>i</i>	Lab, S/N <i>j</i> →																
	NRC A140		PTB 229074		PTB 229075		VNIIIFTRI 79		VNIIIFTRI 89		NMJJ B271		NMJJ B310				
	D_i	U_i		D_{ij}	U_{ij}		D_{ij}	U_{ij}		D_{ij}	U_{ij}		D_{ij}	U_{ij}			
NIST A128	-0.081	0.238		-0.154	0.555	0.009	0.485	-0.087	0.485	-0.195	0.881	-0.351	0.985	-0.214	0.370	-0.210	0.369
NIST A129	-0.114	0.238		-0.186	0.555	-0.023	0.485	-0.120	0.485	-0.228	0.881	-0.383	0.985	-0.247	0.370	-0.243	0.369
NMi-VSL 226246	0.033	0.330		-0.040	0.601	0.123	0.537	0.026	0.537	-0.082	0.910	-0.237	1.011	-0.101	0.435	-0.097	0.434
NPL 221481	-0.034	0.401		-0.107	0.642	0.056	0.583	-0.041	0.583	-0.149	0.938	-0.304	1.036	-0.168	0.491	-0.164	0.490
NPL 221483	-0.063	0.401		-0.135	0.642	0.028	0.583	-0.069	0.583	-0.177	0.938	-0.332	1.036	-0.196	0.491	-0.192	0.490
NPL 221485	0.029	0.401		-0.043	0.642	0.120	0.583	0.023	0.583	-0.085	0.938	-0.240	1.036	-0.104	0.491	-0.100	0.490
NRC A138	0.000	0.502		-0.073	0.710	0.090	0.656	-0.006	0.656	-0.114	0.985	-0.270	1.079	-0.133	0.576	-0.129	0.576
NRC A140	0.073	0.502				0.163	0.656	0.067	0.656	-0.041	0.985	-0.197	1.079	-0.061	0.576	-0.056	0.576
PTB 229074	-0.090	0.423		-0.163	0.656			-0.096	0.598	-0.204	0.948	-0.360	1.045	-0.224	0.509	-0.219	0.508
PTB 229075	0.006	0.423		-0.067	0.656	0.096	0.598			-0.108	0.948	-0.264	1.045	-0.127	0.509	-0.123	0.508
VNIIIFTRI 79	0.114	0.848		0.041	0.985	0.204	0.948	0.108	0.948			-0.155	1.278	-0.019	0.894	-0.015	0.894
VNIIIFTRI 89	0.270	0.956		0.197	1.079	0.360	1.045	0.264	1.045	0.155	1.278			0.136	0.997	0.140	0.996
NMJJ B271	0.134	0.283		0.061	0.576	0.224	0.509	0.127	0.509	0.019	0.894	-0.136	0.997			0.004	0.400
NMJJ B310	0.129	0.282		0.056	0.576	0.219	0.508	0.123	0.508	0.015	0.894	-0.140	0.996	-0.004	0.400		

CCT-K1 : Nominal temperature, $T_{90} = 4.225$ K
Degrees of equivalence, D_i , and expanded uncertainties ($k = 2$), U_i , expressed in mK



Key comparison CCT-K1

MEASURAND : Temperature T_{90} / K

NOMINAL TEMPERATURE : $T_{90} = 4.478$ K

Laboratory individual measurements

Lab: Laboratory

S/N i : serial number of Thermometer i

T_i : temperature value for Thermometer i

u_i : standard uncertainty in the calibration of Thermometer i

u_{comp} : standard uncertainty in the comparison measurements

$$u_{\text{comp}} = 0.112 \text{ mK}$$

Lab, S/N i	T_i / mK	u_i / mK
NIST A128	4.477545	0.042
NIST A129	4.477546	0.042
NMi-VSL 226246	4.477459	0.206
NPL 221481	4.477410	0.210
NPL 221483	4.477489	0.201
NPL 221485	4.477502	0.201
NRC A138	4.477701	0.550
NRC A140	4.478134	0.550
PTB 229074	4.477475	0.180
PTB 229075	4.477406	0.180
VNIIFTRI 79	4.477674	0.413
VNIIFTRI 89	4.477791	0.468

Key comparison CCT-K1

MEASURAND : Temperature T_{90} / K

KEY COMPARISON REFERENCE VALUE: $T_R = 4.477522$ K


Matrix of equivalence

Lab, S/N i	Lab, S/N j \longrightarrow													
	NIST A128		NIST A129		NMi-VSL 226246		NPL 221481		NPL 221483		NPL 221485			
	D_i	U_i		D_{ij}	U_{ij}		D_{ij}	U_{ij}		D_{ij}	U_{ij}			
NIST A128	0.023	0.240		-0.001	0.339	0.086	0.526	0.135	0.533	0.056	0.519	0.043	0.519	
NIST A129	0.024	0.240	0.001	0.339		0.087	0.526	0.136	0.533	0.057	0.519	0.044	0.519	
NMi-VSL 226246	-0.063	0.468	-0.086	0.526	-0.087	0.526		0.049	0.668	-0.030	0.656	-0.043	0.656	
NPL 221481	-0.112	0.476	-0.135	0.533	-0.136	0.533	-0.049	0.668		-0.078	0.662	-0.091	0.662	
NPL 221483	-0.033	0.460	-0.056	0.519	-0.057	0.519	0.030	0.656	0.078	0.662		-0.013	0.651	
NPL 221485	-0.020	0.460	-0.043	0.519	-0.044	0.519	0.043	0.656	0.091	0.662	0.013	0.651		
NRC A138	0.179	1.123	0.156	1.148	0.155	1.148	0.242	1.216	0.290	1.219	0.212	1.213	0.199	1.213
NRC A140	0.612	1.123	0.589	1.148	0.588	1.148	0.675	1.216	0.724	1.219	0.645	1.213	0.632	1.213
PTB 229074	-0.047	0.424	-0.070	0.487	-0.071	0.487	0.016	0.632	0.064	0.638	-0.014	0.626	-0.027	0.626
PTB 229075	-0.116	0.424	-0.139	0.487	-0.140	0.487	-0.053	0.632	-0.005	0.638	-0.083	0.626	-0.096	0.626
VNIIFTRI 79	0.152	0.856	0.129	0.889	0.128	0.889	0.215	0.976	0.264	0.980	0.186	0.972	0.172	0.972
VNIIFTRI 89	0.269	0.962	0.246	0.991	0.245	0.991	0.332	1.070	0.381	1.073	0.302	1.066	0.289	1.066

Matrix of equivalence (Continued)

Lab, S/N <i>i</i>	Lab, S/N <i>j</i> \longrightarrow														
	NRC A138		NRC A140		PTB 229074		PTB 229075		VNIIIFTRI 79		VNIIIFTRI 89				
	D_i	U_i		D_{ij}	U_{ij}		D_{ij}	U_{ij}		D_{ij}	U_{ij}				
	$/ \text{mK}$			$/ \text{mK}$			$/ \text{mK}$			$/ \text{mK}$					
NIST A128	0.023	0.240		-0.156	1.148	-0.589	1.148	0.070	0.487	0.139	0.487	-0.129	0.889	-0.246	0.991
NIST A129	0.024	0.240		-0.155	1.148	-0.588	1.148	0.071	0.487	0.140	0.487	-0.128	0.889	-0.245	0.991
NMi-VSL 226246	-0.063	0.468		-0.242	1.216	-0.675	1.216	-0.016	0.632	0.053	0.632	-0.215	0.976	-0.332	1.070
NPL 221481	-0.112	0.476		-0.290	1.219	-0.724	1.219	-0.064	0.638	0.005	0.638	-0.264	0.980	-0.381	1.073
NPL 221483	-0.033	0.460		-0.212	1.213	-0.645	1.213	0.014	0.626	0.083	0.626	-0.186	0.972	-0.302	1.066
NPL 221485	-0.020	0.460		-0.199	1.213	-0.632	1.213	0.027	0.626	0.096	0.626	-0.172	0.972	-0.289	1.066
NRC A138	0.179	1.123				-0.433	1.588	0.226	1.200	0.295	1.200	0.026	1.412	-0.090	1.478
NRC A140	0.612	1.123		0.433	1.588		0.659	1.200	0.728	1.200	0.460	1.412	0.343	1.478	
PTB 229074	-0.047	0.424		-0.226	1.200	-0.659	1.200			0.069	0.600	-0.200	0.956	-0.316	1.051
PTB 229075	-0.116	0.424		-0.295	1.200	-0.728	1.200	-0.069	0.600			-0.268	0.956	-0.385	1.051
VNIIIFTRI 79	0.152	0.856		-0.026	1.412	-0.460	1.412	0.200	0.956	0.268	0.956			-0.117	1.288
VNIIIFTRI 89	0.269	0.962		0.090	1.478	-0.343	1.478	0.316	1.051	0.385	1.051	0.117	1.288		

CCT-K1 : Nominal temperature, $T_{90} = 4.478$ K
Degrees of equivalence, D_i , and expanded uncertainties ($k = 2$), U_i , expressed in mK



Key comparison CCT-K1

MEASURAND : Temperature T_{90} / K

NOMINAL TEMPERATURE : $T_{90} = 5.000$ K

Laboratory individual measurements

Lab: Laboratory

S/N i : serial number of Thermometer i

T_i : temperature value for Thermometer i

u_i : standard uncertainty in the calibration of Thermometer i

u_{comp} : standard uncertainty in the comparison measurements

$$u_{\text{comp}} = 0.114 \text{ mK}$$

Lab, S/N i	T_i / mK	u_i / mK
NIST A128	5.000483	0.067
NIST A129	5.000483	0.067
NMi-VSL 226246	5.000468	0.206
NPL 221481	5.000220	0.210
NPL 221483	5.000395	0.201
NPL 221485	5.000332	0.201
NRC A138	5.001124	0.550
NRC A140	5.001270	0.550
PTB 229074	5.000397	0.270
PTB 229075	5.000352	0.250
VNIIFTRI 79	5.000647	0.421
VNIIFTRI 89	5.000713	0.474

Key comparison CCT-K1

MEASURAND : Temperature T_{90} / K

KEY COMPARISON REFERENCE VALUE: $T_R = 5.000458$ K


Matrix of equivalence

Lab, S/N <i>i</i>	Lab, S/N <i>j</i> \longrightarrow												
	NIST A128		NIST A129		NMi-VSL 226246		NPL 221481		NPL 221483		NPL 221485		
	D_i	U_i		D_{ij}	U_{ij}		D_{ij}	U_{ij}		D_{ij}	U_{ij}		
NIST A128	0.025	0.264		0.000	0.374	0.015	0.539	0.264	0.546	0.088	0.532	0.151	0.532
NIST A129	0.025	0.264		0.000	0.374	0.015	0.539	0.263	0.546	0.088	0.532	0.151	0.532
NMi-VSL 226246	0.010	0.470		-0.015	0.539	-0.015	0.539	0.248	0.670	0.073	0.659	0.136	0.659
NPL 221481	-0.238	0.478		-0.264	0.546	-0.263	0.546	-0.248	0.670	-0.175	0.665	-0.113	0.665
NPL 221483	-0.063	0.462		-0.088	0.532	-0.088	0.532	-0.073	0.659	0.175	0.665	0.063	0.654
NPL 221485	-0.126	0.462		-0.151	0.532	-0.151	0.532	-0.136	0.659	0.113	0.665	-0.063	0.654
NRC A138	0.666	1.123		0.640	1.154	0.641	1.154	0.655	1.218	0.904	1.221	0.728	1.215
NRC A140	0.812	1.123		0.787	1.154	0.787	1.154	0.802	1.218	1.051	1.221	0.875	1.215
PTB 229074	-0.061	0.586		-0.087	0.643	-0.086	0.643	-0.071	0.751	0.177	0.756	0.002	0.746
PTB 229075	-0.106	0.550		-0.131	0.610	-0.131	0.610	-0.116	0.723	0.132	0.728	-0.043	0.718
VNIIFTRI 79	0.189	0.873		0.163	0.912	0.164	0.912	0.178	0.992	0.427	0.996	0.251	0.988
VNIIFTRI 89	0.255	0.975		0.229	1.010	0.230	1.010	0.245	1.083	0.493	1.086	0.318	1.079

Matrix of equivalence (Continued)

Lab, S/N <i>i</i>	Lab, S/N <i>j</i> \longrightarrow														
	NRC A138		NRC A140		PTB 229074		PTB 229075		VNIIIFTRI 79		VNIIIFTRI 89				
	D_i	U_i		D_{ij}	U_{ij}		D_{ij}	U_{ij}		D_{ij}	U_{ij}				
	$/ \text{mK}$			$/ \text{mK}$			$/ \text{mK}$			$/ \text{mK}$					
NIST A128	0.025	0.264		-0.640	1.154	-0.787	1.154	0.087	0.643	0.131	0.610	-0.163	0.912	-0.229	1.010
NIST A129	0.025	0.264		-0.641	1.154	-0.787	1.154	0.086	0.643	0.131	0.610	-0.164	0.912	-0.230	1.010
NMi-VSL 226246	0.010	0.470		-0.655	1.218	-0.802	1.218	0.071	0.751	0.116	0.723	-0.178	0.992	-0.245	1.083
NPL 221481	-0.238	0.478		-0.904	1.221	-1.051	1.221	-0.177	0.756	-0.132	0.728	-0.427	0.996	-0.493	1.086
NPL 221483	-0.063	0.462		-0.728	1.215	-0.875	1.215	-0.002	0.746	0.043	0.718	-0.251	0.988	-0.318	1.079
NPL 221485	-0.126	0.462		-0.791	1.215	-0.938	1.215	-0.065	0.746	-0.020	0.718	-0.314	0.988	-0.380	1.079
NRC A138	0.666	1.123				-0.147	1.589	0.727	1.267	0.772	1.251	0.477	1.423	0.411	1.488
NRC A140	0.812	1.123		0.147	1.589			0.873	1.267	0.918	1.251	0.624	1.423	0.558	1.488
PTB 229074	-0.061	0.586		-0.727	1.267	-0.873	1.267			0.045	0.804	-0.250	1.052	-0.316	1.138
PTB 229075	-0.106	0.550		-0.772	1.251	-0.918	1.251	-0.045	0.804			-0.294	1.032	-0.361	1.119
VNIIIFTRI 79	0.189	0.873		-0.477	1.423	-0.624	1.423	0.250	1.052	0.294	1.032			-0.066	1.309
VNIIIFTRI 89	0.255	0.975		-0.411	1.488	-0.558	1.488	0.316	1.138	0.361	1.119	0.066	1.309		

CCT-K1 : Nominal temperature, $T_{90} = 5.000$ K
Degrees of equivalence, D_i , and expanded uncertainties ($k = 2$), U_i , expressed in mK



Key comparison CCT-K1

MEASURAND : Temperature T_{90} / K

NOMINAL TEMPERATURE : $T_{90} = 5.948$ K

Laboratory individual measurements

Lab: Laboratory

S/N i : serial number of Thermometer i

T_i : temperature value for Thermometer i

u_i : standard uncertainty in the calibration of Thermometer i

u_{comp} : standard uncertainty in the comparison measurements

$$u_{\text{comp}} = 0.118 \text{ mK}$$

Lab, S/N i	T_i / mK	u_i / mK
NIST A128	5.947948	0.064
NIST A129	5.948148	0.064
NMi-VSL 226246	5.948420	0.206
NPL 221481	5.948030	0.210
NPL 221483	5.948374	0.201
NPL 221485	5.947935	0.201
NRC A138	5.948869	0.550
NRC A140	5.949112	0.550
PTB 229074	5.948358	0.273
PTB 229075	5.948217	0.252
VNIIFTRI 79	5.948466	0.436
VNIIFTRI 89	5.948584	0.485

Key comparison CCT-K1

MEASURAND : Temperature T_{90} / K

KEY COMPARISON REFERENCE VALUE: $T_R = 5.948165$ K

Matrix of equivalence


Lab, S/N i	Lab, S/N j \longrightarrow														
	NIST A128		NIST A129		NMi-VSL 226246		NPL 221481		NPL 221483		NPL 221485				
	D_i	U_i		D_{ij}	U_{ij}		D_{ij}	U_{ij}		D_{ij}	U_{ij}				
	/ mK	/ mK		/ mK	/ mK		/ mK	/ mK		/ mK	/ mK				
NIST A128	-0.217	0.268			-0.200	0.379	-0.472	0.544	-0.082	0.551	-0.426	0.537	0.013	0.537	
NIST A129	-0.017	0.268		0.200	0.379		-0.271	0.544	0.118	0.551	-0.226	0.537	0.213	0.537	
NMi-VSL 226246	0.255	0.474		0.472	0.544	0.271	0.544		0.389	0.675	0.046	0.664	0.485	0.664	
NPL 221481	-0.135	0.481		0.082	0.551	-0.118	0.551	-0.389	0.675		-0.344	0.670	0.095	0.670	
NPL 221483	0.209	0.466		0.426	0.537	0.226	0.537	-0.046	0.664	0.344	0.670		0.439	0.659	
NPL 221485	-0.230	0.466		-0.013	0.537	-0.213	0.537	-0.485	0.664	-0.095	0.670	-0.439	0.659		
NRC A138	0.704	1.125		0.921	1.156	0.721	1.156	0.449	1.220	0.838	1.224	0.495	1.217	0.934	1.217
NRC A140	0.947	1.125		1.164	1.156	0.963	1.156	0.692	1.220	1.081	1.224	0.738	1.217	1.177	1.217
PTB 229074	0.193	0.594		0.410	0.652	0.209	0.652	-0.062	0.760	0.327	0.765	-0.016	0.755	0.423	0.755
PTB 229075	0.052	0.557		0.269	0.618	0.068	0.618	-0.203	0.731	0.186	0.736	-0.157	0.726	0.282	0.726
VNIIFTRI 79	0.301	0.904		0.518	0.943	0.317	0.943	0.046	1.021	0.435	1.024	0.092	1.017	0.531	1.017
VNIIFTRI 89	0.419	0.999		0.636	1.034	0.435	1.034	0.164	1.106	0.553	1.109	0.210	1.102	0.649	1.102

Matrix of equivalence (Continued)

Lab, S/N <i>i</i>	Lab, S/N <i>j</i> \longrightarrow														
	NRC A138		NRC A140		PTB 229074		PTB 229075		VNIIIFTRI 79		VNIIIFTRI 89				
	D_i	U_i		D_{ij}	U_{ij}		D_{ij}	U_{ij}		D_{ij}	U_{ij}				
	$/ \text{mK}$			$/ \text{mK}$			$/ \text{mK}$			$/ \text{mK}$					
NIST A128	-0.217	0.268		-0.921	1.156	-1.164	1.156	-0.410	0.652	-0.269	0.618	-0.518	0.943	-0.636	1.034
NIST A129	-0.017	0.268		-0.721	1.156	-0.963	1.156	-0.209	0.652	-0.068	0.618	-0.317	0.943	-0.435	1.034
NMi-VSL 226246	0.255	0.474		-0.449	1.220	-0.692	1.220	0.062	0.760	0.203	0.731	-0.046	1.021	-0.164	1.106
NPL 221481	-0.135	0.481		-0.838	1.224	-1.081	1.224	-0.327	0.765	-0.186	0.736	-0.435	1.024	-0.553	1.109
NPL 221483	0.209	0.466		-0.495	1.217	-0.738	1.217	0.016	0.755	0.157	0.726	-0.092	1.017	-0.210	1.102
NPL 221485	-0.230	0.466		-0.934	1.217	-1.177	1.217	-0.423	0.755	-0.282	0.726	-0.531	1.017	-0.649	1.102
NRC A138	0.704	1.125				-0.243	1.591	0.511	1.272	0.652	1.255	0.403	1.443	0.285	1.504
NRC A140	0.947	1.125		0.243	1.591		0.754	1.272	0.895	1.255	0.646	1.443	0.528	1.504	
PTB 229074	0.193	0.594		-0.511	1.272	-0.754	1.272			0.141	0.815	-0.108	1.082	-0.226	1.163
PTB 229075	0.052	0.557		-0.652	1.255	-0.895	1.255	-0.141	0.815			-0.249	1.062	-0.367	1.144
VNIIIFTRI 79	0.301	0.904		-0.403	1.443	-0.646	1.443	0.108	1.082	0.249	1.062			-0.118	1.347
VNIIIFTRI 89	0.419	0.999		-0.285	1.504	-0.528	1.504	0.226	1.163	0.367	1.144	0.118	1.347		

CCT-K1 : Nominal temperature, $T_{90} = 5.948$ K

Degrees of equivalence, D_i , and expanded uncertainties ($k = 2$), U_i , expressed in mK



Key comparison CCT-K1

MEASURAND : Temperature T_{90} / K

NOMINAL TEMPERATURE : $T_{90} = 7.202$ K

Laboratory individual measurements

Lab: Laboratory

S/N i : serial number of Thermometer i

T_i : temperature value for Thermometer i

u_i : standard uncertainty in the calibration of Thermometer i

u_{comp} : standard uncertainty in the comparison measurements

$$u_{\text{comp}} = 0.122 \text{ mK}$$

Lab, S/N i	T_i / mK	u_i / mK
NIST A128	7.201615	0.070
NIST A129	7.201513	0.070
NMi-VSL 226246	7.201679	0.206
NPL 221481	7.201400	0.210
NPL 221483	7.201415	0.201
NPL 221485	7.201191	0.201
NRC A138	7.202363	0.550
NRC A140	7.202306	0.550
PTB 229074	7.201593	0.277
PTB 229075	7.201538	0.256
VNIIFTRI 79	7.201591	0.456
VNIIFTRI 89	7.201839	0.501

Key comparison CCT-K1

MEASURAND : Temperature T_{90} / K

KEY COMPARISON REFERENCE VALUE: $T_R = 7.201544$ K


Matrix of equivalence

Lab, S/N i	Lab, S/N j \longrightarrow																
	NIST A128		NIST A129		NMi-VSL 226246		NPL 221481		NPL 221483		NPL 221485						
	D_i	U_i		D_{ij}	U_{ij}		D_{ij}	U_{ij}		D_{ij}	U_{ij}		D_{ij}	U_{ij}		D_{ij}	U_{ij}
NIST A128	0.071	0.282			0.102	0.399	-0.064	0.555	0.215	0.562	0.200	0.549	0.424	0.549			
NIST A129	-0.031	0.282		-0.102	0.399		-0.166	0.555	0.114	0.562	0.098	0.549	0.322	0.549			
NMi-VSL 226246	0.135	0.478		0.064	0.555	0.166	0.555		0.279	0.682	0.263	0.671	0.488	0.671			
NPL 221481	-0.144	0.486		-0.215	0.562	-0.114	0.562	-0.279	0.682			-0.016	0.677	0.209	0.677		
NPL 221483	-0.129	0.471		-0.200	0.549	-0.098	0.549	-0.263	0.671	0.016	0.677			0.224	0.666		
NPL 221485	-0.353	0.471		-0.424	0.549	-0.322	0.549	-0.488	0.671	-0.209	0.677	-0.224	0.666				
NRC A138	0.819	1.127		0.748	1.162	0.849	1.162	0.684	1.224	0.963	1.227	0.947	1.221	1.172	1.221		
NRC A140	0.762	1.127		0.691	1.162	0.793	1.162	0.628	1.224	0.907	1.227	0.891	1.221	1.115	1.221		
PTB 229074	0.049	0.605		-0.022	0.668	0.080	0.668	-0.086	0.771	0.193	0.776	0.178	0.767	0.402	0.767		
PTB 229075	-0.006	0.567		-0.077	0.633	0.025	0.633	-0.141	0.742	0.139	0.747	0.123	0.737	0.347	0.737		
VNIIFTRI 79	0.047	0.945		-0.024	0.986	0.078	0.986	-0.088	1.059	0.191	1.062	0.175	1.055	0.400	1.055		
VNIIFTRI 89	0.295	1.031		0.224	1.069	0.326	1.069	0.160	1.136	0.439	1.140	0.424	1.133	0.648	1.133		

Matrix of equivalence (Continued)

Lab, S/N <i>i</i>	Lab, S/N <i>j</i> \longrightarrow															
	NRC A138		NRC A140		PTB 229074		PTB 229075		VNIIIFTRI 79		VNIIIFTRI 89					
	D_i	U_i		D_{ij}	U_{ij}		D_{ij}	U_{ij}		D_{ij}	U_{ij}					
	$/ \text{mK}$			$/ \text{mK}$			$/ \text{mK}$			$/ \text{mK}$						
NIST A128	0.071	0.282		-0.748	1.162	-0.691	1.162	0.022	0.668	0.077	0.633	0.024	0.986	-0.224	1.069	
NIST A129	-0.031	0.282		-0.849	1.162	-0.793	1.162	-0.080	0.668	-0.025	0.633	-0.078	0.986	-0.326	1.069	
NMi-VSL 226246	0.135	0.478		-0.684	1.224	-0.628	1.224	0.086	0.771	0.141	0.742	0.088	1.059	-0.160	1.136	
NPL 221481	-0.144	0.486		-0.963	1.227	-0.907	1.227	-0.193	0.776	-0.139	0.747	-0.191	1.062	-0.439	1.140	
NPL 221483	-0.129	0.471		-0.947	1.221	-0.891	1.221	-0.178	0.767	-0.123	0.737	-0.175	1.055	-0.424	1.133	
NPL 221485	-0.353	0.471		-1.172	1.221	-1.115	1.221	-0.402	0.767	-0.347	0.737	-0.400	1.055	-0.648	1.133	
NRC A138	0.819	1.127				0.056	1.594	0.770	1.279	0.824	1.261	0.772	1.470	0.524	1.527	
NRC A140	0.762	1.127				-0.056	1.594		0.713	1.279	0.768	1.261	0.715	1.470	0.467	1.527
PTB 229074	0.049	0.605				-0.770	1.279	-0.713		0.055	0.829	0.002	1.122	-0.246	1.195	
PTB 229075	-0.006	0.567				-0.824	1.261	-0.768	1.261	-0.055	0.829		-0.053	1.102	-0.301	1.176
VNIIIFTRI 79	0.047	0.945				-0.772	1.470	-0.715	1.470	-0.002	1.122	0.053	1.102		-0.248	1.398
VNIIIFTRI 89	0.295	1.031				-0.524	1.527	-0.467	1.527	0.246	1.195	0.301	1.176	0.248	1.398	

CCT-K1 : Nominal temperature, $T_{90} = 7.202$ K
Degrees of equivalence, D_i , and expanded uncertainties ($k = 2$), U_i , expressed in mK



Key comparison CCT-K1

MEASURAND : Temperature T_{90} / K

NOMINAL TEMPERATURE : $T_{90} = 8.296$ K

Laboratory individual measurements

Lab: Laboratory

S/N i : serial number of Thermometer i

T_i : temperature value for Thermometer i

u_i : standard uncertainty in the calibration of Thermometer i

u_{comp} : standard uncertainty in the comparison measurements

$$u_{\text{comp}} = 0.127 \text{ mK}$$

Lab, S/N i	T_i / mK	u_i / mK
NIST A128	8.296546	0.076
NIST A129	8.296561	0.076
NMi-VSL 226246	8.296278	0.206
NPL 221481	8.296020	0.210
NPL 221485	8.295779	0.201
NRC A138	8.297318	0.550
NRC A140	8.297308	0.550
PTB 229074	8.296214	0.280
PTB 229075	8.296135	0.258

Key comparison CCT-K1

MEASURAND : Temperature T_{90} / K

KEY COMPARISON REFERENCE VALUE: $T_R = 8.296372$ K


Matrix of equivalence

Lab, S/N <i>i</i>	Lab, S/N <i>j</i> \longrightarrow											
	NIST A128		NIST A129		NMi-VSL 226246		NPL 221481		NPL 221485		NRC A138	
	D_i	U_i		D_{ij}	U_{ij}		D_{ij}	U_{ij}		D_{ij}	U_{ij}	
NIST A128	0.174	0.296		-0.015	0.418	0.267	0.566	0.526	0.573	0.767	0.560	-0.773 1.167
NIST A129	0.189	0.296	0.015 0.418			0.283	0.566	0.541	0.573	0.782	0.560	-0.757 1.167
NMi-VSL 226246	-0.094	0.483	-0.267 0.566	-0.283 0.566			0.259	0.688	0.499	0.677	-1.040	1.228
NPL 221481	-0.352	0.490	-0.526 0.573	-0.541 0.573	-0.259 0.688				0.241	0.683	-1.299	1.231
NPL 221485	-0.593	0.475	-0.767 0.560	-0.782 0.560	-0.499 0.677	-0.241 0.683				-1.539	1.225	
NRC A138	0.946	1.129	0.773 1.167	0.757 1.167	1.040 1.228	1.299 1.231	1.539 1.225					
NRC A140	0.936	1.129	0.762 1.167	0.747 1.167	1.030 1.228	1.288 1.231	1.529 1.225	-0.010	1.596			
PTB 229074	-0.158	0.615	-0.332 0.682	-0.347 0.682	-0.064 0.782	0.194 0.786	0.435 0.777	-1.104	1.285			
PTB 229075	-0.237	0.575	-0.410 0.647	-0.426 0.647	-0.143 0.751	0.116 0.756	0.356 0.746	-1.183	1.267			

Matrix of equivalence (Continued)

Lab, S/N <i>i</i>	Lab, S/N <i>j</i> \longrightarrow							
	D _{<i>i</i>} U _{<i>i</i>} / mK		NRC A140		PTB 229074		PTB 229075	
	D _{<i>ij</i>}	U _{<i>ij</i>}	D _{<i>ij</i>}	U _{<i>ij</i>}	D _{<i>ij</i>}	U _{<i>ij</i>}	D _{<i>ij</i>}	U _{<i>ij</i>}
NIST A128	0.174	0.296	-0.762	1.167	0.332	0.682	0.410	0.647
NIST A129	0.189	0.296	-0.747	1.167	0.347	0.682	0.426	0.647
NMi-VSL 226246	-0.094	0.483	-1.030	1.228	0.064	0.782	0.143	0.751
NPL 221481	-0.352	0.490	-1.288	1.231	-0.194	0.786	-0.116	0.756
NPL 221485	-0.593	0.475	-1.529	1.225	-0.435	0.777	-0.356	0.746
NRC A138	0.946	1.129	0.010	1.596	1.104	1.285	1.183	1.267
NRC A140	0.936	1.129			1.094	1.285	1.173	1.267
PTB 229074	-0.158	0.615	-1.094	1.285			0.078	0.842
PTB 229075	-0.237	0.575	-1.173	1.267	-0.078	0.842		

CCT-K1 : Nominal temperature, $T_{90} = 8.296$ K
Degrees of equivalence, D_i , and expanded uncertainties ($k = 2$), U_i , expressed in mK



Key comparison CCT-K1

MEASURAND : Temperature T_{90} / K

NOMINAL TEMPERATURE : $T_{90} = 8.400$ K

Laboratory individual measurements

Lab: Laboratory

S/N i : serial number of Thermometer i

T_i : temperature value for Thermometer i

u_i : standard uncertainty in the calibration of Thermometer i

u_{comp} : standard uncertainty in the comparison measurements

$$u_{\text{comp}} = 0.127 \text{ mK}$$

Lab, S/N i	T_i / mK	u_i / mK
NIST A128	8.399816	0.076
NIST A129	8.399757	0.076
NMi-VSL 226246	8.399564	0.206
NPL 221481	8.399350	0.210
NPL 221483	8.399316	0.201
NPL 221485	8.398898	0.201
NRC A138	8.400617	0.550
NRC A140	8.400524	0.550
PTB 229074	8.399608	0.280
PTB 229075	8.399518	0.259
VNIIFTRI 79	8.399714	0.475
VNIIFTRI 89	8.400138	0.515

Key comparison CCT-K1

MEASURAND : Temperature T_{90} / K

KEY COMPARISON REFERENCE VALUE: $T_R = 8.399612$ K

Matrix of equivalence


Lab, S/N i	Lab, S/N j \longrightarrow													
	NIST A128		NIST A129		NMi-VSL 226246		NPL 221481		NPL 221483		NPL 221485			
	D_i	U_i		D_{ij}	U_{ij}		D_{ij}	U_{ij}		D_{ij}	U_{ij}			
NIST A128	0.204	0.296		0.060	0.419	0.253	0.567	0.467	0.573	0.500	0.560	0.918	0.560	
NIST A129	0.145	0.296	-0.060	0.419		0.193	0.567	0.407	0.573	0.441	0.560	0.859	0.560	
NMi-VSL 226246	-0.048	0.483	-0.253	0.567	-0.193	0.567		0.214	0.689	0.248	0.678	0.666	0.678	
NPL 221481	-0.262	0.491	-0.467	0.573	-0.407	0.573	-0.214	0.689		0.034	0.683	0.452	0.683	
NPL 221483	-0.296	0.475	-0.500	0.560	-0.441	0.560	-0.248	0.678	-0.034	0.683		0.418	0.672	
NPL 221485	-0.714	0.475	-0.918	0.560	-0.859	0.560	-0.666	0.678	-0.452	0.683	-0.418	0.672		
NRC A138	1.005	1.129	0.800	1.167	0.860	1.167	1.053	1.228	1.267	1.231	1.301	1.225	1.719	1.225
NRC A140	0.912	1.129	0.708	1.167	0.767	1.167	0.960	1.228	1.174	1.231	1.208	1.225	1.626	1.225
PTB 229074	-0.004	0.616	-0.208	0.683	-0.149	0.683	0.044	0.783	0.258	0.787	0.292	0.778	0.710	0.778
PTB 229075	-0.094	0.576	-0.299	0.648	-0.239	0.648	-0.046	0.752	0.168	0.757	0.202	0.747	0.620	0.747
VNIIFTRI 79	0.102	0.983	-0.103	1.027	-0.043	1.027	0.150	1.096	0.364	1.099	0.397	1.092	0.815	1.092
VNIIFTRI 89	0.526	1.061	0.322	1.101	0.381	1.101	0.574	1.166	0.788	1.169	0.822	1.162	1.240	1.162

Matrix of equivalence (Continued)

Lab, S/N <i>i</i>	Lab, S/N <i>j</i> \longrightarrow														
	NRC A138		NRC A140		PTB 229074		PTB 229075		VNIIIFTRI 79		VNIIIFTRI 89				
	D_i	U_i		D_{ij}	U_{ij}		D_{ij}	U_{ij}		D_{ij}	U_{ij}				
	$/ \text{mK}$			$/ \text{mK}$			$/ \text{mK}$			$/ \text{mK}$					
NIST A128	0.204	0.296		-0.800	1.167	-0.708	1.167	0.208	0.683	0.299	0.648	0.103	1.027	-0.322	1.101
NIST A129	0.145	0.296		-0.860	1.167	-0.767	1.167	0.149	0.683	0.239	0.648	0.043	1.027	-0.381	1.101
NMi-VSL 226246	-0.048	0.483		-1.053	1.228	-0.960	1.228	-0.044	0.783	0.046	0.752	-0.150	1.096	-0.574	1.166
NPL 221481	-0.262	0.491		-1.267	1.231	-1.174	1.231	-0.258	0.787	-0.168	0.757	-0.364	1.099	-0.788	1.169
NPL 221483	-0.296	0.475		-1.301	1.225	-1.208	1.225	-0.292	0.778	-0.202	0.747	-0.397	1.092	-0.822	1.162
NPL 221485	-0.714	0.475		-1.719	1.225	-1.626	1.225	-0.710	0.778	-0.620	0.747	-0.815	1.092	-1.240	1.162
NRC A138	1.005	1.129			0.093	1.597	1.009	1.286	1.099	1.268	0.903	1.497	0.479	1.549	
NRC A140	0.912	1.129		-0.093	1.597		0.916	1.286	1.006	1.268	0.810	1.497	0.386	1.549	
PTB 229074	-0.004	0.616		-1.009	1.286	-0.916	1.286		0.090	0.843	-0.105	1.160	-0.530	1.227	
PTB 229075	-0.094	0.576		-1.099	1.268	-1.006	1.268	-0.090	0.843		-0.196	1.140	-0.620	1.207	
VNIIIFTRI 79	0.102	0.983		-0.903	1.497	-0.810	1.497	0.105	1.160	0.196	1.140		-0.425	1.446	
VNIIIFTRI 89	0.526	1.061		-0.479	1.549	-0.386	1.549	0.530	1.227	0.620	1.207	0.425	1.446		

CCT-K1 : Nominal temperature, $T_{90} = 8.400$ K

Degrees of equivalence, D_i , and expanded uncertainties ($k = 2$), U_i , expressed in mK



Key comparison CCT-K1

MEASURAND : Temperature T_{90} / K

NOMINAL TEMPERATURE : $T_{90} = 9.508$ K

Laboratory individual measurements

Lab: Laboratory

S/N i : serial number of Thermometer i

T_i : temperature value for Thermometer i

u_i : standard uncertainty in the calibration of Thermometer i

u_{comp} : standard uncertainty in the comparison measurements

$$u_{\text{comp}} = 0.131 \text{ mK}$$

Lab, S/N i	T_i / mK	u_i / mK
NIST A128	9.508227	0.086
NIST A129	9.508237	0.086
NMi-VSL 226246	9.508099	0.206
NPL 221481	9.507680	0.210
NPL 221483	9.507531	0.201
NPL 221485	9.507387	0.201
NRC A138	9.509019	0.550
NRC A140	9.508971	0.550
PTB 229074	9.507997	0.284
PTB 229075	9.507851	0.262
VNIIFTRI 79	9.508226	0.492
VNIIFTRI 89	9.508531	0.528

Key comparison CCT-K1

MEASURAND : Temperature T_{90} / K

KEY COMPARISON REFERENCE VALUE: $T_R = 9.50802$ K


Matrix of equivalence

Lab, S/N i	Lab, S/N j \longrightarrow													
	NIST A128		NIST A129		NMi-VSL 226246		NPL 221481		NPL 221483		NPL 221485			
	D_i	U_i		D_{ij}	U_{ij}		D_{ij}	U_{ij}		D_{ij}	U_{ij}			
	/ mK	/ mK		/ mK	/ mK		/ mK	/ mK		/ mK	/ mK			
NIST A128	0.207	0.313		-0.010	0.443	0.128	0.580	0.547	0.586	0.696	0.573	0.841	0.573	
NIST A129	0.217	0.313	0.010	0.443		0.138	0.580	0.557	0.586	0.706	0.573	0.850	0.573	
NMi-VSL 226246	0.079	0.488	-0.128	0.580	-0.138	0.580		0.419	0.695	0.568	0.684	0.712	0.684	
NPL 221481	-0.340	0.495	-0.547	0.586	-0.557	0.586	-0.419	0.695		0.149	0.690	0.293	0.690	
NPL 221483	-0.489	0.480	-0.696	0.573	-0.706	0.573	-0.568	0.684	-0.149	0.690		0.145	0.679	
NPL 221485	-0.633	0.480	-0.841	0.573	-0.850	0.573	-0.712	0.684	-0.293	0.690	-0.145	0.679		
NRC A138	0.999	1.131	0.792	1.173	0.782	1.173	0.920	1.231	1.339	1.234	1.488	1.228	1.633	1.228
NRC A140	0.951	1.131	0.744	1.173	0.734	1.173	0.872	1.231	1.291	1.234	1.440	1.228	1.584	1.228
PTB 229074	-0.023	0.625	-0.230	0.699	-0.240	0.699	-0.102	0.793	0.317	0.798	0.466	0.788	0.611	0.788
PTB 229075	-0.169	0.585	-0.376	0.664	-0.386	0.664	-0.248	0.762	0.171	0.767	0.320	0.757	0.464	0.757
VNIIFTRI 79	0.206	1.019	-0.002	1.066	-0.011	1.066	0.127	1.130	0.546	1.133	0.694	1.127	0.839	1.127
VNIIFTRI 89	0.511	1.089	0.304	1.133	0.294	1.133	0.432	1.193	0.851	1.196	1.000	1.190	1.145	1.190

Matrix of equivalence (Continued)

Lab, S/N <i>i</i>	Lab, S/N <i>j</i> \longrightarrow															
	NRC A138		NRC A140		PTB 229074		PTB 229075		VNIIIFTRI 79		VNIIIFTRI 89					
	D_i	U_i		D_{ij}	U_{ij}		D_{ij}	U_{ij}		D_{ij}	U_{ij}					
	$/ \text{mK}$			$/ \text{mK}$			$/ \text{mK}$			$/ \text{mK}$						
NIST A128	0.207	0.313		-0.792	1.173	-0.744	1.173	0.230	0.699	0.376	0.664	0.002	1.066	-0.304	1.133	
NIST A129	0.217	0.313		-0.782	1.173	-0.734	1.173	0.240	0.699	0.386	0.664	0.011	1.066	-0.294	1.133	
NMi-VSL 226246	0.079	0.488		-0.920	1.231	-0.872	1.231	0.102	0.793	0.248	0.762	-0.127	1.130	-0.432	1.193	
NPL 221481	-0.340	0.495		-1.339	1.234	-1.291	1.234	-0.317	0.798	-0.171	0.767	-0.546	1.133	-0.851	1.196	
NPL 221483	-0.489	0.480		-1.488	1.228	-1.440	1.228	-0.466	0.788	-0.320	0.757	-0.694	1.127	-1.000	1.190	
NPL 221485	-0.633	0.480		-1.633	1.228	-1.584	1.228	-0.611	0.788	-0.464	0.757	-0.839	1.127	-1.145	1.190	
NRC A138	0.999	1.131				0.048	1.599	1.022	1.292	1.168	1.273	0.793	1.522	0.488	1.570	
NRC A140	0.951	1.131				-0.048	1.599		0.974	1.292	1.120	1.273	0.745	1.522	0.440	1.570
PTB 229074	-0.023	0.625				-1.022	1.292	-0.974	1.292		0.146	0.856	-0.229	1.196	-0.534	1.256
PTB 229075	-0.169	0.585				-1.168	1.273	-1.120	1.273	-0.146	0.856		-0.375	1.175	-0.680	1.236
VNIIIFTRI 79	0.206	1.019				-0.793	1.522	-0.745	1.522	0.229	1.196	0.375	1.175		-0.306	1.491
VNIIIFTRI 89	0.511	1.089				-0.488	1.570	-0.440	1.570	0.534	1.256	0.680	1.236	0.306	1.491	

CCT-K1 : Nominal temperature, $T_{90} = 9.508$ K
Degrees of equivalence, D_i , and expanded uncertainties ($k = 2$), U_i , expressed in mK



Key comparison CCT-K1

MEASURAND : Temperature T_{90} / K

NOMINAL TEMPERATURE : $T_{90} = 10.803$ K

Laboratory individual measurements

Lab: Laboratory

S/N i : serial number of Thermometer i

T_i : temperature value for Thermometer i

u_i : standard uncertainty in the calibration of Thermometer i

u_{comp} : standard uncertainty in the comparison measurements

$$u_{\text{comp}} = 0.137 \text{ mK}$$

Lab, S/N i	T_i / mK	u_i / mK
NIST A128	10.803560	0.096
NIST A129	10.803616	0.096
NMi-VSL 226246	10.803584	0.227
NPL 221481	10.803070	0.206
NPL 221483	10.803011	0.200
NPL 221485	10.802842	0.200
NRC A138	10.804196	0.550
NRC A140	10.804185	0.550
PTB 229074	10.803248	0.288
PTB 229075	10.803221	0.265
VNIIFTRI 79	10.803595	0.513
VNIIFTRI 89	10.803927	0.544

Key comparison CCT-K1

MEASURAND : Temperature T_{90} / K

KEY COMPARISON REFERENCE VALUE: $T_R = 10.803390$ K


Matrix of equivalence

Lab, S/N i	Lab, S/N j \longrightarrow													
	NIST A128		NIST A129		NMi-VSL 226246		NPL 221481		NPL 221483		NPL 221485			
	D_i	U_i		D_{ij}	U_{ij}		D_{ij}	U_{ij}		D_{ij}	U_{ij}			
	/ mK	/ mK		/ mK	/ mK		/ mK	/ mK		/ mK	/ mK			
NIST A128	0.170	0.335		-0.056	0.473	-0.025	0.627	0.489	0.598	0.549	0.589	0.718	0.589	
NIST A129	0.226	0.335	0.056	0.473		0.032	0.627	0.546	0.598	0.605	0.589	0.774	0.589	
NMi-VSL 226246	0.194	0.531	0.025	0.627	-0.032	0.627		0.514	0.726	0.574	0.719	0.742	0.719	
NPL 221481	-0.320	0.495	-0.489	0.598	-0.546	0.598	-0.514	0.726		0.060	0.693	0.228	0.693	
NPL 221483	-0.379	0.484	-0.549	0.589	-0.605	0.589	-0.574	0.719	-0.060	0.693		0.169	0.685	
NPL 221485	-0.548	0.484	-0.718	0.589	-0.774	0.589	-0.742	0.719	-0.228	0.693	-0.169	0.685		
NRC A138	0.806	1.134	0.636	1.182	0.580	1.182	0.612	1.252	1.126	1.237	1.185	1.233	1.354	1.233
NRC A140	0.795	1.134	0.625	1.182	0.569	1.182	0.600	1.252	1.114	1.237	1.174	1.233	1.342	1.233
PTB 229074	-0.142	0.638	-0.312	0.720	-0.368	0.720	-0.337	0.830	0.177	0.807	0.237	0.801	0.405	0.801
PTB 229075	-0.169	0.597	-0.338	0.684	-0.395	0.684	-0.363	0.798	0.151	0.775	0.211	0.768	0.379	0.768
VNIIFTRI 79	0.205	1.062	0.036	1.113	-0.021	1.113	0.011	1.187	0.525	1.171	0.585	1.167	0.753	1.167
VNIIFTRI 89	0.537	1.122	0.367	1.171	0.311	1.171	0.342	1.241	0.856	1.226	0.916	1.222	1.085	1.222

Matrix of equivalence (Continued)

Lab, S/N <i>i</i>	Lab, S/N <i>j</i> \longrightarrow														
	NRC A138		NRC A140		PTB 229074		PTB 229075		VNIIIFTRI 79		VNIIIFTRI 89				
	D_i	U_i		D_{ij}	U_{ij}		D_{ij}	U_{ij}		D_{ij}	U_{ij}				
	$/ \text{mK}$			$/ \text{mK}$			$/ \text{mK}$			$/ \text{mK}$					
NIST A128	0.170	0.335		-0.636	1.182	-0.625	1.182	0.312	0.720	0.338	0.684	-0.036	1.113	-0.367	1.171
NIST A129	0.226	0.335		-0.580	1.182	-0.569	1.182	0.368	0.720	0.395	0.684	0.021	1.113	-0.311	1.171
NMi-VSL 226246	0.194	0.531		-0.612	1.252	-0.600	1.252	0.337	0.830	0.363	0.798	-0.011	1.187	-0.342	1.241
NPL 221481	-0.320	0.495		-1.126	1.237	-1.114	1.237	-0.177	0.807	-0.151	0.775	-0.525	1.171	-0.856	1.226
NPL 221483	-0.379	0.484		-1.185	1.233	-1.174	1.233	-0.237	0.801	-0.211	0.768	-0.585	1.167	-0.916	1.222
NPL 221485	-0.548	0.484		-1.354	1.233	-1.342	1.233	-0.405	0.801	-0.379	0.768	-0.753	1.167	-1.085	1.222
NRC A138	0.806	1.134				0.012	1.603	0.948	1.301	0.975	1.281	0.601	1.553	0.269	1.595
NRC A140	0.795	1.134						0.937	1.301	0.963	1.281	0.589	1.553	0.258	1.595
PTB 229074	-0.142	0.638								0.026	0.873	-0.348	1.239	-0.679	1.291
PTB 229075	-0.169	0.597										-0.374	1.218	-0.705	1.271
VNIIIFTRI 79	0.205	1.062											-0.331	1.545	
VNIIIFTRI 89	0.537	1.122													

CCT-K1 : Nominal temperature, $T_{90} = 10.803$ K
Degrees of equivalence, D_i , and expanded uncertainties ($k = 2$), U_i , expressed in mK



Key comparison CCT-K1

MEASURAND : Temperature T_{90} / K

NOMINAL TEMPERATURE : $T_{90} = 12.297$ K

Laboratory individual measurements

Lab: Laboratory

S/N i : serial number of Thermometer i

T_i : temperature value for Thermometer i

u_i : standard uncertainty in the calibration of Thermometer i

u_{comp} : standard uncertainty in the comparison measurements

$$u_{\text{comp}} = 0.145 \text{ mK}$$

Lab, S/N i	T_i / mK	u_i / mK
NIST A128	12.297349	0.106
NIST A129	12.297413	0.106
NMi-VSL 226246	12.297900	0.267
NPL 221481	12.297090	0.199
NPL 221483	12.297224	0.197
NPL 221485	12.296905	0.197
NRC A138	12.297793	0.550
NRC A140	12.297778	0.550
PTB 229074	12.297149	0.292
PTB 229075	12.297099	0.269
VNIIFTRI 79	12.297633	0.536
VNIIFTRI 89	12.297547	0.562

Key comparison CCT-K1

MEASURAND : Temperature T_{90} / K

KEY COMPARISON REFERENCE VALUE: $T_R = 12.297309$ K


Matrix of equivalence

Lab, S/N i	Lab, S/N j \longrightarrow													
	NIST A128		NIST A129		NMi-VSL 226246		NPL 221481		NPL 221483		NPL 221485			
	D_i	U_i		D_{ij}	U_{ij}		D_{ij}	U_{ij}		D_{ij}	U_{ij}			
NIST A128	0.040	0.359		-0.064	0.508	-0.550	0.706	0.259	0.609	0.125	0.607	0.445	0.607	
NIST A129	0.104	0.359	0.064	0.508		-0.486	0.706	0.323	0.609	0.189	0.607	0.509	0.607	
NMi-VSL 226246	0.591	0.608	0.550	0.706	0.486	0.706		0.809	0.782	0.675	0.781	0.995	0.781	
NPL 221481	-0.219	0.492	-0.259	0.609	-0.323	0.609	-0.809	0.782		-0.134	0.693	0.186	0.693	
NPL 221483	-0.085	0.489	-0.125	0.607	-0.189	0.607	-0.675	0.781	0.134	0.693		0.320	0.691	
NPL 221485	-0.404	0.489	-0.445	0.607	-0.509	0.607	-0.995	0.781	-0.186	0.693	-0.320	0.691		
NRC A138	0.484	1.138	0.444	1.193	0.380	1.193	-0.106	1.290	0.703	1.239	0.569	1.238	0.889	1.238
NRC A140	0.469	1.138	0.429	1.193	0.365	1.193	-0.122	1.290	0.688	1.239	0.553	1.238	0.873	1.238
PTB 229074	-0.160	0.653	-0.200	0.745	-0.264	0.745	-0.751	0.892	0.058	0.817	-0.076	0.816	0.244	0.816
PTB 229075	-0.210	0.611	-0.250	0.708	-0.314	0.708	-0.801	0.862	0.009	0.784	-0.126	0.782	0.194	0.782
VNIIFTRI 79	0.324	1.111	0.284	1.168	0.220	1.168	-0.267	1.267	0.543	1.215	0.408	1.214	0.728	1.214
VNIIFTRI 89	0.238	1.161	0.197	1.215	0.134	1.215	-0.353	1.311	0.456	1.261	0.322	1.259	0.642	1.259

Matrix of equivalence (Continued)

Lab, S/N <i>i</i>	Lab, S/N <i>j</i> \longrightarrow														
	NRC A138		NRC A140		PTB 229074		PTB 229075		VNIIIFTRI 79		VNIIIFTRI 89				
	D_i	U_i		D_{ij}	U_{ij}		D_{ij}	U_{ij}		D_{ij}	U_{ij}				
	$/ \text{mK}$			$/ \text{mK}$			$/ \text{mK}$			$/ \text{mK}$					
NIST A128	0.040	0.359		-0.444	1.193	-0.429	1.193	0.200	0.745	0.250	0.708	-0.284	1.168	-0.197	1.215
NIST A129	0.104	0.359		-0.380	1.193	-0.365	1.193	0.264	0.745	0.314	0.708	-0.220	1.168	-0.134	1.215
NMi-VSL 226246	0.591	0.608		0.106	1.290	0.122	1.290	0.751	0.892	0.801	0.862	0.267	1.267	0.353	1.311
NPL 221481	-0.219	0.492		-0.703	1.239	-0.688	1.239	-0.058	0.817	-0.009	0.784	-0.543	1.215	-0.456	1.261
NPL 221483	-0.085	0.489		-0.569	1.238	-0.553	1.238	0.076	0.816	0.126	0.782	-0.408	1.214	-0.322	1.259
NPL 221485	-0.404	0.489		-0.889	1.238	-0.873	1.238	-0.244	0.816	-0.194	0.782	-0.728	1.214	-0.642	1.259
NRC A138	0.484	1.138			0.015	1.609	0.645	1.312	0.695	1.291	0.160	1.590	0.247	1.625	
NRC A140	0.469	1.138		-0.015	1.609		0.629	1.312	0.679	1.291	0.145	1.590	0.231	1.625	
PTB 229074	-0.160	0.653		-0.645	1.312	-0.629	1.312		0.050	0.894	-0.484	1.289	-0.398	1.332	
PTB 229075	-0.210	0.611		-0.695	1.291	-0.679	1.291	-0.050	0.894		-0.534	1.268	-0.448	1.312	
VNIIIFTRI 79	0.324	1.111		-0.160	1.590	-0.145	1.590	0.484	1.289	0.534	1.268		0.086	1.607	
VNIIIFTRI 89	0.238	1.161		-0.247	1.625	-0.231	1.625	0.398	1.332	0.448	1.312	-0.086	1.607		

CCT-K1 : Nominal temperature, $T_{90} = 12.297$ K
Degrees of equivalence, D_i , and expanded uncertainties ($k = 2$), U_i , expressed in mK



Key comparison CCT-K1

MEASURAND : Temperature T_{90} / K

NOMINAL TEMPERATURE : $T_{90} = 13.798$ K

Laboratory individual measurements

Lab: Laboratory

S/N i : serial number of Thermometer i

T_i : temperature value for Thermometer i

u_i : standard uncertainty in the calibration of Thermometer i

u_{comp} : standard uncertainty in the comparison measurements

$$u_{\text{comp}} = 0.153 \text{ mK}$$

Lab, S/N i	T_i / mK	u_i / mK
NIST A128	13.798156	0.111
NIST A129	13.798122	0.111
NMi-VSL 226246	13.798881	0.308
NPL 221481	13.798130	0.191
NPL 221483	13.798170	0.182
NPL 221485	13.797754	0.182
NRC A138	13.798459	0.550
NRC A140	13.798305	0.550
PTB 229074	13.798317	0.297
PTB 229075	13.798209	0.273
VNIIFTRI 79	13.798638	0.560
VNIIFTRI 89	13.798620	0.580

Key comparison CCT-K1

MEASURAND : Temperature T_{90} / K

KEY COMPARISON REFERENCE VALUE: $T_R = 13.798183$ K


Matrix of equivalence

Lab, S/N i	Lab, S/N j \longrightarrow											
	NIST A128		NIST A129		NMi-VSL 226246		NPL 221481		NPL 221483		NPL 221485	
	D_i	U_i		D_{ij}	U_{ij}		D_{ij}	U_{ij}		D_{ij}	U_{ij}	
	/ mK	/ mK		/ mK	/ mK		/ mK	/ mK		/ mK	/ mK	
NIST A128	-0.027	0.378			0.033	0.534	-0.725	0.785	0.026	0.618	-0.014	0.607
NIST A129	-0.061	0.378		-0.033	0.534		-0.758	0.785	-0.008	0.618	-0.048	0.607
NMi-VSL 226246	0.698	0.688		0.725	0.785	0.758	0.785		0.751	0.844	0.711	0.836
NPL 221481	-0.053	0.489		-0.026	0.618	0.008	0.618	-0.751	0.844		-0.040	0.682
NPL 221483	-0.013	0.476		0.014	0.607	0.048	0.607	-0.711	0.836	0.040	0.682	0.376
NPL 221485	-0.429	0.476		-0.402	0.607	-0.368	0.607	-1.127	0.836	-0.376	0.682	-0.416
NRC A138	0.276	1.142		0.303	1.203	0.337	1.203	-0.422	1.333	0.329	1.242	0.289
NRC A140	0.122	1.142		0.149	1.203	0.183	1.203	-0.576	1.333	0.175	1.242	0.135
PTB 229074	0.134	0.668		0.161	0.768	0.194	0.768	-0.564	0.959	0.187	0.828	0.147
PTB 229075	0.026	0.625		0.054	0.730	0.087	0.730	-0.671	0.929	0.079	0.794	0.039
VNIIFTRI 79	0.455	1.161		0.483	1.221	0.516	1.221	-0.242	1.349	0.508	1.260	0.468
VNIIFTRI 89	0.437	1.200		0.464	1.258	0.497	1.258	-0.261	1.383	0.490	1.296	0.450

Matrix of equivalence (Continued)

Lab, S/N <i>i</i>	Lab, S/N <i>j</i> \longrightarrow											
	NRC A138		NRC A140		PTB 229074		PTB 229075		VNIIIFTRI 79		VNIIIFTRI 89	
	D_i	U_i		D_{ij}	U_{ij}		D_{ij}	U_{ij}		D_{ij}	U_{ij}	
	/ mK	/ mK		/ mK	/ mK		/ mK	/ mK		/ mK	/ mK	
NIST A128	-0.027	0.378		-0.303	1.203	-0.149	1.203	-0.161	0.768	-0.054	0.730	-0.483
NIST A129	-0.061	0.378		-0.337	1.203	-0.183	1.203	-0.194	0.768	-0.087	0.730	-0.516
NMi-VSL 226246	0.698	0.688		0.422	1.333	0.576	1.333	0.564	0.959	0.671	0.929	0.242
NPL 221481	-0.053	0.489		-0.329	1.242	-0.175	1.242	-0.187	0.828	-0.079	0.794	-0.508
NPL 221483	-0.013	0.476		-0.289	1.237	-0.135	1.237	-0.147	0.820	-0.039	0.785	-0.468
NPL 221485	-0.429	0.476		-0.705	1.237	-0.551	1.237	-0.563	0.820	-0.455	0.785	-0.884
NRC A138	0.276	1.142				0.154	1.615	0.143	1.323	0.250	1.302	-0.179
NRC A140	0.122	1.142				-0.154	1.615	-0.012	1.323	0.096	1.302	-0.333
PTB 229074	0.134	0.668				-0.143	1.323	0.012	1.323	0.107	0.915	-0.322
PTB 229075	0.026	0.625				-0.250	1.302	-0.096	1.302	-0.107	0.915	-0.429
VNIIIFTRI 79	0.455	1.161				0.179	1.628	0.333	1.628	0.429	1.319	0.410
VNIIIFTRI 89	0.437	1.200				0.160	1.656	0.314	1.656	0.303	1.373	0.019

CCT-K1 : Nominal temperature, $T_{90} = 13.798$ K
Degrees of equivalence, D_i , and expanded uncertainties ($k = 2$), U_i , expressed in mK



Key comparison CCT-K1

MEASURAND : Temperature T_{90} / K

NOMINAL TEMPERATURE : $T_{90} = 15.500$ K

Laboratory individual measurements

Lab: Laboratory

S/N i : serial number of Thermometer i

T_i : temperature value for Thermometer i

u_i : standard uncertainty in the calibration of Thermometer i

u_{comp} : standard uncertainty in the comparison measurements

$$u_{\text{comp}} = 0.153 \text{ mK}$$

Lab, S/N i	T_i / mK	u_i / mK
NIST A128	15.499479	0.119
NIST A129	15.499474	0.119
NMi-VSL 226246	15.500221	0.354
NPL 221481	15.499523	0.194
NPL 221483	15.499571	0.185
NPL 221485	15.499138	0.185
NRC A138	15.499689	0.550
NRC A140	15.499662	0.550
PTB 229074	15.499940	0.302
PTB 229075	15.499849	0.277
VNIIFTRI 79	15.499890	0.560
VNIIFTRI 89	15.499817	0.585

Key comparison CCT-K1

MEASURAND : Temperature T_{90} / K

KEY COMPARISON REFERENCE VALUE: $T_R = 15.499566$ K


Matrix of equivalence

Lab, S/N i	Lab, S/N j \longrightarrow											
	NIST A128		NIST A129		NMi-VSL 226246		NPL 221481		NPL 221483		NPL 221485	
	D_i	U_i		D_{ij}	U_{ij}		D_{ij}	U_{ij}		D_{ij}	U_{ij}	
	/ mK	/ mK		/ mK	/ mK		/ mK	/ mK		/ mK	/ mK	
NIST A128	-0.087	0.388			0.006	0.548	-0.741	0.863	-0.043	0.628	-0.091	0.617
NIST A129	-0.092	0.388		-0.006	0.548		-0.747	0.863	-0.049	0.628	-0.097	0.617
NMi-VSL 226246	0.655	0.771		0.741	0.863	0.747	0.863		0.698	0.916	0.650	0.908
NPL 221481	-0.043	0.494		0.043	0.628	0.049	0.628	-0.698	0.916		-0.048	0.689
NPL 221483	0.005	0.480		0.091	0.617	0.097	0.617	-0.650	0.908	0.048	0.689	0.385
NPL 221485	-0.428	0.480		-0.342	0.617	-0.336	0.617	-1.083	0.908	-0.385	0.689	-0.433
NRC A138	0.123	1.142		0.209	1.206	0.215	1.206	-0.532	1.378	0.166	1.244	0.118
NRC A140	0.096	1.142		0.182	1.206	0.188	1.206	-0.559	1.378	0.139	1.244	0.091
PTB 229074	0.374	0.677		0.460	0.780	0.466	0.780	-0.281	1.026	0.417	0.839	0.369
PTB 229075	0.283	0.633		0.370	0.742	0.376	0.742	-0.371	0.997	0.326	0.803	0.279
VNIIFTRI 79	0.324	1.161		0.411	1.224	0.416	1.224	-0.331	1.394	0.367	1.262	0.319
VNIIFTRI 89	0.251	1.209		0.338	1.269	0.344	1.269	-0.403	1.434	0.294	1.306	0.247

Matrix of equivalence (Continued)

Lab, S/N <i>i</i>	Lab, S/N <i>j</i> \longrightarrow														
	NRC A138		NRC A140		PTB 229074		PTB 229075		VNIIIFTRI 79		VNIIIFTRI 89				
	D_i	U_i		D_{ij}	U_{ij}		D_{ij}	U_{ij}		D_{ij}	U_{ij}				
	$/ \text{mK}$			$/ \text{mK}$			$/ \text{mK}$			$/ \text{mK}$					
NIST A128	-0.087	0.388		-0.209	1.206	-0.182	1.206	-0.460	0.780	-0.370	0.742	-0.411	1.224	-0.338	1.269
NIST A129	-0.092	0.388		-0.215	1.206	-0.188	1.206	-0.466	0.780	-0.376	0.742	-0.416	1.224	-0.344	1.269
NMi-VSL 226246	0.655	0.771		0.532	1.378	0.559	1.378	0.281	1.026	0.371	0.997	0.331	1.394	0.403	1.434
NPL 221481	-0.043	0.494		-0.166	1.244	-0.139	1.244	-0.417	0.839	-0.326	0.803	-0.367	1.262	-0.294	1.306
NPL 221483	0.005	0.480		-0.118	1.239	-0.091	1.239	-0.369	0.831	-0.279	0.794	-0.319	1.256	-0.247	1.301
NPL 221485	-0.428	0.480		-0.551	1.239	-0.524	1.239	-0.802	0.831	-0.712	0.794	-0.753	1.256	-0.680	1.301
NRC A138	0.123	1.142				0.027	1.615	-0.251	1.328	-0.160	1.305	-0.201	1.628	-0.128	1.663
NRC A140	0.096	1.142		-0.027	1.615			-0.278	1.328	-0.188	1.305	-0.228	1.628	-0.156	1.663
PTB 229074	0.374	0.677		0.251	1.328	0.278	1.328			0.090	0.927	0.050	1.344	0.122	1.386
PTB 229075	0.283	0.633		0.160	1.305	0.188	1.305	-0.090	0.927			-0.041	1.322	0.032	1.364
VNIIIFTRI 79	0.324	1.161		0.201	1.628	0.228	1.628	-0.050	1.344	0.041	1.322			0.073	1.676
VNIIIFTRI 89	0.251	1.209		0.128	1.663	0.156	1.663	-0.122	1.386	-0.032	1.364	-0.073	1.676		

CCT-K1 : Nominal temperature, $T_{90} = 15.500$ K
Degrees of equivalence, D_i , and expanded uncertainties ($k = 2$), U_i , expressed in mK



Key comparison CCT-K1

MEASURAND : Temperature T_{90} / K

NOMINAL TEMPERATURE : $T_{90} = 16.999$ K

Laboratory individual measurements

Lab: Laboratory

S/N i : serial number of Thermometer i

T_i : temperature value for Thermometer i

u_i : standard uncertainty in the calibration of Thermometer i

u_{comp} : standard uncertainty in the comparison measurements

$$u_{\text{comp}} = 0.153 \text{ mK}$$

Lab, S/N i	T_i / mK	u_i / mK
NIST A128	16.999286	0.124
NIST A129	16.999271	0.124
NMi-VSL 226246	16.999833	0.394
NPL 221481	16.999290	0.197
NPL 221483	16.999280	0.188
NPL 221485	16.998847	0.188
NRC A138	16.999430	0.550
NRC A140	16.999259	0.550
PTB 229074	16.999859	0.307
PTB 229075	16.999701	0.281
VNIIFTRI 79	16.999602	0.560
VNIIFTRI 89	16.999783	0.589

Key comparison CCT-K1

MEASURAND : Temperature T_{90} / K

KEY COMPARISON REFERENCE VALUE: $T_R = 16.999335$ K


Matrix of equivalence

Lab, S/N i	Lab, S/N j \longrightarrow																
	NIST A128		NIST A129		NMi-VSL 226246		NPL 221481		NPL 221483		NPL 221485						
	D_i	U_i		D_{ij}	U_{ij}		D_{ij}	U_{ij}		D_{ij}	U_{ij}		D_{ij}	U_{ij}		D_{ij}	U_{ij}
NIST A128	-0.049	0.394					0.015	0.557	-0.546	0.933	-0.003	0.635	0.006	0.624	0.439	0.624	
NIST A129	-0.064	0.394		-0.015	0.557				-0.562	0.933	-0.019	0.635	-0.009	0.624	0.424	0.624	
NMi-VSL 226246	0.498	0.845		0.546	0.933	0.562	0.933				0.543	0.982	0.553	0.975	0.985	0.975	
NPL 221481	-0.045	0.499		0.003	0.635	0.019	0.635	-0.543	0.982				0.009	0.695	0.442	0.695	
NPL 221483	-0.055	0.485		-0.006	0.624	0.009	0.624	-0.553	0.975	-0.009	0.695			0.433	0.685		
NPL 221485	-0.488	0.485		-0.439	0.624	-0.424	0.624	-0.985	0.975	-0.442	0.695	-0.433	0.685				
NRC A138	0.095	1.142		0.144	1.208	0.160	1.208	-0.402	1.421	0.141	1.246	0.150	1.240	0.583	1.240		
NRC A140	-0.076	1.142		-0.027	1.208	-0.012	1.208	-0.574	1.421	-0.031	1.246	-0.021	1.240	0.412	1.240		
PTB 229074	0.524	0.686		0.573	0.791	0.588	0.791	0.026	1.089	0.569	0.848	0.579	0.840	1.011	0.840		
PTB 229075	0.366	0.639		0.414	0.751	0.430	0.751	-0.132	1.060	0.411	0.811	0.420	0.802	0.853	0.802		
VNIIFTRI 79	0.267	1.161		0.316	1.226	0.331	1.226	-0.231	1.436	0.312	1.264	0.322	1.258	0.755	1.258		
VNIIFTRI 89	0.448	1.217		0.497	1.279	0.512	1.279	-0.050	1.482	0.493	1.315	0.503	1.310	0.935	1.310		

Matrix of equivalence (Continued)

Lab, S/N <i>i</i>	Lab, S/N <i>j</i> \longrightarrow														
	NRC A138		NRC A140		PTB 229074		PTB 229075		VNIIIFTRI 79		VNIIIFTRI 89				
	D_i	U_i		D_{ij}	U_{ij}		D_{ij}	U_{ij}		D_{ij}	U_{ij}				
	$/ \text{mK}$			$/ \text{mK}$			$/ \text{mK}$			$/ \text{mK}$					
NIST A128	-0.049	0.394		-0.144	1.208	0.027	1.208	-0.573	0.791	-0.414	0.751	-0.316	1.226	-0.497	1.279
NIST A129	-0.064	0.394		-0.160	1.208	0.012	1.208	-0.588	0.791	-0.430	0.751	-0.331	1.226	-0.512	1.279
NMi-VSL 226246	0.498	0.845		0.402	1.421	0.574	1.421	-0.026	1.089	0.132	1.060	0.231	1.436	0.050	1.482
NPL 221481	-0.045	0.499		-0.141	1.246	0.031	1.246	-0.569	0.848	-0.411	0.811	-0.312	1.264	-0.493	1.315
NPL 221483	-0.055	0.485		-0.150	1.240	0.021	1.240	-0.579	0.840	-0.420	0.802	-0.322	1.258	-0.503	1.310
NPL 221485	-0.488	0.485		-0.583	1.240	-0.412	1.240	-1.011	0.840	-0.853	0.802	-0.755	1.258	-0.935	1.310
NRC A138	0.095	1.142				0.171	1.615	-0.428	1.332	-0.270	1.309	-0.171	1.628	-0.352	1.669
NRC A140	-0.076	1.142		-0.171	1.615			-0.600	1.332	-0.441	1.309	-0.343	1.628	-0.524	1.669
PTB 229074	0.524	0.686		0.428	1.332	0.600	1.332			0.158	0.938	0.257	1.348	0.076	1.397
PTB 229075	0.366	0.639		0.270	1.309	0.441	1.309	-0.158	0.938			0.099	1.325	-0.082	1.375
VNIIIFTRI 79	0.267	1.161		0.171	1.628	0.343	1.628	-0.257	1.348	-0.099	1.325			-0.181	1.682
VNIIIFTRI 89	0.448	1.217		0.352	1.669	0.524	1.669	-0.076	1.397	0.082	1.375	0.181	1.682		

CCT-K1 : Nominal temperature, $T_{90} = 16.999$ K
Degrees of equivalence, D_i , and expanded uncertainties ($k = 2$), U_i , expressed in mK



Key comparison CCT-K1

MEASURAND : Temperature T_{90} / K

NOMINAL TEMPERATURE : $T_{90} = 18.597$ K

Laboratory individual measurements

Lab: Laboratory

S/N i : serial number of Thermometer i

T_i : temperature value for Thermometer i

u_i : standard uncertainty in the calibration of Thermometer i

u_{comp} : standard uncertainty in the comparison measurements

$$u_{\text{comp}} = 0.153 \text{ mK}$$

Lab, S/N i	T_i / mK	u_i / mK
NIST A128	18.597374	0.129
NIST A129	18.597342	0.129
NMi-VSL 226246	18.597934	0.437
NPL 221481	18.597330	0.200
NPL 221483	18.597229	0.191
NPL 221485	18.596804	0.191
NRC A138	18.597402	0.550
NRC A140	18.597237	0.550
PTB 229074	18.597946	0.312
PTB 229075	18.597872	0.285
VNIIFTRI 79	18.597658	0.560
VNIIFTRI 89	18.597830	0.593

Key comparison CCT-K1

MEASURAND : Temperature T_{90} / K

KEY COMPARISON REFERENCE VALUE: $T_R = 18.597377$ K


Matrix of equivalence

Lab, S/N i	Lab, S/N j \longrightarrow												
	NIST A128		NIST A129		NMi-VSL 226246		NPL 221481		NPL 221483		NPL 221485		
	D_i	U_i		D_{ij}	U_{ij}		D_{ij}	U_{ij}		D_{ij}	U_{ij}		
	/ mK	/ mK		/ mK	/ mK		/ mK	/ mK		/ mK	/ mK		
NIST A128	-0.003	0.401											
NIST A129	-0.035	0.401		0.032	0.567	-0.560	1.009	0.044	0.643	0.145	0.632	0.570	0.632
NMi-VSL 226246	0.557	0.926		-0.032	0.567	-0.593	1.009	0.012	0.643	0.113	0.632	0.538	0.632
NPL 221481	-0.047	0.503		0.560	1.009	0.593	1.009	0.604	1.054	0.706	1.048	1.131	1.048
NPL 221483	-0.148	0.489		-0.044	0.643	-0.012	0.643	-0.604	1.054	0.101	0.702	0.526	0.702
NPL 221485	-0.573	0.489		-0.145	0.632	-0.113	0.632	-0.706	1.048	-0.101	0.702	0.425	0.692
NRC A138	0.025	1.142		-0.570	0.632	-0.538	0.632	-1.131	1.048	-0.526	0.702	-0.425	0.692
NRC A140	-0.140	1.142		0.028	1.210	0.060	1.210	-0.533	1.470	0.072	1.248	0.173	1.242
PTB 229074	0.569	0.695		-0.137	1.210	-0.105	1.210	-0.697	1.470	-0.093	1.248	0.008	1.242
PTB 229075	0.495	0.647		0.572	0.802	0.605	0.802	0.012	1.158	0.616	0.858	0.718	0.850
VNIIFTRI 79	0.281	1.161		0.498	0.761	0.530	0.761	-0.063	1.130	0.542	0.819	0.643	0.811
VNIIFTRI 89	0.453	1.226		0.283	1.228	0.316	1.228	-0.277	1.485	0.327	1.265	0.429	1.260
				0.456	1.290	0.489	1.290	-0.104	1.536	0.500	1.325	0.602	1.320
												1.026	1.320

Matrix of equivalence (Continued)

Lab, S/N <i>i</i>	Lab, S/N <i>j</i> \longrightarrow															
	NRC A138		NRC A140		PTB 229074		PTB 229075		VNIIIFTRI 79		VNIIIFTRI 89					
	D_i	U_i		D_{ij}	U_{ij}		D_{ij}	U_{ij}		D_{ij}	U_{ij}					
	$/ \text{mK}$			$/ \text{mK}$			$/ \text{mK}$			$/ \text{mK}$						
NIST A128	-0.003	0.401		-0.028	1.210	0.137	1.210	-0.572	0.802	-0.498	0.761	-0.283	1.228	-0.456	1.290	
NIST A129	-0.035	0.401		-0.060	1.210	0.105	1.210	-0.605	0.802	-0.530	0.761	-0.316	1.228	-0.489	1.290	
NMi-VSL 226246	0.557	0.926		0.533	1.470	0.697	1.470	-0.012	1.158	0.063	1.130	0.277	1.485	0.104	1.536	
NPL 221481	-0.047	0.503		-0.072	1.248	0.093	1.248	-0.616	0.858	-0.542	0.819	-0.327	1.265	-0.500	1.325	
NPL 221483	-0.148	0.489		-0.173	1.242	-0.008	1.242	-0.718	0.850	-0.643	0.811	-0.429	1.260	-0.602	1.320	
NPL 221485	-0.573	0.489		-0.598	1.242	-0.433	1.242	-1.142	0.850	-1.068	0.811	-0.854	1.260	-1.026	1.320	
NRC A138	0.025	1.142				0.165	1.615	-0.544	1.336	-0.470	1.312	-0.256	1.628	-0.428	1.675	
NRC A140	-0.140	1.142				-0.165	1.615		-0.709	1.336	-0.635	1.312	-0.420	1.628	-0.593	1.675
PTB 229074	0.569	0.695				0.544	1.336	0.709		0.075	0.949	0.289	1.353	0.116	1.409	
PTB 229075	0.495	0.647				0.470	1.312	0.635	1.312	-0.075	0.949		0.214	1.329	0.041	1.386
VNIIIFTRI 79	0.281	1.161				0.256	1.628	0.420	1.628	-0.289	1.353	-0.214	1.329		-0.173	1.688
VNIIIFTRI 89	0.453	1.226				0.428	1.675	0.593	1.675	-0.116	1.409	-0.041	1.386	0.173	1.688	

CCT-K1 : Nominal temperature, $T_{90} = 18.597$ K
Degrees of equivalence, D_i , and expanded uncertainties ($k = 2$), U_i , expressed in mK



Key comparison CCT-K1

MEASURAND : Temperature T_{90} / K

NOMINAL TEMPERATURE : $T_{90} = 20.299$ K

Laboratory individual measurements

Lab: Laboratory

S/N i : serial number of Thermometer i

T_i : temperature value for Thermometer i

u_i : standard uncertainty in the calibration of Thermometer i

u_{comp} : standard uncertainty in the comparison measurements

$$u_{\text{comp}} = 0.153 \text{ mK}$$

Lab, S/N i	T_i / mK	u_i / mK
NIST A128	20.299107	0.132
NIST A129	20.299108	0.132
NMi-VSL 226246	20.299225	0.475
NPL 221481	20.298690	0.203
NPL 221483	20.298576	0.194
NPL 221485	20.298352	0.194
NRC A138	20.298800	0.550
NRC A140	20.298750	0.550
PTB 229074	20.299355	0.317
PTB 229075	20.299131	0.289
VNIIFTRI 79	20.298975	0.560
VNIIFTRI 89	20.299144	0.598

Key comparison CCT-K1

MEASURAND : Temperature T_{90} / K

KEY COMPARISON REFERENCE VALUE: $T_R = 20.298899$ K


Matrix of equivalence

Lab, S/N <i>i</i>	Lab, S/N <i>j</i> \longrightarrow													
	D _{<i>i</i>} U _{<i>i</i>} / mK		NIST A128		NIST A129		NMi-VSL 226246		NPL 221481		NPL 221483		NPL 221485	
	D _{<i>ij</i>}	U _{<i>ij</i>}	D _{<i>ij</i>}	U _{<i>ij</i>}	D _{<i>ij</i>}	U _{<i>ij</i>}	D _{<i>ij</i>}	U _{<i>ij</i>}	D _{<i>ij</i>}	U _{<i>ij</i>}	D _{<i>ij</i>}	U _{<i>ij</i>}	D _{<i>ij</i>}	U _{<i>ij</i>}
NIST A128	0.208	0.404			-0.001	0.572	-0.119	1.077	0.417	0.650	0.530	0.638	0.754	0.638
NIST A129	0.209	0.404	0.001	0.572			-0.118	1.077	0.418	0.650	0.531	0.638	0.755	0.638
NMi-VSL 226246	0.326	0.998	0.119	1.077	0.118	1.077			0.535	1.120	0.649	1.114	0.873	1.114
NPL 221481	-0.209	0.508	-0.417	0.650	-0.418	0.650	-0.535	1.120			0.114	0.709	0.338	0.709
NPL 221483	-0.323	0.494	-0.530	0.638	-0.531	0.638	-0.649	1.114	-0.114	0.709			0.224	0.699
NPL 221485	-0.547	0.494	-0.754	0.638	-0.755	0.638	-0.873	1.114	-0.338	0.709	-0.224	0.699		
NRC A138	-0.099	1.142	-0.307	1.211	-0.308	1.211	-0.425	1.516	0.110	1.250	0.224	1.244	0.448	1.244
NRC A140	-0.149	1.142	-0.356	1.211	-0.357	1.211	-0.475	1.516	0.060	1.250	0.174	1.244	0.398	1.244
PTB 229074	0.456	0.704	0.249	0.812	0.248	0.812	0.130	1.221	0.665	0.868	0.779	0.860	1.003	0.860
PTB 229075	0.232	0.654	0.024	0.769	0.023	0.769	-0.094	1.193	0.441	0.829	0.555	0.820	0.779	0.820
VNIIFTRI 79	0.076	1.161	-0.132	1.229	-0.133	1.229	-0.251	1.531	0.284	1.268	0.398	1.262	0.622	1.262
VNIIFTRI 89	0.245	1.235	0.037	1.299	0.037	1.299	-0.081	1.588	0.454	1.336	0.568	1.330	0.792	1.330

Matrix of equivalence (Continued)

Lab, S/N <i>i</i>	Lab, S/N <i>j</i> \longrightarrow															
	NRC A138		NRC A140		PTB 229074		PTB 229075		VNIIIFTRI 79		VNIIIFTRI 89					
	D_i	U_i		D_{ij}	U_{ij}		D_{ij}	U_{ij}		D_{ij}	U_{ij}					
	$/ \text{mK}$			$/ \text{mK}$			$/ \text{mK}$			$/ \text{mK}$						
NIST A128	0.208	0.404		0.307	1.211	0.356	1.211	-0.249	0.812	-0.024	0.769	0.132	1.229	-0.037	1.299	
NIST A129	0.209	0.404		0.308	1.211	0.357	1.211	-0.248	0.812	-0.023	0.769	0.133	1.229	-0.037	1.299	
NMi-VSL 226246	0.326	0.998		0.425	1.516	0.475	1.516	-0.130	1.221	0.094	1.193	0.251	1.531	0.081	1.588	
NPL 221481	-0.209	0.508		-0.110	1.250	-0.060	1.250	-0.665	0.868	-0.441	0.829	-0.284	1.268	-0.454	1.336	
NPL 221483	-0.323	0.494		-0.224	1.244	-0.174	1.244	-0.779	0.860	-0.555	0.820	-0.398	1.262	-0.568	1.330	
NPL 221485	-0.547	0.494		-0.448	1.244	-0.398	1.244	-1.003	0.860	-0.779	0.820	-0.622	1.262	-0.792	1.330	
NRC A138	-0.099	1.142				0.050	1.615	-0.555	1.341	-0.331	1.316	-0.174	1.628	-0.344	1.682	
NRC A140	-0.149	1.142				-0.050	1.615		-0.605	1.341	-0.381	1.316	-0.224	1.628	-0.394	1.682
PTB 229074	0.456	0.704				0.555	1.341	0.605		0.224	0.961	0.381	1.358	0.211	1.421	
PTB 229075	0.232	0.654				0.331	1.316	0.381	1.316	-0.224	0.961	0.156	1.333	-0.013	1.398	
VNIIIFTRI 79	0.076	1.161				0.174	1.628	0.224	1.628	-0.381	1.358	-0.156	1.333	-0.170	1.695	
VNIIIFTRI 89	0.245	1.235				0.344	1.682	0.394	1.682	-0.211	1.421	0.013	1.398	0.170	1.695	

CCT-K1 : Nominal temperature, $T_{90} = 20.299$ K
Degrees of equivalence, D_i , and expanded uncertainties ($k = 2$), U_i , expressed in mK



Key comparison CCT-K1

MEASURAND : Temperature T_{90} / K

NOMINAL TEMPERATURE : $T_{90} = 21.575$ K

Laboratory individual measurements

Lab: Laboratory

S/N i : serial number of Thermometer i

T_i : temperature value for Thermometer i

u_i : standard uncertainty in the calibration of Thermometer i

u_{comp} : standard uncertainty in the comparison measurements

$$u_{\text{comp}} = 0.153 \text{ mK}$$

Lab, S/N i	T_i / mK	u_i / mK
NIST A128	21.575775	0.142
NIST A129	21.575754	0.142
NMi-VSL 226246	21.575850	0.514
NPL 221481	21.575120	0.207
NPL 221483	21.574976	0.198
NPL 221485	21.574975	0.198
NRC A138	21.575266	0.550
NRC A140	21.575114	0.550
PTB 229074	21.575934	0.321
PTB 229075	21.575693	0.292
VNIIFTRI 79	21.575227	0.560
VNIIFTRI 89	21.575344	0.602

Key comparison CCT-K1

MEASURAND : Temperature T_{90} / K

KEY COMPARISON REFERENCE VALUE: $T_R = 21.575444$ K


Matrix of equivalence

Lab, S/N i	Lab, S/N j \longrightarrow													
	NIST A128		NIST A129		NMi-VSL 226246		NPL 221481		NPL 221483		NPL 221485			
	D_i	U_i		D_{ij}	U_{ij}		D_{ij}	U_{ij}		D_{ij}	U_{ij}			
	/ mK	/ mK		/ mK	/ mK		/ mK	/ mK		/ mK	/ mK			
NIST A128	0.331	0.417		0.021	0.589	-0.075	1.152	0.655	0.662	0.799	0.651	0.800	0.651	
NIST A129	0.310	0.417	-0.021	0.589		-0.096	1.152	0.634	0.662	0.778	0.651	0.779	0.651	
NMi-VSL 226246	0.406	1.073	0.075	1.152	0.096	1.152		0.730	1.190	0.874	1.184	0.874	1.184	
NPL 221481	-0.324	0.515	-0.655	0.662	-0.634	0.662	-0.730	1.190		0.144	0.718	0.144	0.718	
NPL 221483	-0.468	0.500	-0.799	0.651	-0.778	0.651	-0.874	1.184	-0.144	0.718		0.000	0.708	
NPL 221485	-0.469	0.500	-0.800	0.651	-0.779	0.651	-0.874	1.184	-0.144	0.718	0.000	0.708		
NRC A138	-0.178	1.142	-0.509	1.215	-0.488	1.215	-0.584	1.567	0.146	1.252	0.290	1.247	0.290	1.247
NRC A140	-0.330	1.142	-0.661	1.215	-0.640	1.215	-0.736	1.567	-0.006	1.252	0.138	1.247	0.139	1.247
PTB 229074	0.490	0.711	0.159	0.824	0.180	0.824	0.085	1.288	0.814	0.878	0.959	0.869	0.959	0.869
PTB 229075	0.249	0.660	-0.082	0.781	-0.061	0.781	-0.157	1.260	0.573	0.837	0.717	0.828	0.718	0.828
VNIIFTRI 79	-0.217	1.161	-0.548	1.234	-0.527	1.234	-0.623	1.581	0.107	1.270	0.251	1.264	0.251	1.264
VNIIFTRI 89	-0.100	1.242	-0.431	1.310	-0.410	1.310	-0.506	1.642	0.224	1.344	0.368	1.339	0.368	1.339

Matrix of equivalence (Continued)

Lab, S/N <i>i</i>	Lab, S/N <i>j</i> \longrightarrow														
	NRC A138		NRC A140		PTB 229074		PTB 229075		VNIIIFTRI 79		VNIIIFTRI 89				
	D_i	U_i		D_{ij}	U_{ij}		D_{ij}	U_{ij}		D_{ij}	U_{ij}				
	/ mK	/ mK		/ mK	/ mK		/ mK	/ mK		/ mK	/ mK				
NIST A128	0.331	0.417		0.509	1.215	0.661	1.215	-0.159	0.824	0.082	0.781	0.548	1.234	0.431	1.310
NIST A129	0.310	0.417		0.488	1.215	0.640	1.215	-0.180	0.824	0.061	0.781	0.527	1.234	0.410	1.310
NMi-VSL 226246	0.406	1.073		0.584	1.567	0.736	1.567	-0.085	1.288	0.157	1.260	0.623	1.581	0.506	1.642
NPL 221481	-0.324	0.515		-0.146	1.252	0.006	1.252	-0.814	0.878	-0.573	0.837	-0.107	1.270	-0.224	1.344
NPL 221483	-0.468	0.500		-0.290	1.247	-0.138	1.247	-0.959	0.869	-0.717	0.828	-0.251	1.264	-0.368	1.339
NPL 221485	-0.469	0.500		-0.290	1.247	-0.139	1.247	-0.959	0.869	-0.718	0.828	-0.251	1.264	-0.368	1.339
NRC A138	-0.178	1.142				0.152	1.615	-0.669	1.345	-0.427	1.319	0.039	1.628	-0.078	1.687
NRC A140	-0.330	1.142		-0.152	1.615			-0.820	1.345	-0.579	1.319	-0.113	1.628	-0.230	1.687
PTB 229074	0.490	0.711		0.669	1.345	0.820	1.345			0.241	0.970	0.708	1.361	0.591	1.431
PTB 229075	0.249	0.660		0.427	1.319	0.579	1.319	-0.241	0.970			0.466	1.336	0.349	1.406
VNIIIFTRI 79	-0.217	1.161		-0.039	1.628	0.113	1.628	-0.708	1.361	-0.466	1.336			-0.117	1.700
VNIIIFTRI 89	-0.100	1.242		0.078	1.687	0.230	1.687	-0.591	1.431	-0.349	1.406	0.117	1.700		

CCT-K1 : Nominal temperature, $T_{90} = 21.575$ K
Degrees of equivalence, D_i , and expanded uncertainties ($k = 2$), U_i , expressed in mK



Key comparison CCT-K1

MEASURAND : Temperature T_{90} / K

NOMINAL TEMPERATURE : $T_{90} = 22.677$ K

Laboratory individual measurements

Lab: Laboratory

S/N i : serial number of Thermometer i

T_i : temperature value for Thermometer i

u_i : standard uncertainty in the calibration of Thermometer i

u_{comp} : standard uncertainty in the comparison measurements

$$u_{\text{comp}} = 0.153 \text{ mK}$$

Lab, S/N i	T_i / mK	u_i / mK
NIST A128	22.677193	0.152
NIST A129	22.677155	0.152
NMi-VSL 226246	22.677427	0.542
NPL 221481	22.676930	0.210
NPL 221483	22.676699	0.201
NPL 221485	22.676646	0.201
NRC A138	22.676586	0.550
NRC A140	22.676571	0.550
PTB 229074	22.677453	0.324
PTB 229075	22.677222	0.295
VNIIFTRI 79	22.676869	0.560
VNIIFTRI 89	22.676872	0.605

Key comparison CCT-K1

MEASURAND : Temperature T_{90} / K

KEY COMPARISON REFERENCE VALUE: $T_R = 22.676998$ K


Matrix of equivalence

Lab, S/N i	Lab, S/N j \longrightarrow														
	NIST A128		NIST A129		NMi-VSL 226246		NPL 221481		NPL 221483		NPL 221485				
	D_i	U_i		D_{ij}	U_{ij}		D_{ij}	U_{ij}		D_{ij}	U_{ij}				
NIST A128	0.195	0.431			0.038	0.610	-0.234	1.206	0.263	0.676	0.494	0.665	0.546	0.665	
NIST A129	0.157	0.431		-0.038	0.610		-0.272	1.206	0.225	0.676	0.456	0.665	0.508	0.665	
NMi-VSL 226246	0.429	1.127		0.234	1.206	0.272	1.206		0.497	1.241	0.728	1.235	0.780	1.235	
NPL 221481	-0.068	0.520		-0.263	0.676	-0.225	0.676	-0.497	1.241		0.231	0.725	0.284	0.725	
NPL 221483	-0.299	0.506		-0.494	0.665	-0.456	0.665	-0.728	1.235	-0.231	0.725		0.052	0.715	
NPL 221485	-0.352	0.506		-0.546	0.665	-0.508	0.665	-0.780	1.235	-0.284	0.725	-0.052	0.715		
NRC A138	-0.412	1.142		-0.606	1.221	-0.568	1.221	-0.840	1.604	-0.344	1.255	-0.112	1.249	-0.060	1.249
NRC A140	-0.427	1.142		-0.622	1.221	-0.584	1.221	-0.856	1.604	-0.359	1.255	-0.128	1.249	-0.075	1.249
PTB 229074	0.455	0.717		0.261	0.837	0.299	0.837	0.027	1.335	0.523	0.886	0.755	0.877	0.807	0.877
PTB 229075	0.224	0.665		0.029	0.793	0.068	0.793	-0.205	1.308	0.292	0.844	0.523	0.835	0.576	0.835
VNIIFTRI 79	-0.129	1.161		-0.324	1.239	-0.286	1.239	-0.558	1.618	-0.061	1.272	0.170	1.266	0.222	1.266
VNIIFTRI 89	-0.126	1.248		-0.321	1.320	-0.283	1.320	-0.555	1.681	-0.058	1.352	0.173	1.346	0.225	1.346

Matrix of equivalence (Continued)

Lab, S/N <i>i</i>	Lab, S/N <i>j</i> \longrightarrow																
	NRC A138		NRC A140		PTB 229074		PTB 229075		VNIIIFTRI 79		VNIIIFTRI 89						
	D_i	U_i		D_{ij}	U_{ij}		D_{ij}	U_{ij}		D_{ij}	U_{ij}						
	$/ \text{mK}$			$/ \text{mK}$			$/ \text{mK}$			$/ \text{mK}$							
NIST A128	0.195	0.431		0.606	1.221	0.622	1.221	-0.261	0.837	-0.029	0.793	0.324	1.239	0.321	1.320		
NIST A129	0.157	0.431		0.568	1.221	0.584	1.221	-0.299	0.837	-0.068	0.793	0.286	1.239	0.283	1.320		
NMi-VSL 226246	0.429	1.127		0.840	1.604	0.856	1.604	-0.027	1.335	0.205	1.308	0.558	1.618	0.555	1.681		
NPL 221481	-0.068	0.520		0.344	1.255	0.359	1.255	-0.523	0.886	-0.292	0.844	0.061	1.272	0.058	1.352		
NPL 221483	-0.299	0.506		0.112	1.249	0.128	1.249	-0.755	0.877	-0.523	0.835	-0.170	1.266	-0.173	1.346		
NPL 221485	-0.352	0.506		0.060	1.249	0.075	1.249	-0.807	0.877	-0.576	0.835	-0.222	1.266	-0.225	1.346		
NRC A138	-0.412	1.142				0.016	1.615	-0.867	1.348	-0.636	1.321	-0.282	1.628	-0.285	1.691		
NRC A140	-0.427	1.142			-0.016	1.615		-0.882	1.348	-0.651	1.321	-0.298	1.628	-0.301	1.691		
PTB 229074	0.455	0.717						0.867	1.348	0.882	1.348	0.231	0.978	0.585	1.365	0.582	1.439
PTB 229075	0.224	0.665						0.636	1.321	0.651	1.321			0.353	1.338	0.351	1.414
VNIIIFTRI 79	-0.129	1.161						0.282	1.628	0.298	1.628	-0.585	1.365			-0.003	1.704
VNIIIFTRI 89	-0.126	1.248						0.285	1.691	0.301	1.691	-0.582	1.439	-0.351	1.414		

CCT-K1 : Nominal temperature, $T_{90} = 22.677$ K
Degrees of equivalence, D_i , and expanded uncertainties ($k = 2$), U_i , expressed in mK



Key comparison CCT-K1

MEASURAND : Temperature T_{90} / K

NOMINAL TEMPERATURE : $T_{90} = 23.496$ K

Laboratory individual measurements

Lab: Laboratory

S/N i : serial number of Thermometer i

T_i : temperature value for Thermometer i

u_i : standard uncertainty in the calibration of Thermometer i

u_{comp} : standard uncertainty in the comparison measurements

$$u_{\text{comp}} = 0.153 \text{ mK}$$

Lab, S/N i	T_i / mK	u_i / mK
NIST A128	23.496504	0.162
NIST A129	23.496495	0.162
NMi-VSL 226246	23.496942	0.563
NPL 221481	23.496400	0.213
NPL 221483	23.496171	0.204
NPL 221485	23.496328	0.204
NRC A138	23.496000	0.550
NRC A140	23.496009	0.550
PTB 229074	23.496996	0.327
PTB 229075	23.496714	0.297
VNIIFTRI 79	23.496311	0.560
VNIIFTRI 89	23.496381	0.607

Key comparison CCT-K1

MEASURAND : Temperature T_{90} / K

KEY COMPARISON REFERENCE VALUE: $T_R = 23.496448$ K


Matrix of equivalence

Lab, S/N i	Lab, S/N j \longrightarrow													
	D _{i} U _{i} / mK		NIST A128		NIST A129		NMi-VSL 226246		NPL 221481		NPL 221483		NPL 221485	
	D _{ij} / mK	U _{ij} / mK	D _{ij} / mK	U _{ij} / mK	D _{ij} / mK	U _{ij} / mK	D _{ij} / mK	U _{ij} / mK	D _{ij} / mK	U _{ij} / mK	D _{ij} / mK	U _{ij} / mK	D _{ij} / mK	U _{ij} / mK
NIST A128	0.056	0.446			0.010	0.630	-0.438	1.249	0.104	0.688	0.333	0.677	0.176	0.677
NIST A129	0.047	0.446	-0.010	0.630			-0.448	1.249	0.094	0.688	0.324	0.677	0.167	0.677
NMi-VSL 226246	0.494	1.166	0.438	1.249	0.448	1.249			0.542	1.279	0.771	1.273	0.614	1.273
NPL 221481	-0.048	0.524	-0.104	0.688	-0.094	0.688	-0.542	1.279			0.229	0.731	0.073	0.731
NPL 221483	-0.277	0.510	-0.333	0.677	-0.324	0.677	-0.771	1.273	-0.229	0.731			-0.157	0.721
NPL 221485	-0.120	0.510	-0.176	0.677	-0.167	0.677	-0.614	1.273	-0.073	0.731	0.157	0.721		
NRC A138	-0.448	1.142	-0.504	1.226	-0.495	1.226	-0.942	1.632	-0.401	1.256	-0.171	1.250	-0.328	1.250
NRC A140	-0.439	1.142	-0.495	1.226	-0.486	1.226	-0.934	1.632	-0.392	1.256	-0.162	1.250	-0.319	1.250
PTB 229074	0.548	0.722	0.492	0.848	0.501	0.848	0.054	1.371	0.596	0.892	0.825	0.883	0.668	0.883
PTB 229075	0.266	0.669	0.210	0.804	0.219	0.804	-0.228	1.344	0.314	0.850	0.543	0.841	0.386	0.841
VNIIFTRI 79	-0.137	1.161	-0.193	1.244	-0.184	1.244	-0.631	1.646	-0.090	1.274	0.140	1.268	-0.017	1.268
VNIIFTRI 89	-0.067	1.252	-0.124	1.329	-0.114	1.329	-0.562	1.711	-0.020	1.358	0.210	1.352	0.053	1.352

Matrix of equivalence (Continued)

Lab, S/N <i>i</i>	Lab, S/N <i>j</i> \longrightarrow											
	NRC A138		NRC A140		PTB 229074		PTB 229075		VNIIIFTRI 79		VNIIIFTRI 89	
	D_i	U_i		D_{ij}	U_{ij}		D_{ij}	U_{ij}		D_{ij}	U_{ij}	
	/ mK	/ mK		/ mK	/ mK		/ mK	/ mK		/ mK	/ mK	
NIST A128	0.056	0.446		0.504	1.226	0.495	1.226	-0.492	0.848	-0.210	0.804	0.193
NIST A129	0.047	0.446		0.495	1.226	0.486	1.226	-0.501	0.848	-0.219	0.804	0.184
NMi-VSL 226246	0.494	1.166		0.942	1.632	0.934	1.632	-0.054	1.371	0.228	1.344	0.631
NPL 221481	-0.048	0.524		0.401	1.256	0.392	1.256	-0.596	0.892	-0.314	0.850	0.090
NPL 221483	-0.277	0.510		0.171	1.250	0.162	1.250	-0.825	0.883	-0.543	0.841	-0.140
NPL 221485	-0.120	0.510		0.328	1.250	0.319	1.250	-0.668	0.883	-0.386	0.841	0.017
NRC A138	-0.448	1.142				-0.009	1.615	-0.996	1.351	-0.714	1.323	-0.311
NRC A140	-0.439	1.142				0.009	1.615		-0.987	1.351	-0.705	1.323
PTB 229074	0.548	0.722							0.282	0.984	0.685	1.367
PTB 229075	0.266	0.669								0.403	1.340	0.333
VNIIIFTRI 79	-0.137	1.161									-0.070	1.708
VNIIIFTRI 89	-0.067	1.252										

CCT-K1 : Nominal temperature, $T_{90} = 23.496$ K
Degrees of equivalence, D_i , and expanded uncertainties ($k = 2$), U_i , expressed in mK



Key comparison CCT-K1

MEASURAND : Temperature T_{90} / K

NOMINAL TEMPERATURE : $T_{90} = 24.102$ K

Laboratory individual measurements

Lab: Laboratory

S/N i : serial number of Thermometer i

T_i : temperature value for Thermometer i

u_i : standard uncertainty in the calibration of Thermometer i

u_{comp} : standard uncertainty in the comparison measurements

$$u_{\text{comp}} = 0.153 \text{ mK}$$

Lab, S/N i	T_i / mK	u_i / mK
NIST A128	24.101929	0.168
NIST A129	24.102011	0.168
NMi-VSL 226246	24.102501	0.578
NPL 221481	24.101920	0.215
NPL 221483	24.101771	0.206
NPL 221485	24.101752	0.206
NRC A138	24.101887	0.550
NRC A140	24.101930	0.550
PTB 229074	24.102615	0.329
PTB 229075	24.102251	0.299
VNIIFTRI 79	24.101651	0.560
VNIIFTRI 89	24.101785	0.609

Key comparison CCT-K1

MEASURAND : Temperature T_{90} / K

KEY COMPARISON REFERENCE VALUE: $T_R = 24.101970$ K


Matrix of equivalence

Lab, S/N i	Lab, S/N j \longrightarrow											
	NIST A128		NIST A129		NMi-VSL 226246		NPL 221481		NPL 221483		NPL 221485	
	D_i	U_i		D_{ij}	U_{ij}		D_{ij}	U_{ij}		D_{ij}	U_{ij}	
	/ mK	/ mK		/ mK	/ mK		/ mK	/ mK		/ mK	/ mK	
NIST A128	-0.041	0.454										
NIST A129	0.041	0.454		-0.082	0.643		-0.573	1.279		0.009	0.696	
NMi-VSL 226246	0.531	1.196		0.082	0.643		-0.491	1.279		0.091	0.696	
NPL 221481	-0.050	0.527		0.573	1.279	0.491	1.279		0.581	1.307	0.731	1.301
NPL 221483	-0.199	0.513		-0.009	0.696	-0.091	0.696	-0.581	1.307		0.149	0.735
NPL 221485	-0.218	0.513		-0.158	0.685	-0.240	0.685	-0.731	1.301	-0.149	0.735	0.019
NRC A138	-0.083	1.142		-0.177	0.685	-0.259	0.685	-0.749	1.301	-0.168	0.735	
NRC A140	-0.040	1.142		-0.042	1.229	-0.124	1.229	-0.614	1.653	-0.033	1.258	0.116
PTB 229074	0.645	0.725		0.001	1.229	-0.081	1.229	-0.571	1.653	0.010	1.258	0.159
PTB 229075	0.281	0.671		0.686	0.856	0.604	0.856	0.114	1.398	0.695	0.896	0.845
VNIIFTRI 79	-0.319	1.161		0.322	0.811	0.240	0.811	-0.251	1.371	0.331	0.854	0.480
VNIIFTRI 89	-0.185	1.256		-0.278	1.247	-0.360	1.247	-0.850	1.667	-0.269	1.275	-0.120
				-0.144	1.335	-0.226	1.335	-0.717	1.734	-0.135	1.362	0.014
												0.033
												1.356

Matrix of equivalence (Continued)

Lab, S/N <i>i</i>	Lab, S/N <i>j</i> \longrightarrow														
	NRC A138		NRC A140		PTB 229074		PTB 229075		VNIIIFTRI 79		VNIIIFTRI 89				
	D_i	U_i		D_{ij}	U_{ij}		D_{ij}	U_{ij}		D_{ij}	U_{ij}				
	$/ \text{mK}$			$/ \text{mK}$			$/ \text{mK}$			$/ \text{mK}$					
NIST A128	-0.041	0.454		0.042	1.229	-0.001	1.229	-0.686	0.856	-0.322	0.811	0.278	1.247	0.144	1.335
NIST A129	0.041	0.454		0.124	1.229	0.081	1.229	-0.604	0.856	-0.240	0.811	0.360	1.247	0.226	1.335
NMi-VSL 226246	0.531	1.196		0.614	1.653	0.571	1.653	-0.114	1.398	0.251	1.371	0.850	1.667	0.717	1.734
NPL 221481	-0.050	0.527		0.033	1.258	-0.010	1.258	-0.695	0.896	-0.331	0.854	0.269	1.275	0.135	1.362
NPL 221483	-0.199	0.513		-0.116	1.252	-0.159	1.252	-0.845	0.888	-0.480	0.845	0.120	1.269	-0.014	1.356
NPL 221485	-0.218	0.513		-0.135	1.252	-0.178	1.252	-0.863	0.888	-0.499	0.845	0.101	1.269	-0.033	1.356
NRC A138	-0.083	1.142				-0.043	1.615	-0.728	1.352	-0.364	1.325	0.236	1.628	0.102	1.697
NRC A140	-0.040	1.142		0.043	1.615			-0.685	1.352	-0.321	1.325	0.279	1.628	0.145	1.697
PTB 229074	0.645	0.725		0.728	1.352	0.685	1.352			0.364	0.988	0.964	1.369	0.830	1.450
PTB 229075	0.281	0.671		0.364	1.325	0.321	1.325	-0.364	0.988			0.600	1.341	0.466	1.424
VNIIIFTRI 79	-0.319	1.161		-0.236	1.628	-0.279	1.628	-0.964	1.369	-0.600	1.341			-0.134	1.710
VNIIIFTRI 89	-0.185	1.256		-0.102	1.697	-0.145	1.697	-0.830	1.450	-0.466	1.424	0.134	1.710		

CCT-K1 : Nominal temperature, $T_{90} = 24.102$ K
Degrees of equivalence, D_i , and expanded uncertainties ($k = 2$), U_i , expressed in mK



Key comparison CCT-K1

MEASURAND : Temperature T_{90} / K

NOMINAL TEMPERATURE : $T_{90} = 24.340$ K

Laboratory individual measurements

Lab: Laboratory

S/N i : serial number of Thermometer i

T_i : temperature value for Thermometer i

u_i : standard uncertainty in the calibration of Thermometer i

u_{comp} : standard uncertainty in the comparison measurements

$$u_{\text{comp}} = 0.153 \text{ mK}$$

Lab, S/N i	T_i / mK	u_i / mK
NIST A128	24.340191	0.174
NIST A129	24.340214	0.174
NMi-VSL 226246	24.340765	0.584
NPL 221481	24.340270	0.215
NPL 221485	24.340219	0.206
NRC A138	24.340051	0.550
NRC A140	24.339870	0.550
PTB 229074	24.340847	0.329
PTB 229075	24.340698	0.299

Key comparison CCT-K1

MEASURAND : Temperature T_{90} / K

KEY COMPARISON REFERENCE VALUE: $T_R = 24.340317$ K


Matrix of equivalence

Lab, S/N i	Lab, S/N j \longrightarrow														
	NIST A128		NIST A129		NMi-VSL 226246		NPL 221481		NPL 221485		NRC A138				
	D_i	U_i		D_{ij}	U_{ij}		D_{ij}	U_{ij}		D_{ij}	U_{ij}				
NIST A128	-0.126	0.463			-0.023	0.655	-0.574	1.293	-0.079	0.703	-0.028	0.692	0.140	1.232	
NIST A129	-0.103	0.463		0.023	0.655		-0.552	1.293	-0.056	0.703	-0.006	0.692	0.163	1.232	
NMi-VSL 226246	0.448	1.207		0.574	1.293	0.552	1.293		0.495	1.318	0.546	1.312	0.715	1.662	
NPL 221481	-0.047	0.528		0.079	0.703	0.056	0.703	-0.495	1.318		0.051	0.737	0.219	1.258	
NPL 221485	-0.098	0.514		0.028	0.692	0.006	0.692	-0.546	1.312	-0.051	0.737		0.169	1.252	
NRC A138	-0.266	1.142		-0.140	1.232	-0.163	1.232	-0.715	1.662	-0.219	1.258	-0.169	1.252		
NRC A140	-0.447	1.142		-0.321	1.232	-0.344	1.232	-0.896	1.662	-0.400	1.258	-0.350	1.252	-0.181	1.615
PTB 229074	0.530	0.726		0.656	0.861	0.633	0.861	0.081	1.409	0.577	0.898	0.627	0.890	0.796	1.353
PTB 229075	0.381	0.673		0.507	0.817	0.484	0.817	-0.067	1.382	0.428	0.855	0.479	0.846	0.647	1.325

Matrix of equivalence (Continued)

Lab, S/N <i>i</i>	Lab, S/N <i>j</i> \longrightarrow					
	NRC A140		PTB 229074		PTB 229075	
	D_i	U_i	D_{ij}	U_{ij}	D_{ij}	U_{ij}
NIST A128	-0.126	0.463	0.321	1.232	-0.656	0.861
NIST A129	-0.103	0.463	0.344	1.232	-0.633	0.861
NMi-VSL 226246	0.448	1.207	0.896	1.662	-0.081	1.409
NPL 221481	-0.047	0.528	0.400	1.258	-0.577	0.898
NPL 221485	-0.098	0.514	0.350	1.252	-0.627	0.890
NRC A138	-0.266	1.142	0.181	1.615	-0.796	1.353
NRC A140	-0.447	1.142			-0.977	1.353
PTB 229074	0.530	0.726	0.977	1.353		0.149
PTB 229075	0.381	0.673	0.828	1.325	-0.149	0.990

CCT-K1 : Nominal temperature, $T_{90} = 24.340$ K
Degrees of equivalence, D_i , and expanded uncertainties ($k = 2$), U_i , expressed in mK



Key comparison CCT-K1

MEASURAND : Temperature T_{90} / K

NOMINAL TEMPERATURE : $T_{90} = 24.446$ K

Laboratory individual measurements

Lab: Laboratory

S/N i : serial number of Thermometer i

T_i : temperature value for Thermometer i

u_i : standard uncertainty in the calibration of Thermometer i

u_{comp} : standard uncertainty in the comparison measurements

$$u_{\text{comp}} = 0.153 \text{ mK}$$

Lab, S/N i	T_i / mK	u_i / mK
NIST A128	24.446295	0.174
NIST A129	24.446429	0.174
NMi-VSL 226246	24.446981	0.587
NPL 221481	24.446370	0.216
NPL 221483	24.446221	0.207
NPL 221485	24.446209	0.207
NRC A138	24.446435	0.550
NRC A140	24.446466	0.550
PTB 229074	24.447000	0.330
PTB 229075	24.446767	0.300
VNIIFTRI 79	24.446009	0.560
VNIIFTRI 89	24.446107	0.610

Key comparison CCT-K1

MEASURAND : Temperature T_{90} / K

KEY COMPARISON REFERENCE VALUE: $T_R = 24.446403$ K


Matrix of equivalence

Lab, S/N i			Lab, S/N j			
	D_i	U_i	D_{ij}	U_{ij}	D_{ij}	U_{ij}
	/ mK	/ mK	/ mK	/ mK	/ mK	/ mK
NIST A128	-0.108	0.463				
NIST A129	0.026	0.463	-0.134	0.655	-0.686	1.298
NMi-VSL 226246	0.578	1.212	0.134	0.655	-0.553	1.298
NPL 221481	-0.033	0.529	0.686	1.298	0.553	1.298
NPL 221483	-0.182	0.514	0.075	0.703	-0.058	0.703
NPL 221485	-0.194	0.514	-0.611	1.323	-0.611	1.323
NRC A138	0.032	1.142	-0.074	0.692	-0.207	0.692
NRC A140	0.063	1.142	-0.760	1.317	-0.760	1.317
PTB 229074	0.597	0.727	-0.149	0.738	-0.149	0.738
PTB 229075	0.364	0.673	-0.112	0.727	-0.112	0.727
VNIIFTRI 79	-0.394	1.161	-0.012	0.727	-0.012	0.727
VNIIFTRI 89	-0.296	1.257	0.227	1.252	0.227	1.252
			0.257	1.252	0.257	1.252
			0.791	0.890	0.791	0.890
			0.559	0.847	0.559	0.847
			-0.199	1.270	-0.199	1.270
			1.270	-0.199	1.270	-0.199
			-0.102	1.359	-0.102	1.359
			1.359	-0.102	1.359	-0.102

Matrix of equivalence (Continued)

Lab, S/N <i>i</i>	Lab, S/N <i>j</i> \longrightarrow														
	NRC A138		NRC A140		PTB 229074		PTB 229075		VNIIIFTRI 79		VNIIIFTRI 89				
	D_i	U_i		D_{ij}	U_{ij}		D_{ij}	U_{ij}		D_{ij}	U_{ij}				
	$/ \text{mK}$			$/ \text{mK}$			$/ \text{mK}$			$/ \text{mK}$					
NIST A128	-0.108	0.463		-0.140	1.232	-0.171	1.232	-0.705	0.862	-0.472	0.817	0.286	1.250	0.188	1.340
NIST A129	0.026	0.463		-0.007	1.232	-0.037	1.232	-0.572	0.862	-0.339	0.817	0.419	1.250	0.321	1.340
NMi-VSL 226246	0.578	1.212		0.546	1.665	0.515	1.665	-0.019	1.414	0.214	1.387	0.972	1.679	0.874	1.747
NPL 221481	-0.033	0.529		-0.065	1.258	-0.096	1.258	-0.630	0.899	-0.397	0.856	0.361	1.276	0.263	1.364
NPL 221483	-0.182	0.514		-0.214	1.252	-0.245	1.252	-0.779	0.890	-0.546	0.847	0.212	1.270	0.114	1.359
NPL 221485	-0.194	0.514		-0.227	1.252	-0.257	1.252	-0.791	0.890	-0.559	0.847	0.199	1.270	0.102	1.359
NRC A138	0.032	1.142				-0.030	1.615	-0.565	1.354	-0.332	1.325	0.426	1.628	0.328	1.699
NRC A140	0.063	1.142		0.030	1.615		-0.535	1.354	-0.302	1.325	0.456	1.628	0.358	1.699	
PTB 229074	0.597	0.727		0.565	1.354	0.535	1.354		0.233	0.991	0.991	1.370	0.893	1.452	
PTB 229075	0.364	0.673		0.332	1.325	0.302	1.325	-0.233	0.991		0.758	1.342	0.660	1.426	
VNIIIFTRI 79	-0.394	1.161		-0.426	1.628	-0.456	1.628	-0.991	1.370	-0.758	1.342		-0.098	1.712	
VNIIIFTRI 89	-0.296	1.257		-0.328	1.699	-0.358	1.699	-0.893	1.452	-0.660	1.426	0.098	1.712		

CCT-K1 : Nominal temperature, $T_{90} = 24.446$ K
Degrees of equivalence, D_i , and expanded uncertainties ($k = 2$), U_i , expressed in mK



Key comparison CCT-K1

MEASURAND : Temperature T_{90} / K

NOMINAL TEMPERATURE : $T_{90} = 24.551$ K

Laboratory individual measurements

Lab: Laboratory

S/N i : serial number of Thermometer i

T_i : temperature value for Thermometer i

u_i : standard uncertainty in the calibration of Thermometer i

u_{comp} : standard uncertainty in the comparison measurements

$$u_{\text{comp}} = 0.153 \text{ mK}$$

Lab, S/N i	T_i / mK	u_i / mK
NIST A128	24.551286	0.174
NIST A129	24.551288	0.174
NMi-VSL 226246	24.551814	0.588
NPL 221481	24.551375	0.216
NPL 221483	24.551221	0.207
NPL 221485	24.551122	0.207
NRC A138	24.551544	0.550
NRC A140	24.551501	0.550
PTB 229074	24.551980	0.330
PTB 229075	24.551635	0.300
VNIIFTRI 79	24.550982	0.560
VNIIFTRI 89	24.551080	0.610

Key comparison CCT-K1

MEASURAND : Temperature T_{90} / K

KEY COMPARISON REFERENCE VALUE: $T_R = 24.551354$ K

Matrix of equivalence

Lab, S/N i			Lab, S/N j			
	D_i	U_i	D_{ij}	U_{ij}	D_{ij}	U_{ij}
	/ mK	/ mK	/ mK	/ mK	/ mK	/ mK
NIST A128	-0.068	0.463				
NIST A129	-0.066	0.463	0.002	0.655	-0.528	1.300
NMi-VSL 226246	0.460	1.215	0.528	1.300	0.526	1.300
NPL 221481	0.021	0.529	0.088	0.703	-0.440	1.325
NPL 221483	-0.133	0.515	-0.065	0.693	-0.593	1.319
NPL 221485	-0.232	0.515	-0.164	0.693	-0.166	0.693
NRC A138	0.190	1.142	0.258	1.232	0.256	1.232
NRC A140	0.147	1.142	0.214	1.232	0.212	1.232
PTB 229074	0.626	0.727	0.694	0.862	0.692	0.862
PTB 229075	0.281	0.674	0.348	0.817	0.346	0.817
VNIIFTRI 79	-0.372	1.161	-0.304	1.250	-0.306	1.250
VNIIFTRI 89	-0.274	1.258	-0.207	1.341	-0.209	1.341

Matrix of equivalence (Continued)

Lab, S/N <i>i</i>	Lab, S/N <i>j</i> \longrightarrow														
	NRC A138		NRC A140		PTB 229074		PTB 229075		VNIIIFTRI 79		VNIIIFTRI 89				
	D_i	U_i		D_{ij}	U_{ij}		D_{ij}	U_{ij}		D_{ij}	U_{ij}				
	$/ \text{mK}$			$/ \text{mK}$			$/ \text{mK}$			$/ \text{mK}$					
NIST A128	-0.068	0.463		-0.258	1.232	-0.214	1.232	-0.694	0.862	-0.348	0.817	0.304	1.250	0.207	1.341
NIST A129	-0.066	0.463		-0.256	1.232	-0.212	1.232	-0.692	0.862	-0.346	0.817	0.306	1.250	0.209	1.341
NMi-VSL 226246	0.460	1.215		0.270	1.667	0.314	1.667	-0.166	1.416	0.180	1.389	0.832	1.680	0.735	1.749
NPL 221481	0.021	0.529		-0.169	1.259	-0.126	1.259	-0.606	0.900	-0.260	0.857	0.392	1.276	0.295	1.365
NPL 221483	-0.133	0.515		-0.323	1.252	-0.279	1.252	-0.759	0.891	-0.413	0.848	0.239	1.270	0.142	1.359
NPL 221485	-0.232	0.515		-0.422	1.252	-0.378	1.252	-0.858	0.891	-0.512	0.848	0.140	1.270	0.043	1.359
NRC A138	0.190	1.142				0.043	1.615	-0.437	1.354	-0.091	1.326	0.562	1.628	0.464	1.699
NRC A140	0.147	1.142		-0.043	1.615			-0.480	1.354	-0.134	1.326	0.518	1.628	0.421	1.699
PTB 229074	0.626	0.727		0.437	1.354	0.480	1.354			0.346	0.991	0.998	1.370	0.901	1.453
PTB 229075	0.281	0.674		0.091	1.326	0.134	1.326	-0.346	0.991			0.652	1.342	0.555	1.427
VNIIIFTRI 79	-0.372	1.161		-0.562	1.628	-0.518	1.628	-0.998	1.370	-0.652	1.342			-0.097	1.712
VNIIIFTRI 89	-0.274	1.258		-0.464	1.699	-0.421	1.699	-0.901	1.453	-0.555	1.427	0.097	1.712		

CCT-K1 : Nominal temperature, $T_{90} = 24.551$ K
Degrees of equivalence, D_i , and expanded uncertainties ($k = 2$), U_i , expressed in mK

