

Key comparison CCT-K1 and EURAMET.T-K1

MEASURAND : Temperature T_{90} / K

Equivalence statements

For each temperature in the comparison the key comparison reference value, T_R , is the weighted mean of T_i , the individual temperature values of thermometers.

The weighted mean is calculated using the laboratory uncertainty combined with the comparison uncertainty, to set the weights. T_R is used as the baseline for the comparison, but has no special significance with respect to the ITS-90, and is used without uncertainty.

The degree of equivalence of each temperature T_i with respect to the key comparison reference value, T_R , is given by a pair of terms: $D_i = (T_i - T_R)$ and U_i , its expanded uncertainty at 95 % confidence, both expressed in mK.

U_i includes the uncertainties in the original laboratory calibrations and in the comparison measurements, but not in T_R .

CCT-K1

KEY COMPARISON REFERENCE VALUES, T_R / K

T_R / K	T_R / K	T_R / K	T_R / K
0.649875	1.996554	4.477522	16.999335
0.676928	2.248485	5.000458	18.597377
0.704354	2.600776	5.948165	20.298899
0.761580	2.699911	7.201544	21.575444
0.858421	2.896733	8.296372	22.676998
0.991223	2.996648	8.399612	23.496448
1.031584	3.099398	9.508020	24.101970
1.224991	3.400235	10.803390	24.340317
1.249542	3.429250	12.297309	24.446403
1.503370	3.800903	13.798183	24.551354
1.754822	4.224794	15.499566	

EURAMET.T-K1

EURAMET.T-K1 was linked to CCT-K1 via NPL, PTB and VSL who participated in both comparisons.

Key comparison CCT-K1

MEASURAND : Temperature T_{90} / K

NOMINAL TEMPERATURE : $T_{90} = 0.650$ K

Laboratory individual measurements

Lab: Laboratory

S/N i : serial number of Thermometer i

T_i : temperature value for Thermometer i

u_i : standard uncertainty in the calibration of Thermometer i

u_{comp} : standard uncertainty in the comparison measurements

$$u_{\text{comp}} = 0.205 \text{ mK}$$

Lab, S/N i	T_i / mK	u_i / mK
NIST A128	0.650152	0.139
NIST A129	0.649998	0.139
NMi-VSL 226246	0.650370	0.182
NPL 221481	0.649615	0.254
NPL 221483	0.649625	0.254
NPL 221485	0.649681	0.254
PTB 229074	0.649546	0.180
PTB 229075	0.649551	0.180
VNIIFTRI 79	0.650145	0.540
VNIIFTRI 89	0.650376	0.550

Key comparison CCT-K1

MEASURAND : Temperature T_{90} / K

KEY COMPARISON REFERENCE VALUE: $T_R = 0.649875$ K


Matrix of equivalence

Lab, S/N i	Lab, S/N j \longrightarrow											
	D_i / mK		U_i / mK		D_{ij} / mK		U_{ij} / mK		D_{ij} / mK		U_{ij} / mK	
NIST A128	0.277	0.495			0.154	0.700	-0.218	0.739	0.536	0.819	0.526	0.819
NIST A129	0.123	0.495			-0.154	0.700	-0.372	0.739	0.383	0.819	0.373	0.819
NMi-VSL 226246	0.495	0.548			0.218	0.739	0.372	0.739	0.755	0.852	0.745	0.852
NPL 221481	-0.260	0.653			-0.536	0.819	-0.383	0.819	-0.755	0.852	-0.010	0.923
NPL 221483	-0.250	0.653			-0.526	0.819	-0.373	0.819	-0.745	0.852	0.010	0.923
NPL 221485	-0.194	0.653			-0.470	0.819	-0.316	0.819	-0.689	0.852	0.066	0.923
PTB 229074	-0.329	0.546			-0.606	0.737	-0.452	0.737	-0.824	0.773	-0.069	0.851
PTB 229075	-0.324	0.546			-0.601	0.737	-0.447	0.737	-0.819	0.773	-0.064	0.851
VNIIFTRI 79	0.270	1.155			-0.007	1.257	0.147	1.257	-0.225	1.279	0.529	1.327
VNIIFTRI 89	0.501	1.174			0.224	1.274	0.378	1.274	0.005	1.295	0.760	1.343
					0.224	1.274	0.378	1.274	0.005	1.295	0.760	1.343
					0.224	1.274	0.378	1.274	0.005	1.295	0.760	1.343

Matrix of equivalence (Continued)

Lab, S/N i	Lab, S/N j \longrightarrow											
	D_i / mK		U_i / mK		D_{ij} / mK		U_{ij} / mK		D_{ij} / mK		U_{ij} / mK	
NIST A128	0.277	0.495			0.606	0.737	0.601	0.737	0.007	1.257	-0.224	1.274
NIST A129	0.123	0.495			0.452	0.737	0.447	0.737	-0.147	1.257	-0.378	1.274
NMi-VSL 226246	0.495	0.548			0.824	0.773	0.819	0.773	0.225	1.279	-0.005	1.295
NPL 221481	-0.260	0.653			0.069	0.851	0.064	0.851	-0.529	1.327	-0.760	1.343
NPL 221483	-0.250	0.653			0.079	0.851	0.074	0.851	-0.519	1.327	-0.750	1.343
NPL 221485	-0.194	0.653			0.135	0.851	0.130	0.851	-0.463	1.327	-0.694	1.343
PTB 229074	-0.329	0.546					-0.005	0.772	-0.599	1.278	-0.830	1.294
PTB 229075	-0.324	0.546			0.005	0.772			-0.594	1.278	-0.825	1.294
VNIIFTRI 79	0.270	1.155			0.599	1.278	0.594	1.278			-0.231	1.647
VNIIFTRI 89	0.501	1.174			0.8296	1.2945	0.825	1.294	0.231	1.647		

CCT-K1 : Nominal temperature, $T_{90} = 0.650$ K
Degrees of equivalence, D_i , and expanded uncertainties ($k = 2$), U_i , expressed in mK



Key comparison CCT-K1

MEASURAND : Temperature T_{90} / K

NOMINAL TEMPERATURE : $T_{90} = 0.677$ K

Laboratory individual measurements

Lab: Laboratory

S/N i : serial number of Thermometer i

T_i : temperature value for Thermometer i

u_i : standard uncertainty in the calibration of Thermometer i

u_{comp} : standard uncertainty in the comparison measurements

$$u_{\text{comp}} = 0.201 \text{ mK}$$

Lab, S/N i	T_i / mK	u_i / mK
NIST A128	0.677079	0.121
NIST A129	0.676957	0.121
NMi-VSL 226246	0.677410	0.182
NPL 221481	0.677030	0.245
NPL 221483	0.676970	0.245
NPL 221485	0.676438	0.245
PTB 229074	0.676573	0.180
PTB 229075	0.676440	0.180
VNIIFTRI 79	0.677630	0.533
VNIIFTRI 89	0.677956	0.548

Key comparison CCT-K1

MEASURAND : Temperature T_{90} / K

KEY COMPARISON REFERENCE VALUE: $T_R = 0.676928$ K


Matrix of equivalence

Lab, S/N i	Lab, S/N j \longrightarrow					
	D_i	U_i	D_{ij}	U_{ij}	D_{ij}	U_{ij}
	/ mK	/ mK	/ mK	/ mK	/ mK	/ mK
NIST A128	0.151	0.470	0.122	0.664	-0.330	0.718
NIST A129	0.029	0.470	-0.122	0.664	-0.452	0.718
NMi-VSL 226246	0.482	0.543	0.330	0.718	0.452	0.718
NPL 221481	0.102	0.635	-0.049	0.790	0.073	0.790
NPL 221483	0.042	0.635	-0.109	0.790	0.013	0.790
NPL 221485	-0.490	0.635	-0.641	0.790	-0.440	0.835
PTB 229074	-0.355	0.540	-0.506	0.716	-0.384	0.716
PTB 229075	-0.488	0.540	-0.639	0.716	-0.517	0.716
VNIIFTRI 79	0.702	1.139	0.551	1.232	0.673	1.232
VNIIFTRI 89	1.028	1.167	0.877	1.258	0.999	1.258

Matrix of equivalence (Continued)

Lab, S/N i	Lab, S/N j \longrightarrow					
	D_i	U_i	D_{ij}	U_{ij}	D_{ij}	U_{ij}
	/ mK	/ mK	/ mK	/ mK	/ mK	/ mK
NIST A128	0.151	0.470	0.506	0.716	0.639	0.716
NIST A129	0.029	0.470	0.384	0.716	0.517	0.716
NMi-VSL 226246	0.482	0.543	0.836	0.766	0.970	0.766
NPL 221481	0.102	0.635	0.457	0.833	0.590	0.833
NPL 221483	0.042	0.635	0.397	0.833	0.530	0.833
NPL 221485	-0.490	0.635	-0.135	0.833	-0.002	0.833
PTB 229074	-0.355	0.540	0.133	0.764	-1.057	1.261
PTB 229075	-0.488	0.540	-0.133	0.764	-1.190	1.261
VNIIFTRI 79	0.702	1.139	1.057	1.261	-0.326	1.631
VNIIFTRI 89	1.028	1.167	1.383	1.286	1.516	1.286

CCT-K1 : Nominal temperature, $T_{90} = 0.677$ K
Degrees of equivalence, D_i , and expanded uncertainties ($k = 2$), U_i , expressed in mK



Key comparison CCT-K1

MEASURAND : Temperature T_{90} / K

NOMINAL TEMPERATURE : $T_{90} = 0.704$ K

Laboratory individual measurements

Lab: Laboratory

S/N i : serial number of Thermometer i

T_i : temperature value for Thermometer i

u_i : standard uncertainty in the calibration of Thermometer i

u_{comp} : standard uncertainty in the comparison measurements

$$u_{\text{comp}} = 0.198 \text{ mK}$$

Lab, S/N i	T_i / mK	u_i / mK
NIST A128	0.704853	0.110
NIST A129	0.704179	0.110
NMi-VSL 226246	0.704626	0.182
NPL 221481	0.704350	0.236
NPL 221483	0.704305	0.236
NPL 221485	0.704029	0.236
PTB 229074	0.704040	0.180
PTB 229075	0.703875	0.180
VNIIFTRI 79	0.705043	0.526
VNIIFTRI 89	0.705293	0.545

Key comparison CCT-K1

MEASURAND : Temperature T_{90} / K

KEY COMPARISON REFERENCE VALUE: $T_R = 0.704354$ K


Matrix of equivalence

Lab, S/N i	Lab, S/N j \longrightarrow														
	D_i U_i / mK		D_{ij} U_{ij} / mK		D_{ij} U_{ij} / mK		D_{ij} U_{ij} / mK		D_{ij} U_{ij} / mK						
NIST A128	0.499	0.453		0.674	0.640	0.227	0.702	0.503	0.765	0.548	0.765	0.824	0.765		
NIST A129	-0.175	0.453		-0.674	0.640	-0.447	0.702	-0.171	0.765	-0.126	0.765	0.150	0.765		
NMi-VSL 226246	0.272	0.537		-0.227	0.702	0.447	0.702		0.276	0.817	0.321	0.817	0.598	0.817	
NPL 221481	-0.004	0.616		-0.503	0.765	0.171	0.765	-0.276	0.817		0.045	0.872	0.322	0.872	
NPL 221483	-0.049	0.616		-0.548	0.765	0.126	0.765	-0.321	0.817	-0.045	0.872		0.277	0.872	
NPL 221485	-0.325	0.616		-0.824	0.765	-0.150	0.765	-0.598	0.817	-0.322	0.872	-0.277	0.872		
PTB 229074	-0.314	0.535		-0.813	0.701	-0.139	0.701	-0.586	0.758	-0.310	0.816	-0.265	0.816	0.012	0.816
PTB 229075	-0.479	0.535		-0.978	0.701	-0.304	0.701	-0.752	0.758	-0.476	0.816	-0.431	0.816	-0.154	0.816
VNIIFTRI 79	0.689	1.124		0.190	1.211	0.864	1.211	0.416	1.245	0.693	1.281	0.738	1.281	1.014	1.281
VNIIFTRI 89	0.939	1.160		0.440	1.245	1.114	1.245	0.667	1.278	0.943	1.314	0.988	1.314	1.264	1.314

Matrix of equivalence (Continued)

Lab, S/N i	Lab, S/N j \longrightarrow										
	D_i U_i / mK		D_{ij} U_{ij} / mK		D_{ij} U_{ij} / mK						
NIST A128	0.499	0.453		0.813	0.701	0.978	0.701	-0.190	1.211	-0.440	1.245
NIST A129	-0.175	0.453		0.139	0.701	0.304	0.701	-0.864	1.211	-1.114	1.245
NMi-VSL 226246	0.272	0.537		0.586	0.758	0.752	0.758	-0.416	1.245	-0.667	1.278
NPL 221481	-0.004	0.616		0.310	0.816	0.476	0.816	-0.693	1.281	-0.943	1.314
NPL 221483	-0.049	0.616		0.265	0.816	0.431	0.816	-0.738	1.281	-0.988	1.314
NPL 221485	-0.325	0.616		-0.012	0.816	0.154	0.816	-1.014	1.281	-1.264	1.314
PTB 229074	-0.314	0.535				0.166	0.756	-1.003	1.244	-1.253	1.277
PTB 229075	-0.479	0.535		-0.166	0.756			-1.168	1.244	-1.418	1.277
VNIIFTRI 79	0.689	1.124		1.003	1.244	1.168	1.244			-0.250	1.615
VNIIFTRI 89	0.939	1.160		1.2527	1.2772	1.418	1.277	0.250	1.615		

CCT-K1 : Nominal temperature, $T_{90} = 0.704$ K
Degrees of equivalence, D_i , and expanded uncertainties ($k = 2$), U_i , expressed in mK



Key comparison CCT-K1

MEASURAND : Temperature T_{90} / K

NOMINAL TEMPERATURE : $T_{90} = 0.762$ K

Laboratory individual measurements

Lab: Laboratory

S/N i : serial number of Thermometer i

T_i : temperature value for Thermometer i

u_i : standard uncertainty in the calibration of Thermometer i

u_{comp} : standard uncertainty in the comparison measurements

$$u_{\text{comp}} = 0.190 \text{ mK}$$

Lab, S/N i	T_i / mK	u_i / mK
NIST A128	0.761654	0.082
NIST A129	0.761491	0.082
NMi-VSL 226246	0.761866	0.182
NPL 221481	0.761820	0.218
NPL 221483	0.761770	0.218
NPL 221485	0.761112	0.218
PTB 229074	0.761082	0.180
PTB 229075	0.761437	0.180
VNIIFTRI 79	0.762375	0.511
VNIIFTRI 89	0.762682	0.540

Key comparison CCT-K1

MEASURAND : Temperature T_{90} / K

KEY COMPARISON REFERENCE VALUE: $T_R = 0.761580$ K


Matrix of equivalence

Lab, S/N i	Lab, S/N j \longrightarrow													
	D_i / mK		U_i / mK		D_{ij} / mK		U_{ij} / mK		D_{ij} / mK		U_{ij} / mK			
NIST A128	0.074	0.413			0.163	0.585	-0.213	0.669	-0.166	0.711	-0.116	0.711	0.541	0.711
NIST A129	-0.089	0.413			-0.163	0.585	-0.376	0.669	-0.329	0.711	-0.279	0.711	0.379	0.711
NMi-VSL 226246	0.286	0.526			0.213	0.669	0.376	0.669	0.046	0.781	0.096	0.781	0.754	0.781
NPL 221481	0.240	0.578			0.166	0.711	0.329	0.711	0.050	0.817	0.708	0.817		
NPL 221483	0.190	0.578			0.116	0.711	0.279	0.711	-0.050	0.817			0.658	0.817
NPL 221485	-0.468	0.578			-0.541	0.711	-0.379	0.711	-0.754	0.781	-0.708	0.817	-0.658	0.817
PTB 229074	-0.498	0.523			-0.572	0.667	-0.409	0.667	-0.785	0.742	-0.738	0.780	-0.688	0.780
PTB 229075	-0.143	0.523			-0.217	0.667	-0.054	0.667	-0.430	0.742	-0.383	0.780	-0.333	0.780
VNIIFTRI 79	0.795	1.090			0.721	1.166	0.884	1.166	0.508	1.211	0.555	1.234	0.605	1.234
VNIIFTRI 89	1.102	1.145			1.028	1.218	1.191	1.218	0.815	1.260	0.862	1.283	0.912	1.283
													1.569	1.283

Matrix of equivalence (Continued)

Lab, S/N i	Lab, S/N j \longrightarrow											
	D_i / mK		U_i / mK		D_{ij} / mK		U_{ij} / mK		D_{ij} / mK		U_{ij} / mK	
NIST A128	0.074	0.413			0.572	0.667	0.217	0.667	-0.721	1.166	-1.028	1.218
NIST A129	-0.089	0.413			0.409	0.667	0.054	0.667	-0.884	1.166	-1.191	1.218
NMi-VSL 226246	0.286	0.526			0.785	0.742	0.430	0.742	-0.508	1.211	-0.815	1.260
NPL 221481	0.240	0.578			0.738	0.780	0.383	0.780	-0.555	1.234	-0.862	1.283
NPL 221483	0.190	0.578			0.688	0.780	0.333	0.780	-0.605	1.234	-0.912	1.283
NPL 221485	-0.468	0.578			0.031	0.780	-0.324	0.780	-1.262	1.234	-1.569	1.283
PTB 229074	-0.498	0.523					-0.355	0.740	-1.293	1.210	-1.600	1.259
PTB 229075	-0.143	0.523			0.355	0.740			-0.938	1.210	-1.245	1.259
VNIIFTRI 79	0.795	1.090			1.293	1.210	0.938	1.210			-0.307	1.582
VNIIFTRI 89	1.102	1.145			1.6001	1.2593	1.245	1.259	0.307	1.582		

CCT-K1 : Nominal temperature, $T_{90} = 0.762$ K
Degrees of equivalence, D_i , and expanded uncertainties ($k = 2$), U_i , expressed in mK



Key comparison CCT-K1

MEASURAND : Temperature T_{90} / K

NOMINAL TEMPERATURE : $T_{90} = 0.858$ K

Laboratory individual measurements

Lab: Laboratory

S/N i : serial number of Thermometer i

T_i : temperature value for Thermometer i

u_i : standard uncertainty in the calibration of Thermometer i

u_{comp} : standard uncertainty in the comparison measurements

$$u_{\text{comp}} = 0.177 \text{ mK}$$

Lab, S/N i	T_i / mK	u_i / mK
NIST A128	0.858267	0.076
NIST A129	0.858267	0.076
NMi-VSL 226246	0.858603	0.182
NPL 221481	0.858740	0.186
NPL 221483	0.858680	0.186
NPL 221485	0.858267	0.186
PTB 229074	0.858122	0.180
PTB 229075	0.858288	0.180
VNIIFTRI 79	0.859183	0.486
VNIIFTRI 89	0.859376	0.532

Key comparison CCT-K1

MEASURAND : Temperature T_{90} / K

KEY COMPARISON REFERENCE VALUE: $T_R = 0.858421$ K


Matrix of equivalence

Lab, S/N i	Lab, S/N j \longrightarrow					
	D_i / mK	U_i / mK	D_{ij} / mK	U_{ij} / mK	D_{ij} / mK	U_{ij} / mK
NIST A128	-0.154	0.384		0.000	0.544	-0.336
NIST A129	-0.154	0.384		-0.336	0.636	-0.473
NMi-VSL 226246	0.182	0.507	0.336	0.636	-0.473	0.642
NPL 221481	0.319	0.514	0.473	0.642	0.137	0.722
NPL 221483	0.259	0.514	0.413	0.642	0.077	0.722
NPL 221485	-0.154	0.514	0.000	0.642	0.060	0.727
PTB 229074	-0.299	0.505	-0.145	0.634	0.473	0.727
PTB 229075	-0.133	0.505	0.021	0.634	-0.413	0.727
VNIIFTRI 79	0.762	1.035	0.917	1.104	-0.145	0.720
VNIIFTRI 89	0.955	1.121	1.104	0.917	0.021	0.720
	1.109	1.185	1.109	1.185	0.916	1.155
			0.773	1.231	0.636	1.233
					0.696	1.233
					1.109	1.233

Matrix of equivalence (Continued)

Lab, S/N i	Lab, S/N j \longrightarrow					
	D_i / mK	U_i / mK	D_{ij} / mK	U_{ij} / mK	D_{ij} / mK	U_{ij} / mK
NIST A128	-0.154	0.384	0.145	0.634	-0.021	0.634
NIST A129	-0.154	0.384	0.145	0.634	-0.917	1.104
NMi-VSL 226246	0.182	0.507	0.481	0.715	0.315	0.715
NPL 221481	0.319	0.514	0.619	0.720	0.452	1.155
NPL 221483	0.259	0.514	0.559	0.720	0.503	1.155
NPL 221485	-0.154	0.514	0.145	0.720	-0.021	0.720
PTB 229074	-0.299	0.505			-0.166	0.714
PTB 229075	-0.133	0.505			-1.062	1.151
VNIIFTRI 79	0.762	1.035			-0.895	1.151
VNIIFTRI 89	0.955	1.121			-1.088	1.230
	1.2545	1.2296	0.895	1.151	-0.193	1.526
			0.193	1.526		

CCT-K1 : Nominal temperature, $T_{90} = 0.858$ K
Degrees of equivalence, D_i , and expanded uncertainties ($k = 2$), U_i , expressed in mK



Key comparison CCT-K1

MEASURAND : Temperature T_{90} / K

NOMINAL TEMPERATURE : $T_{90} = 0.991$ K

Laboratory individual measurements

Lab: Laboratory

S/N i : serial number of Thermometer i

T_i : temperature value for Thermometer i

u_i : standard uncertainty in the calibration of Thermometer i

u_{comp} : standard uncertainty in the comparison measurements

$$u_{\text{comp}} = 0.159 \text{ mK}$$

Lab, S/N i	T_i / mK	u_i / mK
NIST A128	0.991179	0.078
NIST A129	0.991012	0.078
NMi-VSL 226246	0.991361	0.182
NPL 221481	0.991590	0.144
NPL 221485	0.991375	0.144
PTB 229074	0.991133	0.180
PTB 229075	0.990995	0.180

Key comparison CCT-K1


MEASURAND : Temperature T_{90} / K

KEY COMPARISON REFERENCE VALUE: $T_R = 0.991223$ K

Matrix of equivalence

Lab, S/N i	Lab, S/N j \longrightarrow															
	NIST A128		NIST A129		NMi-VSL 226246		NPL 221481		NPL 221485		PTB 229074		PTB 229075			
	D_i	U_i	D_{ij}	U_{ij}	D_{ij}	U_{ij}	D_{ij}	U_{ij}	D_{ij}	U_{ij}	D_{ij}	U_{ij}	D_{ij}	U_{ij}		
NIST A128	-0.044	0.354			0.168	0.500	-0.182	0.599	-0.411	0.556	-0.196	0.556	0.046	0.597	0.184	0.597
NIST A129	-0.211	0.354	-0.168	0.500			-0.349	0.599	-0.578	0.556	-0.363	0.556	-0.121	0.597	0.016	0.597
NMi-VSL 226246	0.138	0.483	0.182	0.599	0.349	0.599			-0.229	0.646	-0.014	0.646	0.228	0.681	0.366	0.681
NPL 221481	0.367	0.429	0.411	0.556	0.578	0.556	0.229	0.646			0.215	0.607	0.457	0.644	0.595	0.644
NPL 221485	0.152	0.429	0.196	0.556	0.363	0.556	0.014	0.646	-0.215	0.607			0.242	0.644	0.380	0.644
PTB 229074	-0.090	0.480	-0.046	0.597	0.121	0.597	-0.228	0.681	-0.457	0.644	-0.242	0.644			0.138	0.679
PTB 229075	-0.228	0.480	-0.184	0.597	-0.016	0.597	-0.366	0.681	-0.595	0.644	-0.380	0.644	-0.138	0.679		

CCT-K1 : Nominal temperature, $T_{90} = 0.991$ K
Degrees of equivalence, D_i , and expanded uncertainties ($k = 2$), U_i , expressed in mK



Key comparison CCT-K1

MEASURAND : Temperature T_{90} / K

NOMINAL TEMPERATURE : $T_{90} = 1.032$ K

Laboratory individual measurements

Lab: Laboratory

S/N i : serial number of Thermometer i

T_i : temperature value for Thermometer i

u_i : standard uncertainty in the calibration of Thermometer i

u_{comp} : standard uncertainty in the comparison measurements

$$u_{\text{comp}} = 0.157 \text{ mK}$$

Lab, S/N i	T_i / mK	u_i / mK
NIST A128	1.031395	0.078
NIST A129	1.031369	0.078
NMi-VSL 226246	1.031678	0.182
NPL 221481	1.031835	0.145
NPL 221483	1.031765	0.145
NPL 221485	1.031717	0.145
PTB 229074	1.031415	0.180
PTB 229075	1.031397	0.180
VNIIFTRI 79	1.032281	0.454
VNIIFTRI 89	1.032306	0.526

Key comparison CCT-K1

MEASURAND : Temperature T_{90} / K

KEY COMPARISON REFERENCE VALUE: $T_R = 1.031584$ K


Matrix of equivalence

Lab, S/N i	Lab, S/N j \longrightarrow					
	D_i	U_i	D_{ij}	U_{ij}	D_{ij}	U_{ij}
	/ mK	/ mK	/ mK	/ mK	/ mK	/ mK
NIST A128	-0.189	0.351	0.026	0.496	-0.283	0.595
NIST A129	-0.215	0.351	-0.026	0.496	-0.310	0.595
NMi-VSL 226246	0.094	0.481	0.283	0.595	0.467	0.553
NPL 221481	0.251	0.428	0.440	0.553	0.157	0.644
NPL 221483	0.181	0.428	0.370	0.553	-0.070	0.605
NPL 221485	0.133	0.428	0.322	0.553	-0.118	0.605
PTB 229074	-0.169	0.478	0.021	0.593	-0.263	0.678
PTB 229075	-0.187	0.478	0.002	0.593	-0.281	0.678
VNIIFTRI 79	0.697	0.961	0.886	1.023	0.603	1.074
VNIIFTRI 89	0.722	1.099	0.911	1.153	0.937	1.153

Matrix of equivalence (Continued)

Lab, S/N i	Lab, S/N j \longrightarrow					
	D_i	U_i	D_{ij}	U_{ij}	D_{ij}	U_{ij}
	/ mK	/ mK	/ mK	/ mK	/ mK	/ mK
NIST A128	-0.189	0.351	-0.021	0.593	-0.002	0.593
NIST A129	-0.215	0.351	-0.047	0.593	-0.028	0.593
NMi-VSL 226246	0.094	0.481	0.263	0.678	-0.913	1.023
NPL 221481	0.251	0.428	0.420	0.642	-0.446	1.052
NPL 221483	0.181	0.428	0.350	0.642	-0.516	1.052
NPL 221485	0.133	0.428	0.302	0.642	-0.564	1.052
PTB 229074	-0.169	0.478	0.018	0.676	-0.866	1.073
PTB 229075	-0.187	0.478	-0.018	0.676	-0.884	1.073
VNIIFTRI 79	0.697	0.961	0.866	1.073	-0.909	1.198
VNIIFTRI 89	0.722	1.099	0.8904	1.1983	0.909	1.198

CCT-K1 : Nominal temperature, $T_{90} = 1.032$ K
Degrees of equivalence, D_i , and expanded uncertainties ($k = 2$), U_i , expressed in mK



Key comparison CCT-K1

MEASURAND : Temperature T_{90} / K

NOMINAL TEMPERATURE : $T_{90} = 1.225$ K

Laboratory individual measurements

Lab: Laboratory

S/N i : serial number of Thermometer i

T_i : temperature value for Thermometer i

u_i : standard uncertainty in the calibration of Thermometer i

u_{comp} : standard uncertainty in the comparison measurements

$$u_{\text{comp}} = 0.149 \text{ mK}$$

Lab, S/N i	T_i / mK	u_i / mK
NIST A128	1.224814	0.065
NIST A129	1.224828	0.065
NMi-VSL 226246	1.225123	0.256
NPL 221481	1.225340	0.151
NPL 221485	1.225201	0.151
PTB 229074	1.224998	0.180
PTB 229075	1.224931	0.180

Key comparison CCT-K1


MEASURAND : Temperature T_{90} / K

KEY COMPARISON REFERENCE VALUE: $T_R = 1.224991$ K

Matrix of equivalence

Lab, S/N i	Lab, S/N j \longrightarrow															
	NIST A128		NIST A129		NMi-VSL 226246		NPL 221481		NPL 221485		PTB 229074		PTB 229075			
	D_i	U_i	D_{ij}	U_{ij}	D_{ij}	U_{ij}	D_{ij}	U_{ij}	D_{ij}	U_{ij}	D_{ij}	U_{ij}	D_{ij}	U_{ij}		
NIST A128	-0.177	0.325			-0.014	0.459	-0.309	0.675	-0.526	0.534	-0.388	0.534	-0.185	0.569	-0.118	0.569
NIST A129	-0.163	0.325	0.014	0.459			-0.295	0.675	-0.512	0.534	-0.373	0.534	-0.171	0.569	-0.103	0.569
NMi-VSL 226246	0.132	0.592	0.309	0.675	0.295	0.675			-0.217	0.729	-0.078	0.729	0.124	0.754	0.191	0.754
NPL 221481	0.349	0.425	0.526	0.534	0.512	0.534	0.217	0.729			0.139	0.600	0.341	0.631	0.408	0.631
NPL 221485	0.210	0.425	0.388	0.534	0.373	0.534	0.078	0.729	-0.139	0.600			0.203	0.631	0.270	0.631
PTB 229074	0.007	0.467	0.185	0.569	0.171	0.569	-0.124	0.754	-0.341	0.631	-0.203	0.631			0.067	0.661
PTB 229075	-0.060	0.467	0.118	0.569	0.103	0.569	-0.191	0.754	-0.408	0.631	-0.270	0.631	-0.067	0.661		

CCT-K1 : Nominal temperature, $T_{90} = 1.225$ K
Degrees of equivalence, D_i , and expanded uncertainties ($k = 2$), U_i , expressed in mK



Key comparison CCT-K1

MEASURAND : Temperature T_{90} / K

NOMINAL TEMPERATURE : $T_{90} = 1.250$ K

Laboratory individual measurements

Lab: Laboratory

S/N i : serial number of Thermometer i

T_i : temperature value for Thermometer i

u_i : standard uncertainty in the calibration of Thermometer i

u_{comp} : standard uncertainty in the comparison measurements

$$u_{\text{comp}} = 0.148 \text{ mK}$$

Lab, S/N i	T_i / mK	u_i / mK
NIST A128	1.249254	0.065
NIST A129	1.249283	0.065
NMi-VSL 226246	1.249645	0.256
NPL 221481	1.250010	0.152
NPL 221483	1.249780	0.152
NPL 221485	1.249814	0.152
PTB 229074	1.249406	0.180
PTB 229075	1.249394	0.180
VNIIFTRI 79	1.249940	0.480
VNIIFTRI 89	1.250150	0.570

Key comparison CCT-K1

MEASURAND : Temperature T_{90} / K

KEY COMPARISON REFERENCE VALUE: $T_R = 1.249542$ K


Matrix of equivalence

Lab, S/N i	Lab, S/N j \longrightarrow					
	D_i	U_i	D_{ij}	U_{ij}	D_{ij}	U_{ij}
	/ mK	/ mK	/ mK	/ mK	/ mK	/ mK
NIST A128	-0.288	0.323				
NIST A129	-0.259	0.323				
NMi-VSL 226246	0.103	0.591				
NPL 221481	0.468	0.424				
NPL 221483	0.238	0.424				
NPL 221485	0.272	0.424				
PTB 229074	-0.136	0.466				
PTB 229075	-0.148	0.466				
VNIIFTRI 79	0.398	1.004				
VNIIFTRI 89	0.608	1.178				
NIST A128	-0.288	0.323	-0.029	0.456	-0.390	0.673
NIST A129	-0.259	0.323	0.029	0.456	-0.361	0.673
NMi-VSL 226246	0.103	0.591	0.390	0.673	-0.726	0.533
NPL 221481	0.468	0.424	0.755	0.533	-0.525	0.533
NPL 221483	0.238	0.424	0.525	0.533	-0.496	0.533
NPL 221485	0.272	0.424	0.559	0.533	-0.530	0.533
PTB 229074	-0.136	0.466	0.152	0.567	0.123	0.567
PTB 229075	-0.148	0.466	0.140	0.567	0.111	0.567
VNIIFTRI 79	0.398	1.004	0.686	1.055	0.296	1.166
VNIIFTRI 89	0.608	1.178	0.896	1.221	0.657	1.221
NIST A128	-0.288	0.323	-0.152	0.567	-0.140	0.567
NIST A129	-0.259	0.323	-0.123	0.567	-0.111	0.567
NMi-VSL 226246	0.103	0.591	0.239	0.753	0.250	0.753
NPL 221481	0.468	0.424	0.604	0.630	0.615	0.630
NPL 221483	0.238	0.424	0.374	0.630	0.385	0.630
NPL 221485	0.272	0.424	0.408	0.630	0.419	0.630
PTB 229074	-0.136	0.466	0.012	0.659	-0.534	1.107
PTB 229075	-0.148	0.466	-0.012	0.659	-0.546	1.107
VNIIFTRI 79	0.398	1.004	0.534	1.107	-0.210	1.548
VNIIFTRI 89	0.608	1.178	0.744	1.2665	0.756	1.267

Matrix of equivalence (Continued)

Lab, S/N i	Lab, S/N j \longrightarrow					
	D_i	U_i	D_{ij}	U_{ij}	D_{ij}	U_{ij}
	/ mK	/ mK	/ mK	/ mK	/ mK	/ mK
NIST A128	-0.288	0.323	-0.152	0.567	-0.140	0.567
NIST A129	-0.259	0.323	-0.123	0.567	-0.111	0.567
NMi-VSL 226246	0.103	0.591	0.239	0.753	0.250	0.753
NPL 221481	0.468	0.424	0.604	0.630	0.615	0.630
NPL 221483	0.238	0.424	0.374	0.630	0.385	0.630
NPL 221485	0.272	0.424	0.408	0.630	0.419	0.630
PTB 229074	-0.136	0.466	0.012	0.659	-0.534	1.107
PTB 229075	-0.148	0.466	-0.012	0.659	-0.546	1.107
VNIIFTRI 79	0.398	1.004	0.534	1.107	-0.210	1.548
VNIIFTRI 89	0.608	1.178	0.744	1.2665	0.756	1.267
NIST A128	-0.288	0.323	-0.152	0.567	-0.140	0.567
NIST A129	-0.259	0.323	-0.123	0.567	-0.111	0.567
NMi-VSL 226246	0.103	0.591	0.239	0.753	0.250	0.753
NPL 221481	0.468	0.424	0.604	0.630	0.615	0.630
NPL 221483	0.238	0.424	0.374	0.630	0.385	0.630
NPL 221485	0.272	0.424	0.408	0.630	0.419	0.630
PTB 229074	-0.136	0.466	0.012	0.659	-0.534	1.107
PTB 229075	-0.148	0.466	-0.012	0.659	-0.546	1.107
VNIIFTRI 79	0.398	1.004	0.534	1.107	-0.210	1.548
VNIIFTRI 89	0.608	1.178	0.744	1.2665	0.756	1.267

CCT-K1 : Nominal temperature, $T_{90} = 1.250$ K
Degrees of equivalence, D_i , and expanded uncertainties ($k = 2$), U_i , expressed in mK



Key comparison CCT-K1

MEASURAND : Temperature T_{90} / K

NOMINAL TEMPERATURE : $T_{90} = 1.503$ K

Laboratory individual measurements

Lab: Laboratory

S/N i : serial number of Thermometer i

T_i : temperature value for Thermometer i

u_i : standard uncertainty in the calibration of Thermometer i

u_{comp} : standard uncertainty in the comparison measurements

$$u_{\text{comp}} = 0.137 \text{ mK}$$

Lab, S/N i	T_i / mK	u_i / mK
NIST A128	1.503103	0.062
NIST A129	1.503008	0.062
NMi-VSL 226246	1.503751	0.256
NPL 221481	1.503610	0.160
NPL 221483	1.503650	0.160
NPL 221485	1.503561	0.160
NRC A138	1.503715	0.225
NRC A140	1.503918	0.225
PTB 229074	1.503155	0.180
PTB 229075	1.503100	0.180
VNIIFTRI 79	1.503876	0.510
VNIIFTRI 89	1.504156	0.620

Key comparison CCT-K1

MEASURAND : Temperature T_{90} / K

KEY COMPARISON REFERENCE VALUE: $T_R = 1.503370$ K


Matrix of equivalence

Lab, S/N i	Lab, S/N j \longrightarrow					
	D_i	U_i / mK	D_j	U_j / mK	D_j	U_j / mK
NIST A128	-0.267	0.301	0.095	0.426	-0.507	0.518
NIST A129	-0.362	0.301	-0.426	0.654	-0.602	0.518
NMi-VSL 226246	0.381	0.581	0.648	0.654	0.141	0.717
NPL 221481	0.240	0.421	0.507	0.518	0.101	0.717
NPL 221483	0.280	0.421	0.547	0.518	0.050	0.596
NPL 221485	0.191	0.421	0.457	0.518	0.090	0.596
NRC A138	0.345	0.527	0.611	0.607	0.104	0.675
NRC A140	0.548	0.527	0.815	0.607	0.308	0.675
PTB 229074	-0.215	0.452	0.052	0.543	0.268	0.675
PTB 229075	-0.270	0.452	-0.004	0.543	0.455	0.618
VNIIFTRI 79	0.506	1.055	0.773	1.098	0.511	0.618
VNIIFTRI 89	0.786	1.269	1.052	1.304	0.266	1.136
		1.147	1.304	0.405	1.396	0.505
				0.545	1.337	0.595
					1.337	
						1.337

Matrix of equivalence (Continued)

Lab, S/N i	Lab, S/N j \longrightarrow					
	D_i	U_i / mK	D_j	U_j / mK	D_j	U_j / mK
NIST A128	-0.267	0.301	-0.611	0.607	-0.815	0.607
NIST A129	-0.362	0.301	-0.706	0.607	-0.910	0.607
NMi-VSL 226246	0.381	0.581	0.036	0.784	-0.167	0.784
NPL 221481	0.240	0.421	-0.104	0.675	-0.308	0.675
NPL 221483	0.280	0.421	-0.064	0.675	-0.268	0.675
NPL 221485	0.191	0.421	-0.154	0.675	-0.357	0.675
NRC A138	0.345	0.527	0.2032	0.7451	0.559	0.694
NRC A140	0.548	0.527	-0.559	0.694	0.763	0.694
PTB 229074	-0.215	0.452	-0.615	0.694	-0.818	0.694
PTB 229075	-0.270	0.452	0.161	1.180	-0.042	1.180
VNIIFTRI 79	0.506	1.055	0.441	1.374	0.238	1.374
VNIIFTRI 89	0.786	1.269				

CCT-K1 : Nominal temperature, $T_{90} = 1.503$ K
Degrees of equivalence, D_i , and expanded uncertainties ($k = 2$), U_i , expressed in mK



Key comparison CCT-K1

MEASURAND : Temperature T_{90} / K

NOMINAL TEMPERATURE : $T_{90} = 1.755$ K

Laboratory individual measurements

Lab: Laboratory

S/N i : serial number of Thermometer i

T_i : temperature value for Thermometer i

u_i : standard uncertainty in the calibration of Thermometer i

u_{comp} : standard uncertainty in the comparison measurements

$$u_{\text{comp}} = 0.132 \text{ mK}$$

Lab, S/N i	T_i / mK	u_i / mK
NIST A128	1.754643	0.038
NIST A129	1.754505	0.038
NMi-VSL 226246	1.755116	0.256
NPL 221481	1.755060	0.157
NPL 221483	1.755070	0.157
NPL 221485	1.754965	0.157
NRC A138	1.755172	0.225
NRC A140	1.755291	0.225
PTB 229074	1.754594	0.180
PTB 229075	1.754562	0.180
VNIIFTRI 79	1.755577	0.484
VNIIFTRI 89	1.755761	0.592

Key comparison CCT-K1

MEASURAND : Temperature T_{90} / K

KEY COMPARISON REFERENCE VALUE: $T_R = 1.754822$ K


Matrix of equivalence

Lab, S/N i	Lab, S/N j \longrightarrow					
	D_i	U_i	$/ \text{mK}$	D_j	U_j	$/ \text{mK}$
NIST A128	-0.179	0.275		0.138	0.388	
NIST A129	-0.317	0.275		-0.473	0.638	
NMi-VSL 226246	0.294	0.576		-0.556	0.493	
NPL 221481	0.238	0.410		0.055	0.707	
NPL 221483	0.248	0.410		0.010	0.579	
NPL 221485	0.143	0.410		-0.095	0.579	
NRC A138	0.350	0.522		0.112	0.663	
NRC A140	0.469	0.522		0.207	0.663	
PTB 229074	-0.228	0.446		0.231	0.663	
PTB 229075	-0.260	0.446		0.466	0.606	
VNIIFTRI 79	0.755	1.004		-0.476	0.606	
VNIIFTRI 89	0.939	1.213		-0.371	0.606	
	0.934	1.041	1.073	1.041	0.461	1.158
	1.118	1.243	1.256	1.243	0.517	1.085
				0.701	1.280	0.691
					0.796	1.280

Matrix of equivalence (Continued)

Lab, S/N i	Lab, S/N j \longrightarrow					
	D_i	U_i	$/ \text{mK}$	D_j	U_j	$/ \text{mK}$
NIST A128	-0.179	0.275		-0.649	0.590	
NIST A129	-0.317	0.275		0.049	0.524	
NMi-VSL 226246	0.294	0.576		0.081	0.524	
NPL 221481	0.238	0.410		-0.934	1.041	
NPL 221483	0.248	0.410		-1.118	1.243	
NPL 221485	0.143	0.410		-0.057	0.524	
NRC A138	0.350	0.522		-1.073	1.041	
NRC A140	0.469	0.522		-1.256	1.243	
PTB 229074	-0.228	0.446		0.554	0.729	
PTB 229075	-0.260	0.446		-0.461	1.158	
VNIIFTRI 79	0.755	1.004		-0.645	1.343	
VNIIFTRI 89	0.939	1.213		-0.701	1.280	
	-0.207	0.663	-0.326	0.663	0.371	0.606
	0.1192	0.7377		0.697	0.686	0.403
				0.730	0.686	0.610
				-0.286	1.132	0.686
				-0.470	1.320	0.495
				-0.470	1.320	1.132
				-0.578	0.686	0.697
				-0.033	0.631	0.686
				-0.983	1.099	0.986
				-1.167	1.292	1.099
				-1.015	1.099	1.099
				-1.199	1.292	1.099
				-0.184	1.575	1.099
				0.589	1.320	0.470
				1.167	1.320	1.292
				1.199	1.292	1.099
				0.1838	1.5745	1.099

CCT-K1 : Nominal temperature, $T_{90} = 1.755$ K
Degrees of equivalence, D_i , and expanded uncertainties ($k = 2$), U_i , expressed in mK



Key comparison CCT-K1

MEASURAND : Temperature T_{90} / K

NOMINAL TEMPERATURE : $T_{90} = 1.997$ K

Laboratory individual measurements

Lab: Laboratory

S/N i : serial number of Thermometer i

T_i : temperature value for Thermometer i

u_i : standard uncertainty in the calibration of Thermometer i

u_{comp} : standard uncertainty in the comparison measurements

$$u_{\text{comp}} = 0.127 \text{ mK}$$

Lab, S/N i	T_i / mK	u_i / mK
NIST A128	1.996327	0.039
NIST A129	1.996288	0.039
NMi-VSL 226246	1.996820	0.256
NPL 221481	1.996780	0.154
NPL 221483	1.996840	0.154
NPL 221485	1.996757	0.154
NRC A138	1.996922	0.225
NRC A140	1.996943	0.225
PTB 229074	1.996278	0.180
PTB 229075	1.996315	0.180
VNIIFTRI 79	1.997300	0.460
VNIIFTRI 89	1.997462	0.565

Key comparison CCT-K1

MEASURAND : Temperature T_{90} / K

KEY COMPARISON REFERENCE VALUE: $T_R = 1.996554$ K


Matrix of equivalence

Lab, S/N <i>i</i>	Lab, S/N <i>j</i> \longrightarrow					
	D_i	U_i	$/ \text{mK}$	D_j	U_j	$/ \text{mK}$
NIST A128	-0.227	0.266		0.040	0.376	-0.493
NIST A129	-0.266	0.266		-0.493	0.630	-0.453
NMi-VSL 226246	0.266	0.572		0.376	-0.532	-0.492
NPL 221481	0.226	0.399		-0.493	0.630	-0.479
NPL 221483	0.286	0.399		0.492	0.479	-0.552
NPL 221485	0.203	0.399		-0.040	0.697	0.479
NRC A138	0.368	0.517		0.479	-0.020	-0.470
NRC A140	0.389	0.517		0.513	0.564	0.479
PTB 229074	-0.276	0.441		0.492	0.479	0.564
PTB 229075	-0.239	0.441		0.203	0.564	0.564
VNIIFTRI 79	0.746	0.955		0.226	0.399	0.564
VNIIFTRI 89	0.908	1.159		0.286	0.399	0.564
	1.135	1.189	1.174	1.189	0.642	1.292
					0.682	1.225
					0.622	1.225
					0.705	1.225

Matrix of equivalence (Continued)

Lab, S/N <i>i</i>	Lab, S/N <i>j</i> \longrightarrow					
	D_i	U_i	$/ \text{mK}$	D_j	U_j	$/ \text{mK}$
NIST A128	-0.227	0.266		-0.615	0.581	0.049
NIST A129	-0.266	0.266		-0.655	0.581	0.010
NMi-VSL 226246	0.266	0.572		-0.123	0.771	0.542
NPL 221481	0.226	0.399		0.771	0.515	0.012
NPL 221483	0.286	0.399		0.142	0.653	0.515
NPL 221485	0.203	0.399		0.163	0.594	-0.973
NRC A138	0.368	0.517		0.465	0.594	0.991
NRC A140	0.389	0.517		0.525	0.594	-1.135
PTB 229074	-0.276	0.441		0.442	0.594	1.189
PTB 229075	-0.239	0.441		0.543	1.035	-0.642
VNIIFTRI 79	0.746	0.955		0.705	1.225	1.292
VNIIFTRI 89	0.908	1.159		0.620	1.269	1.225
	0.0209	0.7309		0.665	0.679	0.628
				0.679	0.679	0.679
				-0.357	1.086	-0.519
				1.086	-0.519	1.269
				-0.378	1.086	-0.540
				1.086	-0.540	1.269
				-0.607	0.679	0.679
				0.679	0.679	-0.644
				-0.644	0.679	0.665
				0.679	0.679	0.679
				-0.037	0.623	-0.037
				0.623	-1.022	0.623
				-1.022	1.052	-1.022
				1.052	-1.184	1.052
				-1.184	1.240	-1.184
				1.240	-1.147	1.240
				-1.147	1.240	-1.147
				1.240	-0.162	1.502
				-0.162	1.502	-0.162
				1.502		

CCT-K1 : Nominal temperature, $T_{90} = 1.997$ K
Degrees of equivalence, D_i , and expanded uncertainties ($k = 2$), U_i , expressed in mK



Key comparison CCT-K1

MEASURAND : Temperature T_{90} / K

NOMINAL TEMPERATURE : $T_{90} = 2.248$ K

Laboratory individual measurements

Lab: Laboratory

S/N i : serial number of Thermometer i

T_i : temperature value for Thermometer i

u_i : standard uncertainty in the calibration of Thermometer i

u_{comp} : standard uncertainty in the comparison measurements

$$u_{\text{comp}} = 0.122 \text{ mK}$$

Lab, S/N i	T_i / mK	u_i / mK
NIST A128	2.248374	0.039
NIST A129	2.248385	0.039
NMi-VSL 226246	2.248743	0.225
NPL 221481	2.248710	0.150
NPL 221483	2.248750	0.150
NPL 221485	2.248685	0.150
NRC A138	2.248713	0.225
NRC A140	2.248734	0.225
PTB 229074	2.248321	0.180
PTB 229075	2.248306	0.180
VNIIFTRI 79	2.249248	0.435
VNIIFTRI 89	2.249363	0.538
NMIJ B271	2.248277	0.095
NMIJ B310	2.248308	0.094

Key comparison CCT-K1

MEASURAND : Temperature T_{90} / K

KEY COMPARISON REFERENCE VALUE: $T_R = 2.248485$ K

Matrix of equivalence

Lab, S/N <i>i</i>	Lab, S/N <i>j</i> \longrightarrow															
	NIST A128		NIST A129		NMi-VSL 226246		NPL 221481		NPL 221483		NPL 221485		NRC A138			
	D_i	U_i		D_i	U_i		D_i	U_i		D_i	U_i		D_i	U_i		
NIST A128	-0.111	0.256		-0.011	0.363	-0.368	0.573	-0.336	0.465	-0.376	0.465	-0.311	0.465	-0.339	0.573	
NIST A129	-0.100	0.256	0.011	0.363		-0.358	0.573	-0.325	0.465	-0.365	0.465	-0.300	0.465	-0.328	0.573	
NMI-VSL 226246	0.258	0.512	0.368	0.573	0.358		0.033	0.642	-0.007	0.642	0.057	0.642	0.029	0.724		
NPL 221481	0.225	0.387	0.336	0.465	0.325	0.465	-0.033	0.642		-0.040	0.548	0.024	0.548	-0.003	0.642	
NPL 221483	0.265	0.387	0.376	0.465	0.365	0.465	0.007	0.642	0.040	0.548		0.064	0.548	0.037	0.642	
NPL 221485	0.200	0.387	0.311	0.465	0.300	0.465	-0.057	0.642	-0.024	0.548	-0.064	0.548		-0.028	0.642	
NRC A138	0.228	0.512	0.339	0.573	0.328	0.573	-0.029	0.724	0.003	0.642	-0.037	0.642	0.028	0.642		
NRC A140	0.249	0.512	0.360	0.573	0.349	0.573	-0.009	0.724	0.024	0.642	-0.016	0.642	0.049	0.642	0.021	0.724
PTB 229074	-0.164	0.435	-0.053	0.505	-0.064	0.505	-0.421	0.672	-0.388	0.582	-0.428	0.582	-0.364	0.582	-0.392	0.672
PTB 229075	-0.179	0.435	-0.068	0.505	-0.079	0.505	-0.437	0.672	-0.404	0.582	-0.444	0.582	-0.380	0.582	-0.407	0.672
VNIIFTRI 79	0.763	0.904	0.873	0.939	0.863	0.939	0.505	1.039	0.538	0.983	0.498	0.983	0.562	0.983	0.534	1.039
VNIIFTRI 89	0.878	1.102	0.989	1.132	0.978	1.132	0.621	1.216	0.654	1.169	0.614	1.169	0.678	1.169	0.650	1.216
NMIJ B271	-0.208	0.309	-0.097	0.402	-0.108	0.402	-0.465	0.598	-0.432	0.496	-0.472	0.496	-0.408	0.496	-0.436	0.598
NMIJ B310	-0.177	0.308	-0.066	0.401	-0.077	0.401	-0.435	0.597	-0.402	0.495	-0.442	0.495	-0.378	0.495	-0.405	0.597


MEASURAND : Temperature T_{90} / K

KEY COMPARISON REFERENCE VALUE: $T_R = 2.248485$ K

Matrix of equivalence (Continued)

Lab, S/N <i>i</i>	Lab, S/N <i>j</i> \longrightarrow															
	NRC A140		PTB 229074		PTB 229075		VNIIFTRI 79		VNIIFTRI 89		NMJJ B271		NMJJ B310			
	D_i	U_i		D_i	U_i		D_i	U_i		D_i	U_i		D_i	U_i		
NIST A128	-0.111	0.256	-0.360	0.573	0.053	0.505	0.068	0.505	-0.873	0.939	-0.989	1.132	0.097	0.402	0.066	0.401
NIST A129	-0.100	0.256	-0.349	0.573	0.064	0.505	0.079	0.505	-0.863	0.939	-0.978	1.132	0.108	0.402	0.077	0.401
NMI-VSL 226246	0.258	0.512	0.009	0.724	0.421	0.672	0.437	0.672	-0.505	1.039	-0.621	1.216	0.465	0.598	0.435	0.597
NPL 221481	0.225	0.387	-0.024	0.642	0.388	0.582	0.404	0.582	-0.538	0.983	-0.654	1.169	0.432	0.496	0.402	0.495
NPL 221483	0.265	0.387	0.016	0.642	0.428	0.582	0.444	0.582	-0.498	0.983	-0.614	1.169	0.472	0.496	0.442	0.495
NPL 221485	0.200	0.387	-0.049	0.642	0.364	0.582	0.380	0.582	-0.562	0.983	-0.678	1.169	0.408	0.496	0.378	0.495
NRC A138	0.228	0.512	-0.021	0.724	0.392	0.672	0.407	0.672	-0.534	1.039	-0.650	1.216	0.436	0.598	0.405	0.597
NRC A140	0.249	0.512			0.413	0.672	0.428	0.672	-0.514	1.039	-0.629	1.216	0.457	0.598	0.426	0.597
PTB 229074	-0.164	0.435	-0.413	0.672			0.016	0.615	-0.926	1.003	-1.042	1.185	0.044	0.534	0.014	0.533
PTB 229075	-0.179	0.435	-0.428	0.672	-0.016	0.615			-0.942	1.003	-1.058	1.185	0.028	0.534	-0.002	0.533
VNIIFTRI 79	0.763	0.904	0.514	1.039	0.926	1.003	0.942	1.003			-0.116	1.426	0.970	0.955	0.940	0.955
VNIIFTRI 89	0.878	1.102	0.629	1.216	1.042	1.185	1.058	1.185	0.116	1.426	0.970	0.955			1.056	1.145
NMJJ B271	-0.208	0.309	-0.457	0.598	-0.044	0.534	-0.028	0.534	-0.970	0.955	-1.086	1.145		-0.030	0.436	
NMJJ B310	-0.177	0.308	-0.426	0.597	-0.014	0.533	0.002	0.533	-0.940	0.955	-1.056	1.145	0.030	0.436		

CCT-K1 : Nominal temperature, $T_{90} = 2.248$ K
Degrees of equivalence, D_i , and expanded uncertainties ($k = 2$), U_i , expressed in mK



Key comparison CCT-K1 and EURAMET.T-K1

MEASURAND : Temperature T_{90} / K

NOMINAL TEMPERATURE : $T_{90} = 2.601$ K

CCT-K1

Laboratory individual measurements

Lab: Laboratory

S/N i : serial number of Thermometer i

T_i : temperature value for Thermometer i

u_i : standard uncertainty in the calibration of Thermometer i

u_{comp} : standard uncertainty in the comparison measurements

$$u_{\text{comp}} = 0.115 \text{ mK}$$

Lab, S/N i	T_i / mK	u_i / mK
NIST A128	2.600642	0.040
NIST A129	2.600648	0.040
NMi-VSL 226246	2.600862	0.181
NPL 221481	2.600827	0.146
NPL 221483	2.600842	0.146
NPL 221485	2.600863	0.146
NRC A138	2.600854	0.225
NRC A140	2.600896	0.225
PTB 229074	2.600646	0.180
PTB 229075	2.600798	0.180
VNIIFTRI 79	2.601411	0.406
VNIIFTRI 89	2.601495	0.498
NMIJ B271	2.600793	0.084
NMIJ B310	2.600816	0.082

Key comparison CCT-K1 and EURAMET.T-K1


MEASURAND : Temperature T_{90} / K

KEY COMPARISON REFERENCE VALUE: $T_R = 2.600776$ K

Matrix of equivalence

Lab, S/N i	D_i	U_i / mK
NIST A128	-0.134	0.243
NIST A129	-0.128	0.243
NMi-VSL 226246	0.086	0.429
NPL 221481	0.051	0.372
NPL 221483	0.066	0.372
NPL 221485	0.087	0.372
NRC A138	0.078	0.505
NRC A140	0.120	0.505
PTB 229074	-0.130	0.427
PTB 229075	0.022	0.427
VNIIFTRI 79	0.635	0.844
VNIIFTRI 89	0.719	1.022
NMIJ B271	0.017	0.285
NMIJ B310	0.040	0.282
PTB 229074	0.158	0.326
PTB 229075	-0.090	0.326
NPL 221485	-0.032	0.326
VSL 226246	-0.036	0.326
INRiM 232324	0.124	1.027
INTiBS B178	0.387	1.027

CCT-K1 and EURAMET.T-K1 : Nominal temperature, $T_{90} = 2.601$ K
Degrees of equivalence, D_i , and expanded uncertainties ($k = 2$), U_i , expressed in mK



Key comparison CCT-K1 and EURAMET.T-K1

MEASURAND : Temperature T_{90} / K

NOMINAL TEMPERATURE : $T_{90} = 2.700$ K

CCT-K1

Laboratory individual measurements

Lab: Laboratory

S/N i : serial number of Thermometer i

T_i : temperature value for Thermometer i

u_i : standard uncertainty in the calibration of Thermometer i

u_{comp} : standard uncertainty in the comparison measurements

$$u_{\text{comp}} = 0.115 \text{ mK}$$

Lab, S/N i	T_i / mK	u_i / mK
NIST A128	2.699899	0.040
NIST A129	2.699752	0.040
NMi-VSL 226246	2.700027	0.169
NPL 221481	2.699890	0.147
NPL 221483	2.699940	0.147
NPL 221485	2.699849	0.147
NRC A138	2.699987	0.225
NRC A140	2.700099	0.225
PTB 229074	2.699794	0.180
PTB 229075	2.699936	0.180
VNIIFTRI 79	2.700463	0.402
VNIIFTRI 89	2.700602	0.486
NMIJ B271	2.699928	0.085
NMIJ B310	2.699939	0.082

Key comparison CCT-K1 and EURAMET.T-K1


MEASURAND : Temperature T_{90} / K

KEY COMPARISON REFERENCE VALUE: $T_R = 2.699911$ K

Matrix of equivalence

Lab, S/N i	D_i	U_i / mK
NIST A128	-0.012	0.243
NIST A129	-0.159	0.243
NMi-VSL 226246	0.116	0.408
NPL 221481	-0.021	0.373
NPL 221483	0.029	0.373
NPL 221485	-0.062	0.373
NRC A138	0.076	0.505
NRC A140	0.188	0.505
PTB 229074	-0.117	0.427
PTB 229075	0.025	0.427
VNIIFTRI 79	0.552	0.836
VNIIFTRI 89	0.691	0.999
NMIJ B271	0.017	0.286
NMIJ B310	0.028	0.282
PTB 229074	0.259	0.324
PTB 229075	-0.152	0.324
NPL 221485	-0.042	0.324
VSL 226246	-0.065	0.324
INRiM 232324	0.060	1.027
INTiBS B178	0.438	1.027

CCT-K1 and EURAMET.T-K1 : Nominal temperature, $T_{90} = 2.700$ K
Degrees of equivalence, D_i , and expanded uncertainties ($k = 2$), U_i , expressed in mK



Key comparison CCT-K1

MEASURAND : Temperature T_{90} / K

NOMINAL TEMPERATURE : $T_{90} = 2.897$ K

Laboratory individual measurements

Lab: Laboratory

S/N i : serial number of Thermometer i

T_i : temperature value for Thermometer i

u_i : standard uncertainty in the calibration of Thermometer i

u_{comp} : standard uncertainty in the comparison measurements

$$u_{\text{comp}} = 0.114 \text{ mK}$$

Lab, S/N i	T_i / mK	u_i / mK
NIST A128	2.896679	0.040
NIST A129	2.896544	0.040
NMi-VSL 226246	2.896816	0.144
NPL 221481	2.896790	0.150
NPL 221483	2.896800	0.150
NPL 221485	2.896663	0.150
NRC A138	2.896787	0.225
NRC A140	2.896897	0.225
PTB 229074	2.896512	0.180
PTB 229075	2.896793	0.180
VNIIFTRI 79	2.897272	0.394
VNIIFTRI 89	2.897367	0.462
NMIJ B271	2.896750	0.081
NMIJ B310	2.896850	0.084

Key comparison CCT-K1

MEASURAND : Temperature T_{90} / K

KEY COMPARISON REFERENCE VALUE: $T_R = 2.896733$ K


Matrix of equivalence

Lab, S/N i	Lab, S/N j \longrightarrow															
	NIST A128		NIST A129		NMi-VSL 226246		NPL 221481		NPL 221483		NPL 221485		NRC A138			
	D_i / mK	U_i / mK		D_j / mK	U_j / mK		D_j / mK	U_j / mK		D_j / mK	U_j / mK		D_j / mK	U_j / mK		
NIST A128	-0.054	0.242		0.135	0.343	-0.136	0.440	-0.111	0.448	-0.121	0.448	0.017	0.448	-0.108	0.560	
NIST A129	-0.189	0.242		-0.135	0.343	-0.272	0.440	-0.246	0.448	-0.256	0.448	-0.118	0.448	-0.243	0.560	
NMi-VSL 226246	0.083	0.368		0.136	0.440	0.272	0.440	0.026	0.526	0.016	0.526	0.153	0.526	0.029	0.624	
NPL 221481	0.057	0.377		0.111	0.448	0.246	0.448	-0.026	0.526		-0.010	0.533	0.128	0.533	0.003	0.630
NPL 221483	0.067	0.377		0.121	0.448	0.256	0.448	-0.016	0.526	0.010	0.533		0.138	0.533	0.013	0.630
NPL 221485	-0.070	0.377		-0.017	0.448	0.118	0.448	-0.153	0.526	-0.128	0.533	-0.138	0.533		-0.125	0.630
NRC A138	0.054	0.505		0.108	0.560	0.243	0.560	-0.029	0.624	-0.003	0.630	-0.013	0.630	0.125	0.630	
NRC A140	0.164	0.505		0.218	0.560	0.353	0.560	0.081	0.624	0.107	0.630	0.097	0.630	0.235	0.630	0.110
PTB 229074	-0.221	0.426		-0.168	0.490	-0.033	0.490	-0.304	0.563	-0.279	0.569	-0.289	0.569	-0.151	0.569	-0.275
PTB 229075	0.060	0.426		0.114	0.490	0.249	0.490	-0.023	0.563	0.003	0.569	-0.007	0.569	0.131	0.569	0.006
VNIIFTRI 79	0.539	0.821		0.593	0.856	0.728	0.856	0.457	0.899	0.482	0.903	0.472	0.903	0.610	0.903	0.485
VNIIFTRI 89	0.634	0.952		0.687	0.983	0.823	0.983	0.551	1.021	0.577	1.024	0.567	1.024	0.704	1.024	0.580
NMIJ B271	0.017	0.280		0.070	0.370	0.205	0.370	-0.066	0.462	-0.041	0.470	-0.051	0.470	0.087	0.470	-0.038
NMIJ B310	0.117	0.284		0.170	0.373	0.305	0.373	0.034	0.464	0.059	0.472	0.049	0.472	0.187	0.472	0.062

Matrix of equivalence (Continued)

Lab, S/N i	Lab, S/N j \longrightarrow															
	NRC A140		PTB 229074		PTB 229075		VNIIFTRI 79		VNIIFTRI 89		NMIJ B271		NMIJ B310			
	D_i / mK	U_i / mK		D_j / mK	U_j / mK		D_j / mK	U_j / mK		D_j / mK	U_j / mK		D_j / mK	U_j / mK		
NIST A128	-0.054	0.242		-0.218	0.560	0.168	0.490	-0.114	0.490	-0.593	0.856	-0.687	0.983	-0.070	0.370	-0.170
NIST A129	-0.189	0.242		-0.353	0.560	0.033	0.490	-0.249	0.490	-0.728	0.856	-0.823	0.983	-0.205	0.370	-0.305
NMi-VSL 226246	0.083	0.368		-0.081	0.624	0.304	0.563	0.023	0.563	-0.457	0.899	-0.551	1.021	0.066	0.462	-0.034
NPL 221481	0.057	0.377		-0.107	0.630	0.279	0.569	-0.003	0.569	-0.482	0.903	-0.577	1.024	0.041	0.470	-0.059
NPL 221483	0.067	0.377		-0.097	0.630	0.289	0.569	0.007	0.569	-0.472	0.903	-0.567	1.024	0.051	0.470	-0.049
NPL 221485	-0.070	0.377		-0.235	0.630	0.151	0.569	-0.131	0.569	-0.610	0.903	-0.704	1.024	-0.087	0.470	-0.187
NRC A138	0.054	0.505		-0.110	0.714	0.275	0.661	-0.006	0.661	-0.485	0.963	-0.580	1.078	0.038	0.577	-0.062
NRC A140	0.164	0.505				0.385	0.661	0.104	0.661	-0.375	0.963	-0.470	1.078	0.148	0.577	0.047
PTB 229074	-0.221	0.426		-0.385	0.661		-0.281	0.603		-0.761	0.925	-0.855	1.044	-0.238	0.510	-0.338
PTB 229075	0.060	0.426		-0.104	0.661	0.281	0.603			-0.479	0.925	-0.574	1.044	0.043	0.510	-0.057
VNIIFTRI 79	0.539	0.821		0.375	0.963	0.761	0.925	0.479	0.925		-0.094	1.257	0.523	0.867	0.423	0.868
VNIIFTRI 89	0.634	0.952		0.470	1.078	0.855	1.044	0.574	1.044	0.094	1.257		0.617	0.993	0.517	0.994
NMIJ B271	0.017	0.280		-0.148	0.577	0.238	0.510	-0.043	0.510	-0.523	0.867	-0.617	0.993		-0.100	0.399
NMIJ B310	0.117	0.284		-0.047	0.579	0.338	0.512	0.057	0.512	-0.423	0.868	-0.517	0.994	0.100	0.399	

CCT-K1 : Nominal temperature, $T_{90} = 2.897$ K
Degrees of equivalence, D_i , and expanded uncertainties ($k = 2$), U_i , expressed in mK



Key comparison CCT-K1 and EURAMET.T-K1

MEASURAND : Temperature T_{90} / K

NOMINAL TEMPERATURE : $T_{90} = 2.997$ K

CCT-K1

Laboratory individual measurements

Lab: Laboratory

S/N i : serial number of Thermometer i

T_i : temperature value for Thermometer i

u_i : standard uncertainty in the calibration of Thermometer i

u_{comp} : standard uncertainty in the comparison measurements

$$u_{\text{comp}} = 0.114 \text{ mK}$$

Lab, S/N i	T_i / mK	u_i / mK
NIST A128	2.996572	0.041
NIST A129	2.996576	0.041
NMi-VSL 226246	2.996750	0.131
NPL 221481	2.996710	0.151
NPL 221485	2.996632	0.151
NRC A138	2.996624	0.225
NRC A140	2.996781	0.225
PTB 229074	2.996582	0.180
PTB 229075	2.996720	0.180
NMIJ B271	2.996679	0.082
NMIJ B310	2.996688	0.080

Key comparison CCT-K1 and EURAMET.T-K1


MEASURAND : Temperature T_{90} / K

KEY COMPARISON REFERENCE VALUE: $T_R = 2.996648$ K

Matrix of equivalence

Lab, S/N i	D_i	U_i / mK
NIST A128	-0.076	0.242
NIST A129	-0.072	0.242
NMi-VSL 226246	0.102	0.347
NPL 221481	0.062	0.379
NPL 221485	-0.016	0.379
NRC A138	-0.024	0.504
NRC A140	0.133	0.504
PTB 229074	-0.066	0.426
PTB 229075	0.072	0.426
NMIJ B271	0.031	0.281
NMIJ B310	0.040	0.279
PTB 229074	0.084	0.320
PTB 229075	-0.010	0.320
NPL 221485	-0.071	0.320
VSL 226246	-0.003	0.320
INRiM 232324	0.006	1.027
INTiBS B178	0.205	1.027

**CCT-K1 and EURAMET.T-K1 : Nominal temperature, $T_{90} = 2.997$ K
Degrees of equivalence, D_i , and expanded uncertainties ($k = 2$), U_i , expressed in mK**



Key comparison CCT-K1 and EURAMET.T-K1

MEASURAND : Temperature T_{90} / K

NOMINAL TEMPERATURE : $T_{90} = 3.099$ K

CCT-K1

Laboratory individual measurements

Lab: Laboratory

S/N i : serial number of Thermometer i

T_i : temperature value for Thermometer i

u_i : standard uncertainty in the calibration of Thermometer i

u_{comp} : standard uncertainty in the comparison measurements

$$u_{\text{comp}} = 0.114 \text{ mK}$$

Lab, S/N i	T_i / mK	u_i / mK
NIST A128	3.099314	0.041
NIST A129	3.099213	0.041
NMI-VSL 226246	3.099533	0.130
NPL 221481	3.099520	0.153
NPL 221483	3.099480	0.153
NPL 221485	3.099376	0.153
NRC A138	3.099336	0.225
NRC A140	3.099494	0.225
PTB 229074	3.099343	0.180
PTB 229075	3.099502	0.180
VNIIFTRI 79	3.099793	0.392
VNIIFTRI 89	3.099869	0.451
NMIJ B271	3.099409	0.082
NMIJ B310	3.099440	0.080

Key comparison CCT-K1 and EURAMET.T-K1


MEASURAND : Temperature T_{90} / K

KEY COMPARISON REFERENCE VALUE: $T_R = 3.099398$ K

Matrix of equivalence

Lab, S/N i	D_i	U_i / mK
NIST A128	-0.084	0.242
NIST A129	-0.185	0.242
NMi-VSL 226246	0.135	0.346
NPL 221481	0.122	0.381
NPL 221483	0.082	0.381
NPL 221485	-0.022	0.381
NRC A138	-0.062	0.504
NRC A140	0.096	0.504
PTB 229074	-0.055	0.426
PTB 229075	0.104	0.426
VNIIFTRI 79	0.395	0.816
VNIIFTRI 89	0.471	0.931
NMIJ B271	0.011	0.280
NMIJ B310	0.042	0.278
PTB 229074	0.078	0.318
PTB 229075	-0.015	0.318
NPL 221485	-0.070	0.318
VSL 226246	0.006	0.318
INRiM 232324	0.051	1.027
INTiBS B178	0.202	1.027

**CCT-K1 and EURAMET.T-K1: Nominal temperature, $T_{90} = 3.099$ K
Degrees of equivalence, D_i , and expanded uncertainties ($k = 2$), U_i , expressed in mK**



Key comparison CCT-K1 and EURAMET.T-K1

MEASURAND : Temperature T_{90} / K

NOMINAL TEMPERATURE : $T_{90} = 3.400$ K

CCT-K1

Laboratory individual measurements

Lab: Laboratory

S/N i : serial number of Thermometer i

T_i : temperature value for Thermometer i

u_i : standard uncertainty in the calibration of Thermometer i

u_{comp} : standard uncertainty in the comparison measurements

$$u_{\text{comp}} = 0.113 \text{ mK}$$

Lab, S/N i	T_i / mK	u_i / mK
NIST A128	3.400154	0.042
NIST A129	3.400049	0.042
NMI-VSL 226246	3.400396	0.127
NPL 221481	3.400390	0.157
NPL 221483	3.400290	0.157
NPL 221485	3.400242	0.157
NRC A138	3.400163	0.225
NRC A140	3.400390	0.225
PTB 229074	3.400225	0.180
PTB 229075	3.400352	0.180
VNIIFTRI 79	3.400410	0.396
VNIIFTRI 89	3.400480	0.455
NMIJ B271	3.400259	0.082
NMIJ B310	3.400234	0.081

Key comparison CCT-K1 and EURAMET.T-K1


MEASURAND : Temperature T_{90} / K

KEY COMPARISON REFERENCE VALUE: $T_R = 3.400235$ K

Matrix of equivalence

Lab, S/N i	D_i	U_i / mK
NIST A128	-0.081	0.241
NIST A129	-0.186	0.241
NMi-VSL 226246	0.161	0.341
NPL 221481	0.155	0.386
NPL 221483	0.055	0.386
NPL 221485	0.007	0.386
NRC A138	-0.072	0.504
NRC A140	0.155	0.504
PTB 229074	-0.010	0.425
PTB 229075	0.117	0.425
VNIIFTRI 79	0.175	0.824
VNIIFTRI 89	0.245	0.937
NMIJ B271	0.024	0.279
NMIJ B310	-0.001	0.278
PTB 229074	0.100	0.313
PTB 229075	-0.053	0.313
NPL 221485	-0.058	0.313
VSL 226246	0.011	0.313
INRiM 232324	0.176	1.027
INTiBS B178	0.168	0.313

CCT-K1 and EURAMET.T-K1 : Nominal temperature, $T_{90} = 3.400$ K
Degrees of equivalence, D_i , and expanded uncertainties ($k = 2$), U_i , expressed in mK



Key comparison CCT-K1

MEASURAND : Temperature T_{90} / K

NOMINAL TEMPERATURE : $T_{90} = 3.429$ K

Laboratory individual measurements

Lab: Laboratory

S/N i : serial number of Thermometer i

T_i : temperature value for Thermometer i

u_i : standard uncertainty in the calibration of Thermometer i

u_{comp} : standard uncertainty in the comparison measurements

$$u_{\text{comp}} = 0.113 \text{ mK}$$

Lab, S/N i	T_i / mK	u_i / mK
NIST A128	3.429195	0.042
NIST A129	3.429129	0.042
NMi-VSL 226246	3.429411	0.127
NPL 221481	3.429460	0.157
NPL 221485	3.429250	0.157
NRC A138	3.429182	0.225
NRC A140	3.429382	0.225
PTB 229074	3.429166	0.180
PTB 229075	3.429333	0.180
NMIJ B271	3.429255	0.082
NMIJ B310	3.429244	0.081

Key comparison CCT-K1

MEASURAND : Temperature T_{90} / K

KEY COMPARISON REFERENCE VALUE: $T_R = 3.429250$ K


Matrix of equivalence

Lab, S/N i	Lab, S/N j \longrightarrow													
	D _i U _i / mK		NIST A128		NIST A129		NMi-VSL 226246		NPL 221481		NPL 221485		NRC A138	
	D _{ij} U _{ij} / mK													
NIST A128	-0.055	0.241		0.066	0.340	-0.217	0.417	-0.266	0.455	-0.055	0.455	0.013	0.558	
NIST A129	-0.121	0.241	-0.066	0.340		-0.282	0.417	-0.331	0.455	-0.121	0.455	-0.053	0.558	
NMi-VSL 226246	0.161	0.340	0.217	0.417	0.282	0.417		-0.049	0.515	0.161	0.515	0.229	0.608	
NPL 221481	0.210	0.387	0.266	0.455	0.331	0.455	0.049	0.515		0.210	0.547	0.279	0.635	
NPL 221485	0.000	0.387	0.055	0.455	0.121	0.455	-0.161	0.515	-0.210	0.547		0.068	0.635	
NRC A138	-0.068	0.503	-0.013	0.558	0.053	0.558	-0.229	0.608	-0.279	0.635	-0.068	0.635		
NRC A140	0.132	0.503	0.187	0.558	0.253	0.558	-0.029	0.608	-0.078	0.635	0.132	0.635	0.200	0.712
PTB 229074	-0.084	0.425	-0.029	0.488	0.037	0.488	-0.245	0.544	-0.294	0.575	-0.084	0.575	-0.016	0.659
PTB 229075	0.083	0.425	0.138	0.488	0.204	0.488	-0.078	0.544	-0.127	0.575	0.083	0.575	0.151	0.659
NMIJ B271	0.005	0.279	0.060	0.369	0.126	0.369	-0.156	0.440	-0.205	0.477	0.005	0.477	0.073	0.576
NMIJ B310	-0.006	0.278	0.050	0.368	0.115	0.368	-0.167	0.439	-0.216	0.476	-0.006	0.476	0.062	0.575

Matrix of equivalence (Continued)

Lab, S/N i	Lab, S/N j \longrightarrow											
	D _i U _i / mK		NRC A140		PTB 229074		PTB 229075		NMIIJ B271		NMIIJ B310	
	D _{ij} U _{ij} / mK											
NIST A128	-0.055	0.241	-0.187	0.558	0.029	0.488	-0.138	0.488	-0.060	0.369	-0.050	0.368
NIST A129	-0.121	0.241	-0.253	0.558	-0.037	0.488	-0.204	0.488	-0.126	0.369	-0.115	0.368
NMi-VSL 226246	0.161	0.340	0.029	0.608	0.245	0.544	0.078	0.544	0.156	0.440	0.167	0.439
NPL 221481	0.210	0.387	0.078	0.635	0.294	0.575	0.127	0.575	0.205	0.477	0.216	0.476
NPL 221485	0.000	0.387	-0.132	0.635	0.084	0.575	-0.083	0.575	-0.005	0.477	0.006	0.476
NRC A138	-0.068	0.503	-0.200	0.712	0.016	0.659	-0.151	0.659	-0.073	0.576	-0.062	0.575
NRC A140	0.132	0.503			0.216	0.659	0.049	0.659	0.127	0.576	0.138	0.575
PTB 229074	-0.084	0.425	-0.216	0.659		-0.167	0.601	-0.089	0.508	-0.078	0.508	
PTB 229075	0.083	0.425	-0.049	0.659	0.167	0.601		0.078	0.508	0.089	0.508	
NMIJ B271	0.005	0.279	-0.127	0.576	0.089	0.508	-0.078	0.508		0.011	0.394	
NMIJ B310	-0.006	0.278	-0.138	0.575	0.078	0.508	-0.089	0.508	-0.011	0.394		

CCT-K1 : Nominal temperature, $T_{90} = 3.429$ K
Degrees of equivalence, D_i , and expanded uncertainties ($k = 2$), U_i , expressed in mK



Key comparison CCT-K1 and EURAMET.T-K1

MEASURAND : Temperature T_{90} / K

NOMINAL TEMPERATURE : $T_{90} = 3.801$ K

CCT-K1

Laboratory individual measurements

Lab: Laboratory

S/N i : serial number of Thermometer i

T_i : temperature value for Thermometer i

u_i : standard uncertainty in the calibration of Thermometer i

u_{comp} : standard uncertainty in the comparison measurements

$$u_{\text{comp}} = 0.112 \text{ mK}$$

Lab, S/N i	T_i / mK	u_i / mK
NIST A128	3.800754	0.042
NIST A129	3.800731	0.042
NMI-VSL 226246	3.801097	0.124
NPL 221481	3.801025	0.162
NPL 221483	3.800935	0.162
NPL 221485	3.800972	0.162
NRC A138	3.800894	0.225
NRC A140	3.800958	0.225
PTB 229074	3.800798	0.180
PTB 229075	3.801003	0.180
VNIIFTRI 79	3.801008	0.403
VNIIFTRI 89	3.801137	0.460
NMIJ B271	3.800962	0.085
NMIJ B310	3.800991	0.084

Key comparison CCT-K1 and EURAMET.T-K1


MEASURAND : Temperature T_{90} / K

KEY COMPARISON REFERENCE VALUE: $T_R = 3.800903$ K

Matrix of equivalence

Lab, S/N i	D_i	U_i / mK
NIST A128	-0.149	0.239
NIST A129	-0.172	0.239
NMi-VSL 226246	0.194	0.334
NPL 221481	0.122	0.394
NPL 221483	0.032	0.394
NPL 221485	0.069	0.394
NRC A138	-0.009	0.503
NRC A140	0.055	0.503
PTB 229074	-0.105	0.424
PTB 229075	0.100	0.424
VNIIFTRI 79	0.105	0.836
VNIIFTRI 89	0.234	0.946
NMIJ B271	0.059	0.281
NMIJ B310	0.088	0.280
PTB 229074	0.056	0.306
PTB 229075	-0.032	0.306
NPL 221485	-0.062	0.306
VSL 226246	0.038	0.306
INRiM 232324	0.078	1.026
INTiBS B178	0.172	1.026

**CCT-K1 and EURAMET.T-K1 : Nominal temperature, $T_{90} = 3.801$ K
Degrees of equivalence, D_i , and expanded uncertainties ($k = 2$), U_i , expressed in mK**



Key comparison CCT-K1 and EURAMET.T-K1

MEASURAND : Temperature T_{90} / K

NOMINAL TEMPERATURE : $T_{90} = 4.225$ K

CCT-K1

Laboratory individual measurements

Lab: Laboratory

S/N i : serial number of Thermometer i

T_i : temperature value for Thermometer i

u_i : standard uncertainty in the calibration of Thermometer i

u_{comp} : standard uncertainty in the comparison measurements

$$u_{\text{comp}} = 0.111 \text{ mK}$$

Lab, S/N i	T_i / mK	u_i / mK
NIST A128	4.224713	0.042
NIST A129	4.224680	0.042
NMI-VSL 226246	4.224827	0.122
NPL 221481	4.224760	0.167
NPL 221483	4.224731	0.167
NPL 221485	4.224823	0.167
NRC A138	4.224794	0.225
NRC A140	4.224867	0.225
PTB 229074	4.224704	0.180
PTB 229075	4.224800	0.180
VNIIFTRI 79	4.224908	0.409
VNIIFTRI 89	4.225064	0.465
NMIJ B271	4.224928	0.088
NMIJ B310	4.224923	0.087

Key comparison CCT-K1 and EURAMET.T-K1


MEASURAND : Temperature T_{90} / K

KEY COMPARISON REFERENCE VALUE: $T_R = 4.224794$ K

Matrix of equivalence

Lab, S/N i	D_i	U_i / mK
NIST A128	-0.081	0.238
NIST A129	-0.114	0.238
NMi-VSL 226246	0.033	0.330
NPL 221481	-0.034	0.401
NPL 221483	-0.063	0.401
NPL 221485	0.029	0.401
NRC A138	0.000	0.502
NRC A140	0.073	0.502
PTB 229074	-0.090	0.423
PTB 229075	0.006	0.423
VNIIFTRI 79	0.114	0.848
VNIIFTRI 89	0.270	0.956
NMIJ B271	0.134	0.283
NMIJ B310	0.129	0.282
PTB 229074	0.064	0.300
PTB 229075	-0.027	0.300
NPL 221485	-0.075	0.300
VSL 226246	0.038	0.300
INRiM 232324	-0.253	1.320
INTiBS B178	0.260	1.026

CCT-K1 and EURAMET.T-K1 : Nominal temperature, $T_{90} = 4.225$ K
Degrees of equivalence, D_i , and expanded uncertainties ($k = 2$), U_i , expressed in mK



Key comparison CCT-K1 and EURAMET.T-K1

MEASURAND : Temperature T_{90} / K

NOMINAL TEMPERATURE : $T_{90} = 4.478$ K

CCT-K1

Laboratory individual measurements

Lab: Laboratory

S/N i : serial number of Thermometer i

T_i : temperature value for Thermometer i

u_i : standard uncertainty in the calibration of Thermometer i

u_{comp} : standard uncertainty in the comparison measurements

$$u_{\text{comp}} = 0.112 \text{ mK}$$

Lab, S/N i	T_i / mK	u_i / mK
NIST A128	4.477545	0.042
NIST A129	4.477546	0.042
NMI-VSL 226246	4.477459	0.206
NPL 221481	4.477410	0.210
NPL 221483	4.477489	0.201
NPL 221485	4.477502	0.201
NRC A138	4.477701	0.550
NRC A140	4.478134	0.550
PTB 229074	4.477475	0.180
PTB 229075	4.477406	0.180
VNIIFTRI 79	4.477674	0.413
VNIIFTRI 89	4.477791	0.468

Key comparison CCT-K1 and EURAMET.T-K1


MEASURAND : Temperature T_{90} / K

KEY COMPARISON REFERENCE VALUE: $T_R = 4.477522$ K

Matrix of equivalence

Lab, S/N i	D_i	U_i / mK
NIST A128	0.023	0.240
NIST A129	0.024	0.240
NMi-VSL 226246	-0.063	0.468
NPL 221481	-0.112	0.476
NPL 221483	-0.033	0.460
NPL 221485	-0.020	0.460
NRC A138	0.179	1.123
NRC A140	0.612	1.123
PTB 229074	-0.047	0.424
PTB 229075	-0.116	0.424
VNIIFTRI 79	0.152	0.856
VNIIFTRI 89	0.269	0.962
PTB 229074	0.100	0.302
PTB 229075	-0.117	0.302
NPL 221485	-0.026	0.302
VSL 226246	0.043	0.302
INRiM 232324	-0.161	1.320
INTIBS B178	0.342	1.026

CCT-K1 and EURAMET.T-K1 : Nominal temperature, $T_{90} = 4.478$ K
Degrees of equivalence, D_i , and expanded uncertainties ($k = 2$), U_i , expressed in mK



Key comparison CCT-K1 and EURAMET.T-K1

MEASURAND : Temperature T_{90} / K

NOMINAL TEMPERATURE : $T_{90} = 5.000$ K

CCT-K1

Laboratory individual measurements

Lab: Laboratory

S/N i : serial number of Thermometer i

T_i : temperature value for Thermometer i

u_i : standard uncertainty in the calibration of Thermometer i

u_{comp} : standard uncertainty in the comparison measurements

$$u_{\text{comp}} = 0.114 \text{ mK}$$

Lab, S/N i	T_i / mK	u_i / mK
NIST A128	5.000483	0.067
NIST A129	5.000483	0.067
NMi-VSL 226246	5.000468	0.206
NPL 221481	5.000220	0.210
NPL 221483	5.000395	0.201
NPL 221485	5.000332	0.201
NRC A138	5.001124	0.550
NRC A140	5.001270	0.550
PTB 229074	5.000397	0.270
PTB 229075	5.000352	0.250
VNIIFTRI 79	5.000647	0.421
VNIIFTRI 89	5.000713	0.474

Key comparison CCT-K1 and EURAMET.T-K1


MEASURAND : Temperature T_{90} / K

KEY COMPARISON REFERENCE VALUE: $T_R = 5.000458$ K

Matrix of equivalence

Lab, S/N i	D_i	U_i / mK
NIST A128	0.025	0.264
NIST A129	0.025	0.264
NMi-VSL 226246	0.010	0.470
NPL 221481	-0.238	0.478
NPL 221483	-0.063	0.462
NPL 221485	-0.126	0.462
NRC A138	0.666	1.123
NRC A140	0.812	1.123
PTB 229074	-0.061	0.586
PTB 229075	-0.106	0.550
VNIIFTRI 79	0.189	0.873
VNIIFTRI 89	0.255	0.975
PTB 229074	0.071	0.307
PTB 229075	-0.069	0.307
NPL 221485	-0.024	0.307
VSL 226246	0.021	0.307
INRiM 232324	-0.130	1.321
INTIBS B178	0.414	1.027

CCT-K1 and EURAMET.T-K1 : Nominal temperature, $T_{90} = 5.000$ K
Degrees of equivalence, D_i , and expanded uncertainties ($k = 2$), U_i , expressed in mK



Key comparison CCT-K1 and EURAMET.T-K1

MEASURAND : Temperature T_{90} / K

NOMINAL TEMPERATURE : $T_{90} = 5.948$ K

Laboratory individual measurements

Lab: Laboratory

S/N i : serial number of Thermometer i

T_i : temperature value for Thermometer i

u_i : standard uncertainty in the calibration of Thermometer i

u_{comp} : standard uncertainty in the comparison measurements

$$u_{\text{comp}} = 0.118 \text{ mK}$$

Lab, S/N i	T_i / mK	u_i / mK
NIST A128	5.947948	0.064
NIST A129	5.948148	0.064
NMi-VSL 226246	5.948420	0.206
NPL 221481	5.948030	0.210
NPL 221483	5.948374	0.201
NPL 221485	5.947935	0.201
NRC A138	5.948869	0.550
NRC A140	5.949112	0.550
PTB 229074	5.948358	0.273
PTB 229075	5.948217	0.252
VNIIFTRI 79	5.948466	0.436
VNIIFTRI 89	5.948584	0.485

Key comparison CCT-K1 and EURAMET.T-K1


MEASURAND : Temperature T_{90} / K

KEY COMPARISON REFERENCE VALUE: $T_R = 5.948165$ K

Matrix of equivalence

Lab, S/N i	D_i	U_i / mK
NIST A128	-0.217	0.268
NIST A129	-0.017	0.268
NMI-VSL 226246	0.255	0.474
NPL 221481	-0.135	0.481
NPL 221483	0.209	0.466
NPL 221485	-0.230	0.466
NRC A138	0.704	1.125
NRC A140	0.947	1.125
PTB 229074	0.193	0.594
PTB 229075	0.052	0.557
VNIIFTRI 79	0.301	0.904
VNIIFTRI 89	0.419	0.999
PTB 229074	0.088	0.316
PTB 229075	-0.074	0.316
NPL 221485	-0.005	0.316
VSL 226246	-0.008	0.316
INRIM 232324	0.227	1.321
INTIBS B178	0.400	1.027

CCT-K1 and EURAMET.T-K1 : Nominal temperature, $T_{90} = 5.948$ K
Degrees of equivalence, D_i , and expanded uncertainties ($k = 2$), U_i , expressed in mK



Key comparison CCT-K1 and EURAMET.T-K1

MEASURAND : Temperature T_{90} / K

NOMINAL TEMPERATURE : $T_{90} = 7.202$ K

CCT-K1

Laboratory individual measurements

Lab: Laboratory

S/N i : serial number of Thermometer i

T_i : temperature value for Thermometer i

u_i : standard uncertainty in the calibration of Thermometer i

u_{comp} : standard uncertainty in the comparison measurements

$$u_{\text{comp}} = 0.122 \text{ mK}$$

Lab, S/N i	T_i / mK	u_i / mK
NIST A128	7.201615	0.070
NIST A129	7.201513	0.070
NMi-VSL 226246	7.201679	0.206
NPL 221481	7.201400	0.210
NPL 221483	7.201415	0.201
NPL 221485	7.201191	0.201
NRC A138	7.202363	0.550
NRC A140	7.202306	0.550
PTB 229074	7.201593	0.277
PTB 229075	7.201538	0.256
VNIIFTRI 79	7.201591	0.456
VNIIFTRI 89	7.201839	0.501

Key comparison CCT-K1 and EURAMET.T-K1


MEASURAND : Temperature T_{90} / K

KEY COMPARISON REFERENCE VALUE: $T_R = 7.201544$ K

Matrix of equivalence

Lab, S/N i	D_i	U_i / mK
NIST A128	0.071	0.282
NIST A129	-0.031	0.282
NMI-VSL 226246	0.135	0.478
NPL 221481	-0.144	0.486
NPL 221483	-0.129	0.471
NPL 221485	-0.353	0.471
NRC A138	0.819	1.127
NRC A140	0.762	1.127
PTB 229074	0.049	0.605
PTB 229075	-0.006	0.567
VNIIFTRI 79	0.047	0.945
VNIIFTRI 89	0.295	1.031
PTB 229074	0.099	0.327
PTB 229075	-0.220	0.327
NPL 221481	0.238	0.327
NPL 221485	-0.008	0.327
VSL 226246	-0.109	0.327
INRIM 232324	0.380	1.322
INRIM B190	0.343	1.028
INTIBS B178	0.188	1.028

CCT-K1 and EURAMET.T-K1 : Nominal temperature, $T_{90} = 7.202$ K
Degrees of equivalence, D_i , and expanded uncertainties ($k = 2$), U_i , expressed in mK



Key comparison CCT-K1 and EURAMET.T-K1

MEASURAND : Temperature T_{90} / K

NOMINAL TEMPERATURE : $T_{90} = 8.296$ K

CCT-K1

Laboratory individual measurements

Lab: Laboratory

S/N i : serial number of Thermometer i

T_i : temperature value for Thermometer i

u_i : standard uncertainty in the calibration of Thermometer i

u_{comp} : standard uncertainty in the comparison measurements

$$u_{\text{comp}} = 0.127 \text{ mK}$$

Lab, S/N i	T_i / mK	u_i / mK
NIST A128	8.296546	0.076
NIST A129	8.296561	0.076
NMiVSL 226246	8.296278	0.206
NPL 221481	8.296020	0.210
NPL 221485	8.295779	0.201
NRC A138	8.297318	0.550
NRC A140	8.297308	0.550
PTB 229074	8.296214	0.280
PTB 229075	8.296135	0.258

Key comparison CCT-K1 and EURAMET.T-K1


MEASURAND : Temperature T_{90} / K

KEY COMPARISON REFERENCE VALUE: $T_R = 8.296372$ K

Matrix of equivalence

Lab, S/N i	D_i	U_i / mK
NIST A128	0.174	0.296
NIST A129	0.189	0.296
NMiVSL 226246	-0.094	0.483
NPL 221481	-0.352	0.490
NPL 221485	-0.593	0.475
NRC A138	0.946	1.129
NRC A140	0.936	1.129
PTB 229074	-0.158	0.615
PTB 229075	-0.237	0.575
PTB 229074	0.067	0.338
PTB 229075	-0.174	0.338
NPL 221481	0.185	0.338
NPL 221485	-0.060	0.338
VSL 226246	-0.019	0.338
INRIM 232324	0.194	1.323
INRIM B190	0.107	1.029
INTiBS B178	0.030	1.029

**CCT-K1 and EURAMET.T-K1: Nominal temperature, $T_{90} = 8.296$ K
Degrees of equivalence, D_i , and expanded uncertainties ($k = 2$), U_i , expressed in mK**



Key comparison CCT-K1

MEASURAND : Temperature T_{90} / K

NOMINAL TEMPERATURE : $T_{90} = 8.400$ K

Laboratory individual measurements

Lab: Laboratory

S/N i : serial number of Thermometer i

T_i : temperature value for Thermometer i

u_i : standard uncertainty in the calibration of Thermometer i

u_{comp} : standard uncertainty in the comparison measurements

$$u_{\text{comp}} = 0.127 \text{ mK}$$

Lab, S/N i	T_i / mK	u_i / mK
NIST A128	8.399816	0.076
NIST A129	8.399757	0.076
NMi-VSL 226246	8.399564	0.206
NPL 221481	8.399350	0.210
NPL 221483	8.399316	0.201
NPL 221485	8.398898	0.201
NRC A138	8.400617	0.550
NRC A140	8.400524	0.550
PTB 229074	8.399608	0.280
PTB 229075	8.399518	0.259
VNIIFTRI 79	8.399714	0.475
VNIIFTRI 89	8.400138	0.515

Key comparison CCT-K1

MEASURAND : Temperature T_{90} / K

KEY COMPARISON REFERENCE VALUE: $T_R = 8.399612$ K


Matrix of equivalence

Lab, S/N <i>i</i>	Lab, S/N <i>j</i> \longrightarrow					
	D_i	U_i	$/ \text{mK}$	D_j	U_j	$/ \text{mK}$
NIST A128	0.204	0.296		0.060	0.419	0.253
NIST A129	0.145	0.296		-0.193	0.567	0.193
NMi-VSL 226246	-0.048	0.483		-0.253	0.567	-0.214
NPL 221481	-0.262	0.491		-0.467	0.573	0.407
NPL 221483	-0.296	0.475		-0.500	0.560	-0.034
NPL 221485	-0.714	0.475		-0.918	0.560	-0.452
NRC A138	1.005	1.129		0.800	1.167	1.053
NRC A140	0.912	1.129		0.708	1.167	0.767
PTB 229074	-0.004	0.616		-0.208	0.683	-0.149
PTB 229075	-0.094	0.576		-0.299	0.648	-0.239
VNIIFTRI 79	0.102	0.983		-0.103	1.027	-0.043
VNIIFTRI 89	0.526	1.061		0.322	1.101	0.381

Matrix of equivalence (Continued)

Lab, S/N <i>i</i>	Lab, S/N <i>j</i> \longrightarrow					
	D_i	U_i	$/ \text{mK}$	D_j	U_j	$/ \text{mK}$
NIST A128	0.204	0.296		-0.800	1.167	-0.708
NIST A129	0.145	0.296		-0.860	1.167	-0.767
NMi-VSL 226246	-0.048	0.483		-1.053	1.228	-0.960
NPL 221481	-0.262	0.491		-1.267	1.231	-1.174
NPL 221483	-0.296	0.475		-1.301	1.225	-1.208
NPL 221485	-0.714	0.475		-1.719	1.225	-1.626
NRC A138	1.005	1.129		0.093	1.5965	1.009
NRC A140	0.912	1.129		-0.093	1.5965	0.916
PTB 229074	-0.004	0.616		-1.009	1.286	-0.916
PTB 229075	-0.094	0.576		-1.099	1.268	-1.006
VNIIFTRI 79	0.102	0.983		-0.903	1.497	-0.810
VNIIFTRI 89	0.526	1.061		-0.479	1.549	-0.386

CCT-K1 : Nominal temperature, $T_{90} = 8.400$ K
Degrees of equivalence, D_i , and expanded uncertainties ($k = 2$), U_i , expressed in mK



Key comparison CCT-K1 and EURAMET.T-K1

MEASURAND : Temperature T_{90} / K

NOMINAL TEMPERATURE : $T_{90} = 9.508$ K

CCT-K1

Laboratory individual measurements

Lab: Laboratory

S/N i : serial number of Thermometer i

T_i : temperature value for Thermometer i

u_i : standard uncertainty in the calibration of Thermometer i

u_{comp} : standard uncertainty in the comparison measurements

$$u_{\text{comp}} = 0.131 \text{ mK}$$

Lab, S/N i	T_i / mK	u_i / mK
NIST A128	9.508227	0.086
NIST A129	9.508237	0.086
NMi-VSL 226246	9.508099	0.206
NPL 221481	9.507680	0.210
NPL 221483	9.507531	0.201
NPL 221485	9.507387	0.201
NRC A138	9.509019	0.550
NRC A140	9.508971	0.550
PTB 229074	9.507997	0.284
PTB 229075	9.507851	0.262
VNIIFTRI 79	9.508226	0.492
VNIIFTRI 89	9.508531	0.528

Key comparison CCT-K1 and EURAMET.T-K1


MEASURAND : Temperature T_{90} / K

KEY COMPARISON REFERENCE VALUE: $T_R = 9.50802$ K

Matrix of equivalence

Lab, S/N i	D_i	U_i / mK
NIST A128	0.207	0.313
NIST A129	0.217	0.313
NMi-VSL 226246	0.079	0.488
NPL 221481	-0.340	0.495
NPL 221483	-0.489	0.480
NPL 221485	-0.633	0.480
NRC A138	0.999	1.131
NRC A140	0.951	1.131
PTB 229074	-0.023	0.625
PTB 229075	-0.169	0.585
VNIIFTRI 79	0.206	1.019
VNIIFTRI 89	0.511	1.089
PTB 229074	0.074	0.349
PTB 229075	-0.162	0.349
NPL 221481	0.071	0.349
NPL 221485	-0.033	0.349
VSL 226246	0.050	0.349
INRIM 232324	-0.001	1.323
INRIM B190	-0.071	1.030
INTIBS B178	0.225	1.030

**CCT-K1 and EURAMET.T-K1 : Nominal temperature, $T_{90} = 9.508$ K
Degrees of equivalence, D_i , and expanded uncertainties ($k = 2$), U_i , expressed in mK**



Key comparison CCT-K1 and EURAMET.T-K1

MEASURAND : Temperature T_{90} / K

NOMINAL TEMPERATURE : $T_{90} = 10.803$ K

CCT-K1

Laboratory individual measurements

Lab: Laboratory

S/N i : serial number of Thermometer i

T_i : temperature value for Thermometer i

u_i : standard uncertainty in the calibration of Thermometer i

u_{comp} : standard uncertainty in the comparison measurements

$$u_{\text{comp}} = 0.137 \text{ mK}$$

Lab, S/N i	T_i / mK	u_i / mK
NIST A128	10.803560	0.096
NIST A129	10.803616	0.096
NMI-VSL 226246	10.803584	0.227
NPL 221481	10.803070	0.206
NPL 221483	10.803011	0.200
NPL 221485	10.802842	0.200
NRC A138	10.804196	0.550
NRC A140	10.804185	0.550
PTB 229074	10.803248	0.288
PTB 229075	10.803221	0.265
VNIIFTRI 79	10.803595	0.513
VNIIFTRI 89	10.803927	0.544

Key comparison CCT-K1 and EURAMET.T-K1


MEASURAND : Temperature T_{90} / K

KEY COMPARISON REFERENCE VALUE: $T_R = 10.803390$ K

Matrix of equivalence

Lab, S/N i	D_i	U_i / mK
NIST A128	0.170	0.335
NIST A129	0.226	0.335
NMi-VSL 226246	0.194	0.531
NPL 221481	-0.320	0.495
NPL 221483	-0.379	0.484
NPL 221485	-0.548	0.484
NRC A138	0.806	1.134
NRC A140	0.795	1.134
PTB 229074	-0.142	0.638
PTB 229075	-0.169	0.597
VNIIFTRI 79	0.205	1.062
VNIIFTRI 89	0.537	1.122
PTB 229074	0.086	0.386
PTB 229075	-0.219	0.386
NPL 221481	0.060	0.386
NPL 221485	-0.001	0.386
VSL 226246	0.074	0.386
INRIM 232324	0.035	1.331
INRIM B190	0.024	1.040
INTiBS B178	0.534	1.040

**CCT-K1 and EURAMET.T-K1 : Nominal temperature, $T_{90} = 10.803$ K
Degrees of equivalence, D_i , and expanded uncertainties ($k = 2$), U_i , expressed in mK**



Key comparison CCT-K1 and EURAMET.T-K1

MEASURAND : Temperature T_{90} / K

NOMINAL TEMPERATURE : $T_{90} = 12.297$ K

CCT-K1

Laboratory individual measurements

Lab: Laboratory

S/N i : serial number of Thermometer i

T_i : temperature value for Thermometer i

u_i : standard uncertainty in the calibration of Thermometer i

u_{comp} : standard uncertainty in the comparison measurements

$$u_{\text{comp}} = 0.145 \text{ mK}$$

Lab, S/N i	T_i / mK	u_i / mK
NIST A128	12.297349	0.106
NIST A129	12.297413	0.106
NMi-VSL 226246	12.297900	0.267
NPL 221481	12.297090	0.199
NPL 221483	12.297224	0.197
NPL 221485	12.296905	0.197
NRC A138	12.297793	0.550
NRC A140	12.297778	0.550
PTB 229074	12.297149	0.292
PTB 229075	12.297099	0.269
VNIIFTRI 79	12.297633	0.536
VNIIFTRI 89	12.297547	0.562

Key comparison CCT-K1 and EURAMET.T-K1


MEASURAND : Temperature T_{90} / K

KEY COMPARISON REFERENCE VALUE: $T_R = 12.297309$ K

Matrix of equivalence

Lab, S/N i	D_i	U_i / mK
NIST A128	0.040	0.359
NIST A129	0.104	0.359
NMI-VSL 226246	0.591	0.608
NPL 221481	-0.219	0.492
NPL 221483	-0.085	0.489
NPL 221485	-0.404	0.489
NRC A138	0.484	1.138
NRC A140	0.469	1.138
PTB 229074	-0.160	0.653
PTB 229075	-0.210	0.611
VNIIFTRI 79	0.324	1.111
VNIIFTRI 89	0.238	1.161
PTB 229074	0.048	0.428
PTB 229075	-0.207	0.428
NPL 221481	0.074	0.428
NPL 221485	-0.002	0.428
VSL 226246	0.087	0.428
INRIM 232324	-0.006	1.341
INRIM B190	0.173	1.052
INTIBS B178	0.603	1.052

CCT-K1 and EURAMET.T-K1 : Nominal temperature, $T_{90} = 12.297$ K
Degrees of equivalence, D_i , and expanded uncertainties ($k = 2$), U_i , expressed in mK



Key comparison CCT-K1 and EURAMET.T-K1

MEASURAND : Temperature T_{90} / K

NOMINAL TEMPERATURE : $T_{90} = 13.798$ K

CCT-K1

Laboratory individual measurements

Lab: Laboratory

S/N i : serial number of Thermometer i

T_i : temperature value for Thermometer i

u_i : standard uncertainty in the calibration of Thermometer i

u_{comp} : standard uncertainty in the comparison measurements

$$u_{\text{comp}} = 0.153 \text{ mK}$$

Lab, S/N i	T_i / mK	u_i / mK
NIST A128	13.798156	0.111
NIST A129	13.798122	0.111
NMi-VSL 226246	13.798881	0.308
NPL 221481	13.798130	0.191
NPL 221483	13.798170	0.182
NPL 221485	13.797754	0.182
NRC A138	13.798459	0.550
NRC A140	13.798305	0.550
PTB 229074	13.798317	0.297
PTB 229075	13.798209	0.273
VNIIFTRI 79	13.798638	0.560
VNIIFTRI 89	13.798620	0.580

Key comparison CCT-K1 and EURAMET.T-K1


MEASURAND : Temperature T_{90} / K

KEY COMPARISON REFERENCE VALUE: $T_R = 13.798183$ K

Matrix of equivalence

Lab, S/N i	D_i	U_i / mK
NIST A128	-0.027	0.378
NIST A129	-0.061	0.378
NMI-VSL 226246	0.698	0.688
NPL 221481	-0.053	0.489
NPL 221483	-0.013	0.476
NPL 221485	-0.429	0.476
NRC A138	0.276	1.142
NRC A140	0.122	1.142
PTB 229074	0.134	0.668
PTB 229075	0.026	0.625
VNIIFTRI 79	0.455	1.161
VNIIFTRI 89	0.437	1.200
PTB 229074	0.082	0.472
PTB 229075	-0.168	0.472
NPL 221481	0.050	0.472
NPL 221485	-0.009	0.472
VSL 226246	0.046	0.472
INRIM 232324	-0.272	1.352
INRIM B190	0.375	1.067
INTIBS B178	0.801	1.067

**CCT-K1 and EURAMET.T-K1 : Nominal temperature, $T_{90} = 13.798$ K
Degrees of equivalence, D_i , and expanded uncertainties ($k = 2$), U_i , expressed in mK**



Key comparison CCT-K1 and EURAMET.T-K1

MEASURAND : Temperature T_{90} / K

NOMINAL TEMPERATURE : $T_{90} = 15.500$ K

CCT-K1

Laboratory individual measurements

Lab: Laboratory

S/N i : serial number of Thermometer i

T_i : temperature value for Thermometer i

u_i : standard uncertainty in the calibration of Thermometer i

u_{comp} : standard uncertainty in the comparison measurements

$$u_{\text{comp}} = 0.153 \text{ mK}$$

Lab, S/N i	T_i / mK	u_i / mK
NIST A128	15.499479	0.119
NIST A129	15.499474	0.119
NMi-VSL 226246	15.500221	0.354
NPL 221481	15.499523	0.194
NPL 221483	15.499571	0.185
NPL 221485	15.499138	0.185
NRC A138	15.499689	0.550
NRC A140	15.499662	0.550
PTB 229074	15.499940	0.302
PTB 229075	15.499849	0.277
VNIIFTRI 79	15.499890	0.560
VNIIFTRI 89	15.499817	0.585

Key comparison CCT-K1 and EURAMET.T-K1

MEASURAND : Temperature T_{90} / K

KEY COMPARISON REFERENCE VALUE: $T_R = 15.499566$ K


Matrix of equivalence

Lab, S/N i



	D_i	U_i / mK
NIST A128	-0.087	0.388
NIST A129	-0.092	0.388
NMi-VSL 226246	0.655	0.771
NPL 221481	-0.043	0.494
NPL 221483	0.005	0.480
NPL 221485	-0.428	0.480
NRC A138	0.123	1.142
NRC A140	0.096	1.142
PTB 229074	0.374	0.677
PTB 229075	0.283	0.633
VNIIFTRI 79	0.324	1.161
VNIIFTRI 89	0.251	1.209
PTB 229074	0.144	0.472
PTB 229075	-0.072	0.472
NPL 221481	0.043	0.472
NPL 221485	-0.139	0.472
VSL 226246	0.024	0.472
INRIM 232324	-0.424	1.352
INRIM B190	0.365	1.067
INTiBS B178	0.830	1.067

**CCT-K1 and EURAMET.T-K1: Nominal temperature, $T_{90} = 15.500$ K
Degrees of equivalence, D_i , and expanded uncertainties ($k = 2$), U_i , expressed in mK**



Key comparison CCT-K1 and EURAMET.T-K1

MEASURAND : Temperature T_{90} / K

NOMINAL TEMPERATURE : $T_{90} = 16.999$ K

CCT-K1

Laboratory individual measurements

Lab: Laboratory

S/N i : serial number of Thermometer i

T_i : temperature value for Thermometer i

u_i : standard uncertainty in the calibration of Thermometer i

u_{comp} : standard uncertainty in the comparison measurements

$$u_{\text{comp}} = 0.153 \text{ mK}$$

Lab, S/N i	T_i / mK	u_i / mK
NIST A128	16.999286	0.124
NIST A129	16.999271	0.124
NMi-VSL 226246	16.999833	0.394
NPL 221481	16.999290	0.197
NPL 221483	16.999280	0.188
NPL 221485	16.998847	0.188
NRC A138	16.999430	0.550
NRC A140	16.999259	0.550
PTB 229074	16.999859	0.307
PTB 229075	16.999701	0.281
VNIIFTRI 79	16.999602	0.560
VNIIFTRI 89	16.999783	0.589

Key comparison CCT-K1 and EURAMET.T-K1


MEASURAND : Temperature T_{90} / K

KEY COMPARISON REFERENCE VALUE: $T_R = 16.999335$ K

Matrix of equivalence

Lab, S/N i	D_i	U_i / mK
NIST A128	-0.049	0.394
NIST A129	-0.064	0.394
NMi-VSL 226246	0.498	0.845
NPL 221481	-0.045	0.499
NPL 221483	-0.055	0.485
NPL 221485	-0.488	0.485
NRC A138	0.095	1.142
NRC A140	-0.076	1.142
PTB 229074	0.524	0.686
PTB 229075	0.366	0.639
VNIIFTRI 79	0.267	1.161
VNIIFTRI 89	0.448	1.217
PTB 229074	0.144	0.472
PTB 229075	-0.072	0.472
NPL 221481	0.043	0.472
NPL 221485	-0.139	0.472
VSL 226246	0.024	0.472
INRIM 232324	-0.424	1.352
INRIM B190	0.365	1.067
INTIBS B178	0.830	1.067

CCT-K1 and EURAMET.T-K1: Nominal temperature, $T_{90} = 16.999$ K
Degrees of equivalence, D_i , and expanded uncertainties ($k = 2$), U_i , expressed in mK



Key comparison CCT-K1 and EURAMET.T-K1

MEASURAND : Temperature T_{90} / K

NOMINAL TEMPERATURE : $T_{90} = 18.597$ K

CCT-K1

Laboratory individual measurements

Lab: Laboratory

S/N i : serial number of Thermometer i

T_i : temperature value for Thermometer i

u_i : standard uncertainty in the calibration of Thermometer i

u_{comp} : standard uncertainty in the comparison measurements

$$u_{\text{comp}} = 0.153 \text{ mK}$$

Lab, S/N i	T_i / mK	u_i / mK
NIST A128	18.597374	0.129
NIST A129	18.597342	0.129
NMi-VSL 226246	18.597934	0.437
NPL 221481	18.597330	0.200
NPL 221483	18.597229	0.191
NPL 221485	18.596804	0.191
NRC A138	18.597402	0.550
NRC A140	18.597237	0.550
PTB 229074	18.597946	0.312
PTB 229075	18.597872	0.285
VNIIFTRI 79	18.597658	0.560
VNIIFTRI 89	18.597830	0.593

Key comparison CCT-K1 and EURAMET.T-K1


MEASURAND : Temperature T_{90} / K

KEY COMPARISON REFERENCE VALUE: $T_R = 18.597377$ K

Matrix of equivalence

Lab, S/N i	D_i	U_i / mK
NIST A128	-0.003	0.401
NIST A129	-0.035	0.401
NMI-VSL 226246	0.557	0.926
NPL 221481	-0.047	0.503
NPL 221483	-0.148	0.489
NPL 221485	-0.573	0.489
NRC A138	0.025	1.142
NRC A140	-0.140	1.142
PTB 229074	0.569	0.695
PTB 229075	0.495	0.647
VNIIFTRI 79	0.281	1.161
VNIIFTRI 89	0.453	1.226
PTB 229074	0.279	0.472
PTB 229075	-0.308	0.472
NPL 221481	0.115	0.472
NPL 221485	0.012	0.472
VSL 226246	-0.097	0.472
INRIM 232324	-0.645	1.352
INRIM B190	0.377	1.067
INTIBS B178	0.751	1.067

**CCT-K1 and EURAMET.T-K1 : Nominal temperature, $T_{90} = 18.597$ K
Degrees of equivalence, D_i , and expanded uncertainties ($k = 2$), U_i , expressed in mK**



Key comparison CCT-K1 and EURAMET.T-K1

MEASURAND : Temperature T_{90} / K

NOMINAL TEMPERATURE : $T_{90} = 20.299$ K

CCT-K1

Laboratory individual measurements

Lab: Laboratory

S/N i : serial number of Thermometer i

T_i : temperature value for Thermometer i

u_i : standard uncertainty in the calibration of Thermometer i

u_{comp} : standard uncertainty in the comparison measurements

$$u_{\text{comp}} = 0.153 \text{ mK}$$

Lab, S/N i	T_i / mK	u_i / mK
NIST A128	20.299107	0.132
NIST A129	20.299108	0.132
NMI-VSL 226246	20.299225	0.475
NPL 221481	20.298690	0.203
NPL 221483	20.298576	0.194
NPL 221485	20.298352	0.194
NRC A138	20.298800	0.550
NRC A140	20.298750	0.550
PTB 229074	20.299355	0.317
PTB 229075	20.299131	0.289
VNIIFTRI 79	20.298975	0.560
VNIIFTRI 89	20.299144	0.598

Key comparison CCT-K1 and EURAMET.T-K1


MEASURAND : Temperature T_{90} / K

KEY COMPARISON REFERENCE VALUE: $T_R = 20.298899$ K

Matrix of equivalence

Lab, S/N i	D_i	U_i / mK
NIST A128	0.208	0.404
NIST A129	0.209	0.404
NMi-VSL 226246	0.326	0.998
NPL 221481	-0.209	0.508
NPL 221483	-0.323	0.494
NPL 221485	-0.547	0.494
NRC A138	-0.099	1.142
NRC A140	-0.149	1.142
PTB 229074	0.456	0.704
PTB 229075	0.232	0.654
VNIIFTRI 79	0.076	1.161
VNIIFTRI 89	0.245	1.235
PTB 229074	0.125	0.472
PTB 229075	-0.183	0.472
NPL 221481	0.039	0.472
NPL 221485	-0.037	0.472
VSL 226246	0.057	0.472
INRIM 232324	-0.620	1.352
INRIM B190	0.294	1.067
INTiBS B178	0.767	1.067

**CCT-K1 and EURAMET.T-K1 : Nominal temperature, $T_{90} = 20.299$ K
Degrees of equivalence, D_i , and expanded uncertainties ($k = 2$), U_i , expressed in mK**



Key comparison CCT-K1 and EURAMET.T-K1

MEASURAND : Temperature T_{90} / K

NOMINAL TEMPERATURE : $T_{90} = 21.575$ K

CCT-K1

Laboratory individual measurements

Lab: Laboratory

S/N i : serial number of Thermometer i

T_i : temperature value for Thermometer i

u_i : standard uncertainty in the calibration of Thermometer i

u_{comp} : standard uncertainty in the comparison measurements

$$u_{\text{comp}} = 0.153 \text{ mK}$$

Lab, S/N i	T_i / mK	u_i / mK
NIST A128	21.575775	0.142
NIST A129	21.575754	0.142
NMI-VSL 226246	21.575850	0.514
NPL 221481	21.575120	0.207
NPL 221483	21.574976	0.198
NPL 221485	21.574975	0.198
NRC A138	21.575266	0.550
NRC A140	21.575114	0.550
PTB 229074	21.575934	0.321
PTB 229075	21.575693	0.292
VNIIFTRI 79	21.575227	0.560
VNIIFTRI 89	21.575344	0.602

Key comparison CCT-K1 and EURAMET.T-K1


MEASURAND : Temperature T_{90} / K

KEY COMPARISON REFERENCE VALUE: $T_R = 21.575444$ K

Matrix of equivalence

Lab, S/N i	D_i	U_i / mK
NIST A128	0.331	0.417
NIST A129	0.310	0.417
NMi-VSL 226246	0.406	1.073
NPL 221481	-0.324	0.515
NPL 221483	-0.468	0.500
NPL 221485	-0.469	0.500
NRC A138	-0.178	1.142
NRC A140	-0.330	1.142
PTB 229074	0.490	0.711
PTB 229075	0.249	0.660
VNIIFTRI 79	-0.217	1.161
VNIIFTRI 89	-0.100	1.242
PTB 229074	0.063	0.472
PTB 229075	-0.212	0.472
NPL 221481	0.060	0.472
NPL 221485	-0.084	0.472
VSL 226246	0.172	0.472
INRIM 232324	-0.607	1.352
INRIM B190	0.269	1.067
INTiBS B178	0.620	1.067

CCT-K1 and EURAMET.T-K1: Nominal temperature, $T_{90} = 21.575$ K
Degrees of equivalence, D_i , and expanded uncertainties ($k = 2$), U_i , expressed in mK



Key comparison CCT-K1 and EURAMET.T-K1

MEASURAND : Temperature T_{90} / K

NOMINAL TEMPERATURE : $T_{90} = 22.677$ K

CCT-K1

Laboratory individual measurements

Lab: Laboratory

S/N i : serial number of Thermometer i

T_i : temperature value for Thermometer i

u_i : standard uncertainty in the calibration of Thermometer i

u_{comp} : standard uncertainty in the comparison measurements

$$u_{\text{comp}} = 0.153 \text{ mK}$$

Lab, S/N i	T_i / mK	u_i / mK
NIST A128	22.677193	0.152
NIST A129	22.677155	0.152
NMI-VSL 226246	22.677427	0.542
NPL 221481	22.676930	0.210
NPL 221483	22.676699	0.201
NPL 221485	22.676646	0.201
NRC A138	22.676586	0.550
NRC A140	22.676571	0.550
PTB 229074	22.677453	0.324
PTB 229075	22.677222	0.295
VNIIFTRI 79	22.676869	0.560
VNIIFTRI 89	22.676872	0.605

Key comparison CCT-K1 and EURAMET.T-K1


MEASURAND : Temperature T_{90} / K

KEY COMPARISON REFERENCE VALUE: $T_R = 22.676998$ K

Matrix of equivalence

Lab, S/N i	D_i	U_i / mK
NIST A128	0.195	0.431
NIST A129	0.157	0.431
NMi-VSL 226246	0.429	1.127
NPL 221481	-0.068	0.520
NPL 221483	-0.299	0.506
NPL 221485	-0.352	0.506
NRC A138	-0.412	1.142
NRC A140	-0.427	1.142
PTB 229074	0.455	0.717
PTB 229075	0.224	0.665
VNIIFTRI 79	-0.129	1.161
VNIIFTRI 89	-0.126	1.248
PTB 229074	0.121	0.472
PTB 229075	-0.183	0.472
NPL 221481	0.065	0.472
NPL 221485	-0.076	0.472
VSL 226246	0.073	0.472
INRIM 232324	-0.257	1.352
INRIM B190	0.265	1.067
INTiBS B178	0.852	1.067

**CCT-K1 and EURAMET.T-K1 : Nominal temperature, $T_{90} = 22.677$ K
Degrees of equivalence, D_i , and expanded uncertainties ($k = 2$), U_i , expressed in mK**



Key comparison CCT-K1

MEASURAND : Temperature T_{90} / K

NOMINAL TEMPERATURE : $T_{90} = 23.496$ K

Laboratory individual measurements

Lab: Laboratory

S/N i : serial number of Thermometer i

T_i : temperature value for Thermometer i

u_i : standard uncertainty in the calibration of Thermometer i

u_{comp} : standard uncertainty in the comparison measurements

$$u_{\text{comp}} = 0.153 \text{ mK}$$

Lab, S/N i	T_i / mK	u_i / mK
NIST A128	23.496504	0.162
NIST A129	23.496495	0.162
NMi-VSL 226246	23.496942	0.563
NPL 221481	23.496400	0.213
NPL 221483	23.496171	0.204
NPL 221485	23.496328	0.204
NRC A138	23.496000	0.550
NRC A140	23.496009	0.550
PTB 229074	23.496996	0.327
PTB 229075	23.496714	0.297
VNIIFTRI 79	23.496311	0.560
VNIIFTRI 89	23.496381	0.607

Key comparison CCT-K1

MEASURAND : Temperature T_{90} / K

KEY COMPARISON REFERENCE VALUE: $T_R = 23.496448$ K


Matrix of equivalence

Lab, S/N <i>i</i>	Lab, S/N <i>j</i> \longrightarrow													
	D_i U_i / mK		NIST A128		NIST A129		NMI-VSL 226246		NPL 221481		NPL 221483		NPL 221485	
	D_{ij} / mK	U_{ij} / mK	D_{ij} / mK	U_{ij} / mK	D_{ij} / mK	U_{ij} / mK	D_{ij} / mK	U_{ij} / mK	D_{ij} / mK	U_{ij} / mK	D_{ij} / mK	U_{ij} / mK	D_{ij} / mK	U_{ij} / mK
NIST A128	0.056	0.446	0.010	0.630	-0.438	1.249	0.104	0.688	0.333	0.677	0.176	0.677		
NIST A129	0.047	0.446	-0.010	0.630	-0.448	1.249	0.094	0.688	0.324	0.677	0.167	0.677		
NMI-VSL 226246	0.494	1.166	0.438	1.249	0.448	1.249	0.542	1.279	0.771	1.273	0.614	1.273		
NPL 221481	-0.048	0.524	-0.104	0.688	-0.094	0.688	-0.542	1.279			0.229	0.731	0.073	0.731
NPL 221483	-0.277	0.510	-0.333	0.677	-0.324	0.677	-0.771	1.273	-0.229	0.731			-0.157	0.721
NPL 221485	-0.120	0.510	-0.176	0.677	-0.167	0.677	-0.614	1.273	-0.073	0.731	0.157	0.721		
NRC A138	-0.448	1.142	-0.504	1.226	-0.495	1.226	-0.942	1.632	-0.401	1.256	-0.171	1.250	-0.328	1.250
NRC A140	-0.439	1.142	-0.495	1.226	-0.486	1.226	-0.934	1.632	-0.392	1.256	-0.162	1.250	-0.319	1.250
PTB 229074	0.548	0.722	0.492	0.848	0.501	0.848	0.054	1.371	0.596	0.892	0.825	0.883	0.668	0.883
PTB 229075	0.266	0.669	0.210	0.804	0.219	0.804	-0.228	1.344	0.314	0.850	0.543	0.841	0.386	0.841
VNIIFTRI 79	-0.137	1.161	-0.193	1.244	-0.184	1.244	-0.631	1.646	-0.090	1.274	0.140	1.268	-0.017	1.268
VNIIFTRI 89	-0.067	1.252	-0.124	1.329	-0.114	1.329	-0.562	1.711	-0.020	1.358	0.210	1.352	0.053	1.352

Matrix of equivalence (Continued)

Lab, S/N <i>i</i>	Lab, S/N <i>j</i> \longrightarrow													
	NRC A138		NRC A140		PTB 229074		PTB 229075		VNIIFTRI 79		VNIIFTRI 89			
	D_{ij} / mK	U_{ij} / mK	D_{ij} / mK	U_{ij} / mK	D_{ij} / mK	U_{ij} / mK	D_{ij} / mK	U_{ij} / mK	D_{ij} / mK	U_{ij} / mK	D_{ij} / mK	U_{ij} / mK		
NIST A128	0.056	0.446	0.504	1.226	0.495	1.226	-0.492	0.848	-0.210	0.804	0.193	1.244	0.124	1.329
NIST A129	0.047	0.446	0.495	1.226	0.486	1.226	-0.501	0.848	-0.219	0.804	0.184	1.244	0.114	1.329
NMI-VSL 226246	0.494	1.166	0.942	1.632	0.934	1.632	-0.054	1.371	0.228	1.344	0.631	1.646	0.562	1.711
NPL 221481	-0.048	0.524	0.401	1.256	0.392	1.256	-0.596	0.892	-0.314	0.850	0.090	1.274	0.020	1.358
NPL 221483	-0.277	0.510	0.171	1.250	0.162	1.250	-0.825	0.883	-0.543	0.841	-0.140	1.268	-0.210	1.352
NPL 221485	-0.120	0.510	0.328	1.250	0.319	1.250	-0.668	0.883	-0.386	0.841	0.017	1.268	-0.053	1.352
NRC A138	-0.448	1.142			-0.009	1.615	-0.996	1.351	-0.714	1.323	-0.311	1.628	-0.381	1.695
NRC A140	-0.439	1.142	0.0089	1.6147			-0.987	1.351	-0.705	1.323	-0.302	1.628	-0.372	1.695
PTB 229074	0.548	0.722	0.996	1.351	0.9874	1.3507			0.282	0.984	0.685	1.367	0.615	1.445
PTB 229075	0.266	0.669	0.714	1.323	0.705	1.323	-0.2822	0.9838			0.403	1.340	0.333	1.420
VNIIFTRI 79	-0.137	1.161	0.311	1.628	0.302	1.628	-0.685	1.367	-0.403	1.3399			-0.070	1.708
VNIIFTRI 89	-0.067	1.252	0.381	1.695	0.372	1.695	-0.615	1.445	-0.333	1.420	0.0699	1.7077		

CCT-K1 : Nominal temperature, $T_{90} = 23.496$ K
Degrees of equivalence, D_i , and expanded uncertainties ($k = 2$), U_i , expressed in mK



Key comparison CCT-K1

MEASURAND : Temperature T_{90} / K

NOMINAL TEMPERATURE : $T_{90} = 24.102$ K

Laboratory individual measurements

Lab: Laboratory

S/N i : serial number of Thermometer i

T_i : temperature value for Thermometer i

u_i : standard uncertainty in the calibration of Thermometer i

u_{comp} : standard uncertainty in the comparison measurements

$$u_{\text{comp}} = 0.153 \text{ mK}$$

Lab, S/N i	T_i / mK	u_i / mK
NIST A128	24.101929	0.168
NIST A129	24.102011	0.168
NMi-VSL 226246	24.102501	0.578
NPL 221481	24.101920	0.215
NPL 221483	24.101771	0.206
NPL 221485	24.101752	0.206
NRC A138	24.101887	0.550
NRC A140	24.101930	0.550
PTB 229074	24.102615	0.329
PTB 229075	24.102251	0.299
VNIIFTRI 79	24.101651	0.560
VNIIFTRI 89	24.101785	0.609

Key comparison CCT-K1

MEASURAND : Temperature T_{90} / K

KEY COMPARISON REFERENCE VALUE: $T_R = 24.101970$ K


Matrix of equivalence

Lab, S/N <i>i</i>	Lab, S/N <i>j</i> \longrightarrow					
	D_i	U_i / mK	D_j	U_j / mK	D_j	U_j / mK
NIST A128	-0.041	0.454				
NIST A129	0.041	0.454	-0.082	0.643	-0.573	1.279
NMi-VSL 226246	0.531	1.196	0.082	0.643	-0.491	1.279
NPL 221481	-0.050	0.527	0.573	1.279	0.581	1.307
NPL 221483	-0.199	0.513	-0.009	0.696	-0.581	1.307
NPL 221485	-0.218	0.513	-0.158	0.685	-0.731	1.301
NRC A138	-0.083	1.142	-0.177	0.685	-0.259	0.685
NRC A140	-0.040	1.142	-0.042	1.229	-0.124	1.229
PTB 229074	0.645	0.725	0.001	1.229	-0.614	1.653
PTB 229075	0.281	0.671	0.686	0.856	0.114	1.398
VNIIFTRI 79	-0.319	1.161	0.322	0.811	0.278	1.247
VNIIFTRI 89	-0.185	1.256	0.124	1.229	-0.251	1.371
	-0.144	1.335	-0.278	1.247	-0.850	1.667
	-0.226	1.335	-0.717	1.734	-0.135	1.362
					0.014	1.356
					0.033	1.356

Matrix of equivalence (Continued)

Lab, S/N <i>i</i>	Lab, S/N <i>j</i> \longrightarrow					
	D_i	U_i / mK	D_j	U_j / mK	D_j	U_j / mK
NIST A128	-0.041	0.454				
NIST A129	0.041	0.454	0.042	1.229	-0.001	1.229
NMi-VSL 226246	0.531	1.196	0.124	1.229	0.081	1.229
NPL 221481	-0.050	0.527	0.614	1.653	0.571	1.653
NPL 221483	-0.199	0.513	0.033	1.258	-0.010	1.258
NPL 221485	-0.218	0.513	-0.116	1.252	-0.159	1.252
NRC A138	-0.083	1.142	-0.135	1.252	-0.178	1.252
NRC A140	-0.040	1.142	0.0431	1.6147	-0.043	1.615
PTB 229074	0.645	0.725	0.728	1.352	-0.685	1.352
PTB 229075	0.281	0.671	0.364	1.352	-0.321	1.325
VNIIFTRI 79	-0.319	1.161	0.364	0.988	0.964	1.369
VNIIFTRI 89	-0.185	1.256	-0.236	1.628	0.236	1.628
	-0.102	1.697	-0.145	1.697	-0.830	1.450
	-0.466	1.424	0.1339	1.7102		
					0.145	1.697
					0.830	1.450
					0.466	1.424
					-0.134	1.710

CCT-K1 : Nominal temperature, $T_{90} = 24.102$ K
Degrees of equivalence, D_i , and expanded uncertainties ($k = 2$), U_i , expressed in mK



Key comparison CCT-K1

MEASURAND : Temperature T_{90} / K

NOMINAL TEMPERATURE : $T_{90} = 24.340$ K

Laboratory individual measurements

Lab: Laboratory

S/N i : serial number of Thermometer i

T_i : temperature value for Thermometer i

u_i : standard uncertainty in the calibration of Thermometer i

u_{comp} : standard uncertainty in the comparison measurements

$$u_{\text{comp}} = 0.153 \text{ mK}$$

Lab, S/N i	T_i / mK	u_i / mK
NIST A128	24.340191	0.174
NIST A129	24.340214	0.174
NMiVSL 226246	24.340765	0.584
NPL 221481	24.340270	0.215
NPL 221485	24.340219	0.206
NRC A138	24.340051	0.550
NRC A140	24.339870	0.550
PTB 229074	24.340847	0.329
PTB 229075	24.340698	0.299

Key comparison CCT-K1

MEASURAND : Temperature T_{90} / K

KEY COMPARISON REFERENCE VALUE: $T_R = 24.340317$ K


Matrix of equivalence

Lab, S/N i	Lab, S/N j \longrightarrow																						
	D_i / mK		U_i		D_{ij} / mK		U_{ij}		D_{ij} / mK		U_{ij}		D_{ij} / mK		U_{ij}		D_{ij} / mK		U_{ij}				
NIST A128	-0.126	0.463				-0.023	0.655	-0.574	1.293	-0.079	0.703	-0.028	0.692	0.140	1.232								
NIST A129	-0.103	0.463				0.023	0.655			-0.056	0.703	-0.006	0.692	0.163	1.232								
NMiVSL 226246	0.448	1.207				0.574	1.293	0.552	1.293			0.495	1.318	0.546	1.312	0.715	1.662						
NPL 221481	-0.047	0.528				0.079	0.703	0.056	0.703	-0.495	1.318			0.051	0.737	0.219	1.258						
NPL 221485	-0.098	0.514				0.028	0.692	0.006	0.692	-0.546	1.312	-0.051	0.737			0.169	1.252						
NRC A138	-0.266	1.142				-0.140	1.232	-0.163	1.232	-0.715	1.662	-0.219	1.258	-0.169	1.252								
NRC A140	-0.447	1.142				-0.321	1.232	-0.344	1.232	-0.896	1.662	-0.400	1.258	-0.350	1.252	-0.181	1.615						
PTB 229074	0.530	0.726				0.656	0.861	0.633	0.861	0.081	1.409	0.577	0.898	0.627	0.890	0.796	1.353						
PTB 229075	0.381	0.673				0.507	0.817	0.484	0.817	-0.067	1.382	0.428	0.855	0.479	0.846	0.647	1.325						

Matrix of equivalence (Continued)

Lab, S/N i	Lab, S/N j \longrightarrow												
	D_i / mK		U_i		D_{ij} / mK		U_{ij}		D_{ij} / mK		U_{ij}		
NIST A128	-0.126	0.463				0.321	1.232	-0.656	0.861	-0.507	0.817		
NIST A129	-0.103	0.463				0.344	1.232	-0.633	0.861	-0.484	0.817		
NMiVSL 226246	0.448	1.207				0.896	1.662	-0.081	1.409	0.067	1.382		
NPL 221481	-0.047	0.528				0.400	1.258	-0.577	0.898	-0.428	0.855		
NPL 221485	-0.098	0.514				0.350	1.252	-0.627	0.890	-0.479	0.846		
NRC A138	-0.266	1.142				0.181	1.615	-0.796	1.353	-0.647	1.325		
NRC A140	-0.447	1.142						-0.977	1.353	-0.828	1.325		
PTB 229074	0.530	0.726				0.977	1.3532			0.149	0.990		
PTB 229075	0.381	0.673				0.828	1.325	-0.149	0.9899				

CCT-K1 : Nominal temperature, $T_{90} = 24.340$ K
Degrees of equivalence, D_i , and expanded uncertainties ($k = 2$), U_i , expressed in mK



Key comparison CCT-K1 and EURAMET.T-K1

MEASURAND : Temperature T_{90} / K

NOMINAL TEMPERATURE : $T_{90} = 24.446$ K

CCT-K1

Laboratory individual measurements

Lab: Laboratory

S/N i : serial number of Thermometer i

T_i : temperature value for Thermometer i

u_i : standard uncertainty in the calibration of Thermometer i

u_{comp} : standard uncertainty in the comparison measurements

$$u_{\text{comp}} = 0.153 \text{ mK}$$

Lab, S/N i	T_i / mK	u_i / mK
NIST A128	24.446295	0.174
NIST A129	24.446429	0.174
NMi-VSL 226246	24.446981	0.587
NPL 221481	24.446370	0.216
NPL 221483	24.446221	0.207
NPL 221485	24.446209	0.207
NRC A138	24.446435	0.550
NRC A140	24.446466	0.550
PTB 229074	24.447000	0.330
PTB 229075	24.446767	0.300
VNIIFTRI 79	24.446009	0.560
VNIIFTRI 89	24.446107	0.610

Key comparison CCT-K1 and EURAMET.T-K1


MEASURAND : Temperature T_{90} / K

KEY COMPARISON REFERENCE VALUE: $T_R = 24.446403$ K

Matrix of equivalence

Lab, S/N i	D_i	U_i / mK
NIST A128	-0.108	0.463
NIST A129	0.026	0.463
NMI-VSL 226246	0.578	1.212
NPL 221481	-0.033	0.529
NPL 221483	-0.182	0.514
NPL 221485	-0.194	0.514
NRC A138	0.032	1.142
NRC A140	0.063	1.142
PTB 229074	0.597	0.727
PTB 229075	0.364	0.673
VNIIFTRI 79	-0.394	1.161
VNIIFTRI 89	-0.296	1.257
PTB 229074	0.081	0.472
PTB 229075	-0.147	0.472
NPL 221481	0.090	0.472
NPL 221485	-0.134	0.472
VSL 226246	0.109	0.472
INRIM 232324	0.307	1.352
INRIM B190	0.446	1.067
INTIBS B178	0.844	1.067

**CCT-K1 and EURAMET.T-K1 : Nominal temperature, $T_{90} = 24.446$ K
Degrees of equivalence, D_i , and expanded uncertainties ($k = 2$), U_i , expressed in mK**



Key comparison CCT-K1

MEASURAND : Temperature T_{90} / K

NOMINAL TEMPERATURE : $T_{90} = 24.551$ K

Laboratory individual measurements

Lab: Laboratory

S/N i : serial number of Thermometer i

T_i : temperature value for Thermometer i

u_i : standard uncertainty in the calibration of Thermometer i

u_{comp} : standard uncertainty in the comparison measurements

$$u_{\text{comp}} = 0.153 \text{ mK}$$

Lab, S/N i	T_i / mK	u_i / mK
NIST A128	24.551286	0.174
NIST A129	24.551288	0.174
NMi-VSL 226246	24.551814	0.588
NPL 221481	24.551375	0.216
NPL 221483	24.551221	0.207
NPL 221485	24.551122	0.207
NRC A138	24.551544	0.550
NRC A140	24.551501	0.550
PTB 229074	24.551980	0.330
PTB 229075	24.551635	0.300
VNIIFTRI 79	24.550982	0.560
VNIIFTRI 89	24.551080	0.610

Key comparison CCT-K1

MEASURAND : Temperature T_{90} / K

KEY COMPARISON REFERENCE VALUE: $T_R = 24.551354$ K

Matrix of equivalence

Lab, S/N <i>i</i>	Lab, S/N <i>j</i> \longrightarrow					
	D_i	U_i	$/ \text{mK}$	D_j	U_j	$/ \text{mK}$
NIST A128	-0.068	0.463		-0.002	0.655	
NIST A129	-0.066	0.463		-0.528	1.300	
NMi-VSL 226246	0.460	1.215		0.526	1.300	
NPL 221481	0.021	0.529		0.088	0.703	
NPL 221483	-0.133	0.515		0.086	0.703	
NPL 221485	-0.232	0.515		0.067	0.693	
NRC A138	0.190	1.142		0.164	0.693	
NRC A140	0.147	1.142		0.166	0.693	
PTB 229074	0.626	0.727		0.692	1.319	
PTB 229075	0.281	0.674		0.253	0.738	
VNIIFTRI 79	-0.372	1.161		0.154	0.738	
VNIIFTRI 89	-0.274	1.258		0.253	0.738	
	-0.207	1.341		0.099	0.728	
	-0.209	1.341		0.253	0.738	
	-0.735	1.749		0.140	1.270	
	-0.295	1.365		-0.142	1.359	
	-0.43	1.359		-0.043	1.359	

Matrix of equivalence (Continued)

Lab, S/N <i>i</i>	Lab, S/N <i>j</i> \longrightarrow					
	D_i	U_i	$/ \text{mK}$	D_j	U_j	$/ \text{mK}$
NIST A128	-0.068	0.463		-0.214	1.232	
NIST A129	-0.066	0.463		-0.694	0.862	
NMi-VSL 226246	0.460	1.215		-0.348	0.817	
NPL 221481	0.021	0.529		0.304	1.250	
NPL 221483	-0.133	0.515		0.207	1.341	
NPL 221485	-0.232	0.515		-0.256	1.232	
NRC A138	0.190	1.142		-0.692	0.862	
NRC A140	0.147	1.142		-0.346	0.817	
PTB 229074	0.626	0.727		0.306	1.250	
PTB 229075	0.281	0.674		0.209	1.341	
VNIIFTRI 79	-0.372	1.161		-0.224	1.232	
VNIIFTRI 89	-0.274	1.258		-0.142	1.359	
	-0.43	1.359		-0.043	1.359	
	-0.480	1.354		-0.134	1.326	
	-0.346	0.991		0.518	1.628	
	0.998	1.370		0.421	1.699	
	0.652	1.342		0.043	1.359	
	0.555	1.427		0.464	1.699	
	-0.097	1.712		0.437	1.354	
	-0.464	1.699		0.091	1.453	
	-0.421	1.699		0.134	1.326	
	-0.901	1.453		-0.562	1.628	
	-0.555	1.427		-0.652	1.3423	
	0.0973	1.7119		-0.097	1.712	

CCT-K1 : Nominal temperature, $T_{90} = 24.551$ K
Degrees of equivalence, D_i , and expanded uncertainties ($k = 2$), U_i , expressed in mK

