

# CCEM.RF-K10.CL and EUROMET.EM.RF-K10.CL

Key comparison CCEM.RF-K10.CL

**MEASURAND : Calibration factor**

**TRANSFER STANDARDS : Power sensors PTB 1-3, PTB 1-3-1, PTB 2-6 and PTB 2-6-1**

**FREQUENCY : from 50 MHz to 26 GHz**

$x_i$ : result of measurement carried out by laboratory  $i$

$u_i$ : combined standard uncertainty of  $x_i$

Power sensor PTB 1-3, relabelled as PTB 1-3-1 after repair at PTB in November 2001 (see on page 10 of the Final Report)

Lab $i$	Frequency $\Rightarrow$													
	50 MHz		1 GHz		10 GHz		18 GHz		20 GHz		23 GHz		26 GHz	
	$x_i$	$u_i$	$x_i$	$u_i$	$x_i$	$u_i$	$x_i$	$u_i$	$x_i$	$u_i$	$x_i$	$u_i$	$x_i$	$u_i$
NMIJ	0.9828	0.0014	0.9720	0.0017	0.9438	0.0020	0.9352	0.0058	0.9328	0.0039	0.9262	0.0044	0.9020	0.0069
NRC	0.9837	0.0040	0.9763	0.0040	0.9404	0.0040	0.9378	0.0048	0.9301	0.0048	0.9284	0.0048	0.9258	0.0048
NIST	0.9887	0.0086	0.9741	0.0087	0.9417	0.0088	0.9290	0.0088	0.9259	0.0090	0.9225	0.0092	0.9145	0.0094
METAS	0.9810	0.0060	0.9720	0.0070	0.9410	0.0100	0.9280	0.0100	0.9290	0.0110	0.9280	0.0110	0.9090	0.0120
CSIR-NML	0.9920	0.0070	0.9780	0.0080	0.9460	0.0110	0.9280	0.0130	0.9320	0.0150	0.9310	0.0160	0.9250	0.0190
PTB	0.9877	0.0020	0.9741	0.0025	0.9428	0.0040	0.9304	0.0050	0.9263	0.0055	0.9251	0.0065	0.9227	0.0085
NMIA	0.9880	0.0020	0.9760	0.0025	0.9420	0.0047	0.9330	0.0063	0.9320	0.0093	0.9270	0.0092	0.9200	0.0091
NPL	0.9857	0.0015	0.9740	0.0030	0.9390	0.0044	0.9290	0.0055	0.9310	0.0055	0.9310	0.0066	0.9270	0.0077
MIRS/SIQ	0.9890	0.0030	0.9750	0.0030	0.9380	0.0070	0.9370	0.0080	0.9300	0.0080	0.9260	0.0080	0.9190	0.0080
INRIM	0.9895	0.0052	0.9817	0.0063	0.9344	0.0135	0.9273	0.0138	0.9192	0.0131	0.9115	0.0190	0.9003	0.0187
VNIIFTRI	-	-	0.9800	0.0030	0.9610	0.0060	0.9110	0.0070	0.9080	0.0110	0.9180	0.0120	0.9240	0.0140
PTB	0.9897	0.0020	0.9821	0.0025	0.9362	0.0040	0.8982	0.0050	0.8814	0.0055	0.8578	0.0065	0.8580	0.0085
SPRING Singapore	0.9936	0.0067	0.9821	0.0067	0.9384	0.0074	0.8933	0.0098	0.8840	0.0099	0.8590	0.0130	0.8583	0.0084

Laboratory having reported results for transfer standard PTB 1-3-1

## Key comparison CCEM.RF-K10.CL

Power sensor PTB 2-6, relabelled as PTB 2-6-1 after repair at PTB in November 2001 (see on page 10 of the Final Report)

Lab <i>i</i> ↓	Frequency →													
	50 MHz		1 GHz		10 GHz		18 GHz		20 GHz		23 GHz		26 GHz	
	$x_i$	$u_i$	$x_i$	$u_i$	$x_i$	$u_i$	$x_i$	$u_i$	$x_i$	$u_i$	$x_i$	$u_i$	$x_i$	$u_i$
NMIJ	0.9845	0.0014	0.9764	0.0017	0.9402	0.0019	0.9092	0.0038	0.8973	0.0047	0.8784	0.0044	0.8686	0.0075
NRC	-	-	-	-	-	-	-	-	-	-	-	-	-	-
NIST	0.9910	0.0086	0.9784	0.0087	0.9406	0.0088	0.9074	0.0088	0.8952	0.0090	0.8779	0.0092	0.8778	0.0095
METAS	0.9850	0.0060	0.9770	0.0070	0.9390	0.0100	0.9060	0.0100	0.9000	0.0110	0.8820	0.0100	0.8700	0.0120
CSIR-NML	0.9930	0.0070	0.9820	0.0080	0.9430	0.0110	0.9030	0.0130	0.9010	0.0150	0.8820	0.0160	0.8810	0.0190
PTB	0.9889	0.0020	0.9784	0.0025	0.9376	0.0040	0.9048	0.0050	0.8904	0.0055	0.8769	0.0065	0.8805	0.0085
NMIA	0.9900	0.0020	0.9810	0.0025	0.9400	0.0047	0.9110	0.0063	0.9000	0.0093	0.8810	0.0092	0.8830	0.0093
NPL	0.9872	0.0017	0.9780	0.0032	0.9360	0.0045	0.9070	0.0055	0.8970	0.0056	0.8840	0.0067	0.8890	0.0077
MIRS/SIQ	0.9910	0.0030	0.9790	0.0030	0.9310	0.0070	0.9110	0.0080	0.9020	0.0080	0.8800	0.0080	0.8790	0.0080
INRIM	0.9951	0.0058	0.9857	0.0058	0.9335	0.0153	0.9231	0.0146	0.9000	0.0145	0.9013	0.0207	0.9032	0.0206
VNIIFTRI	-	-	0.9850	0.0060	0.9590	0.0100	0.8850	0.0120	0.8710	0.0220	0.8710	0.0290	0.8950	0.0150
PTB	0.9883	0.0020	0.9780	0.0025	0.9402	0.0040	0.9104	0.0050	0.9075	0.0055	0.8857	0.0065	0.8905	0.0085
SPRING Singapore	0.9892	0.0067	0.9775	0.0067	0.9404	0.0071	0.9071	0.0100	0.9044	0.0103	0.8866	0.0134	0.8945	0.0088

Laboratory having reported results for transfer standard PTB 2-6-1

Lab <i>i</i> ↓	Date of measurement
NMIJ	Jul - Aug 2000
NRC	Aug - Sep 2000
NIST	Oct - Nov 2000
METAS	Nov - Dec 2000
CSIR-NML	Dec 2000 - Jan 2001
PTB	Feb 2001
NMIA	Mar - Apr 2001

Lab <i>i</i> ↓	Date of measurement
NPL	Apr - May 2001
MIRS/SIQ	May - Jun 2001
INRIM	Jun - Jul 2001
VNIIFTRI	Sep - Oct 2001
PTB	Nov 2001
SPRING Singapore	Dec 2001 - Jan 2002

**Key comparison EUROMET.EM.RF-K10.CL**

**MEASURAND : Calibration factor**

**TRANSFER STANDARD : Power sensor PTB 1-03-2**

**FREQUENCY : from 50 MHz to 26 GHz**

$x_{i\text{-EUR}}$  : result of measurement carried out by laboratory  $i$  participant in EUROMET.EM.RF-K10.CL

$u_{i\text{-EUR}}$  : combined standard uncertainty of  $x_{i\text{-EUR}}$

Lab $i$	Frequency $\Rightarrow$													
	50 MHz		1 GHz		10 GHz		18 GHz		20 GHz		23 GHz		26 GHz	
	$x_{i\text{-EUR}}$	$u_{i\text{-EUR}}$	$x_{i\text{-EUR}}$	$u_{i\text{-EUR}}$	$x_{i\text{-EUR}}$	$u_{i\text{-EUR}}$	$x_{i\text{-EUR}}$	$u_{i\text{-EUR}}$	$x_{i\text{-EUR}}$	$u_{i\text{-EUR}}$	$x_{i\text{-EUR}}$	$u_{i\text{-EUR}}$	$x_{i\text{-EUR}}$	$u_{i\text{-EUR}}$
EIM	0.9880	0.0070	0.9840	0.0080	0.9440	0.0100	0.9440	0.0140	0.9060	0.0130	0.9310	0.0140	0.9140	0.0150
SP	0.9900	0.0029	0.9830	0.0032	0.9560	0.0048	0.9400	0.0066	0.9420	0.0066	0.9370	0.0080	0.9420	0.0104
NPL	0.9912	0.0012	0.9823	0.0030	0.9532	0.0044	0.9426	0.0053	0.9385	0.0053	0.9387	0.0064	0.9386	0.0075
AREPA	0.9990	0.0040	0.9910	0.0040	0.9620	0.0080	0.9520	0.0110	0.9540	0.0140	0.9510	0.0170	0.9440	0.0170
NMi-VSL	0.9764	0.0038	0.9679	0.0051	0.9454	0.0070	0.9340	0.0086	0.9329	0.0086	0.9271	0.0087	0.9268	0.0097
LNE	-	-	-	-	-	-	-	-	-	-	-	-	-	-
UME	-	-	-	-	-	-	-	-	-	-	-	-	-	-
PTB	0.9917	0.0020	0.9832	0.0025	0.9601	0.0040	0.9484	0.0050	0.9418	0.0055	0.9440	0.0065	0.9366	0.0085

Key comparison EUROMET.EM.RF-K10.CL

TRANSFER STANDARD : Power sensor PTB 2-06-2

Lab <i>i</i> ↓	Frequency →													
	50 MHz		1 GHz		10 GHz		18 GHz		20 GHz		23 GHz		26 GHz	
	$x_{i\text{-EUR}}$	$u_{i\text{-EUR}}$												
EIM	-	-	-	-	-	-	-	-	-	-	-	-	-	-
SP	0.9870	0.0029	0.9790	0.0032	0.9370	0.0047	0.9040	0.0066	0.8970	0.0066	0.8740	0.0080	0.8820	0.0105
NPL	0.9880	0.0012	0.9777	0.0030	0.9341	0.0044	0.9053	0.0054	0.8944	0.0055	0.8748	0.0066	0.8783	0.0077
AREPA	0.9940	0.0040	0.9850	0.0040	0.9420	0.0070	0.9180	0.0110	0.9090	0.0130	0.8850	0.0150	0.8800	0.0170
NMi-VSL	0.9807	0.0038	0.9708	0.0051	0.9314	0.0069	0.9091	0.0083	0.8962	0.0083	0.8729	0.0082	0.8696	0.0091
LNE	0.9852	0.0023	0.9761	0.0062	0.9494	0.0085	0.9167	0.0119	0.9022	0.0128	0.8918	0.0107	0.8837	0.0101
UME	0.9922	0.0028	0.9820	0.0027	0.9508	0.0039	0.9233	0.0082	0.9140	0.0083	0.9002	0.0110	0.9037	0.0181
PTB	0.9884	0.0020	0.9782	0.0025	0.9388	0.0080	0.9156	0.0050	0.8998	0.0055	0.8783	0.0065	0.8767	0.0085

Lab <i>i</i> ↓	Date of measurement
EIM	May 2002
SP	Jun 2002
NPL	Jul - Aug 2002
AREPA	Aug - Sep 2002
NMi-VSL	Sep - Oct 2002
LNE	Nov 2002
UME	Feb - Mar 2003
PTB	Mar - Apr 2002 Dec 2002 - Jan 2003 Apr 2003

## Key comparison CCEM.RF-K10.CL

**MEASURAND :** Calibration factor

**TRANSFER STANDARDS :** Power sensors PTB 1-3, PTB 1-3-1, PTB 2-6 and PTB 2-6-1

The computation of the key comparison reference values,  $x_R$ , and of their uncertainties,  $u_R$ , is explained in Section 6.2 of the Final Report (pages 14 and 15).

For each frequency and each of the transfer standards PTB 1-3 and PTB 2-6,  $x_R$  is computed as the unweighted arithmetic mean of the participants' results obtained using independant primary power standards with outliers excluded. Its standard uncertainty,  $u_R$ , is the standard deviation of the mean.

For each frequency and each of the repaired transfer standards PTB 1-3-1 and PTB 2-6-1,  $x_R$  is computed using the common measurements carried out by PTB (see on page 15 of the Final Report). Its standard uncertainty,  $u_R$ , is assumed to be equal to the standard uncertainty obtained with the transfer standard before repair.

Frequency ↓	Transfer standard PTB 1-3		Transfer standard PTB 2-6		Transfer standard PTB 1-3 -1		Transfer standard PTB 2-6-1	
	$x_R$	$u_R$	$x_R$	$u_R$	$x_R$	$u_R$	$x_R$	$u_R$
50 MHz	0.9866	0.0010	0.9895	0.0019	0.9895	0.0010	0.9881	0.0019
1 GHz	0.9760	0.0012	0.9804	0.0015	0.9841	0.0012	0.9800	0.0015
10 GHz	0.9416	0.0007	0.9380	0.0012	0.9358	0.0007	0.9398	0.0012
18 GHz	0.9317	0.0014	0.9079	0.0010	0.9004	0.0014	0.9126	0.0010
20 GHz	0.9276	0.0020	0.8960	0.0016	0.8849	0.0020	0.9110	0.0016
23 GHz	0.9252	0.0019	0.8776	0.0021	0.8582	0.0019	0.8861	0.0021
26 GHz	0.9228	0.0022	0.8857	0.0051	0.8607	0.0022	0.8932	0.0051

The degree of equivalence of laboratory  $i$  with respect to the key comparison reference value is given by two terms, both expressed in  $10^{-3}$ :  $D_i = x_i - x_R$  and its expanded uncertainty  $U_i$  ( $k = 2$ ),  $U_i = 2(u_i^2 + u_R^2)^{1/2}$ .

The degree of equivalence between two laboratories  $i$  and  $j$  is given by two terms, both expressed in  $10^{-3}$ :

$D_{ij} = (D_i - D_j)$  and its expanded uncertainty  $U_{ij}$  ( $k = 2$ ),  $U_{ij} = 2(u_i^2 + u_j^2)^{1/2}$ .

## Linking EUROMET.EM.RF-K10.CL to CCEM.RF-K10.CL

The laboratories that are participants in both comparisons (PTB and NPL) are used for linking the results obtained in EUROMET.EM.RF-K10.CL to the results obtained in CCEM.RF-K10.CL. The linking process is explained in section 6.2 of the EUROMET.EM.RF-K10.CL Final Report.

For each transfer standard and a given frequency the degree of equivalence of each laboratory participant in EUROMET.EM.RF-K10.CL with respect to the key comparison reference value is given by a pair of terms both express in  $10^{-3}$ :  $D_i$  and its expanded uncertainty ( $k = 2$ ),  $U_i$ .

The degree of equivalence is computed as explained in section 6.3 of the EUROMET.EM.RF-K10.CL Final Report.

For each transfer standard and a given frequency, the degree of equivalence between two laboratories  $i$  and  $j$  is given by a pair of terms:  $D_{ij}$  and its expanded uncertainty ( $k = 2$ ),  $U_{ij}$ , both express in  $10^{-3}$ :

The pair-wise degrees of equivalence are computed only between two laboratories participants in EUROMET.EM.RF-K10.CL.

It follows that the Matrices of equivalence and the Graphs of equivalence obtained for CCEM.RF-K10.CL are extended with results obtained in EUROMET.EM.RF-K10.CL.

CCEM.RF-K10.CL and EUROMET.EM.RF-K10.CL

#### **MEASURAND : Calibration factor**

**TRANSFER STANDARDS :** Power sensors PTB 1-3, PTB 1-3-1 for CCEM.RF-K10.CL and PTB 1-03-2 for EUROMET.EM-RF-K10.CL

## Matrix of equivalence

**FREQUENCY : 50 MHz**

Lab 1

	$D_i$ $/ 10^{-3}$	$U_i$ $/ 10^{-3}$
NMIJ	-3.8	3.4
NRC	-2.9	8.2
NIST	2.1	17.3
METAS	-5.6	12.2
CSIR-NML	5.4	14.1
PTB	1.1	4.5
NMIA	1.4	4.5
NPL	-0.9	3.6
MIRS/SIQ	2.4	6.3
INRIM	2.9	10.6
SPRING	4.1	13.5

## Lab *j*

NMIJ		NRC		NIST		METAS		CSIR-NML		PTB		NMIA		NPL	
$D_{ij}$ / $10^{-3}$	$U_{ij}$ / $10^{-3}$														
		-0.9	8.5	-5.9	17.4	1.8	12.3	-9.2	14.3	-4.9	4.9	-5.2	4.9	-2.9	4.1
0.9	8.5			-5.0	19.0	2.7	14.4	-8.3	16.1	-4.0	8.9	-4.3	8.9	-2.0	8.5
5.9	17.4	5.0	19.0			7.7	21.0	-3.3	22.2	1.0	17.7	0.7	17.7	3.0	17.5
-1.8	12.3	-2.7	14.4	-7.7	21.0			-11.0	18.4	-6.7	12.6	-7.0	12.6	-4.7	12.4
9.2	14.3	8.3	16.1	3.3	22.2	11.0	18.4			4.3	14.6	4.0	14.6	6.3	14.3
4.9	4.9	4.0	8.9	-1.0	17.7	6.7	12.6	-4.3	14.6			-0.3	5.7	2.0	5.0
5.2	4.9	4.3	8.9	-0.7	17.7	7.0	12.6	-4.0	14.6	0.3	5.7			2.3	5.0
2.9	4.1	2.0	8.5	-3.0	17.5	4.7	12.4	-6.3	14.3	-2.0	5.0	-2.3	5.0		
6.2	6.6	5.3	10.0	0.3	18.2	8.0	13.4	-3.0	15.2	1.3	7.2	1.0	7.2	3.3	6.7
6.7	10.8	5.8	13.1	0.8	20.1	8.5	15.9	-2.5	17.4	1.8	11.1	1.5	11.1	3.8	10.8
7.9	13.7	7.0	15.6	2.0	21.8	9.7	18.0	-1.3	19.4	3.0	14.0	2.7	14.0	5.0	13.7

<b>EIM</b>	<b>-4.1</b>	<b>14.0</b>
<b>SP</b>	<b>-2.1</b>	<b>5.8</b>
<b>NPL</b>	<b>-1.0</b>	<b>2.4</b>
<b>AREPA</b>	<b>6.9</b>	<b>8.0</b>
<b>PTB</b>	<b>-0.4</b>	<b>4.0</b>

Laboratory having used the repaired transfer standard PTB 1-3-1

## **MEASURAND : Calibration factor**

**TRANSFER STANDARDS :** Power sensors PTB 1-3, PTB 1-3-1 for CCEM.RF-K10.CL and PTB 1-03-2 for EUROMET.EM.RF-K10.CL

## Matrix of equivalence

**FREQUENCY : 50 MHz**


Lab <i>i</i>	Lab <i>j</i> →					
	MIRS/SIQ		INRIM		SPRING	
	$D_i$ / 10 <sup>-3</sup>	$U_i$ / 10 <sup>-3</sup>	$D_{ij}$ / 10 <sup>-3</sup>	$U_{ij}$ / 10 <sup>-3</sup>	$D_{ij}$ / 10 <sup>-3</sup>	$U_{ij}$ / 10 <sup>-3</sup>
NMIJ	-3.8	3.4	-6.2	6.6	-6.7	10.8
NRC	-2.9	8.2	-5.3	10.0	-5.8	13.1
NIST	2.1	17.3	-0.3	18.2	-0.8	20.1
METAS	-5.6	12.2	-8.0	13.4	-8.5	15.9
CSIR-NML	5.4	14.1	3.0	15.2	2.5	17.4
PTB	1.1	4.5	-1.3	7.2	-1.8	11.1
NMIA	1.4	4.5	-1.0	7.2	-1.5	11.1
NPL	-0.9	3.6	-3.3	6.7	-3.8	10.8
MIRS/SIQ	2.4	6.3			-0.5	12.0
INRIM	2.9	10.6	0.5	12.0		-1.2
SPRING	4.1	13.5	1.7	14.7	1.2	17.0

<b>EIM</b>	<b>-4.1</b>	14.0				
<b>SP</b>	<b>-2.1</b>	5.8				
<b>NPL</b>	<b>-1.0</b>	2.4				
<b>AREPA</b>	<b>6.9</b>	8.0				
<b>PTB</b>	<b>-0.4</b>	4.0				
						<b>Not computed</b>

		<b>-2.0</b>	15.2	<b>-3.1</b>	14.2	<b>-11.0</b>	16.1	<b>-3.7</b>	14.6
<b>2.0</b>	15.2			<b>-1.1</b>	6.3	<b>-9.0</b>	9.9	<b>-1.7</b>	7.0
<b>3.1</b>	14.2	<b>1.1</b>	6.3			<b>-7.9</b>	8.4	<b>-0.6</b>	4.7
<b>11.0</b>	16.1	<b>9.0</b>	9.9	<b>7.9</b>	8.4			<b>7.3</b>	8.9
<b>3.7</b>	14.6	<b>1.7</b>	7.0	<b>0.6</b>	4.7	<b>-7.3</b>	8.9		

Laboratory having used the repaired transfer standard PTB 1-3-1

**CCEM.RF-K10.CL and EUROMET.EM.RF-K10.CL Calibration factor, frequency: 50 MHz,  
Transfer standard: PTB 1-3**  
**Degrees of equivalence [  $D_i$  and its expanded uncertainty ( $k = 2$ ),  $U_i$  ]**



Open symbols represent values from laboratories in Associate States and Economies of the CGPM  
 Blue triangles: participants in EUROMET.EM.RF-K10.CL

CCEM.RF-K10.CL and EUROMET.EM.RF-K10.CL

#### **MEASURAND : Calibration factor**

**TRANSFER STANDARDS :** Power sensors PTB 1-3, PTB 1-3-1 for CCEM.RF-K10.CL and PTB 1-03-2 for EUROMET.EM.RF-K10.CL

## Matrix of equivalence

**FREQUENCY : 1 GHz**

<b>Lab <i>i</i></b>	<b><math>D_i</math> <math>/ 10^{-3}</math></b>	<b><math>U_i</math> <math>/ 10^{-3}</math></b>
<b>NMIJ</b>	<b>-4.0</b>	<b>4.2</b>
<b>NRC</b>	<b>0.3</b>	<b>8.4</b>
<b>NIST</b>	<b>-1.9</b>	<b>17.6</b>
<b>METAS</b>	<b>-4.0</b>	<b>14.2</b>
<b>CSIR-NML</b>	<b>2.0</b>	<b>16.2</b>
<b>PTB</b>	<b>-1.9</b>	<b>5.5</b>
<b>NMIA</b>	<b>0.0</b>	<b>5.5</b>
<b>NPL</b>	<b>-2.0</b>	<b>6.5</b>
<b>MIRS/SIQ</b>	<b>-1.0</b>	<b>6.5</b>
<b>INRIM</b>	<b>5.7</b>	<b>12.8</b>
<b>VNIIFTRI</b>	<b>4.0</b>	<b>6.5</b>
<b>SPRING</b>	<b>-2.0</b>	<b>13.6</b>

NMJJ		NRC		NIST		METAS		CSIR-NML		PTB		NMIA		NPL		MIRS/SIQ	
$D_{ij}$ / $10^{-3}$	$U_{ij}$ / $10^{-3}$																
		-4.3	8.7	-2.1	17.7	0.0	14.4	-6.0	16.4	-2.1	6.0	-4.0	6.0	-2.0	6.9	-3.0	6.9
4.3	8.7			2.2	19.2	4.3	16.1	-1.7	17.9	2.2	9.4	0.3	9.4	2.3	10.0	1.3	10.0
2.1	17.7	-2.2	19.2			2.1	22.3	-3.9	23.6	0.0	18.1	-1.9	18.1	0.1	18.4	-0.9	18.4
0.0	14.4	-4.3	16.1	-2.1	22.3			-6.0	21.3	-2.1	14.9	-4.0	14.9	-2.0	15.2	-3.0	15.2
6.0	16.4	1.7	17.9	3.9	23.6	6.0	21.3			3.9	16.8	2.0	16.8	4.0	17.1	3.0	17.1
2.1	6.0	-2.2	9.4	0.0	18.1	2.1	14.9	-3.9	16.8			-1.9	7.1	0.1	7.8	-0.9	7.8
4.0	6.0	-0.3	9.4	1.9	18.1	4.0	14.9	-2.0	16.8	1.9	7.1			2.0	7.8	1.0	7.8
2.0	6.9	-2.3	10.0	-0.1	18.4	2.0	15.2	-4.0	17.1	-0.1	7.8	-2.0	7.8			-1.0	8.5
3.0	6.9	-1.3	10.0	0.9	18.4	3.0	15.2	-3.0	17.1	0.9	7.8	-1.0	7.8	1.0	8.5		
9.7	13.1	5.4	14.9	7.6	21.5	9.7	18.8	3.7	20.4	7.6	13.6	5.7	13.6	7.7	14.0	6.7	14.0
8.0	6.9	3.7	10.0	5.9	18.4	8.0	15.2	2.0	17.1	5.9	7.8	4.0	7.8	6.0	8.5	5.0	8.5
2.0	13.8	-2.3	15.6	-0.1	22.0	2.0	19.4	-4.0	20.9	-0.1	14.3	-2.0	14.3	0.0	14.7	-1.0	14.7

<b>EIM</b>	<b>-0.9</b>	16.0
<b>SP</b>	<b>-1.9</b>	6.4
<b>NPL</b>	<b>-2.6</b>	6.0
<b>AREPA</b>	<b>6.1</b>	8.0
<b>PTB</b>	<b>-1.7</b>	5.0

Laboratory having used the repaired transfer standard PTB 1-3-1

**MEASURAND : Calibration factor**

**TRANSFER STANDARDS :** Power sensors PTB 1-3, PTB 1-3-1 for CCEM.RF-K10.CL and PTB 1-03-2 for EUROMET.EM.RF-K10.CL


**Matrix of equivalence**

**FREQUENCY : 1 GHz** (continued)

Lab <i>i</i>	Lab <i>j</i> →																	
	$D_{ij}$ / 10 <sup>-3</sup>	$U_{ij}$ / 10 <sup>-3</sup>	INRIM		VNIIIFTRI		SPRING		EIM		SP		NPL		AREPA		PTB	
NMIJ	-4.0	4.2	-9.7	13.1	-8.0	6.9	-2.0	13.8										
NRC	0.3	8.4	-5.4	14.9	-3.7	10.0	2.3	15.6										
NIST	-1.9	17.6	-7.6	21.5	-5.9	18.4	0.1	22.0										
METAS	-4.0	14.2	-9.7	18.8	-8.0	15.2	-2.0	19.4										
CSIR-NML	2.0	16.2	-3.7	20.4	-2.0	17.1	4.0	20.9										
PTB	-1.9	5.5	-7.6	13.6	-5.9	7.8	0.1	14.3										
NMIA	0.0	5.5	-5.7	13.6	-4.0	7.8	2.0	14.3										
NPL	-2.0	6.5	-7.7	14.0	-6.0	8.5	0.0	14.7										
MIRS/SIQ	-1.0	6.5	-6.7	14.0	-5.0	8.5	1.0	14.7										
INRIM	5.7	12.8			1.7	14.0	7.7	18.4										
VNIIIFTRI	4.0	6.5	-1.7	14.0			6.0	14.7										
SPRING	-2.0	13.6	-7.7	18.4	-6.0	14.7												
EIM	-0.9	16.0							1.0	17.2	1.7	17.1	-7.0	17.9	0.8	16.8		
SP	-1.9	6.4							-1.0	17.2	0.7	8.8	-8.0	10.2	-0.2	8.1		
NPL	-2.6	6.0							-1.7	17.1	-0.7	8.8		-8.7	10.0	-0.9	7.8	
AREPA	6.1	8.0							7.0	17.9	8.0	10.2	8.7	10.0		7.8	9.4	
PTB	-1.7	5.0							-0.8	16.8	0.2	8.1	0.9	7.8	-7.8	9.4		

Laboratory having used the repaired transfer standard PTB 1-3-1

**CCEM.RF-K10.CL and EUROMET.EM.RF-K10.CL Calibration factor, frequency: 1 GHz,  
Transfer standard: PTB 1-3**  
**Degrees of equivalence [  $D_i$  and its expanded uncertainty ( $k = 2$ ),  $U_i$  ]**



Open symbols represent values from laboratories in Associate States and Economies of the CGPM

**Blue triangles:** participants in EUROMET.EM.RF-K10.CL


CCEM.RF-K10.CL and EUROMET.EM.RF-K10.CL

## **MEASURAND : Calibration factor**

**TRANSFER STANDARDS :** Power sensors PTB 1-3, PTB 1-3-1 for CCEM.RF-K10.CL and PTB 1-03-2 for EUROMET.EM.RF-K10.CL

## Matrix of equivalence

FREQUENCY : 10 GHz



<b>EIM</b>	<b>-13.6</b>	20.0
<b>SP</b>	<b>-1.6</b>	9.6
<b>NPL</b>	<b>-4.3</b>	8.8
<b>AREPA</b>	<b>4.4</b>	16.0
<b>PTB</b>	<b>2.5</b>	8.0

**Not computed**

Laboratory having used the repaired transfer standard PTB 1-3-1

MEASURAND : Calibration factor

TRANSFER STANDARDS : Power sensors PTB 1-3, PTB 1-3-1 for CCEM.RF-K10.CL and PTB 1-03-2 for EUROMET.EM.RF-K10.CL

Matrix of equivalence


FREQUENCY : 10 GHz (continued)

Lab <i>i</i>	Lab <i>j</i> →															
	INRIM		VNIIIFTRI		SPRING		EIM		SP		NPL		AREPA		PTB	
	$D_{ij}$ / 10 <sup>-3</sup>	$U_{ij}$ / 10 <sup>-3</sup>														
NMIJ	2.2	4.2														
NRC	-1.2	8.1	9.4	27.3	-17.2	12.6	-0.4	15.3								
NIST	0.1	17.7	6.0	28.2	-20.6	14.4	-3.8	16.8								
METAS	-0.6	20.0	7.3	32.2	-19.3	21.3	-2.5	23.0								
CSIR-NML	4.4	22.0	6.6	33.6	-20.0	23.3	-3.2	24.9								
PTB	1.2	8.1	11.6	34.8	-15.0	25.1	1.8	26.5								
NMIA	0.4	9.5	8.4	28.2	-18.2	14.4	-1.4	16.8								
NPL	-2.6	8.9	7.6	28.6	-19.0	15.2	-2.2	17.5								
MIRS/SIQ	-3.6	14.1	4.6	28.4	-22.0	14.9	-5.2	17.2								
INRIM	-7.2	27.0	3.6	30.4	-23.0	18.4	-6.2	20.4								
VNIIIFTRI	19.4	12.1			-26.6	29.5	-9.8	30.8								
SPRING	2.6	14.9	26.6	29.5		16.8	19.1									
	9.8	30.8	9.8	30.8	-16.8	19.1										
EIM	-13.6	20.0														
SP	-1.6	9.6														
NPL	-4.3	8.8														
AREPA	4.4	16.0														
PTB	2.5	8.0														

Laboratory having used the repaired transfer standard PTB 1-3-1

**CCEM.RF-K10.CL and EUROMET.EM-RF-K10.CL Calibration factor, frequency: 10 GHz,  
Transfer standard: PTB 1- 3**

**Degrees of equivalence [  $D_i$  and its expanded uncertainty ( $k = 2$ ),  $U_i$  ]**



Open symbols represent values from laboratories in Associate States and Economies of the CGPM

**Blue triangles:** participants in EUROMET.EM.RF-K10.CL

CCEM.RF-K10.CL and EUROMET.EM.RF-K10.CL

MEASURAND : Calibration factor

TRANSFER STANDARDS : Power sensors PTB 1-3, PTB 1-3-1 for CCEM.RF-K10.CL and PTB 1-03-2 for EUROMET.EM.RF-K10.CL

Matrix of equivalence

FREQUENCY : 18 GHz

Lab <i>i</i>	$D_i$ / $10^{-3}$	$U_i$ / $10^{-3}$
NMIJ	3.5	11.9
NRC	6.1	10.0
NIST	-2.7	17.8
METAS	-3.7	20.2
CSIR-NML	-3.7	26.2
PTB	-1.3	10.4
NMIA	1.3	12.9
NPL	-2.7	11.4
MIRS/SIQ	5.3	16.2
INRIM	-4.4	27.7
VNIIFTRI	-20.7	14.3
SPRING	-7.1	19.8

Lab <i>j</i>	FREQUENCY : 18 GHz																	
NMIJ	NRC		NIST		METAS		CSIR-NML		PTB		NMIA		NPL		MIRS/SIQ			
$D_{ij}$ / $10^{-3}$	$U_{ij}$ / $10^{-3}$																	
3.5	-2.6	15.1	6.2	21.1	7.2	23.1	7.2	28.5	4.8	15.3	2.2	17.1	6.2	16.0	-1.8	19.8		
6.1	2.6	15.1	8.8	20.0	9.8	22.2	9.8	27.7	7.4	13.9	4.8	15.8	8.8	14.6	0.8	18.7		
-2.7	-6.2	21.1	-8.8	20.0	1.0	26.6	1.0	31.4	-1.4	20.2	-4.0	21.6	0.0	20.8	-8.0	23.8		
-3.7	-7.2	23.1	-9.8	22.2	-1.0	26.6	0.0	32.8	-2.4	22.4	-5.0	23.6	-1.0	22.8	-9.0	25.6		
-3.7	-7.2	28.5	-9.8	27.7	-1.0	31.4	0.0	32.8	-2.4	27.9	-5.0	28.9	-1.0	28.2	-9.0	30.5		
-1.3	-4.8	15.3	-7.4	13.9	1.4	20.2	2.4	22.4	2.4	27.9	-2.6	16.1	1.4	14.9	-6.6	18.9		
1.3	-2.2	17.1	-4.8	15.8	4.0	21.6	5.0	23.6	5.0	28.9	2.6	16.1	4.0	16.7	-4.0	20.4		
12.9	-6.2	16.0	-8.8	14.6	0.0	20.8	1.0	22.8	1.0	28.2	-1.4	14.9	-4.0	16.7	-8.0	19.4		
11.4	1.8	19.8	-0.8	18.7	8.0	23.8	9.0	25.6	9.0	30.5	6.6	18.9	4.0	20.4	8.0	19.4		
-2.7	-7.9	29.9	-10.5	29.2	-1.7	32.7	-0.7	34.1	-0.7	37.9	-3.1	29.4	-5.7	30.3	-1.7	29.7		
14.3	-24.2	18.2	-26.8	17.0	-18.0	22.5	-17.0	24.4	-17.0	29.5	-19.4	17.2	-22.0	18.8	-18.0	17.8		
19.8	-10.6	22.8	-13.2	21.8	-4.4	26.3	-3.4	28.0	-3.4	32.6	-5.8	22.0	-8.4	23.3	-4.4	22.5		

EIM	-3.5	28.0
SP	-7.5	13.2
NPL	-4.9	10.6
AREPA	4.5	11.0
PTB	0.9	10.0

Not computed																		

Laboratory having used the repaired transfer standard PTB 1-3-1

MEASURAND : Calibration factor

TRANSFER STANDARDS : Power sensors PTB 1-3, PTB 1-3-1 for CCEM.RF-K10.CL and PTB 1-03-2 for EUROMET.EM.RF-K10.CL

Matrix of equivalence

FREQUENCY : 18 GHz (continued)

Lab *i*



	$D_i$ / $10^{-3}$	$U_i$ / $10^{-3}$
NMIJ	3.5	11.9
NRC	6.1	10.0
NIST	-2.7	17.8
METAS	-3.7	20.2
CSIR-NML	-3.7	26.2
PTB	-1.3	10.4
NMIA	1.3	12.9
NPL	-2.7	11.4
MIRS/SIQ	5.3	16.2
INRIM	-4.4	27.7
VNIIFTRI	-20.7	14.3
SPRING	-7.1	19.8

Lab *j* →

	INRIM		VNIIIFTRI		SPRING	
	$D_{ij}$ / $10^{-3}$	$U_{ij}$ / $10^{-3}$	$D_{ij}$ / $10^{-3}$	$U_{ij}$ / $10^{-3}$	$D_{ij}$ / $10^{-3}$	$U_{ij}$ / $10^{-3}$
NMIJ	7.9	29.9	24.2	18.2	10.6	22.8
NRC	10.5	29.2	26.8	17.0	13.2	21.8
NIST	1.7	32.7	18.0	22.5	4.4	26.3
METAS	0.7	34.1	17.0	24.4	3.4	28.0
CSIR-NML	0.7	37.9	17.0	29.5	3.4	32.6
PTB	3.1	29.4	19.4	17.2	5.8	22.0
NMIA	5.7	30.3	22.0	18.8	8.4	23.3
NPL	1.7	29.7	18.0	17.8	4.4	22.5
MIRS/SIQ	9.7	31.9	26.0	21.3	12.4	25.3
INRIM			16.3	30.9	2.7	33.9
VNIIFTRI	-16.3	30.9			-13.6	24.1
SPRING	-2.7	33.9	13.6	24.1		

EIM	-3.5	28.0
SP	-7.5	13.2
NPL	-4.9	10.6
AREPA	4.5	11.0
PTB	0.9	10.0


Not computed

	EIM		SP		NPL		AREPA		PTB	
	$D_{ij}$ / $10^{-3}$	$U_{ij}$ / $10^{-3}$								
EIM										
SP										
NPL										
AREPA										
PTB										

Not computed


	<b>4.0</b>	31.0	<b>1.4</b>	29.9	<b>-8.0</b>	30.1	<b>-4.4</b>	29.7
-4.0	31.0		-2.6	16.9	-12.0	17.2	-8.4	16.6
-1.4	29.9	<b>2.6</b>	16.9		-9.4	15.3	-5.8	14.6
8.0	30.1	<b>12.0</b>	17.2	<b>9.4</b>	15.3		3.6	14.9
4.4	29.7	<b>8.4</b>	16.6	<b>5.8</b>	14.6	-3.6	14.9	

Laboratory having used the repaired transfer standard PTB 1-3-1

CCEM.RF-K10.CL and EUROMET.EM.RF-K10.CL Calibration factor, frequency: 18 GHz,

Transfer standard: PTB 1-3

Degrees of equivalence [ $D_i$  and its expanded uncertainty ( $k = 2$ ),  $U_i$ ]



Open symbols represent values from laboratories in Associate States and Economies of the CGPM

Blue triangles: participants in EUROMET.EM.RF-K10.CL

CCEM.RF-K10.CL and EUROMET.EM.RF-K10.CL

#### **MEASURAND : Calibration factor**

**TRANSFER STANDARDS :** Power sensors PTB 1-3, PTB 1-3-1 for CCEM.RF-K10.CL and PTB 1-03-2 for EUROMET.EM-RF-K10.CL

## Matrix of equivalence

FREQUENCY : 20 GHz

Lab <i>i</i>	Eqs. 1-10																			
	NMIJ		NRC		NIST		METAS		CSIR-NML		PTB		NMIA		NPL		MIRS/SIQ			
	$D_{ij}$ / 10 <sup>-3</sup>	$U_{ij}$ / 10 <sup>-3</sup>																		
NMIJ	5.2	8.8			2.7	12.4	6.9	19.6	3.8	23.3	0.8	31.0	6.5	13.5	0.8	20.2	1.8	13.5	2.8	17.8
NRC	2.5	10.4	-2.7	12.4			4.2	20.4	1.1	24.0	-1.9	31.5	3.8	14.6	-1.9	20.9	-0.9	14.6	0.1	18.7
NIST	-1.7	18.4	-6.9	19.6	-4.2	20.4			-3.1	28.4	-6.1	35.0	-0.4	21.1	-6.1	25.9	-5.1	21.1	-4.1	24.1
METAS	1.4	22.4	-3.8	23.3	-1.1	24.0	3.1	28.4			-3.0	37.2	2.7	24.6	-3.0	28.8	-2.0	24.6	-1.0	27.2
CSIR-NML	4.4	30.3	-0.8	31.0	1.9	31.5	6.1	35.0	3.0	37.2			5.7	32.0	0.0	35.3	1.0	32.0	2.0	34.0
PTB	-1.3	11.7	-6.5	13.5	-3.8	14.6	0.4	21.1	-2.7	24.6	-5.7	32.0			-5.7	21.6	-4.7	15.6	-3.7	19.4
NMIA	4.4	19.0	-0.8	20.2	1.9	20.9	6.1	25.9	3.0	28.8	0.0	35.3	5.7	21.6			1.0	21.6	2.0	24.5
NPL	3.4	11.7	-1.8	13.5	0.9	14.6	5.1	21.1	2.0	24.6	-1.0	32.0	4.7	15.6	-1.0	21.6			1.0	19.4
MIRS/SIQ	2.4	16.5	-2.8	17.8	-0.1	18.7	4.1	24.1	1.0	27.2	-2.0	34.0	3.7	19.4	-2.0	24.5	-1.0	19.4		
INRIM	-8.4	26.5	-13.6	27.3	-10.9	27.9	-6.7	31.8	-9.8	34.2	-12.8	39.8	-7.1	28.4	-12.8	32.1	-11.8	28.4	-10.8	30.7
VNIIFTRI	-19.6	22.4	-24.8	23.3	-22.1	24.0	-17.9	28.4	-21.0	31.1	-24.0	37.2	-18.3	24.6	-24.0	28.8	-23.0	24.6	-22.0	27.2
SPRING	-0.9	20.2	-6.1	21.3	-3.4	22.0	0.8	26.8	-2.3	29.6	-5.3	35.9	0.4	22.7	-5.3	27.2	-4.3	22.7	-3.3	25.5

<b>EIM</b>	<b>-34.8</b>	26.0
<b>SP</b>	<b>1.2</b>	13.2
<b>NPL</b>	<b>-2.3</b>	10.6
<b>AREPA</b>	<b>13.2</b>	28.0
<b>PTB</b>	<b>1.0</b>	11.0

Laboratory having used the repaired transfer standard PTB 1-3-1

#### **MEASURAND : Calibration factor**

**TRANSFER STANDARDS :** Power sensors PTB 1-3, PTB 1-3-1 for CCEM.RF-K10.CL and PTB 1-03-2 for EUROMET.EM.RF-K10.CL

## Matrix of equivalence

## **FREQUENCY : 20 GHz (continued)**

<b>Lab <i>i</i></b>	<b><math>D_i</math> <math>/ 10^{-3}</math></b>	<b><math>U_i</math> <math>/ 10^{-3}</math></b>
<b>NMIJ</b>	<b>5.2</b>	8.8
<b>NRC</b>	<b>2.5</b>	10.4
<b>NIST</b>	<b>-1.7</b>	18.4
<b>METAS</b>	<b>1.4</b>	22.4
<b>CSIR-NML</b>	<b>4.4</b>	30.3
<b>PTB</b>	<b>-1.3</b>	11.7
<b>NMIA</b>	<b>4.4</b>	19.0
<b>NPL</b>	<b>3.4</b>	11.7
<b>MIRS/SIQ</b>	<b>2.4</b>	16.5
<b>INRIM</b>	<b>-8.4</b>	26.5
<b>VNIIFTRI</b>	<b>-19.6</b>	22.4
<b>SPRING</b>	<b>-0.9</b>	20.2

INRIM		VNIIFTRI		SPRING	
$D_{ij}$ / $10^{-3}$	$U_{ij}$ / $10^{-3}$	$D_{ij}$ / $10^{-3}$	$U_{ij}$ / $10^{-3}$	$D_{ij}$ / $10^{-3}$	$U_{ij}$ / $10^{-3}$
13.6	27.3	<b>24.8</b>	23.3	<b>6.1</b>	21.3
<b>10.9</b>	27.9	<b>22.1</b>	24.0	<b>3.4</b>	22.0
<b>6.7</b>	31.8	<b>17.9</b>	28.4	<b>-0.8</b>	26.8
<b>9.8</b>	34.2	<b>21.0</b>	31.1	<b>2.3</b>	29.6
<b>12.8</b>	39.8	<b>24.0</b>	37.2	<b>5.3</b>	35.9
<b>7.1</b>	28.4	<b>18.3</b>	24.6	<b>-0.4</b>	22.7
<b>12.8</b>	32.1	<b>24.0</b>	28.8	<b>5.3</b>	27.2
<b>11.8</b>	28.4	<b>23.0</b>	24.6	<b>4.3</b>	22.7
<b>10.8</b>	30.7	<b>22.0</b>	27.2	<b>3.3</b>	25.5
		<b>11.2</b>	34.2	<b>-7.5</b>	32.8
<b>-11.2</b>	34.2			<b>-18.7</b>	29.6
<b>7.5</b>	32.8	<b>18.7</b>	29.6		

<b>EIM</b>	<b>-34.8</b>	26.0
<b>SP</b>	<b>1.2</b>	13.2
<b>NPL</b>	<b>-2.3</b>	10.6
<b>AREPA</b>	<b>13.2</b>	28.0
<b>PTB</b>	<b>1.0</b>	11.0


	<b>-36.0</b>	29.2	<b>-32.5</b>	28.1	<b>-48.0</b>	38.2	<b>-35.8</b>	28.2
<b>36.0</b>	29.2		<b>3.5</b>	16.9	<b>-12.0</b>	31.0	<b>0.2</b>	17.2
<b>32.5</b>	28.1	<b>-3.5</b>	16.9		<b>-15.5</b>	29.9	<b>-3.3</b>	15.3
<b>48.0</b>	38.2	<b>12.0</b>	31.0	<b>15.5</b>	29.9		<b>12.2</b>	30.1
<b>35.8</b>	28.2	<b>-0.2</b>	17.2	<b>3.3</b>	15.3	<b>-12.2</b>	30.1	

Laboratory having used the repaired transfer standard PTB 1-3-1

CCEM.RF-K10.CL and EUROMET.EM.RF-K10.CL Calibration factor, frequency: 20 GHz,

Transfer standard: PTB 1-3

Degrees of equivalence [ $D_i$  and its expanded uncertainty ( $k = 2$ ),  $U_i$ ]



Open symbols represent values from laboratories in Associate States and Economies of the CGPM

Blue triangles: participants in EUROMET.EM.RF-K10.CL

CCEM.RF-K10.CL and EUROMET.EM.RF-K10.CL

#### **MEASURAND : Calibration factor**

**TRANSFER STANDARDS :** Power sensors PTB 1-3, PTB 1-3-1 for CCEM.RF-K10.CL and PTB 1-03-2 for EUROMET.EM-RF-K10.CL

## Matrix of equivalence

FREQUENCY : 23 GHz

Lab <i>i</i>			NMIJ		NRC		NIST		METAS		CSIR-NML		PTB		NMIA		NPL		MIRS/SIQ	
	$D_i$ $/ 10^{-3}$	$U_i$ $/ 10^{-3}$	$D_{ij}$ $/ 10^{-3}$	$U_{ij}$ $/ 10^{-3}$																
NMIJ	1.0	9.6			-2.2	13.0	3.7	20.4	-1.8	23.7	-4.8	33.2	1.1	15.7	-0.8	20.4	-4.8	15.9	0.2	18.3
NRC	3.2	10.3	2.2	13.0			5.9	20.8	0.4	24.0	-2.6	33.4	3.3	16.2	1.4	20.8	-2.6	16.3	2.4	18.7
NIST	-2.7	18.8	-3.7	20.4	-5.9	20.8			-5.5	28.7	-8.5	36.9	-2.6	22.5	-4.5	26.0	-8.5	22.6	-3.5	24.4
METAS	2.8	22.3	1.8	23.7	-0.4	24.0	5.5	28.7			-3.0	38.8	2.9	25.6	1.0	28.7	-3.0	25.7	2.0	27.2
CSIR-NML	5.8	32.2	4.8	33.2	2.6	33.4	8.5	36.9	3.0	38.8			5.9	34.5	4.0	36.9	0.0	34.6	5.0	35.8
PTB	-0.1	13.5	-1.1	15.7	-3.3	16.2	2.6	22.5	-2.9	25.6	-5.9	34.5			-1.9	22.5	-5.9	18.5	-0.9	20.6
NMIA	1.8	18.8	0.8	20.4	-1.4	20.8	4.5	26.0	-1.0	28.7	-4.0	36.9	1.9	22.5			-4.0	22.6	1.0	24.4
NPL	5.8	13.7	4.8	15.9	2.6	16.3	8.5	22.6	3.0	25.7	0.0	34.6	5.9	18.5	4.0	22.6			5.0	20.7
MIRS/SIQ	0.8	16.4	-0.2	18.3	-2.4	18.7	3.5	24.4	-2.0	27.2	-5.0	35.8	0.9	20.6	-1.0	24.4	-5.0	20.7		
INRIM	-13.7	38.2	-14.7	39.0	-16.9	39.2	-11.0	42.2	-16.5	43.9	-19.5	49.7	-13.6	40.2	-15.5	42.2	-19.5	40.2	-14.5	41.2
VNIIFTRI	-7.2	24.3	-8.2	25.6	-10.4	25.8	-4.5	30.2	-10.0	32.6	-13.0	40.0	-7.1	27.3	-9.0	30.2	-13.0	27.4	-8.0	28.8
SPRING	0.8	26.3	-0.2	27.4	-2.4	27.7	3.5	31.9	-2.0	34.1	-5.0	41.2	0.9	29.1	-1.0	31.9	-5.0	29.2	0.0	30.5

EIM	-7.5	28.0
SP	-1.5	16.0
NPL	0.2	12.8
AREPA	12.5	34.0
PTB	5.5	13.0

**Not computed**

Laboratory having used the repaired transfer standard PTB 1-3-1

#### **MEASURAND : Calibration factor**

**TRANSFER STANDARDS :** Power sensors PTB 1-3, PTB 1-3-1 for CCEM.RF-K10.CL and PTB 1-03-2 for EUROMET.EM.RF-K10.CL

## Matrix of equivalence

## **FREQUENCY : 23 GHz (continued)**

## Lab *j* →

<b>Lab <i>i</i></b>	<b><math>D_i</math> <math>/ 10^{-3}</math></b>	<b><math>U_i</math> <math>/ 10^{-3}</math></b>
<b>NMIJ</b>	<b>1.0</b>	9.6
<b>NRC</b>	<b>3.2</b>	10.3
<b>NIST</b>	<b>-2.7</b>	18.8
<b>METAS</b>	<b>2.8</b>	22.3
<b>CSIR-NML</b>	<b>5.8</b>	32.2
<b>PTB</b>	<b>-0.1</b>	13.5
<b>NMIA</b>	<b>1.8</b>	18.8
<b>NPL</b>	<b>5.8</b>	13.7
<b>MIRS/SIQ</b>	<b>0.8</b>	16.4
<b>INRIM</b>	<b>-13.7</b>	38.2
<b>VNIIFTRI</b>	<b>-7.2</b>	24.3
<b>SPRING</b>	<b>0.8</b>	26.3

<b>EIM</b>	<b>-7.5</b>	28.0
<b>SP</b>	<b>-1.5</b>	16.0
<b>NPL</b>	<b>0.2</b>	12.8
<b>AREPA</b>	<b>12.5</b>	34.0
<b>PTB</b>	<b>5.5</b>	13.0

INRIM		VNIIFTRI		SPRING	
$D_{ij}$ / $10^{-3}$	$U_{ij}$ / $10^{-3}$	$D_{ij}$ / $10^{-3}$	$U_{ij}$ / $10^{-3}$	$D_{ij}$ / $10^{-3}$	$U_{ij}$ / $10^{-3}$
14.7	39.0	8.2	25.6	0.2	27.4
16.9	39.2	10.4	25.8	2.4	27.7
11.0	42.2	4.5	30.2	-3.5	31.9
16.5	43.9	10.0	32.6	2.0	34.1
19.5	49.7	13.0	40.0	5.0	41.2
13.6	40.2	7.1	27.3	-0.9	29.1
15.5	42.2	9.0	30.2	1.0	31.9
19.5	40.2	13.0	27.4	5.0	29.2
14.5	41.2	8.0	28.8	0.0	30.5
		-6.5	44.9	-14.5	46.0
6.5	44.9			-8.0	35.4
14.5	46.0	8.0	35.4		

Not computed


Not computed

	<b>-6.0</b>	32.2	<b>-7.7</b>	30.8	<b>-20.0</b>	44.0	<b>-13.0</b>	30.9
<b>6.0</b>	32.2		<b>-1.7</b>	20.5	<b>-14.0</b>	37.6	<b>-7.0</b>	20.6
<b>7.7</b>	30.8	<b>1.7</b>	20.5		<b>-12.3</b>	36.3	<b>-5.3</b>	18.2
<b>20.0</b>	44.0	<b>14.0</b>	37.6	<b>12.3</b>	36.3		<b>7.0</b>	36.4
<b>13.0</b>	30.9	<b>7.0</b>	20.6	<b>5.3</b>	18.2	<b>-7.0</b>	36.4	

Laboratory having used the repaired transfer standard PTB 1-3-1

**CCEM.RF-K10.CL and EUROMET.EM.RF-K10.CL Calibration factor, frequency: 23 GHz, Transfer standard: PTB 1-3**

**Degrees of equivalence [  $D_i$  and its expanded uncertainty ( $k = 2$ ),  $U_i$  ]**



Open symbols represent values from laboratories in Associate States and Economies of the CGPM

**Blue triangles:** participants in EUROMET.EM.RF-K10.CL

CCEM.RF-K10.CL and EUROMET.EM.RF-K10.CL

MEASURAND : Calibration factor

TRANSFER STANDARDS : Power sensors PTB 1-3, PTB 1-3-1 for CCEM.RF-K10.CL and PTB 1-03-2 for EUROMET.EM.RF-K10.CL

Matrix of equivalence

FREQUENCY : 26 GHz

Lab <i>i</i>	$D_i$ / $10^{-3}$	$U_i$ / $10^{-3}$
NMIJ	-20.8	14.5
NRC	3.0	10.6
NIST	-8.3	19.3
METAS	-13.8	24.4
CSIR-NML	2.2	38.3
PTB	-0.1	17.6
NMIA	-2.8	18.7
NPL	4.2	16.0
MIRS/SIQ	-3.8	16.6
INRIM	-22.5	37.7
VNIIFTRI	1.2	28.3
SPRING	-2.4	17.4

Lab <i>j</i>	FREQUENCY : 26 GHz																	
NMIJ	NRC		NIST		METAS		CSIR-NML		PTB		NMIA		NPL		MIRS/SIQ			
$D_{ij}$ / $10^{-3}$	$U_{ij}$ / $10^{-3}$																	
NMIJ	-23.8	16.8	-12.5	23.3	-7.0	27.7	-23.0	40.4	-20.7	21.9	-18.0	22.8	-25.0	20.7	-17.0	21.1		
NRC	23.8	16.8	11.3	21.1	16.8	25.8	0.8	39.2	3.1	19.5	5.8	20.6	-1.2	18.1	6.8	18.7		
NIST	12.5	23.3	-11.3	21.1	5.5	30.5	-10.5	42.4	-8.2	25.3	-5.5	26.2	-12.5	24.3	-4.5	24.7		
METAS	7.0	27.7	-16.8	25.8	-5.5	30.5	-16.0	44.9	-13.7	29.4	-11.0	30.1	-18.0	28.5	-10.0	28.8		
CSIR-NML	23.0	40.4	-0.8	39.2	10.5	42.4	16.0	44.9	2.3	41.6	5.0	42.1	-2.0	41.0	6.0	41.2		
PTB	20.7	21.9	-3.1	19.5	8.2	25.3	13.7	29.4	-2.3	41.6	2.7	24.9	-4.3	22.9	3.7	23.3		
NMIA	18.0	22.8	-5.8	20.6	5.5	26.2	11.0	30.1	-5.0	42.1	-2.7	24.9	-7.0	23.8	1.0	24.2		
NPL	25.0	20.7	1.2	18.1	12.5	24.3	18.0	28.5	2.0	41.0	4.3	22.9	7.0	23.8	8.0	22.2		
MIRS/SIQ	17.0	21.1	-6.8	18.7	4.5	24.7	10.0	28.8	-6.0	41.2	-3.7	23.3	-1.0	24.2	-8.0	22.2		
INRIM	-1.7	39.9	-25.5	38.6	-14.2	41.9	-8.7	44.4	-24.7	53.3	-22.4	41.1	-19.7	41.6	-26.7	40.4	-18.7	40.7
VNIIFTRI	22.0	31.2	-1.8	29.6	9.5	33.7	15.0	36.9	-1.0	47.2	1.3	32.8	4.0	33.4	-3.0	32.0	5.0	32.2
SPRING	18.4	21.7	-5.4	19.3	5.9	25.2	11.4	29.3	-4.6	41.5	-2.3	23.9	0.4	24.8	-6.6	22.8	1.4	23.2

EIM	-23.1	30.0
SP	4.9	20.8
NPL	1.5	15.0
AREPA	6.9	34.0
PTB	-0.5	17.0

Not computed															

Laboratory having used the repaired transfer standard PTB 1-3-1

#### **MEASURAND : Calibration factor**

**TRANSFER STANDARDS :** Power sensors PTB 1-3, PTB 1-3-1 for CCEM.RF-K10.CL and PTB 1-03-2 for EUROMET.EM.RF-K10.CL

## Matrix of equivalence

## **FREQUENCY : 26 GHz (continued)**

<b>Lab <i>i</i></b>	<b><math>D_i</math> <math>/ 10^{-3}</math></b>	<b><math>U_i</math> <math>/ 10^{-3}</math></b>
<b>NMIJ</b>	<b>-20.8</b>	14.5
<b>NRC</b>	<b>3.0</b>	10.6
<b>NIST</b>	<b>-8.3</b>	19.3
<b>METAS</b>	<b>-13.8</b>	24.4
<b>CSIR-NML</b>	<b>2.2</b>	38.3
<b>PTB</b>	<b>-0.1</b>	17.6
<b>NMIA</b>	<b>-2.8</b>	18.7
<b>NPL</b>	<b>4.2</b>	16.0
<b>MIRS/SIQ</b>	<b>-3.8</b>	16.6
<b>INRIM</b>	<b>-22.5</b>	37.7
<b>VNIIFTRI</b>	<b>1.2</b>	28.3
<b>SPRING</b>	<b>-2.4</b>	17.4

INRIM		VNIIIFTRI		SPRING	
$D_{ij}$ / $10^{-3}$	$U_{ij}$ / $10^{-3}$	$D_{ij}$ / $10^{-3}$	$U_{ij}$ / $10^{-3}$	$D_{ij}$ / $10^{-3}$	$U_{ij}$ / $10^{-3}$
1.7	39.9	-22.0	31.2	-18.4	21.7
25.5	38.6	1.8	29.6	5.4	19.3
14.2	41.9	-9.5	33.7	-5.9	25.2
8.7	44.4	-15.0	36.9	-11.4	29.3
24.7	53.3	1.0	47.2	4.6	41.5
22.4	41.1	-1.3	32.8	2.3	23.9
19.7	41.6	-4.0	33.4	-0.4	24.8
26.7	40.4	3.0	32.0	6.6	22.8
18.7	40.7	-5.0	32.2	-1.4	23.2
		-23.7	46.7	-20.1	41.0
23.7	46.7			3.6	32.7
20.1	41.0	-3.6	32.7		


<b>EIM</b>	<b>-23.1</b>	30.0
<b>SP</b>	<b>4.9</b>	20.8
<b>NPL</b>	<b>1.5</b>	15.0
<b>AREPA</b>	<b>6.9</b>	34.0
<b>PTB</b>	<b>-0.5</b>	17.0

	<b>-28.0</b>	36.5	<b>-24.6</b>	33.5	<b>-30.0</b>	45.3	<b>-22.6</b>	34.5
<b>28.0</b>	36.5		<b>3.4</b>	25.6	<b>-2.0</b>	39.9	<b>5.4</b>	26.9
<b>24.6</b>	33.5	<b>-3.4</b>	25.6		<b>-5.4</b>	37.2	<b>2.0</b>	22.7
<b>30.0</b>	45.3	<b>2.0</b>	39.9	<b>5.4</b>	37.2		<b>7.4</b>	38.0
<b>22.6</b>	34.5	<b>-5.4</b>	26.9	<b>-2.0</b>	22.7	<b>-7.4</b>	38.0	

Laboratory having used the repaired transfer standard PTB 1-3-1

**CCEM.RF-K10.CL and EUROMET.EM.RF-K10.CL Calibration factor, frequency: 26 GHz, Transfer standard: PTB 1-3**

**Degrees of equivalence [  $D_i$  and its expanded uncertainty ( $k = 2$ ),  $U_i$  ]**



Open symbols represent values from laboratories in Associate States and Economies of the CGPM

**Blue triangles:** participants in EUROMET.EM.RF-K10.CL

MEASURAND : Calibration factor

TRANSFER STANDARDS : Power sensors PTB 2-6, PTB 2-6-1 for CCEM.RF-K10.CL and PTB 2-06-2 for EUROMET.EM.RF-K10.CL

Matrix of equivalence

FREQUENCY : 50 MHz

Lab <i>i</i>	Lab <i>j</i> →																				
	NMJ		NIST		METAS		CSIR-NML		PTB		NMIA		NPL		MIRS/SIQ		INRIM		SPRING		
	$D_{ij}$ / 10 <sup>-3</sup>	$U_{ij}$ / 10 <sup>-3</sup>		$D_{ij}$ / 10 <sup>-3</sup>	$U_{ij}$ / 10 <sup>-3</sup>		$D_{ij}$ / 10 <sup>-3</sup>	$U_{ij}$ / 10 <sup>-3</sup>		$D_{ij}$ / 10 <sup>-3</sup>	$U_{ij}$ / 10 <sup>-3</sup>		$D_{ij}$ / 10 <sup>-3</sup>	$U_{ij}$ / 10 <sup>-3</sup>		$D_{ij}$ / 10 <sup>-3</sup>	$U_{ij}$ / 10 <sup>-3</sup>		$D_{ij}$ / 10 <sup>-3</sup>	$U_{ij}$ / 10 <sup>-3</sup>	
NMIJ	-5.0	4.7		-6.5	17.4	-0.5	12.3	-8.5	14.3	-4.4	4.9	-5.5	4.9	-2.7	4.3	-6.5	6.6	-10.6	11.9	-6.1	13.7
NIST	1.5	17.6	6.5	17.4		6.0	21.0	-2.0	22.2	2.1	17.7	1.0	17.7	3.8	17.5	0.0	18.2	-4.1	20.7	0.4	21.8
METAS	-4.5	12.6	0.5	12.3	-6.0	21.0		-8.0	18.4	-3.9	12.6	-5.0	12.6	-2.2	12.4	-6.0	13.4	-10.1	16.7	-5.6	18.0
CSIR-NML	3.5	14.5	8.5	14.3	2.0	22.2	8.0	18.4		4.1	14.6	3.0	14.6	5.8	14.4	2.0	15.2	-2.1	18.2	2.4	19.4
PTB	-0.6	5.5	4.4	4.9	-2.1	17.7	3.9	12.6	-4.1	14.6		-1.1	5.7	1.7	5.1	-2.1	7.2	-6.2	12.3	-1.7	14.0
NMIA	0.5	5.5	5.5	4.9	-1.0	17.7	5.0	12.6	-3.0	14.6	1.1	5.7		2.8	5.1	-1.0	7.2	-5.1	12.3	-0.6	14.0
NPL	-2.3	5.1	2.7	4.3	-3.8	17.5	2.2	12.4	-5.8	14.4	-1.7	5.1	-2.8	5.1		-3.8	6.8	-7.9	12.0	-3.4	13.8
MIRS/SIQ	1.5	7.1	6.5	6.6	0.0	18.2	6.0	13.4	-2.0	15.2	2.1	7.2	1.0	7.2	3.8	6.8		-4.1	13.1	0.4	14.7
INRIM	5.6	12.2	10.6	11.9	4.1	20.7	10.1	16.7	2.1	18.2	6.2	12.3	5.1	12.3	7.9	12.0	4.1	13.1		4.5	17.7
SPRING	1.1	13.9	6.1	13.7	-0.4	21.8	5.6	18.0	-2.4	19.4	1.7	14.0	0.6	14.0	3.4	13.8	-0.4	14.7	-4.5	17.7	
SP	-1.9	5.8																			
NPL	-0.9	2.4																			
AREPA	5.1	8.0																			
NMi-VSL	-8.2	7.6																			
LNE	-3.7	4.6																			
UME	3.3	5.6																			
PTB	-0.5	4.0																			

Laboratory having used the repaired transfer standard PTB 2-6-1

## **MEASURAND : Calibration factor**

**TRANSFER STANDARDS :** Power sensors PTB 2-6, PTB 2-6-1 for CCEM.RF-K10.CL and PTB 2-06-2 for EUROMET.EM.RF-K10.CL

## Matrix of equivalence

FREQUENCY : 50 MHZ

**(Continued)**


Lab <i>i</i>	$D_i / 10^{-3}$	$U_i / 10^{-3}$
NMIJ	-5.0	4.7
NIST	1.5	17.6
METAS	-4.5	12.6
CSIR-NML	3.5	14.5
PTB	-0.6	5.5
NMIA	0.5	5.5
NPL	-2.3	5.1
MIRS/SIQ	1.5	7.1
INRIM	5.6	12.2
SPRING	1.1	13.9

<b>SP</b>	<b>-1.9</b>	5.8
<b>NPL</b>	<b>-0.9</b>	2.4
<b>AREPA</b>	<b>5.1</b>	8.0
<b>NMi-VSL</b>	<b>-8.2</b>	7.6
<b>LNE</b>	<b>-3.7</b>	4.6
<b>UME</b>	<b>3.3</b>	5.6
<b>PTB</b>	<b>-0.5</b>	4.0

		<b>-1.0</b>	6.3	<b>-7.0</b>	9.9	<b>6.3</b>	9.6	<b>1.8</b>	7.4	<b>-5.2</b>	8.1	<b>-1.4</b>	7.0
<b>1.0</b>	6.3			<b>-6.0</b>	8.4	<b>7.3</b>	8.0	<b>2.8</b>	5.2	<b>-4.2</b>	6.1	<b>-0.4</b>	4.7
<b>7.0</b>	9.9	<b>6.0</b>	8.4			<b>13.3</b>	11.0	<b>8.8</b>	9.2	<b>1.8</b>	9.8	<b>5.6</b>	8.9
<b>-6.3</b>	9.6	<b>-7.3</b>	8.0	<b>-13.3</b>	11.0			<b>-4.5</b>	8.9	<b>-11.5</b>	9.4	<b>-7.7</b>	8.6
<b>-1.8</b>	7.4	<b>-2.8</b>	5.2	<b>-8.8</b>	9.2	<b>4.5</b>	8.9			<b>-7.0</b>	7.2	<b>-3.2</b>	6.1
<b>5.2</b>	8.1	<b>4.2</b>	6.1	<b>-1.8</b>	9.8	<b>11.5</b>	9.4	<b>7.0</b>	7.2			<b>3.8</b>	6.9
<b>1.4</b>	7.0	<b>0.4</b>	4.7	<b>-5.6</b>	8.9	<b>7.7</b>	8.6	<b>3.2</b>	6.1	<b>-3.8</b>	6.9		

Laboratory having used the repaired transfer standard PTB 2-6-1

**CCEM.RF-K10.CL and EUROMET.EM.RF-K10.CL Calibration factor, frequency: 50 MHz,  
Transfer standard: PTB 2-6**  
**Degrees of equivalence [  $D_i$  and its expanded uncertainty ( $k = 2$ ),  $U_i$  ]**



Open symbols represent values from laboratories in Associate States and Economies of the CGPM  
 Blue triangles: participants in EUROMET.EM.RF-K10.CL

#### **MEASURAND : Calibration factor**

**TRANSFER STANDARDS :** Power sensors PTB 2-6, PTB 2-6-1 for CCEM.RF-K10.CL and PTB 2-06-2 for EUROMET.EM.RF-K10.CL

## **Matrix of equivalence**

FREQUENCY : 1 GHz

The diagram illustrates the transfer of data between two laboratories, *i* and *j*. On the left, a box labeled "Lab *i*" contains a table with data  $D_i$  and  $U_i$ , both scaled by  $10^{-3}$ . An arrow points from this box to a second box on the right labeled "Lab *j*". This second box is labeled "NMIJ" and contains a table with data  $D_{ij}$  and  $U_{ij}$ , also scaled by  $10^{-3}$ .

	$D_i$ $/ 10^{-3}$	$U_i$ $/ 10^{-3}$
NMIJ	-4.0	4.5
NIST	-2.0	17.7
METAS	-3.4	14.3
CSIR-NML	1.6	16.3
PTB	-2.0	5.8
NMIA	0.6	5.8
NPL	-2.4	7.1
MIRS/SIQ	-1.4	6.7
INRIM	5.3	12.0
VNIIFTRI	4.6	12.4
SPRING	-2.5	13.7

<b>SP</b>	<b>-1.1</b>	6.4
<b>NPL</b>	<b>-2.3</b>	6.0
<b>AREPA</b>	<b>5.0</b>	8.0
<b>NMi-VSL</b>	<b>-9.2</b>	10.2
<b>LNE</b>	<b>-4.0</b>	12.4
<b>UME</b>	<b>2.0</b>	5.4
<b>PTB</b>	<b>-1.9</b>	5.0

NMJJ		NIST		METAS		CSIR-NML		PTB		NMIA		NPL		MIRS/SIQ		INRIM		VNIIIFTRI	
$D_{ij}$	$U_{ij}$																		
/10 <sup>-3</sup>																			
		-2.0	17.7	-0.6	14.4	-5.6	16.4	-2.0	6.0	-4.6	6.0	-1.6	7.2	-2.6	6.9	-9.3	12.1	-8.6	12.5
2.0	17.7			1.4	22.3	-3.6	23.6	0.0	18.1	-2.6	18.1	0.4	18.5	-0.6	18.4	-7.3	20.9	-6.6	21.1
0.6	14.4	-1.4	22.3			-5.0	21.3	-1.4	14.9	-4.0	14.9	-1.0	15.4	-2.0	15.2	-8.7	18.2	-8.0	18.4
5.6	16.4	3.6	23.6	5.0	21.3			3.6	16.8	1.0	16.8	4.0	17.2	3.0	17.1	-3.7	19.8	-3.0	20.0
2.0	6.0	0.0	18.1	1.4	14.9	-3.6	16.8			-2.6	7.1	0.4	8.1	-0.6	7.8	-7.3	12.6	-6.6	13.0
4.6	6.0	2.6	18.1	4.0	14.9	-1.0	16.8	2.6	7.1			3.0	8.1	2.0	7.8	-4.7	12.6	-4.0	13.0
1.6	7.2	-0.4	18.5	1.0	15.4	-4.0	17.2	-0.4	8.1	-3.0	8.1			-1.0	8.8	-7.7	13.2	-7.0	13.6
2.6	6.9	0.6	18.4	2.0	15.2	-3.0	17.1	0.6	7.8	-2.0	7.8	1.0	8.8			-6.7	13.1	-6.0	13.4
9.3	12.1	7.3	20.9	8.7	18.2	3.7	19.8	7.3	12.6	4.7	12.6	7.7	13.2	6.7	13.1			0.7	16.7
8.6	12.5	6.6	21.1	8.0	18.4	3.0	20.0	6.6	13.0	4.0	13.0	7.0	13.6	6.0	13.4	-0.7	16.7		
1.5	13.8	-0.5	22.0	0.9	19.4	-4.1	20.9	-0.5	14.3	-3.1	14.3	-0.1	14.8	-1.1	14.7	-7.8	17.7	-7.1	18.0

### Not computed

Laboratory having used the repaired transfer standard PTB 2-6-1

MEASURAND : Calibration factor

TRANSFER STANDARDS : Power sensors PTB 2-6, PTB 2-6-1 for CCEM.RF-K10.CL and PTB 2-06-2 for EUROMET.EM.RF-K10.CL

Matrix of equivalence


FREQUENCY : 1 GHz

(Continued)

Lab <i>i</i>	Lab <i>j</i> →															
	SPRING		SP		NPL		AREPA		NMi-VSL		LNE		UME		PTB	
	$D_{ij}$ / $10^{-3}$	$U_{ij}$ / $10^{-3}$														
NMIJ	-4.0	4.5														
NIST	-2.0	17.7	0.5	22.0												
METAS	-3.4	14.3	-0.9	19.4												
CSIR-NML	1.6	16.3	4.1	20.9												
PTB	-2.0	5.8	0.5	14.3												
NMIA	0.6	5.8	3.1	14.3												
NPL	-2.4	7.1	0.1	14.8												
MIRS/SIQ	-1.4	6.7	1.1	14.7												
INRIM	5.3	12.0	7.8	17.7												
VNIIFTRI	4.6	12.4	7.1	18.0												
SPRING	-2.5	13.7														
SP	-1.1	6.4														
NPL	-2.3	6.0														
AREPA	5.0	8.0														
NMi-VSL	-9.2	10.2	Not computed													
LNE	-4.0	12.4														
UME	2.0	5.4														
PTB	-1.9	5.0														

Laboratory having used the repaired transfer standard PTB 2-6-1

**CCEM.RF-K10.CL and EUROMET.EM.RF-K10.CL Calibration factor, frequency: 1 GHz,  
Transfer standard: PTB 2-6**  
**Degrees of equivalence [  $D_i$  and its expanded uncertainty ( $k = 2$ ),  $U_i$  ]**



Open symbols represent values from laboratories in Associate States and Economies of the CGPM  
 Blue triangles: participants in EUROMET.EM.RF-K10.CL

#### **MEASURAND : Calibration factor**

**TRANSFER STANDARDS :** Power sensors PTB 2-6, PTB 2-6-1 for CCEM.RF-K10.CL and PTB 2-06-2 for EUROMET.EM.RF-K10.CL

## **Matrix of equivalence**

FREQUENCY : 10 GHz

Lab *j* 

	$D_i$ / 10 <sup>-3</sup>	$U_i$ / 10 <sup>-3</sup>
NMIJ	2.2	4.5
NIST	2.6	17.8
METAS	1.0	20.1
CSIR-NML	5.0	22.1
PTB	-0.4	8.4
NMIA	2.0	9.7
NPL	-2.0	9.3
MIRS/SIQ	-7.0	14.2
INRIM	-4.5	30.7
VNIIFTRI	21.0	20.1
SPRING	0.6	14.4

NMJJ		NIST		METAS		CSIR-NML		PTB		NMIA		NPL		MIRS/SIQ		INRIM		VNIIFTRI	
$D_{ij}$ / 10 <sup>-3</sup>	$U_{ij}$ / 10 <sup>-3</sup>																		
		<b>-0.4</b>	18.0	<b>1.2</b>	20.4	<b>-2.8</b>	22.3	<b>2.6</b>	8.9	<b>0.2</b>	10.1	<b>4.2</b>	9.8	<b>9.2</b>	14.5	<b>6.7</b>	30.8	<b>-18.8</b>	20.4
<b>0.4</b>	18.0			<b>1.6</b>	26.6	<b>-2.4</b>	28.2	<b>3.0</b>	19.3	<b>0.6</b>	20.0	<b>4.6</b>	19.8	<b>9.6</b>	22.5	<b>7.1</b>	35.3	<b>-18.4</b>	26.6
<b>-1.2</b>	20.4	<b>-1.6</b>	26.6			<b>-4.0</b>	29.7	<b>1.4</b>	21.5	<b>-1.0</b>	22.1	<b>3.0</b>	21.9	<b>8.0</b>	24.4	<b>5.5</b>	36.6	<b>-20.0</b>	28.3
<b>2.8</b>	22.3	<b>2.4</b>	28.2	<b>4.0</b>	29.7			<b>5.4</b>	23.4	<b>3.0</b>	23.9	<b>7.0</b>	23.8	<b>12.0</b>	26.1	<b>9.5</b>	37.7	<b>-16.0</b>	29.7
<b>-2.6</b>	8.9	<b>-3.0</b>	19.3	<b>-1.4</b>	21.5	<b>-5.4</b>	23.4			<b>-2.4</b>	12.3	<b>1.6</b>	12.0	<b>6.6</b>	16.1	<b>4.1</b>	31.6	<b>-21.4</b>	21.5
<b>-0.2</b>	10.1	<b>-0.6</b>	20.0	<b>1.0</b>	22.1	<b>-3.0</b>	23.9	<b>2.4</b>	12.3			<b>4.0</b>	13.0	<b>9.0</b>	16.9	<b>6.5</b>	32.0	<b>-19.0</b>	22.1
<b>-4.2</b>	9.8	<b>-4.6</b>	19.8	<b>-3.0</b>	21.9	<b>-7.0</b>	23.8	<b>-1.6</b>	12.0	<b>-4.0</b>	13.0			<b>5.0</b>	16.6	<b>2.5</b>	31.9	<b>-23.0</b>	21.9
<b>-9.2</b>	14.5	<b>-9.6</b>	22.5	<b>-8.0</b>	24.4	<b>-12.0</b>	26.1	<b>-6.6</b>	16.1	<b>-9.0</b>	16.9	<b>-5.0</b>	16.6			<b>-2.5</b>	33.7	<b>-28.0</b>	24.4
<b>-6.7</b>	30.8	<b>-7.1</b>	35.3	<b>-5.5</b>	36.6	<b>-9.5</b>	37.7	<b>-4.1</b>	31.6	<b>-6.5</b>	32.0	<b>-2.5</b>	31.9	<b>2.5</b>	33.7			<b>-25.5</b>	36.6
<b>18.8</b>	20.4	<b>18.4</b>	26.6	<b>20.0</b>	28.3	<b>16.0</b>	29.7	<b>21.4</b>	21.5	<b>19.0</b>	22.1	<b>23.0</b>	21.9	<b>28.0</b>	24.4	<b>25.5</b>	36.6		
<b>-1.6</b>	14.7	<b>-2.0</b>	22.6	<b>-0.4</b>	24.5	<b>-4.4</b>	26.2	<b>1.0</b>	16.3	<b>-1.4</b>	17.0	<b>2.6</b>	16.8	<b>7.6</b>	19.9	<b>5.1</b>	33.7	<b>-20.4</b>	24.5

<b>SP</b>	<b>-0.3</b>	9.4
<b>NPL</b>	<b>-3.2</b>	8.8
<b>AREPA</b>	<b>4.6</b>	14.0
<b>NMi-VSL</b>	<b>-6.0</b>	13.8
<b>LNE</b>	<b>12.1</b>	17.0
<b>UME</b>	<b>13.5</b>	7.8
<b>PTB</b>	<b>1.4</b>	16.0

Laboratory having used the repaired transfer standard PTB 2-6-1

MEASURAND : Calibration factor

TRANSFER STANDARDS : Power sensors PTB 2-6, PTB 2-6-1 for CCEM.RF-K10.CL and PTB 2-06-2 for EUROMET.EM.RF-K10.CL

Matrix of equivalence


FREQUENCY : 10 GHz

(Continued)

Lab <i>i</i>	Lab <i>j</i> →															
	SPRING		SP		NPL		AREPA		NMi-VSL		LNE		UME		PTB	
	$D_{ij}$ / 10 <sup>-3</sup>	$U_{ij}$ / 10 <sup>-3</sup>														
NMIJ	2.2	4.5	1.6	14.7												
NIST	2.6	17.8	2.0	22.6												
METAS	1.0	20.1	0.4	24.5												
CSIR-NML	5.0	22.1	4.4	26.2												
PTB	-0.4	8.4	-1.0	16.3												
NMIA	2.0	9.7	1.4	17.0												
NPL	-2.0	9.3	-2.6	16.8												
MIRS/SIQ	-7.0	14.2	-7.6	19.9												
INRIM	-4.5	30.7	-5.1	33.7												
VNIIFTRI	21.0	20.1	20.4	24.5												
SPRING	0.6	14.4														
SP	-0.3	9.4														
NPL	-3.2	8.8														
AREPA	4.6	14.0														
NMi-VSL	-6.0	13.8	Not computed													
LNE	12.1	17.0														
UME	13.5	7.8														
PTB	1.4	16.0														

Laboratory having used the repaired transfer standard PTB 2-6-1

**CCEM.RF-K10.CL and EUROMET.EM.RF-K10.CL Calibration factor, frequency: 10 GHz,  
Transfer standard: PTB 2-6**  
**Degrees of equivalence [  $D_i$  and its expanded uncertainty ( $k = 2$ ),  $U_i$  ]**



Open symbols represent values from laboratories in Associate States and Economies of the CGPM  
Blue triangles: participants in EUROMET.EM.RF-K10.CL

MEASURAND : Calibration factor

TRANSFER STANDARDS : Power sensors PTB 2-6, PTB 2-6-1 for CCEM.RF-K10.CL and PTB 2-06-2 for EUROMET.EM.RF-K10.CL

Matrix of equivalence

FREQUENCY : 18 GHz

Lab <i>i</i>	$D_i$ / $10^{-3}$	$U_i$ / $10^{-3}$
NMIJ	1.3	7.9
NIST	-0.5	17.7
METAS	-1.9	20.1
CSIR-NML	-4.9	26.1
PTB	-3.1	10.2
NMIA	3.1	12.8
NPL	-0.9	11.2
MIRS/SIQ	3.1	16.1
INRIM	15.2	29.3
VNIIFTRI	-22.9	24.1
SPRING	-5.5	20.1

Lab <i>j</i>	FREQUENCY : 18 GHz																					
	NMJJ		NIST		METAS		CSIR-NML		PTB		NMIA		NPL		MIRS/SIQ		INRIM		VNIIIFTRI			
	$D_{ij}$ / $10^{-3}$	$U_{ij}$ / $10^{-3}$																				
NMIJ	1.3	7.9	1.8	19.2	3.2	21.4	6.2	27.1	4.4	12.6	-1.8	14.7	2.2	13.4	-1.8	17.7	-13.9	30.2	24.2	25.2		
NIST	-0.5	17.7	-1.8	19.2	1.4	26.6	4.4	31.4	2.6	20.2	-3.6	21.6	0.4	20.8	-3.6	23.8	-15.7	34.1	22.4	29.8		
METAS	-1.9	20.1	-3.2	21.4	-1.4	26.6		3.0	32.8	1.2	22.4	-5.0	23.6	-1.0	22.8	-5.0	25.6	-17.1	35.4	21.0	31.2	
CSIR-NML	-4.9	26.1	-6.2	27.1	-4.4	31.4	-3.0	32.8		-1.8	27.9	-8.0	28.9	-4.0	28.2	-8.0	30.5	-20.1	39.1	18.0	35.4	
PTB	-3.1	10.2	-4.4	12.6	-2.6	20.2	-1.2	22.4	1.8	27.9		-6.2	16.1	-2.2	14.9	-6.2	18.9	-18.3	30.9	19.8	26.0	
NMIA	3.1	12.8	1.8	14.7	3.6	21.6	5.0	23.6	8.0	28.9	6.2	16.1		4.0	16.7	0.0	20.4	-12.1	31.8	26.0	27.1	
NPL	-0.9	11.2	-2.2	13.4	-0.4	20.8	1.0	22.8	4.0	28.2	2.2	14.9	-4.0	16.7		-4.0	19.4	-16.1	31.2	22.0	26.4	
MIRS/SIQ	3.1	16.1	1.8	17.7	3.6	23.8	5.0	25.6	8.0	30.5	6.2	18.9	0.0	20.4	4.0	19.4		-12.1	33.3	26.0	28.8	
INRIM	15.2	29.3	13.9	30.2	15.7	34.1	17.1	35.4	20.1	39.1	18.3	30.9	12.1	31.8	16.1	31.2	12.1	33.3	38.1	37.8		
VNIIIFTRI	-22.9	24.1	-24.2	25.2	-22.4	29.8	-21.0	31.2	-18.0	35.4	-19.8	26.0	-26.0	27.1	-22.0	26.4	-26.0	28.8	-38.1	37.8		
SPRING	-5.5	20.1	-6.8	21.4	-5.0	26.6	-3.6	28.3	-0.6	32.8	-2.4	22.4	-8.6	23.6	-4.6	22.8	-8.6	25.6	-20.7	35.4	17.4	31.2

SP	-8.4	13.2
NPL	-7.1	10.8
AREPA	5.6	22.0
NMI-VSL	-3.3	16.6
LNE	4.2	23.8
UME	10.9	16.4
PTB	3.1	10.0

Not computed																				
Not computed																				
Not computed																				
Not computed																				

Laboratory having used the repaired transfer standard PTB 2-6-1

MEASURAND : Calibration factor

TRANSFER STANDARDS : Power sensors PTB 2-6, PTB 2-6-1 for CCEM.RF-K10.CL and PTB 2-06-2 for EUROMET.EM.RF-K10.CL

Matrix of equivalence


FREQUENCY : 18 GHz

(Continued)

Lab <i>i</i>	Lab <i>j</i> →															
	SPRING		SP		NPL		AREPA		NMi-VSL		LNE		UME		PTB	
	$D_{ij}$ / $10^{-3}$	$U_{ij}$ / $10^{-3}$														
NMIJ	1.3	7.9	6.8	21.4												
NIST	-0.5	17.7	5.0	26.6												
METAS	-1.9	20.1	3.6	28.3												
CSIR-NML	-4.9	26.1	0.6	32.8												
PTB	-3.1	10.2	2.4	22.4												
NMIA	3.1	12.8	8.6	23.6												
NPL	-0.9	11.2	4.6	22.8												
MIRS/SIQ	3.1	16.1	8.6	25.6												
INRIM	15.2	29.3	20.7	35.4												
VNIIFTRI	-22.9	24.1	-17.4	31.2												
SPRING	-5.5	20.1														
SP	-8.4	13.2														
NPL	-7.1	10.8														
AREPA	5.6	22.0														
NMi-VSL	-3.3	16.6														
LNE	4.2	23.8														
UME	10.9	16.4														
PTB	3.1	10.0														

Laboratory having used the repaired transfer standard PTB 2-6-1

**CCEM.RF-K10.CL and EUROMET.EM.RF-K10.CL Calibration factor, frequency: 18 GHz,  
Transfer standard: PTB 2-6**  
**Degrees of equivalence [  $D_i$  and its expanded uncertainty ( $k = 2$ ),  $U_i$  ]**



Open symbols represent values from laboratories in Associate States and Economies of the CGPM  
 Blue triangles: participants in EUROMET.EM.RF-K10.CL

## **MEASURAND : Calibration factor**

**TRANSFER STANDARDS :** Power sensors PTB 2-6, PTB 2-6-1 for CCEM.RF-K10.CL and PTB 2-06-2 for EUROMET.EM.RF-K10.CL

## Matrix of equivalence

FREQUENCY : 20 GHz

## Lab *j* →

<b>Lab <i>i</i></b>	<b><math>D_i</math> <math>/ 10^{-3}</math></b>	<b><math>U_i</math> <math>/ 10^{-3}</math></b>
NMIJ	1.3	9.9
NIST	-0.8	18.3
METAS	4.0	22.2
CSIR-NML	5.0	30.2
PTB	-5.6	11.5
NMIA	4.0	18.9
NPL	1.0	11.6
MIRS/SIQ	6.0	16.3
INRIM	4.0	29.2
VNIIFTRI	-25.0	44.1
SPRING	-6.6	20.8

NMJ		NIST		METAS		CSIR-NML		PTB		NMIA		NPL		MIRS/SIQ		INRIM		VNIIFTRI	
$D_{ij}$ / 10 <sup>-3</sup>	$U_{ij}$ / 10 <sup>-3</sup>																		
		2.1	20.3	-2.7	23.9	-3.7	31.4	6.9	14.5	-2.7	20.8	0.3	14.6	-4.7	18.6	-2.7	30.5	26.3	45.0
-2.1	20.3			-4.8	28.4	-5.8	35.0	4.8	21.1	-4.8	25.9	-1.8	21.2	-6.8	24.1	-4.8	34.1	24.2	47.5
2.7	23.9	4.8	28.4			-1.0	37.2	9.6	24.6	0.0	28.8	3.0	24.7	-2.0	27.2	0.0	36.4	29.0	49.2
3.7	31.4	5.8	35.0	1.0	37.2			10.6	32.0	1.0	35.3	4.0	32.0	-1.0	34.0	1.0	41.7	30.0	53.3
-6.9	14.5	-4.8	21.1	-9.6	24.6	-10.6	32.0			-9.6	21.6	-6.6	15.7	-11.6	19.4	-9.6	31.0	19.4	45.4
2.7	20.8	4.8	25.9	0.0	28.8	-1.0	35.3	9.6	21.6			3.0	21.7	-2.0	24.5	0.0	34.5	29.0	47.8
-0.3	14.6	1.8	21.2	-3.0	24.7	-4.0	32.0	6.6	15.7	-3.0	21.7			-5.0	19.5	-3.0	31.1	26.0	45.4
4.7	18.6	6.8	24.1	2.0	27.2	1.0	34.0	11.6	19.4	2.0	24.5	5.0	19.5			2.0	33.1	31.0	46.8
2.7	30.5	4.8	34.1	0.0	36.4	-1.0	41.7	9.6	31.0	0.0	34.5	3.0	31.1	-2.0	33.1			29.0	52.7
-26.3	45.0	-24.2	47.5	-29.0	49.2	-30.0	53.3	-19.4	45.4	-29.0	47.8	-26.0	45.4	-31.0	46.8	-29.0	52.7		
-7.9	22.6	-5.8	27.4	-10.6	30.1	-11.6	36.4	-1.0	23.4	-10.6	27.8	-7.6	23.4	-12.6	26.1	-10.6	35.6	18.4	48.6

<b>SP</b>	<b>-0.7</b>	13.2
<b>NPL</b>	<b>-3.3</b>	11.0
<b>AREPA</b>	<b>11.3</b>	26.0
<b>NMi-VSL</b>	<b>-1.5</b>	16.6
<b>LNE</b>	<b>4.5</b>	25.6
<b>UME</b>	<b>16.3</b>	16.6
<b>PTB</b>	<b>2.1</b>	11.0

Laboratory having used the repaired transfer standard PTB 2-6-1

MEASURAND : Calibration factor

TRANSFER STANDARDS : Power sensors PTB 2-6, PTB 2-6-1 for CCEM.RF-K10.CL and PTB 2-06-2 for EUROMET.EM.RF-K10.CL

Matrix of equivalence

FREQUENCY : 20 GHz

(Continued)


Lab <i>i</i>	Lab <i>j</i>	
	SPRING	SPRING
	$D_{ij}$ / $10^{-3}$	$U_{ij}$ / $10^{-3}$
NMIJ	1.3	9.9
NIST	-0.8	18.3
METAS	4.0	22.2
CSIR-NML	5.0	30.2
PTB	-5.6	11.5
NMIA	4.0	18.9
NPL	1.0	11.6
MIRS/SIQ	6.0	16.3
INRIM	4.0	29.2
VNIIFTRI	-25.0	44.1
SPRING	-6.6	20.8

SP	-0.7	13.2			2.6	17.2	-12.0	29.2	0.8	21.2	-5.2	28.8	-17.0	21.2	-2.8	17.2	
NPL	-3.3	11.0			-2.6	17.2	-14.6	28.2	-1.8	19.9	-7.8	27.9	-19.6	19.9	-5.4	15.6	
AREPA	11.3	26.0			12.0	29.2	14.6	28.2	12.8	30.8	6.8	36.5	-5.0	30.8	9.2	28.2	
NMi-VSL	-1.5	16.6	Not computed		-0.8	21.2	1.8	19.9	-12.8	30.8	-6.0	30.5	-17.8	23.5	-3.6	19.9	
LNE	4.5	25.6			5.2	28.8	7.8	27.9	-6.8	36.5	6.0	30.5	-11.8	30.5	2.4	27.9	
UME	16.3	16.6			17.0	21.2	19.6	19.9	5.0	30.8	17.8	23.5	11.8	30.5		14.2	19.9
PTB	2.1	11.0			2.8	17.2	5.4	15.6	-9.2	28.2	3.6	19.9	-2.4	27.9	-14.2	19.9	

Laboratory having used the repaired transfer standard PTB 2-6-1

**CCEM.RF-K10.CL and EUROMET.EM.RF-K10.CL Calibration factor, frequency: 20 GHz,  
Transfer standard: PTB 2-6**  
**Degrees of equivalence [  $D_i$  and its expanded uncertainty ( $k = 2$ ),  $U_i$  ]**



Open symbols represent values from laboratories in Associate States and Economies of the CGPM  
 Blue triangles: participants in EUROMET.EM.RF-K10.CL

## **MEASURAND : Calibration factor**

**TRANSFER STANDARDS :** Power sensors PTB 2-6, PTB 2-6-1 for CCEM.RF-K10.CL and PTB 2-06-2 for EUROMET.EM.RF-K10.CL

## Matrix of equivalence

**FREQUENCY : 23 GHz**

## Lab *j* →

<b>Lab <i>i</i></b>	<b><math>D_i</math> <math>/ 10^{-3}</math></b>	<b><math>U_i</math> <math>/ 10^{-3}</math></b>
NMIJ	<b>0.8</b>	9.8
NIST	<b>0.3</b>	18.9
METAS	<b>4.4</b>	20.4
CSIR-NML	<b>4.4</b>	32.3
PTB	<b>-0.7</b>	13.7
NMIA	<b>3.4</b>	18.9
NPL	<b>6.4</b>	14.0
MIRS/SIQ	<b>2.4</b>	16.5
INRIM	<b>23.7</b>	41.6
VNIIFTRI	<b>-6.6</b>	58.2
SPRING	<b>0.5</b>	27.1

NMJJ		NIST		METAS		CSIR-NML		PTB		NMIA		NPL		MIRS/SIQ		INRIM		VNIIIFTRI	
$D_{ij}$ / 10 <sup>-3</sup>	$U_{ij}$ / 10 <sup>-3</sup>																		
		<b>0.5</b>	20.4	<b>-3.6</b>	21.9	<b>-3.6</b>	33.2	<b>1.5</b>	15.7	<b>-2.6</b>	20.4	<b>-5.6</b>	16.0	<b>-1.6</b>	18.3	<b>-22.9</b>	42.3	<b>7.4</b>	58.7
<b>-0.5</b>	20.4			<b>-4.1</b>	27.2	<b>-4.1</b>	36.9	<b>1.0</b>	22.5	<b>-3.1</b>	26.0	<b>-6.1</b>	22.8	<b>-2.1</b>	24.4	<b>-23.4</b>	45.3	<b>6.9</b>	60.8
<b>3.6</b>	21.9	<b>4.1</b>	27.2			<b>0.0</b>	37.7	<b>5.1</b>	23.9	<b>1.0</b>	27.2	<b>-2.0</b>	24.1	<b>2.0</b>	25.6	<b>-19.3</b>	46.0	<b>11.0</b>	61.4
<b>3.6</b>	33.2	<b>4.1</b>	36.9	<b>0.0</b>	37.7			<b>5.1</b>	34.5	<b>1.0</b>	36.9	<b>-2.0</b>	34.7	<b>2.0</b>	35.8	<b>-19.3</b>	52.3	<b>11.0</b>	66.2
<b>-1.5</b>	15.7	<b>-1.0</b>	22.5	<b>-5.1</b>	23.9	<b>-5.1</b>	34.5			<b>-4.1</b>	22.5	<b>-7.1</b>	18.7	<b>-3.1</b>	20.6	<b>-24.4</b>	43.4	<b>5.9</b>	59.4
<b>2.6</b>	20.4	<b>3.1</b>	26.0	<b>-1.0</b>	27.2	<b>-1.0</b>	36.9	<b>4.1</b>	22.5			<b>-3.0</b>	22.8	<b>1.0</b>	24.4	<b>-20.3</b>	45.3	<b>10.0</b>	60.8
<b>5.6</b>	16.0	<b>6.1</b>	22.8	<b>2.0</b>	24.1	<b>2.0</b>	34.7	<b>7.1</b>	18.7	<b>3.0</b>	22.8			<b>4.0</b>	20.9	<b>-17.3</b>	43.5	<b>13.0</b>	59.5
<b>1.6</b>	18.3	<b>2.1</b>	24.4	<b>-2.0</b>	25.6	<b>-2.0</b>	35.8	<b>3.1</b>	20.6	<b>-1.0</b>	24.4	<b>-4.0</b>	20.9			<b>-21.3</b>	44.4	<b>9.0</b>	60.2
<b>22.9</b>	42.3	<b>23.4</b>	45.3	<b>19.3</b>	46.0	<b>19.3</b>	52.3	<b>24.4</b>	43.4	<b>20.3</b>	45.3	<b>17.3</b>	43.5	<b>21.3</b>	44.4			<b>30.3</b>	71.3
<b>-7.4</b>	58.7	<b>-6.9</b>	60.8	<b>-11.0</b>	61.4	<b>-11.0</b>	66.2	<b>-5.9</b>	59.4	<b>-10.0</b>	60.8	<b>-13.0</b>	59.5	<b>-9.0</b>	60.2	<b>-30.3</b>	71.3		
<b>-0.3</b>	28.2	<b>0.2</b>	32.5	<b>-3.9</b>	33.4	<b>-3.9</b>	41.7	<b>1.2</b>	29.8	<b>-2.9</b>	32.5	<b>-5.9</b>	30.0	<b>-1.9</b>	31.2	<b>-23.2</b>	49.3	<b>7.1</b>	63.9

<b>SP</b>	<b>0.3</b>	16.0
<b>NPL</b>	<b>1.1</b>	13.2
<b>AREPA</b>	<b>11.3</b>	30.0
<b>NMi-VSL</b>	<b>-0.8</b>	16.4
<b>LNE</b>	<b>18.1</b>	21.4
<b>UME</b>	<b>26.5</b>	22.0
<b>PTB</b>	<b>4.6</b>	13.0

Laboratory having used the repaired transfer standard PTB 2-6-1

MEASURAND : Calibration factor

TRANSFER STANDARDS : Power sensors PTB 2-6, PTB 2-6-1 for CCEM.RF-K10.CL and PTB 2-06-2 for EUROMET.EM.RF-K10.CL

Matrix of equivalence

FREQUENCY : 23 GHz

(Continued)


Lab <i>i</i>	Lab <i>j</i>	
	SPRING	
	$D_{ij}$ / $10^{-3}$	$U_{ij}$ / $10^{-3}$
NMIJ	0.8	9.8
NIST	0.3	18.9
METAS	4.4	20.4
CSIR-NML	4.4	32.3
PTB	-0.7	13.7
NMIA	3.4	18.9
NPL	6.4	14.0
MIRS/SIQ	2.4	16.5
INRIM	23.7	41.6
VNIIFTRI	-6.6	58.2
SPRING	0.5	27.1

SP	0.3	16.0			-0.8	20.7	-11.0	34.0	1.1	22.9	-17.8	26.7	-26.2	27.2	-4.3	20.6
NPL	1.1	13.2			0.8	20.7	-10.2	32.8	1.9	21.1	-17.0	25.1	-25.4	25.7	-3.5	18.5
AREPA	11.3	30.0			11.0	34.0	10.2	32.8	12.1	34.2	-6.8	36.9	-15.2	37.2	6.7	32.7
NMi-VSL	-0.8	16.4			-1.1	22.9	-1.9	21.1	-12.1	34.2	-18.9	27.0	-27.3	27.4	-5.4	20.9
LNE	18.1	21.4			17.8	26.7	17.0	25.1	6.8	36.9	18.9	27.0	-8.4	30.7	13.5	25.0
UME	26.5	22.0			26.2	27.2	25.4	25.7	15.2	37.2	27.3	27.4	8.4	30.7	21.9	25.6
PTB	4.6	13.0			4.3	20.6	3.5	18.5	-6.7	32.7	5.4	20.9	-13.5	25.0	-21.9	25.6

Laboratory having used the repaired transfer standard PTB 2-6-1

**CCEM.RF-K10.CL and EUROMET.EM.RF-K10.CL Calibration factor, frequency: 23 GHz,  
Transfer standard: PTB 2-6**  
**Degrees of equivalence [  $D_i$  and its expanded uncertainty ( $k = 2$ ),  $U_i$  ]**



Open symbols represent values from laboratories in Associate States and Economies of the CGPM  
 Blue triangles: participants in EUROMET.EM.RF-K10.CL

## **MEASURAND : Calibration factor**

**TRANSFER STANDARDS :** Power sensors PTB 2-6, PTB 2-6-1 for CCEM.RF-K10.CL and PTB 2-06-2 for EUROMET.EM.RF-K10.CL

## Matrix of equivalence

FREQUENCY : 26 GHz

## Lab *j* →

<b>Lab <i>i</i></b>	<b><math>D_i</math> <math>/ 10^{-3}</math></b>	<b><math>U_i</math> <math>/ 10^{-3}</math></b>
NMIJ	-17.1	18.1
NIST	-7.9	21.6
METAS	-15.7	26.1
CSIR-NML	-4.7	39.3
PTB	-5.2	19.8
NMIA	-2.7	21.2
NPL	3.3	18.5
MIRS/SIQ	-6.7	19.0
INRIM	17.5	42.4
VNIIFTRI	9.3	31.7
SPRING	1.3	20.3

NMJJ		NIST		METAS		CSIR-NML		PTB		NMIA		NPL		MIRS/SIQ		INRIM		VNIIIFTRI	
$D_{ij}$ / 10 <sup>-3</sup>	$U_{ij}$ / 10 <sup>-3</sup>																		
		-9.2	24.2	-1.4	28.3	-12.4	40.9	-11.9	22.7	-14.4	23.9	-20.4	21.5	-10.4	21.9	-34.6	43.8	-26.4	33.5
9.2	24.2			7.8	30.6	-3.2	42.5	-2.7	25.5	-5.2	26.6	-11.2	24.5	-1.2	24.8	-25.4	45.4	-17.2	35.5
1.4	28.3	-7.8	30.6			-11.0	44.9	-10.5	29.4	-13.0	30.4	-19.0	28.5	-9.0	28.8	-33.2	47.7	-25.0	38.4
12.4	40.9	3.2	42.5	11.0	44.9			0.5	41.6	-2.0	42.3	-8.0	41.0	2.0	41.2	-22.2	56.0	-14.0	48.4
11.9	22.7	2.7	25.5	10.5	29.4	-0.5	41.6			-2.5	25.2	-8.5	22.9	1.5	23.3	-22.7	44.6	-14.5	34.5
14.4	23.9	5.2	26.6	13.0	30.4	2.0	42.3	2.5	25.2			-6.0	24.1	4.0	24.5	-20.2	45.2	-12.0	35.3
20.4	21.5	11.2	24.5	19.0	28.5	8.0	41.0	8.5	22.9	6.0	24.1			10.0	22.2	-14.2	44.0	-6.0	33.7
10.4	21.9	1.2	24.8	9.0	28.8	-2.0	41.2	-1.5	23.3	-4.0	24.5	-10.0	22.2			-24.2	44.2	-16.0	34.0
34.6	43.8	25.4	45.4	33.2	47.7	22.2	56.0	22.7	44.6	20.2	45.2	14.2	44.0	24.2	44.2			8.2	51.0
26.4	33.5	17.2	35.5	25.0	38.4	14.0	48.4	14.5	34.5	12.0	35.3	6.0	33.7	16.0	34.0	-8.2	51.0		
18.4	23.1	9.2	25.9	17.0	29.8	6.0	41.9	6.5	24.5	4.0	25.6	-2.0	23.4	8.0	23.8	-16.2	44.8	-8.0	34.8

<b>SP</b>	<b>5.0</b>	21.0
<b>NPL</b>	<b>1.3</b>	15.4
<b>AREPA</b>	<b>3.0</b>	34.0
<b>NMi-VSL</b>	<b>-7.4</b>	18.2
<b>LNE</b>	<b>6.7</b>	20.2
<b>UME</b>	<b>26.7</b>	36.2
<b>PTB</b>	<b>-0.3</b>	17.0

Laboratory having used the repaired transfer standard PTB 2-6-1

MEASURAND : Calibration factor

TRANSFER STANDARDS : Power sensors PTB 2-6, PTB 2-6-1 for CCEM.RF-K10.CL and PTB 2-06-2 for EUROMET.EM.RF-K10.CL

Matrix of equivalence

FREQUENCY : 26 GHz (Continued)


Lab <i>i</i>	Lab <i>j</i>	
	SPRING	
	$D_{ij}$ / $10^{-3}$	$U_{ij}$ / $10^{-3}$
NMIJ	-17.1	18.1
NIST	-7.9	21.6
METAS	-15.7	26.1
CSIR-NML	-4.7	39.3
PTB	-5.2	19.8
NMIA	-2.7	21.2
NPL	3.3	18.5
MIRS/SIQ	-6.7	19.0
INRIM	17.5	42.4
VNIIFTRI	9.3	31.7
SPRING	1.3	20.3

SP	5.0	21.0			3.7	26.0	2.0	40.0	12.4	27.8	-1.7	29.1	-21.7	41.9	5.3	27.0
NPL	1.3	15.4			-3.7	26.0	-1.7	37.3	8.7	23.8	-5.4	25.4	-25.4	39.3	1.6	22.9
AREPA	3.0	34.0			-2.0	40.0	1.7	37.3	10.4	38.6	-3.7	39.5	-23.7	49.7	3.3	38.0
NMi-VSL	-7.4	18.2			-12.4	27.8	-8.7	23.8	-10.4	38.6	-14.1	27.2	-34.1	40.5	-7.1	24.9
LNE	6.7	20.2			1.7	29.1	5.4	25.4	3.7	39.5	14.1	27.2	-20.0	41.5	7.0	26.4
UME	26.7	36.2			21.7	41.9	25.4	39.3	23.7	49.7	34.1	40.5	20.0	41.5	27.0	40.0
PTB	-0.3	17.0			-5.3	27.0	-1.6	22.9	-3.3	38.0	7.1	24.9	-7.0	26.4	-27.0	40.0

Laboratory having used the repaired transfer standard PTB 2-6-1

**CCEM.RF-K10.CL and EUROMET.EM.RF-K10.CL Calibration factor, frequency: 26 GHz,  
Transfer standard: PTB 2- 6**  
**Degrees of equivalence [  $D_i$  and its expanded uncertainty ( $k = 2$ ),  $U_i$  ]**



Open symbols represent values from laboratories in Associate States and Economies of the CGPM  
 Blue triangles: participants in EUROMET.EM.RF-K10.CL