

Key comparison CCQM-K12

MEASURAND: **Mass fraction of creatinine in human serum
Material I (physiological range)**

x_i : **result of measurement carried out by laboratory i**

u_i : **combined standard uncertainty of x_i**

Lab i	x_i / ($\mu\text{g/g}$)	u_i / ($\mu\text{g/g}$)
IRMM	8.360	0.1060
KRISS	8.186	0.0796
LGC	8.193	0.0080
NIST	8.277	0.0319
PTB	8.211	0.0289

Key comparison CCQM-K12

MEASURAND: Mass fraction of creatinine in human serum
Material I (physiological range)

The key comparison reference value calculated as the mean of the participant results, excluding IRMM is: $x_R = 8.217 \mu\text{g/g}$.

The expanded uncertainty of the key comparison reference value at a 95 % level of confidence is: $U_R = 0.066 \mu\text{g/g}$.


The degree of equivalence of each laboratory with respect to the reference value is given by a pair of terms: $D_i = (x_i - x_R)$ and U_i , its expanded uncertainty corresponding to a 95% level of confidence, both expressed in $\mu\text{g/g}$.

The degree of equivalence between two laboratories is given by a pair of terms: $D_{ij} = D_i - D_j = (x_i - x_R) - (x_j - x_R) = x_i - x_j$ and U_{ij} , its expanded uncertainty corresponding to a 95% level of confidence, both expressed in $\mu\text{g/g}$.

Lab *j* →

Lab <i>i</i> ↓	D_i / ($\mu\text{g/g}$)		U_i / ($\mu\text{g/g}$)		IRMM		KRISS		LGC		NIST		PTB	
	D_{ij} / ($\mu\text{g/g}$)	U_{ij} / ($\mu\text{g/g}$)	D_{ij} / ($\mu\text{g/g}$)	U_{ij} / ($\mu\text{g/g}$)	D_{ij} / ($\mu\text{g/g}$)	U_{ij} / ($\mu\text{g/g}$)	D_{ij} / ($\mu\text{g/g}$)	U_{ij} / ($\mu\text{g/g}$)	D_{ij} / ($\mu\text{g/g}$)	U_{ij} / ($\mu\text{g/g}$)	D_{ij} / ($\mu\text{g/g}$)	U_{ij} / ($\mu\text{g/g}$)	D_{ij} / ($\mu\text{g/g}$)	U_{ij} / ($\mu\text{g/g}$)
	IRMM	0.143	0.216					0.174	0.263	0.167	0.213	0.083	0.221	0.149
KRISS	-0.031	0.164	-0.174	0.263			-0.007	0.160	-0.091	0.171	-0.091	0.171	-0.025	0.168
LGC	-0.024	0.071	-0.167	0.213	0.007	0.160			-0.084	0.091	-0.018	0.060	-0.018	0.060
NIST	0.060	0.093	-0.083	0.221	0.091	0.171	0.084	0.091			0.066	0.094	0.066	0.094
PTB	-0.006	0.074	-0.149	0.219	0.025	0.168	0.018	0.060	-0.066	0.094				

CCQM-K12 Creatinine in Human Serum (Material I)
Degrees of equivalence D_i and expanded uncertainty at a 95% level of confidence U_i



CCQM-K12 and EURAMET.QM-K12

Key comparison CCQM-K12

MEASURAND: **Mass fraction of creatinine in human serum
Material II (elevated range)**

x_i : **result of measurement carried out by laboratory i**
 u_i : **combined standard uncertainty of x_i**

Lab i	x_i / ($\mu\text{g/g}$)	u_i / ($\mu\text{g/g}$)
IRMM	18.720	0.2396
KRISS	18.539	0.1627
LGC	18.614	0.0316
NIST	18.708	0.0722
PTB	18.718	0.0650

Key comparison EURAMET.QM-K12

MEASURAND: **Mass fraction of creatinine in human serum**

$x_{i\text{-EUR}}$: **result of measurement carried out by laboratory i participant in EURAMET.QM-K12**
 $u_{i\text{-EUR}}$: **combined standard uncertainty of $x_{i\text{-EUR}}$**

Material A - high level

Lab i	$x_{i\text{-EUR}}$ / ($\mu\text{g/g}$)	$u_{i\text{-EUR}}$ / ($\mu\text{g/g}$)
LGC	53.50	0.27
LNE	54.04	0.22
HSA	54.24	0.29
EXHM/GCSL-EIM	54.509	0.697
PTB	55.06	0.41

Material B - low level

Lab i	$x_{i\text{-EUR}}$ / ($\mu\text{g/g}$)	$u_{i\text{-EUR}}$ / ($\mu\text{g/g}$)
LGC	37.71	0.12
LNE	37.70	0.22
HSA	37.94	0.22
EXHM/GCSL-EIM	38.438	0.4534
PTB	38.28	0.29

Key comparison CCQM-K12

MEASURAND: Mass fraction of creatinine in human serum
Material II (elevated range)

The key comparison reference value calculated as the mean of the participant results, excluding IRMM is: $x_R = 18.645 \mu\text{g/g}$.

The expanded uncertainty of the key comparison reference value at a 95 % level of confidence is: $U_R = 0.135 \mu\text{g/g}$.

The degree of equivalence of each laboratory with respect to the reference value is given by a pair of terms:

$D_i = (x_i - x_R)$ and U_i , its expanded uncertainty corresponding to a 95% level of confidence, both expressed in $\mu\text{g/g}$.

D_i and U_i are also given in relative terms.

The degree of equivalence between two laboratories is given by a pair of terms:

$D_{ij} = D_i - D_j = (x_i - x_R) - (x_j - x_R) = x_i - x_j$ and U_{ij} , its expanded uncertainty corresponding to a 95% level of confidence, both expressed in $\mu\text{g/g}$.

Linking EURAMET.QM-K12 to key comparison CCQM-K12

The linkage is obtained through the common participation of PTB and LGC in both key comparisons and is explained on page 14 of the EURAMET.QM-K12 Final Report. The set of results obtained for Material II in CCQM-K12 was used to link the results obtained for both material A and B in EURAMET.QM-K12, as this was closer in mass fraction nominal value. It leads to the computation of the degrees of equivalence of participants in EURAMET.QM-K12 relative to the CCQM-K12 reference value, and the D_i and U_i values are given in relative terms.

No pair-wise degrees of equivalence involving participants in EURAMET.QM-K12 only have been computed.

Degrees of equivalence D_i and U_i expressed in relative terms

Lab <i>i</i>	D_i/x_R / %	U_i/x_R / %
IRMM	0.40	2.61
KRISS	-0.57	1.78
LGC	-0.16	0.67
NIST	0.34	1.10
PTB	0.39	0.86
LGC	-1.42	1.38
LNE	-0.42	1.25
HSA	-0.06	1.43
EXHM/GCSL-EIM	0.44	2.74
PTB	1.46	1.78
LGC	-0.80	1.01
LNE	-0.82	1.40
HSA	-0.19	1.38
EXHM/GCSL-EIM	1.12	2.51
PTB	0.70	1.72

Black indicates participants in CCQM-K12


Blue indicates participants in EURAMET.QM-K12 (Material A)

Green indicates participants in EURAMET.QM-K12 (Material B)

Key comparison CCQM-K12

MEASURAND: Mass fraction of creatinine in human serum
Material II (elevated range)

Matrix of equivalence in absolute terms

Lab *j* 

Lab *i*

	D_i / (µg/g)	U_i / (µg/g)
IRMM	0.075	0.487
KRISS	-0.106	0.331
LGC	-0.031	0.125
NIST	0.063	0.205
PTB	0.073	0.161

IRMM		KRISS		LGC		NIST		PTB	
D_{ij} / (µg/g)	U_{ij} / (µg/g)	D_{ij} / (µg/g)	U_{ij} / (µg/g)	D_{ij} / (µg/g)	U_{ij} / (µg/g)	D_{ij} / (µg/g)	U_{ij} / (µg/g)	D_{ij} / (µg/g)	U_{ij} / (µg/g)
		0.181	0.574	0.106	0.484	0.012	0.501	0.002	0.496
-0.181	0.574			-0.075	0.326	-0.169	0.353	-0.179	0.345
-0.106	0.484	0.075	0.326			-0.094	0.203	-0.104	0.145
-0.012	0.501	0.169	0.353	0.094	0.203			-0.010	0.214
-0.002	0.496	0.179	0.345	0.104	0.145	0.010	0.214		

CCQM-K12 and EURAMET.QM-K12 Creatinine in Human Serum (Material II)

Degrees of equivalence D_i and expanded uncertainty at a 95% level of confidence U_i expressed in relative terms

