

CCM.P-K4 and EUROMET.M.P-K1.a for nominal pressures 1 Pa, 3 Pa, 10 Pa, 30 Pa, 100 Pa, 300 Pa, and 1000 Pa

CCM.P-K4 and EURAMET.M.P-K4.2010 for nominal pressures 3 Pa, 10 Pa, and 30 Pa

CCM.P-K4, EUROMET.M.P-K1.a, EUROMET.M.P-K1.b and SIM-EUROMET.M.P-BK3 at a pressure close to 1 Pa

This file covers the results of participants in key comparisons CCM.P-K4 and EUROMET.M.P-K1.a linked for 7 nominal pressures in the range 1 Pa to 1000 Pa.

Additional link is carried out with the results of participants in key comparisons EUROMET.M.P-K1.b and SIM-EUROMET.M.P-BK3 for a pressure close to 1 Pa, and in EURAMET.M.P-K4.2010 for nominal pressures of 3 Pa, 10 Pa and 30 Pa.

Results obtained in the framework of key comparisons EUROMET.M.P-K1.b and SIM-EUROMET.M.P-BK3 for the pressure range 0.3 mPa to 0.9 Pa are available from the KCDB Website by clicking on [EUROMET.M.P-K1.b](#) or [SIM-EUROMET.M.P-BK3](#).

Key comparison CCM.P-K4

MEASURAND: Pressure

NOMINAL VALUES: 1 Pa, 3 Pa, 10 Pa, 30 Pa, 100 Pa, 300 Pa and 1000 Pa

p_i : corrected mean gauge reading obtained by laboratory i (see CCM.P-K4 Final Report, page 15)

u_i : combined standard uncertainty of p_i (see CCM.P-K4 Final Report, page 22)

Lab i	Nominal pressure / Pa	p_i / Pa	u_i / Pa	Measurement method	Date of measurement	
CSIRO-NML	1	1.211	0.023	Liquid-Column Manometer	24 Feb 1999 to 3 Mar 1999	
	3	3.301	0.025			
	10	10.566	0.046			
	30	30.906	0.069			
	100	100.831	0.093			
	300	300.57	0.16			
	1000	1000.33	0.16			
CNR-IMGC	1	1.0023	0.0031	Static Expansion System	19 Jan 1999 to 8 Feb 1999	
	3	3.0042	0.0091			
	10	10.015	0.030			
	30	30.057	0.091			
	100	100.176	0.051	Liquid-Column Manometer		
	300	300.132	0.060			
KRISS	1000	1000.178	0.061			
	10	10.003	0.013	Liquid-Column Manometer	15 Jun 1999 to 22 Jun 1999	
	30	30.022	0.014			
	100	100.036	0.021			
	300	300.010	0.023			
	1000	1000.056	0.055			

Lab <i>i</i>	Nominal pressure / Pa	p_i / Pa	u_i / Pa	Measurement method	Date of measurement
NIST	1	1.0003	0.0010	Liquid-Column Manometer(s)	17 Mar to 03 Apr and 08 Jul to 24 Jul 1998, 23 Apr to 06 May and 23 Aug to 10 Sep 1999
	3	3.0008	0.0015		
	10	9.9997	0.0018		
	30	29.9898	0.0026		
	100	99.9909	0.0053		
	300	299.9871	0.0054		
	1000	1000.0308	0.0059		
NPLI	1	0.9952	0.0010	Static Expansion System	01 Jan 1999 to 14 Jan 1999
	3	2.9929	0.0019		
	10	9.9817	0.0058		
	30	29.953	0.017		
	100	99.874	0.055		
	300	299.76	0.16		
	1000	999.12	0.74		
NPL	1	1.0013	0.0031	Static Expansion System	17 Nov 1998 to 25 Nov 1998
	3	3.0008	0.0091		
	10	10.003	0.030		
	30	29.966	0.057		
	100	99.92	0.18		
	300	299.80	0.54		
	1000	1000.7	1.8		
PTB	1	1.0002	0.0015	Static Expansion System	28 May 1998 to 04 Jun 1998
	3	2.9987	0.0038		
	10	9.995	0.013		
	30	30.002	0.033		
	100	100.00	0.11		
	300	299.92	0.32		
	1000	999.7	1.1		

CCM.P-K4 and EUROMET.M.P-K1.a for nominal pressures 1 Pa, 3 Pa, 10 Pa, 30 Pa, 100 Pa, 300 Pa, and 1000 Pa

CCM.P-K4 and EURAMET.M.P-K4.2010 for nominal pressures 3 Pa, 10 Pa, and 30 Pa

CCM.P-K4, EUROMET.M.P-K1.a, EUROMET.M.P-K1.b and SIM-EUROMET.M.P-BK3 at a pressure close to 1 Pa

Key comparison EUROMET.M.P-K1.a

MEASURAND: Pressure

NOMINAL VALUES: 1 Pa, 3 Pa, 10 Pa, 30 Pa, 100 Pa, 300 Pa and 1000 Pa

The laboratory individual measurements of the participants in the key comparison EUROMET.M.P-K1.a are given in Table 7 and 8 on pages 17, 18 and 19 of the EUROMET.M.P-K1.a Final Report.

All measurements used a Static Expansion System and were carried out between September 1998 and April 2004.

Key comparison EUROMET.M.P-K1.b

MEASURAND: Pressure

NOMINAL VALUE: 0.9 Pa

p_i pressure measurement carried out at laboratory i

u_i standard uncertainty of laboratory i

(values taken from Table 13 of the EUROMET.M.P-K1.b Final Report)

Lab i	p_i / Pa	u_i / Pa	Date of measurement
NPL	9.004E-01	2.6E-03	May 2000
BNM-LNE	9.000E-01	5.1E-03	Jun 2000
MIRS/IMT	9.059E-01	2.7E-03	Sep 2000
CNR-IMGC	9.016E-01	2.3E-03	Oct 2000 to Nov 2000
CEM	9.021E-01	6.4E-03	Apr 2001
UME	8.988E-01	2.5E-03	Dec 2001 to Jan 2002
PTB	8.994E-01	1.7E-03	Apr 2000 - Jul 2000 - Mar 2001 - Nov 2001 - Feb 2002

CCM.P-K4 and EUROMET.M.P-K1.a for nominal pressures 1 Pa, 3 Pa, 10 Pa, 30 Pa, 100 Pa, 300 Pa, and 1000 Pa

CCM.P-K4 and EURAMET.M.P-K4.2010 for nominal pressures 3 Pa, 10 Pa, and 30 Pa

CCM.P-K4, EUROMET.M.P-K1.a, EUROMET.M.P-K1.b and SIM-EUROMET.M.P-BK3 at a pressure close to 1 Pa

Key comparison SIM-EUROMET.M.P-BK3

MEASURAND: Pressure

NOMINAL VALUE: 0.9 Pa

d relative difference between the pressures p_{CENAM} and p_{PTB} : $d = [p_{\text{CENAM}} / p_{\text{PTB}}] - 1$

U expanded uncertainty $k = 2$ of d

(values taken from Table 11 of the SIM-EUROMET.M.P-BK3 Final Report)

Date of measurement: April 2002 to June 2002

Nominal pressure	d	U
0.9 Pa	-0.0018	0.0041

Key comparison EUROMET.M.P-K4.2010

MEASURAND: Pressure

NOMINAL VALUES: 1 Pa, 3 Pa, 10 Pa, 30 Pa, 100 Pa, 300 Pa, 1000 Pa, 3000 Pa, 10000 Pa, and 15000 Pa

The laboratory individual measurements of the participants in the key comparison EURAMET.M.P-K4.2010 are given in Table 10.7a on page 23 of the EURAMET.M.P-K4.2010 Final Report.

CCM.P-K4 and EUROMET.M.P-K1.a for nominal pressures 1 Pa, 3 Pa, 10 Pa, 30 Pa, 100 Pa, 300 Pa, and 1000 Pa

CCM.P-K4 and EURAMET.M.P-K4.2010 for nominal pressures 3 Pa, 10 Pa, and 30 Pa

CCM.P-K4, EUROMET.M.P-K1.a, EUROMET.M.P-K1.b and SIM-EUROMET.M.P-BK3 at a pressure close to 1 Pa

Key comparison CCM.P-K4

NOMINAL VALUES: 1 Pa, 3 Pa, 10 Pa, 30 Pa, 100 Pa, 300 Pa and 1000 Pa

For each nominal pressure value, the key comparison reference value, p_R , is obtained from an unweighted mean of the measurement method means. Outliers at pressures up to and including 100 Pa are excluded.

A correction is applied to set p_R numerically equal to the nominal pressure (see CCM.P-K4 Final Report, page 41).

The combined standard uncertainty of p_R is u_R (see CCM.P-K4 Final Report, page 42).

Nominal pressure / Pa	1	3	10	30	100	300	1000
p_R / Pa	1	3	10	30	100	300	1000
u_R / Pa	0.0008	0.0019	0.0066	0.015	0.038	0.11	0.37

The degree of equivalence of each laboratory with respect to the reference value is given by a pair of terms:

$D_i = (p_i - p_R)$ and U_i , its expanded uncertainty at a 95% level of confidence, both expressed in Pa.

The degree of equivalence between two laboratories is given by a pair of terms:

$D_{ij} = D_i - D_j = (p_i - p_j)$ and U_{ij} , its expanded uncertainty at a 95% level of confidence, both expressed in Pa.

Linking EUROMET.M.P-K1.a to CCM.P-K4

The results of EUROMET.M.P-K1.a are linked to those of CCM.P-K4 for 7 pressure nominal values in the range 1 Pa to 1000 Pa, through the common participation of three laboratories: CNR-IMGC, PTB, and NPL. The CNR-IMGC, however, does not participate in the linking process in the pressure range 100 Pa to 1000 Pa, because it did not use a Static Expansion System for that range of measurements in CCM.P-K4. The detailed calculation used in the linking process is presented on pages 20 and 21 of the EUROMET.M.P-K1.a Final Report.

The degrees of equivalence of laboratories having participated in EUROMET.M.P-K1.a only (plus those of CNR-IMGC in the range 100 Pa to 1000 Pa) relative to the CCM.P-K4 key comparison reference values are obtained from the linking process and are given in Table 11 on page 22 of the EUROMET.M.P-K1.a Final Report.

No pair-wise degrees of equivalence are deduced from the measurements carried out in the framework of key comparison EUROMET.M.P-K1.a.

Key comparison EUROMET.M.P-K1.b

NOMINAL PRESSURE: 0.9 Pa

The key comparison reference value, p_{R-EUR} , is computed as the weighted average of the results of the BNM-LNE, CNR-IMGC, NPL, PTB and UME. The weights are inversely proportional to the square of the individual uncertainties u_i .

The standard uncertainty, u_{R-EUR} , of p_{R-EUR} is the standard uncertainty of the weighted average.

$p_{R-EUR} = 9.000E-01$ Pa and $u_{R-EUR} = 1.1E-03$ Pa.

The EUROMET degree of equivalence of each laboratory participant in EUROMET.M.P.K1. b with respect to the reference value p_{R-EUR} is given by a pair of terms: $D_{i-EUR} = [(p_i / p_{R-EUR}) - 1]$ and U_{i-EUR} , its expanded uncertainty ($k = 2$) computed according to Equations 29 and 30 on page 29 of the EUROMET.M.P-K1.b Final report. D_{i-EUR} and U_{i-EUR} are dimensionless.

	D_{i-EUR}	U_{i-EUR}
NPL	0.0005	0.0052
BNM-LNE	0.0000	0.011
MIRS/IMT	0.0065	0.0063
CNR-IMGC	0.0018	0.0044
CEM	0.0023	0.014
UME	-0.0013	0.0051
PTB	-0.0006	0.0029

No pair-wise degrees of equivalence are computed for the key comparison EUROMET.M.P-K1.b.

Linking EUROMET.M.P-K1.b to CCM.P-K4

Taking into account the nominal pressure ranges involved in the key comparisons CCM.P-K4 and EUROMET.M.P-K1.b, the linkage between both can be computed only for a pressure close to 1 Pa.

For a nominal pressure close to 1 Pa, three laboratories, CNR-IMGC, NPL and PTB, are common participants to both comparisons. Examination of their performance in both comparisons leads to the conclusions (see Section 12 of the EUROMET.M.P-K1.b Final Report) that the CCM.P-K4 key comparison reference value and the EUROMET reference value are in agreement. The EUROMET degrees of equivalence relative to the reference value may thus be transferred to CCM.P-K4.

The CCM.P-K4 degrees of equivalence, D_i and U_i , are expressed in Pa, while the EUROMET.M.P-K1.b ones, D_{i-EUR} and U_{i-EUR} , are expressed in relative terms. The transfer to CCM.P-K4 is thus obtained by multiplying by p_{R-EUR} the D_{i-EUR} and U_{i-EUR} values. It follows that at a nominal pressure of 1 Pa, $D_i = (D_{i-EUR} \times p_{R-EUR})$ and $U_i = (U_{i-EUR} \times p_{R-EUR})$ for participants taking part only in EUROMET.M.P-K1.b.

The CCM.P-K4 graph of equivalence can thus be extended to include the BNM-LNE, MIRS/IMT, CEM and UME results.

Linking SIM-EUROMET.M.P-BK3 to EUROMET.M.P-K1.b

The stability of the PTB primary standard between both comparisons was substantiated by the same check standard. It follows it is possible to link the CENAM value to the EUROMET.M.P-K1.b reference value p_{R-EUR} via the PTB primary standard.

By multiplying the ratio $r = (d + 1)$ with the pressure p_{PTB} obtained during EUROMET.M.P-K1.b, one obtains a fictive value for the pressure p_{CENAM} which can be compared with the EUROMET reference value p_{R-EUR} :
 $p_{CENAM} = r p_{PTB}$ with the uncertainty u_{CENAM} given in Equation 27 of the SIM-EUROMET.M.P-BK3 Final Report.

The degrees of equivalence of CENAM relative to the EUROMET.M.P-K1.b reference value is then computed from p_{CENAM} and u_{CENAM} , as if CENAM had participated in EUROMET.M.P-K1.b.

At a nominal pressure close to 1 Pa:

p_{CENAM} / Pa	u_{CENAM} / Pa	$D_{CENAM-EUR}$	$U_{CENAM-EUR}$
8.978E-01	1.063E-03	-0.0024	0.003

No pair-wise degrees of equivalence involving CENAM are computed.

Linking SIM-EUROMET.M.P-BK3 to CCM.P-K4

The conclusions of the linkage between EUROMET.M.P-K1.b and CCM.P-K4 for a nominal pressure of 1 Pa apply to the results of CENAM.

It follows that at a nominal pressure of 1 Pa, $D_{CENAM} = (D_{CENAM-EUR} \times p_{R-EUR})$ and $U_{CENAM} = (U_{CENAM-EUR} \times p_{R-EUR})$.
The CCM.P-K4 graph of equivalence can thus be extended to include the CENAM results.

Linking EURAMET.M.P-K4.2010 to CCM.P-K4

PTB provides the link between key comparisons CCM.P-K4 and EURAMET.M.P-K4.2010 for nominal pressures 3 Pa, 10 Pa, and 30 Pa. The linkage process is described in Section 11 of the EURAMET.M.P-K4.2010 Final Report.

CCM.P-K4, EUROMET.M.P-K1.a, EUROMET.M.P-K1.b and SIM-EUROMET.M.P-BK3

MEASURAND: Pressure

NOMINAL PRESSURE: 1 Pa

Lab *i*
↓

Lab *j* →

Participants in CCM.P-K4

	D_i	U_i / Pa
CSIRO-NML	0.211	0.046
CNR-IMGC	0.0023	0.0054
NIST	0.0003	0.0015
NPLI	-0.0048	0.0023
NPL	0.0013	0.0054
PTB	0.0002	0.0029

CSIRO-NML		CNR-IMGC		NIST		NPLI		NPL		PTB	
D_{ij}	U_{ij} / Pa	D_{ij}	U_{ij} / Pa	D_{ij}	U_{ij} / Pa	D_{ij}	U_{ij} / Pa	D_{ij}	U_{ij} / Pa	D_{ij}	U_{ij} / Pa
		0.209	0.046	0.211	0.046	0.216	0.046	0.210	0.046	0.211	0.046
-0.209	0.046			0.0021	0.0064	0.0071	0.0063	0.0011	0.0085	0.0022	0.0067
-0.211	0.046	-0.0021	0.0064			0.0051	0.0028	-0.0010	0.0063	0.0001	0.0035
-0.216	0.046	-0.0071	0.0063	-0.0051	0.0028			-0.0061	0.0063	-0.0050	0.0035
-0.210	0.046	-0.0011	0.0085	0.0010	0.0063	0.0061	0.0063			0.0011	0.0066
-0.211	0.046	-0.0022	0.0067	-0.0001	0.0035	0.0050	0.0035	-0.0011	0.0066		

Participants in EUROMET.M.P-K1.b

BNM-LNE	0.0000	0.010
MIRS/IMT	0.0059	0.0057
CEM	0.0021	0.013
UME	-0.0012	0.0046

No pair-wise degrees of equivalence are deduced from the measurements carried out in the framework of key comparison EUROMET.M.P-K1.b.

Participant in SIM-EUROMET.M.P-BK3

CENAM	-0.0022	0.003
-------	---------	-------

No pair-wise degrees of equivalence are deduced from the measurements carried out in the framework of key comparison SIM-EUROMET.M.P-BK3.

Participants in EUROMET.M.P-K1.a

BNM-LNE	-0.0056	0.0051
MIKES	0.039	0.044
SP	0.0028	0.038
CEM	0.0031	0.0090
OMH	0.00082	0.082
UME	0.0062	0.0060
NMI-VSL	0.0036	0.035

No pair-wise degrees of equivalence are deduced from the measurements carried out in the framework of key comparison EUROMET.M.P-K1.a.

CCM.P-K4, EUROMET.M.P-K1.a, and EURAMET.M.P-K4.2010

MEASURAND: Pressure

NOMINAL PRESSURE: 3 Pa

Lab <i>i</i>			Lab <i>j</i>									
	D_i	U_i	D_{ij}	U_{ij}	D_{ij}	U_{ij}	D_{ij}	U_{ij}	D_{ij}	U_{ij}	D_{ij}	U_{ij}
	/ Pa	/ Pa	/ Pa	/ Pa	/ Pa	/ Pa	/ Pa	/ Pa	/ Pa	/ Pa	/ Pa	/ Pa
CSIRO-NML	0.301	0.052			0.297	0.054	0.300	0.052	0.308	0.052	0.300	0.054
CNR-IMGC	0.004	0.016	-0.297	0.054			0.003	0.018	0.011	0.018	0.003	0.025
NIST	0.0008	0.0037	-0.300	0.052	-0.003	0.018			0.0079	0.0048	0.000	0.018
NPLI	-0.0071	0.0049	-0.308	0.052	-0.011	0.018	-0.0079	0.0048			-0.008	0.018
NPL	0.001	0.016	-0.300	0.054	-0.003	0.025	0.000	0.018	0.008	0.018		0.002
PTB	-0.0013	0.0075	-0.302	0.052	-0.006	0.019	-0.0022	0.0081	0.0057	0.0084	-0.002	0.019

BNM-LNE	-0.017	0.013
MIKES	0.041	0.052
SP	0.0056	0.068
CEM	0.0020	0.024
OMH	-0.093	0.23
UME	0.018	0.018
NMi-VSL	-0.033	0.036

No pair-wise degrees of equivalence are deduced from the measurements carried out in the framework of key comparison EUROMET.M.P-K1.a.

CMI	0.008	0.026
INRIM	0.010	0.026
LNE	0.003	0.026
MIKES	0.022	0.072

No pair-wise degrees of equivalence are deduced from the measurements carried out in the framework of key comparison EURAMET.M.P-K4.2010.

CCM.P-K4, EUROMET.M.P-K1.a, and EURAMET.M.P-K4.2010

NOMINAL PRESSURE: 10 Pa

Lab <i>i</i>	↓		Lab <i>j</i>	→										
	<i>D_i</i>	<i>U_i</i>		CSIRO-NML	CNR-IMGC	KRISS	NIST	NPLI	NPL	PTB				
	/ Pa	/ Pa		<i>D_{ij}</i>	<i>U_{ij}</i>									
CSIRO-NML	0.57	0.10			0.55	0.11	0.56	0.10	0.57	0.10	0.58	0.10	0.56	0.11
CNR-IMGC	0.015	0.053		-0.55	0.11		0.012	0.065	0.015	0.060	0.033	0.061	0.012	0.084
KRISS	0.003	0.023		-0.56	0.10	-0.012	0.065		0.003	0.027	0.021	0.029	0.000	0.065
NIST	0.000	0.013		-0.57	0.10	-0.015	0.060	-0.003	0.027		0.018	0.012	-0.003	0.059
NPLI	-0.018	0.016		-0.58	0.10	-0.033	0.061	-0.021	0.029	-0.018	0.012		-0.021	0.060
NPL	0.003	0.053		-0.56	0.11	-0.012	0.084	0.000	0.065	0.003	0.059	0.021	0.060	0.008
PTB	-0.005	0.025		-0.57	0.10	-0.020	0.064	-0.008	0.036	-0.005	0.025	0.013	0.027	-0.008
BNM-LNE	-0.040	0.041												
MIKES	0.043	0.082												
SP	-0.0080	0.13												
CEM	-0.0072	0.079												
OMH	-0.40	0.92												
UME	0.051	0.058												
NMi-VSL	0.0058	0.037												
CMI	-0.010	0.028												
INRIM	-0.003	0.028												
LNE	-0.010	0.028												
MIKES	0.001	0.073												

No pair-wise degrees of equivalence are deduced from the measurements carried out in the framework of key comparison EUROMET.M.P-K1.a.

No pair-wise degrees of equivalence are deduced from the measurements carried out in the framework of key comparison EURAMET.M.P-K4.2010.

CCM.P-K4, EUROMET.M.P-K1.a, and EURAMET.M.P-K4.2010

NOMINAL PRESSURE: 30 Pa

Lab <i>i</i>			Lab <i>j</i>									
	D_i	U_i	D_{ij}	U_{ij}	D_{ij}	U_{ij}	D_{ij}	U_{ij}	D_{ij}	U_{ij}	D_{ij}	U_{ij}
	/ Pa	/ Pa	/ Pa	/ Pa	/ Pa	/ Pa	/ Pa	/ Pa	/ Pa	/ Pa	/ Pa	/ Pa
CSIRO-NML	0.91	0.15			0.85	0.23	0.88	0.15	0.92	0.15	0.95	0.15
CNR-IMGC	0.06	0.16	-0.85	0.23			0.03	0.18	0.07	0.18	0.10	0.18
KRISS	0.022	0.035	-0.88	0.15	-0.03	0.18			0.032	0.028	0.069	0.044
NIST	-0.010	0.029	-0.92	0.15	-0.07	0.18	-0.032	0.028			0.069	0.044
NPLI	-0.047	0.041	-0.95	0.15	-0.10	0.18	-0.069	0.044	-0.037	0.035		
NPL	-0.03	0.10	-0.94	0.18	-0.09	0.21	-0.06	0.12	-0.02	0.11		
PTB	0.002	0.062	-0.90	0.16	-0.06	0.19	-0.020	0.070	0.012	0.064	0.049	0.072
BNM-LNE	-0.078	0.11										
MIKES	0.050	0.20										
SP	-0.055	0.32										
CEM	-0.037	0.20										
OMH	2.0	3.2										
UME	0.11	0.14										
NMi-VSL	0.064	0.042										
CMI	-0.016	0.028										
INRIM	-0.004	0.028										
LNE	-0.014	0.028										
MIKES	-0.007	0.074										

No pair-wise degrees of equivalence are deduced from the measurements carried out in the framework of key comparison EUROMET.M.P-K1.a.

No pair-wise degrees of equivalence are deduced from the measurements carried out in the framework of key comparison EURAMET.M.P-K4.2010.

CCM.P-K4 and EUROMET.M.P-K1.a

NOMINAL PRESSURE: 100 Pa

Lab <i>i</i>			Lab <i>j</i>													
	<i>D_i</i>	<i>U_i</i>	<i>D_{ij}</i>	<i>U_{ij}</i>												
	/ Pa	/ Pa	/ Pa	/ Pa	/ Pa	/ Pa	/ Pa	/ Pa	/ Pa	/ Pa	/ Pa	/ Pa	/ Pa	/ Pa		
CSIRO-NML	0.83	0.20			0.65	0.21	0.79	0.19	0.84	0.19	0.96	0.21	0.91	0.40	0.83	0.28
CNR-IMGC	0.18	0.11	-0.65	0.21			0.14	0.13	0.18	0.11	0.30	0.15	0.25	0.37	0.17	0.23
KRISS	0.036	0.081	-0.79	0.19	-0.14	0.13			0.045	0.043	0.16	0.11	0.11	0.36	0.03	0.21
NIST	-0.009	0.074	-0.84	0.19	-0.18	0.11	-0.045	0.043			0.12	0.11	0.07	0.36	-0.01	0.21
NPLI	-0.13	0.11	-0.96	0.21	-0.30	0.15	-0.16	0.11	-0.12	0.11			-0.05	0.37	-0.13	0.23
NPL	-0.08	0.30	-0.91	0.40	-0.25	0.37	-0.11	0.36	-0.07	0.36	0.05	0.37			-0.08	0.42
PTB	0.00	0.19	-0.83	0.28	-0.17	0.23	-0.03	0.21	0.01	0.21	0.13	0.23	0.08	0.42		
BNM-LNE	-0.17	0.38														
MIKES	0.077	0.70														
SP	-0.15	0.96														
CEM	-0.15	0.63														
OMH	2.5	3.1														
UME	0.32	0.47														
NMi-VSL	0.096	0.069														
CNR-IMGC	-0.080	0.40														

No pair-wise degrees of equivalence are deduced from the measurements carried out in the framework of key comparison EUROMET.M.P-K1.a.

CCM.P-K4 and EUROMET.M.P-K1.a

NOMINAL PRESSURE: 300 Pa

Lab <i>i</i>			Lab <i>j</i>									
	D_i	U_i	D_{ij}	U_{ij}	D_{ij}	U_{ij}	D_{ij}	U_{ij}	D_{ij}	U_{ij}	D_{ij}	U_{ij}
	/ Pa	/ Pa	/ Pa	/ Pa	/ Pa	/ Pa	/ Pa	/ Pa	/ Pa	/ Pa	/ Pa	/ Pa
CSIRO-NML	0.57	0.34			0.44	0.33	0.56	0.31	0.58	0.31	0.81	0.44
CNR-IMGC	0.13	0.24	-0.44	0.33			0.12	0.17	0.15	0.13	0.38	0.33
KRISS	0.01	0.22	-0.56	0.31	-0.12	0.17			0.02	0.05	0.25	0.31
NIST	-0.01	0.22	-0.58	0.31	-0.15	0.13	-0.02	0.05			0.23	0.31
NPLI	-0.24	0.33	-0.81	0.44	-0.38	0.33	-0.25	0.31	-0.23	0.31		
NPL	-0.20	0.90	-0.8	1.1	-0.3	1.1	-0.2	1.1	-0.2	1.1	0.0	1.1
PTB	-0.08	0.56	-0.65	0.70	-0.21	0.64	-0.09	0.63	-0.06	0.63	0.17	0.70

BNM-LNE	-0.30	0.57
MIKES	-0.10	1.3
SP	-0.30	1.3
CEM	-0.30	1.3
CNR-IMGC	0.13	0.61

No pair-wise degrees of equivalence are deduced from the measurements carried out in the framework of key comparison EUROMET.M.P-K1.a.


NOMINAL PRESSURE: 1000 Pa

Lab <i>i</i>			Lab <i>j</i>									
	D_i	U_i	D_{ij}	U_{ij}	D_{ij}	U_{ij}	D_{ij}	U_{ij}	D_{ij}	U_{ij}	D_{ij}	U_{ij}
	/ Pa	/ Pa	/ Pa	/ Pa	/ Pa	/ Pa	/ Pa	/ Pa	/ Pa	/ Pa	/ Pa	/ Pa
CSIRO-NML	0.33	0.78			0.15	0.34	0.28	0.33	0.30	0.32	1.2	1.5
CNR-IMGC	0.18	0.74	-0.15	0.34			0.12	0.18	0.15	0.13	1.1	1.5
KRISS	0.06	0.74	-0.28	0.33	-0.12	0.18			0.03	0.11	0.9	1.5
NIST	0.03	0.73	-0.30	0.32	-0.15	0.13	-0.03	0.11			0.9	1.5
NPLI	-0.9	1.4	-1.2	1.5	-1.1	1.5	-0.9	1.5	-0.9	1.5		
NPL	0.7	3.0	0.4	3.5	0.5	3.5	0.6	3.5	0.7	3.5	1.6	3.8
PTB	-0.3	1.9	-0.6	2.2	-0.4	2.2	-0.3	2.2	-0.3	2.2	0.6	2.6


BNM-LNE	-0.38	1.9
MIKES	0.32	4.5
SP	0.51	2.3
CEM	-0.38	4.2
CNR-IMGC	0.38	2.0

No pair-wise degrees of equivalence are deduced from the measurements carried out in the framework of key comparison EUROMET.M.P-K1.a.


CCM.P-K4, EUROMET.M.P-K1.b, SIM-EUROMET.M.P.BK3, and EUROMET.M.P-K1.a, pressure ~ 1 Pa
Degrees of equivalence [D_i and U_i ($k = 2$)]




CCM.P-K4, EUROMET.M.P-K1.a, and EURAMET.M.P-K4.2010, nominal pressure 3 Pa
Degrees of equivalence [D_i and U_i ($k = 2$)]




CCM.P-K4, EUROMET.M.P-K1.a, and EURAMET.M.P-K4.2010, nominal pressure 10 Pa
Degrees of equivalence [D_i and U_i ($k = 2$)]




CCM.P-K4, EUROMET.M.P-K1.a, and EURAMET.M.P-K4.2010, nominal pressure 30 Pa
Degrees of equivalence [D_i and U_i ($k = 2$)]



CCM.P-K4 and EUROMET.M.P-K1.a, nominal pressure 100 Pa
Degrees of equivalence [D_i and $U_i(k=2)$]



CCM.P-K4 and EUROMET.M.P-K1.a, nominal pressure 300 Pa
Degrees of equivalence [D_i and U_i ($k = 2$)]



CCM.P-K4 and EUROMET.M.P-K1.a, nominal pressure 1000 Pa
Degrees of equivalence [D_i and U_i ($k = 2$)]

