

Calibrations between AOS and OP

d are differential time corrections to be added to [$UTC(AOS) - UTC(OP)$],
and $u(d)$ are estimated uncertainties for the periods of comparisons.

| Date | <i>d</i> /ns | <i>u</i> (<i>d</i>)/ns | Reference | Note |
|---------------|--------------|--------------------------|-------------------------------------|------|
| November 2009 | -0.6 | 3.0 | BIPM Report-2010/04 | (1) |

The uncertainties given in this table are conservative. They are mainly driven by the uncertainty due to the ‘round-trip’ reproducibility at the OP.

Notes

- (1) TTS-3 receiver (s/n: 0002) at AOS