

## Calibrations between SU and OP

$d$  are differential time corrections to be added to  $[UTC(SU) - UTC(OP)]$ ,  
and  $u(d)$  are estimated uncertainties for the periods of comparisons.

Date	$d/ns$	$u(d)/ns$	Reference	Note
June 2009	203.7	5.0	<a href="#">BIPM Report-2010/04</a>	(1)
June 2009	198.0	5.0	<a href="#">BIPM Report-2010/04</a>	(2)
June 2009	198.4	5.0	<a href="#">BIPM Report-2010/04</a>	(3)
June 2009	198.4	5.0	<a href="#">BIPM Report-2010/04</a>	(4)
June 2009	197.7	5.0	<a href="#">BIPM Report-2010/04</a>	(5)

The uncertainties given in this table are conservative. They are mainly driven by the uncertainty due to the ‘round-trip’ reproducibility at the OP.

### Notes

- (1) TTS-3 receiver (s/n: 0026) at SU
- (2) TTS-3 receiver (s/n: 0030) at SU
- (3) TTS-3 receiver (s/n: 0031) at SU
- (4) TTS-3 receiver (s/n: 0032) at SU
- (5) TTS-3 receiver (s/n: 0033) at SU