

Calibrations between APL and OP

d are differential time corrections to be added to [$UTC(APL) - UTC(OP)$],
and $u(d)$ are estimated uncertainties for the periods of comparisons.

Date	d/ns	$u(d)/ns$	Reference	Note
December 2003	-6.5	3.0	BIPM Report-2004/06	(1)

The uncertainties given in this table are conservative. They are mainly driven by the uncertainty due to the ‘round-trip’ reproducibility at the OP.

Notes

- (1) GPSCV receiver (s/n: TFS112) at APL