

Calibrations between AUS and OP

d are differential time corrections to be added to [*UTC*(AUS) – *UTC*(OP)],
and *u*(*d*) are estimated uncertainties for the periods of comparisons.

Date	<i>d</i> /ns	<i>u</i> (<i>d</i>)/ns	Reference	Note
September 1996	38.0	10.0	Rapport BIPM-1997/01	(1)
September 1996	56.0	10.0	Rapport BIPM-1997/01	(2)
November 2004	8.9	3.0	Rapport BIPM-2008/02	(3)

The uncertainties given in this table are conservative. They are mainly driven by the uncertainty due to the ‘round-trip’ reproducibility at the OP.

Notes

- (1) Trimble 5000 receiver (s/n: 2903A00102) at AUS
- (2) FTS 8400 receiver (s/n: 4058) at AUS
- (3) Topcon Euro-80 receiver (s/n: 8RQRFKXT534) at AUS