

Calibrations between BEV and OP

d are differential time corrections to be added to [$UTC(\text{BEV}) - UTC(\text{OP})$],
and $u(d)$ are estimated uncertainties for the periods of comparisons.

| Date | d/ns | $u(d)/\text{ns}$ | Reference | Note |
|--------------|---------------|------------------|-------------------------------------|------|
| October 2006 | -1.7 | 3.0 | BIPM Report-2010/02 | (1) |

The uncertainties given in this table are conservative. They are mainly driven by the uncertainty due to the ‘round-trip’ reproducibility at the OP.

Notes

- (1) TTS-2 receiver (s/n: 024) at BEV