

Calibrations between BIM (*) and OP

d are differential time corrections to be added to [$UTC(\text{BIM}) - UTC(\text{OP})$],
and $u(d)$ are estimated uncertainties for the periods of comparisons.

Date	d/ns	$u(d)/\text{ns}$	Reference	Note
November 2006	283.7	3.0	BIPM Report-2010/02	(1)

The uncertainties given in this table are conservative. They are mainly driven by the uncertainty due to the ‘round-trip’ reproducibility at the OP.

Notes

(1) TTR-6 receiver (s/n: 467) at BIM

(*) BIM: formerly NMC.