

Calibrations between CH and OP

d are differential time corrections to be added to $[UTC(CH) - UTC(OP)]$,
and $u(d)$ are estimated uncertainties for the periods of comparisons.

Date	d/ns	$u(d)/ns$	Reference	Note
October 1997	2.0	3.0	Rapport BIPM-1998/01	(1)
November/December 2003	32.4	3.0	Rapport BIPM-2004/08	(1)

The uncertainties given in this table are conservative. They are mainly driven by the uncertainty due to the 'round-trip' reproducibility at the OP.

Notes

- (1) TTR-5a receiver (s/n: 275) at CH