

## Calibrations between CNMP and OP

*d* are differential time corrections to be added to [ $UTC(\text{CNMP}) - UTC(\text{OP})$ ],  
and  $u(d)$  are estimated uncertainties for the periods of comparisons.

Date	<i>d</i> /ns	$u(d)$ /ns	Reference	Note
May 2005	0.4	4.0	<a href="#">BIPM Report-2008/03</a>	(1)

The uncertainties given in this table are conservative. They are mainly driven by the uncertainty due to the ‘round-trip’ reproducibility at the OP.

### Notes

- (1) TTS-2 receiver (s/n: 029) at CNMP