

Calibrations between IGNA (*) and OP

d are differential time corrections to be added to [$UTC(IGNA) - UTC(OP)$],
and $u(d)$ are estimated uncertainties for the periods of comparisons.

Date	<i>d</i> /ns	<i>u(d)</i> /ns	Reference	Note
August 2004	-5.8	4.0	BIPM Report-2008/03	(1)

The uncertainties given in this table are conservative. They are mainly driven by the uncertainty due to the ‘round-trip’ reproducibility at the OP.

Notes

(1) TTS-2 receiver at IGMA

(*) IGNA: formerly IGMA.