

Calibrations between NICT (*) and OP

d are differential time corrections to be added to $[UTC(NICT) - UTC(OP)]$,
and $u(d)$ are estimated uncertainties for the periods of comparisons.

Date	d/ns	$u(d)/ns$	Reference	Note
February 1992	-2.0	5.0	Rapport BIPM-1993/04	(1)
November 1996	-2.0	3.0	Rapport BIPM-1997/01	(1)
August 2002	-14.0	4.0	BIPM Report-2003/05	(2)
August 2002	-15.0	4.0	BIPM Report-2003/05	(3)
November 2003	4.8	3.0	BIPM Report-2004/06	(4)
January 2005	-13.2	3.0	BIPM Report-2008/02	(4)

The uncertainties given in this table are conservative. They are mainly driven by the uncertainty due to the 'round-trip' reproducibility at the OP.

Notes

- (1) TTR-5A receiver (s/n: 184) at CRL
 - (2) TTR-6 receiver (s/n: 451) at CRL
 - (3) R100/40T receiver (s/n: 0017) at CRL
 - (4) Euro-80 receiver (s/n: 8PN45EETDKW) at NICT
- (*) NICT: formerly CRL.