

Calibrations between NML and OP

d are differential time corrections to be added to $[UTC(NML) - UTC(OP)]$,
and $u(d)$ are estimated uncertainties for the periods of comparisons.

Date	d/ns	$u(d)/ns$	Reference	Note
September 1996	27.0	3.0	Rapport BIPM-1997/01	(1)
October 2002	9.0	4.0	BIPM Report-2003/05	(2)

The uncertainties given in this table are conservative. They are mainly driven by the uncertainty due to the 'round-trip' reproducibility at the OP.

Notes

- (1) TTR6 receiver (s/n: 267) at NML
- (2) Topcon Euro-80 receiver (s/n: 8RQRFKXT534) at NML