

Calibrations between SU and OP

d are differential time corrections to be added to $[UTC(SU) - UTC(OP)]$,
and $u(d)$ are estimated uncertainties for the periods of comparisons.

Date	d/ns	$u(d)/ns$	Reference	Note
June 2009	-2.2	3.0	BIPM Report-2010/04	(1)
June 2009	-10.2	3.0	BIPM Report-2010/04	(2)
June 2009	-3.2	3.0	BIPM Report-2010/04	(3)
June 2009	-2.7	3.0	BIPM Report-2010/04	(4)
June 2009	-2.4	3.0	BIPM Report-2010/04	(5)
June 2009	-2.9	3.0	BIPM Report-2010/04	(6)

The uncertainties given in this table are conservative. They are mainly driven by the uncertainty due to the 'round-trip' reproducibility at the OP.

Notes

- (1) TTR-6 receiver (s/n: 414) at SU
- (2) TTS-3 receiver (s/n: 0026) at SU
- (3) TTS-3 receiver (s/n: 0030) at SU
- (4) TTS-3 receiver (s/n: 0031) at SU
- (5) TTS-3 receiver (s/n: 0032) at SU
- (6) TTS-3 receiver (s/n: 0033) at SU