

Calibrations between TCC and OP

d are differential time corrections to be added to [$UTC(TCC) - UTC(OP)$],
and $u(d)$ are estimated uncertainties for the periods of comparisons.

Date	d/ns	$u(d)/ns$	Reference	Note
May 2004	-4.5	4.0	BIPM Report-2008/03	(1)

The uncertainties given in this table are conservative. They are mainly driven by the uncertainty due to the ‘round-trip’ reproducibility at the OP.

Notes

- (1) TTR-6 receiver (s/n: 443) at TCC