

Calibrations between TP and OP

d are differential time corrections to be added to $[UTC(TP) - UTC(OP)]$,
and $u(d)$ are estimated uncertainties for the periods of comparisons.

Date	<i>d</i> /ns	<i>u</i> (<i>d</i>)/ns	Reference	Note
July 1992	-12.0	1.5	[1]	
October 2006	-15.6	3.0	BIPM Report-2010/02	(1)

The uncertainties given in this table are conservative. They are mainly driven by the uncertainty due to the 'round-trip' reproducibility at the OP.

References

[1] Lewandowski

Notes

(1) GTR-50 receiver (s/n: 002) at TP