

Calibrations between TUG and OP

d are differential time corrections to be added to $[UTC(TUG) - UTC(OP)]$,
and $u(d)$ are estimated uncertainties for the periods of comparisons.

Date	d/ns	$u(d)/ns$	Reference	Note
October 1986	3.4	2.0	[1]	(1)
April 1991	11.6	1.0	Rapport BIPM-1991/06	(2)
October 1994	2.0	2.0	Rapport BIPM-1994/12	(2)
July 1997	-10.0	3.0	Rapport BIPM-1997/05	(2)
November 1997	-5.0	2.0	Rapport BIPM-1998/01	(2)
March 1998	-6.0	2.0	Rapport BIPM-1998/07	(3)
June 1998	-3.0	3.0	[2]	(2)

The uncertainties given in this table are conservative. They are mainly driven by the uncertainty due to the 'round-trip' reproducibility at the OP.

References

- [1] Lewandowski, Weiss
[2] Lewandowski, Konate, 4th trip

Notes

- (1) STI receiver at TUG
(2) NBS receiver (sn/: 03) at TUG
(3) NBS receiver (sn/: 12) at TUG