

CMI  
Czech Republic

BIM, NCM  
Bulgaria

Supplementary comparison of  
measurements of current transformers  
**EURAMET.EM-S30**

Final report

EURAMET Project No 1081

Date: October 2009

Author(s): Emil Dimitrov  
Ginka Kumanova  
Renata Styblíková  
Karel Draxler  
Erik Dierikx

## Table of contents

1.	Introduction .....	3
2.	Participants and organisation of the comparison .....	3
2.1.	List of participants .....	3
2.2.	Comparison schedule .....	4
3.	Travelling standard and measurement instructions .....	5
3.1.	Description of the standard(s).....	5
3.2.	Quantities to be measured and conditions of measurement .....	7
3.3.	Measurement instructions.....	7
3.4.	Deviations from the protocol.....	7
4.	Methods of measurement .....	8
4.1.	Method of CMI .....	8
4.2.	Method of BIM .....	8
5.	Measurement results .....	9
5.1.	Results of the participating institutes .....	9
5.2.	Calculation of the reference value and its uncertainty .....	14
5.3.	Degrees of equivalence .....	14
6.	Summary and conclusions.....	20
Annex A.	Methods of measurement .....	21
Annex B.	Uncertainty budgets .....	23
Annex C.	Comparison protocol.....	27
Annex D.	Optional measurements.....	31

## 1. Introduction

A supplementary comparison was organized between CMI, Czech Republic and BIM, Bulgaria, in the field of current transformer ratio measurements. This comparison is registered as EURAMET project nr. 1081 and EURAMET.EM-S30 in the BIPM key comparison database.

The aim of the comparison was to demonstrate the improvement and extension of the calibration and measurement capabilities (CMC's) of BIM in this working field and to support the submission of an improved and extended CMC for BIM in Appendix C of the CIPM mutual recognition arrangement (MRA).

This comparison was organized in the framework of Phare project BG 2005/017-353.02.02, Lot 1, and was financed in this framework by the EU. This project run from March 2008 to the end of February 2009.

## 2. Participants and organisation of the comparison

### 2.1. List of participants

Participants:

Name: Czech Metrology Institute  
Acronym: CMI  
Address: Laboratory of Fundamental Metrology  
Department of Electromagnetic Quantities  
V Botanice 4  
150 72 Prague 5  
Country: Czech Republic  
Contact person: Mrs. Renata Styblíková  
Tel.: +420 257 288 335  
E-mail: rstyblikova@cmi.cz

Name: Bulgarian Institute of Metrology, National Centre of Metrology  
Acronym: BIM  
Address: General Directorate "National Centre of Metrology"  
Section "Electric Energy Measurements"  
52-B G.M. Dimitrov Blvd.  
1040 Sofia  
Country: Bulgaria  
Contact person: Mr. Emil Dimitrov  
Tel.: +359 29702792  
E-mail: e.dimitrov@bim.government.bg

Comparison coordinator:

Mr. E. Dierikx, VSL Dutch Metrology Institute, The Netherlands  
Tel.: +31 15 269 16 88  
E-mail: edierikx@nmi.nl

## 2.2. Comparison schedule

The travelling standards were circulated in a schedule A-B-A, as shown in the tables below. The behaviour of the travelling CMI transformers during the comparison was determined from the measurements of CMI. The behaviour of the travelling BIM transformer during the comparison was determined from the measurements of BIM.

In the original schedule, each participant was allowed two weeks to perform the measurements and to ship the standards to the other participant. However, it appeared that more time was required to complete the measurements. The actual schedules are given here below.

CMI transformers		
Participant	Measurements from	to
CMI	21-10-2008	18-11-2008
BIM	04-12-2008	06-01-2009
CMI	30-01-2009	09-02-2009

BIM transformer		
Participant	Measurements from	to
BIM	30-10-2008	14-11-2008
CMI	07-12-2008	21-01-2009
BIM	03-02-2009	03-02-2009

The transport of the travelling standards was arranged by door-to-door parcel service.

### 3. Travelling standard and measurement instructions

#### 3.1. Description of the standard(s)

Description, relevant technical data:


1. Tettex 4720 – for ratios (0.1; 1; 5 and 100) A/5 A; burden max. 5 VA; accuracy  $\pm 0.01\%$  and  $\pm 0.5$  min  
Serial number: 116 426  
Owner: CMI
2. CLA 2.2 – 500 A/5 A; class 0.2; burden 5 VA  
Serial number: 269954/08  
Owner: CMI
3. CLA 3.2 – 1000 A/5 A; class 0.1; burden 5 VA  
Serial number: 269955/08  
Owner: CMI
4. CLB 10 – 2000 A/5 A; class 0.1; burden 5 VA  
Serial number: 269956/08  
Owner: CMI
5. I 523 – 4000 A/5 A; class 0.05; nominal burden 15 VA  
Serial number: 18/1981  
Owner: BIM



Figure 1. CMI Tettex transformer



Figure 2. CMI Tettex transformer



**Figure 3.** CMI transformers  
(from left to right CLA 3.2, CLB 10 and CLA 2.2)



**Figure 5.** CMI transformers



**Figure 4.** BIM transformer: I 523

### 3.2. Quantities to be measured and conditions of measurement

The quantities to be measured are the current error and the phase displacement.

$\varepsilon_X$  Current error: The error which a transformer introduces into the measurement of a current and which arises from the fact that the actual transformation ratio is not equal to the rated transformation ratio. The current error expressed in ppm is given by the formula:

$$\varepsilon_X \text{ (ppm)} = 10^6 \cdot (K_n I_s - I_p) / I_p$$

$K_n$  Rated transformation ratio.

$I_p$  Actual primary current.

$I_s$  Actual secondary current when  $I_p$  is flowing under the conditions of measurement.

$I/I_n$  Excitation current, expressed in percent of rated current.

$\delta_X$  Phase displacement: The difference in phase, expressed in minutes, between the primary and secondary current vectors, the direction of the vectors being chosen in such a way that the angle is zero for a perfect transformer. The phase displacement is said to be positive when the secondary current vector leads the primary current vector.

The measurements were to be performed at a frequency of 50 Hz.

The measurements were to be performed at an ambient temperature of  $(23 \pm 1)$  °C.

The comparison was made by measuring the following ratios: primary (4000, 2000, 1000, 500, 100, 5, 1 and 0.5) A / secondary 5 A. All ratios had to be measured with a burden of 5 VA at unity power factor. The measurements were performed at 120 %, 100 %, 20 % and 5 % of the rated current. For ratios with primary current less than 5 A, the point at 5 % was replaced by 10 %.

### 3.3. Measurement instructions

The participants were asked to follow their usual measurement procedure corresponding to their best measurement capabilities.

### 3.4. Deviations from the protocol

- The protocol listed the ratio 0.1 A/ 5A as one of the points to be measured. However, CMI was not ready to perform this measurement. Therefore, it was decided to replace this measurement with 0.5 A/5 A.
- There was some deviation from the original schedule in the protocol. The actual schedule of measurements is given in section 2.2 of this report.

## **4. Methods of measurement**

### **4.1. Method of CMI**

Comparison against standard current comparator Tettex 4764 (for range of primary current 5 A up to 4 kA) or Tettex 4761 (for range of primary current 1 A and 0,5 A). Errors were measured by a commercial test set Tettex 2767; used burden Tettex 3691 (set value 5 VA real).

The reference values for both standards (current comparators) can be linked to the results of the international comparison EUROMET.EM-S11 (EUROMET project nr. 612).

### **4.2. Method of BIM**

Comparison against standard current transformer with errors measured on commercial test set Tettex 2767.

- For the the 5 A to 4000 A ratio, the BIM reference standard was an electronically compensated current comparator Tettex 4764.
- For the 0.5 A to 1 A ratio, the BIM reference standard was a passive current transformer MWB type NJ 0.5b.
- Both reference standards were calibrated at PTB in 2006.



## 5. Measurement results

### 5.1. Results of the participating institutes



Reported ambient conditions:

	Temperature °C	Relative Humidity %
BIM	$23.0 \pm 1.0$	$45 \pm 5$
CMI	$23.0 \pm 0.5$	$40 \pm 5$



Burden: 5 VA;  $\cos \phi = 1$

Frequency: 50 Hz



The reported results from both laboratories are given in the graphs and tables on the following pages.





$k$ (A/A)	$I/I_n$ (%)	CMI		CMI		BIM		BIM	
		$\varepsilon$ ( $\times 10^{-6}$ )	$U(\varepsilon)$ ( $\times 10^{-6}$ )	$\delta$ (minutes)	$U(\delta)$ (minutes)	$\varepsilon$ ( $\times 10^{-6}$ )	$U(\varepsilon)$ ( $\times 10^{-6}$ )	$\delta$ (minutes)	$U(\delta)$ (minutes)
0.5/5	120	57	21	-0.51	0.09	31	52	-0.19	0.27
0.5/5	100	56	21	-0.46	0.09	31	52	-0.16	0.27
0.5/5	20	71	20	0.08	0.14	52	52	0.41	0.27
0.5/5	10	88	22	0.31	0.17	63	52	0.60	0.27





$k$ (A/A)	$I/I_n$ (%)	CMI		CMI		BIM		BIM	
		$\varepsilon$ ( $\times 10^{-6}$ )	$U(\varepsilon)$ ( $\times 10^{-6}$ )	$\delta$ (minutes)	$U(\delta)$ (minutes)	$\varepsilon$ ( $\times 10^{-6}$ )	$U(\varepsilon)$ ( $\times 10^{-6}$ )	$\delta$ (minutes)	$U(\delta)$ (minutes)
1/5	120	65	20	-0.38	0.13	31	52	-0.24	0.27
1/5	100	63	20	-0.33	0.13	35	52	-0.22	0.27
1/5	20	77	19	0.20	0.17	61	52	0.34	0.27
1/5	10	95	22	0.43	0.16	73	52	0.42	0.27





$k$ (A/A)	$I/I_n$ (%)	CMI		CMI		BIM		BIM	
		$\varepsilon$ ( $\times 10^{-6}$ )	$U(\varepsilon)$ ( $\times 10^{-6}$ )	$\delta$ (minutes)	$U(\delta)$ (minutes)	$\varepsilon$ ( $\times 10^{-6}$ )	$U(\varepsilon)$ ( $\times 10^{-6}$ )	$\delta$ (minutes)	$U(\delta)$ (minutes)
5/5	120	59	30	-0.34	0.09	46	32	0.02	0.11
5/5	100	56	30	-0.29	0.09	56	32	-0.13	0.11
5/5	20	70	28	0.24	0.14	84	32	0.43	0.11
5/5	5	97	25	0.59	0.17	169	52	0.79	0.21





$k$ (A/A)	$I/I_n$ (%)	CMI		CMI		BIM		BIM	
		$\varepsilon$ ( $\times 10^{-6}$ )	$U(\varepsilon)$ ( $\times 10^{-6}$ )	$\delta$ (minutes)	$U(\delta)$ (minutes)	$\varepsilon$ ( $\times 10^{-6}$ )	$U(\varepsilon)$ ( $\times 10^{-6}$ )	$\delta$ (minutes)	$U(\delta)$ (minutes)
100/5	120	79	25	-0.39	0.08	74	32	-0.33	0.11
100/5	100	78	27	-0.35	0.08	75	32	-0.32	0.11
100/5	20	75	31	0.17	0.12	81	32	0.25	0.11
100/5	5	108	28	0.53	0.15	146	52	0.76	0.21





		CMI		CMI		BIM		BIM	
$k$ (A/A)	$I/I_n$ (%)	$\varepsilon$ ( $\times 10^{-6}$ )	$U(\varepsilon)$ ( $\times 10^{-6}$ )	$\delta$ (minutes)	$U(\delta)$ (minutes)	$\varepsilon$ ( $\times 10^{-6}$ )	$U(\varepsilon)$ ( $\times 10^{-6}$ )	$\delta$ (minutes)	$U(\delta)$ (minutes)
500/5	120	926	28	2.99	0.15	925	32	2.95	0.11
500/5	100	904	29	3.08	0.15	904	32	3.08	0.11
500/5	20	595	29	4.54	0.16	597	32	4.75	0.11
500/5	5	-79	60	7.99	0.21	-20	52	8.05	0.21



		CMI		CMI		BIM		BIM	
$k$ (A/A)	$I/I_n$ (%)	$\varepsilon$ ( $\times 10^{-6}$ )	$U(\varepsilon)$ ( $\times 10^{-6}$ )	$\delta$ (minutes)	$U(\delta)$ (minutes)	$\varepsilon$ ( $\times 10^{-6}$ )	$U(\varepsilon)$ ( $\times 10^{-6}$ )	$\delta$ (minutes)	$U(\delta)$ (minutes)
1 000/5	120	255	28	1.82	0.08	252	32	1.67	0.11
1 000/5	100	234	28	1.92	0.08	231	32	1.77	0.11
1 000/5	20	-72	31	3.18	0.25	-72	32	3.15	0.11
1 000/5	5	-730	41	6.37	0.74	-755	52	6.18	0.21



$k$ (A/A)	$I/I_n$ (%)	CMI		CMI		BIM		BIM	
		$\varepsilon$ ( $\times 10^{-6}$ )	$U(\varepsilon)$ ( $\times 10^{-6}$ )	$\delta$ (minutes)	$U(\delta)$ (minutes)	$\varepsilon$ ( $\times 10^{-6}$ )	$U(\varepsilon)$ ( $\times 10^{-6}$ )	$\delta$ (minutes)	$U(\delta)$ (minutes)
2 000/5	120	365	48	2.13	0.17	466	32	1.92	0.11
2 000/5	100	335	51	2.32	0.17	445	32	2.13	0.11
2 000/5	20	3	61	3.88	0.19	130	32	4.04	0.11
2 000/5	5	-654	70	6.21	0.22	-383	52	6.22	0.21



$k$ (A/A)	$I/I_n$ (%)	CMI		CMI		BIM		BIM	
		$\varepsilon$ ( $\times 10^{-6}$ )	$U(\varepsilon)$ ( $\times 10^{-6}$ )	$\delta$ (minutes)	$U(\delta)$ (minutes)	$\varepsilon$ ( $\times 10^{-6}$ )	$U(\varepsilon)$ ( $\times 10^{-6}$ )	$\delta$ (minutes)	$U(\delta)$ (minutes)
4 000/5	120	-93	17	0.50	0.08	-87	32	0.40	0.11
4 000/5	100	-92	17	0.53	0.08	-83	32	0.48	0.11
4 000/5	20	-73	18	0.81	0.08	-48	32	0.73	0.11
4 000/5	5	-66	22	0.97	0.12	-42	52	0.92	0.21

## 5.2. Calculation of the reference value and its uncertainty

For the current ratios in the range from 1/5 A/A to 1000/5 A/A, CMI participated in the supplementary comparison EUROMET.EM-S11 with 15 participants. To allow a comparison of the BIM results with the results of the participants in EUROMET.EM-S11, the BIM results are, thus, referred to the reference value of this comparison, as described in Sect. 5.3 below.

For the current ratios 0.5/5 A/A, 2000/5 A/A and 4000/5 A/A, there are no results available from previous comparisons. For this reason, only the degree of equivalence between the two laboratories are determined for these ratios.

## 5.3. Degrees of equivalence

The degree of equivalence ( $D_{ij}$ ) ( $i$  is for rated transformation ratio and  $j$  is excitation current) of BIM, NCM with respect to the reference values of the comparison EUROMET.EM-S11 is found from the difference between the BIM result and the CMI result, corrected for the difference of the CMI results to the reference values of EUROMET.EM-S11, for both the ratio error  $\varepsilon$  and the phase displacement  $\delta$ :

$$D_{ij}(\varepsilon) = \varepsilon_{BIM} - \varepsilon_{CMI} + \chi_{\varepsilon CMI}$$

$$D_{ij}(\delta) = \delta_{BIM} - \delta_{CMI} + \chi_{\delta CMI}$$

where

$\chi_{\varepsilon CMI}$  ( $\chi_{\delta CMI}$ ) is  $\varepsilon_{CMI}$  ( $\delta_{CMI}$ ) deviation from the comparison reference values of EUROMET.EM-S11.

The relevant uncertainties  $u(D_{ij})$  ( $k=1$ ) are given by:

$$u(D_{ij}(\varepsilon)) = \sqrt{u(\varepsilon_{BIM})^2 + u(\varepsilon_{CMI})^2 + u(\varepsilon_{CMI S11})^2}$$

$$u(D_{ij}(\delta)) = \sqrt{u(\delta_{BIM})^2 + u(\delta_{CMI})^2 + u(\delta_{CMI S11})^2}$$

where it is assumed that there is no significant correlation between the results of BIM and the results of CMI.

The expanded uncertainties  $U(D_{ij})$  ( $k=2$ ) are:

$$U(D_{ij}(\varepsilon)) = 2 \cdot u(D_{ij}(\varepsilon))$$



$$U(D_{ij}(\delta)) = 2 \cdot u(D_{ij}(\delta))$$

For the current ratios 0.5/5 A/A, 2000/5 A/A and 4000/5 A/A the degree of equivalence ( $D_{ij}$ ) are:



$$D_{ij} \left( \epsilon \right) = \epsilon_{BIM} - \epsilon_{CMI}$$

$$D_{ij} \left( \delta \right) = \delta_{BIM} - \delta_{CMI}.$$

The degrees of equivalence are given in the tables and graphs on the following pages.





$k$ (A/A)	$I/I_n$ (%)	$D(\varepsilon)$ ( $\times 10^{-6}$ )	$U(D(\varepsilon))$ ( $\times 10^{-6}$ )	$D(\delta)$ (minutes)	$U(D(\delta))$ (minutes)
0.5/5	120	-26	56	0.32	0.28
0.5/5	100	-25	56	0.30	0.29
0.5/5	20	-19	56	0.33	0.30
0.5/5	10	-25	56	0.29	0.32





$k$ (A/A)	$I/I_n$ (%)	$D(\varepsilon)$ ( $\times 10^{-6}$ )	$U(D(\varepsilon))$ ( $\times 10^{-6}$ )	$D(\delta)$ (minutes)	$U(D(\delta))$ (minutes)
1/5	120	-35	57	0.13	0.30
1/5	100	-29	57	0.11	0.30
1/5	20	-17	57	0.14	0.33
1/5	10	-23	61	-0.02	0.32









$k$ (A/A)	$I/I_n$ (%)	$D(\varepsilon)$ ( $\times 10^{-6}$ )	$U(D(\varepsilon))$ ( $\times 10^{-6}$ )	$D(\delta)$ (minutes)	$U(D(\delta))$ (minutes)
5/5	120	-14	45	0.37	0.15
5/5	100	0	45	0.17	0.15
5/5	20	14	44	0.19	0.19
5/5	5	72	60	0.20	0.28





$k$ (A/A)	$I/I_n$ (%)	$D(\varepsilon)$ ( $\times 10^{-6}$ )	$U(D(\varepsilon))$ ( $\times 10^{-6}$ )	$D(\delta)$ (minutes)	$U(D(\delta))$ (minutes)
100/5	120	-4	42	0.06	0.14
100/5	100	-3	44	0.04	0.14
100/5	20	7	47	0.09	0.18
100/5	5	38	61	0.24	0.28





$k$ (A/A)	$I/I_n$ (%)	$D(\varepsilon)$ ( $\times 10^{-6}$ )	$U(D(\varepsilon))$ ( $\times 10^{-6}$ )	$D(\delta)$ (minutes)	$U(D(\delta))$ (minutes)
500/5	120	3	44	-0.04	0.19
500/5	100	4	45	0.00	0.19
500/5	20	5	46	0.21	0.21
500/5	5	63	81	0.06	0.31



$k$ (A/A)	$I/I_n$ (%)	$D(\varepsilon)$ ( $\times 10^{-6}$ )	$U(D(\varepsilon))$ ( $\times 10^{-6}$ )	$D(\delta)$ (minutes)	$U(D(\delta))$ (minutes)
1 000/5	120	0	44	-0.14	0.14
1 000/5	100	0	44	-0.14	0.14
1 000/5	20	4	47	-0.03	0.28
1 000/5	5	-22	68	-0.19	0.77



$k$ (A/A)	$I/I_n$ (%)	$D(\varepsilon)$ ( $\times 10^{-6}$ )	$U(D(\varepsilon))$ ( $\times 10^{-6}$ )	$D(\delta)$ (minutes)	$U(D(\delta))$ (minutes)
2 000/5	120	101	58	-0.21	0.20
2 000/5	100	110	60	-0.19	0.20
2 000/5	20	127	69	0.16	0.22
2 000/5	5	271	88	0.01	0.30



$k$ (A/A)	$I/I_n$ (%)	$D(\varepsilon)$ ( $\times 10^{-6}$ )	$U(D(\varepsilon))$ ( $\times 10^{-6}$ )	$D(\delta)$ (minutes)	$U(D(\delta))$ (minutes)
4 000/5	120	6	36	-0.10	0.13
4 000/5	100	9	36	-0.05	0.13
4 000/5	20	25	36	-0.08	0.14
4 000/5	5	24	57	-0.05	0.24

## **6. Summary and conclusions**

For all of the current ratio error measurements, except for the measurements at 2 kA, there is a good agreement between BIM, NCM (Bulgaria) and CMI (Czech Republic). The cause of this significant discrepancy was investigated through additional measurements provided by CMI. The results of these measurements are given in Annex D.

For the current phase displacement measurements, the agreement on several measurement points is marginal or there are even small discrepancies. This should also be investigated in further detail.

## **7. References**

- [1] IEC 60044-1:2002 Edition 1.2 Instrument Transformers – Part 1: Current Transformers
- [2] BIPM, IEC, IFCC, ISO, IUPAP, OIML, Guide to the Expression of Uncertainty in Measurement, International Organization for Standardization, Geneva, First Edition 1993
- [3] EUROMET.EM-S11: EUROMET Projects 473 and 612 – Comparison of the measurement of current transformers (CTs). Final report.


# Annex A. Methods of measurement

## A.1 Detailed description of the BIM measurement set-up and method of measurement

The measurements were performed using an electronic measurement system. The transformer test set measures the transfer standard (marked as DUT) errors by dividing the differential current into two components. These values are evaluated automatically as current error (CE) and phase displacement (PE) by the test set electronics. The display "In %" on the test set indicates the percentage of the nominal ratio.

The burden used is an electronic current burden which compensates the ohmic resistance of the measurement circuit.


A schematic diagram is given below.



Schematic diagram of the BIM set-up

The power supply runs for 6 minutes and must then be switched off for 24 minutes. Within the 6 minutes, a series of 10 measurements is performed. This cycle of 30 minutes has been repeated 10 times for each ratio.

## A.2 Detailed description of the CMI measurement set-up and method of measurement




**Basic calibration layout in CMI**

In above shown connection a set of SEVERAL measurements was performed.

For the CLA, CLB and Tettex transformers:

3 sets in October - November, 2 sets in January - February

For the transfer standard Tettex 4720 each measurement was performed twice, at first at the position "1" and then at the position "2" (with commutated primary current) – see figure below. Results of these two measurements were averaged and this mean value was taken as a result of one measurement.



**Principle of primary current commutation by measurement of the Tettex 4720 transformer**

For the I523 transformer a set of 10 measurements was performed.

## Annex B. Uncertainty budgets

### B.1 Detailed uncertainty budget for BIM

Uncertainty Budget: Current error in percent 0,5A/5A 100%

Source of uncertainty	Distribution	Value	Type	$u(x_i)$	$\nu_i$	$c_i$	$u(y)$
1 repeatability	Normal	-0.00786	A	6.56E-05	9	1	6.56E-05
2 calibration of the reference standard	Normal	0.011	B	0.0025	50	1	0.0025
3 error in the bridge	Rectangular	0	B	0.000187	infinity	1	0.000187
4 drift of the reference standard	Rectangular	0	B	0.000289	infinity	1	0.000289
5 resolution of test set	Rectangular	0	B	2.89E-05	infinity	1	2.89E-05
6 error due to burden setting	Rectangular	0	B	5.77E-05	infinity	1	5.77E-05
7 error due to current setting	Rectangular	0	B	7.22E-05	infinity	1	7.22E-05
8 circuit and temperature influence	Rectangular	0	B	5.77E-05	infinity	1	5.77E-05
<b>y</b>	Normal	0.00314		0.002527	52.18759		
				$k =$		2.0491	
Result =		0.0031			$U =$		0.0052

Uncertainty Budget:Phase displacement in minutes 0,5A/5A 100%

Source of uncertainty	Distribution	Value	Type	$u(x_i)$	$\nu_i$	$c_i$	$u(y)$
1 repeatability	Normal	-0.3611	A	0.0002161	9	1	0.000216
2 calibration of the reference standard	Normal	0.2	B	0.1	50	1	0.1
3 error in the bridge	Rectangular	0	B	0.006399	infinity	1	0.006399
4 drift of the reference standard	Rectangular	0	B	0.0866	infinity	1	0.0866
5 resolution of test set	Rectangular	0	B	0.00289	infinity	1	0.00289
6 error due to burden setting	Rectangular	0	B	0.00204	infinity	1	0.00204
7 error due to current setting	Rectangular	0	B	0.002448	infinity	1	0.002448
8 circuit and temperature influence	Rectangular	0	B	0.00204	infinity	1	0.00204
<b>y</b>	Normal	-0.1611		0.1325263	infinity		
				$k =$		2.0000	
Result =		-0.16			$U =$		0.27

**Uncertainty Budget:**

**Current error in percent 1kA/5A 100%**

	Source of uncertainty	Distribution	Value	Type	$u(x_i)$	$\nu_i$	$c_i$	$u_i(y)$
1	repeatability	Normal	0.023101	A	3.82E-06	9	1	3.82E-06
2	calibration of the reference standard	Normal	0	B	0.0015	50	1	0.0015
3	error in the bridge	Rectangular	0	B	0.000213	infinity	1	0.000213
4	drift of the reference standard	Rectangular	0	B	0.00031	infinity	1	0.00031
5	resolution of test set	Rectangular	0	B	2.89E-05	infinity	1	2.89E-05
6	error due to burden setting	Rectangular	0	B	5.77E-05	infinity	1	5.77E-05
7	error due to current setting	Rectangular	0	B	7.22E-05	infinity	1	7.22E-05
8	circuit and temperature influence	Rectangular	0	B	5.77E-05	infinity	1	5.77E-05
<b>y</b>		Normal	0.023101		0.001551	57.08771		
						$k =$	2.0448	
			Result = 0.0231			$U =$	0.0032	

**Uncertainty Budget:Phase displacement in minutes 1kA/5A 100%**

	Source of uncertainty	Distribution	Value	Type	$u(x_i)$	$\nu_i$	$c_i$	$u_i(y)$
1	repeatability	Normal	1.77407	A	0.0002458	9	1	0.000246
2	calibration of the reference standard	Normal	0	B	0.05	50	1	0.05
3	error in the bridge	Rectangular	0	B	0.008846	infinity	1	0.008846
4	drift of the reference standard	Rectangular	0	B	0.0154	infinity	1	0.0154
5	resolution of test set	Rectangular	0	B	0.00289	infinity	1	0.00289
6	error due to burden setting	Rectangular	0	B	0.00204	infinity	1	0.00204
7	error due to current setting	Rectangular	0	B	0.002448	infinity	1	0.002448
8	circuit and temperature influence	Rectangular	0	B	0.00204	infinity	1	0.00204
<b>y</b>		Normal	1.77407		0.0532742	64.4403153		
						$k =$	2.0395	
			Result = 1.77			$U =$	0.11	



**Uncertainty Budget: Current error in percent 4kA/5A 100%**

	Source of uncertainty	Distribution	Value	Type	$u(x_i)$	$\nu_i$	$c_i$	$u_i(y)$
1	repeatability	Normal	-0.00827	A	8.37E-06	9	1	8.37E-06
2	calibration of the reference standard	Normal	0	B	0.0015	50	1	0.0015
3	error in the bridge	Rectangular	0	B	0.000188	infinity	1	0.000188
4	drift of the reference standard	Rectangular	0	B	0.00031	infinity	1	0.00031
5	resolution of test set	Rectangular	0	B	2.89E-05	infinity	1	2.89E-05
6	error due to burden setting	Rectangular	0	B	5.77E-05	infinity	1	5.77E-05
7	error due to current setting	Rectangular	0	B	7.22E-05	infinity	1	7.22E-05
8	circuit and temperature influence	Rectangular	0	B	5.77E-05	infinity	1	5.77E-05
<b>y</b>		Normal	-0.00827		0.001547	56.61524		
						$k =$	2.0451	
			Result = -0.0083			$U =$	0.0032	

**Uncertainty Budget:Phase displacement in minutes 4kA/5A 100%**

	Source of uncertainty	Distribution	Value	Type	$u(x_i)$	$\nu_i$	$c_i$	$u_i(y)$
1	repeatability	Normal	0.47874	A	0.0009608	9	1	0.000961
2	calibration of the reference standard	Normal	0	B	0.05	50	1	0.05
3	error in the bridge	Rectangular	0	B	0.006603	infinity	1	0.006603
4	drift of the reference standard	Rectangular	0	B	0.0154	infinity	1	0.0154
5	resolution of test set	Rectangular	0	B	0.00289	infinity	1	0.00289
6	error due to burden setting	Rectangular	0	B	0.00204	infinity	1	0.00204
7	error due to current setting	Rectangular	0	B	0.002448	infinity	1	0.002448
8	circuit and temperature influence	Rectangular	0	B	0.00204	infinity	1	0.00204
<b>y</b>		Normal	0.47874		0.0529561	62.9150174		
						$k =$	2.0405	
			Result = 0.48			$U =$	0.11	

## B.2 Detailed uncertainty budget for CMI

Uncertainty budget was calculated for current transformer I523, Ser. No. 18 1981, transformation ratio 4 kA/5 A

Transformation ratio: 4 kA/5 A

Burden: 5 VA real

Point of measurement:  $I_1 = 100 \% I_N$

Used standards: Current comparator Tettex 4764, Ser. No. 143 473

Errors:  $\pm 10$  ppm and  $\pm 0,05'$

Automatic transformer test set Tettex 2767, Ser. No. 141 727

Limits of errors:

Ratio error:  $\pm 0,5 \% \text{ RDG} \pm (0,5 \% \text{ FS or } 10 \text{ ppm})$  - greater of both values applies

Used range: 0,1999 %

Phase displacement:  $\pm 0,5 \% \text{ RDG} \pm (0,5 \% \text{ FS or } 0,034')$  - greater of both values applies

Used range: 1,999' and 19,99'

Programmable electronic current burden Tettex 3691, Ser. No. 141 474

### Ratio error uncertainty

Source of uncertainty	value (ppm)	type	probable distribution	divisor	$C_i$	$u_i$ (ppm)	$V_i$
standard current comparator	10	B	rectangular	$\sqrt{3}$	1	5.8	$\infty$
transformer test set	11	B	rectangular	$\sqrt{3}$	1	6.1	$\infty$
Burden	2	B	rectangular	$\sqrt{3}$	1	1.2	$\infty$
circuit configuration	2	B	rectangular	$\sqrt{3}$	1	1.2	$\infty$
repeatability	1	A	normal	1	1	0.9	9
combined uncertainty						8.6	
expanded uncertainty, $k = 2$				$U(\epsilon_i)$ (ppm)		17.2	

### Phase displacement uncertainty

Source of uncertainty	value ( ' )	type	probable distribution	divisor	$C_i$	$u_i$ ( ' )	$V_i$
standard current comparator	0.050	B	rectangular	$\sqrt{3}$	1	0.03	$\infty$
transformer test set	0.037	B	rectangular	$\sqrt{3}$	1	0.02	$\infty$
Burden	0.010	B	rectangular	$\sqrt{3}$	1	0.01	$\infty$
circuit configuration	0.010	B	rectangular	$\sqrt{3}$	1	0.01	$\infty$
repeatability	0.011	A	normal	1	1	0.01	9
combined uncertainty						0.04	
expanded uncertainty, $k = 2$				$U(\delta_i)$ ( ' )		0.08	

#### Explanatory notes:

- $C_i$  sensitivity coefficient
- $u_i$  standard uncertainty
- $V_i$  degrees of freedom

## **Annex C. Comparison protocol**

### **Comparison of the measurements of current transformers (CTs)**

#### **Introduction**

This comparison is organized in the framework of, and financed by:  
Phare project BG 2005/017-353.02.02, LOT 1.  
This project runs from March 2008 to the end of February 2009.

Current transformer measurements made by the participating NMIs support a large number of measurements made in the electrical generation, supply and distribution industries in their own countries. They also support many transformer manufacturers.

#### **Participants**

BIM - NCM, Sofia, Bulgaria

Address:

Bulgarian Institute of Metrology

General Directorate "National Centre of Metrology"

Section "Electric Energy Measurements"

52-B G.M. Dimitrov Blvd.

1040 Sofia

Emil Dimitrov

Tel. +359 29702792

E-mail: [e.dimitrov@bim.government.bg](mailto:e.dimitrov@bim.government.bg)

CMI, Czech Metrological Institute, Prague, Czech Republic

Address:

Czech Metrology Institute

Laboratory of Instrument Transformers

V Botanice 4

CZ-150 72 Prague 5

Renata Styblíková

Tel: +420 257 288 335

[rstyblikova@cmi.cz](mailto:rstyblikova@cmi.cz)

## Quantities to be measured

$\varepsilon_X$  Current error: The error which a transformer introduces into the measurement of a current and which arises from the fact that the actual transformation ratio is not equal to the rated transformation ratio. The current error expressed in ppm is given by the formula:

$$\text{Current error (ppm)} = 10^6 \cdot (K_n I_s - I_p) / I_p$$

$K_n$  Rated transformation ratio.

$I_p$  Actual primary current.

$I_s$  Actual secondary current when  $I_p$  is flowing under the conditions of measurement.

$I/I_n$  Excitation current, expressed in percent of rated current.

$\delta_X$  Phase displacement: The difference in phase, expressed in minutes, between the primary and secondary current vectors, the direction of the vectors being so chosen that the angle is zero for a perfect transformer. The phase displacement is said to be positive when the secondary current vector leads the primary current vector.

## Co-ordinator

The co-ordinator of this comparison in the Phare project is:

Erik Dierikx

NMI Van Swinden Laboratorium BV

Tel: +31 15 269 1688

[edierikx@nmi.nl](mailto:edierikx@nmi.nl)

## Support group of the comparison

The support group of this comparison consists of:

The contact persons from the participating laboratories and the co-ordinator of the comparison.

## Travelling standards

1. Tettex 4720 – for ratios (0.1; 1; 5 and 100) A/5 A; burden max. 5 VA; accuracy  $\pm 0.01$  % and  $\pm 0.5$  min  
Serial number: 116 426  
Owner: CMI
2. CLA 2.2 – 500 A/5 A; class 0.2; burden 5 VA  
Serial number: 269954/08  
Owner: CMI
3. CLA 3.2 – 1000 A/5 A; class 0.1; burden 5 VA  
Serial number: 269955/08  
Owner: CMI

4. CLB 10 – 2000 A/5 A; class 0.1; burden 5 VA  
 Serial number: 269956/08  
 Owner: CMI

The insurance value of CMI owned transformers is 1 500 EUR

5. I 523 – 4000 A/5 A; class 0.05; nominal burden 15 VA  
 Serial number: 18/1981  
 Insurance value: 1 000 EUR  
 Owner: NCM

### Transportation information

CMI and NCM will be responsible for arranging transportation to each other. CMI and NCM should inform each other when the travelling standards are sent.

Each laboratory is responsible for proper insurance of the standards during the stay in its laboratory and during transport to the other laboratory.

### Circulation scheme and timetable

Including time for transportation:

For the CMI transformers:

Laboratory	Period of measurements
CMI, Czech Republic	20 October 2008 to 10 November 2008
NCM, Bulgaria	10 November 2008 to 1 December 2008
CMI, Czech Republic	1 December 2008 to 7 December 2008

For the NCM transformer:

Laboratory	Period of measurements
NCM, Bulgaria	20 October 2008 to 10 November 2008
CMI, Czech Republic	10 November 2008 to 1 December 2008
NCM, Bulgaria	1 December 2008 to 7 December 2008

### Measurements

Measurements should be made at frequency 50 Hz.

Measurements should be made at  $(23 \pm 1) ^\circ\text{C}$ .

Comparison will be made by measuring following ratios: 4000, 2000, 1000, 500, 100, 5, 1 & 0.1/ secondary 5 A. All ratios should be measured with a burden 5 VA at unity power factor. Measurements will be performed with the following % of rated current: 120, 100, 20 and 5. For ratios with primary current less than 5 A the point 5 % should be replaced with 10 %.

## **Report**

The report should be sent to the co-ordinator as soon as possible (or at least within 6 weeks) after completing the measurements. In addition to the results, the report should contain a brief description of the measurement technique and all relevant conditions.

The report should contain an uncertainty budget. The uncertainty calculations should comply with the requirements of the GUM for the calculation of the uncertainty of measurement. The report should include the degrees of freedom and the complete budget of uncertainty.

The co-ordinator, as independent party, will collect all the measurement reports and will keep them confidential until all reports have been received. When all reports have been collected, they will be sent to NCM who will do the data analysis and who will prepare the final report. The co-ordinator will observe that the original data (including uncertainties) are not changed anymore.

,by the end of January 2009, the Report - Draft A, will be prepared. The Draft B Report has to be completed before the end of the project, which is 28 February 2009.

## **Computation of the reference value**

The reference values in this comparison will be the results from the linking laboratory. Therefore, for most of the measurements, CMI will be the linking laboratory. Only for the ratio 0.1 A/ 5 A, NCM will be the linking laboratory.

Emil Dimitrov  
September 2008

## **Annex A**

Uncertainty budget table

## **Annex B**

Summary of results form (Table)

## Annex D. Optional measurements

The CLB 10,  $k_I = 2 \text{ kA/5 A}$  at a burden 5 VA real is measured in November 2008 and in February 2009.

After result processing in VSL pilot lab CMI tried to explore a source of big result differences between CMI and BIM. It has emerged that values and measured errors depend strongly on primary cable position in the transformer opening. For this reason in March an error dependence on primary cable position has been explored (see Fig. 1, CMI primary conductor was formed by two round parallel cables).

Results of these measurements are given in Table 1. Results for the "position 0" are given partly by results measured in February and partly by supplementary results measured later in March.




Fig. 1

It can be assumed from results that by the measurements performed in CMI in November the primary conductor was placed probably in the top part of the transformer opening. The results of BIM (the last column in Table 1) correspond rather to the "position 1" – primary cables on the bottom of the transformer opening which was confirmed by BIM laboratory that CLB 10 measurements was performed with copper bus placed on the bottom of the transformer opening.

We assume that the error dependence on primary conductor position depends on small cross section of the transformer core (toroidal core dimensions:  $\varnothing = 123/108 \text{ mm}$ ,  $h = 15 \text{ mm}$ ).

I [%]	CMI			BIM
	position 0 $\varepsilon_I$ (ppm)	position 1 $\varepsilon_I$ (ppm)	position 2 $\varepsilon_I$ (ppm)	$\varepsilon_I$ (ppm)
120	397	467	355	466
100	369	441	319	445
20	45	188	-14	130
5	-602	-390	-645	-383

I [%]	$\delta_I$ (')	$\delta_I$ (')	$\delta_I$ (')	$\delta_I$ (')
120	2,23	2,25	1,70	1,92
100	2,42	2,45	1,88	2,13
20	3,97	3,93	3,37	4,04
5	6,35	6,15	5,45	6,22

Table 1.