

# International comparison CCQM-K74: Nitrogen dioxide, 10 $\mu\text{mol/mol}$

## Final Report

Edgar Flores<sup>\*1</sup>, Faraz Idrees<sup>1</sup>, Philippe Moussay<sup>1</sup>, Joële Viallon<sup>1</sup>, Robert Wielgosz<sup>1</sup>, Teresa Fernández<sup>2</sup>, Sergio Ramírez<sup>2</sup>, Andrés Rojo<sup>2</sup>, Uehara Shinji<sup>3</sup>, Jari Waldén<sup>4</sup>, Michela Segà<sup>5</sup>, Oh Sang-Hyub<sup>6</sup>, Tatiana Macé<sup>7</sup>, Cedric Couret<sup>8</sup>, HAN Qiao<sup>9</sup>, Damian Smeulders<sup>10</sup>, Franklin R. Guenther<sup>11</sup>, William J. Thorn III<sup>11</sup>, James Tshilongo<sup>12</sup>, Napo Godwill Ntsasa<sup>12</sup>, Viliam Štovicík<sup>13</sup>, Miroslava Valková<sup>13</sup>, Leonid Konopelko<sup>14</sup>, Elena Gromova<sup>14</sup>, Gerard Nieuwenkamp<sup>15</sup>, Rob. M. Wessel<sup>15</sup>, Martin Milton<sup>16</sup>, Alice Harling<sup>16</sup>, Gergely Vargha<sup>16</sup>, Dirk Tuma<sup>17</sup>, Anka Kohl<sup>17</sup> and Gert Schulz<sup>17</sup>.

<sup>1</sup>Bureau International des Poids et Mesures, Pavillon de Breteuil, F-92312 Sèvres Cedex, France.

<sup>2</sup>Centro Español de metrología (CEM), Calle Alfar, 2, 28760 Tres Cantos (Madrid), Spain.

<sup>3</sup>Chemicals Evaluation and Research Institute, Japan, 1600 Shimotakano, Sugito-machi, Kitakatsushika-gun, Saitama 345-0043, Japan.

<sup>4</sup>Finnish Meteorological Institute (MIKES-FMI), P.O. BOX 503, FI-00101 Helsinki, Finland.

<sup>5</sup>INRIM-Istituto Nazionale di Ricerca Metrologica, Strada delle Cacce 91, I-10135 Torino, Italy

<sup>6</sup>KRISS, 1 Doryong-Dong, Yuseong-Gu, Daejeon 305-340, Republic of Korea

<sup>7</sup>Laboratoire National de métrologie et d'Essais (LNE), 1, rue Gaston Boissier, 75 724 Paris Cedex 15, France

<sup>8</sup>Federal Office of Metrology METAS, Gas Analysis Laboratory, Lindenweg 50, 3003 Bern-Wabern, Switzerland

<sup>9</sup>National Institute of Metrology (NIM), China, No.18, Bei-San-Huan Dong Str., Beijing 100013, China.

<sup>10</sup>National Measurement Institute Australia (NMIA)

<sup>11</sup>National Institute of Standards and Technology (NIST), 100 Bureau Drive, Gaithersburg, MD 20899-8393, USA

<sup>12</sup>National Metrology Institute of South Africa (NMISA), CSIR, Building 4 West, Meiring Naude Road Brummeria, 0184, Pretoria, South Africa

<sup>13</sup>Slovak Institute of Metrology (SMU), Karloveská 63, SK-842 55 Bratislava, Slovak Republic

<sup>14</sup>D.I.Mendeleyev Institute for Metrology (VNIIM), 19 Moskovsky pr., St. Petersburg, 190005 Russia

<sup>15</sup>VSL, Thijssseweg 11 2629 JA Delft The Netherlands

<sup>16</sup>NPL, Hampton Road, Teddington, Middx, TW11 0LW, UK.

<sup>17</sup>Federal Institute for Materials Research and Testing (BAM), Germany.

### Coordinating laboratories:

Bureau International des Poids et Mesures (BIPM)  
VSL

### Study coordinator: Edgar Flores (BIPM)

Correspondence to be addressed to: Edgar Flores [edgar.flores@bipm.org](mailto:edgar.flores@bipm.org)  
(Tel: + 33 1 45 07 70 92)

**Field:** Amount of substance

**Organizing Body:** CCQM

## Index

|     |   |    |
|-----|---|----|
| 1.  | RATIONALE FOR COMPARISON  | 4  |
| 2.  | QUANTITIES AND UNITS  | 4  |
| 3.  | SCHEDULE  | 4  |
| 4.  | MEASUREMENT STANDARDS   | 4  |
| 5.  | REFERENCE VALUES FOR CYLINDERS  | 20 |
| 6.  | MEASUREMENT PROTOCOL  | 20 |
| 7.  | MEASUREMENT METHODS   | 21 |
| 8.  | RESULTS   | 21 |
| 9.  | CONCLUSION  | 23 |
| 10. | 'HOW FAR THE LIGHT SHINES' STATEMENT  | 23 |
|     | ANNEX 1- BIPM VALUE ASSIGNMENT PROCEDURE  | 27 |
| 1.  | Description of the facility   | 27 |
| 2.  | Measurement protocol of the BIPM  | 30 |
| 3.  | BIPM measurement uncertainties and analyser response  | 31 |
| 4.  | Covariance between two dynamically generated gas mixtures   | 34 |
| 5.  | The Key Comparison Reference Values and their standard uncertainties                                    | 35 |
| 6.  | FT-IR analysis of gas standards   | 35 |
| 6.1 | FT-IR Spectra acquisition procedure   | 35 |
| 6.2 | Quantitative analysis of nitric acid  | 36 |
| 7.  | Uncertainty budget  | 37 |
| 8.  | Regression analysis   | 37 |
| 9.  | Determination and validation of analysis functions  | 37 |
|     | ANNEX 2- REPORT OF PROPOSED $U(KCRV)$ FOR THE DRAFT B REPORT OF CCQM-K74: NITROGEN DIOXIDE, 10 MMOL/MOL | 38 |
| 1.  | Impurity analysis   | 38 |
| 2.  | Reaction of $NO_2$ and $HNO_3$ in the permeation facility   | 43 |
| 3.  | Cylinder stability/Sampling   | 45 |
| 4.  | Flow  | 45 |

|  |            |
|--|------------|
| <b>5. Modified BIPM uncertainty budget</b>                                     | <b>45</b>  |
| 5.1 Previous uncertainty budget:   | 45         |
| <b>5.2 Additional uncertainty components as a result of this study:</b>        | <b>47</b>  |
| <b>5.3 New uncertainty budget:</b>   | <b>48</b>  |
| <b>6. Covariance between two dynamically generated gas mixtures</b>            | <b>50</b>  |
| <b>7. Determination of the calibration function</b>                            | <b>51</b>  |
| <b>8. The Key Comparison Reference Values and their standard uncertainties</b> | <b>51</b>  |
| <b>ANNEX 3 - MEASUREMENT REPORTS OF PARTICIPANTS</b>                           | <b>57</b>  |
| <b>Centro Español de metrología (CEM)</b>                                      | <b>57</b>  |
| <b>Finnish Meteorological Institute (MIKES-FMI)</b>                            | <b>59</b>  |
| <b>Istituto Nazionale di Ricerca Metrologica (INRIM)</b>                       | <b>62</b>  |
| <b>Korea Research Institute of Standards and Science (KRISS)</b>               | <b>68</b>  |
| <b>Laboratoire National de métrologie et d'Essais (LNE)</b>                    | <b>70</b>  |
| <b>Federal Office of Metrology (METAS)</b>                                     | <b>75</b>  |
| <b>National Measurement Institute Australia (NMIA)</b>                         | <b>84</b>  |
| <b>National Institute of Metrology (NIM)</b>                                   | <b>86</b>  |
| <b>National Institute of Standards and Technology (NIST)</b>                   | <b>90</b>  |
| <b>National Physical Laboratory (NPL)</b>                                      | <b>92</b>  |
| <b>Chemicals Evaluation and Research Institute (CERI)</b>                      | <b>96</b>  |
| <b>National Metrology Institute of South Africa (NMISA)</b>                    | <b>99</b>  |
| <b>Slovak Institute of Metrology (SMU)</b>                                     | <b>101</b> |
| <b>Mendeleyev Institute for Metrology (VNIIM)</b>                              | <b>105</b> |
| <b>Federal Institute for Materials Research and Testing (BAM)</b>              | <b>110</b> |
| <b>Dutch Metrology Institute (VSL)</b>   | <b>112</b> |
| <b>Bureau International des Poids et Mesures (BIPM)</b>                        | <b>115</b> |

## 1. Rationale for comparison

There is a high international priority attached to activities which reduce NO<sub>x</sub> in the atmosphere. The current level of permitted emissions is typically between 50 µmol/mol and 100 µmol/mol, but lower values are expected in the future. Currently, ambient air quality monitoring regulations also require the measurement of NO<sub>x</sub> mole fractions of 0.2 µmol/mol. The production of accurate standards at these mole fractions requires either dilution of a stable higher concentration gas standard or production by a dynamic technique, for example, one based on permeation tubes.

The CCQM-K74 comparison is designed to evaluate the level of comparability of laboratories' measurement capabilities for nitrogen dioxide (NO<sub>2</sub>) at a nominal mole fraction of 10 µmol/mol.

## 2. Quantities and Units

In this protocol the measurand was the mole fraction of nitrogen dioxide in nitrogen\*, with measurement results being expressed in mol/mol and its multiples µmol/mol or nmol/mol.

(\*the nitrogen balance gas contains nominally 1000 µmol/mol of oxygen)

## 3. Schedule

The revised schedule for the project was as follows:

|                             |  |
|-----------------------------|--|
| June 2009                   | Shipment of cylinders to the BIPM                        |
| June 2009 – August 2009     | Analysis of mixtures at the BIPM                         |
| September 2009              | Shipment of cylinders from the BIPM to participants      |
| October 2009 – January 2010 | Analysis of mixtures by the participants                 |
| February 2010               | Shipment of cylinders back from participants to the BIPM |
| March 2010 – May 2010       | 2nd set of analysis of mixtures at the BIPM              |
| February 2010 – May 2010    | Reports of the participants                              |
| July 2010                   | Distribution of Draft A of this report                   |

## 4. Measurement standards

The gas mixtures circulated as part of the comparison were prepared by VSL. The nitrogen dioxide gas mixtures were contained in passivated aluminium cylinders of 5 L. The cylinders were pressurized at about 12 MPa.

The nitrogen dioxide gas standards were produced by gravimetric preparation in accordance with the International Standard ISO 6142:2001<sup>1</sup>.

Each cylinder was value assigned by the BIPM dynamic gas facility as described in ANNEX 1, before and after the participant's measurements. The VSL and BIPM values and measurements are given in Table 1 and Table 2 where:

---

<sup>1</sup> ISO 6142:2001: Gas analysis-Preparation of calibration gas mixtures-Gravimetric method.

|                                   |   |
|-----------------------------------|---|
| $x_{\text{VSL}}$                  | is the value assigned by VSL based on gravimetric preparation;  |
| $u_{\text{prep}}(x_{\text{VSL}})$ | the standard uncertainty of the VSL values with contributions due to gravimetry and purity analysis;                                  |
| $u_{\text{ver}}(x_{\text{VSL}})$  | the standard uncertainty including contributions from verification associated with the assigned value $x_{\text{VSL}}$ ;              |
| $x_{\text{KCRV}}$                 | The Key Comparison Reference Value (KCRV) that is the first BIPM measurement result (prior to sending out cylinders to participants); |
| $u(x_{\text{KCRV}})$              | the standard uncertainty of the KCRV;   |
| $x_{\text{BIPM2}}$                | the second BIPM measurement result (on return of cylinders from participants);  |
| $u(x_{\text{BIPM2}})$             | the standard uncertainty of the second BIPM measurement result;   |

From previous studies carried out by the BIPM and VSL it was expected that the mixtures would contain certain amounts of  $\text{HNO}_3$ . Analysis of the gas mixtures at the BIPM using FT-IR spectroscopy confirmed the presence of and permitted the quantification of nitric acid in the gas mixtures. Table 4 lists the nitric acid mole fractions found in the gas standards. To verify the stability of the gas mixtures the purity analysis was repeated when the gas mixtures were returned to the BIPM provided that the participants returned the cylinders with the minimum gas pressure required as described in the comparison protocol (see Table 3).

Table 4 lists:

|                          |   |
|--------------------------|---|
| Cylinder                 | the identification code of the cylinder received by the participating laboratory;   |
| $x_{\text{HNO}_3(1)}$    | the mole fraction of nitric acid measured in the standard by the BIPM (prior to sending standards to participants);                   |
| $u(x_{\text{HNO}_3(1)})$ | the standard uncertainty associated with the nitric acid mole fraction measurement;   |
| $x_{\text{HNO}_3(2)}$    | the mole fraction of nitric acid measured in the standard by the BIPM (following return of standards to the BIPM);                    |
| $u(x_{\text{HNO}_3(2)})$ | is the standard uncertainty associated with the nitric acid determination by FT-IR spectroscopy after the participant's measurements. |

Figure 1 plots the nitrogen dioxide mole fraction reported by VSL for each gas standard. In Figure 1 the error bars represent the standard uncertainty associated with the certified value. The average of nitrogen dioxide mole fractions calculated from VSL values based on gravimetric preparation data was  $10.60 \mu\text{mol/mol}$  with a standard uncertainty of  $105 \text{ nmol/mol}$ .

The nitrogen dioxide mole fractions measured by the BIPM before and after measurements by participants are shown in Figure 2 and Figure 3. In Figure 2 the error bars represent the standard uncertainty associated with the BIPM measurement results including contributions from the dynamic preparation of nitrogen dioxide gas mixtures, NO<sub>2</sub> losses in the permeation system of the BIPM and an observed drift in the nitrogen dioxide mole fractions measured by the BIPM before and after the participant's measurements. For further information see ANNEX 2- Report of Proposed  $u(KCRV)$  for the Draft B report of CCQM-K74: Nitrogen dioxide, 10 µmol/mol. The error bars in Figure 3 represent the standard uncertainty associated with the BIPM measurement results including the contributions from the dynamic preparation of nitrogen dioxide gas mixtures.

The average of the BIPM measurement results for all cylinders was 10.334 µmol/mol with a standard deviation of 72 nmol/mol covering all measured values. The amount of nitric acid found in each cylinder was consistent with the difference between the gravimetric preparation value and BIPM's analytical value for the nitrogen dioxide amount fraction, and accounts for the conversion of nitrogen dioxide to nitric acid (reacting with residual water and oxygen in the gas standards) and limited by the amount of water present. Figure 5 and Figure 6 plot the nitric acid mole fractions measured in each gas standard. The amount of nitric acid was measured before and after the measurements of the participating laboratories. Changes in the mole fractions of nitrogen dioxide and nitric acid in each cylinder during the period of the comparison were well within the measurement uncertainties of these values. The uncertainty budget for the key comparison reference value contains a component which covers any change in value due to instability of the gas transfer standard.

The BIPM was unable to perform a second measurement of nitric acid content in gas mixtures 930659, 930649 and 930654, as the participating laboratories that had made measurements on these cylinders had not followed the comparison protocol and returned the cylinders with insufficient gas to make these measurements. Cylinders 930650 and 930722 were not returned to the BIPM on time and no additional measurements could be made on these cylinders.

## VSL preparation values

| Certificate number | Preparation date | Number of Cylinder | Assigned NO <sub>2</sub> mole fraction<br>$x_{\text{VSL}}$<br>( $\mu\text{mol/mol}$ ) | Gravimetric standard uncertainty<br>$u_{\text{prep}}(x_{\text{VSL}})$<br>( $\mu\text{mol/mol}$ ) | Certified standard uncertainty<br>$u_{\text{ver}}(x_{\text{VSL}})$<br>( $\mu\text{mol/mol}$ ) |
|--------------------|------------------|--------------------|---|--|---|
| 3221115-02         | 24/02/2009       | #930659-PRM        | 10.604  | 0.003  | 0.105   |
| 3221115-05         | 18/03/2009       | #930650-PRM        | 10.617  | 0.002  | 0.105   |
| 3221115-22         | 09/04/2009       | #930655-PRM        | 10.608  | 0.003  | 0.105   |
| 3221115-23         | 09/04/2009       | #930662-PRM        | 10.606  | 0.003  | 0.105   |
| 3221115-21         | 08/04/2009       | #930649-PRM        | 10.609  | 0.003  | 0.105   |
| 3221115-16         | 03/04/2009       | #930671-PRM        | 10.597  | 0.003  | 0.105   |
| 3221115-06         | 19/03/2009       | #930660-PRM        | 10.614  | 0.002  | 0.105   |
| 3221115-07         | 19/03/2009       | #930667-PRM        | 10.603  | 0.002  | 0.105   |
| 3221115-15         | 02/04/2009       | #930661-PRM        | 10.603  | 0.003  | 0.105   |
| 3221115-03         | 25/02/2009       | #930673-PRM        | 10.605  | 0.003  | 0.105   |
| 3221115-17         | 03/04/2009       | #930675-PRM        | 10.609  | 0.003  | 0.105   |
| 3221115-13         | 01/04/2009       | #930654-PRM        | 10.604  | 0.002  | 0.105   |
| 3221115-09         | 20/03/2009       | #930674-PRM        | 10.608  | 0.002  | 0.105   |
| 3221115-25         | 10/04/2009       | #930676-PRM        | 10.602  | 0.003  | 0.105   |
| 3221115-18         | 03/04/2009       | #930713-PRM        | 10.597  | 0.003  | 0.105   |
| 3221115-20         | 08/04/2009       | #930722-PRM        | 10.620  | 0.003  | 0.105   |
| 3221115-11         | 01/04/2009       | #930697-PRM        | 10.600  | 0.003  | 0.105   |

Table 1. Characteristics of gravimetric mixtures as provided by VSL.

## BIPM measurement results

| Number of Cylinder | Measurement date 1st measurement | 1st BIPM NO <sub>2</sub> mole fraction measurement $x_{KCRV}$<br>μmol/mol | Standard uncertainty $u(x_{KCRV})$<br>μmol/mol | Measurement date 2nd measurement | 2nd BIPM assigned NO <sub>2</sub> mole fraction $x_{BIPM2}$<br>μmol/mol | Standard uncertainty $u(x_{BIPM2})$<br>μmol/mol | $\Delta x = (x_{BIPM2} - x_{KCRV})$<br>μmol/mol | $u(\Delta x)$<br>μmol/mol | $2u(\Delta x)$<br>μmol/mol |
|--------------------|----------------------------------|---|--|----------------------------------|---|---|---|---------------------------|----------------------------|
| #930659-PRM        | 19/08/2009                       | 10.226  | 0.042  | *                                |   |   |   |                           |                            |
| #930650-PRM        | 20/08/2009                       | 10.227  | 0.042  | **                               |   |   |   |                           |                            |
| #930655-PRM        | 18/08/2009                       | 10.347  | 0.042  | 01/04/2010                       | 10.351  | 0.035   | 0.004   | 0.055                     | 0.109                      |
| #930662-PRM        | 18/08/2009                       | 10.378  | 0.042  | 30/03/2010                       | 10.353  | 0.035   | -0.025  | 0.054                     | 0.109                      |
| #930649-PRM        | 18/08/2009                       | 10.347  | 0.042  | *                                |   |   |   |                           |                            |
| #930671-PRM        | 21/08/2009                       | 10.351  | 0.041  | 30/03/2010                       | 10.323  | 0.035   | -0.028  | 0.054                     | 0.109                      |
| #930660-PRM        | 21/08/2009                       | 10.431  | 0.041  | 02/04/2010                       | 10.400  | 0.035   | -0.031  | 0.054                     | 0.109                      |
| #930667-PRM        | 01/09/2009                       | 10.183  | 0.042  | 07/04/2010                       | 10.151  | 0.035   | -0.032  | 0.054                     | 0.109                      |
| #930661-PRM        | 19/08/2009                       | 10.270  | 0.042  | 31/03/2010                       | 10.265  | 0.035   | -0.005  | 0.054                     | 0.109                      |
| #930673-PRM        | 26/08/2009                       | 10.417  | 0.041  | 31/03/2010                       | 10.401  | 0.035   | -0.016  | 0.054                     | 0.109                      |
| #930675-PRM        | 27/08/2009                       | 10.378  | 0.041  | 07/04/2010                       | 10.384  | 0.035   | 0.006   | 0.054                     | 0.109                      |
| #930654-PRM        | 25/08/2009                       | 10.299  | 0.041  | *                                |   |   |   |                           |                            |
| #930674-PRM        | 25/08/2009                       | 10.370  | 0.041  | 30/03/2010                       | 10.343  | 0.035   | -0.027  | 0.054                     | 0.109                      |
| #930676-PRM        | 28/08/2009                       | 10.435  | 0.042  | 31/03/2010                       | 10.421  | 0.035   | -0.014  | 0.055                     | 0.109                      |
| #930713-PRM        | 29/08/2009                       | 10.320  | 0.042  | 01/04/2010                       | 10.284  | 0.035   | -0.036  | 0.055                     | 0.110                      |
| #930722-PRM        | 28/08/2009                       | 10.350  | 0.042  | **                               |   |   |   |                           |                            |
| #930697-PRM        | 25/08/2009                       | 10.343  | 0.041  | 31/03/2010                       | 10.343  | 0.024   | 0.000   | 0.048                     | 0.096                      |

Table 2. Results of BIPM NO<sub>2</sub> mole fraction measurements. \* Insufficient gas for second measurement \*\* Standard unavailable for the second measurement.



## Departure and return pressure of the gas standards

| Lab   | Certification date | Number of Cylinder | Date of return | pressure on departure<br>Mpa | pressure on return<br>Mpa |
|-------|--------------------|--------------------|----------------|------------------------------|---------------------------|
| NPL   | 24/02/2009         | #930659-PRM        | 26/01/2010     | 9.0                          | 4.0*                      |
| NIM   | 18/03/2009         | #930650-PRM        | 16/04/2010**   | 10.0                         | 6.0                       |
| SMU   | 09/04/2009         | #930655-PRM        | 02/02/2010     | 10.0                         | 7.5                       |
| NMIA  | 09/04/2009         | #930662-PRM        | 26/02/2010     | 9.5                          | 7.0                       |
| NMISA | 08/04/2009         | #930649-PRM        | 24/02/2010     | 10.5                         | 2.5*                      |
| CERI  | 03/04/2009         | #930671-PRM        | 01/03/2010     | 10.0                         | 7.5                       |
| METAS | 19/03/2009         | #930660-PRM        | 02/02/2010     | 9.0                          | 7.8                       |
| INRIM | 19/03/2009         | #930667-PRM        | 02/02/2010     | 10.0                         | 6.3                       |
| KRISS | 02/04/2009         | #930661-PRM        | 26/02/2010     | 9.5                          | 5.5                       |
| FMI   | 25/02/2009         | #930673-PRM        | 16/02/2010     | 8.7                          | 7.0                       |
| LNE   | 03/04/2009         | #930675-PRM        | 02/02/2010     | 10.0                         | 8.0                       |
| NIST  | 01/04/2009         | #930654-PRM        | 02/02/2010     | 10.2                         | 4.0*                      |
| VSL   | 20/03/2009         | #930674-PRM        | 02/02/2010     | 10.2                         | 7.2                       |
| CEM   | 10/04/2009         | #930676-PRM        | 02/03/2010     | 10.0                         | 6.2                       |
| VNIIM | 03/04/2009         | #930713-PRM        | 16/03/2010     | 10.0                         | 7.2                       |
| BAM   | 08/04/2009         | #930722-PRM        | ***            | 9.8                          | **                        |
| BIPM  | 20/03/2009         | #930697-PRM        | In place       | 10.0                         | 5.0                       |

Table 3. Departure and return pressure of the gas standards after being measured by the participating laboratories.

\* Insufficient gas for 2nd series of BIPM measurements ( $\leq 5$ MPa). \*\* Standard returned to the BIPM behind schedule. \*\*\* Standard not yet returned to the BIPM.

BIPM HNO<sub>3</sub> Measurements

| Cylinder    | Measurement | $x_{\text{HNO}_3(1)}$   | $u(x_{\text{HNO}_3(1)})$ | Measurement | $x_{\text{HNO}_3(2)}$   | $u(x_{\text{HNO}_3(2)})$ | $\Delta x =$                                  |                     |                     |
|-------------|-------------|-------------------------|--------------------------|-------------|-------------------------|--------------------------|---|---------------------|---------------------|
|             | date        | ( $\mu\text{mol/mol}$ ) | ( $\mu\text{mol/mol}$ )  | date        | ( $\mu\text{mol/mol}$ ) | ( $\mu\text{mol/mol}$ )  | $(x_{\text{HNO}_3(2)} - x_{\text{HNO}_3(1)})$ | $u(\Delta x)$       | $2u(\Delta x)$      |
|             |             |                         |                          |             |                         |                          | $\mu\text{mol/mol}$                           | $\mu\text{mol/mol}$ | $\mu\text{mol/mol}$ |
| #930659-PRM | 12/08/2009  | 0.348                   | 0.027                    | *           |                         |                          |   |                     |                     |
| #930650-PRM | 10/08/2009  | 0.214                   | 0.023                    | 07/05/2010  | 0.240                   | 0.024                    | 0.026   | 0.033               | 0.066               |
| #930655-PRM | 30/07/2009  | 0.114                   | 0.021                    | 20/04/2010  | 0.133                   | 0.021                    | 0.019   | 0.030               | 0.059               |
| #930662-PRM | 28/07/2009  | 0.155                   | 0.022                    | 01/05/2010  | 0.140                   | 0.021                    | -0.015  | 0.030               | 0.061               |
| #930649-PRM | 11/08/2009  | 0.237                   | 0.024                    | *           |                         |                          |   |                     |                     |
| #930671-PRM | 06/08/2009  | 0.165                   | 0.022                    | 03/05/2010  | 0.184                   | 0.022                    | 0.019   | 0.031               | 0.062               |
| #930660-PRM | 11/08/2009  | 0.137                   | 0.021                    | 20/04/2010  | 0.152                   | 0.022                    | 0.014   | 0.030               | 0.060               |
| #930667-PRM | 05/08/2009  | 0.345                   | 0.027                    | 22/04/2010  | 0.363                   | 0.028                    | 0.018   | 0.039               | 0.077               |
| #930661-PRM | 30/07/2009  | 0.199                   | 0.023                    | 21/04/2010  | 0.240                   | 0.024                    | 0.041   | 0.033               | 0.065               |
| #930673-PRM | 04/08/2009  | 0.119                   | 0.021                    | 29/04/2010  | 0.114                   | 0.021                    | -0.005  | 0.030               | 0.059               |
| #930675-PRM | 07/08/2009  | 0.148                   | 0.021                    | 21/04/2010  | 0.144                   | 0.021                    | -0.004  | 0.030               | 0.061               |
| #930654-PRM | 11/08/2009  | 0.234                   | 0.023                    | *           |                         |                          |   |                     |                     |
| #930674-PRM | 06/08/2009  | 0.131                   | 0.021                    | 30/05/2010  | 0.124                   | 0.021                    | -0.007  | 0.030               | 0.060               |
| #930676-PRM | 11/08/2009  | 0.081                   | 0.020                    | 01/05/2010  | 0.099                   | 0.021                    | 0.018   | 0.029               | 0.058               |
| #930713-PRM | 12/08/2009  | 0.197                   | 0.022                    | 21/04/2010  | 0.184                   | 0.022                    | -0.013  | 0.032               | 0.063               |
| #930722-PRM | 31/07/2009  | 0.145                   | 0.021                    | **          |                         |                          |   |                     |                     |
| #930697-PRM | 28/07/2009  | 0.141                   | 0.021                    | 30/04/2010  | 0.172                   | 0.022                    | 0.031   | 0.031               | 0.061               |

Table 4. Nitric acid mole fraction measured in cylinder gas standards by the BIPM using FT-IR spectroscopy. \* Insufficient gas for second measurement. \*\* Standard returned to the BIPM behind schedule. \*\*\* Standard not yet returned to the BIPM.

| 1st BIPM<br>NO <sub>2</sub> mole<br>fraction<br>measurement<br>$x_{\text{KCRV}}$<br>$\mu\text{mol/mol}$ | Standard<br>uncertainty<br>$u(x_{\text{KCRV}})$<br>$\mu\text{mol/mol}$ | 2nd BIPM<br>NO <sub>2</sub> mole<br>fraction<br>measurement<br>$x_{\text{BIPM2}}$<br>$\mu\text{mol/mol}$ | Standard<br>uncertainty<br>$u(x_{\text{BIPM2}})$<br>$\mu\text{mol/mol}$ | $x_{\text{KCRV}}$<br>+ | $u(x_{\text{KCRV}} +$<br>$x_{\text{HNO3(1)}})$ | $2u$<br>( $x_{\text{KCRV}} +$<br>$x_{\text{HNO3(1)}})$ | $x_{\text{BIPM2}}$<br>+ | $u(x_{\text{BIPM2}} +$<br>$x_{\text{HNO3(2)}})$ | $2u$<br>( $x_{\text{BIPM2}} +$<br>$x_{\text{HNO3(2)}})$ |
|---|--|--|---|------------------------|--|--|-------------------------|---|---|
| 10.226  | 0.042  | *  |   | 10.574                 | 0.050  | 0.099  |                         |   |   |
| 10.227  | 0.042  | **   |   | 10.441                 | 0.047  | 0.095  |                         |   |   |
| 10.347  | 0.042  | 10.351   | 0.035   | 10.461                 | 0.046  | 0.093  | 10.484                  | 0.041   | 0.082   |
| 10.378  | 0.042  | 10.353   | 0.035   | 10.533                 | 0.047  | 0.094  | 10.493                  | 0.041   | 0.082   |
| 10.347  | 0.042  | *  |   | 10.584                 | 0.048  | 0.095  |                         |   |   |
| 10.351  | 0.041  | 10.323   | 0.035   | 10.516                 | 0.047  | 0.094  | 10.507                  | 0.042   | 0.083   |
| 10.431  | 0.041  | 10.400   | 0.035   | 10.568                 | 0.047  | 0.093  | 10.552                  | 0.041   | 0.083   |
| 10.183  | 0.042  | 10.151   | 0.035   | 10.528                 | 0.050  | 0.099  | 10.514                  | 0.045   | 0.089   |
| 10.270  | 0.042  | 10.265   | 0.035   | 10.469                 | 0.047  | 0.095  | 10.505                  | 0.042   | 0.085   |
| 10.417  | 0.041  | 10.401   | 0.035   | 10.536                 | 0.046  | 0.093  | 10.515                  | 0.041   | 0.082   |
| 10.378  | 0.041  | 10.384   | 0.035   | 10.526                 | 0.047  | 0.093  | 10.528                  | 0.041   | 0.082   |
| 10.299  | 0.041  | *  |   | 10.533                 | 0.048  | 0.095  |                         |   |   |
| 10.370  | 0.041  | 10.343   | 0.035   | 10.501                 | 0.047  | 0.093  | 10.467                  | 0.041   | 0.082   |
| 10.435  | 0.042  | 10.421   | 0.035   | 10.516                 | 0.047  | 0.093  | 10.520                  | 0.041   | 0.081   |
| 10.320  | 0.042  | 10.284   | 0.035   | 10.517                 | 0.048  | 0.096  | 10.468                  | 0.042   | 0.083   |
| 10.350  | 0.042  | **   |   | 10.495                 | 0.047  | 0.094  |                         |   |   |
| 10.343  | 0.041  | 10.343   | 0.024   | 10.484                 | 0.047  | 0.093  | 10.515                  | 0.032   | 0.065   |

Table 5. Summation of Nitrogen Dioxide and Nitric Acid Mole fractions for each standard based on BIPM measurements.

\* Insufficient gas for second measurement. \*\* The standard wasn't available for 2nd measurement




Figure 1. Nitrogen dioxide mole fraction values provided by VSL based on static gravimetric preparation. The error bar represents the standard uncertainty ( $k=1$ ) associated with the certified value.




Figure 2. First series of nitrogen dioxide mole fraction measurements by the BIPM ( $x_{KCRV}$ ), prior to sending standards to participating laboratories. The error bars represent the standard uncertainty ( $k=1$ ) associated with the BIPM measurement results including contributions from the dynamic preparation of nitrogen dioxide gas mixtures,  $\text{NO}_2$  losses in the permeation system of the BIPM and an observed drift in the nitrogen dioxide mole fractions measured by the BIPM before and after the participant's measurements. For further information see ANNEX 2- Report of Proposed  $u(KCRV)$  for the Draft B report of CCQM-K74: Nitrogen dioxide, 10  $\mu\text{mol/mol}$ .




Figure 3. Second series of nitrogen dioxide mole fraction measurements by the BIPM, after return of standards from participating laboratories. The error bars represent the standard uncertainty ( $k=1$ ) associated with the BIPM measurement results including the contributions from the dynamic preparation of nitrogen dioxide gas mixtures.




Figure 4. Difference between the BIPM series of measurements for each standard. The error bar represents the expanded uncertainty at a 95 % level of confidence.




Figure 5. First series of nitric acid mole fraction measurements by the BIPM, prior to sending standards to participating laboratories. The error bar represents the standard uncertainty ( $k=1$ ) associated with the FT-IR measurements.






Figure 6. Second series of nitric acid mole fraction measurements by the BIPM, after return of standards from participating laboratories. The error bar represents the standard uncertainty ( $k=1$ ) associated with the FT-IR measurements.




Figure 7. Difference between the first and second series of nitric acid mole fraction measurements by the BIPM. The error bar represents the expanded uncertainty at a 95 % level of confidence.




Figure 8. Summation of nitrogen dioxide and nitric acid mole fractions in each standard based on BIPM measurements. Red: First measurements. Black: Second measurements. The error bar represents the expanded uncertainty at a 95 % level of confidence.

## 5. Reference Values for Cylinders

During the 24th and 25th meetings of the CCQM GAWG it was agreed that the key comparison reference value for CCQM-K74 was to be based on BIPM measurement results prior to distribution of gas standards to participants. The BIPM's measurements clearly indicate the presence of nitric acid in the gas mixtures ranging from 100 nmol/mol to 350 nmol/mol. The gravimetric preparation values provided by VSL were not used as reference values for the comparison of nitrogen dioxide as these do not account for the presence of nitric acid in the standards, arising from the conversion of NO<sub>2</sub> to nitric acid through the reaction with oxygen and residual water in the cylinders. The current hypothesis is that the water must have been present on the cylinder coatings. In the current version of the report, laboratory results are compared to BIPM values, since they correctly account for the presence of nitric acid in the gas mixtures. Furthermore, the agreement between the summation of nitric acid and nitrogen dioxide mole fractions with the initial amount of nitrogen dioxide (prior to any reaction) expected from static gravimetric preparation values further confirms the hypothesis of the loss mechanism of NO<sub>2</sub> in the cylinders. For each cylinder, the Key Comparison Reference Value is the NO<sub>2</sub> mole fraction assigned by the BIPM (first measurement).

Following the CCQM GAWG guidance, it was decided that the standard uncertainty of the Key Comparison Reference Value (KCRV) was to be calculated from the following equation

$$u_{\text{KCRV}}(x_{\text{NO}_2}) = \sqrt{\left(u(x_{\text{BIPM}})\right)^2 + \left(u(x_{\text{NO}_2\text{Losses}})\right)^2 + \left(u(x_{\text{NO}_2\text{Drift}})\right)^2} \quad (1)$$

where  $u(x_{\text{BIPM}})$  is the uncertainty associated with the value assigned by the BIPM,  $u(x_{\text{NO}_2\text{Losses}})$  the uncertainty contribution due to NO<sub>2</sub> losses equivalent to 5.7 nmol/mol and  $u(x_{\text{NO}_2\text{Drift}})$  the uncertainty contribution due to observed drift of NO<sub>2</sub> estimated to be 21 nmol/mol. This leads to an overall standard uncertainty of the KCRV of 0.041 µmol/mol. A full discussion of the uncertainty of the KCRV is included in Annex 2 of this report.

## 6. Measurement protocol

The measurement protocol requested participants to provide the value and uncertainty of the nitrogen dioxide mole fraction measured by the laboratory, a complete uncertainty budget and a description of their gas analysis procedure. The procedure employed for the gas analysis was the responsibility of the participating laboratory. (See ANNEX 3 - Measurement reports of participants).

## 7. Measurement methods

The measurement methods and calibration methods used by the participating organizations in this comparison are listed in Table 6.

| Laboratory | Measurement method       | Calibration method                                   | Traceability               |
|------------|--------------------------|--|----------------------------|
| BIPM       | FT-IR, UV                | GLS, linear, Bracketing, Permeation-dynamic mixtures | own permeation tube system |
| NMIA       | FT-IR                    | Bracketing (8-12) $\mu\text{mol/mol}$                | own gravimetric standards  |
| BAM        | FT-IR                    | 2 mixtures   | own gravimetric standards  |
| NIM        | FT-IR, Chemiluminescence | 5 mixtures   | own gravimetric standards  |
| NPL        | UV                       | 1 mixtures   | own gravimetric standards  |
| VSL        | UV                       | 4 mixtures   | own gravimetric standards  |
| CEM        | Chemiluminescence        | GLS, linear, 3 mixtures                              | NPL                        |
| CERI       | Chemiluminescence        | Bracketing, 2 mixtures                               | own gravimetric standards  |
| FMI        | Chemiluminescence        | GPT, 1 mixture                                       | NPL                        |
| INRIM      | Chemiluminescence        | GLS, linear, 3 mixtures                              | own gravimetric standards  |
| KRISS      | Chemiluminescence        | 6 mixtures   | own gravimetric standards  |
| LNE        | Chemiluminescence        | Permeation-dynamic mixtures                          | own permeation tube system |
| METAS      | Chemiluminescence        | Permeation-dynamic mixtures                          | own permeation tube system |
| NIST       | Chemiluminescence        | 6 mixtures   | own gravimetric standards  |
| NMISA      | Chemiluminescence        | Bracketing (10-100) $\mu\text{mol/mol}$              | own gravimetric standards  |
| SMU        | Chemiluminescence        | 5 mixtures   | own gravimetric standards  |
| VNIIM      | Chemiluminescence        | 2 mixtures   | own gravimetric standards  |

Table 6. Measurement and calibration methods used by participating laboratories.

## 8. Results

The reported nitrogen dioxide mole fractions by participating laboratories are shown in Figure 9. The evaluation of the level of consistency between the participating laboratories was performed by comparison with the BIPM measurements (1st series, the  $x_{\text{KCRV}}$ ) for each cylinder. This decision was taken due to the systematic difference found between the values based on static gravimetric preparation (Table 1) and the BIPM measurements, caused by the presence of nitric acid in the cylinders (Table 2).

The permeation of nitric acid from  $\text{NO}_2$  permeation tubes was detected and quantified by the BIPM, and the BIPM's values are corrected to avoid systematic errors caused by this issue. This is fully described in ANNEX 1- BIPM Value assignment procedure.

The consistency between the participating laboratory's results and the Key Comparison Reference Value (the first BIPM value), is presented in terms of a degree of equivalence ( $D$ ) expressed quantitatively in two terms: its deviation from the KCRV value and the uncertainty of this deviation (at 95 % level of confidence).

The degree of equivalence is defined as:

$$D = x_{NMI} - x_{KCRV} \quad (2)$$

where  $x_{NMI}$  denotes the amount of substance fraction as measured by the participating National Metrology Institute (NMI) and  $x_{KCRV}$  the reference value given by the BIPM.

The uncertainty in the reference value,  $x_{KCRV}$ , is given by the following expression:

$$u_{KCRV}(x_{NO_2}) = \sqrt{u(x_{BIPM})^2 + u(x_{NO_2Losses})^2 + u(x_{NO_2Drift})^2} \quad (3)$$

where  $u(x_{BIPM})$  is the uncertainty associated with the value assigned by the BIPM following the procedure described in ANNEX 1- BIPM Value assignment procedure and  $u(x_{NO_2Losses})$  and  $u(x_{NO_2Drift})$  the uncertainty contributions due to NO<sub>2</sub> losses and the observed drift in NO<sub>2</sub> detailed in ANNEX 2- Report of Proposed  $u(KCRV)$  for the Draft B report of CCQM-K74: Nitrogen dioxide, 10 µmol/mol.

The combined standard uncertainty associated with the deviation from the KCRV can be expressed as:

$$u(D) = \sqrt{u_{NMI}^2 + u_{KCRV}^2} \quad (4)$$

and the expanded uncertainty, at 95 % confidence level

$$U(D) = k \cdot u(D) \quad (5)$$

where  $k$  denotes the coverage factor, taken as  $k = 2$  (normal distribution, approximately 95 % level of confidence).

The degrees of equivalence are listed in Table 7 where:

|               |  |
|---------------|--|
| Laboratory    | is the acronym of the participating national metrology institute;  |
| Cylinder      | the identification code of the cylinder received by the participating laboratory;                              |
| $x_{KCRV}$    | the assigned amount of substance fraction of a component by the BIPM (1st series of BIPM measurement results); |
| $u(x_{KCRV})$ | the uncertainty of the BIPM measurement result   |
| $x_{Lab}$     | the result as reported by the participating laboratory;  |

|                     |  |
|---------------------|--|
| $u(x_{\text{Lab}})$ | the standard uncertainty associated with the reported value $x_{\text{Lab}}$ ;   |
| $D$                 | degree of equivalence calculated as the difference in amount of substance fraction as measured by the laboratory and $x_{\text{BIPM}}$ the BIPM value; and |
| $U(D)$              | the expanded uncertainty of the degree of equivalence;   |

The BIPM's reported result is based on the second analysis of a cylinder (#930697-PRM) and compared to the first measurement made on this gas mixture.

The graph of equivalence, based on the difference in nitrogen dioxide reported values by participating laboratories and key comparison reference value, is plotted in Figure 10.

## 9. Conclusion

The results of the comparison indicate consistency of the majority of measurement results within limits of  $\pm 3\%$ . This can be compared to a relative standard uncertainty in the key comparison reference value of  $0.4\%$ . The results of only one laboratory lie significantly outside these limits. Its results were based on calibration with nitrogen monoxide and gas phase titration, rather than with statically or dynamically prepared nitrogen dioxide mixtures, as was the case for all other participants.

A full interpretation of the results of the comparison needs to take into account the presence of nitric acid (in the range  $100\text{ nmol/mol}$  to  $350\text{ nmol/mol}$ ) in the cylinders circulated as part of the comparison, as well as the possible presence of nitric acid in the primary standards used by participating laboratories. Furthermore, a number of measurement techniques may not differentiate between nitric acid and nitrogen dioxide. This can occur when using chemiluminescence instrumentation, in which the thermal conversion of nitrogen species to nitrogen monoxide will convert all reactive nitrogen species, nitrogen dioxide as well as nitric acid, into nitrogen monoxide.

The BIPM's FT-IR measurements and dynamic system for the generation of standard gas mixtures has allowed both nitrogen dioxide and nitric acid mole fractions to be quantified, and confirmed by UV absorption measurements, in both the transfer cylinders and permeation tube systems, and were used as the key comparison reference value in the comparison.

## 10. 'How far the light shines' statement

The following 'How far the light shines' statement is proposed:

The results of this key comparison can be used to support CMC claims for analytical capabilities for  $\text{NO}_2$  in nitrogen and synthetic air mixtures in the range  $10\text{ }\mu\text{mol/mol}$  to  $1000\text{ }\mu\text{mol/mol}$ .

| Laboratory | BIPM        |                   |                      | Participants     |                     |                                       |        |                     |
|------------|-------------|-------------------|----------------------|------------------|---------------------|---------------------------------------|--------|---------------------|
|            | Cylinder    | $x_{\text{KCRV}}$ | $u(x_{\text{KCRV}})$ | $x_{\text{Lab}}$ | $u(x_{\text{Lab}})$ | $D(x_{\text{Lab}} - x_{\text{KCRV}})$ | $u(D)$ | $U(D)$<br>( $k=2$ ) |
| NPL        | #930659-PRM | 10.226            | 0.042                | 10.331           | 0.040               | 0.105                                 | 0.058  | 0.115               |
| NIM        | #930650-PRM | 10.227            | 0.042                | 10.150           | 0.050               | -0.077                                | 0.065  | 0.130               |
| SMU        | #930655-PRM | 10.347            | 0.042                | 10.100           | 0.060               | -0.247                                | 0.073  | 0.146               |
| NMIA       | #930662-PRM | 10.378            | 0.042                | 10.740           | 0.315               | 0.362                                 | 0.318  | 0.635               |
| NMISA      | #930649-PRM | 10.347            | 0.042                | 10.690           | 0.185               | 0.343                                 | 0.190  | 0.379               |
| CERI       | #930671-PRM | 10.351            | 0.041                | 10.400           | 0.190               | 0.049                                 | 0.194  | 0.389               |
| METAS      | #930660-PRM | 10.431            | 0.041                | 10.630           | 0.080               | 0.199                                 | 0.090  | 0.180               |
| INRIM      | #930667-PRM | 10.183            | 0.042                | 9.990            | 0.100               | -0.193                                | 0.108  | 0.217               |
| KRISS      | #930661-PRM | 10.270            | 0.042                | 10.450           | 0.155               | 0.180                                 | 0.160  | 0.321               |
| FMI        | #930673-PRM | 10.417            | 0.041                | 9.880            | 0.150               | -0.537                                | 0.156  | 0.311               |
| LNE        | #930675-PRM | 10.378            | 0.041                | 10.260           | 0.065               | -0.118                                | 0.077  | 0.154               |
| NIST       | #930654-PRM | 10.299            | 0.041                | 10.280           | 0.050               | -0.019                                | 0.065  | 0.130               |
| VSL        | #930674-PRM | 10.370            | 0.041                | 10.510           | 0.105               | 0.140                                 | 0.113  | 0.226               |
| CEM        | #930676-PRM | 10.435            | 0.042                | 10.720           | 0.110               | 0.285                                 | 0.118  | 0.235               |
| VNIM       | #930713-PRM | 10.320            | 0.042                | 10.550           | 0.080               | 0.230                                 | 0.090  | 0.181               |
| BAM        | #930722-PRM | 10.350            | 0.042                | 10.530           | 0.375               | 0.180                                 | 0.377  | 0.755               |
| BIPM       | #930697-PRM | 10.343            | 0.041                | 10.343           | 0.024               | 0.000                                 | 0.048  | 0.096               |

Table 7. Laboratory results for nitrogen dioxide measurements ( $\mu\text{mol/mol}$ ).






Figure 9. Nitrogen dioxide mole fractions as reported by the participating laboratories. The error bar represents the standard uncertainties reported by participants.




Figure 10. Difference between participants' results and the KCRV values for nitrogen dioxide mole fractions. The error bar represents the expanded uncertainty at a 95 % level of confidence.

## ANNEX 1- BIPM Value assignment procedure

### 1. Description of the facility

The BIPM-NO<sub>2</sub> primary gas facility combines gravimetry with dynamic generation of gas mixtures. The facility includes a magnetic suspension balance, a flow control system for the dynamic generation of gas mixtures and a flow control system for nitrogen dioxide gas standards in cylinders. Both the gas cylinder and dynamic sources of NO<sub>2</sub> mixtures are ultimately connected to a continuous gas analyser ABB Limas 11 (AO2020), and to the spectrometer FT-IR Thermo-Nicolet Nexus (See *Figure 11*).

The operation and automation of the ensemble of instruments (NO<sub>2</sub> FT-IR facility-ABB Limas 11-FT-IR) is achieved through a LabView<sup>®</sup> programme developed by members of the BIPM Chemistry Department. Through a graphical user interface the programme facilitates the setting and monitoring of all relevant instrumental parameters, automated control of complex procedures, the recording of mass measurements and NO<sub>2</sub> analyser readings and related data and the graphical real-time display of many of the instrument readings.




Figure 11: Schematic of the BIPM NO<sub>2</sub> facility

### The magnetic suspension balance.

The magnetic suspension balance (MSB; Rubotherm, Germany) is central to the system. An electromagnet is suspended from the base of the weighing pan. Below this electromagnet there is a long vertical glass vessel; the measurement cell of the MSB. At

the top of the glass vessel there is a permanent magnet which is held in place by the electromagnet attached to the balance.




Figure 12: Schematic of the BIPM NO<sub>2</sub> facility permeation tube chamber and magnetic suspension balance.

The position of the permanent magnet is detected electronically and maintained by servo-control of the current of the electromagnet. An NO<sub>2</sub> permeation tube is suspended from the permanent magnet. Thus, the balance measures the mass of the permeation tube without being mechanically in contact with it, since the balance and the weighing load are separated by a layer of glass. The coupling between the permeation tube and the balance is purely magnetic and the sensitive balance is protected from the highly corrosive NO<sub>2</sub> gas and the occasionally elevated temperatures and gas flows surrounding the permeation tube. This facilitates continuous monitoring of the mass loss of the permeation tube, which is located in a temperature controlled environment by means of a double glass-walled jacket containing water circulating at a constant temperature, controlled by a remote thermostat. At constant temperature, the tube emits NO<sub>2</sub> through its permeable fluoropolymer membrane at a constant rate. The balance is a high-resolution comparator (model AT20, Mettler, USA) with a range of 0 g to 22 g and 2 µg resolution. The balance is configured with two mass pieces (see *Figure 12*) which are used to perform an external calibration of the balance. The term external calibration is used to distinguish it from the internal calibration of the balance performed with stainless steel mass standards. The two external calibration mass pieces have nominally

the same volume but different mass, as one is made of titanium (Ti) and the other of stainless steel (SS). Briefly, the external calibration mass pieces are used to correct for an effect on the mass measurements arising from changes in the density of the ambient atmosphere surrounding the balance itself. Since it was important to know the mass difference of the Ti and SS pieces with a small uncertainty, the mass and volume of the pieces were calibrated in collaboration with the BIPM Mass Department.

### **The flow control system for the magnetic suspension balance**

To generate primary mixtures using the MSB, a well characterized flow of NO<sub>2</sub>-free gas (nitrogen) is required. Once the flow control system receives a pre-selected gas it delivers two well characterized flows to the balance.

The total gas flow is characterized by means of a molbloc<sup>®</sup>/molbox<sup>®</sup> facility<sup>2</sup>, which was calibrated at the LNE. An electronic digital pressure controller is used to maintain the pressure of the incoming gas entering the molbloc at about 2700 hPa, the optimal pressure to minimize the uncertainty of the molbloc flow measurement (~0.1 %).

The gas flow is then introduced into a gas purifier that removes the remaining water and oxygen that may leak into the gas. The gas flow is then divided into two streams, a carrier and a diluent, both regulated by two mass flow controllers (MFCs).

The flow of the carrier stream is set at a constant value, 100 mL·min<sup>-1</sup>, mixing with the NO<sub>2</sub> emerging at constant rate from the permeation tube. The pressure conditions of the permeation chamber are controlled by an electronic digital pressure device to avoid any buoyancy variation.

The gas mixture of the carrier line is then diluted by a larger flow, the diluent stream, varied within the range 0.3 L·min<sup>-1</sup> to 5 L·min<sup>-1</sup> in order to dynamically generate primary NO<sub>2</sub> mixtures in nitrogen (or air) at various concentrations in the range 1 µmol·mol<sup>-1</sup> to 15 µmol·mol<sup>-1</sup>.

Permeation tubes with permeation rates in the range 5000 ng·min<sup>-1</sup> to 10 000 ng·min<sup>-1</sup> are used for this purpose.

### **The flow control system for NO<sub>2</sub> gas standards**

The third module, namely the flow control system for NO<sub>2</sub> gas standards, enables comparison between the dynamically generated gas mixtures and cylinder standards of NO<sub>2</sub> in nitrogen contained in high pressure cylinders (and, alternatively, comparison between various cylinder mixtures). This comparison is achieved via the response of the NO<sub>2</sub> analyser, whether ABB Limas 11 or FT-IR. The continuous gas analyser ABB Limas 11 (part of the AO2020 series) operates according to the NDUV (Non Dispersive Ultraviolet Absorption) measurement principle. The measuring effect is specific radiation absorption of the measured gas component in the UV spectra region to detect NO<sub>2</sub>. The FT-IR analyser is a Thermo Nicolet Nexus model enclosed in an isolation box, as described in Section 6.1.

---

<sup>2</sup> A molbox<sup>®</sup> facility is a support unit for making gas flow measurements using molbloc mass flow elements. The molbox<sup>®</sup> hardware reads calibration data off the molbloc<sup>®</sup> facility and measures molbloc<sup>®</sup> upstream and downstream pressure using built-in high precision Reference Pressure Transducers (RPTs). The key molbloc<sup>®</sup>L measurement is the differential pressure across the element, which is roughly proportional to the mass flow rate through it. The molbloc<sup>®</sup> elements are calibrated to be used at an absolute pressure which remains nearly constant, while the differential pressure varies with flow rate.

The flow control system enables the sequential sampling of up to 15 standards contained in cylinders by means of a 16 position valve (MPV-16). V2 is a 4-port 2-position valve. It is used to select which sample stream, from either the MSB or from a cylinder, is directed to the analysers, the other stream being directed to waste, without perturbing the flow of either stream.

## 2. Measurement protocol of the BIPM

On receipt by the BIPM, all cylinders were allowed to equilibrate at laboratory temperature for one week. All cylinders were rolled for 60 minutes to ensure homogeneity of the mixture.

Each cylinder was connected to one inlet of a 16-inlet automatic gas sampler connected to the FT-IR spectrometer and to the BIPM NO<sub>2</sub> dynamic generation facility.

The pressure reducers of each cylinder were flushed nine times with the mixture. The cylinder valves were then closed leaving the high pressure side of the pressure reducer at the cylinder pressure and the low pressure side of the pressure reducer at ~300 kPa. The cylinders were left to stand for at least 24 hours, to allow conditioning of the pressure reducers.

Immediately prior to an analysis, each cylinder valve was opened again and the pressure reducer flushed three times. The suite of cylinders was analysed sequentially.

For the FT-IR spectra acquisition, 120 scans were co-added over a period of 2 minutes to provide one single beam spectrum of a sample. This single beam spectrum was then ratioed with a similar spectrum of ultra-pure nitrogen collected under similar conditions to provide an absorbance spectrum of the gas sample (relative to ultra-pure nitrogen).

For each analyser, a calibration line was evaluated using the Generalized Least Squares approach described by ISO 6143:2001<sup>3</sup>.

The assigned BIPM nitrogen dioxide value was then equal to the predicted value from a calibration line calculated from a set of dynamic nitrogen dioxide primary gas mixtures obtained from the BIPM Nitrogen Dioxide (NO<sub>2</sub>) Primary Facility.

---

<sup>3</sup> ISO 6143:2001: Gas analysis- Comparison methods for determining and checking the composition of calibration gas mixtures.

### 3. BIPM measurement uncertainties and analyser response

The mole fractions of the dynamically produced gas mixtures obtained with the BIPM facility were calculated by the expression below:

$$x_{\text{NO}_2} = \left( \frac{P \times V_m}{q_v \times M_{\text{NO}_2}} \right) - \left( \frac{M_{\text{HNO}_3} \times x_{\text{HNO}_3}}{M_{\text{NO}_2}} \right) - \sum \left( \frac{M_{\text{imp}} \times x_{\text{imp}}}{M_{\text{NO}_2}} \right) \quad (6)$$

where:

$x_{\text{NO}_2}$  is the NO<sub>2</sub> mole fraction in  $\mu\text{mol}\cdot\text{mol}^{-1}$ ;

$P$  is the NO<sub>2</sub> permeation rate in  $\text{ng}\cdot\text{min}^{-1}$ ;

$V_m = 22.4038 \text{ L}\cdot\text{mol}^{-1}$ , is the molar volume of air/N<sub>2</sub> at standard conditions (273.15 K, 101.3 kPa);

$M_{\text{NO}_2} = 46.0055 \text{ g}\cdot\text{mol}^{-1}$ , is the molar mass of NO<sub>2</sub>;

$q_v$  is the total flow of N<sub>2</sub> given by the molbloc®/molbox® facility;

$x_{\text{HNO}_3}$  is the HNO<sub>3</sub> mole fraction in  $\mu\text{mol}/\text{mol}$  measured by FT-IR spectroscopy;

$M_{\text{HNO}_3} = 60.005 \text{ g}\cdot\text{mol}^{-1}$  is the molar mass of HNO<sub>3</sub>;

$x_{\text{imp}}$  are the mole fractions in  $\mu\text{mol}/\text{mol}$  of other impurities measured by FT-IR Spectroscopy; and

$M_{\text{imp}}$  are the molar mass of the impurities;

Applying the uncertainty propagation law and assuming no correlation between the input quantities, the following uncertainty expression was developed:

$$u^2(x_{\text{NO}_2}) = \left( \frac{\partial x_{\text{NO}_2}}{\partial P} \right)^2 \times u^2(P) + \left( \frac{\partial x_{\text{NO}_2}}{\partial V_m} \right)^2 \times u^2(V_m) + \left( \frac{\partial x_{\text{NO}_2}}{\partial M_{\text{NO}_2}} \right)^2 \times u^2(M_{\text{NO}_2}) + \left( \frac{\partial x_{\text{NO}_2}}{\partial q_v} \right)^2 \times u^2(F) + \left( \frac{\partial x_{\text{NO}_2}}{\partial x_{\text{HNO}_3}} \right)^2 \times u^2(x_{\text{HNO}_3}) + \left( \frac{\partial x_{\text{NO}_2}}{\partial M_{\text{HNO}_3}} \right)^2 \times u^2(M_{\text{HNO}_3}) + \left( \frac{\partial x_{\text{NO}_2}}{\partial x_{\text{imp}}} \right)^2 \times u^2(x_{\text{imp}}) + \left( \frac{\partial x_{\text{NO}_2}}{\partial M_{\text{imp}}} \right)^2 \times u^2(M_{\text{imp}}) \quad (7)$$

The permeation standard uncertainty, considering a permeation device with a permeation rate equivalent to  $P \approx 8357 \text{ ng}\cdot\text{min}^{-1}$ , was estimated to be  $u(P) \approx 4.18 \text{ ng}\cdot\text{min}^{-1}$ , assuming a rectangular distribution of the probability that  $P$  lies within the interval  $8357 \pm 6.17 \text{ ng}\cdot\text{min}^{-1}$ .

The uncertainty in the NO<sub>2</sub> molar mass of  $0.00047 \text{ g}\cdot\text{mol}^{-1}$ , 0.001 % relative, can be derived from the IUPAC Table of Atomic Weights.

The molar volume  $V_m$  of a real gas at standard conditions ( $T = 273.15 \text{ K}$ ,  $p = 101.325 \text{ kPa}$ ) is given by the formula

$$V_m = \frac{ZRT}{p} \quad (8)$$

where  $Z$  is the compressibility factor and  $R$  is the gas constant,  $8.314\,472\text{ J}\cdot\text{mol}^{-1}\cdot\text{K}^{-1}$ , with a relative uncertainty  $u(R)$  of  $1.8 \times 10^{-6}$ . Since they are defined by convention there is no uncertainty in  $T$  and  $p$ .

The compressibility factor of nitrogen obtained from the NIST Refprop database is  $Z_{N_2} = 0.9995434$  with a relative uncertainty  $u(Z)$  of  $15 \times 10^{-6}$ .

Thus the molar volume of nitrogen and its standard uncertainty are

$$V_{mN_2} = 22.4037\text{ L}\cdot\text{mol}^{-1}$$

$$u(V_{mN_2}) = 0.0003\text{ L}\cdot\text{mol}^{-1}, \text{ or } 1.5 \times 10^{-5} \text{ relative.}$$

The BIPM measured the flow in its system by using molblocs. These were calibrated by the LNE on 27 April 2009. The uncertainty of the BIPM's flow measurements is dominated by on the calibration uncertainty. The uncertainty in the flow measurements  $u(q_v)$  was taken from the LNE calibration certificate N° K20869/1. No additional component for the stability of the flow instrument was added, since the time between calibration and the first measurements were short, and no significant deviation between the first and second series of BIPM measurement results was observed for stable cylinder gas standards. The expanded relative uncertainty ( $k = 2$ ) quoted in the calibration certificate is 0.2 % at the flows used in the comparison. In correspondence between the BIPM and the LNE, the LNE confirmed<sup>4</sup> the relative expanded uncertainties quoted in their CMCs, comparison results and the calibration certificates to be as follows:

- 0.22 % to 0.40 % in LNE's CMCs
- 0.19 % to 0.26 % in the Euramet (1)
- 0.18 % to 0.27 % in the Calibration Certificate K20869/1.

The uncertainty in the calculated nitric acid mole fraction,  $x_{HNO_3}$ , obtained by FT-IR spectroscopy, is given by:

$$u(x_{HNO_3}) \approx \sqrt{(0.02)^2 + (0.015x)^2 + (0.05x)^2} \quad (9)$$

where  $x$  is the mole fraction of nitric acid predicted by FT-IR in the gas mixtures. A future publication will give a detailed description of the measuring methodology and quantification process by FT-IR for the determination of nitric acid.

As for  $NO_2$ , the uncertainty in nitric acid molar mass,  $0.000561\text{ g}\cdot\text{mol}^{-1}$  (0.0009 % relative), was derived from the IUPAC Table of Atomic Weights.

It follows that the uncertainty budget for a  $NO_2$  mixture having a nominal concentration of  $\sim 10.0\text{ }\mu\text{mol}\cdot\text{mol}^{-1}$  is as tabulated below in Table 8 using nitrogen as diluent gas:

<sup>4</sup> Private communication with Jean Barbe from LNE.



| Quantity                     | Estimate<br>$x_i$  | Assumed<br>distribution | Standard<br>uncertainty<br>$u(x_i)$                    | Sensitivity<br>coefficient<br>$c_i = \partial x_{\text{NO}_2} / \partial x$ | Uncertainty<br>contribution<br>$u(y)$<br>$\text{mol} \cdot \text{mol}^{-1}$ | Index<br>% |
|------------------------------|--|-------------------------|--|---|---|------------|
| $P$                          | 8.3573<br>$10^{-6} \cdot \text{g} \cdot \text{min}^{-1}$ | Normal                  | 4.18<br>$10^{-9} \cdot \text{g} \cdot \text{min}^{-1}$ | 1.1   | 4.5<br>$10^{-9}$  | 2.2        |
| $V_m$                        | 22.4038<br>$\text{L} \cdot \text{mol}^{-1}$              | Normal                  | 340.00<br>$10^{-6} \text{L} \cdot \text{mol}^{-1}$     | 400<br>$10^{-9}$  | 140<br>$10^{-12}$   | 0.0        |
| $q_{v \text{ molbloc1}}$     | 452<br>$10^{-3} \cdot \text{L} \cdot \text{min}^{-1}$    | Normal                  | 455.21<br>$10^{-6} \text{L} \cdot \text{min}^{-1}$     | -20<br>$10^{-6}$  | -9.1<br>$10^{-9}$   | 8.8        |
| $M_{\text{NO}_2}$            | 46.0055<br>$\text{g} \cdot \text{mol}^{-1}$              | Normal                  | 1.40<br>$10^{-3} \text{g} \cdot \text{mol}^{-1}$       | -190<br>$10^{-9}$   | -270<br>$10^{-12}$  | 0.0        |
| $x_{\text{HNO}_3}$           | 0.104<br>$10^{-6} \text{mol} \cdot \text{mol}^{-1}$      | Normal                  | 0.021<br>$10^{-6} \text{mol} \cdot \text{mol}^{-1}$    | -1.4  | -29<br>$10^{-9}$  | 88.5       |
| $x_{\text{N}_2\text{O}_4}$   | 0<br>$\text{mol} \cdot \text{mol}^{-1}$                  | Normal                  | 866<br>$10^{-12} \text{mol} \cdot \text{mol}^{-1}$     | -2.0  | -1.7<br>$10^{-9}$   | 0.3        |
| $x_{\text{N}_2\text{O}_3}$   | 0<br>$\text{mol} \cdot \text{mol}^{-1}$                  | Normal                  | 307<br>$10^{-12} \text{mol} \cdot \text{mol}^{-1}$     | -1.7  | -510<br>$10^{-12}$  | 0.0        |
| $x_{\text{N}_2\text{O}_5}$   | 0<br>$\text{mol} \cdot \text{mol}^{-1}$                  | Normal                  | 361<br>$10^{-12} \text{mol} \cdot \text{mol}^{-1}$     | -2.3  | -850<br>$10^{-12}$  | 0.0        |
| $x_{\text{HONO}}$            | 0<br>$\text{mol} \cdot \text{mol}^{-1}$                  | Normal                  | 520<br>$10^{-12} \text{mol} \cdot \text{mol}^{-1}$     | -1.0  | -530<br>$10^{-12}$  | 0.0        |
| $x_{\text{HO}_2\text{NO}_2}$ | 0<br>$\text{mol} \cdot \text{mol}^{-1}$                  | Normal                  | 572<br>$10^{-12} \text{mol} \cdot \text{mol}^{-1}$     | 1.7   | -980<br>$10^{-12}$  | 0.1        |
| $M_{\text{HNO}_3}$           | 63.013<br>$\text{g} \cdot \text{mol}^{-1}$               | Normal                  | 1.17<br>$10^{-3} \text{g} \cdot \text{mol}^{-1}$       | -2.3<br>$10^{-9}$   | -2.6<br>$10^{-12}$  | 0.0        |

| Quantity          | Value   | Standard<br>Uncertainty                   |
|-------------------|---|---|
| $x_{\text{NO}_2}$ | 8.86<br>$\mu\text{mol} \cdot \text{mol}^{-1}$ | 30<br>$\text{nmol} \cdot \text{mol}^{-1}$ |

Table 8. Uncertainty budget for a  $\text{NO}_2/\text{N}_2$  primary mixture generated with the BIPM facility.

Note: the molar masses  $M_{\text{N}_2\text{O}_4}$ ,  $M_{\text{N}_2\text{O}_3}$ ,  $M_{\text{N}_2\text{O}_5}$ ,  $M_{\text{HONO}}$ ,  $M_{\text{HO}_2\text{NO}_2}$  were not included in this budget as they represent negligible uncertainty contributions.

The degrees of freedom were numerous, so a coverage factor  $k = 2$  was assumed appropriate for the expanded uncertainty. The main uncertainty contributors remain the mole fraction determination of nitric acid and the gas flow measurements. Figure 13 illustrates the new uncertainties in  $x_{\text{NO}_2}$  for the dynamic generation of  $\text{NO}_2$  in nitrogen mixtures over the mole fraction range  $8 \mu\text{mol}/\text{mol}$  to  $12 \mu\text{mol}/\text{mol}$ , using a permeation tube with permeation rate of  $8357 \text{ng} \cdot \text{min}^{-1}$  and flows in the range  $350 \text{mL} \cdot \text{min}^{-1}$  to  $450 \text{mL} \cdot \text{min}^{-1}$ . The uncertainty is almost a constant and can be fitted by a linear function of the mole fraction. A least squares fit was made using the Excel LINEST function. The standard uncertainties in  $x_{\text{NO}_2}$  can be modelled by the following linear function (numerical values in  $\mu\text{mol}/\text{mol}$ ):

$$u(x_{\text{NO}_2}) = 0.001036x + 0.020818 \quad (10)$$




Figure 13. Standard uncertainty of dynamically generated NO<sub>2</sub> mixtures on the BIPM NO<sub>2</sub> facility over a range of  $x_{\text{NO}_2} = (8\text{-}12) \mu\text{mol/mol}$ .

#### 4. Covariance between two dynamically generated gas mixtures

Non-zero covariances,  $u(x_{\text{NO}_2,i}, x_{\text{NO}_2,j})$  were included in the uncertainty calculations because all dynamic mixtures were derived from the same BIPM facility and an error in the analyte content of the one gas is considered to propagate to all gas mixtures in a positive correlated fashion. The covariance between two calibration gas mixtures  $i$  and  $j$  is described as follows:

$$u(x_{\text{NO}_2,i}, x_{\text{NO}_2,j}) = \gamma \left[ u(x_{\text{NO}_2,i}) \right]^2, \quad (11)$$

Where  $u(x_{\text{NO}_2,i})$  is the standard uncertainty of the more concentrated mixture as given by equation 10,

$$\gamma = \frac{q_j}{q_i} \quad (12)$$

is the dilution factor of the total gas flows  $q_j$  and  $q_i$  (with  $q_j < q_i$ ). Note that as the NO<sub>2</sub> calibration gas mixtures generated with the facility are distributed in a small range of mole fractions (typically 8 nmol/mol to 12 nmol/mol), the dilution factor is often close to 1, and the covariances often close to the variances  $u(x_{\text{NO}_2,i})^2$ .

## 5. The Key Comparison Reference Values and their standard uncertainties

For each cylinder, the Key Comparison Reference Value is the NO<sub>2</sub> mole fraction assigned by the BIPM following the calibration procedure described above.

Following the CCQM GAWG guidance, it was decided that the standard uncertainty of the Key Comparison Reference Value (KCRV) can be quantified by the following equation

$$u_{\text{KCRV}}(x_{\text{NO}_2}) = \sqrt{(u(x_{\text{BIPM}}))^2 + (u(x_{\text{NO}_2\text{Losses}}))^2 + (u(x_{\text{NO}_2\text{Drift}}))^2} \quad (13)$$

where  $u(x_{\text{BIPM}})$  is the uncertainty associated with the value assigned by the BIPM following the procedure described above,  $u(x_{\text{NO}_2\text{Losses}})$  the uncertainty contribution due to NO<sub>2</sub> losses equivalent to 5.7 nmol/mol and  $u(x_{\text{NO}_2\text{Drift}})$  the uncertainty contribution due to observed drift in NO<sub>2</sub> estimated in 21 nmol/mol. The additional uncertainties described above increase the KCRV uncertainties from about 0.021 μmol/mol (as reported in the draft A report) to about 0.041 μmol/mol.

## 6. FT-IR analysis of gas standards

Analysis of all gas standards was undertaken to quantify nitric acid within the gas standards, and to compare these with the impurities and their uncertainties reported by the participating laboratories.

### 6.1 FT-IR Spectra acquisition procedure

A ThermoNicolet Nexus FT-IR spectrometer was configured with a MCT-high D\* liquid N<sub>2</sub>-cooled mid-infrared detector and a 6.4 m path-length multipass White cell (Gemini Scientific Instruments, USA) for the purposes of quantitative analysis for gas reference standards. The White cell has wetted surfaces of electropolished stainless steel and gold (mirror coatings) to minimize surface interactions with reactive gas phase species. To keep the internal optical path of the spectrometer free of any interference species this ensemble has been placed in stainless steel enclosure which is constantly purged with ultra high purity nitrogen (dewpoint ~-95°C, i.e. ~200 nmol·mol<sup>-1</sup> H<sub>2</sub>O) flowing at ~15 L·min<sup>-1</sup>.

The gas sample, from either the Rubotherm MSB or from a high pressure cylinder, flows from the NO<sub>2</sub> facility sampling manifold through the White cell, and then to waste. The sample flow rate is controlled immediately downstream of the White cell at ~400 mL·min<sup>-1</sup>. The sample pressure and temperature are measured in real time by

means of a calibrated barometer (Series 6000 Digital Pressure Transducer, Mensor, USA) and a calibrated 100  $\Omega$  RTD temperature probe attached to the White cell.

The spectrometer user interface is by means of the IMACC software. IMACC allows the automatic setting of all instrument parameters into Thermo's proprietary Omnic software for control, spectra acquisition and on-line analysis.

For the acquisition of high quality spectra suitable for quantitative analysis, 120 scans are co-added over a period of 2 minutes to provide one single beam spectrum of a sample. This single beam spectrum was then ratioed with a similar spectrum of ultra-pure nitrogen collected under similar conditions to provide an absorbance spectrum of the gas sample (relative to ultra-pure nitrogen).

The White cell has a volume of  $\sim 750$  mL and the sample flows at  $\sim 400$  mL $\cdot$ min $^{-1}$ . Assuming perfect mixing in the cell we estimate that an initial sample at time  $t = 0$  s has been 99.9 % replaced after 10 min of flow, and 99.9999 % replaced after 20 min. Accordingly, to ensure complete exchange of sample, spectrum acquisition started at  $t = 0$  but only the measured spectra obtained after flowing the sample through the White cell for 35 min were used for the mole fraction determination. We also empirically verified that after 30 min of flow, the sample was completely exchanged, within the bounds of measurement uncertainty.

The absorbance spectra of gas reference standards obtained following this procedure had a very high signal: noise ratio, with the level of noise in the baseline being typically  $\sim 2 \times 10^{-4}$  abs $_{10}$  peak-peak. By comparison the main NO $_2$  peak had absorbance in the range (0.04–0.16) abs $_{10}$ .

From times series analysis the uncertainty in the response of the FT-IR spectrometer was estimated in 20 nmol/mol for a 2 minutes average time.

## 6.2 Quantitative analysis of nitric acid

The determination of nitric acid was assessed configuring the FT-IR facility with a multi pass white cell with an optical path of (48 $\pm$ 1.2) m. Spectra were analysed by a non-linear least-square fitting of the measured absorption spectra with synthetic spectra using the program NLM4 (Non Linear MALT). NLM4 included the calculation of synthetic spectra from the HITRAN database of infrared absorption line parameters using the core of the program MALT (an acronym for Multiple Atmospheric Layer Transmission) software developed at the University of Wollongong described in detail by Griffith in 1996(2). The program convolved a stick spectrum calculated from the line parameters with the temperature, pressure, path length, resolution and instrument line shape function specified by the user. Spectra were calculated iteratively from an initial estimate of all input parameters following a modified Levenberg-Marquart algorithm until a least squares best fit to the measured spectrum was obtained. Gas concentrations in the sample were iteratively adjusted during the fit. The quality of the fit could be improved by choosing a proper spectra window of the measured spectrum. Spectra which had been acquired across a total wavelength range of 1660 cm $^{-1}$  to 1850 cm $^{-1}$  were fitted on spectral windows according to the impurities of interest, in this case nitric acid.

## 7. Uncertainty budget

Table 9 below summarizes the uncertainty sources and presents the final combined uncertainty associated with the FT-IR/MATL/CLS measurements of nitric acid at a mole fraction ( $x$ ) ranging from 100 nmol/mol to 250 nmol/mol with a FT-IR white cell with a 48 m optical path.

| Type A               | $\mu\text{mol/mol}$                        |
|----------------------|--|
| Stability            | 0.020                                      |
| Type B               |  |
| MALT                 | $0.015x$                                   |
| HITRAN               | $0.05x$                                    |
| Combined uncertainty | $\sqrt{(0.02)^2 + (0.015x)^2 + (0.05x)^2}$ |

Table 9: uncertainty budget associated with the FT-IR spectrometer used as an absolute method of quantification to determine the concentration of  $\text{HNO}_3$  in nitrogen.

## 8. Regression analysis

The procedure outlined in ISO 6143:2001 (Gas analysis-Comparison methods for determining and checking the composition of calibration gas mixtures) was used for the analysis of the data from the comparison. This required:

- the determination of the analysis function  $x=G(y)$  which expressed analyte contents in relation to corresponding measured responses;
- the validation of the analysis function; and
- the prediction of the mole fraction values from the measured responses and comparison to BIPM and NMI's values.

## 9. Determination and validation of analysis functions

All calculations were performed with B\_LEAST, a computer program which implemented the methodology of ISO 6143:2001, and takes into consideration uncertainties in both axes for regression analysis.

Validation studies performed by the BIPM to be published shortly will confirm the linearity of the FT-IR response in the  $x_{\text{NO}_2}$  range 4.5  $\mu\text{mol/mol}$  to 15.5  $\mu\text{mol/mol}$ .

## ANNEX 2- Report of Proposed $u$ (KCRV) for the Draft B report of CCQM-K74: Nitrogen dioxide, 10 $\mu\text{mol/mol}$

### Introduction and summary

During the 24th meeting of the CCQM GAWG in November 2010 the BIPM, in its role as coordinating laboratory of the key comparison CCQM-K74, was asked to investigate additional sources of uncertainty in its measurement results which had been proposed as reference values for the key comparison. The additional sources of uncertainty investigated were:

1. impurity analysis uncertainties;
2. reaction of  $\text{NO}_2$  to  $\text{HNO}_3$  in the BIPM permeation facility;
3. stability of the gas concentration of the transfer standards (cylinders);
4. contributions from flow measurements;

Version 0.1 of this report, which summarized the results of these investigations, was presented during the 25th CCQM GAWG meeting held in April 2011. During the meeting the value reported for the uncertainty of nitric acid concentrations was questioned. Further review of Version 0.1 of the report by the BIPM has confirmed that incorrect values for the nitric acid concentration were used in both Draft A and Version 0.1 of this report due to an error in the application of equation (8) in the Draft A report. This has been corrected in the current version (0.2) of the report, confirming that the major contributions to the uncertainty of the BIPM measurement results arise from the determination of nitric acid concentrations and the stability of gas mixtures in the transfer standards (cylinders). The inclusion of these components increases the standard uncertainty of the BIPM measurement results and by consequence the standard uncertainties of the proposed key comparison reference values from about 0.02  $\mu\text{mol/mol}$  (as reported in Draft A) to about 0.04  $\mu\text{mol/mol}$ .

### 1. Impurity analysis

The BIPM undertook additional measurements to verify the purity of the gas permeating from the  $\text{NO}_2$  permeation tube. This required: identification of the most likely possible impurities; determination of the limits of detection for the impurities; and measurements of high concentration  $\text{NO}_2$  mixtures in order to confirm the absence/presence of such impurities, taking into account their limits of detection.

A survey of literature on  $\text{NO}_2$  permeation devices was carried out. A publication from 1977 details the results of purity analysis of a  $\text{NO}_2$  permeation device (3). Based on this publication, it was concluded that possible impurities permeating from the tube could include:  $\text{N}_2\text{O}_4$ ,  $\text{N}_2\text{O}_5$ ,  $\text{HONO}$ ,  $\text{HO}_2\text{NO}_2$ ,  $\text{N}_2\text{O}_3$  and  $\text{N}_2\text{O}_4$ . The possibility of detecting such impurities with the BIPM FT-IR facility was evaluated by comparing their integrated line intensities against  $\text{HNO}_3$  in three spectral regions 800  $\text{cm}^{-1}$ , 1200  $\text{cm}^{-1}$  and 1600  $\text{cm}^{-1}$  (see Figure 14 and Table 10). The calculated differences in integrated line intensity, which are listed in Table 11, were subsequently used to calculate limits of detection for each potential impurity by using a ratio to the detection limit of  $\text{HNO}_3$ . The limit of detection for  $\text{HNO}_3$  was determined by calculating the standard deviation of the instrument response measuring  $\sim 600$   $\text{nmol/mol}$  of  $\text{HNO}_3$  contained in a 120

$\mu\text{mol/mol}$   $\text{NO}_2/\text{N}_2$  gas mixture generated with the  $\text{NO}_2$  VICI permeation device. The effective limits of detection at nominally  $10 \mu\text{mol/mol}$  were calculated by dividing the calculated detection limits at high concentration by twelve, as this scale directly with the  $\text{NO}_2$  concentration (see *Table 12*). Finally, the standard uncertainty of the impurities at low concentration were calculated assuming a rectangular distribution by the following equation  $(A/12)/\sqrt{3}$ , except for  $\text{N}_2\text{O}_4$  and  $\text{HNO}_3$  that were experimentally measured.




Figure 14. Integrated band intensity  $/(\text{cm}^{-1}/(\text{molecule}/\text{cm}^2))$  of  $\text{HNO}_3$  (4),  $\text{N}_2\text{O}_3$  (5),  $\text{N}_2\text{O}_5$  (6), HONO (7),  $\text{HO}_2\text{NO}_2$  (8) and  $\text{N}_2\text{O}_4$  (9).

| Molecule               | Integrated band intensity<br>$/(\text{cm}^{-1}/(\text{molecule}/\text{cm}^2))$ |  |  |  |
|------------------------|--|--|--|--|
|                        | Spectral region<br>$820 \text{ cm}^{-1} - 950 \text{ cm}^{-1}$                 | Spectral region<br>$1160 \text{ cm}^{-1} - 1240 \text{ cm}^{-1}$ | Spectral region<br>$1240 \text{ cm}^{-1} - 1400 \text{ cm}^{-1}$ | Spectral region<br>$1640 \text{ cm}^{-1} - 1770 \text{ cm}^{-1}$ |
| $\text{HNO}_3$         | 2.267E-17  | 1.59E-18   | 5.095E-17  | 5.693E-17  |
| HONO                   | 1.50E-17   | 1.00E-17   |  |  |
| $\text{N}_2\text{O}_3$ |  |  |  | 6.13E-17   |
| $\text{N}_2\text{O}_5$ |  |  | 3.82E-17   |  |
| $\text{N}_2\text{O}_4$ |  |  | 5.93E-17   | 9.60E-17   |
| PNA                    |  |  | 1.00E-18   |  |

Table 10. Integrated band intensities of the molecules  $\text{HNO}_3$ ,  $\text{N}_2\text{O}_3$ ,  $\text{N}_2\text{O}_4$ ,  $\text{N}_2\text{O}_5$ , HONO, and PNA in different spectral regions.

| Molecule                      | Relative difference (%) |
|-------------------------------|-------------------------|
| N <sub>2</sub> O <sub>3</sub> | -6.58                   |
| N <sub>2</sub> O <sub>5</sub> | -25.02                  |
| HONO                          | -80.37                  |
| PNA                           | -98.04                  |

Table 11: Relative difference of the integrated band intensities of the molecules N<sub>2</sub>O<sub>3</sub>, N<sub>2</sub>O<sub>5</sub>, HONO, and PNA with respect to HNO<sub>3</sub> in the region 1240 cm<sup>-1</sup> to 1400 cm<sup>-1</sup>.

| I                               | II  | III  | IV   | V   | VI   |
|---------------------------------|---|--|--|---|--|
| IR active impurity              | Detection limit (LOD) of impurity in a 120 μmol·mol <sup>-1</sup> NO <sub>2</sub> /N <sub>2</sub> gas mixture (nmol·mol <sup>-1</sup> ) (A) | Measured mole fraction of impurity in a 120 μmol·mol <sup>-1</sup> NO <sub>2</sub> /N <sub>2</sub> gas mixture (nmol·mol <sup>-1</sup> ) | Effective limit of detection of impurity in a 10 μmol·mol <sup>-1</sup> NO <sub>2</sub> /N <sub>2</sub> gas mixture (nmol·mol <sup>-1</sup> ) (A/12) | Measured mole fraction of impurity in a 10 μmol·mol <sup>-1</sup> NO <sub>2</sub> /N <sub>2</sub> gas mixture (nmol·mol <sup>-1</sup> ) | Assigned standard uncertainty in a 10 μmol·mol <sup>-1</sup> NO <sub>2</sub> /N <sub>2</sub> gas mixture (nmol·mol <sup>-1</sup> ) |
| HNO <sub>3</sub>                | 6.000   | 588  | 0.500  | 80  | 20.431   |
| N <sub>2</sub> O <sub>3</sub>   | 6.390   | 0  | 0.533  | 0   | 0.307  |
| N <sub>2</sub> O <sub>5</sub>   | 7.500   | 0  | 0.625  | 0   | 0.361  |
| HONO                            | 10.800  | 0  | 0.900  | 0   | 0.520  |
| HO <sub>2</sub> NO <sub>2</sub> | 11.880  | 0  | 0.990  | 0   | 0.572  |
| N <sub>2</sub> O <sub>4</sub>   | 1.500   | 60   | 0.125  | 0   | 0.866  |

Table 12: I: molecules identified as possible impurities in the nitrogen dioxide gas mixtures generated by the BIPM NO<sub>2</sub> primary facility using a permeation tube. II: limit of detection of the possible impurities at 120 μmol/mol of NO<sub>2</sub>/N<sub>2</sub>. The limit of detections for HNO<sub>3</sub> and N<sub>2</sub>O<sub>4</sub> were experimentally calculated using a long-path FT-IR gas cell. III: measured mole fraction of impurity present in 120 μmol/mol of NO<sub>2</sub>/N<sub>2</sub> with the long-path gas cell. IV: calculated effective limit of detection of each impurity at 10 μmol/mol of NO<sub>2</sub>/N<sub>2</sub>. V: measured mole fraction of each impurity at 10 μmol/mol of NO<sub>2</sub>/N<sub>2</sub>. VI: Assigned standard uncertainty considering a rectangular distribution (A/12/√3).

As previous purity studies did not allow the BIPM to confirm the existence of any other impurity apart from HNO<sub>3</sub>, highly concentrated NO<sub>2</sub> gas mixtures were analysed by long-path FT-IR. The mixtures were generated using a NO<sub>2</sub> VICI permeation tube that is a similar permeation device to the one used during the CCQM-K74 comparison.

The preparation of the new gas mixtures was undertaken in a new stainless-steel leak-free permeation chamber. The new permeation chamber had a volume three times smaller than the magnetic suspension chamber and it was specially designed for the purpose. The chamber was placed in the Rubotherm temperature controlled bath and was permanently flushed with a well characterized nitrogen flow as shown in the simplified scheme of Figure 15.

Due to the fact that a carrier gas flow rate of 30 ml/min to 100 ml/min of nitrogen was required for the preparation of the high concentration mixtures, 10 to 13 hours of constant measurements were necessary to obtain a stable response of the FT-IR for each mole fraction. The background stability was achieved by means of the FT-IR enclosure box flushed with 5 l/min of pure nitrogen. Temperature stability was achieved by the Rubotherm system bath and flow stability was accomplished by means of



molbloc/molbox measurements and a mass flow controller that was specially designed for small flow rates.

Figure 16 plots the absorbance spectrum of a gas mixture generated with the NO<sub>2</sub> VICI permeation device. NO<sub>2</sub>, HNO<sub>3</sub>, H<sub>2</sub>O, and CO<sub>2</sub> can be easily identified in the spectrum as well as a detected impurity present in the spectral region 1230 cm<sup>-1</sup> to 1280 cm<sup>-1</sup>. This impurity was identified as N<sub>2</sub>O<sub>4</sub>.

The mole fraction of NO<sub>2</sub> was determined using the software MALT and a Classical Least-Square analysis in the region 2820 cm<sup>-1</sup> to 2940 cm<sup>-1</sup>, a region where NO<sub>2</sub> absorbs less strongly but is free of HNO<sub>3</sub> and H<sub>2</sub>O interferences. The NO<sub>2</sub> mole fraction was verified by the ABB LIMAS analyser which was calibrated for this mole fraction range by means of a gas dilution facility and a highly concentrated gas reference standard. According to the MALT/CLS analysis, the gas mixture contained (106 ± 3) µmol/mol of NO<sub>2</sub> [(120 ± 1.2) µmol/mol according to the ABB LIMAS] and (588 ± 73) nmol/mol HNO<sub>3</sub>.

N<sub>2</sub>O<sub>4</sub> was quantified in the spectral region 1230 cm<sup>-1</sup> to 1280 cm<sup>-1</sup> using its relative area and by relating this to the HNO<sub>3</sub> absorption band located in 1286 cm<sup>-1</sup> to 1360 cm<sup>-1</sup> (see Figure 17). The N<sub>2</sub>O<sub>4</sub> calculated mole fraction was (60 ± 6) nmol/mol. A conservative standard uncertainty of 10 % was assigned to the mole fraction determination. Similarly, the N<sub>2</sub>O<sub>4</sub> mole fraction was calculated at different NO<sub>2</sub> mole fractions to observe how this changed with changing NO<sub>2</sub> mole fractions (Figure 18).




Figure 15. Scheme of the purity analysis experiment for NO<sub>2</sub> permeation tubes.




Figure 16. Infrared absorbance spectrum of the purity analysis of a 120 µmol/mol NO<sub>2</sub>/N<sub>2</sub> gas mixture generated using a NO<sub>2</sub> permeation tube commercialized by VICI .




Figure 17. Infrared absorbance spectra of the purity analysis of a 120  $\mu\text{mol/mol}$   $\text{NO}_2/\text{N}_2$  gas mixture generated using a  $\text{NO}_2$  permeation tube commercialized by VICI in the region  $1120\text{ cm}^{-1}$  to  $1460\text{ cm}^{-1}$ .




Figure 18.  $\text{N}_2\text{O}_4$  mole fraction at different  $\text{NO}_2$  concentrations of gas mixtures generated by a  $\text{NO}_2$  VICI permeation tube.

## 2. Reaction of NO<sub>2</sub> and HNO<sub>3</sub> in the permeation facility

The possibility of conversion of NO<sub>2</sub> to HNO<sub>3</sub> within the Rubotherm permeation chamber and pneumatic connections was investigated by FT-IR measurements. NO<sub>2</sub> and HNO<sub>3</sub> mole fraction changes were measured when a constant NO<sub>2</sub>/N<sub>2</sub> gas mixture was passed through the Rubotherm system. The NO<sub>2</sub>/N<sub>2</sub> gas mixtures were, as described in section 1, generated by a NO<sub>2</sub> VICI permeation tube placed into a permeation chamber located in a temperature bath. Figure 19 shows the experimental configuration.

Figure 20 plots the relative difference in nitrogen dioxide when the same gas mixture was passed through the Rubotherm system (A) or went directly to the FT-IR gas cell without passing through the Rubotherm system (B).

The quantitative analysis of the absorbance spectra was performed using a line A spectrum as background until minute 6000 (1200 spectra) where a second background spectrum (in the same position A) was measured. The infrared spectra collection was carried out every 5 minutes. It was noticed that when line A gas mixtures were analysed the relative nitrogen dioxide mole fraction dropped on average 20 nmol/mol. However, when the standard uncertainty due to the instrument response (20 nmol/mol) is plotted in (Figure 20 *i*) the relative difference in the nitrogen dioxide mole fractions measured is equivalent to the measurement uncertainty (Figure 20 *ii*). The nitric acid variations due to changing the flow path (Figure 20 *iii*) again resulted in smaller changes in observed concentration relative to the uncertainty of the instrument response to nitric acid (20 nmol/mol) (Figure 20 *iv*).

Therefore, it is concluded that there is no evidence for the reaction of NO<sub>2</sub> to HNO<sub>3</sub> in the gas phase within the Rubotherm system and that the source of the total HNO<sub>3</sub> measured arises from permeation from the tube. However, an uncertainty component is retained to cover the maximum changes in concentrations observed in these experiments, which was 20 nmol/mol.




Figure 19. Scheme of the experiment for testing possible absorption/desorption of nitrogen dioxide due to the permeation chamber walls and pneumatic connections.




Figure 20. Nitrogen dioxide and nitric acid time series produced by IMACC from gas mixtures A and B (see scheme of Figure 19). The uncertainty bars in the nitrogen dioxide plot on the right are equivalent to the standard uncertainty of the response of the instrument (20 nmol/mol).

### 3. Cylinder stability/Sampling

The differences between the first and the second nitrogen dioxide value assigned by the BIPM to all cylinders used in the comparison was reanalysed, with the maximum difference observed in the values taken as the basis of an uncertainty component describing the stability of the cylinder over the period of the comparison. The contribution from the gas cylinder stability was evaluated using the equation

$$u^2(x_i) = \frac{(b_+ + b_-)^2}{12} \quad (14)$$

where  $b_+$  and  $b_-$  are the upper and the lower boundaries for the nitrogen dioxide difference.

The maximum difference between BIPM assigned values for any particular cylinder was used so that,  $b_+$  and  $b_-$  were considered to be 37 nmol/mol each, and by applying equation (14),  $u_{\text{Drift}}(x_{\text{NO}_2}) = 21$  nmol/mol.

### 4. Flow

The BIPM measured the flow in its system by using molblocs. These were calibrated by the LNE on 27 April 2009. The uncertainty of the BIPM's flow measurements is dominated by and based on calibration. The uncertainty in the flow measurements was taken from the LNE calibration certificate N° K20869/1. No additional component for the stability of the flow instrument was added, since the time between calibration and the first measurements were short, and no significant deviation between the first and second series of BIPM measurement results was observed for stable cylinder gas standards.

The expanded relative uncertainty ( $k=2$ ) quoted in the calibration certificate is 0.2 % at the flows used in the comparison.

In correspondence between the BIPM and the LNE, the LNE confirmed<sup>5</sup> the relative expanded uncertainties quoted in their CMCs, comparison results and the calibration certificates to be as follows:

- 0.22 % to 0.40 % in LNE's CMCs
- 0.19 % to 0.26 % in the Euramet (1)
- 0.18 % to 0.27 % in the Calibration Certificate K20869/1.

### 5. Modified BIPM uncertainty budget

#### 5.1 Previous uncertainty budget:

The mole fractions of the dynamically produced gas mixtures obtained with the BIPM facility in the Draft A report was calculated by the expression:

<sup>5</sup> Private communication with Jean Barbe from LNE.

$$x_{\text{NO}_2} = \left( \frac{P \times V_m}{q_v \times M_{\text{NO}_2}} \right) - \left( \frac{M_{\text{HNO}_3} \times x_{\text{HNO}_3}}{M_{\text{NO}_2}} \right) \quad (15)$$

where:

$x_{\text{NO}_2}$  is the NO<sub>2</sub> mole fraction in µmol/mol;

$P$  is the NO<sub>2</sub> permeation rate in ng·min<sup>-1</sup>;

$V_m = 22.4038 \text{ L} \cdot \text{mol}^{-1}$ , is the molar volume of air/N<sub>2</sub> at standard conditions (273.15 K, 101.3 kPa);

$M_{\text{NO}_2} = 46.0055 \text{ g} \cdot \text{mol}^{-1}$ , is the molar mass of NO<sub>2</sub>;

$q_v$  is the total flow of N<sub>2</sub> given by the molbloc®/molbox® facility;

$x_{\text{HNO}_3}$  is the HNO<sub>3</sub> mole fraction in µmol/mol measured by FT-IR spectroscopy.

$M_{\text{HNO}_3} = 63.005 \text{ g} \cdot \text{mol}^{-1}$  is the molar mass of HNO<sub>3</sub>.

Table 13 shows the Draft A uncertainty budget for the nitrogen dioxide mole fraction determination updated by introducing the correct contribution of the nitric acid mole fraction calculation given by equation 8 of the Draft A report.

| Quantity           | Estimate<br>$x_i$                               | Assumed distribution | Standard uncertainty<br>$u(x_i)$                 | Sensitivity coefficient<br>$c_i = \partial x_{\text{NO}_2} / \partial x$ | Uncertainty contribution<br>$u(y)$<br>mol·mol <sup>-1</sup> | Index % |
|--------------------|---|----------------------|--|--|---|---------|
| $P$                | 8.3573<br>10 <sup>-6</sup> ·g·min <sup>-1</sup> | Normal               | 4.18<br>10 <sup>-9</sup> ·g·min <sup>-1</sup>    | 1.1  | 4.5<br>10 <sup>-9</sup>                                     | 2.2     |
| $V_m$              | 22.4038<br>L·mol <sup>-1</sup>                  | Normal               | 340.00<br>10 <sup>-6</sup> L·mol <sup>-1</sup>   | 400<br>10 <sup>-9</sup>  | 140<br>10 <sup>-12</sup>                                    | 0.0     |
| $q_v$ molbloc1     | 452<br>10 <sup>-3</sup> ·L·min <sup>-1</sup>    | Normal               | 455.21<br>10 <sup>-6</sup> L·min <sup>-1</sup>   | -20<br>10 <sup>-6</sup>  | -8.2<br>10 <sup>-9</sup>                                    | 7.3     |
| $M_{\text{NO}_2}$  | 46.0055<br>g·mol <sup>-1</sup>                  | Normal               | 1.40<br>10 <sup>-3</sup> g·mol <sup>-1</sup>     | -190<br>10 <sup>-9</sup>   | -91<br>10 <sup>-12</sup>                                    | 0.0     |
| $x_{\text{HNO}_3}$ | 0.104<br>10 <sup>-6</sup> mol·mol <sup>-1</sup> | Normal               | 0.021<br>10 <sup>-6</sup> ·mol·mol <sup>-1</sup> | -1.4   | -29<br>10 <sup>-9</sup>                                     | 90.4    |
| $M_{\text{HNO}_3}$ | 63.005<br>g·mol <sup>-1</sup>                   | Normal               | 1.17<br>10 <sup>-3</sup> g·mol <sup>-1</sup>     | -2.3<br>10 <sup>-9</sup>   | -1.3<br>10 <sup>-12</sup>                                   | 0.0     |

| Quantity | Value                          | Standard Uncertainty         |
|----------|--------------------------------|------------------------------|
| $C$      | 8.86<br>µmol·mol <sup>-1</sup> | 30<br>nmol·mol <sup>-1</sup> |

Table 13. Uncertainty budget for a NO<sub>2</sub> /N<sub>2</sub> primary mixture generated with the BIPM facility, before the introduction of additional components described in this report. The degrees of freedom were numerous, so a coverage factor  $k = 2$  was assumed appropriate for the expanded uncertainty.

In the revised version of the Draft A uncertainty, the contribution of the mole fraction determination of nitric acid became the main uncertainty contributor. Figure 21

illustrates the updated uncertainties in  $x_{\text{NO}_2}$  for the dynamic generation of  $\text{NO}_2$  in nitrogen mixtures over the mole fraction range  $8 \mu\text{mol/mol}$  to  $12 \mu\text{mol/mol}$ , using a permeation tube with permeation rate  $8357 \text{ ng}\cdot\text{min}^{-1}$  and flows ranging over  $350 \text{ mL}\cdot\text{min}^{-1}$  to  $450 \text{ mL}\cdot\text{min}^{-1}$ . A least squares fit of the absolute standard uncertainty  $u(x_{\text{NO}_2})$ , was made using the Excel LINEST function. The standard uncertainties in  $x_{\text{NO}_2}$  can be modelled by the following numerical equation (values given in  $\mu\text{mol/mol}$ ):

$$u(x_{\text{NO}_2}) = 0.001036x + 0.020818 \quad (16)$$




Figure 21. Standard uncertainty of dynamically generated  $\text{NO}_2$  mixtures on the BIPM  $\text{NO}_2$  Facility over a range of  $x_{\text{NO}_2} = (8\text{-}12) \mu\text{mol/mol}$ , before the introduction of additional components described in this report.

## 5.2 Additional uncertainty components as a result of this study:

### Impurities:

$\text{N}_2\text{O}_4$  was the only additional impurity that was detected at a concentration above its limit of detection in the highly concentrated gas mixtures. However, as this is the dimer of  $\text{NO}_2$  it is clear that the concentration of  $\text{N}_2\text{O}_4$  is determined by the equilibrium between the species and the concentration of  $\text{NO}_2$ . Extrapolation of the measurements of  $\text{N}_2\text{O}_4$  as a function of  $\text{NO}_2$  concentration (see Figure 18) leads to the conclusion that  $\text{N}_2\text{O}_4$  mole fractions will be smaller than a few  $\text{nmol/mol}$  when the  $\text{NO}_2$  mole fraction is  $10 \mu\text{mol/mol}$ . Therefore, in a  $10 \mu\text{mol/mol}$  nitrogen dioxide gas mixture the most probable concentration of  $\text{N}_2\text{O}_4$  was taken to be zero, with a conservative uncertainty based on the understanding that the  $\text{N}_2\text{O}_4$  fraction could not be greater than  $1.5 \text{ nmol/mol}$ . The associated standard uncertainty was therefore calculated to be  $u_{\text{imp}}(x_{\text{N}_2\text{O}_4}) = 1.5/\sqrt{3} = 0.866 \text{ nmol/mol}$ .

NO<sub>2</sub> losses:

A maximum variation of 20 nmol/mol was found as a result of the investigation of Section 2 where losses of NO<sub>2</sub> produced by the Rubotherm system were studied. The uncertainty contribution due to this issue in NO<sub>2</sub> is  $u_{NO_2 \text{ losses}}(x_{NO_2}) = 10/\sqrt{3} = 5.7$  nmol/mol.

Cylinder stability:

The contribution of the cylinder stability, is  $u_{\text{Drift}}(x_{NO_2}) = 21$  nmol/mol.

## 5.3 New uncertainty budget:

The measured equation proposed for the mole fraction determination for Draft B is:

$$x_{NO_2} = \left( \frac{P \times V_m}{q_v \times M_{NO_2}} \right) - \left( \frac{M_{HNO_3} \times x_{HNO_3}}{M_{NO_2}} \right) - \sum \left( \frac{M_{imp} \times x_{imp}}{M_{NO_2}} \right) \quad (17)$$

where:

$x_{NO_2}$  is the NO<sub>2</sub> mole fraction in  $\mu\text{mol} \cdot \text{mol}^{-1}$ ;

$P$  is the NO<sub>2</sub> permeation rate in  $\text{ng} \cdot \text{min}^{-1}$ ;

$V_m = 22.4038 \text{ L} \cdot \text{mol}^{-1}$ , is the molar volume of air/N<sub>2</sub> at standard conditions (273.15 K, 101.3 kPa);

$M_{NO_2} = 46.0055 \text{ g} \cdot \text{mol}^{-1}$ , is the molar mass of NO<sub>2</sub>;

$q_v$  is the total flow of N<sub>2</sub> given by the molbloc®/molbox® facility;

$x_{HNO_3}$  is the HNO<sub>3</sub> mole fraction in  $\mu\text{mol}/\text{mol}$  measured by FT-IR spectroscopy;

$M_{HNO_3} = 60.005 \text{ g} \cdot \text{mol}^{-1}$  is the molar mass of HNO<sub>3</sub>;

$x_{imp}$  are the mole fractions in  $\mu\text{mol}/\text{mol}$  of the other impurities measured by FT-IR Spectroscopy; and

$M_{imp}$  are the molar mass of the impurities;

The new uncertainty contributions of the impurities are described in Table 14. The uncertainty in the flow measurements was taken from the LNE calibration certificate N° K20869/1.

It follows that the uncertainty budget for a NO<sub>2</sub> mixture having a nominal concentration of  $\sim 8.8 \mu\text{mol}/\text{mol}$  is as shown as follows using nitrogen as the diluent gas:



| Quantity       | Estimate<br>$x_i$                                    | Assumed distribution | Standard uncertainty<br>$u(x_i)$                   | Sensitivity coefficient<br>$c_i = \partial x_{NO_2} / \partial x$ | Uncertainty contribution<br>$u(y)$<br>$\text{mol}\cdot\text{mol}^{-1}$ | Index % |
|----------------|--|----------------------|--|---|--|---------|
| $P$            | 8.3573<br>$10^{-6}\cdot\text{g}\cdot\text{min}^{-1}$ | Normal               | 4.18<br>$10^{-9}\cdot\text{g}\cdot\text{min}^{-1}$ | 1.1   | 4.5<br>$10^{-9}$   | 2.2     |
| $V_m$          | 22.4038<br>$\text{L}\cdot\text{mol}^{-1}$            | Normal               | 340.00<br>$10^{-6}\text{L}\cdot\text{mol}^{-1}$    | 400<br>$10^{-9}$  | 140<br>$10^{-12}$  | 0.0     |
| $q_v$ molbloc1 | 452<br>$10^{-3}\cdot\text{L}\cdot\text{min}^{-1}$    | Normal               | 455.21<br>$10^{-6}\text{L}\cdot\text{min}^{-1}$    | -20<br>$10^{-6}$  | -9.1<br>$10^{-9}$  | 8.8     |
| $M_{NO_2}$     | 46.0055<br>$\text{g}\cdot\text{mol}^{-1}$            | Normal               | 1.40<br>$10^{-3}\text{g}\cdot\text{mol}^{-1}$      | -190<br>$10^{-9}$   | -270<br>$10^{-12}$   | 0.0     |
| $x_{HNO_3}$    | 0.104<br>$10^{-6}\text{mol}\cdot\text{mol}^{-1}$     | Normal               | 0.021<br>$10^{-6}\text{mol}\cdot\text{mol}^{-1}$   | -1.4  | -29<br>$10^{-9}$   | 88.5    |
| $x_{N_2O_4}$   | 0<br>$\text{mol}\cdot\text{mol}^{-1}$                | Normal               | 866<br>$10^{-12}\text{mol}\cdot\text{mol}^{-1}$    | -2.0  | -1.7<br>$10^{-9}$  | 0.3     |
| $x_{N_2O_3}$   | 0<br>$\text{mol}\cdot\text{mol}^{-1}$                | Normal               | 307<br>$10^{-12}\text{mol}\cdot\text{mol}^{-1}$    | -1.7  | -510<br>$10^{-12}$   | 0.0     |
| $x_{N_2O_5}$   | 0<br>$\text{mol}\cdot\text{mol}^{-1}$                | Normal               | 361<br>$10^{-12}\text{mol}\cdot\text{mol}^{-1}$    | -2.3  | -850<br>$10^{-12}$   | 0.0     |
| $x_{HONO}$     | 0<br>$\text{mol}\cdot\text{mol}^{-1}$                | Normal               | 520<br>$10^{-12}\text{mol}\cdot\text{mol}^{-1}$    | -1.0  | -530<br>$10^{-12}$   | 0.0     |
| $x_{HO_2NO_2}$ | 0<br>$\text{mol}\cdot\text{mol}^{-1}$                | Normal               | 572<br>$10^{-12}\text{mol}\cdot\text{mol}^{-1}$    | 1.7   | -980<br>$10^{-12}$   | 0.1     |
| $M_{HNO_3}$    | 63.013<br>$\text{g}\cdot\text{mol}^{-1}$             | Normal               | 1.17<br>$10^{-3}\text{g}\cdot\text{mol}^{-1}$      | -2.3<br>$10^{-9}$   | -2.6<br>$10^{-12}$   | 0.0     |

| Quantity   | Value                                       | Standard Uncertainty                    |
|------------|---|---|
| $x_{NO_2}$ | 8.86<br>$\mu\text{mol}\cdot\text{mol}^{-1}$ | 30<br>$\text{nmol}\cdot\text{mol}^{-1}$ |

Table 14. Uncertainty budget for a  $\text{NO}_2/\text{N}_2$  primary mixture generated with the BIPM facility. Note: the molar masses  $M_{N_2O_4}$ ,  $M_{N_2O_3}$ ,  $M_{N_2O_5}$ ,  $M_{HONO}$ ,  $M_{HO_2NO_2}$  were not included in this budget as they represent negligible uncertainty contributions.

The degrees of freedom were numerous, so a coverage factor  $k = 2$  was assumed appropriate for the expanded uncertainty. The main uncertainty contributors remain the mole fraction determination of nitric acid and the gas flow measurements. Figure 13 illustrates the new uncertainties in  $x_{NO_2}$  for the dynamic generation of  $\text{NO}_2$  in nitrogen mixtures over the mole fraction range  $8\ \mu\text{mol}/\text{mol}$  to  $12\ \mu\text{mol}/\text{mol}$ , using a permeation tube with permeation rate of  $8357\ \text{ng}\cdot\text{min}^{-1}$  and flows in the range  $350\ \text{mL}\cdot\text{min}^{-1}$  to  $450\ \text{mL}\cdot\text{min}^{-1}$ . The uncertainty is almost a constant and can be fitted by a linear function of the mole fraction. A least squares fit was made using the Excel LINEST function. The standard uncertainties in  $x_{NO_2}$  can be modelled by the following linear function (numerical values in  $\mu\text{mol}/\text{mol}$ ):

$$u(x_{\text{NO}_2}) = 0.001036x + 0.020818 \quad (18)$$




Figure 22. Standard uncertainty of dynamically generated NO<sub>2</sub> mixtures on the BIPM NO<sub>2</sub> facility over a range of  $x_{\text{NO}_2} = (8\text{-}12)$  µmol/mol.

## 6. Covariance between two dynamically generated gas mixtures

Non-zero covariances,  $u(x_{\text{NO}_2,i}, x_{\text{NO}_2,j})$  were included in the uncertainty calculations because all dynamic mixtures were derived from the same BIPM facility and an error in the analyte content of the one gas is considered to propagate to all gas mixtures in a positive correlated fashion. The covariance between two calibration gas mixtures  $i$  and  $j$  is described as follows:

$$u(x_{\text{NO}_2,i}, x_{\text{NO}_2,j}) = \gamma [u(x_{\text{NO}_2,i})]^2, \quad (19)$$

Where

$u(x_{\text{NO}_2,i})$  is the standard uncertainty of the more concentrated mixture as given by equation 4,

$$\gamma = q_j / q_i \quad (20)$$

is the dilution factor of the total gas flows  $q_j$  and  $q_i$  (with  $q_j < q_i$ ). Note that as the NO<sub>2</sub> calibration gas mixtures generated with the facility are distributed in a small range of mole fractions (typically 8 nmol/mol to 12 nmol/mol), the dilution factor is often close to 1, and the covariances often close to the variances  $u(x_{\text{NO}_2,i})^2$ .

## 7. Determination of the calibration function

As in the draft A report, a calibration line was calculated using the Generalized Least Square fitting procedure described in the standard ISO 6143:2001. Uncertainties were calculated according to the budget described in Table 14, and covariances between the calibration mixtures were calculated according to Equation 19.

For each cylinder, the calibration line was calculated using a number  $n$  of calibration points, with  $n$  ranging from 12 to 36 depending on the measurement conditions. Note that due to the covariance values, this had a negligible effect on the uncertainties of the calibration parameters.

For each cylinder, a predicted NO<sub>2</sub> mole fraction  $x_{\text{BIPM}}$  was calculated from the FT-IR response using the calibration parameters.

## 8. The Key Comparison Reference Values and their standard uncertainties

For each cylinder, the Key Comparison Reference Value is the NO<sub>2</sub> mole fraction  $x_{\text{BIPM}}$  assigned by the BIPM following the calibration procedure described above. The additional uncertainties described in this report do not change the KCRVs calculated in the draft A report but only their associated uncertainties, as can be seen in Table 15 and Table 16.

Following the CCQM GAWG guidance, it was decided that the standard uncertainty of the Key Comparison Reference Value (KCRV) can be quantified by the following equation:

$$u_{\text{KCRV}}(x_{\text{NO}_2}) = \sqrt{\left(u(x_{\text{BIPM}})\right)^2 + \left(u(x_{\text{NO}_2\text{Losses}})\right)^2 + \left(u(x_{\text{NO}_2\text{Drift}})\right)^2} \quad (21)$$

where  $u(x_{\text{BIPM}})$  is the uncertainty associated with the value assigned by the BIPM following the procedure described above,  $u(x_{\text{NO}_2\text{Losses}})$  the uncertainty contribution due to NO<sub>2</sub> losses equivalent to 5.7 nmol/mol and  $u(x_{\text{NO}_2\text{Drift}})$  the uncertainty contribution due to observed drift in NO<sub>2</sub> estimated in 21 nmol/mol.

The additional uncertainties described in this report increase the KCRV uncertainties from about 0.021 µmol/mol as reported in the draft A report to about 0.041 µmol/mol, as can be seen in Table 15 and Table 16.

Corresponding degrees of equivalence and associated uncertainties are also listed in Table 15 and Table 16 and plotted in Figure 23 and Figure 24.

## Draft A measurement results

| Laboratory | Cylinder    | FT-IR response    |                      | BIPM              |                      | Participants     |                     |                                      |        |                     |
|------------|-------------|-------------------|----------------------|-------------------|----------------------|------------------|---------------------|--------------------------------------|--------|---------------------|
|            |             | $y_{\text{BIPM}}$ | $u(y_{\text{BIPM}})$ | $x_{\text{BIPM}}$ | $u(x_{\text{BIPM}})$ | $x_{\text{Lab}}$ | $u(x_{\text{Lab}})$ | $D(x_{\text{Lab}}, x_{\text{BIPM}})$ | $u(D)$ | $U(D)$<br>( $k=2$ ) |
| NPL        | #930659-PRM | 10.715            | 0.020                | 10.227            | 0.021                | 10.331           | 0.040               | 0.104                                | 0.045  | 0.090               |
| NIM        | #930650-PRM | 10.716            | 0.020                | 10.228            | 0.021                | 10.150           | 0.050               | -0.078                               | 0.054  | 0.108               |
| SMU        | #930655-PRM | 10.832            | 0.020                | 10.347            | 0.021                | 10.100           | 0.060               | -0.247                               | 0.064  | 0.127               |
| NMIA       | #930662-PRM | 10.864            | 0.020                | 10.378            | 0.021                | 10.740           | 0.315               | 0.362                                | 0.316  | 0.631               |
| NMISA      | #930649-PRM | 10.832            | 0.020                | 10.347            | 0.021                | 10.690           | 0.185               | 0.343                                | 0.186  | 0.372               |
| CERI       | #930671-PRM | 10.843            | 0.020                | 10.352            | 0.021                | 10.400           | 0.190               | 0.048                                | 0.191  | 0.382               |
| METAS      | #930660-PRM | 10.904            | 0.020                | 10.432            | 0.021                | 10.630           | 0.080               | 0.198                                | 0.083  | 0.165               |
| INRIM      | #930667-PRM | 10.731            | 0.020                | 10.183            | 0.021                | 11.470           | 0.070               | 1.287                                | 0.073  | 0.146               |
| KRISS      | #930661-PRM | 10.766            | 0.020                | 10.270            | 0.021                | 10.450           | 0.155               | 0.180                                | 0.156  | 0.313               |
| FMI        | #930673-PRM | 10.984            | 0.020                | 10.418            | 0.021                | 9.880            | 0.150               | -0.538                               | 0.151  | 0.303               |
| LNE        | #930675-PRM | 10.944            | 0.020                | 10.378            | 0.021                | 10.260           | 0.065               | -0.118                               | 0.068  | 0.136               |
| NIST       | #930654-PRM | 10.859            | 0.020                | 10.299            | 0.021                | 10.280           | 0.050               | -0.019                               | 0.054  | 0.108               |
| VSL        | #930674-PRM | 10.933            | 0.020                | 10.370            | 0.021                | 10.510           | 0.105               | 0.140                                | 0.107  | 0.214               |
| CEM        | #930676-PRM | 10.991            | 0.020                | 10.435            | 0.023                | 10.720           | 0.110               | 0.285                                | 0.112  | 0.225               |
| VNIIM      | #930713-PRM | 10.896            | 0.020                | 10.321            | 0.029                | 10.550           | 0.080               | 0.229                                | 0.085  | 0.170               |
| BAM        | #930722-PRM | 10.903            | 0.020                | 10.350            | 0.023                | 10.530           | 0.375               | 0.180                                | 0.376  | 0.751               |
| BIPM       | #930697-PRM | 10.905            | 0.020                | 10.343            | 0.021                | 10.343           | 0.024               | 0.000                                | 0.032  | 0.064               |

Table 15. Results of the key comparison presented in the draft A report. All values are given in  $\mu\text{mol/mol}$  apart from FT-IR response which is in arbitrary units

## Draft A - measurement results




Figure 23. Difference between participants' results and BIPM reported values for nitrogen dioxide mole fractions. The error bar represents the expanded uncertainty at a 95 % level of confidence.

## Draft B - measurement results

| Laboratory | Cylinder    | FT-IR response                               |                      | BIPM              |                      | Participants     |                     | Differences                           |        |                     |
|------------|-------------|--|----------------------|-------------------|----------------------|------------------|---------------------|---------------------------------------|--------|---------------------|
|            |             | $y_{\text{BIPM}}$<br>( $\mu\text{mol/mol}$ ) | $u(y_{\text{BIPM}})$ | $x_{\text{BIPM}}$ | $u(x_{\text{BIPM}})$ | $x_{\text{Lab}}$ | $u(x_{\text{Lab}})$ | $D(x_{\text{Lab}} - x_{\text{BIPM}})$ | $u(D)$ | $U(D)$<br>( $k=2$ ) |
| NPL        | #930659-PRM | 10.715                                       | 0.020                | 10.226            | 0.042                | 10.331           | 0.040               | 0.105                                 | 0.058  | 0.115               |
| NIM        | #930650-PRM | 10.716                                       | 0.020                | 10.227            | 0.042                | 10.150           | 0.050               | -0.077                                | 0.065  | 0.130               |
| SMU        | #930655-PRM | 10.832                                       | 0.020                | 10.347            | 0.042                | 10.100           | 0.060               | -0.247                                | 0.073  | 0.146               |
| NMIA       | #930662-PRM | 10.864                                       | 0.020                | 10.378            | 0.042                | 10.740           | 0.315               | 0.362                                 | 0.318  | 0.635               |
| NMISA      | #930649-PRM | 10.832                                       | 0.020                | 10.347            | 0.042                | 10.690           | 0.185               | 0.343                                 | 0.190  | 0.379               |
| CERI       | #930671-PRM | 10.843                                       | 0.020                | 10.351            | 0.041                | 10.400           | 0.190               | 0.049                                 | 0.194  | 0.389               |
| METAS      | #930660-PRM | 10.904                                       | 0.020                | 10.431            | 0.041                | 10.630           | 0.080               | 0.199                                 | 0.090  | 0.180               |
| INRIM      | #930667-PRM | 10.731                                       | 0.020                | 10.183            | 0.042                | 9.990            | 0.100               | -0.193                                | 0.108  | 0.217               |
| KRISS      | #930661-PRM | 10.766                                       | 0.020                | 10.270            | 0.042                | 10.450           | 0.155               | 0.180                                 | 0.160  | 0.321               |
| FMI        | #930673-PRM | 10.984                                       | 0.020                | 10.417            | 0.041                | 9.880            | 0.150               | -0.537                                | 0.156  | 0.311               |
| LNE        | #930675-PRM | 10.944                                       | 0.020                | 10.378            | 0.041                | 10.260           | 0.065               | -0.118                                | 0.077  | 0.154               |
| NIST       | #930654-PRM | 10.859                                       | 0.020                | 10.299            | 0.041                | 10.280           | 0.050               | -0.019                                | 0.065  | 0.130               |
| VSL        | #930674-PRM | 10.933                                       | 0.020                | 10.370            | 0.041                | 10.510           | 0.105               | 0.140                                 | 0.113  | 0.226               |
| CEM        | #930676-PRM | 10.991                                       | 0.020                | 10.435            | 0.042                | 10.720           | 0.110               | 0.285                                 | 0.118  | 0.235               |
| VNIIM      | #930713-PRM | 10.896                                       | 0.020                | 10.320            | 0.042                | 10.550           | 0.080               | 0.230                                 | 0.090  | 0.181               |
| BAM        | #930722-PRM | 10.903                                       | 0.020                | 10.350            | 0.042                | 10.530           | 0.375               | 0.180                                 | 0.377  | 0.755               |
| BIPM       | #930697-PRM | 10.905                                       | 0.020                | 10.343            | 0.041                | 10.343           | 0.024               | 0.000                                 | 0.048  | 0.096               |

Table 16. Results of the key comparison taking into account the additional uncertainties presented in this report<sup>6</sup>. All values are given in  $\mu\text{mol/mol}$  apart from FT-IR response which is in arbitrary units (BIPM reported uncertainty has not been changed from the Draft A value. The corrected standard uncertainty for the BIPM reported value which would result from following the procedure outlined in this document would be  $0.035 \mu\text{mol/mol}$ .)

<sup>6</sup> Note: 1 nmol/mol is the maximum observed difference between Draft A and Draft B assigned values by BIPM due to B\_Least rounding.

## Draft B - measurement results




Figure 24. Difference between participants' results and BIPM reported values for nitrogen dioxide mole fractions. The error bar represents the expanded uncertainty at a 95 % level of confidence.

- [1] Hughes E E, Rook H L, Deardorff E R, Margeson J H and Fuerst R G 1977 Performance of a NO<sub>2</sub> permeation device. *Analytical chemistry* **49** (12) 1823–1829
- [2] Rothman L S *et al* 2005 The HITRAN 2004 molecular spectroscopic database *Journal of Quantitative Spectroscopy & Radiative Transfer* **96** 139–204
- [3] Kagann R H and Maki A G 1984 Infrared absorption intensities for N<sub>2</sub>O<sub>3</sub> *Journal of Quantitative Spectroscopy and Radiative Transfer* **31** 173–176
- [4] Newnham D, Ballard J and Page M 1993 Infrared absorbance cross-sections for dinitrogen pentoxide vapour *Journal of Quantitative Spectroscopy and Radiative Transfer* **50** 571–577
- [5] Barney W S *et al* 2000 Infrared Absorption Cross-Section Measurements for Nitrous Acid (HONO) at Room Temperature *The Journal of Physical Chemistry* **104** (8) 1692–1699
- [6] Höjer S, May R D and Miller C E 1999 Intensities of the 1397 cm<sup>-1</sup> (ν<sub>3</sub>) band of HO<sub>2</sub>NO<sub>2</sub> and feasibility of atmospheric detection *Journal of Quantitative Spectroscopy and Radiative Transfer* **55** 273–278
- [7] Hurtmans D, Herman M and Vander Auwera J 1993 Integrated band intensities in N<sub>2</sub>O<sub>4</sub> in the infrared range *Journal of Quantitative Spectroscopy and Radiative Transfer* **50** 595–602
- [8] Probert C, Johns A I and Metaxiotou Z 2008 An intercomparison of low flow gas facilities at eleven European laboratories using a Molbloc transfer package: EURAMET.MM.FF-S3 (EURAMET Project no 806) *Metrologia* **45**, Tec. Sup. 07012



## ANNEX 3 - Measurement reports of participants

### Centro Español de metrología (CEM)

#### A1. General information

|                                    |  |     |                  |
|------------------------------------|--|-----|------------------|
| Institute                          | CENTRO ESPAÑOL DE METROLOGÍA (CEM)                                     |     |                  |
| Address                            | CALLE ALFAR, 2<br>28760 TRES CANTOS (MADRID)<br>SPAIN                  |     |                  |
| Contact person                     | TERESA E. FERNÁNDEZ VICENTE  |     |                  |
| Telephone                          | + 34 918 074 751   | Fax | + 34 918 074 807 |
| Email*                             | <a href="mailto:tefernandez@cem.mityc.es">tefernandez@cem.mityc.es</a> |     |                  |
| Serial number of cylinder received | 930676 (D650059)   |     |                  |
| Cylinder pressure as received      | ≈ 95 bar   |     |                  |

#### A2. Results

| Nitrogen dioxide mole fraction          | Expanded uncertainty                       | Coverage factor |
|---|--|-----------------|
| $x_{\text{NO}_2}$ / $\mu\text{mol/mol}$ | $U(x_{\text{NO}_2})$ / $\mu\text{mol/mol}$ |                 |
| 10,72                                   | 0,22                                       | 2               |

#### A3. Uncertainty Budget

Please provide a complete uncertainty budget.

The mathematical model used to calculate the uncertainty in the composition of mixture analysed is a linear combination of the sources of uncertainty due to the instrument used and the repeatability of the measurements. This leads to:

$$u = \sqrt{u_{B\_least}^2 + u_r^2}$$

where  $u_{B\_least}$  is the largest uncertainty among the obtained uncertainties by means of the B\_least software (linear fit regression) and  $u_r$  is the standard deviation of the mean of the results obtained during the period of measurements (usually from 3 to 5 days).

Table 1 summarizes the uncertainty budget.

| Uncertainty source   | Assumed distribution | Standard uncertainty / $\mu\text{mol/mol}$ | Sensitivity coefficient | Contribution to standard uncertainty / $\mu\text{mol/mol}$ |
|--|----------------------|--|-------------------------|--|
| $u_{B\_least}$   | normal               | 0,11                                       | 1                       | 0,11   |
| $u_r$  | normal               | 0,0064                                     | 1                       | 0,0064   |
| <b>Combined standard uncertainty / <math>\mu\text{mol/mol}</math></b>              |                      |  |                         | <b>0,11</b>  |
| <b>Expanded uncertainty, <math>k^7 = 2</math> / <math>\mu\text{mol/mol}</math></b> |                      |  |                         | <b>0,22</b>  |

**Table 1.** Detailed uncertainty budget.

#### A4. Description of the procedure used during the gas analysis

Please describe in detail the analytical method(s) used for gas analysis<sup>8</sup>.

The mixture was analysed by means of a Thermo 42i chemiluminescence NO-NO<sub>2</sub>-NO<sub>x</sub> analyser. Three standards were used with the compositions specified in Table 2:

| Species                       | Amount Fraction<br>NPL1272 / $\mu\text{mol/mol}$ | Amount Fraction<br>NPL1273 / $\mu\text{mol/mol}$ | Amount Fraction<br>NPL1274 / $\mu\text{mol/mol}$ |
|-------------------------------|--|--|--|
| <b>Nitrogen Dioxide</b>       | (5,01 ± 0,10)                                    | (10,00 ± 0,15)                                   | (15,01 ± 0,22)                                   |
| <b>Oxygen (not certified)</b> | 13   | 22   | 33   |
| <b>Nitrogen</b>               | Balance  | Balance  | Balance  |

**Table 2.** Primary reference gas mixtures used.

All mixtures were prepared gravimetrically and analysed by the Non-Dispersive Ultraviolet (NDUV) technique.

Upon arrival the sample cylinder was rolled and stored in the laboratory under laboratory reference conditions. A pressure reducer was connected to the standards and the sample cylinder. The reducers were carefully flushed as prescribed in International Standard ISO 16664:2004 (Gas analysis – Handling of calibration gases and calibration gas mixtures – Guidelines).

The standards and the sample cylinder were connected to an automatic gas sampler connected to the specific analyser in the increasing order of concentration and the gas outlet pressure was 2 bar. The cylinders were analysed sequentially in three measurement cycles and the data reported came from three independent working days.

B\_least software based on International Standard ISO 6143:2001 (Gas analysis – Comparison methods for determining and checking the composition of calibration gas mixtures) was used to certify the sample cylinder. A calibration curve was fitted to the mean value from three measurement cycles for each standard and the drift among cycles was used as standard uncertainty, because it turned out to be significant. The method used resulted in a calculated mole fraction and standard uncertainty for the sample cylinder. In all cases a linear function was used with a goodness of fit less than 2.

#### A5. Complementary information on the cylinder

Please report the value of the pressure left in the cylinder before shipment to the BIPM:

65 bar approximately.

If any other component other than NO<sub>2</sub>, nitrogen and oxygen was detected and/or quantified please report its mole fraction in the table below:

<sup>7</sup> The coverage factor shall be based on approximately 95 % confidence.

<sup>8</sup> The choice of the procedure used for gas analysis is the responsibility of the participating laboratory. Nevertheless, for a proper evaluation of the data, it is necessary that the calibration method, as well as the way in which the calibration mixtures have been prepared is reported to the co-ordinators.

## Finnish Meteorological Institute (MIKES-FMI)

### A1. General information

|                                    |  |     |               |
|------------------------------------|--|-----|---------------|
| Institute                          | Finnish Metrological Institute (MIKES-FMI)   |     |               |
| Address                            | P.O. BOX 503<br>FI-00101 HELSINKI<br>FINLAND |     |               |
| Contact person                     | Jari Waldén                                  |     |               |
| Telephone                          | +358505914615                                | Fax | +358919295403 |
| Email*                             | Jari.walden@fmi.fi                           |     |               |
| Serial number of cylinder received | APEX 930 673                                 |     |               |
| Cylinder pressure as received      | 83 bar                                       |     |               |

### A2. Results

| Nitrogen dioxide mole fraction        | Expanded uncertainty                     | Coverage factor |
|---------------------------------------|--|-----------------|
| $x_{\text{NO}_2} / \mu\text{mol/mol}$ | $U(x_{\text{NO}_2}) / \mu\text{mol/mol}$ |                 |
| 9.88                                  | 0.30                                     | 2               |

### A3. Uncertainty Budget

Please provide a complete uncertainty budget.

The measurement equation for calculating the concentration of the CCQM-K74 is following:

$$C(\text{NO}_2)_{\text{K74}} = \frac{[\text{NO}_2]_{\text{K74}} \cdot C(\text{NO})_{\text{PRM}}}{\text{eff} \cdot [\text{NO}]_{\text{PRM}}} \quad (1)$$

Where  $[\text{NO}_2]_{\text{K74}}$  is the diluted concentration of the gas standard K74 ( $C(\text{NO}_2)_{\text{K74}}$ ), where  $[\text{NO}_2]_{\text{K74}} = [\text{NO}_x] - [\text{NO}]$  is measured with the analyser APNA-360. The NO<sub>2</sub>-converter efficiency of the analyser, *eff*, was defined as 98.0 %. The concentration of the PRM,  $C(\text{NO})_{\text{PRM}}$ , was known from the certificate as well as the expanded uncertainty. The PRM was purchased from NPL. The expanded uncertainty can be calculated using the formula of propagation of error.

$$u_c^2 = \sum_{i=1}^n \left( \frac{\partial f}{\partial w_i} \right)^2 u_i^2 + 2 \sum_{i=1}^{N-1} \sum_{j=i+1}^N \frac{\partial f}{\partial w_i} \frac{\partial f}{\partial w_j} u_i u_j \rho_{ij} \quad (2)$$

The combined variance of equation (1) can be expressed as:

$$\begin{aligned} u(NO_2)_{c,K74}^2 &= \left( \frac{[NO_2]_{K74}}{eff \cdot [NO]_{PRM}} \right)^2 \cdot u^2(C(NO)_{PRM}) + \left( \frac{C(NO)_{PRM}}{eff \cdot [NO]_{PRM}} \right)^2 \cdot u^2([NO_2]_{K74}) + \left( -\frac{[NO_2]_{K74} \cdot C(NO)_{PRM}}{eff^2 \cdot [NO]_{PRM}} \right)^2 \cdot u^2(eff) \\ &+ \left( -\frac{[NO_2]_{K74} \cdot C(NO)_{PRM}}{eff \cdot [NO]_{PRM}^2} \right)^2 \cdot u^2([NO]_{PRM}) + 2 \cdot \left\{ \left( \frac{C(NO)_{PRM}}{eff \cdot [NO]_{PRM}} \right) \cdot \left[ \left( \frac{[NO_2]_{K74}}{eff \cdot [NO]_{PRM}} \right) \cdot u([NO_2]_{K74}) \cdot u(C(NO)_{PRM}) \cdot r_{[NO_2]_{K74}, C(NO)_{PRM}} \right. \right. \\ &+ \left. \left. \left( -\frac{[NO_2]_{K74} \cdot C(NO)_{PRM}}{eff^2 \cdot [NO]_{PRM}} \right) \cdot u([NO_2]_{K74}) \cdot u(eff) \cdot r_{[NO_2]_{K74}, eff} + \left( -\frac{[NO_2]_{K74} \cdot C(NO)_{PRM}}{eff \cdot [NO]_{PRM}^2} \right) \cdot u([NO_2]_{K74}) \cdot u([NO]_{PRM}) \cdot r_{[NO_2]_{K74}, [NO]_{PRM}} \right] \right. \\ &+ \left. \left( \frac{[NO_2]_{K74}}{eff \cdot [NO]_{PRM}} \right) \left[ \left( -\frac{[NO_2]_{K74} \cdot C(NO)_{PRM}}{eff^2 \cdot [NO]_{PRM}} \right) u(C(NO)_{PRM}) u(eff) r_{C(NO)_{PRM}, eff} + \left( -\frac{[NO_2]_{K74} \cdot C(NO)_{PRM}}{eff \cdot [NO]_{PRM}^2} \right) u(C(NO)_{PRM}) u([NO]_{PRM}) r_{C(NO)_{PRM}, [NO]_{PRM}} \right. \right. \\ &+ \left. \left. \left( -\frac{[NO_2]_{K74} \cdot C(NO)_{PRM}}{eff^2 \cdot [NO]_{PRM}} \right) \cdot \left( -\frac{[NO_2]_{K74} \cdot C(NO)_{PRM}}{eff \cdot [NO]_{PRM}^2} \right) \cdot u(eff) \cdot u([NO]_{PRM}) \cdot r_{eff, [NO]_{PRM}} \right\} \quad (3) \end{aligned}$$

The standard uncertainties are listed below:

$$u(C(NO)_{PRM}) = 0.25 \% \quad (4)$$

From the certificate of the PRM

$$u([NO_2]_{K74}) = 1.1 \% \quad (5)$$

From the measured value. Uncertainty of the analyser includes: calibration of the analyser, converter efficiency, linearity, repeatability at zero and at measured concentration

$$u(eff) = 0.5 \% \quad (6)$$

is the uncertainty of the converter efficiency

$$u(NO)_{PRM} = 0.9 \% \quad (7)$$

From the measured value. Uncertainty of the analyser includes: calibration of the analyser, linearity, repeatability at zero and at measured concentration.

The correlation factors are analysed next:

$$r_{[NO_2]_{K74}, C(NO)_{PRM}} = 0, \quad (8)$$

No correlation between the concentration of PRM and the concentration of NO<sub>2</sub> was observed at the CCQM-K74 standard

$$r_{[NO_2]_{K74}, eff} = -0.004 (\% / nmol / mol) \quad (9)$$

Analysed of the converter efficiency as a function of NO<sub>2</sub> concentration during GPT

$$r_{[NO_2]_{K74}, [NO]_{PRM}} = 0 \quad (10)$$

No correlation between the CCQM-K74 and PRM

$$r_{C(NO)_{PRM}, eff} = 0.25 (\% nmol / mol) \quad (11)$$

Linked with the uncertainty of the PRM

$$r_{C(NO)_{PRM},[NO]_{PRM}}=0.01 \quad (12)$$

Dilution ratio of the dilution method

$$r_{eff,[NO]_{PRM}}=0.004(\%/nmol/mol) \quad (13)$$

Analysed of the converter efficiency as a function of NO concentration during GPT

As a result from the measurements and the values applied in eqs (4) to (13) into eq. (3) the relative combined standard uncertainty was calculated as 1.5 % and 0.151  $\mu\text{mol/mol}$  as an absolute value.

#### A4. Description of the procedure used during the gas analysis

Please describe in detail the analytical method(s) used for gas analysis<sup>9</sup>.

The method used for the analysis of the cylinder is the one supported by the laboratory in routine work. The method is based on comparison of the results between the dilutions of the known concentration of the gas standard with the result of the unknown standard at the same dilution stage. The known concentration of the gas standard, PRM, by NPL of 49.86  $\mu\text{mol/mol}$  NO in nitrogen with the relative expanded uncertainty of 0.5 % and the unknown standard was the KC standard (APEX 930 673). Since the gas compound of the KC gas compound was nitrogen dioxide the converter efficiency of the gas analyser, based on chemiluminescence method, was defined prior the measurements. The converter efficiency of the gas analyser used, APNA-360 by Horiba, was defined by the gas phase titration method. In addition the linearity of the APNA-360 analyser was defined in the range covering the concentration range for both of the measurements. The concentration of the KC gas standard was calculated according to equation 1.

#### A5. Complementary information on the cylinder

Please report the value of the pressure left in the cylinder before shipment to the BIPM:

67 bar

If any other component other than NO<sub>2</sub>, nitrogen and oxygen was detected and/or quantified please report its mole fraction in the table below:

| Component | Mole fraction / nmol/mol | Expanded uncertainty | Coverage factor | Measurement technique   |
|-----------|--------------------------|----------------------|-----------------|---|
| NO        | 2.                       | 0.5 nmol/mol         | 2               | Direct measurement by TEI42C analyser bypassing the NO <sub>2</sub> converter |

<sup>9</sup> The choice of the procedure used for gas analysis is the responsibility of the participating laboratory. Nevertheless, for a proper evaluation of the data, it is necessary that the calibration method, as well as the way in which the calibration mixtures have been prepared is reported to the co-ordinators.

## Istituto Nazionale di Ricerca Metrologica (INRIM)

### A1. General information

|                                    |  |     |                  |
|------------------------------------|--|-----|------------------|
| Institute                          | INRIM-Istituto Nazionale di Ricerca Metrologica  |     |                  |
| Address                            | Strada delle Cacce 91<br>I-10135 Torino<br>Italy |     |                  |
| Contact person                     | Michela Segà                                     |     |                  |
| Telephone                          | +39 011 3919 948                                 | Fax | +39 011 3919 937 |
| Email*                             | m.sega@inrim.it                                  |     |                  |
| Serial number of cylinder received | 930667   |     |                  |
| Cylinder pressure as received      | 90 bar   |     |                  |

### A2. Results

| Nitrogen dioxide mole fraction        | Expanded uncertainty                     | Coverage factor |
|---------------------------------------|--|-----------------|
| $x_{\text{NO}_2} / \mu\text{mol/mol}$ | $U(x_{\text{NO}_2}) / \mu\text{mol/mol}$ |                 |
| 11,47                                 | 0,14                                     | 2               |

Tab. 1: INRIM results

### A3. Uncertainty Budget

Please provide a complete uncertainty budget.

The contributions to the combined standard uncertainty of the results are due to the calibration curve and to repeatability of readings of sample measurements. From each of the four calibration curves a NO<sub>2</sub> concentration value with its combined standard uncertainty was estimated. The final result is the mean of these four values and its combined standard uncertainty was calculated by pooling the four variances.

#### Determination of calibration curves

The calibration curves were determined by means of an Excel worksheet, developed at INRIM, based on the Weighted Least Squares method, which calculates a linear correction to be applied to the instrument readings according to the following equation:

$$x = y + d(y) = y + \alpha_0 + \alpha_1 y \quad (1)$$

where  $x$  is concentration of the analyte in the reference gas mixtures,  $y$  is the instrument output and  $d(y) = \alpha_0 + \alpha_1 y$  is the correction. The measurands are the polynomial coefficients  $\alpha_0$  and  $\alpha_1$ . The estimation algorithm takes care of different sources of uncertainty: the reference gas mixtures uncertainty, the repeatability of the instrument, the lack of fit, the instrument resolution. These sources are merged together in the Excel worksheet for calibration curves calculation, hence it is very difficult to separate each contribution. For detailed information see the reference: Plassa M., Mosca M., Segà M. "Carbon Dioxide Determination for High Accuracy Weighings" in: *Proceedings of the 16<sup>th</sup> International Conference IMEKO TC3/APMF '98*, Myung Sai Chung Ed.; Taejon, Korea, **1998**, pp. 183-191.

Being the reference gas mixtures prepared at INRIM by diluting the same pre-mixture, a correlation coefficient of 0,9 was adopted in the calculation.

Calibration curve data are summarized in the following tables (Tab. 2-5):

| $\alpha$   |          | $u_c(\alpha)$ | $\psi_\alpha$ |           |
|------------|----------|---------------|---------------|-----------|
| $\alpha_0$ | 1,1E-1   | 1,8E-01       | 3,27E-02      | -2,97E-03 |
| $\alpha_1$ | -2,4E-02 | 1,7E-02       | -2,97E-03     | 2,89E-04  |

Tab. 2: calibration curve parameters of 29/12/09 (first set)

| $\alpha$   |          | $u_c(\alpha)$ | $\psi_\alpha$ |           |
|------------|----------|---------------|---------------|-----------|
| $\alpha_0$ | 6,3E-03  | 2,0E-01       | 3,84E-02      | -3,69E-03 |
| $\alpha_1$ | -1,6E-02 | 1,9E-02       | -3,69E-03     | 3,70E-04  |

Tab. 3: calibration curve parameters of 30/12/09 (second set)

| $\alpha$   |          | $u_c(\alpha)$ | $\psi_\alpha$ |           |
|------------|----------|---------------|---------------|-----------|
| $\alpha_0$ | -6,7E-02 | 2,1E-01       | 4,28E-02      | -3,62E-03 |
| $\alpha_1$ | -1,7E-02 | 1,8E-02       | -3,62E-03     | 3,22E-04  |

Tab. 4: calibration curve parameters of 04/01/10 (third set)

| $\alpha$   |          | $u_c(\alpha)$ | $\psi_\alpha$ |           |
|------------|----------|---------------|---------------|-----------|
| $\alpha_0$ | 3,3E-02  | 1,7E-01       | 2,79E-02      | -2,51E-03 |
| $\alpha_1$ | -2,0E-02 | 1,6E-02       | -2,51E-03     | 2,44E-04  |

Tab. 5: calibration curve parameters of 05/01/10 (fourth set)

After the calibration process  $\alpha_0$  and  $\alpha_1$  being known, if a set of  $n_r$  instrument readings, arranged in a vector  $r$ , are to be corrected by the calibration algorithm, the matrix  $R$  can be defined, whose columns are the first two powers of  $r$ .

$$R = \begin{pmatrix} r^0 & r \end{pmatrix}$$

The correction vector  $d(r)$  can be computed from  $d(r) = R\alpha$ , where  $\alpha$  is the vector of the coefficients  $\alpha_0$  and  $\alpha_1$ . The corrected readings are:

$$q = d(r) + r \quad (2)$$

The covariance matrix of the readings is  $\psi_r = s^2 I$ , where  $s$  is the repeatability standard uncertainty of the instrument and  $I$  an identity matrix. The covariance matrix  $\psi_d$  of  $d$  can be estimated starting from the law of propagation of uncertainty:

$$\psi_d = \nabla_\alpha(d) \psi_\alpha \nabla_\alpha(d)^T + \nabla_r(d) \psi_r \nabla_r(d)^T$$

where the symbol  $\nabla_z(w)$  means the Jacobian matrix, i.e. the matrix derivative, of the vector  $w$  with respect to the vector  $z$  and  $\psi_\alpha$  is the variance-covariance matrix of the coefficients  $\alpha_0$  and  $\alpha_1$ .

#### A4. Description of the procedure used during the gas analysis

Please describe in detail the analytical method(s) used for gas analysis<sup>10</sup>.

The analysis was carried out by means of a chemiluminescence analyser CLD Thermo 42i having resolution of 0,01  $\mu\text{mol mol}^{-1}$ . The data are visualized on the instrument display and manually recorded. For its calibration, a set of three gas mixtures, having the characteristics reported in table 6, were prepared at INRIM by gravimetry. The mixtures were prepared in aluminium alloy cylinders of 5L by diluting with synthetic air N 57 a pre-mixture of NO at the concentration of 99,49  $\mu\text{mol/mol}$  ( $U=0,80 \mu\text{mol/mol}$ ,  $k=2$ ) in  $\text{N}_2$  purchased from NPL (UK). The gravimetric preparation was carried out following the weighing scheme A-B-B-A.

For the uncertainty evaluation of the gas mixtures prepared at INRIM, the following sources were taken into account: molar masses of parent gases, purity of parent gases, composition of the pre-mixture, weighing of parent gases, covariances between the input quantities.

<sup>10</sup> The choice of the procedure used for gas analysis is the responsibility of the participating laboratory. Nevertheless, for a proper evaluation of the data, it is necessary that the calibration method, as well as the way in which the calibration mixtures have been prepared is reported to the co-ordinators.

The gravimetric preparation was checked twice by comparison with two different sets of three mixtures of NO<sub>2</sub> in synthetic air or purified air: the first set was made by a mixture purchased from NPL and two from an ISO 17025 accredited laboratory; the second set was made by three mixtures purchased from an ISO 17025 accredited laboratory. The analyses confirmed also the complete oxidation of NO into NO<sub>2</sub> during the mixture preparation process.

A mixture of NO<sub>2</sub> at the concentration of 10,07 µmol/mol ( $U=0,15$  µmol/mol,  $k=2$ ) in synthetic air, purchased from NPL, was used as a quality control standard to validate the calibration curve obtained with INRIM mixtures. The results were satisfactory: the certified value and the analytical values obtained at INRIM are in agreement within the declared uncertainties.

As for purity, the data certified by the producers were used.

| Mixture number | Cylinder number | NO <sub>2</sub> molar fraction $\chi$<br>µmol/mol | $U(\chi)$ ( $k=2$ )<br>µmol/mol |
|----------------|-----------------|---|---------------------------------|
| 026            | D69 6430        | 11,88   | 0,05                            |
| 028            | D56 6403        | 10,21   | 0,04                            |
| 030            | D56 6405        | 8,24  | 0,03                            |

Tab. 6: calibration mixtures

The measurements were carried out at a flow of approximately 35 L h<sup>-1</sup>. It was previously proved that small flow variations do not affect the measurement value. The instrument readings were collected after the signal stabilization, i.e. 2 minutes.

No correction for ambient pressure was made because the instrument had been calibrated every day in which measurements were carried out according to the following measurement protocol:

Standard n. 1, Sample, Standard N. 2, Sample, Standard N. 3, Sample, (repeated 3 times). The control standard was analysed at the beginning and at the end of the entire sequence.

No correction for ambient temperature was made.

Four different calibration curves were determined, one for each measurement day and they were used to estimate the final result for NO<sub>2</sub>, according to section A3.

The result reported in Table 1 for the estimated concentration of NO<sub>2</sub> if compared with the nominal value of the comparison shows a great discrepancy. Considering that the checking of INRIM calibration standards showed a good comparability between the gravimetric values and the analytical values, and also the NPL mixture of NO<sub>2</sub> used to validate the calibration curves showed a comparability within the declared uncertainty, a possible matrix effect could be the reason of such discrepancy. Indeed, INRIM standard mixtures, the mixtures used for their analytical verification and the NPL mixture used as a control standard have a matrix of synthetic air or purified air, instead the mixture analysed for the comparison has a matrix of nitrogen. This effect was not investigated during the measurements, but some additional measurements are being carried out at INRIM in order to check this aspect.

## A5. Complementary information on the cylinder

Please report the value of the pressure left in the cylinder before shipment to the BIPM: 70 bar

If any other component other than NO<sub>2</sub>, nitrogen and oxygen was detected and/or quantified please report its mole fraction in the table below:

| Component | Mole fraction /<br>nmol/mol | Expanded uncertainty | Coverage factor | Measurement<br>technique |
|-----------|-----------------------------|----------------------|-----------------|--------------------------|
|           |                             |                      |                 |                          |
|           |                             |                      |                 |                          |
|           |                             |                      |                 |                          |



## Result form CCQM-K74-R: Addendum to INRIM report

### 1. Introduction

In section A 5 it was stated that "The result reported in Table 1 for the estimated concentration of NO<sub>2</sub> if compared with the nominal value of the comparison shows a great discrepancy. Considering that the checking of INRIM calibration standards showed a good comparability between the gravimetric values and the analytical values, and also the NPL mixture of NO<sub>2</sub> used to validate the calibration curves showed a comparability within the declared uncertainty, a possible matrix effect could be the reason of such discrepancy. Indeed, INRIM standard mixtures, the mixtures used for their analytical verification and the NPL mixture used as a control standard have a matrix of synthetic air or purified air, instead the mixture analysed for the comparison has a matrix of nitrogen. This effect was not investigated during the measurements, but some additional measurements are being carried out at INRIM in order to check this aspect."

Additional measurement were carried out at INRIM to evaluate the possible bias due to the different concentration of O<sub>2</sub> in the CCQM-K74 mixture (0,001 mol/mol) and in the INRIM standards (0,164 mol/mol) used to calibrate the chemiluminescence analyser for the determination of NO<sub>2</sub> mole fraction in the CCQM-K74 cylinder.

### 2. Evaluation of sensitivity of INRIM chemiluminescence analyser

Gravimetric mixtures containing NO<sub>2</sub> at mole fraction of about 10 µmol/mol and different mole fractions of oxygen prepared at INRIM having composition reported in table 7, were analysed.

| Mixture number     | Cylinder number | NO <sub>2</sub> mole fraction $\chi$<br>µmol/mol | $U(\chi)$ (k=2)<br>µmol/mol | O <sub>2</sub> mole fraction<br>mol/mol |
|--------------------|-----------------|--|-----------------------------|---|
| 027                | D69 6406        | 10,20  | 0,04                        | 0,164                                   |
| 028 <sup>(*)</sup> | D56 6403        | 10,21  | 0,04                        | 0,164                                   |
| 032                | D56 6404        | 10,17  | 0,04                        | 0,045                                   |
| 031                | D56 6402        | 9,94   | 0,04                        | 0,009                                   |
| 033                | D56 6409        | 9,77   | 0,04                        | 0,001                                   |

Tab. 7: INRIM mixtures used to check the sensitivity of INRIM chemiluminescence analyser having different O<sub>2</sub> mole fractions. (\*)One of the 3 mixtures used to calibrate the analyser during the CCQM-K74 measurements (see table 6).

The analyses showed, as expected, a bias affecting the determination of NO<sub>2</sub> mole fraction which is related to the concentration of O<sub>2</sub> present in the mixtures. At the increasing of O<sub>2</sub> concentration, being the nominal mole fraction of NO<sub>2</sub> almost the same, its value, analytically determined by means of the chemiluminescence analyser, increases following a non-linear model.

The sensitivity factor of the analyser itself at the different O<sub>2</sub> mole fractions was calculated by dividing the analyser response by the NO<sub>2</sub> mole fraction. Also the data obtained during the measurements for the CCQM comparison were analysed and the analyser sensitivity for O<sub>2</sub> at 0,164 mol/mol (oxygen mole fraction of INRIM mixtures used for the analyser calibration in the comparison) was determined. The analyser showed comparable values of sensitivity during both the CCQM-K74 analyses and the additional measurements carried out after the comparison, confirming that its performances had remained stable over the time period. Comparing the values, the sensitivity for the analyser for the CCQM-K74 mixture was hence calculated considering an O<sub>2</sub> concentration of 0,001 mol/mol.

### 3. Determination of the correction

With the same calibration standards used for the CCQM comparison, listed in table 6, INRIM mixture n. 033 containing 0,001 mol/mol of O<sub>2</sub> was analysed following the same protocol used for the comparison and an analytical value for the NO<sub>2</sub> mole fraction,  $x_{\text{analytical\_INRIM}}$ , was determined. The same quality control standard was also used.

Table 8 reports the calibration curve parameters; table 9 reports the analytical value  $x_{\text{analytical\_INRIM}}$ , the corresponding gravimetric value  $x_{\text{grav\_INRIM}}$  for mixture n. 033 and their standard uncertainties.

|            | $\alpha$ | $u_c(\alpha)$ | $\psi_\alpha$ |           |
|------------|----------|---------------|---------------|-----------|
| $\alpha_0$ | -5,7E-01 | 2,5E-01       | 6,26E-02      | -6,20E-03 |
| $\alpha_1$ | -8,0E-02 | 2,5E-02       | -6,20E-03     | 6,40E-04  |

Tab. 8: calibration curve parameters of 09/06/10

|                                | NO <sub>2</sub> molar fraction $\chi$<br>μmol/mol | $u(\chi)$<br>μmol/mol |
|--------------------------------|---|-----------------------|
| $x_{\text{analytical\_INRIM}}$ | 11,25   | 0,07                  |
| $x_{\text{grav\_INRIM}}$       | 9,77  | 0,02                  |

Tab. 9: NO<sub>2</sub> molar fractions in INRIM mixture n. 033 with their standard uncertainties

Two different methods were used to calculate the correction for the  $x_{\text{NO}_2}$  value determined during the CCQM-K74 measurements: a multiplicative one, based on the analyser sensitivity, and a additive one based on the difference between the analytical value and the gravimetric one for INRIM mixture n. 033. Due to it being easiest method of estimating the uncertainty of the correction, the additive method was chosen.

The difference  $x_{\text{analytical\_INRIM}} - x_{\text{grav\_INRIM}}$  (1,48 μmol/mol), is the correction for the analytical bias due to the effect of O<sub>2</sub>.

The corrected mole fraction of NO<sub>2</sub>,  $\hat{x}_{\text{NO}_2}$ , is calculated according to eq. 3:

$$\hat{x}_{\text{NO}_2} = x_{\text{NO}_2} - (x_{\text{analytical\_INRIM}} - x_{\text{grav\_INRIM}}) \quad (3)$$

where

$x_{\text{NO}_2}$  : is the mole fraction of NO<sub>2</sub> analytically determined by INRIM in the analysis of the CCQM-K74 cylinder, reported in table 1;

$x_{\text{grav\_INRIM}}$  : is the mole fraction of NO<sub>2</sub> calculated from the gravimetric preparation of INRIM mixture n. 033 containing 0,001 mol/mol of O<sub>2</sub>

$x_{\text{analytical\_INRIM}}$  : is the mole fraction of NO<sub>2</sub> analytically determined in INRIM mixture n. 033 with the parameters reported in table 8.

#### 4. Uncertainty Budget

The combined standard uncertainty of  $\hat{x}_{\text{NO}_2}$  is calculated according to eq. 4:

$$u(\hat{x}_{\text{NO}_2}) = \sqrt{u^2(x_{\text{NO}_2}) + u^2(x_{\text{grav\_INRIM}}) + x_{\text{analytical\_INRIM}}^2 - 2u^2(x_{\text{grav\_INRIM}})} \quad (4)$$

where the last term is due to the covariances between the two series of results  $x_{\text{NO}_2}$  and  $x_{\text{analytical\_INRIM}}$  which have been determined using the same set of calibration gas mixtures. The value of  $u(x_{\text{grav\_INRIM}})$  was used for the covariance contribution as it represents a typical value of standard uncertainty for the gravimetrically prepared mixtures at INRIM (see also table 6).

The values of the different contributions are shown in table 10.

| Quantity                       | Standard uncertainty<br>$\mu\text{mol/mol}$  |
|--------------------------------|--|
| $x_{\text{NO}_2}$              | 0,07   |
| $x_{\text{grav\_INRIM}}$       | 0,02   |
| $x_{\text{analytical\_INRIM}}$ | 0,07   |
|                                | Covariance<br>$\mu\text{mol}^2/\text{mol}^2$ |
| $x_{\text{grav\_INRIM}}$       | -0,0008                                      |

Tab. 10: uncertainty contributions for  $u_c(x_{\text{NO}_2})$

## 5. Results

Table 11 reports the corrected results for  $\hat{x}_{\text{NO}_2}$  with its expanded uncertainty.

| Nitrogen dioxide mole fraction              | Expanded uncertainty                           | Coverage factor |
|---|--|-----------------|
| $\hat{x}_{\text{NO}_2} / \mu\text{mol/mol}$ | $U(\hat{x}_{\text{NO}_2}) / \mu\text{mol/mol}$ |                 |
| 9,99  | 0,20   | 2               |

Tab. 11: corrected results for  $\hat{x}_{\text{NO}_2}$

The corrected value obtained using the multiplicative factor is  $\hat{x}_{\text{NO}_2} = 10,03 \mu\text{mol/mol}$ , which is comparable with the value reported in table 11, within the associated uncertainty.

## Korea Research Institute of Standards and Science (KRISS)

### A1. General information

|                                    |  |     |                 |
|------------------------------------|--|-----|-----------------|
| Institute                          | KRISS  |     |                 |
| Address                            | 1 Doryong-Dong, Yuseong-Gu, Daejeon 305-340, Korea |     |                 |
| Contact person                     | Oh Sang-Hyub                                       |     |                 |
| Telephone                          | +82-42-868-5341                                    | Fax | +82-42-868-5344 |
| Email*                             | shoh@kriss.re.kr                                   |     |                 |
| Serial number of cylinder received | D650044  |     |                 |
| Cylinder pressure as received      | about 85 bar                                       |     |                 |

### A2. Results

| Nitrogen dioxide mole fraction<br>$x_{\text{NO}_2}$ / $\mu\text{mol/mol}$ | Expanded uncertainty<br>$U(x_{\text{NO}_2})$ / $\mu\text{mol/mol}$ | Coverage factor |
|---|--|-----------------|
| 10.45   | 0.31   | 2               |

### A3. Uncertainty Budget

Please provide a complete uncertainty budget.

| Analyte         | Relative standard uncertainties / % |          |           | Expanded uncertainty / % | Coverage factor |
|-----------------|-------------------------------------|----------|-----------|--------------------------|-----------------|
|                 | Gravimetry                          | Analysis | Stability |                          |                 |
| NO <sub>2</sub> | 0.07                                | 0.25     | 1.45      | 2.95                     | 2               |

### A4. Description of the procedure used during the gas analysis

Please describe in detail the analytical method(s) used for gas analysis<sup>11</sup>.

<sup>11</sup> The choice of the procedure used for gas analysis is the responsibility of the participating laboratory. Nevertheless, for a proper evaluation of the data, it is necessary that the calibration method, as well as the way in which the calibration mixtures have been prepared is reported to the co-ordinators.

The NO<sub>2</sub> was analysed using a chemiluminescent NO/NO<sub>x</sub> analyser (Thermo Model 42i). 6 PRMs, about 10 µmol/mol, were connected to multi-position valve of a computer operated gas sampling system (lab. made). A gas sampling system was used to deliver the sample stream to the NO/NO<sub>x</sub> analyser and collect responses. Analysis and stability uncertainties were calculated from these data. The K -74 cylinder and 1 PRM (D727632) were analysed likewise. Sample flow (300 ml/min) to the instrument was controlled by MFC (Bronkhorst) during analysis and each sample was purged for three minutes as follows.

PSM(15 minutes) – Nitrogen(3 minutes) – K-74(15 minutes) – Nitrogen(3 minutes) – PSM(15 minutes) ...

#### A5. Complementary information on the cylinder

Please report the value of the pressure left in the cylinder before shipment to the BIPM:

About 55 bar

If any other component other than NO<sub>2</sub>, nitrogen and oxygen was detected and/or quantified please report its mole fraction in the table below:

| Component | Mole fraction / nmol/mol | Expanded uncertainty | Coverage factor | Measurement technique |
|-----------|--------------------------|----------------------|-----------------|-----------------------|
|           |                          |                      |                 |                       |
|           |                          |                      |                 |                       |
|           |                          |                      |                 |                       |

## Laboratoire National de métrologie et d'Essais (LNE)

### A1. General information

|                                    |  |     |                  |
|------------------------------------|--|-----|------------------|
| Institute                          | Laboratoire National de métrologie et d'Essais (LNE) |     |                  |
| Address                            | 1, rue Gaston Boissier<br>75 724 Paris Cedex 15      |     |                  |
| Contact person                     | Tatiana Macé   |     |                  |
| Telephone                          | 33 1 40 43 38 53                                     | Fax | 33 1 40 43 37 37 |
| Email*                             | tatiana.mace@lne.fr                                  |     |                  |
| Serial number of cylinder received | D650058  |     |                  |
| Cylinder pressure as received      | 95 bars  |     |                  |

### A2. Results

| Nitrogen dioxide mole fraction          | Expanded uncertainty                       | Coverage factor |
|---|--|-----------------|
| $x_{\text{NO}_2}$ / $\mu\text{mol/mol}$ | $U(x_{\text{NO}_2})$ / $\mu\text{mol/mol}$ |                 |
| 10,26                                   | 0,13                                       | k=2             |

### A3. Uncertainty Budget

Please provide a complete uncertainty budget.

First step :

The first step consists in the estimation of the standard uncertainty on each diluted value.

An example of an uncertainty budget on one of the 9 obtained values is given in the following table.

| Uncertainty source                                   | $x_i$<br>(nmol/mol) | Assumed distribution | Standard uncertainty<br>$u(x_i)$<br>(nmol/mol) | Sensitivity coefficient<br>$c_i$ | Contribution to standard uncertainty<br>$u_i(y)$ |
|--|---------------------|----------------------|--|----------------------------------|--|
| Concentration of the reference gas mixture generated | 299.8               | -                    | 0.9578   | 0.993                            | 0.9511   |

|   |     |             |       |       |        |
|---|-----|-------------|-------|-------|--------|
| by permeation ( $C_1$ )   |     |             |       |       |        |
| Reading for the reference gas mixture generated by permeation ( $L_1$ ) | 304 | rectangular | 0.577 | 0.980 | 0.5655 |
| Reading for the diluted unknown gas mixture ( $L_2$ )                   | 302 | rectangular | 0.577 | 0.986 | 0.5690 |

|  |                    |                  |
|--|--------------------|------------------|
| Concentration of the diluted unknown gas mixture ( $C_2$ ) | 297.84<br>nmol/mol | 1.24<br>nmol/mol |
|--|--------------------|------------------|

Then, the standard uncertainty is calculated for each concentration of the unknown gas mixture  $C_2$  as described in the following example.

| Uncertainty source  | $x_i$<br>(nmol/mol) | Assumed<br>distribution | Standard<br>uncertainty<br>$u(x_i)$ | Sensitivity<br>coefficient<br>$c_i$ | Contribution<br>to standard<br>uncertainty<br>$u_i(y)$ |
|---|---------------------|-------------------------|-------------------------------------|-------------------------------------|--|
| Concentration of the diluted unknown gas mixture ( $C_2$ )                            | 297.84              | -                       | 1.24                                | 34.40                               | 42.66  |
| Flowrate of the $\text{NO}_2$ unknown gas mixture ( $\text{NO}_2$ cylinder) ( $D_1$ ) | 75.126              | -                       | 0.19                                | 132.43                              | 25.16  |
| Flowrate of the dilution gas (nitrogen) ( $D_2$ )                                     | 2509.40             | -                       | 6.27                                | 3.96                                | 24.83  |

|   |                      |                  |
|---|----------------------|------------------|
| Concentration of the unknown gas mixture ( $C_2'$ ) | 10246.43<br>nmol/mol | 55.5<br>nmol/mol |
|---|----------------------|------------------|

The standard uncertainties obtained for the 9 values are summarized in the following table.

| Date       | Concentrations of the unknown gas mixture ( $C_2'$ ) (nmol/mol) | $u(C_2')$<br>(nmol/mol) |
|------------|---|-------------------------|
| 13/10/2009 | 10210.967   | 54.2                    |
| 13/10/2009 | 10240.679   | 54.0                    |
| 13/10/2009 | 10241.613   | 54.0                    |
| 16/10/2009 | 10295.842   | 54.7                    |
| 16/10/2009 | 10268.717   | 55.0                    |
| 16/10/2009 | 10307.627   | 55.1                    |
| 22/10/2009 | 10246.432   | 55.5                    |
| 22/10/2009 | 10249.799   | 55.8                    |
| 22/10/2009 | 10281.519   | 55.8                    |

The mean standard uncertainty is calculated as follows:

$$u_{\text{mean}} = \sqrt{\frac{\sum u^2(C_2')}{n}} = 54.91 \text{ nmol / mol}$$

Second step :

The second step in the calculation consists of the standard deviation on the mean of the 9 obtained values.

$$\sigma = 30.61 \text{ nmol / mol}$$



Third step :

The third step in the calculation consists of the expanded uncertainty on the mean concentration of the unknown gas mixture as follows.

$$U(x_{NO_2}) = 2 \times \sqrt{u_{mean}^2 + \sigma^2} = 0.13 \mu\text{mol} / \text{mol}$$

**A4. Description of the procedure used during the gas analysis**

Please describe in detail the analytical method(s) used for gas analysis<sup>12</sup>.

Reference Instrument :

A 42C (TEI) analyser based on the principle of chemiluminescence was used to measure the NO<sub>2</sub> concentrations.

Calibration Standard :

A reference dynamic gas mixture of NO<sub>2</sub> in nitrogen (at about 300 nmol/mol) was generated by the LNE reference method which is the permeation method.

Description of the analytical procedure :

A reference gas mixture was generated by permeation at a concentration (C<sub>1</sub> near 300 nmol/mol) slightly higher than the concentration of the unknown gas mixture and injected into the analyser : the response of the analyser was recorded (L<sub>1</sub>).

The unknown gas mixture at about 10 μmol/mol was diluted to about 300 nmol/mol : this dynamic unknown gas mixture was then injected into the analyser and the response (L<sub>2</sub>) was recorded.

The NO<sub>2</sub> concentration of the diluted unknown gas mixture (C<sub>2</sub>) was equal to :

$$C_2 = \frac{C_1 \times L_2}{L_1}$$

The NO<sub>2</sub> concentration of the unknown gas mixture C<sub>2</sub>' was :

---

<sup>12</sup> The choice of the procedure used for gas analysis is the responsibility of the participating laboratory. Nevertheless, for a proper evaluation of the data, it is necessary that the calibration method, as well as the way in which the calibration mixtures have been prepared is reported to the co-ordinators.

$$C'_2 = \frac{C_2 \times (D_1 + D_2)}{D_1}$$

With :  $C_2$  the concentration of the diluted unknown gas mixture  
 $D_1$  the flowrate of the NO<sub>2</sub> unknown gas mixture (NO<sub>2</sub> cylinder)  
 $D_2$  the flowrate of the dilution gas (nitrogen)

This procedure was carried out 3 times on 3 different days.

The NO<sub>2</sub> concentration is the mean of the 9 obtained values.

#### A5. Complementary information on the cylinder

Please report the value of the pressure left in the cylinder before shipment to the BIPM:

The pressure left in the cylinder before shipment to the BIPM was about 80 bar.

If any other component other than NO<sub>2</sub>, nitrogen and oxygen was detected and/or quantified please report its mole fraction in the table below:

| Component | Mole fraction / nmol/mol | Expanded uncertainty | Coverage factor | Measurement technique |
|-----------|--------------------------|----------------------|-----------------|-----------------------|
|           |                          |                      |                 |                       |
|           |                          |                      |                 |                       |
|           |                          |                      |                 |                       |

## Federal Office of Metrology (METAS)

### A1. General information

|                                    |  |     |                  |
|------------------------------------|--|-----|------------------|
| Institute                          | METAS  |     |                  |
| Address                            | <b>Federal Office of Metrology METAS</b><br>Gas Analysis Laboratory<br>Lindenweg 50, 3003 Bern-Wabern, Switzerland |     |                  |
| Contact person                     | Cedric Couret  |     |                  |
| Telephone                          | +41 31 32 33 381   | Fax | +41 31 32 33 210 |
| Email*                             | cedric.couret@metas.ch   |     |                  |
| Serial number of cylinder received | 930660   |     |                  |
| Cylinder pressure as received      | 96 bar   |     |                  |

### A2. Results

Measurement #1

| Nitrogen dioxide mole fraction        | Expanded uncertainty                     | Coverage factor |
|---------------------------------------|--|-----------------|
| $x_{\text{NO}_2} / \mu\text{mol/mol}$ | $U(x_{\text{NO}_2}) / \mu\text{mol/mol}$ |                 |
| 10.63                                 | 0.17                                     | 2               |

## Measurement #2

| Nitrogen dioxide mole fraction          | Expanded uncertainty                       | Coverage factor |
|---|--|-----------------|
| $x_{\text{NO}_2}$ / $\mu\text{mol/mol}$ | $U(x_{\text{NO}_2})$ / $\mu\text{mol/mol}$ |                 |
| 10.62                                   | 0.17                                       | 2               |

## Measurement #3

| Nitrogen dioxide mole fraction          | Expanded uncertainty                       | Coverage factor |
|---|--|-----------------|
| $x_{\text{NO}_2}$ / $\mu\text{mol/mol}$ | $U(x_{\text{NO}_2})$ / $\mu\text{mol/mol}$ |                 |
| 10.63                                   | 0.16                                       | 2               |

## Result

| Nitrogen dioxide mole fraction          | Expanded uncertainty                       | Coverage factor |
|---|--|-----------------|
| $x_{\text{NO}_2}$ / $\mu\text{mol/mol}$ | $U(x_{\text{NO}_2})$ / $\mu\text{mol/mol}$ |                 |
| 10.63                                   | 0.16                                       | 2               |

**A3. Uncertainty Budget**

Please provide a complete uncertainty budget.

**Model Equation for measurement 2 :**

$$V_{\text{norm}1} = V_{\text{mol}1} * f_{5000};$$

$$V_{\text{norm}2} = V_{\text{mol}2} * f_{5000};$$

$$V_{\text{norm}3} = V_{\text{mol}3} * f_{5000};$$

$$V_{\text{mfc}mean} = (V_{\text{mfc}1} + V_{\text{mfc}2} + V_{\text{mfc}3}) / 3;$$

$$p = (V_{\text{norm}1} - V_{\text{norm}mean}) * (V_{\text{mfc}1} - V_{\text{mfc}mean}) + (V_{\text{norm}2} - V_{\text{norm}mean}) * (V_{\text{mfc}2} - V_{\text{mfc}mean}) + (V_{\text{norm}3} - V_{\text{norm}mean}) * (V_{\text{mfc}3} - V_{\text{mfc}mean});$$

$$q = (V_{\text{norm}1} - V_{\text{norm}mean})^2 + (V_{\text{norm}2} - V_{\text{norm}mean})^2 + (V_{\text{norm}3} - V_{\text{norm}mean})^2;$$

$$b = p/q;$$

$$a = V_{mfc} \text{mean} - b * V_{norm} \text{mean};$$

$$V_{norm} \text{bouteille} = (V_{mfc} \text{bouteille} - a) / b;$$

$$V_{dilution} \text{norm} = V_{dilution} * f_{5000} \text{bis};$$

$$X_{ppbA1} = ((qmC * 486.9)) / (qv_1 * f_{10000});$$

$$X_{ppbA2} = ((qmC * 486.9)) / (qv_2 * f_{10000});$$

$$X_{ppbA3} = ((qmC * 486.9)) / (qv_3 * f_{10000});$$

$$X_{ppb} \text{mean} A = (X_{ppbA1} + X_{ppbA2} + X_{ppbA3}) / 3 - X_{NOx};$$

$$\text{Anz}_{\text{mean}A} = (\text{Anz}_{ppbA1} + \text{Anz}_{ppbA2} + \text{Anz}_{ppbA3}) / 3;$$

$$p2 = (X_{ppbA1} - X_{ppb} \text{mean} A) * (\text{Anz}_{ppbA1} - X_{NOx} - \text{Anz}_{\text{mean}A}) + (X_{ppbA2} - X_{ppb} \text{mean} A) * (\text{Anz}_{ppbA2} - X_{NOx} - \text{Anz}_{\text{mean}A}) + (X_{ppbA3} - X_{ppb} \text{mean} A) * (\text{Anz}_{ppbA3} - X_{NOx} - \text{Anz}_{\text{mean}A});$$

$$q2 = (X_{ppbA1} - X_{ppb} \text{mean} A)^2 + (X_{ppbA2} - X_{ppb} \text{mean} A)^2 + (X_{ppbA3} - X_{ppb} \text{mean} A)^2;$$

$$b2 = p2 / q2;$$

$$a2 = \text{Anz}_{\text{mean}A} - b2 * X_{ppb} \text{mean} A;$$

$$C_{lu} = (\text{Anz}_{Res} - X_{NO2} - a2) / b2$$

$$C_{bouteille} = C_{lu} * V_{norm} \text{bouteille} / (V_{norm} \text{bouteille} - V_{dilution} \text{norm});$$

### List of Quantities:

| Quantity                    | Unit     | Definition   |
|-----------------------------|----------|--|
| $V_{norm} \text{mean}$      | ml/min   | Average flow normal  |
| $V_{norm} 1$                | ml/min   | Flow 1 normal  |
| $V_{norm} 2$                | ml/min   | Flow 2 normal  |
| $V_{norm} 3$                | ml/min   | Flow 3 normal  |
| $V_{mol} 1$                 | ml/min   | Flow 1 molbloc   |
| $f_{5000}$                  |          | Correction factor for flow calibration 5000 ml/min molbloc |
| $V_{mol} 2$                 | ml/min   | Flow 2 molbloc   |
| $V_{mol} 3$                 | ml/min   | Flow 3 molbloc   |
| $V_{mfc} \text{mean}$       | ml/min   | Average flow Mass flow meter                               |
| $V_{mfc} 1$                 | ml/min   | Flow 1 mass flow meter overflow                            |
| $V_{mfc} 2$                 | ml/min   | Flow 2 mass flow meter overflow                            |
| $V_{mfc} 3$                 | ml/min   | Flow 3 mass flow meter overflow                            |
| b                           |          | Slope of calibration function for flow                     |
| p                           |          | Numerator for slope of flow calibration function           |
| q                           |          | Denominator for slope of flow calibration function         |
| a                           |          | Inter of calibration function for flow                     |
| $V_{norm} \text{bouteille}$ | ml/min   | Test mixture flow normal                                   |
| $V_{mfc} \text{bouteille}$  | ml/min   | Test mixture flow mass flow meter                          |
| $V_{dilution} \text{norm}$  | ml/min   | Dilution flow normal                                       |
| $V_{dilution}$              | ml/min   | Dilution flow  |
| $f_{5000} \text{bis}$       |          | Correction factor for flow calibration 5000 ml/min molbloc |
| $C_{bouteille}$             | nmol/mol | Nitrogen dioxide mole fraction in the test mixture         |
| $C_{lu}$                    | nmol/mol | Nitrogen dioxide mole fraction diluted                     |
| $X_{ppbA1}$                 | nmol/mol | Nitrogen dioxide mole fraction balance 1                   |
| qmC                         | ng/min   | Mass flow NO <sub>2</sub> permeation unit                  |

| Quantity        | Unit     | Definition  |
|-----------------|----------|---|
| $q_{v1}$        | ml/min   | Flow balance 1  |
| $f_{10000}$     |          | Correction factor for flow calibration 10000 ml/min molbloc |
| $X_{ppbA2}$     | nmol/mol | Nitrogen dioxide mole fraction balance 2                    |
| $q_{v2}$        | ml/min   | Flow balance 2  |
| $X_{ppbA3}$     | nmol/mol | Nitrogen dioxide mole fraction balance 3                    |
| $q_{v3}$        | ml/min   | Flow balance 3  |
| $X_{ppb,meanA}$ | nmol/mol | Nitrogen dioxide mole fraction balance average              |
| $X_{NOx}$       | nmol/mol | Nitrogen dioxide mole fraction air zero balance             |
| $Anz_{meanA}$   |          | Display instrument average                                  |
| $Anz_{ppbA1}$   |          | Display instrument 1  |
| $Anz_{ppbA2}$   |          | Display instrument 2  |
| $Anz_{ppbA3}$   |          | Display instrument 3  |
| $b_2$           |          | Slope of calibration function for instrument indication     |
| $p_2$           |          | Numerator for slope of indication calibration function      |
| $q_2$           |          | Denominator for slope of indication calibration function    |
| $a_2$           |          | Intercept of calibration function for instrument indication |
| $AnzRes$        |          | Display test mixture  |
| $X_{NO2}$       | nmol/mol | Nitrogen dioxide mole fraction dilution                     |

|   |   |
|---|---|
| <b>V<sub>mol1</sub>:</b><br>Type B normal distribution<br>Value: 1140.840929 ml/min<br>Expanded Uncertainty: 0.308797961 ml/min<br>Coverage Factor: 2 | <b>f<sub>5000</sub>:</b><br>Type B normal distribution<br>Value: 1<br>Expanded Uncertainty: 0.15 %<br>Coverage Factor: 2                                      |
| <b>V<sub>mfc1</sub>:</b><br>Type B normal distribution<br>Value: 99.10032386 ml/min<br>Expanded Uncertainty: 0.136235451 ml/min<br>Coverage Factor: 2 | <b>V<sub>mfc</sub>bouteille:</b><br>Type B normal distribution<br>Value: 99.45063673 ml/min<br>Expanded Uncertainty: 0.137777628 ml/min<br>Coverage Factor: 2 |
| <b>V<sub>dilution</sub>:</b><br>Type B normal distribution<br>Value: 1101.371378 ml/min<br>Expanded Uncertainty: 0.30 ml/min<br>Coverage Factor: 2    | <b>f<sub>5000bis</sub>:</b><br>Type B normal distribution<br>Value: 1<br>Expanded Uncertainty: 0.15 %<br>Coverage Factor: 2                                   |
| <b>qmC:</b><br>Type B normal distribution<br>Value: 1035.46 ng/min<br>Expanded Uncertainty: 1 %<br>Coverage Factor: 2                                 | <b>qv<sub>1</sub>:</b><br>Type B normal distribution<br>Value: 1399.46341 ml/min<br>Expanded Uncertainty: 0.35224 ml/min<br>Coverage Factor: 2                |
| <b>f<sub>10000</sub>:</b><br>Type B normal distribution<br>Value: 1<br>Expanded Uncertainty: 0.15 %<br>Coverage Factor: 2                             | <b>X<sub>NOx</sub>:</b><br>Type B normal distribution<br>Value: 0.13 nmol/mol<br>Expanded Uncertainty: 0.02 nmol/mol<br>Coverage Factor: 2                    |
| <b>Anz<sub>ppbA1</sub>:</b><br>Type B normal distribution<br>Value: 363.31<br>Expanded Uncertainty: 0.35<br>Coverage Factor: 2                        | <b>AnzRes:</b><br>Type B normal distribution<br>Value: 377.21<br>Expanded Uncertainty: 0.26<br>Coverage Factor: 2   |
| <b>X<sub>NO2</sub>:</b><br>Type B normal distribution<br>Value: 0.1107 nmol/mol<br>Expanded Uncertainty: 0.02 nmol/mol<br>Coverage Factor: 2          |   |

## Input Correlation

|                           | f <sub>5000</sub> | f <sub>5000bis</sub> | f <sub>10000</sub> | V <sub>mol1</sub> | V <sub>mol2</sub> | V <sub>mol3</sub> | V <sub>mfc1</sub> | V <sub>mfc2</sub> | V <sub>mfc3</sub> | V <sub>mfcbouteille</sub> | V <sub>dilution</sub> | qmC | qv <sub>1</sub> | qv <sub>2</sub> | qv <sub>3</sub> | Anz <sub>ppbA1</sub> | Anz <sub>ppbA2</sub> | Anz <sub>ppbA3</sub> | AnzRes |
|---------------------------|-------------------|----------------------|--------------------|-------------------|-------------------|-------------------|-------------------|-------------------|-------------------|---------------------------|-----------------------|-----|-----------------|-----------------|-----------------|----------------------|----------------------|----------------------|--------|
| f <sub>5000</sub>         | 1                 | 0.95                 | 0.9                |                   |                   |                   |                   |                   |                   |                           |                       |     |                 |                 |                 |                      |                      |                      |        |
| f <sub>5000bis</sub>      | 0.95              | 1                    | 0.9                |                   |                   |                   |                   |                   |                   |                           |                       |     |                 |                 |                 |                      |                      |                      |        |
| f <sub>10000</sub>        | 0.9               | 0.9                  | 1                  |                   |                   |                   |                   |                   |                   |                           |                       |     |                 |                 |                 |                      |                      |                      |        |
| V <sub>mol1</sub>         |                   |                      |                    | 1                 | 0.9               | 0.9               | 0.9               | 0.9               | 0.9               | 0.9                       | 0.9                   |     | 0.5             | 0.5             | 0.5             |                      |                      |                      |        |
| V <sub>mol2</sub>         |                   |                      |                    | 0.9               | 1                 | 0.9               | 0.9               | 0.9               | 0.9               | 0.9                       | 0.9                   |     | 0.5             | 0.5             | 0.5             |                      |                      |                      |        |
| V <sub>mol3</sub>         |                   |                      |                    | 0.9               | 0.9               | 1                 | 0.9               | 0.9               | 0.9               | 0.9                       | 0.9                   |     | 0.5             | 0.5             | 0.5             |                      |                      |                      |        |
| V <sub>mfc1</sub>         |                   |                      |                    | 0.9               | 0.9               | 0.9               | 1                 | 0.9               | 0.9               | 0.9                       | 0.9                   |     | 0.5             | 0.5             | 0.5             |                      |                      |                      |        |
| V <sub>mfc2</sub>         |                   |                      |                    | 0.9               | 0.9               | 0.9               | 0.9               | 1                 | 0.9               | 0.9                       | 0.9                   |     | 0.5             | 0.5             | 0.5             |                      |                      |                      |        |
| V <sub>mfc3</sub>         |                   |                      |                    | 0.9               | 0.9               | 0.9               | 0.9               | 0.9               | 1                 | 0.9                       | 0.9                   |     | 0.5             | 0.5             | 0.5             |                      |                      |                      |        |
| V <sub>mfcbouteille</sub> |                   |                      |                    | 0.9               | 0.9               | 0.9               | 0.9               | 0.9               | 0.9               | 1                         | 0.9                   |     | 0.5             | 0.5             | 0.5             |                      |                      |                      |        |
| V <sub>dilution</sub>     |                   |                      |                    | 0.9               | 0.9               | 0.9               | 0.9               | 0.9               | 0.9               | 0.9                       | 1                     |     | 0.5             | 0.5             | 0.5             |                      |                      |                      | -1     |
| qmC                       |                   |                      |                    |                   |                   |                   |                   |                   |                   |                           |                       | 1   |                 |                 |                 | 1                    | 1                    | 1                    |        |
| qv <sub>1</sub>           |                   |                      |                    | 0.5               | 0.5               | 0.5               | 0.5               | 0.5               | 0.5               | 0.5                       | 0.5                   |     | 1               |                 |                 | -0.9                 |                      |                      |        |
| qv <sub>2</sub>           |                   |                      |                    | 0.5               | 0.5               | 0.5               | 0.5               | 0.5               | 0.5               | 0.5                       | 0.5                   |     |                 | 1               |                 |                      | -0.9                 |                      |        |
| qv <sub>3</sub>           |                   |                      |                    | 0.5               | 0.5               | 0.5               | 0.5               | 0.5               | 0.5               | 0.5                       | 0.5                   |     |                 |                 | 1               |                      |                      | -0.9                 |        |
| Anz <sub>ppbA1</sub>      |                   |                      |                    |                   |                   |                   |                   |                   |                   |                           |                       | 1   | -0.9            |                 |                 | 1                    |                      |                      |        |
| Anz <sub>ppbA2</sub>      |                   |                      |                    |                   |                   |                   |                   |                   |                   |                           |                       | 1   |                 | -0.9            |                 |                      | 1                    |                      |        |
| Anz <sub>ppbA3</sub>      |                   |                      |                    |                   |                   |                   |                   |                   |                   |                           |                       | 1   |                 |                 | -0.9            |                      |                      | 1                    |        |



**Example Uncertainty Budgets (measurement #2):**C<sub>bouteille</sub>: Nitrogen dioxide mole fraction in the test mixture

| Quantity                   | Value            | Standard Uncertainty | Distribution | Sensitivity Coefficient | Uncertainty Contribution | Index  |
|----------------------------|------------------|----------------------|--------------|-------------------------|--------------------------|--------|
| V <sub>mol1</sub>          | 1140.841 ml/min  | 0.154 ml/min         | normal       | -180                    | -28 nmol/mol             | 1.1 %  |
| f <sub>5000</sub>          | 1.000000         | 0.000750             | normal       | -290·10 <sup>3</sup>    | -220 nmol/mol            | 54.4 % |
| V <sub>mol2</sub>          | 1143.759 ml/min  | 0.141 ml/min         | normal       | -87                     | -12 nmol/mol             | 0.2 %  |
| V <sub>mol3</sub>          | 1146.875 ml/min  | 0.145 ml/min         | normal       | not valid!              | 1.8 nmol/mol             | 0.0 %  |
| V <sub>mfc1</sub>          | 99.1003 ml/min   | 0.0681 ml/min        | normal       | 360                     | 25 nmol/mol              | 0.8 %  |
| V <sub>mfc2</sub>          | 100.5622 ml/min  | 0.0672 ml/min        | normal       | 170                     | 12 nmol/mol              | 0.2 %  |
| V <sub>mfc3</sub>          | 102.1276 ml/min  | 0.0721 ml/min        | normal       | -25                     | -1.8 nmol/mol            | 0.0 %  |
| V <sub>normbouteille</sub> | 1141.540 ml/min  | 0.871 ml/min         |              |                         |                          |        |
| V <sub>mfcbouteille</sub>  | 99.4506 ml/min   | 0.0689 ml/min        | normal       | -510                    | -35 nmol/mol             | 1.7 %  |
| V <sub>dilutionnorm</sub>  | 1101.371 ml/min  | 0.840 ml/min         |              |                         |                          |        |
| V <sub>dilution</sub>      | 1101.371 ml/min  | 0.150 ml/min         | normal       | 260                     | 40 nmol/mol              | 0.2 %  |
| f <sub>5000bis</sub>       | 1.000000         | 0.000750             | normal       | 290·10 <sup>3</sup>     | 220 nmol/mol             | 11.2 % |
| qmC                        | 1035.46 ng/min   | 5.18 ng/min          | normal       | 10                      | 53 nmol/mol              | 34.9 % |
| qv <sub>1</sub>            | 1399.463 ml/min  | 0.176 ml/min         | normal       | -3.6                    | -0.64 nmol/mol           | 0.0 %  |
| f <sub>10000</sub>         | 1.000000         | 0.000750             | normal       | -11000                  | -8.0 nmol/mol            | 0.9 %  |
| qv <sub>2</sub>            | 1322.228 ml/min  | 0.174 ml/min         | normal       | -2.6                    | -0.45 nmol/mol           | 0.0 %  |
| qv <sub>3</sub>            | 1271.393 ml/min  | 0.199 ml/min         | normal       | -1.7                    | -0.33 nmol/mol           | 0.0 %  |
| X <sub>NOx</sub>           | 0.1300 nmol/mol  | 0.0100 nmol/mol      | normal       | -29                     | -0.29 nmol/mol           | 0.0 %  |
| Anz <sub>ppbA1</sub>       | 363.310          | 0.175                | normal       | -14                     | -2.4 nmol/mol            | -1.7 % |
| Anz <sub>ppbA2</sub>       | 384.820          | 0.225                | normal       | -8.8                    | -2.0 nmol/mol            | -1.4 % |
| Anz <sub>ppbA3</sub>       | 400.370          | 0.170                | normal       | -5.2                    | -0.89 nmol/mol           | -0.6 % |
| AnzRes                     | 377.210          | 0.130                | normal       | 28                      | 3.6 nmol/mol             | -1.8 % |
| X <sub>NO2</sub>           | 0.1107 nmol/mol  | 0.0100 nmol/mol      | normal       | -28                     | -0.28 nmol/mol           | 0.0 %  |
| C <sub>bouteille</sub>     | 10617.7 nmol/mol | 85.3 nmol/mol        |              |                         |                          |        |

The uncertainty budget has been calculated using GUM Workbench Pro software (version 2.3.2.36 beta). The main contributions to the combined standard uncertainty are:


- the standard uncertainty of the NO<sub>2</sub> mass flow from the permeation unit
- the standard uncertainties of the flow measurements

#### A4. Description of the procedure used during the gas analysis

Please describe in detail the analytical method(s) used for gas analysis<sup>13</sup>.

A commercial chemiluminescence trace level NO<sub>2</sub>-analyser (Thermo 42i -TL) was calibrated with NO<sub>2</sub> calibration standards in the range from 360 nmol·mol<sup>-1</sup> to 396 nmol·mol<sup>-1</sup> NO<sub>2</sub> in air. The calibration standards were produced by the METAS primary micro gravimetric standard and a NO<sub>2</sub> permeation unit with purity ≥ 99 %. The total zero air dilution gas flow was measured by a molbox-molbloc system. The NO<sub>2</sub> mass flow of the permeation unit was approx. 1035 ng·min<sup>-1</sup> at 25 °C. This value is an average over 5 days and was used for the 3 calibrations.


The test mixture 930660 was dynamically diluted with zero air by a factor of about 0.036 such that the expected amount of substance fraction of the sample lies within the validated and calibrated range of the analyser and of the METAS micro-gravimetric Standard. The flow of the test mixture was set using a capillary column. The upstream pressure at the capillary column was kept constant by a pressure regulator Parker Veriflo (Hastelloy C-22 alloy).



The flow of the test mixture was calculated by subtracting the dilution gas flow from the total flow going to the instrument. The flow of the dilution gas was regulated with a mass flow controller and measured using a molbox-molbloc system. The total flow was calculated by measuring the “over-flow” of the instrument with a mass flow meter (MFM) as shown in the following scheme.

---

<sup>13</sup> The choice of the procedure used for gas analysis is the responsibility of the participating laboratory. Nevertheless, for a proper evaluation of the data, it is necessary that the calibration method, as well as the way in which the calibration mixtures have been prepared is reported to the co-ordinators.



The linear interpolation was determined experimentally by changing the total flow such that the overflow during the test mixture analysis lies within calibrated range of the mass flow meter.

The resulting gas mixture was measured with the NO<sub>2</sub>-analyser and the amount of substance fraction calculated by linear interpolation in agreement with ISO 6143:2001(E).

The molbox-molbloc system was calibrated with the METAS primary standard for low gas flows.

**A5. Complementary information on the cylinder**

Please report the value of the pressure left in the cylinder before shipment to the BIPM: 78 bar

If any component other than NO<sub>2</sub>, nitrogen and oxygen was detected and/or quantified please report its mole fraction in the table below:

| Component | Mole fraction / nmol/mol | Expanded uncertainty | Coverage factor | Measurement technique |
|-----------|--------------------------|----------------------|-----------------|-----------------------|
| NO        | <1ppb                    |                      |                 | Chemiluminescence     |
|           |                          |                      |                 |                       |
|           |                          |                      |                 |                       |

## National Measurement Institute Australia (NMIA)

### A1. General information

|                                    |   |     |                 |
|------------------------------------|---|-----|-----------------|
| Institute                          | NMIA – National Measurement Institute Australia |     |                 |
| Address                            | Bradfield Road<br>Lindfield NSW 2070 AUSTRALIA  |     |                 |
| Contact person                     | Damian Smeulders                                |     |                 |
| Telephone                          | +61 2 84673534                                  | Fax | +61 2 8467 3752 |
| Email*                             | damian.smeulders@measurement.gov.au             |     |                 |
| Serial number of cylinder received | 930662  |     |                 |
| Cylinder pressure as received      | 100 bar   |     |                 |

### A2. Results

| Nitrogen dioxide mole fraction<br>$x_{\text{NO}_2}$ / $\mu\text{mol/mol}$ | Expanded uncertainty<br>$U(x_{\text{NO}_2})$ / $\mu\text{mol/mol}$ | Coverage factor |
|---|--|-----------------|
| 10.74   | 0.63   | 2               |

### A3. Uncertainty Budget

Combined standard uncertainty:  $u = 0.32 \mu\text{mol/mol}$

Expanded uncertainty:  $U = 0.63 \mu\text{mol/mol}$

Contributions to uncertainty:

Gravimetric uncertainty: 0.018

Mixture stability and conversion to  $\text{NO}_2$ : 0.075

Instrument contributions:

Repeatability: 0.30

Resolution: 0.020

Difference due to spectral regions: 0.050

### A4. Description of the procedure used during the gas analysis

A Nicolet FT-IR was used to acquire the spectra of the standards and unknown sample. The spectra were run at resolutions of  $0.5\text{cm}^{-1}$  and  $0.25\text{cm}^{-1}$  with an aperture setting of 2. 100 scans were obtained for each analysis. The background spectra were collected on the evacuated cell. Spectra were collected on a static gas sample with a temperature of  $60^\circ\text{C}$  at a pressure of 650 Torr.

The strong bands in the region  $1530\text{ cm}^{-1}$  to  $1670\text{ cm}^{-1}$  and the weaker bands in the region  $2840\text{ cm}^{-1}$  to  $2940\text{ cm}^{-1}$  were both used for quantitation. The analyses of the standards and sample were repeated three times at each resolution with evacuation and flushing of the cell between tests. The analysis procedure was repeated on several occasions over a two week period.

Four closely bracketed calibration standards containing  $\text{NO}_2$  over the concentration range  $8\text{ }\mu\text{mol/mol}$  to  $12\text{ }\mu\text{mol/mol}$  were used to determine the concentration of  $\text{NO}_2$  in the cylinder from the BIPM. Standards were made in uncoated, but passivated 5L Luxfer aluminium cylinders with SS valves. Standards were manufactured from nitrogen oxide that was converted to nitrogen dioxide in the presence of oxygen. Oxygen in the final mixtures was present at approximately  $1000\text{ }\mu\text{mol/mol}$ .

#### **A5. Complementary information on the cylinder**

Please report the value of the pressure left in the cylinder before shipment to the BIPM: 78 bar

If any other component other than  $\text{NO}_2$ , nitrogen and oxygen was detected and/or quantified please report its mole fraction in the table below:

| Component | Mole fraction / nmol/mol | Expanded uncertainty | Coverage factor | Measurement technique |
|-----------|--------------------------|----------------------|-----------------|-----------------------|
|           |                          |                      |                 |                       |
|           |                          |                      |                 |                       |
|           |                          |                      |                 |                       |

## National Institute of Metrology (NIM)

### Lab Information

Lab Code: 58

Lab Name: National Institute of Metrology (NIM), China

Contact point: Dr. Qiao HAN

Email: [hanqiao@nim.ac.cn](mailto:hanqiao@nim.ac.cn)

Tel.: +86-10-84252300 Fax.: +86-10-84252306

Date of Receiving the Comparison Cylinder: November, 2009

Cylinder No.: 930650

### Measurement of NO<sub>2</sub> by 42C NO-NO<sub>2</sub>-NO<sub>x</sub> Analyzer

| Group # | Date<br>(dd/mm/yy) | Result<br>(µmol/mol) | Standard deviation<br>(% relative) | Number of<br>replicates |
|---------|--------------------|----------------------|------------------------------------|-------------------------|
| 1       | 26/01/10           | 10.14                | 0.13 %                             | 6                       |
| 2       | 28/01/10           | 10.12                | 0.40 %                             | 4                       |
| 3       | 01/02/10           | 10.18                | 0.27 %                             | 6                       |
| 4       | 02/02/10           | 10.17                | 0.17 %                             | 4                       |
| 5       | 03/02/10           | 10.14                | 0.45 %                             | 3                       |

### Measurement of NO<sub>2</sub> by FT-IR

| Group # | Date<br>(dd/mm/yy) | Result<br>(µmol/mol) | Standard deviation<br>(% relative) | Number of<br>replicates |
|---------|--------------------|----------------------|------------------------------------|-------------------------|
| 1       | 27/01/10           | 10.14                | 0.55 %                             | 4                       |

### Result

| Component       | Result         | Expanded Uncertainty | Coverage factor |
|-----------------|----------------|----------------------|-----------------|
| NO <sub>2</sub> | 10.15 µmol/mol | 0.10 µmol/mol        | 2*              |

\*The coverage factor was based on approximately 95 % confidence.

## Method description forms

### Reference Method:

NO<sub>2</sub> was analysed by 42C NO-NO<sub>2</sub>-NO<sub>x</sub> Analyser (Thermo Environmental Instruments made in the USA), which is based on Chemiluminescence principle, with a measurement range from 0 ppm to 100 ppm.

Simultaneously, a FT-IR (Nicolet 5700) with a changeable gas sample cell from Thermo Co was used to compare and check the results from 42C NO-NO<sub>2</sub>-NO<sub>x</sub> Analyser.

In the FT-IR system, the detector is MCT/A (mercury cadmium tellurium). Minimum resolution is  $0.125\text{cm}^{-1}$ . Beamsplitter is KBr. Optical path length of sample cell is a multi-reflection system changeable from 4m~48m.

The pressure in the cell was controlled at about 50 kPa by the gas inlet system when 10 ppm  $\text{NO}_2$  is analysed. The temperature in the gas cell was about  $20\text{ }^\circ\text{C}$ . The optical path length is fixed at about 48m, and the resolution is selected at  $4\text{cm}^{-1}$  with a 256 times scanning.

### Calibration Standards:

All of the references we used were prepared by the gravimetric method according to ISO 6142-2001 in our lab.

The pure gases were checked before using to make sure that their purities were good enough and the impurities had no effect on the quality of reference gas mixtures. The pure gases included  $\text{N}_2$ ,  $\text{O}_2$  and  $\text{NO}$ . The  $\text{NO}_2$  reference gas mixtures came from the gas mixture of  $\text{NO}$  in nitrogen with some  $\text{O}_2$  added into it. The concentration of  $\text{O}_2$  was about 1000 ppm in the  $\text{NO}_2$  reference gas mixtures.

The parent gases were filled into a 4 litre or 5 litre aluminium cylinder, which received a special treatment. More than 10g parent gas was filled into the cylinder at least. The cylinder was weighed before and after the filling using a balance with a sensitivity of 1 mg.

The concentration of reference gas was calculated according to the following equation.

$$x_i = \frac{\sum_{A=1}^P \left( \frac{x_{i,A} \cdot m_A}{\sum_{i=1}^n x_{i,A} \cdot M_i} \right)}{\sum_{A=1}^P \left( \frac{m_A}{\sum_{i=1}^n x_{i,A} \cdot M_i} \right)}$$

The uncertainty of reference gas included the contributions from the gravimetric method and from stability. The uncertainty from stability was evaluated based on short-time and long-time testing. The uncertainty from gravimetric method was calculated according to the following equation.

$$u^2(x_i) = \sum_{r=1}^q \left( \frac{\partial f_i}{\partial f_r} \right)^2 \cdot u^2(y_r) + 2 \sum_{r=1}^{q-1} \sum_{s=r+1}^q \left( \frac{\partial f_i}{\partial f_r} \right) \left( \frac{\partial f_i}{\partial f_s} \right) \cdot u(y_r, y_s)$$

Mass of parent gas filled, molecular weight and mole fraction of compound were the main sources of the uncertainty of the gravimetric method.

The reference gases used were listed in the following table:

| Cylinder Number | Component and assigned value(x)<br>μmol/mol | Relative standard<br>uncertainty (u(x)) |
|-----------------|---|---|
| 499707          | 9.950                                       | 0.36 %                                  |
| CAL017588       | 9.857                                       | 0.36 %                                  |
| CAL017592       | 9.915                                       | 0.36 %                                  |
| CAL017596       | 9.958                                       | 0.36 %                                  |
| 499694          | 10.395                                      | 0.36 %                                  |

### Instrument Calibration:

When testing sample, “A-B-A-B-A” type calibration procedure were used, That means the sample gas and reference gases were measured in the order of Reference – Sample – Reference – Sample – Reference. Single point calibration was used to calculate the concentration of target compound in sample cylinder.

### Sample handling:

When the package including the comparison cylinder arrived at the lab, it was in good condition. The box was unpacked and the comparison cylinder was stored at room temperature. A SS regulator was connected to the cylinder.

When testing NO<sub>2</sub> with the 42C NO-NO<sub>2</sub>-NO<sub>x</sub> Analyser, the reference and sample gases were directly introduced into the analyser through a “T” type tube by the pump inside the instrument used. The flow rate was about 0.5~1L/min controlled by a flow controller. Another outlet of the “T” tube was vented to the atmosphere. There was a pressure regulator between the cylinder and the inlet of the “T” tube to control the total gas flow rate and to make sure that about 100mL/min vented to the atmosphere. The venting flow rate was read from a flow meter.

When FT-IR was used, the gas was introduced into the gas cell by a needle valve, which was between the gas cell and the regulator on the cylinder. The gas cell was vacuumed by a turbo pump before filling it with the gas.

### Evaluation of measurement uncertainty

The contributions of measurement uncertainty were from reference gas signal readings of the sample gas and reference gas repeatability in one day or one group and reproducibility in different days or groups.

$$u(c_{CCQM}) = \sqrt{u^2(c_{PRM}) + u^2(H_{CCQM}) + u^2(H_{PRM}) + u^2(f_{int ra}) + u^2(f_{int er})}$$

Here, *u* means relative standard uncertainty.



$u(c_{CCQM})$ : Measurement uncertainty of concentration of the target component in the comparison sample gas cylinder.

$u(H_{CCQM})$ : Uncertainty of signal reading of the sample gas from 42C NO-NO<sub>2</sub>-NO<sub>x</sub> Analyser.

$u(H_{PRM})$ : Uncertainty of signal reading of the reference gas from 42C NO-NO<sub>2</sub>-NO<sub>x</sub> Analyser.

For the  $H_{CCQM}$  and  $H_{PRM}$ , the relative standard uncertainty could be calculated from the relative standard deviation (RSD) of the signal reading. The relative standard uncertainty is RSD/sqrt(n), where n is the number of signal reading.

$u(c_{PRM})$ : Uncertainty of concentration of the reference gas, which was combined by the uncertainty from gravimetric method according to ISO 6142-2001 and the uncertainty from the stability of the reference gas.

$u(f_{intra})$ : Uncertainty of repeatability in one day or one group. The relative standard uncertainty of  $f_{intra}$  was calculated from the relative standard deviation (RSD) of repeating test in one day or one group. The relative standard uncertainty is RSD/sqrt(n), where n is the number of the repeating test.

$u(f_{inter})$ : Uncertainty of reproducibility in different days or groups. The relative standard uncertainty  $f_{inter}$  was calculated from the relative standard deviation (RSD) of repeating test in different days or groups. The relative standard uncertainty is RSD/sqrt(n), where n is the number of the repeating test.

Uncertainties estimation for the results is listed in following table.

|                                   |  |        |
|-----------------------------------|--|--------|
| Uncertainty sources               | $u(c_{PRM})$                                     | 0.36 % |
|                                   | $u(H_{CCQM})$                                    | 0.10 % |
|                                   | $u(H_{PRM})$                                     | 0.10 % |
|                                   | $u(f_{intra})$                                   | 0.26 % |
|                                   | $u(f_{inter})$                                   | 0.12 % |
| Uncertainty of measurement result | $u(c_{CCQM})$                                    | 0.48 % |
|                                   | Relative expanded uncertainty<br>( $k=2$ , 95 %) | 1.0 %  |

## National Institute of Standards and Technology (NIST)

### A1. General information

|                                    |  |     |              |
|------------------------------------|--|-----|--------------|
| Institute                          | National Institute of Standards and Technology             |     |              |
| Address                            | 100 Bureau Drive<br>Gaithersburg, MD 20899-8393, USA       |     |              |
| Contact person                     | Franklin R. Guenther, William J. Thorn III                 |     |              |
| Telephone                          | 301-975-3939   | Fax | 301-977-8392 |
| Email*                             | <a href="mailto:fguenther@nist.gov">fguenther@nist.gov</a> |     |              |
| Serial number of cylinder received | APEX930654   |     |              |
| Cylinder pressure as received      | 8.0 MPa  |     |              |

### A2. Results

| Nitrogen dioxide mole fraction          | Expanded uncertainty                       | Coverage factor |
|---|--|-----------------|
| $x_{\text{NO}_2}$ / $\mu\text{mol/mol}$ | $U(x_{\text{NO}_2})$ / $\mu\text{mol/mol}$ |                 |
| 10.28                                   | 0.10                                       | 2               |

### A3. Uncertainty Budget

Please provide a complete uncertainty budget.

- 1) Primary Standard Uncertainty = 0.009 to 0.015  $\mu\text{mol/mol}$  (Gaussian)
- 2) Instrument precision = 0.01  $\mu\text{mol/mol}$  (Gaussian)
- 3) Instrument drift = 0.015 to 0.052  $\mu\text{mol/mol}$  (Gaussian)
- 4) Nitric Acid Determination = 0.06  $\mu\text{mol/mol}$  (Gaussian)

Propagation of these core measurement uncertainties resulted in the following table of results against the 6 primary standards:

|            | NOx    | std err | Uncert |
|------------|--------|---------|--------|
| APEX930654 | 10.258 | 0.036   | 0.072  |
| APEX930654 | 10.385 | 0.067   | 0.134  |
| APEX930654 | 10.308 | 0.043   | 0.087  |

|                           |               |              |              |
|---------------------------|---------------|--------------|--------------|
| <b>APEX930654</b>         | <b>10.354</b> | <b>0.034</b> | <b>0.069</b> |
| <b>APEX930654</b>         | <b>10.370</b> | <b>0.044</b> | <b>0.088</b> |
| <b>APEX930654</b>         | <b>10.370</b> | <b>0.069</b> | <b>0.138</b> |
|                           | <b>NOx</b>    |              |              |
| <b>average</b>            | <b>10.34</b>  | <b>0.021</b> |              |
| <b>(Max-Min)/sqrt(12)</b> |               | <b>0.037</b> |              |
| <b>Stderr</b>             |               | <b>0.042</b> |              |
| <b>Uncert</b>             |               | <b>0.08</b>  |              |

The NOx value for the cylinder then was adjusted by the analysed value for nitric acid of  $(0.06 \pm 0.06)$   $\mu\text{mol/mol}$  to get  $(10.28 \pm 0.10)$   $\mu\text{mol/mol}$ .

#### **A4. Description of the procedure used during the gas analysis**

Please describe in detail the analytical method(s) used for gas analysis<sup>14</sup>.

The cylinder was measured for total NOx using chemiluminescence by comparison to six Nitric Oxide primary gas standards. To assure close to 100 % conversion efficiency, NIST used two thermal converters in series. During the analysis, NIST nitrogen dioxide working standards were also compared as controls.

The analysis of the comparison cylinder (APEX930654) was completed over 6 days of analytical comparisons. Each day one of the NIST nitric acid primary standards was compared to a nitrogen dioxide working standard and the comparison cylinders. This resulted in 6 independent analytical values for the comparison cylinder. These 6 independent values were combined by averaging the results into one value, and combining the uncertainties.

The nitric acid value for the cylinder was estimated by directing a gas stream from the cylinder through a trap consisting of a fluorocarbon filter housing containing two 47 mm diameter nylon membrane filters, and on to the chemiluminescence instrument. The value from the instrument without the nylon membranes in the gas stream, and then without the nylon membranes in the gas stream is noted. The difference between these two numbers is the nitric acid value. As this technique has the potential of underreporting nitric acid, or underreporting nitric oxide; the uncertainty is estimated at  $0.06$   $\mu\text{mol/mol}$ . Attempts to place a value for nitric acid by FT-IR failed due to a detection limit of  $0.5$   $\mu\text{mol/mol}$ .

The final value for nitrogen dioxide equals the NOx value determined minus the value for the nitric acid.

#### **A5. Complementary information on the cylinder**

Please report the value of the pressure left in the cylinder before shipment to the BIPM: 4 Mpa

If any other component other than NO<sub>2</sub>, nitrogen and oxygen was detected and/or quantified please report its mole fraction in the table below:

<sup>14</sup> The choice of the procedure used for gas analysis is the responsibility of the participating laboratory. Nevertheless, for a proper evaluation of the data, it is necessary that the calibration method, as well as the way in which the calibration mixtures have been prepared is reported to the co-ordinators.

| Component        | Mole fraction / nmol/mol | Expanded uncertainty | Coverage factor | Measurement technique |
|------------------|--------------------------|----------------------|-----------------|-----------------------|
| HNO <sub>3</sub> | 0.06                     | 0.06                 | 2               | chemiluminescence     |

## National Physical Laboratory (NPL)

CCQM-K74 Report from NPL (March 2010)

The unknown cylinder circulated by BIPM to NPL was D65 0042 (930659).

Pressure on receipt (07/09/2009) = 80 bar  
 Pressure on despatch (25/01/2010) = 40 bar

### Introduction

The measurements of cylinder 930659 were made in December 2009 by direct comparison with a nitrogen dioxide in nitrogen standard containing similar nitrogen dioxide (10ppm) and oxygen (1000ppm) amount fractions to the unknown cylinder (NPL 1275R).

The direct comparison measurements were performed, on two separate days, using a ABB Limas UV 11 photometer.

### Preparation of NPL 1275R

The standard NPL 1275R was prepared by the following series of dilutions from nominally pure nitric oxide:

| Parents   | Daughter                                      |        |
|---|---|--------|
| pure NO + N <sub>2</sub>  | 50 mmol/mol NO/N <sub>2</sub>                 | 74R2   |
| 50 mmol/mol NO/N <sub>2</sub> + 9 % O <sub>2</sub> /N <sub>2</sub> + N <sub>2</sub> | 4000 µmol/mol NO <sub>2</sub> /N <sub>2</sub> | 464R   |
| 4000 µmol/mol NO <sub>2</sub> /N <sub>2</sub> + N <sub>2</sub>                      | 800 µmol/mol NO <sub>2</sub> /N <sub>2</sub>  | 1206R2 |
| 800 µmol/mol + 9 % O <sub>2</sub> /N <sub>2</sub> + N <sub>2</sub>                  | 100 µmol/mol NO <sub>2</sub> /N <sub>2</sub>  | 1117R4 |
| 100 µmol/mol NO <sub>2</sub> /N <sub>2</sub> + N <sub>2</sub>                       | 10 µmol/mol NO <sub>2</sub> /N <sub>2</sub>   | 1275R  |

The full uncertainty for the final step is given as a Table and shows the standard uncertainty in the prepared amount fraction of NO<sub>2</sub> of 13 nmol/mol.

The best estimate of drift in this value is based on the use of this type of cylinder (BOC Spectraseal) for standards of NO<sub>2</sub> at similar amount fractions. This leads to a standard uncertainty of 10 nmol/mol.

### Measurements on 7/12/2009

| Sequence | Result µmol/mol |
|----------|-----------------|
| Z        |                 |
| SU       | 10.3167         |
| US       | 10.2871         |
| Z        |                 |
| SU       | 10.3643         |
| US       | 10.3447         |

|             |                |
|-------------|----------------|
| Z           |                |
| SU          | 10.3052        |
| US          | 10.3768        |
| Z           |                |
| SU          | 10.2935        |
| US          | 10.3626        |
| <b>Mean</b> | <b>10.3314</b> |
| SD          | <b>0.035</b>   |

where, Z indicates a measurement of “zero” nitrogen, S indicates a measurement of the standard (NPL1275R) and U a measurement of the unknown (D65 0042).

Measurements on 8/12/2009

| Sequence  | Result<br>$\mu\text{mol/mol}$ |
|-----------|-------------------------------|
| Z         |                               |
| SU        | 10.39822                      |
| US        | 10.35001                      |
| Z         |                               |
| SU        | 10.26746                      |
| US        | 10.33609                      |
| Mean      | <b>10.33795</b>               |
| <b>SD</b> | <b>0.054</b>                  |

Estimation of matrix effect in NDUV analyser

There is a small matrix effect in the NDUV analyser due to the presence of oxygen in the nitrogen. This effect was estimated by carrying out a comparison of the unknown with a standard (NPL 1126R2), which had 26  $\mu\text{mol/mol}$  of oxygen. The comparison was carried out using the sequence described in the previous section. The results were

| Date      | Result<br>$\mu\text{mol/mol}$ | SD<br>$\mu\text{mol/mol}$ |
|-----------|-------------------------------|---------------------------|
| 7-12-2009 | 10.308                        | 0.035                     |
| 8-12-2009 | 10.283                        | 0.020                     |

The mean of these results (10.29  $\mu\text{mol/mol}$ ) is 40 nmol/mol lower than the results of the comparison reported in the previous section with a “matching” standard. We therefore attribute an uncertainty of 10 nmol/mol (95 % C/I) to the “uncorrected matrix effect” in the analyser.

Identification of Trace Contamination in the Travelling Standard

Analysis of D65 0042 by FT-IR indicated the presence of approximately 200 nmol/mol of nitric acid. We are unable to measure the cross-sensitivity of the NDUV analyser. From information about the spectroscopy of NO<sub>2</sub> and HNO<sub>3</sub> we estimate a cross-sensitivity of no more than 0.1. After using a factor of square root(3), we estimate a standard uncertainty due to possible cross-sensitivity of 12 nmol/mol.

Uncertainty

| Source of uncertainty                  | Estimation<br>Method | Standard uncertainty<br>nmol/mol |
|--|----------------------|----------------------------------|
| Gravimetric preparation of standard    | A                    | 13                               |
| Drift in gravimetric value of standard | B                    | 10                               |
| Repeatability of analysis              | A                    | 35                               |

|  |   |    |
|--|---|----|
| Interference of HNO <sub>3</sub> in analyser | B | 12 |
| Matrix effect (oxygen) in analyser           | B | 5  |

---

Combined uncertainty 41

The expanded uncertainty ( $k=2$ ) is 80 nmol/mol.

Final Result

Amount fraction of nitrogen dioxide in D65 0042 = 10.33 +/- 0.08 μmol/mol

| Gravimetric composition of cylinder NPL 1275R |                            |             |       |
|---|----------------------------|-------------|-------|
| Component                                     | μmol/mol                   | uncertainty | % u/c |
| N2  | 998990.899                 | 1.371       |       |
| O2  | 998.129                    | 0.821       | 0.08  |
| NO2   | 10.001                     | 0.013       | 0.13  |
| Ar  | 0.584                      | 0.094       | 16    |
| H2O   | 0.006                      | 0.0019      | 30    |
| CxHy  | 0.005                      | 0.0045      | 85    |
| CO  | 0.003                      | 0.0014      | 43    |
| N2O   | 0.002                      | 0.0003      | 11    |
| CO2   | 0.002                      | 0.0003      | 16    |
| methane                                       | 0.001                      | 0.0009      | 65    |
| H2  | 0.001                      | 0.0009      | 86    |
| SO2   | 0.001                      | 0.0002      | 38    |
| NO  | 0.000                      | 0.0002      | 46    |
| propane                                       | 0.000                      | 0.0000      | 41    |
| INPUTS  |                            |             |       |
| =====   |                            |             |       |
| File  | m / g                      | u(m) / g    |       |
| NPL1117R4                                     | 146.219                    | 0.03        |       |
| PureN2-BI                                     | 1321.995                   | 0.03        |       |
| *****   | ***** NPL1117R4.txt ****   | *****       |       |
| Component                                     | mol/mol                    | uncertainty |       |
| N2  | 0.989858553                | 1.1E-05     |       |
| O2  | 0.010035802                | 8.1E-06     |       |
| NO2   | 0.000100557                | 1.3E-07     |       |
| Ar  | 1.3431E-06                 | 8.3E-07     |       |
| N2O   | 2.21E-08                   | 2.5E-09     |       |
| H2O   | 0.000000018                | 6.6E-09     |       |
| CO2   | 1.64E-08                   | 2.5E-09     |       |
| CxHy  | 8.4E-09                    | 4.4E-09     |       |
| methane                                       | 4.9E-09                    | 9.0E-10     |       |
| CO  | 4.7E-09                    | 1.4E-09     |       |
| SO2   | 1.7E-09                    | 7.0E-10     |       |
| H2  | 1.5E-09                    | 9.0E-10     |       |
| NO  | 5E-10                      | 2.0E-10     |       |
| propane                                       | 1E-10                      | 0.0E+00     |       |
| *****   | ***** PureN2-BIP+.txt **** | *****       |       |
| Component                                     | mol/mol                    | uncertainty |       |
| Ar  | 0.00000005                 | 0.00000005  |       |
| CO  | 0.000000003                | 1.5E-09     |       |
| O2  | 0.000000005                | 2.5E-09     |       |
| CxHy  | 0.000000005                | 5E-09       |       |
| H2O   | 0.000000005                | 2E-09       |       |
| N2  | 0.999999479                | 8.735E-07   |       |
| NO  | 5E-10                      | 3E-10       |       |
| SO2   | 5E-10                      | 3E-10       |       |
| methane                                       | 0.000000001                | 1E-09       |       |
| H2  | 0.000000001                | 1E-09       |       |



## Chemicals Evaluation and Research Institute (CERI)

**A1. General information**

|                                    |   |     |                 |
|------------------------------------|---|-----|-----------------|
| Institute                          | Chemicals Evaluation and Research Institute, Japan                          |     |                 |
| Address                            | 1600 Shimotakano, Sugito-machi, Kitakatsushika-gun, Saitama 345-0043, Japan |     |                 |
| Contact person                     | Shinji Uehara   |     |                 |
| Telephone                          | +81-480-37-2601   | Fax | +81-480-37-2521 |
| Email*                             | <a href="mailto:uehara-shinji@ceri.jp">uehara-shinji@ceri.jp</a>            |     |                 |
| Serial number of cylinder received | APEX930671  |     |                 |
| Cylinder pressure as received      | 9.7MPa  |     |                 |

**A2. Results**

## Measurement #1

| Component       | Date (dd/mm/yy) | Result ( $\mu\text{mol/mol}$ ) | Standard deviation (% relative) | Number of replicates |
|-----------------|-----------------|--------------------------------|---------------------------------|----------------------|
| NO <sub>2</sub> | 09/12/09        | 10.35                          | 0.39                            | 3                    |

## Measurement #2

| Component       | Date (dd/mm/yy) | Result ( $\mu\text{mol/mol}$ ) | Standard deviation (% relative) | Number of replicates |
|-----------------|-----------------|--------------------------------|---------------------------------|----------------------|
| NO <sub>2</sub> | 10/12/09        | 10.42                          | 0.23                            | 3                    |

## Measurement #3

| Component       | Date (dd/mm/yy) | Result ( $\mu\text{mol/mol}$ ) | Standard deviation (% relative) | Number of replicates |
|-----------------|-----------------|--------------------------------|---------------------------------|----------------------|
| NO <sub>2</sub> | 10/12/09        | 10.42                          | 0.25                            | 3                    |

## Results

| Nitrogen dioxide mole fraction<br>$x_{\text{NO}_2}/\mu\text{mol/mol}$ | Expanded uncertainty<br>$U(x_{\text{NO}_2})/\mu\text{mol/mol}$ | Coverage factor |
|---|--|-----------------|
| 10.40   | <b>0.38</b>  | 2               |



### A3. Uncertainty Budget

| Source                                       | Estimate                    | Distribution | Divisor | Standard uncertainty | Sensitivity coefficient | Contribution |
|--|-----------------------------|--------------|---------|----------------------|-------------------------|--------------|
| Repeatability of measurement                 | 0.1501 $\mu\text{mol/mol}$  | normal       | 2       | 0.07505              | 1                       | 0.07505      |
| Calibration curve                            | 0.03372 $\mu\text{mol/mol}$ | normal       | 2       | 0.01686              | 1                       | 0.01686      |
| Mole fraction of high concentration standard | 0.05952 $\mu\text{mol/mol}$ | normal       | 2       | 0.02976              | 1.118                   | 0.03327      |
| Mole fraction of low concentration standard  | 0.05399 $\mu\text{mol/mol}$ | normal       | 2       | 0.02700              | 0.1183                  | 0.003194     |
| NO in the CCQM-K74 gas mixture               | 0.0008 $\mu\text{mol/mol}$  | normal       | 2       | 0.0004               | 1                       | 0.0004       |
| Stability                                    | 0.34 $\mu\text{mol/mol}$    | normal       | 2       | 0.17                 | 1                       | 0.17         |
| Combined standard uncertainty                |                             |              |         |                      |                         | 0.1896       |
| Expanded uncertainty ( $k=2$ )               |                             |              |         |                      |                         | 0.38         |

### A4. Description of the procedure used during the gas analysis

#### Calibration Standards

| Serial number of cylinder | Mole fraction of $\text{NO}_2$ $\mu\text{mol/mol}$ | Expanded uncertainty ( $k=2$ ) $\mu\text{mol/mol}$ | Nominal mole fraction of $\text{O}_2$ $\mu\text{mol/mol}$ |
|---------------------------|--|--|---|
| CPB19013                  | 10.184   | 0.05615  | 1000  |
| CPB21194                  | 8.398  | 0.05601  | 1000  |

Calibration standards were prepared by gravimetric dilution of pure NO and pure  $\text{O}_2$ . That procedure is as follows,




Fig. Procedure of preparation

#### The analytical method used for gas analysis

Calibration method: Bracketing

Traceability: Own standards


Instrument: Chemiluminescence analyser made by Thermo Fisher Scientific  
(Model 42i-HL)

Catalyst of converter: stainless-steel

Measurement Mode: Manual NOx

This instrument has three modes. (Auto mode, Manual NO mode and Manual NOx mode) NO<sub>2</sub> cannot be analysed in “Manual NOx” mode. NOx was regarded as NO<sub>2</sub> in the report. NO<sub>2</sub> can be estimated by subtracting output value of NO from output value of NOx in “Auto mode”. But the observed value of NO is bigger than the accurate one in this mode. Therefore, the uncertainty increases. So “Manual NOx” mode was selected.

**Comparisons were made by the following sequence:**



Where

$H_i$ : measurement of high concentration standard ( $i=1,2$ )

$L_i$ : measurement of low concentration standard ( $i=1,2$ )

$K_i$ : measurement of the CCQM-K74 gas mixture ( $i=1,2,3$ )

**Configuration of analysis system:**

Gas cylinder → Regulator → Manual 4-way valve → Instrument (Converter → Detector)

**The mole fractions of NO<sub>2</sub> in the CCQM-K74 gas mixture were calculated as follows:**

$$\frac{1}{3} \left\{ (H - L) \left( \frac{K_1^v - L_1^v}{H_1^v - L_1^v} + \frac{K_2^v - L_1^v}{H_2^v - L_1^v} + \frac{K_3^v - L_2^v}{H_2^v - L_2^v} \right) + 3B \right\}$$

Where

$K_i^v$ : output value from measurement  $K_i$

$H_i^v$ : output value from measurement  $H_i$

$L_i^v$ : output value from measurement  $L_i$

$K_i^v$ : output value from measurement  $K_i$

$H$ : mole fraction of high concentration standard

$L$ : mole fraction of low concentration standard

**A5. Complementary information on the cylinder**

The value of the pressure left in the cylinder before shipment to the BIPM: 7.4MPa

If any other component other than NO<sub>2</sub>, nitrogen and oxygen was detected and/or quantified please report its mole fraction in the table below:

| Component | Mole fraction /nmol/mol | Expanded uncertainty | Coverage factor | Measurement technique                  |
|-----------|-------------------------|----------------------|-----------------|--|
| NO        | 0.0036                  | 0.0008               | 2               | Chemiluminescence analyser □ NO mode □ |

## National Metrology Institute of South Africa (NMISA)

### A1. General information

|                                    |  |     |                 |
|------------------------------------|--|-----|-----------------|
| Institute                          | National Metrology Institute of South Africa (NMISA)   |     |                 |
| Address                            | CSIR, Building 4 West<br>Meiring Naude Road<br>Brummeria<br>0184<br>Pretoria<br>South Africa |     |                 |
| Contact person                     | Angelique Botha  |     |                 |
| Telephone                          | +27 12 841 3800  | Fax | +27 12 841 2131 |
| Email*                             | abotha@nmisa.org   |     |                 |
| Serial number of cylinder received | <b>D650032</b>   |     |                 |
| Cylinder pressure as received      | <b>96 bar</b>  |     |                 |

### A2. Results

| Nitrogen dioxide mole fraction<br><br>$x_{\text{NO}_2}$ / $\mu\text{mol/mol}$ | Expanded uncertainty<br><br>$U(x_{\text{NO}_2})$ / $\mu\text{mol/mol}$ | Coverage factor |
|---|--|-----------------|
| <b>10,69</b>  | <b>0,37</b>  | <b>k = 2</b>    |

### A3. Uncertainty Budget

Please provide a complete uncertainty budget.

The budget of the standard uncertainties for the comparison sample is:

| Parameter                    | Standard uncertainty                        |
|------------------------------|---|
| Gravimetric uncertainty      | - Weighing uncertainty<br>- Purity analysis |
| Verification uncertainty (U) | 0,14 % rel.                                 |
| Stability uncertainty (U)    | 2,05 % rel.                                 |
|                              | 0,77 % rel.                                 |

|                            |            |
|----------------------------|------------|
| Regression uncertainty (U) | 0,06% rel. |
|----------------------------|------------|

#### A4. Description of the procedure used during the gas analysis

Please describe in detail the analytical method(s) used for gas analysis<sup>15</sup>.

The NO<sub>2</sub> content of sample D650032 was analysed using a Fischer Rosemount Chemiluminescence analyser calibrated with primary standard mixtures of NO<sub>2</sub>-in-nitrogen over the concentration range of 10 ppm to 100 ppm prepared by NMISA.

#### A5. Complementary information on the cylinder

Please report the value of the pressure left in the cylinder before shipment to the BIPM:

The cylinder was left with **23 bar**.

If any other component other than NO<sub>2</sub>, nitrogen and oxygen was detected and/or quantified please report its mole fraction in the table below:

| Component | Mole fraction / nmol/mol | Expanded uncertainty | Coverage factor | Measurement technique |
|-----------|--------------------------|----------------------|-----------------|-----------------------|
|           |                          |                      |                 |                       |
|           |                          |                      |                 |                       |
|           |                          |                      |                 |                       |

<sup>15</sup> The choice of the procedure used for gas analysis is the responsibility of the participating laboratory. Nevertheless, for a proper evaluation of the data, it is necessary that the calibration method, as well as the way in which the calibration mixtures have been prepared is reported to the co-ordinators.

**Slovak Institute of Metrology (SMU)****A-1 General information**

|                                    |  |     |                  |
|------------------------------------|--|-----|------------------|
| Institute                          | SLOVAK INSTITUTE OF METROLOGY                            |     |                  |
| Address                            | Karloveská 63<br>SK-842 55 Bratislava<br>Slovak Republic |     |                  |
| Contact persons                    | RNDr. Viliam Štovčík, PhD. , Ing. Miroslava Valková      |     |                  |
| Telephone                          | +421 2 60294 523   | Fax | +421 2 60294 561 |
| Email*                             | stovcik@smu.gov.sk, valkova@smu.gov.sk                   |     |                  |
| Serial number of cylinder received | D 650038   |     |                  |
| Cylinder pressure as received      | 9 MPa  |     |                  |

## A-2 Results

| Measuring Month | Nitrogen dioxide mole fraction        | Expanded uncertainty                     | Coverage Factor |
|-----------------|---------------------------------------|--|-----------------|
|                 | $x_{NO_2}$<br>[ $\mu\text{mol/mol}$ ] | $U(x_{NO_2})$<br>[ $\mu\text{mol/mol}$ ] |                 |
| October         | 10.09                                 | 0.13                                     | 2               |
| November        | 10.11                                 | 0.13                                     | 2               |
| December        | 10.11                                 | 0.11                                     | 2               |
| $\bar{X}$       | <b>10.10</b>                          | <b>0.12</b>                              | 2               |

## A-3 Uncertainty Budget

Combined uncertainty of the NO<sub>2</sub> mole fraction (analysed by chemiluminescence method) was calculated according to the formula:

$$U_C(x) = 2 * \sqrt{u_A(x_{NO_2})^2 + u_B(x_{NO_2})^2}$$

Where:

$u_A(x)$ : square deviation of measure (SD)

$$SD = \sqrt{\frac{\sum_{i=1}^n (x_i - \bar{X})^2}{(n-1)}}$$

$n= 20$ ; number of replicates

$$u_B(x) = \sqrt{(x_{\max} - x_{\min})^2 + d^2 + u_{cal}^2}$$

$d= 0.01 \mu\text{mol/mol}$ ; discernment of analyzer

$$u_{cal} = 0.05 \mu\text{mol} / \text{mol}$$

$u_{cal}$  is the standard uncertainty of  $10 \mu\text{mol/mol}$  NO<sub>2</sub> PSM used for the analyzer calibration

To estimate result uncertainty from 3 measurement results we have kept "Standard Practice for Conducting an Interlaboratory Study to Determine the Precision of a Test Method" (Annual Book of ASTM Standards E 691-87) with some approximations.

$$s_R = \sqrt{s_x^2 + s_r \frac{n-1}{n}}$$

$$s_r = \sqrt{\frac{\sum_{i=1}^p u(\bar{x}_i)^2}{p}}$$

$$s_x = \frac{\max(\Delta x)}{\sqrt{3}}$$

$$\Delta x = \bar{x}_1 - \bar{x}_2$$

$p$  – Number of measurements  
 $n$  – Number of replicates

Final result is average from 3 measurement results

$$\bar{X} = \frac{\sum_{i=1}^p \bar{x}_i}{p}$$

As final standard uncertainty we assigned to this result  $\max(s_R \text{ or } s_r)$

$$u(\bar{X}) = \max(s_r; s_R)$$

Expanded uncertainty ( $k=2$ ) of final result

$$U(\bar{X}) = 2 \cdot u(\bar{X})$$

#### A-4 Description of the procedure used during the gas analysis

The concentration of NO<sub>2</sub> was determined by chemiluminescence method in gas laboratory. Thermo Vision 42C analyzer was used for measurements. SMU own PSM prepared gravimetrically in according to ISO 6142 was used for the calibration of the analyzer before each measurement. NO<sub>2</sub> content was measured after receiving the cylinder from pilot laboratory and then two times after 4 weeks. 5 out of different PSMs in the mole fraction range (1-10) μmol/mol NO<sub>2</sub> were used to make a calibration curve and afterwards from calibration curve were constated an unknown quantity of assay. The same fabrics was used for monitoring NO component (NO was found as an impurity). Each measurement was performed under stable conditions:

Pressure: 172.5 mmHg  
 Flow: 0.720 l per min.  
 Integration time: 30 sec

Before measuring process the cylinders were homogenized and stabilized for laboratory temperature about 60 minutes.

The laboratory conditions:

Temperature: 20-23°C

Pressure: 98. 9-101.325 KPa

Rel. humidity: 20%-35%.

#### A-5 Complementary information on the cylinder

Measurement of the NO component was executed by the same method as well as for than NO<sub>2</sub> and its mole fraction are in the table below:

| Component | Mole fraction<br>μmol/mol |
|-----------|---------------------------|
| NO        | 0.065                     |

The value of the pressure left in the cylinder before shipment to the BIPM:

$p = 70,0 \text{ MPa}$

RNDr. V. Štovčík, PhD., Ing. M.Valková  
Laboratory of gases SMU

Ing. S. Musil, PhD.  
Deputy of research SMU



## Mendeleev Institute for Metrology (VNIIM)

### A1. General information

|                                    |  |     |                   |
|------------------------------------|--|-----|-------------------|
| Institute                          | <b>D.I.Mendeleev Institute for Metrology (VNIIM)</b> |     |                   |
| Address                            | 19 Moskovsky pr.,<br>St. Petersburg, 190005 Russia   |     |                   |
| Contact person                     | Leonid Konopelko, Elena Gromova                      |     |                   |
| Telephone                          | +7 812 327 57 11                                     | Fax | + 7 812 327 97 76 |
| Email*                             | lkonop@b10.vniim.ru                                  |     |                   |
| Serial number of cylinder received | 930713   |     |                   |
| Cylinder pressure as received      | 10 MPa   |     |                   |

---

<sup>1</sup> CCQM-K74 is fully described in the document: Protocol for CCQM-K74 and CCQM-P110 – Nitrogen dioxide in nitrogen (10 µmol/mol) page 5.

## A2. Results

|   |  |                 |
|---|--|-----------------|
| Nitrogen dioxide mole fraction          | Expanded uncertainty                       | Coverage factor |
| $x_{\text{NO}_2}$ / $\mu\text{mol/mol}$ | $U(x_{\text{NO}_2})$ / $\mu\text{mol/mol}$ |                 |
| <b>10.55</b>                            | <b>0.16</b>                                | <b>2</b>        |

## A3. Uncertainty Budget

Combined standard uncertainty of NO<sub>2</sub> mole fraction was calculated on the base of the following constituents:

- uncertainty of NO<sub>2</sub> mole fraction in the calibration gas mixtures (including uncertainty of weighing of parent gases and pre-mixtures, uncertainty of the purity of the parent gases);
- uncertainty due to sorption of the component by cylinder inner walls and long-term instability of standard gas mixtures;
- standard deviation of the measurement results of NO<sub>2</sub> mole fraction.

Uncertainty budget for NO<sub>2</sub> mole fraction in gas mixture in the cylinder 930713 is given below.

| Nº                                   | Source of uncertainty   | Type of evaluation | Standard uncertainty, % relative |
|--------------------------------------|---|--------------------|----------------------------------|
| 1                                    | Preparation of the calibration gas mixtures   | A, B               | 0,1                              |
| 2                                    | Sorption of the component by cylinder inner walls and long-term instability of standard gas mixture | B                  | 0,5                              |
| 3                                    | Standard deviation of the measurement results of NO <sub>2</sub> mole fraction                      | A                  | 0,54                             |
| <b>Combined standard uncertainty</b> |   |                    | <b>0,74</b>                      |
| <b>Expanded uncertainty (k=2)</b>    |   |                    | <b>1,5</b>                       |

## A4. Description of the procedure used during the gas analysis

### Reference Method:

The mole fraction of NO<sub>2</sub> was determined by chemiluminescence method. The measurements were carry out by chemiluminescence analyzer AC 30M (Environnement S.A., France) which is a part of equipment of the State Primary Standard of mole fraction and mass concentration of components in gas media GET 154-01.

### Calibration standards:

There were two Primary Standard Gas Mixtures (calibration gas mixtures) of NO<sub>2</sub> 10  $\mu\text{mol/mol}$  in Nitrogen prepared gravimetrically in accordance with the ISO 6142:2001. Preparation of the calibration mixtures was carried out using high purity NO<sub>2</sub> and high purity Nitrogen in 3 stages:

1<sup>st</sup> stage - preparation of the first gas pre-mixtures NO<sub>2</sub>/N<sub>2</sub> with NO<sub>2</sub> mole fraction of about 0,01 mol/mol.

2<sup>nd</sup> stage – preparation of the second gas pre-mixtures NO<sub>2</sub>/N<sub>2</sub> with NO<sub>2</sub> mole fraction of about 300 μmol/mol (the Nitrogen balance gas contains nominally 0,03 mol/mol of Oxygen),

3<sup>rd</sup> stage – preparation of calibration gas mixtures NO<sub>2</sub>/N<sub>2</sub> with NO<sub>2</sub> mole fraction of about 10 μmol/mol (the Nitrogen balance gas contains nominally 1000 μmol/mol of Oxygen).

All gas mixtures were prepared in aluminum cylinders with Experis treatment for NO<sub>2</sub>. The gas mixtures were verified according to the ISO 6143:2001.

#### **Instrument calibration:**

The single calibration point method was used for instrument calibration.

There were five independent measurements made under repeatability conditions with five independent calibrations. Measurement sequence was in order of standard-sample- standard-sample- etc.

#### **Sample handling:**

Prior to measurements the cylinders were rolling to ensure homogeneity of the mixtures for two hours and then stored at room temperature at least 24 hours.

Stainless steel metering valves were connected with the cylinders. The valves were cleaned at least 10 times by sequential purging over a two day period. Then they were left under pressure for 24 hours. Directly before the measurements the valves were purged another 3 times.

After each measurement, pure Nitrogen was flushed through the analyzer. Both sample and calibration mixtures responses were collected.

#### **A5. Complementary information on the cylinder**

The value of the pressure left in the cylinder before shipment to the BIPM:

**7.2 MPa**

#### **Authors:**

L.A. Konopelko  
Y.A. Kustikov  
V.V. Pankratov  
A.V. Malginov  
E.V. Gromova

The single calibration point method was used for instrument calibration. There were four independent measurements made under repeatability conditions with four independent calibrations. Measurement sequence was in order of standard-sample- standard-sample- etc.

**Sample handling:**

Prior to measurements the cylinders were rolling to ensure homogeneity of the mixtures for two hours and then stored at room temperature at least 24 hours.

A stainless steel metering valve was connected with the sample cylinder. The valve was cleaned at least 10 times by sequential purging over a two day period. Directly before the measurements the valve was purged another 3 times.

A pressure regulator for high concentration gas pre-mixture of NO<sub>2</sub> in Nitrogen was purged for 5 times. Then it was left under pressure for 24 hours and directly before measurements it was purged another 3 times.

After each measurement, pure Nitrogen was flushed through the analyzer. Both sample and dynamic mixtures responses were collected.

## Appendix A

The results obtained in additional investigations with alternative reference standards  
(Dynamic volumetric method)

## Results

| Nitrogen dioxide mole fraction<br>$x_{\text{NO}_2}$ / $\mu\text{mol/mol}$ | Expanded uncertainty<br>$U(x_{\text{NO}_2})$ / $\mu\text{mol/mol}$ | Coverage factor |
|---|--|-----------------|
| 10,59   | 0,14   | 2               |

## Uncertainty Budget

| No                                      | Source of uncertainty   | Type of evaluation | Relative standard uncertainty, % |
|---|---|--------------------|----------------------------------|
| 1                                       | Preparation of standard gas mixture<br>317,9 $\mu\text{mol/mol}$                                    | A, B               | 0,07                             |
| 2                                       | Sorption of the component by cylinder inner walls and long-term instability of standard gas mixture | B                  | 0,10                             |
| 3                                       | Dynamic dilution  | B                  | 0,42                             |
| 4                                       | Standard deviation of measurement results   | A                  | 0,50                             |
| Total relative standard uncertainty     |   |                    | 0,66                             |
| Relative expanded uncertainty ( $k=2$ ) |   |                    | 1,3                              |

## Description of the procedure used during the gas analysis

## Reference Method:

AC 30 M (Environnement S.A.) chemiluminescence analyzer was used to measure the  $\text{NO}_2$ .

## Calibration Standards:

There was a gas mixture with  $\text{NO}_2$  mole fraction of 320  $\mu\text{mol/mol}$  used for the dynamic dilution and prepared independent from the static gravimetric mixtures pointed in the main part of the report.

Investigations of the stability of the gas mixture during one month were made.

Then, a gas mixture of  $\text{NO}_2$  in Nitrogen at about 10  $\mu\text{mol/mol}$  was generated by diluting the gravimetric gas mixture with  $\text{NO}_2$  mole fraction of 320  $\mu\text{mol/mol}$  with Nitrogen by using thermal mass-flow controller system. Flow controllers were calibrated by Gas Flow Calibrator Cal=Trak SL-800 (Sierra Instruments, Inc.).

## Federal Institute for Materials Research and Testing (BAM)

|                                    |  |     |                  |
|------------------------------------|--|-----|------------------|
| Institute                          | BAM Federal Institute for Materials Research and Testing |     |                  |
| Address                            | Unter den Eichen 87<br>12205 Berlin<br>Germany           |     |                  |
| Contact person                     | Dirk Tuma  |     |                  |
| Telephone                          | +49 30 8104 4113   | Fax | +49 30 8104 3207 |
| Email*                             | dirk.tuma@bam.de   |     |                  |
| Serial number of cylinder received | 930722 TC1   |     |                  |
| Cylinder pressure as received      | 98 bar   |     |                  |

### A2. Results

| Nitrogen dioxide mole fraction        | Expanded uncertainty                     | Coverage factor |
|---------------------------------------|--|-----------------|
| $x_{\text{NO}_2} / \mu\text{mol/mol}$ | $U(x_{\text{NO}_2}) / \mu\text{mol/mol}$ | $k$             |
| 10.53                                 | 0.75                                     | 2               |

### A3. Uncertainty Budget

Please provide a complete uncertainty budget.

Three terms go with the calculation of  $U$ :  $u_{\text{meas}}$  = standard deviation from the measurement (three scans) of the sample;  $u_{\text{cal}}$  = standard deviation from the measurement of the two calibration gases (three scans for each calibration gas, the value for  $u_{\text{cal}}$  is the arithmetic mean);  $u_{\text{intrinsic}}$  = an intrinsic uncertainty of the calibration gas of 2 % (considers stability)

$$U(x_{\text{NO}_2}) = \sqrt{(u_{\text{meas}})^2 + (u_{\text{cal}})^2 + (u_{\text{intrinsic}})^2}$$

### A4. Description of the procedure used during the gas analysis

Please describe in detail the analytical method(s) used for gas analysis<sup>16</sup>.

<sup>16</sup> The choice of the procedure used for gas analysis is the responsibility of the participating laboratory. Nevertheless, for a proper evaluation of the data, it is necessary that the calibration method, as well as the way in which the calibration mixtures have been prepared is reported to the co-ordinators.

The analytical method was FT-IR spectroscopy (cf. protocol P-110 for details).

Prior to the analysis, the cell was evacuated and flushed with dry nitrogen to remove any impurities that absorb infrared energy. A 10-m gas cell was employed.

Background was collected in the evacuated cell. Subsequently, the gas pressure was adjusted to a constant pressure of 100 kPa, and several test runs provided identical spectra.

Sample spectra of each gas were recorded three times maintaining equal conditions. For each measurement, a new gas portion was loaded from the cylinder. A background scan was done before each sample scan. The analysis resorted to the area below the first overtone, i.e., approximately between 2800 and 2950  $\text{cm}^{-1}$ .

Two calibration gases were employed;  $x(\text{cal. sample \# 1}) = 12.28 \mu\text{mol/mol}$ ,  $x(\text{cal. sample \# 2}) = 9.43 \mu\text{mol/mol}$ . The calibration samples were prepared gravimetrically. The  $\text{NO}_2$  mole fraction of the test sample was calculated via linear interpolation (bracketing method).

Date of analysis: 02-06-2010

#### A5. Complementary information on the cylinder

Please report the value of the pressure left in the cylinder before shipment to the BIPM:

Cylinder pressure: 87 bar

If any other component other than  $\text{NO}_2$ , nitrogen and oxygen was detected and/or quantified please report its mole fraction in the table below:

| Component | Mole fraction / nmol/mol | Expanded uncertainty | Coverage factor | Measurement technique |
|-----------|--------------------------|----------------------|-----------------|-----------------------|
|           |                          |                      |                 |                       |
|           |                          |                      |                 |                       |
|           |                          |                      |                 |                       |

## Dutch Metrology Institute (VSL)

### A1. General information

|                                    |   |     |                  |
|------------------------------------|---|-----|------------------|
| Institute                          | VSL   |     |                  |
| Address                            | Thijsseweg 11<br>2629 JA Delft<br>The Netherlands |     |                  |
| Contact person                     | G. Nieuwenkamp MSc                                |     |                  |
| Telephone                          | +31 15 26 91 682                                  | Fax | +31 15 26 12 971 |
| Email*                             | gnieuwenkamp@vsl.nl                               |     |                  |
| Serial number of cylinder received | APEX 930674                                       |     |                  |
| Cylinder pressure as received      | 10.7 MPa  |     |                  |

---

<sup>1</sup> Protocol A is fully described in the document: Pilot Study CCQM-P110 – Nitrogen dioxide in Nitrogen (10  $\mu\text{mol/mol}$ ) page 5.



| Date of measurement | Measurement technique | Nitrogen dioxide mole fraction<br>$x_{\text{NO}_2}$ / $\mu\text{mol/mol}$ | Expanded uncertainty<br>$U(x_{\text{NO}_2})$ / $\mu\text{mol/mol}$ | Coverage factor |
|---------------------|-----------------------|---|--|-----------------|
| 2009-12-17          | UV                    | 10.505  | 0.22   | 2               |
| 2009-12-18          | UV                    | 10.484  | 0.21   | 2               |
| 2009-12-24          | UV                    | 10.528  | 0.21   | 2               |
| 2009-12-31          | CLD                   | 10.524  | 0.20   | 2               |
| <b>average</b>      |                       | <b>10.51</b>  | <b>0.21</b>  | <b>2</b>        |

### A3. Uncertainty Budget

Please provide a complete uncertainty budget.

| Uncertainty source               | Relative standard deviation<br>$u(x_j)$                      | Coverage factor<br>k |
|----------------------------------|--|----------------------|
| Gravimetric standards            | 0.03 %   |                      |
| Stability related uncertainty    | 0.5 %  |                      |
| Monitor response                 | 0.21 %   |                      |
|                                  | <b>Relative Expanded uncertainty<br/><math>U(x_j)</math></b> |                      |
| Calculated with ISO6143 software | 2.0 %  | 2                    |

### A4. Description of the procedure used during the gas analysis

Please describe in detail the analytical method(s) used for gas analysis<sup>2</sup>.

The NO<sub>2</sub> content is analyzed by using an ABB LIMAS UV monitor. Calibration is performed with

$\mu\text{mol/mol}$ . A linear curve model has been applied. The cylinder is equipped with a stainless steel pressure regulator and flushed prior to use. One single pressure regulator is used for all cylinders, after the analysis the regulator is connected to the next cylinder. A flow of 500 ml/min, controlled by a Bronkhorst mass flow controller (low  $\Delta p$  type, with Kalrez seals), is led to the monitor. After 30 minutes of flushing, the average response over the next 15 minutes is recorded. Linear curve fitting is performed using ISO6143 software.

An additional measurement is performed with a Thermo 17C NO-NO<sub>2</sub>-NH<sub>3</sub> CLD analyser. The same reference materials and the same method have been used, only with a flow of 600 ml/min. The sample is led over a stainless steel convertor, operated at 750 °C. All nitrogen containing species are converted to NO and analysed as N<sub>t</sub> (= total nitrogen).

#### A5. Complementary information on the cylinder

Please report the value of the pressure left in the cylinder before shipment to the BIPM:

Cylinder pressure (2010-01-04) : 7.2 MPa

If any other component other than NO<sub>2</sub>, nitrogen and oxygen was detected and/or quantified please report its mole fraction in the table below:

| Component | Mole fraction /<br>nmol/mol | Expanded uncertainty | Coverage factor | Measurement<br>technique |
|-----------|-----------------------------|----------------------|-----------------|--------------------------|
|           |                             |                      |                 |                          |
|           |                             |                      |                 |                          |
|           |                             |                      |                 |                          |

## Bureau International des Poids et Mesures (BIPM)

### A1. General information

|                                    |  |                            |
|------------------------------------|--|----------------------------|
| Institute                          | Bureau International des Poids et Mesures (BIPM) |                            |
| Address                            | 92312 Sèvres Cedex, France.                      |                            |
| Contact person                     | Edgar Flores                                     |                            |
| Telephone                          | + 33 1 45 07 70 92<br>21                         | Fax :+ 33 1 45 34 20<br>21 |
| Email*                             | edgar.flores@bipm.org                            |                            |
| Serial number of cylinder received | 930697   |                            |
| Cylinder pressure as received      | 11MPa  |                            |

### A2. Results

The BIPM result is given in the following table:

| Nitrogen dioxide mole fraction<br>$x_{\text{NO}_2} / \mu\text{mol/mol}$ | Expanded uncertainty<br>$U(x_{\text{NO}_2}) / \mu\text{mol/mol}$ | Coverage factor |
|---|--|-----------------|
| 10.343  | 0.048  | 2               |

Note: In the version Draft A of this report erroneously the preliminary result 10.329  $\mu\text{mol/mol}$  was reported as the nitrogen dioxide mole fraction of the standard 930697 being the correct value, 10.343  $\mu\text{mol/mol}$ , used in all calculations of this report.

**A3. Uncertainty Budget**

Please provide a complete uncertainty budget.

The uncertainty budget of the BIPM-NO<sub>2</sub> facility is presented in section 3 of ANNEX 1.

**A4. Description of the procedure used during the gas analysis**

Please describe in detail the analytical method(s) used for gas analysis<sup>17</sup>.

The method used for the analysis of the cylinder was based on primary reference mixtures generated by the BIPM-NO<sub>2</sub> facility. The BIPM-NO<sub>2</sub> facility comprises a magnetic suspension balance, a flow control system for the dynamic generation of the gas mixtures and a flow control system for static nitrogen dioxide gas standards. Both, static and dynamic sources of NO<sub>2</sub> mixtures are ultimately connected to a continuous gas analyser ABB Limas 11 (AO2020), and to a FT-IR spectrometer. The operation and automation of the ensemble of instruments (NO<sub>2</sub> FT-IR facility-ABB Limas 11-FT-IR) is achieved through a LabView<sup>®</sup> programme. Through a graphical user interface the program facilitates the setting and monitoring of all relevant instrumental parameters, automated control of complex procedures, the recording of mass measurements and NO<sub>2</sub> analyser readings and related data to file and the graphical real-time display of many of the instrument readings.

Nitric acid was the main impurity in the nitrogen dioxide gas mixtures generated by the BIPM NO<sub>2</sub> facility and this was corrected by quantifying the mole fraction of nitric acid directly using FT-IR spectroscopy with traceability to line parameters within the HITRAN database. The determination of nitric acid was assessed using a (48 ±1.2) m multipath gas cell in the FT-IR system.

**A5. Complementary information on the cylinder**

Please report the value of the pressure left in the cylinder before shipment to the BIPM:

If any other component other than NO<sub>2</sub>, nitrogen and oxygen was detected and/or quantified please report its mole fraction in the table below:

| Component        | Mole fraction / nmol/mol | Expanded uncertainty nmol/mol | Coverage factor | Measurement technique |
|------------------|--------------------------|-------------------------------|-----------------|-----------------------|
| HNO <sub>3</sub> | 141                      | 85                            | 2               | FT-IR Spectroscopy    |
|                  |                          |                               |                 |                       |
|                  |                          |                               |                 |                       |

<sup>17</sup> The choice of the procedure used for gas analysis is the responsibility of the participating laboratory. Nevertheless, for a proper evaluation of the data, it is necessary that the calibration method, as well as the way in which the calibration mixtures have been prepared is reported to the co-ordinators.

## Bibliography

- (1) AIJaZM C. Probert: An intercomparison of low flow gas facilities at eleven European laboratories using a Molbloc transfer package: EURAMET.MM.FF-S3 (EURAMET Project no 806) *Metrologia* 45, Tec. Sup. (2008).
- (2) DWT Griffith: Synthetic calibration and quantitative analysis of gas phase FTIR spectra. *Applied spectroscopy* 50 (1996) 59-70
- (3) HLR E. E. Hughes, E. R. Deardorff, J. H. Margeson, R. G. Fuerst: Performance of a NO<sub>2</sub> permeation device. *Analytical chemistry* 49 (12) (1977) 1823–29.
- (4) DJ L.S. Rothman, A. Barbe, D. Chris Benner, M. Birk, L.R. Brown, M.R. Carleer, C. Chackerian Jr., K. Chancea, L.H. Couderth, V. Danai, V.M. Devic, J.-M. Flaudh,2, R.R. Gamachej, A. Goldmank, J.-M. Hartmannh,2, K.W. JucksI, A.G. Makim, J. Y. Mandini, S.T. Massien, J. Orphalh,2, A. Perrinh,2, C.P. Rinslando, M.A.H. Smitho, J. Tennysonp, R.N. Tolchenovp, R.A. Tothe, J. Vander Auweraf, P. Varanasiq, G. Wagnerd: The HITRAN 2004 molecular spectroscopic database. *Journal of Quantitative Spectroscopy & Radiative Transfer* 96 (2005) 139–204.
- (5) KaAGM Robert H: Infrared absorption intensities for N<sub>2</sub>O<sub>3</sub> *Journal of Quantitative Spectroscopy and Radiative Transfer* 31 (1984) 173-76
- (6) JBaMP D. Newnham: Infrared absorbance cross-sections for dinitrogen pentoxide vapour *Journal of Quantitative Spectroscopy and Radiative Transfer* Volume 50 (1993) 571-77
- (7) LMW William S. Barney, Matthew J. Lakin, Theo Brauers, Jochen Stutz,§ and Barbara J. Finlayson-Pitts: Infrared Absorption Cross-Section Measurements for Nitrous Acid (HONO) at Room Temperature. *The Journal of Physical Chemistry* 104 (8) (2000) 1692–99.
- (8) RDMaCEM Svante Höjer: Intensities of the 1397 cm<sup>-1</sup> (ν<sub>3</sub>) band of HO<sub>2</sub>NO<sub>2</sub> and feasibility of atmospheric detection *Journal of Quantitative Spectroscopy and Radiative Transfer* 55 (1999) 273-78.
- (9) MHaJVA D. Hurtmans: Integrated band intensities in N<sub>2</sub>O<sub>4</sub> in the infrared range *Journal of Quantitative Spectroscopy and Radiative Transfer* 50 (1993) 595-602